



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

JUNE 2023

MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

July 31, 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01 BERTHA GANTER - FORT MCKAY JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	June 1, 2023	Last Cal Date:	May 1, 2023
Start time (MST):	10:00	End time (MST):	13:46
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.19	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC486642			
Removed Cal Gas Conc:	49.19	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000343	0.999958	Backgd or Offset:	19.4	19.1
Calibration intercept:	0.027099	0.086959	Coeff or Slope:	0.893	0.883

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4918	81.3	799.9	806.9	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.5	----
high point	4918	81.3	799.9	800.5	0.999
second point	4959	40.7	400.4	399.4	1.003
third point	4979	20.3	199.7	200.0	0.999
as left zero	5000	0.0	0.0	0.3	----
as left span	4918	81.3	799.9	800.2	1.000
Average Correction Factor					1.000

Baseline Corr As found:	806.70	Previous response	800.24	*% change	0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 1, 2023	Previous Calibration:	May 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:00	End Time (MST):	13:46
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999995	≥0.995
799.9	800.5	0.9993			
400.4	399.4	1.0026	Slope	0.999958	0.90 - 1.10
199.7	200.0	0.9987			
			Intercept	0.086959	+/-30



SO2 Calibration Plot

Date: June 1, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: June 27, 2023 Last Cal Date: May 16, 2023
 Start time (MST): 8:59 End time (MST): 13:45
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153461
 Converter make: CD Nova Converter serial #: 470
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993792	1.002650	Backgd or Offset: 2.27	2.30
Calibration intercept:	0.220000	0.159995	Coeff or Slope: 0.919	0.919

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4921	78.4	80.0	79.4	1.009
as found 2nd point	4960	39.2	40.0	40.0	1.002
as found 3rd point	4980	19.6	20.0	20.0	1.005
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	78.4	80.0	80.3	0.996
second point	4960	39.2	40.0	40.5	0.988
third point	4980	19.6	20.0	20.0	1.000
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	78.4	80.0	79.9	1.001
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	0.995
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.3 Prev response: 79.72 *% change: -0.5%
 Baseline Corr 2nd AF pt: 39.9 AF Slope: 0.991363 AF Intercept: 0.179997
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999988

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 16, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:59	End Time (MST):	13:45
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999970	≥0.995
80.0	80.3	0.9962			
40.0	40.5	0.9876	Slope	1.002650	0.90 - 1.10
20.0	20.0	0.9999			
			Intercept	0.159995	+/-3



TRS Calibration Plot

Date: June 27, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: June 27, 2023 Last Cal Date: May 16, 2023
 Start time (MST): 8:59 End time (MST): 11:00
 Reason: As Found

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1200326167
 Converter make: Thermo Converter Converter serial #: N/A
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001375		Backgd or Offset:	1.92	
Calibration intercept:	0.081613		Coeff or Slope:	1.001	

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4921	78.4	80.0	81.4	0.980
as found 2nd point	4960	39.2	40.0	40.6	0.980
as found 3rd point	4980	19.6	20.0	20.0	0.990
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

SO₂ Scrubber Check

Date of last scrubber change:	March 21, 2022	Ave Corr Factor	
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 81.6 Prev response: 80.17 *% change: 1.8%
 Baseline Corr 2nd AF pt: 40.8 AF Slope: 1.021137 AF Intercept: -0.280002
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999993

* = > +/-5% change initiates investigation


Notes: Inlet filter change completed after as founds. Replaced the kicker assembly.

Calibration Performed By: Rene Chamberland

H₂S Calibration Plot

Date: June 27, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	June 28, 2023	Last Cal Date:	May 16, 2023
Start time (MST):	9:10	End time (MST):	13:06
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.10	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	<u>CC511749</u>			
Removed Cal Gas Conc:	5.10	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	<u>N/A</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167
Converter make:	Thermo Converter	Converter serial #:	N/A
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001375	0.997140	Backgd or Offset:	1.92	1.91
Calibration intercept:	0.081613	0.156826	Coeff or Slope:	1.001	0.992

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4922	78.4	80.0	80.0	1.000
second point	4960	39.2	40.0	39.9	1.002
third point	4980	19.6	20.0	20.0	1.000
as left zero	5000	0.0	0.0	0.5	----
as left span	4922	78.4	80.0	78.3	1.021
SO2 Scrubber Check	4919	81.3	813.0		----
Date of last scrubber change:	March 21, 2022			Ave Corr Factor	1.001
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: Kicker assembly was replaced yesterday. Scrubber check completed after zero. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	May 16, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:10	End Time (MST):	13:06
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999983	
80.0	80.0	0.9997			≥0.995
40.0	39.9	1.0025	Slope	0.997140	
20.0	20.0	0.9999			0.90 - 1.10
			Intercept	0.156826	+/-3



H₂S Calibration Plot

Date: June 28, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	June 1, 2023	Last Cal Date:	May 1, 2023
Start time (MST):	10:00	End time (MST):	13:46
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC486642	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	497.7 ppm	CH ₄ Equiv Conc.	1063.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.6 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.7 ppm	CH ₄ Equiv Conc.	1063.1 ppm
Removed C ₃ H ₈ Conc.	205.6 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	5609

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.00E-04	2.89E-04	NMHC SP Ratio:	6.36E-05
CH ₄ Retention time:	14.6	14.6	NMHC Peak Area:	144534
				147590

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.29	17.88	0.967
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	17.29	17.25	1.002
second point	4959	40.7	8.65	8.56	1.011
third point	4980	20.3	4.32	4.26	1.013
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.29	17.36	0.996

Average Correction Factor				1.009
Baseline Corr AF:	17.88	Prev response	17.28	*% change 3.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4918	81.3	9.19	9.42	0.976
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4918	81.3	9.19	9.17	1.002
second point	4959	40.7	4.60	4.59	1.003
third point	4980	20.3	2.30	2.29	1.001
as left zero	5000	0	0.00	0.00	----
as left span	4918	81.3	9.19	9.25	0.994
Average Correction Factor					1.002
Baseline Corr AF:	9.42	Prev response	9.20	*% change	2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	8.46	0.957
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	8.09	8.08	1.002
second point	4959	40.7	4.05	3.97	1.019
third point	4980	20.3	2.02	1.97	1.027
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.11	0.999
Average Correction Factor					1.016
Baseline Corr AF:	8.46	Prev response	8.08	*% change	4.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998379	0.998392
THC Cal Offset:	0.020702	-0.033901
CH ₄ Cal Slope:	0.999517	0.999433
CH ₄ Cal Offset:	-0.010960	-0.034161
NMHC Cal Slope:	0.997203	0.997502
NMHC Cal Offset:	0.031861	0.000659

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 1, 2023	Previous Calibration:	May 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:00	End Time (MST):	13:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999975	≥ 0.995			
17.29	17.25	1.0021						
8.65	8.56	1.0108				Slope	0.998392	0.90 - 1.10
4.32	4.26	1.0132						
			Intercept	-0.033901	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 1, 2023	Previous Calibration:	May 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:00	End Time (MST):	13:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999894	≥ 0.995
8.09	8.08	1.0017			
4.05	3.97	1.0194			
2.02	1.97	1.0273			
			Slope	0.999433	0.90 - 1.10
			Intercept	-0.034161	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 1, 2023	Previous Calibration:	May 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:00	End Time (MST):	13:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995
9.19	9.17	1.0024			
4.60	4.59	1.0029			
2.30	2.29	1.0011			
			Slope	0.997502	0.90 - 1.10
			Intercept	0.000659	+/-0.5



NMHC Calibration Plot

Date: June 1, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
Calibration Date: June 15, 2023 Last Cal Date: May 18, 2023
Start time (MST): 9:03 End time (MST): 13:39
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1P9L Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.84 ppm NO Cal Gas Conc: 50.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.84 ppm Removed Gas NO Conc: 50.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3565
ZAG make/model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1218153357
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.465	1.474	NO bkgnd or offset:	7.4	7.5
NOX coeff or slope:	0.991	0.990	NOX bkgnd or offset:	7.6	7.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	198.1	198.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999073	0.999157
NO _x Cal Offset:	0.200000	0.120000
NO Cal Slope:	0.999971	1.001013
NO Cal Offset:	-0.620000	-1.060000
NO ₂ Cal Slope:	1.001305	1.000108
NO ₂ Cal Offset:	0.605886	0.179968



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.1	----	----
as found span	4920	80.0	813.4	800.6	12.8	809.2	793.0	16.1	1.0052	1.0096
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
high point	4920	80.0	813.4	800.6	12.8	812.8	800.8	12.0	1.0008	0.9998
second point	4960	40.0	406.7	400.3	6.4	406.8	399.5	7.3	0.9998	1.0021
third point	4980	20.0	203.4	200.2	3.2	203.0	198.0	5.0	1.0018	1.0109
as left zero	5000	0.0	0.0	0.0	0.0	0.2	-0.1	0.3	----	----
as left span	4920	80.0	813.4	397.7	415.7	813.6	398.5	415.1	0.9998	0.9981
Average Correction Factor									1.0008	1.0043

Corrected As found	NO _x = 809.2 ppb	NO = 793.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.5%	
Previous Response	NO _x = 812.9 ppb	NO = 800.0 ppb		*Percent Change	NO = -0.9%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.8	395.9	415.7	415.9	0.9995	100.0%
2nd GPT point (200 ppb O3)	798.8	598.6	213.0	213.3	0.9986	100.1%
3rd GPT point (100 ppb O3)	798.8	699.9	111.7	111.8	0.9991	100.1%
Average Correction Factor					0.9991	100.1%

Notes:

Changed the inlet filter after as founds. Adjusted span.

Calibration Performed By:

Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:03	End Time (MST):	13:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
813.4	812.8	1.0008			
406.7	406.8	0.9998			
203.4	203.0	1.0018			
			Slope	0.999157	0.90 - 1.10
			Intercept	0.120000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:03	End Time (MST):	13:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 $0.90 - 1.10$ ± 20
800.6	800.8	0.9998		
400.3	399.5	1.0021		
200.2	198.0	1.0109		
			0.999991	
			1.001013	
			-1.060000	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:03	End Time (MST):	13:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
415.7	415.9	0.9995		
213.0	213.3	0.9986		
111.7	111.8	0.9991		
			1.000000	
			1.000108	
			0.179968	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: June 15, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
Calibration Date: June 19, 2023 Last Cal Date: June 15, 2023
Start time (MST): 9:42 End time (MST): 12:45
Reason: As Found

Calibration Standards

NO Gas Cylinder #: T2Y1P9L Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.84 ppm NO Cal Gas Conc: 50.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.84 ppm Removed Gas NO Conc: 50.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3565
ZAG make/model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1218153357
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.474	1.474	NO bkgnd or offset:	7.5	7.5
NOX coeff or slope:	0.990	0.990	NOX bkgnd or offset:	7.7	7.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	215.5	185.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999157	
NO _x Cal Offset:	0.120000	
NO Cal Slope:	1.001013	
NO Cal Offset:	-1.060000	
NO ₂ Cal Slope:	1.000108	
NO ₂ Cal Offset:	0.179968	



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
as found span	4920	80.0	813.4	800.6	12.8	719.2	705.5	13.7	1.1310	1.1349
as found 2nd	4960	40.0	406.7	400.3	6.4	358.9	351.3	7.6	1.1332	1.1395
as found 3rd	4980	20.0	203.4	200.2	3.2	179.0	174.3	4.7	1.1361	1.1484
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

Average Correction Factor

Corrected As found	NO _x = 719.5 ppb	NO = 705.7 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -13.0%
Previous Response	NO _x = 812.9 ppb	NO = 800.4 ppb			*Percent Change	NO = -13.4%
Baseline Corr 2nd pt	NO _x = 359.2 ppb	NO = 351.5 ppb	As found	NO _x r ² : 0.999999	Nx SI: 0.884680	Nx Int: -0.640
Baseline Corr 3rd pt	NO _x = 179.3 ppb	NO = 174.5 ppb	As found	NO r ² : 0.999990	NO SI: 0.882052	NO Int: -1.240
			As found	NO ₂ r ² : 0.999996	NO ₂ SI: 0.998547	NO ₂ Int: -0.386

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	-0.1	----	----
as found GPT point (400 ppb NO ₂)	704.2	349.4	367.6	366.9	1.0019	99.8%
as found GPT point (200 ppb NO ₂)	704.2	527.6	189.4	188.4	1.0053	99.5%
as found GPT point (100 ppb NO ₂)	704.2	616.6	100.4	99.7	1.0070	99.3%
1st GPT point (400 ppb O ₃)						
2nd GPT point (200 ppb O ₃)						
3rd GPT point (100 ppb O ₃)						

Average Correction Factor

Notes:

Performing multi point as founds. Replacing the sample pump.


Calibration Performed By:

Rene Chamberland

NO_x Calibration Plot

Date: June 19, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
Calibration Date: June 20, 2023 Last Cal Date: June 15, 2023
Start time (MST): 9:34 End time (MST): 13:44
Reason: Maintenance

Calibration Standards

NO Gas Cylinder #: T2Y1P9L Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.84 ppm NO Cal Gas Conc: 50.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.84 ppm Removed Gas NO Conc: 50.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3565
ZAG make/model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1218153357
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:		1.334	NO bkgnd or offset:		6.6
NOX coeff or slope:		0.988	NOX bkgnd or offset:		6.8
NO2 coeff or slope:		1.000	Reaction cell Press:		186.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:		1.000702
NO _x Cal Offset:		-0.180000
NO Cal Slope:		1.001256
NO Cal Offset:		-0.920000
NO ₂ Cal Slope:		0.998452
NO ₂ Cal Offset:		0.436719



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4920	80.0	813.4	800.6	12.8	814.0	801.1	12.9	0.9993	0.9994
second point	4960	40.0	406.7	400.3	6.4	406.5	399.7	6.8	1.0005	1.0016
third point	4980	20.0	203.4	200.2	3.2	203.3	198.4	5.0	1.0003	1.0089
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
as left span	4920	80.0	813.4	398.6	414.8	813.5	398.1	415.4	0.9999	1.0014
Average Correction Factor									1.0000	1.0033

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.2	397.2	414.8	414.3	1.0012	99.9%
2nd GPT point (200 ppb O3)	799.2	597.9	214.1	214.6	0.9977	100.2%
3rd GPT point (100 ppb O3)	799.2	699.9	112.1	112.7	0.9947	100.5%
Average Correction Factor					0.9979	100.2%

Notes: Performing maintenance calibration after pump swap. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 20, 2023	Previous Calibration:	June 15, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:34	End Time (MST):	13:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
813.4	814.0	0.9993			
406.7	406.5	1.0005			
203.4	203.3	1.0003			
			Slope	1.000702	0.90 - 1.10
			Intercept	-0.180000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 20, 2023	Previous Calibration:	June 15, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:34	End Time (MST):	13:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 $0.90 - 1.10$ ± 20
800.6	801.1	0.9994		
400.3	399.7	1.0016		
200.2	198.4	1.0089		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 20, 2023	Previous Calibration:	June 15, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:34	End Time (MST):	13:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
414.8	414.3	1.0012		
214.1	214.6	0.9977		
112.1	112.7	0.9947		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: June 20, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: June 2, 2023 Last Cal Date: May 4, 2023
 Start time (MST): 9:37 End time (MST): 12:36
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 1107
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999429	1.001229	Backgd or Offset:	3.2	3.2
Calibration intercept:	0.500000	0.260000	Coeff or Slope:	1.010	1.010

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	5000	855.5	400.0	400.3	0.999
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.1	----
high point	5000	858.8	400.0	400.6	0.999
second point	5000	743.1	200.0	200.8	0.996
third point	5000	653.5	100.0	100.4	0.996
as left zero	5000	0.0	0.0	0.2	----
as left span	5000	858.8	400.0	404.8	0.988
Average Correction Factor					0.997

Baseline Corr As found:	400.2	Previous response	400.3	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 2, 2023	Previous Calibration:	May 4, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:37	End Time (MST):	12:36
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
400.0	400.6	0.9985		
200.0	200.8	0.9960	Slope	0.90 - 1.10
100.0	100.4	0.9960		
			Intercept	+/- 5



O₃ Calibration Plot

Date: June 2, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
 Calibration Date: June 2, 2023 Last Cal Date: May 15, 2023
 Start time (MST): 11:38 End time (MST): 12:07

Analyzer Make: API T640 S/N: 326
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	24.6	23.5	24.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	736.1	738	736.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	5.09	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: June 2, 2023		Last Cal Date: May 15, 2023		
	PM w/o HEPA: 13.6		PM w/ HEPA: 0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		_____			<0.2 ug/m3
Disposable Filter Changed:		_____			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: Removing the T640 instrument. Flow, temperature, and pressure all within limits. Leak check passed. PMT peak test did not display any results.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
 Calibration Date: June 2, 2023 Last Cal Date: May 15, 2023
 Start time (MST): 12:47 End time (MST): 13:29

Analyzer Make: API T640 S/N: 324
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	25.8	24.7	25.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	738.8	738	738.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.01	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: June 2, 2023		Last Cal Date:		
	PM w/o HEPA: 206		PM w/ HEPA: 0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.9		10.9	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		_____			<0.2 ug/m3
Disposable Filter Changed:		_____			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: Installing a new T640 instrument. Flow, temperature, and pressure all within limits. Leak check passed. PMT peak test within limits. Inlet head cleaned.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	June 26, 2023	Last Cal Date:	May 11, 2023
Start time (MST):	10:29	End time (MST):	13:41
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	<u>3040</u>	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>3040</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999480	1.001481	Backgd or Offset:	-0.012	-0.012
Calibration intercept:	0.161861	0.163854	Coeff or Slope:	0.992	0.992

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.6	40.8	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.6	40.6	0.998
second point	4966	33.3	20.2	20.8	0.976
third point	4983	16.7	10.2	10.3	0.984
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.5	40.2	1.008
Average Correction Factor					0.986

Baseline Corr As found:	40.70	Prev response:	40.70	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 26, 2023	Previous Calibration:	May 11, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:29	End Time (MST):	13:41
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999847	≥0.995
40.6	40.6	0.9982			
20.2	20.8	0.9759	Slope	1.001481	0.90 - 1.10
10.2	10.3	0.9839			
			Intercept	0.163854	+/-1.5



CO Calibration Plot

Date: June 26, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	June 12, 2023	Last Cal Date:	May 12, 2023
Start time (MST):	9:43	End time (MST):	13:22
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>60,200</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
N2 Gen Make/Model:	Peak Sci NG5000		Serial Number:	7220900034

Analyzer Information

Analyzer make: Teledyne API 360	Analyzer serial #: 442
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002095	1.001960	Backgd or Offset:	0.037	0.045
Calibration intercept:	-6.580000	-4.460000	Coeff or Slope:	0.880	0.874

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-0.8	----
as found span	2920	80.0	1605.3	1604.6	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	-0.1	----
high point	2920	80.0	1605.3	1606.1	1.000
second point	2960	40.0	802.7	797.9	1.006
third point	2980	20.0	401.3	393.1	1.021
as left zero	3000	0.0	0.0	-0.1	----
as left span	2960	40.0	802.7	784.7	1.023
Average Correction Factor					1.009

Baseline Corr As found:	1605.40	Prev response:	1602.12	*% change:	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	June 12, 2023	Previous Calibration	May 12, 2023
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS01
Start Time (MST)	9:43	End Time (MST)	13:22
Analyzer make	Teledyne API 360	Analyzer serial #	442

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999966	≥0.995
1605.3	1606.1	0.9995			
802.7	797.9	1.0060	Slope	1.001960	0.90 - 1.10
401.3	393.1	1.0209			
			Intercept	-4.460000	+/-10



CO₂ Calibration Plot

Date: June 12, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	June 21, 2023	Last Cal Date:	May 30, 2023
Start time (MST):	9:10	End time (MST):	14:30
NH3 Cal Date:	June 22, 2023	Last Cal Date:	May 30, 2023
Start time (MST):	9:15	End time (MST):	12:59
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	50.84	ppm	NO Gas Cylinder #:	T2Y1P9L
NO Cal Gas Conc:	50.04	ppm	NO Cal Gas Expiry:	March 3, 2028
Removed NOX Conc:	50.84	ppm	Removed Cylinder #:	NA
Removed NO Conc:	50.04	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	72.93	ppm	NH3 Gas Cylinder #:	CC281298
			NH3 Cal Gas Expiry:	February 28, 2023
Removed NH3 Conc:	72.93	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	Teledyne API T700		Serial Number:	3565
ZAG make/model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	475
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	5.00
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	512

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.986	0.989	TN coefficient:	0.984	0.995
NOX coefficient:	0.986	0.991	NO bkgnd:	-3.0	-0.5
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-2.5	-0.3
NH3 coefficient:	0.941	0.941	TN bkgnd:	1.4	0.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998848	0.998174
NO _x Cal Offset:	-0.520000	-1.280000
NO Cal Slope:	0.998730	0.999572
NO Cal Offset:	-1.160000	-2.180000
NO ₂ Cal Slope:	0.999748	1.004664
NO ₂ Cal Offset:	0.471532	0.067495
NH3 Cal Slope:	0.998065	0.998759
NH3 Cal Offset:	-0.656397	1.195770
TN Cal Slope:	1.001095	1.001474
TN Cal Offset:	-0.131851	1.255511



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.7	0.0	----	----
as found NO	4920	80.0	813.4	813.4	----	814.0	813.8	0.3	0.999	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high NO point	4920	80.0	813.4	813.4	----	812.0	811.3	0.4	1.002	----
NO/O3 point	4920	80.0	813.4	813.4	----	813.0	812.0	0.8	1.001	----
as found NH3	3413	86.4	1800.6	----	1800.6	1806.3	----	1801.5	0.997	1.000
new NH3 cyl rp							----			
first NH3	3413	86.4	1800.6	----	1800.6	1806.3	----	1801.5	0.997	1.000
second NH3	3452	48.0	1000.2	----	1000.2	997.0	----	994.0	1.003	1.006
third NH3	3476	24.0	500.1	----	500.1	507.6	----	506.1	0.985	0.988
Average Correction Factor									1.0012	0.9980

Corrected As found TN = 814.7 ppb NO_x = 814.5 ppb NH3 = 1801.5 ppb

Previous Response TN = 814.2 ppb NO_x = 812.0 ppb NH3 = 1796.5 ppb

NH3 Previous Converter Efficiency = 94.1%

NH3 Current Converter Efficiency = 94.1%

*Percent Change TN = 0.1%

*Percent Change NO_x = 0.3%

*Percent Change NH3 = 0.3%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.7	0.8	-0.4	----	----
as found span	4920	80.0	813.4	800.6	813.4	812.0	796.1	804.9	1.0018	1.0057
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	-0.1	----	----
high point	4920	80.0	813.4	800.6	813.4	811.3	799.4	812.0	1.0026	1.0016
second point	4960	40.0	406.7	400.3	406.7	404.0	396.2	404.8	1.0067	1.0104
third point	4980	20.0	203.4	200.2	203.4	200.6	196.3	201.3	1.0138	1.0197
Average Correction Factor									1.0077	1.0105

Baseline Corr As fnd	TN = 805.3 ppb	NO _x = 811.3 ppb	NO = 795.3 ppb	*Percent Change	TN = -1.1%
Previous Response	TN = 814.2 ppb	NO _x = 812.0 ppb	NO = 798.5 ppb	*Percent Change	NO _x = -0.1%
				*Percent Change	NO = -0.4%

* = > +/-5% change initiates investigation

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	0.0	----	----
calibration zero	----	----	0.0	0.0	----	----
1st GPT point (400 ppb O3)	794.9	398.9	408.8	410.6	0.9956	100.4%
2nd GPT point (200 ppb O3)	794.9	594.0	213.7	215.2	0.9930	100.7%
3rd GPT point (100 ppb O3)	794.9	694.8	112.9	113.3	0.9965	100.4%
Average Correction Factor					0.9950	100.5%

Notes: Changed the inlet filter after as founds. Adjusted the NO_x/TN zero and span. Used the 2nd GPT reference point due to drift.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TN Calibration Summary


Version-05-2023

Station Information

Calibration Date:	June 22, 2023	Previous Calibration:	May 30, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:10	End Time (MST):	14:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
1800.6	1806.3	0.9969			
1000.2	997.0	1.0032			
500.1	507.6	0.9852			
			Slope	1.001474	0.90 - 1.10
			Intercept	1.255511	+/-20





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	June 22, 2023	Previous Calibration:	May 30, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:10	End Time (MST):	14:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
1800.6	1801.5	0.9995			
1000.2	994.0	1.0062			
500.1	506.1	0.9881			
			Slope	0.998759	0.90 - 1.10
			Intercept	1.195770	+/-20





Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-05-2023


Station Information

Calibration Date:	June 21, 2023	Previous Calibration:	May 30, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:10	End Time (MST):	14:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.4	811.3	1.0026		
406.7	404.0	1.0067		
203.4	200.6	1.0138		
			0.999990	
			0.998174	
			-1.280000	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	June 21, 2023	Previous Calibration:	May 30, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:10	End Time (MST):	14:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
800.6	799.4	1.0016			
400.3	396.2	1.0104			
200.2	196.3	1.0197			
			Slope	0.999572	0.90 - 1.10
			Intercept	-2.180000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	June 21, 2023	Previous Calibration:	May 30, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:10	End Time (MST):	14:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
408.8	410.6	0.9956			0.999997
213.7	215.2	0.9930			1.004664
112.9	113.3	0.9965			0.067495



NO_x Calibration Plot

Date: June 21, 2023


Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: June 22, 2023

Location: Bertha Ganter-Fort McKay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	June 29, 2023	Last Cal Date:	May 11, 2023
Start time (MST):	10:38	End time (MST):	16:04
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T750		Serial Number:	282
ZAG Make/Model:	API T751		Serial Number:	321

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000972	0.999231	Backgd or Offset:	17.9	17.8
Calibration intercept:	-1.065310	-0.605160	Coeff or Slope:	0.793	0.783

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.2	801.6	813.0	0.986
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.2	801.6	801.0	1.001
second point	4960	40.1	400.8	398.9	1.005
third point	4980	20.0	199.9	198.9	1.005
as left zero	5000	0.0	0.0	-0.2	----
as left span	4920	80.2	801.6	785.0	1.021
Average Correction Factor					1.004

Baseline Corr As found:	813.00	Previous response	801.36	*% change	1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 29, 2023	Previous Calibration:	May 11, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:38	End Time (MST):	16:04
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999994	≥0.995
801.6	801.0	1.0008			
400.8	398.9	1.0048	Slope	0.999231	0.90 - 1.10
199.9	198.9	1.0051			
			Intercept	-0.605160	+/-30



SO2 Calibration Plot

Date: June 29, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mildred Lake Station number: AMS02
 Calibration Date: June 29, 2023 Last Cal Date: May 3, 2023
 Start time (MST): 10:00 End time (MST): 15:39
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.29 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345191
 Removed Cal Gas Conc: 5.29 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1185
 ZAG Make/Model: API T701 Serial Number: 5608

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 12113311966
 Converter make: Global G150 Converter serial #: 2022-198
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006965	1.006394	Backgd or Offset:	1.80	1.72
Calibration intercept:	-0.239199	-0.119196	Coeff or Slope:	0.833	0.797

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4924	75.6	80.0	83.4	0.959
as found 2nd point	4962	37.8	40.0	41.7	0.959
as found 3rd point	4981	18.9	20.0	20.4	0.980
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4924	75.6	80.0	80.4	0.995
second point	4962	37.8	40.0	40.2	0.995
third point	4981	18.9	20.0	19.8	1.010
as left zero	5000	0.0	0.0	0.0	----
as left span	4924	75.6	80.0	80.6	0.992
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:	12-Sep-22			Ave Corr Factor	1.000
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 83.4 Prev response: 80.31 *% change: 3.7%
 Baseline Corr 2nd AF pt: 41.7 AF Slope: 1.044541 AF Intercept: -0.179166
 Baseline Corr 3rd AF pt: 20.4 AF Correlation: 0.999964

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after MPAF's. Scrubber check passed after cal zero. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 29, 2023	Previous Calibration:	May 3, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:00	End Time (MST):	15:39
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	≥0.995
80.0	80.4	0.9949		
40.0	40.2	0.9949	Slope	0.90 - 1.10
20.0	19.8	1.0099		
			Intercept	+/-3



H₂S Calibration Plot

Date: June 29, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	June 29, 2023	Last Cal Date:	May 30, 2023
Start time (MST):	10:38	End time (MST):	16:04
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
C ₃ H ₈ Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
Removed C ₃ H ₈ Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T750	Serial Number:	282
ZAG make/model:	Teledyne API T751	Serial Number:	321

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.90E-04	2.88E-04	NMHC SP Ratio:	4.54E-04
CH ₄ Retention time:	14.6	14.6	NMHC Peak Area:	194026
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				4.52E-05
				194709
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	17.28	0.973
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.85	0.998
second point	4960	40.1	8.41	8.40	1.001
third point	4980	20.0	4.19	4.22	0.994
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.48	1.021
Average Correction Factor					0.998

Baseline Corr AF:	17.28	Prev response	16.82	*% change	2.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	9.06	0.971
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.83	0.996
second point	4960	40.1	4.40	4.43	0.993
third point	4980	20.0	2.19	2.24	0.981
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.70	1.011
Average Correction Factor					0.990
Baseline Corr AF:	9.06	Prev response	8.81	*% change	2.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	8.22	0.976
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.02	1.000
second point	4960	40.1	4.01	3.98	1.009
third point	4980	20.0	2.00	1.98	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	7.78	1.031
Average Correction Factor					1.006
Baseline Corr AF:	8.22	Prev response	8.01	*% change	2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000855	1.001404
THC Cal Offset:	-0.012905	0.001307
CH ₄ Cal Slope:	1.001225	0.999914
CH ₄ Cal Offset:	-0.018850	-0.014249
NMHC Cal Slope:	1.000634	1.002789
NMHC Cal Offset:	0.005745	0.015957

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	June 29, 2023	Previous Calibration:	May 30, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:38	End Time (MST):	16:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995			
16.82	16.85	0.9981						
8.41	8.40	1.0011				Slope	1.001404	0.90 - 1.10
4.19	4.22	0.9944						
			Intercept	0.001307	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	June 29, 2023	Previous Calibration:	May 30, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:38	End Time (MST):	16:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999976	≥ 0.995			
8.02	8.02	1.0004						
4.01	3.98	1.0092				Slope	0.999914	0.90 - 1.10
2.00	1.98	1.0095						
			Intercept	-0.014249	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	June 29, 2023	Previous Calibration:	May 30, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:38	End Time (MST):	16:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999983	≥ 0.995			
8.80	8.83	0.9961						
4.40	4.43	0.9934				Slope	1.002789	0.90 - 1.10
2.19	2.24	0.9809						
			Intercept	0.015957	± 0.5			



NMHC Calibration Plot

Date: June 29, 2023

Location: Mildred Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	June 9, 2023	Last Cal Date:	May 4, 2023
Start time (MST):	6:27	End time (MST):	9:13
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.02	ppm	Cal Gas Exp Date:	September 9, 2028
Cal Gas Cylinder #:	CC470284			
Removed Cal Gas Conc:	50.02	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701		Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003456	1.001756	Backgd or Offset:	22.1	22.1
Calibration intercept:	0.000000	0.520000	Coeff or Slope:	0.869	0.860

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4920	80.0	800.3	810.7	0.987
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	80.0	800.3	801.4	0.999
second point	4960	40.0	400.2	403.3	0.992
third point	4980	20.0	200.1	200.5	0.998
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	80.0	800.3	800.4	1.000
Average Correction Factor					0.996

Baseline Corr As found:	810.80	Previous response	803.09	*% change	1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No Maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 9, 2023	Previous Calibration:	May 4, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:27	End Time (MST):	9:13
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999986	≥0.995
800.3	801.4	0.9987			
400.2	403.3	0.9922	Slope	1.001756	0.90 - 1.10
200.1	200.5	0.9979			
			Intercept	0.520000	+/-30



SO2 Calibration Plot

Date: June 9, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: June 22, 2023 Last Cal Date: May 16, 2023
 Start time (MST): 6:05 End time (MST): 10:35
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.42 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345266
 Removed Cal Gas Conc: 5.42 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701H Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1008841400
 Converter make: Global Converter serial #: 2022-200
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003759	1.001341	Backgd or Offset: 2.0	1.8
Calibration intercept:	-0.177711	-0.017797	Coeff or Slope: 1.067	1.095

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4926	74.1	80.3	78.6	1.018
as found 2nd point	4963	37.0	40.1	39.4	1.010
as found 3rd point	4982	18.5	20.1	19.4	1.018
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4926	74.1	80.3	80.4	0.999
second point	4963	37.0	40.1	40.2	0.998
third point	4982	18.5	20.1	20.0	1.003
as left zero	5000	0.0	0.0	0.0	----
as left span	4926	74.1	80.3	80.0	1.004
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	1.000
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 78.9 Prev response: 80.45 *% change: -2.0%
 Baseline Corr 2nd AF pt: 39.7 AF Slope: 0.982700 AF Intercept: -0.238123
 Baseline Corr 3rd AF pt: 19.7 AF Correlation: 0.999980

* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrator zero. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 22, 2023	Previous Calibration:	May 16, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:05	End Time (MST):	10:35
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999998	
80.3	80.4	0.9990			≥0.995
40.1	40.2	0.9977	Slope	1.001341	
20.1	20.0	1.0026			0.90 - 1.10
			Intercept	-0.017797	+/-3



H₂S Calibration Plot

Date: June 22, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	June 9, 2023	Last Cal Date:	May 4, 2023
Start time (MST):	6:27	End time (MST):	9:12
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC470284	Cal Gas Expiry Date:	September 9, 2028
CH ₄ Cal Gas Conc.	497.8 ppm	CH ₄ Equiv Conc.	1062.9 ppm
C ₃ H ₈ Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.8 ppm	CH ₄ Equiv Conc.	1062.9 ppm
Removed C ₃ H ₈ Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	1.860E-04	1.860E-04	NMHC SP Ratio:	3.870E-05
CH ₄ Retention time:	11.8	11.8	NMHC Peak Area:	233712
				233712

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.01	17.08	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.01	17.10	0.995
second point	4960	40.0	8.50	8.53	0.997
third point	4980	20.0	4.25	4.25	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.01	17.12	0.993

Average Correction Factor				0.997
Baseline Corr AF:	17.08	Prev response	16.97	*% change 0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	
as found span	4920	80.0	9.04	9.04	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.04	9.05	0.999
second point	4960	40.0	4.52	4.51	1.002
third point	4980	20.0	2.26	2.24	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.04	9.06	0.998
Average Correction Factor					1.004
Baseline Corr AF:	9.04	Prev response	9.04	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	7.96	8.05	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	7.96	8.05	0.989
second point	4960	40.0	3.98	4.02	0.991
third point	4980	20.0	1.99	2.01	0.991
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	7.96	8.07	0.987
Average Correction Factor					0.990
Baseline Corr AF:	8.05	Prev response	7.94	*% change	1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999063	1.005850
THC Cal Offset:	-0.016000	-0.014000
CH ₄ Cal Slope:	0.997102	1.010733
CH ₄ Cal Offset:	-0.002000	-0.002000
NMHC Cal Slope:	1.000411	1.001548
NMHC Cal Offset:	-0.010000	-0.012000

Notes: Hydrogen and Nitrogen changed out. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 9, 2023	Previous Calibration:	May 4, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:27	End Time (MST):	9:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
17.01	17.10	0.9945						
8.50	8.53	0.9969				Slope	1.005850	0.90 - 1.10
4.25	4.25	1.0004						
			Intercept	-0.014000	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 9, 2023	Previous Calibration:	May 4, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:27	End Time (MST):	9:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>		
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995		
7.96	8.05	0.9894		Slope	1.010733	0.90 - 1.10	
3.98	4.02	0.9906			Intercept	-0.002000	+/-0.5
1.99	2.01	0.9906					





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 9, 2023	Previous Calibration:	May 4, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:27	End Time (MST):	9:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	≥ 0.995			
9.04	9.05	0.9991						
4.52	4.51	1.0024				Slope	1.001548	0.90 - 1.10
2.26	2.24	1.0092						
			Intercept	-0.012000	± 0.5			



NMHC Calibration Plot

Date: June 9, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: June 5, 2023
Start time (MST): 6:45
Reason: Routine
Station number: AMS04
Last Cal Date: May 19, 2023
End time (MST): 11:56

Calibration Standards

NO Gas Cylinder #: T36RH1F
NOX Cal Gas Conc: 51.16 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 51.16 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: August 18, 2023
NO Cal Gas Conc: 50.91 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.91 ppm
NO gas Diff:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 723

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.055	1.146	NO bkgnd or offset:	-1.2	-2.6
NOX coeff or slope:	1.056	1.149	NOX bkgnd or offset:	2.7	2.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	7.6	7.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999316	1.001047
NO _x Cal Offset:	-0.653591	1.266147
NO Cal Slope:	1.001680	0.997960
NO Cal Offset:	-1.693638	0.425945
NO ₂ Cal Slope:	1.001373	1.007354
NO ₂ Cal Offset:	-0.003271	1.181889



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-2.5	-1.1	-1.3	----	----
as found span	4922	78.1	799.1	795.2	3.9	733.9	729.4	4.6	1.0888	1.0902
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.5	0.6	-0.1	----	----
high point	4922	78.1	799.1	795.2	3.9	800.7	793.9	7.0	0.9980	1.0016
second point	4961	39.1	400.1	398.1	2.0	402.5	398.2	4.3	0.9939	0.9998
third point	4981	19.5	199.5	198.5	1.0	201.5	198.0	3.5	0.9901	1.0027
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.8	-0.5	----	----
as left span	4922	78.1	799.1	381.4	417.7	798.2	378.2	419.9	1.0011	1.0085
Average Correction Factor									0.9940	1.0014

Corrected As found	NO _x = 736.4 ppb	NO = 730.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -8.4%
Previous Response	NO _x = 797.9 ppb	NO = 794.8 ppb		*Percent Change	NO = -8.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	791.8	378.0	417.7	421.4	0.9912	100.9%
2nd GPT point (200 ppb O3)	791.8	589.8	205.9	209.0	0.9852	101.5%
3rd GPT point (100 ppb O3)	791.8	690.8	104.9	108.3	0.9687	103.2%
Average Correction Factor					0.9817	101.9%

Notes: Zero and Span adjusted. Diagnostics similar to last month. Moly converter was replaced last month.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 19, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:45	End Time (MST):	11:56
Analyzer make:	API T200	Analyzer serial #:	723

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.5	----	Correlation Coefficient	≥0.995	
799.1	800.7	0.9980			
400.1	402.5	0.9939			
199.5	201.5	0.9901			
			Slope	1.001047	0.90 - 1.10
			Intercept	1.266147	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 19, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:45	End Time (MST):	11:56
Analyzer make:	API T200	Analyzer serial #:	723

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.6	----	Correlation Coefficient	≥0.995	
795.2	793.9	1.0016			
398.1	398.2	0.9998			
198.5	198.0	1.0027			
			Slope	0.997960	0.90 - 1.10
			Intercept	0.425945	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 19, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:45	End Time (MST):	11:56
Analyzer make:	API T200	Analyzer serial #:	723

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 $0.90 - 1.10$ ± 20
417.7	421.4	0.9912		
205.9	209.0	0.9852		
104.9	108.3	0.9687		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: June 5, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: June 23, 2023 Last Cal Date: May 4, 2023
 Start time (MST): 6:56 End time (MST): 9:28
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 2961
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998743	1.000343	Backgd or Offset:	-2.6	-2.6
Calibration intercept:	1.020000	-0.060000	Coeff or Slope:	1.008	1.008

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.5	----
as found span	5000	984.4	400.0	398.2	1.005
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.2	----
high point	5000	986.2	400.0	400.0	1.000
second point	5000	817.2	200.0	200.1	1.000
third point	5000	707.8	100.0	100.1	0.999
as left zero	5000	0.0	0.0	0.2	----
as left span	5000	986.3	400.0	399.7	1.001
Average Correction Factor					1.000

Baseline Corr As found:	398.7	Previous response	400.5	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No Maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 23, 2023	Previous Calibration:	May 4, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:56	End Time (MST):	9:28
Analyzer make:	API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999999	≥0.995
400.0	400.0	1.0000			
200.0	200.1	0.9995	Slope	1.000343	0.90 - 1.10
100.0	100.1	0.9990			
			Intercept	-0.060000	+/- 5



O₃ Calibration Plot

Date: June 23, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: June 23, 2023 Last Cal Date: May 23, 2023
 Start time (MST): 6:04 End time (MST): 6:26

Analyzer Make: API T640 S/N: 322
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Deltacal S/N: 1451
 Temp/RH standard: Deltacal S/N: 1451

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	16.3	16.8	16.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	728.1	729.3	728.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	4.70	5.00	<input checked="" type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: June 23, 2023	Last Cal Date: May 23, 2023			
	PM w/o HEPA: 14.9	PM w/ HEPA: 0			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		May 23, 2023			<0.2 ug/m3
Disposable Filter Changed:		May 23, 2023			

Annual Maintenance

Date Sample Tube Cleaned: May 23, 2023
 Date RH/T Sensor Cleaned: May 23, 2023

Flow adjusted. Head cleaned.

Notes:

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS05
MANNIX
JUNE 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Mannix Station number: AMS05
 Calibration Date: June 15, 2023 Last Cal Date: May 5, 2023
 Start time (MST): 8:50 End time (MST): 12:20
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.02 ppm Cal Gas Exp Date: January 12, 2029
 Cal Gas Cylinder #: XC026809B
 Removed Cal Gas Conc: 50.02 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 621
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 1008841399
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.990561	1.000214	Backgd or Offset: 8.8	8.9
Calibration intercept:	0.040000	-0.540000	Coeff or Slope: 0.920	0.908

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4920	80.0	800.3	810.3	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4920	80.0	800.3	800.3	1.000
second point	4960	40.0	400.2	399.6	1.001
third point	4980	20.0	200.1	198.4	1.008
as left zero	5000	0.0	0.0	0.8	----
as left span	4920	80.0	800.3	799.4	1.001
Average Correction Factor					1.003

Baseline Corr As found: 810.00 Previous response 792.81 *% change 2.1%
 Baseline Corr 2nd AF pt: NA AF Slope: AF Intercept:
 Baseline Corr 3rd AF pt: NA AF Correlation:

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 5, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	8:50	End Time (MST):	12:20
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.4	----	Correlation Coefficient	0.999993	
800.3	800.3	1.0000			≥0.995
400.2	399.6	1.0014	Slope	1.000214	
200.1	198.4	1.0085			0.90 - 1.10
			Intercept	-0.540000	+/-30



SO2 Calibration Plot

Date: June 15, 2023

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mannix Station number: AMS05
 Calibration Date: June 9, 2023 Last Cal Date: May 18, 2023
 Start time (MST): 8:39 End time (MST): 12:28
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.92 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002433
 Removed Cal Gas Conc: 4.92 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1845
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1203169745
 Converter make: Global Converter serial #: 2022-196
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005754	1.003756	Backgd or Offset:	2.25
Calibration intercept:	0.020710	0.040615	Coeff or Slope:	0.893
				0.821

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4919	81.3	80.0	89.2	0.897
as found 2nd point	4960	40.7	40.0	44.5	0.900
as found 3rd point	4980	20.3	20.0	22.1	0.904
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4919	81.3	80.0	80.3	0.996
second point	4960	40.7	40.0	40.3	0.994
third point	4980	20.3	20.0	20.1	0.994
as left zero	5000	0.0	0.0	0.2	----
as left span	4919	81.3	80.0	79.1	1.011
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.995
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 89.2 Prev response: 80.48 *% change: 9.8%
 Baseline Corr 2nd AF pt: 44.5 AF Slope: 1.115601 AF Intercept: -0.099272
 Baseline Corr 3rd AF pt: 22.1 AF Correlation: 0.999994

* = > +/-5% change initiates investigation

Notes: As found span is 10% high, suspecting the change in humidity has caused this as the instrument diagnostics are all good. Changed the inlet filter after as founds. Ran a SO2 scrubber test after calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 9, 2023	Previous Calibration:	May 18, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	8:39	End Time (MST):	12:28
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999998	≥0.995
80.0	80.3	0.9962			
40.0	40.3	0.9936	Slope	1.003756	0.90 - 1.10
20.0	20.1	0.9937			
			Intercept	0.040615	+/-3



H₂S Calibration Plot

Date: June 9, 2023

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Mannix	Station number:	AMS05
Calibration Date:	June 15, 2023	Last Cal Date:	May 8, 2023
Start time (MST):	8:49	End time (MST):	12:20
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.66E-04	2.66E-04	NMHC SP Ratio:	4.41E-05
CH ₄ Retention time:	15.20	15.20	NMHC Peak Area:	207349
				206739

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.03	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.25	0.998
second point	4960	40.0	8.61	8.63	0.998
third point	4980	20.0	4.31	4.33	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.32	0.995

Average Correction Factor				0.997
Baseline Corr AF:	17.03	Prev response	17.25	*% change -1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4920	80	9.15	9.12	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	80	9.15	9.16	0.999
second point	4960	40	4.57	4.60	0.995
third point	4980	20	2.29	2.31	0.988
as left zero	5000	0	0.00	0.00	----
as left span	4920	80	9.15	9.21	0.994
Average Correction Factor					0.994
Baseline Corr AF:	9.12	Prev response	9.18	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	7.91	1.022
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.09	0.998
second point	4960	40.0	4.04	4.03	1.002
third point	4980	20.0	2.02	2.01	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.11	0.996
Average Correction Factor					1.001
Baseline Corr AF:	7.91	Prev response	8.08	*% change	-2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001881	1.001244
THC Cal Offset:	-0.005800	0.006000
CH ₄ Cal Slope:	1.000679	1.001712
CH ₄ Cal Offset:	-0.008200	-0.006600
NMHC Cal Slope:	1.002855	1.000793
NMHC Cal Offset:	0.003000	0.013000

Notes: Changed the inlet filter and the N2 cylinder after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 8, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	8:49	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
17.23	17.25	0.9985						
8.61	8.63	0.9983				Slope	1.001244	0.90 - 1.10
4.31	4.33	0.9953						
			Intercept	0.006000	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 8, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	8:49	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995			
8.08	8.09	0.9984						
4.04	4.03	1.0023				Slope	1.001712	0.90 - 1.10
2.02	2.01	1.0028						
			Intercept	-0.006600	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 8, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	8:49	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999991	≥ 0.995			
9.15	9.16	0.9985						
4.57	4.60	0.9947				Slope	1.000793	0.90 - 1.10
2.29	2.31	0.9883						
			Intercept	0.013000	± 0.5			



NMHC Calibration Plot

Date: June 15, 2023

Location: Mannix





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	June 5, 2023	Last Cal Date:	May 17, 2023
Start time (MST):	9:05	End time (MST):	12:41
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.78	ppm	Cal Gas Exp Date:	September 9, 2024
Cal Gas Cylinder #:	AAL070632			
Removed Cal Gas Conc:	49.78	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	689
ZAG Make/Model:	API T701		Serial Number:	3566

Analyzer Information

Analyzer make: Thermo 43i
 Analyzer Range 0 - 1000 ppb
 Analyzer serial #: 1160290013

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998732	1.003750	Backgd or Offset:	17.2	17.2
Calibration intercept:	1.940243	1.159413	Coeff or Slope:	0.901	0.901

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4920	80.3	799.5	801.3	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.3	799.5	802.9	0.996
second point	4960	40.2	400.2	403.9	0.991
third point	4980	20.1	200.1	202.9	0.986
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	80.3	799.5	804.4	0.994
Average Correction Factor					0.991

Baseline Corr As found:	801.50	Previous response	800.39	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 17, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:05	End Time (MST):	12:41
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	0.999990	
799.5	802.9	0.9957			≥0.995
400.2	403.9	0.9909	Slope	1.003750	
200.1	202.9	0.9863			0.90 - 1.10
			Intercept	1.159413	+/-30



SO2 Calibration Plot

Date: June 5, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: June 12, 2023 Last Cal Date: May 12, 2023
 Start time (MST): 8:53 End time (MST): 13:15
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.38 ppm Cal Gas Exp Date: March 2, 2023
 Cal Gas Cylinder #: EY0000809
 Removed Cal Gas Conc: 5.38 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3566
 ZAG Make/Model: API T701 H Serial Number: 689

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1218153358
 Converter make: CDN-101 Converter serial #: 517
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996460	0.994772	Backgd or Offset: 1.91	1.92
Calibration intercept:	0.320162	0.337244	Coeff or Slope: 1.107	1.116

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.3	79.9	79.2	1.009
as found 2nd point	4963	37.2	40.0	40.0	1.001
as found 3rd point	4981	18.6	20.0	20.2	0.991
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	74.3	79.9	79.8	1.002
second point	4963	37.2	40.0	40.1	0.998
third point	4981	18.6	20.0	20.6	0.972
as left zero	5000	0.0	0.0	0.5	----
as left span	4926	74.3	79.9	79.0	1.012
SO2 Scrubber Check	4920	80.3	803.0	0.2	----

Date of last scrubber change:	December 20, 2021	Ave Corr Factor	0.991
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 79.2 Prev response: 79.98 *% change: -1.0%
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 0.989628 AF Intercept: 0.217291
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999964

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter and the converter after multi point as founds. Had the converter from global installed previously, swapped it out with a CD-NOVA converter. Ran a SO2 scrubber check after calibrator zero. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 12, 2023	Previous Calibration:	May 12, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:53	End Time (MST):	13:15
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999946	≥0.995
79.9	79.8	1.0018			
40.0	40.1	0.9981	Slope	0.994772	0.90 - 1.10
20.0	20.6	0.9716			
			Intercept	0.337244	+/-3



TRS Calibration Plot

Date: June 12, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	June 5, 2023	Last Cal Date:	May 17, 2023
Start time (MST):	9:05	End time (MST):	12:40
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH ₄ Cal Gas Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Ref.	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	261

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.57E-04	3.85E-04	NMHC SP Ratio:	5.96E-05
CH ₄ Retention time:	14.6	14.8	NMHC Peak Area:	152338
				150828

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.12	16.55	1.035
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	17.13	1.000
second point	4960	40.2	8.57	8.54	1.003
third point	4980	20.1	4.29	4.30	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.12	17.16	0.998

Average Correction Factor	1.000
---------------------------	-------

Baseline Corr AF:	16.55	Prev response	17.16	*% change	-3.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF Limit= 0.95-1.05
as found zero	5000	0	0.00	0.00	----
as found span	4920	80.3	9.07	9.01	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	80.3	9.07	9.05	1.002
second point	4960	40.2	4.54	4.54	1.000
third point	4980	20.1	2.27	2.29	0.991
as left zero	5000	0	0.00	0.00	----
as left span	4920	80.3	9.07	9.07	1.000
Average Correction Factor					0.998
Baseline Corr AF:	9.01	Prev response	9.10	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF Limit= 0.95-1.05
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.06	7.55	1.067
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.06	8.08	0.997
second point	4960	40.2	4.03	4.01	1.007
third point	4980	20.1	2.02	2.01	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.06	8.09	0.996
Average Correction Factor					1.002
Baseline Corr AF:	7.55	Prev response	8.06	*% change	-6.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002022	0.999886
THC Cal Offset:	0.005131	-0.000855
CH ₄ Cal Slope:	1.001093	1.003290
CH ₄ Cal Offset:	-0.003602	-0.013600
NMHC Cal Slope:	1.003075	0.996899
NMHC Cal Offset:	0.008332	0.012344

As found CH₄ is 6.7% low, checked all the diagnostics and chromatograms and couldn't find any issues.

Notes: Will monitor the daily zero spans and complete maintenance if the CH₄ continues dropping. Changed the inlet filter after as founds and adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 17, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:05	End Time (MST):	12:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995			
17.12	17.13	0.9996						
8.57	8.54	1.0033				Slope	0.999886	0.90 - 1.10
4.29	4.30	0.9965						
			Intercept	-0.000855	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 17, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:05	End Time (MST):	12:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999970	≥0.995			
8.06	8.08	0.9967						
4.03	4.01	1.0067				Slope	1.003290	0.90 - 1.10
2.02	2.01	1.0037						
			Intercept	-0.013600	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 17, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:05	End Time (MST):	12:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999990	≥ 0.995			
9.07	9.05	1.0022						
4.54	4.54	1.0003				Slope	0.996899	0.90 - 1.10
2.27	2.29	0.9907						
			Intercept	0.012344	± 0.5			



NMHC Calibration Plot

Date: June 5, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	June 19, 2023	Last Cal Date:	June 5, 2023
Start time (MST):	9:17	End time (MST):	14:23
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH ₄ Cal Gas Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Ref.	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	5608

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.85E-04	2.54E-04	NMHC SP Ratio:	6.02E-05
CH ₄ Retention time:	14.8	12.1	NMHC Peak Area:	150828
				115463

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.12	16.63	1.030
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	17.23	0.994
second point	4960	40.2	8.57	8.44	1.016
third point	4980	20.1	4.29	4.19	1.023
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.12	17.17	0.998

Average Correction Factor	1.011
---------------------------	-------

Baseline Corr AF:	16.63	Prev response	17.12	*% change	-2.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4920	80.3	9.07	8.97	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	80.3	9.07	9.13	0.993
second point	4960	40.2	4.54	4.52	1.004
third point	4980	20.1	2.27	2.25	1.008
as left zero	5000	0	0.00	0.00	----
as left span	4920	80.3	9.07	9.14	0.992
Average Correction Factor					1.002
Baseline Corr AF:	8.97	Prev response	9.05	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.06	7.66	1.052
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.06	8.10	0.995
second point	4960	40.2	4.03	3.92	1.029
third point	4980	20.1	2.02	1.94	1.041
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.06	8.02	1.004
Average Correction Factor					1.022
Baseline Corr AF:	7.66	Prev response	8.07	*% change	-5.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999886	1.007754
THC Cal Offset:	-0.000855	-0.087832
CH ₄ Cal Slope:	1.003290	1.007409
CH ₄ Cal Offset:	-0.013600	-0.063374
NMHC Cal Slope:	0.996899	1.007709
NMHC Cal Offset:	0.012344	-0.024058

Notes: CH₄ channel started dips yesterday. Optimized gas flows. Found ZAG leaking. Changed around 11:00 MST and left to warm up before calibrating. Use zero chrome and use flat baseline options removed.

Calibration Performed By: Ryan Power



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 19, 2023	Previous Calibration:	June 5, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:17	End Time (MST):	14:23
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999839	≥ 0.995			
17.12	17.23	0.9935						
8.57	8.44	1.0156				Slope	1.007754	0.90 - 1.10
4.29	4.19	1.0234						
			Intercept	-0.087832	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 19, 2023	Previous Calibration:	June 5, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:17	End Time (MST):	14:23
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999619	≥0.995			
8.06	8.10	0.9945						
4.03	3.92	1.0291				Slope	1.007409	0.90 - 1.10
2.02	1.94	1.0410						
			Intercept	-0.063374	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 19, 2023	Previous Calibration:	June 5, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:17	End Time (MST):	14:23
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999957	≥ 0.995			
9.07	9.13	0.9930						
4.54	4.52	1.0042				Slope	1.007709	0.90 - 1.10
2.27	2.25	1.0083						
			Intercept	-0.024058	± 0.5			



NMHC Calibration Plot

Date: June 19, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Patricia McInnes
Calibration Date: June 26, 2023
Start time (MST): 9:10
Reason: Routine
Station number: AMS06
Last Cal Date: May 1, 2023
End time (MST): 13:39

Calibration Standards

NO Gas Cylinder #: T30YCWN
NOX Cal Gas Conc: 47.94 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 47.94 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: April 11, 2025
NO Cal Gas Conc: 46.39 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 46.39 ppm
NO gas Diff:
Serial Number: 3566
Serial Number: 689

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1172750022

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.835	0.815	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.996	0.986	NOX bkgnd or offset:	3.9	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	156.0	156.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001069	0.996744
NO _x Cal Offset:	1.979986	2.176487
NO Cal Slope:	0.998555	1.000845
NO Cal Offset:	1.080026	1.122676
NO ₂ Cal Slope:	1.005730	1.000605
NO ₂ Cal Offset:	0.795849	-0.357261



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.2	----	----
as found span	4914	86.2	826.5	799.7	26.7	857.4	820.6	37.0	0.9639	0.9746
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.4	0.4	0.0	----	----
high point	4914	86.2	826.5	799.7	26.7	825.0	801.2	23.9	1.0018	0.9982
second point	4957	43.1	413.2	399.9	13.4	415.1	401.6	13.5	0.9955	0.9957
third point	4978	21.6	207.1	200.4	6.7	210.3	202.5	7.7	0.9849	0.9897
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
as left span	4923	76.9	737.3	313.7	423.6	824.6	402.4	422.1	0.8942	0.7796
Average Correction Factor									0.9940	0.9945

Corrected As found	NO _x = 857.3 ppb	NO = 820.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 3.3%
Previous Response	NO _x = 829.3 ppb	NO = 799.7 ppb		*Percent Change	NO = 2.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.1	402.2	423.6	423.6	1.0000	100.0%
2nd GPT point (200 ppb O3)	799.1	602.1	223.7	223.7	1.0001	100.0%
3rd GPT point (100 ppb O3)	799.1	702.2	123.6	122.7	1.0075	99.3%
Average Correction Factor					1.0025	99.7%

Notes:

Span adjusted. No maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 26, 2023	Previous Calibration:	May 1, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:10	End Time (MST):	13:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.5	825.0	1.0018		
413.2	415.1	0.9955		
207.1	210.3	0.9849		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 26, 2023	Previous Calibration:	May 1, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:10	End Time (MST):	13:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999996	≥0.995
799.7	801.2	0.9982			
399.9	401.6	0.9957	Slope	1.000845	0.90 - 1.10
200.4	202.5	0.9897			
			Intercept	1.122676	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 26, 2023	Previous Calibration:	May 1, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:10	End Time (MST):	13:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
423.6	423.6	1.0000		
223.7	223.7	1.0001		
123.6	122.7	1.0075		
			0.999994	
			1.000605	
			-0.357261	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: June 26, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Patricia McInnes Station number: AMS06
 Calibration Date: June 13, 2023 Last Cal Date: May 15, 2023
 Start time (MST): 10:04 End time (MST): 13:15
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3566
 ZAG Make/Model: API T701H Serial Number: 689

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1300156234
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001971	1.002286	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	0.480000	0.600000	Coeff or Slope:	1.019	1.019

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	-0.9	----
as found span	5000	1303.0	400.0	400.3	0.999
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	-0.6	----
high point	5000	1303.0	400.0	400.6	0.999
second point	5000	966.5	200.0	202.6	0.987
third point	5000	794.3	100.0	101.4	0.986
as left zero	5000	800.0	0.0	0.6	----
as left span	5000	1303.0	400.0	402.9	0.993
Average Correction Factor					0.991

Baseline Corr As found:	401.2	Previous response	401.3	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 13, 2023	Previous Calibration:	May 15, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:04	End Time (MST):	13:15
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.6	----	Correlation Coefficient	0.999943	
400.0	400.6	0.9985			≥0.995
200.0	202.6	0.9872	Slope	1.002286	
100.0	101.4	0.9862			0.90 - 1.10
			Intercept	0.600000	+/- 5



O₃ Calibration Plot

Date: June 13, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: June 5, 2023 Last Cal Date: May 17, 2023
 Start time (MST): 12:53 End time (MST): 13:56

Analyzer Make: API T640 S/N: 766
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT FP-25 S/N: 388755
 Temp/RH standard: ALICAT FP-25 S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	22.4	22.0	22.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	725.5	726.2	725.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.16	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>June 5, 2023</u>		Last Cal Date: <u>May 17, 2023</u>		
	PM w/o HEPA: <u>12.9</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	13	13	10.8	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: <u>9.4</u>		w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:	<u>June 5, 2023</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>June 5, 2023</u>				

Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023
 Date RH/T Sensor Cleaned: April 13, 2023

Notes: "Perform span dust check" alarm was on upon arrival. Was unable to complete PMT Peak test as no readings were showing with the span dust. Cleaned the chamber and adjusted the PMT peak. Leak check passed, no other adjustments made.

Calibration by: Max Farrell



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: June 28, 2023 Last Cal Date: June 5, 2023
 Start time (MST): 8:50 End time (MST): 12:00

Analyzer Make: API T640 S/N: 766
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT FP-25 S/N: 388744
 Temp/RH standard: ALICAT FP-25 S/N: 388744

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	27.3	26.6	27.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	723.2	724.6	723.2	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	0.32	0.32	5.00	<input checked="" type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: June 28, 2023 Last Cal Date: June 5, 2023
 PM w/o HEPA: _____ PM w/ HEPA: _____ <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: _____ w/ HEPA: _____
 Date Optical Chamber Cleaned: June 5, 2023 <0.2 ug/m3
 Disposable Filter Changed: June 5, 2023

Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023
 Date RH/T Sensor Cleaned: April 13, 2023

Notes: As found flow was low at 0.3 LPM, pump power output at 100%. Replaced pump but only half flow restored, troubleshooting revealed internal DFU filter was plugged, replaced filter.

Calibration by: Kelly Baragar



Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	June 26, 2023	Last Cal Date:	May 24, 2023
Start time (MST):	9:10	End time (MST):	13:08
NH3 Cal Date:	June 27, 2023	Last Cal Date:	May 24, 2023
Start time (MST):	8:30	End time (MST):	15:58
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	NA
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	77.8	ppm	NH3 Gas Cylinder #:	CC710812
			NH3 Cal Gas Expiry:	March 30, 2023
Removed NH3 Conc:	77.8	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	API T700		Serial Number:	3566
ZAG make/model:	API T701		Serial Number:	689

Analyzer Information

Analyzer model: Teledyne API T201	Analyzer serial #: 152
Converter model: Teledyne API T501	Converter serial #: 147
NH3 Range (ppb): 0 - 2000 ppb	Reaction cell Press: 6.50
NOX Range (ppb): 0 - 1000 ppb	Sample Flow: 25.3

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.794	0.794	TN coefficient:	0.794	0.794
NOX coefficient:	0.796	0.796	NO bkgrnd:	-0.1	-0.1
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	0.0	0.0
NH3 coefficient:	0.908	0.908	TN bkgrnd:	0.0	0.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995223	1.005776
NO _x Cal Offset:	1.876743	1.234716
NO Cal Slope:	0.994331	0.998846
NO Cal Offset:	1.802690	1.222396
NO ₂ Cal Slope:	1.004951	1.005535
NO ₂ Cal Offset:	1.630262	0.912555
NH3 Cal Slope:	1.008694	1.017396
NH3 Cal Offset:	6.507686	0.876681
TN Cal Slope:	1.013896	1.023007
TN Cal Offset:	6.739810	0.746567



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
as found NO	4914	86.2	826.5	826.5	----	822.8	824.0	-1.3	1.004	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
high NO point	4914	86.2	826.5	826.5	----	822.8	824.0	-1.3	1.004	----
GPT point	4914	86.2	826.5	826.5	----	827.0	825.1	1.7	0.999	----
as found NH3	3419	81.0	1800.5	----	1800.5	1843.3	----	1833.4	0.977	0.982
new NH3 cyl rp							----			
first NH3	3419	81.0	1800.5	----	1800.5	1843.3	----	1833.4	0.977	0.982
second NH3	3455	45.0	1000.3	----	1000.3	1021.5	----	1015.8	0.979	0.985
third NH3	3478	22.5	500.1	----	500.1	515.2	----	512.7	0.971	0.975
Average Correction Factor									1.0019	0.9807

Corrected As found TN = 823 ppb NO_x = 824.1 ppb NH3 = 1833.5 ppb

Previous Response TN = 844.7 ppb NO_x = 824.4 ppb NH3 = 1822.7 ppb

NH3 Previous Converter Efficiency = 90.8%

NH3 Current Converter Efficiency = 90.8%

*Percent Change TN = -2.6%

*Percent Change NO_x = 0.0%

*Percent Change NH3 = 0.6%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.8	0.7	----	----
as found span	4914	86.2	826.5	799.7	826.5	829.7	796.0	824.9	0.9961	1.0047
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	0.1	0.2	----	----
high point	4914	86.2	826.5	799.7	826.5	831.7	800.0	828.4	0.9937	0.9997
second point	4957	43.1	413.2	399.9	413.2	417.6	399.6	417.7	0.9895	1.0007
third point	4978	21.6	207.1	200.4	207.1	211.1	203.6	211.4	0.9811	0.9844
Average Correction Factor									0.9881	0.9949

Baseline Corr As fnd	TN = 824.2 ppb	NO _x = 829.6 ppb	NO = 795.2 ppb	*Percent Change	TN = -2.5%
Previous Response	TN = 844.7 ppb	NO _x = 824.4 ppb	NO = 797.0 ppb	*Percent Change	NO _x = 0.6%
				*Percent Change	NO = -0.2%

* = > +/-5% change initiates investigation

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	-0.7	----	----
calibration zero	----	----	0.0	-0.3	----	----
1st GPT point (400 ppb O3)	799.6	399.0	427.3	429.7	0.9945	100.6%
2nd GPT point (200 ppb O3)	799.6	603.3	223.0	226.4	0.9851	101.5%
3rd GPT point (100 ppb O3)	799.6	702.1	124.2	126.7	0.9804	102.0%
Average Correction Factor					0.9867	101.4%

Notes: No_x portion of the calibration done today. No adjustments done.

Calibration Performed By: Maddison Baragar



Wood Buffalo Environmental Association

TN Calibration Summary


Version-05-2023

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 24, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:30	End Time (MST):	15:58
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.5	1843.3	0.9768		
1000.3	1021.5	0.9792		
500.1	515.2	0.9706		
			0.999991	
			1.023007	
			0.746567	





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 24, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:30	End Time (MST):	15:58
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient 0.999990	≥0.995	
1800.5	1833.4	0.9821			
1000.3	1015.8	0.9847			
500.1	512.7	0.9754			
			Slope	1.017396	0.90 - 1.10
			Intercept	0.876681	+/-20





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	June 26, 2023	Previous Calibration:	May 24, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:10	End Time (MST):	13:08
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.5	831.7	0.9937		
413.2	417.6	0.9895		
207.1	211.1	0.9811		
			0.999984	
			1.005776	
			1.234716	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	June 26, 2023	Previous Calibration:	May 24, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:10	End Time (MST):	13:08
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999980	≥0.995
799.7	800.0	0.9997			
399.9	399.6	1.0007	Slope	0.998846	0.90 - 1.10
200.4	203.6	0.9844			
			Intercept	1.222396	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	June 26, 2023	Previous Calibration:	May 24, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:10	End Time (MST):	13:08
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999954	≥0.995
427.3	429.7	0.9945			
223.0	226.4	0.9851	Slope	1.005535	0.90 - 1.10
124.2	126.7	0.9804			
			Intercept	0.912555	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: June 26, 2023


Location: Patricia McInnes



NH₃ Calibration Plot

Date: June 27, 2023

Location: Patricia McInnes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Athabasca Valley Station number: AMS07
Calibration Date: June 16, 2023 Last Cal Date: May 15, 2023
Start time (MST): 9:05 End time (MST): 12:38
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.52 ppm Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #: CC282115
Removed Cal Gas Conc: 50.52 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3805
ZAG Make/Model: API 701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1507864683
Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004627	1.003815	Backgd or Offset:	2.70	2.70
Calibration intercept:	1.624552	1.384231	Coeff or Slope:	0.845	0.840

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	79.3	801.2	803.9	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	79.3	801.2	804.9	0.995
second point	4960	39.6	400.2	404.1	0.990
third point	4980	19.8	200.1	203.1	0.985
as left zero	5000	0.0	0.0	0.0	----
as left span	4921	79.2	800.2	804.1	0.995
Average Correction Factor					0.990

Baseline Corr As found: 804.00 Previous response: 806.53 *% change: -0.3%
Baseline Corr 2nd AF pt: NA AF Slope: AF Intercept:
Baseline Corr 3rd AF pt: NA AF Correlation:

* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 16, 2023	Previous Calibration:	May 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:05	End Time (MST):	12:38
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999989	
801.2	804.9	0.9954			≥0.995
400.2	404.1	0.9902	Slope	1.003815	
200.1	203.1	0.9851			0.90 - 1.10
			Intercept	1.384231	+/-30



SO2 Calibration Plot

Date: June 16, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: June 26, 2023 Last Cal Date: May 10, 2023
 Start time (MST): 9:55 End time (MST): 14:55
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.25 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC504080
 Removed Cal Gas Conc: 5.25 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3805
 ZAG Make/Model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE Analyzer serial #: 1180540018
 Converter make: CDN-101 Converter serial #: 551
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002951	1.003525	Backgd or Offset: 2.29	2.33
Calibration intercept:	-0.222144	-0.042085	Coeff or Slope: 0.876	0.885

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4925	75.5	79.3	79.6	0.995
as found 2nd point	4962	37.7	39.6	39.4	1.002
as found 3rd point	4981	18.9	19.8	19.5	1.013
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4925	75.5	79.3	79.5	0.998
second point	4962	37.7	39.6	39.9	0.993
third point	4981	18.9	19.9	19.7	1.008
as left zero	5000	0.0	0.0	0.1	----
as left span	4925	75.5	79.3	79.3	1.000
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	0.999
Date of last converter efficiency test:	April 22, 2022			92.6% efficiency	

Baseline Corr As found: 79.7 Prev response: 79.28 *% change: 0.5%
 Baseline Corr 2nd AF pt: 39.5 AF Slope: 1.006555 AF Intercept: -0.302251
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999970

* = > +/-5% change initiates investigation

Notes: Scrubber check passed. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 26, 2023	Previous Calibration:	May 10, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:55	End Time (MST):	14:55
Analyzer make:	CDN-101	Analyzer serial #:	551

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999978	
79.3	79.5	0.9976			≥0.995
39.6	39.9	0.9927	Slope	1.003525	
19.9	19.7	1.0080			0.90 - 1.10
			Intercept	-0.042085	+/-3



TRS Calibration Plot

Date: June 26, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	June 16, 2023	Last Cal Date:	May 15, 2023
Start time (MST):	9:05	End time (MST):	12:38
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	0.000280	0.000288	NMHC SP Ratio: 4.75E-05	5.28E-05
CH ₄ Retention time:	13.8	14.0	NMHC Peak Area:	191368 172128

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.3	17.05	16.00	1.066
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.3	17.05	17.12	0.996
second point	4960	39.6	8.52	8.53	0.998
third point	4980	19.8	4.26	4.33	0.983
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.12	0.995

				Average Correction Factor	0.992
Baseline Corr AF:	16.00	Prev response	17.05	*% change	-6.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.3	9.10	8.25	1.103
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.3	9.10	9.18	0.991
second point	4960	39.6	4.55	4.55	0.999
third point	4980	19.8	2.27	2.34	0.972
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	9.09	9.14	0.994
Average Correction Factor					0.987
Baseline Corr AF:	8.25	Prev response	9.12	*% change	-10.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.3	7.95	7.75	1.026
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.3	7.95	7.94	1.001
second point	4960	39.6	3.97	3.98	0.998
third point	4980	19.8	1.98	1.99	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.94	7.97	0.996
Average Correction Factor					0.998
Baseline Corr AF:	7.75	Prev response	7.93	*% change	-2.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999351	1.002897
THC Cal Offset:	0.013892	0.018700
CH ₄ Cal Slope:	0.998818	0.998369
CH ₄ Cal Offset:	-0.004193	0.007616
NMHC Cal Slope:	1.000193	1.006726
NMHC Cal Offset:	0.014085	0.011084

Notes: NM channel was 11% low. No maintenance done today, adjusted span. Will monitor.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 16, 2023	Previous Calibration:	May 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:05	End Time (MST):	12:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999980	≥ 0.995			
17.05	17.12	0.9958						
8.52	8.53	0.9983				Slope	1.002897	0.90 - 1.10
4.26	4.33	0.9826						
			Intercept	0.018700	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 16, 2023	Previous Calibration:	May 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:05	End Time (MST):	12:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	≥0.995
7.95	7.94	1.0013			
3.97	3.98	0.9977			
1.98	1.99	0.9954			
			Slope	0.998369	0.90 - 1.10
			Intercept	0.007616	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 16, 2023	Previous Calibration:	May 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:05	End Time (MST):	12:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999931	≥ 0.995			
9.10	9.18	0.9912						
4.55	4.55	0.9991				Slope	1.006726	0.90 - 1.10
2.27	2.34	0.9717						
			Intercept	0.011084	± 0.5			



NMHC Calibration Plot

Date: June 16, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	June 28, 2023	Last Cal Date:	June 16, 2023
Start time (MST):	10:48	End time (MST):	16:13
Reason:	Maintenance	Investigation to NM sensitivity drop observed in daily spans	

Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	0.000280	0.000288	NMHC SP Ratio: 4.75E-05	5.28E-05
CH ₄ Retention time:	13.8	14.0	NMHC Peak Area:	191368 172128

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.21	1.051
as found 2nd point	4960	39.6	8.52	8.19	1.040
as found 3rd point	4980	19.8	4.26	4.03	1.056
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	16.93	1.006
second point	4960	39.6	8.52	8.45	1.008
third point	4980	19.8	4.26	4.23	1.008
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	16.88	1.009
Average Correction Factor					1.007
Baseline Corr AF:	16.21	Prev response	17.10	*% change	-5.5%
Baseline Corr 2nd AF:	8.2	AF Slope:	0.952684	AF Intercept:	0.008873
Baseline Corr 3rd AF:	4.0	AF Correlation:	0.999956	* => +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.09	8.41	1.081
as found 2nd point	4960	39.6	4.55	4.29	1.060
as found 3rd point	4980	19.8	2.27	2.07	1.097
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	9.09	8.98	1.012
second point	4960	39.6	4.55	4.47	1.017
third point	4980	19.8	2.27	2.24	1.016
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	9.09	8.91	1.020
Average Correction Factor					1.015
Baseline Corr AF:	8.41	Prev response	9.16	*% change	-9.0%
Baseline Corr 2nd AF:	4.3	AF Slope:	0.927375	AF Intercept:	0.004029
Baseline Corr 3rd AF:	2.1	AF Correlation:	0.999818	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.94	7.80	1.018
as found 2nd point	4960	39.6	3.97	3.90	1.018
as found 3rd point	4980	19.8	1.98	1.96	1.013
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.94	7.94	1.000
second point	4960	39.6	3.97	3.98	0.998
third point	4980	19.8	1.98	1.99	0.998
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.94	7.96	0.997
Average Correction Factor					0.999
Baseline Corr AF:	7.80	Prev response	7.93	*% change	-1.7%
Baseline Corr 2nd AF:	3.90	AF Slope:	0.981968	AF Intercept:	0.004044
Baseline Corr 3rd AF:	1.96	AF Correlation:	0.999998	* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002897	0.993911
THC Cal Offset:	0.018700	-0.005538
CH ₄ Cal Slope:	0.998369	1.000237
CH ₄ Cal Offset:	0.007616	0.002841
NMHC Cal Slope:	1.006726	0.988285
NMHC Cal Offset:	0.011084	-0.007979

Notes: Steady decline in sensitivity noted in daily spans, replaced pump. Also found the sample line worn through by vibration, replaced sample line.

Calibration Performed By: Kelly Baragar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	June 16, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:48	End Time (MST):	16:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
17.03	16.93	1.0062						
8.52	8.45	1.0079				Slope	0.993911	0.90 - 1.10
4.26	4.23	1.0077						
			Intercept	-0.005538	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	June 16, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:48	End Time (MST):	16:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥0.995
7.94	7.94	0.9997			
3.97	3.98	0.9979			
1.98	1.99	0.9979			
			Slope	1.000237	0.90 - 1.10
			Intercept	0.002841	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	June 16, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:48	End Time (MST):	16:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995
9.09	8.98	1.0120			
4.55	4.47	1.0165			
2.27	2.24	1.0165			
			Slope	0.988285	0.90 - 1.10
			Intercept	-0.007979	+/-0.5



NMHC Calibration Plot

Date: June 28, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	June 12, 2023	Last Cal Date:	May 8, 2023
Start time (MST):	9:13	End time (MST):	14:18
Reason:	Routine		

Calibration Standards

NO Gas Cylinder #:	T2Y1KA4	Cal Gas Expiry Date:	November 30, 2023
NOX Cal Gas Conc:	50.92 ppm	NO Cal Gas Conc:	49.92 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	50.92 ppm	Removed Gas NO Conc:	49.92 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API T701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.048	1.048	NO bkgnd or offset:	7.3
NOX coeff or slope:	0.995	0.995	NOX bkgnd or offset:	7.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	201.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.991417	0.997768
NO _x Cal Offset:	1.497144	1.378516
NO Cal Slope:	0.991499	0.998277
NO Cal Offset:	0.973228	0.974621
NO ₂ Cal Slope:	1.004055	1.001546
NO ₂ Cal Offset:	0.235209	0.752238



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.2	----	----
as found span	4920	80.2	816.7	800.7	16.0	818.6	800.3	18.2	0.9977	1.0005
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
high point	4920	80.2	816.7	800.7	16.0	815.3	799.5	15.8	1.0017	1.0015
second point	4960	40.1	408.4	400.4	8.0	410.6	402.1	8.6	0.9946	0.9956
third point	4980	20.0	203.7	199.7	4.0	205.0	200.5	4.4	0.9936	0.9959
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
as left span	4920	80.2	816.7	395.1	421.6	821.0	395.9	424.7	0.9948	0.9979
Average Correction Factor									0.9966	0.9977

Corrected As found	NO _x = 818.3 ppb	NO = 800.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.9%
Previous Response	NO _x = 811.2 ppb	NO = 794.9 ppb		*Percent Change	NO = 0.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.6	392.0	421.6	422.6	0.9977	100.2%
2nd GPT point (200 ppb O3)	797.6	594.7	218.9	220.6	0.9925	100.8%
3rd GPT point (100 ppb O3)	797.6	695.4	118.2	119.7	0.9878	101.2%
Average Correction Factor					0.9927	100.7%

Notes:

No maintenance or adjustments done.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 12, 2023	Previous Calibration:	May 8, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:13	End Time (MST):	14:18
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	≥0.995	
816.7	815.3	1.0017			
408.4	410.6	0.9946			
203.7	205.0	0.9936			
			Slope	0.997768	0.90 - 1.10
			Intercept	1.378516	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 12, 2023	Previous Calibration:	May 8, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:13	End Time (MST):	14:18
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.7	799.5	1.0015		
400.4	402.1	0.9956		
199.7	200.5	0.9959		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 12, 2023	Previous Calibration:	May 8, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:13	End Time (MST):	14:18
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
421.6	422.6	0.9977			
218.9	220.6	0.9925			
118.2	119.7	0.9878			
			Slope	1.001546	0.90 - 1.10
			Intercept	0.752238	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: June 12, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: June 6, 2023 Last Cal Date: May 3, 2023
 Start time (MST): 7:48 End time (MST): 12:17
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: T700 Serial Number: 3805
 ZAG Make/Model: T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1152220023
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000257	0.999971	Backgd or Offset:	-1.3	-3.5
Calibration intercept:	1.380000	0.480000	Coeff or Slope:	1.502	1.504

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-3.0	----
as found span	5000	1414.8	400.0	393.3	1.017
as found 2nd point	5000	1038.6	200.0	197.0	1.015
as found 3rd point	5000	855.5	100.0	98.4	1.016
calibrator zero	5000	0.0	0.0	0.5	----
high point	5000	1415.7	400.0	400.9	0.998
second point	5000	1039.9	200.0	199.2	1.004
third point	5000	856.2	100.0	101.3	0.987
as left zero	5000	0.0	0.0	0.3	----
as left span	5000	1416.0	400.0	401.0	0.998
Average Correction Factor					0.996

Baseline Corr As found:	396.3	Previous response	401.5	*% change	-1.3%
Baseline Corr 2nd AF pt:	-196.3	AF Slope:	0.989286	AF Intercept:	-1.700000
Baseline Corr 3rd AF pt:	-98.6	AF Correlation:	0.999950		

* = > +/-5% change initiates investigation

Notes: Big rain storms past 2 days. No water present in the trap of the manifold. Multi point as found was stable and linear. Filter was very dirty. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 6, 2023	Previous Calibration:	May 3, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	7:48	End Time (MST):	12:17
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999972	≥0.995
400.0	400.9	0.9978			
200.0	199.2	1.0040	Slope	0.999971	0.90 - 1.10
100.0	101.3	0.9872			
			Intercept	0.480000	+/- 5



O₃ Calibration Plot

Date: June 6, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Athabasca Valley Station number: AMS 07
 Calibration Date: June 29, 2023 Last Cal Date: May 23, 2023
 Start time (MST): 11:53 End time (MST): 13:11

Analyzer Make: API T640 S/N: 871
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	31.6	31.3	31.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	729.5	732.43	729.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.93	4.86	4.93	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>June 29, 2023</u>		Last Cal Date: <u>May 23, 2023</u>		
	PM w/o HEPA: <u>2.5</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.2	10.2	11	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>May 23, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>May 23, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022
 Date RH/T Sensor Cleaned: December 5, 2022

Notes:

Temp, flow and pressure checked. Leak check passed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	June 29, 2023	Last Cal Date:	May 15, 2023
Start time (MST):	9:44	End time (MST):	13:04
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date:	December 12, 2026
Cal Gas Cylinder #:	LL66942			
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 700H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	Start	Finish		Start	Finish
Calibration slope:	0.991945	0.999915	Backgd or Offset:	3.910	4.147
Calibration intercept:	0.060555	0.018540	Coeff or Slope:	1.079	1.086

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4933	66.7	40.0	40.2	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.0	40.0	1.001
second point	4967	33.3	20.0	20.1	0.995
third point	4983	16.7	10.0	10.0	1.000
as left zero	5000	0.0	0.0	0.0	----
as left span	4933	66.7	40.0	40.0	1.001
Average Correction Factor					0.999

Baseline Corr As found:	40.01	Prev response:	39.76	*% change:	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No Maintenance done. Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 29, 2023	Previous Calibration:	May 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:44	End Time (MST):	13:04
Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999989	
40.0	40.0	1.0006			≥0.995
20.0	20.1	0.9950	Slope	0.999915	
10.0	10.0	1.0001			0.90 - 1.10
			Intercept	0.018540	+/-1.5



CO Calibration Plot

Date: June 29, 2023

Location: Athabasca Valley





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	June 28, 2023	Last Cal Date:	May 8, 2023
Start time (MST):	5:55	End time (MST):	8:22
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC196697			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3060
ZAG Make/Model:	Teledyne API T701		Serial Number:	260

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999415	0.996931	Backgd or Offset:	1.50	1.60
Calibration intercept:	-1.203236	-0.082970	Coeff or Slope:	0.877	0.915

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4920	80.3	800.4	765.1	1.046
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.3	800.4	798.4	1.002
second point	4960	40.2	400.7	397.9	1.007
third point	4980	20.1	200.4	200.4	1.000
as left zero	5000	0.0	0.0	0.0	----
as left span	4920	80.3	800.4	803.9	0.996
Average Correction Factor					1.003

Baseline Corr As found:	765.20	Previous response	798.71	*% change	-4.4%
-------------------------	--------	-------------------	--------	-----------	-------

** = > +/-5% change initiates investigation*

Notes: Remote calibration done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	May 8, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	5:55	End Time (MST):	8:22
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999991	≥0.995
800.4	798.4	1.0025			
400.7	397.9	1.0070	Slope	0.996931	0.90 - 1.10
200.4	200.4	0.9998			
			Intercept	-0.082970	+/-30



SO2 Calibration Plot

Date: June 28, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: June 29, 2023 Last Cal Date: May 29, 2023
 Start time (MST): 8:15 End time (MST): 12:25
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002276
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 260

Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 1203169744
 Converter make: CDN-101 Converter serial #: 14639
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.976855	1.005570	Backgd or Offset:	0.93	0.94
Calibration intercept:	0.498472	0.218839	Coeff or Slope:	0.694	0.707

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4920	80.5	80.0	78.3	1.026
as found 2nd point	4960	40.2	40.0	39.3	1.025
as found 3rd point	4980	20.1	20.0	19.6	1.035
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4920	80.5	80.0	80.7	0.991
second point	4960	40.2	40.0	40.4	0.989
third point	4980	20.1	20.0	20.2	0.989
as left zero	5000	0.0	0.0	0.4	----
as left span	4920	80.5	80.0	79.9	1.001
SO2 Scrubber Check	4919.7	80.3	803.0	0.1	----
Date of last scrubber change:	March 7, 2022			Ave Corr Factor	0.990
Date of last converter efficiency test:	March 15, 2022			100.7% efficiency	

Baseline Corr As found: 78.0 Prev response: 78.66 *% change: -0.8%
 Baseline Corr 2nd AF pt: 39.0 AF Slope: 0.975716 AF Intercept: 0.238308
 Baseline Corr 3rd AF pt: 19.3 AF Correlation: 0.999992

* = > +/-5% change initiates investigation

Notes: Remote calibration done. Sox scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 29, 2023	Previous Calibration:	May 29, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:15	End Time (MST):	12:25
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999995	
80.0	80.7	0.9914			≥0.995
40.0	40.4	0.9890	Slope	1.005570	
20.0	20.2	0.9891			0.90 - 1.10
			Intercept	0.218839	+/-3



TRS Calibration Plot

Date: June 29, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
Calibration Date: June 27, 2023 Last Cal Date: May 15, 2023
Start time (MST): 8:15 End time (MST): 11:34
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC363447 Cal Gas Expiry Date: February 2, 2024
NOX Cal Gas Conc: 48.80 ppm NO Cal Gas Conc: 48.80 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.80 ppm Removed Gas NO Conc: 48.80 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3060
ZAG make/model: Teledyne API T701H Serial Number: 260

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1426262592
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.813	1.933	NO bkgnd or offset:	7.6	8.1
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	7.7	8.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	268.4	268.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995016	0.999486
NO _x Cal Offset:	-0.320000	-0.060000
NO Cal Slope:	0.998443	1.002670
NO Cal Offset:	-0.920000	-1.000000
NO ₂ Cal Slope:	0.994665	0.991397
NO ₂ Cal Offset:	-1.310360	0.179814



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as found span	4918	82.0	800.3	800.3	0.0	751.4	749.5	1.9	1.0651	1.0678
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
high point	4918	82.0	800.3	800.3	0.0	800.0	802.2	-2.2	1.0004	0.9977
second point	4959	41.0	400.2	400.2	0.0	399.5	399.1	0.3	1.0017	1.0027
third point	4980	20.5	200.1	200.1	0.0	200.1	198.9	1.1	0.9999	1.0059
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
as left span	4918	82.0	800.3	419.6	380.7	799.5	422.3	377.4	1.0010	0.9937
Average Correction Factor									1.0007	1.0021

Corrected As found	NO _x = 751.5 ppb	NO = 749.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -5.9%	
Previous Response	NO _x = 796.0 ppb	NO = 798.2 ppb		*Percent Change	NO = -6.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	801.2	420.5	380.7	376.8	1.0104	99.0%
2nd GPT point (200 ppb O3)	801.2	617.2	184.0	184.8	0.9957	100.4%
3rd GPT point (100 ppb O3)	801.2	712.7	88.5	86.8	1.0196	98.1%
Average Correction Factor					1.0085	99.2%

Notes:

Remote Calibration done after filter changed. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 15, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:15	End Time (MST):	11:34
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999999	≥0.995
800.3	800.0	1.0004			
400.2	399.5	1.0017			
200.1	200.1	0.9999			
			Slope	0.999486	0.90 - 1.10
			Intercept	-0.060000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 15, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:15	End Time (MST):	11:34
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.3	802.2	0.9977			
400.2	399.1	1.0027			
200.1	198.9	1.0059			
			Slope	1.002670	0.90 - 1.10
			Intercept	-1.000000	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 15, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:15	End Time (MST):	11:34
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>			
0.0	-0.1	----	Correlation Coefficient	0.999913	≥0.995			
380.7	376.8	1.0104						
184.0	184.8	0.9957				Slope	0.991397	0.90 - 1.10
88.5	86.8	1.0196						
			Intercept	0.179814	+/-20			



NO_x Calibration Plot

Date: June 27, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: June 28, 2023 Last Cal Date: May 10, 2023
 Start time (MST): 13:50 End time (MST): 16:07
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 260

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3872
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007686	1.012486	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	-2.020000	-1.460000	Coeff or Slope:	1.036	1.036

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	5000	913.0	400.0	404.4	0.989
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.6	----
high point	5000	914.7	400.0	404.6	0.989
second point	5000	786.4	200.0	199.9	1.001
third point	5000	701.3	100.0	97.8	1.022
as left zero	5000	0.0	0.0	0.7	----
as left span	5000	963.3	400.0	404.9	0.988
Average Correction Factor					1.004

Baseline Corr As found:	404.3	Previous response	401.1	*% change	0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: remote calibration done. No adjustments

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	May 10, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	13:50	End Time (MST):	16:07
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999882	≥0.995
400.0	404.6	0.9886			
200.0	199.9	1.0005	Slope	1.012486	0.90 - 1.10
100.0	97.8	1.0225			
			Intercept	-1.460000	+/- 5



O₃ Calibration Plot

Date: June 28, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
 Calibration Date: June 29, 2023 Last Cal Date: May 29, 2023
 Start time (MST): 14:20 End time (MST): 14:59

Analyzer Make: API S/N: 216
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1212
 Temp/RH standard: Delta Cal S/N: 1212

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	30.0	30.0	30.0	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	729.6	721.3	729.6	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	4.94	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>June 29, 2023</u>		Last Cal Date: <u>May 29, 2023</u>		
	PM w/o HEPA: <u>0</u>		PM w/ HEPA: <u>0.1</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>8.4</u>	w/ HEPA: <u>0.0</u>		<0.2 ug/m3
Date Optical Chamber Cleaned:		<u>March 13, 2023</u>			
Disposable Filter Changed:		<u>March 13, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: July 14, 2022
 Date RH/T Sensor Cleaned: July 14, 2022

Notes: No adjustments made.

Calibration by: Matthew Courtoreille,



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	June 30, 2023	Last Cal Date:	May 17, 2023
Start time (MST):	10:32	End time (MST):	13:34
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API T701H		Serial Number:	260

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999682	0.988086	Backgd or Offset:	-0.014	-0.014
Calibration intercept:	0.126989	0.136987	Coeff or Slope:	0.968	0.968

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.15	----
as found span	4933	66.7	40.4	40.0	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4934	66.7	40.4	40.0	1.010
second point	4967	33.3	20.2	20.2	0.998
third point	4983	16.7	10.1	10.1	1.005
as left zero	5000	0.0	0.0	-0.1	----
as left span	2960	40.0	40.4	40.1	1.009
Average Correction Factor					1.004

Baseline Corr As found:	40.15	Prev response:	40.54	*% change:	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Remote calibration, sample inlet filter changed out before calibration. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 30, 2023	Previous Calibration:	May 17, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	10:32	End Time (MST):	13:34
Analyzer make:	API T300	Analyzer serial #:	3505

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999963
40.4	40.0	1.0101		
20.2	20.2	0.9975	Slope	0.988086
10.1	10.1	1.0050		
			Intercept	0.136987
				± 1.5



CO Calibration Plot

Date: June 30, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	June 28, 2023	Last Cal Date:	May 17, 2023
Start time (MST):	11:32	End time (MST):	13:51
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3060
N2 Gen Make/Model:	NG 5000		Serial Number:	771048318

Analyzer Information

Analyzer make: Teledyne API T360	Analyzer serial #: 289
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996425	0.986205	Backgd or Offset:	-0.010	0.008
Calibration intercept:	-0.680000	-4.300000	Coeff or Slope:	1.018	1.002

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-1.8	----
as found span	2920	80.0	1605.9	1583.0	1.014
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	-0.1	----
high point	2920	80.0	1605.9	1585.0	1.013
second point	2960	40.0	802.9	775.0	1.036
third point	2980	20.0	401.5	394.4	1.018
as left zero	3000	0.0	0.0	0.3	----
as left span					
Average Correction Factor					1.022

Baseline Corr As found:	1584.80	Prev response:	1599.45	*% change:	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Remote calibration done. Zero adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	June 28, 2023	Previous Calibration	May 17, 2023
Station Name	Fort Chipewyan	Station Number	AMS08
Start Time (MST)	11:32	End Time (MST)	13:51
Analyzer make	Teledyne API T360	Analyzer serial #	289

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999844	≥0.995
1605.9	1585.0	1.0132	Slope	0.986205	0.90 - 1.10
802.9	775.0	1.0360	Intercept	-4.300000	+/-20
401.5	394.4	1.0179			



CO₂ Calibration Plot

Date: June 28, 2023

Location: Fort Chipewyan





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	June 8, 2023	Last Cal Date:	May 8, 2023
Start time (MST):	8:43	End time (MST):	11:50
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.96	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC151285			
Removed Cal Gas Conc:	49.96	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3812
ZAG Make/Model:	API T701		Serial Number:	4888

Analyzer Information

Analyzer make: Thermo 43i	Analyzer serial #: 1118148498
Analyzer Range 0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999100	0.995921	Backgd or Offset:	9.9	9.6
Calibration intercept:	-0.528972	0.510287	Coeff or Slope:	0.986	0.972

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4919	80.2	801.5	809.6	0.990
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4919	80.2	801.5	798.7	1.003
second point	4959	40.1	400.8	399.6	1.003
third point	4980	20.0	199.8	199.7	1.001
as left zero	5000	0.0	0.0	0.3	----
as left span	4919	80.2	801.5	799.6	1.002
Average Correction Factor					1.002

Baseline Corr As found:	809.40	Previous response	800.24	*% change	1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 8, 2023	Previous Calibration:	May 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:43	End Time (MST):	11:50
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.4	----	Correlation Coefficient	1.000000	
801.5	798.7	1.0035			≥0.995
400.8	399.6	1.0029	Slope	0.995921	
199.8	199.7	1.0007			0.90 - 1.10
			Intercept	0.510287	+/-30



SO2 Calibration Plot

Date: June 8, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Barge Landing Station number: AMS09
 Calibration Date: June 7, 2023 Last Cal Date: May 3, 2023
 Start time (MST): 8:05 End time (MST): 12:09
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.87 ppm Cal Gas Exp Date: September 2, 2024
 Cal Gas Cylinder #: EY0002346
 Removed Cal Gas Conc: 4.87 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3812
 ZAG Make/Model: API T701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1331259320
 Converter make: CDN-101 Converter serial #: 519
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005434	1.007720	Backgd or Offset: 2.82	2.77
Calibration intercept:	0.019021	-0.060968	Coeff or Slope: 1.147	1.134

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4918	82.1	80.0	79.2	1.008
as found 2nd point	4959	41.1	40.0	39.3	1.016
as found 3rd point	4979	20.5	20.0	19.7	1.009
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4918	82.1	80.0	80.5	0.993
second point	4959	41.1	40.0	40.3	0.993
third point	4979	20.5	20.0	20.1	0.993
as left zero	5000	0.0	0.0	-0.1	----
as left span	4918	82.1	80.0	81.8	0.978
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:	28-Feb-23			Ave Corr Factor	0.993
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.3 Prev response: 80.42 *% change: -1.4%
 Baseline Corr 2nd AF pt: 39.4 AF Slope: 0.991283 AF Intercept: -0.160828
 Baseline Corr 3rd AF pt: 19.8 AF Correlation: 0.999981

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SOx scrubber check done after calibrator zero. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 7, 2023	Previous Calibration:	May 3, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:05	End Time (MST):	12:09
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999999	
80.0	80.5	0.9933			≥0.995
40.0	40.3	0.9933	Slope	1.007720	
20.0	20.1	0.9935			0.90 - 1.10
			Intercept	-0.060968	+/-3



TRS Calibration Plot

Date: June 7, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	June 8, 2023	Last Cal Date:	May 8, 2023
Start time (MST):	8:43	End time (MST):	11:50
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 100 ppm		
NMHC Range (ppm):	0 - 50 ppm	CH ₄ Range (ppm):	0 - 50 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.50E-04	2.49E-04	NMHC SP Ratio:	4.87E-05
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	187500
				188395

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	17.23	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.16	0.998
second point	4960	40.1	8.56	8.58	0.997
third point	4980	20.0	4.27	4.30	0.994
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.22	0.994

				Average Correction Factor	0.996
Baseline Corr AF:	17.23	Prev response	17.17	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.0	----
as found span	4919	80.2	9.14	9.20	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.16	0.998
second point	4960	40.1	4.57	4.59	0.995
third point	4980	20	2.28	2.30	0.991
as left zero	5000	0	0.00	0.00	----
as left span	4919	80.2	9.14	9.20	0.994
Average Correction Factor					0.995
Baseline Corr AF:	9.20	Prev response	9.18	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	8.03	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	8.00	0.997
second point	4960	40.1	3.99	3.99	0.999
third point	4980	20.0	1.99	2.00	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	8.02	0.995
Average Correction Factor					0.998
Baseline Corr AF:	8.03	Prev response	7.99	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.003421	1.002132
THC Cal Offset:	-0.007132	0.007266
CH ₄ Cal Slope:	1.001195	1.002397
CH ₄ Cal Offset:	-0.006136	-0.001332
NMHC Cal Slope:	1.004627	1.001775
NMHC Cal Offset:	0.001202	0.008598

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 8, 2023	Previous Calibration:	May 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:43	End Time (MST):	11:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
17.12	17.16	0.9976						
8.56	8.58	0.9971				Slope	1.002132	0.90 - 1.10
4.27	4.30	0.9938						
			Intercept	0.007266	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 8, 2023	Previous Calibration:	May 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:43	End Time (MST):	11:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
7.98	8.00	0.9975						
3.99	3.99	0.9994				Slope	1.002397	0.90 - 1.10
1.99	2.00	0.9972						
			Intercept	-0.001332	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 8, 2023	Previous Calibration:	May 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:43	End Time (MST):	11:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
9.14	9.16	0.9978						
4.57	4.59	0.9953				Slope	1.001775	0.90 - 1.10
2.28	2.30	0.9909						
			Intercept	0.008598	± 0.5			



NMHC Calibration Plot

Date: June 8, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Barge Landing
Calibration Date: June 6, 2023
Start time (MST): 8:21
Reason: Routine
Station number: AMS09
Last Cal Date: May 23, 2023
End time (MST): 12:50

Calibration Standards

NO Gas Cylinder #: DT0036634
NOX Cal Gas Conc: 50.00 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 50.00 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: January 28, 2024
NO Cal Gas Conc: 49.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 49.70 ppm
NO gas Diff:
Serial Number: 3812
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1426262593

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.175	1.181	NO bkgnd or offset:	10.7	10.8
NOX coeff or slope:	0.995	0.994	NOX bkgnd or offset:	11.0	11.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	174.6	172.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003622	0.998838
NO _x Cal Offset:	0.329752	0.888950
NO Cal Slope:	1.005383	1.001441
NO Cal Offset:	-0.991769	-0.512202
NO ₂ Cal Slope:	0.999175	0.999242
NO ₂ Cal Offset:	0.249142	0.437176



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.2	0.3	----	----
as found span	4919	80.5	805.1	800.3	4.8	804.3	795.5	8.7	1.001	1.006
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	-0.1	0.3	----	----
high point	4919	80.5	805.1	800.3	4.8	804.6	801.0	3.7	1.001	0.999
second point	4959	40.2	402.1	399.7	2.4	403.1	399.8	3.3	0.997	1.000
third point	4979	20.1	201.0	199.8	1.2	202.2	199.0	3.2	0.994	1.004
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as left span	4919	80.5	805.1	439.8	365.3	805.2	439.7	365.5	1.000	1.000
Average Correction Factor									0.997	1.001

Corrected As found	NO _x = 804.2 ppb	NO = 795.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.5%	
Previous Response	NO _x = 808.3 ppb	NO = 803.6 ppb		*Percent Change	NO = -1.0%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.3	436.8	365.3	365.4	1.000	100.0%
2nd GPT point (200 ppb O3)	797.3	660.1	142.0	142.6	0.996	100.4%
3rd GPT point (100 ppb O3)	797.3	728.5	73.6	74.0	0.995	100.5%
Average Correction Factor					0.997	100.3%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 6, 2023	Previous Calibration:	May 23, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:21	End Time (MST):	12:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
805.1	804.6	1.0006		
402.1	403.1	0.9974		
201.0	202.2	0.9942		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 6, 2023	Previous Calibration:	May 23, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:21	End Time (MST):	12:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	801.0	0.9991		
399.7	399.8	0.9996		
199.8	199.0	1.0042		
			0.999998	
			1.001441	
			-0.512202	





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 6, 2023	Previous Calibration:	May 23, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:21	End Time (MST):	12:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995			
365.3	365.4	0.9998						
142.0	142.6	0.9960				Slope	0.999242	0.90 - 1.10
73.6	74.0	0.9950						
			Intercept	0.437176	+/-20			



NO_x Calibration Plot

Date: June 6, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Barge Landing Station number: AMS 09
Calibration Date: June 7, 2023 Last Cal Date: May 23, 2023
Start time (MST): 9:10 End time (MST): 9:39
Analyzer Make: API T640 S/N: 321
Particulate Fraction: PM2.5
Flow Meter Make/Model: DeltaCal S/N: 1451
Temp/RH standard: DeltaCal S/N: 1451

Monthly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Measured</u>	<u>As left</u>	<u>Adjusted</u>	<u>(Limits)</u>
T (°C)	21.1	22.30	21.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	728.1	729	728.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.04	5.06	5.04	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: June 7, 2023	Last Cal Date: May 23, 2023			
	PM w/o HEPA: 28.0	PM w/ HEPA: 0.0			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Post maintenance</u>	<u>As left</u>	<u>Adjusted</u>	<u>(Limits)</u>
PMT Peak Test				<input type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		May 23, 2023			<0.2 ug/m3
Disposable Filter Changed:		May 23, 2023			

Annual Maintenance

Date Sample Tube Cleaned: November 15, 2022
Date RH/T Sensor Cleaned: November 15, 2022

Notes: Inlet head looks good. No adjustments made. Leak check passed.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS11 LOWER CAMP JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	June 27, 2023	Last Cal Date:	May 15, 2023
Start time (MST):	9:08	End time (MST):	11:51
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.25	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC2216			
Removed Cal Gas Conc:	49.25	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3807
ZAG Make/Model:	Teledyne API T701		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	100841398
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997114	1.004834	Backgd or Offset:	14.6	14.7
Calibration intercept:	-0.568654	-0.648706	Coeff or Slope:	1.034	1.034

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4919	81.3	800.8	802.6	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.3	----
high point	4919	81.3	800.8	804.9	0.995
second point	4959	40.7	400.9	400.4	1.001
third point	4980	20.3	199.9	200.2	0.999
as left zero	5000	0.0	0.0	0.4	----
as left span	4919	81.3	800.8	805.9	0.994
Average Correction Factor					0.998

Baseline Corr As found:	802.50	Previous response	797.88	*% change	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 15, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:08	End Time (MST):	11:51
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999986	
800.8	804.9	0.9949			≥0.995
400.9	400.4	1.0013	Slope	1.004834	
199.9	200.2	0.9987			0.90 - 1.10
			Intercept	-0.648706	+/-30



SO2 Calibration Plot

Date: June 27, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	June 28, 2023	Last Cal Date:	May 16, 2023
Start time (MST):	10:30	End time (MST):	13:59
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.429	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC501097			
Removed Cal Gas Conc:	5.429	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3807
ZAG Make/Model:	API T701H		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003
Converter make:	NA	Converter serial #:	NA
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004895	1.012197	Backgd or Offset:	13.3 13.6
Calibration intercept:	0.035134	0.054446	Coeff or Slope:	1.001 1.001

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4926	73.6	79.9	81.5	0.978
as found 2nd point	4963	36.8	40.0	40.8	0.975
as found 3rd point	4982	18.6	20.2	20.1	0.995
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	73.6	79.9	80.9	0.988
second point	4963	36.8	40.0	40.7	0.982
third point	4982	18.6	20.2	20.3	0.995
as left zero	5000	0.0	0.0	0.2	----
as left span	4926	73.6	79.9	81.4	0.982
SO2 Scrubber Check	4919	81.1	811.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.988
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	81.7	Prev response:	80.35	*% change:	1.7%
Baseline Corr 2nd AF pt:	41.0	AF Slope:	1.023929	AF Intercept:	-0.306388
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999967		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	May 16, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:30	End Time (MST):	13:59
Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999977	
79.9	80.9	0.9879			≥0.995
40.0	40.7	0.9818	Slope	1.012197	
20.2	20.3	0.9948			0.90 - 1.10
			Intercept	0.054446	+/-3



H₂S Calibration Plot

Date: June 28, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	June 27, 2023	Last Cal Date:	May 15, 2023
Start time (MST):	9:08	End time (MST):	11:51
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.5 ppm		
Diff between cyl (CH ₄):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701	Serial Number:	3807
		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.99E-04	2.99E-04	NMHC SP Ratio:	5.79E-05
CH ₄ Retention time:	14.0	14.0	NMHC Peak Area:	158468
				158468

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	17.35	17.33	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	17.35	17.30	1.003
second point	4959	40.7	8.69	8.60	1.010
third point	4980	20.3	4.33	4.30	1.008
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	17.35	17.43	0.995

Average Correction Factor				1.007
Baseline Corr AF:	17.33	Prev response	17.29	*% change 0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	9.19	9.18	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	9.19	9.17	1.002
second point	4959	40.7	4.60	4.57	1.008
third point	4980	20.3	2.29	2.29	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	9.19	9.24	0.994
Average Correction Factor					1.004
Baseline Corr AF:	9.18	Prev response	9.16	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	8.16	8.14	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	8.16	8.13	1.004
second point	4959	40.7	4.09	4.04	1.012
third point	4980	20.3	2.04	2.01	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	8.16	8.19	0.996
Average Correction Factor					1.009
Baseline Corr AF:	8.14	Prev response	8.12	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.996470	0.997114
THC Cal Offset:	-0.001598	-0.018987
CH ₄ Cal Slope:	0.994977	0.996346
CH ₄ Cal Offset:	-0.000697	-0.012088
NMHC Cal Slope:	0.997386	0.997920
NMHC Cal Offset:	-0.000502	-0.006899

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 15, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:08	End Time (MST):	11:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995			
17.35	17.30	1.0029						
8.69	8.60	1.0096				Slope	0.997114	0.90 - 1.10
4.33	4.30	1.0075						
			Intercept	-0.018987	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 15, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:08	End Time (MST):	11:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999981	≥0.995
8.16	8.13	1.0038			
4.09	4.04	1.0115			
2.04	2.01	1.0119			
			Slope	0.996346	0.90 - 1.10
			Intercept	-0.012088	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 15, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:08	End Time (MST):	11:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999989	≥ 0.995			
9.19	9.17	1.0019						
4.60	4.57	1.0078				Slope	0.997920	0.90 - 1.10
2.29	2.29	1.0036						
			Intercept	-0.006899	± 0.5			



NMHC Calibration Plot

Date: June 27, 2023

Location: Lower Camp





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13 FORT MCKAY SOUTH JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
Calibration Date: June 26, 2023 Last Cal Date: May 17, 2023
Start time (MST): 9:42 End time (MST): 12:35
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.55 ppm Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #: CC260812
Removed Cal Gas Conc: 50.55 ppm Rem Gas Exp Date: N/A
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 2448
ZAG Make/Model: API 701 Serial Number: 1117

Analyzer Information

Analyzer make: API T100 Analyzer serial #: 599
Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001326	1.001127	Backgd or Offset:	80.5	86.3
Calibration intercept:	-1.757849	-2.638171	Coeff or Slope:	0.725	0.714

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	1.6	----
as found span	4921	79.1	799.7	812.0	0.985
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4921	79.1	799.7	799.0	1.001
second point	4961	39.5	399.3	396.3	1.008
third point	4980	19.8	200.2	195.2	1.026
as left zero	5000	0.0	0.0	-0.2	----
as left span	4921	79.1	799.7	798.8	1.001
Average Correction Factor					1.011

Baseline Corr As found:	810.40	Previous response	798.99	*% change	1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 26, 2023	Previous Calibration:	May 17, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	9:42	End Time (MST):	12:35
Analyzer make:	API T100	Analyzer serial #:	599

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999960	≥0.995
799.7	799.0	1.0009			
399.3	396.3	1.0076	Slope	1.001127	0.90 - 1.10
200.2	195.2	1.0255			
			Intercept	-2.638171	+/-30



SO2 Calibration Plot

Date: June 26, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	June 29, 2023	Last Cal Date:	May 2, 2023
Start time (MST):	7:09	End time (MST):	11:08
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.34	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC500241			
Removed Cal Gas Conc:	5.34	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2448
ZAG Make/Model:	Teledyne API 701		Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017
Converter make:	CDN-101	Converter serial #:	521
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005466	1.004617	Backgd or Offset:	3.68	3.54
Calibration intercept:	-0.262184	-0.182243	Coeff or Slope:	1.116	1.066

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4925	75.5	80.6	80.7	0.998
as found 2nd point	4962	37.7	40.3	40.2	0.999
as found 3rd point	4981	18.9	20.2	20.1	0.999
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4925	75.5	80.6	80.9	0.997
second point	4962	37.7	40.3	40.2	1.002
third point	4981	18.9	20.2	19.9	1.014
as left zero	5000	0.0	0.0	0.2	----
as left span	4925	75.5	80.6	80.7	0.999
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:	20-Mar-20		Ave Corr Factor		1.004
Date of last converter efficiency test:	NA		efficiency		

Baseline Corr As found:	80.8	Prev response:	80.80	*% change:	0.0%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.002208	AF Intercept:	-0.122275
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999999		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021


Station Information

Calibration Date:	June 29, 2023	Previous Calibration:	May 2, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	7:09	End Time (MST):	13:36
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999977	
80.6	80.9	0.9966			≥0.995
40.3	40.2	1.0016	Slope	1.004617	
20.2	19.9	1.0144			0.90 - 1.10
			Intercept	-0.182243	+/-3







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	June 23, 2023	Last Cal Date:	May 17, 2023
Start time (MST):	8:30	End time (MST):	9:58
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API 701	Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.21E-04	2.21E-04	NMHC SP Ratio:	5.06E-04
CH ₄ Retention time:	12.8	12.8	NMHC Peak Area:	179419

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	17.05	16.95	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	16.90	1.009
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.009
Baseline Corr AF:	16.95	Prev response	16.98	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.1	9.08	9.04	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.1	9.08	9.00	1.009
second point	4960				
third point					
as left zero					
as left span	4921	79.1	9.08		
Average Correction Factor					1.009
Baseline Corr AF:	9.04	Prev response	9.04	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	7.97	7.91	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.90	1.008
second point	4960				
third point					
as left zero					
as left span	4921	79.1	7.97		
Average Correction Factor					1.008
Baseline Corr AF:	7.91	Prev response	7.94	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000879	0.991488
THC Cal Offset:	-0.078771	0.000000
CH ₄ Cal Slope:	1.001767	0.992118
CH ₄ Cal Offset:	-0.042148	0.000000
NMHC Cal Slope:	0.999952	0.991044
NMHC Cal Offset:	-0.036961	0.000000

Notes:


Changed cylinder after as founds.

Calibration Performed By: Sean Bala

NMHC Calibration Plot

Date: June 23, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	June 26, 2023	Last Cal Date:	May 17, 2023
Start time (MST):	9:42	End time (MST):	12:35
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API 701	Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.21E-04	2.22E-04	NMHC SP Ratio:	5.06E-04
CH ₄ Retention time:	12.8	13.0	NMHC Peak Area:	179419
				177982

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	17.05	16.95	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.02	1.001
second point	4961	39.5	8.51	8.38	1.016
third point	4980	19.8	4.27	4.12	1.037
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	17.05	17.04	1.000

Average Correction Factor	1.018
---------------------------	-------

Baseline Corr AF:	16.95	Prev response	16.98	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF Limit= 0.95-1.05
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.1	9.08	9.04	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.1	9.08	9.06	1.002
second point	4960	39.5	4.53	4.49	1.011
third point	4980	19.8	2.27	2.20	1.032
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.1	9.08	9.08	0.999
Average Correction Factor					1.015
Baseline Corr AF:	9.04	Prev response	9.04	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF Limit= 0.95-1.05
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	7.97	7.90	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.96	1.001
second point	4960	39.5	3.98	3.89	1.022
third point	4980	19.8	1.99	1.91	1.043
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	7.97	7.96	1.001
Average Correction Factor					1.022
Baseline Corr AF:	7.90	Prev response	7.94	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000879	1.000853
THC Cal Offset:	-0.078771	-0.082577
CH ₄ Cal Slope:	1.001767	1.001309
CH ₄ Cal Offset:	-0.042148	-0.048553
NMHC Cal Slope:	0.999952	1.000342
NMHC Cal Offset:	-0.036961	-0.034760

Notes:

Changed inlet filter after as founds. Adjusted span.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 26, 2023	Previous Calibration:	May 17, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	9:42	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999891	≥ 0.995			
17.05	17.02	1.0013						
8.51	8.38	1.0156				Slope	1.000853	0.90 - 1.10
4.27	4.12	1.0370						
			Intercept	-0.082577	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 26, 2023	Previous Calibration:	May 17, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	9:42	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999816	≥0.995
7.97	7.96	1.0010			
3.98	3.89	1.0220			
1.99	1.91	1.0431			
			Slope	1.001309	0.90 - 1.10
			Intercept	-0.048553	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 26, 2023	Previous Calibration:	May 17, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	9:42	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999933	≥ 0.995
9.08	9.06	1.0017			
4.53	4.49	1.0106			
2.27	2.20	1.0322			
			Slope	1.000342	0.90 - 1.10
			Intercept	-0.034760	+/-0.5



NMHC Calibration Plot

Date: June 26, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort McKay South
Calibration Date: June 5, 2023
Start time (MST): 9:14
Reason: Routine
Station number: AMS 13
Last Cal Date: May 30, 2023
End time (MST): 13:39

Calibration Standards

NO Gas Cylinder #: T2Y1P76
NOX Cal Gas Conc: 50.98 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 50.98 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 49.32 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 49.32 ppm
NO gas Diff:
Serial Number: 2448
Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1410661329

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.168	1.095	NO bkgnd or offset:	9.2	8.9
NOX coeff or slope:	0.988	0.991	NOX bkgnd or offset:	9.6	10.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	130.9	158.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998875	0.997811
NO _x Cal Offset:	-1.831289	-1.870975
NO Cal Slope:	1.003248	1.002005
NO Cal Offset:	-2.765445	-2.905185
NO ₂ Cal Slope:	0.991692	0.997514
NO ₂ Cal Offset:	-1.272971	0.257066



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.1	0.1	1.1	----	----
as found span	4919	81.1	826.9	800.0	26.9	881.2	854.6	26.3	0.9384	0.9361
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	-0.2	0.3	----	----
high point	4919	81.1	826.9	800.0	26.9	824.2	800.0	24.1	1.0033	0.9999
second point	4960	40.6	413.9	400.4	13.5	410.1	397.0	13.0	1.0093	1.0086
third point	4980	20.3	207.0	200.2	6.7	202.7	195.0	7.7	1.0210	1.0268
as left zero	5000	0.0	0.0	0.0	0.0	-1.0	-0.2	-0.8	----	----
as left span	4919	80.1	816.8	374.0	442.8	830.8	384.4	446.4	0.9832	0.9730
Average Correction Factor									1.0112	1.0118

Corrected As found	NO _x = 880.1 ppb	NO = 854.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 6.4%
Previous Response	NO _x = 824.1 ppb	NO = 799.8 ppb		*Percent Change	NO = 6.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:
					NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	796.5	380.6	442.8	441.8	1.0023	99.8%
2nd GPT point (200 ppb O ₃)	796.5	590.2	233.2	233.5	0.9988	100.1%
3rd GPT point (100 ppb O ₃)	796.5	694.2	129.2	128.7	1.0041	99.6%
Average Correction Factor					1.0017	99.8%

Notes: Adjusted zero and span. Used 2nd NO reference point due to drift.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 30, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:14	End Time (MST):	13:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
826.9	824.2	1.0033			
413.9	410.1	1.0093			
207.0	202.7	1.0210			
			Slope	0.997811	0.90 - 1.10
			Intercept	-1.870975	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 30, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:14	End Time (MST):	13:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999948	≥0.995
800.0	800.0	0.9999			
400.4	397.0	1.0086	Slope	1.002005	0.90 - 1.10
200.2	195.0	1.0268			
			Intercept	-2.905185	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 30, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:14	End Time (MST):	13:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	≥0.995	
442.8	441.8	1.0023			
233.2	233.5	0.9988			
129.2	128.7	1.0041			
			Slope	0.997514	0.90 - 1.10
			Intercept	0.257066	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: June 5, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: June 2, 2023 Last Cal Date: May 8, 2023
 Start time (MST): 8:53 End time (MST): 11:52
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3871
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003343	0.997857	Backgd or Offset:	3.4	2.4
Calibration intercept:	0.440000	1.000000	Coeff or Slope:	0.966	0.963

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-1.2	----
as found span	5000	969.9	400.0	397.9	1.005
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.4	----
high point	5000	980.6	400.0	399.8	1.001
second point	5000	838.0	200.0	201.0	0.995
third point	5000	735.3	100.0	101.3	0.987
as left zero	5000	0.0	0.0	-0.1	----
as left span	5000	979.1	400.0	401.6	0.996
Average Correction Factor					0.994

Baseline Corr As found:	399.1	Previous response	401.8	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Adjusted zero only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 2, 2023	Previous Calibration:	May 8, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:53	End Time (MST):	11:52
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.4	----	Correlation Coefficient	0.999989	
400.0	399.8	1.0005			≥0.995
200.0	201.0	0.9950	Slope	0.997857	
100.0	101.3	0.9872			0.90 - 1.10
			Intercept	1.000000	+/- 5



O₃ Calibration Plot

Date: June 2, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	June 29, 2023	Last Cal Date:	May 9, 2023
Start time (MST):	9:31	End time (MST):	10:31
Analyzer Make:	API T640	S/N:	319
Particulate Fraction:	PM2.5		
Flow Meter Make/Model:	Delta Cal	S/N:	141229
Temp/RH standard:	Delta Cal	S/N:	141229

Monthly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Measured</u>	<u>As left</u>	<u>Adjusted</u>	<u>(Limits)</u>
T (°C)	25.0	25.8	25.0	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	730.1	731.0	730.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.98	4.90	4.98	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>June 29, 2023</u>	Last Cal Date: <u>May 9, 2023</u>			
	PM w/o HEPA: <u>47.6</u>	PM w/ HEPA: <u>0.1</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Post maintenance</u>	<u>As left</u>	<u>Adjusted</u>	<u>(Limits)</u>
PMT Peak Test	9.0	10.9	10.9	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>4.4</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>June 29, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>June 29, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned:	<u>June 29, 2023</u>
Date RH/T Sensor Cleaned:	<u>June 29, 2023</u>

Inlet head clean and inspected. PMT peak test was adjusted.

Notes:

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC
JUNE 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 2, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:50	End Time (MST):	13:01
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999993	
800.2	813.2	0.9840			≥0.995
399.6	405.1	0.9864	Slope	1.016533	
199.8	201.4	0.9921			0.90 - 1.10
			Intercept	-0.680611	+/-30



SO2 Calibration Plot

Date: June 5, 2023

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Anzac Station number: AMS14
 Calibration Date: June 12, 2023 Last Cal Date: May 11, 2023
 Start time (MST): 9:47 End time (MST): 14:17
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.38 ppm Cal Gas Exp Date: February 3, 2023
 Cal Gas Cylinder #: EY0000859
 Removed Cal Gas Conc: 5.38 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 5239
 ZAG Make/Model: API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1180540019
 Converter make: CD Nova CDN-101 Converter serial #: 503
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997841	0.989547	Backgd or Offset: 5.36	5.39
Calibration intercept:	0.098854	0.039166	Coeff or Slope: 0.972	0.972

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4925	74.3	80.0	79.7	1.004
as found 2nd point	4962	37.2	40.0	39.5	1.016
as found 3rd point	4981	18.6	20.0	19.5	1.032
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4925	74.3	80.0	79.3	1.008
second point	4962	37.2	40.0	39.5	1.014
third point	4981	18.6	20.0	19.6	1.021
as left zero	5000	0.0	0.0	0.3	----
as left span	4925	74.3	80.0	79.6	1.004
SO2 Scrubber Check	4920	80.0	800.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	1.014
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 79.6 Prev response: 79.88 *% change: -0.4%
 Baseline Corr 2nd AF pt: 39.4 AF Slope: 0.997122 AF Intercept: -0.200957
 Baseline Corr 3rd AF pt: 19.4 AF Correlation: 0.999933

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 12, 2023	Previous Calibration:	May 11, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	9:47	End Time (MST):	14:17
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540019

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999950	≥0.995
80.0	79.3	1.0083			
40.0	39.5	1.0135	Slope	0.989547	0.90 - 1.10
20.0	19.6	1.0212			
			Intercept	0.039166	+/-3



TRS Calibration Plot

Date: June 12, 2023

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	June 5, 2023	Last Cal Date:	May 2, 2023
Start time (MST):	9:50	End time (MST):	13:01
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	5239
ZAG make/model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.90E-04	3.90E-04	NMHC SP Ratio:	4.53E-05
CH ₄ Retention time:	12.20	12.20	NMHC Peak Area:	201206

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	17.12	17.09	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	17.12	17.05	1.004
second point	4960	40.0	8.55	8.57	0.998
third point	4980	20.0	4.28	4.26	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.1	17.12	17.18	0.997

				Average Correction Factor	1.002
Baseline Corr AF:	17.09	Prev response	17.05	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	9.12	9.08	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	9.12	9.08	1.005
second point	4960	40.0	4.56	4.56	1.000
third point	4980	20.0	2.28	2.26	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.1	9.12	9.12	1.001
Average Correction Factor					1.004
Baseline Corr AF:	9.08	Prev response	9.08	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	8.00	8.01	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	8.00	7.97	1.004
second point	4960	40.0	3.99	4.01	0.996
third point	4980	20.0	2.00	2.00	0.999
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.1	8.00	8.06	0.992
Average Correction Factor					0.999
Baseline Corr AF:	8.01	Prev response	7.97	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.996430	0.996014
THC Cal Offset:	-0.007805	0.012807
CH ₄ Cal Slope:	0.996993	0.996447
CH ₄ Cal Offset:	-0.002038	0.010369
NMHC Cal Slope:	0.995937	0.995647
NMHC Cal Offset:	-0.005767	0.002638

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 2, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:50	End Time (MST):	13:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999988	≥ 0.995			
17.12	17.05	1.0042						
8.55	8.57	0.9981				Slope	0.996014	0.90 - 1.10
4.28	4.26	1.0029						
			Intercept	0.012807	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 2, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:50	End Time (MST):	13:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999980	≥ 0.995			
8.00	7.97	1.0036						
3.99	4.01	0.9956				Slope	0.996447	0.90 - 1.10
2.00	2.00	0.9986						
			Intercept	0.010369	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 2, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:50	End Time (MST):	13:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999991	≥ 0.995			
9.12	9.08	1.0048						
4.56	4.56	1.0000				Slope	0.995647	0.90 - 1.10
2.28	2.26	1.0067						
			Intercept	0.002638	± 0.5			



NMHC Calibration Plot

Date: June 5, 2023

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name: Anzac Station number: AMS 14
 Calibration Date: June 19, 2023 Last Cal Date: June 5, 2023
 Start time (MST): 11:04 End time (MST): 13:33
 Reason: Cylinder Change Support gas cylinder change out.

Calibration Standards

Gas Cert Reference: CC279389 Cal Gas Expiry Date: January 5, 2025
 CH₄ Cal Gas Conc. 499.3 ppm CH₄ Equiv Conc. 1068.8 ppm
 C₃H₈ Cal Gas Conc. 207.1 ppm
 Removed Gas Cert: NA Removed Gas Expiry: NA
 Removed CH₄ Conc. 499.3 ppm CH₄ Equiv Conc. 1068.8 ppm
 Removed C₃H₈ Conc. 207.1 ppm
 Diff between cyl (CH₄): Diff between cyl (NM):
 Calibrator Model: API T700 Serial Number: 5239
 ZAG make/model: API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 1118148494
 THC Range (ppm): 0 - 20 ppm
 NMHC Range (ppm): 0 - 10 ppm CH₄ Range (ppm): 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.90E-04	NA	NMHC SP Ratio:	4.53E-05
CH ₄ Retention time:	12.20	NA	NMHC Peak Area:	201206
				NA

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	17.12	17.24	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	17.12	17.15	0.999
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	0.999
Baseline Corr AF:	17.24	Prev response	17.07	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	9.12	9.11	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	9.12	9.06	1.007
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.007
Baseline Corr AF:	9.11	Prev response	9.09	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	8.00	8.13	0.984
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	8.00	8.09	0.989
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.989
Baseline Corr AF:	8.13	Prev response	7.98	*% change	1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.996014	1.001446
THC Cal Offset:	0.012807	0.000000
CH ₄ Cal Slope:	0.996447	1.011299
CH ₄ Cal Offset:	0.010369	0.000000
NMHC Cal Slope:	0.995647	0.992809
NMHC Cal Offset:	0.002638	0.000000


Notes: Changed out the Hydrogen and Nitrogen cylinders after as founds. Second span run was comparable to the as found points. No changes were made onto the analyzer.

Calibration Performed By: Mohammed Kashif

NMHC Calibration Plot

Date: June 19, 2023

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Anzac
Calibration Date: June 8, 2023
Start time (MST): 9:34
Reason: Routine
Station number: AMS 14
Last Cal Date: May 1, 2023
End time (MST): 14:41

Calibration Standards

NO Gas Cylinder #: T2Y1P8D
NOX Cal Gas Conc: 50.92 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 50.92 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API 701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.05 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.05 ppm
NO gas Diff:
Serial Number: 5239
Serial Number: 357

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1426262592

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.361	1.375	NO bkgnd or offset:	3.7	3.7
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.7	3.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	162.7	162.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995080	1.000547
NO _x Cal Offset:	-0.466631	-0.105430
NO Cal Slope:	0.996609	1.001604
NO Cal Offset:	-1.790505	-1.609467
NO ₂ Cal Slope:	1.001233	1.002084
NO ₂ Cal Offset:	0.590521	1.232321



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.3	----	----
as found span	4921	78.6	800.5	786.8	13.7	792.9	776.1	16.8	1.0096	1.0138
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.4	0.0	0.4	----	----
high point	4921	78.6	800.5	786.8	13.7	800.9	787.1	13.9	0.9995	0.9997
second point	4961	39.3	400.2	393.4	6.8	400.7	392.2	8.5	0.9988	1.0030
third point	4980	19.6	199.6	196.2	3.4	198.7	192.9	5.8	1.0046	1.0172
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.0	0.3	----	----
as left span	4921	78.6	800.5	389.5	411.0	800.6	389.6	411.0	0.9999	0.9999
Average Correction Factor									1.0010	1.0066

Corrected As found	NO _x = 792.8 ppb	NO = 776.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.4%
Previous Response	NO _x = 796.1 ppb	NO = 782.4 ppb		*Percent Change	NO = -0.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	782.6	385.3	411.0	412.6	0.9961	100.4%
2nd GPT point (200 ppb O3)	782.6	582.3	214.0	216.1	0.9902	101.0%
3rd GPT point (100 ppb O3)	782.6	683.2	113.1	115.4	0.9799	102.1%
Average Correction Factor					0.9887	101.1%

Notes:

Sample inlet filter changed after as founds. Adjusted span only.

Mohammed Kashif

CALS_277



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 8, 2023	Previous Calibration:	May 1, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:34	End Time (MST):	14:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
800.5	800.9	0.9995			0.999996
400.2	400.7	0.9988			1.000547
199.6	198.7	1.0046			-0.105430





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 8, 2023	Previous Calibration:	May 1, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:34	End Time (MST):	14:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
786.8	787.1	0.9997		
393.4	392.2	1.0030		
196.2	192.9	1.0172		
			0.999979	
			1.001604	
			-1.609467	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 8, 2023	Previous Calibration:	May 1, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:34	End Time (MST):	14:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
411.0	412.6	0.9961		
214.0	216.1	0.9902		
113.1	115.4	0.9799		



NO_x Calibration Plot

Date: June 8, 2023

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 7, 2023	Previous Calibration:	May 2, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	9:59	End Time (MST):	12:47
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999988	
400.0	397.9	1.0053			≥0.995
200.0	200.2	0.9990	Slope	0.993171	
100.0	100.8	0.9921			0.90 - 1.10
			Intercept	1.020000	+/- 5



O₃ Calibration Plot

Date: June 7, 2023

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Anzac Station number: AMS 14
 Calibration Date: June 14, 2023 Last Cal Date: May 30, 2023
 Start time (MST): 11:10 End time (MST): 14:35

Analyzer Make: API T640 S/N: 825
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388744
 Temp/RH standard: Alicat FP-25 S/N: 388744

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-62.6	10.9	11.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	706.3	707.0	706.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.32	6.23	5.00	<input checked="" type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>June 14, 2023</u>	Last Cal Date: <u>May 30, 2023</u>			
	PM w/o HEPA: <u>9.0</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test		10.8	10.8	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>March 30, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>March 30, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: June 21, 2022
 Date RH/T Sensor Cleaned: June 21, 2022

Notes: As found pump operating at 100% capacity affected by ambient temperature readings that were significantly out, repaired temperature sensor and calibrated flow. Cleaned optics and performed dust test, no adjustments required. Did not do final leak check due to rain.

Calibration by: Kelly Baragar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS17
WAPASU
JUNE 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	June 5, 2023	Last Cal Date:	May 15, 2023
Start time (MST):	10:35	End time (MST):	13:36
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.38 ppm	Cal Gas Exp Date:	January 12, 2029
Cal Gas Cylinder #:	ALM066507		
Removed Cal Gas Conc:	50.38 ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	2449
ZAG Make/Model:	API 701H	Serial Number:	359

Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 1218153459
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998882	0.998368	Backgd or Offset:	12.4	12.4
Calibration intercept:	-2.039793	-2.660243	Coeff or Slope:	1.099	1.099

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4921	79.4	800.0	793.4	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	79.4	800.0	797.8	1.003
second point	4960	39.7	400.0	394.0	1.015
third point	4980	19.8	199.5	194.8	1.024
as left zero	5000	0.0	0.0	-0.2	----
as left span	4920	79.4	800.1	797.1	1.004
Average Correction Factor					1.014

Baseline Corr As found:	793.70	Previous response	797.04	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 15, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:35	End Time (MST):	13:36
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999941	≥0.995
800.0	797.8	1.0027			
400.0	394.0	1.0153	Slope	0.998368	0.90 - 1.10
199.5	194.8	1.0242			
			Intercept	-2.660243	+/-30



SO2 Calibration Plot

Date: June 5, 2023

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	June 21, 2023	Last Cal Date:	May 23, 2023
Start time (MST):	9:45	End time (MST):	14:47
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.076	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511852			
Removed Cal Gas Conc:	5.076	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2449
ZAG Make/Model:	API T701H		Serial Number:	359

Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583
Converter make:	n/a	Converter serial #:	n/a
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.987854	0.997854	Backgd or Offset:	12.7
Calibration intercept:	0.100766	-0.199211	Coeff or Slope:	1.085
				13.2
				1.115

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4921	78.8	80.0	79.5	1.008
as found 2nd point	4961	39.4	40.0	40.0	1.002
as found 3rd point	4980	19.7	20.0	19.8	1.015
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	78.8	80.0	79.8	1.003
second point	4961	39.4	40.0	39.5	1.013
third point	4980	19.7	20.0	19.5	1.026
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	78.8	80.0	77.7	1.030
SO2 Scrubber Check	4921	79.4	800.0	0.3	----
Date of last scrubber change:	n/a			Ave Corr Factor	1.014
Date of last converter efficiency test:	n/a				efficiency

Baseline Corr As found:	79.4	Prev response:	79.13	*% change:	0.3%
Baseline Corr 2nd AF pt:	39.9	AF Slope:	0.993425	AF Intercept:	0.080804
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999982		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 21, 2023	Previous Calibration:	May 23, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:45	End Time (MST):	14:47
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999934	
80.0	79.8	1.0025			≥0.995
40.0	39.5	1.0125	Slope	0.997854	
20.0	19.5	1.0257			0.90 - 1.10
			Intercept	-0.199211	+/-3



H₂S Calibration Plot

Date: June 21, 2023

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	June 5, 2023	Last Cal Date:	May 15, 2023
Start time (MST):	10:35	End time (MST):	13:36
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	ALM066507	Cal Gas Expiry Date:	January 12, 2029
CH4 Cal Gas Conc.	<u>503.5</u> ppm	CH4 Equiv Conc.	1076.3 ppm
C3H8 Cal Gas Conc.	<u>208.3</u> ppm		
Removed Gas Cert:	n/a	Removed Gas Expiry:	n/a
Removed CH4 Conc.	<u>503.5</u> ppm	CH4 Equiv Conc.	1076.3 ppm
Removed C3H8 Conc.	<u>208.3</u> ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	2449
ZAG Make/Model:	API 701H	Serial Number:	359

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1218153352
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004482	0.998402	Background:	3.070	3.210
Calibration intercept:	-0.015906	-0.031959	Coefficient:	4.296	4.483

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.02	----
as found span	4921	79.4	17.09	16.31	1.048
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.03	----
high point	4921	79.4	17.09	17.07	1.001
second point	4960	39.7	8.55	8.44	1.013
third point	4980	19.8	4.26	4.19	1.018
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.4	17.09	17.01	1.005
Average Correction Factor					1.011
Baseline Corr As found:	16.33	Previous response	17.15	*% change	-5.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 15, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:35	End Time (MST):	13:36
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.03	----	Correlation Coefficient	0.999935	≥0.995
17.09	17.07	1.0011			
8.55	8.44	1.0129	Slope	0.998402	0.90 - 1.10
4.26	4.19	1.0180			
			Intercept	-0.031959	+/-1.5



THC Calibration Plot

Date: June 5, 2023

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Wapasu
Calibration Date: June 6, 2023
Start time (MST): 9:55
Reason: Removal
Station number: AMS17
Last Cal Date: May 18, 2023
End time (MST): 12:54

Calibration Standards

NO Gas Cylinder #: T375YK8
NOX Cal Gas Conc: 49.11 ppm
Removed Cylinder #: T375YK8
Removed Gas NOX Conc: 49.11 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: April 13, 2025
NO Cal Gas Conc: 48.07 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 48.07 ppm
NO gas Diff:
Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 833

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.836	N/A	NO bkgnd or offset:	0.1	N/A
NOX coeff or slope:	0.828	N/A	NOX bkgnd or offset:	-0.4	N/A
NO2 coeff or slope:	1.000	N/A	Reaction cell Press:	4.4	N/A

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000389	
NO _x Cal Offset:	-0.960000	
NO Cal Slope:	1.000187	
NO Cal Offset:	-1.540000	
NO ₂ Cal Slope:	0.997323	
NO ₂ Cal Offset:	-0.628631	



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	3.6	-0.4	4.0	----	----
as found span	4917	83.2	817.2	799.9	17.3	816.8	798.8	18.0	1.0005	1.0014
as found 2nd	4958	41.6	408.6	399.9	8.7	407.3	397.5	9.9	1.0032	1.0061
as found 3rd	4979	20.8	204.3	200.0	4.3	202.4	196.2	6.2	1.0094	1.0192
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

Average Correction Factor

Corrected As found	NO _x = 813.2 ppb	NO = 799.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.4%
Previous Response	NO _x = 816.5 ppb	NO = 798.5 ppb		*Percent Change	NO = 0.1%
Baseline Corr 2nd pt	NO _x = 403.7 ppb	NO = 397.9 ppb	As found	NO _x r ² : 0.999961	Nx SI: 0.996599
Baseline Corr 3rd pt	NO _x = 198.8 ppb	NO = 196.6 ppb	As found	NO r ² : 0.999981	NO SI: 1.000273
			As found	NO ₂ r ² : 0.999945	NO ₂ SI: 0.992145
					NO ₂ Int: 3.495

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	4.0	----	----
as found GPT point (400 ppb NO ₂)	795.2	400.5	412.0	413.1	0.9974	100.3%
as found GPT point (200 ppb NO ₂)	795.2	599.5	213.0	212.9	1.0005	100.0%
as found GPT point (100 ppb NO ₂)	795.2	699.7	112.8	116.0	0.9725	102.8%
1st GPT point (400 ppb O ₃)						
2nd GPT point (200 ppb O ₃)						
3rd GPT point (100 ppb O ₃)						

Average Correction Factor


Notes: Removal calibration performed due to elevated zero baseline, the instrument needs further investigation and maintenance.

Calibration Performed By: Aswin Sasi Kumar

NO_x Calibration Plot

Date: June 6, 2023

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Wapasu
Calibration Date: June 7, 2023
Start time (MST): 10:08
Reason: Install
Station number: AMS17
Last Cal Date: N/A
End time (MST): 15:04

Calibration Standards

NO Gas Cylinder #: T375YK8
NOX Cal Gas Conc: 49.11 ppm
Removed Cylinder #: T375YK8
Removed Gas NOX Conc: 49.11 ppm
NOX gas Diff: T375YK8
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: April 13, 2025
NO Cal Gas Conc: 48.07 ppm
Removed Gas Exp Date: T375YK8
Removed Gas NO Conc: 48.07 ppm
NO gas Diff: T375YK8
Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: Thermo Scientific 42iQ
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 12300522720

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	N/A	0.932	NO bkgnd or offset:	N/A	2.4
NOX coeff or slope:	N/A	0.984	NOX bkgnd or offset:	N/A	2.4
NO2 coeff or slope:	N/A	1.000	Reaction cell Press:	N/A	141.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	N/A	1.003075
NO _x Cal Offset:	N/A	-4.520000
NO Cal Slope:	N/A	1.003530
NO Cal Offset:	N/A	-5.460000
NO ₂ Cal Slope:	N/A	1.006000
NO ₂ Cal Offset:	N/A	-0.618680



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.0	----	----
high point	4917	83.2	817.2	799.9	17.3	818.0	800.5	17.5	0.9990	0.9992
second point	4958	41.6	408.6	399.9	8.7	401.5	391.6	9.9	1.0177	1.0213
third point	4979	20.8	204.3	200.0	4.3	196.8	190.8	6.0	1.0381	1.0481
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.3	0.1	----	----
as left span	4917	83.2	817.2	392.7	424.5	812.0	385.2	427.0	1.0064	1.0193
Average Correction Factor									1.0183	1.0229

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.4	390.2	424.5	426.1	0.9963	100.4%
2nd GPT point (200 ppb O3)	797.4	609.2	205.5	208.0	0.9880	101.2%
3rd GPT point (100 ppb O3)	797.4	701.7	113.0	110.9	1.0190	98.1%
Average Correction Factor					1.0011	99.9%

Notes:

Install calibration, zero and span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 7, 2023	Previous Calibration:	N/A
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:08	End Time (MST):	15:04
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient 0.999845	≥0.995	
817.2	818.0	0.9990			
408.6	401.5	1.0177			
204.3	196.8	1.0381			
			Slope	1.003075	0.90 - 1.10
			Intercept	-4.520000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 7, 2023	Previous Calibration:	N/A
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:08	End Time (MST):	15:04
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	0.999778	
799.9	800.5	0.9992			≥0.995
399.9	391.6	1.0213	Slope	1.003530	
200.0	190.8	1.0481			0.90 - 1.10
			Intercept	-5.460000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 7, 2023	Previous Calibration:	N/A
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:08	End Time (MST):	15:04
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits			
0.0	0.0	----	Correlation Coefficient	0.999911	≥0.995			
424.5	426.1	0.9963						
205.5	208.0	0.9880				Slope	1.006000	0.90 - 1.10
113.0	110.9	1.0190						
			Intercept	-0.618680	+/-20			


NO₂ Calibration Curve



NO_x Calibration Plot

Date: June 7, 2023

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Wapasu Station number: AMS17
 Calibration Date: June 1, 2023 Last Cal Date: May 5, 2023
 Start time (MST): 10:15 End time (MST): 13:18
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 2449
 ZAG Make/Model: API T701H Serial Number: 359

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 3870
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008857	1.004914	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.600000	-0.460000	Coeff or Slope:	1.020	1.020

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	5000	1077.3	400.0	401.0	0.998
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.4	----
high point	5000	1077.3	400.0	401.9	0.995
second point	5000	900.3	200.0	200.2	0.999
third point	5000	789.5	100.0	99.1	1.009
as left zero	5000	0.0	0.0	0.3	----
as left span	5000	1077.3	400.0	406.4	0.984
Average Correction Factor					1.001

Baseline Corr As found:	401.0	Previous response	402.9	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 1, 2023	Previous Calibration:	May 5, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:15	End Time (MST):	13:18
Analyzer make:	API T400	Analyzer serial #:	3870

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999979	≥0.995
400.0	401.9	0.9953			
200.0	200.2	0.9990	Slope	1.004914	0.90 - 1.10
100.0	99.1	1.0091			
			Intercept	-0.460000	+/- 5



O₃ Calibration Plot

Date: June 1, 2023

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Wapasu Station number: AMS 17
 Calibration Date: June 21, 2023 Last Cal Date: May 23, 2023
 Start time (MST): 13:31 End time (MST): 14:09

Analyzer Make: API T640 S/N: 1183
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	15.5	15.1	15.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	712.8	715.8	712.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>June 21, 2023</u>	Last Cal Date: <u>May 23, 2023</u>			
	PM w/o HEPA: <u>2.9</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>March 23, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>March 23, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: Temp, pressure and flow checked. Leak check passed.

Calibration by: Aswin Sasi Kumar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	June 15, 2023	Last Cal Date:	May 8, 2023
Start time (MST):	10:37	End time (MST):	13:48
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.40	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC463851			
Removed Cal Gas Conc:	49.40	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API 701H		Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000576	1.002775	Backgd or Offset:	22.6	22.6
Calibration intercept:	-1.343541	-1.563172	Coeff or Slope:	0.808	0.808

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5009	0.0	0.0	-0.2	----
as found span	4919	81.0	800.3	800.2	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4919	81.0	800.3	802.1	0.998
second point	4959	40.5	400.2	398.0	1.005
third point	4979	20.2	199.6	197.4	1.011
as left zero	5000	0.0	0.0	-0.2	----
as left span	4919	81.0	800.3	799.1	1.001
Average Correction Factor					1.005

Baseline Corr As found:	800.40	Previous response	799.40	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 8, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:37	End Time (MST):	13:48
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999975	≥0.995
800.3	802.1	0.9977			
400.2	398.0	1.0055	Slope	1.002775	0.90 - 1.10
199.6	197.4	1.0112			
			Intercept	-1.563172	+/-30



SO2 Calibration Plot

Date: June 15, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: June 27, 2023 Last Cal Date: May 16, 2023
 Start time (MST): 10:00 End time (MST): 14:55
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.479 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC500395
 Removed Cal Gas Conc: 5.479 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
 ZAG Make/Model: Teledyne API T701 Serial Number: 360

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359
 Converter make: CD Nova CDN-101 Converter serial #: 555
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001444	1.003297	Backgd or Offset: 2.59	2.70
Calibration intercept:	0.240955	0.101073	Coeff or Slope: 1.159	1.189

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4927	73.0	80.0	78.1	1.026
as found 2nd point	4964	36.5	40.0	38.3	1.047
as found 3rd point	4983	18.3	20.0	18.6	1.084
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4927	73.0	80.0	80.3	0.996
second point	4964	36.5	40.0	40.5	0.987
third point	4983	18.3	20.0	19.9	1.007
as left zero	5000	0.0	0.0	0.3	----
as left span	4927	73.0	80.0	80.5	0.994
SO2 Scrubber Check	4923	77.1	771.0	0.0	----
Date of last scrubber change:	17-Dec-21			Ave Corr Factor	0.997
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 78.0 Prev response: 80.35 *% change: -3.0%
 Baseline Corr 2nd AF pt: 38.2 AF Slope: 0.978430 AF Intercept: -0.478368
 Baseline Corr 3rd AF pt: 18.5 AF Correlation: 0.999748

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Scrubber check completed after calibrator zero. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 16, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:00	End Time (MST):	14:55
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999947	≥0.995
80.0	80.3	0.9962			
40.0	40.5	0.9875	Slope	1.003297	0.90 - 1.10
20.0	19.9	1.0074			
			Intercept	0.101073	+/-3



TRS Calibration Plot

Date: June 27, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	June 15, 2023	Last Cal Date:	May 13, 2023
Start time (MST):	10:37	End time (MST):	13:48
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.15E-04	3.17E-04	NMHC SP Ratio:	5.66E-05
CH ₄ Retention time:	15.20	15.20	NMHC Peak Area:	161883
				153659

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919	81.0	17.28	16.80	1.029
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4919	81.0	17.28	17.30	0.999
second point	4959	40.5	8.64	8.63	1.001
third point	4979	20.2	4.31	4.32	0.998
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.27	1.000

				Average Correction Factor	1.000
Baseline Corr AF:	16.79	Prev response	17.27	*% change	-2.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	8.74	1.049
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.18	0.999
second point	4959	40.5	4.58	4.59	0.999
third point	4979	20.2	2.29	2.31	0.991
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81	9.17	9.14	1.004
Average Correction Factor					0.996
Baseline Corr AF:	8.74	Prev response	9.16	*% change	-4.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919	81.0	8.11	8.06	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4919	81.0	8.11	8.12	0.999
second point	4959	40.5	4.06	4.04	1.005
third point	4979	20.2	2.02	2.01	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.13	0.998
Average Correction Factor					1.003
Baseline Corr AF:	8.05	Prev response	8.11	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999792	1.000273
THC Cal Offset:	-0.006990	0.004621
CH ₄ Cal Slope:	1.000313	0.999876
CH ₄ Cal Offset:	-0.010212	-0.002412
NMHC Cal Slope:	0.998770	1.000625
NMHC Cal Offset:	0.004221	0.007033

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 13, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:37	End Time (MST):	13:48
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.01	----	Correlation Coefficient	0.999997	≥ 0.995			
17.28	17.30	0.9989						
8.64	8.63	1.0015				Slope	1.000273	0.90 - 1.10
4.31	4.32	0.9982						
			Intercept	0.004621	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 13, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:37	End Time (MST):	13:48
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999981	≥0.995
8.11	8.12	0.9991			
4.06	4.04	1.0047			
2.02	2.01	1.0062			
			Slope	0.999876	0.90 - 1.10
			Intercept	-0.002412	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 13, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:37	End Time (MST):	13:48
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
9.17	9.18	0.9987						
4.58	4.59	0.9986				Slope	1.000625	0.90 - 1.10
2.29	2.31	0.9912						
			Intercept	0.007033	± 0.5			



NMHC Calibration Plot

Date: June 15, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: June 28, 2023 Last Cal Date: May 10, 2023
Start time (MST): 10:55 End time (MST): 16:13
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2XX7ME Cal Gas Expiry Date: January 14, 2024
NOX Cal Gas Conc: 50.48 ppm NO Cal Gas Conc: 49.22 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.48 ppm Removed Gas NO Conc: 49.22 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 2658
ZAG make/model: Teledyne API 701H Serial Number: 360

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1336160088
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.035	0.964	NO bkgnd or offset:	2.9	2.7
NOX coeff or slope:	0.986	0.987	NOX bkgnd or offset:	2.9	2.7
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	220.1	225.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000165	1.001389
NO _x Cal Offset:	0.109832	0.070063
NO Cal Slope:	0.999810	1.001010
NO Cal Offset:	-1.249778	-1.269997
NO ₂ Cal Slope:	0.999469	1.004754
NO ₂ Cal Offset:	1.142324	0.176637



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as found span	4919	81.3	820.8	800.3	20.5	883.5	858.8	24.7	0.9290	0.9318
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
high point	4919	81.3	820.8	800.3	20.5	822.0	800.3	21.6	0.9985	1.0000
second point	4959	40.7	410.9	400.7	10.3	411.6	399.7	11.9	0.9984	1.0024
third point	4980	20.3	204.9	199.8	5.1	205.1	197.0	8.1	0.9992	1.0143
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.3	----	----
as left span	4919	81.3	820.8	381.6	439.2	823.9	382.1	441.8	0.9962	0.9986
Average Correction Factor									0.9987	1.0056

Corrected As found	NO _x = 883.4 ppb	NO = 858.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 7.1%
Previous Response	NO _x = 821.0 ppb	NO = 798.9 ppb		*Percent Change	NO = 7.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.2	380.5	439.2	441.4	0.9950	100.5%
2nd GPT point (200 ppb O3)	799.2	594.6	225.1	226.4	0.9942	100.6%
3rd GPT point (100 ppb O3)	799.2	698.5	121.2	122.0	0.9933	100.7%
Average Correction Factor					0.9942	100.6%

Notes: Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	May 10, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:55	End Time (MST):	16:13
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
820.8	822.0	0.9985		
410.9	411.6	0.9984		
204.9	205.1	0.9992		
			1.000000	
			1.001389	
			0.070063	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	May 10, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:55	End Time (MST):	16:13
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.3	800.3	1.0000			
400.7	399.7	1.0024			
199.8	197.0	1.0143			
			Slope	1.001010	0.90 - 1.10
			Intercept	-1.269997	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	May 10, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:55	End Time (MST):	16:13
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
439.2	441.4	0.9950			
225.1	226.4	0.9942			
121.2	122.0	0.9933			
			Slope	1.004754	0.90 - 1.10
			Intercept	0.176637	+/-20



NO_x Calibration Plot

Date: June 28, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: June 28, 2023 Last Cal Date: May 4, 2023
 Start time (MST): 11:15 End time (MST): 16:29
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T750 Serial Number: 282
 ZAG Make/Model: Teledyne API T751H Serial Number: 321

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 825
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998229	1.005457	Backgd or Offset:	1.000	1.000
Calibration intercept:	-0.340000	-0.380000	Coeff or Slope:	0.987	0.995

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	1.2	----
as found span	4888	1096.9	400.0	399.9	1.000
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	-0.1	----
high point	4888	1101.7	400.0	402.0	0.995
second point	4888	863.9	200.0	200.4	0.998
third point	4888	741.4	100.0	100.0	1.000
as left zero	5000	800.0	0.0	-0.7	----
as left span	4812	1097.9	400.0	407.9	0.981
Average Correction Factor					0.998

Baseline Corr As found:	398.7	Previous response	399.0	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Using portable calibrator for the calibration. Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	May 4, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	11:15	End Time (MST):	16:29
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999997	≥0.995
400.0	402.0	0.9950			
200.0	200.4	0.9980	Slope	1.005457	0.90 - 1.10
100.0	100.0	1.0000			
			Intercept	-0.380000	+/- 5



O₃ Calibration Plot

Date: June 28, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Stony Mountain Station number: AMS 18
 Calibration Date: June 28, 2023 Last Cal Date: May 16, 2023
 Start time (MST): 15:27 End time (MST): 16:08

Analyzer Make: API T640 S/N: 1162
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	18.8	19.2	18.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	700.0	705	700.0	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.98	5.07	4.98	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: _____ Last Cal Date: May 16, 2023
 PM w/o HEPA: _____ PM w/ HEPA: _____ <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	N/A		N/A	<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: N/A w/ HEPA: N/A
 Date Optical Chamber Cleaned: March 22, 2023 <0.2 ug/m3
 Disposable Filter Changed: March 22, 2023

Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022
 Date RH/T Sensor Cleaned: August 30, 2022

Notes: No adjustments needed. Leak check not performed since there was no HEPA filter on-site.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	June 8, 2023	Last Cal Date:	May 9, 2023
Start time (MST):	10:55	End time (MST):	13:47
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,050	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	3,050	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701		Serial Number:	360

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	Start	Finish		Start	Finish
Calibration slope:	1.000645	0.999102	Backgd or Offset:	-0.009	-0.010
Calibration intercept:	0.127800	0.057760	Coeff or Slope:	0.905	0.906

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.7	40.8	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.7	40.7	1.001
second point	4966	33.3	20.3	20.5	0.992
third point	4983	16.7	10.2	10.2	0.998
as left zero	3000	0.0	0.0	0.0	----
as left span	2960	40.0	40.7	40.7	0.999
Average Correction Factor					0.997

Baseline Corr As found:	40.72	Prev response:	40.84	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Adjusted zero and span.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 8, 2023	Previous Calibration:	May 9, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:55	End Time (MST):	13:47
Analyzer make:	API T300	Analyzer serial #:	3504

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999973	≥0.995
40.7	40.7	1.0010			
20.3	20.5	0.9915	Slope	0.999102	0.90 - 1.10
10.2	10.2	0.9978			
			Intercept	0.057760	+/-1.5



CO Calibration Plot

Date: June 8, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	June 14, 2023	Last Cal Date:	May 12, 2023
Start time (MST):	10:24	End time (MST):	13:46
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2026
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
N2 Gen Make/Model:	Peak Scientific		Serial Number:	771048317

Analyzer Information

Analyzer make:	API T360	Analyzer serial #:	283
Analyzer Range	0 - 2,000 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997386	0.997863	Backgd or Offset:	-0.081	-0.081
Calibration intercept:	-0.780000	-0.740000	Coeff or Slope:	1.085	1.091

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-0.8	----
as found span	2920	80.0	1605.9	1564.8	1.026
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	-0.1	----
high point	2920	80.0	1605.9	1600.0	1.004
second point	2960	40.0	802.9	806.2	0.996
third point	2980	20.0	401.5	395.2	1.016
as left zero	3000	0.0	0.0	0.6	----
as left span	2930	80.0	1600.5	1608.2	0.995
Average Correction Factor					1.005

Baseline Corr As found:	1565.60	Prev response:	1600.89	*% change:	-2.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	June 14, 2023	Previous Calibration	May 12, 2023
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:24	End Time (MST)	13:46
Analyzer make	API T360	Analyzer serial #	283

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999959	
1605.9	1600.0	1.0037			≥0.995
802.9	806.2	0.9959	Slope	0.997863	
401.5	395.2	1.0159			0.90 - 1.10
			Intercept	-0.740000	+/-10



CO₂ Calibration Plot

Date: June 14, 2023

Location: Stony Mountain





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS19
FIREBAG
JUNE 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	June 20, 2023	Last Cal Date:	May 1, 2023
Start time (MST):	10:35	End time (MST):	13:43
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.29	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC716618			
Removed Cal Gas Conc:	49.29	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1607
ZAG Make/Model:	API T701		Serial Number:	1118

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994974	0.998790	Backgd or Offset:	10.2	10.3
Calibration intercept:	0.198788	-0.461656	Coeff or Slope:	0.992	0.992

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	-0.3	----
as found span	4919	81.1	799.5	798.1	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.0	-0.3	----
high point	4919	81.1	799.5	798.1	1.002
second point	4959	40.6	400.3	399.3	1.002
third point	4980	20.3	200.1	199.2	1.005
as left zero	4999	0.0	0.0	-0.2	----
as left span	4919	81.1	799.5	799.6	1.000
Average Correction Factor					1.003

Baseline Corr As found:	798.40	Previous response	795.65	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 20, 2023	Previous Calibration:	May 1, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:35	End Time (MST):	13:43
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	1.000000	≥0.995
799.5	798.1	1.0017			
400.3	399.3	1.0024	Slope	0.998790	0.90 - 1.10
200.1	199.2	1.0045			
			Intercept	-0.461656	+/-30



SO2 Calibration Plot

Date: June 20, 2023

Location: Firebag





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Firebag Station number: AMS19
 Calibration Date: June 12, 2023 Last Cal Date: May 2, 2023
 Start time (MST): 10:13 End time (MST): 13:43
 Reason: As Found

Calibration Standards

Cal Gas Concentration: 5.114 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517427
 Removed Cal Gas Conc: 5.114 ppm Rem Gas Exp Date: n/a
 Removed Gas Cyl #: n/a Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1607
 ZAG Make/Model: Teledyne API T701 Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1336160090
 Converter make: Global Converter serial #: 2022-222
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008056		Backgd or Offset: 3.11	NA
Calibration intercept:	-0.061725		Coeff or Slope: 0.983	NA

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	1.4	----
as found span	4922	78.2	80.0	77.5	1.051
as found 2nd point	4961	39.1	40.0	39.5	1.050
as found 3rd point	4980	19.6	20.0	20.4	1.055
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
SO2 Scrubber Check	4922	78.3	800.2		----
Date of last scrubber change:	January 18, 2023			Ave Corr Factor	
Date of last converter efficiency test:	n/a				efficiency

Baseline Corr As found: 76.1 Prev response: 80.56 *% change: -5.9%
 Baseline Corr 2nd AF pt: 38.1 AF Slope: 0.951886 AF Intercept: 1.379494
 Baseline Corr 3rd AF pt: 19.0 AF Correlation: 0.999998

* = > +/-5% change initiates investigation


Notes: As founds completed. Scrubber check passed. Lamp voltage above upper alarm limit, will replace the lamp June 13.

Calibration Performed By: Braiden Boutillier

H₂S Calibration Plot

Date: June 12, 2023

Location: Firebag





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Firebag	Station number:	AMS19
Calibration Date:	June 13, 2023	Last Cal Date:	June 12, 2023
Start time (MST):	10:15	End time (MST):	15:17
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.114	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517427			
Removed Cal Gas Conc:	5.114	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1607
ZAG Make/Model:	Teledyne API T701		Serial Number:	1118

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090
Converter make:	Global	Converter serial #:	2022-222
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007755	1.009200	Backgd or Offset:	4.94
Calibration intercept:	0.000000	-0.001760	Coeff or Slope:	1.040
				4.21
				1.068

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	78.2	80.0	80.7	0.991
second point	4961	39.1	40.0	40.4	0.990
third point	4980	19.6	20.0	20.2	0.992
as left zero	5000	0.0	0.0	0.1	----
as left span	4922	78.2	80.0	80.8	0.990

SO₂ Scrubber Check

Date of last scrubber change:	January 18, 2023	Ave Corr Factor	0.991
Date of last converter efficiency test:	n/a	efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: Changed lamp. Adjusted lamp voltage and PMT voltage. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 13, 2023	Previous Calibration:	June 12, 2023
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	10:15	End Time (MST):	15:17
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.0	----	Correlation Coefficient	0.999999	≥0.995
80.0	80.7	0.9911			
40.0	40.4	0.9899			
20.0	20.2	0.9925			
			Slope	1.009200	0.90 - 1.10
			Intercept	-0.001760	+/-3



H₂S Calibration Plot

Date: June 13, 2023

Location: Firebag





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Firebag Station number: AMS19
 Calibration Date: June 27, 2023 Last Cal Date: June 13, 2023
 Start time (MST): 9:36 End time (MST): 13:28
 Reason: Maintenance

Calibration Standards

Cal Gas Concentration: 5.114 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517427
 Removed Cal Gas Conc: 5.114 ppm Rem Gas Exp Date: n/a
 Removed Gas Cyl #: n/a Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1607
 ZAG Make/Model: Teledyne API T701 Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1336160090
 Converter make: Global Converter serial #: 2022-222
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009200	1.007771	Backgd or Offset: 4.21	6.34
Calibration intercept:	-0.001760	-0.101745	Coeff or Slope: 1.068	1.137

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	1.7	----
as found span	4922	78.2	80.0	78.0	1.048
as found 2nd point	4961	39.1	40.0	39.6	1.055
as found 3rd point	4980	19.6	20.0	20.4	1.072
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4922	78.2	80.0	80.4	0.995
second point	4961	39.1	40.0	40.4	0.990
third point	4980	19.6	20.0	20.1	0.997
as left zero	5000	0.0	0.0	0.1	----
as left span	4922	78.2	80.0	80.3	0.996

SO₂ Scrubber Check

Date of last scrubber change:	January 18, 2023	Ave Corr Factor	0.994
Date of last converter efficiency test:	n/a	efficiency	

Baseline Corr As found: 76.3 Prev response: 80.71 *% change: -5.8%
 Baseline Corr 2nd AF pt: 37.9 AF Slope: 0.955456 AF Intercept: 1.479510
 Baseline Corr 3rd AF pt: 18.7 AF Correlation: 0.999962

* = > +/-5% change initiates investigation

Notes: Calibrated due to baseline shift. Adjusted zero and span. No investigation needed as the baseline shift is most likely due to high humidity.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	June 13, 2023
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	9:36	End Time (MST):	13:28
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient	≥0.995
80.0	80.4	0.9948		
40.0	40.4	0.9899	Slope	0.90 - 1.10
20.0	20.1	0.9974		
			Intercept	+/-3



H₂S Calibration Plot

Date: June 27, 2023

Location: Firebag





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Firebag Station number: AMS19
 Calibration Date: June 30, 2023 Last Cal Date: June 27, 2023
 Start time (MST): 10:17 End time (MST): 10:55
 Reason: Removal problems with lamp caused baseline and sensitivity drift

Calibration Standards

Cal Gas Concentration: 5.114 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517427
 Removed Cal Gas Conc: 5.114 ppm Rem Gas Exp Date: n/a
 Removed Gas Cyl #: n/a Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1607
 ZAG Make/Model: Teledyne API T701 Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1336160090
 Converter make: Global Converter serial #: 2022-222
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009200		Backgd or Offset: 6.34	N/A
Calibration intercept:	-0.001760		Coeff or Slope: 1.137	N/A

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	9.4	----
as found span	4922	78.2	80.0	45.0	2.247
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

SO₂ Scrubber Check

Date of last scrubber change:	January 18, 2023	Ave Corr Factor	
Date of last converter efficiency test:	n/a	efficiency	

Baseline Corr As found: 35.6 Prev response: 80.71 *% change: -126.7%
 Baseline Corr 2nd AF pt: NA AF Slope: NA AF Intercept: NA
 Baseline Corr 3rd AF pt: NA AF Correlation: NA

* = > +/-5% change initiates investigation


Notes: Lamp decayed to the point where readings and baseline were largely out of compliance. Replacing with green tagged 43i-TLE.

Calibration Performed By: Braiden Boutillier

H₂S Calibration Plot

Date: June 30, 2023

Location: Firebag





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Firebag	Station number:	AMS19
Calibration Date:	June 30, 2023	Last Cal Date:	June 27, 2023
Start time (MST):	11:01	End time (MST):	13:54
Reason:	Install		

Calibration Standards

Cal Gas Concentration:	5.114	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517427			
Removed Cal Gas Conc:	5.114	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1607
ZAG Make/Model:	Teledyne API T701		Serial Number:	1118

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032
Converter make:	Global	Converter serial #:	2022-222
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009200	1.002911	Backgd or Offset:	NA
Calibration intercept:	-0.001760	-0.281624	Coeff or Slope:	NA
				2.54
				1.122

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4922	78.2	80.0	80.1	0.998
second point	4961	39.1	40.0	39.5	1.012
third point	4980	19.6	20.0	19.8	1.013
as left zero	5000	0.0	0.0	-0.1	----
as left span	4922	78.2	80.0	79.9	1.001

SO₂ Scrubber Check

Date of last scrubber change:	January 18, 2023	Ave Corr Factor	1.008
Date of last converter efficiency test:	n/a	efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: Installed 43i-TLE (1151680032). Removed 43i-TLE (1336160090) due to increased lamp voltage overnight leading to shifted baseline. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 30, 2023	Previous Calibration:	June 27, 2023
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	11:01	End Time (MST):	13:54
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999952	≥0.995
80.0	80.1	0.9985			
40.0	39.5	1.0124	Slope	1.002911	0.90 - 1.10
20.0	19.8	1.0125			
			Intercept	-0.281624	+/-3



H₂S Calibration Plot

Date: June 30, 2023

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	June 20, 2023	Last Cal Date:	May 24, 2023
Start time (MST):	10:35	End time (MST):	13:43
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1607
ZAG Make/Model:	API T701	Serial Number:	1118

Analyzer Information

Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089
Analyzer Range:	0 - 20 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994758	1.003320	Background:	2.37	2.69
Calibration intercept:	0.018270	-0.013586	Coefficient:	3.846	3.908

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.21	----
as found span	4919	81.1	17.31	17.31	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.00	0.00	----
high point	4919	81.1	17.31	17.35	0.997
second point	4959	40.6	8.66	8.69	0.997
third point	4980	20.3	4.33	4.31	1.006
as left zero	5000	0.0	0.00	0.05	----
as left span	4919	81.1	17.31	17.49	0.989
Average Correction Factor					1.000
Baseline Corr As found:	17.10	Previous response	17.23	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted zero.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 20, 2023	Previous Calibration:	May 24, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:35	End Time (MST):	13:43
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.00	----	Correlation Coefficient	0.999994	
17.31	17.35	0.9974			≥0.995
8.66	8.69	0.9970	Slope	1.003320	
4.33	4.31	1.0057			0.90 - 1.10
			Intercept	-0.013586	+/-1.5



THC Calibration Plot

Date: June 20, 2023

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Firebag
Calibration Date: June 22, 2023
Start time (MST): 10:07
Reason: Routine
Station number: AMS 19
Last Cal Date: May 4, 2023
End time (MST): 14:50

Calibration Standards

NO Gas Cylinder #: T2Y1K63
NOX Cal Gas Conc: 51.12 ppm
Removed Cylinder #: n/a
Removed Gas NOX Conc: 51.12 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 30, 2023
NO Cal Gas Conc: 49.40 ppm
Removed Gas Exp Date: n/a
Removed Gas NO Conc: 49.40 ppm
NO gas Diff:
Serial Number: 1607
Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1410661309

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.078	1.094	NO bkgnd or offset:	7.5	7.6
NOX coeff or slope:	0.995	0.994	NOX bkgnd or offset:	7.5	7.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	210.3	211.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999367	0.998084
NO _x Cal Offset:	0.856331	1.096129
NO Cal Slope:	1.001140	0.999369
NO Cal Offset:	-0.011041	0.508726
NO ₂ Cal Slope:	0.998524	1.001853
NO ₂ Cal Offset:	-0.233199	-0.516592



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.0	0.0	0.0	-0.2	0.2	----	----
as found span	4919	81.0	828.1	800.3	27.9	819.0	789.8	29.1	1.0112	1.0133
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	4999	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4919	81.0	828.1	800.3	27.9	827.0	799.8	27.6	1.0014	1.0006
second point	4960	40.5	414.0	400.1	13.9	415.2	401.3	13.9	0.9972	0.9970
third point	4980	20.2	206.5	199.6	6.9	208.1	200.0	8.1	0.9924	0.9978
as left zero	4999	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as left span	4919	81.0	828.1	362.4	465.8	828.0	363.5	465.0	1.0002	0.9969
Average Correction Factor									0.9970	0.9985

Corrected As found	NO _x = 819.0 ppb	NO = 790.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.2%	
Previous Response	NO _x = 828.5 ppb	NO = 801.2 ppb		*Percent Change	NO = -1.4%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.5	361.6	465.8	466.9	0.9976	100.2%
2nd GPT point (200 ppb O3)	799.5	579.9	247.5	245.7	1.0072	99.3%
3rd GPT point (100 ppb O3)	799.5	691.1	136.3	136.4	0.9990	100.1%
Average Correction Factor					1.0012	99.9%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 22, 2023	Previous Calibration:	May 4, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:07	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
828.1	827.0	1.0014			
414.0	415.2	0.9972			
206.5	208.1	0.9924			
			Slope	0.998084	0.90 - 1.10
			Intercept	1.096129	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 22, 2023	Previous Calibration:	May 4, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:07	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	799.8	1.0006		
400.1	401.3	0.9970		
199.6	200.0	0.9978		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 22, 2023	Previous Calibration:	May 4, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:07	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
465.8	466.9	0.9976		
247.5	245.7	1.0072		
136.3	136.4	0.9990		
			0.999966	
			1.001853	
			-0.516592	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: June 22, 2023

Location: Firebag





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS20 MACKAY RIVER JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	MackKay River	Station number:	AMS20
Calibration Date:	June 8, 2023	Last Cal Date:	May 2, 2023
Start time (MST):	7:15	End time (MST):	9:59
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.22 ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC306868		
Removed Cal Gas Conc:	49.22 ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	1220
ZAG Make/Model:	Teledyne API 701	Serial Number:	4522

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007790	0.995624	Backgd or Offset: 18.8	18.2
Calibration intercept:	2.430853	0.891225	Coeff or Slope: 0.974	0.945

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4919	81.3	800.3	819.2	0.977
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4919	81.3	800.3	797.7	1.003
second point	4959	40.7	400.7	399.0	1.004
third point	4980	20.3	199.8	201.3	0.993
as left zero	5000	0.0	0.0	0.3	----
as left span	4919	81.3	800.3	799.5	1.001
Average Correction Factor					1.000

Baseline Corr As found:	819.20	Previous response	808.93	*% change	1.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 8, 2023	Previous Calibration:	May 2, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:15	End Time (MST):	9:59
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999991	≥0.995
800.3	797.7	1.0032			
400.7	399.0	1.0042	Slope	0.995624	0.90 - 1.10
199.8	201.3	0.9927			
			Intercept	0.891225	+/-30



SO2 Calibration Plot

Date: June 8, 2023

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	MacKay River	Station number:	AMS20
Calibration Date:	June 7, 2023	Last Cal Date:	May 25, 2023
Start time (MST):	7:24	End time (MST):	11:18
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.12	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC515997			
Removed Cal Gas Conc:	5.12	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1220
ZAG Make/Model:	Teledyne API 701		Serial Number:	4522

Analyzer Information

Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	12124313139
Converter make:	Global	Converter serial #:	2022-226
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006091	1.002952	Backgd or Offset:	1.9	1.10
Calibration intercept:	-0.292887	-0.033000	Coeff or Slope:	0.619	0.605

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.7	----
as found span	4922	78.1	80.0	81.8	0.969
as found 2nd point	4961	39.0	39.9	40.6	0.967
as found 3rd point	4980	19.5	20.0	20.2	0.956
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4922	78.1	80.0	80.3	0.997
second point	4961	39.0	40.0	39.8	1.004
third point	4980	19.5	20.0	20.3	0.986
as left zero	5000	0.0	0.0	0.0	----
as left span	4922	78.1	80.0	80.1	0.999
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:	May 25, 2023			Ave Corr Factor	0.996
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	82.5	Prev response:	80.17	*% change:	2.8%
Baseline Corr 2nd AF pt:	41.3	AF Slope:	1.030464	AF Intercept:	-0.560003
Baseline Corr 3rd AF pt:	20.9	AF Correlation:	0.999985		

* = > +/-5% change initiates investigation

Notes: Scrubber check completed after calibrator zero. Adjusted zero and span.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 7, 2023	Previous Calibration:	May 25, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:24	End Time (MST):	11:18
Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	12124313139

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999961	≥0.995
80.0	80.3	0.9967			
40.0	39.8	1.0042	Slope	1.002952	0.90 - 1.10
20.0	20.3	0.9860			
			Intercept	-0.033000	+/-3



H₂S Calibration Plot

Date: June 7, 2023

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	MacKay River	Station number:	AMS20
Calibration Date:	June 8, 2023	Last Cal Date:	May 2, 2023
Start time (MST):	7:15	End time (MST):	9:58
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC306868	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.40</u> ppm	CH4 Equiv Conc.	1066.45 ppm
C3H8 Cal Gas Conc.	<u>206.20</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.40</u> ppm	CH4 Equiv Conc.	1066.45 ppm
Removed C3H8 Conc.	<u>206.20</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	1220
ZAG Make/Model:	Teledyne API 701	Serial Number:	4522

Analyzer Information

Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727
Analyzer Range:	0 - 20 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004428	1.002852	Background:	3.250	3.220
Calibration intercept:	-0.093194	-0.024234	Coefficient:	5.488	5.430

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.03	----
as found span	4919	81.3	17.34	17.55	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	-0.05	----
high point	4919	81.3	17.34	17.34	1.000
second point	4959	40.7	8.68	8.73	0.994
third point	4980	20.3	4.33	4.32	1.002
as left zero	5000	0.0	0.00	-0.06	----
as left span	4919	81.3	17.34	17.28	1.003
Average Correction Factor					0.999
Baseline Corr As found:	17.58	Previous response	17.32	*% change	1.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Hydrogen cylinder changed. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 8, 2023	Previous Calibration:	May 2, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:15	End Time (MST):	9:58
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.05	----	Correlation Coefficient	0.999978	≥0.995
17.34	17.34	1.0000			
8.68	8.73	0.9944	Slope	1.002852	0.90 - 1.10
4.33	4.32	1.0022			
			Intercept	-0.024234	+/-1.5



THC Calibration Plot

Date: June 8, 2023

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: MacKay River Station number: AMS20
Calibration Date: June 19, 2023 Last Cal Date: May 11, 2023
Start time (MST): 7:00 End time (MST): 11:16
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 1220
ZAG make/model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.459	1.503	NO bkgnd or offset:	4.0	4.2
NOX coeff or slope:	0.990	0.992	NOX bkgnd or offset:	4.0	4.2
NO2 coeff or slope:	0.995	0.995	Reaction cell Press:	178.6	178.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.981506	0.997289
NO _x Cal Offset:	1.850414	2.290060
NO Cal Slope:	0.983542	0.998247
NO Cal Offset:	0.691620	1.091419
NO ₂ Cal Slope:	1.000887	1.002522
NO ₂ Cal Offset:	-0.270425	-0.129742



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
as found span	4917	83.3	819.5	800.3	19.2	795.9	777.7	18.4	1.0296	1.0291
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
high point	4917	83.3	819.5	800.3	19.2	818.9	799.9	19.0	1.0007	1.0005
second point	4956	41.7	410.4	400.8	9.6	411.4	400.4	11.0	0.9977	1.0011
third point	4979	20.8	204.6	199.9	4.8	209.3	202.6	6.7	0.9777	0.9864
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
as left span	4917	83.3	819.5	444.4	375.1	816.4	440.0	376.4	1.0037	1.0100
Average Correction Factor									0.9920	0.9960

Corrected As found	NO _x = 795.9 ppb	NO = 777.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.3%	
Previous Response	NO _x = 806.2 ppb	NO = 787.8 ppb		*Percent Change	NO = -1.3%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.3	440.4	375.1	376.1	0.9972	100.3%
2nd GPT point (200 ppb O3)	796.3	612.3	203.2	203.3	0.9993	100.1%
3rd GPT point (100 ppb O3)	796.3	701.7	113.8	113.6	1.0014	99.9%
Average Correction Factor					0.9993	100.1%

Notes:

Span adjusted. No Maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 19, 2023	Previous Calibration:	May 11, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	7:00	End Time (MST):	11:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
819.5	818.9	1.0007			
410.4	411.4	0.9977			
204.6	209.3	0.9777			
			Correlation Coefficient	0.999963	≥0.995
			Slope	0.997289	0.90 - 1.10
			Intercept	2.290060	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 19, 2023	Previous Calibration:	May 11, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	7:00	End Time (MST):	11:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
800.3	799.9	1.0005			
400.8	400.4	1.0011			
199.9	202.6	0.9864			
			Slope	0.998247	0.90 - 1.10
			Intercept	1.091419	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 19, 2023	Previous Calibration:	May 11, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:00	End Time (MST):	11:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
375.1	376.1	0.9972		
203.2	203.3	0.9993		
113.8	113.6	1.0014		
			0.999996	
			1.002522	
			-0.129742	



NO_x Calibration Plot

Date: June 19, 2023

Location: MacKay River





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS21
CONKLIN
JUNE 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	June 5, 2023	Last Cal Date:	May 3, 2023
Start time (MST):	10:25	End time (MST):	13:12
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.93	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	<u>CC259455</u>			
Removed Cal Gas Conc:	49.93	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3810
ZAG Make/Model:	Teledyne API 701		Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.015485	0.999546	Backgd or Offset:	28.4	27.3
Calibration intercept:	1.176454	0.235763	Coeff or Slope:	0.914	0.883

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5005	0.0	0.0	-0.2	----
as found span	4920	80.2	800.8	825.7	0.970
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5005	0.0	0.0	0.0	----
high point	4920	80.2	800.8	800.5	1.000
second point	4960	40.1	400.4	400.9	0.999
third point	4980	20.0	200.1	200.3	0.999
as left zero	5005	0.0	0.0	0.0	----
as left span	4920	80.2	800.8	800.0	1.001
Average Correction Factor					0.999

Baseline Corr As found:	825.90	Previous response	814.42	*% change	1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Adjusted the span only.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 3, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:25	End Time (MST):	13:12
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999999	
800.8	800.5	1.0004			≥0.995
400.4	400.9	0.9988	Slope	0.999546	
200.1	200.3	0.9991			0.90 - 1.10
			Intercept	0.235763	+/-30



SO2 Calibration Plot

Date: June 5, 2023

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Conklin Station number: AMS21
 Calibration Date: June 28, 2023 Last Cal Date: May 30, 2023
 Start time (MST): 8:31 End time (MST): 12:03
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.00 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC501204
 Removed Cal Gas Conc: 5.03 ppm Rem Gas Exp Date: April 16, 2022
 Removed Gas Cyl #: CC505493 Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3810
 ZAG Make/Model: API 701H Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1236656116
 Converter make: CD-Nova 101 Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000000	1.001571	Backgd or Offset:	2.9
Calibration intercept:	0.000000	0.320000	Coeff or Slope:	0.974
				2.4
				0.974

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	4921	79.5	80.0	80.0	0.995
as found 2nd point	4960	39.8	40.0	40.0	0.991
as found 3rd point	4980	19.9	20.0	19.9	0.986
new cylinder response	4920	80.0	80.0		

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.0	80.0	80.3	0.996
second point	4960	40.0	40.0	40.6	0.985
third point	4980	20.0	20.0	20.5	0.976
as left zero	5000	0.0	0.0	0.3	----
as left span	4920	80.0	80.0	80.1	0.999
SO2 Scrubber Check	4920	80.2	802.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.986
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 80.4 Prev response: 79.97 *% change: 0.5%
 Baseline Corr 2nd AF pt: 40.4 AF Slope: 1.004858 AF Intercept: -0.302369
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999993

* = > +/-5% change initiates investigation

Notes: Adjusted the zero only.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	May 30, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:31	End Time (MST):	12:03
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999960	≥0.995
80.0	80.3	0.9963			
40.0	40.6	0.9852	Slope	1.001571	0.90 - 1.10
20.0	20.5	0.9756			
			Intercept	0.320000	+/-3



TRS Calibration Plot

Date: June 28, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	June 5, 2023	Last Cal Date:	May 24, 2023
Start time (MST):	10:25	End time (MST):	13:12
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.22E-04	2.26E-04	NMHC SP Ratio:	4.83E-05
CH ₄ Retention time:	12.00	12.00	NMHC Peak Area:	189107
				184655

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	16.83	1.017
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.14	0.999
second point	4960	40.1	8.56	8.57	0.999
third point	4980	20.0	4.27	4.30	0.993
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.16	0.998
Average Correction Factor					0.997
Baseline Corr AF:	16.83	Prev response	17.14	*% change	-1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	8.95	1.021
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.14	1.000
second point	4960	40.1	4.57	4.58	0.998
third point	4980	20.0	2.28	2.30	0.991
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.15	0.999
Average Correction Factor					0.996
Baseline Corr AF:	8.95	Prev response	9.15	*% change	-2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.88	1.013
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.99	0.999
second point	4960	40.1	3.99	3.99	1.001
third point	4980	20.0	1.99	2.00	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	8.01	0.997
Average Correction Factor					0.998
Baseline Corr AF:	7.88	Prev response	7.99	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999175	1.000150
THC Cal Offset:	0.027981	0.011178
CH ₄ Cal Slope:	0.999308	1.000639
CH ₄ Cal Offset:	0.008750	0.001348
NMHC Cal Slope:	0.998771	0.999397
NMHC Cal Offset:	0.019631	0.010629

Notes: Adjusted the span only.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 24, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:25	End Time (MST):	13:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
17.13	17.14	0.9993						
8.56	8.57	0.9992				Slope	1.000150	0.90 - 1.10
4.27	4.30	0.9930						
			Intercept	0.011178	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 24, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:25	End Time (MST):	13:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
7.99	7.99	0.9990						
3.99	3.99	1.0005				Slope	1.000639	0.90 - 1.10
1.99	2.00	0.9958						
			Intercept	0.001348	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	June 5, 2023	Previous Calibration:	May 24, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:25	End Time (MST):	13:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995
9.14	9.14	0.9999			
4.57	4.58	0.9978			
2.28	2.30	0.9905			
			Slope	0.999397	0.90 - 1.10
			Intercept	0.010629	+/-0.5



NMHC Calibration Plot

Date: June 5, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	June 21, 2023	Last Cal Date:	June 5, 2023
Start time (MST):	10:00	End time (MST):	12:35
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
C3H8 Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
Removed C3H8 Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.26E-04	2.26E-04	NMHC SP Ratio:	4.95E-05
CH4 Retention time:	12.00	12.00	NMHC Peak Area:	184655
				183139

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.06	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.12	1.000
second point	4960	40.1	8.56	8.53	1.004
third point	4980	20.0	4.27	4.29	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.13	1.000
Average Correction Factor					1.000
Baseline Corr AF:	17.06	Prev response	17.14	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.07	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.13	1.001
second point	4960	40.1	4.57	4.57	1.001
third point	4980	20.0	2.28	2.30	0.991
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.14	1.000
Average Correction Factor					0.997
Baseline Corr AF:	9.07	Prev response	9.14	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.99	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.99	1.000
second point	4960	40.1	3.99	3.96	1.008
third point	4980	20.0	1.99	1.99	1.001
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.99	0.999
Average Correction Factor					1.003
Baseline Corr AF:	7.99	Prev response	7.99	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000150	0.999095
THC Cal Offset:	0.011178	0.002575
CH ₄ Cal Slope:	1.000639	0.999624
CH ₄ Cal Offset:	0.001348	-0.006854
NMHC Cal Slope:	0.999397	0.998558
NMHC Cal Offset:	0.010629	0.009229

Notes: No maintenance done just recalibration; due to dips. Adjusted the span.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	June 21, 2023	Previous Calibration:	June 5, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:00	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999993	≥ 0.995			
17.13	17.12	1.0003						
8.56	8.53	1.0036				Slope	0.999095	0.90 - 1.10
4.27	4.29	0.9955						
			Intercept	0.002575	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	June 21, 2023	Previous Calibration:	June 5, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:00	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999980	≥ 0.995
7.99	7.99	1.0000			
3.99	3.96	1.0078			
1.99	1.99	1.0008			
			Slope	0.999624	0.90 - 1.10
			Intercept	-0.006854	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	June 21, 2023	Previous Calibration:	June 5, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:00	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	≥ 0.995			
9.14	9.13	1.0006						
4.57	4.57	1.0006				Slope	0.998558	0.90 - 1.10
2.28	2.30	0.9905						
			Intercept	0.009229	± 0.5			



NMHC Calibration Plot

Date: June 21, 2023

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Conklin
Calibration Date: June 27, 2023
Start time (MST): 9:12
Reason: Routine
Station number: AMS21
Last Cal Date: May 16, 2023
End time (MST): 13:20

Calibration Standards

NO Gas Cylinder #: T2Y1P1H
NOX Cal Gas Conc: 51.09 ppm
Removed Cylinder #: n/a
Removed Gas NOX Conc: 51.09 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.39 ppm
Removed Gas Exp Date: n/a
Removed Gas NO Conc: 50.39 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 691

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1501663731

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.144	1.112	NO bkgnd or offset:	11.6	11.1
NOX coeff or slope:	1.001	0.999	NOX bkgnd or offset:	11.8	11.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	220.7	201.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001335	0.998659
NO _x Cal Offset:	2.164394	1.463814
NO Cal Slope:	1.000766	1.000266
NO Cal Offset:	1.181398	0.781352
NO ₂ Cal Slope:	1.002657	0.995256
NO ₂ Cal Offset:	-0.146701	0.164082



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
as found span	4921	79.4	811.2	800.1	11.1	835.6	821.8	13.8	0.9709	0.9736
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4921	79.4	811.2	800.1	11.1	811.0	800.9	10.0	1.0003	0.9990
second point	4960	39.7	405.7	400.1	5.6	407.0	400.9	6.1	0.9968	0.9981
third point	4980	19.8	202.3	199.6	2.8	205.2	201.5	3.7	0.9860	0.9903
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as left span	4921	79.4	811.2	382.0	429.2	810.6	383.7	426.8	1.0008	0.9956
Average Correction Factor									0.9943	0.9958

Corrected As found	NO _x = 835.8 ppb	NO = 821.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 2.5%
Previous Response	NO _x = 814.5 ppb	NO = 801.9 ppb		*Percent Change	NO = 2.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI: ;

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.1	379.0	429.2	427.4	1.0042	99.6%
2nd GPT point (200 ppb O3)	797.1	590.5	217.7	216.5	1.0056	99.4%
3rd GPT point (100 ppb O3)	797.1	696.0	112.2	112.3	0.9992	100.1%
Average Correction Factor					1.0030	99.7%

Notes:

Adjusted the span only.

Calibration Performed By:

Denny Ray Estador



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 16, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:12	End Time (MST):	13:20
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
811.2	811.0	1.0003		
405.7	407.0	0.9968		
202.3	205.2	0.9860		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 16, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:12	End Time (MST):	13:20
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
800.1	800.9	0.9990			0.999995
400.1	400.9	0.9981			1.000266
199.6	201.5	0.9903			0.781352

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 16, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:12	End Time (MST):	13:20
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
429.2	427.4	1.0042		
217.7	216.5	1.0056		
112.2	112.3	0.9992		
			0.999996	
			0.995256	
			0.164082	



NO_x Calibration Plot

Date: June 27, 2023

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	June 7, 2023	Last Cal Date:	May 9, 2023
Start time (MST):	9:16	End time (MST):	12:13
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	Teledyne API T700	Serial Number:	3810
ZAG Make/Model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734
Analyzer Range:	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002194	0.997914	Backgd or Offset:	-1.2	-2.0
Calibration intercept:	0.656000	1.340000	Coeff or Slope:	1.002	1.002

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-1.5	----
as found span	5000	941.5	400.0	397.0	1.008
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.4	----
high point	5000	941.5	400.0	400.0	1.000
second point	5000	804.0	200.0	201.5	0.993
third point	5000	703.6	100.0	102.0	0.980
as left zero	5000	0.0	0.0	-0.1	----
as left span	5000	936.0	400.0	402.0	0.995
Average Correction Factor					0.991

Baseline Corr As found:	398.5	Previous response	401.5	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Adjusted the zero only.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 7, 2023	Previous Calibration:	May 9, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:16	End Time (MST):	12:13
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999974	≥0.995
400.0	400.0	1.0000			
200.0	201.5	0.9926	Slope	0.997914	0.90 - 1.10
100.0	102.0	0.9804			
			Intercept	1.340000	+/- 5



O₃ Calibration Plot

Date: June 7, 2023

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	June 27, 2023	Last Cal Date:	May 24, 2023
Start time (MST):	9:30	End time (MST):	10:34
Analyzer Make:	API T640X	S/N:	1597
Particulate Fraction:	PM2.5		
Flow Meter Make/Model:	Alicat	S/N:	388751
Temp/RH standard:	Alicat	S/N:	388751

Monthly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Measured</u>	<u>As left</u>	<u>Adjusted</u>	<u>(Limits)</u>
T (°C)	21.6	21.01	21.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	704.7	706.83	704.7	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.09	5.17	5.09	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>June 27, 2023</u>	Last Cal Date: <u>May 24, 2023</u>			
	PM w/o HEPA: <u>21.4</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Post maintenance</u>	<u>As left</u>	<u>Adjusted</u>	<u>(Limits)</u>
PMT Peak Test	11.8	12	11	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>13.1</u>	w/ HEPA: <u>0</u>		<0.2 ug/m3
Date Optical Chamber Cleaned:		<u>June 27, 2023</u>			
Disposable Filter Changed:		<u>June 27, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned:	<u>June 27, 2023</u>
Date RH/T Sensor Cleaned:	<u>June 27, 2023</u>

Notes: No adjustments made for the monthly test. Adjusted PMT Peak Test. Inspected inlet head; relatively clean.

Calibration by: Denny Ray Estador



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER
JUNE 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	June 14, 2023	Last Cal Date:	May 17, 2023
Start time (MST):	9:44	End time (MST):	12:50
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.11	ppm	Cal Gas Exp Date:	January 18, 2029
Cal Gas Cylinder #:	CC281519			
Removed Cal Gas Conc:	50.11	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000578	1.007567	Backgd or Offset:	21.3	21.3
Calibration intercept:	0.744403	1.922834	Coeff or Slope:	0.998	0.998

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.9	----
as found span	4920	79.8	799.8	806.0	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.9	----
high point	4920	79.8	799.8	806.8	0.991
second point	4960	39.9	399.9	406.7	0.983
third point	4980	20.0	200.4	204.0	0.983
as left zero	5000	0.0	0.0	0.9	----
as left span	4920	79.8	799.8	808.8	0.989
Average Correction Factor					0.986

Baseline Corr As found:	805.10	Previous response	800.99	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 14, 2023	Previous Calibration:	May 17, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:44	End Time (MST):	12:50
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.9	----	Correlation Coefficient	0.999985	≥0.995
799.8	806.8	0.9913	Slope	1.007567	0.90 - 1.10
399.9	406.7	0.9832	Intercept	1.922834	+/-30
200.4	204.0	0.9825			



SO2 Calibration Plot

Date: June 14, 2023

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Janvier Station number: AMS22
 Calibration Date: June 13, 2023 Last Cal Date: May 24, 2023
 Start time (MST): 10:20 End time (MST): 14:45
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.03 ppm Cal Gas Exp Date: April 16, 2022
 Cal Gas Cylinder #: DT0018680
 Removed Cal Gas Conc: 5.03 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680031
 Converter make: CDN-101 Converter serial #: 587
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997220	1.000081	Backgd or Offset:	6.33
Calibration intercept:	0.161030	0.260853	Coeff or Slope:	1.022
				0.973

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	3.1	----
as found span	4920	79.5	80.0	86.0	0.965
as found 2nd point	4960	39.8	40.0	44.8	0.960
as found 3rd point	4980	19.9	20.0	24.4	0.940
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	79.5	80.0	80.1	0.999
second point	4960	39.8	40.0	40.5	0.989
third point	4980	19.9	20.0	20.5	0.977
as left zero	5000	0.0	0.0	0.3	----
as left span	4920	79.5	80.0	80.2	0.997
SO2 Scrubber Check	4920	79.8	798.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.988
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 82.9 Prev response: 79.92 *% change: 3.6%
 Baseline Corr 2nd AF pt: 41.7 AF Slope: 1.034373 AF Intercept: 3.360257
 Baseline Corr 3rd AF pt: 21.3 AF Correlation: 0.999950

* = > +/-5% change initiates investigation

Notes: Inlet filter changed after as founds. Scrubber check passed. Adjusted zero and span.

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021


Station Information

Calibration Date:	June 13, 2023	Previous Calibration:	May 24, 2023
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	10:20	End Time (MST):	14:45
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999949	
80.0	80.1	0.9986			≥0.995
40.0	40.5	0.9887	Slope	1.000081	
20.0	20.5	0.9766			0.90 - 1.10
			Intercept	0.260853	+/-3







Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Janvier	Station number:	AMS22
Calibration Date:	June 19, 2023	Last Cal Date:	June 13, 2023
Start time (MST):	11:30	End time (MST):	N/A
Reason:	Removal		

Calibration Standards

Cal Gas Concentration:	5.03	ppm	Cal Gas Exp Date:	April 16, 2022
Cal Gas Cylinder #:	DT0018680			
Removed Cal Gas Conc:	5.03	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031
Converter make:	CDN-101	Converter serial #:	587
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000081		Backgd or Offset:	9.13
Calibration intercept:	0.260853		Coeff or Slope:	0.973
				N/A
				N/A

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	3.4	----
as found span	4920	79.5	80.0	83.2	1.002
as found 2nd point	4960	39.8	40.0	44.0	0.986
as found 3rd point	4980	19.9	20.0	23.3	1.006
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

SO2 Scrubber Check

Date of last scrubber change:	Ave Corr Factor
Date of last converter efficiency test:	efficiency

Baseline Corr As found:	79.8	Prev response:	80.25	*% change:	-0.6%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	0.998940	AF Intercept:	3.500826
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999902		

* = > +/-5% change initiates investigation


Notes: Removing instrument due to elevated zeroes.

Calibration Performed By: Aswin Sasi Kumar

TRS Calibration Plot

Date: June 19, 2023

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Janvier	Station number:	AMS22
Calibration Date:	June 20, 2023	Last Cal Date:	N/A
Start time (MST):	10:33	End time (MST):	13:02
Reason:	Install		

Calibration Standards

Cal Gas Concentration:	5.03	ppm	Cal Gas Exp Date:	April 16, 2022
Cal Gas Cylinder #:	DT0018680			
Removed Cal Gas Conc:	5.03	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326169
Converter make:	CDN-101	Converter serial #:	587
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	N/A	1.004939	Backgd or Offset:	N/A
Calibration intercept:	N/A	0.140790	Coeff or Slope:	N/A
				1.05
				0.954

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	79.5	80.0	80.4	0.995
second point	4960	39.8	40.0	40.6	0.986
third point	4980	19.9	20.0	20.3	0.986
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	79.5	80.0	80.5	0.994
SO2 Scrubber Check	4920	79.8	798.0		----
Date of last scrubber change:				Ave Corr Factor	0.989
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 20, 2023	Previous Calibration:	N/A
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	10:33	End Time (MST):	13:02
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326169

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999976	≥0.995
80.0	80.4	0.9948			
40.0	40.6	0.9862	Slope	1.004939	0.90 - 1.10
20.0	20.3	0.9862			
			Intercept	0.140790	+/-3



TRS Calibration Plot

Date: June 20, 2023

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	June 14, 2023	Last Cal Date:	May 17, 2023
Start time (MST):	9:45	End time (MST):	12:50
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH ₄ Cal Gas Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
C ₃ H ₈ Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
Removed C ₃ H ₈ Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.160E-04	2.150E-04	NMHC SP Ratio:	4.48E-05
CH ₄ Retention time:	13.40	13.40	NMHC Peak Area:	204301
				201202

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	17.10	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.19	0.999
second point	4960	39.9	8.59	8.64	0.994
third point	4980	20.0	4.30	4.31	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.14	1.002

Average Correction Factor				0.998
Baseline Corr AF:	17.10	Prev response	17.19	*% change -0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4920	79.8	9.15	9.00	1.017
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	79.8	9.15	9.13	1.002
second point	4960	39.9	4.57	4.57	1.001
third point	4980	20.0	2.29	2.28	1.008
as left zero	5000	0	0.00	0.00	----
as left span	4920	79.8	9.15	9.09	1.006
Average Correction Factor					1.003
Baseline Corr AF:	9.00	Prev response	9.13	*% change	-1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	8.10	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.06	0.996
second point	4960	39.9	4.01	4.07	0.987
third point	4980	20.0	2.01	2.03	0.991
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.05	0.997
Average Correction Factor					0.991
Baseline Corr AF:	8.10	Prev response	8.06	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001793	1.001387
THC Cal Offset:	-0.014598	0.007200
CH ₄ Cal Slope:	1.004345	1.004460
CH ₄ Cal Offset:	-0.005170	0.011429
NMHC Cal Slope:	0.999528	0.998704
NMHC Cal Offset:	-0.009829	-0.004029

Notes: Changed the inlet filter after and the H2 cylinder as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 14, 2023	Previous Calibration:	May 17, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:45	End Time (MST):	12:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999991	≥ 0.995
17.17	17.19	0.9990			
8.59	8.64	0.9941			
4.30	4.31	0.9997			
			Slope	1.001387	0.90 - 1.10
			Intercept	0.007200	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 14, 2023	Previous Calibration:	May 17, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:45	End Time (MST):	12:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999974	≥ 0.995
8.03	8.06	0.9957			
4.01	4.07	0.9866			
2.01	2.03	0.9907			
			Slope	1.004460	0.90 - 1.10
			Intercept	0.011429	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 14, 2023	Previous Calibration:	May 17, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:45	End Time (MST):	12:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
9.15	9.13	1.0019						
4.57	4.57	1.0005				Slope	0.998704	0.90 - 1.10
2.29	2.28	1.0076						
			Intercept	-0.004029	± 0.5			



NMHC Calibration Plot

Date: June 14, 2023

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: June 16, 2023 Last Cal Date: May 3, 2023
Start time (MST): 8:29 End time (MST): 15:39
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC424183 Cal Gas Expiry Date: April 16, 2023
NOX Cal Gas Conc: 48.60 ppm NO Cal Gas Conc: 48.60 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.60 ppm Removed Gas NO Conc: 48.60 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3806
ZAG make/model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1218153460
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.303	1.338	NO bkgnd or offset:	3.4	3.5
NOX coeff or slope:	0.988	0.988	NOX bkgnd or offset:	4.3	4.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	189.8	189.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999560	0.999180
NO _x Cal Offset:	1.487988	-2.628921
NO Cal Slope:	0.999544	1.003252
NO Cal Offset:	0.468701	-4.029566
NO ₂ Cal Slope:	0.998379	0.994339
NO ₂ Cal Offset:	0.058523	0.394612



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.9	-0.3	-0.6	----	----
as found span	4918	82.3	799.9	799.9	0.0	775.1	774.0	1.1	1.0320	1.0335
as found 2nd	4959	41.2	400.4	400.4	0.0	383.8	382.6	1.2	1.0434	1.0466
as found 3rd	4980	20.6	200.2	200.2	0.0	189.8	188.5	1.3	1.0548	1.0621
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.8	-0.2	1.1	----	----
high point	4918	82.3	799.9	799.9	0.0	798.8	800.8	-1.9	1.0014	0.9989
second point	4959	41.2	400.4	400.4	0.0	394.4	394.6	-0.2	1.0153	1.0148
third point	4980	20.6	200.2	200.2	0.0	194.9	193.8	1.0	1.0272	1.0331
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
as left span	4918	82.3	799.9	386.3	413.6	802.2	389.2	413.0	0.9971	0.9926
Average Correction Factor									1.0147	1.0156

Corrected As found	NO _x = 776.0 ppb	NO = 774.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -3.2%
Previous Response	NO _x = 801.0 ppb	NO = 800.0 ppb		*Percent Change	NO = -3.3%
Baseline Corr 2nd pt	NO _x = 384.7 ppb	NO = 382.9 ppb	As found	NO _x r ² : 0.999960	Nx SI: 0.971078
Baseline Corr 3rd pt	NO _x = 190.7 ppb	NO = 188.8 ppb	As found	NO r ² : 0.999930	NO SI: 0.969448
			As found	NO ₂ r ² : 0.999994	NO ₂ SI: 1.002022
					NO ₂ Int: -0.301

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	-0.6	----	----
as found GPT point (400 ppb NO ₂)	772.7	373.2	399.5	400.0	0.9988	100.1%
as found GPT point (200 ppb NO ₂)	772.7	578.4	194.3	194.1	1.0010	99.9%
as found GPT point (100 ppb NO ₂)	772.7	673.0	99.7	100.2	0.9950	100.5%
1st GPT point (400 ppb O ₃)	797.0	383.4	413.6	412.0	1.0039	99.6%
2nd GPT point (200 ppb O ₃)	797.0	594.0	203.0	201.9	1.0054	99.5%
3rd GPT point (100 ppb O ₃)	797.0	693.2	103.8	102.9	1.0087	99.1%
Average Correction Factor					1.0060	99.4%

Notes: Changed sample inlet filter and zero/span valve after multi point as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 16, 2023	Previous Calibration:	May 3, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	8:29	End Time (MST):	15:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.8	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.9	798.8	1.0014		
400.4	394.4	1.0153		
200.2	194.9	1.0272		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 16, 2023	Previous Calibration:	May 3, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	8:29	End Time (MST):	15:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.9	800.8	0.9989		
400.4	394.6	1.0148		
200.2	193.8	1.0331		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 16, 2023	Previous Calibration:	May 3, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	8:29	End Time (MST):	15:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	1.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
413.6	412.0	1.0039		
203.0	201.9	1.0054		
103.8	102.9	1.0087		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: June 16, 2023

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	June 22, 2023	Last Cal Date:	May 30, 2023
Start time (MST):	9:25	End time (MST):	10:50
Reason:	Removal		

Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	Teledyne API T700	Serial Number:	3806
ZAG Make/Model:	Teledyne API T701	Serial Number:	201

Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #:	3869
Analyzer Range:	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006657		Backgd or Offset:	-2.0	
Calibration intercept:	0.560000		Coeff or Slope:	1.011	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	-32.4	----
as found span	4895	909.7	400.0	405.0	0.988
as found 2nd point	4895	757.2	200.0	204.0	0.980
as found 3rd point	4895	656.9	100.0	102.5	0.976
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

Average Correction Factor					
Baseline Corr As found:	437.4	Previous response	403.2	*% change	7.8%
Baseline Corr 2nd AF pt:	236.4	AF Slope:	1.076657	AF Intercept:	-18.640000
Baseline Corr 3rd AF pt:	134.9	AF Correlation:	0.995352		

* = > +/-5% change initiates investigation


Notes: Removal calibration, will replace instrument due to poor performance.

Calibration Performed By: Denny Ray Estador

O₃ Calibration Plot

Date: June 22, 2023

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: June 22, 2023 Last Cal Date: May 30, 2023
 Start time (MST): 10:50 End time (MST): 13:05
 Reason: Install

Calibration Standards

O₃ generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701H Serial Number: 201

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 7046
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		1.000371	Backgd or Offset:		-0.2
Calibration intercept:		1.760000	Coeff or Slope:		1.018

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.2	----
high point	4895	904.3	400.0	401.0	0.998
second point	4895	756.7	200.0	203.0	0.985
third point	4895	656.1	100.0	103.1	0.970
as left zero	5000	800.0	0.0	0.0	----
as left span	4895	904.3	400.0	405.5	0.986
Average Correction Factor					0.984

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Install calibration. Adjusted both zero and span.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 22, 2023	Previous Calibration:	May 30, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:50	End Time (MST):	13:05
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999928	
400.0	401.0	0.9975			≥0.995
200.0	203.0	0.9852	Slope	1.000371	
100.0	103.1	0.9699			0.90 - 1.10
			Intercept	1.760000	+/- 5



O₃ Calibration Plot

Date: June 22, 2023

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: June 13, 2023 Last Cal Date: May 30, 2023
 Start time (MST): 10:45 End time (MST): 11:35

Analyzer Make: Teledyne API T640 S/N: 325
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	22.9	23.2	22.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	711.5	711.80	711.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.04	5.030	5.04	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>June 13, 2023</u>	Last Cal Date: <u>May 30, 2023</u>			
	PM w/o HEPA: <u>110</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>April 27, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>April 27, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: October 6, 2022
 Date RH/T Sensor Cleaned: October 6, 2022

Notes: Verified flow, temperature, and pressure. No adjustments made. Leak test passed.

Calibration by: Devin Russell



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23 FORT HILLS JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 6, 2023	Previous Calibration:	May 10, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:10	End Time (MST):	12:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999995	
799.1	802.4	0.9959			≥0.995
400.1	400.8	0.9981	Slope	1.004713	
200.0	199.2	1.0042			0.90 - 1.10
			Intercept	-0.844401	+/-30



SO2 Calibration Plot

Date: June 6, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Hills Station number: AMS23
 Calibration Date: June 29, 2023 Last Cal Date: May 25, 2023
 Start time (MST): 9:56 End time (MST): 14:00
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.20 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517372
 Removed Cal Gas Conc: 5.20 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 451
 ZAG Make/Model: API T701 Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 43iQ TLE Analyzer serial #: 1300156232
 Converter make: CDN-101 Converter serial #: 594
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005029	1.006026	Backgd or Offset: 1.92	1.93
Calibration intercept:	-0.038092	0.002053	Coeff or Slope: 1.132	1.132

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4923	77.0	80.0	77.3	1.034
as found 2nd point	4962	38.5	40.0	39.7	1.005
as found 3rd point	4981	19.2	19.9	19.6	1.013
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4923	77.0	80.0	80.4	0.995
second point	4962	38.5	40.0	40.4	0.990
third point	4981	19.2	19.9	20.1	0.992
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.0	80.0	81.3	0.984
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.993
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 77.4 Prev response: 80.37 *% change: -3.8%
 Baseline Corr 2nd AF pt: 39.8 AF Slope: 0.967321 AF Intercept: 0.281225
 Baseline Corr 3rd AF pt: 19.7 AF Correlation: 0.999752

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 29, 2023	Previous Calibration:	May 25, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:56	End Time (MST):	14:00
Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	1300156232

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥ 0.995
80.0	80.4	0.9951		
40.0	40.4	0.9900	Slope	0.90 - 1.10
19.9	20.1	0.9924		
			Intercept	+/-3



TRS Calibration Plot

Date: June 29, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	June 6, 2023	Last Cal Date:	May 10, 2023
Start time (MST):	9:10	End time (MST):	12:00
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
Removed C ₃ H ₈ Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.32E-04	2.32E-04	NMHC SP Ratio:	5.06E-05
CH ₄ Retention time:	13.0	13.0	NMHC Peak Area:	181940

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	17.09	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.12	1.004
second point	4960	40.2	8.61	8.52	1.010
third point	4980	20.1	4.30	4.25	1.013
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.19	17.10	1.005

Average Correction Factor				1.009
Baseline Corr AF:	17.09	Prev response	17.18	*% change -0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.10	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.11	1.006
second point	4960	40.2	4.59	4.58	1.001
third point	4980	20.1	2.29	2.29	0.999
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	9.16	9.09	1.008
Average Correction Factor					1.002
Baseline Corr AF:	9.10	Prev response	9.15	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	8.00	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.01	1.003
second point	4960	40.2	4.02	3.94	1.021
third point	4980	20.1	2.01	1.96	1.029
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.03	8.02	1.002
Average Correction Factor					1.017
Baseline Corr AF:	8.00	Prev response	8.03	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999810	0.995947
THC Cal Offset:	-0.013170	-0.024349
CH ₄ Cal Slope:	1.003198	0.998403
CH ₄ Cal Offset:	-0.029231	-0.034620
NMHC Cal Slope:	0.996713	0.993893
NMHC Cal Offset:	0.016062	0.009871

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 6, 2023	Previous Calibration:	May 10, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:10	End Time (MST):	12:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999988	≥ 0.995
17.19	17.12	1.0044			
8.61	8.52	1.0104			
4.30	4.25	1.0128			
			Slope	0.995947	0.90 - 1.10
			Intercept	-0.024349	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 6, 2023	Previous Calibration:	May 10, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:10	End Time (MST):	12:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999887	≥ 0.995
8.03	8.01	1.0027			
4.02	3.94	1.0212			
2.01	1.96	1.0285			
			Slope	0.998403	0.90 - 1.10
			Intercept	-0.034620	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 6, 2023	Previous Calibration:	May 10, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:10	End Time (MST):	12:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999993	≥ 0.995			
9.16	9.11	1.0059						
4.59	4.58	1.0014				Slope	0.993893	0.90 - 1.10
2.29	2.29	0.9995						
			Intercept	0.009871	± 0.5			



NMHC Calibration Plot

Date: June 6, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Hills
Calibration Date: June 1, 2023
Start time (MST): 9:09
Reason: Routine
Station number: AMS23
Last Cal Date: May 16, 2023
End time (MST): 13:39

Calibration Standards

NO Gas Cylinder #: CC332703
NOX Cal Gas Conc: 49.7 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 49.7 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T750
ZAG make/model: Teledyne API T751H
Cal Gas Expiry Date: January 28, 2024
NO Cal Gas Conc: 49.7 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 49.7 ppm
NO gas Diff:
Serial Number: 275
Serial Number: 307

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1152430007

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.049	1.072	NO bkgnd or offset:	2.9	3
NOX coeff or slope:	0.995	0.995	NOX bkgnd or offset:	3.1	3.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.7	162.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999094	0.998879
NO _x Cal Offset:	1.084161	1.384352
NO Cal Slope:	1.000637	0.999937
NO Cal Offset:	-0.356041	-0.035916
NO ₂ Cal Slope:	1.001549	1.003749
NO ₂ Cal Offset:	-0.718730	0.418698



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.8	0.1	0.7	----	----
as found span	4920	80.5	800.2	800.2	0.0	785.4	780.4	5.1	1.019	1.025
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.7	0.1	0.6	----	----
high point	4920	80.5	800.2	800.2	0.0	799.7	799.6	0.1	1.001	1.001
second point	4960	40.2	399.6	399.6	0.0	402.7	401.1	1.6	0.992	0.996
third point	4980	20.1	199.8	199.8	0.0	200.4	198.5	1.9	0.997	1.006
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
as left span	4920	80.5	800.2	444.7	355.5	801.1	442.6	358.4	0.999	1.005
Average Correction Factor									0.997	1.001

Corrected As found	NO _x = 784.6 ppb	NO = 780.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -2.0%	
Previous Response	NO _x = 800.5 ppb	NO = 800.3 ppb		*Percent Change	NO = -2.6%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.5	443.0	355.5	357.5	0.994	100.6%
2nd GPT point (200 ppb O3)	798.5	622.7	175.8	176.3	0.997	100.3%
3rd GPT point (100 ppb O3)	798.5	709.5	89.0	89.9	0.990	101.0%
Average Correction Factor					0.994	100.6%

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 1, 2023	Previous Calibration:	May 16, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:09	End Time (MST):	13:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.7	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.2	799.7	1.0006		
399.6	402.7	0.9922		
199.8	200.4	0.9970		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 1, 2023	Previous Calibration:	May 16, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:09	End Time (MST):	13:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.2	799.6	1.0007		
399.6	401.1	0.9962		
199.8	198.5	1.0065		





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 1, 2023	Previous Calibration:	May 16, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:09	End Time (MST):	13:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>			
0.0	0.6	----	Correlation Coefficient	0.999994	≥0.995			
355.5	357.5	0.9944						
175.8	176.3	0.9972				Slope	1.003749	0.90 - 1.10
89.0	89.9	0.9900						
			Intercept	0.418698	+/-20			



NO_x Calibration Plot

Date: June 1, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Hills Station number: AMS 23
Calibration Date: June 6, 2023 Last Cal Date: May 10, 2023
Start time (MST): 12:26 End time (MST): 13:06

Analyzer Make: API T640 S/N: 1546
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388755
Temp/RH standard: Alicat FP-25BT S/N: 388755

Monthly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Measured</u>	<u>As left</u>	<u>Adjusted</u>	(Limits)
T (°C)	18.5	18.2	18.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	733.6	732.5	733.6	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	5.08	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>June 6, 2023</u>	Last Cal Date: <u>May 10, 2023</u>			
	PM w/o HEPA: <u>105</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Post maintenance</u>	<u>As left</u>	<u>Adjusted</u>	(Limits)
PMT Peak Test	12	12	10.8	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>117</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>June 6, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>June 6, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: May 10, 2023
Date RH/T Sensor Cleaned: May 10, 2023

Notes: Leak check passed, adjusted the PMT. Instrument is very dirty due to a lot of site activity, completed the quarterly and annual maintenance.

Calibration by: Max Farrell



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25 WASKŌW OHCI PIMÂTISIWIN JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number:	AMS25
Calibration Date:	June 15, 2023	Last Cal Date:	May 10, 2023
Start time (MST):	9:35	End time (MST):	13:06
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.54	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC437219			
Removed Cal Gas Conc:	50.5	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	747
ZAG Make/Model:	API T701		Serial Number:	4765

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000277	0.999805	Backgd or Offset:	9.7	9.7
Calibration intercept:	0.283925	0.523945	Coeff or Slope:	0.983	0.974

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	79.2	800.5	807.0	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	79.2	800.5	800.4	1.000
second point	4960	39.6	400.3	401.7	0.997
third point	4980	19.8	200.1	200.7	0.997
as left zero	5000	0.0	0.0	0.0	----
as left span	4921	79.2	800.5	801.8	0.998
Average Correction Factor					0.998

Baseline Corr As found:	807.00	Previous response	801.03	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Minor adjustment to span, filter changed after As Found.

Calibration Performed By: Ryan Power



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 10, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	9:35	End Time (MST):	13:06
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999996	≥0.995
800.5	800.4	1.0002			
400.3	401.7	0.9965	Slope	0.999805	0.90 - 1.10
200.1	200.7	0.9972			
			Intercept	0.523945	+/-30



SO2 Calibration Plot

Date: June 15, 2023

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS25
 Calibration Date: June 20, 2023 Last Cal Date: May 26, 2023
 Start time (MST): 6:48 End time (MST): 8:05
 Reason: Removal External Converter swap

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC517099
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 747
 ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
 Converter make: Thermo 43C Converter serial #: 328702539
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005984		Backgd or Offset: 3.3	3.3
Calibration intercept:	0.180000		Coeff or Slope: 1.079	1.079

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4920	80.0	79.5	79.1	1.006
as found 2nd point	4960	40.0	39.7	39.9	0.998
as found 3rd point	4980	20.0	19.9	20.1	0.993
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

SO₂ Scrubber Check

Date of last scrubber change:	Ave Corr Factor
Date of last converter efficiency test:	efficiency

Baseline Corr As found: 79.0 Prev response: 80.11 *% change: -1.4%
 Baseline Corr 2nd AF pt: 39.8 AF Slope: 0.993614 AF Intercept: 0.260000
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999978

* = > +/-5% change initiates investigation


Notes: External Converter swap for Sulphur Study.

Calibration Performed By: Melissa Lemay

H₂S Calibration Plot

Date: June 20, 2023

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS25
 Calibration Date: June 20, 2023 Last Cal Date: May 26, 2023
 Start time (MST): 8:45 End time (MST): 11:35
 Reason: Install External Converter swap

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC517099
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 747
 ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
 Converter make: Global G-150 Converter serial #: 2022-219
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005984	1.012312	Backgd or Offset:	3.3	3.25
Calibration intercept:	0.180000	0.360000	Coeff or Slope:	1.079	1.079

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4920	80.0	79.5	80.8	0.983
second point	4960	40.0	39.7	40.6	0.979
third point	4980	20.0	19.9	20.4	0.974
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	80.0	800.0	823.8	0.971
SO2 Scrubber Check	4921	79.2	800.0	0.0	----
Date of last scrubber change:	20-Jun-23			Ave Corr Factor	0.979
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: NA Prev response: NA *% change: NA
 Baseline Corr 2nd AF pt: NA AF Slope: NA AF Intercept: NA
 Baseline Corr 3rd AF pt: NA AF Correlation: NA

* = > +/-5% change initiates investigation

Notes: External Converter swap for Sulphur Study. Zero adjusted. SOx scrubber checked after the calibrator zero.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 20, 2023	Previous Calibration:	May 26, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	8:45	End Time (MST):	11:35
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999998	
79.5	80.8	0.9834			≥0.995
39.7	40.6	0.9785	Slope	1.012312	
19.9	20.4	0.9737			0.90 - 1.10
			Intercept	0.360000	+/-3



H₂S Calibration Plot

Date: June 20, 2023

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS25
 Calibration Date: June 30, 2023 Last Cal Date: June 20, 2023
 Start time (MST): 5:35 End time (MST): 9:23
 Reason: Maintenance Calibration after Sulphur Study

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC517099
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 747
 ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
 Converter make: Global G-150 Converter serial #: 2022-219
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.012312	0.996347	Backgd or Offset:	3.25
Calibration intercept:	0.360000	0.040000	Coeff or Slope:	1.079

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4920	80.0	79.5	78.7	1.008
as found 2nd point	4960	40.0	39.7	39.8	0.996
as found 3rd point	4980	20.0	19.9	19.7	1.003
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	80.0	79.5	79.1	1.005
second point	4960	40.0	39.7	39.8	0.998
third point	4980	20.0	19.9	19.9	0.998
as left zero	5000	0.0	0.0	0.0	----
as left span	4920	80.0	800.0	799.3	1.001
SO2 Scrubber Check	4921	79.2	800.0	0.0	----
Date of last scrubber change:	20-Jun-23			Ave Corr Factor	1.000
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 78.8 Prev response: 80.79 *% change: -2.5%
 Baseline Corr 2nd AF pt: 39.9 AF Slope: 0.992032 AF Intercept: 0.040000
 Baseline Corr 3rd AF pt: 19.8 AF Correlation: 0.999950

* = > +/-5% change initiates investigation

Notes: Calibration after the Sulphur Study. SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 30, 2023	Previous Calibration:	June 20, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	5:35	End Time (MST):	9:23
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999980	≥0.995
79.5	79.1	1.0045			
39.7	39.8	0.9982	Slope	0.996347	0.90 - 1.10
19.9	19.9	0.9982			
			Intercept	0.040000	+/-3



H₂S Calibration Plot

Date: June 30, 2023

Location: Waskow ohci Pimatisiwin





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS26 CHRISTINA LAKE JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 1, 2023	Previous Calibration:	May 9, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	11:35	End Time (MST):	17:14
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.4	----	Correlation Coefficient	0.999938	
799.0	799.8	0.9990			≥0.995
399.4	395.4	1.0102	Slope	1.001939	
200.2	195.9	1.0220			0.90 - 1.10
			Intercept	-2.456875	+/-30



SO2 Calibration Plot

Date: June 1, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Christina Lake Station number: AMS26
 Calibration Date: June 14, 2023 Last Cal Date: May 30, 2023
 Start time (MST): 11:17 End time (MST): 15:09
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.89 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002466
 Removed Cal Gas Conc: 4.89 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 2447
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 450i Analyzer serial #: 1180030032
 Converter make: NA Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.035031	1.010753	Backgd or Offset:	35.4 35.0
Calibration intercept:	-0.240256	-0.140775	Coeff or Slope:	1.125 1.093

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4918	81.8	80.0	83.0	0.965
as found 2nd point	4959	40.9	40.0	41.2	0.973
as found 3rd point	4979	20.4	20.0	20.4	0.983
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4918	81.8	80.0	80.9	0.989
second point	4959	40.9	40.0	40.0	1.000
third point	4979	20.4	20.0	19.9	1.003
as left zero	5000	0.0	0.0	0.3	----
as left span	4918	81.8	80.0	80.6	0.993
SO2 Scrubber Check	4919	80.6	806.1	0.0	----
Date of last scrubber change:	27-Feb-19			Ave Corr Factor	0.997
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 82.9 Prev response: 82.57 *% change: 0.4%
 Baseline Corr 2nd AF pt: 41.1 AF Slope: 1.037318 AF Intercept: -0.120291
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999966

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after MAF's. Ran scrubber check after calibrator zero. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 14, 2023	Previous Calibration:	May 30, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	11:17	End Time (MST):	15:09
Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
80.0	80.9	0.9889		
40.0	40.0	1.0000	Slope	0.90 - 1.10
20.0	19.9	1.0027		
			Intercept	+/-3



H₂S Calibration Plot

Date: June 14, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Christina Lake
Calibration Date: June 15, 2023
Start time (MST): 14:53
Reason: Routine
Station number: AMS 26
Last Cal Date: May 31, 2023
End time (MST): 19:44

Calibration Standards

NO Gas Cylinder #: T2Y1P4C
NOX Cal Gas Conc: 50.82 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 50.82 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: November 12, 2023
NO Cal Gas Conc: 50.02 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.02 ppm
NO gas Diff:
Serial Number: 2447
Serial Number: 832

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1173480006

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.795	1.887	NO bkgnd or offset:	3.4	3.6
NOX coeff or slope:	0.997	0.995	NOX bkgnd or offset:	4.0	4.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	202.8	212.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003078	1.003570
NO _x Cal Offset:	-1.760000	-1.960000
NO Cal Slope:	1.002199	1.003127
NO Cal Offset:	-2.660000	-3.060000
NO ₂ Cal Slope:	1.007991	1.001715
NO ₂ Cal Offset:	0.778153	-0.246575



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.1	-0.4	----	----
as found span	4920	80.0	813.1	800.3	12.8	776.4	759.2	17.3	1.0473	1.0542
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	0.0	-0.3	----	----
high point	4920	80.0	813.1	800.3	12.8	814.9	801.3	13.7	0.9978	0.9988
second point	4960	40.0	406.6	400.2	6.4	405.2	396.8	8.4	1.0034	1.0085
third point	4980	20.0	203.3	200.1	3.2	200.4	194.6	5.8	1.0144	1.0282
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	0.0	-0.4	----	----
as left span	4920	80.0	813.1	394.4	418.7	814.8	394.3	420.5	0.9979	1.0003
Average Correction Factor									1.0052	1.0118

Corrected As found	NO _x = 776.9 ppb	NO = 759.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -4.8%
Previous Response	NO _x = 813.9 ppb	NO = 799.4 ppb		*Percent Change	NO = -5.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	797.6	391.7	418.7	419.3	0.9986	100.1%
2nd GPT point (200 ppb O ₃)	797.6	591.9	218.5	218.2	1.0014	99.9%
3rd GPT point (100 ppb O ₃)	797.6	697.7	112.7	113.0	0.9973	100.3%
Average Correction Factor					0.9991	100.1%

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 31, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	14:53	End Time (MST):	19:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.1	814.9	0.9978		
406.6	405.2	1.0034		
203.3	200.4	1.0144		
			0.999981	
			1.003570	
			-1.960000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 31, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	14:53	End Time (MST):	19:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	801.3	0.9988		
400.2	396.8	1.0085		
200.1	194.6	1.0282		
			0.999934	
			1.003127	
			-3.060000	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	May 31, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	14:53	End Time (MST):	19:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
418.7	419.3	0.9986		
218.5	218.2	1.0014		
112.7	113.0	0.9973		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: June 15, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Christina Lake	Station Number:	AMS 26
Calibration Date:	June 15, 2023	Prev Cal Date:	August 24, 2022
Start Time (MST):	14:10	End Time (MST):	15:10
Tower Height (m):	10.0	Reason:	Removal

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	Y18361
WS Calibrator:	MetOne 053	Serial Number:	CA 03845

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	0.0	-100.0%
400	39.4	0.0	-100.0%
600	58.6	0.0	-100.0%
800	77.8	0.0	-100.0%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		#DIV/0!	≥0.9995
Calculated slope		#DIV/0!	0.90 - 1.10
Calculated intercept		#DIV/0!	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	W23733
As Found Declination (deg east of True North):	<u>13.3</u>	As Left Declination (deg east of True North):	<u>13.3</u>
Solar noon time (MST):	12:23	Calc Declination*:	13.3 Degrees
Deadband calc:	333.0 degrees (<i>Limit 4 deg</i>)		* - calculated declination as per NOAA website

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	16.8	---
90	20.4	-19.5%
180	29.7	-42.1%
270	36.2	-65.5%
356	40.8	-88.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope		13.770550	0.90 - 1.10
Calculated intercept		-217.254137	+/- 4

Notes: As founds before removal.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Christina Lake	Station Number:	AMS 26
Calibration Date:	June 15, 2023	Prev Cal Date:	NA
Start Time (MST):	15:20	End Time (MST):	17:59
Tower Height (m):	10.0	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	P22395
WS Calibrator:	MetOne 053	Serial Number:	CA 03845

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	<i>≥0.9995</i>
Calculated slope		0.998858	<i>0.90 - 1.10</i>
Calculated intercept		0.034341	<i>+/- 2</i>

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	N13744
As Found Declination (deg east of True North):	<u>13.3</u>	As Left Declination (deg east of True North):	<u>13.3</u>
Solar noon time (MST):	12:23	Calc Declination*:	13.3 Degrees
Deadband calc:	1.9 degrees (<i>Limit 4 deg</i>)	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	1.8	---
90	89.9	0.0%
180	179.4	-0.2%
270	268.6	-0.4%
354	356.9	0.8%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999057	<i>≥0.9995</i>
Calculated slope		0.999057	<i>0.90 - 1.10</i>
Calculated intercept		-0.353341	<i>+/- 4</i>

Notes: WS and WD install. Bearings good. No issues to note. Realigned tower using solar noon.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Christina Lake	Station Number:	AMS 26	
Calibration Date:	June 30, 2023	Prev Cal Date:	June 15, 2023	
Start Time (MST):	13:03	End Time (MST):	13:55	
Tower Height (m):	10.0	Reason:	As Found	To test the sensor

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	NA
WS Calibrator:	MetOne 053	Serial Number:	NA

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	NA	---
200	20.2	NA	
400	39.4	NA	
600	58.6	NA	
800	77.8	NA	

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope			0.90 - 1.10
Calculated intercept			+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	N13744
As Found Declination (deg east of True North):	<u>13.3</u>	As Left Declination (deg east of True North):	<u>13.3</u>
Solar noon time (MST):	13:27	Calc Declination*:	13.3 Degrees
Deadband calc:	5.1 degrees (<i>Limit 4 deg</i>)		* - calculated declination as per NOAA website

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	2.8	---
90	90.9	0.3%
180	179.7	-0.1%
270	269.1	-0.3%
355	354.7	-0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope		1.009074	0.90 - 1.10
Calculated intercept		-2.058121	+/- 4

Notes: Perfomed WD as founds only since it was not trending similar to nearby stations.

Calibration Performed By: Mohammed Kashif



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27
JACKFISH 2/3
JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	June 1, 2023	Last Cal Date:	May 1, 2023
Start time (MST):	9:33	End time (MST):	12:18
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	<u>SG9133974BAL</u>			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3811
ZAG Make/Model:	API 701		Serial Number:	135

Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	12124313138
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.022106	1.006390	Backgd or Offset:	7.9	7.5
Calibration intercept:	-1.838756	-1.738080	Coeff or Slope:	0.990	0.942

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.5	----
as found span	4921	79.1	800.2	843.0	0.949
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4921	79.1	800.2	804.1	0.995
second point	4961	39.5	399.5	400.1	0.999
third point	4980	19.8	200.3	198.1	1.011
as left zero	5000	0.0	0.0	-0.4	----
as left span	4921	79.1	800.2	804.0	0.995
Average Correction Factor					1.002

Baseline Corr As found:	843.50	Previous response	816.01	*% change	3.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Adjusted the span only.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 1, 2023	Previous Calibration:	May 1, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	9:33	End Time (MST):	12:18
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999984	≥0.995
800.2	804.1	0.9951			
399.5	400.1	0.9986	Slope	1.006390	0.90 - 1.10
200.3	198.1	1.0111			
			Intercept	-1.738080	+/-30



SO2 Calibration Plot

Date: June 1, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Jackfish 2/3 Station number: AMS27
 Calibration Date: June 13, 2023 Last Cal Date: May 18, 2023
 Start time (MST): 9:42 End time (MST): 13:25
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.41 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345023
 Removed Cal Gas Conc: 5.41 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3811
 ZAG Make/Model: API 701 Serial Number: 135

Analyzer Information

Analyzer make: API T101 Analyzer serial #: 621
 Converter make: Converter serial #:
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000198	1.003478	Backgd or Offset: 25.7	27.3
Calibration intercept:	-0.137827	-0.077831	Coeff or Slope: 0.970	0.961

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.6	----
as found span	4926	74.1	80.2	83.0	0.973
as found 2nd point	4963	37.0	40.0	41.7	0.974
as found 3rd point	4982	18.5	20.0	21.0	0.981
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4926	74.1	80.2	80.4	0.997
second point	4963	37.0	40.0	40.1	0.998
third point	4982	18.5	20.0	19.9	1.006
as left zero	5000	0.0	0.0	0.0	----
as left span	4926	74.1	80.2	80.1	1.001
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.000
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 82.4 Prev response: 80.05 *% change: 2.8%
 Baseline Corr 2nd AF pt: 41.1 AF Slope: 1.028423 AF Intercept: 0.522721
 Baseline Corr 3rd AF pt: 20.4 AF Correlation: 0.999995

* = > +/-5% change initiates investigation

Notes: Adjusted both zero and span.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 13, 2023	Previous Calibration:	May 18, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:42	End Time (MST):	13:25
Analyzer make:	API T101	Analyzer serial #:	621

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	0.999995
80.2	80.4	0.9972		
40.0	40.1	0.9984	Slope	1.003478
20.0	19.9	1.0058		
			Intercept	-0.077831
				+/-3



H₂S Calibration Plot

Date: June 13, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Jackfish 2/3
Calibration Date: June 14, 2023
Start time (MST): 9:56
Reason: Maintenance
Station number: AMS27
Last Cal Date: May 25, 2023
End time (MST): 12:47

Calibration Standards

NO Gas Cylinder #: T2Y1P35
NOX Cal Gas Conc: 51.44 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 51.44 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.40 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.40 ppm
NO gas Diff:
Serial Number: 3811
Serial Number: 135

Analyzer Information

Analyzer make: API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 722

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.080	1.080	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.072	1.072	NOX bkgnd or offset:	0.7	0.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.3	3.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.004933	
NO _x Cal Offset:	-2.155062	
NO Cal Slope:	1.007298	
NO Cal Offset:	-2.599343	
NO ₂ Cal Slope:	0.999563	
NO ₂ Cal Offset:	0.617993	



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.3	0.4	----	----
as found span	4921	79.4	816.8	800.3	16.5	817.9	798.6	19.3	0.9987	1.0021
as found 2nd	4921	39.7	411.7	403.3	8.3	409.1	398.7	10.3	1.0063	1.0117
as found 3rd	4980	19.8	203.7	199.6	4.1	200.8	195.8	5.0	1.0145	1.0194
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

Average Correction Factor

Corrected As found	NO _x = 817.8 ppb	NO = 798.9 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -0.1%
Previous Response	NO _x = 818.7 ppb	NO = 803.5 ppb			*Percent Change	NO = -0.6%
Baseline Corr 2nd pt	NO _x = 409.0 ppb	NO = 399.0 ppb	As found	NO _x r ² : 0.999973	Nx SI: 1.002133	Nx Int: -1.834
Baseline Corr 3rd pt	NO _x = 200.7 ppb	NO = 196.1 ppb	As found	NO r ² : 0.999967	NO SI: 0.999076	NO Int: -2.282
			As found	NO ₂ r ² : 0.999936	NO ₂ SI: 0.996666	NO ₂ Int: 1.545

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	0.4	----	----
as found GPT point (400 ppb NO ₂)	804.6	416.5	404.6	404.4	1.0005	99.9%
as found GPT point (200 ppb NO ₂)	804.6	624.9	196.2	196.7	0.9975	100.2%
as found GPT point (100 ppb NO ₂)	804.6	696.0	125.1	128.2	0.9759	102.5%
1st GPT point (400 ppb O ₃)						
2nd GPT point (200 ppb O ₃)						
3rd GPT point (100 ppb O ₃)						

Average Correction Factor

Notes:

Conducted MPAFs for pump replacement.


Calibration Performed By:

Denny Ray Estador

NO_x Calibration Plot

Date: June 14, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Jackfish 2/3
Calibration Date: June 15, 2023
Start time (MST): 9:56
Reason: Maintenance
Station number: AMS27
Last Cal Date: June 14, 2023
End time (MST): 13:05

Calibration Standards

NO Gas Cylinder #: T2Y1P35
NOX Cal Gas Conc: 51.44 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 51.44 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.40 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.40 ppm
NO gas Diff:
Serial Number: 3811
Serial Number: 135

Analyzer Information

Analyzer make: API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 722

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.080	1.099	NO bkgnd or offset:	0.2	-0.9
NOX coeff or slope:	1.072	1.082	NOX bkgnd or offset:	0.7	-0.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	3.9	4.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:		1.000905
NO _x Cal Offset:		-1.315764
NO Cal Slope:		1.008010
NO Cal Offset:		-1.998764
NO ₂ Cal Slope:		0.996976
NO ₂ Cal Offset:		0.633893



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4921	79.4	816.8	800.3	16.5	817.0	805.6	11.5	0.9998	0.9934
second point	4960	39.7	408.5	400.2	8.3	406.5	400.7	5.9	1.0048	0.9988
third point	4980	19.8	203.7	199.6	4.1	201.5	197.0	4.5	1.0110	1.0132
as left zero	5000	0.0	0.0	0.0	0.0	0.7	1.1	-0.4	----	----
as left span	4921	79.4	816.8	417.2	413.4	816.0	416.5	399.6	1.0010	1.0017
Average Correction Factor									1.0052	1.0018

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* => +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	801.1	404.2	413.4	412.5	1.0022	99.8%
2nd GPT point (200 ppb O3)	801.1	615.6	202.0	202.3	0.9986	100.1%
3rd GPT point (100 ppb O3)	801.1	713.0	104.6	105.6	0.9907	100.9%
Average Correction Factor					0.9972	100.3%

Notes:

Adjusted both zero and span. Pump was replaced.

Calibration Performed By:

Denny Ray Estador



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	June 14, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:56	End Time (MST):	13:05
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
816.8	817.0	0.9998		
408.5	406.5	1.0048		
203.7	201.5	1.0110		
			0.999988	
			1.000905	
			-1.315764	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	June 14, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:56	End Time (MST):	13:05
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	805.6	0.9934		
400.2	400.7	0.9988		
199.6	197.0	1.0132		
			0.999972	
			1.008010	
			-1.998764	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	June 14, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:56	End Time (MST):	13:05
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
413.4	412.5	1.0022		
202.0	202.3	0.9986		
104.6	105.6	0.9907		



NO_x Calibration Plot

Date: June 15, 2023

Location: Jackfish 2/3





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS29
SURMONT 2
JUNE 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	June 1, 2023	Last Cal Date:	May 10, 2023
Start time (MST):	10:06	End time (MST):	11:45
Reason:	As Found		

Calibration Standards

Cal Gas Concentration:	<u>49.21</u>	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC356008</u>			
Removed Cal Gas Conc:	<u>49.21</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	5472
ZAG Make/Model:	Teledyne API T701		Serial Number:	4297

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002228		Backgd or Offset:	12.1	12.1
Calibration intercept:	-1.885142		Coeff or Slope:	0.916	0.916

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4919	81.3	800.1	787.1	1.017
as found 2nd point	4959	40.7	400.6	392.3	1.021
as found 3rd point	4979	20.3	199.8	194.7	1.026
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

Average Correction Factor

Baseline Corr As found:	787.40	Previous response	800.00	*% change	-1.6%
Baseline Corr 2nd AF pt:	392.60	AF Slope:	0.984591	AF Intercept:	-1.284908
Baseline Corr 3rd AF pt:	195.00	AF Correlation:	0.999992		

* = > +/-5% change initiates investigation


Notes: MPAF's completed for pump changeout tomorrow.

Calibration Performed By: Braiden Boutilier

SO2 Calibration Plot

Date: June 1, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 6, 2023	Previous Calibration:	June 1, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:00	End Time (MST):	13:48
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	0.999977	
800.1	801.0	0.9989			≥0.995
400.6	398.7	1.0047	Slope	1.001929	
199.8	196.9	1.0148			0.90 - 1.10
			Intercept	-1.605425	+/-30



SO2 Calibration Plot

Date: June 6, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	June 7, 2023	Last Cal Date:	May 9, 2023
Start time (MST):	10:19	End time (MST):	15:15
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	<u>5.391</u>	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	<u>CC508338</u>			
Removed Cal Gas Conc:	<u>5.391</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>CC508338</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	5472
ZAG Make/Model:	Teledyne API T701		Serial Number:	4297

Analyzer Information

Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170
Converter make:	Global	Converter serial #:	2022-223
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008319	0.999622	Backgd or Offset:	0.82
Calibration intercept:	-0.382705	-0.002939	Coeff or Slope:	1.034
				1.046

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4926	74.2	80.0	79.2	1.009
as found 2nd point	4963	37.2	40.1	39.4	1.015
as found 3rd point	4982	18.6	20.1	19.6	1.018
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4926	74.2	80.0	79.9	1.001
second point	4963	37.2	40.1	40.2	0.998
third point	4982	18.6	20.1	20.1	0.998
as left zero	5000	0.0	0.0	0.0	----
as left span	4926	74.2	80.0	79.7	1.004
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.999
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	79.3	Prev response:	80.29	*% change:	-1.2%
Baseline Corr 2nd AF pt:	39.5	AF Slope:	0.991611	AF Intercept:	-0.222254
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999986		

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SOx scrubber check done after calibrator zero. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 7, 2023	Previous Calibration:	May 9, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:19	End Time (MST):	15:15
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999992	≥0.995
80.0	79.9	1.0013			
40.1	40.2	0.9977	Slope	0.999622	0.90 - 1.10
20.1	20.1	0.9977			
			Intercept	-0.002939	+/-3



H₂S Calibration Plot

Date: June 7, 2023


Location: Surmont 2



THC Calibration Plot

Date: June 1, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.1	----	----
as found span	4916	84.2	799.2	799.2	0.0	805.0	804.0	0.9	0.9928	0.9940
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.2	0.2	----	----
high point	4916	84.2	799.2	799.2	0.0	806.0	805.0	0.7	0.9916	0.9928
second point	4958	42.1	399.6	399.6	0.0	404.0	402.7	1.4	0.9891	0.9923
third point	4979	21.1	200.3	200.3	0.0	201.1	199.5	1.7	0.9959	1.0039
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
as left span	4916	84.2	799.2	404.0	395.2	804.0	406.2	398.1	0.9940	0.9946
Average Correction Factor									0.9922	0.9963

Corrected As found	NO _x = 804.9 ppb	NO = 803.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.5%
Previous Response	NO _x = 801.0 ppb	NO = 801.6 ppb		*Percent Change	NO = 0.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	801.0	405.8	395.2	393.8	1.0036	99.6%
2nd GPT point (200 ppb O3)	801.0	615.7	185.3	183.1	1.0120	98.8%
3rd GPT point (100 ppb O3)	801.0	708.3	92.7	92.6	1.0011	99.9%
Average Correction Factor					1.0056	99.5%

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 21, 2023	Previous Calibration:	May 11, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:25	End Time (MST):	14:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.2	806.0	0.9916		
399.6	404.0	0.9891		
200.3	201.1	0.9959		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 21, 2023	Previous Calibration:	May 11, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:25	End Time (MST):	14:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999989	≥0.995
799.2	805.0	0.9928			
399.6	402.7	0.9923	Slope	1.008047	0.90 - 1.10
200.3	199.5	1.0039			
			Intercept	-0.733798	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 21, 2023	Previous Calibration:	May 11, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:25	End Time (MST):	14:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
395.2	393.8	1.0036		
185.3	183.1	1.0120		
92.7	92.6	1.0011		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: June 21, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name:	Surmont 2	Station number:	AMS 29
Calibration Date:	June 21, 2023	Last Cal Date:	April 26, 2023
Start time (MST):	10:13	End time (MST):	11:56
Analyzer Make:	API T640	S/N:	253
Particulate Fraction:	PM2.5		
Flow Meter Make/Model:	Alicat FP-25BT	S/N:	388750
Temp/RH standard:	Alicat FP-25BT	S/N:	388750

Monthly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Measured</u>	<u>As left</u>	<u>Adjusted</u>	(Limits)
T (°C)	12.0	11.36	12.0	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	709.1	710.47	709.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.005	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>June 21, 2023</u>	Last Cal Date: <u>April 26, 2023</u>			
	PM w/o HEPA: <u>2.5</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Post maintenance</u>	<u>As left</u>	<u>Adjusted</u>	(Limits)
PMT Peak Test	11.9	12.0	11.0	<input checked="" type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>2.5</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>June 21, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>June 21, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned:	<u>September 30, 2022</u>
Date RH/T Sensor Cleaned:	<u>October 6, 2022</u>

Notes:

Adjusted PMT peak, both leak checks passed.

Calibration by: Braiden Boutilier



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Surmont 2	Station Number:	AMS 29
Calibration Date:	June 7, 2023	Prev Cal Date:	December 16, 2022
Start Time (MST):	12:22	End Time (MST):	14:20
Tower Height (m):	10.0	Reason:	Routine

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	N10022
WS Calibrator:	MetOne 053	Serial Number:	CA 03988

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.2%
400	39.4	39.5	0.5%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999996	≥0.9995
Calculated slope		0.998856	0.90 - 1.10
Calculated intercept		0.000701	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	P22885
As Found Declination (deg east of True North):	<u>13.2</u>	As Left Declination (deg east of True North):	<u>13.2</u>
Solar noon time (MST):	12:22	Calc Declination*:	13.2 Degrees
Deadband calc:	-4.7 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	-0.2	---
90	84.8	-1.5%
180	178.6	-0.4%
270	274.9	1.4%
357	361.5	1.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.978289	≥0.9995
Calculated slope		3.412775	0.90 - 1.10
Calculated intercept			+/- 4

Notes: WD sensor out of compliance. Will replace sensor on June 8. Tower verified and aligned with solar noon.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Surmont 2	Station Number:	AMS 29
Calibration Date:	June 8, 2023	Prev Cal Date:	June 7, 2023
Start Time (MST):	11:06	End Time (MST):	11:18
Tower Height (m):	10.0	Reason:	Removal

Wind Speed Information

Sensor make/model:	Serial Number:
WS Calibrator:	Serial Number:

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope			0.90 - 1.10
Calculated intercept			+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	P22885
As Found Declination (deg east of True North):	<u>13.2</u>	As Left Declination (deg east of True North):	<u>13.2</u>
Solar noon time (MST):	12:22	Calc Declination*:	13.2 Degrees
Deadband calc:	-5.4 degrees (<i>Limit 4 deg</i>)	* - calculated declination as per NOAA website	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	-0.4	---
90	85.8	-1.2%
180	179.6	-0.1%
270	275.4	1.5%
357	362.0	1.4%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope		0.977306	0.90 - 1.10
Calculated intercept		3.015779	+/- 4

Notes: As founds before removal.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Surmont 2	Station Number:	AMS 29
Calibration Date:	June 8, 2023	Prev Cal Date:	June 7, 2023
Start Time (MST):	11:18	End Time (MST):	12:10
Tower Height (m):	10.0	Reason:	Install

Wind Speed Information

Sensor make/model:	Serial Number:
WS Calibrator:	Serial Number:

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope			0.90 - 1.10
Calculated intercept			+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	A2005
As Found Declination (deg east of True North):	<u>13.2</u>	As Left Declination (deg east of True North):	<u>13.2</u>
Solar noon time (MST):	12:22	Calc Declination*:	13.2 Degrees
Deadband calc:	0.7 degrees (<i>Limit 4 deg</i>)	* - calculated declination as per NOAA website	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	-0.2	---
90	88.6	-0.4%
180	179.6	-0.1%
270	270.0	0.0%
357	356.1	-0.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope		0.999883	0.90 - 1.10
Calculated intercept		0.604273	+/- 4

Notes: WD install. No issues to note. Realigned tower using solar noon.

Calibration Performed By: Braiden Boutilier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS30
ELLS RIVER
JUNE 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	June 2, 2023	Last Cal Date:	May 5, 2023
Start time (MST):	9:16	End time (MST):	12:16
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.53	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC494126			
Removed Cal Gas Conc:	50.53	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3061
ZAG Make/Model:	API T701H		Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008699	1.010127	Backgd or Offset:	8.9	9.5
Calibration intercept:	-2.796032	-2.796062	Coeff or Slope:	0.988	0.988

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	79.2	800.4	805.0	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4921	79.2	800.4	807.0	0.992
second point	4960	39.6	400.2	400.0	1.001
third point	4980	19.8	200.1	197.0	1.016
as left zero	5000	0.0	0.0	-0.3	----
as left span	4921	79.2	800.4	812.0	0.986
Average Correction Factor					1.003

Baseline Corr As found:	805.00	Previous response	804.53	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Adjusted the zero only.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 2, 2023	Previous Calibration:	May 5, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:16	End Time (MST):	12:16
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999956	≥0.995
800.4	807.0	0.9918			
400.2	400.0	1.0006	Slope	1.010127	0.90 - 1.10
200.1	197.0	1.0158			
			Intercept	-2.796062	+/-30



SO2 Calibration Plot

Date: June 2, 2023

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Ells River Station number: AMS30
 Calibration Date: June 12, 2023 Last Cal Date: May 10, 2023
 Start time (MST): 9:16 End time (MST): 13:16
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.08 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002443
 Removed Cal Gas Conc: 5.08 ppm Rem Gas Exp Date:
 Removed Gas Cyl #: Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3061
 ZAG Make/Model: API T701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1410661331
 Converter make: CDN - 101 Converter serial #: 562
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007352	1.003207	Backgd or Offset: 1.58	1.59
Calibration intercept:	-0.159242	-0.139148	Coeff or Slope: 1.107	1.136

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	78.7	80.0	78.0	1.025
as found 2nd point	4961	39.4	40.0	39.1	1.024
as found 3rd point	4980	19.7	20.0	19.3	1.037
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	78.7	80.0	80.2	0.997
second point	4961	39.4	40.0	39.9	1.003
third point	4980	19.7	20.0	19.7	1.016
as left zero	5000	0.0	0.0	0.1	----
as left span	4921	78.7	80.0	80.0	1.000
SO2 Scrubber Check	4921	79.2	800.4	0.0	----
Date of last scrubber change:	N/A		Ave Corr Factor		1.005
Date of last converter efficiency test:	N/A		95.1% efficiency		

Baseline Corr As found: 78.0 Prev response: 80.39 *% change: -3.1%
 Baseline Corr 2nd AF pt: 39.1 AF Slope: 0.976483 AF Intercept: -0.078748
 Baseline Corr 3rd AF pt: 19.3 AF Correlation: 0.999987

* = > +/-5% change initiates investigation

Notes: Adjusted the span only.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 12, 2023	Previous Calibration:	May 10, 2023
Station Name:	Ells River	Station Number:	AMS30
Start Time (MST):	9:16	End Time (MST):	13:16
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999959	
80.0	80.2	0.9971			≥0.995
40.0	39.9	1.0032	Slope	1.003207	
20.0	19.7	1.0161			0.90 - 1.10
			Intercept	-0.139148	+/-3



TRS Calibration Plot

Date: June 12, 2023

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	June 2, 2023	Last Cal Date:	May 26, 2023
Start time (MST):	9:16	End time (MST):	12:16
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm		
Diff between cyl (CH ₄):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701H	Serial Number:	3061
		Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	0.000238	0.000241	NMHC SP Ratio:	4.39E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	207620
				202169

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.66	1.022
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.04	0.999
second point	4960	39.6	8.51	8.41	1.013
third point	4980	19.8	4.26	4.13	1.031
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.06	0.998

Average Correction Factor				1.014
Baseline Corr AF:	16.66	Prev response	17.00	*% change -2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	8.87	1.028
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.11	1.000
second point	4960	39.6	4.56	4.50	1.012
third point	4980	19.8	2.28	2.20	1.036
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.12	0.999
Average Correction Factor					1.016
Baseline Corr AF:	8.87	Prev response	9.14	*% change	-3.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.79	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.93	0.998
second point	4960	39.6	3.96	3.91	1.013
third point	4980	19.8	1.98	1.93	1.026
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.94	0.997
Average Correction Factor					1.012
Baseline Corr AF:	7.79	Prev response	7.86	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000973	1.002469
THC Cal Offset:	-0.047539	-0.074938
CH ₄ Cal Slope:	0.995687	1.003253
CH ₄ Cal Offset:	-0.021757	-0.032957
NMHC Cal Slope:	1.005200	1.001801
NMHC Cal Offset:	-0.024582	-0.041781

Notes: Adjusted the span only.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 2, 2023	Previous Calibration:	May 26, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:16	End Time (MST):	12:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999910	≥ 0.995			
17.03	17.04	0.9994						
8.51	8.41	1.0127				Slope	1.002469	0.90 - 1.10
4.26	4.13	1.0313						
			Intercept	-0.074938	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 2, 2023	Previous Calibration:	May 26, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:16	End Time (MST):	12:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999912	≥ 0.995			
7.91	7.93	0.9982						
3.96	3.91	1.0130				Slope	1.003253	0.90 - 1.10
1.98	1.93	1.0259						
			Intercept	-0.032957	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 2, 2023	Previous Calibration:	May 26, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:16	End Time (MST):	12:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999904	≥ 0.995
9.11	9.11	1.0005			
4.56	4.50	1.0122			
2.28	2.20	1.0361			
			Slope	1.001801	0.90 - 1.10
			Intercept	-0.041781	+/-0.5



NMHC Calibration Plot

Date: June 2, 2023

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Ells River
Calibration Date: June 6, 2023
Start time (MST): 8:49
Reason: Routine
Station number: AMS 30
Last Cal Date: May 11, 2023
End time (MST): 13:20

Calibration Standards

NO Gas Cylinder #: T2Y1P2R
NOX Cal Gas Conc: 50.83 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 50.83 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 49.97 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 49.97 ppm
NO gas Diff:
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 710321429

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.051	1.051	NO bkgnd or offset:	12.7	12.7
NOX coeff or slope:	0.990	0.990	NOX bkgnd or offset:	12.9	12.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	181.2	180.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000351	0.996431
NO _x Cal Offset:	-1.160000	-1.740000
NO Cal Slope:	1.001858	0.993668
NO Cal Offset:	-2.240000	-2.100000
NO ₂ Cal Slope:	0.998225	1.000832
NO ₂ Cal Offset:	0.085999	0.789762



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	----	----
as found span	4920	80.0	813.3	799.5	13.8	809.5	794.3	15.2	1.0047	1.0066
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
high point	4920	80.0	813.3	799.5	13.8	809.8	793.7	16.2	1.0043	1.0073
second point	4960	40.0	406.6	399.8	6.9	401.8	393.3	8.5	1.0120	1.0164
third point	4980	20.0	203.3	199.9	3.4	199.5	194.8	4.7	1.0191	1.0261
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
as left span	4920	80.0	813.3	435.5	377.8	810.0	425.0	384.9	1.0040	1.0248
Average Correction Factor									1.0118	1.0166

Corrected As found	NO _x = 809.5 ppb	NO = 794.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.4%	
Previous Response	NO _x = 812.4 ppb	NO = 798.8 ppb		*Percent Change	NO = -0.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	788.0	424.0	377.8	378.5	0.9980	100.2%
2nd GPT point (200 ppb O3)	788.0	609.8	192.0	193.1	0.9941	100.6%
3rd GPT point (100 ppb O3)	788.0	697.0	104.8	106.6	0.9827	101.8%
Average Correction Factor					0.9916	100.8%

Notes:

No adjustments made.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 6, 2023	Previous Calibration:	May 11, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:49	End Time (MST):	13:20
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.3	809.8	1.0043		
406.6	401.8	1.0120		
203.3	199.5	1.0191		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 6, 2023	Previous Calibration:	May 11, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:49	End Time (MST):	13:20
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
799.5	793.7	1.0073			
399.8	393.3	1.0164			
199.9	194.8	1.0261			
			Slope	0.993668	0.90 - 1.10
			Intercept	-2.100000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 6, 2023	Previous Calibration:	May 11, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:49	End Time (MST):	13:20
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.0	0.0	----	Correlation Coefficient	0.999978	≥0.995			
377.8	378.5	0.9980						
192.0	193.1	0.9941				Slope	1.000832	0.90 - 1.10
104.8	106.6	0.9827						
			Intercept	0.789762	+/-20			



NO_x Calibration Plot

Date: June 6, 2023

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Ells River Station number: AMS 30
 Calibration Date: June 29, 2023 Last Cal Date: June 16, 2023
 Start time (MST): 8:49 End time (MST): 9:45

Analyzer Make: API T640 S/N: 875
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat S/N: 388751
 Temp/RH standard: Alicat S/N: 388751

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	23.6	23.32	23.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	727.1	729.1	727.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.05	5.13	5.05	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>June 29, 2023</u>	Last Cal Date: <u>May 10, 2023</u>			
	PM w/o HEPA: <u>18.8</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	7.2	11.2	11	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>54.7</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>June 29, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>June 29, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: June 29, 2023
 Date RH/T Sensor Cleaned: June 29, 2023

Notes: Adjusted PMT Peak Test. Inlet head still clean.

Calibration by: Denny Ray Estador



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS506
JACKFISH 1
JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	May 18, 2023
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	8:09	End Time (MST):	10:49
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.5	----	Correlation Coefficient	0.999974	
800.2	799.6	1.0008			≥0.995
400.2	398.9	1.0031	Slope	1.001415	
200.1	195.7	1.0223			0.90 - 1.10
			Intercept	-2.176005	+/-30



SO2 Calibration Plot

Date: June 28, 2023

Location: Jackfish 1





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Jackfish 1 Station number: AMS506
 Calibration Date: June 22, 2023 Last Cal Date: May 25, 2023
 Start time (MST): 8:39 End time (MST): 12:24
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.14 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511843
 Removed Cal Gas Conc: 5.14 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 2659
 ZAG Make/Model: API 701 Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1180540020
 Converter make: Global G150 Converter serial #: 2022-218
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.015003	1.015290	Backgd or Offset: 3.33	3.35
Calibration intercept:	-0.378084	-0.438102	Coeff or Slope: 1.066	1.066

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4922	77.8	80.0	79.5	1.002
as found 2nd point	4961	38.9	40.0	39.7	1.000
as found 3rd point	4981	19.4	19.9	19.4	1.012
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4922	77.8	80.0	80.8	0.990
second point	4961	38.9	40.0	40.2	0.995
third point	4981	19.4	19.9	19.6	1.017
as left zero	5000	0.0	0.0	-0.3	----
as left span	4922	77.8	80.0	80.6	0.992
SO2 Scrubber Check	4921	79.2	792.0	0.0	----

Date of last scrubber change:	24-Feb-23	Ave Corr Factor	1.001
Date of last converter efficiency test:	December 1, 2022	efficiency	

Baseline Corr As found: 79.8 Prev response: 80.8 *% change: -1.3%
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 0.998717 AF Intercept: -0.358434
 Baseline Corr 3rd AF pt: 19.7 AF Correlation: 0.999988

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after multi-point as founds. Scrubber test done and passed after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 22, 2023	Previous Calibration:	May 25, 2023
Station Name:	Jackfish 1	Station Number:	AMS506
Start Time (MST):	8:39	End Time (MST):	12:24
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999982	≥0.995
80.0	80.8	0.9899			
40.0	40.2	0.9948	Slope	1.015290	0.90 - 1.10
19.9	19.6	1.0174			
			Intercept	-0.438102	+/-3



H₂S Calibration Plot

Date: June 22, 2023

Location: Jackfish 1





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
as found span	4916	84.4	801.1	799.9	1.2	798.8	794.0	4.9	1.0028	1.0074
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.2	----	----
high point	4916	84.4	801.1	799.9	1.2	800.8	800.4	0.3	1.0003	0.9993
second point	4958	42.2	400.5	400.0	0.6	400.1	398.6	1.5	1.0011	1.0034
third point	4979	21.1	200.3	200.0	0.3	196.9	194.8	2.1	1.0172	1.0266
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
as left span	4916	84.4	801.1	407.3	393.8	790.8	399.9	390.9	1.0130	1.0185
Average Correction Factor									1.0062	1.0098

Corrected As found	NO _x = 799.2 ppb	NO = 794.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.1%	
Previous Response	NO _x = 799.7 ppb	NO = 798.2 ppb		*Percent Change	NO = -0.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	791.1	398.5	393.8	396.2	0.9939	100.6%
2nd GPT point (200 ppb O3)	791.1	599.6	192.7	193.0	0.9983	100.2%
3rd GPT point (100 ppb O3)	791.1	691.3	101.0	101.2	0.9978	100.2%
Average Correction Factor					0.9967	100.3%

Notes: Changed inlet filter after as founds. Adjusted and span only. Used 2nd NO reference point due to drift.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 26, 2023
Station Name:	Jackfish 1	Station Number:	AMS506
Start Time (MST):	8:21	End Time (MST):	12:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
801.1	800.8	1.0003			
400.5	400.1	1.0011			
200.3	196.9	1.0172			
			Slope	1.001347	0.90 - 1.10
			Intercept	-1.568006	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 26, 2023
Station Name:	Jackfish 1	Station Number:	AMS506
Start Time (MST):	8:21	End Time (MST):	12:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999952	≥0.995
799.9	800.4	0.9993			
400.0	398.6	1.0034	Slope	1.002797	0.90 - 1.10
200.0	194.8	1.0266			
			Intercept	-2.507983	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	June 27, 2023	Previous Calibration:	May 26, 2023
Station Name:	Jackfish 1	Station Number:	AMS506
Start Time (MST):	8:21	End Time (MST):	12:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
393.8	396.2	0.9939		
192.7	193.0	0.9983		
101.0	101.2	0.9978		
			0.999995	
			1.006674	
			-0.458123	

NO₂ Calibration Curve





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS508
KIRBY NORTH
JUNE 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

July 31, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 18, 2023	Previous Calibration:	May 18, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	13:45	End Time (MST):	19:17
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999994	≥0.995
799.6	804.0	0.9946			
400.3	400.7	0.9991	Slope	1.005551	0.90 - 1.10
199.7	200.2	0.9973			
			Intercept	-0.648981	+/-30



SO2 Calibration Plot

Date: June 18, 2023

Location: Kirby North





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Kirby North Station number: AMS508
 Calibration Date: June 14, 2023 Last Cal Date: May 17, 2023
 Start time (MST): 10:44 End time (MST): 15:45
 Reason: As Found

Calibration Standards

Cal Gas Concentration: 5.167 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517378
 Removed Cal Gas Conc: 5.167 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3804
 ZAG Make/Model: API T701H Serial Number: 880

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1150840012
 Converter make: Global Converter serial #: 2022-197
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003026		Backgd or Offset:	1.58	
Calibration intercept:	-0.160868		Coeff or Slope:	1.027	

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4923	77.4	80.0	78.9	1.016
as found 2nd point	4961	38.8	40.1	39.5	1.020
as found 3rd point	4981	19.3	19.9	19.5	1.033
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

SO₂ Scrubber Check

Date of last scrubber change:	14-Jun-23	Ave Corr Factor	
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 78.7 Prev response: 80.06 *% change: -1.7%
 Baseline Corr 2nd AF pt: 39.3 AF Slope: 0.985180 AF Intercept: 0.038845
 Baseline Corr 3rd AF pt: 19.3 AF Correlation: 0.999980

* = > +/-5% change initiates investigation


Notes: Completed as founds. Troubleshooting slow response during high humidity. Changed the hydrocarbon kicker. Changed scrubber beads.

Calibration Performed By: Braiden Boutillier

H₂S Calibration Plot

Date: June 14, 2023

Location: Kirby North





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Kirby North	Station number:	AMS508
Calibration Date:	June 15, 2023	Last Cal Date:	June 14, 2023
Start time (MST):	8:20	End time (MST):	11:37
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.167	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	<u>CC517378</u>			
Removed Cal Gas Conc:	5.167	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3804
ZAG Make/Model:	API T701H		Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1150840012
Converter make:	Global	Converter serial #:	2022-197
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003026	1.003887	Backgd or Offset:	1.62	3.65
Calibration intercept:	-0.160868	-0.140994	Coeff or Slope:	1.027	1.066

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4923	77.4	80.0	80.2	0.997
second point	4961	38.8	40.1	40.0	1.002
third point	4981	19.3	19.9	19.9	1.002
as left zero	5000	0.0	0.0	0.1	----
as left span	4923	77.4	80.0	79.8	1.002

SO2 Scrubber Check

Date of last scrubber change:	14-Jun-23	Ave Corr Factor	1.001
Date of last converter efficiency test:		efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: As finds completed June 14. Scrubber check passed. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	June 14, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	8:20	End Time (MST):	11:37
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999995	≥0.995
80.0	80.2	0.9972			
40.1	40.0	1.0024	Slope	1.003887	0.90 - 1.10
19.9	19.9	1.0022			
			Intercept	-0.140994	+/-3



H₂S Calibration Plot

Date: June 15, 2023

Location: Kirby North





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Kirby North Station number: AMS508
 Calibration Date: June 28, 2023 Last Cal Date: June 15, 2023
 Start time (MST): 10:39 End time (MST): 14:58
 Reason: Maintenance

Calibration Standards

Cal Gas Concentration: 5.167 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517378
 Removed Cal Gas Conc: 5.167 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3804
 ZAG Make/Model: API T701H Serial Number: 880

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1150840012
 Converter make: Global Converter serial #: 2022-197
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003887	1.002027	Backgd or Offset: 3.65	2.62
Calibration intercept:	-0.140994	-0.100872	Coeff or Slope: 1.066	1.066

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-1.3	----
as found span	4923	77.4	80.0	78.6	1.001
as found 2nd point	4961	38.8	40.1	38.8	1.000
as found 3rd point	4981	19.3	19.9	18.4	1.012
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4923	77.4	80.0	80.1	0.998
second point	4961	38.8	40.1	39.9	1.005
third point	4981	19.3	19.9	20.0	0.997
as left zero	5000	0.0	0.0	0.1	----
as left span	4923	77.4	80.0	80.0	1.000

SO₂ Scrubber Check

Date of last scrubber change:	14-Jun-23	Ave Corr Factor	1.000
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 79.9 Prev response: 80.15 *% change: -0.3%
 Baseline Corr 2nd AF pt: 40.1 AF Slope: 1.000039 AF Intercept: -1.381294
 Baseline Corr 3rd AF pt: 19.7 AF Correlation: 0.999989

* = > +/-5% change initiates investigation

Notes: Calibrated due to baseline drift. Adjusted zero only. Changed sample inlet filter after MPAF's.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	June 28, 2023	Previous Calibration:	June 15, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	10:39	End Time (MST):	14:58
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥0.995
80.0	80.1	0.9985		
40.1	39.9	1.0050	Slope	0.90 - 1.10
19.9	20.0	0.9972		
			Intercept	+/-3



H₂S Calibration Plot

Date: June 28, 2023

Location: Kirby North





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-1.2	-0.1	-1.1	----	----
as found span	4919	81.0	800.1	794.1	6.0	781.1	776.5	4.6	1.0243	1.0227
as found 2nd	4960	40.5	400.0	397.0	3.0	388.4	384.7	3.7	1.0299	1.0320
as found 3rd	4980	20.2	199.5	198.0	1.5	190.9	188.9	1.9	1.0452	1.0483
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

Average Correction Factor

Corrected As found	NO _x = 782.3 ppb	NO = 776.6 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -2.2%
Previous Response	NO _x = 799.2 ppb	NO = 792.9 ppb			*Percent Change	NO = -2.1%
Baseline Corr 2nd pt	NO _x = 389.6 ppb	NO = 384.8 ppb	As found	NO _x r ² : 0.999983	Nx SI: 0.978796	Nx Int: -2.697
Baseline Corr 3rd pt	NO _x = 192.1 ppb	NO = 189.0 ppb	As found	NO r ² : 0.999950	NO SI: 0.979451	NO Int: -2.658
			As found	NO ₂ r ² : 0.999998	NO ₂ SI: 0.998293	NO ₂ Int: -1.384

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	-1.1	----	----
as found GPT point (400 ppb NO ₂)	774.5	393.3	387.2	385.3	1.0049	99.5%
as found GPT point (200 ppb NO ₂)	774.5	590.6	189.9	188.0	1.0101	99.0%
as found GPT point (100 ppb NO ₂)	774.5	687.0	93.5	91.7	1.0196	98.1%
1st GPT point (400 ppb O ₃)						
2nd GPT point (200 ppb O ₃)						
3rd GPT point (100 ppb O ₃)						

Average Correction Factor

Notes:

RAM issues causing analyzer to crash frequently. Replaced this T200 with a 42iQ.


Calibration Performed By:

Braiden Boutilier

NO_x Calibration Plot

Date: June 18, 2023

Location: Kirby North





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Kirby North	Station number:	AMS508
Calibration Date:	June 19, 2023	Last Cal Date:	June 18, 2023
Start time (MST):	7:50	End time (MST):	12:49
Reason:	Install		

Calibration Standards

NO Gas Cylinder #:	T34ULGL	Cal Gas Expiry Date:	March 8, 2025
NOX Cal Gas Conc:	49.39 ppm	NO Cal Gas Conc:	49.02 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	49.39 ppm	Removed Gas NO Conc:	49.02 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	3804
ZAG make/model:	API 701H	Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>	
NO coeff or slope:	1.553	1.553	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	1.001	1.001	NOX bkgnd or offset:	3.3	3.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	422.94	422.94

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001459	1.002974
NO _x Cal Offset:	-2.051855	-3.331924
NO Cal Slope:	1.000875	1.003524
NO Cal Offset:	-1.873766	-4.693702
NO ₂ Cal Slope:	0.995418	1.000557
NO ₂ Cal Offset:	-0.467650	0.483574



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.1	----	----
high point	4919	81.0	800.1	794.1	6.0	801.0	795.4	5.8	0.9989	0.9984
second point	4960	40.5	400.0	397.0	3.0	395.4	388.8	6.6	1.0117	1.0211
third point	4980	20.2	199.5	198.0	1.5	194.4	191.2	3.2	1.0264	1.0357
as left zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.4	----	----
as left span	4919	81.0	800.1	402.3	397.8	794.3	397.9	396.4	1.0073	1.0111
Average Correction Factor									1.0123	1.0184

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	790.3	398.5	397.8	398.6	0.9980	100.2%
2nd GPT point (200 ppb O3)	790.3	601.6	194.7	194.4	1.0015	99.8%
3rd GPT point (100 ppb O3)	790.3	699.6	96.7	98.6	0.9807	102.0%
Average Correction Factor					0.9934	100.7%

Notes:

42iQ installation. Adjusted zero and span. No other maintenance done.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 19, 2023	Previous Calibration:	June 18, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	7:50	End Time (MST):	12:49
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.1	801.0	0.9989		
400.0	395.4	1.0117		
199.5	194.4	1.0264		
			0.999931	
			1.002974	
			-3.331924	





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 19, 2023	Previous Calibration:	June 18, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	7:50	End Time (MST):	12:49
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
794.1	795.4	0.9984			
397.0	388.8	1.0211			
198.0	191.2	1.0357			
			Slope	1.003524	0.90 - 1.10
			Intercept	-4.693702	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	June 19, 2023	Previous Calibration:	June 18, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	7:50	End Time (MST):	12:49
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
397.8	398.6	0.9980			
194.7	194.4	1.0015			
96.7	98.6	0.9807			
			Slope	1.000557	0.90 - 1.10
			Intercept	0.483574	+/-20



NO_x Calibration Plot

Date: June 19, 2023

Location: Kirby North





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	June 18, 2023	Previous Calibration:	May 18, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	13:45	End Time (MST):	19:17
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.00	----	Correlation Coefficient	0.999899	
17.26	17.31	0.9973			≥0.995
8.64	8.51	1.0156	Slope	1.003015	
4.31	4.26	1.0128			0.90 - 1.10
			Intercept	-0.057756	+/-1.5



THC Calibration Plot

Date: June 18, 2023

Location: Kirby North





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Kirby North	Station Number:	AMS 508
Calibration Date:	July 4, 2023	Prev Cal Date:	June 29, 2022
Start Time (MST):	14:00	End Time (MST):	14:47
Tower Height (m):	10.0	Reason:	Removal

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	X16479
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥ 0.9995
Calculated slope		0.998908	$0.90 - 1.10$
Calculated intercept		0.028386	± 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	X16495
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>NA</u>
Solar noon time (MST):	12:28	Calc Declination*:	14 Degrees
Deadband calc:	-1.5 degrees (<i>Limit 4 deg</i>)	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.8	---
90	91.6	0.5%
180	182.4	0.7%
270	272.4	0.7%
357	359.4	0.7%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999997	≥ 0.9995
Calculated slope		0.995847	$0.90 - 1.10$
Calculated intercept		-1.173951	± 4

Notes: Removals completed for the station move to Kirby South.

Calibration Performed By: Braiden Boutilier



End of Report