

**WOOD BUFFALO ENVIRONMENTAL ASSOCIATION
MONTHLY AMBIENT AIR QUALITY MONITORING REPORT
MAY 2023
REPORT HISTORY**

Original report release data: June 30, 2023

Revised report release date: August 30, 2023

Revision 1 – Submission of H₂S data and Revision of TRS data at Patricia McInnes.

The Patricia McInnes air monitoring station (AMS06) collected H₂S data from January 3, 2023, to June 12, 2023, but it was reported as TRS. On August 30, 2023, the data was resubmitted as H₂S to the Electronic Transfer System (ETS) and the TRS data was revised. The monthly report cover letter, station summary, network summary, and calibration report contained within have been revised to reflect the change.



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

MAY 2023

MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

August 30, 2023

Revision 01

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01 BERTHA GANTER - FORT MCKAY MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 24, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:14	End Time (MST):	16:22
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999994	≥0.995
799.9	800.7	0.9991			
400.4	399.4	1.0026	Slope	1.000343	0.90 - 1.10
199.7	200.4	0.9967			
			Intercept	0.027099	+/-30



SO2 Calibration Plot

Date: May 1, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: May 16, 2023 Last Cal Date: April 21, 2023
 Start time (MST): 8:59 End time (MST): 14:26
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153461
 Converter make: CD Nova Converter serial #: 470
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999078	0.993792	Backgd or Offset: 2.26	2.27
Calibration intercept:	0.160000	0.220000	Coeff or Slope: 0.919	0.919

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4921	78.4	80.0	77.6	1.032
as found 2nd point	4960	39.2	40.0	39.4	1.018
as found 3rd point	4980	19.6	20.0	19.9	1.010
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	78.4	80.0	79.7	1.004
second point	4960	39.2	40.0	40.0	1.000
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.3	----
as left span	4921	78.4	80.0	79.2	1.010
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	1.000
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 77.5 Prev response: 80.08 *% change: -3.3%
 Baseline Corr 2nd AF pt: 39.3 AF Slope: 0.967790 AF Intercept: 0.379996
 Baseline Corr 3rd AF pt: 19.8 AF Correlation: 0.999926

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	April 21, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:59	End Time (MST):	14:26
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	1.000000	≥0.995
80.0	79.7	1.0037			
40.0	40.0	1.0000	Slope	0.993792	0.90 - 1.10
20.0	20.1	0.9949			
			Intercept	0.220000	+/-3



TRS Calibration Plot

Date: May 16, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: May 16, 2023 Last Cal Date: April 21, 2023
 Start time (MST): 8:59 End time (MST): 14:26
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1200326167
 Converter make: Thermo Converter Converter serial #: N/A
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993089	1.001375	Backgd or Offset:	1.90	1.92
Calibration intercept:	0.521594	0.081613	Coeff or Slope:	1.001	1.001

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4921	78.4	80.0	80.7	0.989
as found 2nd point	4960	39.2	40.0	40.3	0.987
as found 3rd point	4980	19.6	20.0	20.1	0.985
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	78.4	80.0	80.1	0.999
second point	4961	39.2	40.0	40.3	0.992
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.6	----
as left span	4921	78.4	80.0	78.7	1.016
SO2 Scrubber Check	4919	81.3	813.0	0.1	----
Date of last scrubber change:	March 21, 2022			Ave Corr Factor	0.995
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 80.9 Prev response: 79.95 *% change: 1.2%
 Baseline Corr 2nd AF pt: 40.5 AF Slope: 1.011277 AF Intercept: -0.160000
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999999

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	April 21, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:59	End Time (MST):	14:26
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999987	≥0.995
80.0	80.1	0.9987			
40.0	40.3	0.9923	Slope	1.001375	0.90 - 1.10
20.0	20.1	0.9949			
			Intercept	0.081613	+/-3



H₂S Calibration Plot

Date: May 16, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	May 1, 2023	Last Cal Date:	April 26, 2023
Start time (MST):	10:14	End time (MST):	16:22
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC486642	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	497.7 ppm	CH ₄ Equiv Conc.	1063.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.6 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.7 ppm	CH ₄ Equiv Conc.	1063.1 ppm
Removed C ₃ H ₈ Conc.	205.6 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	5609

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.82E-04	3.00E-04	NMHC SP Ratio:	6.11E-05
CH ₄ Retention time:	14.6	14.6	NMHC Peak Area:	150450
				144534

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.29	17.04	1.015
as found 2nd point	4959	40.7	8.65	8.42	1.027
as found 3rd point	4980	20.3	4.32	4.19	1.029
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	17.29	17.28	1.000
second point	4959	40.7	8.65	8.64	1.001
third point	4980	20.3	4.32	4.37	0.988
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.29	17.91	0.966

Average Correction Factor				0.997
Baseline Corr AF:	17.04	Prev response	17.27	*% change -1.3%
Baseline Corr 2nd AF:	8.4	AF Slope:	0.986010	AF Intercept: -0.045298
Baseline Corr 3rd AF:	4.2	AF Correlation:	0.999950	* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4918	81.3	9.19	9.08	1.013
as found 2nd point	4959	40.7	4.60	4.53	1.016
as found 3rd point	4980	20.3	2.30	2.28	1.008
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4918	81.3	9.19	9.19	1.001
second point	4959	40.7	4.60	4.63	0.993
third point	4980	20.3	2.30	2.36	0.975
as left zero	5000	0	0.00	0.00	----
as left span	4918	81.3	9.19	9.60	0.957
Average Correction Factor					0.990
Baseline Corr AF:	9.08	Prev response	9.21	*% change	-1.5%
Baseline Corr 2nd AF:	4.5	AF Slope:	0.986667	AF Intercept:	0.001464
Baseline Corr 3rd AF:	2.3	AF Correlation:	0.999992	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	7.96	1.017
as found 2nd point	4959	40.7	4.05	3.89	1.041
as found 3rd point	4980	20.3	2.02	1.92	1.055
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	8.09	8.09	1.000
second point	4959	40.7	4.05	4.01	1.010
third point	4980	20.3	2.02	2.01	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.30	0.975
Average Correction Factor					1.005
Baseline Corr AF:	7.96	Prev response	8.06	*% change	-1.2%
Baseline Corr 2nd AF:	3.89	AF Slope:	0.985108	AF Intercept:	-0.046961
Baseline Corr 3rd AF:	1.92	AF Correlation:	0.999807	* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000528	0.998379
THC Cal Offset:	-0.030303	0.020702
CH ₄ Cal Slope:	0.999037	0.999517
CH ₄ Cal Offset:	-0.028760	-0.010960
NMHC Cal Slope:	1.001827	0.997203
NMHC Cal Offset:	-0.001743	0.031861

Notes: Changed the inlet filter and H₂/N₂ cylinders after as founds. Performed an input board calibration. Adjusted H₂, N₂, and air flow rates. Captured a new zero chromatogram. Adjusted span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 26, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:14	End Time (MST):	16:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995			
17.29	17.28	1.0005						
8.65	8.64	1.0012				Slope	0.998379	0.90 - 1.10
4.32	4.37	0.9879						
			Intercept	0.020702	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 26, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:14	End Time (MST):	16:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999969	≥ 0.995
8.09	8.09	1.0002			
4.05	4.01	1.0103			
2.02	2.01	1.0038			
			Slope	0.999517	0.90 - 1.10
			Intercept	-0.010960	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 26, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:14	End Time (MST):	16:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999944	≥ 0.995			
9.19	9.19	1.0009						
4.60	4.63	0.9932				Slope	0.997203	0.90 - 1.10
2.30	2.36	0.9747						
			Intercept	0.031861	± 0.5			



NMHC Calibration Plot

Date: May 1, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
Calibration Date: May 18, 2023 Last Cal Date: April 18, 2023
Start time (MST): 8:33 End time (MST): 13:28
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1P9L Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.84 ppm NO Cal Gas Conc: 50.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.84 ppm Removed Gas NO Conc: 50.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3565
ZAG make/model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1218153357
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.498	1.465	NO bkgnd or offset:	7.1	7.4
NOX coeff or slope:	0.990	0.991	NOX bkgnd or offset:	7.2	7.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	201.6	201.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999185	0.999073
NO _x Cal Offset:	0.160000	0.200000
NO Cal Slope:	0.999572	0.999971
NO Cal Offset:	-0.480000	-0.620000
NO ₂ Cal Slope:	0.999865	1.001305
NO ₂ Cal Offset:	-0.124708	0.605886



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.5	0.4	0.1	----	----
as found span	4920	80.0	813.4	800.6	12.8	834.3	819.2	15.1	0.9750	0.9773
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.6	0.4	0.2	----	----
high point	4920	80.0	813.4	800.6	12.8	813.0	800.3	12.8	1.0005	1.0004
second point	4960	40.0	406.7	400.3	6.4	406.6	399.8	6.8	1.0003	1.0013
third point	4980	20.0	203.4	200.2	3.2	202.8	198.1	4.7	1.0028	1.0104
as left zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
as left span	4920	80.0	813.4	396.0	417.4	811.5	395.7	415.7	1.0024	1.0009
Average Correction Factor									1.0012	1.0040

Corrected As found	NO _x = 833.8 ppb	NO = 818.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 2.5%
Previous Response	NO _x = 812.9 ppb	NO = 799.8 ppb		*Percent Change	NO = 2.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.5	391.9	417.4	418.5	0.9974	100.3%
2nd GPT point (200 ppb O3)	796.5	591.2	218.1	218.7	0.9973	100.3%
3rd GPT point (100 ppb O3)	796.5	696.3	113.0	114.5	0.9869	101.3%
Average Correction Factor					0.9938	100.6%

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 18, 2023	Previous Calibration:	April 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:33	End Time (MST):	13:28
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.6	----	Correlation Coefficient	≥0.995	
813.4	813.0	1.0005			
406.7	406.6	1.0003			
203.4	202.8	1.0028			
			Slope	0.999073	0.90 - 1.10
			Intercept	0.200000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 18, 2023	Previous Calibration:	April 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:33	End Time (MST):	13:28
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.4	----	Correlation Coefficient	≥0.995	
800.6	800.3	1.0004			
400.3	399.8	1.0013			
200.2	198.1	1.0104			
			Slope	0.999971	0.90 - 1.10
			Intercept	-0.620000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 18, 2023	Previous Calibration:	April 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:33	End Time (MST):	13:28
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
417.4	418.5	0.9974			
218.1	218.7	0.9973			
113.0	114.5	0.9869			
			Slope	1.001305	0.90 - 1.10
			Intercept	0.605886	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 18, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: May 4, 2023 Last Cal Date: April 5, 2023
 Start time (MST): 9:36 End time (MST): 13:02
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 1107
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001057	0.999429	Backgd or Offset:	2.5	3.2
Calibration intercept:	0.440000	0.500000	Coeff or Slope:	1.040	1.010

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.6	----
as found span	5000	855.5	400.0	414.8	0.964
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.5	----
high point	5000	858.8	400.0	400.2	1.000
second point	5000	743.1	200.0	200.6	0.997
third point	5000	653.5	100.0	100.3	0.997
as left zero	5000	0.0	0.0	-0.2	----
as left span	5000	858.8	400.0	402.0	0.995
Average Correction Factor					0.998

Baseline Corr As found:	414.2	Previous response	400.9	*% change	3.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 4, 2023	Previous Calibration:	April 5, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:36	End Time (MST):	13:02
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.5	----	Correlation Coefficient	≥0.995
400.0	400.2	0.9995		
200.0	200.6	0.9970	Slope	0.90 - 1.10
100.0	100.3	0.9970		
			Intercept	+/- 5



O₃ Calibration Plot

Date: May 4, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
 Calibration Date: May 15, 2023 Last Cal Date: April 20, 2023
 Start time (MST): 9:50 End time (MST): 11:50

Analyzer Make: API T640 S/N: 306
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	24.7	24.4	24.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	735.9	739	735.9	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	4.98	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>May 15, 2023</u>	Last Cal Date: <u>April 20, 2023</u>			
	PM w/o HEPA: <u>3.2</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	6	-60	10.9	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>3.2</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>May 15, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>May 15, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: August 31, 2022
 Date RH/T Sensor Cleaned: May 15, 2023

Notes: Verified flow, temperature, and pressure. Leak check passed. PMT peak test was low. Cleaned the optical chamber and RH/T sensor. Replaced the disposable filter. Inlet head cleaned. Post maintenance PMT peak test was displaying -60 peaks. Removing the instrument.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
 Calibration Date: May 15, 2023 Last Cal Date: NA
 Start time (MST): 12:00 End time (MST): 13:15

Analyzer Make: API T640 S/N: 328
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	26.5	25.7	26.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	736.1	738	736.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.02	5.00	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: May 15, 2023 Last Cal Date: NA
 PM w/o HEPA: 60.2 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.9		10.9	<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: _____ w/ HEPA: _____
 Date Optical Chamber Cleaned: _____ <0.2 ug/m3
 Disposable Filter Changed: _____

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: Installing a new T640 instrument. Flow, temperature, and pressure all within limits. Leak check passed. PMT peak test within limits.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-11-2021

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	May 23, 2023	Last Cal Date:	April 27, 2023
Start time (MST):	10:10	End time (MST):	14:30
NH3 Cal Date:	May 23, 2023	Last Cal Date:	April 28, 2023
Start time (MST):	14:55	End time (MST):	18:40
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	50.84	ppm	NO Gas Cylinder #:	T2Y1P9L
NO Cal Gas Conc:	50.04	ppm	NO Cal Gas Expiry:	March 3, 2028
Removed NOX Conc:	50.84	ppm	Removed Cylinder #:	NA
Removed NO Conc:	50.04	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	72.93	ppm	NH3 Gas Cylinder #:	CC281298
			NH3 Cal Gas Expiry:	February 28, 2023
Removed NH3 Conc:	72.93	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	Teledyne API T700		Serial Number:	3565
ZAG make/model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	808
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	5.20
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	472

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.834	0.834	TN coefficient:	0.836	0.836
NOX coefficient:	0.839	0.839	NO bkgnd:	-0.626	-0.626
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-0.402	-0.402
NH3 coefficient:	0.937	0.937	TN bkgnd:	1.871	1.871

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.005241	1.000126
NO _x Cal Offset:	-0.420000	-0.900000
NO Cal Slope:	1.000685	1.000043
NO Cal Offset:	-2.020000	-1.820000
NO ₂ Cal Slope:	1.007356	1.005118
NO ₂ Cal Offset:	1.516016	-0.102116
NH3 Cal Slope:	1.005229	1.002401
NH3 Cal Offset:	-3.943276	-1.184396
TN Cal Slope:	1.009496	1.006409
TN Cal Offset:	-3.914715	-1.167305



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-11-2021

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	-0.1	----	----
as found NO	4920	80.0	813.4	813.4	----	812.7	812.7	0.0	1.001	----
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.1	----	----
high NO point	4920	80.0	813.4	813.4	----	813.9	813.3	0.6	0.999	----
NO/O3 point	4920	80.0	813.4	813.4	----	812.1	811.0	1.2	1.002	----
as found NH3	3413	86.4	1800.6	----	1800.6	1817.2	----	1810.0	0.991	0.995
new NH3 cyl rp							----			
first NH3	3413	86.4	1800.6	----	1800.6	1817.2	----	1810.0	0.991	0.995
second NH3	3452	48.0	1000.2	----	1000.2	990.5	----	986.5	1.010	1.014
third NH3	3476	24.0	500.1	----	500.1	509.7	----	507.5	0.981	0.985
Average Correction Factor									1.0005	0.9980

Corrected As found TN = 812.8 ppb NO_x = 812.8 ppb NH3 = 1810.1 ppb

Previous Response TN = 817.3 ppb NO_x = 817.3 ppb NH3 = 1806.1 ppb

NH3 Previous Converter Efficiency = 93.7%

NH3 Current Converter Efficiency = 93.7%

*Percent Change TN = -0.5%

*Percent Change NO_x = -0.6%

*Percent Change NH3 = 0.2%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

TN Calibration Summary

Version-11-2021


Station Information

Calibration Date:	May 23, 2023	Previous Calibration:	April 27, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:10	End Time (MST):	14:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
1800.6	1817.2	0.9909		
1000.2	990.5	1.0098		
500.1	509.7	0.9811		

TN Calibration Curve





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 23, 2023	Previous Calibration:	April 27, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:10	End Time (MST):	14:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.6	1810.0	0.9948		
1000.2	986.5	1.0139		
500.1	507.5	0.9854		
			0.999822	
			1.002401	
			-1.184396	





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 23, 2023	Previous Calibration:	April 27, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:10	End Time (MST):	14:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.4	813.3	1.0002		
406.7	404.8	1.0047		
203.4	202.0	1.0067		
			0.999993	
			1.000126	
			-0.900000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-11-2021


Station Information

Calibration Date:	May 23, 2023	Previous Calibration:	April 27, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:10	End Time (MST):	14:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
800.6	800.0	1.0008		
400.3	397.0	1.0084		
200.2	196.8	1.0171		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-11-2021


Station Information

Calibration Date:	May 23, 2023	Previous Calibration:	April 27, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:10	End Time (MST):	14:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
411.6	413.8	0.9947		
215.5	215.8	0.9986		
117.0	118.1	0.9907		
			0.999990	
			1.005118	
			-0.102116	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 23, 2023


Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: May 23, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	May 29, 2023	Last Cal Date:	May 23, 2023
Start time (MST):	8:39	End time (MST):	12:00
NH3 Cal Date:	May 29, 2023	Last Cal Date:	May 23, 2023
Start time (MST):	12:23	End time (MST):	14:01
Reason:	Removal		

Calibration Standards

NOX Cal Gas Conc:	50.84	ppm	NO Gas Cylinder #:	T2Y1P9L
NO Cal Gas Conc:	50.04	ppm	NO Cal Gas Expiry:	March 3, 2028
Removed NOX Conc:	50.84	ppm	Removed Cylinder #:	NA
Removed NO Conc:	50.04	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	72.93	ppm	NH3 Gas Cylinder #:	CC281298
			NH3 Cal Gas Expiry:	February 28, 2023
Removed NH3 Conc:	72.93	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	Teledyne API T700		Serial Number:	3565
ZAG make/model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	808
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	5.20
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	486

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.834		TN coefficient:	0.836	
NOX coefficient:	0.839		NO bkgnd:	-0.626	
NO2 coefficient:	1.000		NOX bkgnd:	-0.402	
NH3 coefficient:	0.937		TN bkgnd:	1.871	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000126	1.003836
NO _x Cal Offset:	-0.900000	-1.320000
NO Cal Slope:	1.000043	0.997074
NO Cal Offset:	-1.820000	-1.380000
NO ₂ Cal Slope:	1.005118	1.006204
NO ₂ Cal Offset:	-0.102116	0.774028
NH3 Cal Slope:	1.002401	1.000316
NH3 Cal Offset:	-1.184396	7.361047
TN Cal Slope:	1.006409	1.004132
TN Cal Offset:	-1.167305	7.636840



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	-0.1	----	----
as found NO	4920	80.0	813.4	813.4	----	816.1	816.1	0.0	0.997	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	-0.1	----	----
high NO point	4920	80.0	813.4	813.4	----	816.1	816.1	0.0	0.997	----
NO/O3 point	4920	80.0	813.4	813.4	----	812.5	811.2	1.1	1.001	----
as found NH3	3413	86.4	1800.6	----	1800.6	1809.8	----	1802.8	0.995	0.999
new NH3 cyl rp							----			
first NH3	3413	86.4	1800.6	----	1800.6	1809.8	----	1802.8	0.995	0.999
second NH3	3452	48.0	1000.2	----	1000.2	1019.6	----	1015.4	0.981	0.985
third NH3	3476	24.0	500.1	----	500.1	515.8	----	513.3	0.970	0.974
Average Correction Factor									0.9989	0.9860

Corrected As found TN = 816.2 ppb NO_x = 816.2 ppb NH3 = 1802.9 ppb
 Previous Response TN = 817.5 ppb NO_x = 812.6 ppb NH3 = 1803.8 ppb

*Percent Change TN = -0.2%
 *Percent Change NO_x = 0.4%
 *Percent Change NH3 = 0.0%
 * = > +/-5% change initiates investigation

NH3 Previous Converter Efficiency = 93.7%
 NH3 Current Converter Efficiency =



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	-0.1	----	----
as found span	4920	80.0	813.4	800.6	813.4	816.1	797.3	816.1	0.9967	1.0042
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	-0.1	----	----
high point	4920	80.0	813.4	800.6	813.4	816.1	797.3	816.1	0.9967	1.0042
second point	4960	40.0	406.7	400.3	406.7	405.6	397.8	408.3	1.0028	1.0063
third point	4980	20.0	203.4	200.2	203.4	202.1	196.6	201.5	1.0062	1.0181
Average Correction Factor									1.0019	1.0095

Baseline Corr As fnd	TN = 816.2 ppb	NO _x = 816.2 ppb	NO = 797.5 ppb	*Percent Change	TN = -0.2%
Previous Response	TN = 817.5 ppb	NO _x = 812.6 ppb	NO = 798.9 ppb	*Percent Change	NO _x = 0.4%
				*Percent Change	NO = -0.2%
				* = > +/-5% change initiates investigation	

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	0.2	----	----
calibration zero	----	----	0.0	0.2	----	----
1st GPT point (400 ppb O3)	795.5	393.0	415.3	418.3	0.9928	100.7%
2nd GPT point (200 ppb O3)	795.5	595.6	212.7	215.2	0.9884	101.2%
3rd GPT point (100 ppb O3)	795.5	694.2	114.1	116.1	0.9828	101.8%
Average Correction Factor					0.9880	101.2%

Notes:

Removing the analyzer due to memory issues.

Calibration Performed By:

Rene Chamberland



Wood Buffalo Environmental Association

TN Calibration Summary


Version-05-2023

Station Information

Calibration Date:	May 29, 2023	Previous Calibration:	May 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:39	End Time (MST):	12:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
1800.6	1809.8	0.9949		
1000.2	1019.6	0.9810		
500.1	515.8	0.9695		





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	May 29, 2023	Previous Calibration:	May 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:39	End Time (MST):	12:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
1800.6	1802.8	0.9988		
1000.2	1015.4	0.9850		
500.1	513.3	0.9743		
			0.999899	
			1.000316	
			7.361047	





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	May 29, 2023	Previous Calibration:	May 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:39	End Time (MST):	12:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.4	816.1	0.9967		
406.7	405.6	1.0028		
203.4	202.1	1.0062		
			0.999987	
			1.003836	
			-1.320000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	May 29, 2023	Previous Calibration:	May 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:39	End Time (MST):	12:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.2	----	Correlation Coefficient	0.999988	≥0.995
800.6	797.3	1.0042			
400.3	397.8	1.0063	Slope	0.997074	0.90 - 1.10
200.2	196.6	1.0181			
			Intercept	-1.380000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	May 29, 2023	Previous Calibration:	May 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:39	End Time (MST):	12:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
415.3	418.3	0.9928		
212.7	215.2	0.9884		
114.1	116.1	0.9828		



NO_x Calibration Plot

Date: May 29, 2023

Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: May 29, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	May 30, 2023	Last Cal Date:	May 23, 2023
Start time (MST):	10:03	End time (MST):	14:20
NH3 Cal Date:	May 30, 2023	Last Cal Date:	May 23, 2023
Start time (MST):	14:45	End time (MST):	18:14
Reason:	Install		

Calibration Standards

NOX Cal Gas Conc:	50.84	ppm	NO Gas Cylinder #:	T2Y1P9L
NO Cal Gas Conc:	50.04	ppm	NO Cal Gas Expiry:	March 3, 2028
Removed NOX Conc:	50.84	ppm	Removed Cylinder #:	NA
Removed NO Conc:	50.04	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	72.93	ppm	NH3 Gas Cylinder #:	CC281298
			NH3 Cal Gas Expiry:	February 28, 2023
Removed NH3 Conc:	72.93	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	Teledyne API T700		Serial Number:	3565
ZAG make/model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	475
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	5.00
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	511

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:		0.986	TN coefficient:		0.984
NOX coefficient:		0.986	NO bkgnd:		-3.0
NO2 coefficient:		1.000	NOX bkgnd:		-2.5
NH3 coefficient:		0.941	TN bkgnd:		1.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:		0.998848
NO _x Cal Offset:		-0.520000
NO Cal Slope:		0.998730
NO Cal Offset:		-1.160000
NO ₂ Cal Slope:		0.999748
NO ₂ Cal Offset:		0.471532
NH3 Cal Slope:		0.998065
NH3 Cal Offset:		-0.656397
TN Cal Slope:		1.001095
TN Cal Offset:		-0.131851



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.2	0.3	-0.1	----	----
as found NO	4920	80.0	813.4	813.4	----	812.3	812.4	-0.2	1.001	----
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.3	-0.1	----	----
high NO point	4920	80.0	813.4	813.4	----	812.3	812.4	-0.2	1.001	----
NO/O3 point	4920	80.0	813.4	813.4	----	808.0	808.0	-0.5	1.007	----
as found NH3	3413	86.4	1800.6	----	1800.6	1804.7	----	1798.9	0.998	1.001
new NH3 cyl rp							----			
first NH3	3413	86.4	1800.6	----	1800.6	1804.7	----	1798.9	0.998	1.001
second NH3	3452	48.0	1000.2	----	1000.2	995.7	----	991.9	1.005	1.008
third NH3	3476	24.0	500.1	----	500.1	503.4	----	501.2	0.993	0.998
Average Correction Factor									1.0041	1.0024

Corrected As found TN = 812.1 ppb NO_x = 812.1 ppb NH3 = 1799.0 ppb

Previous Response TN = NA ppb NO_x = NA ppb NH3 = NA ppb

NH3 Previous Converter Efficiency =

NH3 Current Converter Efficiency = 94.1%

*Percent Change TN = NA

*Percent Change NO_x = NA

*Percent Change NH3 = NA

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>		
as found zero												
as found span												
new NO cyl rp												
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.2	----	----		
high point	4920	80.0	813.4	800.6	813.4	812.4	799.5	812.3	1.0013	1.0014		
second point	4960	40.0	406.7	400.3	406.7	405.3	397.0	404.8	1.0035	1.0084		
third point	4980	20.0	203.4	200.2	203.4	201.8	197.9	201.7	1.0077	1.0114		
Average Correction Factor									1.0042	1.0071		
Baseline Corr As fnd	TN =	NA	ppb	NO _x =	NA	ppb	NO =	NA	ppb	*Percent Change	TN =	NA
Previous Response	TN =	NA	ppb	NO _x =	NA	ppb	NO =	NA	ppb	*Percent Change	NO _x =	NA
										*Percent Change	NO =	NA
										<i>* = > +/-5% change initiates investigation</i>		

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero						
calibration zero	----	----	0.0	0.1	----	----
1st GPT point (400 ppb O3)	795.3	395.2	412.9	413.0	0.9998	100.0%
2nd GPT point (200 ppb O3)	795.3	595.7	212.4	213.2	0.9962	100.4%
3rd GPT point (100 ppb O3)	795.3	693.9	114.2	114.9	0.9939	100.6%
Average Correction Factor					0.9966	100.3%

Notes: Installing a new NH3 analyzer. Adjusted both zero and span. Used the 2nd GPT reference point due to drift. Adjusted the NH3 span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TN Calibration Summary

Version-05-2023


Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	May 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:03	End Time (MST):	14:20
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
1800.6	1804.7	0.9977			
1000.2	995.7	1.0045			
500.1	503.4	0.9934			
			Slope	1.001095	0.90 - 1.10
			Intercept	-0.131851	+/-20

TN Calibration Curve





Wood Buffalo Environmental Association

NH₃ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	May 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:03	End Time (MST):	14:20
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.6	1798.9	1.0010		
1000.2	991.9	1.0084		
500.1	501.2	0.9978		
			0.999974	
			0.998065	
			-0.656397	

NH₃ Calibration Curve





Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-05-2023


Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	May 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:03	End Time (MST):	14:20
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.3	----	Correlation Coefficient	≥0.995	
813.4	812.4	1.0013			
406.7	405.3	1.0035			
203.4	201.8	1.0077			
			Slope	0.998848	0.90 - 1.10
			Intercept	-0.520000	+/-20

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	May 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:03	End Time (MST):	14:20
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.6	799.5	1.0014		
400.3	397.0	1.0084		
200.2	197.9	1.0114		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	May 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:03	End Time (MST):	14:20
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
412.9	413.0	0.9998			
212.4	213.2	0.9962			
114.2	114.9	0.9939			
			Slope	0.999748	0.90 - 1.10
			Intercept	0.471532	+/-20



NO_x Calibration Plot

Date: May 30, 2023


Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: May 30, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	May 11, 2023	Last Cal Date:	April 20, 2023
Start time (MST):	9:51	End time (MST):	13:19
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	<u>3040</u>	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>3040</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001765	0.999480	Backgd or Offset:	-0.012	-0.012
Calibration intercept:	0.123811	0.161861	Coeff or Slope:	0.992	0.992

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.6	41.2	0.983
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.6	40.5	1.000
second point	4966	33.3	20.2	20.7	0.977
third point	4983	16.7	10.2	10.3	0.986
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.5	40.3	1.007
Average Correction Factor					0.988

Baseline Corr As found:	41.12	Prev response:	40.75	*% change:	0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 20, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:51	End Time (MST):	13:19
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999833	
40.6	40.5	1.0004			≥0.995
20.2	20.7	0.9773	Slope	0.999480	
10.2	10.3	0.9858			0.90 - 1.10
			Intercept	0.161861	+/-1.5



CO Calibration Plot

Date: May 11, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	May 12, 2023	Last Cal Date:	April 19, 2023
Start time (MST):	9:39	End time (MST):	12:40
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>60,200</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
N2 Gen Make/Model:	Peak Sci NG5000		Serial Number:	7220900034

Analyzer Information

Analyzer make: Teledyne API 360	Analyzer serial #: 442
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000116	1.002095	Backgd or Offset:	0.037	0.037
Calibration intercept:	-5.740000	-6.580000	Coeff or Slope:	0.880	0.880

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-0.7	----
as found span	2920	80.0	1605.3	1597.0	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	-0.3	----
high point	2920	80.0	1605.3	1604.5	1.001
second point	2960	40.0	802.7	796.9	1.007
third point	2980	20.0	401.3	387.8	1.035
as left zero	3000	0.0	0.0	-0.5	----
as left span	2960	40.0	802.7	780.8	1.028
Average Correction Factor					1.014

Baseline Corr As found:	1597.70	Prev response:	1599.78	*% change:	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	May 12, 2023	Previous Calibration	April 19, 2023
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS01
Start Time (MST)	9:39	End Time (MST)	12:40
Analyzer make	Teledyne API 360	Analyzer serial #	442

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999925	≥0.995
1605.3	1604.5	1.0005			
802.7	796.9	1.0072	Slope	1.002095	0.90 - 1.10
401.3	387.8	1.0349			
			Intercept	-6.580000	+/-10



CO₂ Calibration Plot

Date: May 12, 2023

Location: Bertha Ganter-Fort McKay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	May 11, 2023	Last Cal Date:	April 13, 2023
Start time (MST):	9:11 AM	End time (MST):	12:31
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1185
ZAG Make/Model:	API T701		Serial Number:	5608

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999874	1.000972	Backgd or Offset:	18.3	17.9
Calibration intercept:	-0.805548	-1.065310	Coeff or Slope:	0.805	0.793

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4920	80.2	801.6	813.6	0.985
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.2	801.6	801.9	1.000
second point	4960	40.1	400.8	399.7	1.003
third point	4980	20.0	199.9	197.8	1.011
as left zero	5000	0.0	0.0	0.0	----
as left span	4920	80.2	801.6	801.6	1.000
Average Correction Factor					1.004

Baseline Corr As found:	813.80	Previous response	800.74	*% change	1.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 13, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:11	End Time (MST):	12:31
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999990	
801.6	801.9	0.9997			≥0.995
400.8	399.7	1.0028	Slope	1.000972	
199.9	197.8	1.0107			0.90 - 1.10
			Intercept	-1.065310	+/-30



SO2 Calibration Plot

Date: May 11, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mildred Lake Station number: AMS02
 Calibration Date: May 3, 2023 Last Cal Date: April 5, 2023
 Start time (MST): 8:51 End time (MST): 13:08
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.29 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345191
 Removed Cal Gas Conc: 5.29 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1185
 ZAG Make/Model: API T701 Serial Number: 5608

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 12113311966
 Converter make: Global G150 Converter serial #: 2022-198
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.015395	1.006965	Backgd or Offset: 1.75	1.80
Calibration intercept:	-0.059190	-0.239199	Coeff or Slope: 0.823	0.833

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4924	75.6	80.0	80.0	1.001
as found 2nd point	4962	37.8	40.0	39.9	1.005
as found 3rd point	4981	18.9	20.0	19.6	1.025
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4924	75.6	80.0	80.4	0.995
second point	4962	37.8	40.0	40.0	1.000
third point	4981	18.9	20.0	19.6	1.020
as left zero	5000	0.0	0.0	0.0	----
as left span	4924	75.6	80.0	80.0	1.000

SO2 Scrubber Check

Date of last scrubber change:	12-Sep-22	Ave Corr Factor	1.005
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 79.9 Prev response: 81.16 *% change: -1.6%
 Baseline Corr 2nd AF pt: 39.8 AF Slope: 1.000679 AF Intercept: -0.119203
 Baseline Corr 3rd AF pt: 19.5 AF Correlation: 0.999961

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after MPAF's. Ran a SO2 scrubber check after calibrator zero and passed. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 3, 2023	Previous Calibration:	April 5, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	8:51	End Time (MST):	13:08
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999956	
80.0	80.4	0.9949			≥0.995
40.0	40.0	0.9998	Slope	1.006965	
20.0	19.6	1.0202			0.90 - 1.10
			Intercept	-0.239199	+/-3



H₂S Calibration Plot

Date: May 3, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	May 11, 2023	Last Cal Date:	April 13, 2023
Start time (MST):	9:11	End time (MST):	12:31
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
C ₃ H ₈ Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
Removed C ₃ H ₈ Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	5608

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.86E-04	2.90E-04	NMHC SP Ratio:	4.49E-04
CH ₄ Retention time:	14.6	14.6	NMHC Peak Area:	195861
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	16.60	1.013
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.82	1.000
second point	4960	40.1	8.41	8.38	1.003
third point	4980	20.0	4.19	4.17	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.83	0.999
Average Correction Factor					1.003

Baseline Corr AF:	16.60	Prev response	16.81	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	8.72	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.80	1.000
second point	4960	40.1	4.40	4.40	0.999
third point	4980	20.0	2.19	2.20	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.83	0.996
Average Correction Factor					0.998
Baseline Corr AF:	8.72	Prev response	8.80	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	7.88	1.018
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.02	1.000
second point	4960	40.1	4.01	3.98	1.007
third point	4980	20.0	2.00	1.97	1.016
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	8.01	1.002
Average Correction Factor					1.008
Baseline Corr AF:	7.88	Prev response	8.01	*% change	-1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000482	1.000257
THC Cal Offset:	-0.015912	-0.013510
CH ₄ Cal Slope:	1.001482	1.000527
CH ₄ Cal Offset:	-0.021253	-0.017653
NMHC Cal Slope:	0.999608	1.000102
NMHC Cal Offset:	0.004941	0.004544

Notes: Changed inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 13, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:11	End Time (MST):	12:31
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995
16.82	16.82	1.0000			
8.41	8.38	1.0030			
4.19	4.17	1.0051			
			Slope	1.000257	0.90 - 1.10
			Intercept	-0.013510	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 13, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:11	End Time (MST):	12:31
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999977	≥ 0.995
8.02	8.02	1.0004			
4.01	3.98	1.0074			
2.00	1.97	1.0156			
			Slope	1.000527	0.90 - 1.10
			Intercept	-0.017653	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 13, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:11	End Time (MST):	12:31
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
8.80	8.80	0.9996						
4.40	4.40	0.9988				Slope	1.000102	0.90 - 1.10
2.19	2.20	0.9952						
			Intercept	0.004544	± 0.5			



NMHC Calibration Plot

Date: May 11, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	May 30, 2023	Last Cal Date:	May 11, 2023
Start time (MST):	10:00	End time (MST):	14:12
Reason:	Maintenance	troubleshoot elevated daily zero response	

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
C ₃ H ₈ Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
Removed C ₃ H ₈ Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	5608

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.86E-04	2.90E-04	NMHC SP Ratio:	4.49E-04
CH ₄ Retention time:	14.6	14.6	NMHC Peak Area:	195861
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.20	----
as found span	4920	80.2	16.82	16.70	1.007
as found 2nd point	4960	40.1	8.41	8.44	0.996
as found 3rd point	4980	20.0	4.19	4.34	0.967
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.83	0.999
second point	4960	40.1	8.41	8.39	1.003
third point	4980	20.0	4.19	4.18	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.99	0.990
Average Correction Factor					1.002

Baseline Corr AF:	16.50	Prev response	16.81	*% change	-1.9%
Baseline Corr 2nd AF:	8.2	AF Slope:	0.980546	AF Intercept:	0.205223
Baseline Corr 3rd AF:	4.1	AF Correlation:	0.999997	* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	8.62	1.020
as found 2nd point	4960	40.1	4.40	4.32	1.019
as found 3rd point	4980	20.0	2.19	2.16	1.015
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.80	0.999
second point	4960	40.1	4.40	4.41	0.998
third point	4980	20.0	2.19	2.21	0.994
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.92	0.986
Average Correction Factor					0.997
Baseline Corr AF:	8.62	Prev response	8.80	*% change	-2.1%
Baseline Corr 2nd AF:	4.3	AF Slope:	0.979653	AF Intercept:	0.005700
Baseline Corr 3rd AF:	2.2	AF Correlation:	0.999998	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.20	----
as found span	4920	80.2	8.02	8.08	0.993
as found 2nd point	4960	40.1	4.01	4.13	0.971
as found 3rd point	4980	20.0	2.00	2.18	0.919
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.03	0.999
second point	4960	40.1	4.01	3.98	1.009
third point	4980	20.0	2.00	1.97	1.014
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	8.07	0.994
Average Correction Factor					1.007
Baseline Corr AF:	7.88	Prev response	8.01	*% change	-1.6%
Baseline Corr 2nd AF:	3.94	AF Slope:	0.981781	AF Intercept:	0.200123
Baseline Corr 3rd AF:	1.98	AF Correlation:	0.999992	* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000257	1.000855
THC Cal Offset:	-0.013510	-0.012905
CH ₄ Cal Slope:	1.000527	1.001225
CH ₄ Cal Offset:	-0.017653	-0.018850
NMHC Cal Slope:	1.000102	1.000634
NMHC Cal Offset:	0.004544	0.005745

Notes: Changed zero air generator after as founds to improve daily zero's, span adjustment performed.

Calibration Performed By: Kelly Baragar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	May 11, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:00	End Time (MST):	14:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
16.82	16.83	0.9993						
8.41	8.39	1.0029				Slope	1.000855	0.90 - 1.10
4.19	4.18	1.0034						
			Intercept	-0.012905	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	May 11, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:00	End Time (MST):	14:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999969	≥ 0.995			
8.02	8.03	0.9995						
4.01	3.98	1.0087				Slope	1.001225	0.90 - 1.10
2.00	1.97	1.0141						
			Intercept	-0.018850	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	May 11, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:00	End Time (MST):	14:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
8.80	8.80	0.9990						
4.40	4.41	0.9977				Slope	1.000634	0.90 - 1.10
2.19	2.21	0.9938						
			Intercept	0.005745	± 0.5			



NMHC Calibration Plot

Date: May 30, 2023

Location: Mildred Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	May 4, 2023	Last Cal Date:	April 13, 2023
Start time (MST):	8:35	End time (MST):	11:18
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.02	ppm	Cal Gas Exp Date:	September 9, 2028
Cal Gas Cylinder #:	CC470284			
Removed Cal Gas Conc:	50.02	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701		Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001885	1.003456	Backgd or Offset:	22.1	22.1
Calibration intercept:	-0.100000	0.000000	Coeff or Slope:	0.869	0.869

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4920	80.0	800.3	803.5	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.5	----
high point	4920	80.0	800.3	802.9	0.997
second point	4960	40.0	400.2	402.6	0.994
third point	4980	20.0	200.1	199.4	1.003
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	80.0	800.3	803.8	0.996
Average Correction Factor					0.998

Baseline Corr As found:	803.40	Previous response	801.73	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: no maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 4, 2023	Previous Calibration:	April 13, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:35	End Time (MST):	11:18
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999991	≥0.995
800.3	802.9	0.9968			
400.2	402.6	0.9939	Slope	1.003456	0.90 - 1.10
200.1	199.4	1.0034			
			Intercept	0.000000	+/-30



SO2 Calibration Plot

Date: May 4, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: May 16, 2023 Last Cal Date: April 27, 2023
 Start time (MST): 6:03 End time (MST): 7:30
 Reason: Removal

Calibration Standards

Cal Gas Concentration: 5.42 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345266
 Removed Cal Gas Conc: 5.42 ppm Rem Gas Exp Date: January 4, 2025
 Removed Gas Cyl #: CC345266 Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701H Serial Number: 362

Analyzer Information

Analyzer make: Thermo 450i Analyzer serial #: 1336160094
 Converter make: NA Converter serial #: NA
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996073		Backgd or Offset: 19.0	19.0
Calibration intercept:	0.142229		Coeff or Slope: 1.070	1.070

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4926	74.1	80.3	80.4	1.000
as found 2nd point	4963	37.0	40.1	40.5	0.993
as found 3rd point	4982	18.5	20.1	20.0	1.008
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

SO₂ Scrubber Check

Date of last scrubber change:	Ave Corr Factor
Date of last converter efficiency test:	efficiency

Baseline Corr As found: 80.3 Prev response: 80.15 *% change: 0.2%
 Baseline Corr 2nd AF pt: 40.4 AF Slope: 1.000770 AF Intercept: 0.102267
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999971

* = > +/-5% change initiates investigation


Notes: Removed Due to putting in a H₂S external converter and trace level SO₂ instead.

Calibration Performed By: Melissa Lemay

H₂S Calibration Plot

Date: May 16, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	May 16, 2023	Last Cal Date:	
Start time (MST):	8:45	End time (MST):	11:25
Reason:	Install		

Calibration Standards

Cal Gas Concentration:	5.42	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC345266			
Removed Cal Gas Conc:	5.42	ppm	Rem Gas Exp Date:	January 4, 2025
Removed Gas Cyl #:	CC345266		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701H		Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400
Converter make:	Global	Converter serial #:	2022-200
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		1.003759	Backgd or Offset:		2.0
Calibration intercept:		-0.177711	Coeff or Slope:		1.067

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4926	74.1	80.3	80.4	0.999
second point	4963	37.0	40.1	40.3	0.995
third point	4982	18.5	20.1	19.7	1.018
as left zero	5000	0.0	0.0	-0.2	----
as left span	4926	74.1	80.3	79.9	1.005
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	1.004
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: Install Due to going to external H₂S converters. Sox scrubber checked after the calibrator zero. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:45	End Time (MST):	11:25
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥0.995
80.3	80.4	0.9990		
40.1	40.3	0.9952	Slope	0.90 - 1.10
20.1	19.7	1.0179		
			Intercept	+/-3



H₂S Calibration Plot

Date: May 16, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	May 4, 2023	Last Cal Date:	April 6, 2023
Start time (MST):	8:35	End time (MST):	11:17
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC470284	Cal Gas Expiry Date:	September 9, 2028
CH ₄ Cal Gas Conc.	497.8 ppm	CH ₄ Equiv Conc.	1062.9 ppm
C ₃ H ₈ Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.8 ppm	CH ₄ Equiv Conc.	1062.9 ppm
Removed C ₃ H ₈ Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	1.840E-04	1.860E-04	NMHC SP Ratio:	3.820E-05
CH ₄ Retention time:	11.8	11.8	NMHC Peak Area:	236377
				233712

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.01	16.76	1.015
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.01	16.98	1.002
second point	4960	40.0	8.50	8.48	1.003
third point	4980	20.0	4.25	4.21	1.010
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.01	16.96	1.003

Average Correction Factor				1.005
Baseline Corr AF:	16.76	Prev response	17.00	*% change -1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	
as found span	4920	80.0	9.04	8.94	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.04	9.04	1.000
second point	4960	40.0	4.52	4.51	1.002
third point	4980	20.0	2.26	2.24	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.04	9.04	1.000
Average Correction Factor					1.004
Baseline Corr AF:	8.94	Prev response	9.02	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	7.96	7.83	1.017
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	7.96	7.94	1.003
second point	4960	40.0	3.98	3.97	1.003
third point	4980	20.0	1.99	1.98	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	7.96	7.92	1.006
Average Correction Factor					1.004
Baseline Corr AF:	7.83	Prev response	7.96	*% change	-1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001213	0.999063
THC Cal Offset:	-0.032000	-0.016000
CH ₄ Cal Slope:	1.001406	0.997102
CH ₄ Cal Offset:	-0.012000	-0.002000
NMHC Cal Slope:	0.999779	1.000411
NMHC Cal Offset:	-0.020000	-0.010000

Notes: Span adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 4, 2023	Previous Calibration:	April 6, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:35	End Time (MST):	11:17
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995			
17.01	16.98	1.0016						
8.50	8.48	1.0028				Slope	0.999063	0.90 - 1.10
4.25	4.21	1.0099						
			Intercept	-0.016000	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 4, 2023	Previous Calibration:	April 6, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:35	End Time (MST):	11:17
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	≥0.995			
7.96	7.94	1.0031						
3.98	3.97	1.0031				Slope	0.997102	0.90 - 1.10
1.99	1.98	1.0057						
			Intercept	-0.002000	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 4, 2023	Previous Calibration:	April 6, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:35	End Time (MST):	11:17
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995
9.04	9.04	1.0002			
4.52	4.51	1.0024			
2.26	2.24	1.0092			
			Slope	1.000411	0.90 - 1.10
			Intercept	-0.010000	+/-0.5



NMHC Calibration Plot

Date: May 4, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: May 5, 2023
Start time (MST): 6:48
Reason: Routine
Station number: AMS04
Last Cal Date: April 4, 2023
End time (MST): 11:30

Calibration Standards

NO Gas Cylinder #: T36RH1F
NOX Cal Gas Conc: 51.16 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 51.16 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: August 18, 2023
NO Cal Gas Conc: 50.91 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.91 ppm
NO gas Diff:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 723

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.014	1.032	NO bkgnd or offset:	-2.1	-0.3
NOX coeff or slope:	1.013	1.032	NOX bkgnd or offset:	-1.7	1.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	7.6	7.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999617	0.998201
NO _x Cal Offset:	-0.133690	-0.313645
NO Cal Slope:	0.998232	1.000170
NO Cal Offset:	-0.993777	-1.493235
NO ₂ Cal Slope:	0.984274	0.998883
NO ₂ Cal Offset:	-0.102857	1.152412



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	3.0	1.0	1.9	----	----
as found span	4922	78.1	799.1	795.2	3.9	785.1	779.8	5.4	1.0178	1.0197
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.4	0.4	-0.8	----	----
high point	4922	78.1	799.1	795.2	3.9	797.3	795.0	2.3	1.0023	1.0002
second point	4961	39.1	400.1	398.1	2.0	399.1	395.1	4.0	1.0024	1.0076
third point	4981	19.5	199.5	198.5	1.0	198.9	195.6	3.3	1.0030	1.0150
as left zero	5000	0.0	0.0	0.0	0.0	-0.7	0.1	-0.8	----	----
as left span	4922	78.1	799.1	382.7	416.4	801.2	399.3	401.9	0.9974	0.9584
Average Correction Factor									1.0026	1.0076

Corrected As found	NO _x = 782.1 ppb	NO = 778.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -2.1%	
Previous Response	NO _x = 798.7 ppb	NO = 792.8 ppb		*Percent Change	NO = -1.8%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	792.4	379.9	416.4	415.8	1.0015	99.9%
2nd GPT point (200 ppb O3)	792.4	583.2	213.1	215.8	0.9875	101.3%
3rd GPT point (100 ppb O3)	792.4	688.0	108.3	110.8	0.9775	102.3%
Average Correction Factor					0.9888	101.1%

Notes:

Zero and Span adjusted. No maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 5, 2023	Previous Calibration:	April 4, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:48	End Time (MST):	11:30
Analyzer make:	API T200	Analyzer serial #:	723

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.4	----	Correlation Coefficient	≥0.995	
799.1	797.3	1.0023			
400.1	399.1	1.0024			
199.5	198.9	1.0030			
			Slope	0.998201	0.90 - 1.10
			Intercept	-0.313645	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 5, 2023	Previous Calibration:	April 4, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:48	End Time (MST):	11:30
Analyzer make:	API T200	Analyzer serial #:	723

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
795.2	795.0	1.0002		
398.1	395.1	1.0076		
198.5	195.6	1.0150		
			0.999972	
			1.000170	
			-1.493235	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 5, 2023	Previous Calibration:	April 4, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:48	End Time (MST):	11:30
Analyzer make:	API T200	Analyzer serial #:	723

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.8	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
416.4	415.8	1.0015		
213.1	215.8	0.9875		
108.3	110.8	0.9775		



NO_x Calibration Plot

Date: May 5, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
Calibration Date: May 18, 2023 Last Cal Date: May 5, 2023
Start time (MST): 5:50 End time (MST): 9:30
Reason: As Found Moly converter changed

Calibration Standards

NO Gas Cylinder #: T36RH1F Cal Gas Expiry Date: August 18, 2023
NOX Cal Gas Conc: 51.16 ppm NO Cal Gas Conc: 50.91 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 51.16 ppm Removed Gas NO Conc: 50.91 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: API T700 Serial Number: 3808
ZAG make/model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: API T200 Analyzer serial #: 723
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.032	1.032	NO bkgnd or offset:	-0.3	-0.3
NOX coeff or slope:	1.032	1.032	NOX bkgnd or offset:	1.8	1.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	7.7	7.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998201	
NO _x Cal Offset:	-0.313645	
NO Cal Slope:	1.000170	
NO Cal Offset:	-1.493235	
NO ₂ Cal Slope:	0.998883	
NO ₂ Cal Offset:	1.152412	



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.4	-0.1	0.4	----	----
as found span	4922	78.1	799.1	795.2	3.9	784.1	779.3	4.8	1.0191	1.0204
as found 2nd	4961	39.1	400.1	398.1	2.0	391.6	387.3	4.2	1.0216	1.0279
as found 3rd	4981	19.5	199.5	198.5	1.0	195.0	192.0	3.0	1.0231	1.0340
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

Average Correction Factor

Corrected As found	NO _x = 783.7 ppb	NO = 779.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.7%
Previous Response	NO _x = 797.4 ppb	NO = 793.8 ppb		*Percent Change	NO = -1.9%
Baseline Corr 2nd pt	NO _x = 391.2 ppb	NO = 387.4 ppb	As found	NO _x r ² : 0.999996	Nx SI: 0.981013
Baseline Corr 3rd pt	NO _x = 194.6 ppb	NO = 192.1 ppb	As found	NO r ² : 0.999979	NO SI: 0.980801
			As found	NO ₂ r ² : 0.999914	NO ₂ SI: 0.992749
					NO ₂ Int: 2.088

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	0.4	----	----
as found GPT point (400 ppb NO ₂)	780.8	377.5	407.2	405.3	1.0047	99.5%
as found GPT point (200 ppb NO ₂)	780.8	571.5	213.2	215.0	0.9917	100.8%
as found GPT point (100 ppb NO ₂)	780.8	678.6	106.1	108.9	0.9743	102.6%
1st GPT point (400 ppb O ₃)						
2nd GPT point (200 ppb O ₃)						
3rd GPT point (100 ppb O ₃)						

Average Correction Factor

Notes:

As founds to change the Moly converter. Due to drifting during the GPT the 2nd NO ref point used.


Calibration Performed By:

Melissa Lemay

NO_x Calibration Plot

Date: May 18, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: May 19, 2023
Start time (MST): 5:15
Reason: Routine
Station number: AMS04
Last Cal Date: May 18, 2023
End time (MST): 9:38

Calibration Standards

NO Gas Cylinder #: T36RH1F
NOX Cal Gas Conc: 51.16 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 51.16 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: August 18, 2023
NO Cal Gas Conc: 50.91 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.91 ppm
NO gas Diff:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 723

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.032	1.055	NO bkgnd or offset:	-0.3	-0.5
NOX coeff or slope:	1.032	1.056	NOX bkgnd or offset:	1.8	2.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	7.7	7.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998201	0.999316
NO _x Cal Offset:	-0.313645	-0.653591
NO Cal Slope:	1.000170	1.001680
NO Cal Offset:	-1.493235	-1.693638
NO ₂ Cal Slope:	0.998883	1.001373
NO ₂ Cal Offset:	1.152412	-0.003271



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4922	78.1	799.1	795.2	3.9	798.3	795.7	2.5	1.0010	0.9994
second point	4961	39.1	400.1	398.1	2.0	398.7	396.2	2.5	1.0034	1.0048
third point	4981	19.5	199.5	198.5	1.0	198.0	195.5	2.6	1.0076	1.0155
as left zero	5000	0.0	0.0	0.0	0.0	-0.9	0.3	-1.2	----	----
as left span	4922	78.1	799.1	387.4	411.7	793.8	382.4	411.4	1.0067	1.0131
Average Correction Factor									1.0040	1.0066

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	791.4	383.6	411.7	412.4	0.9983	100.2%
2nd GPT point (200 ppb O3)	791.4	586.1	209.2	209.2	1.0000	100.0%
3rd GPT point (100 ppb O3)	791.4	688.9	106.4	106.6	0.9982	100.2%
Average Correction Factor					0.9988	100.1%

Notes:

Calibration after moly converter change. Zero air scrubbers changed.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 19, 2023	Previous Calibration:	May 18, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	5:15	End Time (MST):	9:38
Analyzer make:	API T200	Analyzer serial #:	723

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
799.1	798.3	1.0010			
400.1	398.7	1.0034			
199.5	198.0	1.0076			
			Slope	0.999316	0.90 - 1.10
			Intercept	-0.653591	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 19, 2023	Previous Calibration:	May 18, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	5:15	End Time (MST):	9:38
Analyzer make:	API T200	Analyzer serial #:	723

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
795.2	795.7	0.9994			
398.1	396.2	1.0048			
198.5	195.5	1.0155			
			Slope	1.001680	0.90 - 1.10
			Intercept	-1.693638	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 19, 2023	Previous Calibration:	May 18, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	5:15	End Time (MST):	9:38
Analyzer make:	API T200	Analyzer serial #:	723

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
411.7	412.4	0.9983		
209.2	209.2	1.0000		
106.4	106.6	0.9982		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 19, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: May 4, 2023 Last Cal Date: April 3, 2023
 Start time (MST): 5:58 End time (MST): 8:37
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 2961
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000143	0.998743	Backgd or Offset:	-2.7	-2.6
Calibration intercept:	-0.100000	1.020000	Coeff or Slope:	1.035	1.008

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	5000	978.9	400.0	410.7	0.974
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.8	----
high point	5000	981.1	400.0	400.4	0.999
second point	5000	813.4	200.0	200.9	0.996
third point	5000	704.3	100.0	101.1	0.989
as left zero	5000	0.0	0.0	1.1	----
as left span	5000	981.8	400.0	400.8	0.998
Average Correction Factor					0.995

Baseline Corr As found:	410.6	Previous response	400.0	*% change	2.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No Maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 4, 2023	Previous Calibration:	April 3, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	5:58	End Time (MST):	8:37
Analyzer make:	API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.8	----	Correlation Coefficient	0.999999	≥0.995
400.0	400.4	0.9990			
200.0	200.9	0.9955	Slope	0.998743	0.90 - 1.10
100.0	101.1	0.9891			
			Intercept	1.020000	+/- 5



O₃ Calibration Plot

Date: May 4, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: May 16, 2023 Last Cal Date: April 27, 2023
 Start time (MST): 7:16 End time (MST): 7:53

Analyzer Make: API T640 S/N: 322
 Particulate Fraction: PM2.5

Flow Meter Make/Model: AliCat S/N: 228085
 Temp/RH standard: AliCat S/N: 228085

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	12.9	12.6	12.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	731.1	732.8	731.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5	4.7	5	<input checked="" type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>May 16, 2023</u>	Last Cal Date: <u>April 27, 2023</u>			
	PM w/o HEPA: <u>58</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test		---		<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>March 28, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>March 28, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: September 15, 2022
 Date RH/T Sensor Cleaned: September 15, 2022

Flow adjusted. Head cleaned. Leak check passed.

Notes:

Calibration by: Melissa Lemay



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: May 23, 2023 Last Cal Date: May 16, 2023
 Start time (MST): 9:17 End time (MST): 10:41

Analyzer Make: API T640 S/N: 322
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Deltacal S/N: 1451
 Temp/RH standard: Deltacal S/N: 1451

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.5	10.8	10.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	732.4	733	732.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.03	5.24	5.03	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>May 23, 2023</u>	Last Cal Date: <u>May 16, 2023</u>			
	PM w/o HEPA: <u>24.3</u>	PM w/ HEPA: <u>24.3</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	5.2	10.8	10.8	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>95.4</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>May 23, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>May 23, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: May 23, 2023
 Date RH/T Sensor Cleaned: May 23, 2023

Notes: Initial leak check is not passing, will proceed with the maintenance. PMT peak test was adjusted. Post-maintenance leak check passed.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS05
MANNIX
MAY 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 5, 2023	Previous Calibration:	April 25, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	12:28	End Time (MST):	15:16
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995
800.3	792.8	1.0095			
400.2	396.7	1.0087	Slope	0.990561	0.90 - 1.10
200.1	197.7	1.0120			
			Intercept	0.040000	+/-30



SO2 Calibration Plot

Date: May 5, 2023

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Mannix	Station number:	AMS05
Calibration Date:	May 18, 2023	Last Cal Date:	April 17, 2023
Start time (MST):	10:00	End time (MST):	14:51
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.92	ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	EY0002433			
Removed Cal Gas Conc:	4.92	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1845
ZAG Make/Model:	API T701H		Serial Number:	832

Analyzer Information

Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1203169745
Converter make:	Global	Converter serial #:	2022-196
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005327	1.005754	Backgd or Offset:	2.17
Calibration intercept:	0.060642	0.020710	Coeff or Slope:	0.849
				2.25
				0.893

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4919	81.3	80.0	80.1	0.999
as found 2nd point	4960	40.7	40.0	39.5	1.014
as found 3rd point	4980	20.3	20.0	19.5	1.024
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4919	81.3	80.0	80.5	0.994
second point	4960	40.7	40.0	40.2	0.996
third point	4980	20.3	20.0	20.2	0.989
as left zero	5000	0.0	0.0	0.3	----
as left span	4919	81.3	80.0	79.7	1.004
SO2 Scrubber Check	4920	80.0	800.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	0.993
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	80.1	Prev response:	80.48	*% change:	-0.5%
Baseline Corr 2nd AF pt:	39.5	AF Slope:	1.002610	AF Intercept:	-0.319257
Baseline Corr 3rd AF pt:	19.5	AF Correlation:	0.999915		

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Scrubber check completed after calibrator zero. Adjusted the span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 18, 2023	Previous Calibration:	April 17, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	10:00	End Time (MST):	14:51
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999995
80.0	80.5	0.9937		
40.0	40.2	0.9961	Slope	1.005754
20.0	20.2	0.9888		
			Intercept	0.020710
				+/-3



H₂S Calibration Plot

Date: May 18, 2023

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name: Mannix Station number: AMS05
 Calibration Date: May 4, 2023 Last Cal Date: April 25, 2023
 Start time (MST): 9:20 End time (MST): 11:03
 Reason: Cylinder Change Support gas cylinder change out.

Calibration Standards

Gas Cert Reference: XCO268098 Cal Gas Expiry Date: January 12, 2029
 CH₄ Cal Gas Conc. 504.9 ppm CH₄ Equiv Conc. 1076.6 ppm
 C₃H₈ Cal Gas Conc. 207.9 ppm
 Removed Gas Cert: NA Removed Gas Expiry:
 Removed CH₄ Conc. 504.9 ppm CH₄ Equiv Conc. 1076.6 ppm
 Removed C₃H₈ Conc. 207.9 ppm
 Diff between cyl (CH₄): Diff between cyl (NM):
 Calibrator Model: API T700 Serial Number: 621
 ZAG make/model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 1152430011
 THC Range (ppm): 0 - 20 ppm
 NMHC Range (ppm): 0 - 10 ppm CH₄ Range (ppm): 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.65E-04	NA	NMHC SP Ratio:	4.42E-05 NA
CH ₄ Retention time:	15.00	NA	NMHC Peak Area:	206898 NA

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.41	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.42	0.989
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	0.989
Baseline Corr AF:	17.41	Prev response	17.24	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4920	80	9.15	9.22	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	80	9.15	9.19	0.995
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.995
Baseline Corr AF:	9.22	Prev response	9.17	*% change	0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	8.19	0.987
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.23	0.982
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.982
Baseline Corr AF:	8.19	Prev response	8.07	*% change	1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000899	1.011204
THC Cal Offset:	-0.002400	0.000000
CH ₄ Cal Slope:	1.000594	1.018642
CH ₄ Cal Offset:	-0.009400	0.000000
NMHC Cal Slope:	1.001168	1.004854
NMHC Cal Offset:	0.007000	0.000000


Notes: Changed out the Hydrogen cylinder after as founds.

Calibration Performed By: Mohammed Kashif

NMHC Calibration Plot

Date: May 4, 2023

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Mannix	Station number:	AMS05
Calibration Date:	May 5, 2023	Last Cal Date:	April 25, 2023
Start time (MST):	12:28	End time (MST):	15:16
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701H	Serial Number:	832

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.65E-04	2.65E-04	NMHC SP Ratio:	4.42E-05
CH ₄ Retention time:	15.00	15.00	NMHC Peak Area:	206898
				206898

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.38	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.24	0.999
second point	4960	40.0	8.61	8.69	0.991
third point	4980	20.0	4.31	4.34	0.993
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.38	0.991

Average Correction Factor				0.994
Baseline Corr AF:	17.38	Prev response	17.24	*% change 0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4920	80	9.15	9.20	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	80	9.15	9.16	0.998
second point	4960	40	4.57	4.61	0.992
third point	4980	20	2.29	2.31	0.989
as left zero	5000	0	0.00	0.00	----
as left span	4920	80	9.15	9.21	0.993
Average Correction Factor					0.993
Baseline Corr AF:	9.20	Prev response	9.17	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	8.19	0.987
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.08	1.000
second point	4960	40.0	4.04	4.08	0.991
third point	4980	20.0	2.02	2.03	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.17	0.989
Average Correction Factor					0.996
Baseline Corr AF:	8.19	Prev response	8.07	*% change	1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000899	1.000806
THC Cal Offset:	-0.002400	0.024800
CH ₄ Cal Slope:	1.000594	1.000481
CH ₄ Cal Offset:	-0.009400	0.009000
NMHC Cal Slope:	1.001168	1.001131
NMHC Cal Offset:	0.007000	0.015400

Notes: Sample inlet filter changed after as founds. No adjustment made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 5, 2023	Previous Calibration:	April 25, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	12:28	End Time (MST):	15:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999981	≥ 0.995			
17.23	17.24	0.9991						
8.61	8.69	0.9914				Slope	1.000806	0.90 - 1.10
4.31	4.34	0.9925						
			Intercept	0.024800	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-01-2020


Station Information

Calibration Date:	May 5, 2023	Previous Calibration:	April 25, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	12:28	End Time (MST):	15:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999976	≥ 0.995
8.08	8.08	0.9998			
4.04	4.08	0.9912			
2.02	2.03	0.9973			
			Slope	1.000481	0.90 - 1.10
			Intercept	0.009000	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 5, 2023	Previous Calibration:	April 25, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	12:28	End Time (MST):	15:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999983	≥ 0.995			
9.15	9.16	0.9984						
4.57	4.61	0.9915				Slope	1.001131	0.90 - 1.10
2.29	2.31	0.9887						
			Intercept	0.015400	± 0.5			



NMHC Calibration Plot

Date: May 5, 2023

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Mannix	Station number:	AMS05
Calibration Date:	May 8, 2023	Last Cal Date:	May 5, 2023
Start time (MST):	9:09	End time (MST):	11:50
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.65E-04	2.66E-04	NMHC SP Ratio:	4.42E-05
CH ₄ Retention time:	15.00	15.20	NMHC Peak Area:	206898
				207349

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.26	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.26	0.998
second point	4960	40.0	8.61	8.61	1.000
third point	4980	20.0	4.31	4.31	0.999
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.25	0.998

Average Correction Factor				0.999
Baseline Corr AF:	17.26	Prev response	17.26	*% change 0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4920	80	9.15	9.20	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	80	9.15	9.18	0.997
second point	4960	40	4.57	4.59	0.997
third point	4980	20	2.29	2.30	0.994
as left zero	5000	0	0.00	0.00	----
as left span	4920	80	9.15	9.16	0.998
Average Correction Factor					0.996
Baseline Corr AF:	9.20	Prev response	9.17	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	8.07	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.08	1.000
second point	4960	40.0	4.04	4.02	1.004
third point	4980	20.0	2.02	2.01	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.09	0.998
Average Correction Factor					1.003
Baseline Corr AF:	8.07	Prev response	8.09	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000806	1.001881
THC Cal Offset:	0.024800	-0.005800
CH ₄ Cal Slope:	1.000481	1.000679
CH ₄ Cal Offset:	0.009000	-0.008200
NMHC Cal Slope:	1.001131	1.002855
NMHC Cal Offset:	0.015400	0.003000

Notes: Maintenance. Adjusted span only to fix the dipping issue.

Calibration Performed By: Denny Ray estador and Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	May 5, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:09	End Time (MST):	11:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
17.23	17.26	0.9981						
8.61	8.61	1.0003				Slope	1.001881	0.90 - 1.10
4.31	4.31	0.9992						
			Intercept	-0.005800	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	May 5, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:09	End Time (MST):	11:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999993	≥ 0.995
8.08	8.08	0.9996			
4.04	4.02	1.0040			
2.02	2.01	1.0053			
			Slope	1.000679	$0.90 - 1.10$
			Intercept	-0.008200	± 0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	May 5, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:09	End Time (MST):	11:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
9.15	9.18	0.9969						
4.57	4.59	0.9967				Slope	1.002855	0.90 - 1.10
2.29	2.30	0.9939						
			Intercept	0.003000	+/-0.5			



NMHC Calibration Plot

Date: May 8, 2023

Location: Mannix





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 30, 2023

Revision 01



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 13, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:19	End Time (MST):	12:26
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999976	≥0.995
799.5	799.5	1.0000			
400.2	402.5	0.9944	Slope	0.998732	0.90 - 1.10
200.1	203.7	0.9824			
			Intercept	1.940243	+/-30



SO2 Calibration Plot

Date: May 17, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

H2S Calibration Report

Version-11-2021

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: May 3, 2023 Last Cal Date: April 5, 2023
 Start time (MST): 8:14 End time (MST): 14:10
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.38 ppm Cal Gas Exp Date: March 2, 2023
 Cal Gas Cylinder #: EY0000809
 Removed Cal Gas Conc: 5.38 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3566
 ZAG Make/Model: API T701 H Serial Number: 689

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1218153358
 Converter make: Global G150 Converter serial #: 2022-195
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992200	0.998919	Backgd or Offset:	1.84 1.91
Calibration intercept:	0.277270	0.217144	Coeff or Slope:	1.070 1.107

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.3	79.9	75.8	1.055
as found 2nd point	4963	37.2	40.0	38.2	1.048
as found 3rd point	4981	18.6	20.0	19.2	1.042
new cylinder response					

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	74.3	79.9	80.0	0.999
second point	4963	37.2	40.0	40.3	0.993
third point	4981	18.6	20.0	20.3	0.986
as left zero	5000	0.0	0.0	0.1	----
as left span	4926	74.3	79.9	80.6	0.992
SO2 Scrubber Check	4920	80.3	803.0	0.1	----

Date of last scrubber change:	December 20, 2021	Ave Corr Factor	0.993
Date of last converter efficiency test:		efficiency	

Baseline Corr As found:	75.8	Prev response:	79.60	*% change:	-5.0%
Baseline Corr 2nd AF pt:	38.2	AF Slope:	0.947594	AF Intercept:	0.138294
Baseline Corr 3rd AF pt:	19.2	AF Correlation:	0.999983		

* = > +/-5% change initiates investigation

Notes: As found span is 5.5% low, suspecting the dry ambient air has caused this change. Changed the inlet filter after as founds and hydrated the scrubber beads for 45 minutes. Ran a SO2 scrubber check after the calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

H2S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 3, 2023	Previous Calibration:	April 5, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:14	End Time (MST):	14:10
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999989	≥0.995
79.9	80.0	0.9993			
40.0	40.3	0.9932	Slope	0.998919	0.90 - 1.10
20.0	20.3	0.9860			
			Intercept	0.217144	+/-3



H2S Calibration Plot

Date: May 3, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

H2S Calibration Report

Version-11-2021

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	May 8, 2023	Last Cal Date:	May 3, 2023
Start time (MST):	12:13	End time (MST):	14:10
Reason:	Cylinder Change		

Calibration Standards

Cal Gas Concentration:	5.33	ppm	Cal Gas Exp Date:	February 14, 2025
Cal Gas Cylinder #:	CC506659			
Removed Cal Gas Conc:	5.38	ppm	Rem Gas Exp Date:	March 2, 2023
Removed Gas Cyl #:	EY0000809		Diff between cyl:	-2.3%
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701 H		Serial Number:	689

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	Global G150	Converter serial #:	2022-195
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998919		Backgd or Offset:	1.91
Calibration intercept:	0.217144		Coeff or Slope:	1.107

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.3	79.9	76.9	1.040
as found 2nd point					
as found 3rd point					
new cylinder response	4925	75.1	80.0	75.3	1.063

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

SO2 Scrubber Check

Date of last scrubber change:	December 20, 2021	Ave Corr Factor	
Date of last converter efficiency test:		efficiency	

Baseline Corr As found:	76.9	Prev response:	80.07	*% change:	-4.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation


Notes: Changed the H2S cylinder after as founds. As found is low, suspecting the scrubber beads are starting to go bad, will be back tomorrow to change the beads.

Calibration Performed By: Max Farrell

H2S Calibration Plot

Date: May 8, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

H2S Calibration Report

Version-11-2021

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: May 9, 2023 Last Cal Date: May 3, 2023
 Start time (MST): 8:21 End time (MST): 13:52
 Reason: Maintenance

Calibration Standards

Cal Gas Concentration: 5.33 ppm Cal Gas Exp Date: February 14, 2025
 Cal Gas Cylinder #: CC506659
 Removed Cal Gas Conc: 5.33 ppm Rem Gas Exp Date: February 14, 2025
 Removed Gas Cyl #: Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3566
 ZAG Make/Model: API T701 H Serial Number: 689

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1218153358
 Converter make: Global G150 Converter serial #: 2022-195
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992200	1.001170	Backgd or Offset:	1.91	2.03
Calibration intercept:	0.277270	0.480293	Coeff or Slope:	1.107	1.186

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4925	75.1	80.0	76.7	1.043
as found 2nd point	4963	37.5	40.0	38.5	1.038
as found 3rd point	4981	18.8	20.0	19.3	1.038
new cylinder response					

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4925	75.1	80.0	80.3	0.997
second point	4963	37.5	40.0	40.9	0.977
third point	4981	18.8	20.0	20.9	0.959
as left zero	5000	0.0	0.0	0.7	----
as left span	4925	75.1	80.0	80.3	0.997
SO2 Scrubber Check	4920	80.3	803.0	0.1	----

Date of last scrubber change:	December 20, 2021	Ave Corr Factor	0.977
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 76.7 Prev response: 79.68 *% change: -3.9%
 Baseline Corr 2nd AF pt: 38.5 AF Slope: 0.958319 AF Intercept: 0.080219
 Baseline Corr 3rd AF pt: 19.3 AF Correlation: 0.999991

* = > +/-5% change initiates investigation

Notes: The span has drifted down since the last calibration. Completed multipoint as founds and changed the SO2 scrubber beads. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

H2S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 9, 2023	Previous Calibration:	May 3, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:21	End Time (MST):	13:52
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	≥ 0.995
80.0	80.3	0.9966		
40.0	40.9	0.9769	Slope	0.90 - 1.10
20.0	20.9	0.9586		
			Intercept	+/-3



H2S Calibration Plot

Date: May 9, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

H2S Calibration Report

Version-11-2021

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: May 12, 2023 Last Cal Date: May 9, 2023
 Start time (MST): 9:12 End time (MST): 12:29
 Reason: Maintenance

Calibration Standards

Cal Gas Concentration: 5.33 ppm Cal Gas Exp Date: February 14, 2025
 Cal Gas Cylinder #: CC506659
 Removed Cal Gas Conc: 5.33 ppm Rem Gas Exp Date: February 14, 2025
 Removed Gas Cyl #: Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3566
 ZAG Make/Model: API T701 H Serial Number: 689

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1218153358
 Converter make: Global G150 Converter serial #: 2022-195
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001170	0.996460	Backgd or Offset:	2.03	1.82
Calibration intercept:	0.480293	0.320162	Coeff or Slope:	1.186	1.069

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4925	75.1	80.0	88.8	0.901
as found 2nd point					
as found 3rd point					
new cylinder response					

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4925	75.1	80.0	80.0	1.000
second point	4963	37.5	40.0	40.2	0.994
third point	4981	18.8	20.0	20.4	0.982
as left zero	5000	0.0	0.0	0.1	----
as left span	4925	75.1	80.0	79.8	1.003

SO2 Scrubber Check

Date of last scrubber change:	December 20, 2021	Ave Corr Factor	0.992
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 88.8 Prev response: 80.60 *% change: 9.2%
 Baseline Corr 2nd AF pt: NA AF Slope: NA AF Intercept: NA
 Baseline Corr 3rd AF pt: NA AF Correlation: NA

* = > +/-5% change initiates investigation

Notes: The span has drifted up by 9% since the last calibration, it seems that maybe the smoky weather might have caused the drift instead of the scrubber beads. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

H2S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 12, 2023	Previous Calibration:	May 9, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:12	End Time (MST):	12:29
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999990	≥0.995
80.0	80.0	1.0003			
40.0	40.2	0.9939	Slope	0.996460	0.90 - 1.10
20.0	20.4	0.9821			
			Intercept	0.320162	+/-3



H2S Calibration Plot

Date: May 12, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	May 17, 2023	Last Cal Date:	April 13, 2023
Start time (MST):	8:18	End time (MST):	12:26
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH ₄ Cal Gas Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Ref.	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	261

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.33E-04	3.57E-04	NMHC SP Ratio:	5.86E-05
CH ₄ Retention time:	14	14.6	NMHC Peak Area:	154840
				152338

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.12	16.58	1.033
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	17.17	0.997
second point	4960	40.2	8.57	8.58	1.000
third point	4980	20.1	4.29	4.32	0.992
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.12	17.03	1.006

Average Correction Factor	0.996
---------------------------	-------

Baseline Corr AF:	16.58	Prev response	17.02	*% change	-2.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF Limit= 0.95-1.05
as found zero	5000	0	0.00	0.00	----
as found span	4920	80.3	9.07	8.76	1.035
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	80.3	9.07	9.10	0.996
second point	4960	40.2	4.54	4.56	0.995
third point	4980	20.1	2.27	2.30	0.988
as left zero	5000	0	0.00	0.00	----
as left span	4920	80.3	9.07	9.04	1.003
Average Correction Factor					0.993
Baseline Corr AF:	8.76	Prev response	9.05	*% change	-3.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF Limit= 0.95-1.05
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.06	7.82	1.030
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.06	8.07	0.998
second point	4960	40.2	4.03	4.02	1.004
third point	4980	20.1	2.02	2.02	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.06	7.99	1.009
Average Correction Factor					1.000
Baseline Corr AF:	7.82	Prev response	7.98	*% change	-2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.993826	1.002022
THC Cal Offset:	0.007564	0.005131
CH ₄ Cal Slope:	0.990693	1.001093
CH ₄ Cal Offset:	-0.004180	-0.003602
NMHC Cal Slope:	0.996672	1.003075
NMHC Cal Offset:	0.010746	0.008332

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 13, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:18	End Time (MST):	12:26
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995			
17.12	17.17	0.9974						
8.57	8.58	0.9995				Slope	1.002022	0.90 - 1.10
4.29	4.32	0.9924						
			Intercept	0.005131	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 13, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:18	End Time (MST):	12:26
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999988	≥0.995
8.06	8.07	0.9985			
4.03	4.02	1.0042			
2.02	2.02	0.9972			
			Slope	1.001093	0.90 - 1.10
			Intercept	-0.003602	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 13, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:18	End Time (MST):	12:26
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995			
9.07	9.10	0.9963						
4.54	4.56	0.9954				Slope	1.003075	0.90 - 1.10
2.27	2.30	0.9881						
			Intercept	0.008332	± 0.5			



NMHC Calibration Plot

Date: May 17, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Patricia McInnes
Calibration Date: May 1, 2023
Start time (MST): 9:03
Reason: Routine
Station number: AMS06
Last Cal Date: April 4, 2023
End time (MST): 13:47

Calibration Standards

NO Gas Cylinder #: T26D9MR
NOX Cal Gas Conc: 52.51 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 52.51 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: August 18, 2023
NO Cal Gas Conc: 51.98 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 51.98 ppm
NO gas Diff:
Serial Number: 3566
Serial Number: 689

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1172750022

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.835	0.835	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	155.1	154.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001408	1.001069
NO _x Cal Offset:	2.760164	1.979986
NO Cal Slope:	0.998955	0.998555
NO Cal Offset:	1.840130	1.080026
NO ₂ Cal Slope:	1.006294	1.005730
NO ₂ Cal Offset:	1.349835	0.795849



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as found span	4923	76.9	807.6	799.5	8.2	810.9	797.9	13.0	0.9959	1.0019
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
high point	4923	76.9	807.6	799.5	8.2	809.5	798.9	10.6	0.9977	1.0007
second point	4962	38.5	404.3	400.2	4.1	407.7	401.2	6.5	0.9917	0.9976
third point	4981	19.2	201.6	199.6	2.0	205.7	201.4	4.3	0.9803	0.9911
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
as left span	4923	76.9	807.6	387.2	420.5	806.0	386.4	419.6	1.0020	1.0019
Average Correction Factor									0.9899	0.9965

Corrected As found	NO _x = 811.0 ppb	NO = 797.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.1%	
Previous Response	NO _x = 811.5 ppb	NO = 800.5 ppb		*Percent Change	NO = -0.3%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.7	384.4	420.5	422.9	0.9942	100.6%
2nd GPT point (200 ppb O3)	796.7	592.2	212.7	216.1	0.9840	101.6%
3rd GPT point (100 ppb O3)	796.7	694.1	110.8	112.3	0.9862	101.4%
Average Correction Factor					0.9882	101.2%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 4, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:03	End Time (MST):	13:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
807.6	809.5	0.9977		
404.3	407.7	0.9917		
201.6	205.7	0.9803		
			0.999975	
			1.001069	
			1.979986	





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 4, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:03	End Time (MST):	13:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.5	798.9	1.0007		
400.2	401.2	0.9976		
199.6	201.4	0.9911		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 4, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:03	End Time (MST):	13:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
420.5	422.9	0.9942		
212.7	216.1	0.9840		
110.8	112.3	0.9862		
			0.999966	
			1.005730	
			0.795849	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 1, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Patricia McInnes
Calibration Date: January 0, 1900
Start time (MST): 10:12
Reason: Cylinder Change
Station number: AMS06
Last Cal Date: May 1, 2023
End time (MST): 12:04

Calibration Standards

NO Gas Cylinder #: T30YCWN
NOX Cal Gas Conc: 47.94 ppm
Removed Cylinder #: T26D9MR
Removed Gas NOX Conc: 52.51 ppm
NOX gas Diff: 3.5%
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: April 11, 2025
NO Cal Gas Conc: 46.39 ppm
Removed Gas Exp Date: August 18, 2023
Removed Gas NO Conc: 51.98 ppm
NO gas Diff: 3.6%
Serial Number: 3566
Serial Number: 689

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1172750022

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.835	0.835	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	154.2	154.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001069	
NO _x Cal Offset:	1.979986	
NO Cal Slope:	0.998555	
NO Cal Offset:	1.080026	
NO ₂ Cal Slope:	1.005730	
NO ₂ Cal Offset:	0.795849	



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.1	----	----
as found span	4923	76.9	807.6	799.5	8.2	810.4	797.4	12.9	0.9965	1.0026
as found 2nd										
as found 3rd										
new cyl resp	4914	86.2	826.5	799.7	26.7	859.9	827.0	32.9	0.9611	0.9670
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

Average Correction Factor

Corrected As found	NO _x = 810.2 ppb	NO = 797.2 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 0.0%
Previous Response	NO _x = 810.4 ppb	NO = 799.4 ppb			*Percent Change	NO = -0.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)						
2nd GPT point (200 ppb O3)						
3rd GPT point (100 ppb O3)						

Average Correction Factor

Notes:

Changed the NO cylinder after as founds.


Calibration Performed By:

Max Farrell

NO_x Calibration Plot

Date: January 0, 1900

Location: Patricia McInnes





Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	May 8, 2023	Last Cal Date:	April 27, 2023
Start time (MST):	10:12	End time (MST):	12:03
NH3 Cal Date:	April 20, 2023	Last Cal Date:	March 8, 2023
Start time (MST):	13:00	End time (MST):	14:48
Reason:	Cylinder Change	NO cylinder change	

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	52.5	ppm	Removed Cylinder #:	T26D9MR
Removed NO Conc:	52.0	ppm	Removed cyl Expiry:	August 18, 2023
NOX gas Diff:	3.2%		NO gas Diff:	3.7%
NH3 Cal Gas Conc:	77.8	ppm	NH3 Gas Cylinder #:	CC710812
Removed NH3 Conc:	77.8	ppm	NH3 Cal Gas Expiry:	March 30, 2023
NH3 gas Diff:			Removed Cylinder #:	
Calibrator Model:	API T700		Removed cyl Expiry:	
ZAG make/model:	API T701		Serial Number:	3566
			Serial Number:	689

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	152
Converter model:	API T501	Converter serial #:	147
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	5.70
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.833	0.833	TN coefficient:	0.837	0.837
NOX coefficient:	0.839	0.839	NO bkgnd:	-0.100	-0.1
NO2 coefficient:	1.000	1.000	NOX bkgnd:	0.000	0.0
NH3 coefficient:	0.951	0.951	TN bkgnd:	0.000	0.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.004262	
NO _x Cal Offset:	2.034812	
NO Cal Slope:	1.000294	
NO Cal Offset:	0.654848	
NO ₂ Cal Slope:	1.003690	
NO ₂ Cal Offset:	0.714817	
NH3 Cal Slope:	1.003144	
NH3 Cal Offset:	4.931327	
TN Cal Slope:	1.008443	
TN Cal Offset:	5.257930	



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found NO										
calibrator zero										
high NO point										
NO/O3 point										
as found NH3										
new NH3 cyl rp										
first NH3										
second NH3										
third NH3										

									Average Correction Factor	
--	--	--	--	--	--	--	--	--	---------------------------	--

Corrected As found TN = NA ppb NO_x = NA ppb NH3 = NA ppb

Previous Response TN = NA ppb NO_x = NA ppb NH3 = NA ppb

NH3 Previous Converter Efficiency = 95.1%

NH3 Current Converter Efficiency = 95.1%

*Percent Change TN = **NA**

*Percent Change NO_x = **NA**

*Percent Change NH3 = **NA**

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.2	0.4	1.0	----	----
as found span	4923	76.9	807.6	799.5	807.6	811.2	794.6	811.0	0.9956	1.0061
new NO cyl rp	4914	86.2	826.5	799.7	826.5	857.9	825.0	857.7	0.9633	0.9694
calibrator zero										
high point										
second point										
third point										

		Average Correction Factor				
Baseline Corr As fnd	TN = 810 ppb	NO _x = 811.0 ppb	NO = 794.2 ppb		*Percent Change	TN = -1.2%
Previous Response	TN = 819.7 ppb	NO _x = 813.1 ppb	NO = 800.4 ppb		*Percent Change	NO _x = -0.3%
					*Percent Change	NO = -0.8%
					<i>* = > +/-5% change initiates investigation</i>	

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	-0.2	----	----
calibration zero						
1st GPT point (400 ppb O3)						
2nd GPT point (200 ppb O3)						
3rd GPT point (100 ppb O3)						

		Average Correction Factor				

Notes:

Changed the NO cylinder after as founds.


Calibration Performed By:

Max Farrell

NO_x Calibration Plot

Date: May 8, 2023


Location: Patricia McInnes



NH₃ Calibration Plot

Date: April 20, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	May 24, 2023	Last Cal Date:	April 27, 2023
Start time (MST):	7:26	End time (MST):	11:40
NH3 Cal Date:	May 24, 2023	Last Cal Date:	April 27, 2023
Start time (MST):	12:00	End time (MST):	14:20
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	77.8	ppm	NH3 Gas Cylinder #:	CC710812
			NH3 Cal Gas Expiry:	March 30, 2023
Removed NH3 Conc:	77.8	ppm	Removed Cylinder #:	
NH3 gas Diff:			Removed cyl Expiry:	
Calibrator Model:	API T700		Serial Number:	3566
ZAG make/model:	API T701		Serial Number:	689

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	152
Converter model:	API T501	Converter serial #:	147
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.50
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	25.3

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.833	0.794	TN coefficient:	0.837	0.794
NOX coefficient:	0.839	0.796	NO bkgrnd:	-0.1	-0.1
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	0.0	0.0
NH3 coefficient:	0.951	0.908	TN bkgrnd:	0.0	0.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.004262	0.995223
NO _x Cal Offset:	2.034812	1.876743
NO Cal Slope:	1.000294	0.994331
NO Cal Offset:	0.654848	1.802690
NO ₂ Cal Slope:	1.003690	1.004951
NO ₂ Cal Offset:	0.714817	1.630262
NH3 Cal Slope:	1.003144	1.008694
NH3 Cal Offset:	4.931327	6.507686
TN Cal Slope:	1.008443	1.013896
TN Cal Offset:	5.257930	6.739810



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.9	-0.6	1.5	----	----
as found NO	4914	86.2	826.5	826.5	----	869.3	870.8	-1.4	0.951	----
calibrator zero	5000	0.0	0.0	0.0	0.0	1.6	0.2	1.5	----	----
high NO point	4914	86.2	826.5	826.5	----	827.4	823.6	3.8	0.999	----
NO/O3 point	4914	86.2	826.5	826.5	----	828.5	824.7	3.8	0.998	----
as found NH3	3419	81.0	1800.5	----	1800.5	1730.5	----	1720.5	1.040	1.047
new NH3 cyl rp							----			
first NH3	3419	81.0	1800.5	----	1800.5	1830.5	----	1820.9	0.984	0.989
second NH3	3455	45.0	1000.3	----	1000.3	1020.4	----	1015.1	0.980	0.985
third NH3	3478	22.5	500.1	----	500.1	521.2	----	518.1	0.959	0.965
Average Correction Factor									0.9982	0.9798

Corrected As found TN = 868.4 ppb NO_x = 871.4 ppb NH3 = 1719.0 ppb

Previous Response TN = 838.7 ppb NO_x = 832.0 ppb NH3 = 1811.1 ppb

NH3 Previous Converter Efficiency = 95.1%

NH3 Current Converter Efficiency = 90.8%

*Percent Change TN = 3.4%

*Percent Change NO_x = 4.5%

*Percent Change NH3 = -5.4%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.2	0.9	----	----
as found span	4914	86.2	826.5	799.7	826.5	870.8	838.2	869.3	0.9491	0.9541
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	1.6	----	----
high point	4914	86.2	826.5	799.7	826.5	823.6	796.9	827.4	1.0035	1.0036
second point	4957	43.1	413.2	399.9	413.2	413.8	397.9	416.9	0.9986	1.0050
third point	4978	21.6	207.1	200.4	207.1	209.8	204.5	210.4	0.9872	0.9801
Average Correction Factor									0.9964	0.9962

Baseline Corr As fnd	TN = 868.4 ppb	NO _x = 871.4 ppb	NO = 838.4 ppb	*Percent Change	TN = 3.4%
Previous Response	TN = 838.7 ppb	NO _x = 832.0 ppb	NO = 800.6 ppb	*Percent Change	NO _x = 4.5%
				*Percent Change	NO = 4.5%
				* = > +/-5% change initiates investigation	

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	-0.5	----	----
calibration zero	----	----	0.0	0.2	----	----
1st GPT point (400 ppb O3)	795.4	399.6	422.5	425.1	0.9939	100.6%
2nd GPT point (200 ppb O3)	795.4	600.2	221.9	226.3	0.9806	102.0%
3rd GPT point (100 ppb O3)	795.4	695.7	126.4	129.6	0.9755	102.5%
Average Correction Factor					0.9834	101.7%

Notes: Changed the inlet filter after as founds. Adjusted both NO_x and NH₃ span. As found span is close to 5% high due to the recent NO cylinder change.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TN Calibration Summary


Version-05-2023

Station Information

Calibration Date:	May 24, 2023	Previous Calibration:	April 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:26	End Time (MST):	11:40
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	1.6	----	Correlation Coefficient	≥0.995	
1800.5	1830.5	0.9836			
1000.3	1020.4	0.9803			
500.1	521.2	0.9595			
			Slope	1.013896	0.90 - 1.10
			Intercept	6.739810	+/-20





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	May 24, 2023	Previous Calibration:	April 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:26	End Time (MST):	11:40
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	1.5	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.5	1820.9	0.9888		
1000.3	1015.1	0.9854		
500.1	518.1	0.9652		





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	May 24, 2023	Previous Calibration:	April 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:26	End Time (MST):	11:40
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.5	823.6	1.0035		
413.2	413.8	0.9986		
207.1	209.8	0.9872		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-05-2023

Station Information

Calibration Date:	May 24, 2023	Previous Calibration:	April 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:26	End Time (MST):	11:40
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
799.7	796.9	1.0036			
399.9	397.9	1.0050			
200.4	204.5	0.9801			
			Slope	0.994331	0.90 - 1.10
			Intercept	1.802690	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	May 24, 2023	Previous Calibration:	April 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:26	End Time (MST):	11:40
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥ 0.995 $0.90 - 1.10$ ± 20
422.5	425.1	0.9939		
221.9	226.3	0.9806		
126.4	129.6	0.9755		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 24, 2023


Location: Patricia McInnes



NH₃ Calibration Plot

Date: May 24, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Patricia McInnes Station number: AMS06
 Calibration Date: May 15, 2023 Last Cal Date: April 3, 2023
 Start time (MST): 9:23 End time (MST): 12:35
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3566
 ZAG Make/Model: API T701H Serial Number: 689

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1300156234
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006429	1.001971	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	0.300000	0.480000	Coeff or Slope:	1.019	1.019

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	-1.1	----
as found span	5000	1303.0	400.0	401.4	0.997
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	-0.6	----
high point	5000	1303.0	400.0	400.7	0.998
second point	5000	966.5	200.0	201.5	0.993
third point	5000	794.3	100.0	101.7	0.983
as left zero	5000	800.0	0.0	-0.8	----
as left span	5000	1303.0	400.0	403.6	0.991
Average Correction Factor					0.991

Baseline Corr As found:	402.5	Previous response	402.9	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 3, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:23	End Time (MST):	12:35
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.6	----	Correlation Coefficient	0.999967	≥0.995
400.0	400.7	0.9983			
200.0	201.5	0.9926	Slope	1.001971	0.90 - 1.10
100.0	101.7	0.9833			
			Intercept	0.480000	+/- 5



O₃ Calibration Plot

Date: May 15, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: May 17, 2023 Last Cal Date: April 13, 2023
 Start time (MST): 12:42 End time (MST): 13:42

Analyzer Make: API T640 S/N: 766
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT FP-25 S/N: 388755
 Temp/RH standard: ALICAT FP-25 S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	8.9	8.66	8.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	732.9	733.5	732.9	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	5.11	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>May 17, 2023</u>	Last Cal Date: <u>April 13, 2023</u>			
	PM w/o HEPA: <u>16.8</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10	10	11.1	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>April 13, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>April 13, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023
 Date RH/T Sensor Cleaned: April 13, 2023

Notes: "Perform span dust check" alarm was on upon arrival, completed PMT Peak adjustment. Leak check

Calibration by: Max Farrell



Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	May 24, 2023	Last Cal Date:	April 27, 2023
Start time (MST):	7:26	End time (MST):	11:40
NH3 Cal Date:	May 24, 2023	Last Cal Date:	April 27, 2023
Start time (MST):	12:00	End time (MST):	14:20
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	77.8	ppm	NH3 Gas Cylinder #:	CC710812
			NH3 Cal Gas Expiry:	March 30, 2023
Removed NH3 Conc:	77.8	ppm	Removed Cylinder #:	
NH3 gas Diff:			Removed cyl Expiry:	
Calibrator Model:	API T700		Serial Number:	3566
ZAG make/model:	API T701		Serial Number:	689

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	152
Converter model:	API T501	Converter serial #:	147
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.50
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	25.3

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.833	0.794	TN coefficient:	0.837	0.794
NOX coefficient:	0.839	0.796	NO bkgnd:	-0.1	-0.1
NO2 coefficient:	1.000	1.000	NOX bkgnd:	0.0	0.0
NH3 coefficient:	0.951	0.908	TN bkgnd:	0.0	0.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.004262	0.995223
NO _x Cal Offset:	2.034812	1.876743
NO Cal Slope:	1.000294	0.994331
NO Cal Offset:	0.654848	1.802690
NO ₂ Cal Slope:	1.003690	1.004951
NO ₂ Cal Offset:	0.714817	1.630262
NH3 Cal Slope:	1.003144	1.008694
NH3 Cal Offset:	4.931327	6.507686
TN Cal Slope:	1.008443	1.013896
TN Cal Offset:	5.257930	6.739810



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.9	-0.6	1.5	----	----
as found NO	4914	86.2	826.5	826.5	----	869.3	870.8	-1.4	0.951	----
calibrator zero	5000	0.0	0.0	0.0	0.0	1.6	0.2	1.5	----	----
high NO point	4914	86.2	826.5	826.5	----	827.4	823.6	3.8	0.999	----
NO/O3 point	4914	86.2	826.5	826.5	----	828.5	824.7	3.8	0.998	----
as found NH3	3419	81.0	1800.5	----	1800.5	1730.5	----	1720.5	1.040	1.047
new NH3 cyl rp							----			
first NH3	3419	81.0	1800.5	----	1800.5	1830.5	----	1820.9	0.984	0.989
second NH3	3455	45.0	1000.3	----	1000.3	1020.4	----	1015.1	0.980	0.985
third NH3	3478	22.5	500.1	----	500.1	521.2	----	518.1	0.959	0.965
Average Correction Factor									0.9982	0.9798

Corrected As found TN = 868.4 ppb NO_x = 871.4 ppb NH3 = 1719.0 ppb

Previous Response TN = 838.7 ppb NO_x = 832.0 ppb NH3 = 1811.1 ppb

NH3 Previous Converter Efficiency = 95.1%

NH3 Current Converter Efficiency = 90.8%

*Percent Change TN = 3.4%

*Percent Change NO_x = 4.5%

*Percent Change NH3 = -5.4%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.2	0.9	----	----
as found span	4914	86.2	826.5	799.7	826.5	870.8	838.2	869.3	0.9491	0.9541
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	1.6	----	----
high point	4914	86.2	826.5	799.7	826.5	823.6	796.9	827.4	1.0035	1.0036
second point	4957	43.1	413.2	399.9	413.2	413.8	397.9	416.9	0.9986	1.0050
third point	4978	21.6	207.1	200.4	207.1	209.8	204.5	210.4	0.9872	0.9801
Average Correction Factor									0.9964	0.9962

Baseline Corr As fnd	TN = 868.4 ppb	NO _x = 871.4 ppb	NO = 838.4 ppb	*Percent Change	TN = 3.4%
Previous Response	TN = 838.7 ppb	NO _x = 832.0 ppb	NO = 800.6 ppb	*Percent Change	NO _x = 4.5%
				*Percent Change	NO = 4.5%

* = > +/-5% change initiates investigation

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	-0.5	----	----
calibration zero	----	----	0.0	0.2	----	----
1st GPT point (400 ppb O3)	795.4	399.6	422.5	425.1	0.9939	100.6%
2nd GPT point (200 ppb O3)	795.4	600.2	221.9	226.3	0.9806	102.0%
3rd GPT point (100 ppb O3)	795.4	695.7	126.4	129.6	0.9755	102.5%
Average Correction Factor					0.9834	101.7%

Notes: Changed the inlet filter after as founds. Adjusted both NO_x and NH₃ span. As found span is close to 5% high due to the recent NO cylinder change.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TN Calibration Summary


Version-05-2023

Station Information

Calibration Date:	May 24, 2023	Previous Calibration:	April 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:26	End Time (MST):	11:40
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	1.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.5	1830.5	0.9836		
1000.3	1020.4	0.9803		
500.1	521.2	0.9595		
			0.999953	
			1.013896	
			6.739810	





Wood Buffalo Environmental Association

NH₃ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	May 24, 2023	Previous Calibration:	April 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:26	End Time (MST):	11:40
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	1.5	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.5	1820.9	0.9888		
1000.3	1015.1	0.9854		
500.1	518.1	0.9652		
			0.999956	
			1.008694	
			6.507686	

NH₃ Calibration Curve





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	May 24, 2023	Previous Calibration:	April 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:26	End Time (MST):	11:40
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.5	823.6	1.0035		
413.2	413.8	0.9986		
207.1	209.8	0.9872		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	May 24, 2023	Previous Calibration:	April 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:26	End Time (MST):	11:40
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
799.7	796.9	1.0036			
399.9	397.9	1.0050			
200.4	204.5	0.9801			
			Slope	0.994331	0.90 - 1.10
			Intercept	1.802690	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	May 24, 2023	Previous Calibration:	April 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:26	End Time (MST):	11:40
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
422.5	425.1	0.9939		
221.9	226.3	0.9806		
126.4	129.6	0.9755		



NO_x Calibration Plot

Date: May 24, 2023


Location: Patricia McInnes



NH₃ Calibration Plot

Date: May 24, 2023

Location: Patricia McInnes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 19, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:36	End Time (MST):	11:19
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999984	≥0.995
801.2	805.6	0.9945			
400.2	405.0	0.9880	Slope	1.004627	0.90 - 1.10
200.1	203.6	0.9826			
			Intercept	1.624552	+/-30



SO2 Calibration Plot

Date: May 15, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: May 10, 2023 Last Cal Date: April 26, 2023
 Start time (MST): 8:30 End time (MST): 13:05
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.25 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC504080
 Removed Cal Gas Conc: 4.94 ppm Rem Gas Exp Date: February 9, 2024
 Removed Gas Cyl #: EY0002277 Diff between cyl: -5.8%
 Calibrator Make/Model: API T700 Serial Number: 3805
 ZAG Make/Model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE Analyzer serial #: 1180540018
 Converter make: CDN-101 Converter serial #: 551
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993201	1.002951	Backgd or Offset: 2.17	2.29
Calibration intercept:	0.141603	-0.222144	Coeff or Slope: 0.829	0.876

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4918	81.6	80.6	80.1	1.005
as found 2nd point	4959	40.8	40.3	39.8	1.010
as found 3rd point	4980	20.4	20.2	19.6	1.023
new cylinder response	4925	75.5	79.3	74.5	1.065

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4925	75.5	79.3	79.4	0.999
second point	4962	37.7	39.6	39.5	1.003
third point	4981	18.9	19.9	19.4	1.024
as left zero	5000	0.0	0.0	0.1	----
as left span	4925	75.5	79.3	78.9	1.005
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	1.008
Date of last converter efficiency test:	April 22, 2022			92.6% efficiency	

Baseline Corr As found: 80.2 Prev response: 80.22 *% change: 0.0%
 Baseline Corr 2nd AF pt: 39.9 AF Slope: 0.995895 AF Intercept: -0.278401
 Baseline Corr 3rd AF pt: 19.7 AF Correlation: 0.999977

* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after calibrator zero. Span adjusted. Cal gas cylinder repalced after MPAFS.

Calibration Performed By: Karina Fenwick



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 10, 2023	Previous Calibration:	April 26, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:30	End Time (MST):	13:05
Analyzer make:	CDN-101	Analyzer serial #:	551

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999959	
79.3	79.4	0.9989			≥0.995
39.6	39.5	1.0028	Slope	1.002951	
19.9	19.4	1.0235			0.90 - 1.10
			Intercept	-0.222144	+/-3



TRS Calibration Plot

Date: May 10, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	May 8, 2023	Last Cal Date:	April 19, 2023
Start time (MST):	6:08	End time (MST):	7:18
Reason:	Cylinder Change Nitrogen Cylincer change		

Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	December 29, 2028
CH4 Cal Gas Conc.	501.2 ppm	CH4 Equiv Conc.	1075.1 ppm
C3H8 Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	501.2 ppm	CH4 Equiv Conc.	1075.1 ppm
Removed C3H8 Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	0.000278	0.000278	NMHC SP Ratio: 4.49E-05	4.49E-05
CH4 Retention time:	13.8	13.8	NMHC Peak Area: 202584	202584

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.3	17.05	16.60	1.027
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.3	17.05	16.61	1.027
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	1.027
Baseline Corr AF:	16.60	Prev response	17.04	*% change	-2.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.3	9.10	8.72	1.044
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.3	9.10	8.74	1.041
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.041
Baseline Corr AF:	8.72	Prev response	9.11	*% change	-4.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.3	7.95	7.88	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.3	7.95	7.87	1.010
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.010
Baseline Corr AF:	7.88	Prev response	7.94	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999083	0.974167
THC Cal Offset:	0.005889	0.000000
CH ₄ Cal Slope:	0.998961	0.990117
CH ₄ Cal Offset:	-0.002189	0.000000
NMHC Cal Slope:	0.999315	0.960239
NMHC Cal Offset:	0.010081	0.000000

Notes:


Nitrogen Cylinder Change.

Calibration Performed By: Melissa Lemay

NMHC Calibration Plot

Date: May 8, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	May 15, 2023	Last Cal Date:	April 19, 2023
Start time (MST):	8:36	End time (MST):	11:18
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	0.000278	0.000280	NMHC SP Ratio: 4.49E-05	4.75E-05
CH ₄ Retention time:	13.8	13.8	NMHC Peak Area:	202584 191368

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.3	17.05	16.40	1.040
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.3	17.05	17.05	1.000
second point	4960	39.6	8.52	8.52	0.999
third point	4980	19.8	4.26	4.29	0.992
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.00	1.002

Average Correction Factor				0.997
Baseline Corr AF:	16.40	Prev response	17.04	*% change -3.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.3	9.10	8.52	1.068
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.3	9.10	9.11	0.999
second point	4960	39.6	4.55	4.57	0.995
third point	4980	19.8	2.27	2.30	0.988
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	9.09	9.08	1.001
Average Correction Factor					0.994
Baseline Corr AF:	8.52	Prev response	9.11	*% change	-6.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.3	7.95	7.88	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.3	7.95	7.94	1.001
second point	4960	39.6	3.97	3.95	1.005
third point	4980	19.8	1.98	1.98	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.94	7.93	1.001
Average Correction Factor					1.003
Baseline Corr AF:	7.88	Prev response	7.94	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999083	0.999351
THC Cal Offset:	0.005889	0.013892
CH ₄ Cal Slope:	0.998961	0.998818
CH ₄ Cal Offset:	-0.002189	-0.004193
NMHC Cal Slope:	0.999315	1.000193
NMHC Cal Offset:	0.010081	0.014085

Notes: NM channel moved a bit since last month, will adjust span and monitor.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 19, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:36	End Time (MST):	11:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
17.05	17.05	1.0000						
8.52	8.52	0.9995				Slope	0.999351	0.90 - 1.10
4.26	4.29	0.9925						
			Intercept	0.013892	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 19, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:36	End Time (MST):	11:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	≥0.995
7.95	7.94	1.0011			
3.97	3.95	1.0050			
1.98	1.98	1.0024			
			Slope	0.998818	0.90 - 1.10
			Intercept	-0.004193	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 19, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:36	End Time (MST):	11:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999989	≥ 0.995			
9.10	9.11	0.9991						
4.55	4.57	0.9947				Slope	1.000193	0.90 - 1.10
2.27	2.30	0.9882						
			Intercept	0.014085	± 0.5			



NMHC Calibration Plot

Date: May 15, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.2	----	----
as found span	4920	80.2	816.7	800.7	16.0	810.7	793.7	17.1	1.0074	1.0088
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.5	0.1	0.3	----	----
high point	4920	80.2	816.7	800.7	16.0	810.3	794.0	16.4	1.0079	1.0084
second point	4960	40.1	408.4	400.4	8.0	408.1	399.6	8.6	1.0007	1.0019
third point	4980	20.0	203.7	199.7	4.0	203.6	199.0	4.6	1.0004	1.0034
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
as left span	4920	80.2	816.7	401.9	414.8	811.9	392.5	419.3	1.0059	1.0239
Average Correction Factor									1.0030	1.0046

Corrected As found	NO _x = 810.6 ppb	NO = 793.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.0%	
Previous Response	NO _x = 818.9 ppb	NO = 802.6 ppb		*Percent Change	NO = -1.1%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	790.8	392.0	414.8	416.9	0.9951	100.5%
2nd GPT point (200 ppb O ₃)	790.8	590.0	216.8	217.6	0.9965	100.4%
3rd GPT point (100 ppb O ₃)	790.8	688.5	118.3	119.2	0.9928	100.7%
Average Correction Factor					0.9948	100.5%

Notes:

No maintenance or adjustments done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 21, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	7:15	End Time (MST):	11:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.5	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
816.7	810.3	1.0079		
408.4	408.1	1.0007		
203.7	203.6	1.0004		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 21, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	7:15	End Time (MST):	11:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.7	794.0	1.0084			
400.4	399.6	1.0019			
199.7	199.0	1.0034			
			Slope	0.991499	0.90 - 1.10
			Intercept	0.973228	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 21, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	7:15	End Time (MST):	11:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
414.8	416.9	0.9951		
216.8	217.6	0.9965		
118.3	119.2	0.9928		
			0.999998	
			1.004055	
			0.235209	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 8, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: May 3, 2023 Last Cal Date: April 20, 2023
 Start time (MST): 9:28 End time (MST): 12:38
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: T700 Serial Number: 3805
 ZAG Make/Model: T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1152220023
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998171	1.000257	Backgd or Offset:	-2.8	-1.3
Calibration intercept:	0.120000	1.380000	Coeff or Slope:	1.435	1.502

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.8	----
as found span	5000	1403.4	400.0	381.4	1.049
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.1	----
high point	5000	1415.0	400.0	400.7	0.998
second point	5000	1035.1	200.0	202.5	0.988
third point	5000	852.2	100.0	102.4	0.977
as left zero	5000	0.0	0.0	-1.3	----
as left span	5000	1406.5	400.0	398.6	1.004
Average Correction Factor					0.987

Baseline Corr As found:	380.6	Previous response	399.4	*% change	-4.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Installed last month. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 3, 2023	Previous Calibration:	April 20, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:28	End Time (MST):	12:38
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999950	≥0.995
400.0	400.7	0.9983			
200.0	202.5	0.9877	Slope	1.000257	0.90 - 1.10
100.0	102.4	0.9766			
			Intercept	1.380000	+/- 5



O₃ Calibration Plot

Date: May 3, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Athabasca Valley Station number: AMS 07
 Calibration Date: May 23, 2023 Last Cal Date:
 Start time (MST): 7:20 End time (MST): 7:52

Analyzer Make: API T640 S/N: 871
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.9	10.7	10.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	737.5	740.2	737.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5	5.11	5	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: May 23, 2023 Last Cal Date: _____
 PM w/o HEPA: 101 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.2	10.2	11	<input checked="" type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: _____ w/ HEPA: _____
 Date Optical Chamber Cleaned: May 23, 2023 <0.2 ug/m3
 Disposable Filter Changed: May 23, 2023

Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022
 Date RH/T Sensor Cleaned: December 5, 2022

Notes: Install Due to power issue with removed T640. PMT adjusted. No other adjustments done. Leak check passed.

Calibration by: Melissa Lemay



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	May 15, 2023	Last Cal Date:	April 25, 2023
Start time (MST):	5:55	End time (MST):	8:39
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date:	December 12, 2026
Cal Gas Cylinder #:	LL66942			
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 700H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994802	0.991945	Backgd or Offset:	3.896	3.910
Calibration intercept:	0.030541	0.060555	Coeff or Slope:	1.079	1.079

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.0	39.8	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.0	39.7	1.008
second point	4967	33.3	20.0	20.0	0.999
third point	4983	16.7	10.0	10.0	1.002
as left zero	5000	0.0	0.0	0.0	----
as left span	4933	66.7	40.0	39.7	1.008
Average Correction Factor					1.003

Baseline Corr As found:	39.74	Prev response:	39.84	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No Maintenance done. Zero adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 25, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	5:55	End Time (MST):	8:39
Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999974	
40.0	39.7	1.0081			≥0.995
20.0	20.0	0.9989	Slope	0.991945	
10.0	10.0	1.0021			0.90 - 1.10
			Intercept	0.060555	+/-1.5



CO Calibration Plot

Date: May 15, 2023

Location: Athabasca Valley





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	May 8, 2023	Last Cal Date:	April 12, 2023
Start time (MST):	11:11	End time (MST):	15:22
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC196697			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3060
ZAG Make/Model:	Teledyne API T701		Serial Number:	260

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003016	0.999415	Backgd or Offset:	1.29	1.50
Calibration intercept:	1.535287	-1.203236	Coeff or Slope:	1.010	0.877

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4920	80.3	800.4	922.0	0.868
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	80.3	800.4	799.5	1.001
second point	4960	40.2	400.7	398.0	1.007
third point	4980	20.1	200.4	198.4	1.010
as left zero	5000	0.0	0.0	-0.1	----
as left span	4920	80.3	800.4	797.3	1.004
Average Correction Factor					1.006

Baseline Corr As found:	921.70	Previous response	804.33	*% change	12.7%
					<i>* = > +/-5% change initiates investigation</i>

Notes: Sample inlet filter changed after as founds. Adjusted the coefficient on analyzer. Adjustments made to zero and span. Changed the calibrator serial # on cal sheet.

Calibration Performed By: Morgan Voyageur, Matthew C



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 12, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	11:11	End Time (MST):	15:22
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	≥0.995
800.4	799.5	1.0011		
400.7	398.0	1.0068	Slope	0.90 - 1.10
200.4	198.4	1.0098		
			Intercept	+/-30



SO2 Calibration Plot

Date: May 8, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: May 29, 2023 Last Cal Date: April 26, 2023
 Start time (MST): 9:45 End time (MST): 13:41
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002276
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 260

Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 1203169744
 Converter make: CDN-101 Converter serial #: 14639
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994424	0.976855	Backgd or Offset:	0.92	0.93
Calibration intercept:	0.158770	0.498472	Coeff or Slope:	0.694	0.694

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4920	80.5	80.0	79.1	1.015
as found 2nd point	4960	40.2	40.0	39.4	1.022
as found 3rd point	4980	20.1	20.0	19.9	1.019
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4920	80.5	80.0	78.5	1.019
second point	4960	40.2	40.0	39.8	1.004
third point	4980	20.1	20.0	20.1	0.994
as left zero	5000	0.0	0.0	0.4	----
as left span	4920	80.5	80.0	79.1	1.011
SO2 Scrubber Check	4919.7	80.3	803.0	0.1	----
Date of last scrubber change:	March 7, 2022			Ave Corr Factor	1.006
Date of last converter efficiency test:	March 15, 2022			100.7% efficiency	

Baseline Corr As found: 78.8 Prev response: 79.72 *% change: -1.2%
 Baseline Corr 2nd AF pt: 39.1 AF Slope: 0.984860 AF Intercept: 0.218396
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999987

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustment made.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 29, 2023	Previous Calibration:	April 26, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	9:45	End Time (MST):	13:41
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999957	≥0.995
80.0	78.5	1.0192			
40.0	39.8	1.0039	Slope	0.976855	0.90 - 1.10
20.0	20.1	0.9940			
			Intercept	0.498472	+/-3



TRS Calibration Plot

Date: May 29, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Chipewyan
Calibration Date: May 15, 2023
Start time (MST): 8:28
Reason: Routine
Station number: AMS08
Last Cal Date: April 13, 2023
End time (MST): 12:54

Calibration Standards

NO Gas Cylinder #: CC363447
NOX Cal Gas Conc: 48.80 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.80 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: February 2, 2024
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3060
Serial Number: 260

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1426262592

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.882	1.813	NO bkgnd or offset:	7.9	7.6
NOX coeff or slope:	1.000	0.993	NOX bkgnd or offset:	8.1	7.7
NO2 coeff or slope:	0.995	1.000	Reaction cell Press:	241.1	241.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996544	0.995016
NO _x Cal Offset:	1.320000	-0.320000
NO Cal Slope:	0.995845	0.998443
NO Cal Offset:	0.740000	-0.920000
NO ₂ Cal Slope:	1.003157	0.994665
NO ₂ Cal Offset:	0.613431	-1.310360



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as found span	4918	82.0	800.3	800.3	0.0	829.1	826.7	2.4	0.9653	0.9681
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
high point	4918	82.0	800.3	800.3	0.0	796.3	798.7	-2.4	1.0050	1.0020
second point	4959	41.0	400.2	400.2	0.0	397.4	398.0	-0.6	1.0069	1.0054
third point	4980	20.5	200.1	200.1	0.0	198.5	197.9	0.6	1.0080	1.0110
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4918	82.0	800.3	432.4	367.9	796.3	437.5	358.7	1.0050	0.9884
Average Correction Factor									1.0067	1.0062

Corrected As found	NO _x = 829.2 ppb	NO = 826.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 3.7%
Previous Response	NO _x = 798.9 ppb	NO = 797.7 ppb		*Percent Change	NO = 3.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.1	427.2	367.9	365.2	1.0074	99.3%
2nd GPT point (200 ppb O3)	795.1	619.0	176.1	173.3	1.0162	98.4%
3rd GPT point (100 ppb O3)	795.1	709.6	85.5	82.4	1.0376	96.4%
Average Correction Factor					1.0204	98.0%

Notes: Sample inlet filter changed after as founds. Adjustment made to span high point.

Calibration Performed By: Morgan Voyageur, Matthew C



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 13, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:28	End Time (MST):	12:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	796.3	1.0050		
400.2	397.4	1.0069		
200.1	198.5	1.0080		
			0.999998	
			0.995016	
			-0.320000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 13, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:28	End Time (MST):	12:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.3	798.7	1.0020			
400.2	398.0	1.0054			
200.1	197.9	1.0110			
			Slope	0.998443	0.90 - 1.10
			Intercept	-0.920000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 13, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:28	End Time (MST):	12:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999944	≥0.995
367.9	365.2	1.0074			
176.1	173.3	1.0162			
85.5	82.4	1.0376			
			Slope	0.994665	0.90 - 1.10
			Intercept	-1.310360	+/-20



NO_x Calibration Plot

Date: May 15, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: May 10, 2023 Last Cal Date: April 12, 2023
 Start time (MST): 7:39 End time (MST): 10:33
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 260

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3872
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.013200	1.007686	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	-0.960000	-2.020000	Coeff or Slope:	1.036	1.036

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	NA	0.0	-0.3	----
as found span	5000	963.6	400.0	399.8	1.001
as found 2nd point					
as found 3rd point					
calibrator zero	5000	NA	0.0	0.4	----
high point	5000	961.7	400.0	402.4	0.994
second point	5000	810.3	200.0	197.9	1.011
third point	5000	701.3	100.0	96.6	1.035
as left zero	5000	NA	0.0	0.3	----
as left span	5000	963.3	400.0	405.2	0.987
Average Correction Factor					1.013

Baseline Corr As found:	400.1	Previous response	404.3	*% change	-1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
 Calibration Date: May 29, 2023 Last Cal Date: April 26, 2023
 Start time (MST): 13:33 End time (MST): 15:21

Analyzer Make: API S/N: 216
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1212
 Temp/RH standard: Delta Cal S/N: 1212

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	18.8	18.5	18.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	726.1	717.3	726.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	4.49	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>May 29, 2023</u>		Last Cal Date: _____		
	PM w/o HEPA: <u>0.1</u>		PM w/ HEPA: <u>0.0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>8.4</u>		w/ HEPA: <u>0.0</u>	
Date Optical Chamber Cleaned:		<u>March 13, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>March 13, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: July 14, 2022
 Date RH/T Sensor Cleaned: July 14, 2022

Notes: No adjustments made.

Calibration by: Morgan Voyageur



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	May 17, 2023	Last Cal Date:	April 21, 2023
Start time (MST):	7:41	End time (MST):	10:25
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API T701H		Serial Number:	260

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997452	0.999682	Backgd or Offset:	-0.014	-0.014
Calibration intercept:	0.118914	0.126989	Coeff or Slope:	1.005	0.968

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.00	----
as found span	4933	66.7	40.4	43.0	0.940
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4934	66.7	40.4	40.4	1.000
second point	4967	33.3	20.2	20.5	0.983
third point	4983	16.7	10.1	10.2	0.988
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.4	40.3	1.001
Average Correction Factor					0.991

Baseline Corr As found:	42.99	Prev response:	40.44	*% change:	5.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Adjusted span. installed new calibrator prior.

Calibration Performed By: Morgan Voyageur, Matthew Courtoreille



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 21, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	7:41	End Time (MST):	10:25
Analyzer make:	API T300	Analyzer serial #:	3505

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	≥ 0.995
40.4	40.4	0.9999		
20.2	20.5	0.9834	Slope	0.90 - 1.10
10.1	10.2	0.9883		
			Intercept	± 1.5



CO Calibration Plot

Date: May 17, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	May 17, 2023	Last Cal Date:	April 21, 2023
Start time (MST):	11:14	End time (MST):	16:54
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3060
N2 Gen Make/Model:	NG 5000		Serial Number:	771048318

Analyzer Information

Analyzer make: Teledyne API T360	Analyzer serial #: 289
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992226	0.996425	Backgd or Offset:	0.006	-0.01
Calibration intercept:	-4.080000	-0.680000	Coeff or Slope:	1.018	1.018

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	0.6	----
as found span	2920	80.0	1605.9	1713.7	0.937
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.2	----
high point	2920	80.0	1605.9	1604.4	1.001
second point	2960	40.0	802.9	785.4	1.022
third point	2980	20.0	401.5	407.5	0.985
as left zero	3000	0.0	0.0	26.0	----
as left span					
Average Correction Factor					1.003

Baseline Corr As found:	1713.10	Prev response:	1589.30	*% change:	7.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. As found points are high due to changing calibrator. Slightly adjusted as found and span points.

Calibration Performed By: Morgan Voyageur, Matthew C



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	May 17, 2023	Previous Calibration	April 21, 2023
Station Name	Fort Chipewyan	Station Number	AMS08
Start Time (MST)	11:14	End Time (MST)	16:54
Analyzer make	Teledyne API T360	Analyzer serial #	289

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	0.999795	≥0.995
1605.9	1604.4	1.0009			
802.9	785.4	1.0223	Slope	0.996425	0.90 - 1.10
401.5	407.5	0.9852			
			Intercept	-0.680000	+/-20



CO₂ Calibration Plot

Date: May 17, 2023

Location: Fort Chipewyan





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 11, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:36	End Time (MST):	13:51
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	0.999986	
801.5	801.1	1.0005			≥0.995
400.8	398.0	1.0069	Slope	0.999100	
199.8	199.4	1.0022			0.90 - 1.10
			Intercept	-0.528972	+/-30



SO2 Calibration Plot

Date: May 8, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Barge Landing Station number: AMS09
 Calibration Date: May 3, 2023 Last Cal Date: April 17, 2023
 Start time (MST): 9:47 End time (MST): 14:57
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.87 ppm Cal Gas Exp Date: September 2, 2024
 Cal Gas Cylinder #: EY0002346
 Removed Cal Gas Conc: 4.87 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3812
 ZAG Make/Model: API T701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1331259320
 Converter make: CDN-101 Converter serial #: 519
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008720	1.005434	Backgd or Offset: 2.78	2.82
Calibration intercept:	0.079030	0.019021	Coeff or Slope: 1.137	1.147

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4918	82.1	80.0	77.8	1.028
as found 2nd point	4959	41.1	40.0	39.2	1.021
as found 3rd point	4979	20.5	20.0	19.6	1.019
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4918	82.1	80.0	80.4	0.995
second point	4959	41.1	40.0	40.3	0.993
third point	4979	20.5	20.0	20.1	0.993
as left zero	5000	0.0	0.0	0.0	----
as left span	4918	82.1	80.0	81.1	0.986
SO2 Scrubber Check	4920	80.2	802.0	-0.1	----
Date of last scrubber change:	28-Feb-23			Ave Corr Factor	0.994
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 77.8 Prev response: 80.74 *% change: -3.8%
 Baseline Corr 2nd AF pt: 39.2 AF Slope: 0.972568 AF Intercept: 0.119009
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999986

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SOx scrubber check done after calibrator zero. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021


Station Information

Calibration Date:	May 3, 2023	Previous Calibration:	April 17, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:47	End Time (MST):	14:57
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	1.000000	
80.0	80.4	0.9946			≥0.995
40.0	40.3	0.9933	Slope	1.005434	
20.0	20.1	0.9935			0.90 - 1.10
			Intercept	0.019021	+/-3







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	May 8, 2023	Last Cal Date:	April 11, 2023
Start time (MST):	10:36	End time (MST):	13:51
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 100 ppm		
NMHC Range (ppm):	0 - 50 ppm	CH ₄ Range (ppm):	0 - 50 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.49E-04	2.50E-04	NMHC SP Ratio:	4.79E-05
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	190949
				187500

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	16.99	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.18	0.996
second point	4960	40.1	8.56	8.56	1.000
third point	4980	20.0	4.27	4.28	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.15	0.998

Average Correction Factor				0.998
Baseline Corr AF:	16.99	Prev response	17.07	*% change -0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.0	----
as found span	4919	80.2	9.14	9.03	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.18	0.995
second point	4960	40.1	4.57	4.59	0.995
third point	4980	20	2.28	2.29	0.994
as left zero	5000	0	0.00	0.00	----
as left span	4919	80.2	9.14	9.17	0.996
Average Correction Factor					0.995
Baseline Corr AF:	9.03	Prev response	9.07	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	7.96	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	7.99	0.999
second point	4960	40.1	3.99	3.98	1.004
third point	4980	20.0	1.99	1.99	1.001
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	7.98	1.000
Average Correction Factor					1.001
Baseline Corr AF:	7.96	Prev response	8.00	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.995837	1.003421
THC Cal Offset:	0.017643	-0.007132
CH ₄ Cal Slope:	1.000907	1.001195
CH ₄ Cal Offset:	0.007869	-0.006136
NMHC Cal Slope:	0.991957	1.004627
NMHC Cal Offset:	0.009575	0.001202

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 11, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:36	End Time (MST):	13:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
17.12	17.18	0.9965						
8.56	8.56	0.9998				Slope	1.003421	0.90 - 1.10
4.27	4.28	0.9973						
			Intercept	-0.007132	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 11, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:36	End Time (MST):	13:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	≥ 0.995			
7.98	7.99	0.9987						
3.99	3.98	1.0039				Slope	1.001195	0.90 - 1.10
1.99	1.99	1.0012						
			Intercept	-0.006136	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 11, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:36	End Time (MST):	13:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995			
9.14	9.18	0.9953						
4.57	4.59	0.9953				Slope	1.004627	0.90 - 1.10
2.28	2.29	0.9939						
			Intercept	0.001202	± 0.5			



NMHC Calibration Plot

Date: May 8, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Barge Landing
Calibration Date: May 23, 2023
Start time (MST): 9:39
Reason: Routine
Station number: AMS09
Last Cal Date: April 28, 2023
End time (MST): 14:54

Calibration Standards

NO Gas Cylinder #: DT0036634
NOX Cal Gas Conc: 50.00 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 50.00 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: January 28, 2024
NO Cal Gas Conc: 49.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 49.70 ppm
NO gas Diff:
Serial Number: 3812
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1426262593

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.175	1.175	NO bkgnd or offset:	10.5	10.7
NOX coeff or slope:	0.995	0.995	NOX bkgnd or offset:	10.5	11.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	174.9	174.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997105	1.003622
NO _x Cal Offset:	0.848780	0.329752
NO Cal Slope:	0.996586	1.005383
NO Cal Offset:	-0.413290	-0.991769
NO ₂ Cal Slope:	1.003813	0.999175
NO ₂ Cal Offset:	0.518213	0.249142



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.4	-0.1	0.5	----	----
as found span	4919	80.5	805.1	800.3	4.8	811.0	804.0	7.3	0.993	0.995
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
high point	4919	80.5	805.1	800.3	4.8	808.0	804.0	4.5	0.996	0.995
second point	4959	40.2	402.1	399.7	2.4	404.3	400.3	4.0	0.994	0.998
third point	4979	20.1	201.0	199.8	1.2	202.5	199.2	3.3	0.993	1.003
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
as left span	4919	80.5	805.1	435.7	369.4	804.0	437.2	367.1	1.001	0.996
Average Correction Factor									0.995	0.999

Corrected As found	NO _x = 810.6 ppb	NO = 804.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.9%
Previous Response	NO _x = 803.6 ppb	NO = 797.1 ppb		*Percent Change	NO = 0.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.1	434.5	369.4	369.3	1.000	100.0%
2nd GPT point (200 ppb O3)	799.1	658.9	145.0	145.2	0.999	100.1%
3rd GPT point (100 ppb O3)	799.1	727.2	76.7	77.2	0.994	100.6%
Average Correction Factor					0.998	100.2%

Notes: Changed sample inlet filter after as founds. Adjusted zero.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 23, 2023	Previous Calibration:	April 28, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:39	End Time (MST):	14:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
805.1	808.0	0.9964		
402.1	404.3	0.9945		
201.0	202.5	0.9928		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 23, 2023	Previous Calibration:	April 28, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:39	End Time (MST):	14:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥ 0.995 $0.90 - 1.10$ ± 20
800.3	804.0	0.9953		
399.7	400.3	0.9984		
199.8	199.2	1.0032		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 23, 2023	Previous Calibration:	April 28, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:39	End Time (MST):	14:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.0	0.0	----	Correlation Coefficient	0.999998	≥0.995			
369.4	369.3	1.0004						
145.0	145.2	0.9988				Slope	0.999175	0.90 - 1.10
76.7	77.2	0.9939						
			Intercept	0.249142	+/-20			



NO_x Calibration Plot

Date: May 23, 2023

Location: Barge Landing





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS11 LOWER CAMP MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 11, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:13	End Time (MST):	13:42
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999988	
800.8	798.7	1.0026			≥0.995
400.9	397.5	1.0086	Slope	0.997114	
199.9	198.9	1.0052			0.90 - 1.10
			Intercept	-0.568654	+/-30



SO2 Calibration Plot

Date: May 15, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Lower Camp Station number: AMS11
 Calibration Date: May 16, 2023 Last Cal Date: April 13, 2023
 Start time (MST): 9:15 End time (MST): 13:04
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.429 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC501097
 Removed Cal Gas Conc: 5.429 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3807
 ZAG Make/Model: API T701H Serial Number: 196

Analyzer Information

Analyzer make: Thermo 450iQ Analyzer serial #: CM20080003
 Converter make: NA Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.015352	1.004895	Backgd or Offset: 14.0	13.3
Calibration intercept:	0.293967	0.035134	Coeff or Slope: 1.043	1.001

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4926	73.6	79.9	83.2	0.963
as found 2nd point	4963	36.8	40.0	41.7	0.963
as found 3rd point	4982	18.6	20.2	20.8	0.980
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4926	73.6	79.9	80.4	0.994
second point	4963	36.8	40.0	40.2	0.994
third point	4982	18.6	20.2	20.1	1.005
as left zero	5000	0.0	0.0	-0.1	----
as left span	4926	73.6	79.9	79.6	1.004
SO2 Scrubber Check	4919	81.1	811.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.998
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 83.0 Prev response: 81.44 *% change: 1.9%
 Baseline Corr 2nd AF pt: 41.5 AF Slope: 1.040103 AF Intercept: 0.052205
 Baseline Corr 3rd AF pt: 20.6 AF Correlation: 0.999975

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero.
 Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	April 13, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:15	End Time (MST):	13:04
Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999977	≥0.995
79.9	80.4	0.9940			
40.0	40.2	0.9940	Slope	1.004895	0.90 - 1.10
20.2	20.1	1.0046			
			Intercept	0.035134	+/-3



H₂S Calibration Plot

Date: May 16, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	May 15, 2023	Last Cal Date:	April 11, 2023
Start time (MST):	10:13	End time (MST):	13:42
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
ZAG make/model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.02E-04	2.99E-04	NMHC SP Ratio:	5.86E-05
CH ₄ Retention time:	13.8	14.0	NMHC Peak Area:	156599
				158468

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	17.35	17.44	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	17.35	17.29	1.003
second point	4959	40.7	8.69	8.64	1.005
third point	4980	20.3	4.33	4.32	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	17.35	17.41	0.997

Average Correction Factor				1.004
Baseline Corr AF:	17.44	Prev response	17.29	*% change 0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	9.19	9.27	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	9.19	9.17	1.002
second point	4959	40.7	4.60	4.58	1.005
third point	4980	20.3	2.29	2.30	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	9.19	9.22	0.996
Average Correction Factor					1.002
Baseline Corr AF:	9.27	Prev response	9.14	*% change	1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	8.16	8.17	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	8.16	8.12	1.005
second point	4959	40.7	4.09	4.06	1.006
third point	4980	20.3	2.04	2.03	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	8.16	8.19	0.997
Average Correction Factor					1.005
Baseline Corr AF:	8.17	Prev response	8.15	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998101	0.996470
THC Cal Offset:	-0.029983	-0.001598
CH ₄ Cal Slope:	1.000279	0.994977
CH ₄ Cal Offset:	-0.017886	-0.000697
NMHC Cal Slope:	0.996415	0.997386
NMHC Cal Offset:	-0.012098	-0.000502

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 11, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:13	End Time (MST):	13:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
17.35	17.29	1.0033						
8.69	8.64	1.0054				Slope	0.996470	0.90 - 1.10
4.33	4.32	1.0021						
			Intercept	-0.001598	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 11, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:13	End Time (MST):	13:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	≥0.995			
8.16	8.12	1.0051						
4.09	4.06	1.0055				Slope	0.994977	0.90 - 1.10
2.04	2.03	1.0054						
			Intercept	-0.000697	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 11, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:13	End Time (MST):	13:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995			
9.19	9.17	1.0022						
4.60	4.58	1.0053				Slope	0.997386	0.90 - 1.10
2.29	2.30	0.9997						
			Intercept	-0.000502	± 0.5			



NMHC Calibration Plot

Date: May 15, 2023

Location: Lower Camp





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13 FORT MCKAY SOUTH MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 19, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:37	End Time (MST):	11:36
Analyzer make:	API T100	Analyzer serial #:	599

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999964	
799.7	799.8	0.9999			≥0.995
399.3	397.8	1.0038	Slope	1.001326	
200.2	196.0	1.0214			0.90 - 1.10
			Intercept	-1.757849	+/-30



SO2 Calibration Plot

Date: May 17, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	May 2, 2023	Last Cal Date:	April 3, 2023
Start time (MST):	8:38	End time (MST):	12:35
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.34	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC500241			
Removed Cal Gas Conc:	5.34	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2448
ZAG Make/Model:	Teledyne API 701		Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017
Converter make:	CDN-101	Converter serial #:	521
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003341	1.005466	Backgd or Offset:	3.71	3.68
Calibration intercept:	-0.062227	-0.262184	Coeff or Slope:	1.120	1.116

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4925	75.5	80.6	80.8	0.997
as found 2nd point	4962	37.7	40.3	40.2	0.999
as found 3rd point	4981	18.9	20.2	19.6	1.025
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4925	75.5	80.6	80.9	0.997
second point	4962	37.7	40.3	40.2	1.002
third point	4981	18.9	20.2	19.7	1.025
as left zero	5000	0.0	0.0	0.1	----
as left span	4925	75.5	80.6	81.0	0.995
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:	20-Mar-20		Ave Corr Factor		1.008
Date of last converter efficiency test:	NA		efficiency		

Baseline Corr As found:	80.9	Prev response:	80.83	*% change:	0.1%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.005607	AF Intercept:	-0.342143
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999946		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021


Station Information

Calibration Date:	May 2, 2023	Previous Calibration:	April 3, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:38	End Time (MST):	13:36
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999947	
80.6	80.9	0.9966			≥0.995
40.3	40.2	1.0016	Slope	1.005466	
20.2	19.7	1.0246			0.90 - 1.10
			Intercept	-0.262184	+/-3







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	May 17, 2023	Last Cal Date:	April 19, 2023
Start time (MST):	8:37	End time (MST):	11:36
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API 701	Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.19E-04	2.21E-04	NMHC SP Ratio:	5.04E-04
CH ₄ Retention time:	12.8	12.8	NMHC Peak Area:	179990
				179419

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	17.05	16.97	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.02	1.001
second point	4961	39.5	8.51	8.39	1.014
third point	4980	19.8	4.27	4.12	1.036
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	17.05	17.08	0.998

Average Correction Factor	1.017
---------------------------	-------

Baseline Corr AF:	16.97	Prev response	16.99	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.1	9.08	9.03	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.1	9.08	9.06	1.002
second point	4960	39.5	4.53	4.48	1.012
third point	4980	19.8	2.27	2.20	1.034
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.1	9.08	9.07	1.001
Average Correction Factor					1.016
Baseline Corr AF:	9.03	Prev response	9.05	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	7.97	7.94	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.96	1.000
second point	4960	39.5	3.98	3.91	1.017
third point	4980	19.8	1.99	1.92	1.039
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	7.97	8.01	0.995
Average Correction Factor					1.019
Baseline Corr AF:	7.94	Prev response	7.94	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002569	1.000879
THC Cal Offset:	-0.096376	-0.078771
CH ₄ Cal Slope:	1.003704	1.001767
CH ₄ Cal Offset:	-0.056151	-0.042148
NMHC Cal Slope:	1.001425	0.999952
NMHC Cal Offset:	-0.040562	-0.036961

Notes:

Changed inlet filter after as founds. Adjusted span.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 19, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:37	End Time (MST):	11:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999902	≥ 0.995			
17.05	17.02	1.0013						
8.51	8.39	1.0140				Slope	1.000879	0.90 - 1.10
4.27	4.12	1.0362						
			Intercept	-0.078771	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 19, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:37	End Time (MST):	11:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999868	≥ 0.995			
7.97	7.96	1.0005						
3.98	3.91	1.0168				Slope	1.001767	0.90 - 1.10
1.99	1.92	1.0387						
			Intercept	-0.042148	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 19, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:37	End Time (MST):	11:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999924	≥ 0.995			
9.08	9.06	1.0021						
4.53	4.48	1.0122				Slope	0.999952	0.90 - 1.10
2.27	2.20	1.0340						
			Intercept	-0.036961	± 0.5			



NMHC Calibration Plot

Date: May 17, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as found span	4919	81.1	826.9	800.0	26.9	834.1	805.3	28.8	0.9913	0.9934
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.1	----	----
high point	4919	81.1	826.9	800.0	26.9	824.0	799.4	24.7	1.0035	1.0007
second point	4960	40.6	413.9	400.4	13.5	410.4	396.7	13.7	1.0085	1.0094
third point	4980	20.3	207.0	200.2	6.7	202.6	194.5	8.1	1.0216	1.0294
as left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
as left span	4919	81.1	826.9	375.2	451.7	835.6	381.0	454.5	0.9896	0.9847
Average Correction Factor									1.0112	1.0132

Corrected As found	NO _x = 834.1 ppb	NO = 805.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.0%
Previous Response	NO _x = 825.6 ppb	NO = 799.4 ppb		*Percent Change	NO = 0.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.8	373.0	451.7	450.4	1.0029	99.7%
2nd GPT point (200 ppb O3)	797.8	588.1	236.6	234.4	1.0095	99.1%
3rd GPT point (100 ppb O3)	797.8	692.9	131.8	129.3	1.0195	98.1%
Average Correction Factor					1.0107	99.0%

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 9, 2023	Previous Calibration:	April 12, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:41	End Time (MST):	13:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.9	824.0	1.0035		
413.9	410.4	1.0085		
207.0	202.6	1.0216		
			0.999976	
			0.997742	
			-1.871090	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 9, 2023	Previous Calibration:	April 12, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:41	End Time (MST):	13:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
800.0	799.4	1.0007		
400.4	396.7	1.0094		
200.2	194.5	1.0294		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 9, 2023	Previous Calibration:	April 12, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:41	End Time (MST):	13:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
451.7	450.4	1.0029			
236.6	234.4	1.0095			
131.8	129.3	1.0195			
			Slope	0.997870	0.90 - 1.10
			Intercept	-1.056811	+/-20



NO_x Calibration Plot

Date: May 9, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort McKay South Station number: AMS 13
Calibration Date: May 29, 2023 Last Cal Date: May 9, 2023
Start time (MST): 10:13 End time (MST): 15:52
Reason: Maintenance Changing pump

Calibration Standards

NO Gas Cylinder #: T2Y1P76 Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.98 ppm NO Cal Gas Conc: 49.32 ppm
Removed Cylinder #: N/A Removed Gas Exp Date: N/A
Removed Gas NOX Conc: 50.98 ppm Removed Gas NO Conc: 49.32 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: API T700 Serial Number: 2448
ZAG make/model: API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1410661329
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.203	1.203	NO bkgnd or offset:	9.5	9.5
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	9.5	9.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	195.9	195.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997742	
NO _x Cal Offset:	-1.871090	
NO Cal Slope:	1.001205	
NO Cal Offset:	-2.924978	
NO ₂ Cal Slope:	0.997870	
NO ₂ Cal Offset:	-1.056811	



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.2	-0.1	0.2	----	----
as found span	4919	81.1	826.9	800.0	26.9	716.8	690.9	25.9	1.1536	1.1578
as found 2nd	4960	40.6	413.9	400.4	13.5	356.6	342.6	14.0	1.1607	1.1688
as found 3rd	4980	20.3	207.0	200.2	6.7	176.1	168.1	8.1	1.1753	1.1911
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

				Average Correction Factor			
Corrected As found	NO _x = 716.6 ppb	NO = 691.0 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = -14.9%	
Previous Response	NO _x = 823.1 ppb	NO = 798.0 ppb			*Percent Change	NO = -15.5%	
Baseline Corr 2nd pt	NO _x = 356.4 ppb	NO = 342.7 ppb	As found	NO _x r ² : 0.999970	Nx SI: 0.867747	Nx Int: -1.646	
Baseline Corr 3rd pt	NO _x = 175.9 ppb	NO = 168.2 ppb	As found	NO r ² : 0.999938	NO SI: 0.865378	NO Int: -2.640	
			As found	NO ₂ r ² : 0.999978	NO ₂ SI: 0.989898	NO ₂ Int: -0.586	

GPT Calibration Data

O ₃ Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	0.2	----	----
as found GPT point (400 ppb NO ₂)	690.1	322.1	394.9	390.6	1.0111	98.9%
as found GPT point (200 ppb NO ₂)	690.1	322.1	394.9	390.6	1.0111	98.9%
as found GPT point (100 ppb NO ₂)	690.1	563.3	153.7	150.3	1.0228	97.8%
1st GPT point (400 ppb O ₃)						
2nd GPT point (200 ppb O ₃)						
3rd GPT point (100 ppb O ₃)						

		Average Correction Factor	


Notes: Here to address the low spans. Changed inlet filter and pump after multi-poins as founds. Will let it settle overnight and continue the calibration tomorrow.

Calibration Performed By: Sean Bala

NO_x Calibration Plot

Date: May 29, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4919	81.1	826.9	800.0	26.9	825.0	801.1	23.8	1.0023	0.9986
second point	4960	40.6	413.9	400.4	13.5	410.8	397.8	13.0	1.0076	1.0066
third point	4980	20.3	207.0	200.2	6.7	203.0	195.2	7.7	1.0195	1.0258
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
as left span	4919	80.1	816.8	364.5	452.3	834.7	380.4	454.3	0.9786	0.9583
Average Correction Factor									1.0098	1.0103

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x =	NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO =	NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:	
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	801.1	375.7	452.3	448.1	1.0094	99.1%
2nd GPT point (200 ppb O3)	801.1	593.9	234.1	230.0	1.0179	98.2%
3rd GPT point (100 ppb O3)	801.1	698.1	129.9	126.4	1.0279	97.3%
Average Correction Factor					1.0184	98.2%

Notes:

Adjusted zero and span.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	May 29, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:27	End Time (MST):	12:24
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
826.9	825.0	1.0023			
413.9	410.8	1.0076			
207.0	203.0	1.0195			
			Slope	0.998875	0.90 - 1.10
			Intercept	-1.831289	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	May 29, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:27	End Time (MST):	12:24
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 $0.90 - 1.10$ ± 20
800.0	801.1	0.9986		
400.4	397.8	1.0066		
200.2	195.2	1.0258		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	May 29, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:27	End Time (MST):	12:24
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999959	≥0.995
452.3	448.1	1.0094			
234.1	230.0	1.0179			
129.9	126.4	1.0279			
			Slope	0.991692	0.90 - 1.10
			Intercept	-1.272971	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 30, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: May 8, 2023 Last Cal Date: April 11, 2023
 Start time (MST): 9:46 End time (MST): 13:05
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3871
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997171	1.003343	Backgd or Offset:	3.4	3.4
Calibration intercept:	0.620000	0.440000	Coeff or Slope:	0.963	0.966

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.6	----
as found span	5000	969.9	400.0	398.7	1.003
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.3	----
high point	5000	980.6	400.0	401.5	0.996
second point	5000	838.0	200.0	201.2	0.994
third point	5000	735.3	100.0	101.7	0.983
as left zero	5000	0.0	0.0	-0.4	----
as left span	5000	979.1	400.0	401.8	0.996
Average Correction Factor					0.991

Baseline Corr As found:	399.3	Previous response	399.5	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Span adjusted.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 11, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	9:46	End Time (MST):	13:05
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.3	----	Correlation Coefficient	0.999983	
400.0	401.5	0.9963			≥0.995
200.0	201.2	0.9940	Slope	1.003343	
100.0	101.7	0.9833			0.90 - 1.10
			Intercept	0.440000	+/- 5



O₃ Calibration Plot

Date: May 8, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort McKay South Station number: AMS 13
 Calibration Date: May 9, 2023 Last Cal Date: April 12, 2023
 Start time (MST): 10:52 End time (MST): 11:08

Analyzer Make: API T640 S/N: 319
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 141229
 Temp/RH standard: Delta Cal S/N: 141229

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	20.3	21.7	20.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	734.1	735.0	734.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.91	5.16	5.16	<input checked="" type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>May 9, 2023</u>	Last Cal Date: <u>April 12, 2023</u>			
	PM w/o HEPA: <u>17.9</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>April 12, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>March 22, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Inlet head clean and inspected. Flow adjusted.

Notes:

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC
MAY 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 2, 2023	Previous Calibration:	April 18, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	6:40	End Time (MST):	9:17
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999988	≥0.995
800.2	799.4	1.0010			
399.6	398.2	1.0035	Slope	0.999181	0.90 - 1.10
199.8	197.5	1.0116			
			Intercept	-0.684110	+/-30



SO2 Calibration Plot

Date: May 2, 2023

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Anzac Station number: AMS14
 Calibration Date: May 11, 2023 Last Cal Date: April 5, 2023
 Start time (MST): 7:00 End time (MST): 10:47
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.38 ppm Cal Gas Exp Date: February 3, 2023
 Cal Gas Cylinder #: EY0000859
 Removed Cal Gas Conc: 5.38 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 5239
 ZAG Make/Model: API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1180540019
 Converter make: CD Nova CDN-101 Converter serial #: 503
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002553	0.997841	Backgd or Offset:	5.51	5.36
Calibration intercept:	-0.041067	0.098854	Coeff or Slope:	0.993	0.972

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4925	74.3	80.0	83.1	0.965
as found 2nd point	4962	37.2	40.0	41.3	0.974
as found 3rd point	4981	18.6	20.0	20.4	0.991
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4925	74.3	80.0	79.9	1.001
second point	4962	37.2	40.0	40.2	0.996
third point	4981	18.6	20.0	19.7	1.016
as left zero	5000	0.0	0.0	0.4	----
as left span	4925	74.3	80.0	79.4	1.007
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.004
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 82.9 Prev response: 80.12 *% change: 3.4%
 Baseline Corr 2nd AF pt: 41.1 AF Slope: 1.038574 AF Intercept: -0.101851
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999939

* = > +/-5% change initiates investigation

Notes: Scrubber checked after the calibrator zero. No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 5, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	7:00	End Time (MST):	10:47
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540019

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999942	
80.0	79.9	1.0007			≥0.995
40.0	40.2	0.9959	Slope	0.997841	
20.0	19.7	1.0160			0.90 - 1.10
			Intercept	0.098854	+/-3



TRS Calibration Plot

Date: May 11, 2023

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	May 2, 2023	Last Cal Date:	April 18, 2023
Start time (MST):	6:40	End time (MST):	9:18
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	5252
ZAG make/model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.90E-04	3.90E-04	NMHC SP Ratio:	4.53E-05
CH ₄ Retention time:	12.20	12.20	NMHC Peak Area:	201206
				201206

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	17.12	17.06	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	17.12	17.05	1.004
second point	4960	40.0	8.55	8.53	1.002
third point	4980	20.0	4.28	4.23	1.011
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.1	17.12	17.03	1.005

Average Correction Factor				1.006
Baseline Corr AF:	17.06	Prev response	17.03	*% change 0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	9.12	9.09	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	9.12	9.08	1.005
second point	4960	40.0	4.56	4.54	1.004
third point	4980	20.0	2.28	2.25	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.1	9.12	9.07	1.006
Average Correction Factor					1.007
Baseline Corr AF:	9.09	Prev response	9.11	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	8.00	7.97	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	8.00	7.97	1.004
second point	4960	40.0	3.99	3.99	1.001
third point	4980	20.0	2.00	1.98	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.1	8.00	7.96	1.005
Average Correction Factor					1.004
Baseline Corr AF:	7.97	Prev response	7.99	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.994025	0.996430
THC Cal Offset:	0.010201	-0.007805
CH ₄ Cal Slope:	0.999707	0.996993
CH ₄ Cal Offset:	-0.004033	-0.002038
NMHC Cal Slope:	0.999569	0.995937
NMHC Cal Offset:	-0.007757	-0.005767

Notes: Hydrogen cylinder changed. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 2, 2023	Previous Calibration:	April 18, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	6:40	End Time (MST):	9:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995			
17.12	17.05	1.0042						
8.55	8.53	1.0024				Slope	0.996430	0.90 - 1.10
4.28	4.23	1.0107						
			Intercept	-0.007805	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 2, 2023	Previous Calibration:	April 18, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	6:40	End Time (MST):	9:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	≥0.995			
8.00	7.97	1.0036						
3.99	3.99	1.0011				Slope	0.996993	0.90 - 1.10
2.00	1.98	1.0087						
			Intercept	-0.002038	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 2, 2023	Previous Calibration:	April 18, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	6:40	End Time (MST):	9:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995			
9.12	9.08	1.0048						
4.56	4.54	1.0036				Slope	0.995937	0.90 - 1.10
2.28	2.25	1.0125						
			Intercept	-0.005767	± 0.5			



NMHC Calibration Plot

Date: May 2, 2023

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name: Anzac Station number: AMS 14
 Calibration Date: May 11, 2023 Last Cal Date: May 2, 2023
 Start time (MST): 5:56 End time (MST): 7:02
 Reason: Cylinder Change Nitrogen Cylinder change

Calibration Standards

Gas Cert Reference: CC279389 Cal Gas Expiry Date: January 5, 2025
 CH₄ Cal Gas Conc. 499.3 ppm CH₄ Equiv Conc. 1068.8 ppm
 C₃H₈ Cal Gas Conc. 207.1 ppm
 Removed Gas Cert: NA Removed Gas Expiry: NA
 Removed CH₄ Conc. 499.3 ppm CH₄ Equiv Conc. 1068.8 ppm
 Removed C₃H₈ Conc. 207.1 ppm
 Diff between cyl (CH₄): Diff between cyl (NM):
 Calibrator Model: API T700 Serial Number: 5239
 ZAG make/model: API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 1118148494
 THC Range (ppm): 0 - 20 ppm
 NMHC Range (ppm): 0 - 10 ppm CH₄ Range (ppm): 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.90E-04	3.90E-04	NMHC SP Ratio:	4.53E-05
CH ₄ Retention time:	12.20	12.20	NMHC Peak Area:	201206
				201206

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	17.12	17.15	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	17.12	17.13	1.000
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	1.000
Baseline Corr AF:	17.15	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	9.12	9.15	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	9.12	9.13	0.999
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.999
Baseline Corr AF:	9.15	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	8.00	8.00	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	8.00	8.00	1.000
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.000
Baseline Corr AF:	8.00	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000454	
THC Cal Offset:	0.000000	
CH ₄ Cal Slope:	1.000172	
CH ₄ Cal Offset:	0.000000	
NMHC Cal Slope:	1.000701	
NMHC Cal Offset:	0.000000	


Notes: Nitrogen cylinder change.

Calibration Performed By: Melissa Lemay

NMHC Calibration Plot

Date: May 11, 2023

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Anzac
Calibration Date: May 1, 2023
Start time (MST): 6:45
Reason: Routine
Station number: AMS 14
Last Cal Date: April 14, 2023
End time (MST): 11:31

Calibration Standards

NO Gas Cylinder #: T2Y1P8D
NOX Cal Gas Conc: 50.92 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 50.92 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API 701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.05 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.05 ppm
NO gas Diff:
Serial Number: 5239
Serial Number: 357

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1426262592

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.361	1.361	NO bkgnd or offset:	3.7	3.7
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.7	3.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	162.7	162.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002446	0.995080
NO _x Cal Offset:	-0.745161	-0.466631
NO Cal Slope:	1.004625	0.996609
NO Cal Offset:	-2.149028	-1.790505
NO ₂ Cal Slope:	1.000896	1.001233
NO ₂ Cal Offset:	0.328255	0.590521



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.2	0.2	----	----
as found span	4921	78.6	800.5	786.8	13.7	798.0	784.0	14.0	1.0032	1.0036
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
high point	4921	78.6	800.5	786.8	13.7	796.2	783.1	13.1	1.0054	1.0048
second point	4961	39.3	400.2	393.4	6.8	398.1	389.8	8.3	1.0053	1.0092
third point	4980	19.6	199.6	196.2	3.4	197.2	191.8	5.4	1.0123	1.0230
as left zero	5000	0.0	0.0	0.0	0.0	0.2	-0.1	0.2	----	----
as left span	4921	78.6	800.5	392.1	408.4	794.8	387.9	406.9	1.0072	1.0110
Average Correction Factor									1.0077	1.0123

Corrected As found	NO _x = 797.9 ppb	NO = 784.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.5%	
Previous Response	NO _x = 801.7 ppb	NO = 788.3 ppb		*Percent Change	NO = -0.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	777.7	383.0	408.4	409.3	0.9977	100.2%
2nd GPT point (200 ppb O ₃)	777.7	577.7	213.7	214.6	0.9957	100.4%
3rd GPT point (100 ppb O ₃)	777.7	679.1	112.3	113.5	0.9892	101.1%
Average Correction Factor					0.9942	100.6%

Notes:

No maintenance or adjustments done.

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 14, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	6:45	End Time (MST):	11:31
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.5	796.2	1.0054		
400.2	398.1	1.0053		
199.6	197.2	1.0123		
			0.999996	
			0.995080	
			-0.466631	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 14, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	6:45	End Time (MST):	11:31
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
786.8	783.1	1.0048		
393.4	389.8	1.0092		
196.2	191.8	1.0230		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 14, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	6:45	End Time (MST):	11:31
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
408.4	409.3	0.9977		
213.7	214.6	0.9957		
112.3	113.5	0.9892		
			0.999995	
			1.001233	
			0.590521	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 1, 2023

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	May 2, 2023	Last Cal Date:	April 18, 2023
Start time (MST):	9:15	End time (MST):	11:38
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	API T700	Serial Number:	5239
ZAG Make/Model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595
Analyzer Range:	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000057	1.007857	Backgd or Offset:	1.3	1.3
Calibration intercept:	0.140000	-0.500000	Coeff or Slope:	1.516	1.516

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	5000	884.2	400.0	402.9	0.993
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.8	----
high point	5000	884.9	400.0	402.4	0.994
second point	5000	769.3	200.0	201.5	0.993
third point	5000	669.8	100.0	100.4	0.996
as left zero	5000	0.0	0.0	-0.6	----
as left span	5000	883.1	400.0	405.1	0.987
Average Correction Factor					0.994

Baseline Corr As found:	403.1	Previous response	400.2	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 2, 2023	Previous Calibration:	April 18, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	9:15	End Time (MST):	11:38
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.8	----	Correlation Coefficient	0.999996	≥0.995
400.0	402.4	0.9940			
200.0	201.5	0.9926	Slope	1.007857	0.90 - 1.10
100.0	100.4	0.9960			
			Intercept	-0.500000	+/- 5



O₃ Calibration Plot

Date: May 2, 2023

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Anzac Station number: AMS 14
 Calibration Date: May 30, 2023 Last Cal Date: April 18, 2023
 Start time (MST): 8:26 End time (MST): 8:49

Analyzer Make: API T640 S/N: 825
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388753
 Temp/RH standard: Alicat FP-25 S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	19.9	20.4	19.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	711.4	712.7	711.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5	5.05	5	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: May 30, 2023 Last Cal Date: April 18, 2023
 PM w/o HEPA: 44 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: _____ w/ HEPA: _____
 Date Optical Chamber Cleaned: March 30, 2023 <0.2 ug/m3
 Disposable Filter Changed: March 30, 2023

Annual Maintenance

Date Sample Tube Cleaned: June 21, 2022
 Date RH/T Sensor Cleaned: June 21, 2022

Notes: No adjustments done. Head Cleaned.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS17
WAPASU
MAY 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 13, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:58	End Time (MST):	13:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	0.999963	
800.0	798.5	1.0018			≥0.995
400.0	395.1	1.0125	Slope	0.998882	
199.5	196.3	1.0164			0.90 - 1.10
			Intercept	-2.039793	+/-30



SO2 Calibration Plot

Date: May 15, 2023

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	May 23, 2023	Last Cal Date:	April 21, 2023
Start time (MST):	10:19	End time (MST):	14:09
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.076	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511852			
Removed Cal Gas Conc:	5.076	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2449
ZAG Make/Model:	API T701H		Serial Number:	359

Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583
Converter make:	n/a	Converter serial #:	n/a
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.990854	0.987854	Backgd or Offset:	12.7
Calibration intercept:	0.220786	0.100766	Coeff or Slope:	1.085

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4921	78.8	80.0	79.8	1.004
as found 2nd point	4961	39.4	40.0	39.4	1.018
as found 3rd point	4980	19.7	20.0	19.8	1.015
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4921	78.8	80.0	79.4	1.008
second point	4961	39.4	40.0	39.1	1.023
third point	4980	19.7	20.0	19.8	1.010
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	78.8	80.0	78.6	1.018
SO2 Scrubber Check	4921	79.4	800.0	-0.2	----
Date of last scrubber change:	n/a			Ave Corr Factor	1.014
Date of last converter efficiency test:	n/a				efficiency

Baseline Corr As found:	79.7	Prev response:	79.49	*% change:	0.3%
Baseline Corr 2nd AF pt:	39.3	AF Slope:	0.996426	AF Intercept:	-0.099220
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999943		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 23, 2023	Previous Calibration:	April 21, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:19	End Time (MST):	14:09
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999875	≥0.995
80.0	79.4	1.0076			
40.0	39.1	1.0229	Slope	0.987854	0.90 - 1.10
20.0	19.8	1.0101			
			Intercept	0.100766	+/-3



H₂S Calibration Plot

Date: May 23, 2023

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 15, 2023	Previous Calibration:	April 13, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:58	End Time (MST):	13:00
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.01	----	Correlation Coefficient	0.999932	≥0.995
17.09	17.19	0.9942			
8.55	8.48	1.0080	Slope	1.004482	0.90 - 1.10
4.26	4.29	0.9943			
			Intercept	-0.015906	+/-1.5



THC Calibration Plot

Date: May 15, 2023

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Wapasu
Calibration Date: May 18, 2023
Start time (MST): 9:22
Reason: Routine
Station number: AMS17
Last Cal Date: April 25, 2023
End time (MST): 13:46

Calibration Standards

NO Gas Cylinder #: T375YK8
NOX Cal Gas Conc: 49.11 ppm
Removed Cylinder #: T375YK8
Removed Gas NOX Conc: 49.11 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: April 13, 2025
NO Cal Gas Conc: 48.07 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 48.07 ppm
NO gas Diff:
Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 833

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.836	0.836	NO bkgnd or offset:	0.1	0.1
NOX coeff or slope:	0.828	0.828	NOX bkgnd or offset:	-0.4	-0.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.4	4.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001718	1.000389
NO _x Cal Offset:	-1.660000	-0.960000
NO Cal Slope:	1.001759	1.000187
NO Cal Offset:	-1.940000	-1.540000
NO ₂ Cal Slope:	0.991719	0.997323
NO ₂ Cal Offset:	-0.351144	-0.628631



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	2.3	-0.2	2.5	----	----
as found span	4917	83.2	817.2	799.9	17.3	815.7	799.2	16.7	1.0018	1.0009
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.7	0.1	0.6	----	----
high point	4917	83.2	817.2	799.9	17.3	817.2	799.2	17.9	1.0000	1.0009
second point	4958	41.6	408.6	399.9	8.7	407.5	398.0	9.5	1.0027	1.0049
third point	4979	20.8	204.3	200.0	4.3	201.4	196.6	4.8	1.0144	1.0171
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.0	0.4	----	----
as left span	4917	83.2	817.2	400.5	416.7	811.6	397.8	413.7	1.0069	1.0067
Average Correction Factor									1.0057	1.0076

Corrected As found	NO _x = 813.4 ppb	NO = 799.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.4%
Previous Response	NO _x = 816.9 ppb	NO = 799.4 ppb		*Percent Change	NO = 0.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.6	398.2	416.7	416.0	1.0017	99.8%
2nd GPT point (200 ppb O3)	797.6	598.6	216.3	213.4	1.0136	98.7%
3rd GPT point (100 ppb O3)	797.6	699.9	115.0	113.5	1.0133	98.7%
Average Correction Factor					1.0095	99.1%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 18, 2023	Previous Calibration:	April 25, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:22	End Time (MST):	13:46
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.7	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
817.2	817.2	1.0000		
408.6	407.5	1.0027		
204.3	201.4	1.0144		
			0.999980	
			1.000389	
			-0.960000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 18, 2023	Previous Calibration:	April 25, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:22	End Time (MST):	13:46
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
799.9	799.2	1.0009			
399.9	398.0	1.0049			
200.0	196.6	1.0171			
			Slope	1.000187	0.90 - 1.10
			Intercept	-1.540000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 18, 2023	Previous Calibration:	April 25, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:22	End Time (MST):	13:46
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
416.7	416.0	1.0017			
216.3	213.4	1.0136			
115.0	113.5	1.0133			
			Correlation Coefficient	0.999938	≥0.995
			Slope	0.997323	0.90 - 1.10
			Intercept	-0.628631	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 18, 2023

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	May 5, 2023	Last Cal Date:	April 11, 2023
Start time (MST):	9:55	End time (MST):	12:38
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	API T700	Serial Number:	2449
ZAG Make/Model:	API T701H	Serial Number:	359

Analyzer Information

Analyzer make:	API T400	Analyzer serial #:	3870
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008600	1.008857	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.380000	-0.600000	Coeff or Slope:	1.020	1.020

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	5000	1077.3	400.0	403.9	0.990
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	5000	1077.3	400.0	403.1	0.992
second point	5000	900.3	200.0	201.2	0.994
third point	5000	789.5	100.0	99.6	1.004
as left zero	5000	0.0	0.0	0.0	----
as left span	5000	1077.3	400.0	405.6	0.986
Average Correction Factor					0.997

Baseline Corr As found:	404.0	Previous response	403.1	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 5, 2023	Previous Calibration:	April 11, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:55	End Time (MST):	12:38
Analyzer make:	API T400	Analyzer serial #:	3870

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999992	≥0.995
400.0	403.1	0.9923			
200.0	201.2	0.9940	Slope	1.008857	0.90 - 1.10
100.0	99.6	1.0040			
			Intercept	-0.600000	+/- 5



O₃ Calibration Plot

Date: May 5, 2023

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Wapasu Station number: AMS 17
 Calibration Date: May 23, 2023 Last Cal Date: April 25, 2023
 Start time (MST): 13:27 End time (MST): 13:57

Analyzer Make: API T640 S/N: 1183
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388755
 Temp/RH standard: Alicat FP-25BT S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	6.8	6.3	6.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	717.1	719.7	717.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.15	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>May 23, 2023</u>	Last Cal Date: <u>April 25, 2023</u>			
	PM w/o HEPA: <u>81.8</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.0	10	11.1	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>March 23, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>March 23, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: Adjusted the PMT peak. Leak check passed.

Calibration by: Max Farrell



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 27, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:55	End Time (MST):	13:38
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999989	
800.3	800.1	1.0002			≥0.995
400.2	398.2	1.0050	Slope	1.000576	
199.6	197.3	1.0117			0.90 - 1.10
			Intercept	-1.343541	+/-30



SO2 Calibration Plot

Date: May 8, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: May 16, 2023 Last Cal Date: April 26, 2023
 Start time (MST): 10:22 End time (MST): 15:30
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.479 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC500395
 Removed Cal Gas Conc: 5.479 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
 ZAG Make/Model: Teledyne API T701 Serial Number: 360

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359
 Converter make: CD Nova CDN-101 Converter serial #: 555
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994871	1.001444	Backgd or Offset: 2.60	2.59
Calibration intercept:	0.321057	0.240955	Coeff or Slope: 1.151	1.159

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4927	73.0	80.0	77.7	1.031
as found 2nd point	4964	36.5	40.0	38.9	1.031
as found 3rd point	4983	18.3	20.0	19.0	1.061
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4927	73.0	80.0	80.3	0.996
second point	4964	36.5	40.0	40.4	0.990
third point	4983	18.3	20.0	20.3	0.988
as left zero	5000	0.0	0.0	0.4	----
as left span	4927	73.0	80.0	80.3	0.996
SO2 Scrubber Check	4923	77.1	771.0	0.0	----
Date of last scrubber change:	17-Dec-21			Ave Corr Factor	0.991
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 77.6 Prev response: 79.90 *% change: -3.0%
 Baseline Corr 2nd AF pt: 38.8 AF Slope: 0.972430 AF Intercept: -0.118314
 Baseline Corr 3rd AF pt: 18.9 AF Correlation: 0.999937

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Scrubber check completed after calibrator zero. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	April 26, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:22	End Time (MST):	15:30
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999995	
80.0	80.3	0.9962			≥0.995
40.0	40.4	0.9899	Slope	1.001444	
20.0	20.3	0.9876			0.90 - 1.10
			Intercept	0.240955	+/-3



TRS Calibration Plot

Date: May 16, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	May 8, 2023	Last Cal Date:	April 27, 2023
Start time (MST):	9:55	End time (MST):	13:38
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.06E-04	3.10E-04	NMHC SP Ratio:	5.73E-05
CH ₄ Retention time:	14.60	14.60	NMHC Peak Area:	159925
				155690

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	17.28	17.02	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.27	1.001
second point	4959	40.5	8.64	8.61	1.003
third point	4979	20.2	4.31	4.28	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.32	0.998

Average Correction Factor				1.003
Baseline Corr AF:	17.02	Prev response	17.27	*% change -1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	8.96	1.023
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.16	1.001
second point	4959	40.5	4.58	4.57	1.002
third point	4979	20.2	2.29	2.28	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81	9.17	9.19	0.998
Average Correction Factor					1.002
Baseline Corr AF:	8.96	Prev response	9.17	*% change	-2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	8.11	8.06	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	8.11	1.000
second point	4959	40.5	4.06	4.04	1.004
third point	4979	20.2	2.02	2.00	1.011
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.14	0.997
Average Correction Factor					1.005
Baseline Corr AF:	8.06	Prev response	8.10	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999520	0.999626
THC Cal Offset:	-0.003183	-0.013988
CH ₄ Cal Slope:	0.999834	1.000468
CH ₄ Cal Offset:	-0.012012	-0.012012
NMHC Cal Slope:	0.999279	0.998732
NMHC Cal Offset:	0.008428	-0.001376

Notes: Sample inlet filter changed after as founds. N2 cylinder changed after as founds. Adjusted the span only.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 27, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:55	End Time (MST):	13:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995
17.28	17.27	1.0007			
8.64	8.61	1.0032			
4.31	4.28	1.0064			
			Slope	0.999626	0.90 - 1.10
			Intercept	-0.013988	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 27, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:55	End Time (MST):	13:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999990	≥0.995
8.11	8.11	1.0002			
4.06	4.04	1.0042			
2.02	2.00	1.0113			
			Slope	1.000468	0.90 - 1.10
			Intercept	-0.012012	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 8, 2023	Previous Calibration:	April 27, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:55	End Time (MST):	13:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995
9.17	9.16	1.0012			
4.58	4.57	1.0023			
2.29	2.28	1.0017			
			Slope	0.998732	0.90 - 1.10
			Intercept	-0.001376	+/-0.5



NMHC Calibration Plot

Date: May 8, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	May 13, 2023	Last Cal Date:	May 8, 2023
Start time (MST):	11:03	End time (MST):	14:21
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.10E-04	3.15E-04	NMHC SP Ratio:	5.89E-05
CH ₄ Retention time:	14.60	15.20	NMHC Peak Area:	155690
				161883

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	17.28	17.18	1.006
as found 2nd point	4959	40.5	8.64	8.56	1.010
as found 3rd point	4979	20.2	4.31	4.26	1.012
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.27	1.001
second point	4959	40.5	8.64	8.64	1.000
third point	4979	20.2	4.31	4.29	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.27	1.000

Average Correction Factor				1.002
Baseline Corr AF:	17.18	Prev response	17.26	*% change -0.4%
Baseline Corr 2nd AF:	8.6	AF Slope:	0.994661	AF Intercept: -0.018206
Baseline Corr 3rd AF:	4.3	AF Correlation:	0.999993	* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	9.25	0.992
as found 2nd point	4959	40.5	4.58	4.62	0.993
as found 3rd point	4979	20.2	2.29	2.31	0.992
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.16	1.001
second point	4959	40.5	4.58	4.60	0.997
third point	4979	20.2	2.29	2.28	1.001
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81	9.17	9.14	1.004
Average Correction Factor					1.000
Baseline Corr AF:	9.25	Prev response	9.16	*% change	1.0%
Baseline Corr 2nd AF:	4.6	AF Slope:	1.008316	AF Intercept:	-0.001557
Baseline Corr 3rd AF:	2.3	AF Correlation:	0.999999	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	8.11	7.94	1.022
as found 2nd point	4959	40.5	4.06	3.94	1.030
as found 3rd point	4979	20.2	2.02	1.95	1.036
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	8.11	1.000
second point	4959	40.5	4.06	4.04	1.003
third point	4979	20.2	2.02	2.00	1.010
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.13	0.998
Average Correction Factor					1.005
Baseline Corr AF:	7.94	Prev response	8.10	*% change	-2.1%
Baseline Corr 2nd AF:	3.94	AF Slope:	0.978891	AF Intercept:	-0.016450
Baseline Corr 3rd AF:	1.95	AF Correlation:	0.999978	* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999626	0.999792
THC Cal Offset:	-0.013988	-0.006990
CH ₄ Cal Slope:	1.000468	1.000313
CH ₄ Cal Offset:	-0.012012	-0.010212
NMHC Cal Slope:	0.998732	0.998770
NMHC Cal Offset:	-0.001376	0.004221

Notes: Instrument started dipping this morning. After as founds checked the operation of the actuator. It was a little sticky so replaced it. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 13, 2023	Previous Calibration:	May 8, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:03	End Time (MST):	14:21
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
17.28	17.27	1.0007						
8.64	8.64	0.9999				Slope	0.999792	0.90 - 1.10
4.31	4.29	1.0054						
			Intercept	-0.006990	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 13, 2023	Previous Calibration:	May 8, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:03	End Time (MST):	14:21
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999993	≥ 0.995
8.11	8.11	1.0004			
4.06	4.04	1.0032			
2.02	2.00	1.0103			
			Slope	1.000313	0.90 - 1.10
			Intercept	-0.010212	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 13, 2023	Previous Calibration:	May 8, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:03	End Time (MST):	14:21
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995			
9.17	9.16	1.0015						
4.58	4.60	0.9971				Slope	0.998770	0.90 - 1.10
2.29	2.28	1.0012						
			Intercept	0.004221	± 0.5			



NMHC Calibration Plot

Date: May 13, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	May 25, 2023	Last Cal Date:	May 13, 2023
Start time (MST):	10:05	End time (MST):	11:18
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.15E-04	3.15E-04	NMHC SP Ratio:	5.66E-05
CH ₄ Retention time:	15.20	15.20	NMHC Peak Area:	161883

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919	81.0	17.28	17.11	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4919	81.0	17.28	17.10	1.011
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	1.011
Baseline Corr AF:	17.10	Prev response	17.27	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	8.94	1.026
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	8.91	1.029
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.029
Baseline Corr AF:	8.94	Prev response	9.16	*% change	-2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919	81.0	8.11	8.17	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4919	81.0	8.11	8.19	0.991
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.991
Baseline Corr AF:	8.17	Prev response	8.11	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999792	0.988812
THC Cal Offset:	-0.006990	0.008000
CH ₄ Cal Slope:	1.000313	1.008017
CH ₄ Cal Offset:	-0.010212	0.008000
NMHC Cal Slope:	0.998770	0.971926
NMHC Cal Offset:	0.004221	0.000000


Notes: N2 cylinder change. No change in response.

Calibration Performed By: Aswin Sasi Kumar

NMHC Calibration Plot

Date: May 25, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	May 31, 2023	Last Cal Date:	May 25, 2023
Start time (MST):	10:43	End time (MST):	12:06
Reason:	Cylinder Change H2 Cylinder change		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.8 ppm	CH4 Equiv Conc.	1066.8 ppm
C3H8 Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	500.8 ppm	CH4 Equiv Conc.	1066.8 ppm
Removed C3H8 Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.15E-04	3.15E-04	NMHC SP Ratio:	5.66E-05
CH4 Retention time:	15.20	15.20	NMHC Peak Area:	161883
				161883

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	17.28	16.89	1.023
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4919	81.0	17.28	16.82	1.027
second point					
third point					
as left zero					
as left span					
Average Correction Factor					1.027
Baseline Corr AF:	16.89	Prev response	17.27	*% change	-2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	8.76	1.047
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	8.74	1.049
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.049
Baseline Corr AF:	8.76	Prev response	9.16	*% change	-4.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	8.11	8.13	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4919	81.0	8.11	8.09	1.003
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.003
Baseline Corr AF:	8.13	Prev response	8.11	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999792	0.972841
THC Cal Offset:	-0.006990	0.008000
CH ₄ Cal Slope:	1.000313	0.996184
CH ₄ Cal Offset:	-0.010212	0.008000
NMHC Cal Slope:	0.998770	0.953275
NMHC Cal Offset:	0.004221	0.000000

Notes:

H2 Cylinder change.


Calibration Performed By:

Max Farrell

NMHC Calibration Plot

Date: May 31, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: May 10, 2023 Last Cal Date: April 20, 2023
Start time (MST): 10:16 End time (MST): 15:21
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2XX7ME Cal Gas Expiry Date: January 14, 2024
NOX Cal Gas Conc: 50.48 ppm NO Cal Gas Conc: 49.22 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.48 ppm Removed Gas NO Conc: 49.22 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 2658
ZAG make/model: Teledyne API 701H Serial Number: 360

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1336160088
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.062	1.035	NO bkgnd or offset:	3.0	2.9
NOX coeff or slope:	0.984	0.986	NOX bkgnd or offset:	3.0	2.9
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	230.2	220.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.993566	1.000165
NO _x Cal Offset:	0.129802	0.109832
NO Cal Slope:	0.995555	0.999810
NO Cal Offset:	-1.209770	-1.249778
NO ₂ Cal Slope:	0.999137	0.999469
NO ₂ Cal Offset:	0.306137	1.142324



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
as found span	4919	81.3	820.8	800.3	20.5	841.0	818.1	23.1	0.9759	0.9782
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.1	----	----
high point	4919	81.3	820.8	800.3	20.5	820.9	799.4	21.5	0.9998	1.0011
second point	4959	40.7	410.9	400.7	10.3	411.5	399.0	12.5	0.9986	1.0042
third point	4980	20.3	204.9	199.8	5.1	204.7	197.1	7.6	1.0012	1.0138
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
as left span	4919	81.3	820.8	357.7	463.1	825.4	360.5	465.0	0.9944	0.9921
Average Correction Factor									0.9999	1.0064

Corrected As found	NO _x = 841.0 ppb	NO = 818.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 3.0%
Previous Response	NO _x = 815.6 ppb	NO = 795.5 ppb		*Percent Change	NO = 2.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	799.0	356.4	463.1	463.5	0.9991	100.1%
2nd GPT point (200 ppb O ₃)	799.0	595.1	224.4	225.8	0.9937	100.6%
3rd GPT point (100 ppb O ₃)	799.0	699.6	119.9	122.1	0.9819	101.8%
Average Correction Factor					0.9916	100.9%

Notes: Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 10, 2023	Previous Calibration:	April 20, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:16	End Time (MST):	15:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
820.8	820.9	0.9998			
410.9	411.5	0.9986			
204.9	204.7	1.0012			
			Slope	1.000165	0.90 - 1.10
			Intercept	0.109832	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 10, 2023	Previous Calibration:	April 20, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:16	End Time (MST):	15:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	799.4	1.0011		
400.7	399.0	1.0042		
199.8	197.1	1.0138		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 10, 2023	Previous Calibration:	April 20, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:16	End Time (MST):	15:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
463.1	463.5	0.9991			
224.4	225.8	0.9937			
119.9	122.1	0.9819			
			Slope	0.999469	0.90 - 1.10
			Intercept	1.142324	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 10, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: May 4, 2023 Last Cal Date: April 4, 2023
 Start time (MST): 9:50 End time (MST): 13:29
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
 ZAG Make/Model: Teledyne API T701H Serial Number: 360

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 825
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999114	0.998229	Backgd or Offset:	1.000	1.000
Calibration intercept:	-0.420000	-0.340000	Coeff or Slope:	0.993	0.987

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.2	----
as found span	4888	1096.9	400.0	405.5	0.986
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.1	----
high point	4888	1101.7	400.0	399.3	1.002
second point	4888	863.9	200.0	198.7	1.007
third point	4888	741.4	100.0	99.3	1.007
as left zero	5000	800.0	0.0	0.4	----
as left span	4812	1097.9	400.0	404.0	0.990
Average Correction Factor					1.005

Baseline Corr As found:	405.3	Previous response	399.2	*% change	1.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 4, 2023	Previous Calibration:	April 4, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	9:50	End Time (MST):	13:29
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999992	≥0.995
400.0	399.3	1.0018			
200.0	198.7	1.0065	Slope	0.998229	0.90 - 1.10
100.0	99.3	1.0070			
			Intercept	-0.340000	+/- 5



O₃ Calibration Plot

Date: May 4, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: May 16, 2023 Last Cal Date: April 19, 2023
Start time (MST): 13:55 End time (MST): 14:24
Analyzer Make: API T640 S/N: 1162
Particulate Fraction: PM2.5
Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	13.8	13.65	13.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	704.8	704.54	704.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5	5.016	5	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>May 16, 2023</u>	Last Cal Date: <u>April 19, 2023</u>			
	PM w/o HEPA: <u>29.4</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	N/A		N/A	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>N/A</u>	w/ HEPA: <u>N/A</u>		
Date Optical Chamber Cleaned:		<u>March 22, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>March 22, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022
Date RH/T Sensor Cleaned: August 30, 2022

Notes: No adjustments needed. Leak check passed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	May 9, 2023	Last Cal Date:	April 12, 2023
Start time (MST):	10:05	End time (MST):	13:03
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,050	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	3,050	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701		Serial Number:	360

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011069	1.000645	Backgd or Offset:	-0.012	-0.009
Calibration intercept:	0.019763	0.127800	Coeff or Slope:	0.912	0.905

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.7	41.4	0.984
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4933	66.7	40.7	40.8	0.998
second point	4966	33.3	20.3	20.7	0.984
third point	4983	16.7	10.2	10.3	0.992
as left zero	3000	0.0	0.0	0.1	----
as left span	2920	80.0	81.3	80.6	1.009
Average Correction Factor					0.991

Baseline Corr As found:	41.30	Prev response:	41.16	*% change:	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Adjusted the span only.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 9, 2023	Previous Calibration:	April 12, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:05	End Time (MST):	13:03
Analyzer make:	API T300	Analyzer serial #:	3504

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999944	≥0.995
40.7	40.8	0.9983			
20.3	20.7	0.9838	Slope	1.000645	0.90 - 1.10
10.2	10.3	0.9920			
			Intercept	0.127800	+/-1.5



CO Calibration Plot

Date: May 9, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	May 12, 2023	Previous Calibration	April 19, 2023
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:12	End Time (MST)	13:35
Analyzer make	API T360	Analyzer serial #	283

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	0.999954	≥0.995
1605.9	1599.1	1.0042			
802.9	806.1	0.9961	Slope	0.997386	0.90 - 1.10
401.5	394.7	1.0171			
			Intercept	-0.780000	+/-10



CO₂ Calibration Plot

Date: May 12, 2023

Location: Stony Mountain





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS19
FIREBAG
MAY 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 14, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:20	End Time (MST):	13:51
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999999	≥0.995
799.5	795.4	1.0051			
400.3	399.0	1.0032	Slope	0.994974	0.90 - 1.10
200.1	199.2	1.0045			
			Intercept	0.198788	+/-30



SO2 Calibration Plot

Date: May 1, 2023

Location: Firebag





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Firebag	Station number:	AMS19
Calibration Date:	May 2, 2023	Last Cal Date:	April 3, 2023
Start time (MST):	10:04	End time (MST):	14:42
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.114	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517427			
Removed Cal Gas Conc:	5.114	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1607
ZAG Make/Model:	Teledyne API T701		Serial Number:	1118

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090
Converter make:	Global	Converter serial #:	2022-222
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003054	1.008056	Backgd or Offset:	3.12
Calibration intercept:	0.038381	-0.061725	Coeff or Slope:	0.990
				3.11
				0.983

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4922	78.2	80.0	81.0	0.986
as found 2nd point	4961	39.1	40.0	40.5	0.985
as found 3rd point	4980	19.6	20.0	20.2	0.988
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	78.2	80.0	80.6	0.992
second point	4961	39.1	40.0	40.2	0.995
third point	4980	19.6	20.0	20.1	0.997
as left zero	5000	0.0	0.0	0.0	----
as left span	4922	78.2	80.0	80.4	0.995
SO2 Scrubber Check	4922	78.3	800.2	0.0	----
Date of last scrubber change:	January 18, 2023			Ave Corr Factor	0.995
Date of last converter efficiency test:	n/a			efficiency	

Baseline Corr As found:	81.1	Prev response:	80.26	*% change:	1.0%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.014202	AF Intercept:	-0.101867
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999999		

* = > +/-5% change initiates investigation

Notes: SOx scrubber check done after calibrator zero. Adjusted span. Changed sample inlet filter after MPAF's.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 2, 2023	Previous Calibration:	April 3, 2023
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	10:04	End Time (MST):	14:42
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999997	
80.0	80.6	0.9923			≥0.995
40.0	40.2	0.9948	Slope	1.008056	
20.0	20.1	0.9974			0.90 - 1.10
			Intercept	-0.061725	+/-3



H₂S Calibration Plot

Date: May 2, 2023

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	May 1, 2023	Last Cal Date:	April 14, 2023
Start time (MST):	10:20	End time (MST):	13:51
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1607
ZAG Make/Model:	API T701	Serial Number:	1118

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1336160089
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997928	0.994758	Background:	2.62	2.37
Calibration intercept:	-0.020741	0.018270	Coefficient:	3.815	3.846

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.26	----
as found span	4919	81.1	17.31	16.88	1.025
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.00	0.02	----
high point	4919	81.1	17.31	17.25	1.003
second point	4959	40.6	8.66	8.59	1.009
third point	4980	20.3	4.33	4.35	0.995
as left zero	5000	0.0	0.00	0.04	----
as left span	4919	81.1	17.31	17.40	0.995
Average Correction Factor					1.002
Baseline Corr As found:	17.14	Previous response	17.25	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 14, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:20	End Time (MST):	13:51
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.02	----	Correlation Coefficient	0.999981	≥0.995
17.31	17.25	1.0032			
8.66	8.59	1.0086	Slope	0.994758	0.90 - 1.10
4.33	4.35	0.9953			
			Intercept	0.018270	+/-1.5



THC Calibration Plot

Date: May 1, 2023

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	May 24, 2023	Last Cal Date:	May 1, 2023
Start time (MST):	10:05	End time (MST):	12:40
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1607
ZAG Make/Model:	API T701	Serial Number:	1118

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1336160089
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994758		Background:	2.37	2.37
Calibration intercept:	0.018270		Coefficient:	3.846	3.846

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.32	----
as found span	4919	81.1	17.31	16.91	1.023
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.00	0.06	----
high point	4919	81.1	17.31	17.31	1.000
second point					
third point					
as left zero	5000	0.0	0.00	0.14	----
as left span	4919	81.1	17.31	17.37	0.996
Average Correction Factor					1.000
Baseline Corr As found:	17.23	Previous response	17.23	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation


Notes: Changed hydrogen cylinder. Checked zero and span response after the cylinder was swapped. No adjustment needed.

Calibration Performed By: Braiden Boutillier

THC Calibration Plot

Date: May 24, 2023

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Firebag
Calibration Date: May 4, 2023
Start time (MST): 10:14
Reason: Routine
Station number: AMS 19
Last Cal Date: April 4, 2023
End time (MST): 14:59

Calibration Standards

NO Gas Cylinder #: T2Y1K63
NOX Cal Gas Conc: 51.12 ppm
Removed Cylinder #: n/a
Removed Gas NOX Conc: 51.12 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 30, 2023
NO Cal Gas Conc: 49.40 ppm
Removed Gas Exp Date: n/a
Removed Gas NO Conc: 49.40 ppm
NO gas Diff:
Serial Number: 1607
Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1410661309

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.041	1.078	NO bkgnd or offset:	7.3	7.5
NOX coeff or slope:	0.996	0.995	NOX bkgnd or offset:	7.3	7.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	210.9	210.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.991227	0.999367
NO _x Cal Offset:	0.554416	0.856331
NO Cal Slope:	0.991132	1.001140
NO Cal Offset:	-0.133220	-0.011041
NO ₂ Cal Slope:	0.999545	0.998524
NO ₂ Cal Offset:	-0.476369	-0.233199



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.0	0.0	-0.1	-0.2	0.0	----	----
as found span	4919	81.0	828.1	800.3	27.9	802.0	772.9	29.4	1.0326	1.0354
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	4999	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
high point	4919	81.0	828.1	800.3	27.9	828.0	801.0	26.6	1.0002	0.9991
second point	4960	40.5	414.0	400.1	13.9	415.2	401.0	14.2	0.9972	0.9978
third point	4980	20.2	206.5	199.6	6.9	208.0	199.6	8.4	0.9929	0.9998
as left zero	4999	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as left span	4919	81.0	828.1	359.9	468.3	831.0	360.5	470.8	0.9966	0.9983
Average Correction Factor									0.9967	0.9989

Corrected As found	NO _x = 802.1 ppb	NO = 773.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -2.4%
Previous Response	NO _x = 821.4 ppb	NO = 793.0 ppb		*Percent Change	NO = -2.6%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	801.0	360.6	468.3	467.6	1.0014	99.9%
2nd GPT point (200 ppb O3)	801.0	580.7	248.2	247.2	1.0039	99.6%
3rd GPT point (100 ppb O3)	801.0	691.9	137.0	136.3	1.0049	99.5%
Average Correction Factor					1.0034	99.7%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 4, 2023	Previous Calibration:	April 4, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:14	End Time (MST):	14:59
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
828.1	828.0	1.0002		
414.0	415.2	0.9972		
206.5	208.0	0.9929		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 4, 2023	Previous Calibration:	April 4, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:14	End Time (MST):	14:59
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	801.0	0.9991		
400.1	401.0	0.9978		
199.6	199.6	0.9998		
			0.999999	
			1.001140	
			-0.011041	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 4, 2023	Previous Calibration:	April 4, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:14	End Time (MST):	14:59
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
468.3	467.6	1.0014		
248.2	247.2	1.0039		
137.0	136.3	1.0049		



NO_x Calibration Plot

Date: May 4, 2023

Location: Firebag





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Firebag	Station Number:	AMS 19
Calibration Date:	May 1, 2023	Prev Cal Date:	July 4, 2022
Start Time (MST):	13:40	End Time (MST):	15:12
Tower Height (m):	10.0	Reason:	Routine

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	W15276
WS Calibrator:	MetOne 053	Serial Number:	CA 03845

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.6	0.0%
800	77.8	77.7	-0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥ 0.9995
Calculated slope		1.000283	$0.90 - 1.10$
Calculated intercept		0.005312	± 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	U11347
As Found Declination (deg east of True North):	<u>28</u>	As Left Declination (deg east of True North):	<u>15</u>
Solar noon time (MST):	12:20	Calc Declination*:	15.44 Degrees
Deadband calc:	4.4 degrees (<i>Limit 4 deg</i>)	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
2	3.4	---
90	89.2	-0.2%
180	178.4	-0.4%
270	267.6	-0.7%
357	356.0	-0.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥ 0.9995
Calculated slope		1.007234	$0.90 - 1.10$
Calculated intercept		-0.418403	± 4

Notes: Adjusted tower crossarm to proper orientation. Adjusted using a compass. Deadband out of compliance limits, will return with a replacement sensor.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Firebag	Station Number:	AMS 19
Calibration Date:	May 4, 2023	Prev Cal Date:	May 1, 2023
Start Time (MST):	13:22	End Time (MST):	14:40
Tower Height (m):	10.0	Reason:	Removal

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	W15276
WS Calibrator:	MetOne 053	Serial Number:	CA 03845

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0		---
200	20.2		
400	39.4		
600	58.6		
800	77.8		

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope			0.90 - 1.10
Calculated intercept			+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	U11347
As Found Declination (deg east of True North):	14	As Left Declination (deg east of True North):	15
Solar noon time (MST):	12:20	Calc Declination*:	15.44 Degrees
Deadband calc:	5.2 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
2	3.2	---
90	89.3	-0.2%
180	178.3	-0.5%
270	267.5	-0.7%
357	355.0	-0.6%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope		1.009250	0.90 - 1.10
Calculated intercept		-0.503217	+/- 4

Notes: Deadband out of compliance from the calibration done on May 1, 2023. As founds done before removing the WD sensor.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Firebag	Station Number:	AMS 19
Calibration Date:	May 4, 2023	Prev Cal Date:	May 1, 2023
Start Time (MST):	13:40	End Time (MST):	15:12
Tower Height (m):	10.0	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	W15276
WS Calibrator:	MetOne 053	Serial Number:	CA 03845

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0		---
200	20.2		
400	39.4		
600	58.6		
800	77.8		

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope			0.90 - 1.10
Calculated intercept			+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	A23968
As Found Declination (deg east of True North):	14	As Left Declination (deg east of True North):	15
Solar noon time (MST):	12:20	Calc Declination*:	15.44 Degrees
Deadband calc:	1.6 degrees (<i>Limit 4 deg</i>)		<small>* - calculated declination as per NOAA website</small>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
2	4.0	---
90	88.8	-0.3%
180	181.7	0.5%
270	272.1	0.6%
357	359.4	0.7%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope		0.995367	0.90 - 1.10
Calculated intercept		-0.576136	+/- 4

Notes: Deadband out of compliance from the calibration done on May 1, 2023. New WD sensor installed. Aligned tower using solar noon.

Calibration Performed By: Braiden Boutillier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS20 MACKAY RIVER MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Mackay River	Station number:	AMS20
Calibration Date:	May 2, 2023	Last Cal Date:	April 3, 2023
Start time (MST):	9:55	End time (MST):	12:35
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	<u>49.22</u> ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC306868</u>		
Removed Cal Gas Conc:	<u>49.22</u> ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	1220
ZAG Make/Model:	Teledyne API 701	Serial Number:	4522

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005548	1.007790	Backgd or Offset:	18.9	18.8
Calibration intercept:	0.790939	2.430853	Coeff or Slope:	0.974	0.974

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4919	81.3	800.3	805.6	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4919	81.3	800.3	808.2	0.990
second point	4959	40.7	400.7	406.4	0.986
third point	4980	20.3	199.8	206.4	0.968
as left zero	5000	0.0	0.0	0.3	----
as left span	4919	81.3	800.3	814.5	0.983
Average Correction Factor					0.981

Baseline Corr As found:	805.50	Previous response	805.50	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments required.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 2, 2023	Previous Calibration:	April 3, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	9:55	End Time (MST):	12:35
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999968	
800.3	808.2	0.9902			≥0.995
400.7	406.4	0.9859	Slope	1.007790	
199.8	206.4	0.9681			0.90 - 1.10
			Intercept	2.430853	+/-30



SO2 Calibration Plot

Date: May 2, 2023

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: MacKay River Station number: AMS20
 Calibration Date: May 24, 2023 Last Cal Date: April 4, 2023
 Start time (MST): 9:56 End time (MST): 13:27
 Reason: Removal and cal gas cylinder change

Calibration Standards

Cal Gas Concentration: 5.12 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC515997
 Removed Cal Gas Conc: 4.87 ppm Rem Gas Exp Date: May 5, 2023
 Removed Gas Cyl #: EY0001922 Diff between cyl: -2.5%
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1220
 ZAG Make/Model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Teledyne API T101 Analyzer serial #: 196
 Converter make: Internal Converter serial #: NA
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003004	NA	Backgd or Offset: 46.3	NA
Calibration intercept:	0.079047	NA	Coeff or Slope: 0.987	NA

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.6	----
as found span	4918	82.1	80.0	82.1	0.981
as found 2nd point	4959	41.1	40.0	41.5	0.979
as found 3rd point	4979	20.5	20.0	21.1	0.974
new cylinder response	4922	78.1	80.0	80.1	0.999

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
SO2 Scrubber Check	4919	80.0	800.2		----
Date of last scrubber change:	December 15, 2020			Ave Corr Factor	
Date of last converter efficiency test:					efficiency
Baseline Corr As found:	81.5	Prev response:	80.28	*% change:	1.5%
Baseline Corr 2nd AF pt:	40.9	AF Slope:	1.018723	AF Intercept:	0.679021
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999996		

* = > +/-5% change initiates investigation


Notes: Removal calibration for instrument change out. Scrubber check completed after third point, it passed; and then changed calibration gas cylinder

Calibration Performed By: Mohammed Kashif

H₂S Calibration Plot

Date: May 24, 2023

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	MacKay River	Station number:	AMS20
Calibration Date:	May 25, 2023	Last Cal Date:	NA
Start time (MST):	10:33	End time (MST):	13:25
Reason:	Install		

Calibration Standards

Cal Gas Concentration:	5.12	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC515997			
Removed Cal Gas Conc:	5.12	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1220
ZAG Make/Model:	Teledyne API 701		Serial Number:	4522

Analyzer Information

Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	12124313139
Converter make:	Global	Converter serial #:	2022-226
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.006091	Backgd or Offset:	NA	1.92
Calibration intercept:	NA	-0.292887	Coeff or Slope:	NA	0.619

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4922	78.1	80.0	80.3	0.997
second point	4961	39.0	40.0	39.8	1.004
third point	4980	19.5	20.0	19.8	1.011
as left zero	5000	0.0	0.0	-0.5	----
as left span	4922	78.1	80.0	79.9	1.002
SO2 Scrubber Check	4919	80.0	800.2	-0.1	----
Date of last scrubber change:	May 25, 2023			Ave Corr Factor	1.004
Date of last converter efficiency test:	efficiency				

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: Install calibration. Scrubber check completed after calibrator zero. Adjusted zero and span.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 25, 2023	Previous Calibration:	NA
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	10:33	End Time (MST):	13:25
Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	12124313139

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999992	≥0.995
80.0	80.3	0.9967			
40.0	39.8	1.0042	Slope	1.006091	0.90 - 1.10
20.0	19.8	1.0109			
			Intercept	-0.292887	+/-3



H₂S Calibration Plot

Date: May 25, 2023

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 2, 2023	Previous Calibration:	April 3, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	9:55	End Time (MST):	12:35
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	-0.14	----	Correlation Coefficient	0.999963	≥0.995
17.34	17.31	1.0018			
8.68	8.63	1.0064	Slope	1.004428	0.90 - 1.10
4.33	4.32	1.0029			
			Intercept	-0.093194	+/-1.5



THC Calibration Plot

Date: May 2, 2023

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: MacKay River Station number: AMS20
Calibration Date: May 11, 2023 Last Cal Date: April 19, 2023
Start time (MST): 9:15 End time (MST): 13:21
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 1220
ZAG make/model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.459	1.459	NO bkgnd or offset:	4.0	4.0
NOX coeff or slope:	0.990	0.990	NOX bkgnd or offset:	4.0	4.0
NO2 coeff or slope:	0.995	0.995	Reaction cell Press:	181.9	180.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996899	0.981506
NO _x Cal Offset:	1.929986	1.850414
NO Cal Slope:	0.999776	0.983542
NO Cal Offset:	0.730780	0.691620
NO ₂ Cal Slope:	0.995110	1.000887
NO ₂ Cal Offset:	-1.870675	-0.270425



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.1	----	----
as found span	4917	83.3	819.5	800.3	19.2	807.7	788.3	19.4	1.0146	1.0152
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4917	83.3	819.5	800.3	19.2	805.5	787.7	17.9	1.0173	1.0160
second point	4956	41.7	410.4	400.8	9.6	404.9	394.6	10.3	1.0137	1.0158
third point	4979	20.8	204.6	199.9	4.8	204.9	198.4	6.5	0.9987	1.0073
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4917	83.3	819.5	447.1	372.4	809.1	434.1	375.0	1.0128	1.0299
Average Correction Factor									1.0099	1.0130

Corrected As found	NO _x = 807.8 ppb	NO = 788.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.4%	
Previous Response	NO _x = 818.8 ppb	NO = 800.8 ppb		*Percent Change	NO = -1.6%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	785.6	432.4	372.4	372.6	0.9994	100.1%
2nd GPT point (200 ppb O ₃)	785.6	603.4	201.4	201.2	1.0008	99.9%
3rd GPT point (100 ppb O ₃)	785.6	692.5	112.3	111.6	1.0059	99.4%
Average Correction Factor					1.0020	99.8%

Notes:

No adjustments made.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 19, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	9:15	End Time (MST):	13:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
819.5	805.5	1.0173			
410.4	404.9	1.0137			
204.6	204.9	0.9987			
			Slope	0.981506	0.90 - 1.10
			Intercept	1.850414	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 19, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	9:15	End Time (MST):	13:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	787.7	1.0160		
400.8	394.6	1.0158		
199.9	198.4	1.0073		





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 19, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	9:15	End Time (MST):	13:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.0	0.1	----	Correlation Coefficient	0.999994	≥0.995			
372.4	372.6	0.9994						
201.4	201.2	1.0008				Slope	1.000887	0.90 - 1.10
112.3	111.6	1.0059						
			Intercept	-0.270425	+/-20			



NO_x Calibration Plot

Date: May 11, 2023

Location: MacKay River





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS21
CONKLIN
MAY 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 3, 2023	Previous Calibration:	April 13, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:41	End Time (MST):	12:41
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999991	
800.8	814.0	0.9838			≥0.995
400.4	408.0	0.9814	Slope	1.015485	
200.1	205.7	0.9729			0.90 - 1.10
			Intercept	1.176454	+/-30



SO2 Calibration Plot

Date: May 3, 2023

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Conklin Station number: AMS21
 Calibration Date: May 30, 2023 Last Cal Date: April 26, 2023
 Start time (MST): 9:32 End time (MST): 14:00
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.00 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC501204
 Removed Cal Gas Conc: 5.03 ppm Rem Gas Exp Date: April 16, 2022
 Removed Gas Cyl #: CC505493 Diff between cyl: -0.8%
 Calibrator Make/Model: API T700 Serial Number: 3810
 ZAG Make/Model: API 701H Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1236656116
 Converter make: CD-Nova 101 Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009291	1.000000	Backgd or Offset: 2.8	2.9
Calibration intercept:	-0.082550	0.000000	Coeff or Slope: 0.974	0.974

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	4921	79.5	80.0	81.3	0.978
as found 2nd point	4960	39.8	40.0	41.0	0.966
as found 3rd point	4980	19.9	20.0	20.2	0.970
new cylinder response	4920	80.0	80.0	80.7	0.991

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.0	80.0	80.0	1.000
second point	4960	40.0	40.0	40.0	1.000
third point	4980	20.0	20.0	20.0	1.000
as left zero	5000	0.0	0.0	0.0	----
as left span	4920	80.0	80.0	79.9	1.001
SO2 Scrubber Check	4920	80.2	802.0	-0.3	----

Date of last scrubber change:	Ave Corr Factor	1.000
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 81.7 Prev response: 80.63 *% change: 1.3%
 Baseline Corr 2nd AF pt: 41.4 AF Slope: 1.022026 AF Intercept: -0.260878
 Baseline Corr 3rd AF pt: 20.6 AF Correlation: 0.999953

* = > +/-5% change initiates investigation

Notes: No adjustments made. Replaced H2S gas.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	April 26, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:32	End Time (MST):	14:00
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits			
0.0	0.0	----	Correlation Coefficient	1.000000	≥ 0.995			
80.0	80.0	1.0000						
40.0	40.0	1.0000				Slope	1.000000	0.90 - 1.10
20.0	20.0	1.0000						
			Intercept	0.000000	+/-3			



TRS Calibration Plot

Date: May 30, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	May 3, 2023	Last Cal Date:	April 18, 2023
Start time (MST):	9:41	End time (MST):	12:41
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.16E-04	2.22E-04	NMHC SP Ratio:	4.66E-05
CH ₄ Retention time:	12.00	12.00	NMHC Peak Area:	196084
				189107

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	16.59	1.033
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.13	1.000
second point	4960	40.1	8.56	8.59	0.997
third point	4980	20.0	4.27	4.33	0.987
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.15	0.999
Average Correction Factor					0.994
Baseline Corr AF:	16.59	Prev response	17.16	*% change	-3.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	8.82	1.036
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.14	1.000
second point	4960	40.1	4.57	4.59	0.995
third point	4980	20.0	2.28	2.32	0.984
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.15	0.999
Average Correction Factor					0.993
Baseline Corr AF:	8.82	Prev response	9.16	*% change	-3.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.77	1.028
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.99	1.000
second point	4960	40.1	3.99	3.99	1.000
third point	4980	20.0	1.99	2.01	0.989
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.99	0.999
Average Correction Factor					0.996
Baseline Corr AF:	7.77	Prev response	8.00	*% change	-3.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999574	0.999175
THC Cal Offset:	0.045989	0.027981
CH ₄ Cal Slope:	0.999808	0.999308
CH ₄ Cal Offset:	0.019754	0.008750
NMHC Cal Slope:	0.999495	0.998771
NMHC Cal Offset:	0.026235	0.019631

Notes: Replaced N2 cylinder after the as founds. Adjusted the span only.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	May 3, 2023	Previous Calibration:	April 18, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:41	End Time (MST):	12:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999987	≥ 0.995			
17.13	17.13	0.9998						
8.56	8.59	0.9969				Slope	0.999175	0.90 - 1.10
4.27	4.33	0.9866						
			Intercept	0.027981	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	May 3, 2023	Previous Calibration:	April 18, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:41	End Time (MST):	12:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999991	≥0.995			
7.99	7.99	0.9997						
3.99	3.99	0.9998				Slope	0.999308	0.90 - 1.10
1.99	2.01	0.9889						
			Intercept	0.008750	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	May 3, 2023	Previous Calibration:	April 18, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:41	End Time (MST):	12:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999979	≥ 0.995
9.14	9.14	1.0001			
4.57	4.59	0.9949			
2.28	2.32	0.9841			
			Slope	0.998771	0.90 - 1.10
			Intercept	0.019631	+/-0.5



NMHC Calibration Plot

Date: May 3, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	May 24, 2023	Last Cal Date:	May 3, 2023
Start time (MST):	10:17	End time (MST):	11:55
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.22E-04	2.22E-04	NMHC SP Ratio:	4.83E-05
CH ₄ Retention time:	12.00	12.00	NMHC Peak Area:	189107
				189107

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	16.91	1.013
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	16.85	1.016
Average Correction Factor					
Baseline Corr AF:	16.91	Prev response	17.14	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	8.99	1.017
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	8.94	1.022
Average Correction Factor					
Baseline Corr AF:	8.99	Prev response	9.15	*% change	-1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.92	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.91	1.009
Average Correction Factor					
Baseline Corr AF:	7.92	Prev response	7.99	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999175	
THC Cal Offset:	0.027981	
CH ₄ Cal Slope:	0.999308	
CH ₄ Cal Offset:	0.008750	
NMHC Cal Slope:	0.998771	
NMHC Cal Offset:	0.019631	


Notes: Changed H2 cylinder.

Calibration Performed By: Denny Ray Estador

NMHC Calibration Plot

Date: May 24, 2023

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Conklin
Calibration Date: May 16, 2023
Start time (MST): 9:41
Reason: Routine
Station number: AMS21
Last Cal Date: April 28, 2023
End time (MST): 13:35

Calibration Standards

NO Gas Cylinder #: T2Y1P1H
NOX Cal Gas Conc: 51.09 ppm
Removed Cylinder #: n/a
Removed Gas NOX Conc: 51.09 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.39 ppm
Removed Gas Exp Date: n/a
Removed Gas NO Conc: 50.39 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 691

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1501663731

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.144	1.144	NO bkgnd or offset:	11.6	11.6
NOX coeff or slope:	1.001	1.001	NOX bkgnd or offset:	11.8	11.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	219.5	220.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.010616	1.001335
NO _x Cal Offset:	2.246244	2.164394
NO Cal Slope:	1.008763	1.000766
NO Cal Offset:	1.982980	1.181398
NO ₂ Cal Slope:	1.003351	1.002657
NO ₂ Cal Offset:	-0.204066	-0.146701



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.2	0.1	----	----
as found span	4921	79.4	811.2	800.1	11.1	813.3	800.0	13.4	0.9975	1.0001
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
high point	4921	79.4	811.2	800.1	11.1	813.5	801.3	12.1	0.9972	0.9985
second point	4960	39.7	405.7	400.1	5.6	409.4	402.3	7.1	0.9909	0.9946
third point	4980	19.8	202.3	199.6	2.8	206.7	202.0	4.7	0.9788	0.9879
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4921	79.4	811.2	379.2	432.0	814.0	385.2	428.9	0.9966	0.9845
Average Correction Factor									0.9890	0.9937

Corrected As found	NO _x = 813.3 ppb	NO = 800.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.1%
Previous Response	NO _x = 822.1 ppb	NO = 809.1 ppb		*Percent Change	NO = -1.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI: ;

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.5	378.6	432.0	432.9	0.9980	100.2%
2nd GPT point (200 ppb O3)	799.5	590.9	219.7	220.9	0.9946	100.5%
3rd GPT point (100 ppb O3)	799.5	694.0	116.6	115.8	1.0070	99.3%
Average Correction Factor					0.9999	100.0%

Notes:

No adjustments required.

Calibration Performed By:

Denny Ray Estador



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	April 28, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:41	End Time (MST):	13:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
811.2	813.5	0.9972		
405.7	409.4	0.9909		
202.3	206.7	0.9788		
			0.999973	
			1.001335	
			2.164394	





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	April 28, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:41	End Time (MST):	13:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
800.1	801.3	0.9985			0.999990
400.1	402.3	0.9946			1.000766
199.6	202.0	0.9879			1.181398





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	April 28, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:41	End Time (MST):	13:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
432.0	432.9	0.9980		
219.7	220.9	0.9946		
116.6	115.8	1.0070		



NO_x Calibration Plot

Date: May 16, 2023

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 9, 2023	Previous Calibration:	April 14, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:02	End Time (MST):	12:05
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥0.995
400.0	401.0	0.9975		
200.0	201.9	0.9906	Slope	0.90 - 1.10
100.0	101.4	0.9862		
			Intercept	+/- 5



O₃ Calibration Plot

Date: May 9, 2023

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Conklin Station number: AMS 21
 Calibration Date: May 24, 2023 Last Cal Date: May 23, 2023
 Start time (MST): 11:25 End time (MST): 11:40

Analyzer Make: API T640X S/N: 1597
 Particulate Fraction: PM2.5

Flow Meter Make/Model: DeltaCal S/N: 954
 Temp/RH standard: DeltaCal S/N: 954

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	15.9	15.8	15.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	707.4	708	707.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5	5.03	5	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>May 24, 2023</u>	Last Cal Date: <u>May 23, 2023</u>			
	PM w/o HEPA: <u>20.1</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.9	NA	10.9	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>20.1</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>NA</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>NA</u>			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: No adjustments made. Inspected inlet head; relatively clean.

Calibration by: Denny Ray Estador



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER
MAY 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 26, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:24	End Time (MST):	13:45
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999995	≥0.995
799.8	800.5	0.9991			
399.9	402.0	0.9947	Slope	1.000578	0.90 - 1.10
200.4	201.0	0.9972			
			Intercept	0.744403	+/-30



SO2 Calibration Plot

Date: May 17, 2023

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Janvier Station number: AMS22
 Calibration Date: May 24, 2023 Last Cal Date: April 25, 2023
 Start time (MST): 10:30 End time (MST): 15:18
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.03 ppm Cal Gas Exp Date: April 16, 2022
 Cal Gas Cylinder #: DT0018680
 Removed Cal Gas Conc: 5.03 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680031
 Converter make: CDN-101 Converter serial #: 587
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999363	0.997220	Backgd or Offset: 5.95	6.33
Calibration intercept:	-0.038984	0.161030	Coeff or Slope: 1.071	1.022

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.7	----
as found span	4920	79.5	80.0	84.1	0.959
as found 2nd point	4960	39.8	40.0	42.9	0.949
as found 3rd point	4980	19.9	20.0	21.8	0.949
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	79.5	80.0	79.9	1.001
second point	4960	39.8	40.0	40.1	0.999
third point	4980	19.9	20.0	20.2	0.991
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	79.5	80.0	79.8	1.002
SO2 Scrubber Check	4920	79.8	798.0	0.0	----

Date of last scrubber change:	Ave Corr Factor	0.997
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 83.4 Prev response: 79.90 *% change: 4.2%
 Baseline Corr 2nd AF pt: 42.2 AF Slope: 1.042376 AF Intercept: 0.880046
 Baseline Corr 3rd AF pt: 21.1 AF Correlation: 0.999963

* = > +/-5% change initiates investigation

Notes: Changed out the inlet filter after as founds. Scrubber check passed. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 24, 2023	Previous Calibration:	April 25, 2023
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	10:30	End Time (MST):	15:18
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999997	
80.0	79.9	1.0011			≥0.995
40.0	40.1	0.9985	Slope	0.997220	
20.0	20.2	0.9911			0.90 - 1.10
			Intercept	0.161030	+/-3



TRS Calibration Plot

Date: May 24, 2023

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	May 17, 2023	Last Cal Date:	April 26, 2023
Start time (MST):	10:24	End time (MST):	13:45
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH ₄ Cal Gas Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
C ₃ H ₈ Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
Removed C ₃ H ₈ Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.150E-04	2.160E-04	NMHC SP Ratio:	4.45E-05
CH ₄ Retention time:	13.20	13.40	NMHC Peak Area:	205602
				204301

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	17.38	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.19	0.999
second point	4960	39.9	8.59	8.59	1.000
third point	4980	20.0	4.30	4.28	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.19	0.999

Average Correction Factor				1.002
Baseline Corr AF:	17.38	Prev response	17.19	*% change 1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4920	79.8	9.15	9.20	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	79.8	9.15	9.14	1.001
second point	4960	39.9	4.57	4.57	1.002
third point	4980	20.0	2.29	2.27	1.012
as left zero	5000	0	0.00	0.00	----
as left span	4920	79.8	9.15	9.09	1.007
Average Correction Factor					1.005
Baseline Corr AF:	9.20	Prev response	9.15	*% change	0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	8.18	0.982
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.06	0.996
second point	4960	39.9	4.01	4.02	0.998
third point	4980	20.0	2.01	2.01	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.10	0.991
Average Correction Factor					0.998
Baseline Corr AF:	8.18	Prev response	8.05	*% change	1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002618	1.001793
THC Cal Offset:	-0.031798	-0.014598
CH ₄ Cal Slope:	1.004358	1.004345
CH ₄ Cal Offset:	-0.017966	-0.005170
NMHC Cal Slope:	1.001328	0.999528
NMHC Cal Offset:	-0.014032	-0.009829

Notes: Changed the inlet filter and N₂/H₂ cylinders after as founds. Adjusted span only.

Calibration Performed By:

Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 26, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:24	End Time (MST):	13:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
17.17	17.19	0.9988						
8.59	8.59	0.9999				Slope	1.001793	0.90 - 1.10
4.30	4.28	1.0062						
			Intercept	-0.014598	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 26, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:24	End Time (MST):	13:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥0.995
8.03	8.06	0.9959			
4.01	4.02	0.9981			
2.01	2.01	1.0001			
			Slope	1.004345	0.90 - 1.10
			Intercept	-0.005170	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 26, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:24	End Time (MST):	13:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	≥ 0.995			
9.15	9.14	1.0013						
4.57	4.57	1.0018				Slope	0.999528	0.90 - 1.10
2.29	2.27	1.0117						
			Intercept	-0.009829	± 0.5			



NMHC Calibration Plot

Date: May 17, 2023

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: May 2, 2023 Last Cal Date: April 27, 2023
Start time (MST): 10:28 End time (MST): 16:02
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC424183 Cal Gas Expiry Date: April 16, 2023
NOX Cal Gas Conc: 48.60 ppm NO Cal Gas Conc: 48.60 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.60 ppm Removed Gas NO Conc: 48.60 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3806
ZAG make/model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Teledyne API T200 Analyzer serial #: 7117
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.012	1.013	NO bkgnd or offset:	4.8	0.6
NOX coeff or slope:	1.002	1.002	NOX bkgnd or offset:	5.9	2.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	5.1	5.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999487	0.999801
NO _x Cal Offset:	0.688544	0.228682
NO Cal Slope:	0.999800	0.999815
NO Cal Offset:	-0.271133	-0.151058
NO ₂ Cal Slope:	1.003470	1.001797
NO ₂ Cal Offset:	0.116060	0.161197



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	2.3	-2.4	4.7	----	----
as found span	4918	82.3	799.9	799.9	0.0	799.9	794.1	5.9	1.0000	1.0073
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4918	82.3	799.9	799.9	0.0	799.8	799.7	0.1	1.0001	1.0003
second point	4959	41.2	400.4	400.4	0.0	400.8	400.0	0.8	0.9991	1.0011
third point	4980	20.6	200.2	200.2	0.0	200.7	200.1	0.7	0.9975	1.0005
as left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.2	0.2	----	----
as left span	4918	82.3	799.9	399.8	400.1	797.4	397.7	399.7	1.0031	1.0053
Average Correction Factor									0.9989	1.0006

Corrected As found	NO _x = 797.6 ppb	NO = 796.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.3%
Previous Response	NO _x = 800.2 ppb	NO = 799.5 ppb		*Percent Change	NO = -0.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.2	396.1	400.1	400.9	0.9980	100.2%
2nd GPT point (200 ppb O3)	796.2	597.2	199.0	199.6	0.9970	100.3%
3rd GPT point (100 ppb O3)	796.2	696.8	99.4	99.9	0.9950	100.5%
Average Correction Factor					0.9967	100.3%

Notes: Changed the inlet filter after as founds. Changed out the ZAG charcoal and purafil scrubbers. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 2, 2023	Previous Calibration:	April 27, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:28	End Time (MST):	16:02
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.9	799.8	1.0001		
400.4	400.8	0.9991		
200.2	200.7	0.9975		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 2, 2023	Previous Calibration:	April 27, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:28	End Time (MST):	16:02
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
799.9	799.7	1.0003			
400.4	400.0	1.0011			
200.2	200.1	1.0005			
			Slope	0.999815	0.90 - 1.10
			Intercept	-0.151058	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 2, 2023	Previous Calibration:	April 27, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:28	End Time (MST):	16:02
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
400.1	400.9	0.9980		
199.0	199.6	0.9970		
99.4	99.9	0.9950		
			0.999999	
			1.001797	
			0.161197	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 2, 2023

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: May 3, 2023 Last Cal Date: NA
Start time (MST): 10:09 End time (MST): 15:42
Reason: Install

Calibration Standards

NO Gas Cylinder #: CC424183 Cal Gas Expiry Date: April 16, 2023
NOX Cal Gas Conc: 48.60 ppm NO Cal Gas Conc: 48.60 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.60 ppm Removed Gas NO Conc: 48.60 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3806
ZAG make/model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1218153460
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:		1.303	NO bkgnd or offset:		3.4
NOX coeff or slope:		0.988	NOX bkgnd or offset:		4.3
NO2 coeff or slope:		1.000	Reaction cell Press:		189.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:		0.999560
NO _x Cal Offset:		1.487988
NO Cal Slope:		0.999544
NO Cal Offset:		0.468701
NO ₂ Cal Slope:		0.998379
NO ₂ Cal Offset:		0.058523



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
high point	4918	82.3	799.9	799.9	0.0	800.1	799.8	0.2	0.9998	1.0001
second point	4959	41.2	400.4	400.4	0.0	403.0	400.9	2.1	0.9937	0.9989
third point	4980	20.6	200.2	200.2	0.0	202.9	201.1	1.8	0.9867	0.9956
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as left span	4918	82.3	799.9	397.1	402.8	803.0	399.0	404.1	0.9961	0.9953
Average Correction Factor									0.9934	0.9982

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* => +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.3	396.5	402.8	402.2	1.0015	99.9%
2nd GPT point (200 ppb O3)	799.3	600.1	199.2	198.8	1.0020	99.8%
3rd GPT point (100 ppb O3)	799.3	701.8	97.5	97.7	0.9980	100.2%
Average Correction Factor					1.0005	100.0%

Notes: Installing a new NO_x analyzer due to unstable zeros. Inlet filter was changed yesterday. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 3, 2023	Previous Calibration:	NA
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:09	End Time (MST):	15:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
799.9	800.1	0.9998			
400.4	403.0	0.9937			
200.2	202.9	0.9867			
			Slope	0.999560	0.90 - 1.10
			Intercept	1.487988	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 3, 2023	Previous Calibration:	NA
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:09	End Time (MST):	15:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.9	799.8	1.0001		
400.4	400.9	0.9989		
200.2	201.1	0.9956		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 3, 2023	Previous Calibration:	NA
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:09	End Time (MST):	15:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
402.8	402.2	1.0015		
199.2	198.8	1.0020		
97.5	97.7	0.9980		
			0.999998	
			0.998379	
			0.058523	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 3, 2023

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: May 30, 2023 Last Cal Date: April 26, 2023
 Start time (MST): 10:08 End time (MST): 13:23
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701 Serial Number: 201

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3869
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007857	1.006657	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	0.800000	0.560000	Coeff or Slope:	1.011	1.011

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	-0.1	----
as found span	4893	899.1	400.0	402.9	0.993
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	-0.1	----
high point	4893	899.1	400.0	402.8	0.993
second point	4893	753.4	200.0	202.5	0.988
third point	4893	655.7	100.0	101.7	0.983
as left zero	5000	800.0	0.0	0.2	----
as left span	4816	899.1	400.0	406.9	0.983
Average Correction Factor					0.988

Baseline Corr As found:	403.0	Previous response	403.9	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	April 26, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:08	End Time (MST):	13:23
Analyzer make:	Teledyne API T400	Analyzer serial #:	3869

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999986	≥0.995
400.0	402.8	0.9930			
200.0	202.5	0.9877	Slope	1.006657	0.90 - 1.10
100.0	101.7	0.9833			
			Intercept	0.560000	+/- 5



O₃ Calibration Plot

Date: May 30, 2023

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: May 30, 2023 Last Cal Date: April 27, 2023
 Start time (MST): 11:15 End time (MST): 13:28

Analyzer Make: Teledyne API T640 S/N: 325
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388750
 Temp/RH standard: Alicat FP-25 S/N: 388750

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	23.8	23.62	23.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	712.5	713.29	712.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	5.080	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>May 30, 2023</u>	Last Cal Date: <u>April 27, 2023</u>			
	PM w/o HEPA: <u>12.6</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>April 27, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>April 27, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: October 6, 2022
 Date RH/T Sensor Cleaned: October 6, 2022

Notes: Verified flow, temperature, and pressure. No adjustments made. Leak test passed.

Calibration by: Braiden Boutilier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23 FORT HILLS MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 10, 2023	Previous Calibration:	April 14, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:24	End Time (MST):	12:35
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.2	----	Correlation Coefficient	0.999999	≥0.995
799.1	800.0	0.9989			
400.1	399.8	1.0006	Slope	1.001510	0.90 - 1.10
200.0	199.6	1.0022			
			Intercept	-0.523913	+/-30



SO2 Calibration Plot

Date: May 10, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Hills Station number: AMS23
 Calibration Date: May 25, 2023 Last Cal Date: April 5, 2023
 Start time (MST): 9:24 End time (MST): 13:29
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.20 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517372
 Removed Cal Gas Conc: 5.20 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 451
 ZAG Make/Model: API T701 Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 43iQ TLE Analyzer serial #: 1300156232
 Converter make: CDN-101 Converter serial #: 594
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003023	1.005029	Backgd or Offset:	1.92	1.92
Calibration intercept:	0.182125	-0.038092	Coeff or Slope:	1.132	1.132

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4923	77.0	80.0	79.3	1.008
as found 2nd point	4962	38.5	40.0	39.7	1.005
as found 3rd point	4981	19.2	19.9	19.7	1.007
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.0	80.0	80.3	0.996
second point	4962	38.5	40.0	40.4	0.990
third point	4981	19.2	19.9	19.8	1.007
as left zero	5000	0.0	0.0	0.7	----
as left span	4923	77.0	80.0	80.4	0.995
SO2 Scrubber Check	4920	80.3	803.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.998
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 79.4 Prev response: 80.43 *% change: -1.3%
 Baseline Corr 2nd AF pt: 39.8 AF Slope: 0.992601 AF Intercept: -0.078272
 Baseline Corr 3rd AF pt: 19.8 AF Correlation: 0.999998

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021


Station Information

Calibration Date:	May 25, 2023	Previous Calibration:	April 5, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:24	End Time (MST):	13:29
Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	1300156232

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999970	
80.0	80.3	0.9963			≥0.995
40.0	40.4	0.9900	Slope	1.005029	
19.9	19.8	1.0075			0.90 - 1.10
			Intercept	-0.038092	+/-3







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	May 10, 2023	Last Cal Date:	April 14, 2023
Start time (MST):	9:24	End time (MST):	12:35
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
Removed C ₃ H ₈ Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.32E-04	2.32E-04	NMHC SP Ratio:	5.06E-05
CH ₄ Retention time:	13.0	13.0	NMHC Peak Area:	181940

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	17.21	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.19	1.000
second point	4960	40.2	8.61	8.56	1.005
third point	4980	20.1	4.30	4.29	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.19	17.15	1.003

Average Correction Factor				1.003
Baseline Corr AF:	17.21	Prev response	17.26	*% change -0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.15	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.14	1.003
second point	4960	40.2	4.59	4.60	0.997
third point	4980	20.1	2.29	2.32	0.990
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	9.16	9.12	1.004
Average Correction Factor					0.997
Baseline Corr AF:	9.15	Prev response	9.20	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	8.07	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.05	0.997
second point	4960	40.2	4.02	3.96	1.015
third point	4980	20.1	2.01	1.98	1.017
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.03	8.02	1.001
Average Correction Factor					1.010
Baseline Corr AF:	8.07	Prev response	8.06	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002839	0.999810
THC Cal Offset:	0.020787	-0.013170
CH ₄ Cal Slope:	1.005380	1.003198
CH ₄ Cal Offset:	-0.017654	-0.029231
NMHC Cal Slope:	1.000647	0.996713
NMHC Cal Offset:	0.038041	0.016062

Notes: Changed the inlet filter and the N2 cylinder after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 10, 2023	Previous Calibration:	April 14, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:24	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	≥ 0.995			
17.19	17.19	1.0001						
8.61	8.56	1.0052				Slope	0.999810	0.90 - 1.10
4.30	4.29	1.0027						
			Intercept	-0.013170	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 10, 2023	Previous Calibration:	April 14, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:24	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999907	≥ 0.995
8.03	8.05	0.9975			
4.02	3.96	1.0147			
2.01	1.98	1.0171			
			Slope	1.003198	0.90 - 1.10
			Intercept	-0.029231	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 10, 2023	Previous Calibration:	April 14, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:24	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999985	≥ 0.995
9.16	9.14	1.0025			
4.59	4.60	0.9973			
2.29	2.32	0.9904			
			Slope	0.996713	0.90 - 1.10
			Intercept	0.016062	+/-0.5



NMHC Calibration Plot

Date: May 10, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Hills
Calibration Date: May 16, 2023
Start time (MST): 9:08
Reason: Routine
Station number: AMS23
Last Cal Date: April 21, 2023
End time (MST): 13:55

Calibration Standards

NO Gas Cylinder #: CC332703
NOX Cal Gas Conc: 49.7 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 49.7 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T750
ZAG make/model: Teledyne API T751H
Cal Gas Expiry Date: January 28, 2024
NO Cal Gas Conc: 49.7 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 49.7 ppm
NO gas Diff:
Serial Number: 275
Serial Number: 307

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1152430007

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.065	1.049	NO bkgnd or offset:	2.9	2.9
NOX coeff or slope:	0.995	0.995	NOX bkgnd or offset:	3.2	3.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	158.3	160.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003237	0.999094
NO _x Cal Offset:	0.084507	1.084161
NO Cal Slope:	1.004153	1.000637
NO Cal Offset:	-1.236008	-0.356041
NO ₂ Cal Slope:	1.006445	1.001549
NO ₂ Cal Offset:	0.589439	-0.718730



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.7	-0.1	0.8	----	----
as found span	4920	80.5	800.2	800.2	0.0	813.3	810.3	3.0	0.984	0.987
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.6	0.0	0.5	----	----
high point	4920	80.5	800.2	800.2	0.0	799.8	800.1	-0.3	1.000	1.000
second point	4960	40.2	399.6	399.6	0.0	402.0	400.5	1.5	0.994	0.998
third point	4980	20.1	199.8	199.8	0.0	200.2	198.4	1.8	0.998	1.007
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4920	80.5	800.2	441.3	358.9	797.1	439.2	357.9	1.004	1.005
Average Correction Factor									0.997	1.002

Corrected As found	NO _x = 812.6 ppb	NO = 810.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.2%
Previous Response	NO _x = 802.8 ppb	NO = 802.3 ppb		*Percent Change	NO = 1.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.6	440.7	358.9	359.5	0.998	100.2%
2nd GPT point (200 ppb O3)	799.6	621.2	178.4	176.9	1.008	99.2%
3rd GPT point (100 ppb O3)	799.6	707.5	92.1	90.6	1.017	98.4%
Average Correction Factor					1.008	99.2%

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	April 21, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:08	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.2	799.8	1.0005		
399.6	402.0	0.9940		
199.8	200.2	0.9980		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	April 21, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:08	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 $0.90 - 1.10$ ± 20
800.2	800.1	1.0001		
399.6	400.5	0.9977		
199.8	198.4	1.0070		
			0.999993	
			1.000637	
			-0.356041	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	April 21, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:08	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits			
0.0	0.5	----	Correlation Coefficient	0.999943	≥0.995			
358.9	359.5	0.9983						
178.4	176.9	1.0085				Slope	1.001549	0.90 - 1.10
92.1	90.6	1.0166						
			Intercept	-0.718730	+/-20			



NO_x Calibration Plot

Date: May 16, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Hills Station number: AMS 23
 Calibration Date: May 10, 2023 Last Cal Date: April 21, 2023
 Start time (MST): 12:41 End time (MST): 14:00

Analyzer Make: API T640 S/N: 1546
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388755
 Temp/RH standard: Alicat FP-25BT S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	24.2	23.27	24.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	735.06	733.4	735.06	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.07	5.14	5.07	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: May 10, 2023 Last Cal Date: April 21, 2023
 PM w/o HEPA: 90.1 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.8	11	11	<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: 80.7 w/ HEPA: 0.0
 Date Optical Chamber Cleaned: May 10, 2023 <0.2 ug/m3
 Disposable Filter Changed: May 10, 2023

Annual Maintenance

Date Sample Tube Cleaned: May 10, 2023
 Date RH/T Sensor Cleaned: May 10, 2023

Notes: Leak check passed, no adjustments needed. Completed quarterly and annual maintenance.

Calibration by: Max Farrell



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25 WASKŌW OHCI PIMÂTISIWIN MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 10, 2023	Previous Calibration:	April 11, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	9:44	End Time (MST):	12:20
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999989	
800.5	800.5	1.0000			≥0.995
400.3	402.3	0.9951	Slope	1.000277	
200.1	199.4	1.0037			0.90 - 1.10
			Intercept	0.283925	+/-30



SO2 Calibration Plot

Date: May 10, 2023

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS25
 Calibration Date: May 17, 2023 Last Cal Date: April 12, 2023
 Start time (MST): 6:05 End time (MST): 10:24
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC517099
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 747
 ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
 Converter make: Thermo 43C Converter serial #: 328702539
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.014757	1.021518	Backgd or Offset: 3.4	3.4
Calibration intercept:	0.200000	0.240000	Coeff or Slope: 1.108	1.108

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4918	81.6	81.1	81.2	0.999
as found 2nd point	4959	40.8	40.5	40.7	0.997
as found 3rd point	4980	20.4	20.3	20.5	0.993
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4920	80.0	79.5	81.4	0.976
second point	4960	40.0	39.7	40.8	0.974
third point	4980	20.0	19.9	20.6	0.964
as left zero	5000	0.0	0.0	0.2	----
as left span	4912	88.3	800.0	826.0	0.969
SO2 Scrubber Check	4924	76.3	800.0	0.0	----
Date of last scrubber change:	19-Jul-10			Ave Corr Factor	0.971
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 81.1 Prev response: 82.45 *% change: -1.7%
 Baseline Corr 2nd AF pt: 40.6 AF Slope: 1.000170 AF Intercept: 0.167624
 Baseline Corr 3rd AF pt: 20.4 AF Correlation: 0.999997

* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 12, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	6:05	End Time (MST):	10:24
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999998	≥0.995
79.5	81.4	0.9761			
39.7	40.8	0.9737	Slope	1.021518	0.90 - 1.10
19.9	20.6	0.9643			
			Intercept	0.240000	+/-3



H₂S Calibration Plot

Date: May 17, 2023

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS25
 Calibration Date: May 26, 2023 Last Cal Date: May 17, 2023
 Start time (MST): 5:50 End time (MST): 9:52
 Reason: Maintenance Calibration and SOx scrubber check after Sulphur Study

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC517099
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 747
 ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
 Converter make: Thermo 43C Converter serial #: 328702539
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.021518	1.005984	Backgd or Offset: 3.4	3.3
Calibration intercept:	0.240000	0.180000	Coeff or Slope: 1.108	1.079

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4920	80.0	79.5	81.6	0.976
as found 2nd point	4960	40.0	39.7	41.0	0.974
as found 3rd point	4980	20.0	19.9	20.6	0.974
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4920	80.0	79.5	80.1	0.992
second point	4960	40.0	39.7	40.2	0.988
third point	4980	20.0	19.9	20.1	0.988
as left zero	5000	0.0	0.0	0.2	----
as left span	4912	88.3	800.0	808.9	0.989
SO2 Scrubber Check	4924	76.3	800.0	0.0	----
Date of last scrubber change:	19-Jul-10			Ave Corr Factor	0.989
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	81.4	Prev response:	81.41	*% change:	0.0%
Baseline Corr 2nd AF pt:	40.8	AF Slope:	1.024394	AF Intercept:	0.240000
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999998		

* = > +/-5% change initiates investigation

Notes: Calibration and SOx scrubber check after the Sulphur Study test. Sox scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 26, 2023	Previous Calibration:	May 17, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	5:50	End Time (MST):	9:52
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999998	
79.5	80.1	0.9920			≥0.995
39.7	40.2	0.9883	Slope	1.005984	
19.9	20.1	0.9883			0.90 - 1.10
			Intercept	0.180000	+/-3



H₂S Calibration Plot

Date: May 26, 2023

Location: Waskow ohci Pimatisiwin





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS26 CHRISTINA LAKE MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 9, 2023	Previous Calibration:	April 27, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	10:11	End Time (MST):	12:54
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999969	≥0.995
799.0	811.6	0.9844			
399.4	404.2	0.9882	Slope	1.017133	0.90 - 1.10
200.2	199.3	1.0046			
			Intercept	-1.819322	+/-30



SO2 Calibration Plot

Date: May 9, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Christina Lake	Station number:	AMS26
Calibration Date:	May 30, 2023	Last Cal Date:	April 25, 2023
Start time (MST):	16:15	End time (MST):	20:04
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.89	ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	<u>EY0002466</u>			
Removed Cal Gas Conc:	4.89	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2447
ZAG Make/Model:	API T701H		Serial Number:	832

Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032
Converter make:	NA	Converter serial #:	NA
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005611	1.035031	Backgd or Offset:	35.4	35.4
Calibration intercept:	-0.360862	-0.240256	Coeff or Slope:	1.125	1.125

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4918	81.8	80.0	82.7	0.964
as found 2nd point	4959	40.9	40.0	41.0	0.969
as found 3rd point	4979	20.4	20.0	20.1	0.978
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4918	81.8	80.0	82.6	0.969
second point	4959	40.9	40.0	41.1	0.973
third point	4979	20.4	20.0	20.4	0.978
as left zero	5000	0.0	0.0	-0.1	----
as left span	4918	81.8	80.0	82.5	0.970
SO2 Scrubber Check	4919	80.6	806.1	-0.1	----
Date of last scrubber change:	27-Feb-19			Ave Corr Factor	0.973
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	83.0	Prev response:	80.09	*% change:	3.5%
Baseline Corr 2nd AF pt:	41.3	AF Slope:	1.038460	AF Intercept:	-0.460261
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999983		

* = > +/-5% change initiates investigation

Notes: Installed new zero air generator. Changed sample inlet filter after MAF's. Ran scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 30, 2023	Previous Calibration:	April 25, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	16:15	End Time (MST):	20:04
Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.2	----	Correlation Coefficient	0.999998	≥0.995
80.0	82.6	0.9686			
40.0	41.1	0.9733	Slope	1.035031	0.90 - 1.10
20.0	20.4	0.9781			
			Intercept	-0.240256	+/-3



H₂S Calibration Plot

Date: May 30, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Christina Lake
Calibration Date: May 31, 2023
Start time (MST): 8:11
Reason: Routine
Station number: AMS 26
Last Cal Date: April 26, 2023
End time (MST): 13:18

Calibration Standards

NO Gas Cylinder #: T2Y1P4C
NOX Cal Gas Conc: 50.82 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 50.82 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: November 12, 2023
NO Cal Gas Conc: 50.02 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.02 ppm
NO gas Diff:
Serial Number: 2447
Serial Number: 832

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1173480006

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.762	1.795	NO bkgnd or offset:	2.9	3.4
NOX coeff or slope:	0.997	0.997	NOX bkgnd or offset:	3.0	4.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	198.6	202.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999958	1.003078
NO _x Cal Offset:	-1.700000	-1.760000
NO Cal Slope:	0.999786	1.002199
NO Cal Offset:	-2.340000	-2.660000
NO ₂ Cal Slope:	1.003094	1.007991
NO ₂ Cal Offset:	0.778854	0.778153



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.8	0.3	0.5	----	----
as found span	4920	80.0	813.1	800.3	12.8	801.4	784.6	16.9	1.0146	1.0200
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4920	80.0	813.1	800.3	12.8	814.7	800.8	13.9	0.9981	0.9994
second point	4960	40.0	406.6	400.2	6.4	405.3	396.9	8.4	1.0031	1.0082
third point	4980	20.0	203.3	200.1	3.2	200.3	195.3	5.1	1.0149	1.0245
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.2	----	----
as left span	4920	80.0	813.1	395.5	417.6	817.9	399.3	418.7	0.9942	0.9905
Average Correction Factor									1.0053	1.0107

Corrected As found	NO _x = 800.6 ppb	NO = 784.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.3%
Previous Response	NO _x = 811.4 ppb	NO = 797.8 ppb		*Percent Change	NO = -1.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	799.6	394.8	417.6	421.4	0.9910	100.9%
2nd GPT point (200 ppb O ₃)	799.6	593.8	218.6	221.2	0.9882	101.2%
3rd GPT point (100 ppb O ₃)	799.6	699.3	113.1	115.8	0.9767	102.4%
Average Correction Factor					0.9853	101.5%

Notes: Changed sample inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 31, 2023	Previous Calibration:	April 26, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	8:11	End Time (MST):	13:18
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient 0.999979	≥0.995	
813.1	814.7	0.9981			
406.6	405.3	1.0031			
203.3	200.3	1.0149			
			Slope	1.003078	0.90 - 1.10
			Intercept	-1.760000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 31, 2023	Previous Calibration:	April 26, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	8:11	End Time (MST):	13:18
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	800.8	0.9994		
400.2	396.9	1.0082		
200.1	195.3	1.0245		
			0.999950	
			1.002199	
			-2.660000	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 31, 2023	Previous Calibration:	April 26, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	8:11	End Time (MST):	13:18
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
417.6	421.4	0.9910		
218.6	221.2	0.9882		
113.1	115.8	0.9767		
			0.999982	
			1.007991	
			0.778153	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 31, 2023

Location: Christina Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27
JACKFISH 2/3
MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 1, 2023	Previous Calibration:	April 4, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	10:41	End Time (MST):	13:12
Analyzer make:	Thero 43iQ	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999976	≥0.995
800.2	817.1	0.9793			
399.5	405.1	0.9863	Slope	1.022106	0.90 - 1.10
200.3	201.4	0.9946			
			Intercept	-1.838756	+/-30



SO2 Calibration Plot

Date: May 1, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Jackfish 2/3 Station number: AMS27
 Calibration Date: May 18, 2023 Last Cal Date: April 27, 2023
 Start time (MST): 6:53 End time (MST): 10:25
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.41 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345023
 Removed Cal Gas Conc: 5.41 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3811
 ZAG Make/Model: API 701 Serial Number: 135

Analyzer Information

Analyzer make: API T101 Analyzer serial #: 621
 Converter make: Converter serial #:
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000342	1.000198	Backgd or Offset: 25.7	25.7
Calibration intercept:	-0.217897	-0.137827	Coeff or Slope: 0.970	0.970

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.1	80.2	81.0	0.990
as found 2nd point	4963	37.0	40.0	40.3	0.993
as found 3rd point	4982	18.5	20.0	19.6	1.021
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4926	74.1	80.2	80.0	1.002
second point	4963	37.0	40.0	40.2	0.996
third point	4982	18.5	20.0	19.5	1.026
as left zero	5000	0.0	0.0	0.1	----
as left span	4926	74.1	80.2	79.7	1.006
SO2 Scrubber Check	4921	79.1	791.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	1.008
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 81.0 Prev response: 79.98 *% change: 1.3%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 1.012745 AF Intercept: -0.277693
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999934

* = > +/-5% change initiates investigation

Notes: No adjustments made.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 18, 2023	Previous Calibration:	April 27, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	6:53	End Time (MST):	10:25
Analyzer make:	API T101	Analyzer serial #:	621

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999928	
80.2	80.0	1.0022			≥0.995
40.0	40.2	0.9959	Slope	1.000198	
20.0	19.5	1.0264			0.90 - 1.10
			Intercept	-0.137827	+/-3



H₂S Calibration Plot

Date: May 18, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.0	0.7	0.3	----	----
as found span	4921	79.4	816.8	800.3	16.5	767.2	744.3	22.9	1.0647	1.0752
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
high point	4921	79.4	816.8	800.3	16.5	819.9	804.3	15.5	0.9962	0.9950
second point	4960	39.7	408.5	400.2	8.3	406.5	400.5	6.0	1.0048	0.9993
third point	4980	19.8	203.7	199.6	4.1	201.3	195.4	5.9	1.0120	1.0215
as left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
as left span	4921	79.4	816.8	415.1	413.0	819.0	417.2	401.7	0.9973	0.9950
Average Correction Factor									1.0043	1.0052

Corrected As found	NO _x = 766.2 ppb	NO = 743.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -6.8%	
Previous Response	NO _x = 818.6 ppb	NO = 799.3 ppb		*Percent Change	NO = -7.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	803.2	406.7	413.0	413.1	0.9998	100.0%
2nd GPT point (200 ppb O3)	803.2	619.9	199.8	200.8	0.9951	100.5%
3rd GPT point (100 ppb O3)	803.2	714.3	105.4	106.5	0.9898	101.0%
Average Correction Factor					0.9949	100.5%

Notes: Adjusted both zero and span. As finds are 7% low; due to low pump issue.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 25, 2023	Previous Calibration:	April 20, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:40	End Time (MST):	14:18
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
816.8	819.9	0.9962		
408.5	406.5	1.0048		
203.7	201.3	1.0120		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 25, 2023	Previous Calibration:	April 20, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:40	End Time (MST):	14:18
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
800.3	804.3	0.9950			
400.2	400.5	0.9993			
199.6	195.4	1.0215			
			Slope	1.007298	0.90 - 1.10
			Intercept	-2.599343	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 25, 2023	Previous Calibration:	April 20, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:40	End Time (MST):	14:18
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
413.0	413.1	0.9998		
199.8	200.8	0.9951		
105.4	106.5	0.9898		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 25, 2023

Location: Jackfish 2/3





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS29
SURMONT 2
MAY 2023**


Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023

SO2 Calibration Plot

Date: May 9, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 10, 2023	Previous Calibration:	May 9, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:15	End Time (MST):	15:20
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.3	----	Correlation Coefficient	0.999980	
800.1	801.0	0.9989			≥0.995
400.6	398.2	1.0060	Slope	1.002228	
199.8	197.2	1.0133			0.90 - 1.10
			Intercept	-1.885142	+/-30



SO2 Calibration Plot

Date: May 10, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 9, 2023	Previous Calibration:	April 12, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	11:15	End Time (MST):	15:52
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999923	≥0.995
80.0	80.5	0.9938			
40.1	39.7	1.0103	Slope	1.008319	0.90 - 1.10
20.1	19.7	1.0180			
			Intercept	-0.382705	+/-3



H₂S Calibration Plot

Date: May 9, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	May 9, 2023	Last Cal Date:	April 13, 2023
Start time (MST):	9:51	End time (MST):	11:31
Reason:	As Found		

Calibration Standards

Gas Cert Reference:	CC356008	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
C3H8 Cal Gas Conc.	<u>205.7</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
Removed C3H8 Conc.	<u>205.7</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5258
ZAG Make/Model:	Teledyne API T701	Serial Number:	4297

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1170050149
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000707		Background:	4.62	4.62
Calibration intercept:	-0.035654		Coefficient:	5.286	5.286

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4918	81.3	17.31	16.32	1.061
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	-0.01	----
as left span	4918	81.3	17.31	16.36	1.058
Average Correction Factor					
Baseline Corr As found:	16.31	Previous response	17.29	*% change	-6.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation


Notes: As founds completed. A large adjustment was needed, potential reasons: the pump needs a changeout, or the hydrogen cylinder level is too low. Both will be tested May 10.

Calibration Performed By: Braiden Boutillier

THC Calibration Plot

Date: May 9, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	May 10, 2023	Last Cal Date:	May 9, 2023
Start time (MST):	10:15	End time (MST):	15:20
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC356008	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
C3H8 Cal Gas Conc.	<u>205.7</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
Removed C3H8 Conc.	<u>205.7</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5258
ZAG Make/Model:	Teledyne API T701	Serial Number:	4297

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1170050149
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000707	1.002670	Background:	4.62	4.88
Calibration intercept:	-0.035654	-0.072027	Coefficient:	5.286	5.485

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.04	----
as found span	4918	81.3	17.31	16.33	1.060
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	-0.05	----
high point	4918	81.3	17.31	17.32	1.000
second point	4959	40.7	8.67	8.55	1.014
third point	4979	20.3	4.32	4.27	1.012
as left zero	5000	0.0	0.00	-0.15	----
as left span	4918	81.3	17.31	17.43	0.993
Average Correction Factor					1.008
Baseline Corr As found:	16.29	Previous response	17.29	*% change	-6.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Changed hydrogen cylinder. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 10, 2023	Previous Calibration:	May 9, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:15	End Time (MST):	15:20
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<i>Limits</i>	
0.00	-0.05	----	Correlation Coefficient	0.999961	
17.31	17.32	0.9997			≥ 0.995
8.67	8.55	1.0137	Slope	1.002670	
4.32	4.27	1.0120			$0.90 - 1.10$
			Intercept	-0.072027	± 1.5



THC Calibration Plot

Date: May 10, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
as found span	4916	84.2	799.2	799.2	0.0	803.0	804.0	-1.0	0.9953	0.9940
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
high point	4916	84.2	799.2	799.2	0.0	801.0	802.0	-0.4	0.9977	0.9965
second point	4958	42.1	399.6	399.6	0.0	400.8	400.3	0.6	0.9970	0.9983
third point	4979	21.1	200.3	200.3	0.0	199.5	197.6	1.9	1.0039	1.0135
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.2	----	----
as left span	4916	84.2	799.2	388.2	411.0	797.8	383.4	414.4	1.0017	1.0125
Average Correction Factor									0.9996	1.0028

Corrected As found	NO _x = 803.0 ppb	NO = 804.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.5%
Previous Response	NO _x = 799.0 ppb	NO = 798.0 ppb		*Percent Change	NO = 0.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.9	384.9	411.0	408.8	1.0054	99.5%
2nd GPT point (200 ppb O3)	795.9	592.1	203.8	203.2	1.0030	99.7%
3rd GPT point (100 ppb O3)	795.9	686.9	109.0	107.1	1.0177	98.3%
Average Correction Factor					1.0087	99.1%

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 26, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:25	End Time (MST):	15:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.2	801.0	0.9977		
399.6	400.8	0.9970		
200.3	199.5	1.0039		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 26, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:25	End Time (MST):	15:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.2	802.0	0.9965		
399.6	400.3	0.9983		
200.3	197.6	1.0135		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 26, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:25	End Time (MST):	15:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
411.0	408.8	1.0054		
203.8	203.2	1.0030		
109.0	107.1	1.0177		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 11, 2023

Location: Surmont 2





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS30 ELLS RIVER MAY 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 5, 2023	Previous Calibration:	April 11, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:53	End Time (MST):	12:35
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999933	≥0.995
800.4	806.0	0.9930			
400.2	399.6	1.0016	Slope	1.008699	0.90 - 1.10
200.1	195.8	1.0220			
			Intercept	-2.796032	+/-30



SO2 Calibration Plot

Date: May 5, 2023

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Ells River Station number: AMS30
 Calibration Date: May 10, 2023 Last Cal Date: April 3, 2023
 Start time (MST): 9:00 End time (MST): 13:00
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.08 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002443
 Removed Cal Gas Conc: 5.08 ppm Rem Gas Exp Date:
 Removed Gas Cyl #: Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3061
 ZAG Make/Model: API T701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1410661331
 Converter make: CDN - 101 Converter serial #: 562
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000494	1.007352	Backgd or Offset:	1.55 1.58
Calibration intercept:	-0.019196	-0.159242	Coeff or Slope:	1.092 1.107

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	78.7	80.0	78.3	1.021
as found 2nd point	4961	39.4	40.0	38.9	1.029
as found 3rd point	4980	19.7	20.0	19.4	1.032
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	78.7	80.0	80.5	0.993
second point	4961	39.4	40.0	40.0	1.001
third point	4980	19.7	20.0	19.9	1.006
as left zero	5000	0.0	0.0	0.1	----
as left span	4921	78.7	80.0	80.0	1.000
SO2 Scrubber Check	4921	79.2	800.4	0.1	----
Date of last scrubber change:	N/A		Ave Corr Factor		1.000
Date of last converter efficiency test:	N/A		95.1% efficiency		

Baseline Corr As found: 78.3 Prev response: 79.98 *% change: -2.2%
 Baseline Corr 2nd AF pt: 38.9 AF Slope: 0.979624 AF Intercept: -0.138713
 Baseline Corr 3rd AF pt: 19.4 AF Correlation: 0.999981

* = > +/-5% change initiates investigation

Notes: Adjusted the span only.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 10, 2023	Previous Calibration:	April 3, 2023
Station Name:	Ells River	Station Number:	AMS30
Start Time (MST):	9:00	End Time (MST):	13:00
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999979	
80.0	80.5	0.9933			≥0.995
40.0	40.0	1.0007	Slope	1.007352	
20.0	19.9	1.0058			0.90 - 1.10
			Intercept	-0.159242	+/-3



TRS Calibration Plot

Date: May 10, 2023

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	May 5, 2023	Last Cal Date:	April 11, 2023
Start time (MST):	9:53	End time (MST):	12:35
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm		
Diff between cyl (CH ₄):		Diff between cyl (THC):	
Diff between cyl (NMHC):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
ZAG make/model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	0.000238	0.000238	NMHC SP Ratio:	4.39E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	207620
				207620

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	17.02	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.02	1.000
second point	4960	39.6	8.51	8.45	1.008
third point	4980	19.8	4.26	4.17	1.021
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.10	0.996

Average Correction Factor				1.010
Baseline Corr AF:	17.02	Prev response	16.95	*% change 0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	9.15	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.15	0.996
second point	4960	39.6	4.56	4.55	1.002
third point	4980	19.8	2.28	2.24	1.018
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.20	0.991
Average Correction Factor					1.005
Baseline Corr AF:	9.15	Prev response	9.09	*% change	0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.87	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.87	1.005
second point	4960	39.6	3.96	3.90	1.014
third point	4980	19.8	1.98	1.93	1.025
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.90	1.002
Average Correction Factor					1.015
Baseline Corr AF:	7.87	Prev response	7.86	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998335	1.000973
THC Cal Offset:	-0.048138	-0.047539
CH ₄ Cal Slope:	0.996510	0.995687
CH ₄ Cal Offset:	-0.029356	-0.021757
NMHC Cal Slope:	0.999568	1.005200
NMHC Cal Offset:	-0.018381	-0.024582

Notes: No adjustments made.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 5, 2023	Previous Calibration:	April 11, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:53	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999964	≥ 0.995			
17.03	17.02	1.0004						
8.51	8.45	1.0078				Slope	1.000973	0.90 - 1.10
4.26	4.17	1.0214						
			Intercept	-0.047539	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-01-2020


Station Information

Calibration Date:	May 5, 2023	Previous Calibration:	April 11, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:53	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999963	≥0.995
7.91	7.87	1.0055			
3.96	3.90	1.0143			
1.98	1.93	1.0248			
			Slope	0.995687	0.90 - 1.10
			Intercept	-0.021757	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 5, 2023	Previous Calibration:	April 11, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:53	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999966	≥ 0.995			
9.11	9.15	0.9963						
4.56	4.55	1.0019				Slope	1.005200	0.90 - 1.10
2.28	2.24	1.0180						
			Intercept	-0.024582	± 0.5			



NMHC Calibration Plot

Date: May 5, 2023

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	May 26, 2023	Last Cal Date:	May 5, 2023
Start time (MST):	11:54	End time (MST):	13:07
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
ZAG make/model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	0.000238	0.000238	NMHC SP Ratio:	4.39E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	207620

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.83	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	16.86	1.010

Average Correction Factor					
Baseline Corr AF:	16.83	Prev response	17.00	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	9.09	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.07	1.005
Average Correction Factor					
Baseline Corr AF:	9.09	Prev response	9.14	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.74	1.023
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.79	1.016
Average Correction Factor					
Baseline Corr AF:	7.74	Prev response	7.86	*% change	-1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000973	
THC Cal Offset:	-0.047539	
CH ₄ Cal Slope:	0.995687	
CH ₄ Cal Offset:	-0.021757	
NMHC Cal Slope:	1.005200	
NMHC Cal Offset:	-0.024582	


Notes: Replaced N2 cylinder.

Calibration Performed By: Denny Ray Estador

NMHC Calibration Plot

Date: May 26, 2023

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Ells River
Calibration Date: May 11, 2023
Start time (MST): 8:51
Reason: Routine
Station number: AMS 30
Last Cal Date: April 5, 2023
End time (MST): 13:43

Calibration Standards

NO Gas Cylinder #: T2Y1P2R
NOX Cal Gas Conc: 50.83 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 50.83 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 49.97 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 49.97 ppm
NO gas Diff:
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 710321429

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.029	1.051	NO bkgnd or offset:	12.5	12.7
NOX coeff or slope:	0.992	0.990	NOX bkgnd or offset:	12.5	12.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	183.6	181.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000576	1.000351
NO _x Cal Offset:	-1.140000	-1.160000
NO Cal Slope:	0.998571	1.001858
NO Cal Offset:	-1.840000	-2.240000
NO ₂ Cal Slope:	1.002663	0.998225
NO ₂ Cal Offset:	0.679699	0.085999



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
as found span	4920	80.0	813.3	799.5	13.8	803.0	786.9	16.1	1.0128	1.0160
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4920	80.0	813.3	799.5	13.8	813.0	799.9	13.2	1.0003	0.9995
second point	4960	40.0	406.6	399.8	6.9	405.0	397.1	7.9	1.0040	1.0067
third point	4980	20.0	203.3	199.9	3.4	201.1	195.8	5.2	1.0110	1.0208
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
as left span	4920	80.0	813.3	427.5	385.8	816.7	428.4	388.2	0.9958	0.9979
Average Correction Factor									1.0051	1.0090

Corrected As found	NO _x = 802.9 ppb	NO = 787.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.2%	
Previous Response	NO _x = 812.6 ppb	NO = 796.5 ppb		*Percent Change	NO = -1.2%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	797.5	425.5	385.8	385.0	1.0020	99.8%
2nd GPT point (200 ppb O ₃)	797.5	615.1	196.2	196.3	0.9993	100.1%
3rd GPT point (100 ppb O ₃)	797.5	703.6	107.7	107.4	1.0024	99.8%
Average Correction Factor					1.0012	99.9%

Notes:

Both zero and span was adjusted.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 5, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:51	End Time (MST):	13:43
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.3	813.0	1.0003		
406.6	405.0	1.0040		
203.3	201.1	1.0110		
			0.999991	
			1.000351	
			-1.160000	





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 5, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:51	End Time (MST):	13:43
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.5	799.9	0.9995		
399.8	397.1	1.0067		
199.9	195.8	1.0208		
			0.999964	
			1.001858	
			-2.240000	





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 11, 2023	Previous Calibration:	April 5, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:51	End Time (MST):	13:43
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
385.8	385.0	1.0020		
196.2	196.3	0.9993		
107.7	107.4	1.0024		
			0.999997	
			0.998225	
			0.085999	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 11, 2023

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Ells River Station number: AMS 30
 Calibration Date: May 10, 2023 Last Cal Date: April 11, 2023
 Start time (MST): 9:26 End time (MST): 9:48

Analyzer Make: API T640 S/N: 875
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 954
 Temp/RH standard: Delta Cal S/N: 954

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	18.2	17.5	18.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	732.4	732.5	732.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>May 10, 2023</u>	Last Cal Date: <u>April 11, 2023</u>			
	PM w/o HEPA: <u>9.9</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		_____			<0.2 ug/m3
Disposable Filter Changed:		_____			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: No adjustments made. Inlet head still clean.

Calibration by: Denny Ray Estador



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS506
JACKFISH 1
MAY 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 18, 2023	Previous Calibration:	April 20, 2023
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	9:06	End Time (MST):	11:54
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999980	
800.2	802.1	0.9976			≥0.995
400.2	399.5	1.0016	Slope	1.003957	
200.1	197.0	1.0156			0.90 - 1.10
			Intercept	-1.916022	+/-30



SO2 Calibration Plot

Date: May 18, 2023

Location: Jackfish 1





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Jackfish 1 Station number: AMS506
 Calibration Date: May 25, 2023 Last Cal Date: April 25, 2023
 Start time (MST): 8:16 End time (MST): 12:03
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.14 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511843
 Removed Cal Gas Conc: 5.14 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 2659
 ZAG Make/Model: API 701 Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1180540020
 Converter make: Global G150 Converter serial #: 2022-218
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999860	1.015003	Backgd or Offset: 3.37	3.33
Calibration intercept:	-0.298400	-0.378084	Coeff or Slope: 1.066	1.066

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4922	77.8	80.0	80.3	0.992
as found 2nd point	4961	38.9	40.0	39.9	0.995
as found 3rd point	4981	19.4	19.9	19.4	1.012
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4922	77.8	80.0	80.8	0.990
second point	4961	38.9	40.0	40.3	0.992
third point	4981	19.4	19.9	19.7	1.012
as left zero	5000	0.0	0.0	-0.3	----
as left span	4922	77.8	80.0	79.9	1.001
SO2 Scrubber Check	4921	79.2	792.0	0.0	----
Date of last scrubber change:	24-Feb-23			Ave Corr Factor	0.998
Date of last converter efficiency test:	December 1, 2022			efficiency	

Baseline Corr As found: 80.6 Prev response: 79.67 *% change: 1.2%
 Baseline Corr 2nd AF pt: 40.2 AF Slope: 1.009290 AF Intercept: -0.478254
 Baseline Corr 3rd AF pt: 19.7 AF Correlation: 0.999973

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after multi-point as founds. Scrubber test done and passed after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 25, 2023	Previous Calibration:	April 25, 2023
Station Name:	Jackfish 1	Station Number:	AMS506
Start Time (MST):	8:16	End Time (MST):	12:03
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999989	≥0.995
80.0	80.8	0.9899			
40.0	40.3	0.9923	Slope	1.015003	0.90 - 1.10
19.9	19.7	1.0123			
			Intercept	-0.378084	+/-3



H₂S Calibration Plot

Date: May 25, 2023

Location: Jackfish 1





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.4	0.3	0.1	----	----
as found span	4916	84.4	801.1	799.9	1.2	806.6	800.4	6.1	0.9931	0.9993
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4916	84.4	801.1	799.9	1.2	800.3	799.4	0.8	1.0010	1.0006
second point	4958	42.2	400.5	400.0	0.6	398.8	396.8	2.0	1.0044	1.0080
third point	4979	21.1	200.3	200.0	0.3	196.6	194.0	2.6	1.0187	1.0308
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4916	84.4	801.1	405.9	395.2	788.3	402.3	386.0	1.0162	1.0089
Average Correction Factor									1.0080	1.0131

Corrected As found	NO _x = 806.2 ppb	NO = 800.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.6%
Previous Response	NO _x = 801.5 ppb	NO = 798.1 ppb		*Percent Change	NO = 0.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	791.6	397.6	395.2	394.7	1.0012	99.9%
2nd GPT point (200 ppb O3)	791.6	594.4	198.4	197.9	1.0024	99.8%
3rd GPT point (100 ppb O3)	791.6	685.5	107.3	104.6	1.0256	97.5%
Average Correction Factor					1.0098	99.0%

Notes: Changed inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 26, 2023	Previous Calibration:	April 26, 2023
Station Name:	Jackfish 1	Station Number:	AMS506
Start Time (MST):	8:06	End Time (MST):	12:25
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
801.1	800.3	1.0010			
400.5	398.8	1.0044			
200.3	196.6	1.0187			
			Slope	1.000448	0.90 - 1.10
			Intercept	-1.727986	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 26, 2023	Previous Calibration:	April 26, 2023
Station Name:	Jackfish 1	Station Number:	AMS506
Start Time (MST):	8:06	End Time (MST):	12:25
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999939	≥0.995
799.9	799.4	1.0006			
400.0	396.8	1.0080	Slope	1.001597	0.90 - 1.10
200.0	194.0	1.0308			
			Intercept	-2.987959	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 26, 2023	Previous Calibration:	April 26, 2023
Station Name:	Jackfish 1	Station Number:	AMS506
Start Time (MST):	8:06	End Time (MST):	12:25
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
395.2	394.7	1.0012		
198.4	197.9	1.0024		
107.3	104.6	1.0256		
			0.999949	
			1.000773	
			-1.046494	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 26, 2023

Location: Jackfish 1





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS508
KIRBY NORTH
MAY 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

June 30, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 18, 2023	Previous Calibration:	April 21, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	7:42	End Time (MST):	11:39
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.2	----	Correlation Coefficient	0.999979	≥0.995
799.6	805.0	0.9933			
400.3	399.7	1.0016	Slope	1.006620	0.90 - 1.10
199.7	200.6	0.9953			
			Intercept	-0.948262	+/-30



SO2 Calibration Plot

Date: May 18, 2023

Location: Kirby North





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Kirby North	Station number:	AMS508
Calibration Date:	May 17, 2023	Last Cal Date:	April 20, 2023
Start time (MST):	10:58	End time (MST):	17:13
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.167	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	<u>CC517378</u>			
Removed Cal Gas Conc:	5.167	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3804
ZAG Make/Model:	API T701H		Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1150840012
Converter make:	Global	Converter serial #:	2022-197
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001312	1.003026	Backgd or Offset:	1.60	1.58
Calibration intercept:	-0.100838	-0.160868	Coeff or Slope:	1.046	1.027

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4923	77.4	80.0	79.8	1.001
as found 2nd point	4961	38.8	40.1	39.7	1.007
as found 3rd point	4981	19.3	19.9	19.6	1.012
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.4	80.0	80.2	0.997
second point	4961	38.8	40.1	39.8	1.007
third point	4981	19.3	19.9	19.8	1.007
as left zero	5000	0.0	0.0	0.1	----
as left span	4923	77.4	80.0	79.9	1.001

SO2 Scrubber Check

Date of last scrubber change:	19-Apr-23	Ave Corr Factor	1.004
Date of last converter efficiency test:		efficiency	

Baseline Corr As found:	79.9	Prev response:	79.98	*% change:	-0.1%
Baseline Corr 2nd AF pt:	39.8	AF Slope:	0.999603	AF Intercept:	-0.241018
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999984		

* = > +/-5% change initiates investigation

Notes: Adjusted span. Moved scrubber outside of the Global Converter after MPAF's.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	May 17, 2023	Previous Calibration:	April 20, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	10:58	End Time (MST):	17:13
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999968	≥0.995
80.0	80.2	0.9972			
40.1	39.8	1.0075	Slope	1.003026	0.90 - 1.10
19.9	19.8	1.0072			
			Intercept	-0.160868	+/-3



H₂S Calibration Plot

Date: May 17, 2023

Location: Kirby North





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	May 18, 2023	Previous Calibration:	April 21, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	7:42	End Time (MST):	11:39
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.12	----	Correlation Coefficient	0.999958	≥0.995
17.26	17.27	0.9996			
8.64	8.57	1.0085	Slope	1.009048	0.90 - 1.10
4.31	4.12	1.0459			
			Intercept	-0.160831	+/-1.5



THC Calibration Plot

Date: May 18, 2023

Location: Kirby North





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.5	0.0	0.5	----	----
as found span	4919	81.0	800.1	794.1	6.0	798.2	788.3	9.9	1.0024	1.0074
as found 2nd	4960	40.5	400.0	397.0	3.0	397.0	392.3	4.7	1.0076	1.0120
as found 3rd	4980	20.2	199.5	198.0	1.5	197.0	192.7	4.3	1.0128	1.0277
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

Average Correction Factor

Corrected As found	NO _x = 797.7 ppb	NO = 788.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.2%
Previous Response	NO _x = 799.5 ppb	NO = 793.7 ppb		*Percent Change	NO = -0.7%
Baseline Corr 2nd pt	NO _x = 396.5 ppb	NO = 392.3 ppb	As found	NO _x r ² : 0.999984	Nx SI: 0.997689
Baseline Corr 3rd pt	NO _x = 196.5 ppb	NO = 192.7 ppb	As found	NO r ² : 0.999972	NO SI: 0.994084
			As found	NO ₂ r ² : 0.999985	NO ₂ SI: 1.003303
					NO ₂ Int: 1.154

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	0.5	----	----
as found GPT point (400 ppb NO ₂)	785.8	402.2	389.6	391.6	0.9949	100.5%
as found GPT point (200 ppb NO ₂)	785.8	604.4	187.4	189.9	0.9868	101.3%
as found GPT point (100 ppb NO ₂)	785.8	701.9	89.9	91.7	0.9803	102.0%
1st GPT point (400 ppb O ₃)						
2nd GPT point (200 ppb O ₃)						
3rd GPT point (100 ppb O ₃)						

Average Correction Factor


Notes: Planned to change the analyzer due to a memory error. Could not complete as communication issues occurred with the new analyzer. Deleted datalog on the DAS (potential fix for memory issue). Will finish the work May 16.

Calibration Performed By: Braiden Boutilier

NO_x Calibration Plot

Date: May 15, 2023

Location: Kirby North





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.4	0.0	-0.4	----	----
high point	4919	81.0	800.1	794.1	6.0	800.0	793.7	6.7	1.0001	1.0005
second point	4960	40.5	400.0	397.0	3.0	397.9	395.1	2.8	1.0053	1.0049
third point	4980	20.2	199.5	198.0	1.5	196.0	194.1	1.9	1.0180	1.0203
as left zero	5000	0.0	0.0	0.0	0.0	-1.5	0.2	-1.6	----	----
as left span	4919	81.0	800.1	406.1	394.0	793.6	404.9	388.7	1.0082	1.0030
Average Correction Factor									1.0078	1.0086

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	791.9	403.9	394.0	392.1	1.0048	99.5%
2nd GPT point (200 ppb O3)	791.9	620.6	177.3	174.8	1.0143	98.6%
3rd GPT point (100 ppb O3)	791.9	711.9	86.0	85.9	1.0011	99.9%
Average Correction Factor					1.0067	99.3%

Notes: Calibration completed after as founds and maintenance done yesterday. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	May 15, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	9:58	End Time (MST):	14:18
Analyzer make:	API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.1	800.0	1.0001		
400.0	397.9	1.0053		
199.5	196.0	1.0180		
			0.999980	
			1.001459	
			-2.051855	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	May 15, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	9:58	End Time (MST):	14:18
Analyzer make:	API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
794.1	793.7	1.0005		
397.0	395.1	1.0049		
198.0	194.1	1.0203		
			0.999973	
			1.000875	
			-1.873766	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	May 16, 2023	Previous Calibration:	May 15, 2023
Station Name:	Kirby North	Station Number:	AMS508
Start Time (MST):	9:58	End Time (MST):	14:18
Analyzer make:	API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
394.0	392.1	1.0048		
177.3	174.8	1.0143		
86.0	85.9	1.0011		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: May 16, 2023

Location: Kirby North





End of Report