



**WOOD BUFFALO  
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive  
Fort McMurray, AB T9K 0K4  
P: 780.799.4420 E: info@wbea.org  
[wbea.org](http://wbea.org)

Wood Buffalo Environmental Association

# JULY 2023

# MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

August 31, 2023

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS01 BERTHA GANTER - FORT MCKAY JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	July 14, 2023	Last Cal Date:	June 1, 2023
Start time (MST):	9:56	End time (MST):	14:53
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC418809			
Removed Cal Gas Conc:	49.19	ppm	Rem Gas Exp Date:	February 23, 2025
Removed Gas Cyl #:	CC486642		Diff between cyl:	-1.8%
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999958	1.000680	Backgd or Offset:	19.1	19.2
Calibration intercept:	0.086959	-0.433310	Coeff or Slope:	0.883	0.892

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4918	81.3	799.9	802.4	0.997
as found 2nd point					
as found 3rd point					
new cylinder response	4918	81.3	800.3	788.4	1.015
calibrator zero	5000	0.0	0.0	0.5	----
high point	4918	81.3	800.3	801.0	0.999
second point	4959	40.7	400.6	399.5	1.003
third point	4979	20.3	199.8	198.9	1.005
as left zero	5000	0.0	0.0	0.4	----
as left span	4918	81.3	800.3	801.0	0.999
Average Correction Factor					1.002

Baseline Corr As found:	802.20	Previous response	799.99	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter and SO<sub>2</sub>/HC cylinder after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 14, 2023	Previous Calibration:	June 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:56	End Time (MST):	14:53
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999993	≥0.995
800.3	801.0	0.9991			
400.6	399.5	1.0027	Slope	1.000680	0.90 - 1.10
199.8	198.9	1.0046			
			Intercept	-0.433310	+/-30



SO2 Calibration Plot

Date: July 14, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Bertha Ganter-Fort McKay      Station number: AMS01  
 Calibration Date: July 18, 2023      Last Cal Date: June 27, 2023  
 Start time (MST): 9:41      End time (MST): 15:29  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.10 ppm      Cal Gas Exp Date: September 16, 2024  
 Cal Gas Cylinder #: CC511749  
 Removed Cal Gas Conc: 5.10 ppm      Rem Gas Exp Date: N/A  
 Removed Gas Cyl #: N/A      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 3565  
 ZAG Make/Model: Teledyne API T701      Serial Number: 5609

### Analyzer Information

Analyzer make: Thermo 43i-TLE      Analyzer serial #: 1218153461  
 Converter make: CD Nova      Converter serial #: 470  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002650	1.000507	Backgd or Offset: 2.30	2.28
Calibration intercept:	0.159995	0.159999	Coeff or Slope: 0.919	0.919

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4921	78.4	80.0	78.7	1.019
as found 2nd point	4960	39.2	40.0	39.4	1.020
as found 3rd point	4980	19.6	20.0	19.8	1.020
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	78.4	80.0	80.2	0.997
second point	4960	39.2	40.0	40.2	0.995
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.3	----
as left span	4921	78.4	80.0	79.2	1.010
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	0.996
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 78.5      Prev response: 80.37      \*% change: -2.4%  
 Baseline Corr 2nd AF pt: 39.2      AF Slope: 0.981363      AF Intercept: 0.180000  
 Baseline Corr 3rd AF pt: 19.6      AF Correlation: 0.999999

\* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 27, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:41	End Time (MST):	15:29
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999998	≥0.995
80.0	80.2	0.9974			
40.0	40.2	0.9950	Slope	1.000507	0.90 - 1.10
20.0	20.1	0.9949			
			Intercept	0.159999	+/-3



TRS Calibration Plot

Date: July 18, 2023

Location: Bertha Ganter-Fort McKay







# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Bertha Ganter-Fort McKay      Station number: AMS01  
 Calibration Date: July 18, 2023      Last Cal Date: June 28, 2023  
 Start time (MST): 9:41      End time (MST): 15:29  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.10 ppm      Cal Gas Exp Date: September 16, 2024  
 Cal Gas Cylinder #: CC511749  
 Removed Cal Gas Conc: 5.10 ppm      Rem Gas Exp Date: N/A  
 Removed Gas Cyl #: N/A      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 3565  
 ZAG Make/Model: Teledyne API T701      Serial Number: 5609

### Analyzer Information

Analyzer make: Thermo 43iQTL      Analyzer serial #: 1200326167  
 Converter make: Thermo Converter      Converter serial #: N/A  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997140	1.002857	Backgd or Offset:	1.91	1.95
Calibration intercept:	0.156826	0.056779	Coeff or Slope:	0.992	1.014

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4922	78.4	80.0	80.0	0.997
as found 2nd point	4960	39.2	40.0	39.6	1.005
as found 3rd point	4980	19.6	20.0	19.7	1.005
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	78.4	80.0	80.2	0.997
second point	4960	39.2	40.0	40.3	0.993
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.4	----
as left span	4922	78.4	80.0	77.2	1.036
SO2 Scrubber Check	4919	81.3	813.0	-0.1	----
Date of last scrubber change:	March 21, 2022			Ave Corr Factor	0.995
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 80.2      Prev response: 79.89      \*% change: 0.4%  
 Baseline Corr 2nd AF pt: 39.8      AF Slope: 1.003195      AF Intercept: -0.323189  
 Baseline Corr 3rd AF pt: 19.9      AF Correlation: 0.999981

\* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 28, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:41	End Time (MST):	15:29
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999993	
80.0	80.2	0.9972			≥0.995
40.0	40.3	0.9925	Slope	1.002857	
20.0	20.1	0.9949			0.90 - 1.10
			Intercept	0.056779	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 18, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	July 14, 2023	Last Cal Date:	June 1, 2023
Start time (MST):	9:56	End time (MST):	14:53
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH <sub>4</sub> Cal Gas Conc.	497.2 ppm	CH <sub>4</sub> Equiv Conc.	1061.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	CC486642	Removed Gas Expiry:	February 23, 2025
Removed CH <sub>4</sub> Conc.	497.7 ppm	CH <sub>4</sub> Equiv Conc.	1063.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.6 ppm	Diff between cyl (THC):	2.3%
Diff between cyl (CH <sub>4</sub> ):	0.7%	Diff between cyl (NM):	3.8%
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	5609

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.89E-04	2.90E-04	NMHC SP Ratio:	6.23E-05
CH <sub>4</sub> Retention time:	14.6	14.6	NMHC Peak Area:	147590
				149639

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.29	17.09	1.012
as found 2nd point					
as found 3rd point					
new cylinder response	4918	81.3	17.27	17.47	0.988
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	17.27	17.29	0.999
second point	4959	40.7	8.64	8.51	1.016
third point	4980	20.3	4.31	4.24	1.017
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.27	17.28	0.999

Average Correction Factor				1.010
Baseline Corr AF:	17.09	Prev response	17.23	*% change -0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4918	81.3	9.19	9.05	1.016
as found 2nd point					
as found 3rd point					
new cylinder response	4918	81.3	9.18	9.39	0.978
calibrator zero	5000	0	0.00	0.00	----
high point	4918	81.3	9.18	9.19	0.999
second point	4959	40.7	4.60	4.55	1.010
third point	4980	20.3	2.29	2.28	1.004
as left zero	5000	0	0.00	0.00	----
as left span	4918	81.3	9.18	9.20	0.998
Average Correction Factor					1.005
Baseline Corr AF:	9.05	Prev response	9.17	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	8.04	1.007
as found 2nd point					
as found 3rd point					
new cylinder response	4918	81.3	8.09	8.08	1.000
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	8.09	8.09	0.999
second point	4959	40.7	4.05	3.96	1.022
third point	4980	20.3	2.02	1.96	1.031
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.08	1.000
Average Correction Factor					1.017
Baseline Corr AF:	8.04	Prev response	8.05	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998392	1.001595
THC Cal Offset:	-0.033901	-0.058682
CH <sub>4</sub> Cal Slope:	0.999433	1.002288
CH <sub>4</sub> Cal Offset:	-0.034161	-0.042956
NMHC Cal Slope:	0.997502	1.000424
NMHC Cal Offset:	0.000659	-0.014727

Notes: Changed the inlet filter, SO<sub>2</sub>/HC cylinder, and H<sub>2</sub> cylinder after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 14, 2023	Previous Calibration:	June 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:56	End Time (MST):	14:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999914	$\geq 0.995$			
17.27	17.29	0.9990						
8.64	8.51	1.0157				Slope	1.001595	0.90 - 1.10
4.31	4.24	1.0167						
			Intercept	-0.058682	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 14, 2023	Previous Calibration:	June 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:56	End Time (MST):	14:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999830	$\geq 0.995$
8.09	8.09	0.9991			
4.05	3.96	1.0218			
2.02	1.96	1.0310			
			Slope	1.002288	0.90 - 1.10
			Intercept	-0.042956	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 14, 2023	Previous Calibration:	June 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:56	End Time (MST):	14:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999963	$\geq 0.995$			
9.18	9.19	0.9994						
4.60	4.55	1.0105				Slope	1.000424	0.90 - 1.10
2.29	2.28	1.0045						
			Intercept	-0.014727	$\pm 0.5$			






NMHC Calibration Plot

Date: July 14, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Bertha Ganter-Fort McKay      Station number: AMS01  
Calibration Date: July 11, 2023      Last Cal Date: June 20, 2023  
Start time (MST): 9:20      End time (MST): 14:08  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T2Y1P9L      Cal Gas Expiry Date: December 11, 2023  
NOX Cal Gas Conc: 50.84 ppm      NO Cal Gas Conc: 50.04 ppm  
Removed Cylinder #: NA      Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 50.84 ppm      Removed Gas NO Conc: 50.04 ppm  
NOX gas Diff:      NO gas Diff:  
Calibrator Model: Teledyne API T700      Serial Number: 3565  
ZAG make/model: Teledyne API T701      Serial Number: 5609

### Analyzer Information

Analyzer make: Thermo 42i      Analyzer serial #: 1218153357  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.334	1.324	NO bkgnd or offset:	6.6	6.9
NOX coeff or slope:	0.988	0.991	NOX bkgnd or offset:	6.8	7.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	186.2	186.5

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000702	0.998356
NO <sub>x</sub> Cal Offset:	-0.180000	-0.420000
NO Cal Slope:	1.001256	0.998644
NO Cal Offset:	-0.920000	-1.180000
NO <sub>2</sub> Cal Slope:	0.998452	1.002441
NO <sub>2</sub> Cal Offset:	0.436719	0.783625



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.4	0.3	0.1	----	----
as found span	4920	80.0	813.4	800.6	12.8	817.3	803.2	14.0	0.9953	0.9968
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
high point	4920	80.0	813.4	800.6	12.8	811.9	798.9	13.0	1.0019	1.0022
second point	4960	40.0	406.7	400.3	6.4	405.5	398.1	7.4	1.0030	1.0056
third point	4980	20.0	203.4	200.2	3.2	202.0	197.6	4.4	1.0067	1.0130
as left zero	5000	0.0	0.0	0.0	0.0	0.2	-0.1	0.3	----	----
as left span	4920	80.0	813.4	404.8	408.6	812.0	402.5	409.5	1.0018	1.0058
Average Correction Factor									1.0039	1.0069

Corrected As found	NO <sub>x</sub> = 816.9 ppb	NO = 802.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.4%
Previous Response	NO <sub>x</sub> = 813.8 ppb	NO = 800.7 ppb		*Percent Change	NO = 0.3%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.2	400.4	408.6	410.1	0.9963	100.4%
2nd GPT point (200 ppb O3)	796.2	602.9	206.1	207.6	0.9928	100.7%
3rd GPT point (100 ppb O3)	796.2	700.3	108.7	110.4	0.9846	101.6%
Average Correction Factor					0.9912	100.9%

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 20, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:20	End Time (MST):	14:08
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
813.4	811.9	1.0019			
406.7	405.5	1.0030			
203.4	202.0	1.0067			
			Slope	0.998356	0.90 - 1.10
			Intercept	-0.420000	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 20, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:20	End Time (MST):	14:08
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
800.6	798.9	1.0022			
400.3	398.1	1.0056			
200.2	197.6	1.0130			
			Slope	0.998644	0.90 - 1.10
			Intercept	-1.180000	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 20, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:20	End Time (MST):	14:08
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
408.6	410.1	0.9963		
206.1	207.6	0.9928		
108.7	110.4	0.9846		
			0.999990	
			1.002441	
			0.783625	


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: July 11, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Bertha Ganter-Fort McKay      Station number: AMS01  
 Calibration Date: July 4, 2023      Last Cal Date: June 2, 2023  
 Start time (MST): 9:40      End time (MST): 12:21  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 3565  
 ZAG Make/Model: Teledyne API T701      Serial Number: 5609

### Analyzer Information

Analyzer make: Teledyne API T400      Analyzer serial #: 1107  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001229	1.001486	Backgd or Offset:	3.2	3.2
Calibration intercept:	0.260000	0.240000	Coeff or Slope:	1.010	1.010

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	5000	855.5	400.0	400.9	0.998
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.1	----
high point	5000	858.8	400.0	400.6	0.999
second point	5000	743.1	200.0	201.1	0.995
third point	5000	653.5	100.0	100.2	0.998
as left zero	5000	0.0	0.0	-0.3	----
as left span	5000	858.8	400.0	405.0	0.988
Average Correction Factor					0.997

Baseline Corr As found:	401.1	Previous response	400.8	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 2, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:40	End Time (MST):	12:21
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
400.0	400.6	0.9985		
200.0	201.1	0.9945	Slope	0.90 - 1.10
100.0	100.2	0.9980		
			Intercept	+/- 5



O<sub>3</sub> Calibration Plot

Date: July 4, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Fort McKay - Bertha Ganter      Station number: AMS 01  
 Calibration Date: July 18, 2023      Last Cal Date: June 2, 2023  
 Start time (MST): 12:25      End time (MST): 14:07

Analyzer Make: Teledyne API T640      S/N: 324  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal      S/N: 1450  
 Temp/RH standard: Delta Cal      S/N: 1450

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	27.7	26.6	27.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	734.1	734	734.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	5.08	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>July 18, 2023</u>		Last Cal Date: <u>June 2, 2023</u>		
	PM w/o HEPA: <u>71.1</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.1	10.5	11.1	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>63.6</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>July 18, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 18, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_  
 Date RH/T Sensor Cleaned: July 18, 2023

Notes: Flow, temperature, and pressure all within limits. Leak checks passed. PMT peak voltage adjusted from 1358 to 1362. Inlet head cleaned. Disposable filter changed. Optical chamber and RH/T sensor cleaned.

Calibration by: Rene Chamberland



# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	July 17, 2023	Last Cal Date:	June 26, 2023
Start time (MST):	10:03	End time (MST):	13:07
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	<u>3040</u>	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>3040</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

### Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001481	1.002385	Backgd or Offset:	-0.012	-0.012
Calibration intercept:	0.163854	0.147828	Coeff or Slope:	0.992	0.992

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.6	40.8	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.6	40.7	0.997
second point	4966	33.3	20.2	20.7	0.978
third point	4983	16.7	10.2	10.3	0.986
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.5	40.1	1.010
Average Correction Factor					0.987

Baseline Corr As found:	40.73	Prev response:	40.78	*% change:	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 26, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:03	End Time (MST):	13:07
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999899
40.6	40.7	0.9970		
20.2	20.7	0.9782	Slope	1.002385
10.2	10.3	0.9858		
			Intercept	0.147828
				<i>+/-1.5</i>



CO Calibration Plot

Date: July 17, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	July 10, 2023	Last Cal Date:	June 12, 2023
Start time (MST):	9:34	End time (MST):	13:03
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>60,200</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
N2 Gen Make/Model:	Peak Sci NG5000		Serial Number:	7220900034

### Analyzer Information

Analyzer make: Teledyne API 360	Analyzer serial #: 442
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001960	1.000422	Backgd or Offset:	0.045	0.045
Calibration intercept:	-4.460000	-4.480000	Coeff or Slope:	0.874	0.874

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	0.2	----
as found span	2920	80.0	1605.3	1604.9	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	-0.1	----
high point	2920	80.0	1605.3	1603.2	1.001
second point	2960	40.0	802.7	797.9	1.006
third point	2980	20.0	401.3	391.6	1.025
as left zero	3000	0.0	0.0	0.0	----
as left span	2960	40.0	802.7	786.5	1.021
Average Correction Factor					1.011

Baseline Corr As found:	1604.70	Prev response:	1604.02	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Cycled power to the analyzer. No adjustments made.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date	July 10, 2023	Previous Calibration	June 12, 2023
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS01
Start Time (MST)	9:34	End Time (MST)	13:03
Analyzer make	Teledyne API 360	Analyzer serial #	442

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999963	≥0.995
1605.3	1603.2	1.0013			
802.7	797.9	1.0060	Slope	1.000422	0.90 - 1.10
401.3	391.6	1.0249			
			Intercept	-4.480000	+/-10






CO<sub>2</sub> Calibration Plot

Date: July 10, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## TN - NO<sub>x</sub> - NH<sub>3</sub> Calibration Report

Version-05-2023

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	July 12, 2023	Last Cal Date:	June 21, 2023
Start time (MST):	9:50	End time (MST):	14:38
NH3 Cal Date:	July 13, 2023	Last Cal Date:	June 22, 2023
Start time (MST):	9:58	End time (MST):	16:46
Reason:	Routine		

### Calibration Standards

NOX Cal Gas Conc:	50.84	ppm	NO Gas Cylinder #:	T2Y1P9L
NO Cal Gas Conc:	50.04	ppm	NO Cal Gas Expiry:	March 3, 2023
Removed NOX Conc:	50.84	ppm	Removed Cylinder #:	NA
Removed NO Conc:	50.04	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	74.90	ppm	NH3 Gas Cylinder #:	CC744566
			NH3 Cal Gas Expiry:	December 21, 2023
Removed NH3 Conc:	72.93	ppm	Removed Cylinder #:	CC281298
NH3 gas Diff:	3.7%		Removed cyl Expiry:	February 28, 2023
Calibrator Model:	Teledyne API T700		Serial Number:	3565
ZAG make/model:	Teledyne API T701		Serial Number:	5609

### Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	475
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	5.00
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	512

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.989	0.987	TN coefficient:	0.995	0.989
NOX coefficient:	0.991	0.987	NO bkgnd:	-0.5	-0.9
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-0.3	-0.3
NH3 coefficient:	0.941	0.973	TN bkgnd:	0.5	1.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.998174	1.000126
NO <sub>x</sub> Cal Offset:	-1.280000	-1.500000
NO Cal Slope:	0.999572	1.000428
NO Cal Offset:	-2.180000	-2.080000
NO <sub>2</sub> Cal Slope:	1.004664	0.999045
NO <sub>2</sub> Cal Offset:	0.067495	0.023464
NH3 Cal Slope:	0.998759	1.002916
NH3 Cal Offset:	1.195770	-5.175346
TN Cal Slope:	1.001474	1.005585
TN Cal Offset:	1.255511	-5.152839



# Wood Buffalo Environmental Association

## TN - NOX - NH<sub>3</sub> Calibration Report

Version-05-2023

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
as found NO	4920	80.0	813.4	813.4	----	813.3	812.8	0.5	1.000	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
high NO point	4920	80.0	813.4	813.4	----	812.0	813.0	-0.8	1.002	----
NO/O3 point	4920	80.0	813.4	813.4	----	811.4	809.3	2.4	1.003	----
as found NH3	3413	86.4	1800.6	----	1800.6	1796.1	----	1791.3	1.003	1.005
new NH3 cyl rp	3416	84.1	1799.7	----	1799.7	1862.2	----	1858.0	0.966	0.969
first NH3	3416	84.1	1799.7	----	1799.7	1805.9	----	1801.2	0.997	0.999
second NH3	3453	46.7	999.4	----	999.4	1001.6	----	998.6	0.998	1.001
third NH3	3477	23.4	500.8	----	500.8	490.3	----	489.1	1.021	1.024
<b>Average Correction Factor</b>									<b>1.0021</b>	<b>1.0079</b>

Corrected As found    TN = 813.5 ppb    NO<sub>x</sub> = 813.1 ppb    NH3 = 1791.2 ppb

Previous Response    TN = 815.9 ppb    NO<sub>x</sub> = 810.7 ppb    NH3 = 1799.6 ppb

NH3 Previous Converter Efficiency = 94.1%

NH3 Current Converter Efficiency = 97.3%

\*Percent Change    TN = -0.3%

\*Percent Change    NO<sub>x</sub> = 0.3%

\*Percent Change    NH3 = -0.5%

\* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## NO<sub>x</sub> - NO - NO<sub>2</sub> Calibration Report

Version-05-2023

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.5	-1.3	-0.6	----	----
as found span	4920	80.0	813.4	800.6	813.4	814.3	801.3	818.1	0.9989	0.9992
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	-0.1	----	----
high point	4920	80.0	813.4	800.6	813.4	813.0	800.0	812.0	1.0005	1.0008
second point	4960	40.0	406.7	400.3	406.7	403.9	397.2	404.6	1.0070	1.0079
third point	4980	20.0	203.4	200.2	203.4	200.8	196.2	200.9	1.0127	1.0202
Average Correction Factor									1.0068	1.0096

Baseline Corr As fnd	TN = 818.7 ppb	NO <sub>x</sub> = 814.8 ppb	NO = 802.6 ppb	*Percent Change	TN = 0.3%
Previous Response	TN = 815.9 ppb	NO <sub>x</sub> = 810.7 ppb	NO = 798.1 ppb	*Percent Change	NO <sub>x</sub> = 0.5%
				*Percent Change	NO = 0.6%
				<i>* = &gt; +/-5% change initiates investigation</i>	

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	0.8	----	----
calibration zero	----	----	0.0	0.0	----	----
1st GPT point (400 ppb O3)	795.4	398.1	410.1	409.7	1.0010	99.9%
2nd GPT point (200 ppb O3)	795.4	602.8	205.4	205.3	1.0005	100.0%
3rd GPT point (100 ppb O3)	795.4	697.4	110.8	110.7	1.0009	99.9%
Average Correction Factor					1.0008	99.9%

Notes: Changed the inlet filter after as founds. Adjusted the NO<sub>x</sub>/TN zero and span. Used the 2nd GPT reference point due to drift. Changed out the NH<sub>3</sub> cylinder. Adjusted the NH<sub>3</sub> span.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## TN Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 21, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:50	End Time (MST):	14:38
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999946	≥0.995
1799.7	1805.9	0.9966			
999.4	1001.6	0.9978	Slope	1.005585	0.90 - 1.10
500.8	490.3	1.0213			
			Intercept	-5.152839	+/-20





# Wood Buffalo Environmental Association

## NH<sub>3</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 21, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:50	End Time (MST):	14:38
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1799.7	1801.2	0.9992		
999.4	998.6	1.0008		
500.8	489.1	1.0238		
			0.999948	
			1.002916	
			-5.175346	





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	July 12, 2023	Previous Calibration:	June 21, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:50	End Time (MST):	14:38
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.4	813.0	1.0005		
406.7	403.9	1.0070		
203.4	200.8	1.0127		
			0.999983	
			1.000126	
			-1.500000	





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-05-2023


### Station Information

Calibration Date:	July 12, 2023	Previous Calibration:	June 21, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:50	End Time (MST):	14:38
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ 0.90 - 1.10 +/-20
800.6	800.0	1.0008		
400.3	397.2	1.0079		
200.2	196.2	1.0202		
			0.999969	
			1.000428	
			-2.080000	

**NO Calibration Curve**







# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	July 12, 2023	Previous Calibration:	June 21, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:50	End Time (MST):	14:38
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
410.1	409.7	1.0010			
205.4	205.3	1.0005			
110.8	110.7	1.0009			
			Slope	0.999045	0.90 - 1.10
			Intercept	0.023464	+/-20



# NO<sub>x</sub> Calibration Plot

Date: July 12, 2023


Location: Bertha Ganter-Fort McKay



# NH<sub>3</sub> Calibration Plot

Date: July 13, 2023

Location: Bertha Ganter-Fort McKay





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS02 MILDRED LAKE JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	July 12, 2023	Last Cal Date:	June 29, 2023
Start time (MST):	9:25	End time (MST):	12:37
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1185
ZAG Make/Model:	API T750		Serial Number:	4891

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999231	1.000728	Backgd or Offset:	17.8	17.8
Calibration intercept:	-0.605160	-1.304958	Coeff or Slope:	0.783	0.783

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4920	80.2	801.6	800.0	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.2	801.6	802.0	1.000
second point	4960	40.1	400.8	397.9	1.007
third point	4980	20.0	199.9	198.3	1.008
as left zero	5000	0.0	0.0	-0.1	----
as left span	4920	80.2	801.6	795.0	1.008
Average Correction Factor:					1.005

Baseline Corr As found:	800.20	Previous response	800.43	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 12, 2023	Previous Calibration:	June 29, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:25	End Time (MST):	12:37
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999981	
801.6	802.0	0.9996			≥0.995
400.8	397.9	1.0074	Slope	1.000728	
199.9	198.3	1.0082			0.90 - 1.10
			Intercept	-1.304958	+/-30



SO2 Calibration Plot

Date: July 12, 2023

Location: Mildred Lake





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Mildred Lake      Station number: AMS02  
 Calibration Date: July 20, 2023      Last Cal Date: June 29, 2023  
 Start time (MST): 9:17      End time (MST): 15:25  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.29 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC345191  
 Removed Cal Gas Conc: 5.29 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 1185  
 ZAG Make/Model: API T701      Serial Number: 5608

### Analyzer Information

Analyzer make: Thermo 43iQTL      Analyzer serial #: 12113311966  
 Converter make: Global G150      Converter serial #: 2022-198  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006394	0.994964	Backgd or Offset: 1.72	1.77
Calibration intercept:	-0.119196	-0.019203	Coeff or Slope: 0.797	0.821

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4924	75.6	80.0	79.0	1.013
as found 2nd point	4962	37.8	40.0	39.1	1.023
as found 3rd point	4981	18.9	20.0	19.4	1.031
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4924	75.6	80.0	79.5	1.006
second point	4962	37.8	40.0	40.0	1.000
third point	4981	18.9	20.0	19.7	1.015
as left zero	5000	0.0	0.0	0.2	----
as left span	4924	75.6	80.0	78.1	1.024
SO2 Scrubber Check	4920	80.2	802.0	0.1	----
Date of last scrubber change:	12-Sep-22			Ave Corr Factor	1.007
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.0      Prev response: 80.38      \*% change: -1.8%  
 Baseline Corr 2nd AF pt: 39.1      AF Slope: 0.988535      AF Intercept: -0.219214  
 Baseline Corr 3rd AF pt: 19.4      AF Correlation: 0.999960

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after MPAF's. Scrubber check passed after MPAF's. Adjusted span.

Calibration Performed By: Braiden Boutillier





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 20, 2023	Previous Calibration:	June 29, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:17	End Time (MST):	15:25
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999975	≥0.995
80.0	79.5	1.0062			
40.0	40.0	0.9998	Slope	0.994964	0.90 - 1.10
20.0	19.7	1.0151			
			Intercept	-0.019203	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 20, 2023

Location: Mildred Lake





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	July 12, 2023	Last Cal Date:	June 29, 2023
Start time (MST):	9:25	End time (MST):	12:37
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH <sub>4</sub> Cal Gas Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1048.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1048.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.88E-04	2.97E-04	NMHC SP Ratio:	4.52E-05
CH <sub>4</sub> Retention time:	14.6	14.8	NMHC Peak Area:	194709
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	16.48	1.021
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.82	1.000
second point	4960	40.1	8.41	8.38	1.003
third point	4980	20.0	4.19	4.17	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.84	0.999
Average Correction Factor					1.003

Baseline Corr AF:	16.48	Prev response	16.84	*% change	-2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	8.69	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.80	0.999
second point	4960	40.1	4.40	4.40	1.000
third point	4980	20.0	2.19	2.21	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.81	0.998
Average Correction Factor					0.998
Baseline Corr AF:	8.69	Prev response	8.84	*% change	-1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	7.79	1.030
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.02	1.000
second point	4960	40.1	4.01	3.98	1.008
third point	4980	20.0	2.00	1.97	1.016
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	8.04	0.998
Average Correction Factor					1.008
Baseline Corr AF:	7.79	Prev response	8.01	*% change	-2.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001404	1.000332
THC Cal Offset:	0.001307	-0.014309
CH <sub>4</sub> Cal Slope:	0.999914	1.000556
CH <sub>4</sub> Cal Offset:	-0.014249	-0.018253
NMHC Cal Slope:	1.002789	1.000114
NMHC Cal Offset:	0.015957	0.003745

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	July 12, 2023	Previous Calibration:	June 29, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:25	End Time (MST):	12:37
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	$\geq 0.995$			
16.82	16.82	0.9999						
8.41	8.38	1.0035				Slope	1.000332	0.90 - 1.10
4.19	4.17	1.0048						
			Intercept	-0.014309	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	July 12, 2023	Previous Calibration:	June 29, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:25	End Time (MST):	12:37
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999975	$\geq 0.995$
8.02	8.02	1.0004			
4.01	3.98	1.0077			
2.00	1.97	1.0162			
			Slope	1.000556	0.90 - 1.10
			Intercept	-0.018253	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	July 12, 2023	Previous Calibration:	June 29, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:25	End Time (MST):	12:37
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	$\geq 0.995$			
8.80	8.80	0.9995						
4.40	4.40	0.9999				Slope	1.000114	0.90 - 1.10
2.19	2.21	0.9947						
			Intercept	0.003745	$\pm 0.5$			



NMHC Calibration Plot

Date: July 12, 2023

Location: Mildred Lake







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	July 20, 2023	Last Cal Date:	July 12, 2023
Start time (MST):	15:00	End time (MST):	16:34
Reason:	Cylinder Change		

### Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH <sub>4</sub> Cal Gas Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1048.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1048.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.97E-04	2.97E-04	NMHC SP Ratio:	4.58E-05
CH <sub>4</sub> Retention time:	14.8	14.8	NMHC Peak Area:	192143
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	17.00	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	17.28	0.973
Average Correction Factor					

Baseline Corr AF:	17.00	Prev response	16.81	*% change	1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	8.93	0.985
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	9.07	0.970
<b>Average Correction Factor</b>					
Baseline Corr AF:	8.93	Prev response	8.80	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	8.07	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	8.21	0.977
<b>Average Correction Factor</b>					
Baseline Corr AF:	8.07	Prev response	8.01	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000332	
THC Cal Offset:	-0.014309	
CH <sub>4</sub> Cal Slope:	1.000556	
CH <sub>4</sub> Cal Offset:	-0.018253	
NMHC Cal Slope:	1.000114	
NMHC Cal Offset:	0.003745	


Notes: Changed Hydrogen cylinder.

Calibration Performed By: Braiden Boutillier

NMHC Calibration Plot

Date: July 20, 2023

Location: Mildred Lake





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS04 BUFFALO VIEWPOINT JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	July 19, 2023	Last Cal Date:	June 9, 2023
Start time (MST):	6:42	End time (MST):	10:02
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.87	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC446753			
Removed Cal Gas Conc:	50.02	ppm	Rem Gas Exp Date:	September 9, 2028
Removed Gas Cyl #:	CC470284		Diff between cyl:	-0.8%
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701		Serial Number:	362

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001756	1.000328	Backgd or Offset:	22.1	22.1
Calibration intercept:	0.520000	0.714976	Coeff or Slope:	0.860	0.860

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4920	80.0	800.3	806.0	0.993
as found 2nd point					
as found 3rd point					
new cylinder response	4921	78.6	799.7	798.6	1.001
calibrator zero	5000	0.0	0.0	0.3	----
high point	4921	78.6	799.7	800.5	0.999
second point	4961	39.3	399.8	400.9	0.997
third point	4980	19.6	199.4	200.6	0.994
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	78.6	799.7	796.8	1.004
Average Correction Factor					0.997

Baseline Corr As found:	805.80	Previous response	802.25	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Calibration gas changed. No adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 19, 2023	Previous Calibration:	June 9, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:42	End Time (MST):	10:02
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995
799.7	800.5	0.9991			
399.8	400.9	0.9973	Slope	1.000328	0.90 - 1.10
199.4	200.6	0.9941			
			Intercept	0.714976	+/-30



SO2 Calibration Plot

Date: July 19, 2023

Location: Buffalo Viewpoint







# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Buffalo Viewpoint      Station number: AMS04  
 Calibration Date: July 11, 2023      Last Cal Date: June 22, 2023  
 Start time (MST): 5:45      End time (MST): 9:43  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.42 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC345266  
 Removed Cal Gas Conc: 5.42 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3808  
 ZAG Make/Model: API T701H      Serial Number: 362

### Analyzer Information

Analyzer make: Thermo 43i-LTE      Analyzer serial #: 1008841400  
 Converter make: Global      Converter serial #: 2022-200  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001341	0.994226	Backgd or Offset:	1.8	1.8
Calibration intercept:	-0.017797	-0.017926	Coeff or Slope:	1.095	1.095

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.1	80.3	80.0	1.004
as found 2nd point	4963	37.0	40.1	40.1	1.000
as found 3rd point	4982	18.5	20.1	19.8	1.013
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4926	74.1	80.3	79.8	1.007
second point	4963	37.0	40.1	40.0	1.003
third point	4982	18.5	20.1	19.8	1.013
as left zero	5000	0.0	0.0	0.0	----
as left span	4926	74.1	80.3	79.4	1.012
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	1.007
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 80.0      Prev response: 80.41      \*% change: -0.5%  
 Baseline Corr 2nd AF pt: 40.1      AF Slope: 0.996930      AF Intercept: -0.037869  
 Baseline Corr 3rd AF pt: 19.8      AF Correlation: 0.999986

\* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 22, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	5:45	End Time (MST):	9:43
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999990	≥0.995
80.3	79.8	1.0066			
40.1	40.0	1.0027	Slope	0.994226	0.90 - 1.10
20.1	19.8	1.0127			
			Intercept	-0.017926	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 11, 2023

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	July 19, 2023	Last Cal Date:	June 9, 2023
Start time (MST):	6:42	End time (MST):	10:01
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH <sub>4</sub> Cal Gas Conc.	497.2 ppm	CH <sub>4</sub> Equiv Conc.	1058.2 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:	CC470284	Removed Gas Expiry:	September 9, 2028
Removed CH <sub>4</sub> Conc.	497.8 ppm	CH <sub>4</sub> Equiv Conc.	1062.9 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.5 ppm	Diff between cyl (THC):	0.0%
Diff between cyl (CH <sub>4</sub> ):	0.0%	Diff between cyl (NM):	0.0%
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	1.860E-04	1.840E-04	NMHC SP Ratio:	3.870E-05
CH <sub>4</sub> Retention time:	11.8	11.8	NMHC Peak Area:	233712
				225808

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.01	16.97	1.002
as found 2nd point					
as found 3rd point					
new cylinder response	4921	78.6	16.64	16.60	1.002
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	16.64	16.64	1.000
second point	4961	39.3	8.32	8.32	1.000
third point	4980	19.6	4.15	4.14	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	16.64	16.62	1.001

Average Correction Factor				1.000
Baseline Corr AF:	16.97	Prev response	17.09	*% change -0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	
as found span	4920	80.0	9.04	8.94	1.011
as found 2nd point					
as found 3rd point					
new cylinder response	4921	78.6	8.82	8.72	1.011
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	8.82	8.82	1.000
second point	4961	39.3	4.41	4.42	0.998
third point	4980	19.6	2.20	2.20	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	8.82	8.81	1.001
Average Correction Factor					0.999
Baseline Corr AF:	8.94	Prev response	9.04	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	7.96	8.03	0.992
as found 2nd point					
as found 3rd point					
new cylinder response	4921	78.6	7.82	7.88	0.992
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	7.82	7.83	0.998
second point	4961	39.3	3.91	3.90	1.002
third point	4980	19.6	1.95	1.94	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	7.82	7.81	1.001
Average Correction Factor					1.002
Baseline Corr AF:	8.03	Prev response	8.05	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.005850	1.000429
THC Cal Offset:	-0.014000	-0.003535
CH <sub>4</sub> Cal Slope:	1.010733	1.002052
CH <sub>4</sub> Cal Offset:	-0.002000	-0.007900
NMHC Cal Slope:	1.001548	1.000156
NMHC Cal Offset:	-0.012000	0.002367

Notes: Span Adjusted. Calibration gas changed.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 19, 2023	Previous Calibration:	June 9, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:42	End Time (MST):	10:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	$\geq 0.995$			
16.64	16.64	0.9998						
8.32	8.32	0.9996				Slope	1.000429	0.90 - 1.10
4.15	4.14	1.0020						
			Intercept	-0.003535	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 19, 2023	Previous Calibration:	June 9, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:42	End Time (MST):	10:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	≥0.995
7.82	7.83	0.9983			
3.91	3.90	1.0020			
1.95	1.94	1.0047			
			Slope	1.002052	0.90 - 1.10
			Intercept	-0.007900	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 19, 2023	Previous Calibration:	June 9, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:42	End Time (MST):	10:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	$\geq 0.995$			
8.82	8.82	1.0000						
4.41	4.42	0.9976				Slope	1.000156	0.90 - 1.10
2.20	2.20	0.9997						
			Intercept	0.002367	$\pm 0.5$			






NMHC Calibration Plot

Date: July 19, 2023

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Buffalo Viewpoint  
Calibration Date: July 10, 2023  
Start time (MST): 6:20  
Reason: Routine

Station number: AMS04  
Last Cal Date: June 5, 2023  
End time (MST): 11:37

### Calibration Standards

NO Gas Cylinder #: T36RH1F  
NOX Cal Gas Conc: 51.16 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 51.16 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701

Cal Gas Expiry Date: August 18, 2023  
NO Cal Gas Conc: 50.91 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 50.91 ppm  
NO gas Diff:  
Serial Number: 3808  
Serial Number: 362

### Analyzer Information

Analyzer make: API T200  
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 723

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.146	1.194	NO bkgnd or offset:	-2.6	-2.7
NOX coeff or slope:	1.149	1.188	NOX bkgnd or offset:	-2.7	-1.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	7.5	7.5

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.001047	0.997031
NO <sub>x</sub> Cal Offset:	1.266147	0.445643
NO Cal Slope:	0.997960	1.000259
NO Cal Offset:	0.425945	-0.474006
NO <sub>2</sub> Cal Slope:	1.007354	0.987013
NO <sub>2</sub> Cal Offset:	1.181889	-0.851422



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
as found span	4922	78.1	799.1	795.2	3.9	772.4	761.9	10.4	1.0346	1.0437
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	0.4	-0.5	----	----
high point	4922	78.1	799.1	795.2	3.9	796.5	795.2	1.4	1.0033	1.0000
second point	4961	39.1	400.1	398.1	2.0	400.7	397.8	2.8	0.9984	1.0008
third point	4981	19.5	199.5	198.5	1.0	199.3	196.9	2.5	1.0010	1.0083
as left zero	5000	0.0	0.0	0.0	0.0	-0.5	0.2	-0.7	----	----
as left span	4922	78.1	799.1	795.2	3.9	794.8	782.7	412.1	1.0054	0.9927
Average Correction Factor									1.0009	1.0030

Corrected As found	NO <sub>x</sub> = 772.7 ppb	NO = 762.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -3.7%
Previous Response	NO <sub>x</sub> = 801.2 ppb	NO = 794.0 ppb		*Percent Change	NO = -4.2%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.7	379.4	419.2	413.1	1.0148	98.5%
2nd GPT point (200 ppb O3)	794.7	586.8	211.8	208.0	1.0183	98.2%
3rd GPT point (100 ppb O3)	794.7	689.7	108.9	106.3	1.0245	97.6%
Average Correction Factor					1.0192	98.1%

Notes:

Zero and Span adjusted. No Maintenance done.

Calibration Performed By:

Melissa Lemay



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	June 5, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:20	End Time (MST):	11:37
Analyzer make:	API T200	Analyzer serial #:	723

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
799.1	796.5	1.0033			
400.1	400.7	0.9984			
199.5	199.3	1.0010			
			Slope	0.997031	0.90 - 1.10
			Intercept	0.445643	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	June 5, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:20	End Time (MST):	11:37
Analyzer make:	API T200	Analyzer serial #:	723

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.4	----	Correlation Coefficient	≥0.995	
795.2	795.2	1.0000			
398.1	397.8	1.0008			
198.5	196.9	1.0083			
			Slope	1.000259	0.90 - 1.10
			Intercept	-0.474006	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	June 5, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:20	End Time (MST):	11:37
Analyzer make:	API T200	Analyzer serial #:	723

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.5	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
419.2	413.1	1.0148		
211.8	208.0	1.0183		
108.9	106.3	1.0245		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: July 10, 2023

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	July 26, 2023	Last Cal Date:	July 10, 2023
Start time (MST):	6:50	End time (MST):	10:35
Reason:	Removal		unstable Zero

### Calibration Standards

NO Gas Cylinder #:	T36RH1F	Cal Gas Expiry Date:	August 18, 2023
NOX Cal Gas Conc:	51.16 ppm	NO Cal Gas Conc:	50.91 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	51.16 ppm	Removed Gas NO Conc:	50.91 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

### Analyzer Information

Analyzer make:	API T200	Analyzer serial #:	723
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.194	1.194	NO bkgnd or offset:	-2.7	-2.7
NOX coeff or slope:	1.188	1.188	NOX bkgnd or offset:	-1.6	-1.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	7.4	7.4

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.997031	
NO <sub>x</sub> Cal Offset:	0.445643	
NO Cal Slope:	1.000259	
NO Cal Offset:	-0.474006	
NO <sub>2</sub> Cal Slope:	0.987013	
NO <sub>2</sub> Cal Offset:	-0.851422	





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.6	0.1	----	----
as found span	4922	78.1	799.1	795.2	3.9	757.2	754.0	3.2	1.0553	1.0546
as found 2nd	4961	39.1	400.1	398.1	2.0	381.2	375.9	5.3	1.0495	1.0591
as found 3rd	4981	19.5	199.5	198.5	1.0	190.6	186.8	3.8	1.0467	1.0628
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

### Average Correction Factor

Corrected As found	NO <sub>x</sub> = 757.7 ppb	NO = 754.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -5.2%
Previous Response	NO <sub>x</sub> = 797.2 ppb	NO = 794.9 ppb		*Percent Change	NO = -5.3%
Baseline Corr 2nd pt	NO <sub>x</sub> = 381.7 ppb	NO = 376.5 ppb	As found	NO <sub>x</sub> r <sup>2</sup> : 0.999985	Nx SI: 0.947698
Baseline Corr 3rd pt	NO <sub>x</sub> = 191.1 ppb	NO = 187.4 ppb	As found	NO r <sup>2</sup> : 0.999996	NO SI: 0.949189
			As found	NO <sub>2</sub> r <sup>2</sup> : 0.999997	NO <sub>2</sub> SI: 0.987921
					NO <sub>2</sub> Int: 0.328

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	0.1	----	----
as found GPT point (400 ppb NO2)	753.8	360.8	396.9	392.4	1.0115	98.9%
as found GPT point (200 ppb NO2)	753.8	557.6	200.1	197.9	1.0111	98.9%
as found GPT point (100 ppb NO2)	753.8	657.1	100.6	100.1	1.0050	99.5%
1st GPT point (400 ppb O3)						
2nd GPT point (200 ppb O3)						
3rd GPT point (100 ppb O3)						

### Average Correction Factor

Notes:

Removed due to unstable zero

Calibration Performed By:

Melissa Lemay



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 26, 2023	Previous Calibration:	July 10, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:50	End Time (MST):	10:35
Analyzer make:	API T200	Analyzer serial #:	723

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
			Correlation Coefficient	≥0.995
			Slope	0.90 - 1.10
			Intercept	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 26, 2023	Previous Calibration:	July 10, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:50	End Time (MST):	10:35
Analyzer make:	API T200	Analyzer serial #:	723

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
			Correlation Coefficient	≥0.995
			Slope	0.90 - 1.10
			Intercept	+/-20





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 26, 2023	Previous Calibration:	July 10, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:50	End Time (MST):	10:35
Analyzer make:	API T200	Analyzer serial #:	723

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
			Correlation Coefficient	≥0.995
			Slope	0.90 - 1.10
			Intercept	+/-20



NO<sub>x</sub> Calibration Plot

Date: July 26, 2023

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Buffalo Viewpoint  
Calibration Date: July 27, 2023  
Start time (MST): 6:00  
Reason: Install  
Station number: AMS04  
Last Cal Date:  
End time (MST): 10:51

### Calibration Standards

NO Gas Cylinder #: T36RH1F  
NOX Cal Gas Conc: 51.16 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 51.16 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701  
Cal Gas Expiry Date: August 18, 2023  
NO Cal Gas Conc: 50.91 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 50.91 ppm  
NO gas Diff:  
Serial Number: 3808  
Serial Number: 362

### Analyzer Information

Analyzer make: API T200  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 721

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:		1.071	NO bkgnd or offset:		-0.6
NOX coeff or slope:		1.066	NOX bkgnd or offset:		-0.3
NO2 coeff or slope:		1.000	Reaction cell Press:		4.5

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.997031	1.002547
NO <sub>x</sub> Cal Offset:	0.445643	-0.233343
NO Cal Slope:	1.000259	1.001796
NO Cal Offset:	-0.474006	-1.333988
NO <sub>2</sub> Cal Slope:	0.987013	0.990498
NO <sub>2</sub> Cal Offset:	-0.851422	0.120026



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.0	0.4	----	----
high point	4922	78.1	799.1	795.2	3.9	801.3	795.8	5.5	0.9973	0.9992
second point	4961	39.1	400.1	398.1	2.0	400.2	397.3	2.9	0.9997	1.0020
third point	4981	19.5	199.5	198.5	1.0	199.5	195.9	3.6	1.0000	1.0134
as left zero	5000	0.0	0.0	0.0	0.0	0.5	0.6	0.0	----	----
as left span	4922	78.1	799.1	381.9	417.2	791.7	381.6	410.1	1.0094	1.0008
Average Correction Factor									0.9990	1.0049

Corrected As found	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	<i>* =&gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = NA
Previous Response	NO <sub>x</sub> = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
					As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.1	380.8	417.2	413.6	1.0087	99.1%
2nd GPT point (200 ppb O3)	794.1	587.9	210.1	207.8	1.0111	98.9%
3rd GPT point (100 ppb O3)	794.1	689.2	108.8	107.8	1.0093	99.1%
Average Correction Factor					1.0097	99.0%

Notes: Installed due to old analyzer having an unstable zero. Span adjusted.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:00	End Time (MST):	10:51
Analyzer make:	API T200	Analyzer serial #:	721

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.1	801.3	0.9973		
400.1	400.2	0.9997		
199.5	199.5	1.0000		
			0.999997	
			1.002547	
			-0.233343	







# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:00	End Time (MST):	10:51
Analyzer make:	API T200	Analyzer serial #:	721

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
795.2	795.8	0.9992		
398.1	397.3	1.0020		
198.5	195.9	1.0134		
			0.999986	
			1.001796	
			-1.333988	





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:00	End Time (MST):	10:51
Analyzer make:	API T200	Analyzer serial #:	721

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
417.2	413.6	1.0087		
210.1	207.8	1.0111		
108.8	107.8	1.0093		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: July 27, 2023

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Buffalo Viewpoint      Station number: AMS04  
 Calibration Date: July 20, 2023      Last Cal Date: June 23, 2023  
 Start time (MST): 6:25      End time (MST): 8:52  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: API T700      Serial Number: 3808  
 ZAG Make/Model: API T701      Serial Number: 362

### Analyzer Information

Analyzer make: API T400      Analyzer serial #: 2961  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000343	0.996714	Backgd or Offset:	-2.6	-2.6
Calibration intercept:	-0.060000	0.200000	Coeff or Slope:	1.008	1.008

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	5000	985.2	400.0	397.4	1.007
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.0	----
high point	5000	984.3	400.0	398.7	1.003
second point	5000	816.3	200.0	199.9	1.001
third point	5000	708.4	100.0	99.9	1.001
as left zero	5000	0.0	0.0	0.0	----
as left span	5000	985.9	400.0	399.6	1.001
Average Correction Factor					1.002

Baseline Corr As found:	397.8	Previous response	400.1	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No Maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 20, 2023	Previous Calibration:	June 23, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:25	End Time (MST):	8:52
Analyzer make:	API T400	Analyzer serial #:	2961

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999998	
400.0	398.7	1.0033			≥0.995
200.0	199.9	1.0005	Slope	0.996714	
100.0	99.9	1.0010			0.90 - 1.10
			Intercept	0.200000	+/- 5



O<sub>3</sub> Calibration Plot

Date: July 20, 2023

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04  
 Calibration Date: July 11, 2023 Last Cal Date: June 23, 2023  
 Start time (MST): 9:43 End time (MST): 10:10

Analyzer Make: API T640 S/N: 322  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Deltacal S/N: 1451  
 Temp/RH standard: Deltacal S/N: 1451

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	20.5	20.6	20.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	728	726.6	728	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	4.90	5.02	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: July 11, 2023 Last Cal Date: June 23, 2023  
 PM w/o HEPA: 3.9 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: \_\_\_\_\_ w/ HEPA: \_\_\_\_\_  
 Date Optical Chamber Cleaned: May 23, 2023 <0.2 ug/m3  
 Disposable Filter Changed: May 23, 2023

### Annual Maintenance

Date Sample Tube Cleaned: May 23, 2023  
 Date RH/T Sensor Cleaned: May 23, 2023

No adjustments done. Head cleaned.

Notes:

Calibration by: Melissa Lemay



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS05  
MANNIX  
JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Mannix	Station number:	AMS05
Calibration Date:	July 11, 2023	Last Cal Date:	June 15, 2023
Start time (MST):	9:32	End time (MST):	13:10
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.02	ppm	Cal Gas Exp Date:	January 12, 2029
Cal Gas Cylinder #:	XC026809B			
Removed Cal Gas Conc:	50.02	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	621
ZAG Make/Model:	API T701H		Serial Number:	832

### Analyzer Information

Analyzer make: Thermo 43i	Analyzer serial #: 1008841399
Analyzer Range 0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000214	1.000771	Backgd or Offset:	8.9	9.0
Calibration intercept:	-0.540000	0.040000	Coeff or Slope:	0.908	0.941

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4920	80.0	800.3	775.6	1.032
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4920	80.0	800.3	801.1	0.999
second point	4960	40.0	400.2	400.5	0.999
third point	4980	20.0	200.1	199.8	1.001
as left zero	5000	0.0	0.0	0.6	----
as left span	4920	80.0	800.3	799.4	1.001
Average Correction Factor					1.000

Baseline Corr As found:	775.10	Previous response	799.95	*% change	-3.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 15, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:32	End Time (MST):	13:10
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.4	----	Correlation Coefficient	0.999999	
800.3	801.1	0.9990			≥0.995
400.2	400.5	0.9992	Slope	1.000771	
200.1	199.8	1.0014			0.90 - 1.10
			Intercept	0.040000	+/-30



SO2 Calibration Plot

Date: July 11, 2023

Location: Mannix





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Mannix Station number: AMS05  
 Calibration Date: July 13, 2023 Last Cal Date: June 9, 2023  
 Start time (MST): 9:03 End time (MST): 13:38  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 4.92 ppm Cal Gas Exp Date: February 9, 2024  
 Cal Gas Cylinder #: EY0002433  
 Removed Cal Gas Conc: 4.92 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 1845  
 ZAG Make/Model: API T701H Serial Number: 832

### Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1203169745  
 Converter make: Global Converter serial #: 2022-196  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003756	0.996187	Backgd or Offset:	2.07 2.13
Calibration intercept:	0.040615	0.180584	Coeff or Slope:	0.821 0.842

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4919	81.3	80.0	79.1	1.013
as found 2nd point	4960	40.7	40.0	39.5	1.016
as found 3rd point	4980	20.3	20.0	19.7	1.019
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4919	81.3	80.0	79.8	1.002
second point	4960	40.7	40.0	40.2	0.996
third point	4980	20.3	20.0	20.1	0.994
as left zero	5000	0.0	0.0	0.5	----
as left span	4919	81.3	80.0	77.7	1.030
SO2 Scrubber Check	4920	80.0	800.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.997
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 79.0 Prev response: 80.34 \*% change: -1.7%  
 Baseline Corr 2nd AF pt: 39.4 AF Slope: 0.987899 AF Intercept: 0.020656  
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999995

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 9, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:03	End Time (MST):	13:38
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1203169745

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999992	≥0.995
80.0	79.8	1.0024			
40.0	40.2	0.9961	Slope	0.996187	0.90 - 1.10
20.0	20.1	0.9937			
			Intercept	0.180584	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 13, 2023

Location: Mannix





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Mannix	Station number:	AMS05
Calibration Date:	July 11, 2023	Last Cal Date:	June 15, 2023
Start time (MST):	9:32	End time (MST):	13:10
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH <sub>4</sub> Cal Gas Conc.	504.9 ppm	CH <sub>4</sub> Equiv Conc.	1076.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	504.9 ppm	CH <sub>4</sub> Equiv Conc.	1076.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.66E-04	2.60E-05	NMHC SP Ratio:	4.43E-05
CH <sub>4</sub> Retention time:	15.20	15.00	NMHC Peak Area:	206739
				210847

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.55	0.981
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.22	1.001
second point	4960	40.0	8.61	8.63	0.998
third point	4980	20.0	4.31	4.31	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.30	0.996

Average Correction Factor				1.000
Baseline Corr AF:	17.55	Prev response	17.25	*% change 1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4920	80	9.15	9.32	0.981
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	80	9.15	9.13	1.002
second point	4960	40	4.57	4.59	0.997
third point	4980	20	2.29	2.30	0.995
as left zero	5000	0	0.00	0.00	----
as left span	4920	80	9.15	9.20	0.995
Average Correction Factor					0.998
Baseline Corr AF:	9.32	Prev response	9.17	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	8.23	0.982
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.09	0.999
second point	4960	40.0	4.04	4.04	1.001
third point	4980	20.0	2.02	2.01	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.11	0.996
Average Correction Factor					1.002
Baseline Corr AF:	8.23	Prev response	8.09	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001244	0.999419
THC Cal Offset:	0.006000	0.005000
CH <sub>4</sub> Cal Slope:	1.001712	1.001938
CH <sub>4</sub> Cal Offset:	-0.006600	-0.007400
NMHC Cal Slope:	1.000793	0.997195
NMHC Cal Offset:	0.013000	0.012400

Notes: Changed the inlet filter and the H2 cylinder after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell





# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 15, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:32	End Time (MST):	13:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	$\geq 0.995$			
17.23	17.22	1.0006						
8.61	8.63	0.9985				Slope	0.999419	0.90 - 1.10
4.31	4.31	0.9999						
			Intercept	0.005000	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 15, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:32	End Time (MST):	13:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	$\geq 0.995$			
8.08	8.09	0.9986						
4.04	4.04	1.0005				Slope	1.001938	0.90 - 1.10
2.02	2.01	1.0058						
			Intercept	-0.007400	$\pm 0.5$			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 15, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:32	End Time (MST):	13:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999989	$\geq 0.995$
9.15	9.13	1.0025			
4.57	4.59	0.9967			
2.29	2.30	0.9947			
			Slope	0.997195	0.90 - 1.10
			Intercept	0.012400	+/-0.5



NMHC Calibration Plot

Date: July 11, 2023

Location: Mannix





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS06**  
**PATRICIA MCINNES**  
**JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	July 4, 2023	Last Cal Date:	June 5, 2023
Start time (MST):	9:20	End time (MST):	12:31
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.78	ppm	Cal Gas Exp Date:	September 9, 2024
Cal Gas Cylinder #:	AAL070632			
Removed Cal Gas Conc:	49.78	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	5608

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003750	1.006537	Backgd or Offset:	17.2	17.2
Calibration intercept:	1.159413	1.758830	Coeff or Slope:	0.901	0.901

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4920	80.3	799.5	803.4	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.3	799.5	805.7	0.992
second point	4960	40.2	400.2	405.1	0.988
third point	4980	20.1	200.1	205.2	0.975
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	80.3	799.5	804.0	0.994
Average Correction Factor					0.985

Baseline Corr As found:	803.60	Previous response	803.62	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Ryan Power



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 5, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:20	End Time (MST):	12:31
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999977	
799.5	805.7	0.9923			≥0.995
400.2	405.1	0.9880	Slope	1.006537	
200.1	205.2	0.9752			0.90 - 1.10
			Intercept	1.758830	+/-30



SO2 Calibration Plot

Date: July 4, 2023

Location: Patricia McInnes







# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	July 12, 2023	Last Cal Date:	June 12, 2023
Start time (MST):	8:40	End time (MST):	15:05
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.38	ppm	Cal Gas Exp Date:	March 2, 2023
Cal Gas Cylinder #:	EY0000809			
Removed Cal Gas Conc:	5.38	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

### Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994772	1.004924	Backgd or Offset:	1.92
Calibration intercept:	0.337244	0.356984	Coeff or Slope:	1.116
				2.06
				1.187

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4926	74.3	79.9	75.6	1.059
as found 2nd point	4963	37.2	40.0	38.3	1.048
as found 3rd point	4981	18.6	20.0	19.2	1.048
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4926	74.3	79.9	80.6	0.992
second point	4963	37.2	40.0	40.7	0.983
third point	4981	18.6	20.0	20.6	0.972
as left zero	5000	0.0	0.0	0.6	----
as left span	4926	74.3	79.9	79.6	1.004
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	0.982
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	75.5	Prev response:	79.86	*% change:	-5.8%
Baseline Corr 2nd AF pt:	38.2	AF Slope:	0.944164	AF Intercept:	0.258323
Baseline Corr 3rd AF pt:	19.1	AF Correlation:	0.999965		

\* = > +/-5% change initiates investigation

Notes: Problem with the zero air during as founds, swapped out the zero air generator and continued the calibration. As founds are low but diagnostics are normal. Changed the inlet filter after as founds.

Ran a SO2 scrubber check after calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021


### Station Information

Calibration Date:	July 12, 2023	Previous Calibration:	June 12, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:40	End Time (MST):	15:05
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999982	≥0.995
79.9	80.6	0.9918			
40.0	40.7	0.9834	Slope	1.004924	0.90 - 1.10
20.0	20.6	0.9716			
			Intercept	0.356984	+/-3







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	July 4, 2023	Last Cal Date:	June 19, 2023
Start time (MST):	9:20	End time (MST):	12:30
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH <sub>4</sub> Cal Gas Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.3 ppm		
Removed Gas Ref.	N/A	Removed Gas Expiry:	N/A
Removed CH <sub>4</sub> Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	5608

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.54E-04	2.63E-04	NMHC SP Ratio:	7.86E-05
CH <sub>4</sub> Retention time:	12.1	12.5	NMHC Peak Area:	115463
				125821

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.12	17.90	0.957
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	17.24	0.993
second point	4960	40.2	8.57	8.50	1.009
third point	4980	20.1	4.29	4.25	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.12	17.24	0.993

Average Correction Factor	1.004
---------------------------	-------

Baseline Corr AF:	17.90	Prev response	17.17	*% change	4.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF Limit= 0.95-1.05
as found zero	5000	0	0.00	0.00	----
as found span	4920	80.3	9.07	10.00	0.907
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	80.3	9.07	9.15	0.991
second point	4960	40.2	4.54	4.56	0.995
third point	4980	20.1	2.27	2.28	0.996
as left zero	5000	0	0.00	0.00	----
as left span	4920	80.3	9.07	9.07	1.000
Average Correction Factor					0.994
Baseline Corr AF:	10.00	Prev response	9.11	*% change	8.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF Limit= 0.95-1.05
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.06	7.89	1.021
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.06	8.09	0.996
second point	4960	40.2	4.03	3.93	1.025
third point	4980	20.1	2.02	1.97	1.024
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.06	8.17	0.986
Average Correction Factor					1.015
Baseline Corr AF:	7.89	Prev response	8.05	*% change	-2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.007754	1.007103
THC Cal Offset:	-0.087832	-0.053449
CH <sub>4</sub> Cal Slope:	1.007409	1.004333
CH <sub>4</sub> Cal Offset:	-0.063374	-0.044776
NMHC Cal Slope:	1.007709	1.009703
NMHC Cal Offset:	-0.024058	-0.008474

Notes: NM spans slowly rising since ZAG changed out last month. Seems to have settled. Analyzer flows were also adjusted last month, putting 8.9% change down to these maintenance items. New zero chrome ran and applied. Span adjusted. Filter changed after As Found.

Calibration Performed By: Ryan Power



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 19, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:20	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999924	$\geq 0.995$			
17.12	17.24	0.9934						
8.57	8.50	1.0091				Slope	1.007103	0.90 - 1.10
4.29	4.25	1.0090						
			Intercept	-0.053449	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 19, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:20	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999750	≥0.995
8.06	8.09	0.9964			
4.03	3.93	1.0254			
2.02	1.97	1.0241			
			Slope	1.004333	0.90 - 1.10
			Intercept	-0.044776	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 19, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:20	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	$\geq 0.995$			
9.07	9.15	0.9906						
4.54	4.56	0.9946				Slope	1.009703	0.90 - 1.10
2.27	2.28	0.9959						
			Intercept	-0.008474	$\pm 0.5$			






NMHC Calibration Plot

Date: July 4, 2023

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	July 10, 2023	Last Cal Date:	July 4, 2023
Start time (MST):	8:40	End time (MST):	10:20
Reason:	Removal		

### Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH <sub>4</sub> Cal Gas Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.3 ppm		
Removed Gas Ref.	N/A	Removed Gas Expiry:	N/A
Removed CH <sub>4</sub> Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	5608

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.63E-04	NA	NMHC SP Ratio:	7.22E-05
CH <sub>4</sub> Retention time:	12.5	NA	NMHC Peak Area:	125821
				NA

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.12	16.55	1.035
as found 2nd point	4960	40.2	8.57	8.13	1.054
as found 3rd point	4980	20.1	4.29	4.05	1.057
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	16.55	1.035
second point	4960	40.2	8.57	8.13	1.054
third point	4980	20.1	4.29	4.05	1.057

as left zero  
as left span

Average Correction Factor	1.049
---------------------------	-------

Baseline Corr AF:	16.55	Prev response	17.17	*% change	-3.7%
Baseline Corr 2nd AF:	8.1	AF Slope:	0.967241	AF Intercept:	-0.065575
Baseline Corr 3rd AF:	4.1	AF Correlation:	0.999891	<i>* = &gt; +/-5% change initiates investigation</i>	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF Limit= 0.95-1.05
as found zero	5000	0	0.00	0.00	----
as found span	4920	80.3	9.07	9.21	0.985
as found 2nd point	4960	40.2	4.54	4.54	1.001
as found 3rd point	4980	20.1	2.27	2.28	0.998
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	80.3	9.07	9.21	0.985
second point	4960	40.2	4.54	4.54	1.001
third point	4980	20.1	2.27	2.28	0.998
as left zero					
as left span					
Average Correction Factor					0.994
Baseline Corr AF:	9.21	Prev response	9.11	*% change	1.0%
Baseline Corr 2nd AF:	4.5	AF Slope:	1.015397	AF Intercept:	-0.026014
Baseline Corr 3rd AF:	2.3	AF Correlation:	0.999925	* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF Limit= 0.95-1.05
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.06	7.34	1.097
as found 2nd point	4960	40.2	4.03	3.60	1.121
as found 3rd point	4980	20.1	2.02	1.78	1.133
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.06	7.34	1.097
second point	4960	40.2	4.03	3.60	1.121
third point	4980	20.1	2.02	1.78	1.133
as left zero					
as left span					
Average Correction Factor					1.117
Baseline Corr AF:	7.34	Prev response	8.05	*% change	-9.6%
Baseline Corr 2nd AF:	3.60	AF Slope:	0.913040	AF Intercept:	-0.039560
Baseline Corr 3rd AF:	1.78	AF Correlation:	0.999833	* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.007754	0.967241
THC Cal Offset:	-0.087832	-0.065575
CH <sub>4</sub> Cal Slope:	1.007409	0.913040
CH <sub>4</sub> Cal Offset:	-0.063374	-0.039560
NMHC Cal Slope:	1.007709	1.015397
NMHC Cal Offset:	-0.024058	-0.026014

Notes:

Removal calibration

Calibration Performed By: Devin Russell



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	July 4, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:40	End Time (MST):	10:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999891	$\geq 0.995$			
17.12	16.55	1.0347						
8.57	8.13	1.0539				Slope	0.967241	0.90 - 1.10
4.29	4.05	1.0572						
			Intercept	-0.065575	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	July 4, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:40	End Time (MST):	10:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999833	<i>≥0.995</i>
8.06	7.34	1.0969			
4.03	3.60	1.1211			
2.02	1.78	1.1334			
			Slope	0.913040	<i>0.90 - 1.10</i>
			Intercept	-0.039560	<i>+/-0.5</i>





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	July 4, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:40	End Time (MST):	10:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999925	$\geq 0.995$			
9.07	9.21	0.9850						
4.54	4.54	1.0007				Slope	1.015397	0.90 - 1.10
2.27	2.28	0.9976						
			Intercept	-0.026014	$\pm 0.5$			



NMHC Calibration Plot

Date: July 10, 2023

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	July 10, 2023	Last Cal Date:	NA
Start time (MST):	11:50	End time (MST):	14:15
Reason:	Install		

### Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH <sub>4</sub> Cal Gas Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.3 ppm		
Removed Gas Ref.	N/A	Removed Gas Expiry:	N/A
Removed CH <sub>4</sub> Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	5608

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1172750022
THC Range (ppm):	0 - 20 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm
NMHC Range (ppm):	0 - 10 ppm		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	NA	2.12E-04	NMHC SP Ratio:	NA
CH <sub>4</sub> Retention time:	NA	14.0	NMHC Peak Area:	NA
				4.63E-05
				195941

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	17.12	1.000
second point	4960	40.2	8.57	8.57	1.000
third point	4980	20.1	4.29	4.36	0.984
as left zero	5000	0.0	0.00	0.01	----
as left span	4920	80.3	17.12	17.09	1.002
Average Correction Factor					0.995

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	80.3	9.07	9.09	0.997
second point	4960	40.2	4.54	4.57	0.993
third point	4980	20.1	2.27	2.34	0.970
as left zero	5000	0	0.00	0.01	----
as left span	4920	80.3	9.07	9.04	1.003
Average Correction Factor					0.987
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.06	8.03	1.003
second point	4960	40.2	4.03	4.00	1.008
third point	4980	20.1	2.02	2.02	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.06	8.05	1.000
Average Correction Factor					1.004
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	NA	0.998360
THC Cal Offset:	NA	0.029442
CH <sub>4</sub> Cal Slope:	NA	0.996256
CH <sub>4</sub> Cal Offset:	NA	-0.001745
NMHC Cal Slope:	NA	1.000342
NMHC Cal Offset:	NA	0.030986

Notes:

Installation calibration. Span adjusted.

Calibration Performed By:

Devin Russell



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	NA
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	11:50	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999980	$\geq 0.995$			
17.12	17.12	1.0001						
8.57	8.57	1.0002				Slope	0.998360	0.90 - 1.10
4.29	4.36	0.9841						
			Intercept	0.029442	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	NA
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	11:50	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999988	$\geq 0.995$			
8.06	8.03	1.0032						
4.03	4.00	1.0084				Slope	0.996256	0.90 - 1.10
2.02	2.02	1.0002						
			Intercept	-0.001745	$\pm 0.5$			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	NA
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	11:50	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750022

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999949	$\geq 0.995$			
9.07	9.09	0.9973						
4.54	4.57	0.9930				Slope	1.000342	0.90 - 1.10
2.27	2.34	0.9703						
			Intercept	0.030986	$\pm 0.5$			



NMHC Calibration Plot

Date: July 10, 2023

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	July 30, 2023	Last Cal Date:	July 10, 2023
Start time (MST):	10:15	End time (MST):	12:14
Reason:	As Found		

### Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH <sub>4</sub> Cal Gas Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.3 ppm		
Removed Gas Ref.	N/A	Removed Gas Expiry:	N/A
Removed CH <sub>4</sub> Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	5608

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.12E-04	2.12E-04	NMHC SP Ratio:	4.63E-05
CH <sub>4</sub> Retention time:	14.0	14.0	NMHC Peak Area:	195941

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.02	----
as found span	4920	80.3	17.12	12.21	<b>1.402</b>
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4920	80.3	17.12	12.40	<b>1.381</b>
second point					
third point					
as left zero					
as left span					
Average Correction Factor					<b>1.381</b>
Baseline Corr AF:	12.20	Prev response	17.12	*% change	<b>-40.4%</b>
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<b>* = &gt; +/-5% change initiates investigation</b>	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF Limit= 0.95-1.05
as found zero	5000	0	0.00	0.01	----
as found span	4920	80.3	9.07	5.87	1.546
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.01	----
high point	4920	80.3	9.07	6.06	1.497
second point					
third point					
as left zero					
as left span					
Average Correction Factor					1.497
Baseline Corr AF:	5.86	Prev response	9.10	*% change	-55.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF Limit= 0.95-1.05
as found zero	5000	0.0	0.00	0.01	----
as found span	4920	80.3	8.06	6.35	1.269
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4920	80.3	8.06	6.34	1.271
second point					
third point					
as left zero					
as left span					
Average Correction Factor					1.271
Baseline Corr AF:	6.34	Prev response	8.02	*% change	-26.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998360	0.723598
THC Cal Offset:	0.029442	0.010000
CH <sub>4</sub> Cal Slope:	0.996256	0.786400
CH <sub>4</sub> Cal Offset:	-0.001745	0.005000
NMHC Cal Slope:	1.000342	0.667580
NMHC Cal Offset:	0.030986	0.005000

Notes: Analyzer baseline shifted down and was dipping. Performed as founds. Span adjustments were attempted but the analyzer kept reverting back to the previous span targets.

Calibration Performed By: Rene Chamberland

NMHC Calibration Plot

Date: July 30, 2023

Location: Patricia McInnes







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	July 31, 2023	Last Cal Date:	July 10, 2023
Start time (MST):	9:00	End time (MST):	12:20
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH <sub>4</sub> Cal Gas Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.3 ppm		
Removed Gas Ref.	N/A	Removed Gas Expiry:	N/A
Removed CH <sub>4</sub> Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	4602

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.12E-04	2.11E-04	NMHC SP Ratio:	4.63E-05
CH <sub>4</sub> Retention time:	14.0	14.0	NMHC Peak Area:	195941
				192120

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4920	80.3	17.12	17.18	0.997
second point	4960	40.2	8.57	8.61	0.996
third point	4980	20.1	4.29	4.36	0.984
as left zero	5000	0.0	0.00	0.02	----
as left span	4920	80.3	17.12	17.08	1.003
Average Correction Factor					0.992

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.01	----
high point	4920	80.3	9.07	9.12	0.994
second point	4960	40.2	4.54	4.60	0.987
third point	4980	20.1	2.27	2.34	0.970
as left zero	5000	0	0.00	0.02	----
as left span	4920	80.3	9.07	9.08	0.999
Average Correction Factor					0.984
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.06	8.06	0.999
second point	4960	40.2	4.03	4.01	1.006
third point	4980	20.1	2.02	2.02	0.999
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.06	8.00	1.007
Average Correction Factor					1.001
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998360	1.001550
THC Cal Offset:	0.029442	0.032027
CH <sub>4</sub> Cal Slope:	0.996256	1.000285
CH <sub>4</sub> Cal Offset:	-0.001745	-0.004951
NMHC Cal Slope:	1.000342	1.002851
NMHC Cal Offset:	0.030986	0.036779

Notes: Analyzer baseline shifted down and was dipping. As founds were completed yesterday. Changed the pump and adjusted the span.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 31, 2023	Previous Calibration:	July 10, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:00	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.01	----	Correlation Coefficient	0.999991	$\geq 0.995$			
17.12	17.18	0.9968						
8.57	8.61	0.9956				Slope	1.001550	0.90 - 1.10
4.29	4.36	0.9839						
			Intercept	0.032027	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 31, 2023	Previous Calibration:	July 10, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:00	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999986	$\geq 0.995$
8.06	8.06	0.9993			
4.03	4.01	1.0057			
2.02	2.02	0.9992			
			Slope	1.000285	0.90 - 1.10
			Intercept	-0.004951	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 31, 2023	Previous Calibration:	July 10, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:00	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999966	$\geq 0.995$
9.07	9.12	0.9944			
4.54	4.60	0.9870			
2.27	2.34	0.9703			
			Slope	1.002851	0.90 - 1.10
			Intercept	0.036779	+/-0.5



NMHC Calibration Plot

Date: July 31, 2023

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Patricia McInnes  
Calibration Date: July 27, 2023  
Start time (MST): 8:08  
Reason: Routine  
Station number: AMS06  
Last Cal Date: June 26, 2023  
End time (MST): 12:36

### Calibration Standards

NO Gas Cylinder #: T30YCWN  
NOX Cal Gas Conc: 47.94 ppm  
Removed Cylinder #: N/A  
Removed Gas NOX Conc: 47.94 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701  
Cal Gas Expiry Date: April 11, 2025  
NO Cal Gas Conc: 46.39 ppm  
Removed Gas Exp Date: N/A  
Removed Gas NO Conc: 46.39 ppm  
NO gas Diff:  
Serial Number: 3566  
Serial Number: 4602

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1172750022

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.815	0.815	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.986	0.986	NOX bkgnd or offset:	3.8	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	156.0	157.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.996744	0.993425
NO <sub>x</sub> Cal Offset:	2.176487	2.677022
NO Cal Slope:	1.000845	0.996185
NO Cal Offset:	1.122676	1.203660
NO <sub>2</sub> Cal Slope:	1.000605	0.998218
NO <sub>2</sub> Cal Offset:	-0.357261	-0.045063



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.6	0.2	0.4	----	----
as found span	4914	86.2	826.5	799.7	26.7	824.1	795.2	28.9	1.0029	1.0057
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.7	0.3	0.4	----	----
high point	4914	86.2	826.5	799.7	26.7	822.7	797.3	25.6	1.0046	1.0030
second point	4957	43.1	413.2	399.9	13.4	414.2	400.4	13.8	0.9977	0.9987
third point	4978	21.6	207.1	200.4	6.7	210.4	201.5	8.9	0.9844	0.9946
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.2	0.1	----	----
as left span	4914	86.2	826.5	402.5	423.9	823.7	399.1	424.7	1.0033	1.0086
Average Correction Factor									0.9955	0.9988

Corrected As found	NO <sub>x</sub> = 823.5 ppb	NO = 795.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.3%
Previous Response	NO <sub>x</sub> = 825.9 ppb	NO = 801.5 ppb		*Percent Change	NO = -0.8%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.0	398.8	423.9	423.5	1.0010	99.9%
2nd GPT point (200 ppb O3)	796.0	598.0	224.7	223.7	1.0046	99.5%
3rd GPT point (100 ppb O3)	796.0	695.7	127.0	126.5	1.0041	99.6%
Average Correction Factor					1.0032	99.7%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	June 26, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:08	End Time (MST):	12:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.7	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.5	822.7	1.0046		
413.2	414.2	0.9977		
207.1	210.4	0.9844		
			0.999973	
			0.993425	
			2.677022	





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	June 26, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:08	End Time (MST):	12:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999993	≥0.995
799.7	797.3	1.0030			
399.9	400.4	0.9987	Slope	0.996185	0.90 - 1.10
200.4	201.5	0.9946			
			Intercept	1.203660	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	June 26, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:08	End Time (MST):	12:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
423.9	423.5	1.0010		
224.7	223.7	1.0046		
127.0	126.5	1.0041		

**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: July 27, 2023

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Patricia McInnes      Station number: AMS06  
 Calibration Date: July 14, 2023      Last Cal Date: June 13, 2023  
 Start time (MST): 9:15      End time (MST): 12:12  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: API T700      Serial Number: 3566  
 ZAG Make/Model: API T701      Serial Number: 4602

### Analyzer Information

Analyzer make: Thermo 49i      Analyzer serial #: 1300156234  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002286	1.001286	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	0.600000	0.500000	Coeff or Slope:	1.019	1.019

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.1	----
as found span	5000	1303.0	400.0	400.1	1.000
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	-0.1	----
high point	5000	1303.0	400.0	400.6	0.999
second point	5000	966.5	200.0	201.4	0.993
third point	5000	794.3	100.0	101.0	0.990
as left zero	5000	800.0	0.0	0.6	----
as left span	5000	1303.0	400.0	402.8	0.993
Average Correction Factor					0.994

Baseline Corr As found:	400.0	Previous response	401.5	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 14, 2023	Previous Calibration:	June 13, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:15	End Time (MST):	12:12
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999988	
400.0	400.6	0.9985			≥0.995
200.0	201.4	0.9930	Slope	1.001286	
100.0	101.0	0.9901			0.90 - 1.10
			Intercept	0.500000	+/- 5



O<sub>3</sub> Calibration Plot

Date: July 14, 2023

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Patricia McInnes      Station number: AMS 06  
 Calibration Date: July 14, 2023      Last Cal Date: June 5, 2023  
 Start time (MST): 13:19      End time (MST): 14:20

Analyzer Make: API T640      S/N: 766  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT FP-25      S/N: 388755  
 Temp/RH standard: ALICAT FP-25      S/N: 388755

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	19.4	19.4	19.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	727.5	728.1	727.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.06	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>July 14, 2023</u>		Last Cal Date: <u>June 5, 2023</u>		
	PM w/o HEPA: <u>105.3</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning :      Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9	10.9	10.9	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: <u>79.4</u>		w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:	<u>July 14, 2023</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>July 14, 2023</u>				

### Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023  
 Date RH/T Sensor Cleaned: April 13, 2023

Notes: "Perform span dust check" alarm was on upon arrival. Cleaned the chamber and changed the filter.  
 PMT peak was on target after the post maintenance. Leak check passed, no adjustments made.

Calibration by: Max Farrell





# Wood Buffalo Environmental Association

## TN - NO<sub>x</sub> - NH<sub>3</sub> Calibration Report

Version-05-2023

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	July 18, 2023	Last Cal Date:	June 26, 2023
Start time (MST):	9:19	End time (MST):	13:45
NH3 Cal Date:	July 18, 2023	Last Cal Date:	June 27, 2023
Start time (MST):	13:45	End time (MST):	17:20
Reason:	Routine		

### Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWVN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	NA
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	77.8	ppm	NH3 Gas Cylinder #:	CC710812
			NH3 Cal Gas Expiry:	March 30, 2023
Removed NH3 Conc:	77.8	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	API T700		Serial Number:	3566
ZAG make/model:	API T701		Serial Number:	4602

### Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	152
Converter model:	Teledyne API T501	Converter serial #:	147
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.50
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	542

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.794	0.794	TN coefficient:	0.794	0.794
NOX coefficient:	0.796	0.796	NO bkgnd:	-0.1	-0.1
NO2 coefficient:	1.000	1.000	NOX bkgnd:	0.0	0.0
NH3 coefficient:	0.908	0.908	TN bkgnd:	0.0	0.0

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.005776	1.012637
NO <sub>x</sub> Cal Offset:	1.234716	1.353007
NO Cal Slope:	0.998846	1.012223
NO Cal Offset:	1.222396	1.240145
NO <sub>2</sub> Cal Slope:	1.005535	1.001251
NO <sub>2</sub> Cal Offset:	0.912555	0.861558
NH3 Cal Slope:	1.017396	1.026272
NH3 Cal Offset:	0.876681	3.727263
TN Cal Slope:	1.023007	1.031709
TN Cal Offset:	0.746567	4.015689



# Wood Buffalo Environmental Association

## TN - NOX - NH<sub>3</sub> Calibration Report

Version-05-2023

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.2	0.0	1.2	----	----
as found NO	4914	86.2	826.5	826.5	----	839.1	839.0	0.0	0.985	----
calibrator zero	5000	0.0	0.0	0.0	0.0	1.2	0.0	1.2	----	----
high NO point	4914	86.2	826.5	826.5	----	839.1	839.0	0.0	0.985	----
GPT point	4914	86.2	826.5	826.5	----	846.4	834.5	11.8	0.976	----
as found NH3	3419	81.0	1800.5	----	1800.5	1859.7	----	1849.8	0.968	0.973
new NH3 cyl rp							----			
first NH3	3419	81.0	1800.5	----	1800.5	1859.7	----	1849.8	0.968	0.973
second NH3	3455	45.0	1000.3	----	1000.3	1038.2	----	1032.3	0.963	0.969
third NH3	3478	22.5	500.1	----	500.1	522.5	----	519.2	0.957	0.963
<b>Average Correction Factor</b>									<b>0.9807</b>	<b>0.9685</b>

Corrected As found      TN = 837.9 ppb      NO<sub>x</sub> = 839.0 ppb      NH3 = 1848.6 ppb

Previous Response      TN = 846.2 ppb      NO<sub>x</sub> = 832.5 ppb      NH3 = 1832.7 ppb

NH3 Previous Converter Efficiency = 90.8%

NH3 Current Converter Efficiency = 90.8%

\*Percent Change      TN = -1.0%

\*Percent Change      NO<sub>x</sub> = 0.8%

\*Percent Change      NH3 = 0.9%

\* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## NO<sub>x</sub> - NO - NO<sub>2</sub> Calibration Report

Version-05-2023

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	0.2	1.2	----	----
as found span	4914	86.2	826.5	799.7	826.5	839.0	807.7	839.1	0.9850	0.9901
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.2	2.9	----	----
high point	4914	86.2	826.5	799.7	826.5	837.9	810.4	839.3	0.9863	0.9868
second point	4957	43.1	413.2	399.9	413.2	419.4	406.0	424.5	0.9853	0.9849
third point	4978	21.6	207.1	200.4	207.1	213.3	205.5	215.4	0.9710	0.9753
<b>Average Correction Factor</b>									<b>0.9809</b>	<b>0.9823</b>

Baseline Corr As fnd	TN = 837.9 ppb	NO <sub>x</sub> = 839.0 ppb	NO = 807.5 ppb	*Percent Change	TN = -1.0%
Previous Response	TN = 846.2 ppb	NO <sub>x</sub> = 832.5 ppb	NO = 800.0 ppb	*Percent Change	NO <sub>x</sub> = 0.8%
				*Percent Change	NO = 0.9%

\* = > +/-5% change initiates investigation

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	-0.3	----	----
calibration zero	----	----	0.0	-0.3	----	----
1st GPT point (400 ppb O3)	807.0	401.5	432.2	433.0	0.9982	100.2%
2nd GPT point (200 ppb O3)	807.0	607.3	226.4	227.9	0.9935	100.7%
3rd GPT point (100 ppb O3)	807.0	700.2	133.5	136.0	0.9818	101.9%
<b>Average Correction Factor</b>					<b>0.9912</b>	<b>100.9%</b>

Notes:

No adjustments or maintenance performed.

Calibration Performed By:

Maddison Baragar



# Wood Buffalo Environmental Association

## TN Calibration Summary

Version-05-2023


### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 26, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	13:45	End Time (MST):	17:20
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	1.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.5	1859.7	0.9682		
1000.3	1038.2	0.9635		
500.1	522.5	0.9571		
			0.999988	
			1.031709	
			4.015689	

**TN Calibration Curve**





# Wood Buffalo Environmental Association

## NH<sub>3</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 26, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	13:45	End Time (MST):	17:20
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	1.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.5	1849.8	0.9734		
1000.3	1032.3	0.9690		
500.1	519.2	0.9632		
			0.999990	
			1.026272	
			3.727263	





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 26, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:19	End Time (MST):	13:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
826.5	837.9	0.9863			
413.2	419.4	0.9853			
207.1	213.3	0.9710			
			Slope	1.012637	0.90 - 1.10
			Intercept	1.353007	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-05-2023


### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 26, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:19	End Time (MST):	13:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
799.7	810.4	0.9868			
399.9	406.0	0.9849			
200.4	205.5	0.9753			
			Slope	1.012223	0.90 - 1.10
			Intercept	1.240145	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-05-2023


### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 26, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:19	End Time (MST):	13:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.3	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ $0.90 - 1.10$ $\pm 20$
432.2	433.0	0.9982		
226.4	227.9	0.9935		
133.5	136.0	0.9818		
			0.999960	
			1.001251	
			0.861558	

**NO<sub>2</sub> Calibration Curve**






NO<sub>x</sub> Calibration Plot

Date: July 18, 2023


Location: Patricia McInnes



# NH<sub>3</sub> Calibration Plot

Date: July 18, 2023

Location: Patricia McInnes





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS07 ATHABASCA VALLEY JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	July 6, 2023	Last Cal Date:	June 16, 2023
Start time (MST):	9:17	End time (MST):	16:58
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.52	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC282115			
Removed Cal Gas Conc:	50.52	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 701H		Serial Number:	198

### Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003815	0.995398	Backgd or Offset:	2.70	2.61
Calibration intercept:	1.384231	2.183198	Coeff or Slope:	0.840	0.834

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	79.3	801.2	805.9	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	79.3	801.2	798.4	1.004
second point	4960	39.6	400.2	402.2	0.995
third point	4980	19.8	200.1	203.1	0.985
as left zero	5000	0.0	0.0	0.0	----
as left span	4921	79.2	800.2	804.1	0.995
Average Correction Factor					0.994

Baseline Corr As found:	806.00	Previous response	805.64	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	June 16, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:17	End Time (MST):	16:58
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	0.999964	
801.2	798.4	1.0035			≥0.995
400.2	402.2	0.9949	Slope	0.995398	
200.1	203.1	0.9851			0.90 - 1.10
			Intercept	2.183198	+/-30



SO2 Calibration Plot

Date: July 6, 2023

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Athabasca Valley Station number: AMS07  
 Calibration Date: July 11, 2023 Last Cal Date: June 26, 2023  
 Start time (MST): 9:05 End time (MST): 13:45  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.25 ppm Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC504080  
 Removed Cal Gas Conc: 5.25 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 3805  
 ZAG Make/Model: API T701H Serial Number: 198

### Analyzer Information

Analyzer make: Thermo 43i LTE Analyzer serial #: 1180540018  
 Converter make: CDN-101 Converter serial #: 551  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003525	1.008281	Backgd or Offset: 2.33	2.33
Calibration intercept:	-0.042085	-0.182074	Coeff or Slope: 0.885	0.885

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4925	75.5	79.3	80.0	0.990
as found 2nd point	4962	37.7	39.6	39.8	0.992
as found 3rd point	4981	18.9	19.8	19.7	1.002
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4925	75.5	79.3	79.8	0.994
second point	4962	37.7	39.6	39.9	0.993
third point	4981	18.9	19.9	19.5	1.018
as left zero	5000	0.0	0.0	0.0	----
as left span	4925	75.5	79.3	79.5	0.998
SO2 Scrubber Check	4921	79.2	800.2	0.1	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	1.002
Date of last converter efficiency test:	April 22, 2022			92.6% efficiency	

Baseline Corr As found: 80.1 Prev response: 79.50 \*% change: 0.7%  
 Baseline Corr 2nd AF pt: 39.9 AF Slope: 1.011456 AF Intercept: -0.222210  
 Baseline Corr 3rd AF pt: 19.8 AF Correlation: 0.999989

\* = > +/-5% change initiates investigation

Notes: Scrubber check passed. No adjustment needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 26, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:05	End Time (MST):	13:45
Analyzer make:	CDN-101	Analyzer serial #:	551

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999952	≥0.995
79.3	79.8	0.9939			
39.6	39.9	0.9927	Slope	1.008281	0.90 - 1.10
19.9	19.5	1.0183			
			Intercept	-0.182074	+/-3






TRS Calibration Plot

Date: July 11, 2023

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	July 6, 2023	Last Cal Date:	June 28, 2023
Start time (MST):	9:00	End time (MST):	10:55
Reason:	Removal		

### Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	December 29, 2028
CH <sub>4</sub> Cal Gas Conc.	501.2 ppm	CH <sub>4</sub> Equiv Conc.	1075.1 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	501.2 ppm	CH <sub>4</sub> Equiv Conc.	1075.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	0.000288	NA	NMHC SP Ratio:	5.28E-05
CH <sub>4</sub> Retention time:	14.0	NA	NMHC Peak Area:	172128
				NA

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.31	1.044
as found 2nd point	4960	39.6	8.52	8.15	1.044
as found 3rd point	4980	19.8	4.26	4.11	1.035
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

Average Correction Factor					
Baseline Corr AF:	16.31	Prev response	16.92	*% change	-3.7%
Baseline Corr 2nd AF:	8.2	AF Slope:	0.957020	AF Intercept:	0.014074
Baseline Corr 3rd AF:	4.1	AF Correlation:	0.999994	* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.09	8.16	<b>1.114</b>
as found 2nd point	4960	39.6	4.55	4.09	<b>1.112</b>
as found 3rd point	4980	19.8	2.27	2.08	1.094
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
<b>Average Correction Factor</b>					
Baseline Corr AF:	8.16	Prev response	8.98	<b>*% change</b>	<b>-10.0%</b>
Baseline Corr 2nd AF:	4.1	AF Slope:	0.896083	AF Intercept:	0.017237
Baseline Corr 3rd AF:	2.1	AF Correlation:	0.999975	<b>* = &gt; +/-5% change initiates investigation</b>	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.94	8.15	0.974
as found 2nd point	4960	39.6	3.97	4.07	0.977
as found 3rd point	4980	19.8	1.98	2.04	0.975
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
<b>Average Correction Factor</b>					
Baseline Corr AF:	8.15	Prev response	7.94	<b>*% change</b>	2.6%
Baseline Corr 2nd AF:	4.07	AF Slope:	1.027057	AF Intercept:	-0.003563
Baseline Corr 3rd AF:	2.04	AF Correlation:	0.999997	<b>* = &gt; +/-5% change initiates investigation</b>	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.993911	
THC Cal Offset:	-0.005538	
CH <sub>4</sub> Cal Slope:	1.000237	
CH <sub>4</sub> Cal Offset:	0.002841	
NMHC Cal Slope:	0.988285	
NMHC Cal Offset:	-0.007979	


Notes: Removal calibration performed due to dips, more maintenance needed at the WBEAC.

Calibration Performed By: Aswin Sasi Kumar

NMHC Calibration Plot

Date: July 6, 2023

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	July 6, 2023	Last Cal Date:	NA
Start time (MST):	11:45	End time (MST):	16:58
Reason:	Install		

### Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	December 29, 2028
CH <sub>4</sub> Cal Gas Conc.	501.2 ppm	CH <sub>4</sub> Equiv Conc.	1075.1 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	501.2 ppm	CH <sub>4</sub> Equiv Conc.	1075.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	NA	0.000259	NMHC SP Ratio:	NA
CH <sub>4</sub> Retention time:	NA	14.2	NMHC Peak Area:	NA
				5.19E-05
				175309

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4921	79.2	17.03	17.03	1.000
second point	4960	39.6	8.52	8.51	1.000
third point	4980	19.8	4.26	4.31	0.989
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	16.85	1.010

				Average Correction Factor	0.996
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	9.09	9.09	1.000
second point	4960	39.6	4.55	4.54	1.002
third point	4980	19.8	2.27	2.32	0.981
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	9.09	8.95	1.016
Average Correction Factor					0.994
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00		----
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4921	79.2	7.94	7.94	1.000
second point	4960	39.6	3.97	3.97	0.999
third point	4980	19.8	1.98	1.99	0.998
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.94	7.91	1.004
Average Correction Factor					0.999
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		0.998642
THC Cal Offset:		0.023460
CH <sub>4</sub> Cal Slope:		0.999172
CH <sub>4</sub> Cal Offset:		0.007041
NMHC Cal Slope:		0.998192
NMHC Cal Offset:		0.015619

Notes: Install calibration, "zero chromatogram" and "use flat baseline" turned on. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	NA
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:45	End Time (MST):	16:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.01	----	Correlation Coefficient	0.999992	$\geq 0.995$			
17.03	17.03	1.0000						
8.52	8.51	1.0004				Slope	0.998642	0.90 - 1.10
4.26	4.31	0.9890						
			Intercept	0.023460	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	NA
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:45	End Time (MST):	16:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.00	0.01	----	Correlation Coefficient	≥0.995	
7.94	7.94	0.9998			
3.97	3.97	0.9992			
1.98	1.99	0.9984			
			Slope	0.999172	0.90 - 1.10
			Intercept	0.007041	+/-0.5







# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	NA
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:45	End Time (MST):	16:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999966	$\geq 0.995$
9.09	9.09	1.0001			
4.55	4.54	1.0020			
2.27	2.32	0.9814			
			Slope	0.998192	0.90 - 1.10
			Intercept	0.015619	+/-0.5



NMHC Calibration Plot

Date: July 6, 2023

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	July 10, 2023	Last Cal Date:	July 6, 2023
Start time (MST):	9:08	End time (MST):	14:18
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	December 29, 2028
CH <sub>4</sub> Cal Gas Conc.	501.2 ppm	CH <sub>4</sub> Equiv Conc.	1075.1 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	501.2 ppm	CH <sub>4</sub> Equiv Conc.	1075.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	0.000259	0.000268	NMHC SP Ratio:	5.19E-05
CH <sub>4</sub> Retention time:	14.2	14.4	NMHC Peak Area:	175309
				6.03E-05
				150713

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	15.87	1.073
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4921	79.2	17.03	17.11	0.995
second point	4960	39.6	8.52	8.55	0.996
third point	4980	19.8	4.26	4.26	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.07	0.997

				Average Correction Factor	0.997
Baseline Corr AF:	15.87	Prev response	17.03	*% change	-7.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.09	8.16	<b>1.114</b>
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	9.09	9.14	0.995
second point	4960	39.6	4.55	4.55	0.998
third point	4980	19.8	2.27	2.26	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	9.09	9.11	0.998
Average Correction Factor					1.000
Baseline Corr AF:	8.16	Prev response	9.09	*% change	<b>-11.4%</b>
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<b>* = &gt; +/-5% change initiates investigation</b>	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.94	7.71	1.030
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4921	79.2	7.94	7.97	0.996
second point	4960	39.6	3.97	4.00	0.993
third point	4980	19.8	1.98	2.00	0.992
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.94	7.96	0.997
Average Correction Factor					0.994
Baseline Corr AF:	7.71	Prev response	7.94	*% change	<b>-3.0%</b>
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<b>* = &gt; +/-5% change initiates investigation</b>	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998642	1.004709
THC Cal Offset:	0.023460	-0.003742
CH <sub>4</sub> Cal Slope:	0.999172	1.003246
CH <sub>4</sub> Cal Offset:	0.007041	0.010640
NMHC Cal Slope:	0.998192	1.006250
NMHC Cal Offset:	0.015619	-0.014182

Notes: Follow-up calibration on the recently installed instrument since span was 11% low during nightly z/s.  
Zero chromatogram setting turned off. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	July 6, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:08	End Time (MST):	14:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999997	$\geq 0.995$
17.03	17.11	0.9952			
8.52	8.55	0.9961			
4.26	4.26	0.9997			
			Slope	1.004709	0.90 - 1.10
			Intercept	-0.003742	+/-0.5





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	July 6, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:08	End Time (MST):	14:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999999	$\geq 0.995$
7.94	7.97	0.9957			
3.97	4.00	0.9930			
1.98	2.00	0.9919			
			Slope	1.003246	0.90 - 1.10
			Intercept	0.010640	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	July 6, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:08	End Time (MST):	14:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999989	$\geq 0.995$			
9.09	9.14	0.9946						
4.55	4.55	0.9982				Slope	1.006250	0.90 - 1.10
2.27	2.26	1.0066						
			Intercept	-0.014182	$\pm 0.5$			



NMHC Calibration Plot

Date: July 10, 2023

Location: Athabasca Valley







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	July 21, 2023	Last Cal Date:	July 10, 2023
Start time (MST):	9:33	End time (MST):	16:03
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	December 29, 2028
CH <sub>4</sub> Cal Gas Conc.	501.2 ppm	CH <sub>4</sub> Equiv Conc.	1075.1 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	501.2 ppm	CH <sub>4</sub> Equiv Conc.	1075.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	0.000268	0.000263	NMHC SP Ratio: 6.03E-05	4.97E-05
CH <sub>4</sub> Retention time:	14.4	14.4	NMHC Peak Area:	150713
				182990

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	17.51	0.973
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4921	79.2	17.03	16.93	1.006
second point	4960	39.6	8.52	8.35	1.019
third point	4980	19.8	4.26	4.21	1.011
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	16.50	1.032

Average Correction Factor				1.012
Baseline Corr AF:	17.51	Prev response	17.11	*% change 2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.09	9.47	0.960
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	9.09	8.99	1.011
second point	4960	39.6	4.55	4.41	1.032
third point	4980	19.8	2.27	2.23	1.020
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	9.09	8.62	1.055
Average Correction Factor					1.021
Baseline Corr AF:	9.47	Prev response	9.13	*% change	3.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.94	8.04	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4921	79.2	7.94	7.94	1.000
second point	4960	39.6	3.97	3.95	1.006
third point	4980	19.8	1.98	1.98	1.001
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.94	7.88	1.008
Average Correction Factor					1.002
Baseline Corr AF:	8.04	Prev response	7.98	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.004709	0.993293
THC Cal Offset:	-0.003742	-0.024935
CH <sub>4</sub> Cal Slope:	1.003246	0.998509
CH <sub>4</sub> Cal Offset:	0.010640	0.000842
NMHC Cal Slope:	1.006250	0.988725
NMHC Cal Offset:	-0.014182	-0.025977

Notes: NM was 11% high during nightly z/s cycle. Zero chromatogram feature turned on for baseline. Span adjusted. As left NM span is 5.5% low.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 21, 2023	Previous Calibration:	July 10, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:33	End Time (MST):	16:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.01	----	Correlation Coefficient	0.999943	$\geq 0.995$			
17.03	16.93	1.0060						
8.52	8.35	1.0194				Slope	0.993293	0.90 - 1.10
4.26	4.21	1.0108						
			Intercept	-0.024935	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 21, 2023	Previous Calibration:	July 10, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:33	End Time (MST):	16:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999988	$\geq 0.995$
7.94	7.94	1.0003			
3.97	3.95	1.0055			
1.98	1.98	1.0009			
			Slope	0.998509	0.90 - 1.10
			Intercept	0.000842	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 21, 2023	Previous Calibration:	July 10, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:33	End Time (MST):	16:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999872	$\geq 0.995$
9.09	8.99	1.0110			
4.55	4.41	1.0320			
2.27	2.23	1.0197			
			Slope	0.988725	0.90 - 1.10
			Intercept	-0.025977	+/-0.5



NMHC Calibration Plot

Date: July 21, 2023

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Athabasca Valley      Station number: AMS07  
Calibration Date: July 13, 2023      Last Cal Date: June 12, 2023  
Start time (MST): 9:02      End time (MST): 14:03  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T2Y1KA4      Cal Gas Expiry Date: November 30, 2023  
NOX Cal Gas Conc: 50.92 ppm      NO Cal Gas Conc: 49.92 ppm  
Removed Cylinder #: NA      Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 50.92 ppm      Removed Gas NO Conc: 49.92 ppm  
NOX gas Diff:      NO gas Diff:  
Calibrator Model: API T700      Serial Number: 3805  
ZAG make/model: API T701H      Serial Number: 198

### Analyzer Information

Analyzer make: Thermo 42i      Analyzer serial #: 1160120024  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.048	1.065	NO bkgnd or offset:	7.3	7.5
NOX coeff or slope:	0.995	0.995	NOX bkgnd or offset:	7.5	7.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	201.6	207.8

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.997768	1.001112
NO <sub>x</sub> Cal Offset:	1.378516	1.259152
NO Cal Slope:	0.998277	1.001431
NO Cal Offset:	0.974621	0.895212
NO <sub>2</sub> Cal Slope:	1.001546	1.003989
NO <sub>2</sub> Cal Offset:	0.752238	0.417819



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.3	----	----
as found span	4920	80.2	816.7	800.7	16.0	804.9	786.7	18.3	1.0147	1.0178
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
high point	4920	80.2	816.7	800.7	16.0	818.0	802.0	16.0	0.9984	0.9984
second point	4960	40.1	408.4	400.4	8.0	411.7	403.2	8.5	0.9919	0.9929
third point	4980	20.0	203.7	199.7	4.0	205.5	201.0	4.5	0.9911	0.9934
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
as left span	4920	80.2	816.7	394.6	422.1	824.0	395.7	428.4	0.9912	0.9972
Average Correction Factor									0.9938	0.9949

Corrected As found	NO <sub>x</sub> = 804.5 ppb	NO = 786.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -1.5%	
Previous Response	NO <sub>x</sub> = 816.3 ppb	NO = 800.3 ppb		*Percent Change	NO = -1.7%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	800.1	394.0	422.1	423.8	0.9961	100.4%
2nd GPT point (200 ppb O <sub>3</sub> )	800.1	599.9	216.2	218.5	0.9897	101.0%
3rd GPT point (100 ppb O <sub>3</sub> )	800.1	699.8	116.3	117.0	0.9944	100.6%
Average Correction Factor					0.9934	100.7%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 12, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:02	End Time (MST):	14:03
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	≥0.995	
816.7	818.0	0.9984			
408.4	411.7	0.9919			
203.7	205.5	0.9911			
			Slope	1.001112	0.90 - 1.10
			Intercept	1.259152	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 12, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:02	End Time (MST):	14:03
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.7	802.0	0.9984		
400.4	403.2	0.9929		
199.7	201.0	0.9934		

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 12, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:02	End Time (MST):	14:03
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
422.1	423.8	0.9961		
216.2	218.5	0.9897		
116.3	117.0	0.9944		
			0.999987	
			1.003989	
			0.417819	


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: July 13, 2023

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Athabasca Valley      Station number: AMS07  
 Calibration Date: July 7, 2023      Last Cal Date: June 6, 2023  
 Start time (MST): 10:00      End time (MST): 14:45  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: T700      Serial Number: 3805  
 ZAG Make/Model: T701H      Serial Number: 198

### Analyzer Information

Analyzer make: Thermo 49i      Analyzer serial #: 1152220023  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999971	1.002371	Backgd or Offset:	-3.5	-2.6
Calibration intercept:	0.480000	0.860000	Coeff or Slope:	1.504	1.522

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	5000	1414.8	400.0	394.1	1.015
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.5	----
high point	5000	1415.7	400.0	401.7	0.996
second point	5000	1039.9	200.0	201.3	0.994
third point	5000	856.2	100.0	101.6	0.984
as left zero	5000	0.0	0.0	0.3	----
as left span	5000	1416.0	400.0	400.2	1.000
Average Correction Factor					0.991

Baseline Corr As found:	394.1	Previous response	400.5	*% change	-1.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 7, 2023	Previous Calibration:	June 6, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:00	End Time (MST):	14:45
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.5	----	Correlation Coefficient	0.999996	
400.0	401.7	0.9958			≥0.995
200.0	201.3	0.9935	Slope	1.002371	
100.0	101.6	0.9843			0.90 - 1.10
			Intercept	0.860000	+/- 5



O<sub>3</sub> Calibration Plot

Date: July 7, 2023

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Athabasca Valley Station number: AMS 07  
Calibration Date: July 24, 2023 Last Cal Date: June 29, 2023  
Start time (MST): 11:30 End time (MST): 14:33  
Analyzer Make: API T640 S/N: 871  
Particulate Fraction: PM2.5  
Flow Meter Make/Model: Alicat FP-25BT S/N: 388748  
Temp/RH standard: Alicat FP-25BT S/N: 388748

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	29.8	29.09	31.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	732	735.03	729.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.10	5.10	4.93	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: July 24, 2023	Last Cal Date: June 29, 2023			
	PM w/o HEPA: 14.8	PM w/ HEPA: 0.0			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.2	10.1	10.1	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: 20.1	w/ HEPA: 0.0		
Date Optical Chamber Cleaned:		July 24, 2023			<0.2 ug/m3
Disposable Filter Changed:		July 24, 2023			

### Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022  
Date RH/T Sensor Cleaned: December 5, 2022

Notes: Temp, flow and pressure checked. Leak check passed. Chamber cleaned, filter swapped. No issues.

Calibration by: Aswin Sasi Kumar





# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	July 24, 2023	Last Cal Date:	June 29, 2023
Start time (MST):	9:44	End time (MST):	13:07
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date:	December 12, 2026
Cal Gas Cylinder #:	LL66942			
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 700H		Serial Number:	198

### Analyzer Information

Analyzer make: Thermo 48i-LTE      Analyzer serial #: 1408761381  
 Analyzer Range: 0 - 50 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999915	1.000886	Backgd or Offset:	4.147	4.137
Calibration intercept:	0.018540	0.076548	Coeff or Slope:	1.086	1.086

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.0	40.1	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.0	40.1	0.998
second point	4967	33.3	20.0	20.2	0.991
third point	4983	16.7	10.0	10.1	0.993
as left zero	5000	0.0	0.0	0.0	----
as left span	4933	66.7	40.0	40.0	1.001
Average Correction Factor					0.994

Baseline Corr As found:	40.03	Prev response:	40.04	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No Maintenance done. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 24, 2023	Previous Calibration:	June 29, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:44	End Time (MST):	13:07
Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999985	
40.0	40.1	0.9983			≥0.995
20.0	20.2	0.9905	Slope	1.000886	
10.0	10.1	0.9931			0.90 - 1.10
			Intercept	0.076548	+/-1.5



CO Calibration Plot

Date: July 24, 2023

Location: Athabasca Valley





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS08 FORT CHIPEWYAN JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	July 14, 2023	Last Cal Date:	June 28, 2023
Start time (MST):	11:34	End time (MST):	14:16
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC196697			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3060
ZAG Make/Model:	Teledyne API T701		Serial Number:	260

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996931	1.003156	Backgd or Offset:	1.60	1.53
Calibration intercept:	-0.082970	-0.363935	Coeff or Slope:	0.915	0.915

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4920	80.3	800.4	800.9	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.6	----
high point	4920	80.3	800.4	803.4	0.996
second point	4960	40.2	400.7	400.0	1.002
third point	4980	20.1	200.4	200.4	1.000
as left zero	5000	0.0	0.0	0.6	----
as left span	4920	80.3	800.4	805.2	0.994
Average Correction Factor					0.999

Baseline Corr As found:	800.60	Previous response	797.84	*% change	0.3%
-------------------------	--------	-------------------	--------	-----------	------

\* = > +/-5% change initiates investigation

Notes: Sampled inlet filters after as founds. No adjustments made.

Calibration Performed By: Matthew C



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 14, 2023	Previous Calibration:	June 28, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	11:34	End Time (MST):	14:16
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.6	----	Correlation Coefficient	≥0.995
800.4	803.4	0.9962		
400.7	400.0	1.0017	Slope	0.90 - 1.10
200.4	200.4	0.9998		
			Intercept	+/-30



SO2 Calibration Plot

Date: July 14, 2023

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	July 14, 2023	Last Cal Date:	June 29, 2023
Start time (MST):	15:00	End time (MST):	18:22
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	4.97	ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	EY0002276			
Removed Cal Gas Conc:	4.97	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
ZAG Make/Model:	Teledyne API T701		Serial Number:	260

### Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744
Converter make:	CDN-101	Converter serial #:	14639
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005570	1.011142	Backgd or Offset:	0.94
Calibration intercept:	0.218839	0.198873	Coeff or Slope:	0.707

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4920	80.5	80.0	80.2	1.003
as found 2nd point	4960	40.2	40.0	40.4	0.999
as found 3rd point	4980	20.1	20.0	20.1	1.014
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4920	80.5	80.0	81.2	0.985
second point	4960	40.2	40.0	40.4	0.989
third point	4980	20.1	20.0	20.4	0.979
as left zero	5000	0.0	0.0	0.5	----
as left span	4920	80.5	80.0	80.0	1.000
SO2 Scrubber Check	4919.7	80.3	803.0	0.2	----
Date of last scrubber change:	March 7, 2022			Ave Corr Factor	0.985
Date of last converter efficiency test:	March 15, 2022			100.7% efficiency	

Baseline Corr As found:	79.8	Prev response:	80.67	*% change:	-1.1%
Baseline Corr 2nd AF pt:	40.0	AF Slope:	0.998569	AF Intercept:	0.338765
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999981		

\* = > +/-5% change initiates investigation

Notes: Sampled inlet filter changed after as founds. Scrubber check passed. No adjustments made.

Calibration Performed By: Matthew C





# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 14, 2023	Previous Calibration:	June 29, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	15:00	End Time (MST):	18:22
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	$\geq 0.995$	
80.0	81.2	0.9853			
40.0	40.4	0.9890			
20.0	20.4	0.9794			
			Slope	1.011142	0.90 - 1.10
			Intercept	0.198873	+/-3



TRS Calibration Plot

Date: July 14, 2023

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Fort Chipewyan Station number: AMS08  
Calibration Date: July 13, 2023 Last Cal Date: June 27, 2023  
Start time (MST): 7:41 End time (MST): 12:16  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: CC363447 Cal Gas Expiry Date: February 2, 2024  
NOX Cal Gas Conc: 48.80 ppm NO Cal Gas Conc: 48.80 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 48.80 ppm Removed Gas NO Conc: 48.80 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 3060  
ZAG make/model: Teledyne API T701H Serial Number: 260

### Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1426262592  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.933	1.933	NO bkgnd or offset:	8.1	8.1
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	8.2	8.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	268.1	268.1

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.999486	0.991475
NO <sub>x</sub> Cal Offset:	-0.060000	0.520000
NO Cal Slope:	1.002670	0.992674
NO Cal Offset:	-1.000000	-0.500000
NO <sub>2</sub> Cal Slope:	0.991397	0.995889
NO <sub>2</sub> Cal Offset:	0.179814	0.340396



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.6	0.2	0.4	----	----
as found span	4918	82.0	800.3	800.3	0.0	796.3	793.8	2.3	1.0050	1.0082
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.7	0.2	0.4	----	----
high point	4918	82.0	800.3	800.3	0.0	794.0	794.2	-0.2	1.0080	1.0077
second point	4959	41.0	400.2	400.2	0.0	397.5	396.7	0.8	1.0067	1.0087
third point	4980	20.5	200.1	200.1	0.0	198.5	197.2	1.3	1.0080	1.0146
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.2	0.3	----	----
as left span	4918	82.0	800.3	423.0	377.3	794.7	418.4	376.2	1.0071	1.0110
Average Correction Factor									1.0075	1.0103

Corrected As found	NO <sub>x</sub> = 795.7 ppb	NO = 793.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.5%	
Previous Response	NO <sub>x</sub> = 799.8 ppb	NO = 801.5 ppb		*Percent Change	NO = -1.0%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	792.6	415.3	377.3	376.4	1.0024	99.8%
2nd GPT point (200 ppb O <sub>3</sub> )	792.6	609.8	182.8	181.5	1.0072	99.3%
3rd GPT point (100 ppb O <sub>3</sub> )	792.6	705.2	87.4	87.9	0.9943	100.6%
Average Correction Factor					1.0013	99.9%

Notes: sampled inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Matthew C



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 27, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	7:41	End Time (MST):	12:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient Slope Intercept	0.999999 0.991475 0.520000	≥0.995 0.90 - 1.10 +/-20
800.3	794.0	1.0080			
400.2	397.5	1.0067			
200.1	198.5	1.0080			





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 27, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	7:41	End Time (MST):	12:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	≥0.995	
800.3	794.2	1.0077			
400.2	396.7	1.0087			
200.1	197.2	1.0146			
			Slope	0.992674	0.90 - 1.10
			Intercept	-0.500000	+/-20

NO Calibration Curve





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 27, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	7:41	End Time (MST):	12:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
377.3	376.4	1.0024		
182.8	181.5	1.0072		
87.4	87.9	0.9943		



NO<sub>x</sub> Calibration Plot

Date: July 13, 2023

Location: Fort Chipewyan







# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Fort Chipewyan      Station number: AMS08  
 Calibration Date: July 10, 2023      Last Cal Date: June 28, 2023  
 Start time (MST): 8:51      End time (MST): 11:57  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 3252  
 ZAG Make/Model: Teledyne API T701      Serial Number: 260

### Analyzer Information

Analyzer make: Teledyne API T400      Analyzer serial #: 3872  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.012486	1.010943	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	-1.460000	-1.940000	Coeff or Slope:	1.036	1.036

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	NA	0.0	-0.2	----
as found span	5000	913.0	400.0	403.3	0.992
as found 2nd point					
as found 3rd point					
calibrator zero	5000	NA	0.0	0.5	----
high point	5000	914.7	400.0	403.7	0.991
second point	5000	786.4	200.0	198.9	1.006
third point	5000	701.3	100.0	96.8	1.033
as left zero	5000	NA	0.0	0.0	----
as left span	5000	963.3	400.0	404.5	0.989
Average Correction Factor					1.010

Baseline Corr As found:	403.5	Previous response	403.5	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: remote calibration done. No adjustments

Calibration Performed By: Morgan Voyageur



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	June 28, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:51	End Time (MST):	11:57
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999833	
400.0	403.7	0.9908			≥0.995
200.0	198.9	1.0055	Slope	1.010943	
100.0	96.8	1.0331			0.90 - 1.10
			Intercept	-1.940000	+/- 5



O<sub>3</sub> Calibration Plot

Date: July 10, 2023

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Fort Chipewyan Station number: AMS 08  
 Calibration Date: July 25, 2023 Last Cal Date: June 29, 2023  
 Start time (MST): 11:37 End time (MST): 13:39

Analyzer Make: Teledyne API T640 S/N: 216  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: FP-25 S/N: 388747  
 Temp/RH standard: FP-25 S/N: 388747

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	21.3	21.14	21.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	730.9	734.27	730.9	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.05	4.96	5.05	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: July 25, 2023	Last Cal Date: June 29, 2023			
	PM w/o HEPA: 18.8	PM w/ HEPA: 0.0			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.2	10.3	11.2	<input checked="" type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: 24.5	w/ HEPA: 0.0		
Date Optical Chamber Cleaned:		July 25, 2023			<0.2 ug/m3
Disposable Filter Changed:		July 25, 2023			

### Annual Maintenance

Date Sample Tube Cleaned: July 25, 2023  
 Date RH/T Sensor Cleaned: July 25, 2023

Notes: Verified flow, temperature, and pressure. Leak test passed. PMT peak voltage adjusted. Optical chamber, RH/T sensor, and sample tube cleaned. Disposable filter changed. Inlet head cleaned

Calibration by: Mohammed Kashif



# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	July 25, 2023	Last Cal Date:	June 30, 2023
Start time (MST):	8:03	End time (MST):	11:00
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API T701H		Serial Number:	260

### Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<b><u>Start</u></b>	<b><u>Finish</u></b>		<b><u>Start</u></b>	<b><u>Finish</u></b>
Calibration slope:	0.988086	0.996203	Backgd or Offset:	-0.014	-0.014
Calibration intercept:	0.136987	0.030984	Coeff or Slope:	0.968	0.968

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.11	----
as found span	4933	66.7	40.4	40.2	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4934	66.7	40.4	40.2	1.006
second point	4967	33.3	20.2	20.4	0.992
third point	4983	16.7	10.1	10.1	0.999
as left zero	5000	0.0	0.0	-0.1	----
as left span	2960	40.0	40.4	40.1	1.009
Average Correction Factor					0.999

Baseline Corr As found:	40.33	Prev response:	40.08	*% change:	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustments made.

Calibration Performed By: Matthew C



# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 25, 2023	Previous Calibration:	June 30, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:03	End Time (MST):	11:00
Analyzer make:	API T300	Analyzer serial #:	3505

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999916	≥0.995
40.4	40.2	1.0058			
20.2	20.4	0.9916	Slope	0.996203	0.90 - 1.10
10.1	10.1	0.9990			
			Intercept	0.030984	+/-1.5



CO Calibration Plot

Date: July 25, 2023

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	July 26, 2023	Last Cal Date:	July 25, 2023
Start time (MST):	13:50	End time (MST):	16:48
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3252
ZAG Make/Model:	API T701H		Serial Number:	260

### Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996203	1.002649	Backgd or Offset:	-0.014	-0.013
Calibration intercept:	0.030984	-0.072958	Coeff or Slope:	0.968	0.999

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.27	----
as found span	4933	66.7	40.4	37.8	1.070
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4934	66.7	40.4	40.4	1.000
second point	4967	33.3	20.2	20.4	0.992
third point	4983	16.7	10.1	9.9	1.027
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.4	40.2	1.005
Average Correction Factor					1.006

Baseline Corr As found:	38.05	Prev response:	40.30	*% change:	-5.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Inlet filter was changed out yesterday. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland





# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 26, 2023	Previous Calibration:	July 25, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	13:50	End Time (MST):	16:48
Analyzer make:	API T300	Analyzer serial #:	3505

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999896	≥0.995
40.4	40.4	1.0001			
20.2	20.4	0.9916	Slope	1.002649	0.90 - 1.10
10.1	9.9	1.0274			
			Intercept	-0.072958	+/-1.5



CO Calibration Plot

Date: July 26, 2023

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	July 25, 2023	Last Cal Date:	June 28, 2023
Start time (MST):	11:15	End time (MST):	12:45
Reason:	As Found		

### Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3060
N2 Gen Make/Model:	NG 5000		Serial Number:	771048318

### Analyzer Information

Analyzer make: Teledyne API T360	Analyzer serial #: 289
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.986205		Backgd or Offset:	0.008	0.008
Calibration intercept:	-4.300000		Coeff or Slope:	1.002	1.002

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	0.4	----
as found span	2920	80.0	1605.9	1587.0	1.012
as found 2nd point	2960	40.0	802.9	779.7	1.030
as found 3rd point	2980	20.0	401.5	397.4	1.010
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

#### Average Correction Factor

Baseline Corr As found:	1586.60	Prev response:	1579.41	*% change:	0.5%
Baseline Corr 2nd AF pt:	779.3	AF Slope:	0.986931	AF Intercept:	-2.260000
Baseline Corr 3rd AF pt:	397.0	AF Correlation:	0.999892		

\* = > +/-5% change initiates investigation


Notes: Performed multi-point as founds. Changed out the sample pump.

Calibration Performed By: Rene Chamberland

CO<sub>2</sub> Calibration Plot

Date: July 25, 2023

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	July 26, 2023	Last Cal Date:	June 28, 2023
Start time (MST):	10:10	End time (MST):	13:45
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
N2 Gen Make/Model:	NG 5000		Serial Number:	771048318

### Analyzer Information

Analyzer make: Teledyne API T360	Analyzer serial #: 289
Analyzer Range 0 - 2,000 ppm	

	<b>Start</b>	<b>Finish</b>		<b>Start</b>	<b>Finish</b>
Calibration slope:	0.986205	0.997030	Backgd or Offset:	0.008	0.008
Calibration intercept:	-4.300000	-4.080000	Coeff or Slope:	1.002	1.016

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.7	----
high point	2920	80.0	1605.9	1604.1	1.001
second point	2960	40.0	802.9	780.0	1.029
third point	2980	20.0	401.5	400.8	1.002
as left zero	3000	0.0	0.0	1.4	----
as left span	2960	40.0	802.9	776.0	1.035
Average Correction Factor					1.011

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: As founds and pump change were done yesterday. Changed the inlet filter. Adjusted the span. Performed a linearity adjustment on the 3rd point.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date	July 26, 2023	Previous Calibration	June 28, 2023
Station Name	Fort Chipewyan	Station Number	AMS08
Start Time (MST)	10:10	End Time (MST)	13:45
Analyzer make	Teledyne API T360	Analyzer serial #	289

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.7	----	Correlation Coefficient	0.999739	≥0.995
1605.9	1604.1	1.0011			
802.9	780.0	1.0294	Slope	0.997030	0.90 - 1.10
401.5	400.8	1.0017			
			Intercept	-4.080000	+/-20



CO<sub>2</sub> Calibration Plot

Date: July 26, 2023

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Fort Chipewyan	Station Number:	AMS 08
Calibration Date:	July 25, 2023	Prev Cal Date:	June 29, 2022
Start Time (MST):	12:23	End Time (MST):	14:00
Tower Height (m):	9.6	Reason:	Routine

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	P22394
WS Calibrator:	MetOne 053	Serial Number:	CA 05231

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.2%
400	39.4	39.3	-0.3%
600	58.6	58.4	-0.3%
800	77.8	77.6	-0.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	0.999999	1.000000	<i>≥0.9995</i>
Calculated slope	0.998873	1.002580	<i>0.90 - 1.10</i>
Calculated intercept	-0.027395	-0.001028	<i>+/- 2</i>

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	E4853
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	14:27	Calc Declination*:	13.75 Degrees
Deadband calc:	-0.7 degrees ( <i>Limit 4 deg</i> )	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.0	---
90	89.2	-0.2%
180	181.4	0.4%
270	273.2	0.9%
357	357.7	0.2%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	0.999915	0.999961	<i>≥0.9995</i>
Calculated slope	0.990857	0.993820	<i>0.90 - 1.10</i>
Calculated intercept	-0.704563	0.199340	<i>+/- 4</i>

Notes: Annual met calibration.

Calibration Performed By: Rene Chamberland





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS09 BARGE LANDING JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:56	End Time (MST):	12:12
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995
801.5	802.5	0.9987			
400.8	402.1	0.9966	Slope	1.000739	0.90 - 1.10
199.8	200.9	0.9947			
			Intercept	0.671341	+/-30



SO2 Calibration Plot

Date: July 11, 2023

Location: Barge Landing





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	July 4, 2023	Last Cal Date:	June 7, 2023
Start time (MST):	9:21	End time (MST):	13:14
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	4.87	ppm	Cal Gas Exp Date:	September 2, 2024
Cal Gas Cylinder #:	EY0002346			
Removed Cal Gas Conc:	4.87	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3812
ZAG Make/Model:	API T701		Serial Number:	4888

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320
Converter make:	CDN-101	Converter serial #:	519
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007720	1.000859	Backgd or Offset:	2.77	2.77
Calibration intercept:	-0.060968	-0.220903	Coeff or Slope:	1.134	1.134

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4918	82.1	80.0	78.9	1.012
as found 2nd point	4959	41.1	40.0	39.3	1.016
as found 3rd point	4979	20.5	20.0	19.4	1.024
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4918	82.1	80.0	79.9	1.001
second point	4959	41.1	40.0	39.7	1.008
third point	4979	20.5	20.0	19.7	1.014
as left zero	5000	0.0	0.0	-0.1	----
as left span	4918	82.1	80.0	80.3	0.996
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:	28-Feb-23			Ave Corr Factor	1.008
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	79.0	Prev response:	80.52	*% change:	-1.9%
Baseline Corr 2nd AF pt:	39.4	AF Slope:	0.988714	AF Intercept:	-0.220946
Baseline Corr 3rd AF pt:	19.5	AF Correlation:	0.999989		

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SOx scrubber check done after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021


### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 7, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:21	End Time (MST):	13:14
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999986	
80.0	79.9	1.0008			≥0.995
40.0	39.7	1.0083	Slope	1.000859	
20.0	19.7	1.0137			0.90 - 1.10
			Intercept	-0.220903	+/-3







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	July 11, 2023	Last Cal Date:	June 8, 2023
Start time (MST):	8:46	End time (MST):	12:12
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	497.6 ppm	CH <sub>4</sub> Equiv Conc.	1067.1 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	497.6 ppm	CH <sub>4</sub> Equiv Conc.	1067.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 100 ppm		
NMHC Range (ppm):	0 - 50 ppm	CH <sub>4</sub> Range (ppm):	0 - 50 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.49E-04	2.52E-04	NMHC SP Ratio:	4.85E-05
CH <sub>4</sub> Retention time:	15.2	15.4	NMHC Peak Area:	188395
				188466

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	17.11	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.17	0.997
second point	4960	40.1	8.56	8.56	1.000
third point	4980	20.0	4.27	4.28	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.15	0.998

				Average Correction Factor	0.998
Baseline Corr AF:	17.11	Prev response	17.16	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.0	----
as found span	4919	80.2	9.14	9.18	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.17	0.997
second point	4960	40.1	4.57	4.57	1.000
third point	4980	20	2.28	2.28	0.997
as left zero	5000	0	0.00	0.00	----
as left span	4919	80.2	9.14	9.14	1.000
Average Correction Factor					0.998
Baseline Corr AF:	9.18	Prev response	9.16	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	7.93	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	8.00	0.998
second point	4960	40.1	3.99	4.00	0.999
third point	4980	20.0	1.99	2.00	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	8.02	0.996
Average Correction Factor					0.998
Baseline Corr AF:	7.93	Prev response	8.00	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002132	1.002740
THC Cal Offset:	0.007266	-0.004535
CH <sub>4</sub> Cal Slope:	1.002397	1.002125
CH <sub>4</sub> Cal Offset:	-0.001332	-0.000132
NMHC Cal Slope:	1.001775	1.003064
NMHC Cal Offset:	0.008598	-0.003803

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:46	End Time (MST):	12:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	$\geq 0.995$			
17.12	17.17	0.9971						
8.56	8.56	0.9998				Slope	1.002740	0.90 - 1.10
4.27	4.28	0.9971						
			Intercept	-0.004535	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:46	End Time (MST):	12:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥0.995
7.98	8.00	0.9977			
3.99	4.00	0.9989			
1.99	2.00	0.9967			
			Slope	1.002125	0.90 - 1.10
			Intercept	-0.000132	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:46	End Time (MST):	12:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	$\geq 0.995$			
9.14	9.17	0.9968						
4.57	4.57	1.0003				Slope	1.003064	0.90 - 1.10
2.28	2.28	0.9974						
			Intercept	-0.003803	$\pm 0.5$			



NMHC Calibration Plot

Date: July 11, 2023

Location: Barge Landing





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Barge Landing  
Calibration Date: July 5, 2023  
Start time (MST): 9:19  
Reason: Routine  
Station number: AMS09  
Last Cal Date: June 6, 2023  
End time (MST): 13:45

### Calibration Standards

NO Gas Cylinder #: DT0036634  
NOX Cal Gas Conc: 50.00 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 50.00 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701  
Cal Gas Expiry Date: January 28, 2024  
NO Cal Gas Conc: 49.70 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 49.70 ppm  
NO gas Diff:  
Serial Number: 3812  
Serial Number: 4888

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1426262593

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.181	1.169	NO bkgnd or offset:	10.8	10.7
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	11.0	10.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	172.2	175.6

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.998838	0.997858
NO <sub>x</sub> Cal Offset:	0.888950	0.708793
NO Cal Slope:	1.001441	0.999541
NO Cal Offset:	-0.512202	-0.472540
NO <sub>2</sub> Cal Slope:	0.999242	1.000792
NO <sub>2</sub> Cal Offset:	0.437176	-0.264400



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
as found span	4919	80.5	805.1	800.3	4.8	815.4	807.4	8.0	0.987	0.991
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.2	----	----
high point	4919	80.5	805.1	800.3	4.8	803.6	799.5	4.2	1.002	1.001
second point	4959	40.2	402.1	399.7	2.4	402.4	399.1	3.3	0.999	1.001
third point	4979	20.1	201.0	199.8	1.2	202.2	198.7	3.5	0.994	1.006
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
as left span	4919	80.5	805.1	436.0	369.1	802.5	432.0	370.5	1.003	1.009
Average Correction Factor									0.998	1.003

Corrected As found	NO <sub>x</sub> = 815.8 ppb	NO = 807.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 1.3%
Previous Response	NO <sub>x</sub> = 805.0 ppb	NO = 800.9 ppb		*Percent Change	NO = 0.8%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	796.1	431.8	369.1	369.1	1.000	100.0%
2nd GPT point (200 ppb O <sub>3</sub> )	796.1	656.8	144.1	144.3	0.999	100.1%
3rd GPT point (100 ppb O <sub>3</sub> )	796.1	725.7	75.2	74.7	1.007	99.3%
Average Correction Factor					1.002	99.8%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 5, 2023	Previous Calibration:	June 6, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:19	End Time (MST):	13:45
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
805.1	803.6	1.0018		
402.1	402.4	0.9992		
201.0	202.2	0.9942		
			0.999994	
			0.997858	
			0.708793	







# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 5, 2023	Previous Calibration:	June 6, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:19	End Time (MST):	13:45
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
800.3	799.5	1.0009			
399.7	399.1	1.0014			
199.8	198.7	1.0057			
			Slope	0.999541	0.90 - 1.10
			Intercept	-0.472540	+/-20





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 5, 2023	Previous Calibration:	June 6, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:19	End Time (MST):	13:45
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	-0.2	----	Correlation Coefficient	0.999997	≥0.995
369.1	369.1	1.0001			
144.1	144.3	0.9988			
75.2	74.7	1.0071			
			Slope	1.000792	0.90 - 1.10
			Intercept	-0.264400	+/-20



NO<sub>x</sub> Calibration Plot

Date: July 5, 2023

Location: Barge Landing





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Barge Landing Station number: AMS 09  
 Calibration Date: July 5, 2023 Last Cal Date: June 7, 2023  
 Start time (MST): 11:51 End time (MST): 12:08

Analyzer Make: API T640 S/N: 321  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: DeltaCal S/N: 1451  
 Temp/RH standard: DeltaCal S/N: 1451

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	25.4	24.1	25.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	730.3	731.5	730.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	5.08	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>July 5, 2023</u>	Last Cal Date: <u>June 7, 2023</u>			
	PM w/o HEPA: <u>13.5</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>May 23, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>May 23, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: November 15, 2022  
 Date RH/T Sensor Cleaned: November 15, 2022

Notes: Inlet head looks good. No adjustments made. Leak check passed.

Calibration by: Sean Bala



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS11 LOWER CAMP JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 27, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:05	End Time (MST):	13:01
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.7	----	Correlation Coefficient	0.999996	
800.8	805.4	0.9942			≥0.995
400.9	402.2	0.9968	Slope	1.004778	
199.9	201.5	0.9923			0.90 - 1.10
			Intercept	0.370932	+/-30



SO2 Calibration Plot

Date: July 17, 2023

Location: Lower Camp







# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Lower Camp      Station number: AMS11  
 Calibration Date: July 18, 2023      Last Cal Date: June 28, 2023  
 Start time (MST): 9:20      End time (MST): 13:20  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.429 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC501097  
 Removed Cal Gas Conc: 5.429 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3807  
 ZAG Make/Model: API T701H      Serial Number: 196

### Analyzer Information

Analyzer make: Thermo 450iQ      Analyzer serial #: CM20080003  
 Converter make: NA      Converter serial #: NA  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.012197	1.017077	Backgd or Offset: 13.6	13.5
Calibration intercept:	0.054446	0.133554	Coeff or Slope: 1.001	1.001

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	73.6	79.9	81.8	0.977
as found 2nd point	4963	36.8	40.0	40.6	0.984
as found 3rd point	4982	18.6	20.2	20.3	0.995
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	73.6	79.9	81.4	0.982
second point	4963	36.8	40.0	40.8	0.979
third point	4982	18.6	20.2	20.7	0.976
as left zero	5000	0.0	0.0	0.1	----
as left span	4926	73.6	79.9	81.5	0.981
SO2 Scrubber Check	4919	81.1	811.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.979
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 81.8      Prev response: 80.95      \*% change: 1.0%  
 Baseline Corr 2nd AF pt: 40.6      AF Slope: 1.024650      AF Intercept: -0.206638  
 Baseline Corr 3rd AF pt: 20.3      AF Correlation: 0.999970

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 28, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:20	End Time (MST):	13:20
Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999999	
79.9	81.4	0.9818			≥0.995
40.0	40.8	0.9794	Slope	1.017077	
20.2	20.7	0.9755			0.90 - 1.10
			Intercept	0.133554	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 18, 2023

Location: Lower Camp





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	July 17, 2023	Last Cal Date:	June 27, 2023
Start time (MST):	10:05	End time (MST):	13:01
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH <sub>4</sub> Cal Gas Conc.	502.0 ppm	CH <sub>4</sub> Equiv Conc.	1067.1 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	502.0 ppm	CH <sub>4</sub> Equiv Conc.	1067.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.5 ppm		
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701	Serial Number:	3807
		Serial Number:	196

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.99E-04	2.99E-04	NMHC SP Ratio:	5.79E-05
CH <sub>4</sub> Retention time:	14.0	14.0	NMHC Peak Area:	158468
				158468

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	17.35	17.24	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	17.35	17.22	1.007
second point	4959	40.7	8.69	8.58	1.013
third point	4980	20.3	4.33	4.28	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	17.35	17.34	1.001

				Average Correction Factor	1.011
Baseline Corr AF:	17.24	Prev response	17.28	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	9.19	9.15	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	9.19	9.12	1.007
second point	4959	40.7	4.60	4.56	1.010
third point	4980	20.3	2.29	2.28	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	9.19	9.20	0.998
Average Correction Factor					1.008
Baseline Corr AF:	9.15	Prev response	9.16	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	8.16	8.09	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	8.16	8.10	1.007
second point	4959	40.7	4.09	4.02	1.016
third point	4980	20.3	2.04	2.00	1.018
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	8.16	8.14	1.003
Average Correction Factor					1.014
Baseline Corr AF:	8.09	Prev response	8.12	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997114	0.992518
THC Cal Offset:	-0.018987	-0.014592
CH <sub>4</sub> Cal Slope:	0.996346	0.993056
CH <sub>4</sub> Cal Offset:	-0.012088	-0.015088
NMHC Cal Slope:	0.997920	0.992065
NMHC Cal Offset:	-0.006899	0.000895

Notes: Changed sample inlet filter and N2 cylinder after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 27, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:05	End Time (MST):	13:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999991	$\geq 0.995$
17.35	17.22	1.0075			
8.69	8.58	1.0127			
4.33	4.28	1.0122			
			Slope	0.992518	0.90 - 1.10
			Intercept	-0.014592	+/-0.5





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 27, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:05	End Time (MST):	13:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999977	$\geq 0.995$
8.16	8.10	1.0074			
4.09	4.02	1.0158			
2.04	2.00	1.0180			
			Slope	0.993056	0.90 - 1.10
			Intercept	-0.015088	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 27, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:05	End Time (MST):	13:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	$\geq 0.995$			
9.19	9.12	1.0075						
4.60	4.56	1.0095				Slope	0.992065	0.90 - 1.10
2.29	2.28	1.0071						
			Intercept	0.000895	$\pm 0.5$			






NMHC Calibration Plot

Date: July 17, 2023

Location: Lower Camp





# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Lower Camp	Station Number:	AMS 11
Calibration Date:	July 21, 2023	Prev Cal Date:	July 15, 2022
Start Time (MST):	14:41	End Time (MST):	16:31
Tower Height (m):	10.0	Reason:	Routine

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	N11710
WS Calibrator:	MetOne 053	Serial Number:	CA05231

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.3	-0.3%
600	58.6	58.5	0.0%
800	77.8	77.7	-0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	0.999998	0.999999	<i>≥0.9995</i>
Calculated slope	0.998863	1.000455	<i>0.90 - 1.10</i>
Calculated intercept	0.064097	0.029806	<i>+/- 2</i>

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	N12035
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:32	Calc Declination*:	13.64 Degrees
Deadband calc:	3.8 degrees ( <i>Limit 4 deg</i> )	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	2.4	---
90	89.8	-0.1%
180	178.8	-0.3%
270	271.1	0.3%
354.5	355.7	0.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	0.999918	0.999950	<i>≥0.9995</i>
Calculated slope	0.998903	1.001241	<i>0.90 - 1.10</i>
Calculated intercept	-0.302646	-0.872405	<i>+/- 4</i>

Notes: WS/WD calibration. All points are within limit. Cross arm was aligned using compass after calibration.

Calibration Performed By: Mohammed Kashif



## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS13 FORT MCKAY SOUTH JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 25, 2023	Previous Calibration:	June 26, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:31	End Time (MST):	11:33
Analyzer make:	API T100	Analyzer serial #:	599

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999959	
799.7	800.2	0.9994			≥0.995
399.3	397.9	1.0035	Slope	1.002083	
200.2	195.7	1.0229			0.90 - 1.10
			Intercept	-1.997771	+/-30



SO2 Calibration Plot

Date: July 25, 2023

Location: Fort McKay South





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Fort McKay South      Station number: AMS13  
 Calibration Date: July 6, 2023      Last Cal Date: June 29, 2023  
 Start time (MST): 8:48      End time (MST): 12:50  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.34 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC500241  
 Removed Cal Gas Conc: 5.34 ppm      Rem Gas Exp Date:  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 2448  
 ZAG Make/Model: Teledyne API 701      Serial Number: 1117

### Analyzer Information

Analyzer make: Thermo 43i TLE      Analyzer serial #: 1180540017  
 Converter make: CDN-101      Converter serial #: 521  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004617	0.995542	Backgd or Offset: 3.54	3.54
Calibration intercept:	-0.182243	-0.262162	Coeff or Slope: 1.066	1.066

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4925	75.5	80.6	78.6	1.024
as found 2nd point	4962	37.7	40.3	39.1	1.027
as found 3rd point	4981	18.9	20.2	19.3	1.040
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4925	75.5	80.6	80.1	1.007
second point	4962	37.7	40.3	39.8	1.012
third point	4981	18.9	20.2	19.5	1.035
as left zero	5000	0.0	0.0	0.1	----
as left span	4925	75.5	80.6	79.4	1.015
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:	20-Mar-20		Ave Corr Factor		1.018
Date of last converter efficiency test:	NA		efficiency		

Baseline Corr As found:	78.7	Prev response:	80.82	*% change:	-2.7%
Baseline Corr 2nd AF pt:	39.2	AF Slope:	0.977254	AF Intercept:	-0.242156
Baseline Corr 3rd AF pt:	19.4	AF Correlation:	0.999983		

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	June 29, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:48	End Time (MST):	12:50
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999946	
80.6	80.1	1.0066			≥0.995
40.3	39.8	1.0117	Slope	0.995542	
20.2	19.5	1.0352			0.90 - 1.10
			Intercept	-0.262162	+/-3






TRS Calibration Plot

Date: July 6, 2023

Location: Fort McKay South





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	July 25, 2023	Last Cal Date:	June 26, 2023
Start time (MST):	8:31	End time (MST):	11:33
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH <sub>4</sub> Cal Gas Conc.	503.6 ppm	CH <sub>4</sub> Equiv Conc.	1077.5 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	503.6 ppm	CH <sub>4</sub> Equiv Conc.	1077.5 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API 701	Serial Number:	1117

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.22E-04	2.21E-04	NMHC SP Ratio:	5.10E-04
CH <sub>4</sub> Retention time:	13.0	12.8	NMHC Peak Area:	177982
				175502

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	17.05	17.00	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.06	0.999
second point	4961	39.5	8.51	8.41	1.013
third point	4980	19.8	4.27	4.12	1.035
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	17.05	17.04	1.000

Average Correction Factor	1.016
---------------------------	-------

Baseline Corr AF:	17.00	Prev response	16.98	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.1	9.08	8.98	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.1	9.08	9.09	0.999
second point	4960	39.5	4.53	4.49	1.010
third point	4980	19.8	2.27	2.20	1.032
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.1	9.08	9.06	1.002
Average Correction Factor					1.014
Baseline Corr AF:	8.98	Prev response	9.05	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	7.97	8.03	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.98	0.998
second point	4960	39.5	3.98	3.92	1.016
third point	4980	19.8	1.99	1.92	1.040
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	7.97	7.98	0.998
Average Correction Factor					1.018
Baseline Corr AF:	8.03	Prev response	7.93	*% change	1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000853	1.003313
THC Cal Offset:	-0.082577	-0.083173
CH <sub>4</sub> Cal Slope:	1.001309	1.003948
CH <sub>4</sub> Cal Offset:	-0.048553	-0.045749
NMHC Cal Slope:	1.000342	1.002735
NMHC Cal Offset:	-0.034760	-0.038763

Notes:

Changed inlet filter after as founds. Adjusted span.

Calibration Performed By:

Sean Bala



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 25, 2023	Previous Calibration:	June 26, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:31	End Time (MST):	11:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999891	$\geq 0.995$
17.05	17.06	0.9990			
8.51	8.41	1.0126	Slope	1.003313	0.90 - 1.10
4.27	4.12	1.0355			
			Intercept	-0.083173	$\pm 0.5$





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 25, 2023	Previous Calibration:	June 26, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:31	End Time (MST):	11:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999845	$\geq 0.995$
7.97	7.98	0.9985			
3.98	3.92	1.0163			
1.99	1.92	1.0398			
			Slope	1.003948	0.90 - 1.10
			Intercept	-0.045749	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 25, 2023	Previous Calibration:	June 26, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:31	End Time (MST):	11:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999917	$\geq 0.995$			
9.08	9.09	0.9994						
4.53	4.49	1.0104				Slope	1.002735	0.90 - 1.10
2.27	2.20	1.0322						
			Intercept	-0.038763	$\pm 0.5$			



NMHC Calibration Plot

Date: July 25, 2023

Location: Fort McKay South





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Fort McKay South  
Calibration Date: July 17, 2023  
Start time (MST): 9:24  
Reason: Routine  
Station number: AMS 13  
Last Cal Date: June 5, 2023  
End time (MST): 13:36

### Calibration Standards

NO Gas Cylinder #: T2Y1P76  
NOX Cal Gas Conc: 50.98 ppm  
Removed Cylinder #: N/A  
Removed Gas NOX Conc: 50.98 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701  
Cal Gas Expiry Date: December 11, 2023  
NO Cal Gas Conc: 49.32 ppm  
Removed Gas Exp Date: N/A  
Removed Gas NO Conc: 49.32 ppm  
NO gas Diff:  
Serial Number: 2448  
Serial Number: 1117

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1410661329

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.095	1.113	NO bkgnd or offset:	8.9	9.1
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	10.3	10.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	158.1	159.0

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.997811	1.000258
NO <sub>x</sub> Cal Offset:	-1.870975	-1.931590
NO Cal Slope:	1.002005	1.002649
NO Cal Offset:	-2.905185	-2.105499
NO <sub>2</sub> Cal Slope:	0.997514	1.001659
NO <sub>2</sub> Cal Offset:	0.257066	-0.999973





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	0.2	-0.5	----	----
as found span	4919	81.1	826.9	800.0	26.9	814.8	788.3	26.4	1.0148	1.0148
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.4	-0.4	----	----
high point	4919	81.1	826.9	800.0	26.9	826.0	801.1	24.8	1.0011	0.9986
second point	4960	40.6	413.9	400.4	13.5	411.5	398.5	13.0	1.0059	1.0048
third point	4980	20.3	207.0	200.2	6.7	202.9	195.9	7.1	1.0200	1.0221
as left zero	5000	0.0	0.0	0.0	0.0	-0.6	0.3	-0.9	----	----
as left span	4919	80.1	816.8	374.2	442.6	830.6	384.3	446.3	0.9834	0.9738
Average Correction Factor									1.0090	1.0085

Corrected As found	NO <sub>x</sub> = 815.1 ppb	NO = 788.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -1.0%	
Previous Response	NO <sub>x</sub> = 823.2 ppb	NO = 798.7 ppb		*Percent Change	NO = -1.3%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	797.6	381.9	442.6	442.7	0.9998	100.0%
2nd GPT point (200 ppb O <sub>3</sub> )	797.6	595.5	229.0	228.1	1.0041	99.6%
3rd GPT point (100 ppb O <sub>3</sub> )	797.6	697.0	127.5	126.1	1.0113	98.9%
Average Correction Factor					1.0051	99.5%

Notes: Adjusted span only. Used 2nd NO reference point due to drift.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 5, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:24	End Time (MST):	13:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.9	826.0	1.0011		
413.9	411.5	1.0059		
207.0	202.9	1.0200		
			0.999974	
			1.000258	
			-1.931590	

NO<sub>x</sub> Calibration Curve





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 5, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:24	End Time (MST):	13:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.0	801.1	0.9986		
400.4	398.5	1.0048		
200.2	195.9	1.0221		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 5, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:24	End Time (MST):	13:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
442.6	442.7	0.9998		
229.0	228.1	1.0041		
127.5	126.1	1.0113		
			0.999991	
			1.001659	
			-0.999973	


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: July 17, 2023

Location: Fort McKay South





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Fort McKay South      Station number: AMS13  
 Calibration Date: July 11, 2023      Last Cal Date: June 2, 2023  
 Start time (MST): 9:40      End time (MST): 13:05  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 2448  
 ZAG Make/Model: Teledyne API T701      Serial Number: 1117

### Analyzer Information

Analyzer make: Teledyne API T400      Analyzer serial #: 3871  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997857	0.996829	Backgd or Offset:	2.4	2.4
Calibration intercept:	1.000000	0.980000	Coeff or Slope:	0.963	0.963

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	5000	977.0	400.0	397.9	1.005
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.2	----
high point	5000	977.0	400.0	399.3	1.002
second point	5000	838.0	200.0	200.8	0.996
third point	5000	735.9	100.0	101.4	0.986
as left zero	5000	0.0	0.0	-0.5	----
as left span	5000	977.0	400.0	400.7	0.998
Average Correction Factor					0.995

Baseline Corr As found:	397.7	Previous response	400.1	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustments made.

Calibration Performed By: Devin Russell



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 2, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	9:40	End Time (MST):	13:05
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	0.999982	≥0.995
400.0	399.3	1.0018			
200.0	200.8	0.9960	Slope	0.996829	0.90 - 1.10
100.0	101.4	0.9862			
			Intercept	0.980000	+/- 5



O<sub>3</sub> Calibration Plot

Date: July 11, 2023

Location: Fort McKay South









## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS14  
ANZAC  
JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	June 5, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:48	End Time (MST):	12:43
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.7	----	Correlation Coefficient	≥0.995
800.2	816.4	0.9801		
399.6	406.6	0.9828	Slope	0.90 - 1.10
199.8	202.2	0.9881		
			Intercept	+/-30



SO2 Calibration Plot

Date: July 6, 2023

Location: Anzac





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	July 2, 2023	Last Cal Date:	NA
Start time (MST):	11:26	End time (MST):	16:10
Reason:	Install		

### Calibration Standards

Cal Gas Concentration:	5.38	ppm	Cal Gas Exp Date:	February 3, 2023
Cal Gas Cylinder #:	EY0000859			
Removed Cal Gas Conc:	5.38	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	5239
ZAG Make/Model:	API 701H		Serial Number:	357

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582
Converter make:	CD Nova CDN-101	Converter serial #:	503
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	0.998548	Backgd or Offset:	NA	2.20
Calibration intercept:	NA	-0.100891	Coeff or Slope:	NA	0.960

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4925	74.3	80.0	80.0	0.999
second point	4962	37.2	40.0	39.4	1.016
third point	4981	18.6	20.0	19.8	1.011
as left zero	5000	0.0	0.0	0.3	----
as left span	4925	74.3	80.0	78.8	1.015
SO2 Scrubber Check	4920	80.0	800.0	-0.4	----
Date of last scrubber change:				Ave Corr Factor	1.009
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

\* = > +/-5% change initiates investigation

Notes: Install calibration. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 2, 2023	Previous Calibration:	NA
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:26	End Time (MST):	16:10
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999888	≥0.995
80.0	80.0	0.9995			
40.0	39.4	1.0161	Slope	0.998548	0.90 - 1.10
20.0	19.8	1.0109			
			Intercept	-0.100891	+/-3



TRS Calibration Plot

Date: July 2, 2023

Location: Anzac







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	July 6, 2023	Last Cal Date:	June 5, 2023
Start time (MST):	9:48	End time (MST):	12:43
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	5239
ZAG make/model:	API 701H	Serial Number:	357

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.90E-04	3.90E-04	NMHC SP Ratio:	4.53E-05
CH <sub>4</sub> Retention time:	12.20	12.20	NMHC Peak Area:	201206
				201206

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	17.12	17.15	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	17.12	17.16	0.998
second point	4960	40.0	8.55	8.59	0.995
third point	4980	20.0	4.28	4.27	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.1	17.12	17.18	0.997

Average Correction Factor				0.998
Baseline Corr AF:	17.15	Prev response	17.07	*% change 0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	9.12	9.10	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	9.12	9.10	1.003
second point	4960	40.0	4.56	4.56	0.999
third point	4980	20.0	2.28	2.26	1.008
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.1	9.12	9.11	1.001
Average Correction Factor					1.003
Baseline Corr AF:	9.10	Prev response	9.09	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	8.00	8.05	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	8.00	8.06	0.992
second point	4960	40.0	3.99	4.03	0.990
third point	4980	20.0	2.00	2.01	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.1	8.00	8.07	0.992
Average Correction Factor					0.992
Baseline Corr AF:	8.05	Prev response	7.98	*% change	0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.996014	1.002824
THC Cal Offset:	0.012807	-0.002175
CH <sub>4</sub> Cal Slope:	0.996447	1.007994
CH <sub>4</sub> Cal Offset:	0.010369	-0.000017
NMHC Cal Slope:	0.995647	0.998090
NMHC Cal Offset:	0.002638	-0.001358

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	June 5, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:48	End Time (MST):	12:43
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	$\geq 0.995$
17.12	17.16	0.9977			
8.55	8.59	0.9949			
4.28	4.27	1.0017			
			Slope	1.002824	0.90 - 1.10
			Intercept	-0.002175	+/-0.5





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	June 5, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:48	End Time (MST):	12:43
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	$\geq 0.995$			
8.00	8.06	0.9924						
3.99	4.03	0.9902				Slope	1.007994	0.90 - 1.10
2.00	2.01	0.9946						
			Intercept	-0.000017	$\pm 0.5$			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	June 5, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:48	End Time (MST):	12:43
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999991	$\geq 0.995$			
9.12	9.10	1.0026						
4.56	4.56	0.9987				Slope	0.998090	0.90 - 1.10
2.28	2.26	1.0080						
			Intercept	-0.001358	$\pm 0.5$			



NMHC Calibration Plot

Date: July 6, 2023

Location: Anzac





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name: Anzac Station number: AMS 14  
 Calibration Date: July 28, 2023 Last Cal Date: July 6, 2023  
 Start time (MST): 9:47 End time (MST): 11:45  
 Reason: Cylinder Change Nitrogen Cylinder change

### Calibration Standards

Gas Cert Reference: CC279389 Cal Gas Expiry Date: January 5, 2025  
 CH<sub>4</sub> Cal Gas Conc. 499.3 ppm CH<sub>4</sub> Equiv Conc. 1068.8 ppm  
 C<sub>3</sub>H<sub>8</sub> Cal Gas Conc. 207.1 ppm  
 Removed Gas Cert: NA Removed Gas Expiry: NA  
 Removed CH<sub>4</sub> Conc. 499.3 ppm CH<sub>4</sub> Equiv Conc. 1068.8 ppm  
 Removed C<sub>3</sub>H<sub>8</sub> Conc. 207.1 ppm  
 Diff between cyl (CH<sub>4</sub>): Diff between cyl (NM):  
 Calibrator Model: API T700 Serial Number: 5239  
 ZAG make/model: API 701H Serial Number: 357

### Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 1118148494  
 THC Range (ppm): 0 - 20 ppm  
 NMHC Range (ppm): 0 - 10 ppm CH<sub>4</sub> Range (ppm): 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.90E-04	NA	NMHC SP Ratio:	4.53E-05
CH <sub>4</sub> Retention time:	12.20	NA	NMHC Peak Area:	201206
				NA

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	17.12	17.23	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	17.12	17.20	0.995
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.995
Baseline Corr AF:	17.23	Prev response	17.17	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	9.12	9.17	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	9.12	9.15	0.997
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.997
Baseline Corr AF:	9.17	Prev response	9.10	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.1	8.00	8.06	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.1	8.00	8.05	0.994
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.994
Baseline Corr AF:	8.06	Prev response	8.06	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002824	1.004659
THC Cal Offset:	-0.002175	0.000000
CH <sub>4</sub> Cal Slope:	1.007994	1.006423
CH <sub>4</sub> Cal Offset:	-0.000017	0.000000
NMHC Cal Slope:	0.998090	1.003112
NMHC Cal Offset:	-0.001358	0.000000

Notes: Nitrogen cylinder change.


Calibration Performed By: Mohammed Kashif



NMHC Calibration Plot

Date: July 28, 2023

Location: Anzac





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Anzac  
Calibration Date: July 7, 2023  
Start time (MST): 9:58  
Reason: Routine  
Station number: AMS 14  
Last Cal Date: June 8, 2023  
End time (MST): 14:46

### Calibration Standards

NO Gas Cylinder #: T2Y1P8D  
NOX Cal Gas Conc: 50.92 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 50.92 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API 701H  
Cal Gas Expiry Date: December 11, 2023  
NO Cal Gas Conc: 50.05 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 50.05 ppm  
NO gas Diff:  
Serial Number: 5239  
Serial Number: 357

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1426262592

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.375	1.375	NO bkgnd or offset:	3.7	3.7
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.7	3.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	162.1	163.9

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000547	0.995865
NO <sub>x</sub> Cal Offset:	-0.105430	0.033705
NO Cal Slope:	1.001604	0.996593
NO Cal Offset:	-1.609467	-1.310286
NO <sub>2</sub> Cal Slope:	1.002084	1.000260
NO <sub>2</sub> Cal Offset:	1.232321	1.094049



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.2	0.3	----	----
as found span	4921	78.6	800.5	786.8	13.7	797.4	781.6	15.8	1.0039	1.0067
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.4	0.0	0.3	----	----
high point	4921	78.6	800.5	786.8	13.7	797.3	783.4	13.9	1.0040	1.0044
second point	4961	39.3	400.2	393.4	6.8	398.8	390.4	8.4	1.0035	1.0076
third point	4980	19.6	199.6	196.2	3.4	198.2	192.7	5.5	1.0072	1.0182
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.3	----	----
as left span	4921	78.6	800.5	393.8	406.7	794.7	387.6	407.2	1.0073	1.0161
Average Correction Factor									1.0049	1.0101

Corrected As found	NO <sub>x</sub> = 797.3 ppb	NO = 781.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.4%	
Previous Response	NO <sub>x</sub> = 800.9 ppb	NO = 786.5 ppb		*Percent Change	NO = -0.6%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	778.0	385.0	406.7	407.3	0.9985	100.2%
2nd GPT point (200 ppb O <sub>3</sub> )	778.0	576.3	215.4	217.3	0.9912	100.9%
3rd GPT point (100 ppb O <sub>3</sub> )	778.0	677.5	114.2	115.9	0.9851	101.5%
Average Correction Factor					0.9916	100.9%

Notes:

Sample inlet filter changed after as founds. No adjustments made.

Mohammed Kashif

CALS\_307



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 7, 2023	Previous Calibration:	June 8, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:58	End Time (MST):	14:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.5	797.3	1.0040		
400.2	398.8	1.0035		
199.6	198.2	1.0072		





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 7, 2023	Previous Calibration:	June 8, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:58	End Time (MST):	14:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
786.8	783.4	1.0044		
393.4	390.4	1.0076		
196.2	192.7	1.0182		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 7, 2023	Previous Calibration:	June 8, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:58	End Time (MST):	14:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
406.7	407.3	0.9985		
215.4	217.3	0.9912		
114.2	115.9	0.9851		
			0.999979	
			1.000260	
			1.094049	


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: July 7, 2023

Location: Anzac









# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	June 7, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	9:46	End Time (MST):	12:31
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999999	≥0.995
400.0	396.1	1.0098			
200.0	198.2	1.0091	Slope	0.989000	0.90 - 1.10
100.0	99.6	1.0040			
			Intercept	0.500000	+/- 5



O<sub>3</sub> Calibration Plot

Date: July 11, 2023

Location: Anzac





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Anzac Station number: AMS 14  
 Calibration Date: July 6, 2023 Last Cal Date: June 14, 2023  
 Start time (MST): 12:55 End time (MST): 13:39

Analyzer Make: API T640 S/N: 825  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388749  
 Temp/RH standard: Alicat FP-25 S/N: 388749

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	21.6	23.1	21.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	716.4	717.5	716.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	4.98	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>July 6, 2023</u>	Last Cal Date: <u>June 14, 2023</u>			
	PM w/o HEPA: <u>27.7</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.7	10	10.9	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>23.1</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>July 6, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 6, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: July 6, 2023  
 Date RH/T Sensor Cleaned: July 6, 2023

PMT adjusted after cleaning. Leak check passed. Completed quarterly and annual maintenance.  
 Notes: Head Cleaned.

Calibration by: Mohammed Kashif



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS17  
WAPASU  
JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 5, 2023	Previous Calibration:	June 5, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:50	End Time (MST):	14:30
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999980
800.0	805.6	0.9930		
400.0	400.6	0.9986	Slope	1.008014
199.5	197.9	1.0082		
			Intercept	-1.662517
				+/-30



SO2 Calibration Plot

Date: July 5, 2023

Location: Wapasu





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	July 17, 2023	Last Cal Date:	June 21, 2023
Start time (MST):	10:42	End time (MST):	16:06
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.076	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511852			
Removed Cal Gas Conc:	5.076	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2449
ZAG Make/Model:	API T701H		Serial Number:	359

### Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583
Converter make:	n/a	Converter serial #:	n/a
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997854	0.996997	Backgd or Offset:	13.2
Calibration intercept:	-0.199211	-0.319220	Coeff or Slope:	1.115
				1.134

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4921	78.8	80.0	79.5	1.010
as found 2nd point	4961	39.4	40.0	41.1	0.980
as found 3rd point	4980	19.7	20.0	20.7	0.980
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4921	78.8	80.0	79.6	1.005
second point	4961	39.4	40.0	39.2	1.020
third point	4980	19.7	20.0	19.7	1.015
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	78.8	80.0	77.1	1.038
SO2 Scrubber Check	4921	79.4	800.0	0.1	----
Date of last scrubber change:		n/a		Ave Corr Factor	1.014
Date of last converter efficiency test:		n/a			efficiency

Baseline Corr As found:	79.2	Prev response:	79.63	*% change:	-0.5%
Baseline Corr 2nd AF pt:	40.8	AF Slope:	0.989138	AF Intercept:	0.780827
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999710		

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 21, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:42	End Time (MST):	16:06
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999950	≥0.995
80.0	79.6	1.0050			
40.0	39.2	1.0203	Slope	0.996997	0.90 - 1.10
20.0	19.7	1.0153			
			Intercept	-0.319220	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 17, 2023

Location: Wapasu





# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	July 5, 2023	Last Cal Date:	June 5, 2023
Start time (MST):	9:50	End time (MST):	14:30
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	ALM066507	Cal Gas Expiry Date:	January 12, 2029
CH4 Cal Gas Conc.	<u>503.5</u> ppm	CH4 Equiv Conc.	1076.3 ppm
C3H8 Cal Gas Conc.	<u>208.3</u> ppm		
Removed Gas Cert:	n/a	Removed Gas Expiry:	n/a
Removed CH4 Conc.	<u>503.5</u> ppm	CH4 Equiv Conc.	1076.3 ppm
Removed C3H8 Conc.	<u>208.3</u> ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	2449
ZAG Make/Model:	API 701H	Serial Number:	359

### Analyzer Information

Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352
Analyzer Range:	0 - 20 ppm		

Calibration slope:	<u>Start</u> 0.998402	<u>Finish</u> 0.994932	Background:	<u>Start</u> 3.210	<u>Finish</u> 3.300
Calibration intercept:	-0.031959	-0.001771	Coefficient:	4.483	4.460

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.15	----
as found span	4921	79.4	17.09	15.54	1.100
as found 2nd point	4960	39.7	8.55	7.65	1.118
as found 3rd point	4980	19.8	4.26	3.74	1.140
new cylinder response					
calibrator zero	5000	0.0	0.00	0.03	----
high point	4921	79.4	17.09	17.02	1.004
second point	4960	39.7	8.55	8.49	1.007
third point	4980	19.8	4.26	4.21	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.4	17.09	17.06	1.002
Average Correction Factor					1.008
Baseline Corr As found:	15.69	Previous response	17.03	*% change	-8.5%
Baseline Corr 2nd AF pt:	7.79	AF Slope:	0.918313	AF Intercept:	-0.169055
Baseline Corr 3rd AF pt:	3.89	AF Correlation:	0.999985		

\* = > +/-5% change initiates investigation

Notes: Inlet filter, H2 cylinder, and pump replaced after as founds. Zero and span adjusted.

Calibration Performed By: Devin Russell



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 5, 2023	Previous Calibration:	June 5, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:50	End Time (MST):	14:30
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.03	----	Correlation Coefficient	0.999987	
17.09	17.02	1.0043			≥0.995
8.55	8.49	1.0071	Slope	0.994932	
4.26	4.21	1.0122			0.90 - 1.10
			Intercept	-0.001771	+/-1.5



THC Calibration Plot

Date: July 5, 2023

Location: Wapasu





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Wapasu  
Calibration Date: July 25, 2023  
Start time (MST): 10:07  
Reason: Install  
Station number: AMS17  
Last Cal Date: June 7, 2023  
End time (MST): 17:56

### Calibration Standards

NO Gas Cylinder #: T375YK8  
NOX Cal Gas Conc: 49.11 ppm  
Removed Cylinder #: T375YK8  
Removed Gas NOX Conc: 49.11 ppm  
NOX gas Diff: T375YK8  
Calibrator Model: API T700  
ZAG make/model: API T701H  
Cal Gas Expiry Date: April 13, 2025  
NO Cal Gas Conc: 48.07 ppm  
Removed Gas Exp Date: T375YK8  
Removed Gas NO Conc: 48.07 ppm  
NO gas Diff: T375YK8  
Serial Number: 2449  
Serial Number: 359

### Analyzer Information

Analyzer make: Thermo Scientific 42iQ  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 12300522720

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.932	1.205	NO bkgnd or offset:	2.4	10.1
NOX coeff or slope:	0.984	0.996	NOX bkgnd or offset:	2.4	10.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	141.3	430.0

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.003075	0.988306
NO <sub>x</sub> Cal Offset:	-4.520000	-0.640000
NO Cal Slope:	1.003530	0.986171
NO Cal Offset:	-5.460000	-1.460000
NO <sub>2</sub> Cal Slope:	1.006000	1.024572
NO <sub>2</sub> Cal Offset:	-0.618680	-0.186089



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.4	0.3	0.1	----	----
as found span	4917	83.2	817.2	799.9	17.3	793.6	776.7	16.9	1.0297	1.0299
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.1	----	----
high point	4917	83.2	817.2	799.9	17.3	806.7	787.5	19.2	1.0130	1.0157
second point	4958	41.6	408.6	399.9	8.7	405.0	394.3	10.7	1.0089	1.0143
third point	4979	20.8	204.3	200.0	4.3	198.8	192.5	6.3	1.0277	1.0388
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.3	0.1	----	----
as left span	4917	83.2	817.2	393.9	423.3	813.0	388.7	424.3	1.0051	1.0133
Average Correction Factor									1.0165	1.0229

Corrected As found	NO <sub>x</sub> = 793.2 ppb	NO = 776.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -2.8%
Previous Response	NO <sub>x</sub> = 815.2 ppb	NO = 797.2 ppb		*Percent Change	NO = -2.7%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.3	392.3	423.3	433.9	0.9756	102.5%
2nd GPT point (200 ppb O3)	798.3	593.1	222.5	227.0	0.9802	102.0%
3rd GPT point (100 ppb O3)	798.3	695.2	120.4	123.3	0.9765	102.4%
Average Correction Factor					0.9774	102.3%

Notes:

Zero and span adjusted after pump changout.

Calibration Performed By:

Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 25, 2023	Previous Calibration:	June 7, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:07	End Time (MST):	17:56
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
817.2	806.7	1.0130		
408.6	405.0	1.0089		
204.3	198.8	1.0277		
			0.999971	
			0.988306	
			-0.640000	







# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 25, 2023	Previous Calibration:	June 7, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:07	End Time (MST):	17:56
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.9	787.5	1.0157		
399.9	394.3	1.0143		
200.0	192.5	1.0388		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 25, 2023	Previous Calibration:	June 7, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:07	End Time (MST):	17:56
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient 0.999992	≥0.995	
423.3	433.9	0.9756			
222.5	227.0	0.9802			
120.4	123.3	0.9765			
			Slope	1.024572	0.90 - 1.10
			Intercept	-0.186089	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: July 25, 2023

Location: Wapasu





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	July 12, 2023	Last Cal Date:	June 1, 2023
Start time (MST):	9:41	End time (MST):	13:20
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	API T700	Serial Number:	2449
ZAG Make/Model:	API T701H	Serial Number:	359

### Analyzer Information

Analyzer make:	API T400	Analyzer serial #:	3870
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004914	1.000171	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.460000	-0.580000	Coeff or Slope:	1.020	1.020

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.5	----
as found span	5000	1077.3	400.0	397.5	1.006
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.2	----
high point	5000	1077.3	400.0	399.6	1.001
second point	5000	900.3	200.0	199.5	1.003
third point	5000	789.5	100.0	98.9	1.011
as left zero	5000	0.0	0.0	0.0	----
as left span	5000	1077.3	400.0	404.3	0.989
Average Correction Factor					1.005

Baseline Corr As found:	398.0	Previous response	401.5	*% change	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 12, 2023	Previous Calibration:	June 1, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:41	End Time (MST):	13:20
Analyzer make:	API T400	Analyzer serial #:	3870

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999995	≥0.995
400.0	399.6	1.0010			
200.0	199.5	1.0025	Slope	1.000171	0.90 - 1.10
100.0	98.9	1.0111			
			Intercept	-0.580000	+/- 5



O<sub>3</sub> Calibration Plot

Date: July 12, 2023

Location: Wapasu





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Wapasu Station number: AMS 17  
 Calibration Date: July 27, 2023 Last Cal Date: June 21, 2023  
 Start time (MST): 12:23 End time (MST): 13:33

Analyzer Make: API T640 S/N: 1183  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748  
 Temp/RH standard: Alicat FP-25BT S/N: 388748

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	16.9	16.6	16.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	719.4	722.4	719.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.46	4.51	5.00	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: July 27, 2023 Last Cal Date: June 21, 2023  
 PM w/o HEPA: 5.7 PM w/ HEPA: 0.0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.9	10.9	10.9	<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: 7 w/ HEPA: 0.0  
 Date Optical Chamber Cleaned: July 27, 2023 <0.2 ug/m3  
 Disposable Filter Changed: July 27, 2023

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_  
 Date RH/T Sensor Cleaned: \_\_\_\_\_

Notes: Temp, pressure and flow checked. Flow was low, changing out filter increased pump output to expected levels. Leak check passed.

Calibration by: Aswin Sasi Kumar



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS18 STONY MOUNTAIN JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023







# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 15, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:51	End Time (MST):	13:08
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999972	
800.3	802.1	0.9977			≥0.995
400.2	403.9	0.9908	Slope	1.003290	
199.6	198.4	1.0061			0.90 - 1.10
			Intercept	-0.043536	+/-30



SO2 Calibration Plot

Date: July 18, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Stony Mountain      Station number: AMS18  
 Calibration Date: July 19, 2023      Last Cal Date: June 27, 2023  
 Start time (MST): 9:47      End time (MST): 15:30  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.479 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC500395  
 Removed Cal Gas Conc: 5.479 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 2658  
 ZAG Make/Model: Teledyne API T701      Serial Number: 360

### Analyzer Information

Analyzer make: Thermo 43i-TLE      Analyzer serial #: 1218153359  
 Converter make: CD Nova CDN-101      Converter serial #: 555  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003297	1.003873	Backgd or Offset: 2.70	2.66
Calibration intercept:	0.101073	0.080889	Coeff or Slope: 1.189	1.189

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4927	73.0	80.0	79.7	1.006
as found 2nd point	4964	36.5	40.0	39.4	1.020
as found 3rd point	4983	18.3	20.0	19.2	1.055
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4927	73.0	80.0	80.5	0.994
second point	4964	36.5	40.0	40.0	1.000
third point	4983	18.3	20.0	20.2	0.992
as left zero	5000	0.0	0.0	0.4	----
as left span	4927	73.0	80.0	80.3	0.996
SO2 Scrubber Check	4923	77.1	771.0	0.0	----
Date of last scrubber change:	17-Dec-21			Ave Corr Factor	0.995
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.5      Prev response: 80.36      \*% change: -1.1%  
 Baseline Corr 2nd AF pt: 39.2      AF Slope: 0.997007      AF Intercept: -0.278730  
 Baseline Corr 3rd AF pt: 19.0      AF Correlation: 0.999834

\* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Scrubber check completed after calibrator zero. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 19, 2023	Previous Calibration:	June 27, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	9:47	End Time (MST):	15:30
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999977	≥0.995
80.0	80.5	0.9937			
40.0	40.0	0.9998	Slope	1.003873	0.90 - 1.10
20.0	20.2	0.9925			
			Intercept	0.080889	+/-3



TRS Calibration Plot

Date: July 19, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	July 18, 2023	Last Cal Date:	June 15, 2023
Start time (MST):	9:51	End time (MST):	13:08
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH <sub>4</sub> Cal Gas Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.17E-04	3.21E-04	NMHC SP Ratio:	5.97E-05
CH <sub>4</sub> Retention time:	15.20	15.20	NMHC Peak Area:	153659
				147573

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.02	----
as found span	4919	81.0	17.28	16.82	1.028
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4919	81.0	17.28	17.26	1.001
second point	4959	40.5	8.64	8.62	1.003
third point	4979	20.2	4.31	4.30	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.27	1.001

				Average Correction Factor	1.002
Baseline Corr AF:	16.80	Prev response	17.29	*% change	-2.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	8.82	1.040
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.15	1.002
second point	4959	40.5	4.58	4.57	1.003
third point	4979	20.2	2.29	2.29	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81	9.17	9.12	1.006
Average Correction Factor					1.001
Baseline Corr AF:	8.82	Prev response	9.18	*% change	-4.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.02	----
as found span	4919	81.0	8.11	8.00	1.014
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4919	81.0	8.11	8.11	1.000
second point	4959	40.5	4.06	4.05	1.003
third point	4979	20.2	2.02	2.01	1.008
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.15	0.995
Average Correction Factor					1.004
Baseline Corr AF:	7.98	Prev response	8.11	*% change	-1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000273	0.998376
THC Cal Offset:	0.004621	0.001211
CH <sub>4</sub> Cal Slope:	0.999876	0.999482
CH <sub>4</sub> Cal Offset:	-0.002412	-0.003013
NMHC Cal Slope:	1.000625	0.997360
NMHC Cal Offset:	0.007033	0.003625

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar





# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 15, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:51	End Time (MST):	13:08
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999998	$\geq 0.995$
17.28	17.26	1.0012			
8.64	8.62	1.0028			
4.31	4.30	1.0019			
			Slope	0.998376	0.90 - 1.10
			Intercept	0.001211	+/-0.5





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 15, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:51	End Time (MST):	13:08
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999990	≥0.995
8.11	8.11	1.0001			
4.06	4.05	1.0029			
2.02	2.01	1.0077			
			Slope	0.999482	0.90 - 1.10
			Intercept	-0.003013	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 15, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:51	End Time (MST):	13:08
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	$\geq 0.995$			
9.17	9.15	1.0021						
4.58	4.57	1.0034				Slope	0.997360	0.90 - 1.10
2.29	2.29	0.9969						
			Intercept	0.003625	$\pm 0.5$			



NMHC Calibration Plot

Date: July 18, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Stony Mountain Station number: AMS 18  
Calibration Date: July 26, 2023 Last Cal Date: June 28, 2023  
Start time (MST): 10:38 End time (MST): 16:01  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T2XX7ME Cal Gas Expiry Date: January 14, 2024  
NOX Cal Gas Conc: 50.48 ppm NO Cal Gas Conc: 49.22 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 50.48 ppm Removed Gas NO Conc: 49.22 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 2658  
ZAG make/model: Teledyne API 701H Serial Number: 360

### Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1336160088  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.964	0.993	NO bkgnd or offset:	2.7	2.8
NOX coeff or slope:	0.987	0.984	NOX bkgnd or offset:	2.7	2.8
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	225.1	230.6

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.001389	0.998716
NO <sub>x</sub> Cal Offset:	0.070063	-0.069984
NO Cal Slope:	1.001010	1.001009
NO Cal Offset:	-1.269997	-1.469781
NO <sub>2</sub> Cal Slope:	1.004754	1.000734
NO <sub>2</sub> Cal Offset:	0.176637	-0.286696



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
as found span	4919	81.3	820.8	800.3	20.5	800.0	774.7	25.5	1.0259	1.0330
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
high point	4919	81.3	820.8	800.3	20.5	819.7	800.3	19.5	1.0013	1.0000
second point	4959	40.7	410.9	400.7	10.3	410.4	399.1	11.2	1.0013	1.0039
third point	4980	20.3	204.9	199.8	5.1	204.2	196.8	7.5	1.0036	1.0154
as left zero	5000	0.0	0.0	0.0	0.0	0.3	0.0	0.3	----	----
as left span	4919	81.3	820.8	362.9	457.9	822.3	361.2	461.1	0.9981	1.0046
Average Correction Factor									1.0021	1.0064

Corrected As found	NO <sub>x</sub> = 799.8 ppb	NO = 774.7 ppb	<i>* =&gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -2.8%
Previous Response	NO <sub>x</sub> = 822.0 ppb	NO = 799.8 ppb			*Percent Change	NO = -3.2%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.4	360.0	457.9	458.0	0.9998	100.0%
2nd GPT point (200 ppb O3)	797.4	595.0	222.9	223.0	0.9995	100.1%
3rd GPT point (100 ppb O3)	797.4	698.5	119.4	118.5	1.0075	99.3%
Average Correction Factor					1.0022	99.8%

Notes:

Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 26, 2023	Previous Calibration:	June 28, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:38	End Time (MST):	16:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
820.8	819.7	1.0013			
410.9	410.4	1.0013			
204.9	204.2	1.0036			
			Slope	0.998716	0.90 - 1.10
			Intercept	-0.069984	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 26, 2023	Previous Calibration:	June 28, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:38	End Time (MST):	16:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	800.3	1.0000		
400.7	399.1	1.0039		
199.8	196.8	1.0154		

NO Calibration Curve







# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 26, 2023	Previous Calibration:	June 28, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:38	End Time (MST):	16:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
457.9	458.0	0.9998			
222.9	223.0	0.9995			
119.4	118.5	1.0075			
			Slope	1.000734	0.90 - 1.10
			Intercept	-0.286696	+/-20



NO<sub>x</sub> Calibration Plot

Date: July 26, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Stony Mountain      Station number: AMS18  
 Calibration Date: July 26, 2023      Last Cal Date: June 28, 2023  
 Start time (MST): 10:58      End time (MST): 16:15  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T750      Serial Number: 282  
 ZAG Make/Model: Teledyne API T751H      Serial Number: 321

### Analyzer Information

Analyzer make: API T400      Analyzer serial #: 825  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005457	1.007600	Backgd or Offset:	1.000	1.300
Calibration intercept:	-0.380000	-1.380000	Coeff or Slope:	0.995	1.002

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	-1.0	----
as found span	4888	1096.9	400.0	399.9	1.000
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.2	----
high point	4888	1101.7	400.0	402.5	0.994
second point	4888	863.9	200.0	199.2	1.004
third point	4888	741.4	100.0	97.9	1.021
as left zero	5000	800.0	0.0	0.3	----
as left span	4812	1097.9	400.0	407.9	0.981
Average Correction Factor					1.006

Baseline Corr As found:	400.9	Previous response	401.8	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Using portable calibrator for the calibration. Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 26, 2023	Previous Calibration:	June 28, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:58	End Time (MST):	16:15
Analyzer make:	API T400	Analyzer serial #:	825

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999929	
400.0	402.5	0.9938			≥0.995
200.0	199.2	1.0040	Slope	1.007600	
100.0	97.9	1.0215			0.90 - 1.10
			Intercept	-1.380000	+/- 5



O<sub>3</sub> Calibration Plot

Date: July 26, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Stony Mountain Station number: AMS 18  
 Calibration Date: July 26, 2023 Last Cal Date: June 28, 2023  
 Start time (MST): 14:23 End time (MST): 16:02

Analyzer Make: API T640 S/N: 1162  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748  
 Temp/RH standard: Alicat FP-25BT S/N: 388748

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	16.3	16.07	16.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	697.3	697.05	697.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	5.00	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: July 26, 2023	Last Cal Date: May 16, 2023			
	PM w/o HEPA: 2.8	PM w/ HEPA: 0.0			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.0	10.0	10.0	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: 2.4	w/ HEPA: 0.0		
Date Optical Chamber Cleaned:		July 26, 2023			<0.2 ug/m3
Disposable Filter Changed:		July 26, 2023			

### Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022  
 Date RH/T Sensor Cleaned: August 30, 2022

Notes: No adjustments needed.

Calibration by: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	July 4, 2023	Last Cal Date:	June 8, 2023
Start time (MST):	10:37	End time (MST):	13:49
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,050	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	3,050	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701		Serial Number:	360

### Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	<b>Start</b>	<b>Finish</b>		<b>Start</b>	<b>Finish</b>
Calibration slope:	0.999102	0.999550	Backgd or Offset:	-0.010	-0.010
Calibration intercept:	0.057760	0.039792	Coeff or Slope:	0.906	0.906

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4933	66.7	40.7	40.6	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.7	40.6	1.001
second point	4966	33.3	20.3	20.5	0.990
third point	4983	16.7	10.2	10.2	1.003
as left zero	3000	0.0	0.0	0.0	----
as left span	2960	40.0	40.7	40.7	0.999
Average Correction Factor					0.998

Baseline Corr As found:	40.57	Prev response:	40.71	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 8, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:37	End Time (MST):	13:49
Analyzer make:	API T300	Analyzer serial #:	3504

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999955
40.7	40.6	1.0012		
20.3	20.5	0.9901	Slope	0.999550
10.2	10.2	1.0027		
			Intercept	0.039792
				<i>+/-1.5</i>






CO Calibration Plot

Date: July 4, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	July 5, 2023	Last Cal Date:	June 14, 2023
Start time (MST):	10:36	End time (MST):	14:14
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2026
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
N2 Gen Make/Model:	Peak Scientific		Serial Number:	771048317

### Analyzer Information

Analyzer make:	API T360	Analyzer serial #:	283
Analyzer Range	0 - 2,000 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997863	0.999115	Backgd or Offset:	-0.081	-0.103
Calibration intercept:	-0.740000	-4.820000	Coeff or Slope:	1.091	1.109

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	1.4	----
as found span	2920	80.0	1605.9	1644.2	0.977
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	-0.1	----
high point	2920	80.0	1605.9	1601.0	1.003
second point	2960	40.0	802.9	798.0	1.006
third point	2980	20.0	401.5	389.6	1.030
as left zero	3000	0.0	0.0	0.8	----
as left span	2930	80.0	1600.5	1602.0	0.999
Average Correction Factor					1.013

Baseline Corr As found:	1642.80	Prev response:	1601.69	*% change:	2.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date	July 5, 2023	Previous Calibration	June 14, 2023
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:36	End Time (MST)	14:14
Analyzer make	API T360	Analyzer serial #	283

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999951	≥0.995
1605.9	1601.0	1.0030			
802.9	798.0	1.0062	Slope	0.999115	0.90 - 1.10
401.5	389.6	1.0305			
			Intercept	-4.820000	+/-10



CO<sub>2</sub> Calibration Plot

Date: July 5, 2023

Location: Stony Mountain





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS19  
FIREBAG  
JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	July 19, 2023	Last Cal Date:	June 20, 2023
Start time (MST):	9:58	End time (MST):	15:13
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.29	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC716618			
Removed Cal Gas Conc:	49.29	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1607
ZAG Make/Model:	API T701		Serial Number:	1118

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998790	0.995832	Backgd or Offset:	10.3	10.2
Calibration intercept:	-0.461656	-0.101381	Coeff or Slope:	0.992	0.992

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	-0.2	----
as found span	4919	81.1	799.5	796.6	1.004
as found 2nd point	4959	40.6	400.3	399.6	1.002
as found 3rd point	4980	20.3	200.1	199.6	1.003
new cylinder response					
calibrator zero	4999	0.0	0.0	-0.1	----
high point	4919	81.1	799.5	795.8	1.005
second point	4959	40.6	400.3	399.2	1.003
third point	4980	20.3	200.1	198.7	1.007
as left zero	4999	0.0	0.0	-0.1	----
as left span	4919	81.1	799.5	797.0	1.003
Average Correction Factor					1.005

Baseline Corr As found:	796.80	Previous response	798.04	*% change	-0.2%
Baseline Corr 2nd AF pt:	399.80	AF Slope:	0.996633	AF Intercept:	0.118409
Baseline Corr 3rd AF pt:	199.80	AF Correlation:	0.999999		

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 19, 2023	Previous Calibration:	June 20, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:58	End Time (MST):	15:13
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999998	
799.5	795.8	1.0046			≥0.995
400.3	399.2	1.0027	Slope	0.995832	
200.1	198.7	1.0071			0.90 - 1.10
			Intercept	-0.101381	+/-30






SO2 Calibration Plot

Date: July 19, 2023

Location: Firebag





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Firebag	Station number:	AMS19
Calibration Date:	July 18, 2023	Last Cal Date:	June 30, 2023
Start time (MST):	10:02	End time (MST):	14:08
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.114	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517427			
Removed Cal Gas Conc:	5.114	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1607
ZAG Make/Model:	Teledyne API T701		Serial Number:	1118

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032
Converter make:	Global	Converter serial #:	2022-222
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002911	1.004340	Backgd or Offset:	2.54
Calibration intercept:	-0.281624	-0.181643	Coeff or Slope:	1.122

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4922	78.2	80.0	81.2	0.983
as found 2nd point	4961	39.1	40.0	40.4	0.985
as found 3rd point	4980	19.6	20.0	20.1	0.988
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4922	78.2	80.0	80.1	0.998
second point	4961	39.1	40.0	40.1	0.997
third point	4980	19.6	20.0	19.9	1.007
as left zero	5000	0.0	0.0	0.0	----
as left span	4922	78.2	80.0	79.8	1.002
SO2 Scrubber Check	4922	78.3	800.2	0.0	----
Date of last scrubber change:	January 18, 2023			Ave Corr Factor	1.001
Date of last converter efficiency test:	n/a			efficiency	

Baseline Corr As found:	81.4	Prev response:	79.93	*% change:	1.8%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.018061	AF Intercept:	-0.261941
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999997		

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter and the external pump after MPAF's. Ran SOx scrubber check after MPAF's. No adjustments needed.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 30, 2023
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	10:02	End Time (MST):	14:08
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999995	≥0.995
80.0	80.1	0.9985			
40.0	40.1	0.9973	Slope	1.004340	0.90 - 1.10
20.0	19.9	1.0075			
			Intercept	-0.181643	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 18, 2023

Location: Firebag





# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	July 19, 2023	Last Cal Date:	June 20, 2023
Start time (MST):	9:58	End time (MST):	15:13
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1607
ZAG Make/Model:	API T701	Serial Number:	1118

### Analyzer Information

Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089
Analyzer Range:	0 - 20 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003320	0.999589	Background:	2.69	3.75
Calibration intercept:	-0.013586	-0.108328	Coefficient:	3.908	3.716

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.32	----
as found span	4919	81.1	17.31	17.54	0.987
as found 2nd point	4959	40.6	8.66	8.77	0.988
as found 3rd point	4980	20.3	4.33	4.64	0.933
new cylinder response					
calibrator zero	4999	0.0	0.00	-0.04	----
high point	4919	81.1	17.31	17.23	1.004
second point	4959	40.6	8.66	8.50	1.019
third point	4980	20.3	4.33	4.17	1.040
as left zero	5000	0.0	0.00	-0.20	----
as left span	4919	81.1	17.31	17.15	1.009
Average Correction Factor					1.021
Baseline Corr As found:	17.22	Previous response	17.35	*% change	-0.8%
Baseline Corr 2nd AF pt:	8.45	AF Slope:	0.993498	AF Intercept:	0.293316
Baseline Corr 3rd AF pt:	4.32	AF Correlation:	0.999856		

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Changed internal pump, adjusted zero and span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 19, 2023	Previous Calibration:	June 20, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:58	End Time (MST):	15:13
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	-0.04	----	Correlation Coefficient	0.999925	≥0.995
17.31	17.23	1.0044			
8.66	8.50	1.0193	Slope	0.999589	0.90 - 1.10
4.33	4.17	1.0400			
			Intercept	-0.108328	+/-1.5



THC Calibration Plot

Date: July 19, 2023

Location: Firebag





# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	July 24, 2023	Last Cal Date:	July 19, 2023
Start time (MST):	10:28	End time (MST):	13:24
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1607
ZAG Make/Model:	API T701	Serial Number:	1118

### Analyzer Information

Analyzer make: Thermo 51i-LT                      Analyzer serial #: 1336160089  
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999589	1.005161	Background:	3.75	3.00
Calibration intercept:	-0.108328	-0.081786	Coefficient:	3.716	3.775

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.81	----
as found span	4919	81.1	17.31	16.25	1.065
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.00	-0.08	----
high point	4919	81.1	17.31	17.32	0.999
second point	4959	40.6	8.66	8.61	1.006
third point	4980	20.3	4.33	4.28	1.013
as left zero	5000	0.0	0.00	-0.02	----
as left span	4919	81.1	17.31	17.39	0.995
Average Correction Factor					1.006
Baseline Corr As found:	17.06	Previous response	17.19	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Adjusted zero and span after baseline drift.

Calibration Performed By: Braiden Boutillier





# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 24, 2023	Previous Calibration:	July 19, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:28	End Time (MST):	13:24
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<i>Limits</i>	
0.00	-0.08	----	Correlation Coefficient	0.999998	
17.31	17.32	0.9991			≥0.995
8.66	8.61	1.0063	Slope	1.005161	
4.33	4.28	1.0127			0.90 - 1.10
			Intercept	-0.081786	+/-1.5



THC Calibration Plot

Date: July 24, 2023

Location: Firebag





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Firebag  
Calibration Date: July 27, 2023  
Start time (MST): 9:40  
Reason: Routine  
Station number: AMS 19  
Last Cal Date: June 22, 2023  
End time (MST): 14:01

### Calibration Standards

NO Gas Cylinder #: T2Y1K63  
NOX Cal Gas Conc: 51.12 ppm  
Removed Cylinder #: n/a  
Removed Gas NOX Conc: 51.12 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701  
Cal Gas Expiry Date: November 30, 2023  
NO Cal Gas Conc: 49.40 ppm  
Removed Gas Exp Date: n/a  
Removed Gas NO Conc: 49.40 ppm  
NO gas Diff:  
Serial Number: 1607  
Serial Number: 1118

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1410661309

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.094	1.106	NO bkgnd or offset:	7.6	7.7
NOX coeff or slope:	0.994	0.993	NOX bkgnd or offset:	7.6	7.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	211.5	214.0

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.998084	0.996456
NO <sub>x</sub> Cal Offset:	1.096129	1.235903
NO Cal Slope:	0.999369	0.999440
NO Cal Offset:	0.508726	0.208843
NO <sub>2</sub> Cal Slope:	1.001853	0.999338
NO <sub>2</sub> Cal Offset:	-0.516592	0.066857



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as found span	4919	81.0	828.1	800.3	27.9	821.0	790.3	30.6	1.0087	1.0126
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	4999	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
high point	4919	81.0	828.1	800.3	27.9	826.0	800.0	26.0	1.0026	1.0004
second point	4960	40.5	414.0	400.1	13.9	414.1	400.0	14.2	0.9998	1.0002
third point	4980	20.2	206.5	199.6	6.9	208.2	200.0	8.2	0.9919	0.9978
as left zero	4999	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
as left span	4919	81.0	828.1	363.5	464.7	828.0	362.2	465.7	1.0002	1.0035
Average Correction Factor									0.9981	0.9995

Corrected As found	NO <sub>x</sub> = 821.1 ppb	NO = 790.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.8%	
Previous Response	NO <sub>x</sub> = 827.7 ppb	NO = 800.3 ppb		*Percent Change	NO = -1.2%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.9	361.1	464.7	464.3	1.0008	99.9%
2nd GPT point (200 ppb O3)	797.9	580.3	245.5	245.9	0.9982	100.2%
3rd GPT point (100 ppb O3)	797.9	690.6	135.2	134.6	1.0042	99.6%
Average Correction Factor					1.0011	99.9%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	June 22, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:40	End Time (MST):	14:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
828.1	826.0	1.0026			
414.0	414.1	0.9998			
206.5	208.2	0.9919			
			Slope	0.996456	0.90 - 1.10
			Intercept	1.235903	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	June 22, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:40	End Time (MST):	14:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	800.0	1.0004		
400.1	400.0	1.0002		
199.6	200.0	0.9978		
			1.000000	
			0.999440	
			0.208843	

NO Calibration Curve





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	June 22, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:40	End Time (MST):	14:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
464.7	464.3	1.0008			
245.5	245.9	0.9982			
135.2	134.6	1.0042			
			Slope	0.999338	0.90 - 1.10
			Intercept	0.066857	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: July 27, 2023

Location: Firebag







## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS20 MACKAY RIVER JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: MacKay River Station number: AMS20  
 Calibration Date: July 6, 2023 Last Cal Date: June 8, 2023  
 Start time (MST): 7:30 End time (MST): 10:03  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 49.22 ppm Cal Gas Exp Date: February 23, 2025  
 Cal Gas Cylinder #: CC306868  
 Removed Cal Gas Conc: 49.22 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1220  
 ZAG Make/Model: Teledyne API 701 Serial Number: 4522

### Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 1501301450  
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995624	0.996723	Backgd or Offset: 18.2	18.2
Calibration intercept:	0.891225	2.331293	Coeff or Slope: 0.945	0.945

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.8	----
as found span	4919	81.3	800.3	800.6	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.5	----
high point	4919	81.3	800.3	799.5	1.001
second point	4959	40.7	400.7	401.3	0.998
third point	4980	20.3	199.8	204.2	0.979
as left zero	5000	0.0	0.0	0.4	----
as left span	4919	81.3	800.3	799.9	1.000
Average Correction Factor					0.993

Baseline Corr As found: 799.80 Previous response 797.66 \*% change 0.3%  
 Baseline Corr 2nd AF pt: NA AF Slope: AF Intercept:  
 Baseline Corr 3rd AF pt: NA AF Correlation:

\* = > +/-5% change initiates investigation

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	June 8, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:30	End Time (MST):	10:03
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.5	----	Correlation Coefficient	0.999968	
800.3	799.5	1.0010			≥0.995
400.7	401.3	0.9984	Slope	0.996723	
199.8	204.2	0.9786			0.90 - 1.10
			Intercept	2.331293	+/-30



SO2 Calibration Plot

Date: July 6, 2023

Location: MacKay River





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: MacKay River      Station number: AMS20  
 Calibration Date: July 5, 2023      Last Cal Date: June 7, 2023  
 Start time (MST): 7:10      End time (MST): 11:09  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.12 ppm      Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC515997  
 Removed Cal Gas Conc: 5.12 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 1220  
 ZAG Make/Model: Teledyne API 701      Serial Number: 4522

### Analyzer Information

Analyzer make: Thermo 43iQ TLE      Analyzer serial #: 12124313139  
 Converter make: Global      Converter serial #: 2022-226  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002952	0.996518	Backgd or Offset:	1.1	1.03
Calibration intercept:	-0.033000	0.267198	Coeff or Slope:	0.605	0.571

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4922	78.1	80.0	85.4	0.936
as found 2nd point	4961	39.0	39.9	42.7	0.935
as found 3rd point	4980	19.5	20.0	21.6	0.925
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4922	78.1	80.0	79.9	1.002
second point	4961	39.0	40.0	40.3	0.992
third point	4980	19.5	20.0	20.3	0.986
as left zero	5000	0.0	0.0	0.1	----
as left span	4922	78.1	80.0	79.7	1.004
SO2 Scrubber Check	4919	80.0	800.2	-0.1	----
Date of last scrubber change:	May 25, 2023			Ave Corr Factor	0.993
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 85.4      Prev response: 80.18      \*% change: **6.1%**  
 Baseline Corr 2nd AF pt: 42.7      AF Slope: 1.066761      AF Intercept: 0.120683  
 Baseline Corr 3rd AF pt: 21.6      AF Correlation: 0.999988

\* = > +/-5% change initiates investigation

Notes: Diagnostics similar to last month. Sox Scrubber check completed after calibrator zero. Span Adjusted.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 5, 2023	Previous Calibration:	June 7, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:10	End Time (MST):	11:09
Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	12124313139

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999973	
80.0	79.9	1.0017			≥0.995
40.0	40.3	0.9917	Slope	0.996518	
20.0	20.3	0.9860			0.90 - 1.10
			Intercept	0.267198	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 5, 2023

Location: MacKay River





# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name:	MacKay River	Station number:	AMS20
Calibration Date:	July 6, 2023	Last Cal Date:	June 8, 2023
Start time (MST):	7:30	End time (MST):	10:02
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC306868	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.40</u> ppm	CH4 Equiv Conc.	1066.45 ppm
C3H8 Cal Gas Conc.	<u>206.20</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.40</u> ppm	CH4 Equiv Conc.	1066.45 ppm
Removed C3H8 Conc.	<u>206.20</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	1220
ZAG Make/Model:	Teledyne API 701	Serial Number:	4522

### Analyzer Information

Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727
Analyzer Range:	0 - 20 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002852	0.989494	Background:	3.220	3.040
Calibration intercept:	-0.024234	0.030619	Coefficient:	5.430	5.384

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.14	----
as found span	4919	81.3	17.34	17.22	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	-0.01	----
high point	4919	81.3	17.34	17.18	1.009
second point	4959	40.7	8.68	8.60	1.009
third point	4980	20.3	4.33	4.38	0.988
as left zero	5000	0.0	0.00	-0.03	----
as left span	4919	81.3	17.34	17.17	1.010
Average Correction Factor					1.002
Baseline Corr As found:	17.36	Previous response	17.36	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No maintenance done. Zero adjusted.

Calibration Performed By: Melissa Lemay





# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	June 8, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:30	End Time (MST):	10:02
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.01	----	Correlation Coefficient	0.999961	≥0.995
17.34	17.18	1.0093			
8.68	8.60	1.0090	Slope	0.989494	0.90 - 1.10
4.33	4.38	0.9885			
			Intercept	0.030619	+/-1.5



THC Calibration Plot

Date: July 6, 2023

Location: MacKay River







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
as found span	4917	83.3	819.5	800.3	19.2	817.3	797.4	19.9	1.0026	1.0036
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
high point	4917	83.3	819.5	800.3	19.2	813.3	794.4	19.0	1.0076	1.0074
second point	4956	41.7	410.4	400.8	9.6	409.5	398.9	10.7	1.0023	1.0049
third point	4979	20.8	204.6	199.9	4.8	208.2	201.5	6.7	0.9829	0.9918
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.1	----	----
as left span	4917	83.3	819.5	442.0	377.5	818.3	437.3	381.1	1.0014	1.0107
Average Correction Factor									0.9976	1.0014

Corrected As found	NO <sub>x</sub> = 817.3 ppb	NO = 797.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.3%	
Previous Response	NO <sub>x</sub> = 819.5 ppb	NO = 800.0 ppb		*Percent Change	NO = -0.3%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	793.1	434.8	377.5	377.0	1.0012	99.9%
2nd GPT point (200 ppb O <sub>3</sub> )	793.1	606.2	206.1	204.5	1.0076	99.2%
3rd GPT point (100 ppb O <sub>3</sub> )	793.1	697.4	114.9	113.9	1.0084	99.2%
Average Correction Factor					1.0057	99.4%

Notes:

No adjustments and maintenance done.

Calibration Performed By:

Melissa Lemay



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 19, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	7:14	End Time (MST):	11:07
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
819.5	813.3	1.0076		
410.4	409.5	1.0023		
204.6	208.2	0.9829		





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 19, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	7:14	End Time (MST):	11:07
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	794.4	1.0074		
400.8	398.9	1.0049		
199.9	201.5	0.9918		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 19, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:14	End Time (MST):	11:07
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999980	≥0.995
377.5	377.0	1.0012			
206.1	204.5	1.0076			
114.9	113.9	1.0084			
			Slope	0.998430	0.90 - 1.10
			Intercept	-0.419338	+/-20



NO<sub>x</sub> Calibration Plot

Date: July 4, 2023

Location: MacKay River







## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS21  
CONKLIN  
JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	July 13, 2023	Last Cal Date:	June 5, 2023
Start time (MST):	9:55	End time (MST):	12:50
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.93	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	<u>CC259455</u>			
Removed Cal Gas Conc:	49.93	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3810
ZAG Make/Model:	Teledyne API 701		Serial Number:	691

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999546	0.997605	Backgd or Offset:	27.3	27.2
Calibration intercept:	0.235763	0.415762	Coeff or Slope:	0.883	0.883

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5005	0.0	0.0	-0.2	----
as found span	4920	80.2	800.8	796.5	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5005	0.0	0.0	-0.2	----
high point	4920	80.2	800.8	799.1	1.002
second point	4960	40.1	400.4	400.0	1.001
third point	4980	20.0	200.1	200.8	0.997
as left zero	5005	0.0	0.0	-0.1	----
as left span	4920	80.2	800.8	798.0	1.004
Average Correction Factor					1.000

Baseline Corr As found:	796.70	Previous response	800.72	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No adjustments made.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 5, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:55	End Time (MST):	12:50
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999997	≥0.995
800.8	799.1	1.0022			
400.4	400.0	1.0011	Slope	0.997605	0.90 - 1.10
200.1	200.8	0.9966			
			Intercept	0.415762	+/-30



SO2 Calibration Plot

Date: July 13, 2023

Location: Conklin





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	July 27, 2023	Last Cal Date:	June 28, 2023
Start time (MST):	8:14	End time (MST):	11:50
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.00	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC501204			
Removed Cal Gas Conc:	5.03	ppm	Rem Gas Exp Date:	April 16, 2022
Removed Gas Cyl #:	CC505493		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3810
ZAG Make/Model:	API 701H		Serial Number:	691

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116
Converter make:	CD-Nova 101	Converter serial #:	NA
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001571	1.007143	Backgd or Offset:	2.4	2.4
Calibration intercept:	0.320000	0.200000	Coeff or Slope:	0.974	0.958

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.0	80.5	81.0	0.994
as found 2nd point	4960	40.0	40.2	40.7	0.989
as found 3rd point	4980	20.0	20.1	20.3	0.991
new cylinder response	4920	80.0	80.0		

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.0	80.0	80.7	0.991
second point	4960	40.0	40.0	40.6	0.985
third point	4980	20.0	20.0	20.4	0.980
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	80.0	80.0	80.2	0.998
SO2 Scrubber Check	4920	80.2	802.0		----

Date of last scrubber change:	Ave Corr Factor	0.986
Date of last converter efficiency test:	efficiency	

Baseline Corr As found:	81.0	Prev response:	80.93	*% change:	0.1%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.006532	AF Intercept:	0.060000
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999993		

\* = > +/-5% change initiates investigation

Notes: Adjusted the span only.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	June 28, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:14	End Time (MST):	11:50
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	$\geq 0.995$	
80.0	80.7	0.9913			
40.0	40.6	0.9852			
20.0	20.4	0.9804			
			Slope	1.007143	$0.90 - 1.10$
			Intercept	0.200000	$\pm 3$



TRS Calibration Plot

Date: July 27, 2023

Location: Conklin





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	July 13, 2023	Last Cal Date:	June 21, 2023
Start time (MST):	9:55	End time (MST):	12:50
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	497.9 ppm	CH <sub>4</sub> Equiv Conc.	1067.7 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	497.9 ppm	CH <sub>4</sub> Equiv Conc.	1067.7 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.26E-04	2.27E-04	NMHC SP Ratio:	4.99E-05
CH <sub>4</sub> Retention time:	12.00	12.00	NMHC Peak Area:	183139
				180824

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	16.98	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.12	1.000
second point	4960	40.1	8.56	8.54	1.002
third point	4980	20.0	4.27	4.29	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.16	0.998
Average Correction Factor					0.999
Baseline Corr AF:	16.98	Prev response	17.11	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.04	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.13	1.001
second point	4960	40.1	4.57	4.57	0.999
third point	4980	20.0	2.28	2.30	0.991
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.16	0.998
Average Correction Factor					0.997
Baseline Corr AF:	9.04	Prev response	9.14	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.94	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.99	1.000
second point	4960	40.1	3.99	3.97	1.006
third point	4980	20.0	1.99	1.99	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	8.00	0.998
Average Correction Factor					1.002
Baseline Corr AF:	7.94	Prev response	7.98	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999095	0.999162
THC Cal Offset:	0.002575	0.004574
CH <sub>4</sub> Cal Slope:	0.999624	0.999695
CH <sub>4</sub> Cal Offset:	-0.006854	-0.004854
NMHC Cal Slope:	0.998558	0.998696
NMHC Cal Offset:	0.009229	0.009428

Notes: Adjusted the span only.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 21, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:55	End Time (MST):	12:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	$\geq 0.995$			
17.13	17.12	1.0003						
8.56	8.54	1.0024				Slope	0.999162	0.90 - 1.10
4.27	4.29	0.9955						
			Intercept	0.004574	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 21, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:55	End Time (MST):	12:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999989	$\geq 0.995$
7.99	7.99	1.0000			
3.99	3.97	1.0058			
1.99	1.99	1.0003			
			Slope	0.999695	0.90 - 1.10
			Intercept	-0.004854	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 21, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:55	End Time (MST):	12:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999994	$\geq 0.995$			
9.14	9.13	1.0006						
4.57	4.57	0.9995				Slope	0.998696	0.90 - 1.10
2.28	2.30	0.9914						
			Intercept	0.009428	$\pm 0.5$			



NMHC Calibration Plot

Date: July 13, 2023

Location: Conklin





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	July 18, 2023	Last Cal Date:	July 13, 2023
Start time (MST):	9:26	End time (MST):	11:05
Reason:	Cylinder Change		

### Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
C3H8 Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
Removed C3H8 Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.27E-04	2.27E-04	NMHC SP Ratio:	5.05E-05
CH4 Retention time:	12.00	12.00	NMHC Peak Area:	180824
				180824

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.12	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.53	0.977
Average Correction Factor					
Baseline Corr AF:	17.12	Prev response	17.12	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.16	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.43	0.969
<b>Average Correction Factor</b>					
Baseline Corr AF:	9.16	Prev response	9.14	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.97	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	8.10	0.986
<b>Average Correction Factor</b>					
Baseline Corr AF:	7.97	Prev response	7.98	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999162	
THC Cal Offset:	0.004574	
CH <sub>4</sub> Cal Slope:	0.999695	
CH <sub>4</sub> Cal Offset:	-0.004854	
NMHC Cal Slope:	0.998696	
NMHC Cal Offset:	0.009428	

Notes:

Replaced N2 cylinder.


Calibration Performed By:

Denny Ray Estador

NMHC Calibration Plot

Date: July 18, 2023

Location: Conklin







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Conklin  
Calibration Date: July 28, 2023  
Start time (MST): 8:18  
Reason: Routine  
Station number: AMS21  
Last Cal Date: June 27, 2023  
End time (MST): 12:10

### Calibration Standards

NO Gas Cylinder #: T2Y1P1H  
NOX Cal Gas Conc: 51.09 ppm  
Removed Cylinder #: n/a  
Removed Gas NOX Conc: 51.09 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701H  
Cal Gas Expiry Date: December 11, 2023  
NO Cal Gas Conc: 50.39 ppm  
Removed Gas Exp Date: n/a  
Removed Gas NO Conc: 50.39 ppm  
NO gas Diff:  
Serial Number: 3810  
Serial Number: 691

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1501663731

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.112	1.115	NO bkgnd or offset:	11.1	11.2
NOX coeff or slope:	0.999	1.000	NOX bkgnd or offset:	11.3	11.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	201.3	203.7

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.998659	0.998617
NO <sub>x</sub> Cal Offset:	1.463814	2.003897
NO Cal Slope:	1.000266	0.999267
NO Cal Offset:	0.781352	0.881047
NO <sub>2</sub> Cal Slope:	0.995256	0.998137
NO <sub>2</sub> Cal Offset:	0.164082	-0.798489



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as found span	4921	79.4	811.2	800.1	11.1	811.0	799.5	11.6	1.0003	1.0008
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4921	79.4	811.2	800.1	11.1	811.1	800.0	10.9	1.0002	1.0002
second point	4960	39.7	405.7	400.1	5.6	408.2	401.1	7.2	0.9938	0.9976
third point	4980	19.8	202.3	199.6	2.8	206.0	201.2	4.8	0.9822	0.9918
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as left span	4921	79.4	811.2	374.7	436.5	810.0	377.0	433.0	1.0015	0.9940
Average Correction Factor									0.9921	0.9965

Corrected As found	NO <sub>x</sub> = 811.1 ppb	NO = 799.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.1%	
Previous Response	NO <sub>x</sub> = 811.6 ppb	NO = 801.1 ppb		*Percent Change	NO = -0.2%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI: ;	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.4	373.0	436.5	435.4	1.0026	99.7%
2nd GPT point (200 ppb O3)	798.4	586.1	223.4	221.6	1.0082	99.2%
3rd GPT point (100 ppb O3)	798.4	690.5	119.0	117.3	1.0146	98.6%
Average Correction Factor					1.0085	99.2%

Notes:

Adjusted the span only.

Calibration Performed By:

Denny Ray Estador



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 28, 2023	Previous Calibration:	June 27, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:18	End Time (MST):	12:10
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
811.2	811.1	1.0002			
405.7	408.2	0.9938			
202.3	206.0	0.9822			
			Slope	0.998617	0.90 - 1.10
			Intercept	2.003897	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 28, 2023	Previous Calibration:	June 27, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:18	End Time (MST):	12:10
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.1	800.0	1.0002		
400.1	401.1	0.9976		
199.6	201.2	0.9918		
			0.999994	
			0.999267	
			0.881047	

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 28, 2023	Previous Calibration:	June 27, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:18	End Time (MST):	12:10
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
436.5	435.4	1.0026		
223.4	221.6	1.0082		
119.0	117.3	1.0146		
			0.999983	
			0.998137	
			-0.798489	



NO<sub>x</sub> Calibration Plot

Date: July 28, 2023

Location: Conklin





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Conklin  
 Calibration Date: July 24, 2023  
 Start time (MST): 10:22  
 Reason: Routine

Station number: AMS21  
 Last Cal Date: June 7, 2023  
 End time (MST): 13:05

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700  
 ZAG Make/Model: Teledyne API 701H

Serial Number: 3810  
 Serial Number: 691

### Analyzer Information

Analyzer make: Thermo 49i  
 Analyzer Range 0 - 500 ppb

Analyzer serial #: 1501663734

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997914	1.012514	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	1.340000	1.160000	Coeff or Slope:	1.002	1.002

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-1.3	----
as found span	5000	951.2	400.0	405.0	0.988
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	5000	950.9	400.0	405.2	0.987
second point	5000	804.0	200.0	205.3	0.974
third point	5000	703.6	100.0	103.0	0.971
as left zero	5000	0.0	0.0	-0.3	----
as left span	5000	936.0	400.0	410.1	0.975
Average Correction Factor					0.977

Baseline Corr As found:	406.3	Previous response	400.5	*% change	1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

*\* = > +/-5% change initiates investigation*

Notes: No adjustments made.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 24, 2023	Previous Calibration:	June 7, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:22	End Time (MST):	13:05
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999938	≥0.995
400.0	405.2	0.9872			
200.0	205.3	0.9742	Slope	1.012514	0.90 - 1.10
100.0	103.0	0.9709			
			Intercept	1.160000	+/- 5






O<sub>3</sub> Calibration Plot

Date: July 24, 2023

Location: Conklin





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Conklin Station number: AMS 21  
 Calibration Date: July 27, 2023 Last Cal Date: June 27, 2023  
 Start time (MST): 8:33 End time (MST): 9:29

Analyzer Make: API T640X S/N: 1597  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat S/N: 388751  
 Temp/RH standard: Alicat S/N: 388751

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	12.2	12.03	12.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	713.2	715.28	713.2	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.05	5.01	5.05	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>July 27, 2023</u>	Last Cal Date: <u>June 27, 2023</u>			
	PM w/o HEPA: <u>5.4</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11	11.1	11.1	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>3.3</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>July 27, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 27, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_  
 Date RH/T Sensor Cleaned: \_\_\_\_\_

Notes: No adjustments made for both monthly and quarterly maintenance.

Calibration by: Denny Ray Estador



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS22  
JANVIER  
JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	July 17, 2023	Last Cal Date:	June 14, 2023
Start time (MST):	10:07	End time (MST):	13:02
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.11	ppm	Cal Gas Exp Date:	January 18, 2029
Cal Gas Cylinder #:	CC281519			
Removed Cal Gas Conc:	50.11	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007567	1.001279	Backgd or Offset:	21.3	21.4
Calibration intercept:	1.922834	2.123935	Coeff or Slope:	0.998	0.989

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.9	----
as found span	4920	79.8	799.8	809.7	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	1.0	----
high point	4920	79.8	799.8	802.0	0.997
second point	4960	39.9	399.9	404.2	0.989
third point	4980	20.0	200.4	203.2	0.986
as left zero	5000	0.0	0.0	0.9	----
as left span	4920	79.8	799.8	804.6	0.994
Average Correction Factor					0.991

Baseline Corr As found:	808.80	Previous response	807.76	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 14, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:07	End Time (MST):	13:02
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.0	----	Correlation Coefficient	0.999985	
799.8	802.0	0.9972			≥0.995
399.9	404.2	0.9893	Slope	1.001279	
200.4	203.2	0.9864			0.90 - 1.10
			Intercept	2.123935	+/-30



SO2 Calibration Plot

Date: July 17, 2023

Location: Janvier





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Janvier Station number: AMS22  
 Calibration Date: July 25, 2023 Last Cal Date: June 20, 2023  
 Start time (MST): 9:04 End time (MST): 13:28  
 Reason: Maintenance

### Calibration Standards

Cal Gas Concentration: 5.03 ppm Cal Gas Exp Date: April 16, 2022  
 Cal Gas Cylinder #: DT0018680  
 Removed Cal Gas Conc: 5.03 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806  
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

### Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 1200326169  
 Converter make: CDN-101 Converter serial #: 587  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004939	1.003363	Backgd or Offset:	1.05	0.96
Calibration intercept:	0.140790	0.020959	Coeff or Slope:	0.954	0.880

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4920	79.5	80.0	78.3	1.024
as found 2nd point	4960	39.8	40.0	40.0	1.006
as found 3rd point	4980	19.9	20.0	20.3	0.996
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	79.5	80.0	80.3	0.996
second point	4960	39.8	40.0	40.2	0.996
third point	4980	19.9	20.0	20.0	1.001
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	79.5	80.0	79.9	1.001
SO2 Scrubber Check	4920	79.8	798.0	0.0	----

Date of last scrubber change:	Ave Corr Factor	0.998
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 78.1 Prev response: 80.52 \*% change: -3.1%  
 Baseline Corr 2nd AF pt: 39.8 AF Slope: 0.975079 AF Intercept: 0.561232  
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999880

\* = > +/-5% change initiates investigation

Notes: Changed the SO2 scrubbers after multipoint as founds. Adjusted the span only. Ran a SO2 scrubber check after the third point.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 25, 2023	Previous Calibration:	June 20, 2023
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	9:04	End Time (MST):	13:28
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326169

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999995	
80.0	80.3	0.9961			≥0.995
40.0	40.2	0.9960	Slope	1.003363	
20.0	20.0	1.0010			0.90 - 1.10
			Intercept	0.020959	+/-3






TRS Calibration Plot

Date: July 25, 2023

Location: Janvier





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	July 17, 2023	Last Cal Date:	June 14, 2023
Start time (MST):	10:07	End time (MST):	13:02
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH <sub>4</sub> Cal Gas Conc.	502.8 ppm	CH <sub>4</sub> Equiv Conc.	1075.9 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH <sub>4</sub> Conc.	502.8 ppm	CH <sub>4</sub> Equiv Conc.	1075.9 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.150E-04	2.180E-04	NMHC SP Ratio:	4.55E-05
CH <sub>4</sub> Retention time:	13.4	13.4	NMHC Peak Area:	201202
				210352

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	17.64	0.973
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.27	0.994
second point	4960	39.9	8.59	8.63	0.995
third point	4980	20.0	4.30	4.31	0.999
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.23	0.997

Average Correction Factor				0.996
Baseline Corr AF:	17.64	Prev response	17.20	*% change 2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4920	79.8	9.15	9.58	0.955
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4920	79.8	9.15	9.13	1.002
second point	4960	39.9	4.57	4.56	1.003
third point	4980	20.0	2.29	2.27	1.011
as left zero	5000	0	0.00	0.00	----
as left span	4920	79.8	9.15	9.11	1.004
Average Correction Factor					1.005
Baseline Corr AF:	9.58	Prev response	9.13	*% change	4.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	8.06	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.14	0.986
second point	4960	39.9	4.01	4.07	0.986
third point	4980	20.0	2.01	2.04	0.985
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.11	0.989
Average Correction Factor					0.986
Baseline Corr AF:	8.06	Prev response	8.08	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001387	1.006073
THC Cal Offset:	0.007200	-0.009020
CH <sub>4</sub> Cal Slope:	1.004460	1.014274
CH <sub>4</sub> Cal Offset:	0.011429	0.000208
NMHC Cal Slope:	0.998704	0.998891
NMHC Cal Offset:	-0.004029	-0.009028

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 14, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:07	End Time (MST):	13:02
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	$\geq 0.995$			
17.17	17.27	0.9943						
8.59	8.63	0.9949				Slope	1.006073	0.90 - 1.10
4.30	4.31	0.9990						
			Intercept	-0.009020	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 14, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:07	End Time (MST):	13:02
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	$\geq 0.995$
8.03	8.14	0.9859			
4.01	4.07	0.9861			
2.01	2.04	0.9854			
			Slope	1.014274	0.90 - 1.10
			Intercept	0.000208	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 14, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:07	End Time (MST):	13:02
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999994	$\geq 0.995$			
9.15	9.13	1.0019						
4.57	4.56	1.0025				Slope	0.998891	0.90 - 1.10
2.29	2.27	1.0112						
			Intercept	-0.009028	$\pm 0.5$			



NMHC Calibration Plot

Date: July 17, 2023

Location: Janvier





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Janvier Station number: AMS 22  
Calibration Date: July 19, 2023 Last Cal Date: N/A  
Start time (MST): 9:57 End time (MST): 14:05  
Reason: Install

### Calibration Standards

NO Gas Cylinder #: CC424183 Cal Gas Expiry Date: April 16, 2023  
NOX Cal Gas Conc: 48.60 ppm NO Cal Gas Conc: 48.60 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 48.60 ppm Removed Gas NO Conc: 48.60 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 3806  
ZAG make/model: Teledyne API T701 Serial Number: 4890

### Analyzer Information

Analyzer make: API T200 Analyzer serial #: 833  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	N/A	0.808	NO bkgnd or offset:	N/A	-5.6
NOX coeff or slope:	N/A	0.799	NOX bkgnd or offset:	N/A	-3.9
NO2 coeff or slope:	N/A	1.000	Reaction cell Press:	N/A	5.0

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	N/A	0.997788
NO <sub>x</sub> Cal Offset:	N/A	2.208515
NO Cal Slope:	N/A	0.999902
NO Cal Offset:	N/A	1.168443
NO <sub>2</sub> Cal Slope:	N/A	0.994081
NO <sub>2</sub> Cal Offset:	N/A	-0.360935





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	1.0	0.7	0.3	----	----
high point	4918	82.3	799.9	799.9	0.0	799.6	800.6	-0.9	1.0004	0.9991
second point	4959	41.2	400.4	400.4	0.0	402.8	402.3	0.5	0.9942	0.9954
third point	4980	20.6	200.2	200.2	0.0	202.9	201.5	1.4	0.9867	0.9936
as left zero	5000	0.0	0.0	0.0	0.0	2.1	2.1	0.1	----	----
as left span	4918	82.3	799.9	372.0	427.9	793.5	369.6	423.9	1.0081	1.0065
Average Correction Factor									0.9938	0.9960

Corrected As found	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = NA
Previous Response	NO <sub>x</sub> = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
					As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.6	370.7	427.9	425.6	1.0054	99.5%
2nd GPT point (200 ppb O3)	798.6	594.0	204.6	201.8	1.0139	98.6%
3rd GPT point (100 ppb O3)	798.6	695.1	103.5	102.5	1.0098	99.0%
Average Correction Factor					1.0097	99.0%

Notes:

Replaced the previous instrument with this one due to stability issues. Adjusted zero and span.

Calibration Performed By:

Max Farrell



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 19, 2023	Previous Calibration:	N/A
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:57	End Time (MST):	14:05
Analyzer make:	API T200	Analyzer serial #:	833

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	1.0	----	Correlation Coefficient	≥0.995	
799.9	799.6	1.0004			
400.4	402.8	0.9942			
200.2	202.9	0.9867			
			Slope	0.997788	0.90 - 1.10
			Intercept	2.208515	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 19, 2023	Previous Calibration:	N/A
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:57	End Time (MST):	14:05
Analyzer make:	API T200	Analyzer serial #:	833

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient	0.999997	≥0.995
799.9	800.6	0.9991			
400.4	402.3	0.9954	Slope	0.999902	0.90 - 1.10
200.2	201.5	0.9936			
			Intercept	1.168443	+/-20

NO Calibration Curve





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 19, 2023	Previous Calibration:	N/A
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:57	End Time (MST):	14:05
Analyzer make:	API T200	Analyzer serial #:	833

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
427.9	425.6	1.0054		
204.6	201.8	1.0139		
103.5	102.5	1.0098		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: July 19, 2023

Location: Janvier





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Janvier Station number: AMS 22  
 Calibration Date: July 26, 2023 Last Cal Date: June 22, 2023  
 Start time (MST): 10:58 End time (MST): 14:30  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806  
 ZAG Make/Model: Teledyne API T701H Serial Number: 201

### Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 7046  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000371	0.999714	Backgd or Offset:	-0.2	-0.1
Calibration intercept:	1.760000	2.000000	Coeff or Slope:	1.018	0.953

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.6	----
as found span	4895	904.3	400.0	429.1	0.932
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	-0.1	----
high point	4895	904.3	400.0	400.6	0.999
second point	4895	756.7	200.0	203.7	0.982
third point	4895	656.1	100.0	103.6	0.965
as left zero	5000	800.0	0.0	0.5	----
as left span	4895	904.3	400.0	406.5	0.984
Average Correction Factor					0.982

Baseline Corr As found:	428.5	Previous response	401.9	*% change	6.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. There was a power outage during as founds and the as found span is 7% high, diagnostics are all good and the instrument was installed last month. See docit note for more info. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 26, 2023	Previous Calibration:	June 22, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:58	End Time (MST):	14:30
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	≥0.995
400.0	400.6	0.9985		
200.0	203.7	0.9818	Slope	0.90 - 1.10
100.0	103.6	0.9653		
			Intercept	+/- 5



O<sub>3</sub> Calibration Plot

Date: July 26, 2023

Location: Janvier







# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Janvier Station number: AMS 22  
 Calibration Date: July 26, 2023 Last Cal Date: June 13, 2023  
 Start time (MST): 9:48 End time (MST): 10:20

Analyzer Make: Teledyne API T640 S/N: 325  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450  
 Temp/RH standard: Delta Cal S/N: 1450

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	17.8	18.01	17.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	712.8	712.6	712.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.03	4.94	5.03	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>July 26, 2023</u>	Last Cal Date: <u>June 13, 2023</u>			
	PM w/o HEPA: <u>9.4</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11	11	11	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>4.9</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>July 26, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 26, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: July 26, 2023  
 Date RH/T Sensor Cleaned: July 26, 2023

### Notes:

Completed annual, monthly and quarterly calibrations. No adjustments needed. Leak check passed.

Calibration by: Max Farrell



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS23 FORT HILLS JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	July 10, 2023	Last Cal Date:	June 6, 2023
Start time (MST):	10:03	End time (MST):	13:00
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.76	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC281425			
Removed Cal Gas Conc:	49.76	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	451
ZAG Make/Model:	API T701		Serial Number:	5611

### Analyzer Information

Analyzer make: Thermo 43i	Analyzer serial #: 1160290012
Analyzer Range 0 - 1000 ppb	

	<u><i>Start</i></u>	<u><i>Finish</i></u>		<u><i>Start</i></u>	<u><i>Finish</i></u>
Calibration slope:	1.004713	1.002525	Backgd or Offset:	17.8	17.8
Calibration intercept:	-0.844401	-0.604149	Coeff or Slope:	1.040	1.040

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4920	80.3	799.1	798.2	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.3	799.1	800.7	0.998
second point	4960	40.2	400.1	400.5	0.999
third point	4980	20.1	200.0	199.1	1.005
as left zero	5000	0.0	0.0	0.0	----
as left span	4920	80.3	799.1	799.8	0.999
<b>Average Correction Factor</b>					<b>1.001</b>

Baseline Corr As found:	798.40	Previous response	802.02	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

*\* = > +/-5% change initiates investigation*

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	June 6, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	10:03	End Time (MST):	13:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	0.999997	
799.1	800.7	0.9980			≥0.995
400.1	400.5	0.9989	Slope	1.002525	
200.0	199.1	1.0047			0.90 - 1.10
			Intercept	-0.604149	+/-30



SO2 Calibration Plot

Date: July 10, 2023

Location: Fort Hills





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	July 24, 2023	Last Cal Date:	June 29, 2023
Start time (MST):	10:13	End time (MST):	14:20
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.20	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517372			
Removed Cal Gas Conc:	5.20	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	451
ZAG Make/Model:	API T701		Serial Number:	5611

### Analyzer Information

Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	1300156232
Converter make:	CDN-101	Converter serial #:	594
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006026	1.006312	Backgd or Offset:	1.93	1.19
Calibration intercept:	0.002053	0.042041	Coeff or Slope:	1.132	1.124

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4923	77.0	80.0	78.0	1.024
as found 2nd point	4962	38.5	40.0	39.2	1.018
as found 3rd point	4981	19.2	19.9	19.5	1.018
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4923	77.0	80.0	80.4	0.995
second point	4962	38.5	40.0	40.6	0.985
third point	4981	19.2	19.9	20.1	0.992
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.0	80.0	81.8	0.978
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.991
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	78.1	Prev response:	80.49	*% change:	-3.1%
Baseline Corr 2nd AF pt:	39.3	AF Slope:	0.976033	AF Intercept:	0.001396
Baseline Corr 3rd AF pt:	19.6	AF Correlation:	0.999987		

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 24, 2023	Previous Calibration:	June 29, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	10:13	End Time (MST):	14:20
Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	1300156232

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999961	≥0.995
80.0	80.4	0.9951			
40.0	40.6	0.9852	Slope	1.006312	0.90 - 1.10
19.9	20.1	0.9924			
			Intercept	0.042041	+/-3



TRS Calibration Plot

Date: July 24, 2023

Location: Fort Hills







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	July 10, 2023	Last Cal Date:	June 6, 2023
Start time (MST):	10:02	End time (MST):	13:00
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1070.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH <sub>4</sub> Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1070.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.32E-04	2.32E-04	NMHC SP Ratio:	5.06E-05
CH <sub>4</sub> Retention time:	13.0	13.0	NMHC Peak Area:	181940

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	17.05	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.08	1.006
second point	4960	40.2	8.61	8.50	1.012
third point	4980	20.1	4.30	4.25	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.19	17.08	1.007

				Average Correction Factor	1.010
Baseline Corr AF:	17.05	Prev response	17.10	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.07	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.08	1.009
second point	4960	40.2	4.59	4.56	1.006
third point	4980	20.1	2.29	2.30	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	9.16	9.07	1.009
Average Correction Factor					1.004
Baseline Corr AF:	9.07	Prev response	9.11	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	7.98	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.00	1.004
second point	4960	40.2	4.02	3.94	1.020
third point	4980	20.1	2.01	1.95	1.031
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.03	8.00	1.004
Average Correction Factor					1.018
Baseline Corr AF:	7.98	Prev response	7.99	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.995947	0.993826
THC Cal Offset:	-0.024349	-0.020141
CH <sub>4</sub> Cal Slope:	0.998403	0.997507
CH <sub>4</sub> Cal Offset:	-0.034620	-0.033221
NMHC Cal Slope:	0.993893	0.990511
NMHC Cal Offset:	0.009871	0.013680

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	June 6, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	10:02	End Time (MST):	13:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999989	$\geq 0.995$
17.19	17.08	1.0064			
8.61	8.50	1.0123			
4.30	4.25	1.0124			
			Slope	0.993826	0.90 - 1.10
			Intercept	-0.020141	+/-0.5





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	June 6, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	10:02	End Time (MST):	13:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999908	≥0.995			
8.03	8.00	1.0038						
4.02	3.94	1.0196				Slope	0.997507	0.90 - 1.10
2.01	1.95	1.0306						
			Intercept	-0.033221	+/-0.5			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	June 6, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	10:02	End Time (MST):	13:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999989	$\geq 0.995$			
9.16	9.08	1.0087						
4.59	4.56	1.0056				Slope	0.990511	0.90 - 1.10
2.29	2.30	0.9969						
			Intercept	0.013680	$\pm 0.5$			



NMHC Calibration Plot

Date: July 10, 2023

Location: Fort Hills





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	July 21, 2023	Last Cal Date:	July 10, 2023
Start time (MST):	10:30	End time (MST):	12:30
Reason:	Cylinder Change		

### Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1070.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH <sub>4</sub> Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1070.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T750	Serial Number:	282
ZAG make/model:	API T751H	Serial Number:	321

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.32E-04	2.32E-04	NMHC SP Ratio:	5.06E-05
CH <sub>4</sub> Retention time:	13.0	13.0	NMHC Peak Area:	181940

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	17.33	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.19	17.27	0.995

Average Correction Factor					
Baseline Corr AF:	17.33	Prev response	17.07	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.25	0.990
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	9.16	9.23	0.993
Average Correction Factor					
Baseline Corr AF:	9.25	Prev response	9.09	*% change	1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	8.08	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.03	8.05	0.998
Average Correction Factor					
Baseline Corr AF:	8.08	Prev response	7.98	*% change	1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.993826	
THC Cal Offset:	-0.020141	
CH <sub>4</sub> Cal Slope:	0.997507	
CH <sub>4</sub> Cal Offset:	-0.033221	
NMHC Cal Slope:	0.990511	
NMHC Cal Offset:	0.013680	

Notes: Changed the N2 cylinder after as founds.


Calibration Performed By: Rene Chamberland



NMHC Calibration Plot

Date: July 21, 2023

Location: Fort Hills





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Fort Hills  
Calibration Date: July 21, 2023  
Start time (MST): 10:14  
Reason: Routine  
Station number: AMS23  
Last Cal Date: June 1, 2023  
End time (MST): 15:26

### Calibration Standards

NO Gas Cylinder #: CC332703  
NOX Cal Gas Conc: 49.7 ppm  
Removed Cylinder #: N/A  
Removed Gas NOX Conc: 49.7 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701  
Cal Gas Expiry Date: January 28, 2024  
NO Cal Gas Conc: 49.7 ppm  
Removed Gas Exp Date: N/A  
Removed Gas NO Conc: 49.7 ppm  
NO gas Diff:  
Serial Number: 451  
Serial Number: 5611

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1152430007

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.072	1.094	NO bkgnd or offset:	3	3.4
NOX coeff or slope:	0.995	0.995	NOX bkgnd or offset:	3.2	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	166.3	165.6

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.998879	0.999493
NO <sub>x</sub> Cal Offset:	1.384352	0.844323
NO Cal Slope:	0.999937	0.999823
NO Cal Offset:	-0.035916	0.004005
NO <sub>2</sub> Cal Slope:	1.003749	1.006741
NO <sub>2</sub> Cal Offset:	0.418698	0.163045



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.9	0.4	0.5	----	----
as found span	4920	80.5	800.2	800.2	0.0	782.7	777.2	5.4	1.022	1.030
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.1	----	----
high point	4920	80.5	800.2	800.2	0.0	799.8	799.6	0.2	1.000	1.001
second point	4960	40.2	399.6	399.6	0.0	401.9	400.7	1.1	0.994	0.997
third point	4980	20.1	199.8	199.8	0.0	200.4	199.1	1.3	0.997	1.003
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	-0.1	----	----
as left span	4920	80.5	800.2	447.3	352.9	797.3	442.9	354.4	1.004	1.010
Average Correction Factor									0.997	1.000

Corrected As found	NO <sub>x</sub> = 781.8 ppb	NO = 776.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -2.4%	
Previous Response	NO <sub>x</sub> = 800.7 ppb	NO = 800.1 ppb		*Percent Change	NO = -3.0%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.7	443.8	352.9	355.4	0.993	100.7%
2nd GPT point (200 ppb O3)	796.7	622.8	173.9	175.3	0.992	100.8%
3rd GPT point (100 ppb O3)	796.7	708.2	88.5	89.3	0.991	100.9%
Average Correction Factor					0.992	100.8%

Notes: Changed the inlet filter after as founds. Adjusted both zero and span. Used the 2nd GPT reference points due to drift.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 21, 2023	Previous Calibration:	June 1, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	10:14	End Time (MST):	15:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.2	799.8	1.0005			
399.6	401.9	0.9942			
199.8	200.4	0.9970			
			Slope	0.999493	0.90 - 1.10
			Intercept	0.844323	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 21, 2023	Previous Calibration:	June 1, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	10:14	End Time (MST):	15:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.2	799.6	1.0007		
399.6	400.7	0.9972		
199.8	199.1	1.0035		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 21, 2023	Previous Calibration:	June 1, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	10:14	End Time (MST):	15:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	1.000000	≥0.995
352.9	355.4	0.9930			
173.9	175.3	0.9920			
88.5	89.3	0.9910			
			Slope	1.006741	0.90 - 1.10
			Intercept	0.163045	+/-20



NO<sub>x</sub> Calibration Plot

Date: July 21, 2023

Location: Fort Hills





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Fort Hills Station number: AMS 23  
 Calibration Date: July 24, 2023 Last Cal Date: June 6, 2023  
 Start time (MST): 13:54 End time (MST): 14:38

Analyzer Make: API T640 S/N: 1546  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388755  
 Temp/RH standard: Alicat FP-25BT S/N: 388755

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	30	30.79	30	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	732.8	732.08	732.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	4.99	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>July 24, 2023</u>	Last Cal Date: <u>June 6, 2023</u>			
	PM w/o HEPA: <u>29.5</u>	PM w/ HEPA: <u>0.1</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.8	11.3	11	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>30.3</u>	w/ HEPA: <u>0.1</u>		
Date Optical Chamber Cleaned:		<u>July 24, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 24, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: May 10, 2023  
 Date RH/T Sensor Cleaned: May 10, 2023

Notes: Leak check passed, adjusted the PMT. Instrument is very dirty due to a lot of site activity and smoke in the area, completed the quarterly maintenance and the monthly checks.

Calibration by: Max Farrell





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS25 WASKŌW OHCI PIMÂTISIWIN JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number:	AMS25
Calibration Date:	July 24, 2023	Last Cal Date:	June 15, 2023
Start time (MST):	6:48	End time (MST):	10:00
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.54	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC437219			
Removed Cal Gas Conc:	50.54	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	747
ZAG Make/Model:	API T701		Serial Number:	4765

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999805	1.005116	Backgd or Offset:	9.7	9.7
Calibration intercept:	0.523945	-0.536058	Coeff or Slope:	0.974	0.988

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	79.2	800.5	789.5	1.014
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	79.2	800.5	804.5	0.995
second point	4960	39.6	400.3	401.3	0.998
third point	4980	19.8	200.1	200.0	1.001
as left zero	5000	0.0	0.0	0.3	----
as left span	4921	79.2	800.5	802.3	0.998
Average Correction Factor					0.998

Baseline Corr As found:	789.50	Previous response	800.89	*% change	-1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Span adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 24, 2023	Previous Calibration:	June 15, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	6:48	End Time (MST):	10:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	0.999996	
800.5	804.5	0.9951			≥0.995
400.3	401.3	0.9975	Slope	1.005116	
200.1	200.0	1.0007			0.90 - 1.10
			Intercept	-0.536058	+/-30



SO2 Calibration Plot

Date: July 24, 2023

Location: Waskow ohci Pimatisiwin





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number:	AMS25
Calibration Date:	July 18, 2023	Last Cal Date:	June 30, 2023
Start time (MST):	6:42	End time (MST):	11:37
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	4.97	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC517099			
Removed Cal Gas Conc:	4.97	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	747
ZAG Make/Model:	API T701		Serial Number:	261

### Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146
Converter make:	Global G-150	Converter serial #:	2022-219
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996347	0.997066	Backgd or Offset:	3.25
Calibration intercept:	0.040000	-0.060000	Coeff or Slope:	1.079

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4920	80.0	79.5	79.1	1.003
as found 2nd point	4960	40.0	39.7	39.6	1.001
as found 3rd point	4980	20.0	19.9	19.9	0.993
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	80.0	79.5	79.1	1.005
second point	4960	40.0	39.7	39.7	1.001
third point	4980	20.0	19.9	19.7	1.008
as left zero	5000	0.0	0.0	0.0	----
as left span	4920	80.0	800.0	800.0	1.000
SO2 Scrubber Check	4921	79.2	800.0	0.1	----
Date of last scrubber change:	20-Jun-23			Ave Corr Factor	1.005
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	79.2	Prev response:	79.21	*% change:	0.0%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.996059	AF Intercept:	0.000000
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999992		

\* = > +/-5% change initiates investigation

Notes: Calibration after the Sulphur Study. SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 30, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	6:42	End Time (MST):	11:37
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999991	≥0.995
79.5	79.1	1.0045			
39.7	39.7	1.0007	Slope	0.997066	0.90 - 1.10
19.9	19.7	1.0083			
			Intercept	-0.060000	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 18, 2023

Location: Waskow ohci Pimatisiwin





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS26 CHRISTINA LAKE JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023







# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 1, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	8:34	End Time (MST):	11:20
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999943	≥0.995
799.0	810.2	0.9861			
399.4	402.3	0.9929	Slope	1.015687	0.90 - 1.10
200.2	197.6	1.0132			
			Intercept	-2.538773	+/-30



SO2 Calibration Plot

Date: July 13, 2023

Location: Christina Lake





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Christina Lake	Station number:	AMS26
Calibration Date:	July 14, 2023	Last Cal Date:	June 14, 2023
Start time (MST):	6:35	End time (MST):	10:35
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	4.89	ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	<u>EY0002466</u>			
Removed Cal Gas Conc:	4.89	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2447
ZAG Make/Model:	API T701H		Serial Number:	832

### Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032
Converter make:	NA	Converter serial #:	NA
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010753	1.001325	Backgd or Offset:	35.0      35.1
Calibration intercept:	-0.140775	-0.160901	Coeff or Slope:	1.093      1.093

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4918	81.8	80.0	82.2	0.977
as found 2nd point	4959	40.9	40.0	41.4	0.973
as found 3rd point	4979	20.4	20.0	20.4	0.993
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4918	81.8	80.0	79.9	1.001
second point	4959	40.9	40.0	40.0	1.000
third point	4979	20.4	20.0	19.8	1.008
as left zero	5000	0.0	0.0	0.0	----
as left span	4918	81.8	80.0	80.7	0.991
SO2 Scrubber Check	4919	80.6	806.1	0.0	----
Date of last scrubber change:	27-Feb-19			Ave Corr Factor	1.003
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	81.9	Prev response:	80.72	*% change:	1.4%
Baseline Corr 2nd AF pt:	41.1	AF Slope:	1.025322	AF Intercept:	0.199445
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999969		

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after MAF's. Ran scrubber check after calibrator zero. Adjusted zero only.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 14, 2023	Previous Calibration:	June 14, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	6:35	End Time (MST):	10:35
Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999996	<b>≥0.995</b>
80.0	79.9	1.0013			
40.0	40.0	1.0000	Slope	1.001325	<b>0.90 - 1.10</b>
20.0	19.8	1.0078			
			Intercept	-0.160901	<b>+/-3</b>



H<sub>2</sub>S Calibration Plot

Date: July 14, 2023

Location: Christina Lake





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Christina Lake  
Calibration Date: July 12, 2023  
Start time (MST): 11:37  
Reason: As Found  
Station number: AMS 26  
Last Cal Date: June 15, 2023  
End time (MST): 15:20

### Calibration Standards

NO Gas Cylinder #: T2Y1P4C  
NOX Cal Gas Conc: 50.82 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 50.82 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701H  
Cal Gas Expiry Date: November 12, 2023  
NO Cal Gas Conc: 50.02 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 50.02 ppm  
NO gas Diff:  
Serial Number: 2447  
Serial Number: 832

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1173480006

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.887	1.887	NO bkgnd or offset:	3.6	3.6
NOX coeff or slope:	0.995	0.995	NOX bkgnd or offset:	4.2	4.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	212.8	225.5

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.003570	
NO <sub>x</sub> Cal Offset:	-1.960000	
NO Cal Slope:	1.003127	
NO Cal Offset:	-3.060000	
NO <sub>2</sub> Cal Slope:	1.001715	
NO <sub>2</sub> Cal Offset:	-0.246575	



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.1	-0.5	----	----
as found span	4920	80.0	813.1	800.3	12.8	762.0	748.1	13.9	1.0671	1.0698
as found 2nd	4960	40.0	406.6	400.2	6.4	381.4	373.4	8.0	1.0660	1.0717
as found 3rd	4980	20.0	203.3	200.1	3.2	188.6	183.5	5.2	1.0778	1.0904
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

### Average Correction Factor

Corrected As found	NO <sub>x</sub> = 762.6 ppb	NO = 748.2 ppb	<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -6.7%
Previous Response	NO <sub>x</sub> = 814.1 ppb	NO = 799.8 ppb			*Percent Change	NO = -6.9%
Baseline Corr 2nd pt	NO <sub>x</sub> = 382.0 ppb	NO = 373.5 ppb	As found	NO <sub>x</sub> r <sup>2</sup> : 0.999993	Nx SI: 0.938579	Nx Int: -1.040
Baseline Corr 3rd pt	NO <sub>x</sub> = 189.2 ppb	NO = 183.6 ppb	As found	NO r <sup>2</sup> : 0.999976	NO SI: 0.936268	NO Int: -1.600
			As found	NO <sub>2</sub> r <sup>2</sup> : 0.999994	NO <sub>2</sub> SI: 1.002544	NO <sub>2</sub> Int: -0.182

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	-0.5	----	----
as found GPT point (400 ppb NO <sub>2</sub> )	750.7	364.8	398.7	399.5	0.9980	100.2%
as found GPT point (200 ppb NO <sub>2</sub> )	750.7	553.4	210.1	210.2	0.9995	100.0%
as found GPT point (100 ppb NO <sub>2</sub> )	750.7	654.0	109.5	110.2	0.9936	100.6%
1st GPT point (400 ppb O <sub>3</sub> )						
2nd GPT point (200 ppb O <sub>3</sub> )						
3rd GPT point (100 ppb O <sub>3</sub> )						

### Average Correction Factor

Notes: As found span was low, investigated issue and found pump was the issue. Performed MPAF's for pump change out.


Calibration Performed By: Mohammed Kashif



NO<sub>x</sub> Calibration Plot

Date: July 12, 2023

Location: Christina Lake





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Christina Lake  
Calibration Date: July 13, 2023  
Start time (MST): 11:20  
Reason: Maintenance  
Station number: AMS26  
Last Cal Date: June 15, 2023  
End time (MST): 15:39

### Calibration Standards

NO Gas Cylinder #: T2Y1P4C  
NOX Cal Gas Conc: 50.82 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 50.82 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701H  
Cal Gas Expiry Date: November 12, 2023  
NO Cal Gas Conc: 50.02 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 50.02 ppm  
NO gas Diff:  
Serial Number: 2447  
Serial Number: 832

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1173480006

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.887	1.396	NO bkgnd or offset:	3.6	2.6
NOX coeff or slope:	0.995	0.991	NOX bkgnd or offset:	4.2	3.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	225.5	164.7

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:		0.999944
NO <sub>x</sub> Cal Offset:		-1.520000
NO Cal Slope:		1.003527
NO Cal Offset:		-2.600000
NO <sub>2</sub> Cal Slope:		0.993893
NO <sub>2</sub> Cal Offset:		-0.796010



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
high point	4920	80.0	813.1	800.3	12.8	812.2	801.6	10.6	1.0011	0.9984
second point	4960	40.0	406.6	400.2	6.4	404.6	398.3	6.4	1.0048	1.0047
third point	4980	20.0	203.3	200.1	3.2	200.0	195.3	4.6	1.0164	1.0245
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	0.0	-0.4	----	----
as left span	4920	80.0	813.1	391.8	421.3	814.6	397.0	417.5	0.9982	0.9870
Average Correction Factor									1.0075	1.0092

Corrected As found	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = NA
Previous Response	NO <sub>x</sub> = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
					As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	801.9	393.4	421.3	418.7	1.0062	99.4%
2nd GPT point (200 ppb O3)	801.9	593.6	221.1	217.7	1.0156	98.5%
3rd GPT point (100 ppb O3)	801.9	701.2	113.5	111.6	1.0170	98.3%
Average Correction Factor					1.0130	98.7%

Notes: Installed new pump and changed sample inlet filter. Adjusted both zero and span. No other maintenance needed.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 15, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	11:20	End Time (MST):	15:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.1	812.2	1.0011		
406.6	404.6	1.0048		
203.3	200.0	1.0164		





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 15, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	11:20	End Time (MST):	15:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	801.6	0.9984		
400.2	398.3	1.0047		
200.1	195.3	1.0245		
			0.999954	
			1.003527	
			-2.600000	





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 15, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	11:20	End Time (MST):	15:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
421.3	418.7	1.0062		
221.1	217.7	1.0156		
113.5	111.6	1.0170		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: July 13, 2023

Location: Christina Lake





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS27**  
**JACKFISH 2/3**  
**JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Jackfish 2/3      Station number: AMS 27  
 Calibration Date: July 7, 2023      Last Cal Date: June 1, 2023  
 Start time (MST): 9:26      End time (MST): 12:06  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 50.58 ppm      Cal Gas Exp Date: December 29, 2028  
 Cal Gas Cylinder #: SG9133974BAL  
 Removed Cal Gas Conc: 50.58 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3811  
 ZAG Make/Model: API 701      Serial Number: 135

### Analyzer Information

Analyzer make: Thero 43iQ      Analyzer serial #: 12124313138  
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006390	1.008391	Backgd or Offset:	7.5	7.5
Calibration intercept:	-1.738080	-1.738506	Coeff or Slope:	0.942	0.942

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	79.1	800.2	802.8	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4921	79.1	800.2	806.0	0.993
second point	4961	39.5	399.5	400.2	0.998
third point	4980	19.8	200.3	198.7	1.008
as left zero	5000	0.0	0.0	-0.1	----
as left span	4921	79.1	800.2	807.0	0.992
Average Correction Factor					1.000

Baseline Corr As found:	802.90	Previous response	803.53	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No adjustments made.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 7, 2023	Previous Calibration:	June 1, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	9:26	End Time (MST):	12:06
Analyzer make:	Thero 43iQ	Analyzer serial #:	12124313138

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥0.995
800.2	806.0	0.9928		
399.5	400.2	0.9984	Slope	0.90 - 1.10
200.3	198.7	1.0081		
			Intercept	+/-30



SO2 Calibration Plot

Date: July 7, 2023

Location: Jackfish 2/3





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Jackfish 2/3      Station number: AMS27  
 Calibration Date: July 19, 2023      Last Cal Date: June 13, 2023  
 Start time (MST): 7:49      End time (MST): 11:20  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.41 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC345023  
 Removed Cal Gas Conc: 5.41 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3811  
 ZAG Make/Model: API 701      Serial Number: 135

### Analyzer Information

Analyzer make: API T101      Analyzer serial #: 621  
 Converter make:      Converter serial #:  
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003478	1.011028	Backgd or Offset: 27.3	27.3
Calibration intercept:	-0.077831	0.082512	Coeff or Slope: 0.961	0.961

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.1	80.2	82.0	0.978
as found 2nd point	4963	37.0	40.0	41.0	0.976
as found 3rd point	4982	18.5	20.0	20.2	0.991
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4926	74.1	80.2	81.0	0.990
second point	4963	37.0	40.0	40.9	0.979
third point	4982	18.5	20.0	20.2	0.991
as left zero	5000	0.0	0.0	0.4	----
as left span	4926	74.1	80.2	80.2	1.000
SO2 Scrubber Check	4921	79.1	791.0	-0.2	----
Date of last scrubber change:				Ave Corr Factor	0.986
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 82.0      Prev response: 80.38      \*% change: 2.0%  
 Baseline Corr 2nd AF pt: 41.0      AF Slope: 1.024003      AF Intercept: -0.097351  
 Baseline Corr 3rd AF pt: 20.2      AF Correlation: 0.999984

\* = > +/-5% change initiates investigation

Notes: No adjustments made.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 19, 2023	Previous Calibration:	June 13, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	7:49	End Time (MST):	11:20
Analyzer make:	API T101	Analyzer serial #:	621

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	0.999956
80.2	81.0	0.9898		
40.0	40.9	0.9788	Slope	1.011028
20.0	20.2	0.9908		
			Intercept	0.082512
				+/-3



H<sub>2</sub>S Calibration Plot

Date: July 19, 2023

Location: Jackfish 2/3





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Jackfish 2/3  
Calibration Date: July 20, 2023  
Start time (MST): 7:52  
Reason: Routine  
Station number: AMS27  
Last Cal Date: June 15, 2023  
End time (MST): 14:08

### Calibration Standards

NO Gas Cylinder #: T2Y1P35  
NOX Cal Gas Conc: 51.44 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 51.44 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701  
Cal Gas Expiry Date: December 11, 2023  
NO Cal Gas Conc: 50.40 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 50.40 ppm  
NO gas Diff:  
Serial Number: 3811  
Serial Number: 135

### Analyzer Information

Analyzer make: API T200  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 722

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.099	1.082	NO bkgnd or offset:	-0.9	0.3
NOX coeff or slope:	1.082	1.081	NOX bkgnd or offset:	-0.2	1.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.9	5.5

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000905	1.000962
NO <sub>x</sub> Cal Offset:	-1.315764	-1.136298
NO Cal Slope:	1.008010	1.001486
NO Cal Offset:	-1.998764	-2.340129
NO <sub>2</sub> Cal Slope:	0.996976	1.003914
NO <sub>2</sub> Cal Offset:	0.633893	0.816669



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.2	0.6	0.6	----	----
as found span	4921	79.4	816.8	800.3	16.5	815.2	804.1	11.1	1.0020	0.9953
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	-0.1	----	----
high point	4921	79.4	816.8	800.3	16.5	816.4	800.0	16.5	1.0005	1.0004
second point	4960	39.7	408.5	400.2	8.3	408.8	398.0	10.8	0.9992	1.0055
third point	4980	19.8	203.7	199.6	4.1	200.8	195.0	5.8	1.0145	1.0235
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.7	-0.7	----	----
as left span	4921	79.4	816.8	421.8	385.4	812.5	417.4	395.0	1.0053	1.0105
Average Correction Factor									1.0047	1.0098

Corrected As found	NO <sub>x</sub> = 814.0 ppb	NO = 803.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.3%	
Previous Response	NO <sub>x</sub> = 816.2 ppb	NO = 804.7 ppb		*Percent Change	NO = -0.1%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.8	427.9	385.4	387.1	0.9956	100.4%
2nd GPT point (200 ppb O3)	796.8	618.2	195.1	197.6	0.9874	101.3%
3rd GPT point (100 ppb O3)	796.8	706.2	107.1	109.0	0.9827	101.8%
Average Correction Factor					0.9886	101.2%

Notes: Adjusted zero only at first but third point of GPT was failing hence, went back to adjust the span and performed the calibration process again at around 11:15 MST.

Calibration Performed By: Denny Ray Estador





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 20, 2023	Previous Calibration:	June 15, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	7:52	End Time (MST):	14:08
Analyzer make:	API T200	Analyzer serial #:	722

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
816.8	816.4	1.0005		
408.5	408.8	0.9992		
203.7	200.8	1.0145		
			0.999984	
			1.000962	
			-1.136298	





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 20, 2023	Previous Calibration:	June 15, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	7:52	End Time (MST):	14:08
Analyzer make:	API T200	Analyzer serial #:	722

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ <i>0.90 - 1.10</i> <i>+/-20</i>
800.3	800.0	1.0004		
400.2	398.0	1.0055		
199.6	195.0	1.0235		
			0.999966	
			1.001486	
			-2.340129	

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 20, 2023	Previous Calibration:	June 15, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	7:52	End Time (MST):	14:08
Analyzer make:	API T200	Analyzer serial #:	722

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
385.4	387.1	0.9956		
195.1	197.6	0.9874		
107.1	109.0	0.9827		



NO<sub>x</sub> Calibration Plot

Date: July 20, 2023

Location: Jackfish 2/3





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS29  
SURMONT 2  
JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 6, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:56	End Time (MST):	13:05
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999984	≥0.995
800.1	804.0	0.9952			
400.6	400.7	0.9997	Slope	1.005842	0.90 - 1.10
199.8	197.9	1.0097			
			Intercept	-1.525687	+/-30



SO2 Calibration Plot

Date: July 17, 2023

Location: Surmont 2









# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 10, 2023	Previous Calibration:	June 7, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:50	End Time (MST):	14:52
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999998	≥0.995
80.0	80.0	1.0000			
40.1	40.1	1.0002	Slope	1.001618	0.90 - 1.10
20.1	19.9	1.0077			
			Intercept	-0.122896	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 10, 2023

Location: Surmont 2





# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	July 17, 2023	Last Cal Date:	June 2, 2023
Start time (MST):	9:56	End time (MST):	13:05
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC356008	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
C3H8 Cal Gas Conc.	<u>205.7</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
Removed C3H8 Conc.	<u>205.7</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5472
ZAG Make/Model:	Teledyne API T701	Serial Number:	4297

### Analyzer Information

Analyzer make: Thermo 51i-LT                      Analyzer serial #: 1170050149  
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003409	1.000217	Background:	3.72	3.64
Calibration intercept:	-0.030830	-0.075664	Coefficient:	4.074	3.987

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.02	----
as found span	4918	81.3	17.31	17.71	0.978
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	-0.02	----
high point	4918	81.3	17.31	17.27	1.003
second point	4959	40.6	8.65	8.54	1.012
third point	4979	20.3	4.32	4.20	1.029
as left zero	5000	0.0	0.00	-0.06	----
as left span	4918	81.3	17.31	17.35	0.998
Average Correction Factor					1.015
Baseline Corr As found:	17.70	Previous response	17.34	*% change	2.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 17, 2023	Previous Calibration:	June 2, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:56	End Time (MST):	13:05
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<i>Limits</i>	
0.00	-0.02	----	Correlation Coefficient	0.999959	
17.31	17.27	1.0026			≥0.995
8.65	8.54	1.0124	Slope	1.000217	
4.32	4.20	1.0291			0.90 - 1.10
			Intercept	-0.075664	+/-1.5



THC Calibration Plot

Date: July 17, 2023

Location: Surmont 2







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.2	----	----
as found span	4916	84.2	799.2	799.2	0.0	799.8	798.2	1.5	0.9992	1.0012
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.3	----	----
high point	4916	84.2	799.2	799.2	0.0	801.0	800.0	0.3	0.9977	0.9990
second point	4958	42.1	399.6	399.6	0.0	401.1	400.2	0.9	0.9963	0.9985
third point	4979	21.1	200.3	200.3	0.0	200.5	197.6	2.8	0.9989	1.0135
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
as left span	4916	84.2	799.2	394.9	404.3	802.0	396.2	405.7	0.9965	0.9967
Average Correction Factor									0.9976	1.0037

Corrected As found	NO <sub>x</sub> = 799.5 ppb	NO = 798.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.8%	
Previous Response	NO <sub>x</sub> = 806.2 ppb	NO = 804.9 ppb		*Percent Change	NO = -0.9%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.6	395.3	404.3	404.7	0.9990	100.1%
2nd GPT point (200 ppb O3)	799.6	603.5	196.1	196.1	1.0000	100.0%
3rd GPT point (100 ppb O3)	799.6	701.5	98.1	99.4	0.9869	101.3%
Average Correction Factor					0.9953	100.5%

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 21, 2023	Previous Calibration:	June 21, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:40	End Time (MST):	14:20
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.2	801.0	0.9977		
399.6	401.1	0.9963		
200.3	200.5	0.9989		





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 21, 2023	Previous Calibration:	June 21, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:40	End Time (MST):	14:20
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.2	800.0	0.9990		
399.6	400.2	0.9985		
200.3	197.6	1.0135		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 21, 2023	Previous Calibration:	June 21, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:40	End Time (MST):	14:20
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	≥0.995	
404.3	404.7	0.9990			
196.1	196.1	1.0000			
98.1	99.4	0.9869			
			Slope	0.999330	0.90 - 1.10
			Intercept	0.616986	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: July 21, 2023

Location: Surmont 2







## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS30  
ELLS RIVER  
JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 2, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:26	End Time (MST):	12:20
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.2	----	Correlation Coefficient	0.999941	≥0.995
800.4	807.1	0.9917			
400.2	398.9	1.0033	Slope	1.010041	0.90 - 1.10
200.1	196.9	1.0163			
			Intercept	-3.016022	+/-30






SO2 Calibration Plot

Date: July 4, 2023

Location: Ells River





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Ells River Station number: AMS30  
 Calibration Date: July 5, 2023 Last Cal Date: June 12, 2023  
 Start time (MST): 9:23 End time (MST): 13:15  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.08 ppm Cal Gas Exp Date: February 9, 2024  
 Cal Gas Cylinder #: EY0002443  
 Removed Cal Gas Conc: 5.08 ppm Rem Gas Exp Date:  
 Removed Gas Cyl #: Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 3061  
 ZAG Make/Model: API T701H Serial Number: 358

### Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1410661331  
 Converter make: CDN - 101 Converter serial #: 562  
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003207	1.008782	Backgd or Offset:	1.59 1.61
Calibration intercept:	-0.139148	-0.159281	Coeff or Slope:	1.136 1.136

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	78.7	80.0	79.9	1.001
as found 2nd point	4961	39.4	40.0	39.9	1.003
as found 3rd point	4980	19.7	20.0	19.9	1.006
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	78.7	80.0	80.6	0.992
second point	4961	39.4	40.0	40.1	0.998
third point	4980	19.7	20.0	19.9	1.006
as left zero	5000	0.0	0.0	0.1	----
as left span	4921	78.7	80.0	80.2	0.997
SO2 Scrubber Check	4921	79.2	800.4	0.1	----
Date of last scrubber change:	N/A			Ave Corr Factor	0.999
Date of last converter efficiency test:	N/A			efficiency	

Baseline Corr As found: 79.9 Prev response: 80.08 \*% change: -0.2%  
 Baseline Corr 2nd AF pt: 39.9 AF Slope: 0.999493 AF Intercept: -0.059157  
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999997

\* = > +/-5% change initiates investigation

Notes: No adjustments made.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 5, 2023	Previous Calibration:	June 12, 2023
Station Name:	Ells River	Station Number:	AMS30
Start Time (MST):	9:23	End Time (MST):	13:15
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999981	
80.0	80.6	0.9921			≥0.995
40.0	40.1	0.9982	Slope	1.008782	
20.0	19.9	1.0058			0.90 - 1.10
			Intercept	-0.159281	+/-3



TRS Calibration Plot

Date: July 5, 2023

Location: Ells River





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	July 4, 2023	Last Cal Date:	June 2, 2023
Start time (MST):	9:26	End time (MST):	12:20
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH <sub>4</sub> Cal Gas Conc.	499.7 ppm	CH <sub>4</sub> Equiv Conc.	1075.0 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	499.7 ppm	CH <sub>4</sub> Equiv Conc.	1075.0 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	209.2 ppm		
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701H	Serial Number:	3061
		Serial Number:	358

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	0.000241	0.000241	NMHC SP Ratio:	4.51E-05
CH <sub>4</sub> Retention time:	14.2	14.2	NMHC Peak Area:	202169
				202169

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.97	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.01	1.001
second point	4960	39.6	8.51	8.41	1.012
third point	4980	19.8	4.26	4.11	1.035
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.11	0.995

				Average Correction Factor	1.016
Baseline Corr AF:	16.97	Prev response	16.99	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	9.18	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.22	0.989
second point	4960	39.6	4.56	4.55	1.003
third point	4980	19.8	2.28	2.21	1.029
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.26	0.984
Average Correction Factor					1.007
Baseline Corr AF:	9.18	Prev response	9.09	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.80	1.015
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.80	1.015
second point	4960	39.6	3.96	3.87	1.023
third point	4980	19.8	1.98	1.90	1.042
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.85	1.009
Average Correction Factor					1.027
Baseline Corr AF:	7.80	Prev response	7.91	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002469	1.001214
THC Cal Offset:	-0.074938	-0.074338
CH <sub>4</sub> Cal Slope:	1.003253	0.986547
CH <sub>4</sub> Cal Offset:	-0.032957	-0.025355
NMHC Cal Slope:	1.001801	1.013966
NMHC Cal Offset:	-0.041781	-0.048783

Notes: No adjustments made.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 2, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:26	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999913	$\geq 0.995$			
17.03	17.01	1.0010						
8.51	8.41	1.0121				Slope	1.001214	0.90 - 1.10
4.26	4.11	1.0348						
			Intercept	-0.074338	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 2, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:26	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999952	$\geq 0.995$
7.91	7.80	1.0154			
3.96	3.87	1.0230			
1.98	1.90	1.0415			
			Slope	0.986547	0.90 - 1.10
			Intercept	-0.025355	+/-0.5







# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 4, 2023	Previous Calibration:	June 2, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:26	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999872	$\geq 0.995$			
9.11	9.22	0.9888						
4.56	4.55	1.0026				Slope	1.013966	0.90 - 1.10
2.28	2.21	1.0290						
			Intercept	-0.048783	$\pm 0.5$			



NMHC Calibration Plot

Date: July 4, 2023

Location: Ells River





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	July 21, 2023	Last Cal Date:	July 4, 2023
Start time (MST):	8:44	End time (MST):	14:50
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH <sub>4</sub> Cal Gas Conc.	499.7 ppm	CH <sub>4</sub> Equiv Conc.	1075.0 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	499.7 ppm	CH <sub>4</sub> Equiv Conc.	1075.0 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	209.2 ppm		
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701H	Serial Number:	3061
		Serial Number:	358

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	0.000241	0.000252	NMHC SP Ratio:	4.51E-05
CH <sub>4</sub> Retention time:	14.2	14.4	NMHC Peak Area:	202169
				195284

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	15.13	1.125
as found 2nd point	4960	39.6	8.51	7.53	1.130
as found 3rd point	4980	19.8	4.26	3.70	1.151
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.03	1.000
second point	4960	39.6	8.51	8.43	1.010
third point	4980	19.8	4.26	4.14	1.027
as left zero	5000	0.0	0.00	4.14	----
as left span	4921	79.2	17.03	17.05	0.999
Average Correction Factor					1.012
Baseline Corr AF:	15.13	Prev response	16.97	*% change	-12.2%
Baseline Corr 2nd AF:	7.5	AF Slope:	0.890002	AF Intercept:	-0.040671
Baseline Corr 3rd AF:	3.7	AF Correlation:	0.999964	* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	7.96	<b>1.145</b>
as found 2nd point	4960	39.6	4.56	3.96	<b>1.153</b>
as found 3rd point	4980	19.8	2.28	1.94	<b>1.175</b>
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.11	1.000
second point	4960	39.6	4.56	4.51	1.010
third point	4980	19.8	2.28	2.22	1.028
as left zero	5000	0	0.00	2.22	----
as left span	4921	79.2	9.11	9.13	0.998
Average Correction Factor					1.013
Baseline Corr AF:	7.96	Prev response	9.19	*% change	<b>-15.5%</b>
Baseline Corr 2nd AF:	4.0	AF Slope:	0.874897	AF Intercept:	-0.025924
Baseline Corr 3rd AF:	1.9	AF Correlation:	0.999951	* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.17	<b>1.104</b>
as found 2nd point	4960	39.6	3.96	3.57	<b>1.108</b>
as found 3rd point	4980	19.8	1.98	1.76	<b>1.125</b>
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.91	1.000
second point	4960	39.6	3.96	3.92	1.010
third point	4980	19.8	1.98	1.93	1.026
as left zero	5000	0.0	0.00	1.93	----
as left span	4921	79.2	7.91	7.92	1.000
Average Correction Factor					1.012
Baseline Corr AF:	7.17	Prev response	7.78	*% change	<b>-8.6%</b>
Baseline Corr 2nd AF:	3.57	AF Slope:	0.906873	AF Intercept:	-0.015916
Baseline Corr 3rd AF:	1.76	AF Correlation:	0.999976	* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001214	1.001691
THC Cal Offset:	-0.074338	-0.062138
CH <sub>4</sub> Cal Slope:	0.986547	1.001405
CH <sub>4</sub> Cal Offset:	-0.025355	-0.027557
NMHC Cal Slope:	1.013966	1.001951
NMHC Cal Offset:	-0.048783	-0.034381

Notes: Replaced pump after MPAFs. Calibration started at 13:00 MST.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 21, 2023	Previous Calibration:	July 4, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:44	End Time (MST):	14:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999939	$\geq 0.995$
17.03	17.03	1.0000			
8.51	8.43	1.0100			
4.26	4.14	1.0273			
			Slope	1.001691	0.90 - 1.10
			Intercept	-0.062138	+/-0.5





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 21, 2023	Previous Calibration:	July 4, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:44	End Time (MST):	14:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999945	≥0.995
7.91	7.91	1.0002			
3.96	3.92	1.0097			
1.98	1.93	1.0264			
			Slope	1.001405	0.90 - 1.10
			Intercept	-0.027557	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 21, 2023	Previous Calibration:	July 4, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:44	End Time (MST):	14:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999935	$\geq 0.995$			
9.11	9.11	0.9998						
4.56	4.51	1.0101				Slope	1.001951	0.90 - 1.10
2.28	2.22	1.0281						
			Intercept	-0.034381	+/-0.5			



NMHC Calibration Plot

Date: July 21, 2023

Location: Ells River







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Ells River  
Calibration Date: July 6, 2023  
Start time (MST): 9:16  
Reason: Routine  
Station number: AMS 30  
Last Cal Date: June 6, 2023  
End time (MST): 13:55

### Calibration Standards

NO Gas Cylinder #: T2Y1P2R  
NOX Cal Gas Conc: 50.83 ppm  
Removed Cylinder #: T2Y1P2R  
Removed Gas NOX Conc: 50.83 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701H  
Cal Gas Expiry Date: December 11, 2023  
NO Cal Gas Conc: 49.97 ppm  
Removed Gas Exp Date:  
Removed Gas NO Conc: 49.97 ppm  
NO gas Diff:  
Serial Number: 3061  
Serial Number: 358

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 710321429

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.051	1.031	NO bkgnd or offset:	12.7	12.8
NOX coeff or slope:	0.990	0.994	NOX bkgnd or offset:	12.9	12.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	180.6	184.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.996431	1.001026
NO <sub>x</sub> Cal Offset:	-1.740000	-1.400000
NO Cal Slope:	0.993668	0.999657
NO Cal Offset:	-2.100000	-2.520000
NO <sub>2</sub> Cal Slope:	1.000832	1.005854
NO <sub>2</sub> Cal Offset:	0.789762	1.195147



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
as found span	4920	80.0	813.3	799.5	13.8	830.0	814.5	15.7	0.9799	0.9816
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4920	80.0	813.3	799.5	13.8	813.5	798.1	15.3	0.9997	1.0018
second point	4960	40.0	406.6	399.8	6.9	404.7	395.5	9.2	1.0048	1.0108
third point	4980	20.0	203.3	199.9	3.4	200.9	195.0	5.9	1.0120	1.0250
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as left span	4920	80.0	813.3	421.3	392.0	817.3	422.8	394.5	0.9951	0.9965
Average Correction Factor									1.0055	1.0125

Corrected As found	NO <sub>x</sub> = 829.9 ppb	NO = 814.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 2.6%
Previous Response	NO <sub>x</sub> = 808.6 ppb	NO = 792.4 ppb		*Percent Change	NO = 2.7%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.9	416.7	392.0	394.7	0.9931	100.7%
2nd GPT point (200 ppb O3)	794.9	616.0	192.7	196.0	0.9830	101.7%
3rd GPT point (100 ppb O3)	794.9	703.2	105.5	108.2	0.9747	102.6%
Average Correction Factor					0.9836	101.7%

Notes:

Both zero and span was adjusted.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	June 6, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:16	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.3	813.5	0.9997		
406.6	404.7	1.0048		
203.3	200.9	1.0120		





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	June 6, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:16	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.5	798.1	1.0018		
399.8	395.5	1.0108		
199.9	195.0	1.0250		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	June 6, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:16	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
392.0	394.7	0.9931		
192.7	196.0	0.9830		
105.5	108.2	0.9747		


NO<sub>2</sub> Calibration Curve



NO<sub>x</sub> Calibration Plot

Date: July 6, 2023

Location: Ells River





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Ells River Station number: AMS 30  
 Calibration Date: July 26, 2023 Last Cal Date: June 29, 2023  
 Start time (MST): 9:13 End time (MST): 9:30

Analyzer Make: API T640 S/N: 875  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat S/N: 388751  
 Temp/RH standard: Alicat S/N: 388751

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	15	14.46	15	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	726.5	728.43	726.43	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	4.98	4.99	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: July 26, 2023 Last Cal Date: July 21, 2023  
 PM w/o HEPA: 4.2 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: \_\_\_\_\_ w/ HEPA: \_\_\_\_\_  
 Date Optical Chamber Cleaned: \_\_\_\_\_ <0.2 ug/m3  
 Disposable Filter Changed: \_\_\_\_\_

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_  
 Date RH/T Sensor Cleaned: \_\_\_\_\_

Notes: No adjustments made. Inlet head: Clean!

Calibration by: Denny Ray Estador



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS505**  
**SAWBONES BAY**  
**JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023







# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 14, 2023	Previous Calibration:	June 21, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:21	End Time (MST):	11:17
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321323

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999995	≥0.995
799.8	800.4	0.9993			
399.9	400.1	0.9995	Slope	1.001253	0.90 - 1.10
200.4	199.1	1.0067			
			Intercept	-0.552518	+/-30



SO2 Calibration Plot

Date: July 14, 2023

Location: Sawbones Bay







# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	June 21, 2023	Previous Calibration:	NA
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:49	End Time (MST):	11:00
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321323

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999993	
799.8	800.6	0.9990			≥0.995
399.9	399.8	1.0003	Slope	1.001395	
200.4	198.8	1.0082			0.90 - 1.10
			Intercept	-0.652437	+/-30



SO2 Calibration Plot

Date: June 21, 2023

Location: Sawbones Bay





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Sawbones Bay	Station number:	AMS505
Calibration Date:	June 14, 2023	Last Cal Date:	NA
Start time (MST):	10:19	End time (MST):	13:42
Reason:	Install		

### Calibration Standards

Cal Gas Concentration:	5.15	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517397			
Removed Cal Gas Conc:	5.15	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledybe API T700		Serial Number:	5112
ZAG Make/Model:	Teledybe API T701		Serial Number:	690

### Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	1228021057
Converter make:	Global 150	Converter serial #:	2022-224
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.000362	Backgd or Offset:	NA	2.26
Calibration intercept:	NA	-0.258108	Coeff or Slope:	NA	1.037

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4922	77.7	80.0	79.8	1.003
second point	4961	38.8	40.0	39.8	1.004
third point	4981	19.4	20.0	19.6	1.019
as left zero	5000	0.0	0.0	-0.1	----
as left span	4922	77.7	80.0	80.0	1.000
SO2 Scrubber Check	4922	77.8	778.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.009
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

\* = > +/-5% change initiates investigation

Notes: Changed inlet after calibrator zero. Scrubber check and passed after calibrator zero. Adjusted span only.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	June 14, 2023	Previous Calibration:	NA
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	10:19	End Time (MST):	13:42
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1228021057

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999993	
80.0	79.8	1.0030			≥0.995
40.0	39.8	1.0042	Slope	1.000362	
20.0	19.6	1.0194			0.90 - 1.10
			Intercept	-0.258108	+/-3






H<sub>2</sub>S Calibration Plot

Date: June 14, 2023

Location: Sawbones Bay





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Sawbones Bay	Station number:	AMS505
Calibration Date:	July 12, 2023	Last Cal Date:	June 14, 2023
Start time (MST):	8:30	End time (MST):	12:34
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.15	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517397			
Removed Cal Gas Conc:	5.15	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledybe API T700		Serial Number:	5112
ZAG Make/Model:	Teledybe API T701		Serial Number:	690

### Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	1228021057
Converter make:	Global 150	Converter serial #:	2022-224
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000362	0.994365	Backgd or Offset:	2.26
Calibration intercept:	-0.258108	-0.098257	Coeff or Slope:	1.037

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4922	77.7	80.0	79.1	1.011
as found 2nd point	4961	38.8	40.0	39.7	1.004
as found 3rd point	4981	19.4	20.0	19.6	1.014
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4922	77.7	80.0	79.5	1.007
second point	4961	38.8	40.0	39.6	1.009
third point	4981	19.4	20.0	19.8	1.009
as left zero	5000	0.0	0.0	0.1	----
as left span	4922	77.7	80.0	79.3	1.009
SO2 Scrubber Check	4922	77.8	778.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.008
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	79.2	Prev response:	79.81	*% change:	-0.8%
Baseline Corr 2nd AF pt:	39.8	AF Slope:	0.990222	AF Intercept:	-0.078258
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999983		

\* = > +/-5% change initiates investigation

Notes: Changed inlet after calibrator zero. Scrubber check and passed after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 12, 2023	Previous Calibration:	June 14, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:30	End Time (MST):	12:34
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1228021057

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999999	
80.0	79.5	1.0067			≥0.995
40.0	39.6	1.0092	Slope	0.994365	
20.0	19.8	1.0091			0.90 - 1.10
			Intercept	-0.098257	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 12, 2023

Location: Sawbones Bay







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
high point	4917	83.4	799.6	799.6	0.0	801.0	802.0	-1.4	0.9982	0.9970
second point	4958	41.7	399.8	399.8	0.0	395.9	393.6	2.3	1.0100	1.0159
third point	4979	20.9	200.4	200.4	0.0	193.8	191.6	2.2	1.0340	1.0459
as left zero	5000	0.0	0.0	0.0	0.0	-0.6	0.1	-0.7	----	----
as left span	4916	83.4	799.7	315.6	484.1	795.6	309.0	486.5	1.0052	1.0215
Average Correction Factor									1.0141	1.0196

Corrected As found	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = NA
Previous Response	NO <sub>x</sub> = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
					As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	792.3	308.2	484.1	487.4	0.9932	100.7%
2nd GPT point (200 ppb O3)	792.3	537.9	254.4	261.4	0.9732	102.8%
3rd GPT point (100 ppb O3)	792.3	628.8	163.5	168.8	0.9686	103.2%
Average Correction Factor					0.9784	102.2%

Notes: Adjusted zero and span. The first span response was not settling properly. Decided to go back to zero and purge the calibrator. Used 2nd NO reference point due to drift.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	NA
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	9:19	End Time (MST):	16:28
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	0.999906	≥0.995
799.6	801.0	0.9982			
399.8	395.9	1.0100	Slope	1.004292	0.90 - 1.10
200.4	193.8	1.0340			
			Intercept	-3.829947	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	NA
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	9:19	End Time (MST):	16:28
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999805	≥0.995
799.6	802.0	0.9970			
399.8	393.6	1.0159	Slope	1.006091	0.90 - 1.10
200.4	191.6	1.0459			
			Intercept	-5.309700	+/-20

**NO Calibration Curve**







# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	June 15, 2023	Previous Calibration:	NA
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	9:19	End Time (MST):	16:28
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
484.1	487.4	0.9932			
254.4	261.4	0.9732			
163.5	168.8	0.9686			
			Slope	1.006124	0.90 - 1.10
			Intercept	2.493980	+/-20



NO<sub>x</sub> Calibration Plot

Date: June 15, 2023

Location: Sawbones Bay







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	3.4	-0.2	3.6	----	----
as found span	4917	83.4	799.6	799.6	0.0	810.0	813.8	-3.8	0.9871	0.9825
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
high point	4917	83.4	799.6	799.6	0.0	801.4	805.2	-3.9	0.9977	0.9930
second point	4958	41.7	399.8	399.8	0.0	401.7	403.1	-1.3	0.9954	0.9919
third point	4979	20.9	200.4	200.4	0.0	198.2	198.5	-0.3	1.0111	1.0095
as left zero	5000	0.0	0.0	0.0	0.0	-0.6	0.1	-0.7	----	----
as left span	4916	83.4	799.7	325.3	474.4	801.6	329.0	472.6	0.9977	0.9889
Average Correction Factor									1.0014	0.9982

Corrected As found	NO <sub>x</sub> = 806.6 ppb	NO = 814.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.9%	
Previous Response	NO <sub>x</sub> = 799.2 ppb	NO = 799.1 ppb		*Percent Change	NO = 1.8%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	801.1	326.7	474.4	473.9	1.0011	99.9%
2nd GPT point (200 ppb O <sub>3</sub> )	801.1	555.5	245.6	247.8	0.9911	100.9%
3rd GPT point (100 ppb O <sub>3</sub> )	801.1	656.8	144.3	144.2	1.0007	99.9%
Average Correction Factor					0.9976	100.2%

Notes:

Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By:

Sean Bala



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 15, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:17	End Time (MST):	12:29
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.6	801.4	0.9977		
399.8	401.7	0.9954		
200.4	198.2	1.0111		
			0.999981	
			1.003551	
			-0.870760	

**NO<sub>x</sub> Calibration Curve**





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 15, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:17	End Time (MST):	12:29
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.6	805.2	0.9930		
399.8	403.1	0.9919		
200.4	198.5	1.0095		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 13, 2023	Previous Calibration:	June 15, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:17	End Time (MST):	12:29
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
474.4	473.9	1.0011			
245.6	247.8	0.9911			
144.3	144.2	1.0007			
			Slope	0.999342	0.90 - 1.10
			Intercept	0.567256	+/-20



NO<sub>x</sub> Calibration Plot

Date: July 13, 2023

Location: Sawbones Bay







# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Sawbones Bay	Station Number:	AMS 505
Calibration Date:	June 16, 2023	Prev Cal Date:	NA
Start Time (MST):	12:17	End Time (MST):	13:45
Tower Height (m):		Reason:	Install

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	CA 03845
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.5%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	NA	0.999999	<i>≥0.9995</i>
Calculated slope	NA	0.999594	<i>0.90 - 1.10</i>
Calculated intercept	NA	-0.028293	<i>+/- 2</i>

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	B4693
As Found Declination (deg east of True North):	<u>NA</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:20	Calc Declination*:	12.96 Degrees
Deadband calc:	0.6 degrees ( <i>Limit 4 deg</i> )		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.2	---
90	89.1	-0.3%
180	179.2	-0.2%
270	270.1	0.0%
357	356.6	-0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	NA	0.999994	<i>≥0.9995</i>
Calculated slope	NA	1.000239	<i>0.90 - 1.10</i>
Calculated intercept	NA	0.330283	<i>+/- 4</i>

Notes: Install calibration. WS/WD sensor lines replaced.

Calibration Performed By: Rene Chamberland



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS506  
JACKFISH 1  
JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 18, 2023	Previous Calibration:	June 28, 2023
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	8:44	End Time (MST):	11:50
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.5	----	Correlation Coefficient	0.999988	≥0.995
800.2	802.2	0.9975			
400.2	401.1	0.9976	Slope	1.004172	0.90 - 1.10
200.1	197.8	1.0115			
			Intercept	-1.416078	+/-30



SO2 Calibration Plot

Date: July 18, 2023

Location: Jackfish 1






SO2 Calibration Plot

Date: August 10, 2023

Location: Jackfish 1





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Jackfish 1	Station number:	AMS506
Calibration Date:	July 20, 2023	Last Cal Date:	June 22, 2023
Start time (MST):	9:50	End time (MST):	13:40
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.14	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511843			
Removed Cal Gas Conc:	5.14	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2659
ZAG Make/Model:	API 701		Serial Number:	4427

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020
Converter make:	Global G150	Converter serial #:	2022-218
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.015290	1.014004	Backgd or Offset:	3.35
Calibration intercept:	-0.438102	-0.418127	Coeff or Slope:	1.066

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic-AFzero))
as found zero	5000	0.0	0.0	-0.3	----
as found span	4922	77.8	80.0	78.2	1.019
as found 2nd point	4961	38.9	40.0	39.4	1.007
as found 3rd point	4981	19.4	19.9	19.4	1.012
new cylinder response					

*Limit = 0.90-1.10*

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)
					<i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4922	77.8	80.0	80.7	0.991
second point	4961	38.9	40.0	40.2	0.995
third point	4981	19.4	19.9	19.6	1.017
as left zero	5000	0.0	0.0	-0.3	----
as left span	4922	77.8	80.0	80.4	0.995
SO2 Scrubber Check	4921	79.2	792.0	0.0	----

Date of last scrubber change:	24-Feb-23	Ave Corr Factor	1.001
Date of last converter efficiency test:	December 1, 2022	efficiency	

Baseline Corr As found:	78.5	Prev response:	80.77	*% change:	-2.9%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.981572	AF Intercept:	-0.158727
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999960		

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after multi-point as founds. Scrubber test done and passed after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 20, 2023	Previous Calibration:	June 22, 2023
Station Name:	Jackfish 1	Station Number:	AMS506
Start Time (MST):	9:50	End Time (MST):	13:40
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999983	<b>≥0.995</b>
80.0	80.7	0.9911			
40.0	40.2	0.9948			
19.9	19.6	1.0174			
			Slope	1.014004	<b>0.90 - 1.10</b>
			Intercept	-0.418127	<b>+/-3</b>



H<sub>2</sub>S Calibration Plot

Date: July 20, 2023

Location: Jackfish 1





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Jackfish 1	Station number:	AMS506
Calibration Date:	August 10, 2023	Last Cal Date:	July 20, 2023
Start time (MST):	10:24	End time (MST):	11:56
Reason:	Removal		

### Calibration Standards

Cal Gas Concentration:	5.14	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511843			
Removed Cal Gas Conc:	5.14	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T750		Serial Number:	282
ZAG Make/Model:	API 751H		Serial Number:	321

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020
Converter make:	Global G150	Converter serial #:	2022-218
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.014004		Backgd or Offset:	3.35
Calibration intercept:	-0.418127		Coeff or Slope:	1.066

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	4922	77.8	80.0	80.4	0.990
as found 2nd point	4961	38.9	40.0	40.3	0.983
as found 3rd point	4981	19.4	19.9	20.1	0.973
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

#### SO2 Scrubber Check

Date of last scrubber change:	24-Feb-23	Ave Corr Factor	
Date of last converter efficiency test:	December 1, 2022	efficiency	

Baseline Corr As found:	80.8	Prev response:	80.68	*% change:	0.1%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.009141	AF Intercept:	-0.198029
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999971		

\* = > +/-5% change initiates investigation

Notes:


Removal Calibration.

Calibration Performed By: Sean Bala

H<sub>2</sub>S Calibration Plot

Date: August 10, 2023

Location: Jackfish 1





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name:	Jackfish 1	Station number:	AMS506
Calibration Date:	July 27, 2023	Last Cal Date:	June 27, 2023
Start time (MST):	8:30	End time (MST):	13:01
Reason:	Routine		

### Calibration Standards

NO Gas Cylinder #:	T26811M	Cal Gas Expiry Date:	October 30, 2024
NOX Cal Gas Conc:	<u>47.46</u> ppm	NO Cal Gas Conc:	<u>47.39</u> ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	<u>47.46</u> ppm	Removed Gas NO Conc:	<u>47.39</u> ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	2659
ZAG make/model:	API 701	Serial Number:	4427

### Analyzer Information

Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.150	1.162	NO bkgnd or offset:	3.8	3.8
NOX coeff or slope:	0.989	0.990	NOX bkgnd or offset:	4.0	4.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	173.1	173.1

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.001347	0.999777
NO <sub>x</sub> Cal Offset:	-1.568006	-1.467995
NO Cal Slope:	1.002797	1.000782
NO Cal Offset:	-2.507983	-2.727959
NO <sub>2</sub> Cal Slope:	1.006674	1.000178
NO <sub>2</sub> Cal Offset:	-0.458123	-0.291235



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.2	-0.2	----	----
as found span	4916	84.4	801.1	799.9	1.2	797.4	794.0	3.3	1.0046	1.0074
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.2	----	----
high point	4916	84.4	801.1	799.9	1.2	799.6	798.8	0.8	1.0018	1.0014
second point	4958	42.2	400.5	400.0	0.6	399.6	397.1	2.4	1.0024	1.0072
third point	4979	21.1	200.3	200.0	0.3	196.8	194.2	2.6	1.0177	1.0298
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
as left span	4916	84.4	801.1	422.8	378.3	795.2	416.4	379.5	1.0074	1.0153
Average Correction Factor									1.0073	1.0128

Corrected As found	NO <sub>x</sub> = 797.8 ppb	NO = 794.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.3%
Previous Response	NO <sub>x</sub> = 800.6 ppb	NO = 799.6 ppb		*Percent Change	NO = -0.7%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	793.2	416.1	378.3	378.1	1.0005	100.0%
2nd GPT point (200 ppb O3)	793.2	599.5	194.9	194.6	1.0014	99.9%
3rd GPT point (100 ppb O3)	793.2	691.8	102.6	102.2	1.0037	99.6%
Average Correction Factor					1.0019	99.8%

Notes: Changed inlet filter after as founds. Adjusted and span only.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	June 27, 2023
Station Name:	Jackfish 1	Station Number:	AMS506
Start Time (MST):	8:30	End Time (MST):	13:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
801.1	799.6	1.0018			
400.5	399.6	1.0024			
200.3	196.8	1.0177			
			Slope	0.999777	0.90 - 1.10
			Intercept	-1.467995	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	June 27, 2023
Station Name:	Jackfish 1	Station Number:	AMS506
Start Time (MST):	8:30	End Time (MST):	13:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
799.9	798.8	1.0014			
400.0	397.1	1.0072			
200.0	194.2	1.0298			
			Slope	1.000782	0.90 - 1.10
			Intercept	-2.727959	+/-20

**NO Calibration Curve**







# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 27, 2023	Previous Calibration:	June 27, 2023
Station Name:	Jackfish 1	Station Number:	AMS506
Start Time (MST):	8:30	End Time (MST):	13:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
378.3	378.1	1.0005			
194.9	194.6	1.0014			
102.6	102.2	1.0037			
			Slope	1.000178	0.90 - 1.10
			Intercept	-0.291235	+/-20



NO<sub>x</sub> Calibration Plot

Date: July 27, 2023

Location: Jackfish 1







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
as found span	4916	84.4	801.1	799.9	1.2	796.0	791.9	4.0	1.0064	1.0101
as found 2nd	4958	42.2	400.5	400.0	0.6	397.2	394.0	3.2	1.0084	1.0151
as found 3rd	4979	21.1	200.3	200.0	0.3	195.1	192.1	3.0	1.0265	1.0410
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

### Average Correction Factor

Corrected As found	NO <sub>x</sub> = 796.3 ppb	NO = 792.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.4%
Previous Response	NO <sub>x</sub> = 799.4 ppb	NO = 797.8 ppb		*Percent Change	NO = -0.7%
Baseline Corr 2nd pt	NO <sub>x</sub> = 397.5 ppb	NO = 394.2 ppb	As found	NO <sub>x</sub> r <sup>2</sup> : 0.999975	Nx SI: 0.995540
Baseline Corr 3rd pt	NO <sub>x</sub> = 195.4 ppb	NO = 192.3 ppb	As found	NO r <sup>2</sup> : 0.999941	NO SI: 0.992467
			As found	NO <sub>2</sub> r <sup>2</sup> : 0.999983	NO <sub>2</sub> SI: 1.001405
					NO <sub>2</sub> Int: -0.788

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	-0.1	----	----
as found GPT point (400 ppb NO <sub>2</sub> )	782.2	417.7	365.7	365.8	0.9997	100.0%
as found GPT point (200 ppb NO <sub>2</sub> )	782.2	599.6	183.8	182.8	1.0054	99.5%
as found GPT point (100 ppb NO <sub>2</sub> )	782.2	687.1	96.3	95.0	1.0135	98.7%
1st GPT point (400 ppb O <sub>3</sub> )						
2nd GPT point (200 ppb O <sub>3</sub> )						
3rd GPT point (100 ppb O <sub>3</sub> )						

### Average Correction Factor

Notes:

Removal calibration.


Calibration Performed By:

Sean Bala

NO<sub>x</sub> Calibration Plot

Date: August 10, 2023

Location: Jackfish 1





# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-03-2017

### Station Information

Station Name:	Jackfish 1	Station Number:	AMS 506
Calibration Date:	August 10, 2023	Prev Cal Date:	July 7, 2022
Start Time (MST):	10:46	End Time (MST):	11:39
Barometric Press:	N/A	Station Temp:	21.5 Deg C
Reason:	Removal		

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	Y18362
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Actual Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	Correction factor (Cv/Iv) <i>Limit = 0.95-1.05</i>
0	0.0	0.0	n/a
200	20.2	20.2	0.9981
400	39.4	39.4	0.9990
600	58.6	58.6	0.9992
800	77.8	77.8	0.9994
<b>Average Correction Factor</b>			<b>0.9989</b>

	<i>Start</i>	<i>Finish</i>	<i>Limits</i>
Correl Coeff (r <sup>2</sup> )	0.999999	1.000000	≥0.995
Calculated slope	0.998858	0.999465	0.90 - 1.10
Calculated intercept	0.034341	-0.013446	+/- 2

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	R14654
As Found Declination (deg east of North)	<u>14</u>	As Left Declination (deg east of North)	<u>NA</u>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	Correction factor (Cv/Iv) <i>Limit = 0.95-1.05</i>
0	0.0	n/a
90	91.2	0.9868
180	179.7	1.0017
270	270.2	0.9993
356	357.0	0.9972
<b>Average Correction Factor</b>		<b>0.9962</b>

	<i>Start</i>	<i>Finish</i>	<i>Limits</i>
Correl Coeff (r <sup>2</sup> )	0.999989	0.999990	≥0.995
Calculated slope	1.006027	0.998879	0.90 - 1.10
Calculated intercept	0.130799	-0.218717	+/- 7

Notes: Removal Calibration.

Calibration Performed By: Sean Bala



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS507**  
**KIRBY SOUTH**  
**JULY 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

August 31, 2023







# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	NA
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	13:15	End Time (MST):	16:00
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999990	≥0.995
799.6	801.5	0.9977			
400.3	399.0	1.0034	Slope	1.002607	0.90 - 1.10
199.7	199.0	1.0033			
			Intercept	-0.968916	+/-30



SO2 Calibration Plot

Date: July 6, 2023

Location: Kirby South





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	July 7, 2023	Last Cal Date:	NA
Start time (MST):	10:30	End time (MST):	14:54
Reason:	Install		

### Calibration Standards

Cal Gas Concentration:	5.167	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	<u>CC517378</u>			
Removed Cal Gas Conc:	5.167	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3804
ZAG Make/Model:	API T701H		Serial Number:	880

### Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1150840012
Converter make:	Global	Converter serial #:	2022-197
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.005741	Backgd or Offset:	NA	3.00
Calibration intercept:	NA	-0.280895	Coeff or Slope:	NA	1.069

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4923	77.4	80.0	80.4	0.995
second point	4961	38.8	40.1	39.7	1.010
third point	4981	19.3	19.9	19.5	1.023
as left zero	5000	0.0	0.0	-0.3	----
as left span	4923	77.4	80.0	79.8	1.002
SO2 Scrubber Check	4919	81.3	813.0	-0.1	----
Date of last scrubber change:	14-Jun-23			Ave Corr Factor	1.009
Date of last converter efficiency test:	efficiency				

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

\* = > +/-5% change initiates investigation

Notes: Adjusted zero and span. Scrubber check done after cal zero.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 7, 2023	Previous Calibration:	NA
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	10:30	End Time (MST):	14:54
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
80.0	80.4	0.9948		
40.1	39.7	1.0100	Slope	0.90 - 1.10
19.9	19.5	1.0227		
			Intercept	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 7, 2023

Location: Kirby South





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	July 11, 2023	Last Cal Date:	July 7, 2023
Start time (MST):	10:30	End time (MST):	14:46
Reason:	Maintenance		

### Calibration Standards

Cal Gas Concentration:	5.167	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	<u>CC517378</u>			
Removed Cal Gas Conc:	5.167	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3804
ZAG Make/Model:	API T701H		Serial Number:	880

### Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1150840012
Converter make:	Global	Converter serial #:	2022-197
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005741	1.004463	Backgd or Offset:	3.00	2.00
Calibration intercept:	-0.280895	-0.061147	Coeff or Slope:	1.069	1.069

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-1.0	----
as found span	4923	77.4	80.0	79.3	0.996
as found 2nd point	4961	38.8	40.1	39.0	1.002
as found 3rd point	4981	19.3	19.9	18.8	1.007
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.4	80.0	80.3	0.996
second point	4961	38.8	40.1	40.2	0.997
third point	4981	19.3	19.9	19.9	1.002
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.4	80.0	80.2	0.997

#### SO<sub>2</sub> Scrubber Check

Date of last scrubber change:	14-Jun-23	Ave Corr Factor	0.999
Date of last converter efficiency test:		efficiency	

Baseline Corr As found:	80.3	Prev response:	80.16	*% change:	0.2%
Baseline Corr 2nd AF pt:	40.0	AF Slope:	1.004602	AF Intercept:	-1.141023
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999984		

\* = > +/-5% change initiates investigation

Notes: Adjusted zero due to baseline drift.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 11, 2023	Previous Calibration:	July 7, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	10:30	End Time (MST):	14:46
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999997	
80.0	80.3	0.9960			≥0.995
40.1	40.2	0.9975	Slope	1.004463	
19.9	19.9	1.0022			0.90 - 1.10
			Intercept	-0.061147	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 11, 2023

Location: Kirby South







# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Kirby South Station number: AMS507  
 Calibration Date: July 25, 2023 Last Cal Date: July 11, 2023  
 Start time (MST): 11:30 End time (MST): 19:00  
 Reason: Maintenance

### Calibration Standards

Cal Gas Concentration: 5.167 ppm Cal Gas Exp Date: February 5, 2024  
 Cal Gas Cylinder #: CC517378  
 Removed Cal Gas Conc: 5.167 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 3804  
 ZAG Make/Model: API T701H Serial Number: 880

### Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1150840012  
 Converter make: Global Converter serial #: 2022-197  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004463	1.005318	Backgd or Offset:	2.00	1.64
Calibration intercept:	-0.061147	0.458932	Coeff or Slope:	1.069	1.109

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4923	77.4	80.0	79.9	1.001
as found 2nd point	4961	38.8	40.1	39.9	1.005
as found 3rd point	4981	19.3	19.9	19.6	1.018
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4923	77.4	80.0	80.7	0.991
second point	4961	38.8	40.1	40.9	0.980
third point	4981	19.3	19.9	20.9	0.954
as left zero	5000	0.0	0.0	0.6	----
as left span	4923	77.4	80.0	81.2	0.985

SO2 Scrubber Check

Date of last scrubber change:	25-Jul-23	Ave Corr Factor	0.975
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 79.9 Prev response: 80.27 \*% change: -0.5%  
 Baseline Corr 2nd AF pt: 39.9 AF Slope: 1.000178 AF Intercept: -0.161170  
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999981

\* = > +/-5% change initiates investigation

Notes: Replaced scrubber due to slow response during nightly spans. Adjusted zero and span. Scrubber check passed.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	July 25, 2023	Previous Calibration:	July 11, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	11:30	End Time (MST):	19:00
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999908	≥0.995
80.0	80.7	0.9911			
40.1	40.9	0.9804	Slope	1.005318	0.90 - 1.10
19.9	20.9	0.9542			
			Intercept	0.458932	+/-3



H<sub>2</sub>S Calibration Plot

Date: July 25, 2023

Location: Kirby South







# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	July 6, 2023	Previous Calibration:	NA
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	13:58	End Time (MST):	16:00
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.04	----	Correlation Coefficient	0.999986	≥0.995
17.26	17.21	1.0031			
8.64	8.54	1.0119	Slope	0.999300	0.90 - 1.10
4.31	4.24	1.0166			
			Intercept	-0.060190	+/-1.5



THC Calibration Plot

Date: July 6, 2023

Location: Kirby South





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	July 7, 2023	Last Cal Date:	NA
Start time (MST):	10:30	End time (MST):	15:32
Reason:	Install		

### Calibration Standards

NO Gas Cylinder #:	T34ULGL	Cal Gas Expiry Date:	March 8, 2025
NOX Cal Gas Conc:	49.39 ppm	NO Cal Gas Conc:	49.02 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	49.39 ppm	Removed Gas NO Conc:	49.02 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T750	Serial Number:	282
ZAG make/model:	API 751H	Serial Number:	321

### Analyzer Information

Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	NA	1.429	NO bkgnd or offset:	NA	1.9
NOX coeff or slope:	NA	0.992	NOX bkgnd or offset:	NA	2.0
NO2 coeff or slope:	NA	1.000	Reaction cell Press:	NA	162.66

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	NA	1.001602
NO <sub>x</sub> Cal Offset:	NA	-2.351758
NO Cal Slope:	NA	1.003048
NO Cal Offset:	NA	-3.253489
NO <sub>2</sub> Cal Slope:	NA	1.006879
NO <sub>2</sub> Cal Offset:	NA	1.400873



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.6	0.0	----	----
high point	4919	81.0	800.1	794.1	6.0	800.0	794.7	5.7	1.0001	0.9993
second point	4960	40.5	400.0	397.0	3.0	397.2	393.4	3.8	1.0071	1.0092
third point	4980	20.2	199.5	198.0	1.5	195.9	192.9	3.1	1.0185	1.0266
as left zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.6	0.0	----	----
as left span	4919	81.0	800.1	392.0	408.1	779.8	376.2	403.6	1.0261	1.0421
Average Correction Factor									1.0086	1.0117

Corrected As found	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	<i>* =&gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = NA
Previous Response	NO <sub>x</sub> = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
					As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.7	392.6	408.1	411.4	0.9920	100.8%
2nd GPT point (200 ppb O3)	794.7	597.3	203.4	207.5	0.9802	102.0%
3rd GPT point (100 ppb O3)	794.7	697.8	102.9	106.0	0.9707	103.0%
Average Correction Factor					0.9810	101.9%

Notes: Install calibration for the move to Kirby South. Installed an external pump. Adjusted zero and span. Large span adjustment made due to the large change in flow.

Calibration Performed By: Braiden Boutilier





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 7, 2023	Previous Calibration:	NA
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	10:30	End Time (MST):	15:32
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.1	800.0	1.0001		
400.0	397.2	1.0071		
199.5	195.9	1.0185		
			0.999978	
			1.001602	
			-2.351758	





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	July 7, 2023	Previous Calibration:	NA
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	10:30	End Time (MST):	15:32
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
794.1	794.7	0.9993		
397.0	393.4	1.0092		
198.0	192.9	1.0266		
			0.999949	
			1.003048	
			-3.253489	

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	July 7, 2023	Previous Calibration:	NA
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	10:30	End Time (MST):	15:32
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
408.1	411.4	0.9920		
203.4	207.5	0.9802		
102.9	106.0	0.9707		



NO<sub>x</sub> Calibration Plot

Date: July 7, 2023

Location: Kirby South





# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Kirby South	Station Number:	AMS 507
Calibration Date:	July 7, 2023	Prev Cal Date:	NA
Start Time (MST):	12:15	End Time (MST):	14:15
Tower Height (m):	10.0	Reason:	Install

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	X16479
WS Calibrator:	Young 18802	Serial Number:	03845

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.4%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999999	$\geq 0.9995$
Calculated slope		0.999615	$0.90 - 1.10$
Calculated intercept		-0.025296	$\pm 2$

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	X16495
As Found Declination (deg east of True North):	<u>NA</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:28	Calc Declination*:	14 Degrees
Deadband calc:	0.3 degrees ( <i>Limit 4 deg</i> )		<small>* - calculated declination as per NOAA website</small>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	2.9	---
90	92.5	0.7%
180	182.0	0.6%
270	271.3	0.4%
357	359.6	0.7%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999992	$\geq 0.9995$
Calculated slope		1.001983	$0.90 - 1.10$
Calculated intercept		-2.611256	$\pm 4$

Notes: All points passed. Aligned tower using solar noon and verified with a compass.

Calibration Performed By: Braiden Boutilier



End of Report