



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

AUGUST 2023

MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

September 29, 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01 BERTHA GANTER - FORT MCKAY AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	August 1, 2023	Last Cal Date:	July 14, 2023
Start time (MST):	8:37	End time (MST):	12:28
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC418809			
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000680	0.999194	Backgd or Offset:	19.2	19.2
Calibration intercept:	-0.433310	0.086933	Coeff or Slope:	0.892	0.886

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4918	81.3	800.3	800.1	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.3	----
high point	4918	81.3	800.3	800.1	1.000
second point	4959	40.7	400.6	399.4	1.003
third point	4979	20.3	199.8	200.1	0.999
as left zero	5000	0.0	0.0	0.7	----
as left span	4918	81.3	800.3	800.6	1.000
Average Correction Factor					1.001

Baseline Corr As found:	799.90	Previous response	800.38	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 14, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:37	End Time (MST):	12:28
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999996	≥0.995
800.3	800.1	1.0002			
400.6	399.4	1.0030	Slope	0.999194	0.90 - 1.10
199.8	200.1	0.9986			
			Intercept	0.086933	+/-30



SO2 Calibration Plot

Date: August 1, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: August 28, 2023 Last Cal Date: July 18, 2023
 Start time (MST): 9:08 End time (MST): 15:22
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153461
 Converter make: CD Nova Converter serial #: 470
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000507	1.000364	Backgd or Offset: 2.28	2.29
Calibration intercept:	0.159999	0.140000	Coeff or Slope: 0.919	0.919

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4921	78.4	80.0	78.2	1.026
as found 2nd point	4960	39.2	40.0	39.7	1.013
as found 3rd point	4980	19.6	20.0	19.9	1.015
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	78.4	80.0	80.2	0.997
second point	4960	39.2	40.0	40.1	0.997
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	78.4	80.0	79.7	1.004
SO2 Scrubber Check	4919	81.3	813.0	0.0	----

Date of last scrubber change:	December 17, 2021	Ave Corr Factor	0.997
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 78.0 Prev response: 80.19 *% change: -2.8%
 Baseline Corr 2nd AF pt: 39.5 AF Slope: 0.974934 AF Intercept: 0.379995
 Baseline Corr 3rd AF pt: 19.7 AF Correlation: 0.999950

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 28, 2023	Previous Calibration:	July 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:08	End Time (MST):	15:22
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999997	≥0.995
80.0	80.2	0.9974			
40.0	40.1	0.9975	Slope	1.000364	0.90 - 1.10
20.0	20.1	0.9949			
			Intercept	0.140000	+/-3



TRS Calibration Plot

Date: August 28, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: August 28, 2023 Last Cal Date: July 18, 2023
 Start time (MST): 9:08 End time (MST): 15:22
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1200326167
 Converter make: Thermo Converter Converter serial #: N/A
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002857	1.003143	Backgd or Offset:	1.95 1.74
Calibration intercept:	0.056779	0.096765	Coeff or Slope:	1.014 1.014

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4922	78.4	80.0	82.5	0.967
as found 2nd point	4960	39.2	40.0	41.1	0.968
as found 3rd point	4980	19.6	20.0	19.9	0.995
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	78.4	80.0	80.2	0.997
second point	4960	39.2	40.0	40.5	0.988
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.7	----
as left span	4922	78.4	80.0	78.6	1.018
SO2 Scrubber Check	4919	81.3	813.0	-0.1	----
Date of last scrubber change:	March 21, 2022			Ave Corr Factor	0.993
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 82.7 Prev response: 80.25 *% change: 3.0%
 Baseline Corr 2nd AF pt: 41.3 AF Slope: 1.036640 AF Intercept: -0.443312
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999942

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted zero only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 28, 2023	Previous Calibration:	July 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:08	End Time (MST):	15:22
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	≥0.995
80.0	80.2	0.9972		
40.0	40.5	0.9876	Slope	0.90 - 1.10
20.0	20.1	0.9949		
			Intercept	+/-3



H₂S Calibration Plot

Date: August 28, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	August 1, 2023	Last Cal Date:	July 14, 2023
Start time (MST):	8:37	End time (MST):	12:28
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	5609

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.90E-04	2.91E-04	NMHC SP Ratio:	6.14E-05
CH ₄ Retention time:	14.6	14.6	NMHC Peak Area:	149639
				145218

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.27	17.03	1.014
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	17.27	17.21	1.003
second point	4959	40.7	8.64	8.52	1.015
third point	4980	20.3	4.31	4.25	1.014
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.27	17.25	1.001

Average Correction Factor				1.011
Baseline Corr AF:	17.03	Prev response	17.24	*% change -1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4918	81.3	9.18	8.98	1.023
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4918	81.3	9.18	9.15	1.004
second point	4959	40.7	4.60	4.58	1.004
third point	4980	20.3	2.29	2.31	0.993
as left zero	5000	0	0.00	0.00	----
as left span	4918	81.3	9.18	9.21	0.997
Average Correction Factor					1.000
Baseline Corr AF:	8.98	Prev response	9.17	*% change	-2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	8.06	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	8.09	8.07	1.003
second point	4959	40.7	4.05	3.94	1.027
third point	4980	20.3	2.02	1.94	1.039
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.05	1.005
Average Correction Factor					1.023
Baseline Corr AF:	8.06	Prev response	8.06	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001595	0.997017
THC Cal Offset:	-0.058682	-0.038093
CH ₄ Cal Slope:	1.002288	0.999108
CH ₄ Cal Offset:	-0.042956	-0.047957
NMHC Cal Slope:	1.000424	0.995361
NMHC Cal Offset:	-0.014727	0.009863

Notes: Changed the inlet filter and N2 cylinder after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 14, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:37	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999959	≥ 0.995
17.27	17.21	1.0032			
8.64	8.52	1.0149			
4.31	4.25	1.0141			
			Slope	0.997017	0.90 - 1.10
			Intercept	-0.038093	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 14, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:37	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999795	≥ 0.995
8.09	8.07	1.0026			
4.05	3.94	1.0272			
2.02	1.94	1.0395			
			Slope	0.999108	0.90 - 1.10
			Intercept	-0.047957	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 14, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	8:37	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999991	≥ 0.995			
9.18	9.15	1.0037						
4.60	4.58	1.0036				Slope	0.995361	0.90 - 1.10
2.29	2.31	0.9931						
			Intercept	0.009863	± 0.5			



NMHC Calibration Plot

Date: August 1, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
Calibration Date: August 25, 2023 Last Cal Date: July 11, 2023
Start time (MST): 9:36 End time (MST): 14:50
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1P9L Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.84 ppm NO Cal Gas Conc: 50.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.84 ppm Removed Gas NO Conc: 50.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3565
ZAG make/model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1218153357
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.324	1.348	NO bkgnd or offset:	6.9	6.9
NOX coeff or slope:	0.991	0.992	NOX bkgnd or offset:	7.0	7.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	190.7	189.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998356	1.000155
NO _x Cal Offset:	-0.420000	-0.260000
NO Cal Slope:	0.998644	0.998858
NO Cal Offset:	-1.180000	-1.280000
NO ₂ Cal Slope:	1.002441	1.001610
NO ₂ Cal Offset:	0.783625	0.629963



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as found span	4920	80.0	813.4	800.6	12.8	798.0	781.4	16.6	1.0193	1.0246
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
high point	4920	80.0	813.4	800.6	12.8	813.5	799.0	14.5	0.9999	1.0021
second point	4960	40.0	406.7	400.3	6.4	406.4	398.2	8.2	1.0008	1.0053
third point	4980	20.0	203.4	200.2	3.2	202.6	197.2	5.4	1.0038	1.0150
as left zero	5000	0.0	0.0	0.0	0.0	0.2	-0.1	0.3	----	----
as left span	4920	80.0	813.4	408.8	404.6	812.0	402.5	409.5	1.0018	1.0158
Average Correction Factor									1.0015	1.0075

Corrected As found	NO _x = 797.9 ppb	NO = 781.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.7%	
Previous Response	NO _x = 811.7 ppb	NO = 798.4 ppb		*Percent Change	NO = -2.2%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.5	403.7	404.6	405.6	0.9975	100.2%
2nd GPT point (200 ppb O3)	795.5	593.9	214.4	215.7	0.9940	100.6%
3rd GPT point (100 ppb O3)	795.5	694.1	114.2	115.4	0.9896	101.1%
Average Correction Factor					0.9937	100.6%

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 25, 2023	Previous Calibration:	July 11, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:36	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.4	813.5	0.9999		
406.7	406.4	1.0008		
203.4	202.6	1.0038		
			0.999998	
			1.000155	
			-0.260000	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 25, 2023	Previous Calibration:	July 11, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:36	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.6	799.0	1.0021		
400.3	398.2	1.0053		
200.2	197.2	1.0150		
			0.999988	
			0.998858	
			-1.280000	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 25, 2023	Previous Calibration:	July 11, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:36	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
404.6	405.6	0.9975		
214.4	215.7	0.9940		
114.2	115.4	0.9896		
			0.999994	
			1.001610	
			0.629963	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 25, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: August 10, 2023 Last Cal Date: July 4, 2023
 Start time (MST): 9:43 End time (MST): 13:08
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 1107
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001486	1.001743	Backgd or Offset:	3.2	3.2
Calibration intercept:	0.240000	-0.080000	Coeff or Slope:	1.010	1.010

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	5000	863.1	400.0	399.0	1.003
as found 2nd point	5000	742.5	200.0	199.2	1.004
as found 3rd point	5000	651.7	100.0	99.9	1.001
calibrator zero	5000	0.0	0.0	-0.1	----
high point	5000	863.1	400.0	400.6	0.999
second point	5000	742.5	200.0	200.3	0.999
third point	5000	651.7	100.0	100.1	0.999
as left zero	5000	0.0	0.0	-0.1	----
as left span	5000	863.1	400.0	404.4	0.989
Average Correction Factor					0.999

Baseline Corr As found:	399.0	Previous response	400.8	*% change	-0.5%
Baseline Corr 2nd AF pt:	199.2	AF Slope:	0.997286	AF Intercept:	0.000000
Baseline Corr 3rd AF pt:	99.9	AF Correlation:	0.999999		

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Adjusted the UV source lamp voltage. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 4, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:43	End Time (MST):	13:08
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	1.000000	≥0.995
400.0	400.6	0.9985			
200.0	200.3	0.9985	Slope	1.001743	0.90 - 1.10
100.0	100.1	0.9990			
			Intercept	-0.080000	+/- 5



O₃ Calibration Plot

Date: August 10, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
 Calibration Date: August 23, 2023 Last Cal Date: July 18, 2023
 Start time (MST): 11:29 End time (MST): 12:57

Analyzer Make: Teledyne API T640 S/N: 324
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	19.3	18.6	19.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	736	735	736	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.07	5.00	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: August 23, 2023 Last Cal Date: July 18, 2023
 PM w/o HEPA: 23.8 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: _____ w/ HEPA: _____
 Date Optical Chamber Cleaned: July 18, 2023 <0.2 ug/m3
 Disposable Filter Changed: August 23, 2023

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: July 18, 2023

Notes: Flow, temperature, and pressure all within limits. Leak checks passed. Disposable filter changed.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	August 21, 2023	Last Cal Date:	July 17, 2023
Start time (MST):	10:24	End time (MST):	13:56
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	<u>3040</u>	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>3040</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002385	1.001990	Backgd or Offset:	-0.012	-0.012
Calibration intercept:	0.147828	0.159827	Coeff or Slope:	0.992	0.992

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.6	40.9	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.6	40.7	0.997
second point	4966	33.3	20.2	20.7	0.977
third point	4983	16.7	10.2	10.3	0.982
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.5	40.1	1.010
Average Correction Factor					0.986

Baseline Corr As found:	40.75	Prev response:	40.80	*% change:	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 21, 2023	Previous Calibration:	July 17, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:24	End Time (MST):	13:56
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999877	≥ 0.995
40.6	40.7	0.9974	Slope	1.001990	0.90 - 1.10
20.2	20.7	0.9773	Intercept	0.159827	± 1.5
10.2	10.3	0.9820			



CO Calibration Plot

Date: August 21, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	August 11, 2023	Last Cal Date:	July 10, 2023
Start time (MST):	9:43	End time (MST):	12:52
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>60,200</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
N2 Gen Make/Model:	Peak Sci NG5000		Serial Number:	7220900034

Analyzer Information

Analyzer make: Teledyne API 360	Analyzer serial #: 442
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000422	1.000529	Backgd or Offset:	0.045	0.045
Calibration intercept:	-4.480000	-5.380000	Coeff or Slope:	0.874	0.874

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-0.1	----
as found span	2920	80.0	1605.3	1600.4	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.0	----
high point	2920	80.0	1605.3	1603.5	1.001
second point	2960	40.0	802.7	795.0	1.010
third point	2980	20.0	401.3	390.8	1.027
as left zero	3000	0.0	0.0	0.0	----
as left span	2960	40.0	802.7	784.3	1.023
Average Correction Factor					1.013

Baseline Corr As found:	1600.50	Prev response:	1601.53	*% change:	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	August 11, 2023	Previous Calibration	July 10, 2023
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS01
Start Time (MST)	9:43	End Time (MST)	12:52
Analyzer make	Teledyne API 360	Analyzer serial #	442

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999949	≥0.995
1605.3	1603.5	1.0011	Slope	1.000529	0.90 - 1.10
802.7	795.0	1.0096	Intercept	-5.380000	+/-10
401.3	390.8	1.0270			



CO₂ Calibration Plot

Date: August 11, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	August 23, 2023	Last Cal Date:	July 12, 2023
Start time (MST):	9:48	End time (MST):	14:20
NH3 Cal Date:	August 23, 2023	Last Cal Date:	July 13, 2023
Start time (MST):	14:45	End time (MST):	16:50
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	50.84	ppm	NO Gas Cylinder #:	T2Y1P9L
NO Cal Gas Conc:	50.04	ppm	NO Cal Gas Expiry:	March 3, 2028
Removed NOX Conc:	50.84	ppm	Removed Cylinder #:	NA
Removed NO Conc:	50.04	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	74.90	ppm	NH3 Gas Cylinder #:	CC744566
Removed NH3 Conc:	74.90	ppm	NH3 Cal Gas Expiry:	December 21, 2023
NH3 gas Diff:			Removed Cylinder #:	NA
Calibrator Model:	Teledyne API T700		Removed cyl Expiry:	NA
ZAG make/model:	Teledyne API T701		Serial Number:	3565
			Serial Number:	5609

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	475
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	5.10
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	514

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.987	0.987	TN coefficient:	0.989	0.989
NOX coefficient:	0.987	0.987	NO bkgnd:	-0.9	-0.9
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-0.3	-0.3
NH3 coefficient:	0.973	0.973	TN bkgnd:	1.2	1.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000126	0.998131
NO _x Cal Offset:	-1.500000	-2.040000
NO Cal Slope:	1.000428	0.998944
NO Cal Offset:	-2.080000	-2.460000
NO ₂ Cal Slope:	0.999045	0.999734
NO ₂ Cal Offset:	0.023464	0.074243
NH3 Cal Slope:	1.002916	1.003610
NH3 Cal Offset:	-5.175346	0.576979
TN Cal Slope:	1.005585	1.006104
TN Cal Offset:	-5.152839	0.744017



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-2.2	-2.3	0.1	----	----
as found NO	4920	80.0	813.4	813.4	----	808.0	810.0	-2.2	1.007	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high NO point	4920	80.0	813.4	813.4	----	810.1	811.1	-1.2	1.004	----
NO/O3 point	4920	80.0	813.4	813.4	----	806.0	803.7	2.1	1.009	----
as found NH3	3416	84.1	1799.7	----	1799.7	1811.5	----	1807.0	0.994	0.996
new NH3 cyl rp							----			
first NH3	3416	84.1	1799.7	----	1799.7	1811.5	----	1807.0	0.994	0.996
second NH3	3453	46.7	999.4	----	999.4	1005.3	----	1002.5	0.994	0.997
third NH3	3477	23.4	500.8	----	500.8	506.3	----	504.6	0.989	0.992
Average Correction Factor									1.0067	0.9951

Corrected As found TN = 810.2 ppb NO_x = 812.3 ppb NH3 = 1806.9 ppb

Previous Response TN = 812.8 ppb NO_x = 812.0 ppb NH3 = 1799.8 ppb

NH3 Previous Converter Efficiency = 97.3%

NH3 Current Converter Efficiency = 97.3%

*Percent Change TN = -0.3%

*Percent Change NO_x = 0.0%

*Percent Change NH3 = 0.4%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-2.3	-2.4	-2.2	----	----
as found span	4920	80.0	813.4	800.6	813.4	810.0	797.0	808.0	1.0042	1.0046
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.1	----	----
high point	4920	80.0	813.4	800.6	813.4	811.1	799.0	810.1	1.0029	1.0021
second point	4960	40.0	406.7	400.3	406.7	402.2	395.0	402.5	1.0112	1.0135
third point	4980	20.0	203.4	200.2	203.4	199.5	195.7	199.5	1.0193	1.0228
Average Correction Factor									1.0112	1.0128

Baseline Corr As fnd	TN = 810.2 ppb	NO _x = 812.3 ppb	NO = 799.4 ppb	*Percent Change	TN = -0.3%
Previous Response	TN = 812.8 ppb	NO _x = 812.0 ppb	NO = 798.9 ppb	*Percent Change	NO _x = 0.0%
				*Percent Change	NO = 0.1%

* = > +/-5% change initiates investigation

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	0.2	----	----
calibration zero	----	----	0.0	-0.2	----	----
1st GPT point (400 ppb O3)	791.1	389.0	414.9	414.6	1.0007	99.9%
2nd GPT point (200 ppb O3)	791.1	589.9	214.0	214.5	0.9977	100.2%
3rd GPT point (100 ppb O3)	791.1	691.2	112.7	112.8	0.9991	100.1%
Average Correction Factor					0.9992	100.1%

Notes: Changed the inlet filter after as founds. No adjustments made. Used the 2nd GPT reference point due to drift.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TN Calibration Summary

Version-05-2023


Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 12, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:48	End Time (MST):	14:20
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1799.7	1811.5	0.9935		
999.4	1005.3	0.9941		
500.8	506.3	0.9891		
			0.999997	
			1.006104	
			0.744017	

TN Calibration Curve





Wood Buffalo Environmental Association

NH₃ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 12, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:48	End Time (MST):	14:20
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1799.7	1807.0	0.9960		
999.4	1002.5	0.9969		
500.8	504.6	0.9924		
			0.999998	
			1.003610	
			0.576979	

NH₃ Calibration Curve





Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-05-2023


Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 12, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:48	End Time (MST):	14:20
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.4	811.1	1.0029		
406.7	402.2	1.0112		
203.4	199.5	1.0193		
			0.999971	
			0.998131	
			-2.040000	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 12, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:48	End Time (MST):	14:20
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.6	799.0	1.0021			
400.3	395.0	1.0135			
200.2	195.7	1.0228			
			Slope	0.998944	0.90 - 1.10
			Intercept	-2.460000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 12, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:48	End Time (MST):	14:20
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
414.9	414.6	1.0007			
214.0	214.5	0.9977			
112.7	112.8	0.9991			
			Slope	0.999734	0.90 - 1.10
			Intercept	0.074243	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 23, 2023


Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: August 23, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Fort McKay - Bertha Ganter	Station Number:	AMS 01
Calibration Date:	August 25, 2023	Prev Cal Date:	September 16, 2022
Start Time (MST):	11:43	End Time (MST):	13:17
Tower Height (m):	10.0	Reason:	Routine

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	P10041
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.5%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999999	0.999999	≥0.9995
Calculated slope	0.999594	0.999594	0.90 - 1.10
Calculated intercept	-0.028293	-0.028293	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	R14655
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:28	Calc Declination*:	13.74 Degrees
Deadband calc:	1.5 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	1.1	---
90	90.6	0.2%
180	180.2	0.1%
270	270.7	0.2%
357	356.6	-0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999996	0.999997	≥0.9995
Calculated slope	1.003118	1.003199	0.90 - 1.10
Calculated intercept	-0.399196	-1.004699	+/- 4

Notes: Performing annual MET sensor calibration. Bearings are good. Crossarm declination was aligned with true North using a compass.

Calibration Performed By: Rene Chamberland




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	August 15, 2023	Last Cal Date:	July 12, 2023
Start time (MST):	9:49	End time (MST):	13:58
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1185
ZAG Make/Model:	API T701		Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000728	1.008382	Backgd or Offset:	17.8	18.1
Calibration intercept:	-1.304958	-0.263519	Coeff or Slope:	0.783	0.797

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4920	80.2	801.6	787.2	1.018
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	80.2	801.6	808.0	0.992
second point	4960	40.1	400.8	404.4	0.991
third point	4980	20.0	199.9	200.8	0.996
as left zero	5000	0.0	0.0	-0.1	----
as left span	4920	80.2	801.6	804.0	0.997
Average Correction Factor					0.993

Baseline Corr As found:	787.50	Previous response	800.93	*% change	-1.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 12, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:49	End Time (MST):	13:58
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	≥0.995
801.6	808.0	0.9921		
400.8	404.4	0.9912	Slope	0.90 - 1.10
199.9	200.8	0.9956		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 15, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mildred Lake Station number: AMS02
 Calibration Date: August 30, 2023 Last Cal Date: July 20, 2023
 Start time (MST): 9:16 End time (MST): 17:40
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.29 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345191
 Removed Cal Gas Conc: 5.29 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1185
 ZAG Make/Model: API T701 Serial Number: 5608

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 12113311966
 Converter make: Global G150 Converter serial #: 2022-198
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994964	1.031254	Backgd or Offset: 1.77	1.79
Calibration intercept:	-0.019203	-0.239187	Coeff or Slope: 0.821	0.821

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4924	75.6	80.0	79.9	1.001
as found 2nd point	4962	37.8	40.0	39.3	1.018
as found 3rd point	4981	18.9	20.0	19.3	1.036
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4924	75.6	80.0	82.6	0.968
second point	4962	37.8	40.0	40.4	0.990
third point	4981	18.9	20.0	20.2	0.990
as left zero	5000	0.0	0.0	0.1	----
as left span	4924	75.6	80.0	78.8	1.015
SO2 Scrubber Check	4920	80.2	802.0	-0.1	----
Date of last scrubber change:	12-Sep-22			Ave Corr Factor	0.983
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	79.9	Prev response:	79.57	*% change:	0.4%
Baseline Corr 2nd AF pt:	39.3	AF Slope:	1.000822	AF Intercept:	-0.399208
Baseline Corr 3rd AF pt:	19.3	AF Correlation:	0.999879		

* = > +/-5% change initiates investigation

Notes: Scrubber check done after MPAF's, passed. Ran room air to hydrate scrubber beads as the span was low. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 30, 2023	Previous Calibration:	July 20, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:16	End Time (MST):	17:40
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999808	≥0.995
80.0	82.6	0.9684			
40.0	40.4	0.9900	Slope	1.031254	0.90 - 1.10
20.0	20.2	0.9899			
			Intercept	-0.239187	+/-3



H₂S Calibration Plot

Date: August 30, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	August 15, 2023	Last Cal Date:	July 20, 2023
Start time (MST):	9:49	End time (MST):	13:58
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
C ₃ H ₈ Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
Removed C ₃ H ₈ Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.97E-04	2.95E-04	NMHC SP Ratio:	4.58E-05
CH ₄ Retention time:	14.8	14.6	NMHC Peak Area:	192143
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	17.10	0.984
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.79	1.002
second point	4960	40.1	8.41	8.34	1.008
third point	4980	20.0	4.19	4.17	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	17.02	0.988
Average Correction Factor					1.005

Baseline Corr AF:	17.10	Prev response	16.81	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	8.89	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.78	1.002
second point	4960	40.1	4.40	4.37	1.005
third point	4980	20.0	2.19	2.20	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.83	0.996
Average Correction Factor					1.001
Baseline Corr AF:	8.89	Prev response	8.80	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	8.22	0.976
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.00	1.003
second point	4960	40.1	4.01	3.97	1.011
third point	4980	20.0	2.00	1.97	1.017
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	8.20	0.978
Average Correction Factor					1.010
Baseline Corr AF:	8.22	Prev response	8.01	*% change	2.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000332	0.998327
THC Cal Offset:	-0.014309	-0.018313
CH ₄ Cal Slope:	1.000556	0.997907
CH ₄ Cal Offset:	-0.018253	-0.017457
NMHC Cal Slope:	1.000114	0.997620
NMHC Cal Offset:	0.003745	0.001341

Notes: Changed nitrogen cylinder. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary

Version-06-2022


Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 20, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:49	End Time (MST):	13:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999987	≥ 0.995
16.82	16.79	1.0017			
8.41	8.34	1.0083			
4.19	4.17	1.0060			
			Slope	0.998327	0.90 - 1.10
			Intercept	-0.018313	+/-0.5

THC Calibration Curve





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 20, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:49	End Time (MST):	13:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999975	≥0.995
8.02	8.00	1.0029			
4.01	3.97	1.0107			
2.00	1.97	1.0172			
			Slope	0.997907	0.90 - 1.10
			Intercept	-0.017457	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 20, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:49	End Time (MST):	13:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999991	≥ 0.995			
8.80	8.78	1.0017						
4.40	4.37	1.0054				Slope	0.997620	0.90 - 1.10
2.19	2.20	0.9965						
			Intercept	0.001341	± 0.5			



NMHC Calibration Plot

Date: August 15, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Mildred Lake	Station Number:	AMS 02
Calibration Date:	August 30, 2023	Prev Cal Date:	October 6, 2022
Start Time (MST):	9:30	End Time (MST):	10:11
Tower Height (m):	10.0	Reason:	Routine

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	E5130
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.7	0.2%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥0.9995
Calculated slope		0.998137	0.90 - 1.10
Calculated intercept		0.028665	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	E5130
As Found Declination (deg east of True North):	<u>13</u>	As Left Declination (deg east of True North):	<u>13</u>
Solar noon time (MST):	11:27	Calc Declination*:	13.67 Degrees
Deadband calc:	-2.1 degrees (<i>Limit 4 deg</i>)	* - calculated declination as per NOAA website	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.6	---
90	87.6	-0.7%
180	179.3	-0.2%
270	273.3	0.9%
357	359.7	0.8%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999924	≥0.9995
Calculated slope		0.988811	0.90 - 1.10
Calculated intercept		1.321034	+/- 4

Notes: Verified crossarm aligned with true north. Both sensors were within acceptable limits.

Calibration Performed By: Braiden Boutilier, Devin Russel




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023







Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	August 9, 2023	Last Cal Date:	July 19, 2023
Start time (MST):	5:40	End time (MST):	8:23
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.87	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC446753			
Removed Cal Gas Conc:	50.87	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701		Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000328	0.999657	Backgd or Offset:	22.1	22.1
Calibration intercept:	0.714976	0.974700	Coeff or Slope:	0.860	0.860

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4921	78.6	799.7	799.0	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.5	----
high point	4921	78.6	799.7	800.0	1.000
second point	4961	39.3	399.8	401.5	0.996
third point	4980	19.6	199.4	200.4	0.995
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	78.6	799.7	799.3	1.001
Average Correction Factor					0.997

Baseline Corr As found:	798.60	Previous response	800.72	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 19, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	5:40	End Time (MST):	8:23
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	≥0.995
799.7	800.0	0.9997		
399.8	401.5	0.9958	Slope	0.90 - 1.10
199.4	200.4	0.9951		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 9, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: August 2, 2023 Last Cal Date: July 11, 2023
 Start time (MST): 7:50 End time (MST): 12:15
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.42 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345266
 Removed Cal Gas Conc: 5.42 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701H Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1008841400
 Converter make: Global Converter serial #: 2022-200
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994226	1.000770	Backgd or Offset: 1.8	1.8
Calibration intercept:	-0.017926	-0.097733	Coeff or Slope: 1.095	1.095

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.1	80.3	79.6	1.009
as found 2nd point	4963	37.0	40.1	40.1	1.000
as found 3rd point	4982	18.5	20.1	19.9	1.008
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4926	74.1	80.3	80.2	1.002
second point	4963	37.0	40.1	40.3	0.995
third point	4982	18.5	20.1	19.8	1.013
as left zero	5000	0.0	0.0	-0.1	----
as left span	4926	74.1	80.3	79.5	1.010
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	1.003
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.6 Prev response: 79.84 *% change: -0.3%
 Baseline Corr 2nd AF pt: 40.1 AF Slope: 0.991379 AF Intercept: 0.082088
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999974

* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 2, 2023	Previous Calibration:	July 11, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:50	End Time (MST):	12:15
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999971	≥0.995
80.3	80.2	1.0015			
40.1	40.3	0.9952	Slope	1.000770	0.90 - 1.10
20.1	19.8	1.0127			
			Intercept	-0.097733	+/-3



H₂S Calibration Plot

Date: August 2, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	August 9, 2023	Last Cal Date:	July 19, 2023
Start time (MST):	5:40	End time (MST):	8:22
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	1.840E-04	1.860E-04	NMHC SP Ratio:	3.910E-05
CH ₄ Retention time:	11.8	11.8	NMHC Peak Area:	225808
				222570

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	16.64	16.44	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	16.64	16.64	1.000
second point	4961	39.3	8.32	8.31	1.001
third point	4980	19.6	4.15	4.14	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	16.64	16.64	1.000

Average Correction Factor				1.001
Baseline Corr AF:	16.44	Prev response	16.64	*% change -1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* => +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	
as found span	4921	78.6	8.82	8.71	1.013
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	8.82	8.82	1.000
second point	4961	39.3	4.41	4.41	1.000
third point	4980	19.6	2.20	2.20	1.001
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	8.82	8.82	1.000
Average Correction Factor					1.000
Baseline Corr AF:	8.71	Prev response	8.82	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	7.82	7.74	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	7.82	7.82	1.000
second point	4961	39.3	3.91	3.90	1.002
third point	4980	19.6	1.95	1.94	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	7.82	7.82	1.000
Average Correction Factor					1.002
Baseline Corr AF:	7.74	Prev response	7.82	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000429	1.000360
THC Cal Offset:	-0.003535	-0.005535
CH ₄ Cal Slope:	1.002052	1.000736
CH ₄ Cal Offset:	-0.007900	-0.005903
NMHC Cal Slope:	1.000156	1.000144
NMHC Cal Offset:	0.002367	-0.000833

Notes:

Span Adjusted. No maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 19, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	5:40	End Time (MST):	8:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
16.64	16.64	0.9998						
8.32	8.31	1.0008				Slope	1.000360	0.90 - 1.10
4.15	4.14	1.0020						
			Intercept	-0.005535	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 19, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	5:40	End Time (MST):	8:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
7.82	7.82	0.9996						
3.91	3.90	1.0020				Slope	1.000736	0.90 - 1.10
1.95	1.94	1.0047						
			Intercept	-0.005903	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 19, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	5:40	End Time (MST):	8:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995			
8.82	8.82	1.0000						
4.41	4.41	0.9998				Slope	1.000144	0.90 - 1.10
2.20	2.20	1.0010						
			Intercept	-0.000833	+/-0.5			



NMHC Calibration Plot

Date: August 9, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: August 25, 2023
Start time (MST): 9:39
Reason: Routine
Station number: AMS04
Last Cal Date: July 27, 2023
End time (MST): 14:52

Calibration Standards

NO Gas Cylinder #: T36RH1F
NOX Cal Gas Conc: 51.16 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 51.16 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: August 18, 2023
NO Cal Gas Conc: 50.91 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.91 ppm
NO gas Diff:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 721

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.071	1.121	NO bkgnd or offset:	-0.6	-0.6
NOX coeff or slope:	1.066	1.119	NOX bkgnd or offset:	-0.3	-0.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.5	4.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002547	1.000803
NO _x Cal Offset:	-0.233343	0.426473
NO Cal Slope:	1.001796	1.000459
NO Cal Offset:	-1.333988	-0.993702
NO ₂ Cal Slope:	0.990498	0.991739
NO ₂ Cal Offset:	0.120026	2.285236



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.0	0.2	0.7	----	----
as found span	4922	78.1	799.1	795.2	3.9	759.3	753.6	5.7	1.0524	1.0552
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.6	0.5	0.1	----	----
high point	4922	78.1	799.1	795.2	3.9	800.3	795.3	5.0	0.9985	0.9999
second point	4961	39.1	400.1	398.1	2.0	400.6	396.6	4.0	0.9987	1.0038
third point	4981	19.5	199.5	198.5	1.0	200.0	196.1	3.9	0.9975	1.0124
as left zero	5000	0.0	0.0	0.0	0.0	0.6	0.8	-0.2	----	----
as left span	4922	78.1	799.1	391.7	407.4	793.6	386.1	407.5	1.0069	1.0145
Average Correction Factor									0.9982	1.0054

Corrected As found	NO _x = 758.3 ppb	NO = 753.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -5.6%	
Previous Response	NO _x = 800.9 ppb	NO = 795.3 ppb		*Percent Change	NO = -5.6%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	789.6	386.1	407.4	404.8	1.0064	99.4%
2nd GPT point (200 ppb O3)	789.6	584.7	208.8	211.6	0.9868	101.3%
3rd GPT point (100 ppb O3)	789.6	688.8	104.7	107.6	0.9731	102.8%
Average Correction Factor					0.9888	101.2%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 25, 2023	Previous Calibration:	July 27, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	9:39	End Time (MST):	14:52
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	≥0.995	
799.1	800.3	0.9985			
400.1	400.6	0.9987			
199.5	200.0	0.9975			
			Slope	1.000803	0.90 - 1.10
			Intercept	0.426473	+/-20

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 25, 2023	Previous Calibration:	July 27, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	9:39	End Time (MST):	14:52
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	≥0.995	
795.2	795.3	0.9999			
398.1	396.6	1.0038			
198.5	196.1	1.0124			
			Slope	1.000459	0.90 - 1.10
			Intercept	-0.993702	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 25, 2023	Previous Calibration:	July 27, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	9:39	End Time (MST):	14:52
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
407.4	404.8	1.0064		
208.8	211.6	0.9868		
104.7	107.6	0.9731		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 25, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: August 9, 2023 Last Cal Date: July 20, 2023
 Start time (MST): 8:15 End time (MST): 10:41
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 2961
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996714	0.996114	Backgd or Offset:	-2.6	-2.6
Calibration intercept:	0.200000	0.480000	Coeff or Slope:	1.008	1.008

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	5000	986.3	400.0	398.4	1.004
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.1	----
high point	5000	987.0	400.0	398.7	1.003
second point	5000	817.7	200.0	200.0	1.000
third point	5000	708.4	100.0	100.4	0.996
as left zero	5000	0.0	0.0	0.4	----
as left span	5000	986.8	400.0	400.0	1.000
Average Correction Factor					1.000

Baseline Corr As found:	398.8	Previous response	398.9	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No Maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 20, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:15	End Time (MST):	10:41
Analyzer make:	API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999996	≥0.995
400.0	398.7	1.0033			
200.0	200.0	1.0000	Slope	0.996114	0.90 - 1.10
100.0	100.4	0.9960			
			Intercept	0.480000	+/- 5



O₃ Calibration Plot

Date: August 9, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: August 31, 2023 Last Cal Date: July 11, 2023
 Start time (MST): 6:58 End time (MST): 8:16

Analyzer Make: API T640 S/N: 322
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Deltacal S/N: 1451
 Temp/RH standard: Deltacal S/N: 1451

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	16.7	17	16.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	719.9	721.6	719.9	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.60	4.45	5.01	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: August 31, 2023 Last Cal Date: July 11, 2023
 PM w/o HEPA: 30.6 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.6	10.5	10.5	<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: 33.6 w/ HEPA: 0
 Date Optical Chamber Cleaned: August 31, 2023 <0.2 ug/m3
 Disposable Filter Changed: August 31, 2023

Annual Maintenance

Date Sample Tube Cleaned: May 23, 2023
 Date RH/T Sensor Cleaned: May 23, 2023

Notes: Pump was at 100%. Filter was black. After cleaning pump went down to 35%. Flow at 5. No adjustments done. Leak check passed before and after cleaning.

Calibration by: Melissa Lemay




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS05
MANNIX
AUGUST 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Mannix Station number: AMS05
Calibration Date: August 15, 2023 Last Cal Date: July 11, 2023
Start time (MST): 10:35 End time (MST): 14:49
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.02 ppm Cal Gas Exp Date: January 12, 2029
Cal Gas Cylinder #: XC026809B
Removed Cal Gas Conc: 50.02 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 621
ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 1008841399
Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000771	0.999257	Backgd or Offset:	9.0	8.9
Calibration intercept:	0.040000	0.220000	Coeff or Slope:	0.941	0.930

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4920	80.0	800.3	812.4	0.985
as found 2nd point	4960	40.0	400.2	404.2	0.990
as found 3rd point	4980	20.0	200.1	201.8	0.991
new cylinder response					
calibrator zero	5000	0.0	0.0	0.5	----
high point	4920	80.0	800.3	800.0	1.000
second point	4960	40.0	400.2	400.2	1.000
third point	4980	20.0	200.1	199.7	1.002
as left zero	5000	0.0	0.0	0.5	----
as left span	4920	80.0	800.3	800.4	1.000
Average Correction Factor					1.001

Baseline Corr As found:	812.00	Previous response	800.98	*% change	1.4%
Baseline Corr 2nd AF pt:	403.80	AF Slope:	1.014965	AF Intercept:	-0.680000
Baseline Corr 3rd AF pt:	201.40	AF Correlation:	0.999990		

* = > +/-5% change initiates investigation

Notes: Multi point as founds completed for the NMHC instrument. Changed the inlet filter after as founds.
Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 11, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	10:35	End Time (MST):	14:49
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999999	≥0.995
800.3	800.0	1.0004			
400.2	400.2	0.9999	Slope	0.999257	0.90 - 1.10
200.1	199.7	1.0019			
			Intercept	0.220000	+/-30



SO2 Calibration Plot

Date: August 15, 2023

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mannix Station number: AMS05
 Calibration Date: August 22, 2023 Last Cal Date: July 13, 2023
 Start time (MST): 10:04 End time (MST): 15:01
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.92 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002433
 Removed Cal Gas Conc: 4.92 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1845
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1203169745
 Converter make: Global Converter serial #: 2022-196
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996187	1.000615	Backgd or Offset: 2.13	2.18
Calibration intercept:	0.180584	0.200559	Coeff or Slope: 0.842	0.866

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4919	81.3	80.0	81.9	0.977
as found 2nd point	4960	40.7	40.0	40.5	0.989
as found 3rd point	4980	20.3	20.0	19.9	1.004
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4919	81.3	80.0	80.2	0.997
second point	4960	40.7	40.0	40.4	0.991
third point	4980	20.3	20.0	20.1	0.994
as left zero	5000	0.0	0.0	0.7	----
as left span	4919	81.3	80.0	77.0	1.039
SO2 Scrubber Check	4920	80.0	800.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.994
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 81.9 Prev response: 79.87 *% change: 2.5%
 Baseline Corr 2nd AF pt: 40.5 AF Slope: 1.025466 AF Intercept: -0.319296
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999928

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 22, 2023	Previous Calibration:	July 13, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	10:04	End Time (MST):	15:01
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient	≥0.995
80.0	80.2	0.9974		
40.0	40.4	0.9912	Slope	0.90 - 1.10
20.0	20.1	0.9937		
			Intercept	+/-3



H₂S Calibration Plot

Date: August 22, 2023

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	August 15, 2023	Last Cal Date:	July 11, 2023
Start time (MST):	10:34	End time (MST):	14:49
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.60E-05	2.62E-05	NMHC SP Ratio:	4.34E-05	4.44E-05
CH ₄ Retention time:	15.00	15.00	NMHC Peak Area:	210847	206221
Zero Chromatogram:	ON	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	16.89	1.020
as found 2nd point	4960	40.0	8.61	8.45	1.019
as found 3rd point	4980	20.0	4.31	4.23	1.018
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.22	1.001
second point	4960	40.0	8.61	8.63	0.998
third point	4980	20.0	4.31	4.34	0.993
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.30	0.996
Average Correction Factor					0.997

Baseline Corr AF:	16.89	Prev response	17.22	*% change	-2.0%
Baseline Corr 2nd AF:	8.4	AF Slope:	0.980173	AF Intercept:	0.004800
Baseline Corr 3rd AF:	4.2	AF Correlation:	1.000000	<i>* = +/-5% change initiates investigation</i>	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	9.15	9.01	1.015
as found 2nd point	4960	40.0	4.57	4.53	1.009
as found 3rd point	4980	20.0	2.29	2.28	1.001
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.14	1.001
second point	4960	40.0	4.57	4.61	0.993
third point	4980	20.0	2.29	2.33	0.982
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.15	9.21	0.993
Average Correction Factor					0.992
Baseline Corr AF:	9.01	Prev response	9.13	*% change	-1.4%
Baseline Corr 2nd AF:	4.5	AF Slope:	0.983915	AF Intercept:	0.018800
Baseline Corr 3rd AF:	2.3	AF Correlation:	0.999979	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	7.88	1.025
as found 2nd point	4960	40.0	4.04	3.92	1.031
as found 3rd point	4980	20.0	2.02	1.95	1.037
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.07	1.001
second point	4960	40.0	4.04	4.03	1.004
third point	4980	20.0	2.02	2.01	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.10	0.998
Average Correction Factor					1.003
Baseline Corr AF:	7.88	Prev response	8.09	*% change	-2.6%
Baseline Corr 2nd AF:	3.92	AF Slope:	0.975922	AF Intercept:	-0.013200
Baseline Corr 3rd AF:	1.95	AF Correlation:	0.999986	* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999419	0.998929
THC Cal Offset:	0.005000	0.017200
CH ₄ Cal Slope:	1.001938	0.999731
CH ₄ Cal Offset:	-0.007400	-0.006600
NMHC Cal Slope:	0.997195	0.998220
NMHC Cal Offset:	0.012400	0.023800

Notes: Changed the inlet filter the actuator after as founds. The instrument has been dipping and after an inspection decided to change the actuator. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 11, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:34	End Time (MST):	14:49
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995
17.23	17.22	1.0006			
8.61	8.63	0.9980	Slope	0.998929	0.90 - 1.10
4.31	4.34	0.9932			
			Intercept	0.017200	± 0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 11, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:34	End Time (MST):	14:49
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999996	≥0.995
8.08	8.07	1.0005			
4.04	4.03	1.0035			
2.02	2.01	1.0058			
			Slope	0.999731	0.90 - 1.10
			Intercept	-0.006600	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 11, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:34	End Time (MST):	14:49
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999968	≥ 0.995			
9.15	9.14	1.0006						
4.57	4.61	0.9932				Slope	0.998220	0.90 - 1.10
2.29	2.33	0.9823						
			Intercept	0.023800	± 0.5			



NMHC Calibration Plot

Date: August 15, 2023

Location: Mannix






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	August 17, 2023	Last Cal Date:	July 4, 2023
Start time (MST):	8:44	End time (MST):	14:20
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.78	ppm	Cal Gas Exp Date:	September 9, 2024
Cal Gas Cylinder #:	AAL070632			
Removed Cal Gas Conc:	49.78	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	5608

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006537	1.007062	Backgd or Offset:	17.2
Calibration intercept:	1.758830	-0.099843	Coeff or Slope:	0.901
				17.2
				0.901

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4920	80.3	799.5	794.3	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.3	799.5	805.7	0.992
second point	4960	40.2	400.2	401.0	0.998
third point	4980	20.1	200.1	202.6	0.988
as left zero	5000	0.0	0.0	0.0	----
as left span	4920	80.3	799.5	800.2	0.999
Average Correction Factor					0.993

Baseline Corr As found:	794.40	Previous response	806.45	*% change	-1.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed preventative maintenance on the H2 generator for the THC instrument. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 17, 2023	Previous Calibration:	July 4, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:44	End Time (MST):	14:20
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	≥0.995
799.5	805.7	0.9923		
400.2	401.0	0.9981	Slope	0.90 - 1.10
200.1	202.6	0.9877		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 17, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	August 25, 2023	Last Cal Date:	July 12, 2023
Start time (MST):	8:47	End time (MST):	14:30
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.38	ppm	Cal Gas Exp Date:	March 2, 2023
Cal Gas Cylinder #:	EY0000809			
Removed Cal Gas Conc:	5.38	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004924	0.988760	Backgd or Offset:	2.06	2.06
Calibration intercept:	0.356984	-0.002363	Coeff or Slope:	1.187	1.187

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.3	79.9	76.8	1.041
as found 2nd point	4963	37.2	40.0	38.8	1.032
as found 3rd point	4981	18.6	20.0	20.0	1.001
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4926	74.3	79.9	79.2	1.009
second point	4963	37.2	40.0	39.4	1.016
third point	4981	18.6	20.0	19.5	1.026
as left zero	5000	0.0	0.0	0.2	----
as left span	4926	74.3	79.9	72.8	1.098
SO2 Scrubber Check	4920	80.3	803.0	-0.2	----

Date of last scrubber change:	December 20, 2021	Ave Corr Factor	1.017
Date of last converter efficiency test:		efficiency	

Baseline Corr As found:	76.8	Prev response:	80.69	*% change:	-5.1%
Baseline Corr 2nd AF pt:	38.8	AF Slope:	0.957890	AF Intercept:	0.377982
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999883		

* = > +/-5% change initiates investigation

Notes: Bypassed the converter and ran a 100ppb SO2 point after the as founds, instrument reached the target at a reasonable time. Suspecting that the smoke is having an affect on the scrubber beads again. Installed a hydrator and added little bit of water. Ran a SO2 scrubber test after the calibrator zero. No adjustments made. See docit note for more info.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 25, 2023	Previous Calibration:	July 12, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:47	End Time (MST):	14:30
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999933	≥0.995
79.9	79.2	1.0094			
40.0	39.4	1.0159	Slope	0.988760	0.90 - 1.10
20.0	19.5	1.0264			
			Intercept	-0.002363	+/-3



TRS Calibration Plot

Date: August 25, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	August 28, 2023	Last Cal Date:	August 25, 2023
Start time (MST):	8:25	End time (MST):	12:54
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.38	ppm	Cal Gas Exp Date:	March 2, 2023
Cal Gas Cylinder #:	EY0000809			
Removed Cal Gas Conc:	5.38	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.988760	1.003779	Backgd or Offset:	2.06
Calibration intercept:	-0.002363	0.097067	Coeff or Slope:	1.187
				2.00
				1.158

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.3	79.9	65.8	1.215
as found 2nd point	4963	37.2	40.0	32.7	1.224
as found 3rd point	4981	18.6	20.0	16.0	1.251
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4926	74.3	79.9	80.3	0.996
second point	4963	37.2	40.0	40.3	0.993
third point	4981	18.6	20.0	20.3	0.986
as left zero	5000	0.0	0.0	0.1	----
as left span	4926	74.3	79.9	80.2	0.997
SO2 Scrubber Check	4920	80.3	803.0	0.1	----
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	0.992
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	65.8	Prev response:	79.04	*% change:	-20.1%
Baseline Corr 2nd AF pt:	32.7	AF Slope:	0.824778	AF Intercept:	-0.238684
Baseline Corr 3rd AF pt:	16.0	AF Correlation:	0.999939		

* = > +/-5% change initiates investigation

Notes: Spans are very low due to SO2 scrubbers. After as founds changed the SO2 scrubber beads and installed it vertically outside the converter. Removed the hydrator. Ran a SO2 scrubber test. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 28, 2023	Previous Calibration:	August 25, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:25	End Time (MST):	12:54
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999993	
79.9	80.3	0.9955			≥0.995
40.0	40.3	0.9932	Slope	1.003779	
20.0	20.3	0.9860			0.90 - 1.10
			Intercept	0.097067	+/-3



TRS Calibration Plot

Date: August 28, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	August 17, 2023	Last Cal Date:	July 31, 2023
Start time (MST):	8:44	End time (MST):	14:20
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH ₄ Cal Gas Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Ref.	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.11E-04	2.11E-04	NMHC SP Ratio:	4.72E-05
CH ₄ Retention time:	14.0	14.0	NMHC Peak Area:	192120

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4920	80.3	17.12	17.06	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4920	80.3	17.12	17.04	1.005
second point	4960	40.2	8.57	8.57	1.001
third point	4980	20.1	4.29	4.36	0.984
as left zero	5000	0.0	0.00	0.02	----
as left span	4920	80.3	17.12	17.02	1.006

Average Correction Factor	0.996
---------------------------	-------

Baseline Corr AF:	17.05	Prev response	17.18	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.01	----
as found span	4920	80.3	9.07	9.02	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.01	----
high point	4920	80.3	9.07	9.01	1.006
second point	4960	40.2	4.54	4.55	0.997
third point	4980	20.1	2.27	2.33	0.974
as left zero	5000	0	0.00	0.02	----
as left span	4920	80.3	9.07	9.00	1.007
Average Correction Factor					0.993
Baseline Corr AF:	9.01	Prev response	9.13	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.06	8.04	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.06	8.03	1.003
second point	4960	40.2	4.03	4.01	1.005
third point	4980	20.1	2.02	2.03	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.06	8.02	1.005
Average Correction Factor					1.001
Baseline Corr AF:	8.04	Prev response	8.05	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001550	0.993295
THC Cal Offset:	0.032027	0.049653
CH ₄ Cal Slope:	1.000285	0.996285
CH ₄ Cal Offset:	-0.004951	0.004651
NMHC Cal Slope:	1.002851	0.990348
NMHC Cal Offset:	0.036779	0.045403

Notes: Changed the inlet filter after as founds. Completed preventative maintenance on the H₂ generator and swapped the H₂ cylinder for the Mocon. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary

Version-01-2020


Station Information

Calibration Date:	August 17, 2023	Previous Calibration:	July 31, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:44	End Time (MST):	14:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999975	≥0.995
17.12	17.04	1.0046			
8.57	8.57	1.0005			
4.29	4.36	0.9841			
			Slope	0.993295	0.90 - 1.10
			Intercept	0.049653	+/-0.5

THC Calibration Curve





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-01-2020


Station Information

Calibration Date:	August 17, 2023	Previous Calibration:	July 31, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:44	End Time (MST):	14:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	≥0.995			
8.06	8.03	1.0030						
4.03	4.01	1.0049				Slope	0.996285	0.90 - 1.10
2.02	2.03	0.9953						
			Intercept	0.004651	+/-0.5			

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 17, 2023	Previous Calibration:	July 31, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:44	End Time (MST):	14:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.01	----	Correlation Coefficient	0.999936	≥ 0.995			
9.07	9.01	1.0063						
4.54	4.55	0.9971				Slope	0.990348	0.90 - 1.10
2.27	2.33	0.9741						
			Intercept	0.045403	± 0.5			



NMHC Calibration Plot

Date: August 17, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Patricia McInnes
Calibration Date: August 9, 2023
Start time (MST): 8:16
Reason: Routine
Station number: AMS06
Last Cal Date: July 27, 2023
End time (MST): 12:49

Calibration Standards

NO Gas Cylinder #: T30YCWN
NOX Cal Gas Conc: 47.94 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 47.94 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: April 11, 2025
NO Cal Gas Conc: 46.39 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 46.39 ppm
NO gas Diff:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1172750022

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.815	0.815	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.986	0.986	NOX bkgnd or offset:	3.8	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.2	157.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.993425	0.992318
NO _x Cal Offset:	2.677022	2.277330
NO Cal Slope:	0.996185	0.995327
NO Cal Offset:	1.203660	1.003845
NO ₂ Cal Slope:	0.998218	0.998809
NO ₂ Cal Offset:	-0.045063	-0.587092



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
as found span	4914	86.2	826.5	799.7	26.7	819.6	793.2	26.5	1.0084	1.0082
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.4	0.3	0.0	----	----
high point	4914	86.2	826.5	799.7	26.7	821.4	796.6	24.9	1.0062	1.0039
second point	4957	43.1	413.2	399.9	13.4	413.4	399.5	13.9	0.9996	1.0009
third point	4978	21.6	207.1	200.4	6.7	209.6	201.1	8.5	0.9882	0.9966
as left zero	5000	0.0	0.0	0.0	0.0	0.5	0.5	0.0	----	----
as left span	4914	86.2	826.5	404.9	421.5	818.9	401.3	417.5	1.0092	1.0090
Average Correction Factor									0.9980	1.0005

Corrected As found	NO _x = 819.5 ppb	NO = 793.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.5%	
Previous Response	NO _x = 823.7 ppb	NO = 797.9 ppb		*Percent Change	NO = -0.6%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.8	400.0	421.5	420.8	1.0017	99.8%
2nd GPT point (200 ppb O3)	794.8	600.0	221.5	220.3	1.0055	99.4%
3rd GPT point (100 ppb O3)	794.8	696.5	125.0	123.7	1.0107	98.9%
Average Correction Factor					1.0060	99.4%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:16	End Time (MST):	12:49
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.5	821.4	1.0062		
413.2	413.4	0.9996		
207.1	209.6	0.9882		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:16	End Time (MST):	12:49
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.7	796.6	1.0039		
399.9	399.5	1.0009		
200.4	201.1	0.9966		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 27, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:16	End Time (MST):	12:49
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
421.5	420.8	1.0017		
221.5	220.3	1.0055		
125.0	123.7	1.0107		
			0.999990	
			0.998809	
			-0.587092	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 9, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Patricia McInnes Station number: AMS06
 Calibration Date: August 11, 2023 Last Cal Date: July 14, 2023
 Start time (MST): 9:28 End time (MST): 12:28
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3566
 ZAG Make/Model: API T701 Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1300156234
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001286	1.002629	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	0.500000	0.540000	Coeff or Slope:	1.019	1.019

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	-1.1	----
as found span	5000	1303.0	400.0	401.2	0.997
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	-0.2	----
high point	5000	1303.0	400.0	401.4	0.997
second point	5000	966.5	200.0	200.9	0.996
third point	5000	794.3	100.0	101.9	0.981
as left zero	5000	800.0	0.0	-0.8	----
as left span	5000	1303.0	400.0	403.5	0.991
Average Correction Factor					0.991

Baseline Corr As found:	402.3	Previous response	401.0	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 11, 2023	Previous Calibration:	July 14, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:28	End Time (MST):	12:28
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999979	≥0.995
400.0	401.4	0.9965			
200.0	200.9	0.9955	Slope	1.002629	0.90 - 1.10
100.0	101.9	0.9814			
			Intercept	0.540000	+/- 5



O₃ Calibration Plot

Date: August 11, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: August 24, 2023 Last Cal Date: July 14, 2023
 Start time (MST): 11:58 End time (MST): 13:02

Analyzer Make: API T640 S/N: 766
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT FP-25 S/N: 388755
 Temp/RH standard: ALICAT FP-25 S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	17.7	17.0	17.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	728.9	729.6	728.9	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	4.98	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>August 24, 2023</u>	Last Cal Date: <u>July 14, 2023</u>			
	PM w/o HEPA: <u>31.4</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.0	11.0	11.0	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>21.8</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>August 24, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>August 24, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023
 Date RH/T Sensor Cleaned: April 13, 2023

Notes: Cleaned the chamber and changed the filter. PMT peak was on target. Leak check passed, no adjustments made.

Calibration by: Max Farrell



Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	August 3, 2023	Last Cal Date:	August 2, 2023
Start time (MST):	7:43	End time (MST):	10:53
NH3 Cal Date:	August 3, 2023	Last Cal Date:	August 2, 2023
Start time (MST):	11:15	End time (MST):	14:00
Reason:	Maintenance		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	NA
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	77.8	ppm	NH3 Gas Cylinder #:	CC710812
			NH3 Cal Gas Expiry:	March 30, 2023
Removed NH3 Conc:	77.8	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	API T700		Serial Number:	3566
ZAG make/model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	152
Converter model:	Teledyne API T501	Converter serial #:	147
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.50
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	542

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.794	0.794	TN coefficient:	0.794	0.794
NOX coefficient:	0.796	0.796	NO bkgnd:	-0.1	-0.1
NO2 coefficient:	1.000	1.000	NOX bkgnd:	0.0	0.0
NH3 coefficient:	0.908	0.908	TN bkgnd:	0.0	0.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.012637	1.003439
NO _x Cal Offset:	1.353007	1.455068
NO Cal Slope:	1.012223	0.997870
NO Cal Offset:	1.240145	0.164066
NO ₂ Cal Slope:	1.001251	1.000820
NO ₂ Cal Offset:	0.861558	-0.026033
NH3 Cal Slope:	1.026272	1.002680
NH3 Cal Offset:	3.727263	5.795340
TN Cal Slope:	1.031709	1.008130
TN Cal Offset:	4.015689	5.548239



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.3	0.1	1.2	----	----
as found NO	4914	86.2	826.5	826.5	----	825.9	830.3	-4.3	1.001	----
calibrator zero	5000	0.0	0.0	0.0	0.0	1.3	0.1	1.2	----	----
high NO point	4914	86.2	826.5	826.5	----	825.9	830.3	-4.3	1.001	----
GPT point	4914	86.2	826.5	826.5	----	827.1	825.1	2.2	0.999	----
as found NH3	3419	81.0	1800.5	----	1800.5	1819.4	----	1809.7	0.990	0.995
new NH3 cyl rp							----			
first NH3	3419	81.0	1800.5	----	1800.5	1819.4	----	1809.7	0.990	0.995
second NH3	3455	45.0	1000.3	----	1000.3	1013.1	----	1007.9	0.987	0.992
third NH3	3478	22.5	500.1	----	500.1	516.1	----	514.1	0.969	0.973
Average Correction Factor									0.9999	0.9867

Corrected As found TN = 824.6 ppb NO_x = 830.2 ppb NH3 = 1808.5 ppb

Previous Response TN = 856.7 ppb NO_x = 838.2 ppb NH3 = 1851.5 ppb

NH3 Previous Converter Efficiency = 90.8%

NH3 Current Converter Efficiency = 90.8%

*Percent Change TN = -3.9%

*Percent Change NO_x = -1.0%

*Percent Change NH3 = -2.4%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.8	1.3	----	----
as found span	4914	86.2	826.5	799.7	826.5	830.3	798.4	825.9	0.9954	1.0017
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.8	1.3	----	----
high point	4914	86.2	826.5	799.7	826.5	830.3	798.4	825.9	0.9954	1.0017
second point	4957	43.1	413.2	399.9	413.2	416.1	399.2	414.5	0.9931	1.0017
third point	4978	21.6	207.1	200.4	207.1	211.1	199.3	210.9	0.9811	1.0056
Average Correction Factor									0.9899	1.0030

Baseline Corr As fnd	TN = 824.6 ppb	NO _x = 830.2 ppb	NO = 797.6 ppb		*Percent Change	TN = -3.9%
Previous Response	TN = 856.7 ppb	NO _x = 838.2 ppb	NO = 810.7 ppb		*Percent Change	NO _x = -1.0%
					*Percent Change	NO = -1.6%
<i>* = > +/-5% change initiates investigation</i>						

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05 As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	-0.8	----	----
calibration zero	----	----	0.0	-0.8	----	----
1st GPT point (400 ppb O3)	799.0	396.9	428.8	428.3	1.0012	99.9%
2nd GPT point (200 ppb O3)	799.0	601.8	223.9	225.7	0.9921	100.8%
3rd GPT point (100 ppb O3)	799.0	696.0	129.7	129.8	0.9994	100.1%
Average Correction Factor					0.9976	100.2%

Notes: Multipoint as founds completed yesterday and the pump has been swapped out for preventative maintenance. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TN Calibration Summary

Version-05-2023


Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	August 2, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	11:15	End Time (MST):	14:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	1.3	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
1800.5	1819.4	0.9896		
1000.3	1013.1	0.9874		
500.1	516.1	0.9689		

TN Calibration Curve





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	August 2, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	11:15	End Time (MST):	14:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	1.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.5	1809.7	0.9949		
1000.3	1007.9	0.9924		
500.1	514.1	0.9727		





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	August 2, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:43	End Time (MST):	10:53
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
826.5	830.3	0.9954			
413.2	416.1	0.9931			
207.1	211.1	0.9811			
			Slope	1.003439	0.90 - 1.10
			Intercept	1.455068	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	August 2, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:43	End Time (MST):	10:53
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.8	----	Correlation Coefficient	≥0.995	
799.7	798.4	1.0017			
399.9	399.2	1.0017			
200.4	199.3	1.0056			
			Slope	0.997870	0.90 - 1.10
			Intercept	0.164066	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	August 2, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:43	End Time (MST):	10:53
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.8	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
428.8	428.3	1.0012		
223.9	225.7	0.9921		
129.7	129.8	0.9994		
			0.999960	
			1.000820	
			-0.026033	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 3, 2023


Location: Patricia McInnes



NH₃ Calibration Plot

Date: August 3, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Patricia McInnes	Station Number:	AMS 06
Calibration Date:	August 25, 2023	Prev Cal Date:	August 30, 2022
Start Time (MST):	13:40	End Time (MST):	14:25
Tower Height (m):	10.0	Reason:	Routine

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	B10015
WS Calibrator:	Young 18802	Serial Number:	CA 03845

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.5	-0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999998	≥0.9995
Calculated slope		0.999113	0.90 - 1.10
Calculated intercept		0.030337	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	E4854
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):		Calc Declination*:	Degrees
Deadband calc:	2.6 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	1.9	---
90	87.2	-0.8%
180	177.9	-0.6%
270	269.6	-0.1%
357	356.3	-0.2%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999922	≥0.9995
Calculated slope		1.003011	0.90 - 1.10
Calculated intercept		0.282225	+/- 4

Notes: Bearing are in good condition, checked crossarm alignment using a compass.

Calibration Performed By: Max Farrell




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	August 9, 2023	Last Cal Date:	July 6, 2023
Start time (MST):	8:38	End time (MST):	12:38
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.52	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC282115			
Removed Cal Gas Conc:	50.52	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 701H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995398	0.997338	Backgd or Offset:	2.61	2.61
Calibration intercept:	2.183198	1.503428	Coeff or Slope:	0.834	0.834

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	79.3	801.2	797.2	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	79.3	801.2	799.4	1.002
second point	4960	39.6	400.2	402.8	0.993
third point	4980	19.8	200.1	201.3	0.994
as left zero	5000	0.0	0.0	0.0	----
as left span	4921	79.2	800.2	798.7	1.002
Average Correction Factor					0.997

Baseline Corr As found:	797.30	Previous response	799.70	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 6, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:38	End Time (MST):	12:38
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999977	≥0.995
801.2	799.4	1.0023			
400.2	402.8	0.9934	Slope	0.997338	0.90 - 1.10
200.1	201.3	0.9939			
			Intercept	1.503428	+/-30



SO2 Calibration Plot

Date: August 9, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: August 2, 2023 Last Cal Date: July 11, 2023
 Start time (MST): 9:00 End time (MST): 14:06
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.25 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC504080
 Removed Cal Gas Conc: 5.25 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3805
 ZAG Make/Model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE Analyzer serial #: 1180540018
 Converter make: CDN-101 Converter serial #: 551
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008281	1.005977	Backgd or Offset: 2.33	2.33
Calibration intercept:	-0.182074	-0.002158	Coeff or Slope: 0.885	0.899

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4925	75.5	79.3	78.5	1.008
as found 2nd point	4962	37.7	39.6	38.9	1.015
as found 3rd point	4981	18.9	19.8	19.2	1.028
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4925	75.5	79.3	79.8	0.994
second point	4962	37.7	39.6	39.9	0.993
third point	4981	18.9	19.9	19.8	1.003
as left zero	5000	0.0	0.0	0.2	----
as left span	4925	75.5	79.3	79.5	0.998
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	0.996
Date of last converter efficiency test:	April 22, 2022			92.6% efficiency	

Baseline Corr As found: 78.6 Prev response: 79.74 *% change: -1.5%
 Baseline Corr 2nd AF pt: 39.0 AF Slope: 0.992855 AF Intercept: -0.302193
 Baseline Corr 3rd AF pt: 19.3 AF Correlation: 0.999970

* = > +/-5% change initiates investigation

Notes: Scrubber check passed. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 2, 2023	Previous Calibration:	July 11, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:00	End Time (MST):	14:06
Analyzer make:	CDN-101	Analyzer serial #:	551

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999987	
79.3	79.8	0.9939			≥0.995
39.6	39.9	0.9927	Slope	1.005977	
19.9	19.8	1.0029			0.90 - 1.10
			Intercept	-0.002158	+/-3



TRS Calibration Plot

Date: August 2, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	August 9, 2023	Last Cal Date:	July 21, 2023
Start time (MST):	8:38	End time (MST):	12:39
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	0.000263	0.000256	NMHC SP Ratio: 4.97E-05	4.57E-05
CH ₄ Retention time:	14.4	14.2	NMHC Peak Area:	182990 198820

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	18.04	0.944
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	16.98	1.003
second point	4960	39.6	8.52	8.51	1.000
third point	4980	19.8	4.26	4.28	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.01	1.001

Average Correction Factor				0.999
Baseline Corr AF:	18.04	Prev response	16.89	*% change 6.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.09	9.91	0.918
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	9.09	9.07	1.002
second point	4960	39.6	4.55	4.56	0.998
third point	4980	19.8	2.27	2.31	0.986
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	9.09	9.08	1.001
Average Correction Factor					0.995
Baseline Corr AF:	9.91	Prev response	8.96	*% change	9.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.94	8.13	0.976
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.94	7.91	1.004
second point	4960	39.6	3.97	3.96	1.003
third point	4980	19.8	1.98	1.98	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.94	7.93	1.002
Average Correction Factor					1.004
Baseline Corr AF:	8.13	Prev response	7.93	*% change	2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.993293	0.996468
THC Cal Offset:	-0.024935	0.018660
CH ₄ Cal Slope:	0.998509	0.996465
CH ₄ Cal Offset:	0.000842	-0.000558
NMHC Cal Slope:	0.988725	0.996357
NMHC Cal Offset:	-0.025977	0.019418

Notes: Dips present during night z/s cycle. As founds 9% high. Chromatograms look normal. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary

Version-01-2020


Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 21, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:38	End Time (MST):	12:39
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995
17.03	16.98	1.0030			
8.52	8.51	1.0004			
4.26	4.28	0.9945			
			Slope	0.996468	0.90 - 1.10
			Intercept	0.018660	+/-0.5

THC Calibration Curve





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-01-2020


Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 21, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:38	End Time (MST):	12:39
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	≥0.995
7.94	7.91	1.0036			
3.97	3.96	1.0035			
1.98	1.98	1.0045			
			Slope	0.996465	0.90 - 1.10
			Intercept	-0.000558	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 21, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:38	End Time (MST):	12:39
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999979	≥ 0.995
9.09	9.07	1.0025			
4.55	4.56	0.9978			
2.27	2.31	0.9860			
			Slope	0.996357	0.90 - 1.10
			Intercept	0.019418	+/-0.5



NMHC Calibration Plot

Date: August 9, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	August 29, 2023	Last Cal Date:	August 9, 2023
Start time (MST):	9:22	End time (MST):	10:53
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	0.000256	0.000256	NMHC SP Ratio:	4.57E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	198820
				198820

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.70	0.962
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	0.962
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	9.09	9.66	0.941
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.941
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00		----
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.94	8.05	0.986
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.986
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	N/A	N/A
THC Cal Offset:	N/A	N/A
CH ₄ Cal Slope:	N/A	N/A
CH ₄ Cal Offset:	N/A	N/A
NMHC Cal Slope:	N/A	N/A
NMHC Cal Offset:	N/A	N/A


Notes: No as founds done due to both N2 and H2 being low. N2 and H2 cylinders replaced.

Calibration Performed By: Aswin Sasi Kumar

NMHC Calibration Plot

Date: August 29, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	August 14, 2023	Last Cal Date:	July 13, 2023
Start time (MST):	9:48	End time (MST):	14:55
Reason:	Routine		

Calibration Standards

NO Gas Cylinder #:	T2Y1KA4	Cal Gas Expiry Date:	November 30, 2023
NOX Cal Gas Conc:	50.92 ppm	NO Cal Gas Conc:	49.92 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	50.92 ppm	Removed Gas NO Conc:	49.92 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API T701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024		
NOX Range (ppb):	0 - 1000 ppb				
	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>	
NO coeff or slope:	1.065	1.080	NO bkgnd or offset:	7.5	7.6
NOX coeff or slope:	0.995	0.994	NOX bkgnd or offset:	7.7	7.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	207.8	211.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001112	1.002539
NO _x Cal Offset:	1.259152	0.999372
NO Cal Slope:	1.001431	1.002515
NO Cal Offset:	0.895212	1.115453
NO ₂ Cal Slope:	1.003989	0.998415
NO ₂ Cal Offset:	0.417819	0.146993



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as found span	4920	80.2	816.7	800.7	16.0	808.2	789.9	18.3	1.0105	1.0137
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
high point	4920	80.2	816.7	800.7	16.0	819.0	803.0	16.0	0.9972	0.9971
second point	4960	40.1	408.4	400.4	8.0	411.9	404.0	7.9	0.9914	0.9910
third point	4980	20.0	203.7	199.7	4.0	205.4	201.5	3.9	0.9916	0.9910
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
as left span	4920	80.2	816.7	393.3	423.4	825.2	398.8	426.4	0.9897	0.9862
Average Correction Factor									0.9934	0.9930

Corrected As found	NO _x = 808.2 ppb	NO = 789.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.3%	
Previous Response	NO _x = 818.9 ppb	NO = 802.7 ppb		*Percent Change	NO = -1.6%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	802.2	394.8	423.4	422.7	1.0017	99.8%
2nd GPT point (200 ppb O ₃)	802.2	599.1	219.1	219.3	0.9993	100.1%
3rd GPT point (100 ppb O ₃)	802.2	700.1	118.1	118.2	0.9995	100.1%
Average Correction Factor					1.0002	100.0%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	August 14, 2023	Previous Calibration:	July 13, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:48	End Time (MST):	14:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
816.7	819.0	0.9972			
408.4	411.9	0.9914			
203.7	205.4	0.9916			
			Slope	1.002539	0.90 - 1.10
			Intercept	0.999372	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 14, 2023	Previous Calibration:	July 13, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:48	End Time (MST):	14:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999989	≥0.995
800.7	803.0	0.9971			
400.4	404.0	0.9910	Slope	1.002515	0.90 - 1.10
199.7	201.5	0.9910			
			Intercept	1.115453	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 14, 2023	Previous Calibration:	July 13, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:48	End Time (MST):	14:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
423.4	422.7	1.0017		
219.1	219.3	0.9993		
118.1	118.2	0.9995		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 14, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: August 11, 2023 Last Cal Date: July 7, 2023
 Start time (MST): 9:45 End time (MST): 13:27
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: T700 Serial Number: 3805
 ZAG Make/Model: T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1152220023
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002371	0.999171	Backgd or Offset:	-2.6	-2.6
Calibration intercept:	0.860000	0.320000	Coeff or Slope:	1.522	1.522

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-1.3	----
as found span	5000	1414.8	400.0	398.7	1.003
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.6	----
high point	5000	1415.7	400.0	399.4	1.002
second point	5000	1039.9	200.0	201.0	0.995
third point	5000	856.2	100.0	100.9	0.991
as left zero	5000	0.0	0.0	-0.5	----
as left span	5000	1416.0	400.0	403.9	0.990
Average Correction Factor					0.996

Baseline Corr As found:	400.0	Previous response	401.8	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 11, 2023	Previous Calibration:	July 7, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:45	End Time (MST):	13:27
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.6	----	Correlation Coefficient	0.999973	≥0.995
400.0	399.4	1.0015			
200.0	201.0	0.9950	Slope	0.999171	0.90 - 1.10
100.0	100.9	0.9911			
			Intercept	0.320000	+/- 5



O₃ Calibration Plot

Date: August 11, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Athabasca Valley Station number: AMS 07
 Calibration Date: August 28, 2023 Last Cal Date: July 24, 2023
 Start time (MST): 13:25 End time (MST): 15:25

Analyzer Make: API T640 S/N: 871
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	27.8	26.65	27.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	735.1	737.74	735.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.10	5.07	5.10	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: August 28, 2023	Last Cal Date: July 24, 2023			
	PM w/o HEPA: 57.8	PM w/ HEPA: 0.0			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.5	10.5	10.5	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: 57.8	w/ HEPA: 0.0		
Date Optical Chamber Cleaned:		August 28, 2023			<0.2 ug/m3
Disposable Filter Changed:		August 28, 2023			

Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022
 Date RH/T Sensor Cleaned: December 5, 2022

Notes: Temp, flow and pressure checked. Leak check passed. Chamber cleaned, filter swapped. No issues.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	August 16, 2023	Last Cal Date:	July 24, 2023
Start time (MST):	9:17	End time (MST):	12:35
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date:	December 12, 2026
Cal Gas Cylinder #:	LL66942			
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 700H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000886	0.999458	Backgd or Offset:	4.137	4.225
Calibration intercept:	0.076548	0.036542	Coeff or Slope:	1.086	1.086

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.0	40.2	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.0	40.0	1.001
second point	4967	33.3	20.0	20.1	0.994
third point	4983	16.7	10.0	10.1	0.997
as left zero	5000	0.0	0.0	0.0	----
as left span	4933	66.7	40.0	39.9	1.003
Average Correction Factor					0.997

Baseline Corr As found:	40.08	Prev response:	40.13	*% change:	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Zero slightly adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 16, 2023	Previous Calibration:	July 24, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:17	End Time (MST):	12:35
Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999985	≥ 0.995
40.0	40.0	1.0008	Slope	0.999458	0.90 - 1.10
20.0	20.1	0.9940	Intercept	0.036542	± 1.5
10.0	10.1	0.9971			



CO Calibration Plot

Date: August 16, 2023

Location: Athabasca Valley






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	August 14, 2023	Last Cal Date:	July 14, 2023
Start time (MST):	11:11	End time (MST):	13:46
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC196697			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3060
ZAG Make/Model:	Teledyne API T701		Serial Number:	260

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003156	1.000314	Backgd or Offset:	1.53	4.65
Calibration intercept:	-0.363935	-0.543305	Coeff or Slope:	0.915	0.963

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	2.1	----
as found span	4920	80.3	800.4	758.9	1.055
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	80.3	800.4	801.0	0.999
second point	4960	40.2	400.7	398.0	1.007
third point	4980	20.1	200.4	200.8	0.998
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	80.3	800.4	802.2	0.998
Average Correction Factor					1.001

Baseline Corr As found:	756.80	Previous response	802.54	*% change	-6.0%
-------------------------	--------	-------------------	--------	-----------	-------

* = > +/-5% change initiates investigation

Notes: changed inlet filters after as founds. Adjusted zero and span

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 14, 2023	Previous Calibration:	July 14, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	11:11	End Time (MST):	13:46
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥0.995
800.4	801.0	0.9992		
400.7	398.0	1.0068	Slope	0.90 - 1.10
200.4	200.8	0.9978		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 14, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: August 24, 2023 Last Cal Date: July 14, 2023
 Start time (MST): 12:26 End time (MST): 15:49
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002276
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 260

Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 1203169744
 Converter make: CDN-101 Converter serial #: 14639
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011142	0.999424	Backgd or Offset: 0.94	0.99
Calibration intercept:	0.198873	0.458863	Coeff or Slope: 0.707	0.752

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4920	80.5	80.0	75.0	1.071
as found 2nd point	4960	40.2	40.0	37.5	1.074
as found 3rd point	4980	20.1	20.0	19.0	1.068
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4920	80.5	80.0	80.3	0.996
second point	4960	40.2	40.0	40.6	0.984
third point	4980	20.1	20.0	20.5	0.975
as left zero	5000	0.0	0.0	0.5	----
as left span	4920	80.5	80.0	80.8	0.990
SO2 Scrubber Check	4919.7	80.3	803.0	0.2	----
Date of last scrubber change:	March 7, 2022			Ave Corr Factor	0.985
Date of last converter efficiency test:	March 15, 2022			100.7% efficiency	

Baseline Corr As found: 74.7 Prev response: 81.10 *% change: -8.6%
 Baseline Corr 2nd AF pt: 37.2 AF Slope: 0.933294 AF Intercept: 0.297479
 Baseline Corr 3rd AF pt: 18.7 AF Correlation: 0.999996

* = > +/-5% change initiates investigation

Notes: Sampled inlet filter changed after as founds. Scrubber check passed. Adjusted span.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 24, 2023	Previous Calibration:	July 14, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	12:26	End Time (MST):	15:49
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999975	≥0.995
80.0	80.3	0.9964			
40.0	40.6	0.9842	Slope	0.999424	0.90 - 1.10
20.0	20.5	0.9746			
			Intercept	0.458863	+/-3



TRS Calibration Plot

Date: August 24, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Chipewyan
Calibration Date: August 28, 2023
Start time (MST): 12:13
Reason: As Found
Station number: AMS08
Last Cal Date: July 13, 2023
End time (MST): 16:00

Calibration Standards

NO Gas Cylinder #: CC363447
NOX Cal Gas Conc: 48.80 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.80 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: February 2, 2024
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3060
Serial Number: 260

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1426262592

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.933	1.020	NO bkgnd or offset:	8.1	9.2
NOX coeff or slope:	0.993	0.992	NOX bkgnd or offset:	8.2	10.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	268.1	146.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.991475	
NO _x Cal Offset:	0.520000	
NO Cal Slope:	0.992674	
NO Cal Offset:	-0.500000	
NO ₂ Cal Slope:	0.995889	
NO ₂ Cal Offset:	0.340396	



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.2	0.4	-0.2	----	----
as found span	4918	82.0	800.3	800.3	0.0	734.6	731.5	3.1	1.0895	1.0941
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

Average Correction Factor

Corrected As found	NO _x = 734.4 ppb	NO = 731.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -8.1%
Previous Response	NO _x = 794.0 ppb	NO = 794.0 ppb		*Percent Change	NO = -8.6%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)						
2nd GPT point (200 ppb O3)						
3rd GPT point (100 ppb O3)						

Average Correction Factor


Notes: Sample inlet filter changed after as founds, sample flow was low so pump was changed. Span adjustment limited, adjusted HVPS to reach optimal setpoint. Calibration to be completed on August 29th.

Calibration Performed By: Matthew C

NO_x Calibration Plot

Date: August 28, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
Calibration Date: August 29, 2023 Last Cal Date: July 13, 2023
Start time (MST): 15:05 End time (MST): 19:15
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC363447 Cal Gas Expiry Date: February 2, 2024
NOX Cal Gas Conc: 48.80 ppm NO Cal Gas Conc: 48.80 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.80 ppm Removed Gas NO Conc: 48.80 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3060
ZAG make/model: Teledyne API T701H Serial Number: 260

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1426262592
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.933	1.020	NO bkgnd or offset:	8.1	9.2
NOX coeff or slope:	0.993	0.992	NOX bkgnd or offset:	8.2	10.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	268.1	268.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.991475	0.996773
NO _x Cal Offset:	0.520000	0.340000
NO Cal Slope:	0.992674	0.997515
NO Cal Offset:	-0.500000	-0.520000
NO ₂ Cal Slope:	0.995889	1.009480
NO ₂ Cal Offset:	0.340396	1.358955



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.2	-0.4	----	----
as found span	4918	82.0	800.3	800.3	0.0	804.2	800.5	2.8	0.9952	0.9998
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.2	-0.2	----	----
high point	4918	82.0	800.3	800.3	0.0	797.2	797.4	-0.2	1.0039	1.0037
second point	4959	41.0	400.2	400.2	0.0	401.1	400.2	0.9	0.9977	0.9999
third point	4980	20.5	200.1	200.1	0.0	199.5	197.6	1.9	1.0029	1.0126
as left zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.2	-0.4	----	----
as left span	4918	82.0	800.3	410.7	389.6	793.4	403.9	389.5	1.0087	1.0169
Average Correction Factor									1.0015	1.0054

Corrected As found	NO _x = 804.8 ppb	NO = 800.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.3%
Previous Response	NO _x = 794.0 ppb	NO = 794.0 ppb		*Percent Change	NO = 0.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	791.5	401.9	389.6	393.5	0.9901	101.0%
2nd GPT point (200 ppb O3)	791.5	601.8	189.7	194.9	0.9733	102.7%
3rd GPT point (100 ppb O3)	791.5	699.5	92.0	94.9	0.9694	103.2%
Average Correction Factor					0.9776	102.3%

Notes: sampled inlet filter changed after as founds. No adjustments made. Triggered gen auto NO and by mistake during GPT point.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	August 29, 2023	Previous Calibration:	July 13, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	15:05	End Time (MST):	19:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.4	----	Correlation Coefficient	0.999986	≥0.995
800.3	797.2	1.0039			
400.2	401.1	0.9977	Slope	0.996773	0.90 - 1.10
200.1	199.5	1.0029			
			Intercept	0.340000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 29, 2023	Previous Calibration:	July 13, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	15:05	End Time (MST):	19:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	797.4	1.0037		
400.2	400.2	0.9999		
200.1	197.6	1.0126		
			0.999986	
			0.997515	
			-0.520000	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 29, 2023	Previous Calibration:	July 13, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	15:05	End Time (MST):	19:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
389.6	393.5	0.9901		
189.7	194.9	0.9733		
92.0	94.9	0.9694		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 29, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: August 14, 2023 Last Cal Date: July 10, 2023
 Start time (MST): 8:32 End time (MST): 11:14
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 260

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3872
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010943	1.018771	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	-1.940000	-0.760000	Coeff or Slope:	1.036	1.036

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	NA	0.0	0.3	----
as found span	5000	913.0	400.0	403.4	0.992
as found 2nd point					
as found 3rd point					
calibrator zero	5000	NA	0.0	0.3	----
high point	5000	914.7	400.0	407.3	0.982
second point	5000	786.4	200.0	202.4	0.988
third point	5000	701.3	100.0	100.1	0.999
as left zero	5000	NA	0.0	0.3	----
as left span	5000	963.3	400.0	407.5	0.982
Average Correction Factor					0.990

Baseline Corr As found:	403.1	Previous response	402.4	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed out inlet filter after as found. adjusted span

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 14, 2023	Previous Calibration:	July 10, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:32	End Time (MST):	11:14
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999969	≥0.995
400.0	407.3	0.9821			
200.0	202.4	0.9881	Slope	1.018771	0.90 - 1.10
100.0	100.1	0.9990			
			Intercept	-0.760000	+/- 5



O₃ Calibration Plot

Date: August 14, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
 Calibration Date: August 24, 2023 Last Cal Date: July 25, 2023
 Start time (MST): 14:05 End time (MST): 14:34

Analyzer Make: Teledyne API T640 S/N: 216
 Particulate Fraction: PM2.5

Flow Meter Make/Model: FP-25 S/N: 388747
 Temp/RH standard: FP-25 S/N: 388747

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	18.9	19.1	19.0	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	735.6	728.3	740.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.00	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>July 25, 2023</u>	Last Cal Date: <u>June 29, 2023</u>			
	PM w/o HEPA: <u>18.8</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.1	11.0	11.2	<input type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>24.5</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>July 25, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 25, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: July 25, 2023
 Date RH/T Sensor Cleaned: July 25, 2023

Notes: No adjustment made.

Calibration by: Morgan Voyageur



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	August 30, 2023	Last Cal Date:	July 25, 2023
Start time (MST):	16:10	End time (MST):	18:45
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API T701H		Serial Number:	260

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	Start	Finish		Start	Finish
Calibration slope:	0.996203	0.995357	Backgd or Offset:	-0.014	-0.013
Calibration intercept:	0.030984	0.050947	Coeff or Slope:	0.968	0.999

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.00	----
as found span	4933	66.7	40.4	40.3	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4934	66.7	40.4	40.2	1.004
second point	4967	33.3	20.2	20.2	0.997
third point	4983	16.7	10.1	10.1	1.005
as left zero	5000	0.0	0.0	0.1	----
as left span	2960	40.0	40.4	40.1	1.007
Average Correction Factor					1.002

Baseline Corr As found:	40.25	Prev response:	40.30	*% change:	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustments made.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 30, 2023	Previous Calibration:	July 25, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	16:10	End Time (MST):	18:45
Analyzer make:	API T300	Analyzer serial #:	3505

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	0.999983	
40.4	40.2	1.0043			<i>≥0.995</i>
20.2	20.2	0.9970	Slope	0.995357	
10.1	10.1	1.0050			<i>0.90 - 1.10</i>
			Intercept	0.050947	<i>+/-1.5</i>



CO Calibration Plot

Date: August 30, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	August 28, 2023	Last Cal Date:	July 26, 2023
Start time (MST):	15:57	End time (MST):	19:09
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
N2 Gen Make/Model:	NG 5000		Serial Number:	771048318

Analyzer Information

Analyzer make: Teledyne API T360 Analyzer serial #: 289
 Analyzer Range 0 - 2,000 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997030	0.997557	Backgd or Offset:	0.008	0.008
Calibration intercept:	-4.080000	-1.000000	Coeff or Slope:	1.016	1.019

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	0.5	----
as found span	2920	80.0	1605.9	1584.9	1.013
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	1.4	----
high point	2920	80.0	1605.9	1606.4	1.000
second point	2960	40.0	802.9	786.0	1.022
third point	2980	20.0	401.5	405.6	0.990
as left zero	3000	0.0	0.0	1.9	----
as left span	2960	40.0	802.9	784.5	1.023
Average Correction Factor					1.004

Baseline Corr As found:	1584.40	Prev response:	1597.02	*% change:	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: As founds done and changed inlet filter. Made adjustment to span point but changed to early.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	August 28, 2023	Previous Calibration	July 26, 2023
Station Name	Fort Chipewyan	Station Number	AMS08
Start Time (MST)	15:57	End Time (MST)	19:06
Analyzer make	Teledyne API T360	Analyzer serial #	289

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.4	----	Correlation Coefficient	0.999809	≥0.995
1605.9	1606.4	0.9997			
802.9	786.0	1.0215	Slope	0.997557	0.90 - 1.10
401.5	405.6	0.9898			
			Intercept	-1.000000	+/-20



CO₂ Calibration Plot

Date: August 28, 2023

Location: Fort Chipewyan






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	August 15, 2023	Last Cal Date:	July 11, 2023
Start time (MST):	8:31	End time (MST):	11:55
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.96	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC151285			
Removed Cal Gas Conc:	49.96	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3812
ZAG Make/Model:	API T701		Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000739	0.998231	Backgd or Offset:	9.6	9.7
Calibration intercept:	0.671341	0.050728	Coeff or Slope:	0.986	0.975

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4919	80.2	801.5	808.5	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.5	----
high point	4919	80.2	801.5	800.6	1.001
second point	4959	40.1	400.8	399.1	1.004
third point	4980	20.0	199.8	199.6	1.001
as left zero	5000	0.0	0.0	0.3	----
as left span	4919	80.2	801.5	803.4	0.998
Average Correction Factor					1.002

Baseline Corr As found:	808.20	Previous response	802.75	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 11, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:31	End Time (MST):	11:55
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999996	≥0.995
801.5	800.6	1.0011			
400.8	399.1	1.0041	Slope	0.998231	0.90 - 1.10
199.8	199.6	1.0012			
			Intercept	0.050728	+/-30



SO2 Calibration Plot

Date: August 15, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Barge Landing Station number: AMS09
 Calibration Date: August 8, 2023 Last Cal Date: July 4, 2023
 Start time (MST): 9:30 End time (MST): 13:40
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.87 ppm Cal Gas Exp Date: September 2, 2024
 Cal Gas Cylinder #: EY0002346
 Removed Cal Gas Conc: 4.87 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3812
 ZAG Make/Model: API T701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1331259320
 Converter make: CDN-101 Converter serial #: 519
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000859	1.007720	Backgd or Offset: 2.77	2.77
Calibration intercept:	-0.220903	-0.060968	Coeff or Slope: 1.134	1.134

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4918	82.1	80.0	79.6	1.003
as found 2nd point	4959	41.1	40.0	39.8	1.003
as found 3rd point	4979	20.5	20.0	19.7	1.009
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4918	82.1	80.0	80.5	0.993
second point	4959	41.1	40.0	40.3	0.993
third point	4979	20.5	20.0	20.1	0.993
as left zero	5000	0.0	0.0	0.0	----
as left span	4918	82.1	80.0	80.9	0.988
SO2 Scrubber Check	4920	80.2	802.0	-0.1	----
Date of last scrubber change:	28-Feb-23			Ave Corr Factor	0.993
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.7 Prev response: 79.81 *% change: -0.1%
 Baseline Corr 2nd AF pt: 39.9 AF Slope: 0.997146 AF Intercept: -0.140983
 Baseline Corr 3rd AF pt: 19.8 AF Correlation: 0.999998

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SOx scrubber check done after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 8, 2023	Previous Calibration:	July 4, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:30	End Time (MST):	13:40
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.0	-0.1	----	Correlation Coefficient	0.999999	≥0.995			
80.0	80.5	0.9933						
40.0	40.3	0.9933				Slope	1.007720	0.90 - 1.10
20.0	20.1	0.9935				Intercept	-0.060968	+/-3



TRS Calibration Plot

Date: August 8, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	August 15, 2023	Last Cal Date:	July 11, 2023
Start time (MST):	8:31	End time (MST):	11:55
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 100 ppm		
NMHC Range (ppm):	0 - 50 ppm	CH ₄ Range (ppm):	0 - 50 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.52E-04	2.51E-04	NMHC SP Ratio:	4.85E-05
CH ₄ Retention time:	15.4	15.2	NMHC Peak Area:	188466
				187816

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	17.15	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.14	0.999
second point	4960	40.1	8.56	8.55	1.001
third point	4980	20.0	4.27	4.29	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.15	0.998

Average Correction Factor				0.999
Baseline Corr AF:	17.15	Prev response	17.16	*% change -0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.0	----
as found span	4919	80.2	9.14	9.13	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.14	1.000
second point	4960	40.1	4.57	4.57	1.000
third point	4980	20	2.28	2.30	0.991
as left zero	5000	0	0.00	0.00	----
as left span	4919	80.2	9.14	9.14	1.000
Average Correction Factor					0.997
Baseline Corr AF:	9.13	Prev response	9.16	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	8.02	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	8.00	0.998
second point	4960	40.1	3.99	3.98	1.003
third point	4980	20.0	1.99	1.99	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	8.02	0.996
Average Correction Factor					1.001
Baseline Corr AF:	8.02	Prev response	8.00	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002740	1.000663
THC Cal Offset:	-0.004535	0.001260
CH ₄ Cal Slope:	1.002125	1.002197
CH ₄ Cal Offset:	-0.000132	-0.007135
NMHC Cal Slope:	1.003064	0.999486
NMHC Cal Offset:	-0.003803	0.007995

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 11, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:31	End Time (MST):	11:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
17.12	17.14	0.9989						
8.56	8.55	1.0014				Slope	1.000663	0.90 - 1.10
4.27	4.29	0.9957						
			Intercept	0.001260	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 11, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:31	End Time (MST):	11:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	≥ 0.995			
7.98	8.00	0.9978						
3.99	3.98	1.0029				Slope	1.002197	0.90 - 1.10
1.99	1.99	1.0017						
			Intercept	-0.007135	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 11, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:31	End Time (MST):	11:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995
9.14	9.14	0.9997			
4.57	4.57	0.9999			
2.28	2.30	0.9909			
			Slope	0.999486	0.90 - 1.10
			Intercept	0.007995	+/-0.5



NMHC Calibration Plot

Date: August 15, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Barge Landing
Calibration Date: August 23, 2023
Start time (MST): 9:03
Reason: Routine
Station number: AMS09
Last Cal Date: July 5, 2023
End time (MST): 13:42

Calibration Standards

NO Gas Cylinder #: DT0036634
NOX Cal Gas Conc: 50.00 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 50.00 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: January 28, 2024
NO Cal Gas Conc: 49.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 49.70 ppm
NO gas Diff:
Serial Number: 3812
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1426262593

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.169	1.157	NO bkgnd or offset:	10.7	10.6
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	10.9	10.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	175.6	174.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997858	0.999930
NO _x Cal Offset:	0.708793	0.929269
NO Cal Slope:	0.999541	1.001426
NO Cal Offset:	-0.472540	-0.332141
NO ₂ Cal Slope:	1.000792	1.000531
NO ₂ Cal Offset:	-0.264400	0.173936



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.2	0.2	----	----
as found span	4919	80.5	805.1	800.3	4.8	816.2	809.5	6.6	0.986	0.989
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
high point	4919	80.5	805.1	800.3	4.8	805.4	801.1	4.2	1.000	0.999
second point	4959	40.2	402.1	399.7	2.4	403.7	400.0	3.7	0.996	0.999
third point	4979	20.1	201.0	199.8	1.2	202.7	199.4	3.4	0.992	1.002
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as left span	4919	80.5	805.1	437.1	368.0	804.5	437.5	366.9	1.001	0.999
Average Correction Factor									0.996	1.000

Corrected As found	NO _x = 816.2 ppb	NO = 809.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.5%
Previous Response	NO _x = 804.1 ppb	NO = 799.4 ppb		*Percent Change	NO = 1.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.3	435.1	368.0	368.4	0.999	100.1%
2nd GPT point (200 ppb O3)	798.3	658.4	144.7	144.9	0.999	100.1%
3rd GPT point (100 ppb O3)	798.3	727.3	75.8	76.2	0.995	100.5%
Average Correction Factor					0.998	100.2%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 5, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:03	End Time (MST):	13:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
805.1	805.4	0.9996		
402.1	403.7	0.9959		
201.0	202.7	0.9918		
			0.999994	
			0.999930	
			0.929269	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 5, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:03	End Time (MST):	13:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	801.1	0.9989		
399.7	400.0	0.9991		
199.8	199.4	1.0022		
			0.999999	
			1.001426	
			-0.332141	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 5, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:03	End Time (MST):	13:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
368.0	368.4	0.9990		
144.7	144.9	0.9988		
75.8	76.2	0.9952		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 23, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Barge Landing Station number: AMS 09
Calibration Date: August 23, 2023 Last Cal Date: July 5, 2023
Start time (MST): 9:38 End time (MST): 10:45

Analyzer Make: API T640 S/N: 321
Particulate Fraction: PM2.5

Flow Meter Make/Model: DeltaCal S/N: 1451
Temp/RH standard: DeltaCal S/N: 1451

Monthly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Measured</u>	<u>As left</u>	<u>Adjusted</u>	(Limits)
T (°C)	15.8	16.2	15.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	732.8	733.5	732.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	5.06	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>August 23, 2023</u>	Last Cal Date: <u>July 5, 2023</u>			
	PM w/o HEPA: <u>26.9</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Post maintenance</u>	<u>As left</u>	<u>Adjusted</u>	(Limits)
PMT Peak Test	10.0	10.9	10.9	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>26.6</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>August 23, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>August 23, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: August 23, 2023
Date RH/T Sensor Cleaned: August 23, 2023

Notes: Inlet head looks good. No adjustments made. Leak check passed. PMT adjusted.

Calibration by: Sean Bala



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Barge Landing	Station Number:	AMS 09
Calibration Date:	August 22, 2023	Prev Cal Date:	November 15, 2022
Start Time (MST):	10:45	End Time (MST):	10:59
Tower Height (m):	20.0	Reason:	Removal

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	P19838
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.5%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	1.000000	0.999999	≥ 0.9995
Calculated slope	0.999465	0.999594	<i>0.90 - 1.10</i>
Calculated intercept	-0.013446	-0.028293	<i>+/- 2</i>

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	E4852
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:29	Calc Declination*:	13.72 Degrees
Deadband calc:	<i>4.1 degrees (Limit 4 deg)</i>		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	1.0	---
90	88.1	-0.5%
180	177.1	-0.8%
270	266.3	-1.0%
357	353.9	-0.9%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999975	0.999975	≥ 0.9995
Calculated slope	1.010368	1.011326	<i>0.90 - 1.10</i>
Calculated intercept	-3.799847	0.097530	<i>+/- 4</i>

Notes: Removing the old 20m WS/WD sensors. WD sensor calibration was completed at the shop.

Calibration Performed By: Rene Chamberland & Devin Russell



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Barge Landing	Station Number:	AMS 09
Calibration Date:	August 22, 2023	Prev Cal Date:	November 15, 2022
Start Time (MST):	10:15	End Time (MST):	10:30
Tower Height (m):	20.0	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	B4129
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.5%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥ 0.9995
Calculated slope		0.999594	<i>0.90 - 1.10</i>
Calculated intercept		-0.028293	<i>+/- 2</i>

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	D14061
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:29	Calc Declination*:	13.72 Degrees
Deadband calc:	-1.8 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	-0.7	---
90	88.2	-0.5%
180	178.5	-0.4%
270	268.6	-0.4%
357	358.1	0.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999975	≥ 0.9995
Calculated slope		0.995462	<i>0.90 - 1.10</i>
Calculated intercept		1.663468	<i>+/- 4</i>

Notes: Installing new 20m WS/WD sensors. Crossarm aligned with true north. No issues to note.

Calibration Performed By: Rene Chamberland & Devin Russell.




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS11
LOWER CAMP
AUGUST 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	August 10, 2023	Last Cal Date:	July 17, 2023
Start time (MST):	9:48	End time (MST):	12:42
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.25	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC2216			
Removed Cal Gas Conc:	49.25	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3807
ZAG Make/Model:	Teledyne API T701		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	100841398
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004778	1.001812	Backgd or Offset:	14.7	14.8
Calibration intercept:	0.370932	0.110451	Coeff or Slope:	1.034	1.034

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4919	81.3	800.8	803.3	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.6	----
high point	4919	81.3	800.8	802.6	0.998
second point	4959	40.7	400.9	401.4	0.999
third point	4980	20.3	199.9	200.0	1.000
as left zero	5000	0.0	0.0	0.6	----
as left span	4919	81.3	800.8	805.7	0.994
Average Correction Factor					0.999

Baseline Corr As found:	803.00	Previous response	804.95	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 17, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:48	End Time (MST):	12:42
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.6	----	Correlation Coefficient	≥0.995
800.8	802.6	0.9977		
400.9	401.4	0.9988	Slope	0.90 - 1.10
199.9	200.0	0.9997		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 10, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Lower Camp Station number: AMS11
 Calibration Date: August 29, 2023 Last Cal Date: July 18, 2023
 Start time (MST): 10:31 End time (MST): 14:44
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.429 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC501097
 Removed Cal Gas Conc: 5.429 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3807
 ZAG Make/Model: API T701H Serial Number: 196

Analyzer Information

Analyzer make: Thermo 450iQ Analyzer serial #: CM20080003
 Converter make: NA Converter serial #: NA
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.017077	1.017077	Backgd or Offset: 13.5	14.1
Calibration intercept:	0.133554	-0.466446	Coeff or Slope: 1.001	1.001

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4926	73.6	79.9	81.0	0.993
as found 2nd point	4963	36.8	40.0	41.0	0.987
as found 3rd point	4982	18.6	20.2	20.6	1.005
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.5	----
high point	4926	73.6	79.9	80.8	0.989
second point	4963	36.8	40.0	40.2	0.994
third point	4982	18.6	20.2	20.1	1.005
as left zero	5000	0.0	0.0	-0.2	----
as left span	4926	73.6	79.9	80.4	0.994
SO2 Scrubber Check	4919	81.1	811.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	0.996
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 80.5 Prev response: 81.42 *% change: -1.1%
 Baseline Corr 2nd AF pt: 40.5 AF Slope: 1.008617 AF Intercept: 0.454811
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999968

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Completed SO2 scrubber check after calibrator zero.
 Adjusted zero only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 29, 2023	Previous Calibration:	July 18, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:31	End Time (MST):	14:44
Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.5	----	Correlation Coefficient	≥0.995
79.9	80.8	0.9891		
40.0	40.2	0.9940	Slope	0.90 - 1.10
20.2	20.1	1.0046		
			Intercept	+/-3



H₂S Calibration Plot

Date: August 29, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	August 10, 2023	Last Cal Date:	July 17, 2023
Start time (MST):	9:48	End time (MST):	12:42
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.5 ppm		
Diff between cyl (CH ₄):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701	Serial Number:	3807
		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.99E-04	2.99E-04	NMHC SP Ratio:	5.79E-05
CH ₄ Retention time:	14.0	14.0	NMHC Peak Area:	158468
				158468

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	17.35	17.24	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	17.35	17.29	1.004
second point	4959	40.7	8.69	8.59	1.011
third point	4980	20.3	4.33	4.29	1.011
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	17.35	17.36	1.000

Average Correction Factor				1.009
Baseline Corr AF:	17.24	Prev response	17.21	*% change 0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	9.19	9.14	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	9.19	9.15	1.004
second point	4959	40.7	4.60	4.56	1.010
third point	4980	20.3	2.29	2.28	1.008
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	9.19	9.19	1.000
Average Correction Factor					1.007
Baseline Corr AF:	9.14	Prev response	9.12	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	8.16	8.10	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	8.16	8.13	1.004
second point	4959	40.7	4.09	4.03	1.014
third point	4980	20.3	2.04	2.01	1.014
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	8.16	8.16	1.000
Average Correction Factor					1.010
Baseline Corr AF:	8.10	Prev response	8.09	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.992518	0.996495
THC Cal Offset:	-0.014592	-0.024788
CH ₄ Cal Slope:	0.993056	0.996527
CH ₄ Cal Offset:	-0.015088	-0.015486
NMHC Cal Slope:	0.992065	0.996217
NMHC Cal Offset:	0.000895	-0.009301

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary

Version-01-2020


Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 17, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:48	End Time (MST):	12:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999982	≥ 0.995			
17.35	17.29	1.0037						
8.69	8.59	1.0112				Slope	0.996495	0.90 - 1.10
4.33	4.29	1.0110						
			Intercept	-0.024788	± 0.5			

THC Calibration Curve





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-01-2020


Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 17, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:48	End Time (MST):	12:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999969	≥ 0.995
8.16	8.13	1.0037			
4.09	4.03	1.0135			
2.04	2.01	1.0139			
			Slope	0.996527	0.90 - 1.10
			Intercept	-0.015486	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 17, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:48	End Time (MST):	12:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999990	≥ 0.995			
9.19	9.15	1.0039						
4.60	4.56	1.0095				Slope	0.996217	0.90 - 1.10
2.29	2.28	1.0085						
			Intercept	-0.009301	± 0.5			



NMHC Calibration Plot

Date: August 10, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Lower Camp	Station Number:	AMS 11
Calibration Date:	August 16, 2023	Prev Cal Date:	July 21, 2023
Start Time (MST):	9:17	End Time (MST):	9:45
Tower Height (m):	10.0	Reason:	As Found To test WD sensor.

Wind Speed Information

Sensor make/model:	NA	Serial Number:	NA
WS Calibrator:	NA	Serial Number:	NA

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	NA	---
200	20.2	NA	
400	39.4	NA	
600	58.6	NA	
800	77.8	NA	

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999998		≥0.9995
Calculated slope	0.998863		0.90 - 1.10
Calculated intercept	0.064097		+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	N12035
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:30	Calc Declination*:	13.64 Degrees
Deadband calc:	4.0 degrees (<i>Limit 4 deg</i>)		* - calculated declination as per NOAA website

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	3.0	---
90	90.6	0.2%
180	180.1	0.0%
270	272.5	0.7%
354.5	356.0	0.4%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999918	0.999964	≥0.9995
Calculated slope	0.998903	1.001308	0.90 - 1.10
Calculated intercept	-0.302646	-1.771447	+/- 4

Notes: Performed WD as founds to test sensor since it was not trending with nearby stations. Sensor was not secured in its coupler properly this caused misalignment with true north; readings were off ~54 degrees. Installed the sensor back and realigned crossarm using solar noon as reference.

Calibration Performed By: Mohammed Kashif




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13 FORT MCKAY SOUTH AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
Calibration Date: August 16, 2023 Last Cal Date: July 25, 2023
Start time (MST): 8:48 End time (MST): 12:02
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.55 ppm Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #: CC260812
Removed Cal Gas Conc: 50.55 ppm Rem Gas Exp Date: N/A
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 2448
ZAG Make/Model: API 701 Serial Number: 1117

Analyzer Information

Analyzer make: API T100 Analyzer serial #: 599
Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002083	1.002886	Backgd or Offset:	86.3	86.3
Calibration intercept:	-1.997771	-2.078412	Coeff or Slope:	0.715	0.715

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4921	79.1	799.7	799.4	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	79.1	799.7	801.0	0.998
second point	4961	39.5	399.3	397.3	1.005
third point	4980	19.8	200.2	196.5	1.019
as left zero	5000	0.0	0.0	0.2	----
as left span	4921	79.1	799.7	800.6	0.999
Average Correction Factor					1.007

Baseline Corr As found: 799.30 Previous response 799.35 *% change 0.0%
Baseline Corr 2nd AF pt: NA AF Slope: AF Intercept:
Baseline Corr 3rd AF pt: NA AF Correlation:

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 16, 2023	Previous Calibration:	July 25, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:48	End Time (MST):	12:02
Analyzer make:	API T100	Analyzer serial #:	599

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
799.7	801.0	0.9984		
399.3	397.3	1.0050	Slope	0.90 - 1.10
200.2	196.5	1.0188		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 16, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: August 15, 2023 Last Cal Date: July 6, 2023
 Start time (MST): 8:55 End time (MST): 13:25
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.34 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC500241
 Removed Cal Gas Conc: 5.34 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API 701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1180540017
 Converter make: CDN-101 Converter serial #: 521
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995542	1.004479	Backgd or Offset:	3.54	3.77
Calibration intercept:	-0.262162	-0.502377	Coeff or Slope:	1.066	1.130

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4925	75.5	80.6	76.9	1.046
as found 2nd point	4962	37.7	40.3	38.1	1.051
as found 3rd point	4981	18.9	20.2	18.7	1.068
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4925	75.5	80.6	80.7	0.999
second point	4962	37.7	40.3	39.6	1.017
third point	4981	18.9	20.2	19.6	1.030
as left zero	5000	0.0	0.0	0.0	----
as left span	4925	75.5	80.6	78.3	1.030
SO2 Scrubber Check	4921	79.1	791.0	0.1	----

Date of last scrubber change:	20-Mar-20	Ave Corr Factor	1.015
Date of last converter efficiency test:	NA	efficiency	

Baseline Corr As found: 77.1 Prev response: 80.00 *% change: -3.8%
 Baseline Corr 2nd AF pt: 38.3 AF Slope: 0.957689 AF Intercept: -0.402123
 Baseline Corr 3rd AF pt: 18.9 AF Correlation: 0.999968

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. Span adjusted.

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 6, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:55	End Time (MST):	13:25
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999920	≥0.995
80.6	80.7	0.9991			
40.3	39.6	1.0168	Slope	1.004479	0.90 - 1.10
20.2	19.6	1.0299			
			Intercept	-0.502377	+/-3



TRS Calibration Plot

Date: August 15, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	August 16, 2023	Last Cal Date:	July 25, 2023
Start time (MST):	8:48	End time (MST):	12:02
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API701	Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	11700501330
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.21E-04	2.22E-04	NMHC SP Ratio:	5.17E-04	4.70E-04
CH ₄ Retention time:	12.80	12.80	NMHC Peak Area:	175502	193345
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF	ON

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	17.05	17.08	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.07	0.999
second point	4961	39.5	8.51	8.41	1.012
third point	4980	19.8	4.27	4.13	1.033
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	17.05	17.05	1.000
Average Correction Factor					1.015

Baseline Corr AF:	17.08	Prev response	17.02	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	9.08	9.11	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	9.08	9.09	0.999
second point	4961	39.5	4.53	4.50	1.008
third point	4980	19.8	2.27	2.21	1.027
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	9.08	9.07	1.001
Average Correction Factor					1.011
Baseline Corr AF:	9.11	Prev response	9.07	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	7.97	7.97	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.98	0.998
second point	4961	39.5	3.98	3.91	1.017
third point	4980	19.8	1.99	1.92	1.040
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	7.97	7.98	0.998
Average Correction Factor					1.018
Baseline Corr AF:	7.97	Prev response	7.95	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.003313	1.003320
THC Cal Offset:	-0.083173	-0.078975
CH ₄ Cal Slope:	1.003948	1.004551
CH ₄ Cal Offset:	-0.045749	-0.047351
NMHC Cal Slope:	1.002735	1.002546
NMHC Cal Offset:	-0.038763	-0.032763

Notes: Changed inlet filter after as founds. Used zero chromatogram and flat baseline. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary

Version-06-2022


Station Information

Calibration Date:	August 16, 2023	Previous Calibration:	July 25, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:48	End Time (MST):	12:02
Analyzer make:	Thermo 55i	Analyzer serial #:	11700501330

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999901	≥ 0.995
17.05	17.07	0.9988			
8.51	8.41	1.0121	Slope	1.003320	0.90 - 1.10
4.27	4.13	1.0330			
			Intercept	-0.078975	± 0.5

THC Calibration Curve





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	August 16, 2023	Previous Calibration:	July 25, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:48	End Time (MST):	12:02
Analyzer make:	Thermo 55i	Analyzer serial #:	11700501330

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999833	≥ 0.995
7.97	7.98	0.9978			
3.98	3.91	1.0169			
1.99	1.92	1.0398			
			Slope	1.004563	0.90 - 1.10
			Intercept	-0.047194	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	August 16, 2023	Previous Calibration:	July 25, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:48	End Time (MST):	12:02
Analyzer make:	Thermo 55i	Analyzer serial #:	11700501330

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999941	≥ 0.995
9.08	9.09	0.9993			
4.53	4.50	1.0081			
2.27	2.21	1.0270			
			Slope	1.002557	0.90 - 1.10
			Intercept	-0.032582	+/-0.5



NMHC Calibration Plot

Date: August 16, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	August 29, 2023	Last Cal Date:	August 16, 2023
Start time (MST):	8:53	End time (MST):	10:41
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API701	Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	11700501330
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>	
CH ₄ SP Ratio:	2.21E-04	2.22E-04	NMHC SP Ratio:	5.17E-04	4.70E-04
CH ₄ Retention time:	12.80	12.80	NMHC Peak Area:	175502	193345
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF	ON

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	17.05	17.08	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.08	0.998
second point					
third point					
as left zero					
as left span					
Average Correction Factor					0.998

Baseline Corr AF:	17.08	Prev response	17.02	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	9.08	9.17	0.990
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	9.08	9.16	0.991
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.991
Baseline Corr AF:	9.17	Prev response	9.07	*% change	1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	7.97	7.91	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.92	1.007
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.007
Baseline Corr AF:	7.91	Prev response	7.95	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.003313	1.003320
THC Cal Offset:	-0.083173	-0.078975
CH ₄ Cal Slope:	1.003948	1.004551
CH ₄ Cal Offset:	-0.045749	-0.047351
NMHC Cal Slope:	1.002735	1.002546
NMHC Cal Offset:	-0.038763	-0.032763

Notes:

Changed N2 cylinder.


Calibration Performed By:

Sean Bala

NMHC Calibration Plot

Date: August 29, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort McKay South
Calibration Date: August 24, 2023
Start time (MST): 8:45
Reason: Routine
Station number: AMS 13
Last Cal Date: July 17, 2023
End time (MST): 13:26

Calibration Standards

NO Gas Cylinder #: T2Y1P76
NOX Cal Gas Conc: 50.98 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 50.98 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 49.32 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 49.32 ppm
NO gas Diff:
Serial Number: 2448
Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1410661329

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.113	1.113	NO bkgnd or offset:	9.1	9.5
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	10.5	9.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	159.0	160.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000258	0.996609
NO _x Cal Offset:	-1.931590	-1.710896
NO Cal Slope:	1.002649	0.999177
NO Cal Offset:	-2.105499	-2.764638
NO ₂ Cal Slope:	1.001659	0.995830
NO ₂ Cal Offset:	-0.999973	-0.645143



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-1.1	0.2	-1.3	----	----
as found span	4919	81.1	826.9	800.0	26.9	823.9	796.6	27.3	1.0036	1.0042
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4919	81.1	826.9	800.0	26.9	823.2	797.9	25.2	1.0045	1.0026
second point	4960	40.6	413.9	400.4	13.5	410.0	396.0	14.1	1.0095	1.0112
third point	4980	20.3	207.0	200.2	6.7	202.8	194.5	8.3	1.0205	1.0294
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
as left span	4919	80.1	816.8	372.6	444.2	828.9	380.3	448.5	0.9855	0.9798
Average Correction Factor									1.0115	1.0144

Corrected As found	NO _x = 825.0 ppb	NO = 796.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.0%	
Previous Response	NO _x = 825.2 ppb	NO = 800.0 ppb		*Percent Change	NO = -0.4%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	795.2	377.9	444.2	442.3	1.0044	99.6%
2nd GPT point (200 ppb O ₃)	795.2	585.5	236.6	234.1	1.0108	98.9%
3rd GPT point (100 ppb O ₃)	795.2	689.2	132.9	131.4	1.0116	98.9%
Average Correction Factor					1.0089	99.1%

Notes:

Changed inlet filter after as founds. Adjusted zero only.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 24, 2023	Previous Calibration:	July 17, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:45	End Time (MST):	13:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.9	823.2	1.0045		
413.9	410.0	1.0095		
207.0	202.8	1.0205		
			0.999980	
			0.996609	
			-1.710896	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 24, 2023	Previous Calibration:	July 17, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:45	End Time (MST):	13:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.0	797.9	1.0026		
400.4	396.0	1.0112		
200.2	194.5	1.0294		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 24, 2023	Previous Calibration:	July 17, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:45	End Time (MST):	13:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
444.2	442.3	1.0044			
236.6	234.1	1.0108			
132.9	131.4	1.0116			
			Slope	0.995830	0.90 - 1.10
			Intercept	-0.645143	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 24, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: August 11, 2023 Last Cal Date: July 11, 2023
 Start time (MST): 9:01 End time (MST): 12:26
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3871
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996829	0.997543	Backgd or Offset:	2.4	2.4
Calibration intercept:	0.980000	1.180000	Coeff or Slope:	0.963	0.963

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	5000	977.0	400.0	398.6	1.004
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.4	----
high point	5000	977.0	400.0	399.8	1.001
second point	5000	838.0	200.0	201.1	0.995
third point	5000	735.9	100.0	101.7	0.983
as left zero	5000	0.0	0.0	0.1	----
as left span	5000	977.0	400.0	401.3	0.997
Average Correction Factor					0.993

Baseline Corr As found:	398.6	Previous response	399.7	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustments made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 11, 2023	Previous Calibration:	July 11, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	9:01	End Time (MST):	12:26
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient	≥0.995
400.0	399.8	1.0005		
200.0	201.1	0.9945	Slope	0.90 - 1.10
100.0	101.7	0.9833		
			Intercept	+/- 5



O₃ Calibration Plot

Date: August 11, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort McKay South Station number: AMS 13
 Calibration Date: August 24, 2023 Last Cal Date: July 26, 2023
 Start time (MST): 10:37 End time (MST): 11:42

Analyzer Make: API T640 S/N: 319
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	17.6	17.4	17.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	737.5	738.5	737.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	5.02	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: August 24, 2023	Last Cal Date: July 26, 2023			
	PM w/o HEPA: 9.0	PM w/ HEPA: 0.0			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		June 29, 2023			<0.2 ug/m3
Disposable Filter Changed:		July 26, 2023			

Annual Maintenance

Date Sample Tube Cleaned: June 29, 2023
 Date RH/T Sensor Cleaned: June 29, 2023

Inlet head clean and inspected. Leak check passed. No adjustment made.

Notes:

Calibration by: Sean Bala




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC
AUGUST 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Anzac Station number: AMS 14
Calibration Date: August 9, 2023 Last Cal Date: July 6, 2023
Start time (MST): 8:54 End time (MST): 13:17
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.95 ppm Cal Gas Exp Date: January 5, 2025
Cal Gas Cylinder #: CC279389
Removed Cal Gas Conc: 49.95 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3060
ZAG Make/Model: API T701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 0710321322
Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.020118	1.015152	Backgd or Offset:	25.1	25.2
Calibration intercept:	-0.460015	-1.929538	Coeff or Slope:	0.798	0.798

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.6	----
as found span	4938	80.3	799.3	809.1	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.8	----
high point	4938	80.3	799.3	811.0	0.986
second point	4979	40.2	400.1	402.3	0.994
third point	4998	20.2	201.1	199.8	1.006
as left zero	5000	0.0	0.0	0.7	----
as left span	4938	80.3	799.3	813.0	0.983
Average Correction Factor					0.995

Baseline Corr As found: 808.50 Previous response 814.89 *% change -0.8%
Baseline Corr 2nd AF pt: NA AF Slope: AF Intercept:
Baseline Corr 3rd AF pt: NA AF Correlation:

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 6, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	8:54	End Time (MST):	13:17
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.8	----	Correlation Coefficient	≥0.995
799.3	811.0	0.9855		
400.1	402.3	0.9944	Slope	0.90 - 1.10
201.1	199.8	1.0063		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 9, 2023

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	August 4, 2023	Last Cal Date:	July 2, 2023
Start time (MST):	13:00	End time (MST):	16:25
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.15	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC510379			
Removed Cal Gas Conc:	5.38	ppm	Rem Gas Exp Date:	February 3, 2023
Removed Gas Cyl #:	EY0000859		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API 701H		Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582
Converter make:	CD Nova CDN-101	Converter serial #:	503
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998548	0.991300	Backgd or Offset:	2.20
Calibration intercept:	-0.100891	-0.125436	Coeff or Slope:	0.960
				2.30
				0.992

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4938	77.9	80.0	79.4	1.007
second point	4973	38.9	40.0	39.0	1.025
third point	4997	19.5	20.0	19.6	1.021
as left zero	5000	0.0	0.0	-0.1	----
as left span	4938	77.9	80.0	78.9	1.013
SO2 Scrubber Check	4936	80.3	800.4	-0.1	----
Date of last scrubber change:				Ave Corr Factor	1.017
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: Installed new cal gas cylinder. Changed sample inlet filter. Performed scrubber check after calibrator zero and it passed. Adjusted span only. Could not perform MAF's; refer to docit notes.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 4, 2023	Previous Calibration:	July 2, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	13:00	End Time (MST):	16:25
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999876	≥0.995
80.0	79.4	1.0069			
40.0	39.0	1.0245	Slope	0.991300	0.90 - 1.10
20.0	19.6	1.0210			
			Intercept	-0.125436	+/-3



TRS Calibration Plot

Date: August 4, 2023

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	August 9, 2023	Last Cal Date:	July 6, 2023
Start time (MST):	8:54	End time (MST):	13:17
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.90E-04	3.90E-04	NMHC SP Ratio:	4.53E-05
CH ₄ Retention time:	12.20	12.20	NMHC Peak Area:	201206
				201206

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	17.12	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.22	0.993
second point	4979	40.2	8.56	8.59	0.996
third point	4998	20.2	4.30	4.27	1.008
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	17.10	17.30	0.989

Average Correction Factor				0.999
Baseline Corr AF:	17.12	Prev response	17.15	*% change -0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	9.06	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.08	1.004
second point	4979	40.2	4.56	4.53	1.006
third point	4998	20.2	2.29	2.25	1.018
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	9.11	9.13	0.999
Average Correction Factor					1.010
Baseline Corr AF:	9.06	Prev response	9.09	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	8.07	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	8.14	0.982
second point	4979	40.2	4.00	4.06	0.986
third point	4998	20.2	2.01	2.02	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	7.99	8.17	0.978
Average Correction Factor					0.988
Baseline Corr AF:	8.07	Prev response	8.05	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002824	1.007749
THC Cal Offset:	-0.002175	-0.029955
CH ₄ Cal Slope:	1.007994	1.019842
CH ₄ Cal Offset:	-0.000017	-0.015535
NMHC Cal Slope:	0.998090	0.997110
NMHC Cal Offset:	-0.001358	-0.015020

Notes: Sample inlet filter changed after as founds. Swapped hydrogen cylinder. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 6, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	8:54	End Time (MST):	13:17
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999985	≥ 0.995			
17.10	17.22	0.9934						
8.56	8.59	0.9963				Slope	1.007749	0.90 - 1.10
4.30	4.27	1.0078						
			Intercept	-0.029955	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 6, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	8:54	End Time (MST):	13:17
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999983	≥ 0.995			
7.99	8.14	0.9815						
4.00	4.06	0.9860				Slope	1.019842	0.90 - 1.10
2.01	2.02	0.9960						
			Intercept	-0.015535	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 6, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	8:54	End Time (MST):	13:17
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995
9.11	9.08	1.0040			
4.56	4.53	1.0063			
2.29	2.25	1.0185			
			Slope	0.997110	0.90 - 1.10
			Intercept	-0.015020	+/-0.5



NMHC Calibration Plot

Date: August 9, 2023

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Anzac Station number: AMS 14
Calibration Date: August 15, 2023 Last Cal Date: July 7, 2023
Start time (MST): 7:25 End time (MST): 11:56
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1P8D Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.92 ppm NO Cal Gas Conc: 50.05 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.92 ppm Removed Gas NO Conc: 50.05 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3060
ZAG make/model: Teledyne API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1426262592
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.375	1.375	NO bkgnd or offset:	3.7	3.7
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.7	3.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	163.9	161.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995865	0.986023
NO _x Cal Offset:	0.033705	-0.615739
NO Cal Slope:	0.996593	0.987256
NO Cal Offset:	-1.310286	-2.172778
NO ₂ Cal Slope:	1.000260	1.000092
NO ₂ Cal Offset:	1.094049	0.381307



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.2	----	----
as found span	4932	80.1	813.8	799.9	13.9	803.6	786.6	17.0	1.0127	1.0169
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
high point	4932	80.1	813.8	799.9	13.9	802.0	788.7	13.4	1.0147	1.0142
second point	4980	40.1	406.7	399.8	6.9	400.6	391.1	9.5	1.0153	1.0222
third point	4998	20.1	204.0	200.5	3.5	199.3	193.8	5.6	1.0234	1.0344
as left zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.2	----	----
as left span	4932	80.1	813.8	420.4	393.4	803.4	409.8	393.7	1.0129	1.0258
Average Correction Factor									1.0178	1.0236

Corrected As found	NO _x = 803.6 ppb	NO = 786.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.9%	
Previous Response	NO _x = 810.4 ppb	NO = 795.8 ppb		*Percent Change	NO = -1.2%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	786.8	407.3	393.4	393.5	0.9998	100.0%
2nd GPT point (200 ppb O ₃)	786.8	604.6	196.1	197.3	0.9939	100.6%
3rd GPT point (100 ppb O ₃)	786.8	699.1	101.6	101.7	0.9991	100.1%
Average Correction Factor					0.9976	100.2%

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 7, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	7:25	End Time (MST):	11:56
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
813.8	802.0	1.0147			
406.7	400.6	1.0153			
204.0	199.3	1.0234			
			Slope	0.986023	0.90 - 1.10
			Intercept	-0.615739	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 7, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	7:25	End Time (MST):	11:56
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.9	788.7	1.0142		
399.8	391.1	1.0222		
200.5	193.8	1.0344		
			0.999965	
			0.987256	
			-2.172778	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 7, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	7:25	End Time (MST):	11:56
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
393.4	393.5	0.9998		
196.1	197.3	0.9939		
101.6	101.7	0.9991		
			0.999990	
			1.000092	
			0.381307	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 15, 2023

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	August 1, 2023	Last Cal Date:	July 11, 2023
Start time (MST):	9:53	End time (MST):	13:05
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	API T700	Serial Number:	5239
ZAG Make/Model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595
Analyzer Range:	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.989000	0.999171	Backgd or Offset:	1.3	1.3
Calibration intercept:	0.500000	0.520000	Coeff or Slope:	1.516	1.550

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.6	----
as found span	5000	884.2	400.0	395.7	1.011
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.2	----
high point	5000	884.9	400.0	399.9	1.000
second point	5000	769.3	200.0	200.9	0.996
third point	5000	669.8	100.0	100.5	0.995
as left zero	5000	0.0	0.0	0.5	----
as left span	5000	883.1	400.0	403.0	0.993
Average Correction Factor					0.997

Baseline Corr As found:	395.1	Previous response	396.1	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes:

Sample inlet filter changed after as founds. Adjusted span only.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 11, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	9:53	End Time (MST):	13:05
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999994	≥0.995
400.0	399.9	1.0003			
200.0	200.9	0.9955	Slope	0.999171	0.90 - 1.10
100.0	100.5	0.9950			
			Intercept	0.520000	+/- 5



O₃ Calibration Plot

Date: August 1, 2023

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Anzac Station number: AMS 14
 Calibration Date: August 1, 2023 Last Cal Date: July 6, 2023
 Start time (MST): 12:55 End time (MST): 14:50

Analyzer Make: API T640 S/N: 825
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388749
 Temp/RH standard: Alicat FP-25 S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	21.0	22.4	21.0	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	715.1	716.3	715.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.03	4.85	5.03	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>August 1, 2023</u>		Last Cal Date: <u>July 6, 2023</u>		
	PM w/o HEPA: <u>32.8</u>		PM w/ HEPA: <u>0.0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	0	11	11	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: <u>30.9</u>		w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:	<u>August 1, 2023</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>August 1, 2023</u>				

Annual Maintenance

Date Sample Tube Cleaned: July 6, 2023
 Date RH/T Sensor Cleaned: July 6, 2023

Notes: No adjustments needed. Performed quarterly maintenance. As found PMT leak test is 0, because the analyzer's screen appeared to be frozen and was reading 0.00 µg/m3. Leak check passed. Head Cleaned.

Calibration by: Mohammed Kashif



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Anzac	Station Number:	AMS 14
Calibration Date:	August 21, 2023	Prev Cal Date:	August 9, 2022
Start Time (MST):	11:13	End Time (MST):	11:50
Tower Height (m):	20.0	Reason:	Routine

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	D6359
WS Calibrator:	MetOne 053	Serial Number:	CA05231

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	NA	---
200	20.2	NA	
400	39.4	NA	
600	58.6	NA	
800	77.8	NA	

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope			0.90 - 1.10
Calculated intercept			+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	Z1048
As Found Declination (deg east of True North):	NA	As Left Declination (deg east of True North):	NA
Solar noon time (MST):	13:27	Calc Declination*:	12.03 Degrees
Deadband calc:	0.7 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	2.8	---
90	92.2	0.6%
180	182.1	0.6%
270	271.9	0.5%
356	359.1	0.9%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999994	≥0.9995
Calculated slope		0.999682	0.90 - 1.10
Calculated intercept		-2.362272	+/- 4

Notes: As found WD calibration; could not perform WS calibration due to damaged cable. WS bearings appears to be not in optimal condition on inspection.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Anzac	Station Number:	AMS 14
Calibration Date:	August 23, 2023	Prev Cal Date:	NA
Start Time (MST):	10:52	End Time (MST):	11:20
Tower Height (m):	20.0	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	E5132
WS Calibrator:	MetOne 053	Serial Number:	CA05231

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)		0.999999	≥ 0.9995
Calculated slope		0.998909	$0.90 - 1.10$
Calculated intercept		0.030357	± 2

Wind Direction Information

Sensor make/model:	NA	Serial Number:	NA
As Found Declination (deg east of True North):	<u>NA</u>	As Left Declination (deg east of True North):	<u>NA</u>
Solar noon time (MST):	NA	Calc Declination*:	NA Degrees
Deadband calc:	#VALUE! degrees (<i>Limit 4 deg</i>)	* - calculated declination as per NOAA website	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	NA	---
90	NA	
180	NA	
270	NA	
356	NA	

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)			≥ 0.9995
Calculated slope			$0.90 - 1.10$
Calculated intercept			± 4

Notes: WS installation. Inspected bearings, they appear to be in good condition. Swapped both WS and WD cables.

Calibration Performed By: Mohammed Kashif




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS17
WAPASU
AUGUST 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	August 3, 2023	Last Cal Date:	July 5, 2023
Start time (MST):	10:00	End time (MST):	13:08
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	<u>50.38</u> ppm	Cal Gas Exp Date:	January 12, 2029
Cal Gas Cylinder #:	<u>ALM066507</u>		
Removed Cal Gas Conc:	<u>50.38</u> ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	<u>n/a</u>	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	2449
ZAG Make/Model:	API 701H	Serial Number:	359

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008014	1.003838	Backgd or Offset:	12.6	12.6
Calibration intercept:	-1.662517	-2.099065	Coeff or Slope:	1.111	1.111

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4921	79.4	800.0	799.6	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	79.4	800.0	802.4	0.997
second point	4960	39.7	400.0	397.2	1.007
third point	4980	19.8	199.5	196.9	1.013
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	79.4	800.1	804.2	0.995
Average Correction Factor					1.006

Baseline Corr As found:	799.80	Previous response	804.72	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 5, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:00	End Time (MST):	13:08
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	≥0.995
800.0	802.4	0.9970		
400.0	397.2	1.0072	Slope	0.90 - 1.10
199.5	196.9	1.0133		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 3, 2023

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Wapasu Station number: AMS17
 Calibration Date: August 8, 2023 Last Cal Date: July 17, 2023
 Start time (MST): 9:45 End time (MST): 15:08
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.076 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511852
 Removed Cal Gas Conc: 5.076 ppm Rem Gas Exp Date: n/a
 Removed Gas Cyl #: n/a Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 2449
 ZAG Make/Model: API T701H Serial Number: 359

Analyzer Information

Analyzer make: Thermo 450i Analyzer serial #: 1218153583
 Converter make: n/a Converter serial #: n/a
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996997	0.996998	Backgd or Offset: 13.0	13.1
Calibration intercept:	-0.319220	-0.219239	Coeff or Slope: 1.134	1.116

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	78.8	80.0	83.7	0.955
as found 2nd point	4961	39.4	40.0	41.8	0.955
as found 3rd point	4980	19.7	20.0	21.0	0.948
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	78.8	80.0	79.9	1.001
second point	4961	39.4	40.0	38.9	1.028
third point	4980	19.7	20.0	19.8	1.010
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	78.8	80.0	77.2	1.036
SO2 Scrubber Check	4921	79.4	800.0	-0.1	----
Date of last scrubber change:	n/a			Ave Corr Factor	1.013
Date of last converter efficiency test:	n/a				efficiency

Baseline Corr As found: 83.8 Prev response: 79.44 *% change: 5.2%
 Baseline Corr 2nd AF pt: 41.9 AF Slope: 1.046854 AF Intercept: -0.039164
 Baseline Corr 3rd AF pt: 21.1 AF Correlation: 0.999996

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 8, 2023	Previous Calibration:	July 17, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:45	End Time (MST):	15:08
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999768	≥0.995
80.0	79.9	1.0013			
40.0	38.9	1.0282	Slope	0.996998	0.90 - 1.10
20.0	19.8	1.0101			
			Intercept	-0.219239	+/-3



H₂S Calibration Plot

Date: August 8, 2023

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 5, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:00	End Time (MST):	13:08
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.05	----	Correlation Coefficient	0.999966	≥0.995
17.09	17.03	1.0036			
8.55	8.40	1.0174	Slope	0.999409	0.90 - 1.10
4.26	4.18	1.0197			
			Intercept	-0.080737	+/-1.5



THC Calibration Plot

Date: August 3, 2023

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Wapasu
Calibration Date: August 15, 2023
Start time (MST): 9:30
Reason: Routine
Station number: AMS17
Last Cal Date: July 25, 2023
End time (MST): 14:13

Calibration Standards

NO Gas Cylinder #: T375YK8
NOX Cal Gas Conc: 49.11 ppm
Removed Cylinder #: T375YK8
Removed Gas NOX Conc: 49.11 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: April 13, 2025
NO Cal Gas Conc: 48.07 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 48.07 ppm
NO gas Diff:
Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: Thermo Scientific 42iQ
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 12300522720

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.205	1.212	NO bkgnd or offset:	10.1	9.0
NOX coeff or slope:	0.996	0.992	NOX bkgnd or offset:	10.2	9.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	430.0	383.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.988306	1.001480
NO _x Cal Offset:	-0.640000	-3.500000
NO Cal Slope:	0.986171	0.999458
NO Cal Offset:	-1.460000	-3.960000
NO ₂ Cal Slope:	1.024572	0.998776
NO ₂ Cal Offset:	-0.186089	-0.422855



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.2	----	----
as found span	4917	83.2	817.2	799.9	17.3	801.6	782.2	19.5	1.0194	1.0226
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
high point	4917	83.2	817.2	799.9	17.3	817.3	797.9	19.5	0.9999	1.0025
second point	4958	41.6	408.6	399.9	8.7	402.1	392.6	9.5	1.0162	1.0187
third point	4979	20.8	204.3	200.0	4.3	198.7	192.6	6.1	1.0282	1.0383
as left zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.0	----	----
as left span	4917	83.2	817.2	395.3	421.9	799.4	389.1	410.3	1.0222	1.0158
Average Correction Factor									1.0147	1.0198

Corrected As found	NO _x = 801.6 ppb	NO = 782.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.7%	
Previous Response	NO _x = 807.0 ppb	NO = 787.4 ppb		*Percent Change	NO = -0.7%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.8	392.2	421.9	421.4	1.0012	99.9%
2nd GPT point (200 ppb O3)	796.8	595.6	218.5	217.0	1.0069	99.3%
3rd GPT point (100 ppb O3)	796.8	698.3	115.8	115.2	1.0053	99.5%
Average Correction Factor					1.0045	99.6%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 25, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:30	End Time (MST):	14:13
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
817.2	817.3	0.9999			
408.6	402.1	1.0162			
204.3	198.7	1.0282			
			Slope	1.001480	0.90 - 1.10
			Intercept	-3.500000	+/-20

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 25, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:30	End Time (MST):	14:13
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
799.9	797.9	1.0025			
399.9	392.6	1.0187			
200.0	192.6	1.0383			
			Slope	0.999458	0.90 - 1.10
			Intercept	-3.960000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 25, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:30	End Time (MST):	14:13
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
421.9	421.4	1.0012		
218.5	217.0	1.0069		
115.8	115.2	1.0053		
			0.999989	
			0.998776	
			-0.422855	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 15, 2023

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Wapasu Station number: AMS17
 Calibration Date: August 1, 2023 Last Cal Date: July 12, 2023
 Start time (MST): 10:22 End time (MST): 13:55
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 2449
 ZAG Make/Model: API T701H Serial Number: 359

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 3870
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000171	1.003629	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.580000	-0.960000	Coeff or Slope:	1.020	1.026

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.5	----
as found span	5000	1077.3	400.0	396.8	1.008
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.3	----
high point	5000	1077.3	400.0	400.8	0.998
second point	5000	900.3	200.0	199.5	1.003
third point	5000	789.5	100.0	98.7	1.013
as left zero	5000	0.0	0.0	0.1	----
as left span	5000	1077.3	400.0	406.4	0.984
Average Correction Factor					1.005

Baseline Corr As found:	397.3	Previous response	399.5	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 12, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:22	End Time (MST):	13:55
Analyzer make:	API T400	Analyzer serial #:	3870

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.3	----	Correlation Coefficient	≥0.995
400.0	400.8	0.9980		
200.0	199.5	1.0025	Slope	0.90 - 1.10
100.0	98.7	1.0132		
			Intercept	+/- 5



O₃ Calibration Plot

Date: August 1, 2023

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Wapasu Station number: AMS17
 Calibration Date: August 31, 2023 Last Cal Date: August 1, 2023
 Start time (MST): 10:36 End time (MST): 13:24
 Reason: Maintenance

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 2449
 ZAG Make/Model: API T701H Serial Number: 359

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 3870
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003629	1.009629	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.960000	-0.060000	Coeff or Slope:	1.026	1.026

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.3	----
high point	5000	1077.3	400.0	403.8	0.991
second point	5000	900.3	200.0	201.6	0.992
third point	5000	789.5	100.0	101.4	0.986
as left zero	5000	0.0	0.0	-0.2	----
as left span	5000	1077.3	400.0	406.7	0.984
Average Correction Factor					0.990

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Tested response after replacing a severed external pump line. No as founds done as the T400's sample flow was at 0 CC/min. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 31, 2023	Previous Calibration:	August 1, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:36	End Time (MST):	13:24
Analyzer make:	API T400	Analyzer serial #:	3870

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999996	≥0.995
400.0	403.8	0.9906			
200.0	201.6	0.9921	Slope	1.009629	0.90 - 1.10
100.0	101.4	0.9862			
			Intercept	-0.060000	+/- 5



O₃ Calibration Plot

Date: August 31, 2023

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Wapasu Station number: AMS 17
 Calibration Date: August 15, 2023 Last Cal Date: July 27, 2023
 Start time (MST): 12:37 End time (MST): 13:58

Analyzer Make: API T640 S/N: 1183
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	15.3	15.1	15.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	708.7	710.6	708.7	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.05	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: August 15, 2023	Last Cal Date: July 27, 2023			
	PM w/o HEPA: 7.3	PM w/ HEPA: 0.0			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		July 27, 2023			<0.2 ug/m3
Disposable Filter Changed:		July 27, 2023			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: Temp, pressure and flow checked. Leak check passed.

Calibration by: Aswin Sasi Kumar




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: August 10, 2023 Last Cal Date: July 18, 2023
Start time (MST): 10:25 End time (MST): 13:58
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.40 ppm Cal Gas Exp Date: February 23, 2025
Cal Gas Cylinder #: CC463851
Removed Cal Gas Conc: 49.40 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
ZAG Make/Model: Teledyne API 701H Serial Number: 360

Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: JC1501301453
Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003290	1.004073	Backgd or Offset:	22.6	22.6
Calibration intercept:	-0.043536	-0.542544	Coeff or Slope:	0.808	0.808

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5009	0.0	0.0	-0.3	----
as found span	4919	81.0	800.3	803.2	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4919	81.0	800.3	803.6	0.996
second point	4959	40.5	400.2	400.2	1.000
third point	4979	20.2	199.6	199.6	1.000
as left zero	5000	0.0	0.0	-0.1	----
as left span	4919	81.0	800.3	803.8	0.996
Average Correction Factor					0.999

Baseline Corr As found: 803.50 Previous response 802.87 *% change 0.1%
Baseline Corr 2nd AF pt: NA AF Slope: AF Intercept:
Baseline Corr 3rd AF pt: NA AF Correlation:

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 18, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:25	End Time (MST):	13:58
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999994	≥0.995
800.3	803.6	0.9959			
400.2	400.2	1.0000	Slope	1.004073	0.90 - 1.10
199.6	199.6	1.0000			
			Intercept	-0.542544	+/-30



SO2 Calibration Plot

Date: August 10, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: August 24, 2023 Last Cal Date: July 19, 2023
 Start time (MST): 10:52 End time (MST): 15:12
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.479 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC500395
 Removed Cal Gas Conc: 5.479 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
 ZAG Make/Model: Teledyne API T701 Serial Number: 360

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359
 Converter make: CD Nova CDN-101 Converter serial #: 555
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003873	0.993582	Backgd or Offset: 2.66	2.66
Calibration intercept:	0.080889	0.341175	Coeff or Slope: 1.189	1.189

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4927	73.0	80.0	80.3	0.999
as found 2nd point	4964	36.5	40.0	39.6	1.015
as found 3rd point	4983	18.3	20.0	19.5	1.039
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4927	73.0	80.0	79.8	1.002
second point	4964	36.5	40.0	40.2	0.995
third point	4983	18.3	20.0	20.1	0.997
as left zero	5000	0.0	0.0	0.4	----
as left span	4927	73.0	80.0	79.2	1.010
SO2 Scrubber Check	4923	77.1	771.0	-0.2	----

Date of last scrubber change:	17-Dec-21	Ave Corr Factor	0.998
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 80.1 Prev response: 80.38 *% change: -0.4%
 Baseline Corr 2nd AF pt: 39.4 AF Slope: 1.003725 AF Intercept: -0.238924
 Baseline Corr 3rd AF pt: 19.3 AF Correlation: 0.999860

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Scrubber check completed after calibrator zero. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 24, 2023	Previous Calibration:	July 19, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:52	End Time (MST):	15:12
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999987	≥0.995
80.0	79.8	1.0024			
40.0	40.2	0.9948	Slope	0.993582	0.90 - 1.10
20.0	20.1	0.9974			
			Intercept	0.341175	+/-3



TRS Calibration Plot

Date: August 24, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	August 10, 2023	Last Cal Date:	July 18, 2023
Start time (MST):	10:25	End time (MST):	13:58
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.21E-04	3.26E-04	NMHC SP Ratio:	6.21E-05
CH ₄ Retention time:	15.20	15.40	NMHC Peak Area:	147573
				133711

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919	81.0	17.28	16.21	1.066
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4919	81.0	17.28	17.18	1.006
second point	4959	40.5	8.64	8.62	1.002
third point	4979	20.2	4.31	4.28	1.008
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.22	1.003

				Average Correction Factor	1.005
Baseline Corr AF:	16.20	Prev response	17.25	*% change	-6.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	8.25	1.112
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.08	1.010
second point	4959	40.5	4.58	4.58	1.002
third point	4979	20.2	2.29	2.27	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81	9.17	9.10	1.007
Average Correction Factor					1.006
Baseline Corr AF:	8.25	Prev response	9.15	*% change	-10.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919	81.0	8.11	7.97	1.018
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4919	81.0	8.11	8.10	1.002
second point	4959	40.5	4.06	4.04	1.003
third point	4979	20.2	2.02	2.01	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.12	0.999
Average Correction Factor					1.004
Baseline Corr AF:	7.96	Prev response	8.11	*% change	-1.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998376	0.993729
THC Cal Offset:	0.001211	0.010586
CH ₄ Cal Slope:	0.999482	0.997581
CH ₄ Cal Offset:	-0.003013	0.000983
NMHC Cal Slope:	0.997360	0.990059
NMHC Cal Offset:	0.003625	0.010402

Notes: NM low, chromatograms and diagnostics looks fine. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 18, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:25	End Time (MST):	13:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.01	----	Correlation Coefficient	0.999995	≥ 0.995			
17.28	17.18	1.0061						
8.64	8.62	1.0024				Slope	0.993729	0.90 - 1.10
4.31	4.28	1.0076						
			Intercept	0.010586	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-01-2020


Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 18, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:25	End Time (MST):	13:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999993	≥0.995
8.11	8.10	1.0017			
4.06	4.04	1.0032			
2.02	2.01	1.0072			
			Slope	0.997581	0.90 - 1.10
			Intercept	0.000983	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary

Version-01-2020


Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 18, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:25	End Time (MST):	13:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999977	≥ 0.995			
9.17	9.08	1.0103						
4.58	4.58	1.0017				Slope	0.990059	0.90 - 1.10
2.29	2.27	1.0074						
			Intercept	0.010402	± 0.5			


NMHC Calibration Curve



NMHC Calibration Plot

Date: August 10, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	August 23, 2023	Last Cal Date:	August 10, 2023
Start time (MST):	12:02	End time (MST):	13:50
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.26E-04	3.26E-04	NMHC SP Ratio:	6.89E-05
CH ₄ Retention time:	15.40	15.40	NMHC Peak Area:	133711

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	17.28	17.62	0.981
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.60	0.982
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	0.982
Baseline Corr AF:	17.62	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	9.80	0.936
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.79	0.936
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.936
Baseline Corr AF:	9.80	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	8.11	7.82	1.038
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	7.81	1.039
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.039
Baseline Corr AF:	7.82	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	N/A	N/A
THC Cal Offset:	N/A	N/A
CH ₄ Cal Slope:	N/A	N/A
CH ₄ Cal Offset:	N/A	N/A
NMHC Cal Slope:	N/A	N/A
NMHC Cal Offset:	N/A	N/A

Notes:

N2 Cylinder change


Calibration Performed By:

Aswin Sasi Kumar

NMHC Calibration Plot

Date: August 23, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	August 26, 2023	Last Cal Date:	August 10, 2023
Start time (MST):	10:49	End time (MST):	14:42
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.26E-04	3.37E-04	NMHC SP Ratio:	6.89E-05	7.14E-05
CH ₄ Retention time:	15.4	15.6	NMHC Peak Area:	133711	128380
Zero Chromatogram:	ON	ON	Flat Baseline:	ON	ON

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919	81.0	17.28	17.32	0.998
as found 2nd point	4959	40.5	8.64	8.68	0.996
as found 3rd point	4979	20.2	4.31	4.33	0.997
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.29	1.000
second point	4959	40.5	8.64	8.62	1.002
third point	4979	20.2	4.31	4.32	0.998
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.22	1.003
Average Correction Factor					1.000

Baseline Corr AF:	17.31	Prev response	17.18	*% change	0.7%
Baseline Corr 2nd AF:	8.7	AF Slope:	1.001755	AF Intercept:	0.012422
Baseline Corr 3rd AF:	4.3	AF Correlation:	0.999999	* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	9.61	0.954
as found 2nd point	4959	40.5	4.58	4.82	0.951
as found 3rd point	4979	20.2	2.29	2.41	0.948
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.17	0.999
second point	4959	40.5	4.58	4.57	1.004
third point	4979	20.2	2.29	2.30	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	9.17	9.10	1.007
Average Correction Factor					1.000
Baseline Corr AF:	9.61	Prev response	9.09	*% change	5.4%
Baseline Corr 2nd AF:	4.8	AF Slope:	1.047822	AF Intercept:	0.009525
Baseline Corr 3rd AF:	2.4	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919	81.0	8.11	7.70	1.053
as found 2nd point	4959	40.5	4.06	3.86	1.052
as found 3rd point	4979	20.2	2.02	1.91	1.058
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	8.11	1.000
second point	4959	40.5	4.06	4.05	1.001
third point	4979	20.2	2.02	2.02	1.001
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.12	0.999
Average Correction Factor					1.001
Baseline Corr AF:	7.69	Prev response	8.09	*% change	-5.2%
Baseline Corr 2nd AF:	3.85	AF Slope:	0.948921	AF Intercept:	0.003896
Baseline Corr 3rd AF:	1.90	AF Correlation:	0.999994	* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.993729	0.999923
THC Cal Offset:	0.010586	-0.001977
CH ₄ Cal Slope:	0.997581	0.999791
CH ₄ Cal Offset:	0.000983	-0.001608
NMHC Cal Slope:	0.990059	1.000027
NMHC Cal Offset:	0.010402	-0.000569

Notes: Changed the actuator after as founds. Oddly there was no CH₄ readings after this change. Had to adjust the span multiple times as it dropped after the first adjustment. Points are looking good after the second adjustment.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	August 26, 2023	Previous Calibration:	August 10, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:49	End Time (MST):	14:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
17.28	17.29	0.9998						
8.64	8.62	1.0025				Slope	0.999923	0.90 - 1.10
4.31	4.32	0.9982						
			Intercept	-0.001977	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	August 26, 2023	Previous Calibration:	August 10, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:49	End Time (MST):	14:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	≥0.995
8.11	8.11	1.0002			
4.06	4.05	1.0012			
2.02	2.02	1.0013			
			Slope	0.999791	0.90 - 1.10
			Intercept	-0.001608	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary

Version-06-2022


Station Information

Calibration Date:	August 26, 2023	Previous Calibration:	August 10, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:49	End Time (MST):	14:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999990	≥ 0.995
9.17	9.17	0.9994			
4.58	4.57	1.0039			
2.29	2.30	0.9956			
			Slope	1.000027	0.90 - 1.10
			Intercept	-0.000569	+/-0.5


NMHC Calibration Curve



NMHC Calibration Plot

Date: August 26, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: August 30, 2023 Last Cal Date: July 26, 2023
Start time (MST): 11:05 End time (MST): 16:15
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2XX7ME Cal Gas Expiry Date: January 14, 2024
NOX Cal Gas Conc: 50.48 ppm NO Cal Gas Conc: 49.22 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.48 ppm Removed Gas NO Conc: 49.22 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 2658
ZAG make/model: Teledyne API 701H Serial Number: 360

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1336160088
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.993	0.993	NO bkgnd or offset:	2.8	2.8
NOX coeff or slope:	0.984	0.984	NOX bkgnd or offset:	2.8	2.8
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	230.6	230.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998716	1.000289
NO _x Cal Offset:	-0.069984	-0.209933
NO Cal Slope:	1.001009	0.999824
NO Cal Offset:	-1.469781	-1.429814
NO ₂ Cal Slope:	1.000734	1.011795
NO ₂ Cal Offset:	-0.286696	-0.075280



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.2	----	----
as found span	4919	81.3	820.8	800.3	20.5	806.0	784.6	21.6	1.0183	1.0200
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.2	0.1	----	----
high point	4919	81.3	820.8	800.3	20.5	821.0	799.4	21.6	0.9997	1.0011
second point	4959	40.7	410.9	400.7	10.3	410.7	398.7	11.9	1.0006	1.0050
third point	4980	20.3	204.9	199.8	5.1	204.2	196.5	7.7	1.0036	1.0169
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.3	----	----
as left span	4919	81.3	820.8	387.2	433.6	817.8	381.4	436.4	1.0036	1.0151
Average Correction Factor									1.0013	1.0076

Corrected As found	NO _x = 805.7 ppb	NO = 784.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.7%
Previous Response	NO _x = 819.6 ppb	NO = 799.6 ppb		*Percent Change	NO = -1.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	791.2	378.1	433.6	439.1	0.9874	101.3%
2nd GPT point (200 ppb O3)	791.2	589.0	222.7	224.0	0.9941	100.6%
3rd GPT point (100 ppb O3)	791.2	692.9	118.8	120.7	0.9841	101.6%
Average Correction Factor					0.9886	101.2%

Notes: Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 30, 2023	Previous Calibration:	July 26, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:05	End Time (MST):	16:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
820.8	821.0	0.9997		
410.9	410.7	1.0006		
204.9	204.2	1.0036		
			0.999998	
			1.000289	
			-0.209933	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 30, 2023	Previous Calibration:	July 26, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:05	End Time (MST):	16:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	799.4	1.0011		
400.7	398.7	1.0050		
199.8	196.5	1.0169		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 30, 2023	Previous Calibration:	July 26, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:05	End Time (MST):	16:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
433.6	439.1	0.9874		
222.7	224.0	0.9941		
118.8	120.7	0.9841		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 30, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 26, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:35	End Time (MST):	14:02
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.6	----	Correlation Coefficient	≥0.995
400.0	401.8	0.9955		
200.0	200.6	0.9970	Slope	0.90 - 1.10
100.0	100.6	0.9940		
			Intercept	+/- 5



O₃ Calibration Plot

Date: August 23, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Stony Mountain Station number: AMS 18
 Calibration Date: August 30, 2023 Last Cal Date: July 26, 2023
 Start time (MST): 13:55 End time (MST): 16:00

Analyzer Make: API T640 S/N: 1162
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	24.9	24.7	24.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	696.0	697.3	696.0	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.98	5.05	4.98	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>August 30, 2023</u>	Last Cal Date: <u>July 26, 2023</u>			
	PM w/o HEPA: <u>12.4</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.0	11.1	11.1	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>12.4</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>August 30, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>August 30, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022
 Date RH/T Sensor Cleaned: August 30, 2022

Notes: Adjusted PMT, no other adjustments needed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	August 21, 2023	Last Cal Date:	July 4, 2023
Start time (MST):	11:05	End time (MST):	13:49
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,050	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	3,050	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701		Serial Number:	360

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	Start	Finish		Start	Finish
Calibration slope:	0.999550	0.999692	Backgd or Offset:	-0.010	-0.010
Calibration intercept:	0.039792	0.079759	Coeff or Slope:	0.906	0.906

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4933	66.7	40.7	40.8	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.7	40.7	1.000
second point	4966	33.3	20.3	20.5	0.990
third point	4983	16.7	10.2	10.3	0.994
as left zero	3000	0.0	0.0	0.1	----
as left span	2960	40.0	40.7	40.9	0.994
Average Correction Factor					0.994

Baseline Corr As found:	40.71	Prev response:	40.71	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 21, 2023	Previous Calibration:	July 4, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:05	End Time (MST):	13:49
Analyzer make:	API T300	Analyzer serial #:	3504

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999969
40.7	40.7	1.0000		
20.3	20.5	0.9896	Slope	0.999692
10.2	10.3	0.9939		
			Intercept	0.079759
				<i>+/-1.5</i>



CO Calibration Plot

Date: August 21, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	August 22, 2023	Last Cal Date:	July 5, 2023
Start time (MST):	10:06	End time (MST):	14:34
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2026
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
N2 Gen Make/Model:	Peak Scientific		Serial Number:	771048317

Analyzer Information

Analyzer make:	API T360	Analyzer serial #:	283
Analyzer Range	0 - 2,000 ppm		

	Start	Finish		Start	Finish
Calibration slope:	0.999115	0.994404	Backgd or Offset:	-0.081	-0.103
Calibration intercept:	-4.820000	2.640000	Coeff or Slope:	1.091	1.109

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	2.4	----
as found span	2920	80.0	1605.9	1611.4	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	3.1	----
high point	2920	80.0	1605.9	1599.6	1.004
second point	2960	40.0	802.9	801.3	1.002
third point	2980	20.0	401.5	401.1	1.001
as left zero	3000	0.0	0.0	2.6	----
as left span	2930	80.0	1600.5	1602.0	0.999
Average Correction Factor					1.002

Baseline Corr As found:	1609.00	Prev response:	1599.63	*% change:	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Zero and span adjusted. Third point failed, performed linearity adjustment.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	August 22, 2023	Previous Calibration	July 5, 2023
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:06	End Time (MST)	14:34
Analyzer make	API T360	Analyzer serial #	283

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	3.1	----	Correlation Coefficient	0.999999	≥0.995
1605.9	1599.6	1.0039			
802.9	801.3	1.0020	Slope	0.994404	0.90 - 1.10
401.5	401.1	1.0009			
			Intercept	2.640000	+/-10



CO₂ Calibration Plot

Date: August 22, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	August 31, 2023	Previous Calibration	N/A
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	11:54	End Time (MST)	15:55
Analyzer make	API T360	Analyzer serial #	288

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	2.4	----	Correlation Coefficient	0.999988	≥0.995
1605.9	1611.1	0.9968			
802.9	803.5	0.9993	Slope	1.002574	0.90 - 1.10
401.5	399.7	1.0044			
			Intercept	-0.200000	+/-10



CO₂ Calibration Plot

Date: August 31, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Stony Mountain	Station Number:	AMS 18
Calibration Date:	August 24, 2023	Prev Cal Date:	August 4, 2022
Start Time (MST):	12:40	End Time (MST):	12:51
Tower Height (m):	20.0	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	W23536
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.5%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	<i>≥0.9995</i>
Calculated slope		0.999594	<i>0.90 - 1.10</i>
Calculated intercept		-0.028293	<i>+/- 2</i>

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	D14057
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:27	Calc Declination*:	13.2 Degrees
Deadband calc:	1.1 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	-0.4	---
90	87.7	-0.7%
180	179.3	-0.2%
270	269.7	-0.1%
357	355.5	-0.4%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999982	<i>≥0.9995</i>
Calculated slope		1.000171	<i>0.90 - 1.10</i>
Calculated intercept		1.026670	<i>+/- 4</i>

Notes: Installing new WS/WD sensors on the 20m tower. Crossarm aligned with true north. No issues to note.

Calibration Performed By: Rene Chamberland & Devin Russell



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Stony Mountain	Station Number:	AMS 18
Calibration Date:	August 24, 2023	Prev Cal Date:	August 4, 2022
Start Time (MST):	13:38	End Time (MST):	14:00
Tower Height (m):	20.0	Reason:	Removal

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	A1406
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.5	-0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999999	0.999999	≥0.9995
Calculated slope	0.998857	0.999473	0.90 - 1.10
Calculated intercept	0.029838	0.026227	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	10478
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:27	Calc Declination*:	13.2 Degrees
Deadband calc:	-10.9 degrees (<i>Limit 4 deg</i>)	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.9	---
90	85.8	-1.2%
180	176.6	-1.0%
270	279.6	2.7%
357	368.8	3.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999990	0.999467	≥0.9995
Calculated slope	1.000134	0.960823	0.90 - 1.10
Calculated intercept	-1.328101	4.207460	+/- 4

Notes: Removing old 20m WS/WD sensors. WD sensor outside of limits.

Calibration Performed By: Rene Chamberland & Devin Russell




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS19
FIREBAG
AUGUST 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023







Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Firebag Station number: AMS 19
Calibration Date: August 14, 2023 Last Cal Date: July 19, 2023
Start time (MST): 11:00 End time (MST): 14:15
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.29 ppm Cal Gas Exp Date: February 23, 2025
Cal Gas Cylinder #: CC716618
Removed Cal Gas Conc: 49.29 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 1607
ZAG Make/Model: API T701 Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 1410661308
Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995832	1.001265	Backgd or Offset: 10.2	10.3
Calibration intercept:	-0.101381	0.897402	Coeff or Slope: 0.992	1.002

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	-0.1	----
as found span	4919	81.1	799.5	791.9	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.0	0.2	----
high point	4919	81.1	799.5	801.0	0.998
second point	4959	40.6	400.3	402.1	0.995
third point	4980	20.3	200.1	201.9	0.991
as left zero	4999	0.0	0.0	0.1	----
as left span	4919	81.1	799.5	801.0	0.998
Average Correction Factor					0.995

Baseline Corr As found:	792.00	Previous response	796.03	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 14, 2023	Previous Calibration:	July 19, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:00	End Time (MST):	14:15
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient	≥0.995
799.5	801.0	0.9981		
400.3	402.1	0.9954	Slope	0.90 - 1.10
200.1	201.9	0.9911		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 14, 2023

Location: Firebag





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Firebag Station number: AMS19
 Calibration Date: August 10, 2023 Last Cal Date: July 18, 2023
 Start time (MST): 10:15 End time (MST): 14:56
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.114 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517427
 Removed Cal Gas Conc: 5.114 ppm Rem Gas Exp Date: n/a
 Removed Gas Cyl #: n/a Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1607
 ZAG Make/Model: Teledyne API T701 Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680032
 Converter make: Global Converter serial #: 2022-222
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004340	1.000053	Backgd or Offset:	2.54 2.55
Calibration intercept:	-0.181643	-0.081596	Coeff or Slope:	1.122 1.122

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4922	78.2	80.0	80.4	0.991
as found 2nd point	4961	39.1	40.0	40.0	0.992
as found 3rd point	4980	19.6	20.0	19.9	0.992
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4922	78.2	80.0	79.9	1.001
second point	4961	39.1	40.0	39.9	1.002
third point	4980	19.6	20.0	20.0	1.002
as left zero	5000	0.0	0.0	-0.1	----
as left span	4922	78.2	80.0	79.4	1.007
SO2 Scrubber Check	4922	78.3	800.2	0.0	----
Date of last scrubber change:	January 18, 2023			Ave Corr Factor	1.002
Date of last converter efficiency test:	n/a			efficiency	

Baseline Corr As found: 80.7 Prev response: 80.15 *% change: 0.7%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 1.009057 AF Intercept: -0.321768
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999999

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter and ran SO_x scrubber check after MPAF's. No adjustments needed.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 18, 2023
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	10:15	End Time (MST):	14:56
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	1.000000	≥0.995
80.0	79.9	1.0010			
40.0	39.9	1.0023	Slope	1.000053	0.90 - 1.10
20.0	20.0	1.0024			
			Intercept	-0.081596	+/-3



H₂S Calibration Plot

Date: August 10, 2023

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	August 14, 2023	Last Cal Date:	July 19, 2023
Start time (MST):	11:00	End time (MST):	14:15
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1607
ZAG Make/Model:	API T701	Serial Number:	1118

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1336160089
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999589	0.998012	Background:	3.75	3.11
Calibration intercept:	-0.108328	-0.083128	Coefficient:	3.716	3.765

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.16	----
as found span	4919	81.1	17.31	17.45	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.00	-0.05	----
high point	4919	81.1	17.31	17.22	1.005
second point	4959	40.6	8.66	8.50	1.019
third point	4980	20.3	4.33	4.24	1.022
as left zero	5000	0.0	0.00	-0.06	----
as left span	4919	81.1	17.31	17.34	0.998
Average Correction Factor					1.015
Baseline Corr As found:	17.29	Previous response	17.19	*% change	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 14, 2023	Previous Calibration:	July 19, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:00	End Time (MST):	14:15
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.05	----	Correlation Coefficient	0.999962	≥0.995
17.31	17.22	1.0049			
8.66	8.50	1.0193	Slope	0.998012	0.90 - 1.10
4.33	4.24	1.0221			
			Intercept	-0.083128	+/-1.5



THC Calibration Plot

Date: August 14, 2023

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	August 28, 2023	Last Cal Date:	August 14, 2023
Start time (MST):	9:30	End time (MST):	13:08
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1607
ZAG Make/Model:	API T701	Serial Number:	1118

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1336160089
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998012	1.005644	Background:	3.11	3.78
Calibration intercept:	-0.083128	-0.153944	Coefficient:	3.765	3.779

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.00	-0.07	----
high point	4919	81.1	17.31	17.30	1.000
second point	4959	40.6	8.66	8.49	1.021
third point	4980	20.3	4.33	4.14	1.047
as left zero	5000	0.0	0.00	-0.21	----
as left span	4919	81.1	17.31	17.23	1.004
Average Correction Factor					1.023
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Calibrated after instrument crashed and was offline since August 26. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 28, 2023	Previous Calibration:	August 14, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:30	End Time (MST):	13:08
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	-0.07	----	Correlation Coefficient	0.999887	≥0.995
17.31	17.30	1.0003			
8.66	8.49	1.0205	Slope	1.005644	0.90 - 1.10
4.33	4.14	1.0473			
			Intercept	-0.153944	+/-1.5



THC Calibration Plot

Date: August 28, 2023

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	August 29, 2023	Last Cal Date:	August 28, 2023
Start time (MST):	10:00	End time (MST):	13:10
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1607
ZAG Make/Model:	API T701	Serial Number:	1118

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1336160089
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005644	0.994308	Background:	3.78	3.27
Calibration intercept:	-0.153944	-0.010322	Coefficient:	3.779	3.762

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.00	-0.48	----
as found span	4919	81.1	17.31	16.82	1.029
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.00	0.03	----
high point	4919	81.1	17.31	17.22	1.005
second point	4959	40.6	8.66	8.57	1.011
third point	4980	20.3	4.33	4.27	1.014
as left zero	5000	0.0	0.00	-0.01	----
as left span	4919	81.1	17.31	17.30	1.000
Average Correction Factor					1.010
Baseline Corr As found:	17.30	Previous response	17.25	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Calibrated due to baseline drift. Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 29, 2023	Previous Calibration:	August 28, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:00	End Time (MST):	13:10
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.03	----	Correlation Coefficient	0.999977	≥0.995
17.31	17.22	1.0049			
8.66	8.57	1.0110	Slope	0.994308	0.90 - 1.10
4.33	4.27	1.0142			
			Intercept	-0.010322	+/-1.5



THC Calibration Plot

Date: August 29, 2023

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Firebag
Calibration Date: August 16, 2023
Start time (MST): 10:25
Reason: Routine
Station number: AMS 19
Last Cal Date: July 27, 2023
End time (MST): 14:52

Calibration Standards

NO Gas Cylinder #: T2Y1K63
NOX Cal Gas Conc: 51.12 ppm
Removed Cylinder #: n/a
Removed Gas NOX Conc: 51.12 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 30, 2023
NO Cal Gas Conc: 49.40 ppm
Removed Gas Exp Date: n/a
Removed Gas NO Conc: 49.40 ppm
NO gas Diff:
Serial Number: 1607
Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1410661309

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.106	1.112	NO bkgnd or offset:	7.7	7.8
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	7.7	7.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	214.0	213.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996456	0.998001
NO _x Cal Offset:	1.235903	1.176081
NO Cal Slope:	0.999440	0.999755
NO Cal Offset:	0.208843	0.048770
NO ₂ Cal Slope:	0.999338	1.004673
NO ₂ Cal Offset:	0.066857	0.372474



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.0	0.0	0.1	-0.2	0.3	----	----
as found span	4919	81.0	828.1	800.3	27.9	820.0	790.6	29.2	1.0099	1.0122
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	4999	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
high point	4919	81.0	828.1	800.3	27.9	827.0	800.0	26.6	1.0014	1.0004
second point	4960	40.5	414.0	400.1	13.9	415.3	400.3	15.0	0.9969	0.9995
third point	4980	20.2	206.5	199.6	6.9	208.1	199.6	8.5	0.9924	0.9998
as left zero	4999	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
as left span	4919	81.0	828.1	365.8	462.4	837.0	366.4	470.5	0.9894	0.9983
Average Correction Factor									0.9969	0.9999

Corrected As found	NO _x = 819.9 ppb	NO = 790.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.8%	
Previous Response	NO _x = 826.4 ppb	NO = 800.0 ppb		*Percent Change	NO = -1.2%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	800.0	365.5	462.4	465.2	0.9939	100.6%
2nd GPT point (200 ppb O3)	800.0	583.0	244.9	245.3	0.9982	100.2%
3rd GPT point (100 ppb O3)	800.0	696.7	131.2	133.1	0.9855	101.5%
Average Correction Factor					0.9925	100.8%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	August 16, 2023	Previous Calibration:	July 27, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:25	End Time (MST):	14:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
828.1	827.0	1.0014			
414.0	415.3	0.9969			
206.5	208.1	0.9924			
			Slope	0.998001	0.90 - 1.10
			Intercept	1.176081	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 16, 2023	Previous Calibration:	July 27, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:25	End Time (MST):	14:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	1.000000	≥0.995
800.3	800.0	1.0004			
400.1	400.3	0.9995	Slope	0.999755	0.90 - 1.10
199.6	199.6	0.9998			
			Intercept	0.048770	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	August 16, 2023	Previous Calibration:	July 27, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:25	End Time (MST):	14:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient 0.999981	≥0.995	
462.4	465.2	0.9939			
244.9	245.3	0.9982			
131.2	133.1	0.9855			
			Slope	1.004673	0.90 - 1.10
			Intercept	0.372474	+/-20



NO_x Calibration Plot

Date: August 16, 2023

Location: Firebag






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS20 MACKAY RIVER AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 8, 2023	Previous Calibration:	July 6, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:34	End Time (MST):	11:14
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.4	----	Correlation Coefficient	0.999963	
800.3	800.4	0.9998			≥0.995
400.7	402.0	0.9967	Slope	0.997865	
199.8	204.7	0.9762			0.90 - 1.10
			Intercept	2.431247	+/-30



SO2 Calibration Plot

Date: August 8, 2023

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	MacKay River	Station number:	AMS20
Calibration Date:	August 4, 2023	Last Cal Date:	July 5, 2023
Start time (MST):	6:20	End time (MST):	10:23
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.12	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC515997			
Removed Cal Gas Conc:	5.12	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1220
ZAG Make/Model:	Teledyne API 701		Serial Number:	4522

Analyzer Information

Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	12124313139
Converter make:	Global	Converter serial #:	2022-226
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996518	0.994521	Backgd or Offset:	1.0
Calibration intercept:	0.267198	0.287104	Coeff or Slope:	0.571
				0.97
				0.535

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4922	78.1	80.0	86.0	0.930
as found 2nd point	4961	39.0	39.9	43.1	0.927
as found 3rd point	4980	19.5	20.0	21.6	0.925
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4922	78.1	80.0	79.8	1.003
second point	4961	39.0	40.0	40.1	0.997
third point	4980	19.5	20.0	20.4	0.981
as left zero	5000	0.0	0.0	0.3	----
as left span	4922	78.1	80.0	79.6	1.005
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:	May 25, 2023			Ave Corr Factor	0.994
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	86.0	Prev response:	79.96	*% change:	7.0%
Baseline Corr 2nd AF pt:	43.1	AF Slope:	1.075049	AF Intercept:	0.080869
Baseline Corr 3rd AF pt:	21.6	AF Correlation:	0.999995		

* = > +/-5% change initiates investigation

Notes: Diagnostics similar to last month. Sox Scrubber check completed after calibrator zero. Span Adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 4, 2023	Previous Calibration:	July 5, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	6:20	End Time (MST):	10:23
Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	12124313139

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
80.0	79.8	1.0029		
40.0	40.1	0.9967	Slope	0.90 - 1.10
20.0	20.4	0.9812		
			Intercept	+/-3



H₂S Calibration Plot

Date: August 4, 2023

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name: MacKay River Station number: AMS20
Calibration Date: August 8, 2023 Last Cal Date: July 6, 2023
Start time (MST): 8:34 End time (MST): 11:13
Reason: Routine

Calibration Standards

Gas Cert Reference: CC306868 Cal Gas Expiry Date: February 23, 2025
CH4 Cal Gas Conc. 499.40 ppm CH4 Equiv Conc. 1066.45 ppm
C3H8 Cal Gas Conc. 206.20 ppm
Removed Gas Cert: NA Removed Gas Expiry: NA
Removed CH4 Conc. 499.40 ppm CH4 Equiv Conc. 1066.45 ppm
Removed C3H8 Conc. 206.20 ppm Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 1220
ZAG Make/Model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1501663727
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.989494	0.992565	Background:	3.040	3.090
Calibration intercept:	0.030619	0.123818	Coefficient:	5.384	5.465

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.09	----
as found span	4919	81.3	17.34	16.94	1.024
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.07	----
high point	4919	81.3	17.34	17.32	1.001
second point	4959	40.7	8.68	8.73	0.994
third point	4980	20.3	4.33	4.50	0.962
as left zero	5000	0.0	0.00	0.07	----
as left span	4919	81.3	17.34	17.18	1.009
Average Correction Factor					0.986
Baseline Corr As found:	16.85	Previous response	17.19	*% change	-2.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 8, 2023	Previous Calibration:	July 6, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:34	End Time (MST):	11:13
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.07	----	Correlation Coefficient	0.999942	≥0.995
17.34	17.32	1.0011			
8.68	8.73	0.9944	Slope	0.992565	0.90 - 1.10
4.33	4.50	0.9621			
			Intercept	0.123818	+/-1.5



THC Calibration Plot

Date: August 8, 2023

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: MacKay River Station number: AMS20
Calibration Date: August 3, 2023 Last Cal Date: July 4, 2023
Start time (MST): 7:12 End time (MST): 11:22
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 1220
ZAG make/model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.503	1.503	NO bkgnd or offset:	4.2	4.2
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	4.2	4.2
NO2 coeff or slope:	0.995	0.995	Reaction cell Press:	182.5	182.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.990458	0.986386
NO _x Cal Offset:	2.589991	3.150520
NO Cal Slope:	0.991438	0.988909
NO Cal Offset:	1.451137	1.711850
NO ₂ Cal Slope:	0.998430	1.003108
NO ₂ Cal Offset:	-0.419338	1.350779



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.1	----	----
as found span	4917	83.3	819.5	800.3	19.2	813.8	794.0	19.9	1.0070	1.0079
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.2	----	----
high point	4917	83.3	819.5	800.3	19.2	810.4	792.8	17.7	1.0112	1.0095
second point	4956	41.7	410.4	400.8	9.6	408.2	397.5	10.7	1.0055	1.0084
third point	4979	20.8	204.6	199.9	4.8	208.8	201.9	6.8	0.9801	0.9899
as left zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.2	----	----
as left span	4917	83.3	819.5	447.1	372.4	814.7	438.1	376.6	1.0058	1.0205
Average Correction Factor									0.9989	1.0026

Corrected As found	NO _x = 813.7 ppb	NO = 794.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.1%	
Previous Response	NO _x = 814.2 ppb	NO = 794.9 ppb		*Percent Change	NO = -0.1%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	788.3	435.1	372.4	374.5	0.9943	100.6%
2nd GPT point (200 ppb O3)	788.3	606.4	201.1	202.7	0.9919	100.8%
3rd GPT point (100 ppb O3)	788.3	697.6	109.9	113.4	0.9688	103.2%
Average Correction Factor					0.9850	101.5%

Notes:

No adjustments and maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 4, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	7:12	End Time (MST):	11:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
819.5	810.4	1.0112		
410.4	408.2	1.0055		
204.6	208.8	0.9801		

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 4, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	7:12	End Time (MST):	11:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.3	792.8	1.0095			
400.8	397.5	1.0084			
199.9	201.9	0.9899			
			Slope	0.988909	0.90 - 1.10
			Intercept	1.711850	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 4, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:12	End Time (MST):	11:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999933	≥0.995
372.4	374.5	0.9943			
201.1	202.7	0.9919			
109.9	113.4	0.9688			
			Slope	1.003108	0.90 - 1.10
			Intercept	1.350779	+/-20



NO_x Calibration Plot

Date: August 3, 2023

Location: MacKay River






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS21
CONKLIN
AUGUST 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	August 3, 2023	Last Cal Date:	July 13, 2023
Start time (MST):	9:15	End time (MST):	12:25
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.93	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	<u>CC259455</u>			
Removed Cal Gas Conc:	49.93	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3810
ZAG Make/Model:	Teledyne API 701		Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997605	1.006909	Backgd or Offset:	27.2	26.8
Calibration intercept:	0.415762	0.556149	Coeff or Slope:	0.883	0.883

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5005	0.0	0.0	-0.2	----
as found span	4920	80.2	800.8	798.6	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5005	0.0	0.0	0.0	----
high point	4920	80.2	800.8	807.0	0.992
second point	4960	40.1	400.4	403.0	0.994
third point	4980	20.0	200.1	203.3	0.984
as left zero	5005	0.0	0.0	0.2	----
as left span	4920	80.2	800.8	805.0	0.995
Average Correction Factor					0.990

Baseline Corr As found:	798.80	Previous response	799.34	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Adjusted the zero.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 13, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:15	End Time (MST):	12:25
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999993	
800.8	807.0	0.9924			≥0.995
400.4	403.0	0.9936	Slope	1.006909	
200.1	203.3	0.9843			0.90 - 1.10
			Intercept	0.556149	+/-30



SO2 Calibration Plot

Date: August 3, 2023

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Conklin Station number: AMS21
 Calibration Date: August 24, 2023 Last Cal Date: July 27, 2023
 Start time (MST): 9:15 End time (MST): 13:05
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.00 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC501204
 Removed Cal Gas Conc: 5.03 ppm Rem Gas Exp Date: April 16, 2022
 Removed Gas Cyl #: CC505493 Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3810
 ZAG Make/Model: API 701H Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1236656116
 Converter make: CD-Nova 101 Converter serial #: NA
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007143	1.004857	Backgd or Offset: 2.4	2.4
Calibration intercept:	0.200000	0.280000	Coeff or Slope: 0.958	0.958

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.0	80.5	80.0	1.006
as found 2nd point	4960	40.0	40.2	40.3	0.999
as found 3rd point	4980	20.0	20.1	20.2	0.996
new cylinder response	4920	80.0	80.0		

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.0	80.0	80.5	0.994
second point	4960	40.0	40.0	40.8	0.980
third point	4980	20.0	20.0	20.4	0.980
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	80.0	80.0	79.0	1.013
SO2 Scrubber Check	4920	80.2	802.0	0.0	----

Date of last scrubber change:	Ave Corr Factor	0.985
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 80.0 Prev response: 81.25 *% change: -1.6%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 0.993610 AF Intercept: 0.140000
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999981

* = > +/-5% change initiates investigation

Notes: No adjustments made.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 24, 2023	Previous Calibration:	July 27, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:15	End Time (MST):	13:05
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999953	
80.0	80.5	0.9938			≥0.995
40.0	40.8	0.9804	Slope	1.004857	
20.0	20.4	0.9804			0.90 - 1.10
			Intercept	0.280000	+/-3



TRS Calibration Plot

Date: August 24, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	August 3, 2023	Last Cal Date:	July 13, 2023
Start time (MST):	9:15	End time (MST):	12:25
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.27E-04	2.29E-04	NMHC SP Ratio:	5.05E-05
CH ₄ Retention time:	12.00	12.00	NMHC Peak Area:	180824
				179681

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.08	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.10	1.001
second point	4960	40.1	8.56	8.53	1.004
third point	4980	20.0	4.27	4.29	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.18	0.997
Average Correction Factor					1.000
Baseline Corr AF:	17.08	Prev response	17.12	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.11	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.13	1.001
second point	4960	40.1	4.57	4.57	0.999
third point	4980	20.0	2.28	2.30	0.990
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.15	0.999
Average Correction Factor					0.997
Baseline Corr AF:	9.11	Prev response	9.14	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.97	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.98	1.001
second point	4960	40.1	3.99	3.96	1.010
third point	4980	20.0	1.99	1.99	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	8.03	0.995
Average Correction Factor					1.005
Baseline Corr AF:	7.97	Prev response	7.98	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999162	0.997914
THC Cal Offset:	0.004574	0.005170
CH ₄ Cal Slope:	0.999695	0.998322
CH ₄ Cal Offset:	-0.004854	-0.008057
NMHC Cal Slope:	0.998696	0.997658
NMHC Cal Offset:	0.009428	0.012827

Notes: Replaced H2 cylinder. No adjustments made.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 13, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:15	End Time (MST):	12:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995
17.13	17.10	1.0015			
8.56	8.53	1.0040			
4.27	4.29	0.9960			
			Slope	0.997914	0.90 - 1.10
			Intercept	0.005170	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 13, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:15	End Time (MST):	12:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999979	≥0.995
7.99	7.98	1.0014			
3.99	3.96	1.0096			
1.99	1.99	1.0033			
			Slope	0.998322	0.90 - 1.10
			Intercept	-0.008057	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 13, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:15	End Time (MST):	12:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999990	≥ 0.995			
9.14	9.13	1.0014						
4.57	4.57	0.9993				Slope	0.997658	0.90 - 1.10
2.28	2.30	0.9897						
			Intercept	0.012827	± 0.5			



NMHC Calibration Plot

Date: August 3, 2023

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Conklin
Calibration Date: August 10, 2023
Start time (MST): 8:53
Reason: Routine
Station number: AMS21
Last Cal Date: July 28, 2023
End time (MST): 12:50

Calibration Standards

NO Gas Cylinder #: T2Y1P1H
NOX Cal Gas Conc: 51.09 ppm
Removed Cylinder #: n/a
Removed Gas NOX Conc: 51.09 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.39 ppm
Removed Gas Exp Date: n/a
Removed Gas NO Conc: 50.39 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 691

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1501663731

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.115	1.120	NO bkgnd or offset:	11.2	11.2
NOX coeff or slope:	1.000	1.000	NOX bkgnd or offset:	11.4	11.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	203.7	196.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998617	0.998758
NO _x Cal Offset:	2.003897	1.603828
NO Cal Slope:	0.999267	0.999738
NO Cal Offset:	0.881047	0.741061
NO ₂ Cal Slope:	0.998137	1.000498
NO ₂ Cal Offset:	-0.798489	0.217772



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
as found span	4921	79.4	811.2	800.1	11.1	809.0	796.7	12.4	1.0028	1.0043
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
high point	4921	79.4	811.2	800.1	11.1	811.0	800.3	10.7	1.0003	0.9998
second point	4960	39.7	405.7	400.1	5.6	407.6	401.1	6.5	0.9953	0.9976
third point	4980	19.8	202.3	199.6	2.8	205.4	201.0	4.3	0.9850	0.9928
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as left span	4921	79.4	811.2	365.7	445.5	812.6	374.4	438.3	0.9983	0.9768
Average Correction Factor									0.9935	0.9967

Corrected As found	NO _x = 809.3 ppb	NO = 796.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.3%	
Previous Response	NO _x = 812.1 ppb	NO = 800.4 ppb		*Percent Change	NO = -0.4%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI: ;	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.3	361.9	445.5	445.8	0.9994	100.1%
2nd GPT point (200 ppb O3)	796.3	593.2	214.2	214.7	0.9977	100.2%
3rd GPT point (100 ppb O3)	796.3	696.7	110.7	111.3	0.9947	100.5%
Average Correction Factor					0.9973	100.3%

Notes:

Adjusted the span only.

Calibration Performed By:

Denny Ray Estador



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 28, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:53	End Time (MST):	12:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
811.2	811.0	1.0003		
405.7	407.6	0.9953		
202.3	205.4	0.9850		
			0.999980	
			0.998758	
			1.603828	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 28, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:53	End Time (MST):	12:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.1	800.3	0.9998		
400.1	401.1	0.9976		
199.6	201.0	0.9928		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 28, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:53	End Time (MST):	12:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
445.5	445.8	0.9994			
214.2	214.7	0.9977			
110.7	111.3	0.9947			
			Slope	1.000498	0.90 - 1.10
			Intercept	0.217772	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 10, 2023

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 15, 2023	Previous Calibration:	July 24, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:09	End Time (MST):	12:45
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999946	≥0.995
400.0	401.0	0.9975			
200.0	202.3	0.9886	Slope	1.000257	0.90 - 1.10
100.0	103.0	0.9709			
			Intercept	1.580000	+/- 5



O₃ Calibration Plot

Date: August 15, 2023

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Conklin Station number: AMS 21
 Calibration Date: August 17, 2023 Last Cal Date: July 27, 2023
 Start time (MST): 9:56 End time (MST): 10:37

Analyzer Make: API T640X S/N: 1597
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat S/N: 388751
 Temp/RH standard: Alicat S/N: 388751

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	22.5	22.5	22.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	699.9	702.5	699.9	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.05	5.03	5.05	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: August 17, 2023 Last Cal Date: July 27, 2023
 PM w/o HEPA: 16.3 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11	11.2	11	<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: 2.8 w/ HEPA: 0
 Date Optical Chamber Cleaned: August 17, 2023 <0.2 ug/m3
 Disposable Filter Changed: August 17, 2023

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: No adjustments made for both monthly and quarterly maintenance.

Calibration by: Denny Ray Estador



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Conklin	Station Number:	AMS 21
Calibration Date:	August 30, 2023	Prev Cal Date:	July 22, 2022
Start Time (MST):	10:00	End Time (MST):	10:45
Tower Height (m):	10.0	Reason:	Routine

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	J4337
WS Calibrator:	MetOne 053-120	Serial Number:	R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.2	0.2%
400	39.4	39.4	0.1%
600	58.6	58.5	-0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥0.9995
Calculated slope		0.999979	0.90 - 1.10
Calculated intercept		-0.013577	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	P22886
As Found Declination (deg east of True North):	<u>15</u>	As Left Declination (deg east of True North):	<u>13</u>
Solar noon time (MST):	13:25	Calc Declination*:	13 Degrees
Deadband calc:	-1.4 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	-0.2	---
90	89.1	-0.3%
180	179.4	-0.2%
270	269.0	-0.3%
357	358.2	0.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999985	≥0.9995
Calculated slope		0.996994	0.90 - 1.10
Calculated intercept		0.838364	+/- 4

Notes:

Bearings still good. Confirmed declination with a compass.

Calibration Performed By:

Denny Ray Estador




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER
AUGUST 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 17, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:00	End Time (MST):	13:26
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.1	----	Correlation Coefficient	0.999996	≥0.995
799.8	805.4	0.9930			
399.9	404.3	0.9891	Slope	1.005224	0.90 - 1.10
200.4	204.0	0.9825			
			Intercept	1.843010	+/-30



SO2 Calibration Plot

Date: August 23, 2023

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Janvier Station number: AMS22
 Calibration Date: August 8, 2023 Last Cal Date: July 25, 2023
 Start time (MST): 9:35 End time (MST): 13:52
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.03 ppm Cal Gas Exp Date: April 16, 2022
 Cal Gas Cylinder #: DT0018680
 Removed Cal Gas Conc: 5.03 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 1200326169
 Converter make: CDN-101 Converter serial #: 587
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003363	1.004084	Backgd or Offset: 0.96	1.05
Calibration intercept:	0.020959	0.420700	Coeff or Slope: 0.880	0.965

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	79.5	80.0	73.0	1.096
as found 2nd point	4960	39.8	40.0	35.3	1.134
as found 3rd point	4980	19.9	20.0	17.4	1.151
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	79.5	80.0	80.5	0.994
second point	4960	39.8	40.0	41.0	0.977
third point	4980	19.9	20.0	20.7	0.967
as left zero	5000	0.0	0.0	0.3	----
as left span	4920	79.5	80.0	82.7	0.967
SO2 Scrubber Check	4920	79.8	798.0	0.0	----

Date of last scrubber change:	Ave Corr Factor	0.979
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 73.0 Prev response: 80.27 *% change: -10.0%
 Baseline Corr 2nd AF pt: 35.3 AF Slope: 0.914618 AF Intercept: -0.596957
 Baseline Corr 3rd AF pt: 17.4 AF Correlation: 0.999599

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. Adjusted the span only. Diagnostics are normal, suspecting the change in humidity has caused the shift in span since last calibration.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 8, 2023	Previous Calibration:	July 25, 2023
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	9:35	End Time (MST):	13:52
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326169

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	0.999907
80.0	80.5	0.9936		
40.0	41.0	0.9766	Slope	1.004084
20.0	20.7	0.9671		
			Intercept	0.420700
				+/-3



TRS Calibration Plot

Date: August 8, 2023

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Janvier Station number: AMS22
 Calibration Date: August 29, 2023 Last Cal Date: August 8, 2023
 Start time (MST): 9:46 End time (MST): 14:23
 Reason: Maintenance

Calibration Standards

Cal Gas Concentration: 5.03 ppm Cal Gas Exp Date: April 16, 2022
 Cal Gas Cylinder #: DT0018680
 Removed Cal Gas Conc: 5.03 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 1200326169
 Converter make: CDN-101 Converter serial #: 587
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004084	1.004358	Backgd or Offset:	1.05 1.01
Calibration intercept:	0.420700	-0.338889	Coeff or Slope:	0.965 0.926

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	79.5	80.0	78.8	1.015
as found 2nd point	4960	39.8	40.0	40.1	0.999
as found 3rd point	4980	19.9	20.0	19.6	1.021
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	79.5	80.0	80.2	0.997
second point	4960	39.8	40.0	39.7	1.009
third point	4980	19.9	20.0	19.3	1.037
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	79.5	80.0	80.0	1.000
SO2 Scrubber Check	4920	79.8	798.0	0.0	----

Date of last scrubber change:	Ave Corr Factor	1.014
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 78.8 Prev response: 80.73 *% change: -2.5%
 Baseline Corr 2nd AF pt: 40.1 AF Slope: 0.986651 AF Intercept: 0.081062
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999893

* = > +/-5% change initiates investigation

Notes: Daily spans have been getting lower every day and the response has been very slow. Changed the scrubber beads after multipoint as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 29, 2023	Previous Calibration:	August 8, 2023
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	9:46	End Time (MST):	14:23
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326169

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
80.0	80.2	0.9973		
40.0	39.7	1.0086	Slope	0.90 - 1.10
20.0	19.3	1.0373		
			Intercept	+/-3



TRS Calibration Plot

Date: August 29, 2023

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	August 23, 2023	Last Cal Date:	July 17, 2023
Start time (MST):	10:00	End time (MST):	13:27
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH ₄ Cal Gas Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
C ₃ H ₈ Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
Removed C ₃ H ₈ Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make:		Analyzer serial #:	
THC Range (ppm):	0 - 20 ppm	CH ₄ Range (ppm):	0 - 10 ppm
NMHC Range (ppm):	0 - 10 ppm		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.180E-04	2.190E-04	NMHC SP Ratio:	4.35E-05
CH ₄ Retention time:	13.4	13.6	NMHC Peak Area:	210352
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	17.11	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.12	1.003
second point	4960	39.9	8.59	8.54	1.006
third point	4980	20.0	4.30	4.29	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.24	0.996
Average Correction Factor					1.004

Baseline Corr AF:	17.11	Prev response	17.27	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	9.09	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.15	9.12	1.003
second point	4960	39.9	4.57	4.56	1.003
third point	4980	20.0	2.29	2.28	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.15	9.18	0.996
Average Correction Factor					1.004
Baseline Corr AF:	9.09	Prev response	9.13	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	8.02	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.01	1.002
second point	4960	39.9	4.01	3.98	1.009
third point	4980	20.0	2.01	2.01	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.06	0.996
Average Correction Factor					1.004
Baseline Corr AF:	8.02	Prev response	8.14	*% change	-1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.006073	0.997041
THC Cal Offset:	-0.009020	-0.005391
CH ₄ Cal Slope:	1.014274	0.996940
CH ₄ Cal Offset:	0.000208	-0.003162
NMHC Cal Slope:	0.998891	0.997018
NMHC Cal Offset:	-0.009028	-0.002029

Notes: Changed the inlet filter and the H2 cylinder after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary

Version-06-2022


Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 17, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:00	End Time (MST):	13:27
Analyzer make:		Analyzer serial #:	

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
17.17	17.12	1.0028						
8.59	8.54	1.0058				Slope	0.997041	0.90 - 1.10
4.30	4.29	1.0029						
			Intercept	-0.005391	± 0.5			

THC Calibration Curve





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 17, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:00	End Time (MST):	13:27
Analyzer make:		Analyzer serial #:	

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999984	≥0.995
8.03	8.01	1.0025			
4.01	3.98	1.0089			
2.01	2.01	1.0001			
			Slope	0.996940	0.90 - 1.10
			Intercept	-0.003162	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 17, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:00	End Time (MST):	13:27
Analyzer make:		Analyzer serial #:	

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995			
9.15	9.12	1.0032						
4.57	4.56	1.0032				Slope	0.997018	0.90 - 1.10
2.29	2.28	1.0054						
			Intercept	-0.002029	+/-0.5			



NMHC Calibration Plot

Date: August 23, 2023

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: August 16, 2023 Last Cal Date: July 19, 2023
Start time (MST): 9:58 End time (MST): 14:36
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC424183 Cal Gas Expiry Date: April 16, 2023
NOX Cal Gas Conc: 48.60 ppm NO Cal Gas Conc: 48.60 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.60 ppm Removed Gas NO Conc: 48.60 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3806
ZAG make/model: Teledyne API T701 Serial Number: 201

Analyzer Information

Analyzer make: API T200 Analyzer serial #: 833
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.808	0.808	NO bkgnd or offset:	-5.6	-5.6
NOX coeff or slope:	0.799	0.799	NOX bkgnd or offset:	-3.9	-3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	5.0	5.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997788	0.995773
NO _x Cal Offset:	2.208515	2.689123
NO Cal Slope:	0.999902	0.994314
NO Cal Offset:	1.168443	1.949881
NO ₂ Cal Slope:	0.994081	1.001315
NO ₂ Cal Offset:	-0.360935	-0.151118



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.3	1.7	-0.4	----	----
as found span	4918	82.3	799.9	799.9	0.0	796.9	792.5	4.4	1.0038	1.0093
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	1.9	2.3	-0.4	----	----
high point	4918	82.3	799.9	799.9	0.0	798.7	797.3	1.4	1.0015	1.0033
second point	4959	41.2	400.4	400.4	0.0	402.2	400.5	1.7	0.9956	0.9999
third point	4980	20.6	200.2	200.2	0.0	202.6	200.3	2.3	0.9882	0.9995
as left zero	5000	0.0	0.0	0.0	0.0	1.9	2.7	-0.8	----	----
as left span	4918	82.3	799.9	355.2	444.7	794.9	352.0	442.9	1.0063	1.0091
Average Correction Factor									0.9951	1.0009

Corrected As found	NO _x = 795.6 ppb	NO = 790.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.6%
Previous Response	NO _x = 800.3 ppb	NO = 801.0 ppb		*Percent Change	NO = -1.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.9	350.2	444.7	444.7	1.0000	100.0%
2nd GPT point (200 ppb O3)	794.9	585.0	209.9	211.2	0.9938	100.6%
3rd GPT point (100 ppb O3)	794.9	685.7	109.2	108.7	1.0046	99.5%
Average Correction Factor					0.9995	100.1%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	August 16, 2023	Previous Calibration:	July 19, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:58	End Time (MST):	14:36
Analyzer make:	API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	1.9	----	Correlation Coefficient	≥0.995	
799.9	798.7	1.0015			
400.4	402.2	0.9956			
200.2	202.6	0.9882			
			Slope	0.995773	0.90 - 1.10
			Intercept	2.689123	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	August 16, 2023	Previous Calibration:	July 19, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:58	End Time (MST):	14:36
Analyzer make:	API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	2.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.9	797.3	1.0033		
400.4	400.5	0.9999		
200.2	200.3	0.9995		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 16, 2023	Previous Calibration:	July 19, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:58	End Time (MST):	14:36
Analyzer make:	API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	≥0.995
444.7	444.7	1.0000		
209.9	211.2	0.9938	Slope	0.90 - 1.10
109.2	108.7	1.0046		
			Intercept	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 16, 2023

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: August 18, 2023 Last Cal Date: July 26, 2023
 Start time (MST): 9:49 End time (MST): 13:29
 Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701H Serial Number: 201

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 7046
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999714	1.001114	Backgd or Offset:	-0.1	-0.1
Calibration intercept:	2.000000	2.780000	Coeff or Slope:	0.953	0.922

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.2	----
as found span	4895	904.3	400.0	414.7	0.965
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.6	----
high point	4895	904.3	400.0	402.1	0.995
second point	4895	756.7	200.0	204.2	0.979
third point	4895	656.1	100.0	105.0	0.952
as left zero	5000	800.0	0.0	0.9	----
as left span	4895	904.3	400.0	402.1	0.995
Average Correction Factor					0.976

Baseline Corr As found:	414.5	Previous response	401.9	*% change	3.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 18, 2023	Previous Calibration:	July 26, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:49	End Time (MST):	13:29
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999864	≥0.995
400.0	402.1	0.9948			
200.0	204.2	0.9794	Slope	1.001114	0.90 - 1.10
100.0	105.0	0.9524			
			Intercept	2.780000	+/- 5



O₃ Calibration Plot

Date: August 18, 2023

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: August 23, 2023 Last Cal Date: July 26, 2023
 Start time (MST): 14:00 End time (MST): 14:19

Analyzer Make: Teledyne API T640 S/N: 325
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	20.9	21	20.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	716.1	716.6	716.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	4.95	5.01	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: August 23, 2023 Last Cal Date: July 26, 2023
 PM w/o HEPA: 8.4 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: _____ w/ HEPA: _____
 Date Optical Chamber Cleaned: July 26, 2023 <0.2 ug/m3
 Disposable Filter Changed: July 26, 2023

Annual Maintenance

Date Sample Tube Cleaned: July 26, 2023
 Date RH/T Sensor Cleaned: July 26, 2023

Notes:

No adjustments needed. Leak check passed.

Calibration by: Max Farrell



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Janvier	Station Number:	AMS 22
Calibration Date:	August 24, 2023	Prev Cal Date:	September 30, 2022
Start Time (MST):	11:00	End Time (MST):	11:11
Tower Height (m):	20.0	Reason:	Removal

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	B10017
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.5%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999999	0.999999	≥0.9995
Calculated slope	0.999473	0.999594	0.90 - 1.10
Calculated intercept	0.026227	-0.028293	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	A23969
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:25	Calc Declination*:	13.02 Degrees
Deadband calc:	-2.3 degrees (<i>Limit 4 deg</i>)	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.0	---
90	86.1	-1.1%
180	176.3	-1.0%
270	273.2	0.9%
357	359.3	0.6%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999770	0.999819	≥0.9995
Calculated slope	1.004979	0.986918	0.90 - 1.10
Calculated intercept	-0.912418	2.766146	+/- 4

Notes: Removing old WS/WD sensors, WD sensor was outside of acceptable limits.

Calibration Performed By: Rene Chamberland & Devin Russell



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Janvier	Station Number:	AMS 22
Calibration Date:	August 24, 2023	Prev Cal Date:	September 30, 2022
Start Time (MST):	10:10	End Time (MST):	10:23
Tower Height (m):	20.0	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	P22393
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.5%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥ 0.9995
Calculated slope		0.999594	<i>0.90 - 1.10</i>
Calculated intercept		-0.028293	<i>+/- 2</i>

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	D14054
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:25	Calc Declination*:	13.02 Degrees
Deadband calc:	1.7 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.0	---
90	88.1	-0.5%
180	179.1	-0.2%
270	270.2	0.0%
357	355.3	-0.5%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999979	≥ 0.9995
Calculated slope		1.001480	<i>0.90 - 1.10</i>
Calculated intercept		0.582167	<i>+/- 4</i>

Notes: Installing new WS/WD sensors. Crossarm aligned with true north. No issues to note.

Calibration Performed By: Rene Chamberland & Devin Russell




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23
FORT HILLS
AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	August 10, 2023	Last Cal Date:	July 10, 2023
Start time (MST):	9:26	End time (MST):	12:20
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.76	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC281425			
Removed Cal Gas Conc:	49.76	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	451
ZAG Make/Model:	API T701		Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002525	1.000337	Backgd or Offset:	17.8	17.8
Calibration intercept:	-0.604149	-0.563797	Coeff or Slope:	1.040	1.040

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4920	80.3	799.1	798.2	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	80.3	799.1	798.9	1.000
second point	4960	40.2	400.1	399.8	1.001
third point	4980	20.1	200.0	198.8	1.006
as left zero	5000	0.0	0.0	0.0	----
as left span	4920	80.3	799.1	798.5	1.001
Average Correction Factor					1.002

Baseline Corr As found:	798.40	Previous response	800.51	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 10, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:26	End Time (MST):	12:20
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥0.995
799.1	798.9	1.0002		
400.1	399.8	1.0006	Slope	0.90 - 1.10
200.0	198.8	1.0062		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 10, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Hills Station number: AMS23
 Calibration Date: August 17, 2023 Last Cal Date: July 24, 2023
 Start time (MST): 6:50 End time (MST): 10:40
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.20 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517372
 Removed Cal Gas Conc: 5.20 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 451
 ZAG Make/Model: API T701 Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1300156232
 Converter make: CDN-101 Converter serial #: 594
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006312	1.011740	Backgd or Offset:	1.19
Calibration intercept:	0.042041	0.002125	Coeff or Slope:	1.124

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4923	77.0	80.0	77.0	1.038
as found 2nd point	4962	38.5	40.0	38.7	1.031
as found 3rd point	4981	19.2	19.9	19.2	1.034
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4923	77.0	80.0	80.8	0.990
second point	4962	38.5	40.0	40.8	0.980
third point	4981	19.2	19.9	20.1	0.992
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.0	80.0	81.7	0.979
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.988
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 77.1 Prev response: 80.55 *% change: -4.5%
 Baseline Corr 2nd AF pt: 38.8 AF Slope: 0.963751 AF Intercept: -0.018891
 Baseline Corr 3rd AF pt: 19.3 AF Correlation: 0.999987

* = > +/-5% change initiates investigation

Notes: Averaging time changed from 60sec to 90sec. SOx scrubber checked after the calibrator zero.No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 17, 2023	Previous Calibration:	July 24, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	6:50	End Time (MST):	10:40
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999959	
80.0	80.8	0.9901			≥0.995
40.0	40.8	0.9803	Slope	1.011740	
19.9	20.1	0.9924			0.90 - 1.10
			Intercept	0.002125	+/-3



TRS Calibration Plot

Date: August 17, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	August 10, 2023	Last Cal Date:	July 10, 2023
Start time (MST):	9:26	End time (MST):	12:20
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
Removed C ₃ H ₈ Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.32E-04	2.39E-04	NMHC SP Ratio:	5.06E-05
CH ₄ Retention time:	13.0	13.0	NMHC Peak Area:	181940
				176470

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	16.77	1.025
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.28	0.995
second point	4960	40.2	8.61	8.64	0.996
third point	4980	20.1	4.30	4.30	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.19	17.29	0.994

Average Correction Factor				0.998
Baseline Corr AF:	16.77	Prev response	17.07	*% change -1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	8.92	1.027
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.22	0.994
second point	4960	40.2	4.59	4.63	0.990
third point	4980	20.1	2.29	2.32	0.989
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	9.16	9.19	0.996
Average Correction Factor					0.991
Baseline Corr AF:	8.92	Prev response	9.09	*% change	-1.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	7.85	1.024
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.06	0.996
second point	4960	40.2	4.02	4.01	1.003
third point	4980	20.1	2.01	1.98	1.017
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.03	8.10	0.991
Average Correction Factor					1.005
Baseline Corr AF:	7.85	Prev response	7.98	*% change	-1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.993826	1.005681
THC Cal Offset:	-0.020141	-0.013606
CH ₄ Cal Slope:	0.997507	1.005023
CH ₄ Cal Offset:	-0.033221	-0.021648
NMHC Cal Slope:	0.990511	1.006196
NMHC Cal Offset:	0.013680	0.008043

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 10, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:26	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
17.19	17.28	0.9949						
8.61	8.64	0.9957				Slope	1.005681	0.90 - 1.10
4.30	4.30	1.0020						
			Intercept	-0.013606	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-01-2020


Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 10, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:26	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999967	≥0.995
8.03	8.06	0.9964			
4.02	4.01	1.0029			
2.01	1.98	1.0171			
			Slope	1.005023	0.90 - 1.10
			Intercept	-0.021648	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 10, 2023	Previous Calibration:	July 10, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:26	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995			
9.16	9.22	0.9936						
4.59	4.63	0.9899				Slope	1.006196	0.90 - 1.10
2.29	2.32	0.9887						
			Intercept	0.008043	± 0.5			



NMHC Calibration Plot

Date: August 10, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Hills
Calibration Date: August 1, 2023
Start time (MST): 8:50
Reason: Routine
Station number: AMS23
Last Cal Date: July 21, 2023
End time (MST): 13:52

Calibration Standards

NO Gas Cylinder #: CC332703
NOX Cal Gas Conc: 49.7 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 49.7 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 28, 2024
NO Cal Gas Conc: 49.7 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 49.7 ppm
NO gas Diff:
Serial Number: 451
Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1152430007

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.094	1.107	NO bkgnd or offset:	3.4	3.4
NOX coeff or slope:	0.995	0.995	NOX bkgnd or offset:	3.8	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	165.6	166.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999493	0.999964
NO _x Cal Offset:	0.844323	0.504456
NO Cal Slope:	0.999823	1.001151
NO Cal Offset:	0.004005	-0.335801
NO ₂ Cal Slope:	1.006741	0.997897
NO ₂ Cal Offset:	0.163045	-0.698579



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
as found span	4920	80.5	800.2	800.2	0.0	790.7	788.0	2.7	1.012	1.015
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
high point	4920	80.5	800.2	800.2	0.0	799.8	800.4	-0.5	1.000	1.000
second point	4960	40.2	399.6	399.6	0.0	401.7	401.0	0.7	0.995	0.996
third point	4980	20.1	199.8	199.8	0.0	200.4	198.5	1.9	0.997	1.006
as left zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.2	-0.2	----	----
as left span	4920	80.5	800.2	445.8	354.4	800.6	445.8	354.8	0.999	1.000
Average Correction Factor									0.997	1.001

Corrected As found	NO _x = 791.1 ppb	NO = 788.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.2%	
Previous Response	NO _x = 800.6 ppb	NO = 800.0 ppb		*Percent Change	NO = -1.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	800.2	445.8	354.4	353.2	1.003	99.7%
2nd GPT point (200 ppb O ₃)	800.2	623.9	176.3	174.9	1.008	99.2%
3rd GPT point (100 ppb O ₃)	800.2	710.0	90.2	89.0	1.013	98.7%
Average Correction Factor					1.008	99.2%

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 21, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:50	End Time (MST):	13:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.2	799.8	1.0005		
399.6	401.7	0.9947		
199.8	200.4	0.9970		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 21, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:50	End Time (MST):	13:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.2	800.4	0.9997		
399.6	401.0	0.9964		
199.8	198.5	1.0065		
			0.999991	
			1.001151	
			-0.335801	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 21, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:50	End Time (MST):	13:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
354.4	353.2	1.0034		
176.3	174.9	1.0080		
90.2	89.0	1.0135		
			0.999994	
			0.997897	
			-0.698579	

NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 1, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Hills Station number: AMS 23
 Calibration Date: August 17, 2023 Last Cal Date: July 24, 2023
 Start time (MST): 7:09 End time (MST): 9:34

Analyzer Make: API T640 S/N: 1546
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	15.8	16.4	15.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	725.4	724.8	725.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	5.16	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>August 17, 2023</u>	Last Cal Date: <u>July 24, 2023</u>			
	PM w/o HEPA: <u>11.8</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.8	10.6	10.6	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>22.2</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>August 17, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>August 17, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: May 10, 2023
 Date RH/T Sensor Cleaned: May 10, 2023

Notes: No adjustments done. Leak check passed before cleaning. After cleaning leak check PM10 analyzer reading 1.0ug/m3, Datalogger reading 0.0ug/m3. PM2.5 both analyzer and datalogger reading 0.0ug/m3. Unplugged and left without power for 5mins Both PM10 and 2.5 now reading 0.0ug/m3 on analyzer and datalogger.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25 WASKŌW OHCI PIMÂTISIWIN AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number:	AMS25
Calibration Date:	August 29, 2023	Last Cal Date:	July 24, 2023
Start time (MST):	6:30	End time (MST):	9:33
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.54	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC437219			
Removed Cal Gas Conc:	50.54	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	747
ZAG Make/Model:	API T701		Serial Number:	4765

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005116	0.997792	Backgd or Offset:	9.7	9.7
Calibration intercept:	-0.536058	0.703977	Coeff or Slope:	0.988	0.988

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4921	79.2	800.5	796.1	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4921	79.2	800.5	799.0	1.002
second point	4960	39.6	400.3	401.2	0.998
third point	4980	19.8	200.1	200.1	1.000
as left zero	5000	0.0	0.0	0.5	----
as left span	4921	79.2	800.5	798.7	1.002
Average Correction Factor					1.000

Baseline Corr As found:	795.90	Previous response	804.08	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 29, 2023	Previous Calibration:	July 24, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	6:30	End Time (MST):	9:33
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999996	≥0.995
800.5	799.0	1.0019			
400.3	401.2	0.9978	Slope	0.997792	0.90 - 1.10
200.1	200.1	1.0002			
			Intercept	0.703977	+/-30



SO2 Calibration Plot

Date: August 29, 2023

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS25
 Calibration Date: August 18, 2023 Last Cal Date: July 18, 2023
 Start time (MST): 6:12 End time (MST): 10:38
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC517099
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 747
 ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
 Converter make: Global G-150 Converter serial #: 2022-219
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997066	0.991600	Backgd or Offset: 3.25	3.25
Calibration intercept:	-0.060000	0.080000	Coeff or Slope: 1.079	1.079

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4920	80.0	79.5	79.3	0.999
as found 2nd point	4960	40.0	39.7	39.8	0.993
as found 3rd point	4980	20.0	19.9	19.8	0.993
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.0	79.5	78.8	1.008
second point	4960	40.0	39.7	39.6	1.003
third point	4980	20.0	19.9	19.8	1.003
as left zero	5000	0.0	0.0	0.0	----
as left span	4920	80.0	800.0	796.3	1.005
SO2 Scrubber Check	4921	79.2	800.0	0.0	----
Date of last scrubber change:	20-Jun-23			Ave Corr Factor	1.005
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.5 Prev response: 79.16 *% change: 0.4%
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 1.000374 AF Intercept: -0.100000
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999988

* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 18, 2023	Previous Calibration:	July 18, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	6:12	End Time (MST):	10:38
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999992	≥0.995
79.5	78.8	1.0083			
39.7	39.6	1.0032	Slope	0.991600	0.90 - 1.10
19.9	19.8	1.0032			
			Intercept	0.080000	+/-3



H₂S Calibration Plot

Date: August 18, 2023

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Waskōw ohci Pimâtisiwin	Station Number:	AMS 25
Calibration Date:	August 31, 2023	Prev Cal Date:	October 7, 2022
Start Time (MST):	9:13	End Time (MST):	9:43
Tower Height (m):	10m	Reason:	Removal

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	U11125
WS Calibrator:	MetOne 053	Serial Number:	R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.2	0.2%
400	39.4	39.4	0.1%
600	58.6	58.5	-0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999999	0.999999	<i>≥0.9995</i>
Calculated slope	0.999979	0.999979	<i>0.90 - 1.10</i>
Calculated intercept	-0.013577	-0.013577	<i>+/- 2</i>

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	
As Found Declination (deg east of True North):		As Left Declination (deg east of True North):	
Solar noon time (MST):		Calc Declination*:	Degrees
Deadband calc:	357.0 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0		---
90		
180		
270		
357		

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			<i>≥0.9995</i>
Calculated slope			<i>0.90 - 1.10</i>
Calculated intercept			<i>+/- 4</i>

Notes: WS removed Due to stalling at low speeds

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Waskōw ohci Pimâtisiwin	Station Number:	AMS 25
Calibration Date:	August 31, 2023	Prev Cal Date:	October 7, 2022
Start Time (MST):	9:13	End Time (MST):	9:43
Tower Height (m):	10m	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	Y18364
WS Calibrator:	MetOne 053	Serial Number:	R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.5	-0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥0.9995
Calculated slope		0.999473	0.90 - 1.10
Calculated intercept		0.026227	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	U11345
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):		NA Calc Declination*:	NA Degrees
Deadband calc:	1.1 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.1	---
90	90.7	0.2%
180	178.4	-0.4%
270	269.0	-0.3%
357	356.0	-0.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999972	0.999988	≥0.9995
Calculated slope	1.001522	1.004369	0.90 - 1.10
Calculated intercept	0.998562	-0.221411	+/- 4

Notes: WS installed due to other WS not working properly. WD passed the torque test.

Calibration Performed By: Melissa Lemay




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS26 CHRISTINA LAKE AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 17, 2023	Previous Calibration:	July 13, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	15:09	End Time (MST):	17:42
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.6	----	Correlation Coefficient	≥0.995
799.0	812.6	0.9832		
399.4	403.9	0.9889	Slope	0.90 - 1.10
200.2	198.3	1.0097		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 17, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Christina Lake Station number: AMS26
 Calibration Date: August 17, 2023 Last Cal Date: July 14, 2023
 Start time (MST): 6:33 End time (MST): 10:21
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.89 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002466
 Removed Cal Gas Conc: 4.89 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 2447
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 450i Analyzer serial #: 1180030032
 Converter make: NA Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001325	1.002610	Backgd or Offset:	35.1 35.4
Calibration intercept:	-0.160901	0.119119	Coeff or Slope:	1.093 1.093

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4918	81.8	80.0	82.0	0.980
as found 2nd point	4959	40.9	40.0	40.9	0.988
as found 3rd point	4979	20.4	20.0	20.2	1.008
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4918	81.8	80.0	80.4	0.995
second point	4959	40.9	40.0	40.1	0.998
third point	4979	20.4	20.0	20.1	0.993
as left zero	5000	0.0	0.0	0.5	----
as left span	4918	81.8	80.0	81.7	0.979
SO2 Scrubber Check	4919	80.6	806.1	0.1	----
Date of last scrubber change:	27-Feb-19			Ave Corr Factor	0.995
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 81.6 Prev response: 79.95 *% change: 2.0%
 Baseline Corr 2nd AF pt: 40.5 AF Slope: 1.021895 AF Intercept: 0.119353
 Baseline Corr 3rd AF pt: 19.8 AF Correlation: 0.999945

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after MAF's. Ran scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 17, 2023	Previous Calibration:	July 14, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	6:33	End Time (MST):	10:21
Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999992	≥0.995
80.0	80.4	0.9951			
40.0	40.1	0.9975	Slope	1.002610	0.90 - 1.10
20.0	20.1	0.9927			
			Intercept	0.119119	+/-3



H₂S Calibration Plot

Date: August 17, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.1	-0.6	----	----
as found span	4920	80.0	813.1	800.3	12.8	826.4	814.6	11.7	0.9839	0.9825
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.6	0.0	-0.6	----	----
high point	4920	80.0	813.1	800.3	12.8	809.1	800.2	8.9	1.0050	1.0001
second point	4960	40.0	406.6	400.2	6.4	402.8	397.1	5.7	1.0093	1.0077
third point	4980	20.0	203.3	200.1	3.2	199.6	195.6	4.0	1.0184	1.0229
as left zero	5000	0.0	0.0	0.0	0.0	-0.6	0.0	-0.6	----	----
as left span	4920	80.0	813.1	396.0	417.1	810.0	396.9	413.1	1.0039	0.9978
Average Correction Factor									1.0109	1.0103

Corrected As found	NO _x = 827.2 ppb	NO = 814.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.9%
Previous Response	NO _x = 811.6 ppb	NO = 800.5 ppb		*Percent Change	NO = 1.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	799.5	395.2	417.1	414.1	1.0072	99.3%
2nd GPT point (200 ppb O ₃)	799.5	592.8	219.5	216.6	1.0134	98.7%
3rd GPT point (100 ppb O ₃)	799.5	698.9	113.4	109.6	1.0347	96.6%
Average Correction Factor					1.0184	98.2%

Notes:

Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	August 17, 2023	Previous Calibration:	July 13, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:18	End Time (MST):	15:11
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.6	----	Correlation Coefficient	≥0.995	
813.1	809.1	1.0050			
406.6	402.8	1.0093			
203.3	199.6	1.0184			
			Slope	0.996528	0.90 - 1.10
			Intercept	-1.780000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 17, 2023	Previous Calibration:	July 13, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:18	End Time (MST):	15:11
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 $0.90 - 1.10$ ± 20
800.3	800.2	1.0001		
400.2	397.1	1.0077		
200.1	195.6	1.0229		
			0.999960	
			1.001328	
			-2.380000	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 17, 2023	Previous Calibration:	July 13, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:18	End Time (MST):	15:11
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
417.1	414.1	1.0072		
219.5	216.6	1.0134		
113.4	109.6	1.0347		
			0.999956	
			0.995894	
			-1.805192	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 17, 2023

Location: Christina Lake






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27
JACKFISH 2/3
AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	August 23, 2023	Last Cal Date:	July 7, 2023
Start time (MST):	9:31	End time (MST):	12:20
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	<u>SG9133974BAL</u>			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3811
ZAG Make/Model:	API 701		Serial Number:	135

Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	12124313138
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008391	1.000975	Backgd or Offset:	7.5	7.2
Calibration intercept:	-1.738506	-1.618030	Coeff or Slope:	0.942	0.891

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	79.1	800.2	846.6	0.945
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	79.1	800.2	800.0	1.000
second point	4961	39.5	399.5	398.0	1.004
third point	4980	19.8	200.3	196.8	1.018
as left zero	5000	0.0	0.0	0.0	----
as left span	4921	79.1	800.2	800.2	1.000
Average Correction Factor					1.007

Baseline Corr As found:	846.70	Previous response	805.14	*% change	4.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Adjusted both zero and span.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 23, 2023	Previous Calibration:	July 7, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	9:31	End Time (MST):	12:20
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	≥0.995
800.2	800.0	1.0002		
399.5	398.0	1.0039	Slope	0.90 - 1.10
200.3	196.8	1.0178		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 23, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Jackfish 2/3 Station number: AMS27
 Calibration Date: August 2, 2023 Last Cal Date: July 19, 2023
 Start time (MST): 7:50 End time (MST): 11:30
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.41 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345023
 Removed Cal Gas Conc: 5.41 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3811
 ZAG Make/Model: API 701 Serial Number: 135

Analyzer Information

Analyzer make: API T101 Analyzer serial #: 621
 Converter make: Converter serial #:
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011028	1.006901	Backgd or Offset: 27.3	28.3
Calibration intercept:	0.082512	-0.097803	Coeff or Slope: 0.961	0.944

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4926	74.1	80.2	83.0	0.969
as found 2nd point	4963	37.0	40.0	42.0	0.960
as found 3rd point	4982	18.5	20.0	20.8	0.976
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4926	74.1	80.2	80.7	0.993
second point	4963	37.0	40.0	40.1	0.998
third point	4982	18.5	20.0	20.0	1.001
as left zero	5000	0.0	0.0	0.2	----
as left span	4926	74.1	80.2	80.0	1.002
SO2 Scrubber Check	4921	79.1	791.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	0.998
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 82.7 Prev response: 81.14 *% change: 1.9%
 Baseline Corr 2nd AF pt: 41.7 AF Slope: 1.032695 AF Intercept: 0.322966
 Baseline Corr 3rd AF pt: 20.5 AF Correlation: 0.999956

* = > +/-5% change initiates investigation

Notes: Adjusted both zero and span.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 2, 2023	Previous Calibration:	July 19, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	7:50	End Time (MST):	11:30
Analyzer make:	API T101	Analyzer serial #:	621

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999992	
80.2	80.7	0.9935			≥0.995
40.0	40.1	0.9984	Slope	1.006901	
20.0	20.0	1.0007			0.90 - 1.10
			Intercept	-0.097803	+/-3



H₂S Calibration Plot

Date: August 2, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.6	0.3	----	----
as found span	4921	79.4	816.8	800.3	16.5	808.0	780.0	28.0	1.0109	1.0260
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4921	79.4	816.8	800.3	16.5	816.8	808.2	8.7	1.0000	0.9902
second point	4960	39.7	408.5	400.2	8.3	405.9	401.7	4.2	1.0063	0.9963
third point	4980	19.8	203.7	199.6	4.1	202.0	198.5	3.6	1.0085	1.0055
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.8	-1.0	----	----
as left span	4921	79.4	816.8	430.6	401.7	814.5	428.3	386.2	1.0028	1.0054
Average Correction Factor									1.0049	0.9973

Corrected As found	NO _x = 808.3 ppb	NO = 780.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.0%	
Previous Response	NO _x = 816.5 ppb	NO = 799.1 ppb		*Percent Change	NO = -2.4%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	808.6	423.4	401.7	390.8	1.0279	97.3%
2nd GPT point (200 ppb O3)	808.6	630.2	194.9	190.5	1.0232	97.7%
3rd GPT point (100 ppb O3)	808.6	729.2	95.9	94.5	1.0150	98.5%
Average Correction Factor					1.0220	97.8%

Notes:

Adjusted the span only.

Calibration Performed By:

Denny Ray Estador



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 8, 2023	Previous Calibration:	July 20, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:59	End Time (MST):	13:45
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
816.8	816.8	1.0000		
408.5	405.9	1.0063		
203.7	202.0	1.0085		
			0.999987	
			1.000358	
			-1.195601	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 8, 2023	Previous Calibration:	July 20, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:59	End Time (MST):	13:45
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	808.2	0.9902		
400.2	401.7	0.9963		
199.6	198.5	1.0055		
			0.999979	
			1.010851	
			-1.717915	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 8, 2023	Previous Calibration:	July 20, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:59	End Time (MST):	13:45
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
401.7	390.8	1.0279			
194.9	190.5	1.0232			
95.9	94.5	1.0150			
			Slope	0.972010	0.90 - 1.10
			Intercept	0.660630	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 8, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name: Jackfish 2/3 Station Number: AMS 30
Calibration Date: August 29, 2023 Prev Cal Date: August 31, 2022
Start Time (MST): 10:00 End Time (MST): 10:55
Tower Height (m): 10.0 Reason: Routine

Wind Speed Information

Sensor make/model: Met One 010C-1 Serial Number: X16496
WS Calibrator: MetOne 053-120 Serial Number: R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.2	0.2%
400	39.4	39.4	0.1%
600	58.6	58.5	-0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	<i>≥0.9995</i>
Calculated slope		0.999979	<i>0.90 - 1.10</i>
Calculated intercept		-0.013577	<i>+/- 2</i>

Wind Direction Information

Sensor make/model: Met One 020C-1 Serial Number: X16480
As Found Declination (deg east of True North): 14 As Left Declination (deg east of True North): 13
Solar noon time (MST): 13:24 Calc Declination*: 13 Degrees
Deadband calc: 2.2 degrees (*Limit 4 deg*) * - calculated declination as per NOAA website

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	-0.8	---
90	91.1	0.3%
180	181.8	0.5%
270	271.0	0.3%
357	354.0	-0.8%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999918	<i>≥0.9995</i>
Calculated slope		1.004859	<i>0.90 - 1.10</i>
Calculated intercept		-0.893768	<i>+/- 4</i>

Notes:

Bearings still good. Confirmed declination with a compass.

Calibration Performed By:

Denny Ray Estador




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS29
SURMONT 2
AUGUST 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 8, 2023	Previous Calibration:	July 17, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:56	End Time (MST):	12:56
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥0.995
800.1	801.0	0.9989		
400.6	400.4	1.0005	Slope	0.90 - 1.10
199.8	198.8	1.0051		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 8, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 10, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:50	End Time (MST):	14:48
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits			
0.0	-0.2	----	Correlation Coefficient	1.000000	≥0.995			
80.0	79.9	1.0013						
40.1	40.0	1.0027				Slope	1.001190	0.90 - 1.10
20.1	19.9	1.0077				Intercept	-0.182890	+/-3



H₂S Calibration Plot

Date: August 1, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	August 8, 2023	Last Cal Date:	July 17, 2023
Start time (MST):	9:56	End time (MST):	12:56
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC356008	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
C3H8 Cal Gas Conc.	<u>205.7</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
Removed C3H8 Conc.	<u>205.7</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5472
ZAG Make/Model:	Teledyne API T701	Serial Number:	4297

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1170050149
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000217	1.000971	Background:	3.64	3.64
Calibration intercept:	-0.075664	-0.017870	Coefficient:	3.987	3.987

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.04	----
as found span	4918	81.3	17.31	17.36	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.06	----
high point	4918	81.3	17.31	17.35	0.998
second point	4959	40.6	8.65	8.60	1.005
third point	4979	20.3	4.32	4.23	1.021
as left zero	5000	0.0	0.00	0.03	----
as left span	4918	81.3	17.31	17.42	0.994
Average Correction Factor					1.008
Baseline Corr As found:	17.40	Previous response	17.24	*% change	0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 8, 2023	Previous Calibration:	July 17, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:56	End Time (MST):	12:56
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.06	----	Correlation Coefficient	0.999915	
17.31	17.35	0.9979			≥0.995
8.65	8.60	1.0053	Slope	1.000971	
4.32	4.23	1.0211			0.90 - 1.10
			Intercept	-0.017870	+/-1.5



THC Calibration Plot

Date: August 8, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as found span	4916	84.2	799.2	799.2	0.0	797.0	795.9	1.1	1.0028	1.0041
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
high point	4916	84.2	799.2	799.2	0.0	799.7	798.7	1.0	0.9994	1.0006
second point	4958	42.1	399.6	399.6	0.0	398.7	398.4	0.3	1.0023	1.0030
third point	4979	21.1	200.3	200.3	0.0	200.2	198.0	2.2	1.0004	1.0115
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4916	84.2	799.2	407.4	391.8	801.0	407.1	393.7	0.9977	1.0007
Average Correction Factor									1.0007	1.0050

Corrected As found	NO _x = 797.0 ppb	NO = 795.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.5%
Previous Response	NO _x = 801.1 ppb	NO = 799.9 ppb		*Percent Change	NO = -0.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	792.9	401.1	391.8	392.1	0.9992	100.1%
2nd GPT point (200 ppb O3)	792.9	610.5	182.4	182.8	0.9978	100.2%
3rd GPT point (100 ppb O3)	792.9	701.4	91.5	92.0	0.9946	100.5%
Average Correction Factor					0.9972	100.3%

Notes:

Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 21, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:29	End Time (MST):	14:09
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.2	799.7	0.9994		
399.6	398.7	1.0023		
200.3	200.2	1.0004		
			0.999997	
			1.000454	
			-0.253155	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 21, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:29	End Time (MST):	14:09
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.2	798.7	1.0006		
399.6	398.4	1.0030		
200.3	198.0	1.0115		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 21, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:29	End Time (MST):	14:09
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
391.8	392.1	0.9992		
182.4	182.8	0.9978		
91.5	92.0	0.9946		
			0.999999	
			1.000433	
			0.227914	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 9, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name:	Surmont 2	Station number:	AMS 29
Calibration Date:	August 9, 2023	Last Cal Date:	July 21, 2023
Start time (MST):	9:54	End time (MST):	12:20
Analyzer Make:	API T640	S/N:	253
Particulate Fraction:	PM2.5		
Flow Meter Make/Model:	Alicat FP-25BT	S/N:	388750
Temp/RH standard:	Alicat FP-25BT	S/N:	388750

Monthly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Measured</u>	<u>As left</u>	<u>Adjusted</u>	<u>(Limits)</u>
T (°C)	13.2	14.26	13.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	708.1	709.23	708.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.127	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check:	<u>August 9, 2023</u>	Last Cal Date:	<u>July 21, 2023</u>	
	PM w/o HEPA:	<u>3.3</u>	PM w/ HEPA:	<u>0</u>	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Post maintenance</u>	<u>As left</u>	<u>Adjusted</u>	<u>(Limits)</u>
PMT Peak Test	11.0	10.8	11.2	<input checked="" type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA:	<u>3.3</u>	w/ HEPA:	<u>0.0</u>
Date Optical Chamber Cleaned:			<u>August 9, 2023</u>		<0.2 ug/m3
Disposable Filter Changed:			<u>August 9, 2023</u>		

Annual Maintenance

Date Sample Tube Cleaned:	<u>September 30, 2022</u>
Date RH/T Sensor Cleaned:	<u>October 6, 2022</u>

Adjusted PMT peak, both leak checks passed.

Notes:

Calibration by: Braiden Boutilier




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS30
ELLS RIVER
AUGUST 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 4, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:32	End Time (MST):	11:45
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999951	≥0.995
800.4	799.0	1.0017			
400.2	395.3	1.0125	Slope	0.999617	0.90 - 1.10
200.1	195.5	1.0236			
			Intercept	-2.615868	+/-30



SO2 Calibration Plot

Date: August 1, 2023

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Ells River Station number: AMS30
 Calibration Date: August 16, 2023 Last Cal Date: July 5, 2023
 Start time (MST): 9:21 End time (MST): 13:17
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.08 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002443
 Removed Cal Gas Conc: 5.08 ppm Rem Gas Exp Date:
 Removed Gas Cyl #: Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3061
 ZAG Make/Model: API T701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1410661331
 Converter make: CDN - 101 Converter serial #: 562
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008782	1.003494	Backgd or Offset:	1.61	1.63
Calibration intercept:	-0.159281	-0.199209	Coeff or Slope:	1.136	1.136

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	78.7	80.0	79.0	1.012
as found 2nd point	4961	39.4	40.0	39.7	1.008
as found 3rd point	4980	19.7	20.0	19.5	1.026
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	78.7	80.0	80.1	0.998
second point	4961	39.4	40.0	40.0	1.001
third point	4980	19.7	20.0	19.6	1.021
as left zero	5000	0.0	0.0	0.1	----
as left span	4921	78.7	80.0	80.1	0.998
SO2 Scrubber Check	4921	79.2	800.4	0.1	----
Date of last scrubber change:	N/A			Ave Corr Factor	1.007
Date of last converter efficiency test:	N/A			efficiency	

Baseline Corr As found: 79.0 Prev response: 80.51 *% change: -1.9%
 Baseline Corr 2nd AF pt: 39.7 AF Slope: 0.989347 AF Intercept: -0.079019
 Baseline Corr 3rd AF pt: 19.5 AF Correlation: 0.999974

* = > +/-5% change initiates investigation

Notes: No adjustments made.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 16, 2023	Previous Calibration:	July 5, 2023
Station Name:	Ells River	Station Number:	AMS30
Start Time (MST):	9:21	End Time (MST):	13:17
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits			
0.0	0.0	----	Correlation Coefficient	0.999964	≥0.995			
80.0	80.1	0.9983						
40.0	40.0	1.0007				Slope	1.003494	0.90 - 1.10
20.0	19.6	1.0212						
			Intercept	-0.199209	+/-3			



TRS Calibration Plot

Date: August 16, 2023

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	August 1, 2023	Last Cal Date:	July 21, 2023
Start time (MST):	8:32	End time (MST):	11:45
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm		
Diff between cyl (CH ₄):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701H	Serial Number:	3061
		Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	0.000252	0.000252	NMHC SP Ratio:	4.67E-05
CH ₄ Retention time:	14.4	14.4	NMHC Peak Area:	195284
				195284

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	17.25	0.987
as found 2nd point	4960	39.6	8.51	8.55	0.995
as found 3rd point	4980	19.8	4.26	4.20	1.014
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.18	0.991
second point	4960	39.6	8.51	8.50	1.001
third point	4980	19.8	4.26	4.18	1.018
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.26	0.987

Average Correction Factor				1.004
Baseline Corr AF:	17.25	Prev response	16.99	*% change 1.5%
Baseline Corr 2nd AF:	8.5	AF Slope:	1.014882	AF Intercept: -0.059940
Baseline Corr 3rd AF:	4.2	AF Correlation:	0.999945	* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	9.26	0.984
as found 2nd point	4960	39.6	4.56	4.58	0.997
as found 3rd point	4980	19.8	2.28	2.25	1.015
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.24	0.986
second point	4960	39.6	4.56	4.57	0.998
third point	4980	19.8	2.28	2.25	1.013
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.26	0.984
Average Correction Factor					0.999
Baseline Corr AF:	9.26	Prev response	9.10	*% change	1.8%
Baseline Corr 2nd AF:	4.6	AF Slope:	1.018352	AF Intercept:	-0.040535
Baseline Corr 3rd AF:	2.2	AF Correlation:	0.999911	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.99	0.991
as found 2nd point	4960	39.6	3.96	3.97	0.996
as found 3rd point	4980	19.8	1.98	1.96	1.012
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.93	0.998
second point	4960	39.6	3.96	3.93	1.006
third point	4980	19.8	1.98	1.93	1.023
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	8.00	0.990
Average Correction Factor					1.009
Baseline Corr AF:	7.99	Prev response	7.90	*% change	1.1%
Baseline Corr 2nd AF:	3.97	AF Slope:	1.010866	AF Intercept:	-0.021620
Baseline Corr 3rd AF:	1.96	AF Correlation:	0.999966	* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001691	1.010450
THC Cal Offset:	-0.062138	-0.062141
CH ₄ Cal Slope:	1.001405	1.003947
CH ₄ Cal Offset:	-0.027557	-0.026358
NMHC Cal Slope:	1.001951	1.016086
NMHC Cal Offset:	-0.034381	-0.035983

Notes: No adjustments made.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 21, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:32	End Time (MST):	11:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999940	≥ 0.995			
17.03	17.18	0.9913						
8.51	8.50	1.0015				Slope	1.010450	0.90 - 1.10
4.26	4.18	1.0177						
			Intercept	-0.062141	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-01-2020


Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 21, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:32	End Time (MST):	11:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999950	≥0.995
7.91	7.93	0.9977			
3.96	3.93	1.0061			
1.98	1.93	1.0232			
			Slope	1.003947	0.90 - 1.10
			Intercept	-0.026358	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 1, 2023	Previous Calibration:	July 21, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:32	End Time (MST):	11:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999928	≥ 0.995			
9.11	9.24	0.9858						
4.56	4.57	0.9978				Slope	1.016086	0.90 - 1.10
2.28	2.25	1.0130						
			Intercept	-0.035983	± 0.5			



NMHC Calibration Plot

Date: August 1, 2023

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	August 14, 2023	Last Cal Date:	August 1, 2023
Start time (MST):	9:39	End time (MST):	11:20
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
ZAG make/model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	0.000252	0.000252	NMHC SP Ratio: 4.67E-05	4.67E-05
CH ₄ Retention time:	14.4	14.4	NMHC Peak Area: 195284	195284

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.96	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	16.87	1.009

Average Correction Factor					
Baseline Corr AF:	16.96	Prev response	17.14	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* => +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	8.98	1.015
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	8.94	1.020
Average Correction Factor					
Baseline Corr AF:	8.98	Prev response	9.22	*% change	-2.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.98	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.94	0.997
Average Correction Factor					
Baseline Corr AF:	7.98	Prev response	7.92	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.010450	
THC Cal Offset:	-0.062141	
CH ₄ Cal Slope:	1.003947	
CH ₄ Cal Offset:	-0.026358	
NMHC Cal Slope:	1.016086	
NMHC Cal Offset:	-0.035983	

Notes:

Replaced N2 cylinder.


Calibration Performed By:

Denny Ray Estador

NMHC Calibration Plot

Date: August 14, 2023

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Ells River
Calibration Date: August 9, 2023
Start time (MST): 8:32
Reason: Routine
Station number: AMS 30
Last Cal Date: July 6, 2023
End time (MST): 12:50

Calibration Standards

NO Gas Cylinder #: T2Y1P2R
NOX Cal Gas Conc: 50.83 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 50.83 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 49.97 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 49.97 ppm
NO gas Diff:
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 710321429

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.031	1.099	NO bkgnd or offset:	12.8	13.5
NOX coeff or slope:	0.994	0.989	NOX bkgnd or offset:	12.8	13.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	184.2	188.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001026	1.004722
NO _x Cal Offset:	-1.400000	-1.140000
NO Cal Slope:	0.999657	1.006504
NO Cal Offset:	-2.520000	-2.140000
NO ₂ Cal Slope:	1.005854	1.001553
NO ₂ Cal Offset:	1.195147	0.383629



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
as found span	4920	80.0	813.3	799.5	13.8	773.0	755.2	17.8	1.0521	1.0587
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4920	80.0	813.3	799.5	13.8	816.5	803.7	12.8	0.9961	0.9948
second point	4960	40.0	406.6	399.8	6.9	407.1	399.0	8.1	0.9989	1.0019
third point	4980	20.0	203.3	199.9	3.4	201.7	197.0	4.8	1.0080	1.0146
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.1	----	----
as left span	4920	80.0	813.3	420.3	393.0	824.6	431.6	393.0	0.9863	0.9739
Average Correction Factor									1.0010	1.0038

Corrected As found	NO _x = 773.2 ppb	NO = 755.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -5.1%	
Previous Response	NO _x = 812.7 ppb	NO = 796.7 ppb		*Percent Change	NO = -5.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	804.3	425.1	393.0	393.8	0.9979	100.2%
2nd GPT point (200 ppb O ₃)	804.3	619.8	198.3	199.1	0.9958	100.4%
3rd GPT point (100 ppb O ₃)	804.3	710.5	107.6	108.4	0.9923	100.8%
Average Correction Factor					0.9953	100.5%

Notes:

Adjusted the span only.

Calibration Performed By: Denny Ray Estador



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 6, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:32	End Time (MST):	12:50
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.3	816.5	0.9961		
406.6	407.1	0.9989		
203.3	201.7	1.0080		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 6, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:32	End Time (MST):	12:50
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.5	803.7	0.9948		
399.8	399.0	1.0019		
199.9	197.0	1.0146		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 6, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:32	End Time (MST):	12:50
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
393.0	393.8	0.9979		
198.3	199.1	0.9958		
107.6	108.4	0.9923		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 9, 2023

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Ells River Station number: AMS 30
 Calibration Date: August 25, 2023 Last Cal Date: July 26, 2023
 Start time (MST): 9:14 End time (MST): 9:53

Analyzer Make: API T640 S/N: 875
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat S/N: 388751
 Temp/RH standard: Alicat S/N: 388751

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	14.9	14.1	14.9	<input type="checkbox"/>	
P (mmHg)	736.8	738.77	736.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.02	5.00	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: August 25, 2023 Last Cal Date: July 26, 2023
 PM w/o HEPA: 27.8 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.7	11.6	10.8	<input checked="" type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: 19.4 w/ HEPA: 0
 Date Optical Chamber Cleaned: August 25, 2023 <0.2 ug/m3
 Disposable Filter Changed: August 25, 2023

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: Adjusted PMT. Inlet head: Clean!

Calibration by: Denny Ray Estador



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Ells River	Station Number:	AMS 30
Calibration Date:	August 25, 2023	Prev Cal Date:	October 11, 2022
Start Time (MST):	10:00	End Time (MST):	12:05
Tower Height (m):	10.0	Reason:	Routine

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	A3111
WS Calibrator:	MetOne 053-120	Serial Number:	R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.5	-0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥ 0.9995
Calculated slope		0.999473	$0.90 - 1.10$
Calculated intercept		0.026227	± 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	J2732
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:29	Calc Declination*:	14 Degrees
Deadband calc:	0.3 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	-0.4	---
90	87.7	-0.6%
180	176.8	-0.9%
270	270.0	0.0%
357	356.3	-0.2%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999955	≥ 0.9995
Calculated slope		0.998023	$0.90 - 1.10$
Calculated intercept		1.672019	± 4

Notes:

Bearings still good. Confirmed declination with a compass.

Calibration Performed By:

Denny Ray Estador




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS505
SAWBONES BAY
AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Sawbones Bay Station number: AMS505
 Calibration Date: August 9, 2023 Last Cal Date: July 14, 2023
 Start time (MST): 10:05 End time (MST): 12:52
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 51.4 ppm Cal Gas Exp Date: February 15, 2029
 Cal Gas Cylinder #: EY0000672
 Removed Cal Gas Conc: 51.40 ppm Rem Gas Exp Date: February 15, 2029
 Removed Gas Cyl #: EY0000672 Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 5112
 ZAG Make/Model: Teledyne API T701H Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 0710321323
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001253	1.000823	Backgd or Offset:	19.2	19.2
Calibration intercept:	-0.552518	-0.552184	Coeff or Slope:	0.995	0.995

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4922	77.8	799.8	799.6	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4922	77.8	799.8	800.1	1.000
second point	4961	38.9	399.9	400.1	1.000
third point	4981	19.5	200.4	198.5	1.010
as left zero	5000	0.0	0.0	0.3	----
as left span	4922	77.8	799.8	801.2	0.998
Average Correction Factor					1.003

Baseline Corr As found:	799.50	Previous response	800.27	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

** = > +/-5% change initiates investigation*

Notes: Changed inlet filter. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 9, 2023	Previous Calibration:	July 14, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	10:05	End Time (MST):	12:52
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321323

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient	≥0.995
799.8	800.1	0.9996		
399.9	400.1	0.9995	Slope	0.90 - 1.10
200.4	198.5	1.0098		
			Intercept	+/-30



SO2 Calibration Plot

Date: August 9, 2023

Location: Sawbones Bay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Sawbones Bay Station number: AMS505
 Calibration Date: August 30, 2023 Last Cal Date: July 12, 2023
 Start time (MST): 08:24 End time (MST): 12:31
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517397
 Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 5112
 ZAG Make/Model: Teledyne API T701 Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43iQ Analyzer serial #: 1228021057
 Converter make: Global 150 Converter serial #: 2022-224
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994365	0.998506	Backgd or Offset:	2.26
Calibration intercept:	-0.098257	-0.018155	Coeff or Slope:	1.037
				2.21
				1.021

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4922	77.7	80.0	80.4	0.994
as found 2nd point	4961	38.8	40.0	40.4	0.987
as found 3rd point	4981	19.4	20.0	20.2	0.984
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	77.7	80.0	79.9	1.002
second point	4961	38.8	40.0	39.9	1.002
third point	4981	19.4	20.0	19.9	1.004
as left zero	5000	0.0	0.0	0.1	----
as left span	4922	77.7	80.0	79.6	1.005
SO2 Scrubber Check	4922	77.8	778.0	-0.1	----

Date of last scrubber change:	Ave Corr Factor	1.002
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 80.5 Prev response: 79.49 *% change: 1.3%
 Baseline Corr 2nd AF pt: 40.5 AF Slope: 1.005356 AF Intercept: 0.042111
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999981

* = > +/-5% change initiates investigation

Notes: Changed inlet after calibrator zero. Scrubber check and passed after calibrator zero. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 30, 2023	Previous Calibration:	July 12, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):		End Time (MST):	12:31
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1228021057

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits			
0.0	0.0	----	Correlation Coefficient	1.000000	≥0.995			
80.0	79.9	1.0017						
40.0	39.9	1.0016				Slope	0.998506	0.90 - 1.10
20.0	19.9	1.0040				Intercept	-0.018155	+/-3



H₂S Calibration Plot

Date: August 30, 2023

Location: Sawbones Bay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-1.0	0.1	-1.1	----	----
high point	4917	83.4	799.6	799.6	0.0	800.0	799.6	0.6	0.9995	1.0000
second point	4958	41.7	399.8	399.8	0.0	396.9	398.2	-1.2	1.0074	1.0041
third point	4979	20.9	200.4	200.4	0.0	195.6	195.9	-0.3	1.0245	1.0229
as left zero	5000	0.0	0.0	0.0	0.0	-1.5	0.2	-1.6	----	----
as left span	4916	83.4	799.7	330.8	468.9	797.4	327.5	470.0	1.0029	1.0102
Average Correction Factor									1.0105	1.0090

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.7	327.8	468.9	467.9	1.0021	99.8%
2nd GPT point (200 ppb O3)	796.7	536.3	260.4	259.6	1.0031	99.7%
3rd GPT point (100 ppb O3)	796.7	640.6	156.1	154.2	1.0123	98.8%
Average Correction Factor					1.0058	99.4%

Notes: Calibration done after the external pump was changed on August 3. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 4, 2023	Previous Calibration:	July 13, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	7:57	End Time (MST):	12:21
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-1.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.6	800.0	0.9995		
399.8	396.9	1.0074		
200.4	195.6	1.0245		
			0.999966	
			1.003179	
			-3.190449	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 4, 2023	Previous Calibration:	July 13, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	7:57	End Time (MST):	12:21
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
799.6	799.6	1.0000			
399.8	398.2	1.0041			
200.4	195.9	1.0229			
			Slope	1.001620	0.90 - 1.10
			Intercept	-2.069911	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 4, 2023	Previous Calibration:	July 13, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	7:57	End Time (MST):	12:21
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-1.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
468.9	467.9	1.0021		
260.4	259.6	1.0031		
156.1	154.2	1.0123		
			0.999995	
			1.000764	
			-1.369113	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 4, 2023

Location: Sawbones Bay






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS507
KIRBY SOUTH
AUGUST 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

September 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Kirby South	Station number:	AMS 507
Calibration Date:	August 3, 2023	Last Cal Date:	July 6, 2023
Start time (MST):	12:37	End time (MST):	15:44
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.18	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC303554</u>			
Removed Cal Gas Conc:	49.18	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3804
ZAG Make/Model:	API T701H		Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002607	0.999847	Backgd or Offset:	18.9	19.0
Calibration intercept:	-0.968916	-1.128170	Coeff or Slope:	1.135	1.135

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4919	81.3	799.6	793.4	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4919	81.3	799.6	799.7	1.000
second point	4959	40.7	400.3	396.5	1.010
third point	4980	20.3	199.7	198.5	1.006
as left zero	5000	0.0	0.0	0.3	----
as left span	4919	81.3	799.6	797.5	1.003
Average Correction Factor					1.005

Baseline Corr As found:	793.30	Previous response	800.73	*% change	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 6, 2023
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	12:37	End Time (MST):	15:44
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999970	≥0.995
799.6	799.7	0.9999			
400.3	396.5	1.0097	Slope	0.999847	0.90 - 1.10
199.7	198.5	1.0058			
			Intercept	-1.128170	+/-30



SO2 Calibration Plot

Date: August 3, 2023

Location: Kirby South





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Kirby South Station number: AMS507
 Calibration Date: August 2, 2023 Last Cal Date: July 25, 2023
 Start time (MST): 11:02 End time (MST): 15:50
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.167 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517378
 Removed Cal Gas Conc: 5.167 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3804
 ZAG Make/Model: API T701H Serial Number: 880

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1150840012
 Converter make: Global Converter serial #: 2022-197
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005318	1.001457	Backgd or Offset: 1.64	1.53
Calibration intercept:	0.458932	0.019070	Coeff or Slope: 1.109	1.048

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4923	77.4	80.0	84.5	0.948
as found 2nd point	4961	38.8	40.1	41.7	0.964
as found 3rd point	4981	19.3	19.9	20.8	0.963
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4923	77.4	80.0	80.2	0.997
second point	4961	38.8	40.1	40.0	1.002
third point	4981	19.3	19.9	20.0	0.997
as left zero	5000	0.0	0.0	0.1	----
as left span	4923	77.4	80.0	80.1	0.998
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:	25-Jul-23			Ave Corr Factor	0.999
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 84.4 Prev response: 80.86 *% change: 4.2%
 Baseline Corr 2nd AF pt: 41.6 AF Slope: 1.05729 AF Intercept: -0.180734
 Baseline Corr 3rd AF pt: 20.7 AF Correlation: 0.999911

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Scrubber check done after MPAF's, passed. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	August 2, 2023	Previous Calibration:	July 25, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	11:02	End Time (MST):	15:50
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
80.0	80.2	0.9972		
40.1	40.0	1.0024	Slope	0.90 - 1.10
19.9	20.0	0.9972		
			Intercept	+/-3



H₂S Calibration Plot

Date: August 2, 2023

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 6, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	12:37	End Time (MST):	15:44
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.03	----	Correlation Coefficient	0.999970	≥0.995
17.26	17.26	1.0002			
8.64	8.59	1.0062	Slope	0.999381	0.90 - 1.10
4.31	4.25	1.0135			
			Intercept	-0.016800	+/-1.5



THC Calibration Plot

Date: August 3, 2023

Location: Kirby South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	August 3, 2023	Last Cal Date:	July 4, 2023
Start time (MST):	7:36	End time (MST):	12:35
Reason:	Routine		

Calibration Standards

NO Gas Cylinder #:	T34ULGL	Cal Gas Expiry Date:	March 8, 2025
NOX Cal Gas Conc:	49.39 ppm	NO Cal Gas Conc:	49.02 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	49.39 ppm	Removed Gas NO Conc:	49.02 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	3804
ZAG make/model:	API 701H	Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.429	1.429	1.9	1.9
NOX coeff or slope:	0.992	0.992	2.0	2.0
NO2 coeff or slope:	1.000	1.000	162.66	160.75

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001602	1.006090
NO _x Cal Offset:	-2.351758	-4.471991
NO Cal Slope:	1.003048	1.005987
NO Cal Offset:	-3.253489	-5.373912
NO ₂ Cal Slope:	1.006879	1.007635
NO ₂ Cal Offset:	1.400873	1.168421



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.8	0.0	----	----
as found span	4919	81.0	800.1	794.1	6.0	767.8	763.7	4.1	1.0421	1.0398
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4919	81.0	800.1	794.1	6.0	803.0	796.6	6.2	0.9964	0.9969
second point	4960	40.5	400.0	397.0	3.0	394.9	390.0	4.9	1.0130	1.0180
third point	4980	20.2	199.5	198.0	1.5	192.5	189.5	3.0	1.0365	1.0450
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	0.0	----	----
as left span	4919	81.0	800.1	399.4	400.7	796.2	396.1	400.1	1.0049	1.0084
Average Correction Factor									1.0153	1.0200

Corrected As found	NO _x = 768.6 ppb	NO = 764.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -4.0%	
Previous Response	NO _x = 799.0 ppb	NO = 793.3 ppb		*Percent Change	NO = -3.8%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	792.4	397.7	400.7	403.9	0.9921	100.8%
2nd GPT point (200 ppb O ₃)	792.4	606.0	192.4	197.1	0.9761	102.4%
3rd GPT point (100 ppb O ₃)	792.4	704.6	93.8	95.8	0.9791	102.1%
Average Correction Factor					0.9824	101.8%

Notes:

Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 4, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	7:36	End Time (MST):	12:35
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.1	803.0	0.9964		
400.0	394.9	1.0130		
199.5	192.5	1.0365		

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 4, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	7:36	End Time (MST):	12:35
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
794.1	796.6	0.9969		
397.0	390.0	1.0180		
198.0	189.5	1.0450		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	August 3, 2023	Previous Calibration:	July 4, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	7:36	End Time (MST):	12:35
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
400.7	403.9	0.9921		
192.4	197.1	0.9761		
93.8	95.8	0.9791		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: August 3, 2023

Location: Kirby South





End of Report