



**WOOD BUFFALO  
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive  
Fort McMurray, AB T9K 0K4  
P: 780.799.4420 E: info@wbea.org  
[wbea.org](http://wbea.org)

Wood Buffalo Environmental Association

# SEPTEMBER 2023 MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

October 30, 2023

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS01 BERTHA GANTER - FORT MCKAY SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	September 12, 2023	Last Cal Date:	August 1, 2023
Start time (MST):	9:52	End time (MST):	13:11
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC418809			
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999194	0.999839	Backgd or Offset:	19.2	19.3
Calibration intercept:	0.086933	-0.113750	Coeff or Slope:	0.886	0.890

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4918	81.3	800.3	796.0	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4918	81.3	800.3	800.2	1.000
second point	4959	40.7	400.6	400.4	1.000
third point	4979	20.3	199.8	199.0	1.004
as left zero	5000	0.0	0.0	0.7	----
as left span	4918	81.3	800.3	800.2	1.000
Average Correction Factor					1.002

Baseline Corr As found:	795.80	Previous response	799.71	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	August 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:52	End Time (MST):	13:11
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999998	≥0.995
800.3	800.2	1.0001			
400.6	400.4	1.0005	Slope	0.999839	0.90 - 1.10
199.8	199.0	1.0041			
			Intercept	-0.113750	+/-30



SO2 Calibration Plot

Date: September 12, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Bertha Ganter-Fort McKay      Station number: AMS01  
 Calibration Date: September 11, 2023      Last Cal Date: August 28, 2023  
 Start time (MST): 9:15      End time (MST): 14:50  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.10 ppm      Cal Gas Exp Date: September 16, 2024  
 Cal Gas Cylinder #: CC511749  
 Removed Cal Gas Conc: 5.10 ppm      Rem Gas Exp Date: N/A  
 Removed Gas Cyl #: N/A      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 3565  
 ZAG Make/Model: Teledyne API T701      Serial Number: 5609

### Analyzer Information

Analyzer make: Thermo 43i-TLE      Analyzer serial #: 1218153461  
 Converter make: CD Nova      Converter serial #: 470  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000364	0.995649	Backgd or Offset: 2.29	2.26
Calibration intercept:	0.140000	0.279998	Coeff or Slope: 0.919	0.906

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4921	78.4	80.0	78.7	1.019
as found 2nd point	4960	39.2	40.0	40.0	1.005
as found 3rd point	4980	19.6	20.0	20.1	1.005
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4921	78.4	80.0	79.9	1.001
second point	4960	39.2	40.0	40.2	0.995
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.3	----
as left span	4921	78.4	80.0	81.0	0.988
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	0.997
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 78.5      Prev response: 80.16      \*% change: -2.1%  
 Baseline Corr 2nd AF pt: 39.8      AF Slope: 0.980934      AF Intercept: 0.419995  
 Baseline Corr 3rd AF pt: 19.9      AF Correlation: 0.999939

\* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 11, 2023	Previous Calibration:	August 28, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:15	End Time (MST):	14:50
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999995	≥0.995
80.0	79.9	1.0012			
40.0	40.2	0.9950	Slope	0.995649	0.90 - 1.10
20.0	20.1	0.9949			
			Intercept	0.279998	+/-3



TRS Calibration Plot

Date: September 11, 2023

Location: Bertha Ganter-Fort McKay







# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Bertha Ganter-Fort McKay      Station number: AMS01  
 Calibration Date: September 11, 2023      Last Cal Date: August 28, 2023  
 Start time (MST): 9:15      End time (MST): 14:50  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.10 ppm      Cal Gas Exp Date: September 16, 2024  
 Cal Gas Cylinder #: CC511749  
 Removed Cal Gas Conc: 5.10 ppm      Rem Gas Exp Date: N/A  
 Removed Gas Cyl #: N/A      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 3565  
 ZAG Make/Model: Teledyne API T701      Serial Number: 5609

### Analyzer Information

Analyzer make: Thermo 43iQTL      Analyzer serial #: 1200326167  
 Converter make: Thermo Converter      Converter serial #: N/A  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003143	0.997856	Backgd or Offset: 1.74	1.70
Calibration intercept:	0.096765	0.156776	Coeff or Slope: 1.014	0.997

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4922	78.4	80.0	82.9	0.963
as found 2nd point	4960	39.2	40.0	41.5	0.961
as found 3rd point	4980	19.6	20.0	20.4	0.975
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	78.4	80.0	79.8	1.002
second point	4960	39.2	40.0	40.4	0.990
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.7	----
as left span	4922	78.4	80.0	77.7	1.029
SO2 Scrubber Check	4919	81.3	813.0	0.1	----
Date of last scrubber change:	March 21, 2022			Ave Corr Factor	0.996
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 83.0      Prev response: 80.31      \*% change: 3.2%  
 Baseline Corr 2nd AF pt: 41.6      AF Slope: 1.039213      AF Intercept: -0.183330  
 Baseline Corr 3rd AF pt: 20.5      AF Correlation: 0.999984

\* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 11, 2023	Previous Calibration:	August 28, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:15	End Time (MST):	14:50
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999954	
80.0	79.8	1.0022			≥0.995
40.0	40.4	0.9901	Slope	0.997856	
20.0	20.1	0.9949			0.90 - 1.10
			Intercept	0.156776	+/-3



H<sub>2</sub>S Calibration Plot

Date: September 11, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	September 12, 2023	Last Cal Date:	August 1, 2023
Start time (MST):	9:52	End time (MST):	13:11
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH <sub>4</sub> Cal Gas Conc.	497.2 ppm	CH <sub>4</sub> Equiv Conc.	1061.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	497.2 ppm	CH <sub>4</sub> Equiv Conc.	1061.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	5609

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.91E-04	2.97E-04	NMHC SP Ratio:	6.33E-05
CH <sub>4</sub> Retention time:	14.6	14.6	NMHC Peak Area:	145218
				143189

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.27	17.04	1.013
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	17.27	17.23	1.002
second point	4959	40.7	8.64	8.50	1.017
third point	4980	20.3	4.31	4.23	1.020
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.27	17.28	0.999

Average Correction Factor				1.013
Baseline Corr AF:	17.04	Prev response	17.18	*% change -0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4918	81.3	9.18	9.09	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4918	81.3	9.18	9.15	1.003
second point	4959	40.7	4.60	4.56	1.008
third point	4980	20.3	2.29	2.29	1.003
as left zero	5000	0	0.00	0.00	----
as left span	4918	81.3	9.18	9.19	0.999
Average Correction Factor					1.005
Baseline Corr AF:	9.09	Prev response	9.15	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	7.95	1.017
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	8.09	8.08	1.001
second point	4959	40.7	4.05	3.94	1.028
third point	4980	20.3	2.02	1.94	1.039
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.09	1.000
Average Correction Factor					1.023
Baseline Corr AF:	7.95	Prev response	8.03	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997017	0.998565
THC Cal Offset:	-0.038093	-0.056289
CH <sub>4</sub> Cal Slope:	0.999108	1.000577
CH <sub>4</sub> Cal Offset:	-0.047957	-0.051154
NMHC Cal Slope:	0.995361	0.996780
NMHC Cal Offset:	0.009863	-0.004336

Notes: Changed the inlet filter and H2 cylinder after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	August 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:52	End Time (MST):	13:11
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999926	$\geq 0.995$			
17.27	17.23	1.0021						
8.64	8.50	1.0174				Slope	0.998565	0.90 - 1.10
4.31	4.23	1.0201						
			Intercept	-0.056289	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	August 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:52	End Time (MST):	13:11
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999758	$\geq 0.995$
8.09	8.08	1.0011			
4.05	3.94	1.0283			
2.02	1.94	1.0395			
			Slope	1.000577	0.90 - 1.10
			Intercept	-0.051154	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	August 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:52	End Time (MST):	13:11
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999993	$\geq 0.995$			
9.18	9.15	1.0030						
4.60	4.56	1.0076				Slope	0.996780	0.90 - 1.10
2.29	2.29	1.0031						
			Intercept	-0.004336	$\pm 0.5$			






NMHC Calibration Plot

Date: September 12, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Bertha Ganter-Fort McKay      Station number: AMS01  
Calibration Date: September 6, 2023      Last Cal Date: August 25, 2023  
Start time (MST): 9:57      End time (MST): 14:43  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T2Y1P9L      Cal Gas Expiry Date: December 11, 2023  
NOX Cal Gas Conc: 50.84 ppm      NO Cal Gas Conc: 50.04 ppm  
Removed Cylinder #: NA      Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 50.84 ppm      Removed Gas NO Conc: 50.04 ppm  
NOX gas Diff:      NO gas Diff:  
Calibrator Model: Teledyne API T700      Serial Number: 3565  
ZAG make/model: Teledyne API T701      Serial Number: 5609

### Analyzer Information

Analyzer make: Thermo 42i      Analyzer serial #: 1218153357  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.348	1.357	NO bkgnd or offset:	6.9	7.0
NOX coeff or slope:	0.992	0.991	NOX bkgnd or offset:	7.1	7.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	189.5	188.3

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000155	0.999396
NO <sub>x</sub> Cal Offset:	-0.260000	-0.540000
NO Cal Slope:	0.998858	0.999215
NO Cal Offset:	-1.280000	-1.080000
NO <sub>2</sub> Cal Slope:	1.001610	1.001909
NO <sub>2</sub> Cal Offset:	0.629963	-0.228647



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.4	0.0	0.4	----	----
as found span	4920	80.0	813.4	800.6	12.8	809.0	789.2	19.8	1.0055	1.0145
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.1	----	----
high point	4920	80.0	813.4	800.6	12.8	812.7	799.6	13.0	1.0009	1.0013
second point	4960	40.0	406.7	400.3	6.4	405.5	398.1	7.5	1.0030	1.0056
third point	4980	20.0	203.4	200.2	3.2	202.4	197.9	4.5	1.0047	1.0114
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.0	----	----
as left span	4920	80.0	813.4	402.2	411.2	813.3	403.7	409.5	1.0002	0.9964
Average Correction Factor									1.0029	1.0061

Corrected As found	NO <sub>x</sub> = 808.6 ppb	NO = 789.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.6%
Previous Response	NO <sub>x</sub> = 813.3 ppb	NO = 798.4 ppb		*Percent Change	NO = -1.2%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.7	399.3	411.2	411.8	0.9985	100.1%
2nd GPT point (200 ppb O3)	797.7	595.2	215.3	215.5	0.9991	100.1%
3rd GPT point (100 ppb O3)	797.7	696.0	114.5	114.3	1.0017	99.8%
Average Correction Factor					0.9998	100.0%

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 25, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:57	End Time (MST):	14:43
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
813.4	812.7	1.0009			
406.7	405.5	1.0030			
203.4	202.4	1.0047			
			Slope	0.999396	0.90 - 1.10
			Intercept	-0.540000	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 25, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:57	End Time (MST):	14:43
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.6	799.6	1.0013			
400.3	398.1	1.0056			
200.2	197.9	1.0114			
			Slope	0.999215	0.90 - 1.10
			Intercept	-1.080000	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 25, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:57	End Time (MST):	14:43
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
411.2	411.8	0.9985		
215.3	215.5	0.9991		
114.5	114.3	1.0017		
			0.999999	
			1.001909	
			-0.228647	


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 6, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Bertha Ganter-Fort McKay      Station number: AMS01  
 Calibration Date: September 1, 2023      Last Cal Date: August 10, 2023  
 Start time (MST): 9:40      End time (MST): 12:21  
 Reason: Routine

### Calibration Standards

O<sub>3</sub> generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 3565  
 ZAG Make/Model: Teledyne API T701      Serial Number: 5609

### Analyzer Information

Analyzer make: Teledyne API T400      Analyzer serial #: 1107  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001743	0.999943	Backgd or Offset:	3.2	3.2
Calibration intercept:	-0.080000	0.060000	Coeff or Slope:	1.010	1.010

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	5000	863.1	400.0	400.3	0.999
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.0	----
high point	5000	863.1	400.0	400.0	1.000
second point	5000	742.5	200.0	200.1	1.000
third point	5000	651.7	100.0	100.1	0.999
as left zero	5000	0.0	0.0	0.1	----
as left span	5000	863.1	400.0	404.4	0.989
Average Correction Factor					1.000

Baseline Corr As found:	400.4	Previous response	400.6	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 1, 2023	Previous Calibration:	August 10, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:40	End Time (MST):	12:21
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	1.000000	
400.0	400.0	1.0000			≥0.995
200.0	200.1	0.9995	Slope	0.999943	
100.0	100.1	0.9990			0.90 - 1.10
			Intercept	0.060000	+/- 5



O<sub>3</sub> Calibration Plot

Date: September 1, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01  
 Calibration Date: September 14, 2023 Last Cal Date: August 23, 2023  
 Start time (MST): 9:51 End time (MST): 11:18

Analyzer Make: Teledyne API T640 S/N: 324  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450  
 Temp/RH standard: Delta Cal S/N: 1450

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	17.1	16.1	17.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	736.3	735	736.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.12	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 14, 2023</u>		Last Cal Date: <u>August 23, 2023</u>		
	PM w/o HEPA: <u>70.1</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.5	11.1	11.1	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: <u>75.3</u>		w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:	<u>September 14, 2023</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>September 14, 2023</u>				

### Annual Maintenance

Date Sample Tube Cleaned: September 14, 2023  
 Date RH/T Sensor Cleaned: September 14, 2023

Notes: Flow, temperature, and pressure all within limits. Leak checks passed. Disposable filter changed. PMT peak test verified. Optical chamber and RH/T sensor cleaned. Sample tube cleaned.

Calibration by: Rene Chamberland



# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	September 14, 2023	Last Cal Date:	August 21, 2023
Start time (MST):	9:28	End time (MST):	13:07
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	<u>3040</u>	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>3040</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

### Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<b><u>Start</u></b>	<b><u>Finish</u></b>		<b><u>Start</u></b>	<b><u>Finish</u></b>
Calibration slope:	1.001990	1.000977	Backgd or Offset:	-0.012	-0.012
Calibration intercept:	0.159827	0.127800	Coeff or Slope:	0.992	0.990

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.6	40.9	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.6	40.6	0.999
second point	4966	33.3	20.2	20.6	0.982
third point	4983	16.7	10.2	10.3	0.986
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.5	40.1	1.011
Average Correction Factor					0.989

Baseline Corr As found:	40.81	Prev response:	40.80	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 21, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:28	End Time (MST):	13:07
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999915	
40.6	40.6	0.9987			≥0.995
20.2	20.6	0.9820	Slope	1.000977	
10.2	10.3	0.9858			0.90 - 1.10
			Intercept	0.127800	+/-1.5



CO Calibration Plot

Date: September 14, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	September 5, 2023	Last Cal Date:	August 11, 2023
Start time (MST):	9:56	End time (MST):	13:23
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>60,200</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
N2 Gen Make/Model:	Peak Sci NG5000		Serial Number:	7220900034

### Analyzer Information

Analyzer make: Teledyne API 360	Analyzer serial #: 442
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000529	1.001868	Backgd or Offset:	0.045	0.045
Calibration intercept:	-5.380000	-5.320000	Coeff or Slope:	0.874	0.875

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-0.2	----
as found span	2920	80.0	1605.3	1596.2	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.1	----
high point	2920	80.0	1605.3	1605.8	1.000
second point	2960	40.0	802.7	795.9	1.009
third point	2980	20.0	401.3	391.5	1.025
as left zero	3000	0.0	0.0	0.0	----
as left span	2960	40.0	802.7	784.7	1.023
Average Correction Factor					1.011

Baseline Corr As found:	1596.40	Prev response:	1600.80	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date	September 5, 2023	Previous Calibration	August 11, 2023
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS01
Start Time (MST)	9:56	End Time (MST)	13:23
Analyzer make	Teledyne API 360	Analyzer serial #	442

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999948	≥0.995
1605.3	1605.8	0.9997			
802.7	795.9	1.0085	Slope	1.001868	0.90 - 1.10
401.3	391.5	1.0251			
			Intercept	-5.320000	+/-10






CO<sub>2</sub> Calibration Plot

Date: September 5, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## TN - NO<sub>x</sub> - NH<sub>3</sub> Calibration Report

Version-05-2023

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	September 7, 2023	Last Cal Date:	August 23, 2023
Start time (MST):	9:23	End time (MST):	14:39
NH3 Cal Date:	September 8, 2023	Last Cal Date:	August 23, 2023
Start time (MST):	9:21	End time (MST):	13:30
Reason:	Routine		

### Calibration Standards

NOX Cal Gas Conc:	50.84	ppm	NO Gas Cylinder #:	T2Y1P9L
NO Cal Gas Conc:	50.04	ppm	NO Cal Gas Expiry:	March 3, 2028
Removed NOX Conc:	50.84	ppm	Removed Cylinder #:	NA
Removed NO Conc:	50.04	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	74.90	ppm	NH3 Gas Cylinder #:	CC744566
			NH3 Cal Gas Expiry:	December 21, 2023
Removed NH3 Conc:	74.90	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	Teledyne API T700		Serial Number:	3565
ZAG make/model:	Teledyne API T701		Serial Number:	5609

### Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	475
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	5.00
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	512

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.987	0.996	TN coefficient:	0.989	1.001
NOX coefficient:	0.987	0.999	NO bkgnd:	-0.9	-0.9
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-0.3	-0.3
NH3 coefficient:	0.973	0.995	TN bkgnd:	1.2	1.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.998131	0.999916
NO <sub>x</sub> Cal Offset:	-2.040000	-2.000000
NO Cal Slope:	0.998944	0.999929
NO Cal Offset:	-2.460000	-2.380000
NO <sub>2</sub> Cal Slope:	0.999734	0.997369
NO <sub>2</sub> Cal Offset:	0.074243	0.086604
NH3 Cal Slope:	1.003610	1.003196
NH3 Cal Offset:	0.576979	-1.081485
TN Cal Slope:	1.006104	1.005635
TN Cal Offset:	0.744017	-0.793819



# Wood Buffalo Environmental Association

## TN - NO<sub>x</sub> - NH<sub>3</sub> Calibration Report

Version-05-2023

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NH <sub>3</sub> concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NH <sub>3</sub> concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH <sub>3</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
as found NO	4920	80.0	813.4	813.4	----	816.0	815.0	0.9	0.997	----
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
high NO point	4920	80.0	813.4	813.4	----	812.9	813.0	-0.2	1.001	----
NO/O <sub>3</sub> point	4920	80.0	813.4	813.4	----	814.1	811.0	3.6	0.999	----
as found NH <sub>3</sub>	3416	84.1	1799.7	----	1799.7	1844.9	----	1839.9	0.976	0.978
new NH <sub>3</sub> cyl rp							----			
first NH <sub>3</sub>	3416	84.1	1799.7	----	1799.7	1809.9	----	1805.4	0.994	0.997
second NH <sub>3</sub>	3453	46.7	999.4	----	999.4	1003.0	----	1000.0	0.996	0.999
third NH <sub>3</sub>	3477	23.4	500.8	----	500.8	502.3	----	500.7	0.997	1.000
<b>Average Correction Factor</b>									<b>0.9999</b>	<b>0.9988</b>

Corrected As found      TN = 816.3 ppb      NO<sub>x</sub> = 815.3 ppb      NH<sub>3</sub> = 1839.9 ppb

Previous Response      TN = 819.1 ppb      NO<sub>x</sub> = 809.9 ppb      NH<sub>3</sub> = 1806.8 ppb

NH<sub>3</sub> Previous Converter Efficiency = 97.3%

NH<sub>3</sub> Current Converter Efficiency = 99.5%

\*Percent Change      TN = -0.3%

\*Percent Change      NO<sub>x</sub> = 0.7%

\*Percent Change      NH<sub>3</sub> = 1.8%

\* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## NO<sub>x</sub> - NO - NO<sub>2</sub> Calibration Report

Version-05-2023

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.2	-0.6	----	----
as found span	4920	80.0	813.4	800.6	813.4	805.0	793.3	807.5	1.0105	1.0093
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.3	0.1	----	----
high point	4920	80.0	813.4	800.6	813.4	813.0	799.8	812.9	1.0005	1.0011
second point	4960	40.0	406.7	400.3	406.7	401.9	395.8	402.2	1.0120	1.0114
third point	4980	20.0	203.4	200.2	203.4	200.4	195.6	200.2	1.0148	1.0233
Average Correction Factor									1.0091	1.0119

Baseline Corr As fnd	TN = 808.1 ppb	NO <sub>x</sub> = 805.4 ppb	NO = 793.5 ppb	*Percent Change	TN = -1.4%
Previous Response	TN = 819.1 ppb	NO <sub>x</sub> = 809.9 ppb	NO = 797.3 ppb	*Percent Change	NO <sub>x</sub> = -0.6%
				*Percent Change	NO = -0.5%
				<i>* = &gt; +/-5% change initiates investigation</i>	

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	-0.2	----	----
calibration zero	----	----	0.0	-0.2	----	----
1st GPT point (400 ppb O3)	795.8	395.2	413.4	411.9	1.0036	99.6%
2nd GPT point (200 ppb O3)	795.8	595.5	213.1	213.8	0.9967	100.3%
3rd GPT point (100 ppb O3)	795.8	695.4	113.2	112.6	1.0053	99.5%
Average Correction Factor					1.0019	99.8%

Notes: Changed the inlet filter after as founds. Adjusted NO<sub>x</sub>/NO/TN span. Adjusted NH<sub>3</sub> span.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## TN Calibration Summary

Version-05-2023


### Station Information

Calibration Date:	September 8, 2023	Previous Calibration:	August 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:23	End Time (MST):	14:39
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
1799.7	1809.9	0.9944			
999.4	1003.0	0.9964			
500.8	502.3	0.9969			
			Slope	1.005635	0.90 - 1.10
			Intercept	-0.793819	+/-20

**TN Calibration Curve**





# Wood Buffalo Environmental Association

## NH<sub>3</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	September 8, 2023	Previous Calibration:	August 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:23	End Time (MST):	14:39
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1799.7	1805.4	0.9969		
999.4	1000.0	0.9994		
500.8	500.7	1.0001		
			0.999997	
			1.003196	
			-1.081485	





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:23	End Time (MST):	14:39
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
813.4	813.0	1.0005			
406.7	401.9	1.0120			
203.4	200.4	1.0148			
			Slope	0.999916	0.90 - 1.10
			Intercept	-2.000000	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-05-2023


### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:23	End Time (MST):	14:39
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	≥0.995	
800.6	799.8	1.0011			
400.3	395.8	1.0114			
200.2	195.6	1.0233			
			Slope	0.999929	0.90 - 1.10
			Intercept	-2.380000	+/-20

**NO Calibration Curve**







# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:23	End Time (MST):	14:39
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
413.4	411.9	1.0036			
213.1	213.8	0.9967			
113.2	112.6	1.0053			
			Slope	0.997369	0.90 - 1.10
			Intercept	0.086604	+/-20



NO<sub>x</sub> Calibration Plot

Date: September 7, 2023


Location: Bertha Ganter-Fort McKay



# NH<sub>3</sub> Calibration Plot

Date: September 8, 2023

Location: Bertha Ganter-Fort McKay





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS02 MILDRED LAKE SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	October 20, 2023	Last Cal Date:	September 25, 2023
Start time (MST):	9:27	End time (MST):	12:30
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1185
ZAG Make/Model:	API T701		Serial Number:	4891

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995496	0.999174	Backgd or Offset:	18.1	17.5
Calibration intercept:	0.154391	0.714964	Coeff or Slope:	0.797	0.774

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4920	80.2	801.6	813.0	0.986
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.7	----
high point	4920	80.2	801.6	801.7	1.000
second point	4960	40.1	400.8	401.2	0.999
third point	4980	20.0	199.9	200.5	0.997
as left zero	5000	0.0	0.0	0.5	----
as left span	4920	80.2	801.6	805.2	0.996
Average Correction Factor					0.999

Baseline Corr As found:	812.60	Previous response	798.19	*% change	1.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 25, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:27	End Time (MST):	12:30
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient	1.000000	≥0.995
801.6	801.7	0.9999			
400.8	401.2	0.9991	Slope	0.999174	0.90 - 1.10
199.9	200.5	0.9971			
			Intercept	0.714964	+/-30



SO2 Calibration Plot

Date: October 20, 2023

Location: Mildred Lake





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Mildred Lake      Station number: AMS02  
 Calibration Date: October 25, 2023      Last Cal Date: September 20, 2023  
 Start time (MST): 6:45      End time (MST): 10:54  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.29 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC345191  
 Removed Cal Gas Conc: 5.29 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 1185  
 ZAG Make/Model: API T701      Serial Number: 5608

### Analyzer Information

Analyzer make: Thermo 43iQTL      Analyzer serial #: 12113311966  
 Converter make: Global G150      Converter serial #: 2022-198  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004536	1.002536	Backgd or Offset: 1.78	1.71
Calibration intercept:	0.220806	-0.059198	Coeff or Slope: 0.781	0.754

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4924	75.6	80.0	83.2	0.961
as found 2nd point	4962	37.8	40.0	41.6	0.961
as found 3rd point	4981	18.9	20.0	20.6	0.971
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4924	75.6	80.0	80.1	0.999
second point	4962	37.8	40.0	40.2	0.995
third point	4981	18.9	20.0	19.8	1.010
as left zero	5000	0.0	0.0	0.1	----
as left span	4924	75.6	80.0	80.7	0.991
SO2 Scrubber Check	4920	80.2	802.0	0.1	----
Date of last scrubber change:	20-Sep-23			Ave Corr Factor	1.001
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 83.2      Prev response: 80.57      \*% change: 3.2%  
 Baseline Corr 2nd AF pt: 41.6      AF Slope: 1.040969      AF Intercept: -0.079168  
 Baseline Corr 3rd AF pt: 20.6      AF Correlation: 0.999993

\* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrated zero. Span adjusted.

Calibration Performed By: Melissa Lemay





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	October 25, 2023	Previous Calibration:	September 20, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	6:45	End Time (MST):	10:54
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999981	
80.0	80.1	0.9986			≥0.995
40.0	40.2	0.9949	Slope	1.002536	
20.0	19.8	1.0099			0.90 - 1.10
			Intercept	-0.059198	+/-3



H<sub>2</sub>S Calibration Plot

Date: October 25, 2023

Location: Mildred Lake





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	October 20, 2023	Last Cal Date:	September 25, 2023
Start time (MST):	9:27	End time (MST):	12:30
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH <sub>4</sub> Cal Gas Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1048.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1048.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.99E-04	3.04E-04	NMHC SP Ratio:	4.58E-05	4.59E-05
CH <sub>4</sub> Retention time:	14.6	14.8	NMHC Peak Area:	192179	191592
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	16.67	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.76	1.003
second point	4960	40.1	8.41	8.36	1.006
third point	4980	20.0	4.19	4.16	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.78	1.002
Average Correction Factor					1.006

Baseline Corr AF:	16.67	Prev response	16.78	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	8.78	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.79	1.001
second point	4960	40.1	4.40	4.41	0.997
third point	4980	20.0	2.19	2.21	0.991
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.83	0.997
Average Correction Factor					0.996
Baseline Corr AF:	8.78	Prev response	8.79	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	7.88	1.018
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	7.97	1.007
second point	4960	40.1	4.01	3.94	1.018
third point	4980	20.0	2.00	1.95	1.025
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	7.96	1.008
Average Correction Factor					1.017
Baseline Corr AF:	7.88	Prev response	7.99	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998613	0.996697
THC Cal Offset:	-0.013917	-0.011323
CH <sub>4</sub> Cal Slope:	0.999502	0.994304
CH <sub>4</sub> Cal Offset:	-0.025057	-0.023065
NMHC Cal Slope:	0.998932	0.998828
NMHC Cal Offset:	0.008542	0.011943

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 25, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:27	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	$\geq 0.995$			
16.82	16.76	1.0035						
8.41	8.36	1.0065				Slope	0.996697	0.90 - 1.10
4.19	4.16	1.0073						
			Intercept	-0.011323	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 25, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:27	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999954	$\geq 0.995$
8.02	7.97	1.0066			
4.01	3.94	1.0179			
2.00	1.95	1.0250			
			Slope	0.994304	0.90 - 1.10
			Intercept	-0.023065	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 25, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:27	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999991	$\geq 0.995$			
8.80	8.79	1.0006						
4.40	4.41	0.9965				Slope	0.998828	0.90 - 1.10
2.19	2.21	0.9911						
			Intercept	0.011943	$\pm 0.5$			



NMHC Calibration Plot

Date: October 20, 2023

Location: Mildred Lake







## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS04 BUFFALO VIEWPOINT SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	September 13, 2023	Last Cal Date:	August 9, 2023
Start time (MST):	8:35	End time (MST):	11:20
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.87	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC446753			
Removed Cal Gas Conc:	50.87	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701		Serial Number:	362

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999657	0.994771	Backgd or Offset:	22.1	22.1
Calibration intercept:	0.974700	2.033465	Coeff or Slope:	0.860	0.860

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4921	78.6	799.7	800.5	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	1.0	----
high point	4921	78.6	799.7	796.3	1.004
second point	4961	39.3	399.8	402.6	0.993
third point	4980	19.6	199.4	199.9	0.998
as left zero	5000	0.0	0.0	0.8	----
as left span	4921	78.6	799.7	798.4	1.002
Average Correction Factor					0.998

Baseline Corr As found:	800.10	Previous response	800.44	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 13, 2023	Previous Calibration:	August 9, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:35	End Time (MST):	11:20
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.0	----	Correlation Coefficient	0.999968	≥0.995
799.7	796.3	1.0043			
399.8	402.6	0.9931	Slope	0.994771	0.90 - 1.10
199.4	199.9	0.9976			
			Intercept	2.033465	+/-30



SO2 Calibration Plot

Date: September 13, 2023

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Buffalo Viewpoint      Station number: AMS04  
 Calibration Date: September 1, 2023      Last Cal Date: August 2, 2023  
 Start time (MST): 6:15      End time (MST): 10:13  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.42 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC345266  
 Removed Cal Gas Conc: 5.42 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3808  
 ZAG Make/Model: API T701H      Serial Number: 362

### Analyzer Information

Analyzer make: Thermo 43i-LTE      Analyzer serial #: 1008841400  
 Converter make: Global      Converter serial #: 2022-200  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000770	1.004754	Backgd or Offset:	1.8
Calibration intercept:	-0.097733	-0.037663	Coeff or Slope:	1.095

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4926	74.1	80.3	80.9	0.992
as found 2nd point	4963	37.0	40.1	40.5	0.988
as found 3rd point	4982	18.5	20.1	20.1	0.993
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4926	74.1	80.3	80.6	0.997
second point	4963	37.0	40.1	40.4	0.993
third point	4982	18.5	20.1	20.1	0.998
as left zero	5000	0.0	0.0	0.0	----
as left span	4926	74.1	80.3	80.1	1.003
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	0.996
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 81.0      Prev response: 80.29      \*% change: 0.9%  
 Baseline Corr 2nd AF pt: 40.6      AF Slope: 1.008738      AF Intercept: -0.077596  
 Baseline Corr 3rd AF pt: 20.2      AF Correlation: 0.999995

\* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 1, 2023	Previous Calibration:	August 2, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:15	End Time (MST):	10:13
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999992	≥0.995
80.3	80.6	0.9966			
40.1	40.4	0.9928	Slope	1.004754	0.90 - 1.10
20.1	20.1	0.9976			
			Intercept	-0.037663	+/-3



H<sub>2</sub>S Calibration Plot

Date: September 1, 2023

Location: Buffalo Viewpoint







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	September 13, 2023	Last Cal Date:	August 9, 2023
Start time (MST):	8:35	End time (MST):	11:19
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	NA	Cal Gas Expiry Date:	March 10, 1931
CH4 Cal Gas Conc.	497.2 ppm	CH4 Equiv Conc.	1058.2 ppm
C3H8 Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	497.2 ppm	CH4 Equiv Conc.	1058.2 ppm
Removed C3H8 Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	1.86E-04	1.91E-04	NMHC SP Ratio:	3.96E-05	4.09E-05
CH4 Retention time:	11.8	11.8	NMHC Peak Area:	222570	215814
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON	ON

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	16.64	16.16	1.029
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	16.64	16.62	1.001
second point	4961	39.3	8.32	8.28	1.004
third point	4980	19.6	4.15	4.12	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	16.64	16.63	1.000
Average Correction Factor					1.004

Baseline Corr AF:	16.16	Prev response	16.64	*% change	-2.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	8.82	8.55	1.032
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	8.82	8.81	1.001
second point	4961	39.3	4.41	4.40	1.002
third point	4980	19.6	2.20	2.19	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	8.82	8.80	1.002
Average Correction Factor					1.002
Baseline Corr AF:	8.55	Prev response	8.82	*% change	-3.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	7.82	7.61	1.027
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	7.82	7.82	1.000
second point	4961	39.3	3.91	3.88	1.007
third point	4980	19.6	1.95	1.94	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	7.82	7.83	0.998
Average Correction Factor					1.004
Baseline Corr AF:	7.61	Prev response	7.82	*% change	-2.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000360	0.999330
THC Cal Offset:	-0.005535	-0.015543
CH <sub>4</sub> Cal Slope:	1.000736	1.000443
CH <sub>4</sub> Cal Offset:	-0.005903	-0.009902
NMHC Cal Slope:	1.000144	0.999120
NMHC Cal Offset:	-0.000833	-0.003636

Notes: Span adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 13, 2023	Previous Calibration:	August 9, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:35	End Time (MST):	11:19
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	$\geq 0.995$			
16.64	16.62	1.0010						
8.32	8.28	1.0045				Slope	0.999330	0.90 - 1.10
4.15	4.12	1.0069						
			Intercept	-0.015543	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 13, 2023	Previous Calibration:	August 9, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:35	End Time (MST):	11:19
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999983	<b>≥0.995</b>
7.82	7.82	0.9996			
3.91	3.88	1.0072			
1.95	1.94	1.0047			
			Slope	1.000443	<b>0.90 - 1.10</b>
			Intercept	-0.009902	<b>+/-0.5</b>





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 13, 2023	Previous Calibration:	August 9, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:35	End Time (MST):	11:19
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥0.995
8.82	8.81	1.0011			
4.41	4.40	1.0021	Slope	0.999120	0.90 - 1.10
2.20	2.19	1.0042			
			Intercept	-0.003636	+/-0.5



NMHC Calibration Plot

Date: September 13, 2023

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Buffalo Viewpoint  
Calibration Date: September 8, 2023  
Start time (MST): 5:45  
Reason: Routine  
Station number: AMS04  
Last Cal Date: August 25, 2023  
End time (MST): 10:53

### Calibration Standards

NO Gas Cylinder #: T36RH1F  
NOX Cal Gas Conc: 51.16 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 51.16 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701  
Cal Gas Expiry Date: August 18, 2023  
NO Cal Gas Conc: 50.91 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 50.91 ppm  
NO gas Diff:  
Serial Number: 3808  
Serial Number: 362

### Analyzer Information

Analyzer make: API T200  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 721

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.121	1.164	NO bkgnd or offset:	-0.6	-0.6
NOX coeff or slope:	1.119	1.159	NOX bkgnd or offset:	-0.3	-0.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.4	4.4

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000803	1.001176
NO <sub>x</sub> Cal Offset:	0.426473	0.046144
NO Cal Slope:	1.000459	1.001222
NO Cal Offset:	-0.993702	-0.434177
NO <sub>2</sub> Cal Slope:	0.991739	0.996415
NO <sub>2</sub> Cal Offset:	2.285236	-1.016897



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.8	0.2	0.6	----	----
as found span	4922	78.1	799.1	795.2	3.9	767.9	762.5	5.4	1.0406	1.0429
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.5	0.4	0.1	----	----
high point	4922	78.1	799.1	795.2	3.9	800.2	795.9	4.3	0.9986	0.9991
second point	4961	39.1	400.1	398.1	2.0	400.7	398.5	2.2	0.9984	0.9990
third point	4981	19.5	199.5	198.5	1.0	199.1	197.0	2.1	1.0020	1.0078
as left zero	5000	0.0	0.0	0.0	0.0	0.7	1.1	-0.4	----	----
as left span	4922	78.1	799.1	384.1	415.0	794.9	385.1	409.9	1.0053	0.9974
Average Correction Factor									0.9997	1.0020

Corrected As found	NO <sub>x</sub> = 767.1 ppb	NO = 762.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -4.3%
Previous Response	NO <sub>x</sub> = 800.2 ppb	NO = 794.6 ppb		*Percent Change	NO = -4.2%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.8	385.7	415.0	412.7	1.0056	99.4%
2nd GPT point (200 ppb O3)	796.8	586.5	214.2	213.0	1.0057	99.4%
3rd GPT point (100 ppb O3)	796.8	691.5	109.2	105.9	1.0312	97.0%
Average Correction Factor					1.0141	98.6%

Notes: No Maintenance Done. Span adjusted. Due to drifting during the GPT the 2nd NO ref point used.

Calibration Performed By: Melissa Lemay





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 8, 2023	Previous Calibration:	August 25, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	5:45	End Time (MST):	10:53
Analyzer make:	API T200	Analyzer serial #:	721

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	≥0.995	
799.1	800.2	0.9986			
400.1	400.7	0.9984			
199.5	199.1	1.0020			
			Slope	1.001176	0.90 - 1.10
			Intercept	0.046144	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 8, 2023	Previous Calibration:	August 25, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	5:45	End Time (MST):	10:53
Analyzer make:	API T200	Analyzer serial #:	721

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.4	----	Correlation Coefficient	≥0.995	
795.2	795.9	0.9991			
398.1	398.5	0.9990			
198.5	197.0	1.0078			
			Slope	1.001222	0.90 - 1.10
			Intercept	-0.434177	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 8, 2023	Previous Calibration:	August 25, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	5:45	End Time (MST):	10:53
Analyzer make:	API T200	Analyzer serial #:	721

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
415.0	412.7	1.0056		
214.2	213.0	1.0057		
109.2	105.9	1.0312		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 8, 2023

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Buffalo Viewpoint      Station number: AMS04  
 Calibration Date: September 13, 2023      Last Cal Date: August 9, 2023  
 Start time (MST): 6:30      End time (MST): 8:38  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: API T700      Serial Number: 3808  
 ZAG Make/Model: API T701      Serial Number: 362

### Analyzer Information

Analyzer make: API T400      Analyzer serial #: 2961  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996114	0.997486	Backgd or Offset:	-2.6	-2.6
Calibration intercept:	0.480000	-0.460000	Coeff or Slope:	1.008	1.008

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.9	----
high point	5000	984.5	400.0	398.2	1.005
second point	5000	816.9	200.0	199.6	1.002
third point	5000	707.7	100.0	99.5	1.005
as left zero	5000	0.0	0.0	0.3	----
as left span	5000	989.6	400.0	399.2	1.002
Average Correction Factor					1.004

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No as founds as pump had died and flow is at 0.0cc/min. Pump replaced. No adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 13, 2023	Previous Calibration:	August 9, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:30	End Time (MST):	8:38
Analyzer make:	API T400	Analyzer serial #:	2961

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.9	----	Correlation Coefficient	0.999992	≥0.995
400.0	398.2	1.0045			
200.0	199.6	1.0020	Slope	0.997486	0.90 - 1.10
100.0	99.5	1.0050			
			Intercept	-0.460000	+/- 5



O<sub>3</sub> Calibration Plot

Date: September 13, 2023

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04  
 Calibration Date: September 13, 2023 Last Cal Date: August 31, 2023  
 Start time (MST): 7:09 End time (MST): 8:13

Analyzer Make: API T640 S/N: 322  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Deltacal S/N: 1451  
 Temp/RH standard: Deltacal S/N: 1451

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	11.1	11.3	11.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	723	724.5	723	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.03	4.93	5.03	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 13, 2023</u>	Last Cal Date: <u>August 31, 2023</u>			
	PM w/o HEPA: <u>200</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.5	10.5	10.5	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>21.8</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>September 13, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 13, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: May 23, 2023  
 Date RH/T Sensor Cleaned: May 23, 2023

No adjustments done. Leak check done before and after cleaning.

Notes:

Calibration by: Melissa Lemay





# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Calibration Date:	September 29, 2023	Prev Cal Date:	October 6, 2022
Start Time (MST):	6:48	End Time (MST):	8:13
Tower Height (m):	10m	Reason:	Removal

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	Y4520
WS Calibrator:	MetOne 053	Serial Number:	R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.7	0.2%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	0.999999	0.999998	<b>≥0.9995</b>
Calculated slope	0.999979	0.998443	<b>0.90 - 1.10</b>
Calculated intercept	-0.013577	0.026636	<b>+/- 2</b>

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	V11346
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	12:12:43	Calc Declination*:	13.67 Degrees
Deadband calc:	-1.4 degrees ( <i>Limit 4 deg</i> )	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.0	---
90	90.5	0.1%
180	179.4	-0.2%
270	269.7	-0.1%
357	358.4	0.4%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	0.999994	0.999987	<b>≥0.9995</b>
Calculated slope	1.000909	0.997772	<b>0.90 - 1.10</b>
Calculated intercept	-0.427281	0.200128	<b>+/- 4</b>

Notes: WS was removed, WD was just calibrated.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Calibration Date:	September 29, 2023	Prev Cal Date:	October 6, 2022
Start Time (MST):	6:48	End Time (MST):	8:13
Tower Height (m):	10m	Reason:	Install

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	Y3176
WS Calibrator:	MetOne 053	Serial Number:	R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.2	0.2%
400	39.4	39.4	0.1%
600	58.6	58.8	0.4%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999996	$\geq 0.9995$
Calculated slope		0.998429	<i>0.90 - 1.10</i>
Calculated intercept		-0.012747	<i>+/- 2</i>

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	V11346
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	12:12:43	Calc Declination*:	13.67 Degrees
Deadband calc:	-1.4 degrees ( <i>Limit 4 deg</i> )	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.0	---
90	90.5	0.1%
180	179.4	-0.2%
270	269.7	-0.1%
357	358.4	0.4%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	0.999994	0.999987	$\geq 0.9995$
Calculated slope	1.000909	0.997772	<i>0.90 - 1.10</i>
Calculated intercept	-0.427281	0.200128	<i>+/- 4</i>

Notes: Old WS removed, New WS installed.

Calibration Performed By: Melissa Lemay



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS05  
MANNIX  
SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Mannix	Station number:	AMS05
Calibration Date:	September 15, 2023	Last Cal Date:	August 15, 2023
Start time (MST):	9:07	End time (MST):	12:30
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.02	ppm	Cal Gas Exp Date:	January 12, 2029
Cal Gas Cylinder #:	XC026809B			
Removed Cal Gas Conc:	50.02	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	621
ZAG Make/Model:	API T701H		Serial Number:	832

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999257	0.998615	Backgd or Offset:	8.9	8.9
Calibration intercept:	0.220000	0.320000	Coeff or Slope:	0.930	0.930

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4920	80.0	800.3	797.9	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.7	----
high point	4920	80.0	800.3	799.2	1.001
second point	4960	40.0	400.2	401.3	0.997
third point	4980	20.0	200.1	198.7	1.007
as left zero	5000	0.0	0.0	0.8	----
as left span	4920	80.0	800.3	802.8	0.997
Average Correction Factor					1.002

Baseline Corr As found:	797.50	Previous response	799.95	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 15, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:07	End Time (MST):	12:30
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.7	----	Correlation Coefficient	≥0.995
800.3	799.2	1.0014		
400.2	401.3	0.9972	Slope	0.90 - 1.10
200.1	198.7	1.0069		
			Intercept	+/-30



SO2 Calibration Plot

Date: September 15, 2023

Location: Mannix





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Mannix	Station number:	AMS05
Calibration Date:	September 18, 2023	Last Cal Date:	August 22, 2023
Start time (MST):	9:12	End time (MST):	13:50
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	4.92	ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	EY0002433			
Removed Cal Gas Conc:	4.92	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1845
ZAG Make/Model:	API T701H		Serial Number:	832

### Analyzer Information

Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1203169745
Converter make:	Global	Converter serial #:	2022-196
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000615	0.994760	Backgd or Offset:	2.18
Calibration intercept:	0.200559	0.380518	Coeff or Slope:	0.866
				0.866

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4919	81.3	80.0	84.4	0.949
as found 2nd point	4960	40.7	40.0	40.5	0.991
as found 3rd point	4980	20.3	20.0	20.7	0.970
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4919	81.3	80.0	79.8	1.002
second point	4960	40.7	40.0	40.5	0.989
third point	4980	20.3	20.0	20.3	0.984
as left zero	5000	0.0	0.0	0.5	----
as left span	4919	81.3	80.0	77.3	1.035
SO2 Scrubber Check	4920	80.0	800.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.992
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	84.3	Prev response:	80.24	*% change:	4.8%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.053162	AF Intercept:	-0.438733
Baseline Corr 3rd AF pt:	20.6	AF Correlation:	0.999441		

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 18, 2023	Previous Calibration:	August 22, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:12	End Time (MST):	13:50
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1203169745

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999959	≥0.995
80.0	79.8	1.0024			
40.0	40.5	0.9887	Slope	0.994760	0.90 - 1.10
20.0	20.3	0.9839			
			Intercept	0.380518	+/-3






H<sub>2</sub>S Calibration Plot

Date: September 18, 2023

Location: Mannix





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	September 8, 2023	Last Cal Date:	August 15, 2023
Start time (MST):	12:43	End time (MST):	14:16
Reason:	Cylinder Change N2 Cylinder change		

### Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH4 Cal Gas Conc.	504.9 ppm	CH4 Equiv Conc.	1076.6 ppm
C3H8 Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	504.9 ppm	CH4 Equiv Conc.	1076.6 ppm
Removed C3H8 Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.62E-05	2.62E-05	NMHC SP Ratio:	4.44E-05	4.44E-05
CH4 Retention time:	15.00	15.00	NMHC Peak Area:	206221	206221
Zero Chromatogram:	ON	OFF	Flat Baseline:	OFF	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.35	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.47	0.986
second point					
third point					
as left zero					
as left span					
Average Correction Factor					0.986

Baseline Corr AF:	17.35	Prev response	17.22	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	9.15	9.25	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.25	0.989
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.989
Baseline Corr AF:	9.25	Prev response	9.16	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	8.10	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.22	0.983
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.983
Baseline Corr AF:	8.10	Prev response	8.07	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998929	1.014049
THC Cal Offset:	0.017200	0.000000
CH <sub>4</sub> Cal Slope:	0.999731	1.017033
CH <sub>4</sub> Cal Offset:	-0.006600	0.000000
NMHC Cal Slope:	0.998220	1.010976
NMHC Cal Offset:	0.023800	0.000000


Notes: Changed the N2 cylinder.

Calibration Performed By: Max Farrell

NMHC Calibration Plot

Date: September 8, 2023

Location: Mannix





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	September 15, 2023	Last Cal Date:	August 15, 2023
Start time (MST):	9:07	End time (MST):	12:30
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH <sub>4</sub> Cal Gas Conc.	504.9 ppm	CH <sub>4</sub> Equiv Conc.	1076.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	504.9 ppm	CH <sub>4</sub> Equiv Conc.	1076.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.62E-05	2.62E-05	NMHC SP Ratio:	4.44E-05	4.44E-05
CH <sub>4</sub> Retention time:	15.00	15.00	NMHC Peak Area:	206221	206221
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.23	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.08	1.009
second point	4960	40.0	8.61	8.53	1.009
third point	4980	20.0	4.31	4.27	1.008
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.23	1.000
Average Correction Factor					1.009

Baseline Corr AF:	17.23	Prev response	17.22	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	9.15	9.16	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.07	1.009
second point	4960	40.0	4.57	4.55	1.005
third point	4980	20.0	2.29	2.29	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.15	9.17	0.997
Average Correction Factor					1.005
Baseline Corr AF:	9.16	Prev response	9.16	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	8.07	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.01	1.008
second point	4960	40.0	4.04	3.99	1.014
third point	4980	20.0	2.02	1.99	1.017
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.06	1.003
Average Correction Factor					1.013
Baseline Corr AF:	8.07	Prev response	8.07	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998929	0.991126
THC Cal Offset:	0.017200	0.001000
CH <sub>4</sub> Cal Slope:	0.999731	0.992078
CH <sub>4</sub> Cal Offset:	-0.006600	-0.010800
NMHC Cal Slope:	0.998220	0.990661
NMHC Cal Offset:	0.023800	0.011800

Notes: Changed the inlet filter and the H2 cylinder after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 15, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:07	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	$\geq 0.995$			
17.23	17.08	1.0088						
8.61	8.53	1.0093				Slope	0.991126	0.90 - 1.10
4.31	4.27	1.0078						
			Intercept	0.001000	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 15, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:07	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999990	$\geq 0.995$
8.08	8.01	1.0084			
4.04	3.99	1.0136			
2.02	1.99	1.0169			
			Slope	0.992078	0.90 - 1.10
			Intercept	-0.010800	+/-0.5







# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 15, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:07	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	$\geq 0.995$			
9.15	9.07	1.0089						
4.57	4.55	1.0048				Slope	0.990661	0.90 - 1.10
2.29	2.29	1.0000						
			Intercept	0.011800	$\pm 0.5$			



NMHC Calibration Plot

Date: September 15, 2023

Location: Mannix





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS06**  
**PATRICIA MCINNES**  
**SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 17, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:23	End Time (MST):	12:30
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999982	≥0.995
799.5	800.9	0.9982			
400.2	402.7	0.9939	Slope	1.000605	0.90 - 1.10
200.1	203.6	0.9829			
			Intercept	1.660053	+/-30



SO2 Calibration Plot

Date: September 14, 2023

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	September 22, 2023	Last Cal Date:	August 28, 2023
Start time (MST):	9:15	End time (MST):	13:48
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.38	ppm	Cal Gas Exp Date:	March 2, 2023
Cal Gas Cylinder #:	EY0000809			
Removed Cal Gas Conc:	5.38	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

### Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.988760	1.004636	Backgd or Offset:	2.06	2.05
Calibration intercept:	-0.002363	0.217058	Coeff or Slope:	1.187	1.189

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.3	79.9	76.4	1.046
as found 2nd point	4963	37.2	40.0	38.2	1.048
as found 3rd point	4981	18.6	20.0	19.4	1.032
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	74.3	79.9	80.5	0.993
second point	4963	37.2	40.0	40.4	0.991
third point	4981	18.6	20.0	20.5	0.976
as left zero	5000	0.0	0.0	0.2	----
as left span	4926	74.3	79.9	81.1	0.986
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	0.987
Date of last converter efficiency test:	efficiency				

Baseline Corr As found:	76.4	Prev response:	79.04	*% change:	-3.5%
Baseline Corr 2nd AF pt:	38.2	AF Slope:	0.954453	AF Intercept:	0.098235
Baseline Corr 3rd AF pt:	19.4	AF Correlation:	0.999981		

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 22, 2023	Previous Calibration:	August 28, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:15	End Time (MST):	13:48
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999987	
79.9	80.5	0.9931			≥0.995
40.0	40.4	0.9907	Slope	1.004636	
20.0	20.5	0.9764			0.90 - 1.10
			Intercept	0.217058	+/-3






TRS Calibration Plot

Date: September 22, 2023

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	September 14, 2023	Last Cal Date:	August 17, 2023
Start time (MST):	9:23	End time (MST):	12:30
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH <sub>4</sub> Cal Gas Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	4602

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.11E-04	2.12E-04	NMHC SP Ratio:	4.72E-05	4.80E-05
CH <sub>4</sub> Retention time:	14.0	14.0	NMHC Peak Area:	192120	189064
Zero Chromatogram:	OFF		Flat Baseline:	OFF	

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4920	80.3	17.12	17.01	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	17.15	0.998
second point	4960	40.2	8.57	8.62	0.994
third point	4980	20.1	4.29	4.37	0.980
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.12	17.14	0.999
Average Correction Factor					0.991

Baseline Corr AF:	17.00	Prev response	17.06	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919.7	80.3	9.07	8.98	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	9.07	9.09	0.997
second point	4960	40.2	4.54	4.59	0.988
third point	4980	20.1	2.27	2.35	0.967
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	9.07	9.09	0.997
Average Correction Factor					0.984
Baseline Corr AF:	8.97	Prev response	9.02	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919.7	80.3	8.06	8.03	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	8.06	8.06	0.999
second point	4960	40.2	4.03	4.03	1.001
third point	4980	20.1	2.02	2.03	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	8.06	8.05	1.001
Average Correction Factor					0.999
Baseline Corr AF:	8.03	Prev response	8.03	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.993295	1.000464
THC Cal Offset:	0.049653	0.038421
CH <sub>4</sub> Cal Slope:	0.996285	1.000059
CH <sub>4</sub> Cal Offset:	0.004651	0.001844
NMHC Cal Slope:	0.990348	1.000823
NMHC Cal Offset:	0.045403	0.036578

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 17, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:23	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999976	$\geq 0.995$			
17.12	17.15	0.9982						
8.57	8.62	0.9943				Slope	1.000464	0.90 - 1.10
4.29	4.37	0.9801						
			Intercept	0.038421	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 17, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:23	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	$\geq 0.995$			
8.06	8.06	0.9995						
4.03	4.03	1.0014				Slope	1.000059	0.90 - 1.10
2.02	2.03	0.9953						
			Intercept	0.001844	$\pm 0.5$			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 17, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:23	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999925	$\geq 0.995$
9.07	9.09	0.9970			
4.54	4.59	0.9880			
2.27	2.35	0.9670			
			Slope	1.000823	0.90 - 1.10
			Intercept	0.036578	+/-0.5



NMHC Calibration Plot

Date: September 14, 2023

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Patricia McInnes  
Calibration Date: September 20, 2023  
Start time (MST): 7:57  
Reason: Routine  
Station number: AMS06  
Last Cal Date: August 9, 2023  
End time (MST): 12:41

### Calibration Standards

NO Gas Cylinder #: T30YCWN  
NOX Cal Gas Conc: 47.94 ppm  
Removed Cylinder #: N/A  
Removed Gas NOX Conc: 47.94 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701  
Cal Gas Expiry Date: April 11, 2025  
NO Cal Gas Conc: 46.39 ppm  
Removed Gas Exp Date: N/A  
Removed Gas NO Conc: 46.39 ppm  
NO gas Diff:  
Serial Number: 3566  
Serial Number: 4602

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1172750022

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.815	0.825	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.986	0.987	NOX bkgnd or offset:	3.8	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.2	157.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.992318	0.997367
NO <sub>x</sub> Cal Offset:	2.277330	2.676082
NO Cal Slope:	0.995327	0.999173
NO Cal Offset:	1.003845	1.382870
NO <sub>2</sub> Cal Slope:	0.998809	1.001949
NO <sub>2</sub> Cal Offset:	-0.587092	0.485373





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.6	0.6	0.0	----	----
as found span	4914	86.2	826.5	799.7	26.7	815.7	788.9	26.8	1.0132	1.0137
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.5	0.4	0.1	----	----
high point	4914	86.2	826.5	799.7	26.7	825.9	800.0	25.9	1.0007	0.9997
second point	4957	43.1	413.2	399.9	13.4	415.8	401.3	14.4	0.9938	0.9964
third point	4978	21.6	207.1	200.4	6.7	211.5	202.7	8.7	0.9793	0.9888
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.3	0.0	----	----
as left span	4914	86.2	826.5	400.4	426.0	825.8	403.4	422.4	1.0008	0.9926
Average Correction Factor									0.9913	0.9950

Corrected As found	NO <sub>x</sub> = 815.1 ppb	NO = 788.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.9%	
Previous Response	NO <sub>x</sub> = 822.4 ppb	NO = 797.0 ppb		*Percent Change	NO = -1.1%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	800.1	400.8	426.0	427.0	0.9977	100.2%
2nd GPT point (200 ppb O3)	800.1	606.4	220.4	221.9	0.9933	100.7%
3rd GPT point (100 ppb O3)	800.1	705.9	120.9	121.8	0.9928	100.7%
Average Correction Factor					0.9946	100.5%

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 20, 2023	Previous Calibration:	August 9, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	7:57	End Time (MST):	12:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.5	----	Correlation Coefficient	≥0.995	
826.5	825.9	1.0007			
413.2	415.8	0.9938			
207.1	211.5	0.9793			
			Slope	0.997367	0.90 - 1.10
			Intercept	2.676082	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 20, 2023	Previous Calibration:	August 9, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	7:57	End Time (MST):	12:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.7	800.0	0.9997		
399.9	401.3	0.9964		
200.4	202.7	0.9888		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 20, 2023	Previous Calibration:	August 9, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	7:57	End Time (MST):	12:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
426.0	427.0	0.9977		
220.4	221.9	0.9933		
120.9	121.8	0.9928		
			0.999994	
			1.001949	
			0.485373	


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 20, 2023

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Patricia McInnes      Station number: AMS06  
 Calibration Date: September 5, 2023      Last Cal Date: August 11, 2023  
 Start time (MST): 9:53      End time (MST): 13:03  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: API T700      Serial Number: 3566  
 ZAG Make/Model: API T701      Serial Number: 4602

### Analyzer Information

Analyzer make: Thermo 49i      Analyzer serial #: 1300156234  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002629	1.005743	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	0.540000	-0.480000	Coeff or Slope:	1.019	1.019

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	-0.3	----
as found span	5000	1303.0	400.0	400.5	0.999
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	-0.2	----
high point	5000	1303.0	400.0	401.9	0.995
second point	5000	966.5	200.0	200.7	0.997
third point	5000	794.3	100.0	99.7	1.003
as left zero	5000	800.0	0.0	-0.5	----
as left span	5000	1303.0	400.0	403.4	0.992
Average Correction Factor					0.998

Baseline Corr As found:	400.8	Previous response	401.6	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 11, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:53	End Time (MST):	13:03
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999997	≥0.995
400.0	401.9	0.9953			
200.0	200.7	0.9965	Slope	1.005743	0.90 - 1.10
100.0	99.7	1.0030			
			Intercept	-0.480000	+/- 5



O<sub>3</sub> Calibration Plot

Date: September 5, 2023

Location: Patricia McInnes







# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Patricia McInnes Station number: AMS 06  
 Calibration Date: September 22, 2023 Last Cal Date: August 24, 2023  
 Start time (MST): 12:54 End time (MST): 13:53

Analyzer Make: API T640 S/N: 766  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT FP-25 S/N: 388755  
 Temp/RH standard: ALICAT FP-25 S/N: 388755

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	23.1	23.1	23.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	726.5	727.8	726.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	4.97	5.02	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: September 22, 2023 Last Cal Date: August 24, 2023  
 PM w/o HEPA: 8.5 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.0	11.0	11.0	<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: 4.7 w/ HEPA: 0  
 Date Optical Chamber Cleaned: September 22, 2023 <0.2 ug/m3  
 Disposable Filter Changed: September 22, 2023

### Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023  
 Date RH/T Sensor Cleaned: April 13, 2023

Notes: Cleaned the chamber and changed the filter. PMT peak was on target. Leak check passed, no adjustments made.

Calibration by: Max Farrell



# Wood Buffalo Environmental Association

## TN - NO<sub>x</sub> - NH<sub>3</sub> Calibration Report

Version-05-2023

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	September 12, 2023	Last Cal Date:	August 3, 2023
Start time (MST):	8:00	End time (MST):	13:45
NH3 Cal Date:	September 12, 2023	Last Cal Date:	August 3, 2023
Start time (MST):	13:45	End time (MST):	15:48
Reason:	Routine		

### Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	NA
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	77.8	ppm	NH3 Gas Cylinder #:	CC710812
			NH3 Cal Gas Expiry:	March 30, 2023
Removed NH3 Conc:	77.8	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	API T700		Serial Number:	3566
ZAG make/model:	API T701		Serial Number:	4602

### Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	152
Converter model:	Teledyne API T501	Converter serial #:	147
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.50
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	542

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.794	0.788	TN coefficient:	0.794	0.789
NOX coefficient:	0.796	0.787	NO bkgnd:	-0.1	-0.1
NO2 coefficient:	1.000	1.000	NOX bkgnd:	0.0	0.0
NH3 coefficient:	0.908	0.908	TN bkgnd:	0.0	0.0

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.003439	1.000369
NO <sub>x</sub> Cal Offset:	1.455068	1.915538
NO Cal Slope:	0.997870	1.000758
NO Cal Offset:	0.164066	0.303018
NO <sub>2</sub> Cal Slope:	1.000820	1.006122
NO <sub>2</sub> Cal Offset:	-0.026033	-2.141943
NH3 Cal Slope:	1.002680	0.985048
NH3 Cal Offset:	5.795340	7.896031
TN Cal Slope:	1.008130	0.991160
TN Cal Offset:	5.548239	7.076683



# Wood Buffalo Environmental Association

## TN - NOX - NH<sub>3</sub> Calibration Report

Version-05-2023

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.5	0.1	0.4	----	----
as found NO	4914	86.2	826.5	826.5	----	827.2	831.3	-4.2	0.999	----
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	-0.9	1.2	----	----
high NO point	4914	86.2	826.5	826.5	----	828.3	826.9	1.6	0.998	----
NO/O3 point	4914	86.2	826.5	826.5	----	829.6	826.1	3.5	0.996	----
as found NH3	3419	81.0	1800.5	----	1800.5	1786.0	----	1776.0	1.008	1.014
new NH3 cyl rp							----			
first NH3	3419	81.0	1800.5	----	1800.5	1786.0	----	1776.0	1.008	1.014
second NH3	3455	45.0	1000.3	----	1000.3	1006.5	----	1000.8	0.994	0.999
third NH3	3478	22.5	500.1	----	500.1	507.2	----	505.1	0.986	0.990
<b>Average Correction Factor</b>									<b>0.9970</b>	<b>1.0011</b>

Corrected As found    TN = 826.7 ppb    NO<sub>x</sub> = 831.2 ppb    NH3 = 1775.6 ppb

Previous Response    TN = 838.7 ppb    NO<sub>x</sub> = 830.7 ppb    NH3 = 1811.1 ppb

NH3 Previous Converter Efficiency = 90.8%

NH3 Current Converter Efficiency = 90.8%

\*Percent Change    TN = -1.5%

\*Percent Change    NO<sub>x</sub> = 0.1%

\*Percent Change    NH3 = -2.0%

\* = > +/-5% change initiates investigation





# Wood Buffalo Environmental Association

## TN Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	August 3, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:00	End Time (MST):	13:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	≥0.995	
1800.5	1786.0	1.0081			
1000.3	1006.5	0.9938			
500.1	507.2	0.9859			
			Slope	0.991160	0.90 - 1.10
			Intercept	7.076683	+/-20





# Wood Buffalo Environmental Association

## NH<sub>3</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	August 3, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:00	End Time (MST):	13:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	1.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.5	1776.0	1.0138		
1000.3	1000.8	0.9995		
500.1	505.1	0.9900		
			0.999911	
			0.985048	
			7.896031	





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	August 3, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:00	End Time (MST):	13:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.9	----	Correlation Coefficient	≥0.995	
826.5	826.9	0.9995			
413.2	417.8	0.9891			
207.1	211.2	0.9807			
			Slope	1.000369	0.90 - 1.10
			Intercept	1.915538	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-05-2023


### Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	August 3, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:00	End Time (MST):	13:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	≥0.995	
799.7	800.6	0.9989			
399.9	400.5	0.9984			
200.4	201.0	0.9971			
			Slope	1.000758	0.90 - 1.10
			Intercept	0.303018	+/-20

**NO Calibration Curve**







# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-05-2023


### Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	August 3, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:00	End Time (MST):	13:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-1.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
423.1	424.2	0.9975		
222.0	220.5	1.0069		
123.4	121.2	1.0183		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 12, 2023


Location: Patricia McInnes



# NH<sub>3</sub> Calibration Plot

Date: September 12, 2023

Location: Patricia McInnes





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS07 ATHABASCA VALLEY SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	September 6, 2023	Last Cal Date:	August 9, 2023
Start time (MST):	8:50	End time (MST):	12:35
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.52	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC282115			
Removed Cal Gas Conc:	50.52	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 701H		Serial Number:	198

### Analyzer Information

Analyzer make: Thermo 43i-LTE  
 Analyzer Range 0 - 1000 ppb  
 Analyzer serial #: 1507864683

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997338	0.992516	Backgd or Offset:	2.61	2.61
Calibration intercept:	1.503428	1.942886	Coeff or Slope:	0.834	0.834

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	79.3	801.2	794.4	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	79.3	801.2	795.6	1.007
second point	4960	39.6	400.2	401.9	0.996
third point	4980	19.8	200.1	201.1	0.995
as left zero	5000	0.0	0.0	0.1	----
as left span	4921	79.2	800.2	798.7	1.002
Average Correction Factor					0.999

Baseline Corr As found:	794.50	Previous response	800.57	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:50	End Time (MST):	12:35
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999960	≥0.995
801.2	795.6	1.0070			
400.2	401.9	0.9956	Slope	0.992516	0.90 - 1.10
200.1	201.1	0.9949			
			Intercept	1.942886	+/-30



SO2 Calibration Plot

Date: September 6, 2023

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Athabasca Valley      Station number: AMS07  
 Calibration Date: September 18, 2023      Last Cal Date: August 2, 2023  
 Start time (MST): 9:07      End time (MST): 14:06  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.25 ppm      Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC504080  
 Removed Cal Gas Conc: 5.25 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3805  
 ZAG Make/Model: API T701H      Serial Number: 198

### Analyzer Information

Analyzer make: Thermo 43i LTE      Analyzer serial #: 1180540018  
 Converter make: CDN-101      Converter serial #: 551  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005977	1.012468	Backgd or Offset: 2.33	2.33
Calibration intercept:	-0.002158	-0.002341	Coeff or Slope: 0.899	0.899

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4925	75.5	79.3	80.9	0.979
as found 2nd point	4962	37.7	39.6	40.4	0.977
as found 3rd point	4981	18.9	19.8	20.0	0.987
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4925	75.5	79.3	80.4	0.986
second point	4962	37.7	39.6	39.9	0.993
third point	4981	18.9	19.9	20.1	0.988
as left zero	5000	0.0	0.0	0.2	----
as left span	4925	75.5	79.3	80.2	0.989
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	0.989
Date of last converter efficiency test:	April 22, 2022			92.6% efficiency	

Baseline Corr As found: 81.0      Prev response: 79.74      \*% change: 1.6%  
 Baseline Corr 2nd AF pt: 40.5      AF Slope: 1.022703      AF Intercept: -0.162195  
 Baseline Corr 3rd AF pt: 20.1      AF Correlation: 0.999992

\* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar





# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 18, 2023	Previous Calibration:	August 2, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:07	End Time (MST):	14:06
Analyzer make:	CDN-101	Analyzer serial #:	551

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999983	
79.3	80.4	0.9865			≥0.995
39.6	39.9	0.9927	Slope	1.012468	
19.9	20.1	0.9879			0.90 - 1.10
			Intercept	-0.002341	+/-3



TRS Calibration Plot

Date: September 18, 2023

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	September 6, 2023	Last Cal Date:	August 9, 2023
Start time (MST):	8:50	End time (MST):	12:35
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	December 29, 2028
CH <sub>4</sub> Cal Gas Conc.	501.2 ppm	CH <sub>4</sub> Equiv Conc.	1075.1 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	501.2 ppm	CH <sub>4</sub> Equiv Conc.	1075.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	0.000256	0.000249	NMHC SP Ratio: 4.57E-05	4.17E-05
CH <sub>4</sub> Retention time:	14.2	14.0	NMHC Peak Area:	198820
				218242

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	18.03	0.944
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.03	1.000
second point	4960	39.6	8.52	8.51	1.001
third point	4980	19.8	4.26	4.29	0.993
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	16.97	1.004

Average Correction Factor				0.998
Baseline Corr AF:	18.03	Prev response	16.99	*% change 5.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.09	9.91	0.917
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	9.09	9.09	1.000
second point	4960	39.6	4.55	4.56	0.996
third point	4980	19.8	2.27	2.33	0.977
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	9.09	9.06	1.004
Average Correction Factor					0.991
Baseline Corr AF:	9.91	Prev response	9.08	*% change	8.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.94	8.13	0.977
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.94	7.94	1.000
second point	4960	39.6	3.97	3.95	1.006
third point	4980	19.8	1.98	1.96	1.013
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.94	7.91	1.004
Average Correction Factor					1.006
Baseline Corr AF:	8.13	Prev response	7.91	*% change	2.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.996468	0.999300
THC Cal Offset:	0.018660	0.010060
CH <sub>4</sub> Cal Slope:	0.996465	1.001057
CH <sub>4</sub> Cal Offset:	-0.000558	-0.015758
NMHC Cal Slope:	0.996357	0.997865
NMHC Cal Offset:	0.019418	0.025418

Notes: As found NM 9% high. Chromatograms look normal. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:50	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999996	$\geq 0.995$
17.03	17.03	1.0001			
8.52	8.51	1.0009			
4.26	4.29	0.9934			
			Slope	0.999300	0.90 - 1.10
			Intercept	0.010060	+/-0.5





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:50	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999981	≥0.995			
7.94	7.94	0.9997						
3.97	3.95	1.0063				Slope	1.001057	0.90 - 1.10
1.98	1.96	1.0132						
			Intercept	-0.015758	+/-0.5			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:50	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999959	$\geq 0.995$
9.09	9.09	1.0003			
4.55	4.56	0.9965			
2.27	2.33	0.9767			
			Slope	0.997865	0.90 - 1.10
			Intercept	0.025418	+/-0.5



NMHC Calibration Plot

Date: September 6, 2023

Location: Athabasca Valley







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Athabasca Valley  
Calibration Date: September 15, 2023  
Start time (MST): 8:45  
Reason: Routine  
Station number: AMS07  
Last Cal Date: August 14, 2023  
End time (MST): 14:00

### Calibration Standards

NO Gas Cylinder #: T2Y1KA4  
NOX Cal Gas Conc: 50.92 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 50.92 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701H  
Cal Gas Expiry Date: November 30, 2023  
NO Cal Gas Conc: 49.92 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 49.92 ppm  
NO gas Diff:  
Serial Number: 3805  
Serial Number: 198

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1160120024

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.080	1.075	NO bkgnd or offset:	7.6	7.6
NOX coeff or slope:	0.994	0.995	NOX bkgnd or offset:	7.8	7.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	211.7	210.8

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.002539	1.000426
NO <sub>x</sub> Cal Offset:	0.999372	1.179077
NO Cal Slope:	1.002515	1.000774
NO Cal Offset:	1.115453	0.875112
NO <sub>2</sub> Cal Slope:	0.998415	0.997679
NO <sub>2</sub> Cal Offset:	0.146993	1.130689



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
as found span	4920	80.2	816.7	800.7	16.0	822.0	806.0	16.5	0.9936	0.9934
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.0	----	----
high point	4920	80.2	816.7	800.7	16.0	817.5	801.6	15.8	0.9991	0.9989
second point	4960	40.1	408.4	400.4	8.0	411.0	402.6	8.5	0.9936	0.9944
third point	4980	20.0	203.7	199.7	4.0	205.4	200.9	4.6	0.9916	0.9939
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4920	80.2	816.7	394.7	422.0	825.0	400.0	425.3	0.9900	0.9867
Average Correction Factor									0.9948	0.9957

Corrected As found	NO <sub>x</sub> = 821.9 ppb	NO = 805.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.3%
Previous Response	NO <sub>x</sub> = 819.8 ppb	NO = 803.8 ppb		*Percent Change	NO = 0.3%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	801.2	395.2	422.0	421.6	1.0010	99.9%
2nd GPT point (200 ppb O3)	801.2	604.2	213.0	214.2	0.9946	100.5%
3rd GPT point (100 ppb O3)	801.2	702.4	114.8	116.9	0.9824	101.8%
Average Correction Factor					0.9927	100.7%

Notes:

Span adjusted. Used second GPT reference.

Calibration Performed By:

Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 14, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:45	End Time (MST):	14:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
816.7	817.5	0.9991		
408.4	411.0	0.9936		
203.7	205.4	0.9916		
			0.999991	
			1.000426	
			1.179077	





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 14, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:45	End Time (MST):	14:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.7	801.6	0.9989		
400.4	402.6	0.9944		
199.7	200.9	0.9939		

NO Calibration Curve





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 14, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:45	End Time (MST):	14:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999965	≥0.995
422.0	421.6	1.0010			
213.0	214.2	0.9946	Slope	0.997679	0.90 - 1.10
114.8	116.9	0.9824			
			Intercept	1.130689	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 15, 2023

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	September 20, 2023	Last Cal Date:	August 11, 2023
Start time (MST):	9:53	End time (MST):	15:20
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	T700	Serial Number:	3805
ZAG Make/Model:	T701H	Serial Number:	198

### Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023
Analyzer Range:	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999171	0.997143	Backgd or Offset:	-2.6	-2.6
Calibration intercept:	0.320000	0.600000	Coeff or Slope:	1.522	1.522

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.6	----
as found span	5000	1414.8	400.0	397.5	1.006
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.2	----
high point	5000	1415.7	400.0	399.3	1.002
second point	5000	1039.9	200.0	200.1	1.000
third point	5000	856.2	100.0	100.8	0.992
as left zero	5000	0.0	0.0	-0.9	----
as left span	5000	1416.0	400.0	400.5	0.999
Average Correction Factor					0.998

Baseline Corr As found:	398.1	Previous response	400.0	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 20, 2023	Previous Calibration:	August 11, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:53	End Time (MST):	15:20
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999995	
400.0	399.3	1.0018			≥0.995
200.0	200.1	0.9995	Slope	0.997143	
100.0	100.8	0.9921			0.90 - 1.10
			Intercept	0.600000	+/- 5






O<sub>3</sub> Calibration Plot

Date: September 20, 2023

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Athabasca Valley Station number: AMS 07  
 Calibration Date: September 27, 2023 Last Cal Date: August 28, 2023  
 Start time (MST): 12:16 End time (MST): 13:48

Analyzer Make: API T640 S/N: 871  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748  
 Temp/RH standard: Alicat FP-25BT S/N: 388748

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	21.0	20.3	21	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	728.3	737.74	728.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.07	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 27, 2023</u>	Last Cal Date: <u>August 28, 2023</u>			
	PM w/o HEPA: <u>5.3</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10	10	11.2	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>5.3</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>September 27, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>August 28, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022  
 Date RH/T Sensor Cleaned: December 5, 2022

Notes: Temp, flow and pressure checked. Leak check passed. Chamber cleaned. No issues.

Calibration by: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	September 27, 2023	Last Cal Date:	August 16, 2023
Start time (MST):	10:36	End time (MST):	13:40
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date:	December 12, 2026
Cal Gas Cylinder #:	LL66942			
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 700H		Serial Number:	198

### Analyzer Information

Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	<b><u>Start</u></b>	<b><u>Finish</u></b>		<b><u>Start</u></b>	<b><u>Finish</u></b>
Calibration slope:	0.999458	1.003398	Backgd or Offset:	4.225	4.225
Calibration intercept:	0.036542	0.092559	Coeff or Slope:	1.086	1.086

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.0	40.2	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.0	40.2	0.996
second point	4967	33.3	20.0	20.3	0.985
third point	4983	16.7	10.0	10.2	0.983
as left zero	5000	0.0	0.0	0.1	----
as left span	4933	66.7	40.0	40.2	0.996
<b>Average Correction Factor</b>					<b>0.988</b>

Baseline Corr As found:	40.06	Prev response:	40.04	*% change:	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No adjustment needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 27, 2023	Previous Calibration:	August 16, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:36	End Time (MST):	13:40
Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999956	≥0.995
40.0	40.2	0.9963			
20.0	20.3	0.9851	Slope	1.003398	0.90 - 1.10
10.0	10.2	0.9834			
			Intercept	0.092559	+/-1.5



CO Calibration Plot

Date: September 27, 2023

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Athabasca Valley	Station Number:	AMS 07
Calibration Date:	September 20, 2023	Prev Cal Date:	October 4, 2022
Start Time (MST):	11:56	End Time (MST):	14:57
Tower Height (m):	10.0	Reason:	Routine

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	G3212
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.5	-0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999999	$\geq 0.9995$
Calculated slope		0.999473	<i>0.90 - 1.10</i>
Calculated intercept		0.026227	<i>+/- 2</i>

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	R14656
As Found Declination (deg east of True North):		As Left Declination (deg east of True North):	
Solar noon time (MST):		12:17 Calc Declination*:	13.17 Degrees
Deadband calc:	6.0 degrees ( <i>Limit 4 deg</i> )		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.7	---
90	86.9	-0.9%
180	177.3	-0.8%
270	265.9	<i>-1.1%</i>
357	351.7	<i>-1.5%</i>

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999979	$\geq 0.9995$
Calculated slope		1.014605	<i>0.90 - 1.10</i>
Calculated intercept		0.332308	<i>+/- 4</i>

Notes: Wind speed sensor in good condition, wind direction sensor failed, replaced with sensor s/n:D13602. Install calibration paperwork for new wind direction sensor completed separately.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Athabasca Valley	Station Number:	AMS 07
Calibration Date:	September 20, 2023	Prev Cal Date:	October 4, 2022
Start Time (MST):	11:56	End Time (MST):	14:57
Tower Height (m):	10.0	Reason:	Install

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	
WS Calibrator:	MetOne 053	Serial Number:	

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0		---
200	20.2		
400	39.4		
600	58.6		
800	77.8		

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )			$\geq 0.9995$
Calculated slope			$0.90 - 1.10$
Calculated intercept			$\pm 2$

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	D13602
As Found Declination (deg east of True North):		As Left Declination (deg east of True North):	
Solar noon time (MST):		12:17 Calc Declination*:	13.17 Degrees
Deadband calc:	-2.6 degrees ( <i>Limit 4 deg</i> )		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.8	---
90	86.9	-0.9%
180	180.9	0.3%
270	271.2	0.3%
357	360.4	1.0%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999921	$\geq 0.9995$
Calculated slope		0.989273	$0.90 - 1.10$
Calculated intercept		1.299119	$\pm 4$

Notes: Install calibration for new Wind Direction sensor, swapped sensor S/N: R14656 with sensor S/N: D13602 during annual routine calibration.

Calibration Performed By: Aswin Sasi Kumar



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS08 FORT CHIPEWYAN SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	September 6, 2023	Last Cal Date:	August 14, 2023
Start time (MST):	10:53	End time (MST):	13:31
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC196697			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3060
ZAG Make/Model:	Teledyne API T701		Serial Number:	260

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000314	1.004228	Backgd or Offset:	4.65	4.64
Calibration intercept:	-0.543305	-1.764623	Coeff or Slope:	0.963	0.963

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-2.4	----
as found span	4920	80.3	800.4	799.8	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-2.1	----
high point	4920	80.3	800.4	802.4	0.997
second point	4960	40.2	400.7	399.1	1.004
third point	4980	20.1	200.4	200.9	0.997
as left zero	5000	0.0	0.0	-2.2	----
as left span	4920	80.3	800.4	802.6	0.997

Average Correction Factor	1.000
---------------------------	-------

Baseline Corr As found:	802.20	Previous response	800.09	*% change	0.3%
-------------------------	--------	-------------------	--------	-----------	------

\* = > +/-5% change initiates investigation

Notes: changed inlet filters after as founds. no adjustments

Calibration Performed By: Morgan Voyageur



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 14, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	10:53	End Time (MST):	13:31
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-2.1	----	Correlation Coefficient	0.999987	
800.4	802.4	0.9975			≥0.995
400.7	399.1	1.0040	Slope	1.004228	
200.4	200.9	0.9973			0.90 - 1.10
			Intercept	-1.764623	+/-30



SO2 Calibration Plot

Date: September 6, 2023

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Fort Chipewyan Station number: AMS08  
 Calibration Date: September 21, 2023 Last Cal Date: August 24, 2023  
 Start time (MST): 12:29 End time (MST): 16:20  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: February 9, 2024  
 Cal Gas Cylinder #: EY0002276  
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252  
 ZAG Make/Model: Teledyne API T701 Serial Number: 260

### Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 1203169744  
 Converter make: CDN-101 Converter serial #: 14639  
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999424	1.011429	Backgd or Offset:	0.99	1.00
Calibration intercept:	0.458863	0.338843	Coeff or Slope:	0.752	0.752

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4920	80.5	80.0	80.2	1.003
as found 2nd point	4960	40.2	40.0	40.1	1.006
as found 3rd point	4980	20.1	20.0	20.5	0.994
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4920	80.5	80.0	81.4	0.983
second point	4960	40.2	40.0	40.4	0.989
third point	4980	20.1	20.0	20.7	0.965
as left zero	5000	0.0	0.0	0.6	----
as left span	4920	80.5	80.0	81.6	0.981
SO2 Scrubber Check	4919.7	80.3	803.0	0.1	----
Date of last scrubber change:	March 7, 2022			Ave Corr Factor	0.979
Date of last converter efficiency test:	March 15, 2022			100.7% efficiency	

Baseline Corr As found: 79.8 Prev response: 80.42 \*% change: -0.8%  
 Baseline Corr 2nd AF pt: 39.7 AF Slope: 0.996428 AF Intercept: 0.438656  
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999986

\* = > +/-5% change initiates investigation

Notes: Sampled inlet filter changed after as founds. Scrubber check passed.

Calibration Performed By: Matthew Courtoreille



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 21, 2023	Previous Calibration:	August 24, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	12:29	End Time (MST):	16:20
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999952	≥0.995
80.0	81.4	0.9829			
40.0	40.4	0.9890	Slope	1.011429	0.90 - 1.10
20.0	20.7	0.9652			
			Intercept	0.338843	+/-3



TRS Calibration Plot

Date: September 21, 2023

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Fort Chipewyan Station number: AMS08  
Calibration Date: September 21, 2023 Last Cal Date: August 29, 2023  
Start time (MST): 7:30 End time (MST): 11:41  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: CC363447 Cal Gas Expiry Date: February 2, 2024  
NOX Cal Gas Conc: 48.80 ppm NO Cal Gas Conc: 48.80 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 48.80 ppm Removed Gas NO Conc: 48.80 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 3060  
ZAG make/model: Teledyne API T701H Serial Number: 260

### Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1426262592  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.020	1.078	NO bkgnd or offset:	9.2	9.8
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	10.1	10.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	147.0	145.7

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.996773	0.993131
NO <sub>x</sub> Cal Offset:	0.340000	1.940000
NO Cal Slope:	0.997515	0.994931
NO Cal Offset:	-0.520000	1.260000
NO <sub>2</sub> Cal Slope:	1.009480	0.991090
NO <sub>2</sub> Cal Offset:	1.358955	-1.074507



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.7	0.5	0.2	----	----
as found span	4918	82.0	800.3	800.3	0.0	755.5	754.6	1.2	1.0593	1.0606
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.4	0.0	----	----
high point	4918	82.0	800.3	800.3	0.0	795.8	796.9	-0.1	1.0057	1.0043
second point	4959	41.0	400.2	400.2	0.0	400.6	400.4	0.3	0.9989	0.9994
third point	4980	20.5	200.1	200.1	0.0	202.0	200.8	1.1	0.9905	0.9964
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.1	-0.3	----	----
as left span	4918	82.0	800.3	404.3	396.0	793.1	403.1	389.8	1.0091	1.0030
Average Correction Factor									0.9984	1.0000

Corrected As found	NO <sub>x</sub> = 754.8 ppb	NO = 754.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -5.7%
Previous Response	NO <sub>x</sub> = 798.1 ppb	NO = 797.8 ppb		*Percent Change	NO = -5.8%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.8	401.8	396.0	392.5	1.0089	99.1%
2nd GPT point (200 ppb O3)	797.8	597.4	200.4	196.6	1.0193	98.1%
3rd GPT point (100 ppb O3)	797.8	597.1	200.7	196.6	1.0209	98.0%
Average Correction Factor					1.0164	98.4%

Notes: sampled inlet filter changed after as founds. Made adjustment to high point span.

Calibration Performed By: Matthew C





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 21, 2023	Previous Calibration:	August 29, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:30	End Time (MST):	11:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient Slope Intercept	0.999980 0.993131 1.940000	
800.3	795.8	1.0057			≥0.995
400.2	400.6	0.9989			0.90 - 1.10
200.1	202.0	0.9905			+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 21, 2023	Previous Calibration:	August 29, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:30	End Time (MST):	11:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	796.9	1.0043		
400.2	400.4	0.9994		
200.1	200.8	0.9964		
			0.999993	
			0.994931	
			1.260000	

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 21, 2023	Previous Calibration:	August 29, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:30	End Time (MST):	11:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
396.0	392.5	1.0089		
200.4	196.6	1.0193		
200.7	196.6	1.0209		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 21, 2023

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Fort Chipewyan      Station number: AMS08  
 Calibration Date: September 6, 2023      Last Cal Date: August 14, 2023  
 Start time (MST): 8:19      End time (MST): 10:49  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 3252  
 ZAG Make/Model: Teledyne API T701      Serial Number: 260

### Analyzer Information

Analyzer make: Teledyne API T400      Analyzer serial #: 3872  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.018771	1.018086	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	-0.760000	-0.940000	Coeff or Slope:	1.036	1.036

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	5000	913.0	400.0	406.8	0.983
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.0	----
high point	5000	914.7	400.0	406.8	0.983
second point	5000	786.4	200.0	202.1	0.990
third point	5000	701.3	100.0	100.0	1.000
as left zero	5000	0.0	0.0	0.5	----
as left span	5000	963.3	400.0	408.7	0.979
Average Correction Factor					0.991

Baseline Corr As found:	406.9	Previous response	406.7	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed out inlet filter after as found. No adjustments

Calibration Performed By: Morgan Voyageur



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 14, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:19	End Time (MST):	10:49
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999975
400.0	406.8	0.9833		
200.0	202.1	0.9896	Slope	1.018086
100.0	100.0	1.0000		
			Intercept	-0.940000
				+/- 5



O<sub>3</sub> Calibration Plot

Date: September 6, 2023

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Fort Chipewyan Station number: AMS 08  
 Calibration Date: September 21, 2023 Last Cal Date: August 24, 2023  
 Start time (MST): 14:37 End time (MST): 12:27

Analyzer Make: Teledyne API T640 S/N: 216  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: FP-25 S/N: 388747  
 Temp/RH standard: FP-25 S/N: 388747

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	17.0	17.7	17.0	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	735.9	739.4	735.9	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.10	5.01	<input checked="" type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 20, 2023</u>	Last Cal Date: <u>July 25, 2023</u>			
	PM w/o HEPA: <u>71.2</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.1	11.0	11.2	<input type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>24.5</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>July 25, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 25, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: July 25, 2023  
 Date RH/T Sensor Cleaned: July 25, 2023

Notes: flow was adjusted to 5.01 was 5.03.

Calibration by: Matthew Courtoreille





# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	September 22, 2023	Last Cal Date:	August 30, 2023
Start time (MST):	8:48	End time (MST):	11:58
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API T701H		Serial Number:	260

### Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<b><u>Start</u></b>	<b><u>Finish</u></b>		<b><u>Start</u></b>	<b><u>Finish</u></b>
Calibration slope:	0.995357	1.000563	Backgd or Offset:	-0.013	-0.014
Calibration intercept:	0.050947	0.028916	Coeff or Slope:	0.999	1.006

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.24	----
as found span	4933	66.7	40.4	40.6	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4934	66.7	40.4	40.4	1.000
second point	4967	33.3	20.2	20.3	0.994
third point	4983	16.7	10.1	10.1	0.998
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.4	40.3	1.003
<b>Average Correction Factor</b>					<b>0.997</b>

Baseline Corr As found:	40.36	Prev response:	40.29	*% change:	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Adjustments made to Zero point and high point.

Calibration Performed By: Matthew C



# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 22, 2023	Previous Calibration:	August 30, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:48	End Time (MST):	11:58
Analyzer make:	API T300	Analyzer serial #:	3505

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999990
40.4	40.4	0.9996		
20.2	20.3	0.9941	Slope	1.000563
10.1	10.1	0.9980		
			Intercept	0.028916
				<i>+/-1.5</i>



CO Calibration Plot

Date: September 22, 2023

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	September 22, 2023	Last Cal Date:	August 28, 2023
Start time (MST):	12:00	End time (MST):	14:59
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
N2 Gen Make/Model:	NG 5000		Serial Number:	771048318

### Analyzer Information

Analyzer make: Teledyne API T360	Analyzer serial #: 289
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997557	0.989394	Backgd or Offset:	0.008	0.008
Calibration intercept:	-1.000000	-0.140000	Coeff or Slope:	1.019	1.019

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	0.7	----
as found span	2920	80.0	1605.9	1592.6	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	1.0	----
high point	2920	80.0	1605.9	1593.3	1.008
second point	2960	40.0	802.9	781.6	1.027
third point	2980	20.0	401.5	404.0	0.994
as left zero	3000	0.0	0.0	1.4	----
as left span	2960	40.0	802.9	778.7	1.031
Average Correction Factor					1.010

Baseline Corr As found:	1591.90	Prev response:	1600.94	*% change:	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as found.

Calibration Performed By: Matthew C



# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date	September 22, 2023	Previous Calibration	August 28, 2023
Station Name	Fort Chipewyan	Station Number	AMS08
Start Time (MST)	12:00	End Time (MST)	19:06
Analyzer make	Teledyne API T360	Analyzer serial #	289

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.0	----	Correlation Coefficient	0.999833	≥0.995
1605.9	1593.3	1.0079			
802.9	781.6	1.0273	Slope	0.989394	0.90 - 1.10
401.5	404.0	0.9937			
			Intercept	-0.140000	+/-20



CO<sub>2</sub> Calibration Plot

Date: September 22, 2023

Location: Fort Chipewyan





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS09 BARGE LANDING SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023







# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 15, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:50	End Time (MST):	13:09
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999995	≥0.995
801.5	801.7	0.9997			
400.8	399.1	1.0041	Slope	1.000440	0.90 - 1.10
199.8	199.6	1.0012			
			Intercept	-0.648667	+/-30



SO2 Calibration Plot

Date: September 5, 2023

Location: Barge Landing





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Barge Landing      Station number: AMS09  
 Calibration Date: September 7, 2023      Last Cal Date: August 8, 2023  
 Start time (MST): 9:21      End time (MST): 13:23  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 4.87 ppm      Cal Gas Exp Date: September 2, 2024  
 Cal Gas Cylinder #: EY0002346  
 Removed Cal Gas Conc: 4.87 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3812  
 ZAG Make/Model: API T701      Serial Number: 4888

### Analyzer Information

Analyzer make: Thermo 43i-TLE      Analyzer serial #: 1331259320  
 Converter make: CDN-101      Converter serial #: 519  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007720	1.001003	Backgd or Offset: 2.77	2.77
Calibration intercept:	-0.060968	-0.000950	Coeff or Slope: 1.134	1.134

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4918	82.1	80.0	79.2	1.007
as found 2nd point	4959	41.1	40.0	39.3	1.013
as found 3rd point	4979	20.5	20.0	19.7	1.003
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4918	82.1	80.0	80.0	1.000
second point	4959	41.1	40.0	40.1	0.998
third point	4979	20.5	20.0	20.1	0.993
as left zero	5000	0.0	0.0	0.0	----
as left span	4918	82.1	80.0	80.7	0.991
SO2 Scrubber Check	4920	80.2	802.0	-0.1	----
Date of last scrubber change:	28-Feb-23			Ave Corr Factor	0.997
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.4      Prev response: 80.52      \*% change: -1.4%  
 Baseline Corr 2nd AF pt: 39.5      AF Slope: 0.992283      AF Intercept: -0.220830  
 Baseline Corr 3rd AF pt: 19.9      AF Correlation: 0.999983

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SOx scrubber check done after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:21	End Time (MST):	13:23
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999993	
80.0	80.0	0.9995			≥0.995
40.0	40.1	0.9983	Slope	1.001003	
20.0	20.1	0.9935			0.90 - 1.10
			Intercept	-0.000950	+/-3



TRS Calibration Plot

Date: September 7, 2023

Location: Barge Landing





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	September 5, 2023	Last Cal Date:	August 15, 2023
Start time (MST):	9:50	End time (MST):	13:09
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	497.6 ppm	CH <sub>4</sub> Equiv Conc.	1067.1 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	497.6 ppm	CH <sub>4</sub> Equiv Conc.	1067.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.51E-04	2.52E-04	NMHC SP Ratio:	4.87E-05
CH <sub>4</sub> Retention time:	15.20	15.40	NMHC Peak Area:	187816
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	17.10	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.15	0.999
second point	4960	40.1	8.56	8.54	1.002
third point	4980	20.0	4.27	4.28	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.19	0.996
				Response Correction Factor	0.999

Baseline Corr AF:	17.10	Prev response	17.13	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	9.14	9.15	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.15	0.998
second point	4960	40.1	4.57	4.57	1.000
third point	4980	20.0	2.28	2.29	0.993
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	9.14	9.21	0.992
Average Correction Factor					0.997
Baseline Corr AF:	9.15	Prev response	9.14	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	7.95	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	7.99	0.999
second point	4960	40.1	3.99	3.98	1.004
third point	4980	20.0	1.99	1.99	1.001
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	7.98	1.000
Average Correction Factor					1.001
Baseline Corr AF:	7.95	Prev response	7.99	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000663	1.001164
THC Cal Offset:	0.001260	-0.002739
CH <sub>4</sub> Cal Slope:	1.002197	1.000766
CH <sub>4</sub> Cal Offset:	-0.007135	-0.005136
NMHC Cal Slope:	0.999486	1.001350
NMHC Cal Offset:	0.007995	0.002797

Notes: Changed sample inlet filter after as founds. Do a zero chromatogram and use zero chromatogram. Adjusted span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 15, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:50	End Time (MST):	13:09
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	$\geq 0.995$			
17.12	17.15	0.9985						
8.56	8.54	1.0018				Slope	1.001164	0.90 - 1.10
4.27	4.28	0.9968						
			Intercept	-0.002739	$\pm 0.5$			







# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 15, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:50	End Time (MST):	13:09
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999993	$\geq 0.995$			
7.98	7.99	0.9991						
3.99	3.98	1.0039				Slope	1.000766	0.90 - 1.10
1.99	1.99	1.0007						
			Intercept	-0.005136	$\pm 0.5$			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 15, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:50	End Time (MST):	13:09
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	$\geq 0.995$			
9.14	9.15	0.9981						
4.57	4.57	1.0001				Slope	1.001350	0.90 - 1.10
2.28	2.29	0.9931						
			Intercept	0.002797	$\pm 0.5$			



NMHC Calibration Plot

Date: September 5, 2023

Location: Barge Landing





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Barge Landing  
Calibration Date: September 6, 2023  
Start time (MST): 9:23  
Reason: Routine  
Station number: AMS09  
Last Cal Date: August 23, 2023  
End time (MST): 13:53

### Calibration Standards

NO Gas Cylinder #: DT0036634  
NOX Cal Gas Conc: 50.00 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 50.00 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701  
Cal Gas Expiry Date: January 28, 2024  
NO Cal Gas Conc: 49.70 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 49.70 ppm  
NO gas Diff:  
Serial Number: 3812  
Serial Number: 4888

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1426262593

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.157	1.162	NO bkgnd or offset:	10.6	10.6
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	10.8	10.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	174.9	175.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.999930	1.000087
NO <sub>x</sub> Cal Offset:	0.929269	1.049154
NO Cal Slope:	1.001426	1.000827
NO Cal Offset:	-0.332141	-0.572366
NO <sub>2</sub> Cal Slope:	1.000531	1.000344
NO <sub>2</sub> Cal Offset:	0.173936	0.677077



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.4	-0.3	0.6	----	----
as found span	4919	80.5	805.1	800.3	4.8	804.5	794.9	9.5	1.001	1.007
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.4	-0.1	0.5	----	----
high point	4919	80.5	805.1	800.3	4.8	805.8	800.5	5.4	0.999	1.000
second point	4959	40.2	402.1	399.7	2.4	403.7	399.4	4.3	0.996	1.001
third point	4979	20.1	201.0	199.8	1.2	202.6	198.8	3.8	0.992	1.005
as left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.2	----	----
as left span	4919	80.5	805.1	439.6	365.5	804.2	437.0	367.2	1.001	1.006
Average Correction Factor									0.996	1.002

Corrected As found	NO <sub>x</sub> = 804.1 ppb	NO = 795.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.2%
Previous Response	NO <sub>x</sub> = 806.0 ppb	NO = 801.1 ppb		*Percent Change	NO = -0.7%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.4	436.7	365.5	366.3	0.998	100.2%
2nd GPT point (200 ppb O3)	797.4	660.3	141.9	142.6	0.995	100.5%
3rd GPT point (100 ppb O3)	797.4	726.8	75.4	76.4	0.987	101.3%
Average Correction Factor					0.994	100.7%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 23, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:23	End Time (MST):	13:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
805.1	805.8	0.9991		
402.1	403.7	0.9959		
201.0	202.6	0.9923		





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 23, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:23	End Time (MST):	13:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
800.3	800.5	0.9997			
399.7	399.4	1.0006			
199.8	198.8	1.0052			
			Slope	1.000827	0.90 - 1.10
			Intercept	-0.572366	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 23, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:23	End Time (MST):	13:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.5	----	Correlation Coefficient	0.999999	≥0.995
365.5	366.3	0.9979			
141.9	142.6	0.9953			
75.4	76.4	0.9873			
			Slope	1.000344	0.90 - 1.10
			Intercept	0.677077	+/-20






# NO<sub>x</sub> Calibration Plot

Date: September 6, 2023

Location: Barge Landing







# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Barge Landing Station number: AMS 09  
 Calibration Date: September 29, 2023 Last Cal Date: September 7, 2023  
 Start time (MST): 9:23 End time (MST): 9:33

Analyzer Make: API T640 S/N: 321  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: DeltaCal S/N: 1451  
 Temp/RH standard: DeltaCal S/N: 1451

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	8.3	8.6	8.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	732.3	733.5	732.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	4.98	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 29, 2023</u>	Last Cal Date: <u>September 7, 2023</u>			
	PM w/o HEPA: <u>43.6</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>August 23, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 7, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: August 23, 2023  
 Date RH/T Sensor Cleaned: August 23, 2023

Removal calibration.

Notes:

Calibration by: Sean Bala



# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Barge Landing Station number: AMS 09  
 Calibration Date: September 29, 2023 Last Cal Date: NA  
 Start time (MST): 9:52 End time (MST): 10:27

Analyzer Make: API T640 S/N: 844  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: DeltaCal S/N: 1451  
 Temp/RH standard: DeltaCal S/N: 1451

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	8.6	8.4	8.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	727.3	733.5	727.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	4.89	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 29, 2023</u>	Last Cal Date: <u>NA</u>			
	PM w/o HEPA: <u>36.3</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		_____			<0.2 ug/m3
Disposable Filter Changed:		_____			

### Annual Maintenance

Date Sample Tube Cleaned: August 23, 2023  
 Date RH/T Sensor Cleaned: August 23, 2023

Install calibration.

Notes:

Calibration by: Sean Bala



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS11 LOWER CAMP SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 22, 2023	Previous Calibration:	August 10, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:53	End Time (MST):	12:49
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient	≥0.995	
800.8	805.4	0.9942			
400.9	401.2	0.9993			
199.9	200.3	0.9982			
			Slope	1.005149	0.90 - 1.10
			Intercept	-0.308964	+/-30



SO2 Calibration Plot

Date: September 22, 2023

Location: Lower Camp







# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Lower Camp      Station number: AMS11  
 Calibration Date: September 12, 2023      Last Cal Date: August 29, 2023  
 Start time (MST): 9:08      End time (MST): 13:01  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.429 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC501097  
 Removed Cal Gas Conc: 5.429 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3807  
 ZAG Make/Model: API T701H      Serial Number: 196

### Analyzer Information

Analyzer make: Thermo 450iQ      Analyzer serial #: CM20080003  
 Converter make: NA      Converter serial #: NA  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.017077	1.024520	Backgd or Offset: 14.1	14.2
Calibration intercept:	-0.466446	-0.427099	Coeff or Slope: 1.001	1.001

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4926	73.6	79.9	80.9	0.984
as found 2nd point	4963	36.8	40.0	40.6	0.977
as found 3rd point	4982	18.6	20.2	20.0	0.995
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.5	----
high point	4926	73.6	79.9	81.4	0.982
second point	4963	36.8	40.0	40.6	0.984
third point	4982	18.6	20.2	20.3	0.995
as left zero	5000	0.0	0.0	-0.3	----
as left span	4926	73.6	79.9	81.9	0.976
SO2 Scrubber Check	4919	81.1	811.0	0.2	----
Date of last scrubber change:				Ave Corr Factor	0.987
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 81.2      Prev response: 80.82      \*% change: 0.5%  
 Baseline Corr 2nd AF pt: 40.9      AF Slope: 1.017348      AF Intercept: -0.325940  
 Baseline Corr 3rd AF pt: 20.3      AF Correlation: 0.999964

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Completed SO2 scrubber check after calibrator zero.  
 No adjustments required.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	August 29, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:08	End Time (MST):	13:01
Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.5	----	Correlation Coefficient	0.999995	
79.9	81.4	0.9818			≥0.995
40.0	40.6	0.9842	Slope	1.024520	
20.2	20.3	0.9948			0.90 - 1.10
			Intercept	-0.427099	+/-3



H<sub>2</sub>S Calibration Plot

Date: September 12, 2023

Location: Lower Camp





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	September 22, 2023	Last Cal Date:	August 10, 2023
Start time (MST):	9:53	End time (MST):	12:49
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH <sub>4</sub> Cal Gas Conc.	502.0 ppm	CH <sub>4</sub> Equiv Conc.	1067.1 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	502.0 ppm	CH <sub>4</sub> Equiv Conc.	1067.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.5 ppm		
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (THC):	
Diff between cyl (NMHC):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
ZAG make/model:	API T701	Serial Number:	196

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.99E-04	2.99E-04	NMHC SP Ratio:	5.79E-05
CH <sub>4</sub> Retention time:	14.0	14.0	NMHC Peak Area:	158468

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	17.35	17.18	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	17.35	17.26	1.005
second point	4959	40.7	8.69	8.58	1.012
third point	4980	20.3	4.33	4.28	1.011
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	17.35	17.36	0.999

Average Correction Factor				1.009
Baseline Corr AF:	17.18	Prev response	17.26	*% change -0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	9.19	9.09	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	9.19	9.14	1.005
second point	4959	40.7	4.60	4.55	1.012
third point	4980	20.3	2.29	2.27	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	9.19	9.18	1.001
Average Correction Factor					1.010
Baseline Corr AF:	9.09	Prev response	9.14	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	8.16	8.09	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	8.16	8.13	1.004
second point	4959	40.7	4.09	4.04	1.013
third point	4980	20.3	2.04	2.02	1.010
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	8.16	8.18	0.998
Average Correction Factor					1.009
Baseline Corr AF:	8.09	Prev response	8.12	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.996495	0.995046
THC Cal Offset:	-0.024788	-0.021788
CH <sub>4</sub> Cal Slope:	0.996527	0.995660
CH <sub>4</sub> Cal Offset:	-0.015486	-0.010886
NMHC Cal Slope:	0.996217	0.994824
NMHC Cal Offset:	-0.009301	-0.011701

Notes: Changed sample inlet filter and N<sub>2</sub> cylinder after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 22, 2023	Previous Calibration:	August 10, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:53	End Time (MST):	12:49
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999985	$\geq 0.995$			
17.35	17.26	1.0051						
8.69	8.58	1.0120				Slope	0.995046	0.90 - 1.10
4.33	4.28	1.0113						
			Intercept	-0.021788	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 22, 2023	Previous Calibration:	August 10, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:53	End Time (MST):	12:49
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999979	≥0.995			
8.16	8.13	1.0043						
4.09	4.04	1.0125				Slope	0.995660	0.90 - 1.10
2.04	2.02	1.0104						
			Intercept	-0.010886	+/-0.5			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 22, 2023	Previous Calibration:	August 10, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:53	End Time (MST):	12:49
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999988	$\geq 0.995$			
9.19	9.14	1.0054						
4.60	4.55	1.0117				Slope	0.994824	0.90 - 1.10
2.29	2.27	1.0120						
			Intercept	-0.011701	$\pm 0.5$			






NMHC Calibration Plot

Date: September 22, 2023

Location: Lower Camp





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS13 FORT MCKAY SOUTH SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 16, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:52	End Time (MST):	13:30
Analyzer make:	API T100	Analyzer serial #:	599

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999967	
799.7	801.2	0.9981			≥0.995
399.3	398.8	1.0013	Slope	1.002912	
200.2	196.6	1.0182			0.90 - 1.10
			Intercept	-1.537784	+/-30



SO2 Calibration Plot

Date: September 15, 2023

Location: Fort McKay South





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Fort McKay South      Station number: AMS13  
 Calibration Date: September 19, 2023      Last Cal Date: August 15, 2023  
 Start time (MST): 8:50      End time (MST): 12:44  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.34 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC500241  
 Removed Cal Gas Conc: 5.34 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 2448  
 ZAG Make/Model: Teledyne API 701      Serial Number: 1117

### Analyzer Information

Analyzer make: Thermo 43i TLE      Analyzer serial #: 1180540017  
 Converter make: CDN-101      Converter serial #: 521  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004479	1.004193	Backgd or Offset: 3.77	3.77
Calibration intercept:	-0.502377	-0.442262	Coeff or Slope: 1.130	1.130

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4925	75.5	80.6	81.6	0.986
as found 2nd point	4962	37.7	40.3	40.3	0.994
as found 3rd point	4981	18.9	20.2	19.7	1.014
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4925	75.5	80.6	80.7	0.999
second point	4962	37.7	40.3	39.8	1.012
third point	4981	18.9	20.2	19.5	1.035
as left zero	5000	0.0	0.0	0.0	----
as left span	4925	75.5	80.6	79.2	1.018
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:	20-Mar-20		Ave Corr Factor		1.015
Date of last converter efficiency test:	NA		efficiency		

Baseline Corr As found: 81.8      Prev response: 80.48      \*% change: 1.6%  
 Baseline Corr 2nd AF pt: 40.5      AF Slope: 1.016526      AF Intercept: -0.502263  
 Baseline Corr 3rd AF pt: 19.9      AF Correlation: 0.999938

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 19, 2023	Previous Calibration:	August 15, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:50	End Time (MST):	12:44
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999918	≥0.995
80.6	80.7	0.9991			
40.3	39.8	1.0117	Slope	1.004193	0.90 - 1.10
20.2	19.5	1.0352			
			Intercept	-0.442262	+/-3



TRS Calibration Plot

Date: September 19, 2023

Location: Fort McKay South







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	September 15, 2023	Last Cal Date:	August 16, 2023
Start time (MST):	8:52	End time (MST):	13:30
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH <sub>4</sub> Cal Gas Conc.	503.6 ppm	CH <sub>4</sub> Equiv Conc.	1077.5 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	503.6 ppm	CH <sub>4</sub> Equiv Conc.	1077.5 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API701	Serial Number:	1117

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	11700501330
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.22E-04	2.24E-04	NMHC SP Ratio:	4.70E-04
CH <sub>4</sub> Retention time:	12.80	13.00	NMHC Peak Area:	193345
Zero Chromatogram:	ON	ON	Flat Baseline:	ON
				5.16E-05
				176051
				OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	17.05	17.05	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.07	0.998
second point	4961	39.5	8.51	8.41	1.012
third point	4980	19.8	4.27	4.12	1.035
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	17.05	17.05	1.000
Average Correction Factor					1.015

Baseline Corr AF:	17.05	Prev response	17.02	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	9.08	9.16	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	9.08	9.10	0.998
second point	4961	39.5	4.53	4.51	1.004
third point	4980	19.8	2.27	2.22	1.024
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	9.08	9.07	1.001
Average Correction Factor					1.009
Baseline Corr AF:	9.16	Prev response	9.07	*% change	0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	7.97	7.90	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.97	0.999
second point	4961	39.5	3.98	3.90	1.021
third point	4980	19.8	1.99	1.91	1.047
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	7.97	7.98	0.998
Average Correction Factor					1.022
Baseline Corr AF:	7.90	Prev response	7.96	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.003320	1.003320
THC Cal Offset:	-0.078975	-0.078975
CH <sub>4</sub> Cal Slope:	1.004551	1.004551
CH <sub>4</sub> Cal Offset:	-0.047351	-0.047351
NMHC Cal Slope:	1.002546	1.002546
NMHC Cal Offset:	-0.032763	-0.032763

Notes: Changed inlet filter after as founds. 3rd point was failing need to go back and used zero chromatogram and flat baseline was turned off. Adjusted span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 16, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:52	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	11700501330

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999889	$\geq 0.995$			
17.05	17.07	0.9984						
8.51	8.41	1.0123				Slope	1.003830	0.90 - 1.10
4.27	4.12	1.0350						
			Intercept	-0.083774	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 16, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:52	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	11700501330

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999781	≥0.995
7.97	7.97	0.9991			
3.98	3.90	1.0213			
1.99	1.91	1.0469			
			Slope	1.003588	0.90 - 1.10
			Intercept	-0.053796	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 16, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:52	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	11700501330

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999949	$\geq 0.995$			
9.08	9.10	0.9975						
4.53	4.51	1.0043				Slope	1.004357	0.90 - 1.10
2.27	2.22	1.0243						
			Intercept	-0.029978	$\pm 0.5$			



NMHC Calibration Plot

Date: September 15, 2023

Location: Fort McKay South







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.7	0.2	0.5	----	----
as found span	4919	81.1	826.9	800.0	26.9	823.8	794.0	29.9	1.0037	1.0075
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4919	81.1	826.9	800.0	26.9	825.4	800.3	25.1	1.0018	0.9996
second point	4960	40.6	413.9	400.4	13.5	411.0	397.1	13.9	1.0071	1.0084
third point	4980	20.3	207.0	200.2	6.7	203.0	195.0	8.0	1.0195	1.0268
as left zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.4	-0.3	----	----
as left span	4919	80.1	816.8	372.3	444.5	830.2	382.6	447.5	0.9839	0.9731
Average Correction Factor									1.0095	1.0116

Corrected As found	NO <sub>x</sub> = 823.1 ppb	NO = 793.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.1%	
Previous Response	NO <sub>x</sub> = 822.4 ppb	NO = 796.5 ppb		*Percent Change	NO = -0.3%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	798.5	380.9	444.5	442.2	1.0053	99.5%
2nd GPT point (200 ppb O <sub>3</sub> )	798.5	590.0	235.4	232.2	1.0139	98.6%
3rd GPT point (100 ppb O <sub>3</sub> )	798.5	692.7	132.7	129.0	1.0289	97.2%
Average Correction Factor					1.0160	98.4%

Notes: Changed inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Sean Bala





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 26, 2023	Previous Calibration:	August 24, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:11	End Time (MST):	13:48
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
826.9	825.4	1.0018			
413.9	411.0	1.0071			
207.0	203.0	1.0195			
			Slope	0.999304	0.90 - 1.10
			Intercept	-1.811346	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 26, 2023	Previous Calibration:	August 24, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:11	End Time (MST):	13:48
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999943	≥0.995
800.0	800.3	0.9996			
400.4	397.1	1.0084	Slope	1.002205	0.90 - 1.10
200.2	195.0	1.0268			
			Intercept	-2.825167	+/-20

NO Calibration Curve





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 26, 2023	Previous Calibration:	August 24, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:11	End Time (MST):	13:48
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
444.5	442.2	1.0053		
235.4	232.2	1.0139		
132.7	129.0	1.0289		



NO<sub>x</sub> Calibration Plot

Date: September 26, 2023

Location: Fort McKay South





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Fort McKay South      Station number: AMS13  
 Calibration Date: September 8, 2023      Last Cal Date: August 11, 2023  
 Start time (MST): 9:00      End time (MST): 12:32  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 2448  
 ZAG Make/Model: Teledyne API T701      Serial Number: 1117

### Analyzer Information

Analyzer make: Teledyne API T400      Analyzer serial #: 3871  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997543	0.999886	Backgd or Offset:	2.4	2.4
Calibration intercept:	1.180000	1.020000	Coeff or Slope:	0.963	0.967

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	5000	977.0	400.0	397.2	1.007
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.3	----
high point	5000	977.0	400.0	400.5	0.999
second point	5000	838.0	200.0	201.7	0.992
third point	5000	735.9	100.0	101.5	0.985
as left zero	5000	0.0	0.0	-0.2	----
as left span	5000	977.0	400.0	401.6	0.996
Average Correction Factor					0.992

Baseline Corr As found:	397.5	Previous response	400.2	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 8, 2023	Previous Calibration:	August 11, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	9:00	End Time (MST):	12:32
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999983	
400.0	400.5	0.9988			≥0.995
200.0	201.7	0.9916	Slope	0.999886	
100.0	101.5	0.9852			0.90 - 1.10
			Intercept	1.020000	+/- 5



O<sub>3</sub> Calibration Plot

Date: September 8, 2023

Location: Fort McKay South





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Fort McKay South Station number: AMS 13  
 Calibration Date: September 26, 2023 Last Cal Date: August 24, 2023  
 Start time (MST): 11:03 End time (MST): 11:16

Analyzer Make: API T640 S/N: 319  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1451  
 Temp/RH standard: Delta Cal S/N: 1451

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	19.0	19	19.0	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	727.2	727.5	727.2	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.03	5.01	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: September 26, 2023 Last Cal Date: August 24, 2023  
 PM w/o HEPA: 21.4 PM w/ HEPA: 0.0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>June 29, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 26, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: June 29, 2023  
 Date RH/T Sensor Cleaned: June 29, 2023

Notes: Inlet head clean and inspected. Leak check passed. No adjustment made. Quarterly calibration will be done once filter is available.

Calibration by: Sean Bala





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Fort McKay South Station number: AMS 13  
 Calibration Date: September 29, 2023 Last Cal Date: September 26, 2023  
 Start time (MST): 10:52 End time (MST): 11:02

Analyzer Make: API T640 S/N: 319  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1451  
 Temp/RH standard: Delta Cal S/N: 1451

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	8.8	8.9	8.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	734.1	734.0	734.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	4.97	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 29, 2023</u>		Last Cal Date: <u>September 26, 2023</u>		
	PM w/o HEPA: <u>36.8</u>		PM w/ HEPA: <u>0.0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>June 29, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 26, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: June 29, 2023  
 Date RH/T Sensor Cleaned: June 29, 2023

Removal calibration.

Notes:

Calibration by: Sean Bala



# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Fort McKay South Station number: AMS 13  
 Calibration Date: September 29, 2023 Last Cal Date: NA  
 Start time (MST): 11:08 End time (MST): 11:41

Analyzer Make: API T640 S/N: 1335  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1451  
 Temp/RH standard: Delta Cal S/N: 1451

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	9.40	8.90	9.40	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	734.00	735.00	734.00	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	4.90	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 29, 2023</u>		Last Cal Date: <u>NA</u>		
	PM w/o HEPA: <u>24.4</u>		PM w/ HEPA: <u>0.0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		_____			<0.2 ug/m3
Disposable Filter Changed:		_____			

### Annual Maintenance

Date Sample Tube Cleaned: June 29, 2023  
 Date RH/T Sensor Cleaned: June 29, 2023

Install calibration.

Notes:

Calibration by: Sean Bala



# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Fort McKay South	Station Number:	AMS 13
Calibration Date:	September 15, 2023	Prev Cal Date:	October 5, 2022
Start Time (MST):	11:17	End Time (MST):	13:09
Tower Height (m):	9.5	Reason:	Routine

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	
WS Calibrator:	MetOne 053	Serial Number:	

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.7%
400	39.4	39.4	0.1%
600	58.6	58.7	0.2%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	0.999998	0.999998	$\geq 0.9995$
Calculated slope	1.001527	0.999451	$0.90 - 1.10$
Calculated intercept	-0.014138	-0.052868	$\pm 2$

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:21	Calc Declination*:	13.73 Degrees
Deadband calc:	0.0 degrees ( <i>Limit 4 deg</i> )		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.5	---
90	88.1	-0.5%
180	178.5	-0.4%
270	268.3	-0.5%
357	357.5	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	0.999936	0.999963	$\geq 0.9995$
Calculated slope	1.015820	0.999751	$0.90 - 1.10$
Calculated intercept	-0.298621	0.864506	$\pm 4$

Notes: Veified true north using solar noon.

Calibration Performed By: Sean Bala



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS14  
ANZAC  
SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:23	End Time (MST):	12:34
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	1.0	----	Correlation Coefficient	0.999940	
799.3	812.8	0.9834			≥0.995
400.1	402.7	0.9934	Slope	1.017069	
201.1	200.4	1.0033			0.90 - 1.10
			Intercept	-1.850638	+/-30



SO2 Calibration Plot

Date: September 6, 2023

Location: Anzac





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Anzac Station number: AMS14  
 Calibration Date: September 7, 2023 Last Cal Date: August 4, 2023  
 Start time (MST): 9:47 End time (MST): 14:09  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC510379  
 Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 3060  
 ZAG Make/Model: API 701H Serial Number: 357

### Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153582  
 Converter make: CD Nova CDN-101 Converter serial #: 503  
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.991300	0.992584	Backgd or Offset: 2.30	2.31
Calibration intercept:	-0.125436	-0.245370	Coeff or Slope: 0.992	0.992

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4938	77.9	80.0	79.6	1.002
as found 2nd point	4973	38.9	40.0	39.8	0.999
as found 3rd point	4997	19.5	20.0	19.6	1.011
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4938	77.9	80.0	79.2	1.009
second point	4973	38.9	40.0	39.3	1.017
third point	4997	19.5	20.0	19.5	1.026
as left zero	5000	0.0	0.0	-0.1	----
as left span	4938	77.9	80.0	78.1	1.024
SO2 Scrubber Check	4936	80.3	800.4	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.017
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	79.8	Prev response:	79.16	*% change:	0.8%
Baseline Corr 2nd AF pt:	40.0	AF Slope:	0.998628	AF Intercept:	-0.245351
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999988		

\* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif





# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 4, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	9:47	End Time (MST):	14:09
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999984	≥0.995
80.0	79.2	1.0095			
40.0	39.3	1.0167	Slope	0.992584	0.90 - 1.10
20.0	19.5	1.0262			
			Intercept	-0.245370	+/-3



TRS Calibration Plot

Date: September 7, 2023

Location: Anzac





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	September 6, 2023	Last Cal Date:	August 9, 2023
Start time (MST):	9:23	End time (MST):	12:34
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.90E-04	3.75E-04	NMHC SP Ratio:	4.53E-05
CH <sub>4</sub> Retention time:	12.20	12.20	NMHC Peak Area:	201206
				203038

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	17.55	0.975
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.13	0.999
second point	4979	40.2	8.56	8.52	1.005
third point	4998	20.2	4.30	4.24	1.014
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	17.10	17.13	0.998

Average Correction Factor				1.006
Baseline Corr AF:	17.55	Prev response	17.21	*% change 2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	9.21	0.990
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.12	0.999
second point	4979	40.2	4.56	4.54	1.006
third point	4998	20.2	2.29	2.25	1.018
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	9.11	9.12	0.999
Average Correction Factor					1.008
Baseline Corr AF:	9.21	Prev response	9.07	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	8.34	0.958
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	8.01	0.998
second point	4979	40.2	4.00	3.98	1.004
third point	4998	20.2	2.01	1.99	1.011
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	7.99	8.01	0.997
Average Correction Factor					1.004
Baseline Corr AF:	8.34	Prev response	8.13	*% change	2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.007749	1.002474
THC Cal Offset:	-0.029955	-0.037684
CH <sub>4</sub> Cal Slope:	1.019842	1.003084
CH <sub>4</sub> Cal Offset:	-0.015535	-0.015639
NMHC Cal Slope:	0.997110	1.001701
NMHC Cal Offset:	-0.015020	-0.021844

Notes: Changed sample inlet filter and Nitrogen cylinder after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:23	End Time (MST):	12:34
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999978	$\geq 0.995$			
17.10	17.13	0.9985						
8.56	8.52	1.0048				Slope	1.002474	0.90 - 1.10
4.30	4.24	1.0145						
			Intercept	-0.037684	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:23	End Time (MST):	12:34
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999981	≥0.995			
7.99	8.01	0.9977						
4.00	3.98	1.0040				Slope	1.003084	0.90 - 1.10
2.01	1.99	1.0110						
			Intercept	-0.015639	+/-0.5			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:23	End Time (MST):	12:34
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999974	$\geq 0.995$			
9.11	9.12	0.9995						
4.56	4.54	1.0056				Slope	1.001701	0.90 - 1.10
2.29	2.25	1.0175						
			Intercept	-0.021844	$\pm 0.5$			



NMHC Calibration Plot

Date: September 6, 2023

Location: Anzac







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	September 25, 2023	Last Cal Date:	September 6, 2023
Start time (MST):	10:47	End time (MST):	12:24
Reason:	Cylinder Change		

### Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.75E-04	NA	NMHC SP Ratio:	4.49E-05
CH <sub>4</sub> Retention time:	12.20	NA	NMHC Peak Area:	203038

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	17.17	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.17	0.996
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	0.996
Baseline Corr AF:	17.17	Prev response	17.11	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	9.14	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.14	0.997
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.997
Baseline Corr AF:	9.14	Prev response	9.11	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	8.03	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	8.03	0.995
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.995
Baseline Corr AF:	8.03	Prev response	8.00	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002474	1.004050
THC Cal Offset:	-0.037684	0.000000
CH <sub>4</sub> Cal Slope:	1.003084	1.004817
CH <sub>4</sub> Cal Offset:	-0.015639	0.000000
NMHC Cal Slope:	1.001701	1.003378
NMHC Cal Offset:	-0.021844	0.000000


Notes: Hydrogen cylinder change.

Calibration Performed By: Mohammed Kashif

NMHC Calibration Plot

Date: September 25, 2023

Location: Anzac





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Anzac  
Calibration Date: September 5, 2023  
Start time (MST): 10:22  
Reason: Routine  
Station number: AMS 14  
Last Cal Date: August 15, 2023  
End time (MST): 15:14

### Calibration Standards

NO Gas Cylinder #: T2Y1P8D  
NOX Cal Gas Conc: 50.92 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 50.92 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API 701H  
Cal Gas Expiry Date: December 11, 2023  
NO Cal Gas Conc: 50.05 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 50.05 ppm  
NO gas Diff:  
Serial Number: 3060  
Serial Number: 357

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1426262592

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.375	1.411	NO bkgnd or offset:	3.7	3.8
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.7	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	161.2	160.3

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.986023	1.000191
NO <sub>x</sub> Cal Offset:	-0.615739	-0.708567
NO Cal Slope:	0.987256	1.000876
NO Cal Offset:	-2.172778	-2.586342
NO <sub>2</sub> Cal Slope:	1.000092	1.002633
NO <sub>2</sub> Cal Offset:	0.381307	0.926052



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.2	-0.2	0.4	----	----
as found span	4936	80.2	814.1	800.2	13.9	794.6	777.5	17.0	1.0246	1.0292
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.5	0.0	0.6	----	----
high point	4936	80.2	814.1	800.2	13.9	814.2	799.6	14.6	0.9999	1.0008
second point	4979	40.1	406.8	399.9	7.0	405.5	396.4	9.0	1.0033	1.0088
third point	4999	20.1	203.9	200.4	3.5	202.1	195.4	6.6	1.0090	1.0258
as left zero	5000	0.0	0.0	0.0	0.0	0.5	0.0	0.5	----	----
as left span	4936	80.2	814.1	413.5	400.6	813.5	410.1	403.4	1.0008	1.0083
Average Correction Factor									1.0041	1.0118

Corrected As found	NO <sub>x</sub> = 794.4 ppb	NO = 777.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -1.0%	
Previous Response	NO <sub>x</sub> = 802.1 ppb	NO = 787.8 ppb		*Percent Change	NO = -1.3%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.5	410.8	400.6	402.5	0.9953	100.5%
2nd GPT point (200 ppb O3)	797.5	607.4	204.0	205.4	0.9932	100.7%
3rd GPT point (100 ppb O3)	797.5	707.0	104.4	106.1	0.9841	101.6%
Average Correction Factor					0.9909	100.9%

Notes: Sample inlet filter changed after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 15, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:22	End Time (MST):	15:14
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.5	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
814.1	814.2	0.9999		
406.8	405.5	1.0033		
203.9	202.1	1.0090		





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 15, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:22	End Time (MST):	15:14
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ 0.90 - 1.10 +/-20
800.2	799.6	1.0008		
399.9	396.4	1.0088		
200.4	195.4	1.0258		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 15, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:22	End Time (MST):	15:14
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.6	----	Correlation Coefficient	≥0.995	
400.6	402.5	0.9953			
204.0	205.4	0.9932			
104.4	106.1	0.9841			
			Slope	1.002633	0.90 - 1.10
			Intercept	0.926052	+/-20






NO<sub>x</sub> Calibration Plot

Date: September 5, 2023

Location: Anzac





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Anzac Station number: AMS14  
 Calibration Date: September 1, 2023 Last Cal Date: August 1, 2023  
 Start time (MST): 8:47 End time (MST): 11:53  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: API T700 Serial Number: 3060  
 ZAG Make/Model: API 701H Serial Number: 357

### Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1426262595  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999171	1.004857	Backgd or Offset:	1.3	1.4
Calibration intercept:	0.520000	-2.500000	Coeff or Slope:	1.550	1.604

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	5000	920.5	400.0	386.6	1.035
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.1	----
high point	5000	920.5	400.0	400.7	0.998
second point	5000	793.4	200.0	197.3	1.014
third point	5000	686.0	100.0	95.3	1.049
as left zero	5000	0.0	0.0	0.2	----
as left span	5000	920.5	400.0	405.1	0.987
Average Correction Factor					1.020

Baseline Corr As found:	386.4	Previous response	400.2	*% change	-3.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 1, 2023	Previous Calibration:	August 1, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	8:47	End Time (MST):	11:53
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
400.0	400.7	0.9983		
200.0	197.3	1.0137	Slope	0.90 - 1.10
100.0	95.3	1.0493		
			Intercept	+/- 5



O<sub>3</sub> Calibration Plot

Date: September 1, 2023

Location: Anzac





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Anzac Station number: AMS 14  
 Calibration Date: September 8, 2023 Last Cal Date: August 1, 2023  
 Start time (MST): 11:00 End time (MST): 12:21

Analyzer Make: API T640 S/N: 825  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388749  
 Temp/RH standard: Alicat FP-25 S/N: 388749

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	20.5	22.4	20.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	719.9	719.6	719.9	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	4.81	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 8, 2023</u>		Last Cal Date: <u>August 1, 2023</u>		
	PM w/o HEPA: <u>9.9</u>		PM w/ HEPA: <u>0.0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.9	11.6	10.9	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: <u>9.4</u>		w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:	<u>September 8, 2023</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>September 8, 2023</u>				

### Annual Maintenance

Date Sample Tube Cleaned: July 6, 2023  
 Date RH/T Sensor Cleaned: July 6, 2023

Notes: Performed quarterly maintenance and leak check passed. Adjusted PMT peak test. Head cleaned.

Calibration by: Mohammed Kashif



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS17  
WAPASU  
SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 3, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:51	End Time (MST):	13:41
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999961	≥0.995
800.0	801.0	0.9987			
400.0	396.7	1.0084	Slope	1.001653	0.90 - 1.10
199.5	196.7	1.0143			
			Intercept	-1.759511	+/-30






SO2 Calibration Plot

Date: September 5, 2023

Location: Wapasu





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	September 7, 2023	Last Cal Date:	August 8, 2023
Start time (MST):	10:41	End time (MST):	15:30
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.076	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511852			
Removed Cal Gas Conc:	5.076	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2449
ZAG Make/Model:	API T701H		Serial Number:	359

### Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583
Converter make:	n/a	Converter serial #:	n/a
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996998	0.990569	Backgd or Offset:	13.1
Calibration intercept:	-0.219239	-0.219241	Coeff or Slope:	1.116
				12.2
				1.114

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	78.8	80.0	82.2	0.972
as found 2nd point	4961	39.4	40.0	41.5	0.961
as found 3rd point	4980	19.7	20.0	20.4	0.976
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	78.8	80.0	79.4	1.008
second point	4961	39.4	40.0	38.7	1.033
third point	4980	19.7	20.0	19.5	1.026
as left zero	5000	0.0	0.0	0.5	----
as left span	4921	78.8	80.0	76.5	1.046
SO2 Scrubber Check	4921	79.4	800.0	-0.1	----
Date of last scrubber change:	n/a			Ave Corr Factor	1.022
Date of last converter efficiency test:	n/a				efficiency

Baseline Corr As found:	82.3	Prev response:	79.54	*% change:	3.4%
Baseline Corr 2nd AF pt:	41.6	AF Slope:	1.029710	AF Intercept:	-0.039159
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999953		

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 8, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:41	End Time (MST):	15:30
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient	≥0.995
80.0	79.4	1.0076		
40.0	38.7	1.0335	Slope	0.90 - 1.10
20.0	19.5	1.0257		
			Intercept	+/-3



H<sub>2</sub>S Calibration Plot

Date: September 7, 2023

Location: Wapasu







# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 3, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:51	End Time (MST):	13:41
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.01	----	Correlation Coefficient	0.999884	
17.09	17.12	0.9986			≥0.995
8.55	8.40	1.0177	Slope	1.001763	
4.26	4.18	1.0187			0.90 - 1.10
			Intercept	-0.062338	+/-1.5



THC Calibration Plot

Date: September 5, 2023

Location: Wapasu





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Wapasu  
Calibration Date: September 22, 2023  
Start time (MST): 10:27  
Reason: Routine  
Station number: AMS17  
Last Cal Date: August 15, 2023  
End time (MST): 15:37

### Calibration Standards

NO Gas Cylinder #: T375YK8  
NOX Cal Gas Conc: 49.11 ppm  
Removed Cylinder #: T375YK8  
Removed Gas NOX Conc: 49.11 ppm  
NOX gas Diff: 0  
Calibrator Model: API T700  
ZAG make/model: API T701H  
Cal Gas Expiry Date: April 13, 2025  
NO Cal Gas Conc: 48.07 ppm  
Removed Gas Exp Date: 0  
Removed Gas NO Conc: 48.07 ppm  
NO gas Diff: 0  
Serial Number: 2449  
Serial Number: 359

### Analyzer Information

Analyzer make: Thermo Scientific 42iQ  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 12300522720

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.212	1.269	NO bkgnd or offset:	9	9.0
NOX coeff or slope:	0.992	0.989	NOX bkgnd or offset:	9.1	9.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	383.1	363.7

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.001480	1.000361
NO <sub>x</sub> Cal Offset:	-3.500000	-3.500000
NO Cal Slope:	0.999458	1.002373
NO Cal Offset:	-3.960000	-4.480000
NO <sub>2</sub> Cal Slope:	0.998776	0.998879
NO <sub>2</sub> Cal Offset:	-0.422855	-1.541001





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.2	----	----
as found span	4917	83.2	817.2	799.9	17.3	762.5	746.2	16.3	1.0717	1.0719
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.2	0.1	----	----
high point	4917	83.2	817.2	799.9	17.3	816.4	800.1	16.4	1.0010	0.9997
second point	4958	41.6	408.6	399.9	8.7	401.8	392.7	9.1	1.0169	1.0184
third point	4979	20.8	204.3	200.0	4.3	198.1	192.2	5.9	1.0313	1.0404
as left zero	5000	0.0	0.0	0.0	0.0	0.6	0.4	0.2	----	----
as left span	4917	83.2	817.2	398.3	418.9	805.1	389.9	415.2	1.0150	1.0214
Average Correction Factor									1.0164	1.0195

Corrected As found	NO <sub>x</sub> = 762.5 ppb	NO = 746.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -6.9%	
Previous Response	NO <sub>x</sub> = 814.9 ppb	NO = 795.5 ppb		*Percent Change	NO = -6.6%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.8	396.2	418.9	417.8	1.0026	99.7%
2nd GPT point (200 ppb O3)	797.8	593.7	221.4	218.9	1.0114	98.9%
3rd GPT point (100 ppb O3)	797.8	694.6	120.5	117.0	1.0300	97.1%
Average Correction Factor					1.0147	98.6%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 22, 2023	Previous Calibration:	August 15, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:27	End Time (MST):	15:37
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
817.2	816.4	1.0010		
408.6	401.8	1.0169		
204.3	198.1	1.0313		
			0.999891	
			1.000361	
			-3.500000	





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 22, 2023	Previous Calibration:	August 15, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:27	End Time (MST):	15:37
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999836	≥0.995
799.9	800.1	0.9997			
399.9	392.7	1.0184	Slope	1.002373	0.90 - 1.10
200.0	192.2	1.0404			
			Intercept	-4.480000	+/-20

NO Calibration Curve





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 22, 2023	Previous Calibration:	August 15, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:27	End Time (MST):	15:37
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
418.9	417.8	1.0026		
221.4	218.9	1.0114		
120.5	117.0	1.0300		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 22, 2023

Location: Wapasu





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Wapasu Station number: AMS17  
 Calibration Date: September 14, 2023 Last Cal Date: August 31, 2023  
 Start time (MST): 10:20 End time (MST): 14:08  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: API T700 Serial Number: 2449  
 ZAG Make/Model: API T701H Serial Number: 359

### Analyzer Information

Analyzer make: API T400 Analyzer serial #: 3870  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009629	1.004971	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.060000	-1.120000	Coeff or Slope:	1.026	1.014

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.8	----
as found span	5000	1077.3	400.0	397.8	1.006
as found 2nd point	5000	900.3	200.0	197.4	1.013
as found 3rd point	5000	789.5	100.0	97.7	1.024
calibrator zero	5000	0.0	0.0	-0.3	----
high point	5000	1077.3	400.0	401.1	0.997
second point	5000	900.3	200.0	200.0	1.000
third point	5000	789.5	100.0	98.2	1.018
as left zero	5000	0.0	0.0	0.4	----
as left span	5000	1077.3	400.0	406.0	0.985
Average Correction Factor					1.005

Baseline Corr As found:	398.6	Previous response	403.8	*% change	-1.3%
Baseline Corr 2nd AF pt:	198.2	AF Slope:	0.997171	AF Intercept:	-1.480000
Baseline Corr 3rd AF pt:	98.5	AF Correlation:	0.999986		

\* = > +/-5% change initiates investigation

Notes: Sample pump changed out. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 31, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:20	End Time (MST):	14:08
Analyzer make:	API T400	Analyzer serial #:	3870

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999976	≥0.995
400.0	401.1	0.9973			
200.0	200.0	1.0000	Slope	1.004971	0.90 - 1.10
100.0	98.2	1.0183			
			Intercept	-1.120000	+/- 5



O<sub>3</sub> Calibration Plot

Date: September 14, 2023

Location: Wapasu







# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Wapasu Station number: AMS 17  
 Calibration Date: September 22, 2023 Last Cal Date: August 15, 2023  
 Start time (MST): 13:18 End time (MST): 15:02

Analyzer Make: API T640 S/N: 1183  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748  
 Temp/RH standard: Alicat FP-25BT S/N: 388748

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	22.2	23.0	22.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	713.7	715.2	713.7	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	4.98	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 22, 2023</u>	Last Cal Date: <u>August 15, 2023</u>			
	PM w/o HEPA: <u>9.3</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.2		11.2	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>10.9</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>September 22, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 22, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_  
 Date RH/T Sensor Cleaned: \_\_\_\_\_

Notes: Temp, pressure and flow checked. Leak check passed. Chamber cleaned, filter swapped.

Calibration by: Aswin Sasi Kumar



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS18 STONY MOUNTAIN SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	September 8, 2023	Last Cal Date:	August 10, 2023
Start time (MST):	11:06	End time (MST):	15:33
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.40	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC463851			
Removed Cal Gas Conc:	49.40	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API 701H		Serial Number:	360

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004073	0.999906	Backgd or Offset:	22.6	22.2
Calibration intercept:	-0.542544	-1.183922	Coeff or Slope:	0.808	0.795

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5009	0.0	0.0	0.2	----
as found span	4919	81.0	800.3	808.3	0.990
as found 2nd point	4959	40.5	400.2	403.4	0.992
as found 3rd point	4979	20.2	199.6	201.1	0.993
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4919	81.0	800.3	799.6	1.001
second point	4959	40.5	400.2	398.6	1.004
third point	4979	20.2	199.6	196.8	1.014
as left zero	5000	0.0	0.0	0.4	----
as left span	4919	81.0	800.3	799.7	1.001
Average Correction Factor					1.006

Baseline Corr As found:	808.10	Previous response	803.00	*% change	0.6%
Baseline Corr 2nd AF pt:	403.20	AF Slope:	1.009927	AF Intercept:	-0.241517
Baseline Corr 3rd AF pt:	200.90	AF Correlation:	0.999998		

\* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 8, 2023	Previous Calibration:	August 10, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:06	End Time (MST):	15:33
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999986	≥0.995
800.3	799.6	1.0009			
400.2	398.6	1.0040	Slope	0.999906	0.90 - 1.10
199.6	196.8	1.0143			
			Intercept	-1.183922	+/-30



SO2 Calibration Plot

Date: September 8, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Stony Mountain Station number: AMS18  
 Calibration Date: September 25, 2023 Last Cal Date: August 24, 2023  
 Start time (MST): 9:30 End time (MST): 14:38  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.479 ppm Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC500395  
 Removed Cal Gas Conc: 5.479 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658  
 ZAG Make/Model: Teledyne API T701 Serial Number: 360

### Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359  
 Converter make: CD Nova CDN-101 Converter serial #: 555  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993582	1.005299	Backgd or Offset: 2.66	2.66
Calibration intercept:	0.341175	-0.019011	Coeff or Slope: 1.189	1.173

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4927	73.0	80.0	82.0	0.977
as found 2nd point	4964	36.5	40.0	41.1	0.975
as found 3rd point	4983	18.3	20.0	20.4	0.988
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4927	73.0	80.0	80.4	0.995
second point	4964	36.5	40.0	40.3	0.992
third point	4983	18.3	20.0	19.9	1.007
as left zero	5000	0.0	0.0	0.3	----
as left span	4927	73.0	80.0	77.6	1.031
SO2 Scrubber Check	4923	77.1	771.0	0.1	----
Date of last scrubber change:	17-Dec-21			Ave Corr Factor	0.998
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 81.9 Prev response: 79.82 \*% change: 2.5%  
 Baseline Corr 2nd AF pt: 41.0 AF Slope: 1.024877 AF Intercept: 0.020582  
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999988

\* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Scrubber check completed after calibrator zero. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 25, 2023	Previous Calibration:	August 24, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	9:30	End Time (MST):	14:38
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
80.0	80.4	0.9949		
40.0	40.3	0.9924	Slope	0.90 - 1.10
20.0	19.9	1.0074		
			Intercept	+/-3



TRS Calibration Plot

Date: September 25, 2023

Location: Stony Mountain







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	September 8, 2023	Last Cal Date:	N/A
Start time (MST):	13:00	End time (MST):	15:33
Reason:	Install		

### Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH <sub>4</sub> Cal Gas Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	N/A	2.23E-04	NMHC SP Ratio:	N/A	4.19E-05
CH <sub>4</sub> Retention time:	N/A	12.9	NMHC Peak Area:	N/A	186711
Zero Chromatogram:		OFF	Flat Baseline:		OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.34	0.997
second point	4959	40.5	8.64	8.62	1.003
third point	4979	20.2	4.31	4.26	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.31	0.998
Average Correction Factor					1.004

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.24	0.992
second point	4959	40.5	4.58	4.61	0.995
third point	4979	20.2	2.29	2.30	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	9.17	9.17	1.000
Average Correction Factor					0.994
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	8.11	1.000
second point	4959	40.5	4.06	4.01	1.012
third point	4979	20.2	2.02	1.96	1.030
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.14	0.997
Average Correction Factor					1.014
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		1.004309
THC Cal Offset:		-0.036379
CH <sub>4</sub> Cal Slope:		1.001484
CH <sub>4</sub> Cal Offset:		-0.032618
NMHC Cal Slope:		1.007917
NMHC Cal Offset:		-0.005959

Notes: Install calibration . Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 8, 2023	Previous Calibration:	N/A
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	13:00	End Time (MST):	15:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999979	$\geq 0.995$			
17.28	17.34	0.9966						
8.64	8.62	1.0028				Slope	1.004309	0.90 - 1.10
4.31	4.26	1.0116						
			Intercept	-0.036379	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 8, 2023	Previous Calibration:	N/A
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	13:00	End Time (MST):	15:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999926	≥0.995
8.11	8.11	1.0004			
4.06	4.01	1.0117			
2.02	1.96	1.0303			
			Slope	1.001484	0.90 - 1.10
			Intercept	-0.032618	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 8, 2023	Previous Calibration:	N/A
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	13:00	End Time (MST):	15:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	$\geq 0.995$			
9.17	9.24	0.9923						
4.58	4.61	0.9954				Slope	1.007917	0.90 - 1.10
2.29	2.30	0.9956						
			Intercept	-0.005959	$\pm 0.5$			



NMHC Calibration Plot

Date: September 8, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	September 8, 2023	Last Cal Date:	August 26, 2023
Start time (MST):	11:05	End time (MST):	N/A
Reason:	Removal		

### Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH <sub>4</sub> Cal Gas Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.37E-04	N/A	NMHC SP Ratio:	6.89E-05
CH <sub>4</sub> Retention time:	15.6	N/A	NMHC Peak Area:	133711
Zero Chromatogram:	ON		Flat Baseline:	ON

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.02	----
as found span	4919	81.0	17.28	16.26	1.063
as found 2nd point	4959	40.5	8.64	7.98	1.083
as found 3rd point	4979	20.2	4.31	4.08	1.057
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

			Average Correction Factor		
Baseline Corr AF:	16.24	Prev response	17.28	*% change	-6.4%
Baseline Corr 2nd AF:	8.0	AF Slope:	0.938700	AF Intercept:	-0.010991
Baseline Corr 3rd AF:	4.1	AF Correlation:	0.999854	* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	8.94	1.026
as found 2nd point	4959	40.5	4.58	4.49	1.022
as found 3rd point	4979	20.2	2.29	2.25	1.015
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	8.94	Prev response	9.17	*% change	-2.6%
Baseline Corr 2nd AF:	4.5	AF Slope:	0.974079	AF Intercept:	0.012979
Baseline Corr 3rd AF:	2.3	AF Correlation:	0.999990	* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.02	----
as found span	4919	81.0	8.11	7.33	1.107
as found 2nd point	4959	40.5	4.06	3.49	1.162
as found 3rd point	4979	20.2	2.02	1.82	1.109
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	7.31	Prev response	8.11	*% change	-11.0%
Baseline Corr 2nd AF:	3.47	AF Slope:	0.898987	AF Intercept:	-0.024170
Baseline Corr 3rd AF:	1.80	AF Correlation:	0.999188	* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999923	N/A
THC Cal Offset:	-0.001977	N/A
CH <sub>4</sub> Cal Slope:	0.999791	N/A
CH <sub>4</sub> Cal Offset:	-0.001608	N/A
NMHC Cal Slope:	1.000027	N/A
NMHC Cal Offset:	-0.000569	N/A

Notes: Removing instrument due to dips. Further maintenance to be conducted at the WBEA Centre.


Calibration Performed By: Aswin Sasi Kumar



NMHC Calibration Plot

Date: September 8, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Stony Mountain Station number: AMS 18  
Calibration Date: September 19, 2023 Last Cal Date: August 30, 2023  
Start time (MST): 10:20 End time (MST): 15:04  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T2XX7ME Cal Gas Expiry Date: January 14, 2024  
NOX Cal Gas Conc: 50.48 ppm NO Cal Gas Conc: 49.22 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 50.48 ppm Removed Gas NO Conc: 49.22 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 2658  
ZAG make/model: Teledyne API 701H Serial Number: 360

### Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1336160088  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.993	1.029	NO bkgnd or offset:	2.8	2.9
NOX coeff or slope:	0.984	0.984	NOX bkgnd or offset:	2.8	2.9
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	230.6	235.3

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000289	1.000359
NO <sub>x</sub> Cal Offset:	-0.209933	-0.310049
NO Cal Slope:	0.999824	1.001795
NO Cal Offset:	-1.429814	-1.669844
NO <sub>2</sub> Cal Slope:	1.011795	0.998696
NO <sub>2</sub> Cal Offset:	-0.075280	-0.059273



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
as found span	4919	81.3	820.8	800.3	20.5	810.0	786.2	24.2	1.0133	1.0179
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.2	----	----
high point	4919	81.3	820.8	800.3	20.5	821.0	800.8	20.3	0.9997	0.9993
second point	4959	40.7	410.9	400.7	10.3	410.7	399.2	11.5	1.0006	1.0037
third point	4980	20.3	204.9	199.8	5.1	203.8	196.5	7.3	1.0056	1.0169
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.3	----	----
as left span	4919	81.3	820.8	383.2	437.6	824.0	384.7	439.5	0.9961	0.9960
Average Correction Factor									1.0019	1.0066

Corrected As found	NO <sub>x</sub> = 809.8 ppb	NO = 786.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -1.4%
Previous Response	NO <sub>x</sub> = 820.8 ppb	NO = 798.7 ppb		*Percent Change	NO = -1.6%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.7	380.6	437.6	437.1	1.0011	99.9%
2nd GPT point (200 ppb O3)	797.7	594.0	224.2	223.7	1.0022	99.8%
3rd GPT point (100 ppb O3)	797.7	696.4	121.8	121.3	1.0040	99.6%
Average Correction Factor					1.0024	99.8%

Notes: Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 19, 2023	Previous Calibration:	August 30, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:20	End Time (MST):	15:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
820.8	821.0	0.9997		
410.9	410.7	1.0006		
204.9	203.8	1.0056		
			0.999996	
			1.000359	
			-0.310049	





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 19, 2023	Previous Calibration:	August 30, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:20	End Time (MST):	15:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	800.8	0.9993		
400.7	399.2	1.0037		
199.8	196.5	1.0169		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 19, 2023	Previous Calibration:	August 30, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:20	End Time (MST):	15:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
437.6	437.1	1.0011			
224.2	223.7	1.0022			
121.8	121.3	1.0040			
			Slope	0.998696	0.90 - 1.10
			Intercept	-0.059273	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 19, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Stony Mountain      Station number: AMS18  
 Calibration Date: September 21, 2023      Last Cal Date: August 23, 2023  
 Start time (MST): 11:21      End time (MST): 14:13  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T750      Serial Number: 282  
 ZAG Make/Model: Teledyne API T751H      Serial Number: 321

### Analyzer Information

Analyzer make: API T400      Analyzer serial #: 825  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003086	1.003943	Backgd or Offset:	1.300	1.300
Calibration intercept:	0.360000	-0.340000	Coeff or Slope:	1.002	0.996

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.1	----
as found span	4888	1096.9	400.0	404.5	0.989
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.4	----
high point	4888	1101.7	400.0	401.4	0.997
second point	4888	863.9	200.0	200.7	0.997
third point	4888	741.4	100.0	98.9	1.011
as left zero	5000	800.0	0.0	0.2	----
as left span	4812	1097.9	400.0	413.6	0.967
Average Correction Factor					1.001

Baseline Corr As found:	404.4	Previous response	401.6	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Using portable calibrator for the calibration. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 21, 2023	Previous Calibration:	August 23, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	11:21	End Time (MST):	14:13
Analyzer make:	API T400	Analyzer serial #:	825

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999978	
400.0	401.4	0.9965			≥0.995
200.0	200.7	0.9965	Slope	1.003943	
100.0	98.9	1.0111			0.90 - 1.10
			Intercept	-0.340000	+/- 5



# O<sub>3</sub> Calibration Plot

Date: September 21, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Stony Mountain Station number: AMS 18  
 Calibration Date: September 25, 2023 Last Cal Date: August 30, 2023  
 Start time (MST): 13:16 End time (MST): 14:37

Analyzer Make: API T640 S/N: 1162  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748  
 Temp/RH standard: Alicat FP-25BT S/N: 388748

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	21.3	21.5	21.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	702.4	704.3	702.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.94	5.05	4.94	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 25, 2023</u>	Last Cal Date: <u>August 30, 2023</u>			
	PM w/o HEPA: <u>3.9</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.1		11.1	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>3.9</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>September 25, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 25, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022  
 Date RH/T Sensor Cleaned: August 30, 2022

Notes: No adjustments needed.

Calibration by: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	September 11, 2023	Last Cal Date:	August 21, 2023
Start time (MST):	10:43	End time (MST):	13:00
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,050	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	3,050	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701		Serial Number:	360

### Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	<b><u>Start</u></b>	<b><u>Finish</u></b>		<b><u>Start</u></b>	<b><u>Finish</u></b>
Calibration slope:	0.999692	0.995840	Backgd or Offset:	-0.010	-0.010
Calibration intercept:	0.079759	0.085810	Coeff or Slope:	0.906	0.906

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4933	66.7	40.7	40.6	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.7	40.5	1.004
second point	4966	33.3	20.3	20.5	0.990
third point	4983	16.7	10.2	10.2	1.002
as left zero					
as left span					
Average Correction Factor					0.999

Baseline Corr As found:	40.61	Prev response:	40.76	*% change:	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments needed. Power to the station went out during third point. No as lefts done.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 11, 2023	Previous Calibration:	August 21, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:43	End Time (MST):	13:00
Analyzer make:	API T300	Analyzer serial #:	3504

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999938	
40.7	40.5	1.0042			≥0.995
20.3	20.5	0.9901	Slope	0.995840	
10.2	10.2	1.0017			0.90 - 1.10
			Intercept	0.085810	+/-1.5



CO Calibration Plot

Date: September 11, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	September 12, 2023	Last Cal Date:	August 31, 2023
Start time (MST):	10:20	End time (MST):	14:20
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2026
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
N2 Gen Make/Model:	Peak Scientific		Serial Number:	771048317

### Analyzer Information

Analyzer make: API T360	Analyzer serial #: 288
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000937	0.999777	Backgd or Offset:	0.041	-0.002
Calibration intercept:	-4.800000	-1.960000	Coeff or Slope:	1.009	1.051

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	3.0	----
as found span	2920	80.0	1605.9	1602.0	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	3.3	----
high point	2920	80.0	1605.9	1608.2	0.999
second point	2960	40.0	802.9	792.1	1.014
third point	2980	20.0	401.5	398.2	1.008
as left zero	3000	0.0	0.0	3.5	----
as left span	2930	80.0	1600.5	1609.9	0.994
Average Correction Factor					1.007

Baseline Corr As found:	1599.00	Prev response:	1602.57	*% change:	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Zero and span adjusted. Third point failed, performed linearity adjustment. Span and mid point checked. Zero point re-checked after mid point.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date	September 12, 2023	Previous Calibration	August 31, 2023
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:20	End Time (MST)	14:20
Analyzer make	API T360	Analyzer serial #	288

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	3.3	----	Correlation Coefficient	0.999910	≥0.995
1605.9	1608.2	0.9985			
802.9	792.1	1.0137	Slope	0.999777	0.90 - 1.10
401.5	398.2	1.0082			
			Intercept	-1.960000	+/-10






CO<sub>2</sub> Calibration Plot

Date: September 12, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	September 29, 2023	Last Cal Date:	NA
Start time (MST):	11:20	End time (MST):	16:00
Reason:	Install		

### Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2026
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
N2 Gen Make/Model:	Peak Scientific		Serial Number:	771048317

### Analyzer Information

Analyzer make: API T360	Analyzer serial #: 489
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	0.998830	Backgd or Offset:	NA	-0.002
Calibration intercept:	NA	-5.420000	Coeff or Slope:	NA	1.051

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	1.4	----
high point	2920	80.0	1605.9	1606.7	0.999
second point	2960	40.0	802.9	779.0	1.031
third point	2980	20.0	401.5	398.2	1.008
as left zero	3000	0.0	0.0	5.5	----
as left span	2930	80.0	1600.5	1613.7	0.992
Average Correction Factor					1.013

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Instrument installation calibration, linearity adjustment performed.

Calibration Performed By: Kelly Baragar



# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date	September 29, 2023	Previous Calibration	NA
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	11:20	End Time (MST)	16:00
Analyzer make	API T360	Analyzer serial #	489

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.4	----	Correlation Coefficient	0.999696	≥0.995
1605.9	1606.7	0.9995			
802.9	779.0	1.0307	Slope	0.998830	0.90 - 1.10
401.5	398.2	1.0082			
			Intercept	-5.420000	+/-10



CO<sub>2</sub> Calibration Plot

Date: September 29, 2023

Location: Stony Mountain





# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Stony Mountain	Station Number:	AMS 18
Calibration Date:	September 28, 2023	Prev Cal Date:	August 24, 2023
Start Time (MST):	11:40	End Time (MST):	12:18
Tower Height (m):	20.0	Reason:	Removal

### Wind Speed Information

Sensor make/model:	NA	Serial Number:	NA
WS Calibrator:	NA	Serial Number:	NA

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0			---
200			
400			
600			
800			

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )			≥0.9995
Calculated slope			0.90 - 1.10
Calculated intercept			+/- 2

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	D14057
As Found Declination (deg east of True North):	14	As Left Declination (deg east of True North):	14
Solar noon time (MST):	13:27	Calc Declination*:	13.2 Degrees
Deadband calc:	-1.8 degrees ( <i>Limit 4 deg</i> )	* - calculated declination as per NOAA website	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.4	---
90	86.9	-0.9%
180	178.5	-0.4%
270	269.1	-0.2%
357	359.2	0.6%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999920	≥0.9995
Calculated slope		0.993367	0.90 - 1.10
Calculated intercept		1.758093	+/- 4

Notes: Removing WD sensor on the 20m tower due to suspect data. Crossarm aligned with true north.

Calibration Performed By: Rene, Ryan, & Devin



# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Stony Mountain	Station Number:	AMS 18
Calibration Date:	September 28, 2023	Prev Cal Date:	NA
Start Time (MST):	12:00	End Time (MST):	12:20
Tower Height (m):	20.0	Reason:	Install

### Wind Speed Information

Sensor make/model:	NA	Serial Number:	NA
WS Calibrator:	NA	Serial Number:	NA

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0			---
200			
400			
600			
800			

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )			≥0.9995
Calculated slope			0.90 - 1.10
Calculated intercept			+/- 2

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	C21021
As Found Declination (deg east of True North):	14	As Left Declination (deg east of True North):	14
Solar noon time (MST):	13:27	Calc Declination*:	13.2 Degrees
Deadband calc:	-2.2 degrees ( <i>Limit 4 deg</i> )		* - calculated declination as per NOAA website

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.4	---
90	89.9	0.0%
180	179.4	-0.2%
270	270.3	0.1%
357	359.6	0.7%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999977	≥0.9995
Calculated slope		0.994666	0.90 - 1.10
Calculated intercept		0.477464	+/- 4

Notes: Installing new WD sensors on the 20m tower. Crossarm aligned with true north. No issues to note.

Calibration Performed By: Rene, Ryan & Devin



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS19 FIREBAG SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023







# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	September 28, 2023	Last Cal Date:	August 14, 2023
Start time (MST):	9:43	End time (MST):	14:03
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.29	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC716618			
Removed Cal Gas Conc:	49.29	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1607
ZAG Make/Model:	API T701		Serial Number:	1118

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001265	0.999292	Backgd or Offset:	10.3	10.8
Calibration intercept:	0.897402	0.237822	Coeff or Slope:	1.002	0.975

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.5	----
as found span	4919	81.1	799.5	823.0	0.971
as found 2nd point	4959	40.6	400.3	412.2	0.971
as found 3rd point	4980	20.3	200.1	206.6	0.969
new cylinder response					
calibrator zero	4999	0.0	0.0	-0.1	----
high point	4919	81.1	799.5	798.8	1.001
second point	4959	40.6	400.3	400.9	0.998
third point	4980	20.3	200.1	200.2	1.000
as left zero	4999	0.0	0.0	0.0	----
as left span	4919	81.1	799.5	796.0	1.004
Average Correction Factor					1.000

Baseline Corr As found:	822.50	Previous response	801.38	*% change	2.6%
Baseline Corr 2nd AF pt:	411.70	AF Slope:	1.028697	AF Intercept:	0.572269
Baseline Corr 3rd AF pt:	206.10	AF Correlation:	1.000000		

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 28, 2023	Previous Calibration:	August 14, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:43	End Time (MST):	14:03
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999998	≥0.995
799.5	798.8	1.0008			
400.3	400.9	0.9984	Slope	0.999292	0.90 - 1.10
200.1	200.2	0.9995			
			Intercept	0.237822	+/-30



SO2 Calibration Plot

Date: September 28, 2023

Location: Firebag





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Firebag Station number: AMS19  
 Calibration Date: September 11, 2023 Last Cal Date: August 10, 2023  
 Start time (MST): 10:05 End time (MST): 14:43  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.114 ppm Cal Gas Exp Date: February 5, 2024  
 Cal Gas Cylinder #: CC517427  
 Removed Cal Gas Conc: 5.114 ppm Rem Gas Exp Date: n/a  
 Removed Gas Cyl #: n/a Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1607  
 ZAG Make/Model: Teledyne API T701 Serial Number: 1118

### Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680032  
 Converter make: Global Converter serial #: 2022-222  
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000053	0.993765	Backgd or Offset: 2.55	2.53
Calibration intercept:	-0.081596	-0.161463	Coeff or Slope: 1.122	1.122

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4922	78.2	80.0	80.2	0.995
as found 2nd point	4961	39.1	40.0	40.0	0.995
as found 3rd point	4980	19.6	20.0	19.7	1.007
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4922	78.2	80.0	79.3	1.009
second point	4961	39.1	40.0	39.6	1.010
third point	4980	19.6	20.0	19.8	1.013
as left zero	5000	0.0	0.0	0.0	----
as left span	4922	78.2	80.0	78.5	1.019
SO2 Scrubber Check	4922	78.3	800.2	0.1	----
Date of last scrubber change:	January 18, 2023			Ave Corr Factor	1.010
Date of last converter efficiency test:	n/a			efficiency	

Baseline Corr As found: 80.4 Prev response: 79.90 \*% change: 0.6%  
 Baseline Corr 2nd AF pt: 40.2 AF Slope: 1.006339 AF Intercept: -0.301622  
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999988

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after MPAF's. Ran SO<sub>x</sub> scrubber check after cal zero. No adjustments needed.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 11, 2023	Previous Calibration:	August 10, 2023
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	10:05	End Time (MST):	14:43
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999999	≥0.995
80.0	79.3	1.0086			
40.0	39.6	1.0099	Slope	0.993765	0.90 - 1.10
20.0	19.8	1.0125			
			Intercept	-0.161463	+/-3



H<sub>2</sub>S Calibration Plot

Date: September 11, 2023

Location: Firebag





# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name: Firebag Station number: AMS 19  
 Calibration Date: September 28, 2023 Last Cal Date: August 29, 2023  
 Start time (MST): 9:43 End time (MST): 14:03  
 Reason: Routine

### Calibration Standards

Gas Cert Reference: CC716618 Cal Gas Expiry Date: February 23, 2025  
 CH4 Cal Gas Conc. 500.7 ppm CH4 Equiv Conc. 1066.9 ppm  
 C3H8 Cal Gas Conc. 205.9 ppm  
 Removed Gas Cert: Removed Gas Expiry:  
 Removed CH4 Conc. 500.7 ppm CH4 Equiv Conc. 1066.9 ppm  
 Removed C3H8 Conc. 205.9 ppm Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 1607  
 ZAG Make/Model: API T701 Serial Number: 1118

### Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1336160089  
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994308	0.998771	Background:	3.27	3.48
Calibration intercept:	-0.010322	-0.052126	Coefficient:	3.762	3.796

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.00	0.05	----
as found span	4919	81.1	17.31	17.21	1.006
as found 2nd point	4959	40.6	8.66	8.60	1.007
as found 3rd point	4980	20.3	4.33	4.44	0.976
<i>new cylinder response</i>					
calibrator zero	4999	0.0	0.00	-0.01	----
high point	4919	81.1	17.31	17.27	1.002
second point	4959	40.6	8.66	8.53	1.016
third point	4980	20.3	4.33	4.26	1.016
as left zero	5000	0.0	0.00	-0.03	----
as left span	4919	81.1	17.31	17.41	0.994
Average Correction Factor					1.011
Baseline Corr As found:	17.16	Previous response	17.20	*% change	-0.2%
Baseline Corr 2nd AF pt:	8.55	AF Slope:	0.989495	AF Intercept:	0.078885
Baseline Corr 3rd AF pt:	4.39	AF Correlation:	0.999944		

\* = > +/-5% change initiates investigation

Notes: Multipoint as founds done prior to installing firmware update. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 28, 2023	Previous Calibration:	August 29, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:43	End Time (MST):	14:03
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

### Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	-0.01	----	Correlation Coefficient	0.999948	
17.31	17.27	1.0020			≥0.995
8.66	8.53	1.0157	Slope	0.998771	
4.33	4.26	1.0158			0.90 - 1.10
			Intercept	-0.052126	+/-1.5






THC Calibration Plot

Date: September 28, 2023

Location: Firebag





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Firebag  
Calibration Date: September 29, 2023  
Start time (MST): 9:49  
Reason: Routine  
Station number: AMS 19  
Last Cal Date: August 16, 2023  
End time (MST): 14:53

### Calibration Standards

NO Gas Cylinder #: T2Y1K63  
NOX Cal Gas Conc: 51.12 ppm  
Removed Cylinder #: n/a  
Removed Gas NOX Conc: 51.12 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701  
Cal Gas Expiry Date: November 30, 2023  
NO Cal Gas Conc: 49.40 ppm  
Removed Gas Exp Date: n/a  
Removed Gas NO Conc: 49.40 ppm  
NO gas Diff:  
Serial Number: 1607  
Serial Number: 1118

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1410661309

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.112	1.098	NO bkgnd or offset:	7.8	7.6
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	7.8	8.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	213.1	213.1

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.998001	0.998622
NO <sub>x</sub> Cal Offset:	1.176081	0.176163
NO Cal Slope:	0.999755	1.000897
NO Cal Offset:	0.048770	0.249004
NO <sub>2</sub> Cal Slope:	1.004673	0.998565
NO <sub>2</sub> Cal Offset:	0.372474	-1.744715



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.0	0.0	0.8	-0.2	1.0	----	----
as found span	4919	81.0	828.1	800.3	27.9	847.0	816.0	31.3	0.9777	0.9807
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	4999	0.0	0.0	0.0	0.0	-0.4	0.0	-0.4	----	----
high point	4919	81.0	828.1	800.3	27.9	827.0	801.0	25.8	1.0014	0.9991
second point	4960	40.5	414.0	400.1	13.9	413.6	401.2	12.4	1.0010	0.9973
third point	4980	20.2	206.5	199.6	6.9	207.2	200.0	7.2	0.9967	0.9978
as left zero	4999	0.0	0.0	0.0	0.0	-0.7	0.0	-0.7	----	----
as left span	4919	81.0	828.1	364.1	464.1	827.0	365.1	461.9	1.0014	0.9972
Average Correction Factor									0.9997	0.9981

Corrected As found	NO <sub>x</sub> = 846.2 ppb	NO = 816.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 2.2%
Previous Response	NO <sub>x</sub> = 827.7 ppb	NO = 800.1 ppb		*Percent Change	NO = 2.0%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	800.0	363.8	464.1	462.5	1.0034	99.7%
2nd GPT point (200 ppb O3)	800.0	582.9	245.0	242.0	1.0122	98.8%
3rd GPT point (100 ppb O3)	800.0	691.6	136.3	133.0	1.0245	97.6%
Average Correction Factor					1.0134	98.7%

Notes:

Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By:

Braiden Boutilier



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 29, 2023	Previous Calibration:	August 16, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:49	End Time (MST):	14:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.4	----	Correlation Coefficient	≥0.995	
828.1	827.0	1.0014			
414.0	413.6	1.0010			
206.5	207.2	0.9967			
			Slope	0.998622	0.90 - 1.10
			Intercept	0.176163	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 29, 2023	Previous Calibration:	August 16, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:49	End Time (MST):	14:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
800.3	801.0	0.9991			0.999999
400.1	401.2	0.9973			1.000897
199.6	200.0	0.9978			0.249004

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 29, 2023	Previous Calibration:	August 16, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:49	End Time (MST):	14:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.4	----	Correlation Coefficient	≥0.995	
464.1	462.5	1.0034			
245.0	242.0	1.0122			
136.3	133.0	1.0245			
			Slope	0.998565	0.90 - 1.10
			Intercept	-1.744715	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 29, 2023

Location: Firebag





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS20 MACKAY RIVER SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023







# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	August 8, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:46	End Time (MST):	10:25
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999934	≥0.995
800.3	801.8	0.9981			
400.7	403.2	0.9937	Slope	0.998878	0.90 - 1.10
199.8	206.7	0.9667			
			Intercept	3.251492	+/-30



SO2 Calibration Plot

Date: September 12, 2023

Location: MacKay River





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: MacKay River      Station number: AMS20  
 Calibration Date: September 11, 2023      Last Cal Date: August 4, 2023  
 Start time (MST): 6:58      End time (MST): 11:07  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.12 ppm      Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC515997  
 Removed Cal Gas Conc: 5.12 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 1220  
 ZAG Make/Model: Teledyne API 701      Serial Number: 4522

### Analyzer Information

Analyzer make: Thermo 43iQ TLE      Analyzer serial #: 12124313139  
 Converter make: Global      Converter serial #: 2022-226  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994521	0.995379	Backgd or Offset: 1.0	0.91
Calibration intercept:	0.287104	0.407078	Coeff or Slope: 0.535	0.501

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4922	78.1	80.0	86.7	0.923
as found 2nd point	4961	39.0	39.9	43.6	0.918
as found 3rd point	4980	19.5	20.0	22.1	0.908
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4922	78.1	80.0	80.0	1.000
second point	4961	39.0	40.0	40.2	0.994
third point	4980	19.5	20.0	20.6	0.972
as left zero	5000	0.0	0.0	0.2	----
as left span	4922	78.1	80.0	79.0	1.013
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:	May 25, 2023			Ave Corr Factor	0.989
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	86.6	Prev response:	79.82	*% change:	<b>7.8%</b>
Baseline Corr 2nd AF pt:	43.5	AF Slope:	1.081621	AF Intercept:	0.301040
Baseline Corr 3rd AF pt:	22.0	AF Correlation:	0.999975		

\* = > +/-5% change initiates investigation

Notes: Lamp voltage slowly going up causing the spans to go up. Changing Lamp didn't work, continued calibration. Sox scrubber checked after calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 11, 2023	Previous Calibration:	August 4, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	6:58	End Time (MST):	11:07
Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	12124313139

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999965	≥0.995
80.0	80.0	1.0004			
40.0	40.2	0.9942	Slope	0.995379	0.90 - 1.10
20.0	20.6	0.9717			
			Intercept	0.407078	+/-3



H<sub>2</sub>S Calibration Plot

Date: September 11, 2023

Location: MacKay River





# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name:	MacKay River	Station number:	AMS20
Calibration Date:	September 12, 2023	Last Cal Date:	August 8, 2023
Start time (MST):	7:46	End time (MST):	10:24
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC306868	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.40</u> ppm	CH4 Equiv Conc.	1066.45 ppm
C3H8 Cal Gas Conc.	<u>206.20</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.40</u> ppm	CH4 Equiv Conc.	1066.45 ppm
Removed C3H8 Conc.	<u>206.20</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	1220
ZAG Make/Model:	Teledyne API 701	Serial Number:	4522

### Analyzer Information

Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727
Analyzer Range:	0 - 20 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992565	0.991950	Background:	3.090	3.330
Calibration intercept:	0.123818	0.123233	Coefficient:	5.465	5.583

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.19	----
as found span	4919	81.3	17.34	17.23	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.10	----
high point	4919	81.3	17.34	17.33	1.001
second point	4959	40.7	8.68	8.69	0.999
third point	4980	20.3	4.33	4.48	0.966
as left zero	5000	0.0	0.00	-0.07	----
as left span	4919	81.3	17.34	17.26	1.005
Average Correction Factor					0.989
Baseline Corr As found:	17.04	Previous response	17.33	*% change	-1.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No maintenance done. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	August 8, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:46	End Time (MST):	10:24
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

### Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.10	----	Correlation Coefficient	0.999960	≥0.995
17.34	17.33	1.0005			
8.68	8.69	0.9990	Slope	0.991950	0.90 - 1.10
4.33	4.48	0.9664			
			Intercept	0.123233	+/-1.5






THC Calibration Plot

Date: September 12, 2023

Location: MacKay River





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: MacKay River Station number: AMS20  
Calibration Date: September 14, 2023 Last Cal Date: August 3, 2023  
Start time (MST): 6:40 End time (MST): 10:50  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025  
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 1220  
ZAG make/model: Teledyne API 701 Serial Number: 4522

### Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.503	1.536	NO bkgnd or offset:	4.2	4.3
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	4.2	4.3
NO2 coeff or slope:	0.995	0.995	Reaction cell Press:	182.5	182.5

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.986386	0.996495
NO <sub>x</sub> Cal Offset:	3.150520	3.649884
NO Cal Slope:	0.988909	0.997991
NO Cal Offset:	1.711850	2.330879
NO <sub>2</sub> Cal Slope:	1.003108	1.001543
NO <sub>2</sub> Cal Offset:	1.350779	-0.862665



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
as found span	4917	83.3	819.5	800.3	19.2	800.3	780.6	19.7	1.0239	1.0252
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.3	----	----
high point	4917	83.3	819.5	800.3	19.2	819.0	800.2	18.7	1.0006	1.0001
second point	4956	41.7	410.4	400.8	9.6	413.1	402.6	10.4	0.9935	0.9956
third point	4979	20.8	204.6	199.9	4.8	211.6	204.6	7.0	0.9671	0.9768
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.3	----	----
as left span	4917	83.3	819.5	442.1	377.4	818.7	438.7	380.0	1.0009	1.0077
Average Correction Factor									0.9871	0.9909

Corrected As found	NO <sub>x</sub> = 800.2 ppb	NO = 780.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -1.4%
Previous Response	NO <sub>x</sub> = 811.5 ppb	NO = 793.1 ppb		*Percent Change	NO = -1.6%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.0	437.8	377.4	377.8	0.9988	100.1%
2nd GPT point (200 ppb O3)	796.0	610.0	205.2	204.0	1.0057	99.4%
3rd GPT point (100 ppb O3)	796.0	699.6	115.6	113.6	1.0172	98.3%
Average Correction Factor					1.0072	99.3%

Notes:

No maintenance done. Span adjusted.

Calibration Performed By:

Melissa Lemay



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 3, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	6:40	End Time (MST):	10:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
819.5	819.0	1.0006		
410.4	413.1	0.9935		
204.6	211.6	0.9671		





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 3, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	6:40	End Time (MST):	10:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	800.2	1.0001		
400.8	402.6	0.9956		
199.9	204.6	0.9768		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 3, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	6:40	End Time (MST):	10:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits			
0.0	0.3	----	Correlation Coefficient	0.999949	≥0.995			
377.4	377.8	0.9988						
205.2	204.0	1.0057				Slope	1.001543	0.90 - 1.10
115.6	113.6	1.0172						
			Intercept	-0.862665	+/-20			



NO<sub>x</sub> Calibration Plot

Date: September 14, 2023

Location: MacKay River





# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Mackay River	Station Number:	AMS 20
Calibration Date:	September 12, 2023	Prev Cal Date:	August 23, 2022
Start Time (MST):	9:13	End Time (MST):	9:43
Tower Height (m):	10m	Reason:	Routine

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	Y18363
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.2	0.2%
400	39.4	39.4	0.1%
600	58.6	58.7	0.2%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	0.999999	0.999999	<i>≥0.9995</i>
Calculated slope	0.998857	0.998949	<i>0.90 - 1.10</i>
Calculated intercept	0.030366	-0.013169	<i>+/- 2</i>

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	N9937
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):		NA Calc Declination*:	NA
Deadband calc:	2.9 degrees ( <i>Limit 4 deg</i> )		Degrees

\* - calculated declination as per NOAA website

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.0	---
90	89.0	-0.3%
180	179.2	-0.2%
270	270.5	0.1%
357	354.1	-0.8%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )	0.999946	0.999969	<i>≥0.9995</i>
Calculated slope	1.009899	1.004747	<i>0.90 - 1.10</i>
Calculated intercept	0.928586	-0.011707	<i>+/- 4</i>

Notes: WS/WD passed the Torque test. WD head replaced, due to crack.

Calibration Performed By: Melissa Lemay





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS21  
CONKLIN  
SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 26, 2023	Previous Calibration:	August 3, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:10	End Time (MST):	14:32
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999998	≥0.995
800.8	802.0	0.9986			
400.4	401.4	0.9976	Slope	1.000644	0.90 - 1.10
200.1	201.7	0.9922			
			Intercept	0.775866	+/-30



SO2 Calibration Plot

Date: September 26, 2023

Location: Conklin





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Conklin Station number: AMS21  
 Calibration Date: September 14, 2023 Last Cal Date: August 24, 2023  
 Start time (MST): 9:30 End time (MST): 13:34  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.00 ppm Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC501204  
 Removed Cal Gas Conc: 5.00 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 3810  
 ZAG Make/Model: API 701H Serial Number: 691

### Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1236656116  
 Converter make: CD-Nova 101 Converter serial #: NA  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004857	1.007429	Backgd or Offset: 2.4	2.4
Calibration intercept:	0.280000	0.140000	Coeff or Slope: 0.958	0.958

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.0	80.0	80.5	0.994
as found 2nd point	4960	40.0	40.0	40.4	0.990
as found 3rd point	4980	20.0	20.0	20.3	0.985
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.0	80.0	80.7	0.991
second point	4960	40.0	40.0	40.5	0.988
third point	4980	20.0	20.0	20.3	0.985
as left zero	5000	0.0	0.0	0.3	----
as left span	4920	80.0	80.0	79.7	1.004
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.988
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 80.5 Prev response: 80.67 \*% change: -0.2%  
 Baseline Corr 2nd AF pt: 40.4 AF Slope: 1.005714 AF Intercept: 0.100000  
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999993

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 24, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:30	End Time (MST):	13:34
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999998	
80.0	80.7	0.9913			≥0.995
40.0	40.5	0.9877	Slope	1.007429	
20.0	20.3	0.9852			0.90 - 1.10
			Intercept	0.140000	+/-3



TRS Calibration Plot

Date: September 14, 2023

Location: Conklin





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	September 26, 2023	Last Cal Date:	August 3, 2023
Start time (MST):	11:10	End time (MST):	14:32
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	497.9 ppm	CH <sub>4</sub> Equiv Conc.	1067.7 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	497.9 ppm	CH <sub>4</sub> Equiv Conc.	1067.7 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.29E-04	2.29E-04	NMHC SP Ratio:	5.09E-05
CH <sub>4</sub> Retention time:	12.00	12.00	NMHC Peak Area:	179681
				179681

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	16.95	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.10	1.001
second point	4960	40.1	8.56	8.54	1.003
third point	4980	20.0	4.27	4.28	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.02	1.006
Average Correction Factor					1.001
Baseline Corr AF:	16.95	Prev response	17.09	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0	
as found span	4920	80.2	9.14	8.98	1.018
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.08	1.007
second point	4960	40.1	4.57	4.56	1.003
third point	4980	20.0	2.28	2.29	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.02	1.013
Average Correction Factor					1.002
Baseline Corr AF:	8.98	Prev response	9.13	*% change	-1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.97	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	8.02	0.996
second point	4960	40.1	3.99	3.98	1.003
third point	4980	20.0	1.99	1.99	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.99	0.999
Average Correction Factor					1.000
Baseline Corr AF:	7.97	Prev response	7.96	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997914	0.998108
THC Cal Offset:	0.005170	0.004967
CH <sub>4</sub> Cal Slope:	0.998322	1.004174
CH <sub>4</sub> Cal Offset:	-0.008057	-0.009247
NMHC Cal Slope:	0.997658	0.992770
NMHC Cal Offset:	0.012827	0.013615

Notes: Replaced Nitrogen cylinder. No adjustments made.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 26, 2023	Previous Calibration:	August 3, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:10	End Time (MST):	14:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	$\geq 0.995$			
17.13	17.10	1.0015						
8.56	8.54	1.0027				Slope	0.998108	0.90 - 1.10
4.27	4.28	0.9974						
			Intercept	0.004967	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 26, 2023	Previous Calibration:	August 3, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:10	End Time (MST):	14:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999983	≥0.995
7.99	8.02	0.9958			
3.99	3.98	1.0033			
1.99	1.99	0.9998			
			Slope	1.004174	0.90 - 1.10
			Intercept	-0.009247	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 26, 2023	Previous Calibration:	August 3, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:10	End Time (MST):	14:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999990	$\geq 0.995$			
9.14	9.08	1.0065						
4.57	4.56	1.0028				Slope	0.992770	0.90 - 1.10
2.28	2.29	0.9953						
			Intercept	0.013615	$\pm 0.5$			



NMHC Calibration Plot

Date: September 26, 2023

Location: Conklin





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Conklin  
Calibration Date: September 21, 2023  
Start time (MST): 10:27  
Reason: Routine  
Station number: AMS21  
Last Cal Date: August 10, 2023  
End time (MST): 14:53

### Calibration Standards

NO Gas Cylinder #: T2Y1P1H  
NOX Cal Gas Conc: 51.09 ppm  
Removed Cylinder #: n/a  
Removed Gas NOX Conc: 51.09 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701H  
Cal Gas Expiry Date: December 11, 2023  
NO Cal Gas Conc: 50.39 ppm  
Removed Gas Exp Date: n/a  
Removed Gas NO Conc: 50.39 ppm  
NO gas Diff:  
Serial Number: 3810  
Serial Number: 691

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1501663731

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.120	1.104	NO bkgnd or offset:	11.2	11
NOX coeff or slope:	1.000	1.000	NOX bkgnd or offset:	11.4	11.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	167.6	167.3

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.998758	0.993898
NO <sub>x</sub> Cal Offset:	1.603828	2.303154
NO Cal Slope:	0.999738	0.994967
NO Cal Offset:	0.741061	1.360599
NO <sub>2</sub> Cal Slope:	1.000498	0.989964
NO <sub>2</sub> Cal Offset:	0.217772	-0.688924



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as found span	4921	79.4	811.2	800.1	11.1	829.1	815.4	13.7	0.9785	0.9812
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4921	79.4	811.2	800.1	11.1	807.5	796.9	10.7	1.0046	1.0041
second point	4960	39.7	405.7	400.1	5.6	406.6	399.8	6.8	0.9977	1.0008
third point	4980	19.8	202.3	199.6	2.8	205.6	201.5	4.1	0.9841	0.9903
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
as left span	4921	79.4	811.2	389.2	422.0	811.1	393.9	417.2	1.0002	0.9881
Average Correction Factor									0.9955	0.9984

Corrected As found	NO <sub>x</sub> = 829.2 ppb	NO = 815.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 2.1%
Previous Response	NO <sub>x</sub> = 811.8 ppb	NO = 800.6 ppb		*Percent Change	NO = 1.8%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI: ;

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	800.6	389.7	422.0	417.7	1.0103	99.0%
2nd GPT point (200 ppb O3)	800.6	592.1	219.6	215.8	1.0177	98.3%
3rd GPT point (100 ppb O3)	800.6	697.1	114.6	112.3	1.0206	98.0%
Average Correction Factor					1.0162	98.4%

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 21, 2023	Previous Calibration:	August 10, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:27	End Time (MST):	14:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
811.2	807.5	1.0046			
405.7	406.6	0.9977			
202.3	205.6	0.9841			
			Slope	0.993898	0.90 - 1.10
			Intercept	2.303154	+/-20







# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 21, 2023	Previous Calibration:	August 10, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:27	End Time (MST):	14:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
800.1	796.9	1.0041			
400.1	399.8	1.0008			
199.6	201.5	0.9903			
			Slope	0.994967	0.90 - 1.10
			Intercept	1.360599	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 21, 2023	Previous Calibration:	August 10, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:27	End Time (MST):	14:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
422.0	417.7	1.0103		
219.6	215.8	1.0177		
114.6	112.3	1.0206		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 21, 2023

Location: Conklin







# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 11, 2023	Previous Calibration:	August 15, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:32	End Time (MST):	13:51
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.5	----	Correlation Coefficient	0.999977	≥0.995
400.0	400.6	0.9985			
200.0	201.3	0.9935	Slope	1.001714	0.90 - 1.10
100.0	101.4	0.9862			
			Intercept	0.400000	+/- 5



O<sub>3</sub> Calibration Plot

Date: September 11, 2023

Location: Conklin





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Conklin Station number: AMS 21  
 Calibration Date: September 26, 2023 Last Cal Date: August 18, 2023  
 Start time (MST): 11:20 End time (MST): 12:35

Analyzer Make: API T640 S/N: 326  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat S/N: 388744  
 Temp/RH standard: Alicat S/N: 388744

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	22.5	22.58	22.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	703.3	705.02	703.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	5.127	5.02	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: September 26, 2023 Last Cal Date: August 18, 2023  
 PM w/o HEPA: 14.7 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.8	11.8	10.8	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>13.1</u>	w/ HEPA: <u>0</u>		<0.2 ug/m3
Date Optical Chamber Cleaned:		<u>September 26, 2023</u>			
Disposable Filter Changed:		<u>September 26, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_  
 Date RH/T Sensor Cleaned: \_\_\_\_\_

Notes: Adjusted PMT peak, both leak checks passed.

Calibration by: Braiden Boutilier



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS22  
JANVIER  
SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	September 28, 2023	Last Cal Date:	August 23, 2023
Start time (MST):	9:38	End time (MST):	12:50
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.11	ppm	Cal Gas Exp Date:	January 18, 2029
Cal Gas Cylinder #:	CC281519			
Removed Cal Gas Conc:	50.11	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

### Analyzer Information

Analyzer make: Thermo 43i	Analyzer serial #: 1152430006
Analyzer Range 0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005224	1.005095	Backgd or Offset:	21.4	21.4
Calibration intercept:	1.843010	2.463336	Coeff or Slope:	0.989	0.989

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	1.2	----
as found span	4920	79.8	799.8	804.0	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	1.7	----
high point	4920	79.8	799.8	805.6	0.993
second point	4960	39.9	399.9	405.8	0.985
third point	4980	20.0	200.4	204.0	0.983
as left zero	5000	0.0	0.0	1.2	----
as left span	4920	79.8	799.8	804.6	0.994
<b>Average Correction Factor</b>					<b>0.987</b>

Baseline Corr As found:	802.80	Previous response	805.81	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 28, 2023	Previous Calibration:	August 23, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:38	End Time (MST):	12:50
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	1.7	----	Correlation Coefficient	0.999991	
799.8	805.6	0.9928			≥0.995
399.9	405.8	0.9854	Slope	1.005095	
200.4	204.0	0.9825			0.90 - 1.10
			Intercept	2.463336	+/-30



SO2 Calibration Plot

Date: September 28, 2023

Location: Janvier





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Janvier Station number: AMS22  
 Calibration Date: September 21, 2023 Last Cal Date: August 29, 2023  
 Start time (MST): 8:58 End time (MST): 10:25  
 Reason: Removal

### Calibration Standards

Cal Gas Concentration: 5.03 ppm Cal Gas Exp Date: April 16, 2022  
 Cal Gas Cylinder #: DT0018680  
 Removed Cal Gas Conc: 5.03 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806  
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

### Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 1200326169  
 Converter make: CDN-101 Converter serial #: 587  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004358		Backgd or Offset:	1.01	
Calibration intercept:	-0.338889		Coeff or Slope:	0.926	

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4920	79.5	80.0	74.1	1.081
as found 2nd point	4960	39.8	40.0	35.4	1.134
as found 3rd point	4980	19.9	20.0	17.2	1.171
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

### SO2 Scrubber Check

Date of last scrubber change:	Ave Corr Factor
Date of last converter efficiency test:	efficiency

Baseline Corr As found: 74.0 Prev response: 79.99 \*% change: -8.1%  
 Baseline Corr 2nd AF pt: 35.3 AF Slope: 0.928759 AF Intercept: -0.817046  
 Baseline Corr 3rd AF pt: 17.1 AF Correlation: 0.999168

\* = > +/-5% change initiates investigation


Notes: Completed multipoint as founds to replace the instrument due to linearity and drifting issues.

Calibration Performed By: Max Farrell

TRS Calibration Plot

Date: September 21, 2023

Location: Janvier





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name:	Janvier	Station number:	AMS22
Calibration Date:	September 21, 2023	Last Cal Date:	N/A
Start time (MST):	10:33	End time (MST):	13:31
Reason:	Install		

### Calibration Standards

Cal Gas Concentration:	5.03	ppm	Cal Gas Exp Date:	April 16, 2022
Cal Gas Cylinder #:	DT0018680			
Removed Cal Gas Conc:	5.03	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031
Converter make:	CDN-101	Converter serial #:	587
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	N/A	1.003650	Backgd or Offset:	N/A	3.04
Calibration intercept:	N/A	-0.039106	Coeff or Slope:	N/A	1.161

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	79.5	80.0	80.2	0.997
second point	4960	39.8	40.0	40.2	0.996
third point	4980	19.9	20.0	20.1	0.996
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	79.5	80.0	81.7	0.979
SO2 Scrubber Check	4920	79.8	798.0	0.2	----
Date of last scrubber change:				Ave Corr Factor	0.996
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

\* = > +/-5% change initiates investigation

Notes: Install calibrations after replacement of the analyzer due to linearity and drift issues. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 21, 2023	Previous Calibration:	N/A
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	10:33	End Time (MST):	13:31
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999997	≥0.995
80.0	80.2	0.9973			
40.0	40.2	0.9960	Slope	1.003650	0.90 - 1.10
20.0	20.1	0.9960			
			Intercept	-0.039106	+/-3



TRS Calibration Plot

Date: September 21, 2023

Location: Janvier







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	September 28, 2023	Last Cal Date:	August 23, 2023
Start time (MST):	9:38	End time (MST):	12:50
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH <sub>4</sub> Cal Gas Conc.	502.8 ppm	CH <sub>4</sub> Equiv Conc.	1075.9 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	502.8 ppm	CH <sub>4</sub> Equiv Conc.	1075.9 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.190E-04	2.240E-04	NMHC SP Ratio:	4.39E-05
CH <sub>4</sub> Retention time:	13.6	13.6	NMHC Peak Area:	208229
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	17.20	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.18	1.000
second point	4960	39.9	8.59	8.58	1.000
third point	4980	20.0	4.30	4.29	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.49	0.982
Average Correction Factor					1.001

Baseline Corr AF:	17.20	Prev response	17.12	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	9.15	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.15	9.15	1.000
second point	4960	39.9	4.57	4.57	1.002
third point	4980	20.0	2.29	2.28	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.15	9.28	0.986
Average Correction Factor					1.002
Baseline Corr AF:	9.15	Prev response	9.12	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	8.05	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.03	0.999
second point	4960	39.9	4.01	4.02	0.999
third point	4980	20.0	2.01	2.01	0.999
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.21	0.978
Average Correction Factor					0.999
Baseline Corr AF:	8.05	Prev response	8.00	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997041	1.000655
THC Cal Offset:	-0.005391	-0.005800
CH <sub>4</sub> Cal Slope:	0.996940	1.000614
CH <sub>4</sub> Cal Offset:	-0.003162	0.000436
NMHC Cal Slope:	0.997018	1.000704
NMHC Cal Offset:	-0.002029	-0.007036

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 28, 2023	Previous Calibration:	August 23, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:38	End Time (MST):	12:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	$\geq 0.995$			
17.17	17.18	0.9995						
8.59	8.58	1.0002				Slope	1.000655	0.90 - 1.10
4.30	4.29	1.0022						
			Intercept	-0.005800	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 28, 2023	Previous Calibration:	August 23, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:38	End Time (MST):	12:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	$\geq 0.995$			
8.03	8.03	0.9994						
4.01	4.02	0.9991				Slope	1.000614	0.90 - 1.10
2.01	2.01	0.9991						
			Intercept	0.000436	$\pm 0.5$			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	September 28, 2023	Previous Calibration:	August 23, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:38	End Time (MST):	12:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	$\geq 0.995$			
9.15	9.15	0.9997						
4.57	4.57	1.0016				Slope	1.000704	0.90 - 1.10
2.29	2.28	1.0054						
			Intercept	-0.007036	$\pm 0.5$			



NMHC Calibration Plot

Date: September 28, 2023

Location: Janvier





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Janvier Station number: AMS 22  
Calibration Date: September 19, 2023 Last Cal Date: August 16, 2023  
Start time (MST): 9:38 End time (MST): 14:32  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: CC424183 Cal Gas Expiry Date: April 16, 2023  
NOX Cal Gas Conc: 48.60 ppm NO Cal Gas Conc: 48.60 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 48.60 ppm Removed Gas NO Conc: 48.60 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 3806  
ZAG make/model: Teledyne API T701 Serial Number: 201

### Analyzer Information

Analyzer make: API T200 Analyzer serial #: 833  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.808	0.826	NO bkgnd or offset:	-5.6	-5.6
NOX coeff or slope:	0.799	0.815	NOX bkgnd or offset:	-3.9	-3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	5.0	5.0

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.995773	0.998659
NO <sub>x</sub> Cal Offset:	2.689123	2.728523
NO Cal Slope:	0.994314	0.998929
NO Cal Offset:	1.949881	2.308958
NO <sub>2</sub> Cal Slope:	1.001315	0.997414
NO <sub>2</sub> Cal Offset:	-0.151118	-0.454360



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.5	2.1	-0.5	----	----
as found span	4918	82.3	799.9	799.9	0.0	785.0	779.1	6.0	1.0190	1.0267
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	1.9	2.3	-0.4	----	----
high point	4918	82.3	799.9	799.9	0.0	801.0	801.3	-0.3	0.9986	0.9983
second point	4959	41.2	400.4	400.4	0.0	403.5	402.5	1.0	0.9924	0.9949
third point	4980	20.6	200.2	200.2	0.0	203.2	202.2	1.0	0.9853	0.9901
as left zero	5000	0.0	0.0	0.0	0.0	1.9	2.3	-0.4	----	----
as left span	4918	82.3	799.9	409.7	390.2	797.0	404.9	392.1	1.0036	1.0119
Average Correction Factor									0.9921	0.9944

Corrected As found	NO <sub>x</sub> = 783.5 ppb	NO = 777.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -2.0%
Previous Response	NO <sub>x</sub> = 799.2 ppb	NO = 797.3 ppb		*Percent Change	NO = -2.6%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.3	409.1	390.2	388.8	1.0036	99.6%
2nd GPT point (200 ppb O3)	799.3	599.9	199.4	198.3	1.0055	99.4%
3rd GPT point (100 ppb O3)	799.3	699.5	99.8	99.1	1.0071	99.3%
Average Correction Factor					1.0054	99.5%

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 19, 2023	Previous Calibration:	August 16, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:38	End Time (MST):	14:32
Analyzer make:	API T200	Analyzer serial #:	833

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	1.9	----	Correlation Coefficient	≥0.995	
799.9	801.0	0.9986			
400.4	403.5	0.9924			
200.2	203.2	0.9853			
			Slope	0.998659	0.90 - 1.10
			Intercept	2.728523	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 19, 2023	Previous Calibration:	August 16, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:38	End Time (MST):	14:32
Analyzer make:	API T200	Analyzer serial #:	833

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	2.3	----	Correlation Coefficient	1.000000	
799.9	801.3	0.9983			≥0.995
400.4	402.5	0.9949	Slope	0.998929	
200.2	202.2	0.9901			0.90 - 1.10
			Intercept	2.308958	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 19, 2023	Previous Calibration:	August 16, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:38	End Time (MST):	14:32
Analyzer make:	API T200	Analyzer serial #:	833

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.4	----	Correlation Coefficient	≥0.995	
390.2	388.8	1.0036			
199.4	198.3	1.0055			
99.8	99.1	1.0071			
			Slope	0.997414	0.90 - 1.10
			Intercept	-0.454360	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 19, 2023

Location: Janvier





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Janvier Station number: AMS 22  
 Calibration Date: September 7, 2023 Last Cal Date: August 18, 2023  
 Start time (MST): 9:21 End time (MST): 13:45  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806  
 ZAG Make/Model: Teledyne API T701H Serial Number: 201

### Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 7046  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001114	1.000914	Backgd or Offset:	-0.1	-0.2
Calibration intercept:	2.780000	0.840000	Coeff or Slope:	0.922	1.021

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.7	----
as found span	4895	905.3	400.0	359.6	<b>1.112</b>
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.4	----
high point	4895	905.3	400.0	400.9	0.998
second point	4895	756.7	200.0	201.5	0.993
third point	4895	656.1	100.0	101.2	0.988
as left zero	5000	800.0	0.0	0.9	----
as left span	4895	904.3	400.0	401.6	0.996
Average Correction Factor					0.993

Baseline Corr As found:	358.9	Previous response	403.2	*% change	<b>-12.4%</b>
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Completed as found span before the zero by mistake. Changed the inlet filter after as founds. As founds are 11% low due to the O3 pump that was changed recently on the calibrator. Adjusted the span only. See docit note for more info.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 18, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:21	End Time (MST):	13:45
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999993	
400.0	400.9	0.9978			≥0.995
200.0	201.5	0.9926	Slope	1.000914	
100.0	101.2	0.9881			0.90 - 1.10
			Intercept	0.840000	+/- 5



O<sub>3</sub> Calibration Plot

Date: September 7, 2023

Location: Janvier





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Janvier Station number: AMS 22  
 Calibration Date: September 28, 2023 Last Cal Date: August 23, 2023  
 Start time (MST): 12:36 End time (MST): 12:54

Analyzer Make: Teledyne API T640 S/N: 325  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450  
 Temp/RH standard: Delta Cal S/N: 1450

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	18.2	17.5	18.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	712.2	712.8	712.2	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	5.07	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 28, 2023</u>		Last Cal Date: <u>August 23, 2023</u>		
	PM w/o HEPA: <u>7.4</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>July 26, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 26, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: July 26, 2023  
 Date RH/T Sensor Cleaned: July 26, 2023

### Notes:

No adjustments needed. Leak check passed.

Calibration by: Max Farrell





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS23 FORT HILLS SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 10, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:18	End Time (MST):	10:49
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999994	≥0.995
799.1	800.6	0.9981	Slope	1.002497	0.90 - 1.10
400.1	400.4	0.9991	Intercept	-0.744078	+/-30
200.0	198.6	1.0072			



SO2 Calibration Plot

Date: September 7, 2023

Location: Fort Hills





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Fort Hills Station number: AMS23  
 Calibration Date: September 6, 2023 Last Cal Date: August 17, 2023  
 Start time (MST): 6:55 End time (MST): 11:02  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.20 ppm Cal Gas Exp Date: February 5, 2024  
 Cal Gas Cylinder #: CC517372  
 Removed Cal Gas Conc: 5.20 ppm Rem Gas Exp Date: N/A  
 Removed Gas Cyl #: N/A Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 451  
 ZAG Make/Model: API T701 Serial Number: 5611

### Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1300156232  
 Converter make: CDN-101 Converter serial #: 594  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011740	1.001459	Backgd or Offset:	1.19
Calibration intercept:	0.002125	0.061824	Coeff or Slope:	1.124

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4923	77.0	80.0	78.3	1.020
as found 2nd point	4962	38.5	40.0	39.4	1.013
as found 3rd point	4981	19.2	19.9	19.5	1.018
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.0	80.0	80.0	1.000
second point	4962	38.5	40.0	40.6	0.985
third point	4981	19.2	19.9	19.8	1.007
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.0	80.0	81.0	0.988
SO2 Scrubber Check	4920	80.3	803.0	-0.1	----

Date of last scrubber change:	Ave Corr Factor	0.998
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 78.4 Prev response: 80.94 \*% change: -3.2%  
 Baseline Corr 2nd AF pt: 39.5 AF Slope: 0.980176 AF Intercept: -0.018544  
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999981

\* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021


### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 17, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	6:55	End Time (MST):	11:02
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	$\geq 0.995$	
80.0	80.0	1.0000			
40.0	40.6	0.9852			
19.9	19.8	1.0075			
			Slope	1.001459	0.90 - 1.10
			Intercept	0.061824	+/-3







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	September 7, 2023	Last Cal Date:	August 10, 2023
Start time (MST):	8:18	End time (MST):	10:48
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	500.2 ppm	CH4 Equiv Conc.	1070.6 ppm
C3H8 Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH4 Conc.	500.2 ppm	CH4 Equiv Conc.	1070.6 ppm
Removed C3H8 Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.39E-04	2.34E-04	NMHC SP Ratio:	5.21E-05
CH4 Retention time:	13.0	13.0	NMHC Peak Area:	176470
				180460

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	17.58	0.978
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.28	0.995
second point	4960	40.2	8.61	8.62	0.998
third point	4980	20.1	4.30	4.30	1.001
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.19	17.33	0.992

Average Correction Factor				0.998
Baseline Corr AF:	17.58	Prev response	17.28	*% change 1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* => +/-5% change initiates investigation





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.42	0.972
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.20	0.996
second point	4960	40.2	4.59	4.64	0.988
third point	4980	20.1	2.29	2.33	0.984
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	9.16	9.22	0.993
Average Correction Factor					0.989
Baseline Corr AF:	9.42	Prev response	9.22	*% change	2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	8.16	0.984
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.07	0.995
second point	4960	40.2	4.02	3.98	1.010
third point	4980	20.1	2.01	1.97	1.021
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.03	8.12	0.989
Average Correction Factor					1.009
Baseline Corr AF:	8.16	Prev response	8.05	*% change	1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.005681	1.005421
THC Cal Offset:	-0.013606	-0.016398
CH <sub>4</sub> Cal Slope:	1.005023	1.005916
CH <sub>4</sub> Cal Offset:	-0.021648	-0.032039
NMHC Cal Slope:	1.006196	1.003864
NMHC Cal Offset:	0.008043	0.017642

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 10, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:18	End Time (MST):	10:48
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	$\geq 0.995$			
17.19	17.28	0.9949						
8.61	8.62	0.9985				Slope	1.005421	0.90 - 1.10
4.30	4.30	1.0008						
			Intercept	-0.016398	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary

Version-01-2020


### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 10, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:18	End Time (MST):	10:48
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999915	≥0.995
8.03	8.07	0.9954			
4.02	3.98	1.0104	Slope	1.005916	0.90 - 1.10
2.01	1.97	1.0207			
			Intercept	-0.032039	+/-0.5

### CH<sub>4</sub> Calibration Curve





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 10, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:18	End Time (MST):	10:48
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999980	$\geq 0.995$
9.16	9.20	0.9956			
4.59	4.64	0.9882	Slope	1.003864	0.90 - 1.10
2.29	2.33	0.9840			
			Intercept	0.017642	$\pm 0.5$



NMHC Calibration Plot

Date: September 7, 2023

Location: Fort Hills





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	September 27, 2023	Last Cal Date:	September 7, 2023
Start time (MST):	11:00	End time (MST):	12:12
Reason:	Cylinder Change		

### Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1070.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH <sub>4</sub> Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1070.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.39E-04	2.34E-04	NMHC SP Ratio:	5.21E-05
CH <sub>4</sub> Retention time:	13.0	13.0	NMHC Peak Area:	176470
				180460

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	17.33	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.33	0.992
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	0.992
Baseline Corr AF:	17.33	Prev response	17.27	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.18	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.18	0.998
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.998
Baseline Corr AF:	9.18	Prev response	9.21	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	8.14	0.987
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.15	0.986
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.986
Baseline Corr AF:	8.14	Prev response	8.05	*% change	1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.005421	1.008027
THC Cal Offset:	-0.016398	0.000000
CH <sub>4</sub> Cal Slope:	1.005916	1.014599
CH <sub>4</sub> Cal Offset:	-0.032039	0.000000
NMHC Cal Slope:	1.003864	1.002263
NMHC Cal Offset:	0.017642	0.000000

Notes:


Nitrogen cylinder changed.

Calibration Performed By: Melissa Lemay

NMHC Calibration Plot

Date: September 27, 2023

Location: Fort Hills







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Fort Hills  
Calibration Date: September 5, 2023  
Start time (MST): 7:35  
Reason: Routine  
Station number: AMS23  
Last Cal Date: August 1, 2023  
End time (MST): 11:56

### Calibration Standards

NO Gas Cylinder #: CC332703  
NOX Cal Gas Conc: 49.7 ppm  
Removed Cylinder #: N/A  
Removed Gas NOX Conc: 49.7 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701  
Cal Gas Expiry Date: January 28, 2024  
NO Cal Gas Conc: 49.7 ppm  
Removed Gas Exp Date: N/A  
Removed Gas NO Conc: 49.7 ppm  
NO gas Diff:  
Serial Number: 451  
Serial Number: 5611

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1152430007

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.107	1.076	NO bkgnd or offset:	3.4	3.3
NOX coeff or slope:	0.995	0.991	NOX bkgnd or offset:	3.8	3.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	163.8	163.8

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.999964	0.996239
NO <sub>x</sub> Cal Offset:	0.504456	-0.017022
NO Cal Slope:	1.001151	0.998440
NO Cal Offset:	-0.335801	-1.437143
NO <sub>2</sub> Cal Slope:	0.997897	1.016048
NO <sub>2</sub> Cal Offset:	-0.698579	0.951555



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
as found span	4920	80.5	800.2	800.2	0.0	826.1	821.4	4.7	0.969	0.974
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	-0.2	0.2	----	----
high point	4920	80.5	800.2	800.2	0.0	797.0	797.9	-0.9	1.004	1.003
second point	4960	40.2	399.6	399.6	0.0	398.5	397.5	1.1	1.003	1.005
third point	4980	20.1	199.8	199.8	0.0	198.7	196.4	2.3	1.005	1.017
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as left span	4920	80.5	800.2	455.6	344.6	784.3	436.5	347.8	1.020	1.044
Average Correction Factor									1.004	1.008

Corrected As found	NO <sub>x</sub> = 826.3 ppb	NO = 821.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 3.1%
Previous Response	NO <sub>x</sub> = 800.6 ppb	NO = 800.8 ppb		*Percent Change	NO = 2.5%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	783.7	439.1	344.6	350.5	0.983	101.7%
2nd GPT point (200 ppb O3)	783.7	615.3	168.4	173.1	0.973	102.8%
3rd GPT point (100 ppb O3)	783.7	698.9	84.8	87.4	0.970	103.1%
Average Correction Factor					0.975	102.5%

Notes: No maintenance done. Span adjusted. Due to drifting during the GPT the 2nd NO ref point used.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 1, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:35	End Time (MST):	11:56
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.0	0.0	----	Correlation Coefficient	0.999999	≥0.995			
800.2	797.0	1.0040						
399.6	398.5	1.0027				Slope	0.996239	0.90 - 1.10
199.8	198.7	1.0055						
			Intercept	-0.017022	+/-20			





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 1, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:35	End Time (MST):	11:56
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.2	797.9	1.0028		
399.6	397.5	1.0052		
199.8	196.4	1.0173		

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 1, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:35	End Time (MST):	11:56
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999969	≥0.995
344.6	350.5	0.9832			
168.4	173.1	0.9728			
84.8	87.4	0.9703			
			Slope	1.016048	0.90 - 1.10
			Intercept	0.951555	+/-20



NO<sub>x</sub> Calibration Plot

Date: September 5, 2023

Location: Fort Hills





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Fort Hills Station number: AMS 23  
 Calibration Date: September 7, 2023 Last Cal Date: August 17, 2023  
 Start time (MST): 6:40 End time (MST): 7:40

Analyzer Make: API T640 S/N: 1546  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753  
 Temp/RH standard: Alicat FP-25BT S/N: 388753

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	6.9	6.8	6.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	732.9	732.1	732.9	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.1	5.00	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: September 7, 2023 Last Cal Date: August 17, 2023  
 PM w/o HEPA: 66.1 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9	12	10.7	<input checked="" type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: 65.3 w/ HEPA: 0.0  
 Date Optical Chamber Cleaned: September 7, 2023 <0.2 ug/m3  
 Disposable Filter Changed: September 7, 2023

### Annual Maintenance

Date Sample Tube Cleaned: May 10, 2023  
 Date RH/T Sensor Cleaned: May 10, 2023

Leak check passed before and after cleaning. PMT adjusted. Head cleaned.

Notes:

Calibration by: Melissa Lemay



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS25 WASKŌW OHCI PIMÂTISIWIN SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number:	AMS25
Calibration Date:	September 28, 2023	Last Cal Date:	August 29, 2023
Start time (MST):	7:25	End time (MST):	10:28
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.54	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC437219			
Removed Cal Gas Conc:	50.54	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	747
ZAG Make/Model:	API T701		Serial Number:	4765

### Analyzer Information

Analyzer make: Thermo 43i  
 Analyzer Range 0 - 1000 ppb  
 Analyzer serial #: 1118148497

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997792	0.990940	Backgd or Offset:	9.7	9.7
Calibration intercept:	0.703977	0.604127	Coeff or Slope:	0.988	0.988

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4921	79.2	800.5	790.9	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.5	----
high point	4921	79.2	800.5	793.8	1.008
second point	4960	39.6	400.3	397.4	1.007
third point	4980	19.8	200.1	199.0	1.006
as left zero	5000	0.0	0.0	0.5	----
as left span	4921	79.2	800.5	794.0	1.008
Average Correction Factor					1.007

Baseline Corr As found:	790.60	Previous response	799.46	*% change	-1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 28, 2023	Previous Calibration:	August 29, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	7:25	End Time (MST):	10:28
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	1.000000	≥0.995
800.5	793.8	1.0085			
400.3	397.4	1.0073	Slope	0.990940	0.90 - 1.10
200.1	199.0	1.0058			
			Intercept	0.604127	+/-30



SO2 Calibration Plot

Date: September 28, 2023

Location: Waskow ohci Pimatisiwin





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Waskow ohci Pimatisiwin      Station number: AMS25  
 Calibration Date: September 26, 2023      Last Cal Date: August 18, 2023  
 Start time (MST): 6:45      End time (MST): 11:06  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 4.97 ppm      Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC517099  
 Removed Cal Gas Conc: 4.97 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 747  
 ZAG Make/Model: API T701      Serial Number: 261

### Analyzer Information

Analyzer make: Thermo 43i-LTE      Analyzer serial #: 1170050146  
 Converter make: Global G-150      Converter serial #: 2022-219  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.991600	0.980956	Backgd or Offset: 3.25	3.25
Calibration intercept:	0.080000	0.100000	Coeff or Slope: 1.079	1.079

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4920	80.0	79.5	78.6	1.010
as found 2nd point	4960	40.0	39.7	39.6	1.001
as found 3rd point	4980	20.0	19.9	19.7	1.003
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.0	79.5	78.0	1.019
second point	4960	40.0	39.7	39.1	1.016
third point	4980	20.0	19.9	19.7	1.008
as left zero	5000	0.0	0.0	0.0	----
as left span	4920	80.0	800.0	786.8	1.017
SO2 Scrubber Check	4921	79.2	800.0	-0.1	----
Date of last scrubber change:		20-Jun-23		Ave Corr Factor	1.014
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 78.7      Prev response: 78.87      \*% change: -0.2%  
 Baseline Corr 2nd AF pt: 39.7      AF Slope: 0.990449      AF Intercept: 0.020000  
 Baseline Corr 3rd AF pt: 19.8      AF Correlation: 0.999976

\* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 26, 2023	Previous Calibration:	August 18, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	6:45	End Time (MST):	11:06
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999992	≥0.995
79.5	78.0	1.0187			
39.7	39.1	1.0161	Slope	0.980956	0.90 - 1.10
19.9	19.7	1.0083			
			Intercept	0.100000	+/-3



H<sub>2</sub>S Calibration Plot

Date: September 26, 2023

Location: Waskow ohci Pimatisiwin





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS26 CHRISTINA LAKE SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Christina Lake	Station number:	AMS 26
Calibration Date:	September 19, 2023	Last Cal Date:	August 17, 2023
Start time (MST):	14:16	End time (MST):	16:57
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.56	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC362134</u>			
Removed Cal Gas Conc:	49.56	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2447
ZAG Make/Model:	API T701H		Serial Number:	832

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.018405	1.009149	Backgd or Offset:	16.2	16.1
Calibration intercept:	-2.239175	-2.177877	Coeff or Slope:	0.900	0.900

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4919	80.6	799.0	804.4	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.8	----
high point	4919	80.6	799.0	805.7	0.992
second point	4960	40.3	399.4	399.1	1.001
third point	4980	20.2	200.2	197.1	1.016
as left zero	5000	0.0	0.0	0.7	----
as left span	4919	80.6	799.0	807.6	0.989
Average Correction Factor					1.003

Baseline Corr As found:	803.90	Previous response	811.44	*% change	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 19, 2023	Previous Calibration:	August 17, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	14:16	End Time (MST):	16:57
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.8	----	Correlation Coefficient	0.999937	
799.0	805.7	0.9916			≥0.995
399.4	399.1	1.0008	Slope	1.009149	
200.2	197.1	1.0158			0.90 - 1.10
			Intercept	-2.177877	+/-30



SO2 Calibration Plot

Date: September 19, 2023

Location: Christina Lake





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Christina Lake	Station number:	AMS26
Calibration Date:	September 20, 2023	Last Cal Date:	August 17, 2023
Start time (MST):	6:50	End time (MST):	10:51
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	4.89	ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	<u>EY0002466</u>			
Removed Cal Gas Conc:	4.89	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2447
ZAG Make/Model:	API T701H		Serial Number:	832

### Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032
Converter make:	NA	Converter serial #:	NA
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002610	1.010898	Backgd or Offset:	35.4
Calibration intercept:	0.119119	0.179125	Coeff or Slope:	1.093

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4918	81.8	80.0	79.2	1.014
as found 2nd point	4959	40.9	40.0	39.8	1.013
as found 3rd point	4979	20.4	20.0	19.5	1.039
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4918	81.8	80.0	81.1	0.986
second point	4959	40.9	40.0	40.7	0.983
third point	4979	20.4	20.0	20.0	0.998
as left zero	5000	0.0	0.0	0.3	----
as left span	4918	81.8	80.0	80.6	0.993
SO2 Scrubber Check	4919	80.6	806.1	0.0	----
Date of last scrubber change:	27-Feb-19			Ave Corr Factor	0.989
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	78.9	Prev response:	80.33	*% change:	-1.8%
Baseline Corr 2nd AF pt:	39.5	AF Slope:	0.988332	AF Intercept:	0.118700
Baseline Corr 3rd AF pt:	19.2	AF Correlation:	0.999950		

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after MAF's. Ran scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 20, 2023	Previous Calibration:	August 17, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	6:50	End Time (MST):	10:51
Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999950	≥0.995
80.0	81.1	0.9865			
40.0	40.7	0.9828	Slope	1.010898	0.90 - 1.10
20.0	20.0	0.9977			
			Intercept	0.179125	+/-3



H<sub>2</sub>S Calibration Plot

Date: September 20, 2023

Location: Christina Lake





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Christina Lake  
Calibration Date: September 19, 2023  
Start time (MST): 10:35  
Reason: Routine

Station number: AMS26  
Last Cal Date: August 17, 2023  
End time (MST): 17:46

### Calibration Standards

NO Gas Cylinder #: T2Y1P4C  
NOX Cal Gas Conc: 50.82 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 50.82 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701H

Cal Gas Expiry Date: November 12, 2023  
NO Cal Gas Conc: 50.02 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 50.02 ppm  
NO gas Diff:  
Serial Number: 2447  
Serial Number: 832

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1173480006

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.370	1.370	NO bkgnd or offset:	2.6	2.6
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	3.4	3.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.7	161.0

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.996528	1.000604
NO <sub>x</sub> Cal Offset:	-1.780000	-2.180000
NO Cal Slope:	1.001328	1.004213
NO Cal Offset:	-2.380000	-2.740000
NO <sub>2</sub> Cal Slope:	0.995894	0.995671
NO <sub>2</sub> Cal Offset:	-1.805192	-0.892478



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.2	-0.6	----	----
as found span	4920	80.0	813.1	800.3	12.8	812.6	801.6	11.1	1.0006	0.9984
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.5	0.0	-0.5	----	----
high point	4920	80.0	813.1	800.3	12.8	812.2	802.2	9.8	1.0011	0.9977
second point	4960	40.0	406.6	400.2	6.4	404.0	398.1	5.9	1.0063	1.0052
third point	4980	20.0	203.3	200.1	3.2	199.4	195.2	4.2	1.0195	1.0250
as left zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.1	-0.6	----	----
as left span	4920	80.0	813.1	394.8	418.3	811.7	395.4	416.2	1.0017	0.9985
Average Correction Factor									1.0090	1.0093

Corrected As found	NO <sub>x</sub> = 813.3 ppb	NO = 801.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.6%
Previous Response	NO <sub>x</sub> = 808.5 ppb	NO = 799.0 ppb		*Percent Change	NO = 0.3%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.2	393.7	418.3	415.9	1.0058	99.4%
2nd GPT point (200 ppb O3)	799.2	595.5	216.5	214.2	1.0107	98.9%
3rd GPT point (100 ppb O3)	799.2	700.7	111.3	109.7	1.0146	98.6%
Average Correction Factor					1.0104	99.0%

Notes: Changed sample inlet filter after as founds. No adjustments required.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 19, 2023	Previous Calibration:	August 17, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:35	End Time (MST):	17:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.5	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.1	812.2	1.0011		
406.6	404.0	1.0063		
203.3	199.4	1.0195		
			0.999980	
			1.000604	
			-2.180000	







# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 19, 2023	Previous Calibration:	August 17, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:35	End Time (MST):	17:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ $0.90 - 1.10$ $\pm 20$
800.3	802.2	0.9977		
400.2	398.1	1.0052		
200.1	195.2	1.0250		

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 19, 2023	Previous Calibration:	August 17, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:35	End Time (MST):	17:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.5	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
418.3	415.9	1.0058		
216.5	214.2	1.0107		
111.3	109.7	1.0146		



NO<sub>x</sub> Calibration Plot

Date: September 19, 2023

Location: Christina Lake





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS27**  
**JACKFISH 2/3**  
**SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	September 28, 2023	Last Cal Date:	August 23, 2023
Start time (MST):	7:58	End time (MST):	10:48
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	<u>SG9133974BAL</u>			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3811
ZAG Make/Model:	API 701		Serial Number:	135

### Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	12124313138
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000975	0.990232	Backgd or Offset:	7.2	7.2
Calibration intercept:	-1.618030	-2.057655	Coeff or Slope:	0.891	0.891

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	79.1	800.2	789.4	1.014
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	79.1	800.2	791.0	1.012
second point	4961	39.5	399.5	393.6	1.015
third point	4980	19.8	200.3	193.4	1.036
as left zero	5000	0.0	0.0	0.3	----
as left span	4921	79.1	800.2	792.8	1.009
Average Correction Factor					1.021

Baseline Corr As found:	789.40	Previous response	799.32	*% change	-1.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. No adjustments required.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 28, 2023	Previous Calibration:	August 23, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	7:58	End Time (MST):	10:48
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12124313138

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	0.999961	
800.2	791.0	1.0116			≥0.995
399.5	393.6	1.0151	Slope	0.990232	
200.3	193.4	1.0357			0.90 - 1.10
			Intercept	-2.057655	+/-30



SO2 Calibration Plot

Date: September 28, 2023

Location: Jackfish 2/3





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Jackfish 2/3      Station number: AMS27  
 Calibration Date: September 27, 2023      Last Cal Date: August 2, 2023  
 Start time (MST): 12:02      End time (MST): 16:13  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.41 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC345023  
 Removed Cal Gas Conc: 5.41 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3811  
 ZAG Make/Model: API 701      Serial Number: 135

### Analyzer Information

Analyzer make: API T101      Analyzer serial #: 621  
 Converter make:      Converter serial #:  
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006901	1.002337	Backgd or Offset: 28.3	29.9
Calibration intercept:	-0.097803	-0.137809	Coeff or Slope: 0.944	0.928

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.7	----
as found span	4926	74.1	80.2	83.8	0.965
as found 2nd point	4963	37.0	40.0	41.9	0.972
as found 3rd point	4982	18.5	20.0	20.7	1.001
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4926	74.1	80.2	80.2	1.000
second point	4963	37.0	40.0	40.2	0.996
third point	4982	18.5	20.0	19.6	1.021
as left zero	5000	0.0	0.0	0.3	----
as left span	4926	74.1	80.2	79.6	1.007
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.006
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 83.1      Prev response: 80.63      \*% change: 3.0%  
 Baseline Corr 2nd AF pt: 41.2      AF Slope: 1.039259      AF Intercept: 0.342836  
 Baseline Corr 3rd AF pt: 20.0      AF Correlation: 0.999909

\* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. Adjusted both zero and span.

Calibration Performed By: Mohammed Kashif





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 27, 2023	Previous Calibration:	August 2, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	12:02	End Time (MST):	16:13
Analyzer make:	API T101	Analyzer serial #:	621

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999952
80.2	80.2	0.9997		
40.0	40.2	0.9959	Slope	1.002337
20.0	19.6	1.0212		
			Intercept	-0.137809
				+/-3



H<sub>2</sub>S Calibration Plot

Date: September 27, 2023

Location: Jackfish 2/3







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.9	0.7	----	----
as found span	4921	79.4	816.8	800.3	16.5	796.1	786.3	9.8	1.0260	1.0178
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4921	79.4	816.8	800.3	16.5	814.2	801.9	12.3	1.0032	0.9980
second point	4960	39.7	408.5	400.2	8.3	404.6	396.7	7.8	1.0095	1.0088
third point	4980	19.8	203.7	199.6	4.1	199.2	194.9	4.3	1.0226	1.0241
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.6	-0.6	----	----
as left span	4921	79.4	816.8	420.5	399.2	812.4	416.0	396.3	1.0054	1.0108
Average Correction Factor									1.0118	1.0103

Corrected As found	NO <sub>x</sub> = 796.3 ppb	NO = 787.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -2.5%	
Previous Response	NO <sub>x</sub> = 815.9 ppb	NO = 807.3 ppb		*Percent Change	NO = -2.5%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	798.4	415.7	399.2	396.1	1.0079	99.2%
2nd GPT point (200 ppb O <sub>3</sub> )	798.4	618.9	196.0	195.4	1.0031	99.7%
3rd GPT point (100 ppb O <sub>3</sub> )	798.4	713.0	101.9	98.1	1.0389	96.3%
Average Correction Factor					1.0166	98.4%

Notes:

Changed the sample inlet filter after as founds. Adjusted span only.

Calibration Performed By:

Mohammed Kashif



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 28, 2023	Previous Calibration:	August 8, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	10:49	End Time (MST):	15:19
Analyzer make:	API T200	Analyzer serial #:	722

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
816.8	814.2	1.0032		
408.5	404.6	1.0095		
203.7	199.2	1.0226		





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 28, 2023	Previous Calibration:	August 8, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	10:49	End Time (MST):	15:19
Analyzer make:	API T200	Analyzer serial #:	722

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	801.9	0.9980		
400.2	396.7	1.0088		
199.6	194.9	1.0241		

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 28, 2023	Previous Calibration:	August 8, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	10:49	End Time (MST):	15:19
Analyzer make:	API T200	Analyzer serial #:	722

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
399.2	396.1	1.0079		
196.0	195.4	1.0031		
101.9	98.1	1.0389		



NO<sub>x</sub> Calibration Plot

Date: September 28, 2023

Location: Jackfish 2/3







## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS29  
SURMONT 2  
SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	September 5, 2023	Last Cal Date:	August 8, 2023
Start time (MST):	10:41	End time (MST):	13:37
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	<u>49.21</u>	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC356008</u>			
Removed Cal Gas Conc:	<u>49.21</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	5472
ZAG Make/Model:	Teledyne API T701		Serial Number:	4297

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001658	1.008955	Backgd or Offset:	13.1
Calibration intercept:	-0.685685	-2.265490	Coeff or Slope:	0.942
				13.2
				0.942

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.6	----
as found span	4919	81.3	800.1	802.0	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.5	----
high point	4919	81.3	800.1	806.0	0.993
second point	4959	40.7	400.6	400.7	1.000
third point	4979	20.3	199.8	197.8	1.010
as left zero	5000	0.0	0.0	-0.3	----
as left span	4919	81.3	800.1	806.0	0.993
Average Correction Factor					1.001

Baseline Corr As found:	802.60	Previous response	800.75	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 8, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:41	End Time (MST):	13:37
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.5	----	Correlation Coefficient	0.999978	≥0.995
800.1	806.0	0.9927			
400.6	400.7	0.9997	Slope	1.008955	0.90 - 1.10
199.8	197.8	1.0102			
			Intercept	-2.265490	+/-30



SO2 Calibration Plot

Date: September 5, 2023

Location: Surmont 2





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Surmont 2      Station number: AMS29  
 Calibration Date: September 7, 2023      Last Cal Date: August 1, 2023  
 Start time (MST): 10:14      End time (MST): 15:47  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.391 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC508338  
 Removed Cal Gas Conc: 5.391 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: CC508338      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 5472  
 ZAG Make/Model: Teledyne API T701      Serial Number: 4297

### Analyzer Information

Analyzer make: Thermo 43iQ-TLE      Analyzer serial #: 1200326170  
 Converter make: Global      Converter serial #: 2022-223  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001190	1.005044	Backgd or Offset:	0.82	0.83
Calibration intercept:	-0.182890	-0.142944	Coeff or Slope:	1.049	1.060

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4926	74.2	80.0	79.8	1.000
as found 2nd point	4963	37.2	40.1	39.7	1.005
as found 3rd point	4982	18.6	20.1	20.1	0.988
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4926	74.2	80.0	80.3	0.996
second point	4963	37.2	40.1	40.1	1.000
third point	4982	18.6	20.1	20.0	1.003
as left zero	5000	0.0	0.0	0.0	----
as left span	4926	74.2	80.0	79.3	1.009
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.000
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	80.0	Prev response:	79.91	*% change:	0.1%
Baseline Corr 2nd AF pt:	39.9	AF Slope:	0.998614	AF Intercept:	-0.142639
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999972		

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SOx scrubber check done after calibrator zero. Adjusted span.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 1, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:14	End Time (MST):	15:47
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999998	≥0.995
80.0	80.3	0.9963			
40.1	40.1	1.0002	Slope	1.005044	0.90 - 1.10
20.1	20.0	1.0027			
			Intercept	-0.142944	+/-3



H<sub>2</sub>S Calibration Plot

Date: September 7, 2023

Location: Surmont 2





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Surmont 2      Station number: AMS29  
 Calibration Date: September 19, 2023      Last Cal Date: September 7, 2023  
 Start time (MST): 10:09      End time (MST): 16:20  
 Reason: Maintenance

### Calibration Standards

Cal Gas Concentration: 5.391 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC508338  
 Removed Cal Gas Conc: 5.391 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: CC508338      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 5472  
 ZAG Make/Model: Teledyne API T701      Serial Number: 4297

### Analyzer Information

Analyzer make: Thermo 43iQ-TLE      Analyzer serial #: 1200326170  
 Converter make: Global      Converter serial #: 2022-220  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005044	1.005327	Backgd or Offset: 0.83	1.19
Calibration intercept:	-0.142944	-0.202877	Coeff or Slope: 1.060	1.080

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4926	74.2	80.0	80.3	0.996
second point	4963	37.2	40.1	40.1	1.000
third point	4982	18.6	20.1	19.7	1.018
as left zero	5000	0.0	0.0	-0.1	----
as left span	4926	74.2	80.0	80.1	0.999

### SO<sub>2</sub> Scrubber Check

Date of last scrubber change:	Ave Corr Factor	1.005
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: NA      Prev response: NA      \*% change: NA  
 Baseline Corr 2nd AF pt: NA      AF Slope: NA      AF Intercept: NA  
 Baseline Corr 3rd AF pt: NA      AF Correlation: NA

*\* = > +/-5% change initiates investigation*

Notes: Swapped old converter out with a new one as it was DOA. Calibrated after replacement. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 19, 2023	Previous Calibration:	September 7, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:09	End Time (MST):	16:20
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	≥0.995
80.0	80.3	0.9963		
40.1	40.1	1.0002	Slope	0.90 - 1.10
20.1	19.7	1.0180		
			Intercept	+/-3



H<sub>2</sub>S Calibration Plot

Date: September 19, 2023

Location: Surmont 2





# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	September 5, 2023	Last Cal Date:	August 8, 2023
Start time (MST):	10:31	End time (MST):	13:37
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC356008	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
C3H8 Cal Gas Conc.	<u>205.7</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
Removed C3H8 Conc.	<u>205.7</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5472
ZAG Make/Model:	Teledyne API T701	Serial Number:	4297

### Analyzer Information

Analyzer make: Thermo 51i-LT                      Analyzer serial #: 1170050149  
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000971	1.006607	Background:	3.64	3.64
Calibration intercept:	-0.017870	-0.029039	Coefficient:	3.987	3.987

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.01	----
as found span	4918	81.3	17.31	17.42	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.03	----
high point	4918	81.3	17.31	17.43	0.993
second point	4959	40.6	8.65	8.64	1.001
third point	4979	20.3	4.32	4.27	1.013
as left zero	5000	0.0	0.00	-0.02	----
as left span	4918	81.3	17.31	17.45	0.992
Average Correction Factor					1.002
Baseline Corr As found:	17.43	Previous response	17.31	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes:                                      Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By:              Braiden Boutillier



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 5, 2023	Previous Calibration:	August 8, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:31	End Time (MST):	13:37
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.03	----	Correlation Coefficient	0.999949	
17.31	17.43	0.9933			≥0.995
8.65	8.64	1.0007	Slope	1.006607	
4.32	4.27	1.0129			0.90 - 1.10
			Intercept	-0.029039	+/-1.5



THC Calibration Plot

Date: September 5, 2023

Location: Surmont 2





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	September 6, 2023	Last Cal Date:	August 9, 2023
Start time (MST):	9:45	End time (MST):	15:02
Reason:	Routine		

### Calibration Standards

NO Gas Cylinder #:	T12YYFE	Cal Gas Expiry Date:	October 30, 2024
NOX Cal Gas Conc:	47.46 ppm	NO Cal Gas Conc:	47.46 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	47.46 ppm	Removed Gas NO Conc:	47.46 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	Teledyne API T700	Serial Number:	5472
ZAG make/model:	Teledyne API T701	Serial Number:	4297

### Analyzer Information

Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.370	1.386	NO bkgnd or offset:	1.3	1.4
NOX coeff or slope:	0.995	0.996	NOX bkgnd or offset:	1.4	1.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	176.2	178.0

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000454	0.997951
NO <sub>x</sub> Cal Offset:	-0.253155	0.647509
NO Cal Slope:	1.000066	0.996977
NO Cal Offset:	-0.992422	-0.611801
NO <sub>2</sub> Cal Slope:	1.000433	0.993705
NO <sub>2</sub> Cal Offset:	0.227914	1.353170



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.4	0.0	0.4	----	----
as found span	4916	84.2	799.2	799.2	0.0	788.8	789.0	-0.2	1.0132	1.0129
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.6	0.2	0.4	----	----
high point	4916	84.2	799.2	799.2	0.0	798.0	796.3	1.7	1.0015	1.0036
second point	4958	42.1	399.6	399.6	0.0	400.0	398.2	1.8	0.9990	1.0035
third point	4979	21.1	200.3	200.3	0.0	200.2	197.7	2.5	1.0004	1.0130
as left zero	5000	0.0	0.0	0.0	0.0	0.5	0.1	0.4	----	----
as left span	4916	84.2	799.2	411.4	387.8	792.2	402.9	389.3	1.0088	1.0211
Average Correction Factor									1.0003	1.0067

Corrected As found	NO <sub>x</sub> = 788.4 ppb	NO = 789.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -1.4%	
Previous Response	NO <sub>x</sub> = 799.3 ppb	NO = 798.3 ppb		*Percent Change	NO = -1.2%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	789.8	402.0	387.8	386.4	1.0036	99.6%
2nd GPT point (200 ppb O <sub>3</sub> )	789.8	601.7	188.1	188.3	0.9989	100.1%
3rd GPT point (100 ppb O <sub>3</sub> )	789.8	696.5	93.3	95.3	0.9790	102.1%
Average Correction Factor					0.9939	100.6%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:45	End Time (MST):	15:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	≥0.995	
799.2	798.0	1.0015			
399.6	400.0	0.9990			
200.3	200.2	1.0004			
			Slope	0.997951	0.90 - 1.10
			Intercept	0.647509	+/-20







# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:45	End Time (MST):	15:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
799.2	796.3	1.0036			
399.6	398.2	1.0035			
200.3	197.7	1.0130			
			Correlation Coefficient	0.999992	≥0.995
			Slope	0.996977	0.90 - 1.10
			Intercept	-0.611801	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:45	End Time (MST):	15:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
387.8	386.4	1.0036		
188.1	188.3	0.9989		
93.3	95.3	0.9790		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 6, 2023

Location: Surmont 2







## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS30 ELLS RIVER SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 1, 2023	Previous Calibration:	August 1, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:04	End Time (MST):	11:40
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999952	≥0.995
800.4	804.2	0.9952			
400.2	399.0	1.0031	Slope	1.006471	0.90 - 1.10
200.1	196.1	1.0204			
			Intercept	-2.616011	+/-30



SO2 Calibration Plot

Date: September 1, 2023

Location: Ells River







# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name:	Ells River	Station number:	AMS30
Calibration Date:	September 7, 2023	Last Cal Date:	August 16, 2023
Start time (MST):	8:23	End time (MST):	12:15
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.08	ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	EY0002443			
Removed Cal Gas Conc:	5.08	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3061
ZAG Make/Model:	API T701H		Serial Number:	358

### Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331
Converter make:	CDN - 101	Converter serial #:	562
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003494	1.001492	Backgd or Offset:	1.63	1.60
Calibration intercept:	-0.199209	-0.079138	Coeff or Slope:	1.136	1.136

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	78.7	80.0	80.5	0.993
as found 2nd point	4961	39.4	40.0	40.0	1.001
as found 3rd point	4980	19.7	20.0	19.8	1.011
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	78.7	80.0	80.1	0.998
second point	4961	39.4	40.0	39.9	1.003
third point	4980	19.7	20.0	19.8	1.011
as left zero	5000	0.0	0.0	0.1	----
as left span	4921	78.7	80.0	80.0	1.000
SO2 Scrubber Check	4921	79.2	800.4	0.0	----
Date of last scrubber change:	N/A			Ave Corr Factor	1.004
Date of last converter efficiency test:	N/A			efficiency	

Baseline Corr As found:	80.5	Prev response:	80.04	*% change:	0.6%
Baseline Corr 2nd AF pt:	40.0	AF Slope:	1.007781	AF Intercept:	-0.199243
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999971		

\* = > +/-5% change initiates investigation

Notes: No adjustments made.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 7, 2023	Previous Calibration:	August 16, 2023
Station Name:	Ells River	Station Number:	AMS30
Start Time (MST):	8:23	End Time (MST):	12:15
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999977	
80.0	80.1	0.9983			≥0.995
40.0	39.9	1.0032	Slope	1.001492	
20.0	19.8	1.0109			0.90 - 1.10
			Intercept	-0.079138	+/-3



TRS Calibration Plot

Date: September 7, 2023

Location: Ells River





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	September 1, 2023	Last Cal Date:	August 1, 2023
Start time (MST):	9:04	End time (MST):	11:40
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH <sub>4</sub> Cal Gas Conc.	499.7 ppm	CH <sub>4</sub> Equiv Conc.	1075.0 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	499.7 ppm	CH <sub>4</sub> Equiv Conc.	1075.0 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	209.2 ppm		
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701H	Serial Number:	3061
		Serial Number:	358

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	0.000252	0.000244	NMHC SP Ratio: 4.67E-05	4.53E-05
CH <sub>4</sub> Retention time:	14.4	14.2	NMHC Peak Area: 195284	201064

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.71	1.019
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.01	1.001
second point	4960	39.6	8.51	8.43	1.010
third point	4980	19.8	4.26	4.15	1.027
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.08	0.997

Average Correction Factor				1.013
Baseline Corr AF:	16.71	Prev response	17.14	*% change -2.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	8.89	1.025
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.08	1.003
second point	4960	39.6	4.56	4.50	1.012
third point	4980	19.8	2.28	2.21	1.031
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.15	0.996
Average Correction Factor					1.015
Baseline Corr AF:	8.89	Prev response	9.22	*% change	-3.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.82	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.93	0.998
second point	4960	39.6	3.96	3.93	1.008
third point	4980	19.8	1.98	1.94	1.022
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.94	0.997
Average Correction Factor					1.009
Baseline Corr AF:	7.82	Prev response	7.92	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.010450	1.000637
THC Cal Offset:	-0.062141	-0.057538
CH <sub>4</sub> Cal Slope:	1.003947	1.002936
CH <sub>4</sub> Cal Offset:	-0.026358	-0.025358
NMHC Cal Slope:	1.016086	0.998741
NMHC Cal Offset:	-0.035983	-0.032581

Notes: Adjusted the span only.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 1, 2023	Previous Calibration:	August 1, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:04	End Time (MST):	11:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999948	$\geq 0.995$
17.03	17.01	1.0010			
8.51	8.43	1.0098			
4.26	4.15	1.0268			
			Slope	1.000637	0.90 - 1.10
			Intercept	-0.057538	+/-0.5





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 1, 2023	Previous Calibration:	August 1, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:04	End Time (MST):	11:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999952	$\geq 0.995$
7.91	7.93	0.9985			
3.96	3.93	1.0079			
1.98	1.94	1.0216			
			Slope	1.002936	0.90 - 1.10
			Intercept	-0.025358	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 1, 2023	Previous Calibration:	August 1, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:04	End Time (MST):	11:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999941	$\geq 0.995$			
9.11	9.08	1.0031						
4.56	4.50	1.0117				Slope	0.998741	0.90 - 1.10
2.28	2.21	1.0314						
			Intercept	-0.032581	$\pm 0.5$			









# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Ells River  
Calibration Date: September 6, 2023  
Start time (MST): 8:39  
Reason: Routine  
Station number: AMS 30  
Last Cal Date: August 9, 2023  
End time (MST): 12:55

### Calibration Standards

NO Gas Cylinder #: T2Y1P2R  
NOX Cal Gas Conc: 50.83 ppm  
Removed Cylinder #:  
Removed Gas NOX Conc: 50.83 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API T701H  
Cal Gas Expiry Date: December 11, 2023  
NO Cal Gas Conc: 49.97 ppm  
Removed Gas Exp Date:  
Removed Gas NO Conc: 49.97 ppm  
NO gas Diff:  
Serial Number: 3061  
Serial Number: 358

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 710321429

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.099	1.129	NO bkgnd or offset:	13.5	13.8
NOX coeff or slope:	0.989	0.986	NOX bkgnd or offset:	13.5	13.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	188.1	189.7

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.004722	0.996459
NO <sub>x</sub> Cal Offset:	-1.140000	-1.100000
NO Cal Slope:	1.006504	1.000500
NO Cal Offset:	-2.140000	-2.240000
NO <sub>2</sub> Cal Slope:	1.001553	0.993934
NO <sub>2</sub> Cal Offset:	0.383629	-0.960703



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
as found span	4920	80.0	813.3	799.5	13.8	793.2	777.5	15.7	1.0253	1.0283
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	-0.2	0.2	----	----
high point	4920	80.0	813.3	799.5	13.8	809.9	798.7	11.2	1.0042	1.0010
second point	4960	40.0	406.6	399.8	6.9	403.4	396.7	6.7	1.0080	1.0077
third point	4980	20.0	203.3	199.9	3.4	200.5	195.7	4.7	1.0141	1.0214
as left zero	5000	0.0	0.0	0.0	0.0	0.2	-0.1	0.2	----	----
as left span	4920	80.0	813.3	423.7	389.6	812.6	426.0	386.5	1.0008	0.9946
Average Correction Factor									1.0088	1.0100

Corrected As found	NO <sub>x</sub> = 793.4 ppb	NO = 777.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -2.8%	
Previous Response	NO <sub>x</sub> = 816.0 ppb	NO = 802.6 ppb		*Percent Change	NO = -3.2%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.0	420.2	389.6	387.0	1.0066	99.3%
2nd GPT point (200 ppb O3)	796.0	620.1	189.7	186.4	1.0175	98.3%
3rd GPT point (100 ppb O3)	796.0	707.0	102.8	100.4	1.0235	97.7%
Average Correction Factor					1.0159	98.4%

Notes:

Adjusted the span only.

Calibration Performed By: Denny Ray Estador



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:39	End Time (MST):	12:55
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.3	809.9	1.0042		
406.6	403.4	1.0080		
203.3	200.5	1.0141		





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:39	End Time (MST):	12:55
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.5	798.7	1.0010		
399.8	396.7	1.0077		
199.9	195.7	1.0214		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 6, 2023	Previous Calibration:	August 9, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:39	End Time (MST):	12:55
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
389.6	387.0	1.0066		
189.7	186.4	1.0175		
102.8	100.4	1.0235		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 6, 2023

Location: Ells River





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Ells River Station number: AMS 30  
 Calibration Date: September 8, 2023 Last Cal Date: August 25, 2023  
 Start time (MST): 11:28 End time (MST): 12:20

Analyzer Make: API T640 S/N: 875  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat S/N: 388751  
 Temp/RH standard: Alicat S/N: 388751

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	19.7	19.5	19.7	<input type="checkbox"/>	
P (mmHg)	733.6	735.9	733.6	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.98	4.75	4.98	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: September 8, 2023 Last Cal Date: August 25, 2023  
 PM w/o HEPA: 17.4 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.6	10.2	10.8	<input checked="" type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: 12.6 w/ HEPA: 0  
 Date Optical Chamber Cleaned: September 8, 2023 <0.2 ug/m3  
 Disposable Filter Changed: September 8, 2023

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_  
 Date RH/T Sensor Cleaned: \_\_\_\_\_

Notes: Adjusted PMT peak test. Inlet head still clean.

Calibration by: Denny Ray Estador





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS505 SAWBONES BAY SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 30, 2023





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 22, 2023	Previous Calibration:	August 9, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:28	End Time (MST):	11:18
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321323

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999984	≥0.995
799.8	799.6	1.0003			
399.9	400.1	0.9995	Slope	1.000437	0.90 - 1.10
200.4	197.9	1.0128			
			Intercept	-0.691921	+/-30



SO2 Calibration Plot

Date: September 22, 2023

Location: Sawbones Bay





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Sawbones Bay	Station number:	AMS505
Calibration Date:	September 21, 2023	Last Cal Date:	August 30, 2023
Start time (MST):	8:18	End time (MST):	12:16
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.15	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517397			
Removed Cal Gas Conc:	5.15	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledybe API T700		Serial Number:	5112
ZAG Make/Model:	Teledybe API T701		Serial Number:	690

### Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	1228021057
Converter make:	Global 150	Converter serial #:	2022-224
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998506	1.003503	Backgd or Offset:	2.21
Calibration intercept:	-0.018155	0.081954	Coeff or Slope:	1.021

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4922	77.7	80.0	80.8	0.992
as found 2nd point	4961	38.8	40.0	40.5	0.989
as found 3rd point	4981	19.4	20.0	20.1	0.999
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4922	77.7	80.0	80.4	0.995
second point	4961	38.8	40.0	40.2	0.994
third point	4981	19.4	20.0	20.1	0.994
as left zero	5000	0.0	0.0	0.2	----
as left span	4922	77.7	80.0	80.1	0.999
SO2 Scrubber Check	4922	77.8	778.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.995
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	80.7	Prev response:	79.90	*% change:	1.0%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.009071	AF Intercept:	0.062101
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999992		

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after calibrator zero. Scrubber check and passed after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 21, 2023	Previous Calibration:	August 30, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:18	End Time (MST):	12:16
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1228021057

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	1.000000	
80.0	80.4	0.9955			≥0.995
40.0	40.2	0.9942	Slope	1.003503	
20.0	20.1	0.9940			0.90 - 1.10
			Intercept	0.081954	+/-3



H<sub>2</sub>S Calibration Plot

Date: September 21, 2023

Location: Sawbones Bay









# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.7	0.0	0.6	----	----
as found span	4917	83.4	799.6	799.6	0.0	786.4	783.8	2.6	1.0168	1.0201
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
high point	4917	83.4	799.6	799.6	0.0	799.7	799.9	-0.3	0.9998	0.9996
second point	4958	41.7	399.8	399.8	0.0	398.8	398.2	0.6	1.0026	1.0041
third point	4979	20.9	200.4	200.4	0.0	197.6	197.1	0.5	1.0141	1.0167
as left zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
as left span	4916	83.4	799.7	339.8	459.9	793.6	335.7	457.9	1.0077	1.0123
Average Correction Factor									1.0055	1.0068

Corrected As found	NO <sub>x</sub> = 785.7 ppb	NO = 783.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -1.7%	
Previous Response	NO <sub>x</sub> = 798.9 ppb	NO = 798.8 ppb		*Percent Change	NO = -1.9%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.2	335.3	459.9	459.5	1.0009	99.9%
2nd GPT point (200 ppb O3)	795.2	539.3	255.9	256.5	0.9977	100.2%
3rd GPT point (100 ppb O3)	795.2	643.4	151.8	152.7	0.9941	100.6%
Average Correction Factor					0.9975	100.2%

Notes: Changed inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 20, 2023	Previous Calibration:	August 4, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:36	End Time (MST):	13:00
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
799.6	799.7	0.9998			
399.8	398.8	1.0026			
200.4	197.6	1.0141			
			Slope	1.001507	0.90 - 1.10
			Intercept	-1.530484	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 20, 2023	Previous Calibration:	August 4, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:36	End Time (MST):	13:00
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999981	≥0.995
799.6	799.9	0.9996			
399.8	398.2	1.0041	Slope	1.001793	0.90 - 1.10
200.4	197.1	1.0167			
			Intercept	-1.830393	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 20, 2023	Previous Calibration:	August 4, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:36	End Time (MST):	13:00
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
459.9	459.5	1.0009			
255.9	256.5	0.9977			
151.8	152.7	0.9941			
			Slope	0.999010	0.90 - 1.10
			Intercept	0.464663	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: September 20, 2023

Location: Sawbones Bay





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS507**  
**KIRBY SOUTH**  
**SEPTEMBER 2023**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

October 29, 2023



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Kirby South      Station number: AMS 507  
Calibration Date: September 14, 2023      Last Cal Date: August 3, 2023  
Start time (MST): 9:47      End time (MST): 12:39  
Reason: Routine

### Calibration Standards

Cal Gas Concentration: 49.18 ppm      Cal Gas Exp Date: February 23, 2025  
Cal Gas Cylinder #: CC303554  
Removed Cal Gas Conc: 49.18 ppm      Rem Gas Exp Date: NA  
Removed Gas Cyl #: NA      Diff between cyl:  
Calibrator Make/Model: API T700      Serial Number: 3804  
ZAG Make/Model: API T701H      Serial Number: 880

### Analyzer Information

Analyzer make: Thermo 43iQ      Analyzer serial #: 1182340007  
Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999847	1.004608	Backgd or Offset:	19.0	20.1
Calibration intercept:	-1.128170	-0.669160	Coeff or Slope:	1.135	1.135

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.8	----
as found span	4919	81.3	799.6	804.0	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.4	----
high point	4919	81.3	799.6	803.0	0.996
second point	4959	40.7	400.3	400.7	0.999
third point	4980	20.3	199.7	200.1	0.998
as left zero	5000	0.0	0.0	-0.5	----
as left span	4919	81.3	799.6	803.0	0.996
Average Correction Factor					0.998

Baseline Corr As found: 803.20      Previous response: 798.37      \*% change: 0.6%  
Baseline Corr 2nd AF pt: NA      AF Slope:      AF Intercept:  
Baseline Corr 3rd AF pt: NA      AF Correlation:

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted zero.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 3, 2023
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	9:47	End Time (MST):	12:39
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.4	----	Correlation Coefficient	0.999997	≥0.995
799.6	803.0	0.9958			
400.3	400.7	0.9991	Slope	1.004608	0.90 - 1.10
199.7	200.1	0.9978			
			Intercept	-0.669160	+/-30






SO2 Calibration Plot

Date: September 14, 2023

Location: Kirby South





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	September 14, 2023	Last Cal Date:	August 2, 2023
Start time (MST):	12:39	End time (MST):	17:10
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.167	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	<u>CC517378</u>			
Removed Cal Gas Conc:	5.167	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3804
ZAG Make/Model:	API T701H		Serial Number:	880

### Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1150840012
Converter make:	Global	Converter serial #:	2022-197
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001457	0.992888	Backgd or Offset:	1.53	1.54
Calibration intercept:	0.019070	0.019018	Coeff or Slope:	1.048	1.048

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4923	77.4	80.0	79.5	1.006
as found 2nd point	4961	38.8	40.1	39.5	1.015
as found 3rd point	4981	19.3	19.9	19.7	1.012
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4923	77.4	80.0	79.5	1.006
second point	4961	38.8	40.1	39.7	1.010
third point	4981	19.3	19.9	19.8	1.007
as left zero	5000	0.0	0.0	0.2	----
as left span	4923	77.4	80.0	79.0	1.012
SO2 Scrubber Check	4919	80.0	800.2	-0.1	----
Date of last scrubber change:	25-Jul-23			Ave Corr Factor	1.008
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	79.5	Prev response:	80.11	*% change:	-0.8%
Baseline Corr 2nd AF pt:	39.5	AF Slope:	0.994027	AF Intercept:	-0.120860
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999975		

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Scrubber check done after MPAF's, passed. No adjustments made.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 2, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	12:39	End Time (MST):	17:10
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
80.0	79.5	1.0060		
40.1	39.7	1.0100	Slope	0.90 - 1.10
19.9	19.8	1.0072		
			Intercept	+/-3



H<sub>2</sub>S Calibration Plot

Date: September 14, 2023

Location: Kirby South





# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	September 14, 2023	Last Cal Date:	August 3, 2023
Start time (MST):	9:47	End time (MST):	12:39
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC303554	Cal Gas Expiry Date:	March 23, 2025
CH4 Cal Gas Conc.	496.6 ppm	CH4 Equiv Conc.	1061.7 ppm
C3H8 Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	496.6 ppm	CH4 Equiv Conc.	1061.7 ppm
Removed C3H8 Conc.	205.5 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	3804
ZAG Make/Model:	API T701H	Serial Number:	880

### Analyzer Information

Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005
Analyzer Range:	0 - 20 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999381	0.999942	Background:	2.64	2.83
Calibration intercept:	-0.016800	0.005209	Coefficient:	3.799	3.975

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.07	----
as found span	4919	81.3	17.26	16.50	1.046
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.04	----
high point	4919	81.3	17.26	17.29	0.998
second point	4959	40.7	8.64	8.61	1.004
third point	4980	20.3	4.31	4.30	1.003
as left zero	5000	0.0	0.00	0.01	----
as left span	4919	81.3	17.26	17.18	1.005
Average Correction Factor					1.002
Baseline Corr As found:	16.57	Previous response	17.24	*% change	-4.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	August 3, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	9:47	End Time (MST):	12:39
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.04	----	Correlation Coefficient	0.999980	≥0.995
17.26	17.29	0.9984			
8.64	8.61	1.0038	Slope	0.999942	0.90 - 1.10
4.31	4.30	1.0029			
			Intercept	0.005209	+/-1.5



THC Calibration Plot

Date: September 14, 2023

Location: Kirby South





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	September 15, 2023	Last Cal Date:	August 3, 2023
Start time (MST):	8:45	End time (MST):	13:40
Reason:	Routine		

### Calibration Standards

NO Gas Cylinder #:	T34ULGL	Cal Gas Expiry Date:	March 8, 2025
NOX Cal Gas Conc:	49.39 ppm	NO Cal Gas Conc:	49.02 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	49.39 ppm	Removed Gas NO Conc:	49.02 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	3804
ZAG make/model:	API 701H	Serial Number:	880

### Analyzer Information

Analyzer make: Thermo 42iQ  
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1182340006

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.485	1.485	NO bkgnd or offset:	1.5	1.5
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	1.6	1.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.75	167.97

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.006090	1.009702
NO <sub>x</sub> Cal Offset:	-4.471991	-4.611056
NO Cal Slope:	1.005987	1.011497
NO Cal Offset:	-5.373912	-5.412656
NO <sub>2</sub> Cal Slope:	1.007635	1.004706
NO <sub>2</sub> Cal Offset:	1.168421	1.319502





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
as found span	4919	81.0	800.1	794.1	6.0	804.0	798.9	5.1	0.9952	0.9940
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
high point	4919	81.0	800.1	794.1	6.0	806.0	801.0	5.0	0.9927	0.9914
second point	4960	40.5	400.0	397.0	3.0	395.5	391.9	3.6	1.0114	1.0131
third point	4980	20.2	199.5	198.0	1.5	193.5	190.8	2.7	1.0312	1.0379
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
as left span	4919	81.0	800.1	379.9	420.2	803.0	381.6	420.9	0.9964	0.9956
Average Correction Factor									1.0118	1.0141

Corrected As found	NO <sub>x</sub> = 804.3 ppb	NO = 799.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.5%
Previous Response	NO <sub>x</sub> = 800.5 ppb	NO = 793.5 ppb		*Percent Change	NO = 0.7%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.2	383.0	420.2	422.6	0.9943	100.6%
2nd GPT point (200 ppb O3)	797.2	622.1	181.1	185.2	0.9778	102.3%
3rd GPT point (100 ppb O3)	797.2	716.0	87.2	89.2	0.9775	102.3%
Average Correction Factor					0.9832	101.7%

Notes:

No adjustments made.

Calibration Performed By:

Braiden Boutilier



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 3, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	8:45	End Time (MST):	13:40
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999853	≥0.995
800.1	806.0	0.9927			
400.0	395.5	1.0114	Slope	1.009702	0.90 - 1.10
199.5	193.5	1.0312			
			Intercept	-4.611056	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 3, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	8:45	End Time (MST):	13:40
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
794.1	801.0	0.9914		
397.0	391.9	1.0131		
198.0	190.8	1.0379		
			0.999796	
			1.011497	
			-5.412656	

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	September 15, 2023	Previous Calibration:	August 3, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	8:45	End Time (MST):	13:40
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
420.2	422.6	0.9943		
181.1	185.2	0.9778		
87.2	89.2	0.9775		



NO<sub>x</sub> Calibration Plot

Date: September 15, 2023

Location: Kirby South





End of Report