



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

OCTOBER 2023 MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

November 29, 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01
BERTHA GANTER - FORT MCKAY

OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	October 4, 2023	Last Cal Date:	September 12, 2023
Start time (MST):	7:26	End time (MST):	10:30
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC418809			
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999839	0.998709	Backgd or Offset:	19.3	19.2
Calibration intercept:	-0.113750	0.106791	Coeff or Slope:	0.890	0.887

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4918	81.3	800.3	796.0	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.6	----
high point	4918	81.3	800.3	799.8	1.001
second point	4959	40.7	400.6	399.3	1.003
third point	4979	20.3	199.8	199.6	1.001
as left zero	5000	0.0	0.0	0.7	----
as left span	4918	81.3	800.3	800.7	0.999
Average Correction Factor					1.002

Baseline Corr As found:	795.50	Previous response	800.02	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 12, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	7:26	End Time (MST):	10:30
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.6	----	Correlation Coefficient	0.999996	≥0.995
800.3	799.8	1.0006			
400.6	399.3	1.0032	Slope	0.998709	0.90 - 1.10
199.8	199.6	1.0011			
			Intercept	0.106791	+/-30



SO2 Calibration Plot

Date: October 4, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: October 16, 2023 Last Cal Date: September 11, 2023
 Start time (MST): 9:39 End time (MST): 15:15
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153461
 Converter make: CD Nova Converter serial #: 470
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995649	0.993649	Backgd or Offset: 2.26	2.25
Calibration intercept:	0.279998	0.200001	Coeff or Slope: 0.906	0.906

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4921	78.4	80.0	78.9	1.016
as found 2nd point	4960	39.2	40.0	39.7	1.013
as found 3rd point	4980	19.6	20.0	19.8	1.020
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	78.4	80.0	79.7	1.004
second point	4960	39.2	40.0	39.9	1.002
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.3	----
as left span	4921	78.4	80.0	81.0	0.988
SO2 Scrubber Check	4919	81.3	813.0	0.0	----

Date of last scrubber change:	December 17, 2021	Ave Corr Factor	1.000
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 78.7 Prev response: 79.93 *% change: -1.6%
 Baseline Corr 2nd AF pt: 39.5 AF Slope: 0.984363 AF Intercept: 0.199997
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999993

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 16, 2023	Previous Calibration:	September 11, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:39	End Time (MST):	15:15
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999999	≥0.995
80.0	79.7	1.0037			
40.0	39.9	1.0025	Slope	0.993649	0.90 - 1.10
20.0	20.1	0.9949			
			Intercept	0.200001	+/-3



TRS Calibration Plot

Date: October 16, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: October 16, 2023 Last Cal Date: September 11, 2023
 Start time (MST): 9:39 End time (MST): 15:15
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1200326167
 Converter make: Thermo Converter Converter serial #: N/A
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997856	0.989711	Backgd or Offset:	1.70	1.71
Calibration intercept:	0.156776	0.316799	Coeff or Slope:	0.997	0.997

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4922	78.4	80.0	80.4	0.995
as found 2nd point	4960	39.2	40.0	41.5	0.964
as found 3rd point	4980	19.6	20.0	20.1	0.995
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4922	78.4	80.0	79.3	1.009
second point	4960	39.2	40.0	40.3	0.993
third point	4980	19.6	20.0	20.0	1.000
as left zero	5000	0.0	0.0	0.8	----
as left span	4922	78.4	80.0	77.7	1.029
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	March 21, 2022			Ave Corr Factor	1.000
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 80.4 Prev response: 79.95 *% change: 0.6%
 Baseline Corr 2nd AF pt: 41.5 AF Slope: 1.007343 AF Intercept: 0.256695
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999651

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 16, 2023	Previous Calibration:	September 11, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:39	End Time (MST):	15:15
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999938	
80.0	79.3	1.0085			≥0.995
40.0	40.3	0.9925	Slope	0.989711	
20.0	20.0	0.9999			0.90 - 1.10
			Intercept	0.316799	+/-3



H₂S Calibration Plot

Date: October 16, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	October 4, 2023	Last Cal Date:	September 12, 2023
Start time (MST):	7:26	End time (MST):	10:30
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	5609

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.97E-04	2.98E-04	NMHC SP Ratio:	6.42E-05	6.57E-05
CH ₄ Retention time:	14.6	14.6	NMHC Peak Area:	143189	139930
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.27	17.04	1.013
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	17.27	17.30	0.998
second point	4959	40.7	8.64	8.47	1.020
third point	4980	20.3	4.31	4.21	1.023
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.27	17.23	1.002
Average Correction Factor					1.014

Baseline Corr AF:	17.04	Prev response	17.19	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	9.18	9.09	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	9.18	9.23	0.995
second point	4959.3	40.7	4.60	4.55	1.010
third point	4979.7	20.3	2.29	2.28	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	9.18	9.16	1.002
Average Correction Factor					1.004
Baseline Corr AF:	9.09	Prev response	9.15	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	7.95	1.017
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	8.09	8.07	1.002
second point	4959.3	40.7	4.05	3.92	1.031
third point	4979.7	20.3	2.02	1.94	1.043
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.07	1.002
Average Correction Factor					1.025
Baseline Corr AF:	7.95	Prev response	8.04	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998565	1.002996
THC Cal Offset:	-0.056289	-0.080272
CH ₄ Cal Slope:	1.000577	1.000153
CH ₄ Cal Offset:	-0.051154	-0.055151
NMHC Cal Slope:	0.996780	1.005501
NMHC Cal Offset:	-0.004336	-0.025121

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 12, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	7:26	End Time (MST):	10:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999847	≥ 0.995			
17.27	17.30	0.9979						
8.64	8.47	1.0202				Slope	1.002996	0.90 - 1.10
4.31	4.21	1.0230						
			Intercept	-0.080272	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 12, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	7:26	End Time (MST):	10:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999714	≥ 0.995
8.09	8.07	1.0016			
4.05	3.92	1.0314			
2.02	1.94	1.0427			
			Slope	1.000153	0.90 - 1.10
			Intercept	-0.055151	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 12, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	7:26	End Time (MST):	10:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999925	≥ 0.995			
9.18	9.23	0.9946						
4.60	4.55	1.0105				Slope	1.005501	0.90 - 1.10
2.29	2.28	1.0062						
			Intercept	-0.025121	± 0.5			



NMHC Calibration Plot

Date: October 4, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	October 25, 2023	Last Cal Date:	October 4, 2023
Start time (MST):	10:19	End time (MST):	13:05
Reason:	As Found		

Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	497.2 ppm	CH4 Equiv Conc.	1061.8 ppm
C3H8 Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.2 ppm	CH4 Equiv Conc.	1061.8 ppm
Removed C3H8 Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.98E-04	2.98E-04	NMHC SP Ratio:	6.57E-05	6.57E-05
CH4 Retention time:	14.6	14.6	NMHC Peak Area:	139930	139930
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.12	----
as found span	4918	81.3	17.27	17.15	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.27	17.02	1.015
Average Correction Factor					

Baseline Corr AF:	17.02	Prev response	17.19	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	9.18	8.97	1.023
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	9.18	8.98	1.023
Average Correction Factor					
Baseline Corr AF:	8.97	Prev response	9.15	*% change	-1.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.12	----
as found span	4918	81.3	8.09	8.17	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.04	1.006
Average Correction Factor					
Baseline Corr AF:	8.05	Prev response	8.04	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998565	
THC Cal Offset:	-0.056289	
CH ₄ Cal Slope:	1.000577	
CH ₄ Cal Offset:	-0.051154	
NMHC Cal Slope:	0.996780	
NMHC Cal Offset:	-0.004336	


Notes: Changed out the ZAG.

Calibration Performed By: Rene Chamberland

NMHC Calibration Plot

Date: October 25, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
Calibration Date: October 3, 2023 Last Cal Date: September 6, 2023
Start time (MST): 9:21 End time (MST): 13:49
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1P9L Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.84 ppm NO Cal Gas Conc: 50.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.84 ppm Removed Gas NO Conc: 50.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3565
ZAG make/model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1218153357
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.357	1.378	NO bkgnd or offset:	7.0	7.0
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	7.7	7.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	190.7	191.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999396	0.999045
NO _x Cal Offset:	-0.540000	-0.740000
NO Cal Slope:	0.999215	0.998016
NO Cal Offset:	-1.080000	-1.060000
NO ₂ Cal Slope:	1.001909	1.000561
NO ₂ Cal Offset:	-0.228647	1.020945



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	0.0	-0.4	----	----
as found span	4920	80.0	813.4	800.6	12.8	809.0	789.2	19.8	1.0055	1.0145
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.2	----	----
high point	4920	80.0	813.4	800.6	12.8	812.2	798.6	13.6	1.0015	1.0026
second point	4960	40.0	406.7	400.3	6.4	405.4	397.8	7.7	1.0033	1.0063
third point	4980	20.0	203.4	200.2	3.2	201.7	197.6	4.1	1.0082	1.0130
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
as left span	4920	80.0	813.4	394.6	418.8	812.3	396.5	415.8	1.0014	0.9953
Average Correction Factor									1.0043	1.0073

Corrected As found	NO _x = 809.4 ppb	NO = 789.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.4%	
Previous Response	NO _x = 812.4 ppb	NO = 798.9 ppb		*Percent Change	NO = -1.2%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.4	389.4	418.8	419.3	0.9988	100.1%
2nd GPT point (200 ppb O3)	795.4	595.8	212.4	214.5	0.9902	101.0%
3rd GPT point (100 ppb O3)	795.4	697.4	110.8	112.9	0.9814	101.9%
Average Correction Factor					0.9901	101.0%

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 6, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:21	End Time (MST):	13:49
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
813.4	812.2	1.0015			
406.7	405.4	1.0033			
203.4	201.7	1.0082			
			Slope	0.999045	0.90 - 1.10
			Intercept	-0.740000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 6, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:21	End Time (MST):	13:49
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.6	798.6	1.0026			
400.3	397.8	1.0063	Slope	0.90 - 1.10	
200.2	197.6	1.0130			
			Intercept	-1.060000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 6, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:21	End Time (MST):	13:49
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
418.8	419.3	0.9988		
212.4	214.5	0.9902		
110.8	112.9	0.9814		
			0.999958	
			1.000561	
			1.020945	



NO_x Calibration Plot

Date: October 3, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: October 2, 2023 Last Cal Date: September 1, 2023
 Start time (MST): 9:35 End time (MST): 12:29
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 1107
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999943	1.003114	Backgd or Offset:	3.2	3.2
Calibration intercept:	0.060000	0.080000	Coeff or Slope:	1.010	1.010

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	5000	863.1	400.0	400.8	0.998
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.2	----
high point	5000	863.1	400.0	401.4	0.997
second point	5000	742.5	200.0	200.6	0.997
third point	5000	651.7	100.0	100.3	0.997
as left zero	5000	0.0	0.0	0.0	----
as left span	5000	863.1	400.0	404.7	0.988
Average Correction Factor					0.997

Baseline Corr As found:	400.9	Previous response	400.0	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 2, 2023	Previous Calibration:	September 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:35	End Time (MST):	12:29
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	1.000000	≥0.995
400.0	401.4	0.9965			
200.0	200.6	0.9970	Slope	1.003114	0.90 - 1.10
100.0	100.3	0.9970			
			Intercept	0.080000	+/- 5



O₃ Calibration Plot

Date: October 2, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
 Calibration Date: October 19, 2023 Last Cal Date: September 14, 2023
 Start time (MST): 12:56 End time (MST): 15:27

Analyzer Make: Teledyne API T640 S/N: 324
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	17.6	17.1	17.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	725.4	724	725.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.03	5.09	5.03	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>October 19, 2023</u>		Last Cal Date: <u>September 14, 2023</u>		
	PM w/o HEPA: <u>1.7</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.0	11.2	11.2	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: <u>1.9</u>		w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:	<u>October 19, 2023</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>October 19, 2023</u>				

Annual Maintenance

Date Sample Tube Cleaned: September 14, 2023
 Date RH/T Sensor Cleaned: October 19, 2023

Notes: Flow, temperature, and pressure all within limits. Leak checks passed. Disposable filter changed. PMT peak test verified. Optical chamber and RH/T sensor cleaned.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	October 23, 2023	Last Cal Date:	September 14, 2023
Start time (MST):	10:27	End time (MST):	13:11
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	<u>3040</u>	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>3040</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000977	1.000806	Backgd or Offset:	-0.012	-0.012
Calibration intercept:	0.127800	0.185827	Coeff or Slope:	0.990	0.990

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4933	66.7	40.6	40.6	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.6	40.6	0.998
second point	4966	33.3	20.2	20.7	0.977
third point	4983	16.7	10.2	10.4	0.979
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.5	40.2	1.008
Average Correction Factor					0.985

Baseline Corr As found:	40.59	Prev response:	40.72	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 23, 2023	Previous Calibration:	September 14, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:27	End Time (MST):	13:11
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999863	
40.6	40.6	0.9982			≥0.995
20.2	20.7	0.9768	Slope	1.000806	
10.2	10.4	0.9792			0.90 - 1.10
			Intercept	0.185827	+/-1.5



CO Calibration Plot

Date: October 23, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	October 13, 2023	Last Cal Date:	September 5, 2023
Start time (MST):	9:45	End time (MST):	12:54
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>60,200</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
N2 Gen Make/Model:	Peak Sci NG5000		Serial Number:	7220900034

Analyzer Information

Analyzer make: Teledyne API 360	Analyzer serial #: 442
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001868	1.001376	Backgd or Offset:	0.045	0.045
Calibration intercept:	-5.320000	-5.300000	Coeff or Slope:	0.875	0.876

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-0.2	----
as found span	2920	80.0	1605.3	1596.1	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.2	----
high point	2920	80.0	1605.3	1604.9	1.000
second point	2960	40.0	802.7	796.0	1.008
third point	2980	20.0	401.3	390.9	1.027
as left zero	3000	0.0	0.0	0.2	----
as left span	2960	40.0	802.7	784.2	1.024
Average Correction Factor					1.012

Baseline Corr As found:	1596.30	Prev response:	1603.01	*% change:	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	October 13, 2023	Previous Calibration	September 5, 2023
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS01
Start Time (MST)	9:45	End Time (MST)	12:54
Analyzer make	Teledyne API 360	Analyzer serial #	442

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999946	≥0.995
1605.3	1604.9	1.0003	Slope	1.001376	0.90 - 1.10
802.7	796.0	1.0084	Intercept	-5.300000	+/-10
401.3	390.9	1.0267			



CO₂ Calibration Plot

Date: October 13, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	October 18, 2023	Last Cal Date:	September 7, 2023
Start time (MST):	9:39	End time (MST):	15:05
NH3 Cal Date:	October 19, 2023	Last Cal Date:	September 8, 2023
Start time (MST):	10:04	End time (MST):	15:50
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	50.84	ppm	NO Gas Cylinder #:	T2Y1P9L
NO Cal Gas Conc:	50.04	ppm	NO Cal Gas Expiry:	March 3, 2028
Removed NOX Conc:	50.84	ppm	Removed Cylinder #:	NA
Removed NO Conc:	50.04	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.58	ppm	NH3 Gas Cylinder #:	CC743587
Removed NH3 Conc:	74.90	ppm	NH3 Cal Gas Expiry:	August 22, 2024
NH3 gas Diff:	-3.5%		Removed Cylinder #:	CC744566
Calibrator Model:	Teledyne API T700		Removed cyl Expiry:	December 21, 2023
ZAG make/model:	Teledyne API T701		Serial Number:	3565
			Serial Number:	5609

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	475
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	5.10
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	506

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.996	0.991	TN coefficient:	1.001	0.996
NOX coefficient:	0.999	0.993	NO bkgnd:	-0.9	-0.9
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-0.3	-0.3
NH3 coefficient:	0.995	0.943	TN bkgnd:	1.2	1.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999916	1.000492
NO _x Cal Offset:	-2.000000	-1.780000
NO Cal Slope:	0.999929	1.000500
NO Cal Offset:	-2.380000	-1.880000
NO ₂ Cal Slope:	0.997369	1.000741
NO ₂ Cal Offset:	0.086604	-0.538093
NH3 Cal Slope:	1.003196	0.995734
NH3 Cal Offset:	-1.081485	1.216520
TN Cal Slope:	1.005635	0.998191
TN Cal Offset:	-0.793819	1.291202



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	-0.1	----	----
as found NO	4920	80.0	813.4	813.4	----	812.5	811.9	0.5	1.001	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
high NO point	4920	80.0	813.4	813.4	----	814.0	813.0	1.3	0.999	----
NO/O3 point	4920	80.0	813.4	813.4	----	811.5	809.7	1.8	1.002	----
as found NH3	3416	84.1	1799.7	----	1799.7	1775.2	----	1770.2	1.014	1.017
new NH3 cyl rp	3418	82.2	1798.5	----	1798.5	1715.2	----	1710.4	1.049	1.052
first NH3	3418	82.2	1798.5	----	1798.5	1795.3	----	1790.9	1.002	1.004
second NH3	3454	45.7	1000.0	----	1000.0	1001.3	----	998.7	0.999	1.001
third NH3	3477	22.8	498.9	----	498.9	500.2	----	498.7	0.997	1.000
Average Correction Factor									1.0009	1.0020

Corrected As found TN = 812.8 ppb NO_x = 812.2 ppb NH3 = 1770.3 ppb

Previous Response TN = 817.2 ppb NO_x = 811.4 ppb NH3 = 1804.4 ppb

NH3 Previous Converter Efficiency = 99.5%

NH3 Current Converter Efficiency = 94.3%

*Percent Change TN = -0.5%

*Percent Change NO_x = 0.1%

*Percent Change NH3 = -1.9%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

TN Calibration Summary


Version-05-2023

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 7, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:39	End Time (MST):	15:05
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	0.999995	≥0.995
1798.5	1795.3	1.0018			
1000.0	1001.3	0.9987	Slope	0.998191	0.90 - 1.10
498.9	500.2	0.9973			
			Intercept	1.291202	+/-20





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 7, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:39	End Time (MST):	15:05
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
1798.5	1790.9	1.0043			
1000.0	998.7	1.0013			
498.9	498.7	1.0003			
			Slope	0.995734	0.90 - 1.10
			Intercept	1.216520	+/-20





Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-05-2023


Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 7, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:39	End Time (MST):	15:05
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.4	813.0	1.0005		
406.7	404.0	1.0067		
203.4	200.2	1.0158		
			0.999980	
			1.000492	
			-1.780000	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 7, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:39	End Time (MST):	15:05
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.6	800.0	1.0008			
400.3	398.1	1.0056			
200.2	196.1	1.0207			
			Slope	1.000500	0.90 - 1.10
			Intercept	-1.880000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 7, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:39	End Time (MST):	15:05
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
413.9	413.4	1.0012		
216.3	217.1	0.9963		
115.1	113.4	1.0150		
			0.999966	
			1.000741	
			-0.538093	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 18, 2023


Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: October 19, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

PLUVIO CALIBRATION

Version-07-2017

Station Information

Station Name:	Fort McKay - Bertha Ganter	Station number:	AMS 01
Calibration Date:	October 13, 2023	Last Cal Date:	April 21, 2023
Start time (MST):	12:05	End time (MST):	13:15
Make/Model#:	OTT Pluvio 2	Pluvio Serial #:	363525

Annual Weight Test

Date of Check:	October 13, 2023	Previous Check Date:	April 21, 2023
Weights Serial #:	59781	Passed/Failed:	Pass

Spring Drip Test

Date of Check:	April 21, 2023	Previous Check Date:	October 5, 2022
Start Time:	14:00	Stop Time:	14:30
Nozzle Size Used:	100		
Amount of Water Used:	1 L	Data Logger Total:	24.47 mm
Converted Amount of Water used:	25 mm		

Rainfall (mm) = 1 L / 0.04 m² = 25 mm is what we should be reading.

Fall Drip Test

Date of Check:	October 13, 2023	Previous Check Date:	April 21, 2023
Start Time:	12:45	Stop Time:	13:15
Nozzle Size Used:	100		
Amount of Water Used:	1 L	Data Logger Total:	24 mm
Converted Amount of Water used:	25 mm		

Rainfall (mm) = 1 L / 0.04 m² = 25 mm is what we should be reading.

Notes:

Performed By: _____ Tyler Tracksell



OTT Pluvio2 - Guided accuracy test



See test details below:

Date:	2023-10-13 13:11:13
OTT Pluvio2 type:	OTT Pluvio2
Bucket type:	400 cm ²
Serial number / HW Index:	363525 /f3
Firmware version:	V1.40.1
Intensity units:	mm/min
Measured total weight:	5338.08 gram (133.45 mm;5.254 inch)
Information of applied weight:	1000.00 gram (25.00 mm;0.984 inch)
Measured test weight:	1002.72 gram (25.07 mm;0.987 inch)
Permitted deviation:	+2.00 gram (+-0.05 mm;+-0.002 inch)
Deviation:	2.72 gram (0.07 mm;0.003 inch)
base weight	4335.36 gram (108.38 mm;4.267 inch)

Test passed: Yes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	October 20, 2023	Last Cal Date:	September 25, 2023
Start time (MST):	9:27	End time (MST):	12:30
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1185
ZAG Make/Model:	API T701		Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995496	0.999174	Backgd or Offset:	18.1	17.5
Calibration intercept:	0.154391	0.714964	Coeff or Slope:	0.797	0.774

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4920	80.2	801.6	813.0	0.986
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.7	----
high point	4920	80.2	801.6	801.7	1.000
second point	4960	40.1	400.8	401.2	0.999
third point	4980	20.0	199.9	200.5	0.997
as left zero	5000	0.0	0.0	0.5	----
as left span	4920	80.2	801.6	805.2	0.996
Average Correction Factor					0.999

Baseline Corr As found:	812.60	Previous response	798.19	*% change	1.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 25, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:27	End Time (MST):	12:30
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient	1.000000	≥0.995
801.6	801.7	0.9999			
400.8	401.2	0.9991	Slope	0.999174	0.90 - 1.10
199.9	200.5	0.9971			
			Intercept	0.714964	+/-30



SO2 Calibration Plot

Date: October 20, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mildred Lake Station number: AMS02
 Calibration Date: October 25, 2023 Last Cal Date: September 20, 2023
 Start time (MST): 6:45 End time (MST): 10:54
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.29 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345191
 Removed Cal Gas Conc: 5.29 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1185
 ZAG Make/Model: API T701 Serial Number: 5608

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 12113311966
 Converter make: Global G150 Converter serial #: 2022-198
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004536	1.002536	Backgd or Offset: 1.78	1.71
Calibration intercept:	0.220806	-0.059198	Coeff or Slope: 0.781	0.754

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4924	75.6	80.0	83.2	0.961
as found 2nd point	4962	37.8	40.0	41.6	0.961
as found 3rd point	4981	18.9	20.0	20.6	0.971
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4924	75.6	80.0	80.1	0.999
second point	4962	37.8	40.0	40.2	0.995
third point	4981	18.9	20.0	19.8	1.010
as left zero	5000	0.0	0.0	0.1	----
as left span	4924	75.6	80.0	80.7	0.991
SO2 Scrubber Check	4920	80.2	802.0	0.1	----
Date of last scrubber change:		20-Sep-23		Ave Corr Factor	1.001
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 83.2 Prev response: 80.57 *% change: 3.2%
 Baseline Corr 2nd AF pt: 41.6 AF Slope: 1.040969 AF Intercept: -0.079168
 Baseline Corr 3rd AF pt: 20.6 AF Correlation: 0.999993

* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrated zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 25, 2023	Previous Calibration:	September 20, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	6:45	End Time (MST):	10:54
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	≥0.995
80.0	80.1	0.9986		
40.0	40.2	0.9949	Slope	0.90 - 1.10
20.0	19.8	1.0099		
			Intercept	+/-3



H₂S Calibration Plot

Date: October 25, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	October 20, 2023	Last Cal Date:	September 25, 2023
Start time (MST):	9:27	End time (MST):	12:30
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
C ₃ H ₈ Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
Removed C ₃ H ₈ Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.99E-04	3.04E-04	NMHC SP Ratio:	4.58E-05	4.59E-05
CH ₄ Retention time:	14.6	14.8	NMHC Peak Area:	192179	191592
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	16.67	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.76	1.003
second point	4960	40.1	8.41	8.36	1.006
third point	4980	20.0	4.19	4.16	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.78	1.002
Average Correction Factor					1.006

Baseline Corr AF:	16.67	Prev response	16.78	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	8.78	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.79	1.001
second point	4960	40.1	4.40	4.41	0.997
third point	4980	20.0	2.19	2.21	0.991
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.83	0.997
Average Correction Factor					0.996
Baseline Corr AF:	8.78	Prev response	8.79	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	7.88	1.018
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	7.97	1.007
second point	4960	40.1	4.01	3.94	1.018
third point	4980	20.0	2.00	1.95	1.025
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	7.96	1.008
Average Correction Factor					1.017
Baseline Corr AF:	7.88	Prev response	7.99	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998613	0.996697
THC Cal Offset:	-0.013917	-0.011323
CH ₄ Cal Slope:	0.999502	0.994304
CH ₄ Cal Offset:	-0.025057	-0.023065
NMHC Cal Slope:	0.998932	0.998828
NMHC Cal Offset:	0.008542	0.011943

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 25, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:27	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
16.82	16.76	1.0035						
8.41	8.36	1.0065				Slope	0.996697	0.90 - 1.10
4.19	4.16	1.0073						
			Intercept	-0.011323	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 25, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:27	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999954	≥ 0.995
8.02	7.97	1.0066			
4.01	3.94	1.0179			
2.00	1.95	1.0250			
			Slope	0.994304	0.90 - 1.10
			Intercept	-0.023065	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 25, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	9:27	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999991	≥ 0.995			
8.80	8.79	1.0006						
4.40	4.41	0.9965				Slope	0.998828	0.90 - 1.10
2.19	2.21	0.9911						
			Intercept	0.011943	± 0.5			



NMHC Calibration Plot

Date: October 20, 2023

Location: Mildred Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	October 10, 2023	Last Cal Date:	September 13, 2023
Start time (MST):	9:20	End time (MST):	12:04
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.87	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC446753			
Removed Cal Gas Conc:	50.87	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701		Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994771	1.003929	Backgd or Offset:	22.1	24.1
Calibration intercept:	2.033465	-0.144528	Coeff or Slope:	0.860	0.873

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.8	----
as found span	4921	78.6	799.7	792.7	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4921	78.6	799.7	802.6	0.996
second point	4961	39.3	399.8	401.6	0.996
third point	4980	19.6	199.4	199.9	0.998
as left zero	5000	0.0	0.0	-0.4	----
as left span	4921	78.6	799.7	804.5	0.994
Average Correction Factor					0.997

Baseline Corr As found:	791.90	Previous response	797.59	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No Maintenance done. Zero and span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 10, 2023	Previous Calibration:	September 13, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	9:20	End Time (MST):	12:04
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999999	≥0.995
799.7	802.6	0.9964			
399.8	401.6	0.9956	Slope	1.003929	0.90 - 1.10
199.4	199.9	0.9976			
			Intercept	-0.144528	+/-30



SO2 Calibration Plot

Date: October 10, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: October 13, 2023 Last Cal Date: September 1, 2023
 Start time (MST): 6:15 End time (MST): 10:11
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.42 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345266
 Removed Cal Gas Conc: 5.42 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701H Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1008841400
 Converter make: Global Converter serial #: 2022-200
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004754	0.994226	Backgd or Offset: 1.8	1.8
Calibration intercept:	-0.037663	0.082074	Coeff or Slope: 1.095	1.095

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.1	80.3	80.3	1.000
as found 2nd point	4963	37.0	40.1	40.2	0.998
as found 3rd point	4982	18.5	20.1	19.9	1.008
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	74.1	80.3	79.9	1.005
second point	4963	37.0	40.1	40.1	1.000
third point	4982	18.5	20.1	19.9	1.008
as left zero	5000	0.0	0.0	0.1	----
as left span	4926	74.1	80.3	79.4	1.012
SO2 Scrubber Check	4920	80.0	800.0	-0.1	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	1.004
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 80.3 Prev response: 80.67 *% change: -0.5%
 Baseline Corr 2nd AF pt: 40.2 AF Slope: 1.000487 AF Intercept: -0.037805
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999992

* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 13, 2023	Previous Calibration:	September 1, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:15	End Time (MST):	10:11
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999990	≥0.995
80.3	79.9	1.0053			
40.1	40.1	1.0002	Slope	0.994226	0.90 - 1.10
20.1	19.9	1.0076			
			Intercept	0.082074	+/-3



H₂S Calibration Plot

Date: October 13, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	October 10, 2023	Last Cal Date:	September 13, 2023
Start time (MST):	9:20	End time (MST):	12:03
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	NA	Cal Gas Expiry Date:	March 10, 1931
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	1.91E-04	1.95E-04	NMHC SP Ratio:	4.09E-05
CH ₄ Retention time:	11.8	11.8	NMHC Peak Area:	215814
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	16.64	16.26	1.023
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	16.64	16.63	1.000
second point	4961	39.3	8.32	8.27	1.006
third point	4980	19.6	4.15	4.10	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	16.64	16.57	1.004
Average Correction Factor					1.006

Baseline Corr AF:	16.26	Prev response	16.61	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	8.82	8.59	1.027
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	8.82	8.80	1.002
second point	4961	39.3	4.41	4.39	1.004
third point	4980	19.6	2.20	2.18	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	8.82	8.76	1.007
Average Correction Factor					1.005
Baseline Corr AF:	8.59	Prev response	8.81	*% change	-2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	7.82	7.67	1.019
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	7.82	7.83	0.998
second point	4961	39.3	3.91	3.88	1.007
third point	4980	19.6	1.95	1.92	1.015
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	7.82	7.81	1.001
Average Correction Factor					1.007
Baseline Corr AF:	7.67	Prev response	7.81	*% change	-1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999330	1.000279
THC Cal Offset:	-0.015543	-0.027945
CH ₄ Cal Slope:	1.000443	1.002623
CH ₄ Cal Offset:	-0.009902	-0.020104
NMHC Cal Slope:	0.999120	0.998214
NMHC Cal Offset:	-0.003636	-0.007641

Notes: Hydrogen and nitrogen cylinder changed. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 10, 2023	Previous Calibration:	September 13, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	9:20	End Time (MST):	12:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995			
16.64	16.63	1.0004						
8.32	8.27	1.0059				Slope	1.000279	0.90 - 1.10
4.15	4.10	1.0118						
			Intercept	-0.027945	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 10, 2023	Previous Calibration:	September 13, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	9:20	End Time (MST):	12:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999966	≥0.995
7.82	7.83	0.9983			
3.91	3.88	1.0074			
1.95	1.92	1.0152			
			Slope	1.002623	0.90 - 1.10
			Intercept	-0.020104	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 10, 2023	Previous Calibration:	September 13, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	9:20	End Time (MST):	12:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
8.82	8.80	1.0022						
4.41	4.39	1.0044				Slope	0.998214	0.90 - 1.10
2.20	2.18	1.0089						
			Intercept	-0.007641	± 0.5			



NMHC Calibration Plot

Date: October 10, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: October 18, 2023
Start time (MST): 6:20
Reason: Routine
Station number: AMS04
Last Cal Date: September 8, 2023
End time (MST): 11:39

Calibration Standards

NO Gas Cylinder #: T36RH1F
NOX Cal Gas Conc: 51.16 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 51.16 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: August 18, 2023
NO Cal Gas Conc: 50.91 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.91 ppm
NO gas Diff:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 721

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.164	1.180	NO bkgnd or offset:	-0.6	-0.6
NOX coeff or slope:	1.159	1.178	NOX bkgnd or offset:	-0.3	-0.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.4	4.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001176	1.000587
NO _x Cal Offset:	0.046144	-0.272971
NO Cal Slope:	1.001222	0.998332
NO Cal Offset:	-0.434177	-0.653356
NO ₂ Cal Slope:	0.996415	1.005579
NO ₂ Cal Offset:	-1.016897	-0.636037



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.3	----	----
as found span	4922	78.1	799.1	795.2	3.9	785.7	780.6	5.1	1.0171	1.0187
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.6	-0.3	----	----
high point	4922	78.1	799.1	795.2	3.9	799.9	794.0	5.8	0.9990	1.0015
second point	4961	39.1	400.1	398.1	2.0	398.8	395.7	3.2	1.0032	1.0061
third point	4981	19.5	199.5	198.5	1.0	199.4	196.6	2.9	1.0005	1.0098
as left zero	5000	0.0	0.0	0.0	0.0	1.0	1.2	-0.2	----	----
as left span	4922	78.1	799.1	383.1	416.0	793.6	380.1	413.6	1.0069	1.0079
Average Correction Factor									1.0009	1.0058

Corrected As found	NO _x = 785.3 ppb	NO = 780.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.9%	
Previous Response	NO _x = 800.1 ppb	NO = 795.7 ppb		*Percent Change	NO = -2.0%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	792.3	380.2	416.0	417.9	0.9955	100.5%
2nd GPT point (200 ppb O ₃)	792.3	582.3	213.9	214.2	0.9986	100.1%
3rd GPT point (100 ppb O ₃)	792.3	686.0	110.2	109.9	1.0028	99.7%
Average Correction Factor					0.9990	100.1%

Notes:

No Maintenance Done. Span adjusted.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 8, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:20	End Time (MST):	11:39
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.1	799.9	0.9990		
400.1	398.8	1.0032		
199.5	199.4	1.0005		
			0.999994	
			1.000587	
			-0.272971	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 8, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:20	End Time (MST):	11:39
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.6	----	Correlation Coefficient	≥0.995	
795.2	794.0	1.0015			
398.1	395.7	1.0061			
198.5	196.6	1.0098			
			Slope	0.998332	0.90 - 1.10
			Intercept	-0.653356	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 8, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:20	End Time (MST):	11:39
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
416.0	417.9	0.9955		
213.9	214.2	0.9986		
110.2	109.9	1.0028		
			0.999997	
			1.005579	
			-0.636037	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 18, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: October 10, 2023 Last Cal Date: September 13, 2023
 Start time (MST): 7:00 End time (MST): 9:22
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 2961
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997486	0.995171	Backgd or Offset:	-2.6	-2.6
Calibration intercept:	-0.460000	0.820000	Coeff or Slope:	1.008	1.008

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.8	----
as found span	5000	986.7	400.0	396.9	1.008
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.4	----
high point	5000	986.9	400.0	398.5	1.004
second point	5000	817.4	200.0	200.6	0.997
third point	5000	707.7	100.0	100.4	0.996
as left zero	5000	0.0	0.0	0.8	----
as left span	5000	989.6	400.0	398.8	1.003
Average Correction Factor					0.999

Baseline Corr As found:	397.7	Previous response	398.5	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments or maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 10, 2023	Previous Calibration:	September 13, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:00	End Time (MST):	9:22
Analyzer make:	API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999990	
400.0	398.5	1.0038			≥0.995
200.0	200.6	0.9970	Slope	0.995171	
100.0	100.4	0.9960			0.90 - 1.10
			Intercept	0.820000	+/- 5



O₃ Calibration Plot

Date: October 10, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: October 18, 2023 Last Cal Date: September 13, 2023
 Start time (MST): 6:54 End time (MST): 8:07

Analyzer Make: API T640 S/N: 322
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Deltacal S/N: 1451
 Temp/RH standard: Deltacal S/N: 1451

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	4	4	4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	722.8	724.4	722.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.03	4.92	5.03	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>October 18, 2023</u>	Last Cal Date: <u>September 13, 2023</u>			
	PM w/o HEPA: <u>24</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10	10	10.7	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>12.8</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>October 18, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>October 18, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: October 18, 2023
 Date RH/T Sensor Cleaned: May 23, 2023

PMT adjusted. No other adjustments done. Head cleaned.

Notes:

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS05
MANNIX
OCTOBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 15, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:05	End Time (MST):	12:45
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.8	----	Correlation Coefficient	0.999999	≥0.995
800.3	796.7	1.0045			
400.2	399.0	1.0029	Slope	0.994288	0.90 - 1.10
200.1	200.3	0.9989			
			Intercept	1.060000	+/-30



SO2 Calibration Plot

Date: October 4, 2023

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mannix Station number: AMS05
 Calibration Date: October 16, 2023 Last Cal Date: September 18, 2023
 Start time (MST): 9:33 End time (MST): 14:18
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.92 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002433
 Removed Cal Gas Conc: 4.92 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1845
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1203169745
 Converter make: Global Converter serial #: 2022-196
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994760	0.970189	Backgd or Offset: 2.18	2.18
Calibration intercept:	0.380518	0.340566	Coeff or Slope: 0.866	0.866

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4919	81.3	80.0	81.0	0.989
as found 2nd point	4960	40.7	40.0	40.4	0.994
as found 3rd point	4980	20.3	20.0	19.7	1.019
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4919	81.3	80.0	77.9	1.027
second point	4960	40.7	40.0	39.3	1.019
third point	4980	20.3	20.0	19.7	1.014
as left zero	5000	0.0	0.0	0.9	----
as left span	4919	81.3	80.0	76.4	1.047
SO2 Scrubber Check	4920	80.0	800.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	1.020
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 80.9 Prev response: 79.96 *% change: 1.2%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 1.013614 AF Intercept: -0.179437
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999938

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. There is 3ppb of difference between as found and calibrator span, suspecting this is caused by the zero-air drying out the scrubber beads, this is why no adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 16, 2023	Previous Calibration:	September 18, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:33	End Time (MST):	14:18
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999995	≥0.995
80.0	77.9	1.0269			
40.0	39.3	1.0189	Slope	0.970189	0.90 - 1.10
20.0	19.7	1.0139			
			Intercept	0.340566	+/-3



H₂S Calibration Plot

Date: October 16, 2023

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	October 4, 2023	Last Cal Date:	September 15, 2023
Start time (MST):	9:04	End time (MST):	12:45
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.62E-05	2.64E-05	NMHC SP Ratio:	4.44E-05	4.50E-05
CH ₄ Retention time:	15.00	15.00	NMHC Peak Area:	206221	203435
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.04	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.23	1.000
second point	4960	40.0	8.61	8.61	1.000
third point	4980	20.0	4.31	4.31	0.999
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.39	0.991
Average Correction Factor					1.000

Baseline Corr AF:	17.04	Prev response	17.07	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	9.15	9.03	1.013
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.14	1.001
second point	4960	40.0	4.57	4.60	0.995
third point	4980	20.0	2.29	2.31	0.990
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.15	9.25	0.989
Average Correction Factor					0.995
Baseline Corr AF:	9.03	Prev response	9.07	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	8.00	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.08	0.999
second point	4960	40.0	4.04	4.02	1.005
third point	4980	20.0	2.02	2.00	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.14	0.993
Average Correction Factor					1.005
Baseline Corr AF:	8.00	Prev response	8.00	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.991126	0.999944
THC Cal Offset:	0.001000	0.001800
CH ₄ Cal Slope:	0.992078	1.001075
CH ₄ Cal Offset:	-0.010800	-0.012600
NMHC Cal Slope:	0.990661	0.998844
NMHC Cal Offset:	0.011800	0.014800

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 15, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:04	End Time (MST):	12:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995			
17.23	17.23	0.9999						
8.61	8.61	1.0001				Slope	0.999944	0.90 - 1.10
4.31	4.31	0.9987						
			Intercept	0.001800	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 15, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:04	End Time (MST):	12:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999987	≥0.995
8.08	8.08	0.9994			
4.04	4.02	1.0053			
2.02	2.00	1.0093			
			Slope	1.001075	0.90 - 1.10
			Intercept	-0.012600	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 15, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:04	End Time (MST):	12:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999987	≥ 0.995			
9.15	9.14	1.0005						
4.57	4.60	0.9954				Slope	0.998844	0.90 - 1.10
2.29	2.31	0.9896						
			Intercept	0.014800	± 0.5			



NMHC Calibration Plot

Date: October 4, 2023

Location: Mannix





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	October 19, 2023	Last Cal Date:	September 14, 2023
Start time (MST):	8:52	End time (MST):	12:45
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.78	ppm	Cal Gas Exp Date:	September 9, 2024
Cal Gas Cylinder #:	AAL070632			
Removed Cal Gas Conc:	49.78	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	5608

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000605	0.998819	Backgd or Offset:	17.4	17.4
Calibration intercept:	1.660053	2.459939	Coeff or Slope:	0.911	0.911

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.3	799.5	797.5	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4920	80.3	799.5	799.8	1.000
second point	4960	40.2	400.2	403.5	0.992
third point	4980	20.1	200.1	204.5	0.979
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	80.3	799.5	800.9	0.998
Average Correction Factor					0.990

Baseline Corr As found:	797.50	Previous response	801.61	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 14, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:52	End Time (MST):	12:45
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999963	
799.5	799.8	0.9996			≥0.995
400.2	403.5	0.9919	Slope	0.998819	
200.1	204.5	0.9786			0.90 - 1.10
			Intercept	2.459939	+/-30



SO2 Calibration Plot

Date: October 19, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	October 13, 2023	Last Cal Date:	September 22, 2023
Start time (MST):	9:25	End time (MST):	13:29
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.38	ppm	Cal Gas Exp Date:	March 2, 2023
Cal Gas Cylinder #:	EY0000809			
Removed Cal Gas Conc:	5.38	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004636	0.995918	Backgd or Offset:	2.05	2.05
Calibration intercept:	0.217058	0.297166	Coeff or Slope:	1.189	1.189

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.3	79.9	78.6	1.017
as found 2nd point	4963	37.2	40.0	39.5	1.013
as found 3rd point	4981	18.6	20.0	19.9	1.006
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	74.3	79.9	79.8	1.002
second point	4963	37.2	40.0	40.3	0.993
third point	4981	18.6	20.0	20.4	0.981
as left zero	5000	0.0	0.0	0.2	----
as left span	4926	74.3	79.9	80.7	0.991
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	0.992
Date of last converter efficiency test:	efficiency				

Baseline Corr As found:	78.6	Prev response:	80.53	*% change:	-2.5%
Baseline Corr 2nd AF pt:	39.5	AF Slope:	0.982477	AF Intercept:	0.117533
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999990		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 13, 2023	Previous Calibration:	September 22, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:25	End Time (MST):	13:29
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999971	≥0.995
79.9	79.8	1.0018			
40.0	40.3	0.9932	Slope	0.995918	0.90 - 1.10
20.0	20.4	0.9811			
			Intercept	0.297166	+/-3



TRS Calibration Plot

Date: October 13, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	October 19, 2023	Last Cal Date:	September 14, 2023
Start time (MST):	8:52	End time (MST):	12:45
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH ₄ Cal Gas Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.12E-04	2.15E-04	NMHC SP Ratio:	4.80E-05
CH ₄ Retention time:	14.0	14.0	NMHC Peak Area:	189064
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				186260
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4920	80.3	17.12	16.97	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	17.15	0.998
second point	4960	40.2	8.57	8.61	0.996
third point	4980	20.1	4.29	4.37	0.980
as left zero	5000	0.0	0.00	0.01	----
as left span	4920	80.3	17.12	17.20	0.995
Average Correction Factor					0.991

Baseline Corr AF:	16.96	Prev response	17.17	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919.7	80.3	9.07	9.00	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	9.07	9.09	0.998
second point	4960	40.2	4.54	4.59	0.990
third point	4980	20.1	2.27	2.35	0.967
as left zero	5000	0.0	0.00	0.01	----
as left span	4919.7	80.3	9.07	9.13	0.994
Average Correction Factor					0.985
Baseline Corr AF:	8.98	Prev response	9.11	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919.7	80.3	8.06	7.97	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	8.06	8.06	0.999
second point	4960	40.2	4.03	4.02	1.003
third point	4980	20.1	2.02	2.03	0.994
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	8.06	8.08	0.997
Average Correction Factor					0.999
Baseline Corr AF:	7.97	Prev response	8.06	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000464	1.000103
THC Cal Offset:	0.038421	0.036627
CH ₄ Cal Slope:	1.000059	0.999903
CH ₄ Cal Offset:	0.001844	0.001645
NMHC Cal Slope:	1.000823	1.000180
NMHC Cal Offset:	0.036578	0.035381

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 14, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:52	End Time (MST):	12:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999975	≥ 0.995			
17.12	17.15	0.9984						
8.57	8.61	0.9960				Slope	1.000103	0.90 - 1.10
4.29	4.37	0.9799						
			Intercept	0.036627	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 14, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:52	End Time (MST):	12:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999993	≥ 0.995
8.06	8.06	0.9995			
4.03	4.02	1.0027			
2.02	2.03	0.9943			
			Slope	0.999903	0.90 - 1.10
			Intercept	0.001645	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 14, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:52	End Time (MST):	12:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999928	≥ 0.995
9.07	9.09	0.9976			
4.54	4.59	0.9900			
2.27	2.35	0.9674			
			Slope	1.000180	0.90 - 1.10
			Intercept	0.035381	+/-0.5



NMHC Calibration Plot

Date: October 19, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Patricia McInnes
Calibration Date: October 12, 2023
Start time (MST): 8:50
Reason: Routine
Station number: AMS06
Last Cal Date: September 20, 2023
End time (MST): 13:26

Calibration Standards

NO Gas Cylinder #: T30YCWN
NOX Cal Gas Conc: 47.94 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 47.94 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: April 11, 2025
NO Cal Gas Conc: 46.39 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 46.39 ppm
NO gas Diff:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1172750022

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.825	0.825	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.987	0.987	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.2	157.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997367	1.000493
NO _x Cal Offset:	2.676082	2.095627
NO Cal Slope:	0.999173	1.001945
NO Cal Offset:	1.382870	0.962538
NO ₂ Cal Slope:	1.001949	1.003982
NO ₂ Cal Offset:	0.485373	-0.085539



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.3	-0.1	----	----
as found span	4914	86.2	826.5	799.7	26.7	827.6	798.6	29.0	0.9986	1.0014
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.4	-0.1	----	----
high point	4914	86.2	826.5	799.7	26.7	828.1	802.0	26.3	0.9980	0.9972
second point	4957	43.1	413.2	399.9	13.4	416.3	401.8	14.5	0.9926	0.9952
third point	4978	21.6	207.1	200.4	6.7	211.2	202.4	8.8	0.9807	0.9902
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.5	0.0	----	----
as left span	4914	86.2	826.5	400.6	425.8	824.0	405.2	418.7	1.0030	0.9887
Average Correction Factor									0.9904	0.9942

Corrected As found	NO _x = 827.5 ppb	NO = 798.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.1%
Previous Response	NO _x = 827.0 ppb	NO = 800.5 ppb		*Percent Change	NO = -0.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	801.1	402.0	425.8	427.4	0.9963	100.4%
2nd GPT point (200 ppb O3)	801.1	604.7	223.1	224.0	0.9961	100.4%
3rd GPT point (100 ppb O3)	801.1	703.4	124.4	124.8	0.9970	100.3%
Average Correction Factor					0.9964	100.4%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 20, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:50	End Time (MST):	13:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.5	828.1	0.9980		
413.2	416.3	0.9926		
207.1	211.2	0.9807		
			0.999978	
			1.000493	
			2.095627	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 20, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:50	End Time (MST):	13:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999998	≥0.995
799.7	802.0	0.9972			
399.9	401.8	0.9952	Slope	1.001945	0.90 - 1.10
200.4	202.4	0.9902			
			Intercept	0.962538	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 20, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	8:50	End Time (MST):	13:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
425.8	427.4	0.9963		
223.1	224.0	0.9961		
124.4	124.8	0.9970		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 12, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Patricia McInnes Station number: AMS06
 Calibration Date: October 2, 2023 Last Cal Date: September 5, 2023
 Start time (MST): 9:03 End time (MST): 12:50
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3566
 ZAG Make/Model: API T701 Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1300156234
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005743	1.003714	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	-0.480000	-2.700000	Coeff or Slope:	1.019	1.039

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	-0.7	----
as found span	5000	1303.0	400.0	391.8	1.021
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	-0.3	----
high point	5000	1303.0	400.0	400.0	1.000
second point	5000	966.5	200.0	196.8	1.016
third point	5000	794.3	100.0	95.3	1.049
as left zero	5000	800.0	0.0	-0.8	----
as left span	5000	1303.0	400.0	404.1	0.990
Average Correction Factor					1.022

Baseline Corr As found:	392.5	Previous response	401.8	*% change	-2.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 2, 2023	Previous Calibration:	September 5, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:03	End Time (MST):	12:50
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999837	≥0.995
400.0	400.0	1.0000			
200.0	196.8	1.0163	Slope	1.003714	0.90 - 1.10
100.0	95.3	1.0493			
			Intercept	-2.700000	+/- 5



O₃ Calibration Plot

Date: October 2, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: October 19, 2023 Last Cal Date: September 22, 2023
 Start time (MST): 13:05 End time (MST): 13:47

Analyzer Make: API T640 S/N: 766
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT FP-25 S/N: 388755
 Temp/RH standard: ALICAT FP-25 S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	17.2	16.8	17.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	716.1	717.5	716.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.05	4.82	5.05	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: October 19, 2023 Last Cal Date: September 22, 2023
 PM w/o HEPA: 3.3 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>4.7</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>September 22, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 22, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023
 Date RH/T Sensor Cleaned: April 13, 2023

PMT peak test was completed in September. Leak check passed, no adjustments made.

Notes:

Calibration by: Max Farrell



Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	October 3, 2023	Last Cal Date:	September 12, 2023
Start time (MST):	8:30	End time (MST):	12:45
NH3 Cal Date:	October 3, 2023	Last Cal Date:	September 12, 2023
Start time (MST):	12:50	End time (MST):	14:55
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	NA
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	77.8	ppm	NH3 Gas Cylinder #:	CC710812
			NH3 Cal Gas Expiry:	March 30, 2023
Removed NH3 Conc:	77.8	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	API T700		Serial Number:	3566
ZAG make/model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	152
Converter model:	Teledyne API T501	Converter serial #:	147
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.50
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	542

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.788	0.788	TN coefficient:	0.789	0.789
NOX coefficient:	0.787	0.787	NO bkgnd:	-0.1	-0.1
NO2 coefficient:	1.000	1.000	NOX bkgnd:	0.0	0.0
NH3 coefficient:	0.908	0.908	TN bkgnd:	0.0	0.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000369	1.004061
NO _x Cal Offset:	1.915538	0.254959
NO Cal Slope:	1.000758	1.000231
NO Cal Offset:	0.303018	1.162694
NO ₂ Cal Slope:	1.006122	0.998161
NO ₂ Cal Offset:	-2.141943	-0.986846
NH3 Cal Slope:	0.985421	0.995038
NH3 Cal Offset:	7.387990	8.227208
TN Cal Slope:	0.991067	1.001184
TN Cal Offset:	7.203693	7.279855



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.4	-0.4	0.8	----	----
as found NO	4914	86.2	826.5	826.5	----	833.3	829.5	4.1	0.992	----
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	-0.9	1.0	----	----
high NO point	4914	86.2	826.5	826.5	----	829.4	829.8	-0.5	0.996	----
NO/O3 point	4914	86.2	826.5	826.5	----	830.2	826.4	3.8	0.995	----
as found NH3	3419	81.0	1800.5	----	1800.5	1804.8	----	1794.7	0.998	1.003
new NH3 cyl rp							----			
first NH3	3419	81.0	1800.5	----	1800.5	1804.8	----	1794.7	0.998	1.003
second NH3	3455	45.0	1000.3	----	1000.3	1014.9	----	1009.6	0.986	0.991
third NH3	3478	22.5	500.1	----	500.1	514.1	----	512.1	0.973	0.977
Average Correction Factor									0.9960	0.9902

Corrected As found TN = 832.9 ppb NO_x = 829.9 ppb NH3 = 1793.9 ppb

Previous Response TN = 826.3 ppb NO_x = 828.7 ppb NH3 = 1781.7 ppb

NH3 Previous Converter Efficiency = 90.8%

NH3 Current Converter Efficiency = 90.8%

*Percent Change TN = 0.8%

*Percent Change NO_x = 0.1%

*Percent Change NH3 = 0.7%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.4	0.4	----	----
as found span	4914	86.2	826.5	799.7	826.5	829.5	798.8	833.3	0.9963	1.0012
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.9	-0.5	0.1	----	----
high point	4914	86.2	826.5	799.7	826.5	829.8	799.9	829.4	0.9960	0.9998
second point	4957	43.1	413.2	399.9	413.2	414.8	403.0	418.0	0.9962	0.9922
third point	4978	21.6	207.1	200.4	207.1	210.0	202.6	211.7	0.9863	0.9892
Average Correction Factor									0.9928	0.9938

Baseline Corr As fnd	TN = 832.9 ppb	NO _x = 829.9 ppb	NO = 799.2 ppb	*Percent Change	TN = 0.8%
Previous Response	TN = 826.3 ppb	NO _x = 828.7 ppb	NO = 800.6 ppb	*Percent Change	NO _x = 0.1%
				*Percent Change	NO = -0.2%
				<i>* = > +/-5% change initiates investigation</i>	

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	0.0	----	----
calibration zero	----	----	0.0	-0.4	----	----
1st GPT point (400 ppb O3)	800.8	403.8	423.7	422.5	1.0029	99.7%
2nd GPT point (200 ppb O3)	800.8	605.6	221.9	219.6	1.0106	99.0%
3rd GPT point (100 ppb O3)	800.8	703.6	123.9	122.5	1.0116	98.9%
Average Correction Factor					1.0084	99.2%

Notes:

Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By:

Max Farrell



Wood Buffalo Environmental Association

TN Calibration Summary


Version-05-2023

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 12, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:30	End Time (MST):	12:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
1800.5	1804.8	0.9976		
1000.3	1014.9	0.9856		
500.1	514.1	0.9727		





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 12, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:30	End Time (MST):	12:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	1.0	----	Correlation Coefficient	≥0.995	
1800.5	1794.7	1.0032			
1000.3	1009.6	0.9908			
500.1	512.1	0.9765			
			Slope	0.995038	0.90 - 1.10
			Intercept	8.227208	+/-20





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 12, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:30	End Time (MST):	12:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.9	----	Correlation Coefficient	≥0.995	
826.5	829.8	0.9960			
413.2	414.8	0.9962			
207.1	210.0	0.9863			
			Slope	1.004061	0.90 - 1.10
			Intercept	0.254959	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 12, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:30	End Time (MST):	12:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.5	----	Correlation Coefficient	≥0.995	
799.7	799.9	0.9998			
399.9	403.0	0.9922			
200.4	202.6	0.9892			
			Slope	1.000231	0.90 - 1.10
			Intercept	1.162694	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 12, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:30	End Time (MST):	12:45
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
423.7	422.5	1.0029		
221.9	219.6	1.0106		
123.9	122.5	1.0116		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 3, 2023


Location: Patricia McInnes



NH₃ Calibration Plot

Date: October 3, 2023

Location: Patricia McInnes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	October 19, 2023	Last Cal Date:	September 6, 2023
Start time (MST):	9:10	End time (MST):	12:56
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.52	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC282115			
Removed Cal Gas Conc:	50.52	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 701H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992516	1.001786	Backgd or Offset:	2.61	2.61
Calibration intercept:	1.942886	2.644970	Coeff or Slope:	0.834	0.834

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	79.3	801.2	804.3	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	79.3	801.2	803.3	0.997
second point	4960	39.6	400.2	407.0	0.983
third point	4980	19.8	200.1	204.0	0.981
as left zero	5000	0.0	0.0	0.1	----
as left span	4921	79.2	800.2	804.8	0.994
Average Correction Factor					0.987

Baseline Corr As found:	804.40	Previous response	797.15	*% change	0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 6, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:10	End Time (MST):	12:56
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999935	
801.2	803.3	0.9974			≥0.995
400.2	407.0	0.9832	Slope	1.001786	
200.1	204.0	0.9807			0.90 - 1.10
			Intercept	2.644970	+/-30



SO2 Calibration Plot

Date: October 19, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: October 10, 2023 Last Cal Date: September 18, 2023
 Start time (MST): 9:46 End time (MST): 14:53
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.25 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC504080
 Removed Cal Gas Conc: 5.25 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3805
 ZAG Make/Model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE Analyzer serial #: 1180540018
 Converter make: CDN-101 Converter serial #: 551
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.012468	1.014626	Backgd or Offset: 2.33	2.33
Calibration intercept:	-0.002341	-0.002205	Coeff or Slope: 0.899	0.899

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4925	75.5	79.3	80.3	0.987
as found 2nd point	4962	37.7	39.6	40.0	0.990
as found 3rd point	4981	18.9	19.8	20.1	0.987
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4925	75.5	79.3	80.5	0.985
second point	4962	37.7	39.6	40.2	0.985
third point	4981	18.9	19.9	20.0	0.993
as left zero	5000	0.0	0.0	0.1	----
as left span	4925	75.5	79.3	80.5	0.985
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	0.988
Date of last converter efficiency test:	April 22, 2022			92.6% efficiency	

Baseline Corr As found: 80.3 Prev response: 80.25 *% change: 0.1%
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 1.012901 AF Intercept: -0.022289
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999998

* = > +/-5% change initiates investigation

Notes: Pump changed out after as founds. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 10, 2023	Previous Calibration:	September 18, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:46	End Time (MST):	14:53
Analyzer make:	CDN-101	Analyzer serial #:	551

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999991	
79.3	80.5	0.9852			≥0.995
39.6	40.2	0.9853	Slope	1.014626	
19.9	20.0	0.9928			0.90 - 1.10
			Intercept	-0.002205	+/-3



TRS Calibration Plot

Date: October 10, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	October 19, 2023	Last Cal Date:	September 6, 2023
Start time (MST):	9:10	End time (MST):	12:56
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>	
CH ₄ SP Ratio:	0.000249	0.000247	NMHC SP Ratio:	4.17E-05	4.27E-05
CH ₄ Retention time:	14.0	13.8	NMHC Peak Area:	218242	212776
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.83	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.02	1.001
second point	4960	39.6	8.52	8.54	0.998
third point	4980	19.8	4.26	4.27	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.03	1.000
Average Correction Factor					0.998

Baseline Corr AF:	16.83	Prev response	17.03	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	9.09	8.87	1.025
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	9.09	9.08	1.001
second point	4960	39.6	4.55	4.59	0.990
third point	4980	19.8	2.27	2.32	0.980
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	9.09	9.10	0.999
Average Correction Factor					0.990
Baseline Corr AF:	8.87	Prev response	9.10	*% change	-2.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.94	7.96	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.94	7.94	1.000
second point	4960	39.6	3.97	3.94	1.007
third point	4980	19.8	1.98	1.95	1.017
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.94	7.92	1.002
Average Correction Factor					1.008
Baseline Corr AF:	7.96	Prev response	7.93	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999300	0.999306
THC Cal Offset:	0.010060	0.011259
CH ₄ Cal Slope:	1.001057	1.001244
CH ₄ Cal Offset:	-0.015758	-0.019158
NMHC Cal Slope:	0.997865	0.997363
NMHC Cal Offset:	0.025418	0.030417

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 6, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:10	End Time (MST):	12:56
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
17.03	17.02	1.0005						
8.52	8.54	0.9976				Slope	0.999306	0.90 - 1.10
4.26	4.27	0.9969						
			Intercept	0.011259	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 6, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:10	End Time (MST):	12:56
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999973	≥0.995
7.94	7.94	0.9998			
3.97	3.94	1.0068			
1.98	1.95	1.0173			
			Slope	1.001244	0.90 - 1.10
			Intercept	-0.019158	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 6, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:10	End Time (MST):	12:56
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999944	≥ 0.995			
9.09	9.08	1.0014						
4.55	4.59	0.9902				Slope	0.997363	0.90 - 1.10
2.27	2.32	0.9797						
			Intercept	0.030417	± 0.5			



NMHC Calibration Plot

Date: October 19, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as found span	4920	80.2	816.7	800.7	16.0	816.7	799.2	17.5	1.0000	1.0019
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
high point	4920	80.2	816.7	800.7	16.0	817.1	801.8	15.2	0.9995	0.9986
second point	4960	40.1	408.4	400.4	8.0	411.4	403.7	7.7	0.9926	0.9917
third point	4980	20.0	203.7	199.7	4.0	205.4	201.1	4.3	0.9916	0.9929
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4920	80.2	816.7	398.6	418.1	822.4	403.3	419.2	0.9931	0.9883
Average Correction Factor									0.9946	0.9944

Corrected As found	NO _x = 816.7 ppb	NO = 799.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.2%
Previous Response	NO _x = 818.3 ppb	NO = 802.2 ppb		*Percent Change	NO = -0.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	800.4	398.3	418.1	418.5	0.9991	100.1%
2nd GPT point (200 ppb O3)	800.4	595.5	220.9	219.3	1.0075	99.3%
3rd GPT point (100 ppb O3)	800.4	698.8	117.6	116.9	1.0063	99.4%
Average Correction Factor					1.0043	99.6%

Notes:

No adjustments needed.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 6, 2023	Previous Calibration:	September 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:38	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
816.7	817.1	0.9995			
408.4	411.4	0.9926			
203.7	205.4	0.9916			
			Slope	1.000077	0.90 - 1.10
			Intercept	1.278990	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 6, 2023	Previous Calibration:	September 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:38	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
800.7	801.8	0.9986		
400.4	403.7	0.9917		
199.7	201.1	0.9929		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 6, 2023	Previous Calibration:	September 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:38	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
418.1	418.5	0.9991		
220.9	219.3	1.0075		
117.6	116.9	1.0063		
			0.999976	
			1.000884	
			-0.671765	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 6, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: October 3, 2023 Last Cal Date: September 20, 2023
 Start time (MST): 10:17 End time (MST): 14:05
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: T700 Serial Number: 3805
 ZAG Make/Model: T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1152220023
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997143	1.004229	Backgd or Offset:	-2.6	-2.6
Calibration intercept:	0.600000	0.060000	Coeff or Slope:	1.522	1.535

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-1.8	----
as found span	5000	1414.8	400.0	397.5	1.006
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.7	----
high point	5000	1415.7	400.0	401.5	0.996
second point	5000	1039.9	200.0	200.9	0.996
third point	5000	856.2	100.0	101.5	0.985
as left zero	5000	0.0	0.0	-0.6	----
as left span	5000	1416.0	400.0	404.4	0.989
Average Correction Factor					0.992

Baseline Corr As found:	399.3	Previous response	399.5	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Athabasca Valley Station number: AMS 07
 Calibration Date: October 19, 2023 Last Cal Date: September 27, 2023
 Start time (MST): 12:16 End time (MST): 12:54

Analyzer Make: API T640 S/N: 871
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	18.0	17.4	18	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	725.3	727	725.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	5.05	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>October 19, 2023</u>	Last Cal Date: <u>September 27, 2023</u>			
	PM w/o HEPA: <u>2.2</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>NA</u>	w/ HEPA: <u>NA</u>		
Date Optical Chamber Cleaned:		<u>September 27, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>August 28, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022
 Date RH/T Sensor Cleaned: December 5, 2022

Notes: Temp, flow and pressure checked. Leak check passed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 20, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:17	End Time (MST):	14:05
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.7	----	Correlation Coefficient	0.999981	≥0.995
400.0	401.5	0.9963			
200.0	200.9	0.9955	Slope	1.004229	0.90 - 1.10
100.0	101.5	0.9852			
			Intercept	0.060000	+/- 5



O₃ Calibration Plot

Date: October 3, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Athabasca Valley Station number: AMS 07
 Calibration Date: October 19, 2023 Last Cal Date: September 27, 2023
 Start time (MST): 12:16 End time (MST): 12:54

Analyzer Make: API T640 S/N: 871
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	18.0	17.4	18	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	725.3	727	725.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	5.05	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>October 19, 2023</u>	Last Cal Date: <u>September 27, 2023</u>			
	PM w/o HEPA: <u>2.2</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>NA</u>	w/ HEPA: <u>NA</u>		
Date Optical Chamber Cleaned:		<u>September 27, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>August 28, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022
 Date RH/T Sensor Cleaned: December 5, 2022

Notes:

Temp, flow and pressure checked. Leak check passed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	October 13, 2023	Last Cal Date:	September 27, 2023
Start time (MST):	10:15	End time (MST):	13:21
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date:	December 12, 2026
Cal Gas Cylinder #:	LL66942			
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 700H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003398	0.998346	Backgd or Offset:	4.225	4.336
Calibration intercept:	0.092559	0.098508	Coeff or Slope:	1.086	1.093

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.0	40.1	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.0	40.0	1.001
second point	4967	33.3	20.0	20.1	0.992
third point	4983	16.7	10.0	10.2	0.983
as left zero	5000	0.0	0.0	0.0	----
as left span	4933	66.7	40.0	40.0	1.002
Average Correction Factor					0.992

Baseline Corr As found:	39.99	Prev response:	40.25	*% change:	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 13, 2023	Previous Calibration:	September 27, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:15	End Time (MST):	13:21
Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999965	≥0.995
40.0	40.0	1.0008			
20.0	20.1	0.9925	Slope	0.998346	0.90 - 1.10
10.0	10.2	0.9834			
			Intercept	0.098508	+/-1.5



CO Calibration Plot

Date: October 13, 2023

Location: Athabasca Valley





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	October 11, 2023	Last Cal Date:	September 6, 2023
Start time (MST):	10:57	End time (MST):	13:42
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC196697			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3060
ZAG Make/Model:	Teledyne API T701		Serial Number:	260

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004228	1.003114	Backgd or Offset:	4.64	4.65
Calibration intercept:	-1.764623	-2.624243	Coeff or Slope:	0.963	0.955

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-3.0	----
as found span	4920	80.3	800.4	804.2	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-2.9	----
high point	4920	80.3	800.4	800.8	0.999
second point	4960	40.2	400.7	397.4	1.008
third point	4980	20.1	200.4	200.0	1.002
as left zero	5000	0.0	0.0	-3.0	----
as left span	4920	80.3	800.4	801.1	0.999
Average Correction Factor					1.003

Baseline Corr As found:	807.20	Previous response	802.00	*% change	0.6%
-------------------------	--------	-------------------	--------	-----------	------

* = > +/-5% change initiates investigation

Notes: changed inlet filters after as founds. Adjustment made to span.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 11, 2023	Previous Calibration:	September 6, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	10:57	End Time (MST):	13:42
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-2.9	----	Correlation Coefficient	0.999981	≥0.995
800.4	800.8	0.9995			
400.7	397.4	1.0083	Slope	1.003114	0.90 - 1.10
200.4	200.0	1.0018			
			Intercept	-2.624243	+/-30



SO2 Calibration Plot

Date: October 11, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: October 12, 2023 Last Cal Date: September 21, 2023
 Start time (MST): 12:44 End time (MST): 16:40
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002276
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 260

Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 1203169744
 Converter make: CDN-101 Converter serial #: 14639
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011429	1.016423	Backgd or Offset:	1.00	0.99
Calibration intercept:	0.338843	0.539126	Coeff or Slope:	0.752	0.752

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4920	80.5	80.0	80.3	1.001
as found 2nd point	4960	40.2	40.0	40.6	0.994
as found 3rd point	4980	20.1	20.0	20.5	0.994
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4920	80.5	80.0	81.8	0.978
second point	4960	40.2	40.0	41.2	0.970
third point	4980	20.1	20.0	21.0	0.951
as left zero	5000	0.0	0.0	0.5	----
as left span	4920	80.5	80.0	82.3	0.972
SO2 Scrubber Check	4919.7	80.3	803.0	0.1	----
Date of last scrubber change:	March 7, 2022			Ave Corr Factor	0.966
Date of last converter efficiency test:	March 15, 2022			100.7% efficiency	

Baseline Corr As found: 79.9 Prev response: 81.26 *% change: -1.7%
 Baseline Corr 2nd AF pt: 40.2 AF Slope: 0.998424 AF Intercept: 0.518826
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999983

* = > +/-5% change initiates investigation

Notes: Sampled inlet filter changed after as founds. Scrubber check passed.

Calibration Performed By: Matthew Courtoreille



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 21, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	12:44	End Time (MST):	16:20
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999986	≥0.995
80.0	81.8	0.9781			
40.0	41.2	0.9698	Slope	1.016423	0.90 - 1.10
20.0	21.0	0.9514			
			Intercept	0.539126	+/-3



TRS Calibration Plot

Date: October 12, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
Calibration Date: October 12, 2023 Last Cal Date: September 21, 2023
Start time (MST): 8:00 End time (MST): 12:40
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC363447 Cal Gas Expiry Date: February 2, 2024
NOX Cal Gas Conc: 48.80 ppm NO Cal Gas Conc: 48.80 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.80 ppm Removed Gas NO Conc: 48.80 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3060
ZAG make/model: Teledyne API T701H Serial Number: 260

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1426262592
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.078	1.101	NO bkgnd or offset:	9.8	10
NOX coeff or slope:	0.992	0.994	NOX bkgnd or offset:	10.7	11.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	145.7	145.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.993131	0.993517
NO _x Cal Offset:	1.940000	-0.020000
NO Cal Slope:	0.994931	0.991532
NO Cal Offset:	1.260000	-0.300000
NO ₂ Cal Slope:	0.991090	1.008121
NO ₂ Cal Offset:	-1.074507	1.625692



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-1.4	-0.6	-0.8	----	----
as found span	4918	82.0	800.3	800.3	0.0	780.4	779.0	1.4	1.0255	1.0274
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-1.3	-0.4	-0.8	----	----
high point	4918	82.0	800.3	800.3	0.0	793.9	792.6	1.2	1.0081	1.0097
second point	4959	41.0	400.2	400.2	0.0	399.9	398.3	1.4	1.0007	1.0047
third point	4980	20.5	200.1	200.1	0.0	198.9	197.0	1.9	1.0059	1.0156
as left zero	5000	0.0	0.0	0.0	0.0	-1.3	-0.3	-0.9	----	----
as left span	4918	82.0	800.3	401.2	399.1	790.5	395.4	394.9	1.0124	1.0147
Average Correction Factor									1.0049	1.0100

Corrected As found	NO _x = 781.8 ppb	NO = 779.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.9%
Previous Response	NO _x = 796.8 ppb	NO = 797.5 ppb		*Percent Change	NO = -2.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.1	398.0	399.1	402.1	0.9925	100.8%
2nd GPT point (200 ppb O3)	797.1	603.2	193.9	200.5	0.9671	103.4%
3rd GPT point (100 ppb O3)	797.1	700.9	96.2	99.5	0.9668	103.4%
Average Correction Factor					0.9755	102.5%

Notes: sampled inlet filter changed after as founds. Made adjustment to high point span.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 21, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:00	End Time (MST):	12:40
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-1.3	----	Correlation Coefficient	≥0.995	
800.3	793.9	1.0081			
400.2	399.9	1.0007			
200.1	198.9	1.0059			
			Slope	0.993517	0.90 - 1.10
			Intercept	-0.020000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 21, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:00	End Time (MST):	12:40
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.4	----	Correlation Coefficient	0.999986	≥0.995
800.3	792.6	1.0097			
400.2	398.3	1.0047	Slope	0.991532	0.90 - 1.10
200.1	197.0	1.0156			
			Intercept	-0.300000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 21, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:00	End Time (MST):	12:40
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.8	----	Correlation Coefficient	0.999755	≥0.995
399.1	402.1	0.9925			
193.9	200.5	0.9671	Slope	1.008121	0.90 - 1.10
96.2	99.5	0.9668			
			Intercept	1.625692	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 12, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: October 11, 2023 Last Cal Date: September 6, 2023
 Start time (MST): 8:02 End time (MST): 10:58
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 260

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3872
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.018086	1.002143	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	-0.940000	-0.200000	Coeff or Slope:	1.036	1.036

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.9	----
as found span	5000	913.0	400.0	402.2	0.995
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	1.1	----
high point	5000	914.7	400.0	401.3	0.997
second point	5000	786.4	200.0	199.6	1.002
third point	5000	701.3	100.0	98.7	1.013
as left zero	5000	0.0	0.0	0.5	----
as left span	5000	963.3	400.0	400.6	0.999
Average Correction Factor					1.004

Baseline Corr As found:	401.3	Previous response	406.3	*% change	-1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed out inlet filter after as found. No adjustments

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 11, 2023	Previous Calibration:	September 6, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:02	End Time (MST):	10:58
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.1	----	Correlation Coefficient	0.999952	
400.0	401.3	0.9968			≥0.995
200.0	199.6	1.0020	Slope	1.002143	
100.0	98.7	1.0132			0.90 - 1.10
			Intercept	-0.200000	+/- 5



O₃ Calibration Plot

Date: October 11, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
 Calibration Date: October 11, 2023 Last Cal Date: September 21, 2023
 Start time (MST): 9:38 End time (MST): 12:27

Analyzer Make: Teledyne API T640 S/N: 216
 Particulate Fraction: PM2.5

Flow Meter Make/Model: FP-25 S/N: 388747
 Temp/RH standard: FP-25 S/N: 388747

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	12.7	13.7	12.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	730.9	731.0	730.9	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	4.99	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 21, 2023</u>	Last Cal Date: <u>July25,2023</u>			
	PM w/o HEPA: <u>0</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.1	11.0	11.2	<input type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>24.5</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>July 25, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 25, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: July 25, 2023
 Date RH/T Sensor Cleaned: July 25, 2023

Notes: No adjustment made

Calibration by: Morgan Voyageur



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	October 17, 2023	Last Cal Date:	September 22, 2023
Start time (MST):	8:15	End time (MST):	10:47
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API T701H		Serial Number:	260

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000563	1.001383	Backgd or Offset:	-0.014	-0.014
Calibration intercept:	0.028916	0.066908	Coeff or Slope:	1.006	1.006

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.02	----
as found span	4933	66.7	40.4	40.5	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4934	66.7	40.4	40.5	0.998
second point	4967	33.3	20.2	20.4	0.992
third point	4983	16.7	10.1	10.2	0.992
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.4	40.3	1.003
Average Correction Factor					0.994

Baseline Corr As found:	40.51	Prev response:	40.47	*% change:	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustments made.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 22, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:15	End Time (MST):	10:47
Analyzer make:	API T300	Analyzer serial #:	3505

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999991	
40.4	40.5	0.9979			≥0.995
20.2	20.4	0.9916	Slope	1.001383	
10.1	10.2	0.9922			0.90 - 1.10
			Intercept	0.066908	+/-1.5



CO Calibration Plot

Date: October 17, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	October 17, 2023	Last Cal Date:	September 22, 2023
Start time (MST):	11:00	End time (MST):	14:56
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
N2 Gen Make/Model:	NG 5000		Serial Number:	771048318

Analyzer Information

Analyzer make: Teledyne API T360	Analyzer serial #: 289
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.989394	0.986227	Backgd or Offset:	0.008	0.008
Calibration intercept:	-0.140000	1.960000	Coeff or Slope:	1.019	1.019

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	1.6	----
as found span	2920	80.0	1605.9	1590.3	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	2.1	----
high point	2920	80.0	1605.9	1590.2	1.010
second point	2960	40.0	802.9	780.5	1.029
third point	2980	20.0	401.5	406.6	0.987
as left zero	3000	0.0	0.0	3.0	----
as left span	2960	40.0	802.9	779.6	1.030
Average Correction Factor					1.009

Baseline Corr As found:	1588.70	Prev response:	1588.69	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as found. No Adjustments made.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	October 17, 2023	Previous Calibration	September 22, 2023
Station Name	Fort Chipewyan	Station Number	AMS08
Start Time (MST)	11:00	End Time (MST)	14:56
Analyzer make	Teledyne API T360	Analyzer serial #	289

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	2.1	----	Correlation Coefficient	0.999800	≥0.995
1605.9	1590.2	1.0099			
802.9	780.5	1.0287	Slope	0.986227	0.90 - 1.10
401.5	406.6	0.9874			
			Intercept	1.960000	+/-20



CO₂ Calibration Plot

Date: October 17, 2023

Location: Fort Chipewyan





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	October 3, 2023	Last Cal Date:	September 5, 2023
Start time (MST):	9:07	End time (MST):	11:51
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.96	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC151285			
Removed Cal Gas Conc:	49.96	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3812
ZAG Make/Model:	API T701		Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000440	0.995922	Backgd or Offset:	10.3	10.3
Calibration intercept:	-0.648667	-0.090033	Coeff or Slope:	0.979	0.973

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4919	80.2	801.5	800.6	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4919	80.2	801.5	798.1	1.004
second point	4959	40.1	400.8	399.3	1.004
third point	4980	20.0	199.8	198.5	1.007
as left zero	5000	0.0	0.0	-0.2	----
as left span	4919	80.2	801.5	799.2	1.003
Average Correction Factor					1.005

Baseline Corr As found:	800.90	Previous response	801.19	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 5, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:07	End Time (MST):	11:51
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.0	0.1	----	Correlation Coefficient	0.999999	≥0.995			
801.5	798.1	1.0042						
400.8	399.3	1.0036				Slope	0.995922	0.90 - 1.10
199.8	198.5	1.0068				Intercept	-0.090033	+/-30



SO2 Calibration Plot

Date: October 3, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	October 11, 2023	Last Cal Date:	September 7, 2023
Start time (MST):	11:30	End time (MST):	14:09
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.171	ppm	Cal Gas Exp Date:	August 22, 2026
Cal Gas Cylinder #:	CC511415			
Removed Cal Gas Conc:	5.320	ppm	Rem Gas Exp Date:	August 22, 2026
Removed Gas Cyl #:	CC511710		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3812
ZAG Make/Model:	API T701		Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320
Converter make:	CDN-101	Converter serial #:	519
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000594	0.997327	Backgd or Offset:	2.77	3.06
Calibration intercept:	-0.001606	0.165639	Coeff or Slope:	1.134	1.252

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4923	77.4	80.0	79.8	1.003
second point	4961	38.7	40.0	40.4	0.991
third point	4981	18.8	19.4	19.7	0.987
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.4	80.0	79.5	1.007
SO2 Scrubber Check	4920	80.2	802.0		----
Date of last scrubber change:	28-Feb-23			Ave Corr Factor	0.994
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter. SOx scrubber check done after calibrator zero. Swapped the failed CGA cylinder with a new one. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 11, 2023	Previous Calibration:	September 7, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	11:30	End Time (MST):	14:09
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999935	
80.0	79.8	1.0030			≥0.995
40.0	40.4	0.9907	Slope	0.997327	
19.4	19.7	0.9870			0.90 - 1.10
			Intercept	0.165639	+/-3



TRS Calibration Plot

Date: October 11, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Barge Landing Station number: AMS09
 Calibration Date: October 16, 2023 Last Cal Date: October 11, 2023
 Start time (MST): 9:26 End time (MST): 13:23
 Reason: Maintenance

Calibration Standards

Cal Gas Concentration: 5.171 ppm Cal Gas Exp Date: August 22, 2026
 Cal Gas Cylinder #: CC511415
 Removed Cal Gas Conc: 5.171 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3812
 ZAG Make/Model: API T701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1331259320
 Converter make: CDN-101 Converter serial #: 519
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997327	1.006260	Backgd or Offset:	3.06 2.95
Calibration intercept:	0.165639	-0.200515	Coeff or Slope:	1.252 1.217

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4923	77.4	80.0	81.8	0.977
as found 2nd point	4961	38.7	40.0	41.0	0.974
as found 3rd point	4981	19.3	20.0	20.3	0.978
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4923	77.4	80.0	80.4	0.996
second point	4961	38.7	40.0	40.0	1.001
third point	4981	19.3	20.0	19.8	1.008
as left zero	5000	0.0	0.0	0.1	----
as left span	4923	77.4	80.0	80.0	1.001

SO2 Scrubber Check

Date of last scrubber change:	28-Feb-23	Ave Corr Factor	1.001
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 81.9 Prev response: 79.99 *% change: 2.3%
 Baseline Corr 2nd AF pt: 41.1 AF Slope: 1.023533 AF Intercept: -0.080172
 Baseline Corr 3rd AF pt: 20.4 AF Correlation: 0.999995

* = > +/-5% change initiates investigation

Notes: Purge the regulator and line during as found zero to remedy span response. Multi-point as founds within limits and linear. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021


Station Information

Calibration Date:	October 16, 2023	Previous Calibration:	October 11, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:26	End Time (MST):	13:23
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥ 0.995
80.0	80.4	0.9955		
40.0	40.0	1.0006	Slope	0.90 - 1.10
20.0	19.8	1.0080		
			Intercept	+/-3







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	October 2, 2023	Last Cal Date:	September 5, 2023
Start time (MST):	9:07	End time (MST):	11:51
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.52E-04	2.58E-04	NMHC SP Ratio:	4.87E-05
CH ₄ Retention time:	15.40	15.80	NMHC Peak Area:	187823
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				188320
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	17.07	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.08	1.002
second point	4960	40.1	8.56	8.59	0.996
third point	4980	20.0	4.27	4.28	0.998
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.20	0.995
				Response Correction Factor	0.999

Baseline Corr AF:	17.07	Prev response	17.14	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	9.14	9.24	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.12	1.002
second point	4960	40.1	4.57	4.60	0.993
third point	4980	20.0	2.28	2.29	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	9.14	9.18	0.995
Average Correction Factor					0.997
Baseline Corr AF:	9.24	Prev response	9.15	*% change	0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	7.83	1.020
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	7.96	1.002
second point	4960	40.1	3.99	3.99	1.000
third point	4980	20.0	1.99	1.99	0.999
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	8.01	0.996
Average Correction Factor					1.000
Baseline Corr AF:	7.83	Prev response	7.98	*% change	-2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001164	0.997687
THC Cal Offset:	-0.002739	0.018043
CH ₄ Cal Slope:	1.000766	0.997459
CH ₄ Cal Offset:	-0.005136	0.004658
NMHC Cal Slope:	1.001350	0.998099
NMHC Cal Offset:	0.002797	0.012785

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 2, 2023	Previous Calibration:	September 5, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:07	End Time (MST):	11:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999989	≥ 0.995			
17.12	17.08	1.0022						
8.56	8.59	0.9964				Slope	0.997687	0.90 - 1.10
4.27	4.28	0.9975						
			Intercept	0.018043	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 2, 2023	Previous Calibration:	September 5, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:07	End Time (MST):	11:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
7.98	7.96	1.0024						
3.99	3.99	1.0002				Slope	0.997459	0.90 - 1.10
1.99	1.99	0.9987						
			Intercept	0.004658	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 2, 2023	Previous Calibration:	September 5, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:07	End Time (MST):	11:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999977	≥ 0.995			
9.14	9.12	1.0019						
4.57	4.60	0.9934				Slope	0.998099	0.90 - 1.10
2.28	2.29	0.9965						
			Intercept	0.012785	± 0.5			



NMHC Calibration Plot

Date: October 2, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Barge Landing
Calibration Date: October 13, 2023
Start time (MST): 9:02
Reason: Routine
Station number: AMS09
Last Cal Date: September 6, 2023
End time (MST): 13:09

Calibration Standards

NO Gas Cylinder #: DT0036634
NOX Cal Gas Conc: 50.00 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 50.00 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: January 28, 2024
NO Cal Gas Conc: 49.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 49.70 ppm
NO gas Diff:
Serial Number: 3812
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1426262593

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.162	1.162	NO bkgnd or offset:	10.6	10.6
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	10.9	10.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	175.2	177.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000087	1.001860
NO _x Cal Offset:	1.049154	0.949799
NO Cal Slope:	1.000827	1.002626
NO Cal Offset:	-0.572366	-0.151784
NO ₂ Cal Slope:	1.000344	1.002309
NO ₂ Cal Offset:	0.677077	0.845161



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
as found span	4919	80.5	805.1	800.3	4.8	809.8	802.5	7.4	0.994	0.997
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4919	80.5	805.1	800.3	4.8	806.8	802.1	4.9	0.998	0.998
second point	4959	40.2	402.1	399.7	2.4	404.9	400.9	4.0	0.993	0.997
third point	4979	20.1	201.0	199.8	1.2	203.0	199.9	3.2	0.990	1.000
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.1	----	----
as left span	4919	80.5	805.1	444.1	361.0	806.7	444.3	362.4	0.998	0.999
Average Correction Factor									0.994	0.998

Corrected As found	NO _x = 810.1 ppb	NO = 802.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.5%
Previous Response	NO _x = 806.2 ppb	NO = 800.3 ppb		*Percent Change	NO = 0.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.0	442.8	361.0	362.4	0.996	100.4%
2nd GPT point (200 ppb O3)	799.0	663.5	140.3	141.9	0.989	101.1%
3rd GPT point (100 ppb O3)	799.0	730.2	73.6	75.4	0.977	102.4%
Average Correction Factor					0.987	101.3%

Notes: Changed sample inlet filter after as founds. No adjustment.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 13, 2023	Previous Calibration:	September 6, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:02	End Time (MST):	13:09
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
805.1	806.8	0.9979		
402.1	404.9	0.9930		
201.0	203.0	0.9903		
			0.999991	
			1.001860	
			0.949799	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 13, 2023	Previous Calibration:	September 6, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:02	End Time (MST):	13:09
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 $0.90 - 1.10$ ± 20
800.3	802.1	0.9977		
399.7	400.9	0.9969		
199.8	199.9	0.9996		





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 13, 2023	Previous Calibration:	September 6, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:02	End Time (MST):	13:09
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999979	≥0.995
361.0	362.4	0.9962			
140.3	141.9	0.9889			
73.6	75.4	0.9765			
			Slope	1.002309	0.90 - 1.10
			Intercept	0.845161	+/-20



NO_x Calibration Plot

Date: October 13, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Barge Landing Station number: AMS 09
 Calibration Date: October 19, 2023 Last Cal Date: September 7, 2023
 Start time (MST): 9:33 End time (MST): 10:25

Analyzer Make: API T640 S/N: 844
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 1451
 Temp/RH standard: Alicat FP-25 S/N: 1451

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	11.4	10.9	11.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	717.7	724.9	717.7	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.96	4.99	4.96	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>October 19, 2023</u>	Last Cal Date: <u>September 29, 2023</u>			
	PM w/o HEPA: <u>3.0</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.1	10.8	10.8	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>7.9</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>October 19, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>October 19, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: August 23, 2023
 Date RH/T Sensor Cleaned: August 23, 2023

Notes: Inlet head looks good. No adjustments made. Leak check passed. Filter was changed. PMT Peak test no adjustment.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS11
LOWER CAMP**

OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 22, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:32	End Time (MST):	13:52
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.9	----	Correlation Coefficient	0.999995	≥0.995
800.8	805.0	0.9947			
400.9	402.0	0.9973	Slope	1.003950	0.90 - 1.10
199.9	201.7	0.9913			
			Intercept	0.611030	+/-30



SO2 Calibration Plot

Date: October 12, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	October 11, 2023	Last Cal Date:	September 12, 2023
Start time (MST):	9:00	End time (MST):	14:49
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.429	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC501097			
Removed Cal Gas Conc:	5.429	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3807
ZAG Make/Model:	API T701H		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003
Converter make:	NA	Converter serial #:	NA
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.024520	1.022511	Backgd or Offset:	14.2	14.1
Calibration intercept:	-0.427099	-0.206756	Coeff or Slope:	1.001	1.001

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4926	73.6	79.9	81.3	0.982
as found 2nd point	4963	36.8	40.0	40.5	0.984
as found 3rd point	4982	18.6	20.2	20.3	0.990
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4926	73.6	79.9	81.4	0.982
second point	4963	36.8	40.0	40.9	0.977
third point	4982	18.6	20.2	20.4	0.990
as left zero	5000	0.0	0.0	-0.1	----
as left span	4926	73.6	79.9	81.8	0.977
SO2 Scrubber Check	4919	81.1	811.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	0.983
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	81.4	Prev response:	81.45	*% change:	-0.1%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.019073	AF Intercept:	-0.186359
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999995		

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Completed SO2 scrubber check after calibrator zero.
No adjustments required.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 11, 2023	Previous Calibration:	September 12, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	9:00	End Time (MST):	14:49
Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.3	----	Correlation Coefficient	0.999977	
79.9	81.4	0.9818			≥0.995
40.0	40.9	0.9770	Slope	1.022511	
20.2	20.4	0.9899			0.90 - 1.10
			Intercept	-0.206756	+/-3



H₂S Calibration Plot

Date: October 11, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	October 12, 2023	Last Cal Date:	September 22, 2023
Start time (MST):	10:32	End time (MST):	13:52
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
ZAG make/model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.99E-04	2.99E-04	NMHC SP Ratio:	5.79E-05
CH ₄ Retention time:	14.0	14.0	NMHC Peak Area:	158468

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	17.35	17.23	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	17.35	17.24	1.006
second point	4959	40.7	8.69	8.56	1.015
third point	4980	20.3	4.33	4.24	1.022
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	17.35	17.36	0.999

Average Correction Factor				1.014
Baseline Corr AF:	17.23	Prev response	17.24	*% change -0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	9.19	9.10	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	9.19	9.11	1.009
second point	4959	40.7	4.60	4.51	1.019
third point	4980	20.3	2.29	2.22	1.033
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	9.19	9.17	1.002
Average Correction Factor					1.020
Baseline Corr AF:	9.10	Prev response	9.13	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	8.16	8.13	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	8.16	8.13	1.004
second point	4959	40.7	4.09	4.04	1.011
third point	4980	20.3	2.04	2.02	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	8.16	8.19	0.997
Average Correction Factor					1.008
Baseline Corr AF:	8.13	Prev response	8.12	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.995046	0.994388
THC Cal Offset:	-0.021788	-0.039793
CH ₄ Cal Slope:	0.995660	0.996052
CH ₄ Cal Offset:	-0.010886	-0.009288
NMHC Cal Slope:	0.994824	0.992885
NMHC Cal Offset:	-0.011701	-0.030904

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 22, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:32	End Time (MST):	13:52
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999971	≥ 0.995			
17.35	17.24	1.0064						
8.69	8.56	1.0149				Slope	0.994388	0.90 - 1.10
4.33	4.24	1.0218						
			Intercept	-0.039793	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 22, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:32	End Time (MST):	13:52
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999985	≥0.995
8.16	8.13	1.0039			
4.09	4.04	1.0108			
2.04	2.02	1.0094			
			Slope	0.996052	0.90 - 1.10
			Intercept	-0.009288	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 22, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:32	End Time (MST):	13:52
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999946	≥ 0.995			
9.19	9.11	1.0086						
4.60	4.51	1.0191				Slope	0.992885	0.90 - 1.10
2.29	2.22	1.0330						
			Intercept	-0.030904	± 0.5			



NMHC Calibration Plot

Date: October 12, 2023

Location: Lower Camp





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13 FORT MCKAY SOUTH OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 15, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:55	End Time (MST):	12:25
Analyzer make:	API T100	Analyzer serial #:	599

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.4	----	Correlation Coefficient	0.999945	≥0.995
799.7	801.0	0.9984			
399.3	396.8	1.0063	Slope	1.004043	0.90 - 1.10
200.2	194.8	1.0276			
			Intercept	-3.158122	+/-30



SO2 Calibration Plot

Date: October 18, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: October 12, 2023 Last Cal Date: September 19, 2023
 Start time (MST): 7:55 End time (MST): 11:33
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.34 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC500241
 Removed Cal Gas Conc: 5.34 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API 701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1180540017
 Converter make: CDN-101 Converter serial #: 521
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004193	1.006178	Backgd or Offset: 3.77	3.77
Calibration intercept:	-0.442262	-0.262271	Coeff or Slope: 1.130	1.130

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4925	75.5	80.6	82.3	0.983
as found 2nd point	4962	37.7	40.3	40.7	0.997
as found 3rd point	4981	18.9	20.2	19.9	1.030
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4925	75.5	80.6	81.0	0.995
second point	4962	37.7	40.3	40.1	1.004
third point	4981	18.9	20.2	19.8	1.019
as left zero	5000	0.0	0.0	0.1	----
as left span	4925	75.5	80.6	80.2	1.005
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:	20-Mar-20		Ave Corr Factor		1.006
Date of last converter efficiency test:	NA		efficiency		

Baseline Corr As found: 82.0 Prev response: 80.52 *% change: 1.8%
 Baseline Corr 2nd AF pt: 40.4 AF Slope: 1.020212 AF Intercept: -0.182247
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999842

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 19, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	7:55	End Time (MST):	11:33
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999952	
80.6	81.0	0.9954			≥0.995
40.3	40.1	1.0041	Slope	1.006178	
20.2	19.8	1.0195			0.90 - 1.10
			Intercept	-0.262271	+/-3



TRS Calibration Plot

Date: October 12, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	October 18, 2023	Last Cal Date:	September 15, 2023
Start time (MST):	8:55	End time (MST):	12:25
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API701	Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.24E-04	2.31E-04	NMHC SP Ratio:	5.16E-05	5.16E-05
CH ₄ Retention time:	13.00	13.40	NMHC Peak Area:	176051	1756954
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	17.05	16.84	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.05	1.000
second point	4961	39.5	8.51	8.41	1.012
third point	4980	19.8	4.27	4.14	1.031
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	17.05	17.03	1.001
Average Correction Factor					1.014

Baseline Corr AF:	16.84	Prev response	17.02	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	9.08	9.11	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	9.08	9.08	1.000
second point	4961	39.5	4.53	4.49	1.010
third point	4980	19.8	2.27	2.22	1.024
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	9.08	9.06	1.003
Average Correction Factor					1.011
Baseline Corr AF:	9.11	Prev response	9.07	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	7.97	7.73	1.031
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.98	0.999
second point	4961	39.5	3.98	3.91	1.016
third point	4980	19.8	1.99	1.92	1.038
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	7.97	7.97	0.999
Average Correction Factor					1.018
Baseline Corr AF:	7.73	Prev response	7.96	*% change	-2.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.003320	1.002214
THC Cal Offset:	-0.078975	-0.072976
CH ₄ Cal Slope:	1.004551	1.003429
CH ₄ Cal Offset:	-0.047351	-0.043993
NMHC Cal Slope:	1.002546	1.001224
NMHC Cal Offset:	-0.032763	-0.029785

Notes: Changed inlet filter after as founds. Adjusted span.

Calibration Performed By: Devin Russell Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 15, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:55	End Time (MST):	12:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999915	≥ 0.995			
17.05	17.05	0.9997						
8.51	8.41	1.0124				Slope	1.002214	0.90 - 1.10
4.27	4.14	1.0307						
			Intercept	-0.072976	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 15, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:55	End Time (MST):	12:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999855	≥0.995
7.97	7.98	0.9988			
3.98	3.91	1.0164			
1.99	1.92	1.0382			
			Slope	1.003429	0.90 - 1.10
			Intercept	-0.043993	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 15, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:55	End Time (MST):	12:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999950	≥ 0.995
9.08	9.08	1.0003			
4.53	4.49	1.0097			
2.27	2.22	1.0243			
			Slope	1.001224	0.90 - 1.10
			Intercept	-0.029785	+/-0.5



NMHC Calibration Plot

Date: October 18, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort McKay South
Calibration Date: October 24, 2023
Start time (MST): 8:23
Reason: Routine
Station number: AMS 13
Last Cal Date: September 26, 2023
End time (MST): 13:10

Calibration Standards

NO Gas Cylinder #: T2Y1P76
NOX Cal Gas Conc: 50.98 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 50.98 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 49.32 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 49.32 ppm
NO gas Diff:
Serial Number: 2448
Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1410661329

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.120	1.124	NO bkgnd or offset:	10.1	10.4
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	10.8	10.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	158.4	160.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999304	0.999331
NO _x Cal Offset:	-1.811346	-2.071318
NO Cal Slope:	1.002205	1.002262
NO Cal Offset:	-2.825167	-2.945200
NO ₂ Cal Slope:	0.995992	1.001663
NO ₂ Cal Offset:	-1.479105	-0.500859



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.7	0.0	-0.7	----	----
as found span	4919	81.1	826.9	800.0	26.9	825.8	799.2	26.6	1.0013	1.0009
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4919	81.1	826.9	800.0	26.9	825.2	800.2	24.9	1.0020	0.9997
second point	4960	40.6	413.9	400.4	13.5	410.7	397.1	13.5	1.0078	1.0084
third point	4980	20.3	207.0	200.2	6.7	202.7	194.8	7.9	1.0210	1.0279
as left zero	5000	0.0	0.0	0.0	0.0	0.5	0.4	0.1	----	----
as left span	4919	80.1	816.8	380.1	436.7	828.4	387.8	440.6	0.9861	0.9802
Average Correction Factor									1.0103	1.0120

Corrected As found	NO _x = 826.5 ppb	NO = 799.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.2%
Previous Response	NO _x = 824.5 ppb	NO = 798.9 ppb		*Percent Change	NO = 0.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	794.0	384.2	436.7	437.4	0.9985	100.2%
2nd GPT point (200 ppb O ₃)	794.0	588.4	232.5	231.7	1.0036	99.6%
3rd GPT point (100 ppb O ₃)	794.0	690.8	130.1	129.6	1.0040	99.6%
Average Correction Factor					1.0020	99.8%

Notes: Changed inlet filter after as founds. Adjusted zero and span. Used 2nd NO reference point due to drift.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 24, 2023	Previous Calibration:	September 26, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:23	End Time (MST):	13:10
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
826.9	825.2	1.0020			
413.9	410.7	1.0078			
207.0	202.7	1.0210			
			Slope	0.999331	0.90 - 1.10
			Intercept	-2.071318	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 24, 2023	Previous Calibration:	September 26, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:23	End Time (MST):	13:10
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
800.0	800.2	0.9997			
400.4	397.1	1.0084			
200.2	194.8	1.0279			
			Slope	1.002262	0.90 - 1.10
			Intercept	-2.945200	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 24, 2023	Previous Calibration:	September 26, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:23	End Time (MST):	13:10
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
436.7	437.4	0.9985		
232.5	231.7	1.0036		
130.1	129.6	1.0040		
			0.999990	
			1.001663	
			-0.500859	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 24, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: October 2, 2023 Last Cal Date: September 8, 2023
 Start time (MST): 10:12 End time (MST): 13:13
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3871
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999886	0.998314	Backgd or Offset:	2.4	2.4
Calibration intercept:	1.020000	0.820000	Coeff or Slope:	0.967	0.967

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	5000	977.0	400.0	398.5	1.004
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.2	----
high point	5000	977.0	400.0	399.4	1.002
second point	5000	838.0	200.0	201.7	0.992
third point	5000	735.9	100.0	101.2	0.988
as left zero	5000	0.0	0.0	-0.2	----
as left span	5000	977.0	400.0	402.2	0.995
Average Correction Factor					0.994

Baseline Corr As found:	398.6	Previous response	401.0	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustment made. Graph was taken from MDS.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 2, 2023	Previous Calibration:	September 8, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:12	End Time (MST):	13:13
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999961	≥0.995
400.0	399.4	1.0015			
200.0	201.7	0.9916			
100.0	101.2	0.9881			
			Slope	0.998314	0.90 - 1.10
			Intercept	0.820000	+/- 5



O₃ Calibration Plot

Date: October 2, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: October 6, 2023 Last Cal Date: October 2, 2023
 Start time (MST): 8:56 End time (MST): 11:54
 Reason: Maintenance

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3871
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998314	1.004514	Backgd or Offset:	2.4	2.4
Calibration intercept:	0.820000	1.260000	Coeff or Slope:	0.967	0.967

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	5000	977.0	400.0	401.4	0.997
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.6	----
high point	5000	977.0	400.0	402.6	0.994
second point	5000	838.0	200.0	202.9	0.986
third point	5000	735.9	100.0	102.1	0.979
as left zero	5000	0.0	0.0	0.3	----
as left span	5000	977.0	400.0	402.9	0.993
Average Correction Factor					0.986

Baseline Corr As found:	401.6	Previous response	400.1	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Lamp was adjusted yesterday, O3 ref was reading at 2500mV and now its at 4590mV. No adjustment made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 6, 2023	Previous Calibration:	October 2, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	8:56	End Time (MST):	11:54
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.6	----	Correlation Coefficient	0.999985	
400.0	402.6	0.9935			≥0.995
200.0	202.9	0.9857	Slope	1.004514	
100.0	102.1	0.9794			0.90 - 1.10
			Intercept	1.260000	+/- 5



O₃ Calibration Plot

Date: October 6, 2023

Location: Fort McKay South





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC
OCTOBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	October 3, 2023	Last Cal Date:	September 6, 2023
Start time (MST):	9:18	End time (MST):	12:22
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.95	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC279389			
Removed Cal Gas Conc:	49.95	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API T701H		Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.017069	1.006812	Backgd or Offset:	25.2	25.2
Calibration intercept:	-1.850638	-2.184734	Coeff or Slope:	0.798	0.798

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4938	80.3	799.3	803.7	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.7	----
high point	4938	80.3	799.3	804.4	0.994
second point	4979	40.2	400.1	397.9	1.005
third point	4998	20.2	201.1	198.2	1.014
as left zero	5000	0.0	0.0	0.6	----
as left span	4938	80.3	799.3	805.7	0.992
Average Correction Factor					1.005

Baseline Corr As found:	803.20	Previous response	811.06	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 6, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:18	End Time (MST):	12:22
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient	0.999934	
799.3	804.4	0.9936			≥0.995
400.1	397.9	1.0054	Slope	1.006812	
201.1	198.2	1.0145			0.90 - 1.10
			Intercept	-2.184734	+/-30



SO2 Calibration Plot

Date: October 3, 2023

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Anzac Station number: AMS14
 Calibration Date: October 5, 2023 Last Cal Date: September 7, 2023
 Start time (MST): 10:36 End time (MST): 14:44
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC510379
 Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3060
 ZAG Make/Model: API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153582
 Converter make: CD Nova CDN-101 Converter serial #: 503
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992584	0.997157	Backgd or Offset: 2.31	2.28
Calibration intercept:	-0.245370	-0.105310	Coeff or Slope: 0.992	0.992

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4938	77.9	80.0	80.1	0.996
as found 2nd point	4973	38.9	40.0	40.1	0.992
as found 3rd point	4997	19.5	20.0	19.7	1.006
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4938	77.9	80.0	79.6	1.004
second point	4973	38.9	40.0	39.9	1.001
third point	4997	19.5	20.0	19.6	1.021
as left zero	5000	0.0	0.0	0.1	----
as left span	4938	77.9	80.0	78.2	1.022
SO2 Scrubber Check	4936	80.3	800.4	-0.1	----
Date of last scrubber change:				Ave Corr Factor	1.009
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 80.3 Prev response: 79.14 *% change: 1.4%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 1.005058 AF Intercept: -0.245371
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999982

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 7, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:36	End Time (MST):	14:44
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999971	
80.0	79.6	1.0044			≥0.995
40.0	39.9	1.0014	Slope	0.997157	
20.0	19.6	1.0210			0.90 - 1.10
			Intercept	-0.105310	+/-3



TRS Calibration Plot

Date: October 5, 2023

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	October 3, 2023	Last Cal Date:	September 6, 2023
Start time (MST):	9:18	End time (MST):	12:22
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.75E-04	3.75E-04	NMHC SP Ratio:	4.49E-05
CH ₄ Retention time:	12.20	12.20	NMHC Peak Area:	203038

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	17.13	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.09	1.001
second point	4979	40.2	8.56	8.50	1.008
third point	4998	20.2	4.30	4.23	1.018
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	17.10	17.10	1.000

Average Correction Factor				1.009
Baseline Corr AF:	17.13	Prev response	17.11	*% change 0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	9.08	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.05	1.007
second point	4979	40.2	4.56	4.49	1.015
third point	4998	20.2	2.29	2.23	1.028
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	9.11	9.06	1.006
Average Correction Factor					1.017
Baseline Corr AF:	9.08	Prev response	9.11	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	8.05	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	8.05	0.992
second point	4979	40.2	4.00	4.00	1.000
third point	4998	20.2	2.01	2.00	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	7.99	8.04	0.994
Average Correction Factor					1.000
Baseline Corr AF:	8.05	Prev response	8.00	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002474	1.000528
THC Cal Offset:	-0.037684	-0.040856
CH ₄ Cal Slope:	1.003084	1.008350
CH ₄ Cal Offset:	-0.015639	-0.017069
NMHC Cal Slope:	1.001701	0.994097
NMHC Cal Offset:	-0.021844	-0.024989

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 6, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:18	End Time (MST):	12:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999973	≥ 0.995			
17.10	17.09	1.0005						
8.56	8.50	1.0076				Slope	1.000528	0.90 - 1.10
4.30	4.23	1.0176						
			Intercept	-0.040856	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 6, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:18	End Time (MST):	12:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999977	≥0.995
7.99	8.05	0.9925			
4.00	4.00	0.9998			
2.01	2.00	1.0064			
			Slope	1.008350	0.90 - 1.10
			Intercept	-0.017069	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 6, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:18	End Time (MST):	12:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999965	≥ 0.995			
9.11	9.05	1.0072						
4.56	4.49	1.0150				Slope	0.994097	0.90 - 1.10
2.29	2.23	1.0276						
			Intercept	-0.024989	± 0.5			



NMHC Calibration Plot

Date: October 3, 2023

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	October 16, 2023	Last Cal Date:	October 3, 2023
Start time (MST):	10:48	End time (MST):	12:21
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.75E-04	3.75E-04	NMHC SP Ratio:	4.49E-05
CH ₄ Retention time:	12.20	12.20	NMHC Peak Area:	203038

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	17.08	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.05	1.003
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.003
Baseline Corr AF:	17.08	Prev response	17.07	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	9.07	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.07	1.005
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.005
Baseline Corr AF:	9.07	Prev response	9.03	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	8.01	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	7.98	1.001
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.001
Baseline Corr AF:	8.01	Prev response	8.04	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000528	0.996683
THC Cal Offset:	-0.040856	0.000000
CH ₄ Cal Slope:	1.008350	0.999059
CH ₄ Cal Offset:	-0.017069	0.000000
NMHC Cal Slope:	0.994097	0.994819
NMHC Cal Offset:	-0.024989	0.000000


Notes: Nitrogen cylinder change.

Calibration Performed By: Mohammed Kashif

NMHC Calibration Plot

Date: October 16, 2023

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Anzac Station number: AMS 14
Calibration Date: October 4, 2023 Last Cal Date: September 5, 2023
Start time (MST): 8:48 End time (MST): 13:21
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1P8D Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.92 ppm NO Cal Gas Conc: 50.05 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.92 ppm Removed Gas NO Conc: 50.05 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3060
ZAG make/model: Teledyne API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1426262592
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.411	1.411	NO bkgnd or offset:	3.8	3.8
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.8	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.3	159.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000191	0.996975
NO _x Cal Offset:	-0.708567	-0.888234
NO Cal Slope:	1.000876	0.996206
NO Cal Offset:	-2.586342	-2.526016
NO ₂ Cal Slope:	1.002633	1.005379
NO ₂ Cal Offset:	0.926052	1.537643



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.2	0.3	----	----
as found span	4936	80.2	814.1	800.2	13.9	811.6	793.6	18.0	1.0031	1.0083
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.0	0.3	----	----
high point	4936	80.2	814.1	800.2	13.9	811.4	796.0	15.4	1.0034	1.0053
second point	4979	40.1	406.8	399.9	7.0	404.0	394.3	9.7	1.0070	1.0141
third point	4999	20.1	203.9	200.4	3.5	201.3	194.8	6.4	1.0130	1.0289
as left zero	5000	0.0	0.0	0.0	0.0	0.3	-0.1	0.3	----	----
as left span	4936	80.2	814.1	418.0	396.1	812.8	413.3	399.4	1.0016	1.0114
Average Correction Factor									1.0078	1.0161

Corrected As found	NO _x = 811.5 ppb	NO = 793.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.3%	
Previous Response	NO _x = 813.6 ppb	NO = 798.3 ppb		*Percent Change	NO = -0.6%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	793.9	411.7	396.1	399.1	0.9925	100.8%
2nd GPT point (200 ppb O3)	793.9	612.9	194.9	198.3	0.9829	101.7%
3rd GPT point (100 ppb O3)	793.9	707.2	100.6	103.8	0.9693	103.2%
Average Correction Factor					0.9816	101.9%

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 5, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	8:48	End Time (MST):	13:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
814.1	811.4	1.0034		
406.8	404.0	1.0070		
203.9	201.3	1.0130		
			0.999990	
			0.996975	
			-0.888234	





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 5, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	8:48	End Time (MST):	13:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
800.2	796.0	1.0053		
399.9	394.3	1.0141		
200.4	194.8	1.0289		





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 5, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	8:48	End Time (MST):	13:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
396.1	399.1	0.9925		
194.9	198.3	0.9829		
100.6	103.8	0.9693		



NO_x Calibration Plot

Date: October 4, 2023

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 6, 2023	Previous Calibration:	September 1, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	8:47	End Time (MST):	12:18
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999740	
400.0	402.9	0.9928			≥0.995
200.0	197.7	1.0116	Slope	1.009371	
100.0	95.6	1.0460			0.90 - 1.10
			Intercept	-2.440000	+/- 5



O₃ Calibration Plot

Date: October 6, 2023


Location: Anzac



O₃ Calibration Plot

Date: October 13, 2023

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 13, 2023	Previous Calibration:	October 13, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	16:02	End Time (MST):	18:37
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220026

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999999	≥0.995
400.0	399.4	1.0015			
200.0	199.9	1.0005	Slope	0.997514	0.90 - 1.10
100.0	99.9	1.0010			
			Intercept	0.360000	+/- 5



O₃ Calibration Plot

Date: October 13, 2023

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Anzac Station number: AMS 14
 Calibration Date: October 6, 2023 Last Cal Date: September 8, 2023
 Start time (MST): 9:49 End time (MST): 10:34

Analyzer Make: API T640 S/N: 825
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388749
 Temp/RH standard: Alicat FP-25 S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	5.2	6.3	5.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	718.5	719.6	718.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.98	5.03	4.98	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>October 6, 2023</u>		Last Cal Date: <u>September 8, 2023</u>		
	PM w/o HEPA: <u>3.4</u>		PM w/ HEPA: <u>0.0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.7	10	10.8	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: <u>4</u>		w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:	<u>October 6, 2023</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>October 6, 2023</u>				

Annual Maintenance

Date Sample Tube Cleaned: July 6, 2023
 Date RH/T Sensor Cleaned: July 6, 2023

Notes: Performed quarterly maintenance and leak check passed. Adjusted PMT peak test. Head cleaned.

Calibration by: Mohammed Kashif



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS17
WAPASU
OCTOBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 11, 2023	Previous Calibration:	September 5, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:25	End Time (MST):	13:15
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999957	≥0.995
800.0	801.5	0.9981			
400.0	396.5	1.0089	Slope	1.002238	0.90 - 1.10
199.5	197.0	1.0128			
			Intercept	-1.839260	+/-30



SO2 Calibration Plot

Date: October 11, 2023

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	October 16, 2023	Last Cal Date:	September 7, 2023
Start time (MST):	10:42	End time (MST):	15:25
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.076	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511852			
Removed Cal Gas Conc:	5.076	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2449
ZAG Make/Model:	API T701H		Serial Number:	359

Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583
Converter make:	n/a	Converter serial #:	n/a
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.990569	1.001854	Backgd or Offset:	12.2
Calibration intercept:	-0.219241	0.060776	Coeff or Slope:	1.114

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	78.8	80.0	81.9	0.976
as found 2nd point	4961	39.4	40.0	40.9	0.976
as found 3rd point	4980	19.7	20.0	20.1	0.990
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4921	78.8	80.0	80.5	0.994
second point	4961	39.4	40.0	39.6	1.010
third point	4980	19.7	20.0	20.0	1.000
as left zero	5000	0.0	0.0	0.5	----
as left span	4921	78.8	80.0	78.6	1.018
SO2 Scrubber Check	4921	79.4	800.0	-0.3	----
Date of last scrubber change:	n/a			Ave Corr Factor	1.001
Date of last converter efficiency test:	n/a				efficiency

Baseline Corr As found:	82.0	Prev response:	79.03	*% change:	3.6%
Baseline Corr 2nd AF pt:	41.0	AF Slope:	1.026282	AF Intercept:	-0.219175
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999983		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 16, 2023	Previous Calibration:	September 7, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:42	End Time (MST):	15:25
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient	≥0.995
80.0	80.5	0.9938		
40.0	39.6	1.0100	Slope	0.90 - 1.10
20.0	20.0	1.0000		
			Intercept	+/-3



H₂S Calibration Plot

Date: October 16, 2023

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 11, 2023	Previous Calibration:	September 5, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:25	End Time (MST):	13:15
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<i>Limits</i>	
0.00	-0.07	----	Correlation Coefficient	0.999955	
17.09	17.12	0.9983			≥0.995
8.55	8.42	1.0147	Slope	1.005834	
4.26	4.18	1.0204			0.90 - 1.10
			Intercept	-0.105515	+/-1.5



THC Calibration Plot

Date: October 11, 2023

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Wapasu
Calibration Date: October 18, 2023
Start time (MST): 9:57
Reason: Routine
Station number: AMS17
Last Cal Date: September 22, 2023
End time (MST): 14:43

Calibration Standards

NO Gas Cylinder #: T375YK8
NOX Cal Gas Conc: 49.11 ppm
Removed Cylinder #: T375YK8
Removed Gas NOX Conc: 49.11 ppm
NOX gas Diff: T375YK8
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: April 13, 2025
NO Cal Gas Conc: 48.07 ppm
Removed Gas Exp Date: T375YK8
Removed Gas NO Conc: 48.07 ppm
NO gas Diff: T375YK8
Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: Thermo Scientific 42iQ
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 12300522720

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.269	1.294	NO bkgnd or offset:	9.0	9.2
NOX coeff or slope:	0.989	0.990	NOX bkgnd or offset:	9.1	9.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	363.7	363.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000361	1.002375
NO _x Cal Offset:	-3.500000	-3.420000
NO Cal Slope:	1.002373	1.001687
NO Cal Offset:	-4.480000	-3.940000
NO ₂ Cal Slope:	0.998879	1.008551
NO ₂ Cal Offset:	-1.541001	-0.167241



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	-0.1	----	----
as found span	4917	83.2	817.2	799.9	17.3	792.0	776.0	16.0	1.0318	1.0308
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.3	0.0	----	----
high point	4917	83.2	817.2	799.9	17.3	818.2	800.0	18.1	0.9988	0.9999
second point	4958	41.6	408.6	399.9	8.7	402.3	392.8	9.5	1.0156	1.0182
third point	4979	20.8	204.3	200.0	4.3	199.1	193.3	5.8	1.0261	1.0345
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.5	-0.1	----	----
as left span	4917	83.2	817.2	395.2	422.0	814.5	392.8	421.6	1.0033	1.0060
Average Correction Factor									1.0135	1.0175

Corrected As found	NO _x = 791.8 ppb	NO = 775.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -2.8%
Previous Response	NO _x = 814.0 ppb	NO = 797.3 ppb		*Percent Change	NO = -2.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	801.4	396.7	422.0	425.7	0.9913	100.9%
2nd GPT point (200 ppb O3)	801.4	603.0	215.7	216.8	0.9950	100.5%
3rd GPT point (100 ppb O3)	801.4	701.9	116.8	117.8	0.9916	100.9%
Average Correction Factor					0.9926	100.7%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 22, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:57	End Time (MST):	14:43
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
817.2	818.2	0.9988		
408.6	402.3	1.0156		
204.3	199.1	1.0261		
			0.999893	
			1.002375	
			-3.420000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 22, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:57	End Time (MST):	14:43
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.3	----	Correlation Coefficient	0.999858	≥0.995
799.9	800.0	0.9999			
399.9	392.8	1.0182	Slope	1.001687	0.90 - 1.10
200.0	193.3	1.0345			
			Intercept	-3.940000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 22, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:57	End Time (MST):	14:43
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
422.0	425.7	0.9913		
215.7	216.8	0.9950		
116.8	117.8	0.9916		
			0.999995	
			1.008551	
			-0.167241	

NO₂ Calibration Curve





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Wapasu Station number: AMS17
 Calibration Date: October 4, 2023 Last Cal Date: September 14, 2023
 Start time (MST): 10:10 End time (MST): 13:06
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 2449
 ZAG Make/Model: API T701H Serial Number: 359

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 3870
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004971	1.002371	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-1.120000	-0.440000	Coeff or Slope:	1.014	1.014

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.8	----
as found span	5000	1077.3	400.0	400.2	1.000
as found 2nd point	5000	900.3	200.0		
as found 3rd point	5000	789.5	100.0		
calibrator zero	5000	0.0	0.0	0.3	----
high point	5000	1077.3	400.0	400.8	0.998
second point	5000	900.3	200.0	199.9	1.001
third point	5000	789.5	100.0	98.9	1.011
as left zero	5000	0.0	0.0	0.0	----
as left span	5000	1077.3	400.0	405.9	0.985
Average Correction Factor					1.003

Baseline Corr As found:	401.0	Previous response	400.9	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 14, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:10	End Time (MST):	13:06
Analyzer make:	API T400	Analyzer serial #:	3870

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	≥0.995
400.0	400.8	0.9980		
200.0	199.9	1.0005	Slope	0.90 - 1.10
100.0	98.9	1.0111		
			Intercept	+/- 5



O₃ Calibration Plot

Date: October 4, 2023

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Wapasu Station number: AMS 17
 Calibration Date: October 18, 2023 Last Cal Date: August 15, 2023
 Start time (MST): 11:42 End time (MST): 14:00

Analyzer Make: API T640 S/N: 1183
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.9	10.9	11.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	705.3	708.2	704.9	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.45	4.58	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>September 22, 2023</u>	Last Cal Date: <u>August 15, 2023</u>			
	PM w/o HEPA: <u>3.5</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.2		11.2	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>3.4</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>October 18, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 22, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: Temp, pressure and flow checked. Flow was low, pump swapped after as found checks. Leak check passed. Chamber cleaned. PMT checked.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Wapasu	Station Number:	AMS 17
Calibration Date:	October 16, 2023	Prev Cal Date:	October 18, 2022
Start Time (MST):	12:26	End Time (MST):	13:49
Tower Height (m):	9.5	Reason:	Routine

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	N14664
WS Calibrator:	MetOne 053	Serial Number:	CA 03845

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.2	0.2%
400	39.4	39.4	0.1%
600	58.6	58.6	0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		1.000000	<i>≥0.9995</i>
Calculated slope		0.999465	<i>0.90 - 1.10</i>
Calculated intercept		-0.013446	<i>+/- 2</i>

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	P19942
As Found Declination (deg east of True North):		As Left Declination (deg east of True North):	
Solar noon time (MST):		13:10 Calc Declination*:	13.34 Degrees
Deadband calc:	1.0 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	2.0	---
90	90.8	0.2%
180	179.8	-0.1%
270	269.3	-0.2%
357	358.0	0.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999980	<i>≥0.9995</i>
Calculated slope		1.003932	<i>0.90 - 1.10</i>
Calculated intercept		-1.287608	<i>+/- 4</i>

Notes: No issues.

Calibration Performed By: Aswin Sasi Kumar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 8, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:40	End Time (MST):	13:42
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.6	----	Correlation Coefficient	0.999994	
800.3	799.6	1.0009			≥0.995
400.2	398.5	1.0042	Slope	0.998676	
199.6	198.7	1.0046			0.90 - 1.10
			Intercept	-0.203578	+/-30



SO2 Calibration Plot

Date: October 17, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: October 26, 2023 Last Cal Date: September 25, 2023
 Start time (MST): 10:34 End time (MST): 15:05
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.479 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC500395
 Removed Cal Gas Conc: 5.479 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
 ZAG Make/Model: Teledyne API T701 Serial Number: 360

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359
 Converter make: CD Nova CDN-101 Converter serial #: 555
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005299	1.002584	Backgd or Offset:	2.66
Calibration intercept:	-0.019011	0.401023	Coeff or Slope:	1.173

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4927	73.0	80.0	83.2	0.963
as found 2nd point	4964	36.5	40.0	42.0	0.954
as found 3rd point	4983	18.3	20.0	20.9	0.964
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.5	----
high point	4927	73.0	80.0	80.6	0.992
second point	4964	36.5	40.0	40.6	0.985
third point	4983	18.3	20.0	20.3	0.988
as left zero	5000	0.0	0.0	0.4	----
as left span	4927	73.0	80.0	79.0	1.013
SO2 Scrubber Check	4923	77.1	771.0	-0.1	----

Date of last scrubber change:	17-Dec-21	Ave Corr Factor	0.988
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 83.1 Prev response: 80.40 *% change: 3.3%
 Baseline Corr 2nd AF pt: 41.9 AF Slope: 1.039454 AF Intercept: 0.160278
 Baseline Corr 3rd AF pt: 20.8 AF Correlation: 0.999974

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. Scrubber check completed after calibrator zero. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 26, 2023	Previous Calibration:	September 25, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:34	End Time (MST):	15:05
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999983	
80.0	80.6	0.9925			≥0.995
40.0	40.6	0.9850	Slope	1.002584	
20.0	20.3	0.9876			0.90 - 1.10
			Intercept	0.401023	+/-3



TRS Calibration Plot

Date: October 26, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	October 17, 2023	Last Cal Date:	September 8, 2023
Start time (MST):	10:40	End time (MST):	13:42
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.23E-04	2.24E-04	NMHC SP Ratio:	4.19E-05
CH ₄ Retention time:	12.9	12.7	NMHC Peak Area:	186711
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				5.04E-05
				182001
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	17.28	16.96	1.019
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.24	1.002
second point	4959	40.5	8.64	8.63	1.001
third point	4979	20.2	4.31	4.28	1.008
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.44	0.991
Average Correction Factor					1.004

Baseline Corr AF:	16.96	Prev response	17.32	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	8.92	1.027
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.16	1.001
second point	4959	40.5	4.58	4.60	0.996
third point	4979	20.2	2.29	2.30	0.994
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	9.17	9.27	0.989
Average Correction Factor					0.997
Baseline Corr AF:	8.92	Prev response	9.24	*% change	-3.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	8.11	8.04	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	8.09	1.003
second point	4959	40.5	4.06	4.03	1.008
third point	4979	20.2	2.02	1.98	1.025
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.17	0.993
Average Correction Factor					1.012
Baseline Corr AF:	8.04	Prev response	8.09	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.004309	0.998344
THC Cal Offset:	-0.036379	-0.008797
CH ₄ Cal Slope:	1.001484	0.998456
CH ₄ Cal Offset:	-0.032618	-0.020622
NMHC Cal Slope:	1.007917	0.998233
NMHC Cal Offset:	-0.005959	0.011624

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 8, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:40	End Time (MST):	13:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
17.28	17.24	1.0022						
8.64	8.63	1.0013				Slope	0.998344	0.90 - 1.10
4.31	4.28	1.0083						
			Intercept	-0.008797	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 8, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:40	End Time (MST):	13:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999969	≥0.995			
8.11	8.09	1.0031						
4.06	4.03	1.0077				Slope	0.998456	0.90 - 1.10
2.02	1.98	1.0246						
			Intercept	-0.020622	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 8, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:40	End Time (MST):	13:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999990	≥ 0.995			
9.17	9.16	1.0015						
4.58	4.60	0.9960				Slope	0.998233	0.90 - 1.10
2.29	2.30	0.9943						
			Intercept	0.011624	± 0.5			



NMHC Calibration Plot

Date: October 17, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: October 27, 2023 Last Cal Date: September 19, 2023
Start time (MST): 10:25 End time (MST): 15:36
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2XX7ME Cal Gas Expiry Date: January 14, 2024
NOX Cal Gas Conc: 50.48 ppm NO Cal Gas Conc: 49.22 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.48 ppm Removed Gas NO Conc: 49.22 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 2658
ZAG make/model: Teledyne API 701H Serial Number: 360

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1336160088
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.029	1.072	NO bkgnd or offset:	2.9	3.0
NOX coeff or slope:	0.984	0.985	NOX bkgnd or offset:	2.9	3.0
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	235.3	250.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000359	0.999718
NO _x Cal Offset:	-0.310049	-0.329761
NO Cal Slope:	1.001795	1.000010
NO Cal Offset:	-1.669844	-1.269777
NO ₂ Cal Slope:	0.998696	1.001575
NO ₂ Cal Offset:	-0.059273	0.299687



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as found span	4919	81.3	820.8	800.3	20.5	787.9	767.0	21.0	1.0417	1.0434
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
high point	4919	81.3	820.8	800.3	20.5	820.5	799.6	20.8	1.0003	1.0008
second point	4959	40.7	410.9	400.7	10.3	410.1	399.0	11.1	1.0020	1.0042
third point	4980	20.3	204.9	199.8	5.1	204.1	197.0	7.1	1.0041	1.0143
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
as left span	4919	81.3	820.8	372.6	448.2	824.3	371.2	453.1	0.9957	1.0037
Average Correction Factor									1.0021	1.0065

Corrected As found	NO _x = 788.0 ppb	NO = 767.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -4.2%
Previous Response	NO _x = 820.7 ppb	NO = 800.0 ppb		*Percent Change	NO = -4.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.3	371.6	448.2	449.1	0.9980	100.2%
2nd GPT point (200 ppb O3)	799.3	597.8	222.0	222.7	0.9968	100.3%
3rd GPT point (100 ppb O3)	799.3	701.3	118.5	119.2	0.9940	100.6%
Average Correction Factor					0.9963	100.4%

Notes: Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 27, 2023	Previous Calibration:	September 19, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:25	End Time (MST):	15:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	≥0.995	
820.8	820.5	1.0003			
410.9	410.1	1.0020			
204.9	204.1	1.0041			
			Slope	0.999718	0.90 - 1.10
			Intercept	-0.329761	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 27, 2023	Previous Calibration:	September 19, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:25	End Time (MST):	15:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	799.6	1.0008		
400.7	399.0	1.0042		
199.8	197.0	1.0143		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 27, 2023	Previous Calibration:	September 19, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:25	End Time (MST):	15:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999999	≥0.995
448.2	449.1	0.9980			
222.0	222.7	0.9968			
118.5	119.2	0.9940			
			Slope	1.001575	0.90 - 1.10
			Intercept	0.299687	+/-20



NO_x Calibration Plot

Date: October 27, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: October 23, 2023 Last Cal Date: September 21, 2023
 Start time (MST): 10:30 End time (MST): 14:22
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T750 Serial Number: 282
 ZAG Make/Model: Teledyne API T751H Serial Number: 355

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 825
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003943	0.999771	Backgd or Offset:	1.300	2.500
Calibration intercept:	-0.340000	0.240000	Coeff or Slope:	0.996	1.007

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.6	----
as found span	4888	1096.9	400.0	397.6	1.006
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.8	----
high point	4888	1101.7	400.0	400.1	1.000
second point	4888	863.9	200.0	200.9	0.996
third point	4888	741.4	100.0	99.0	1.010
as left zero	5000	800.0	0.0	-0.6	----
as left span	4812	1097.9	400.0	415.4	0.963
Average Correction Factor					1.002

Baseline Corr As found:	397.0	Previous response	401.2	*% change	-1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

** = > +/-5% change initiates investigation*

Notes: Sample inlet filter changed after as founds. Using portable calibrator for the calibration. Portable zero air was not able to keep pressure, switched to station zero air support for the rest of the calibration. Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 23, 2023	Previous Calibration:	September 21, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:30	End Time (MST):	14:22
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.8	----	Correlation Coefficient	0.999974	≥0.995
400.0	400.1	0.9998			
200.0	200.9	0.9955	Slope	0.999771	0.90 - 1.10
100.0	99.0	1.0101			
			Intercept	0.240000	+/- 5



O₃ Calibration Plot

Date: October 23, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Stony Mountain Station number: AMS 18
 Calibration Date: October 27, 2023 Last Cal Date: September 25, 2023
 Start time (MST): 13:30 End time (MST): 15:17

Analyzer Make: API T640 S/N: 1162
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-5.4	-5.2	-5.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	707.3	704.3	707.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	4.98	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>October 27, 2023</u>	Last Cal Date: <u>September 25, 2023</u>			
	PM w/o HEPA: <u>3.9</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	NA		NA	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>3.9</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>September 25, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 25, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022
 Date RH/T Sensor Cleaned: August 30, 2022

Notes: No adjustments needed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	October 12, 2023	Last Cal Date:	September 11, 2023
Start time (MST):	9:50	End time (MST):	12:50
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,050	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	3,050	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701		Serial Number:	360

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995840	0.991149	Backgd or Offset:	-0.010	-0.010
Calibration intercept:	0.085810	0.191815	Coeff or Slope:	0.906	0.901

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.7	41.0	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4933	66.7	40.7	40.4	1.007
second point	4966	33.3	20.3	20.6	0.988
third point	4983	16.7	10.2	10.3	0.990
as left zero	5000	0.0	0.0	0.1	----
as left span	4933	66.7	40.7	40.7	0.999
Average Correction Factor					0.995

Baseline Corr As found:	40.84	Prev response:	40.61	*% change:	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 11, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:50	End Time (MST):	12:50
Analyzer make:	API T300	Analyzer serial #:	3504

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999901	≥0.995
40.7	40.4	1.0072			
20.3	20.6	0.9876	Slope	0.991149	0.90 - 1.10
10.2	10.3	0.9900			
			Intercept	0.191815	+/-1.5



CO Calibration Plot

Date: October 12, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	October 5, 2023	Last Cal Date:	September 29, 2023
Start time (MST):	10:17	End time (MST):	14:45
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2026
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
N2 Gen Make/Model:	Peak Scientific		Serial Number:	771048317

Analyzer Information

Analyzer make: API T360	Analyzer serial #: 489
Analyzer Range 0 - 2,000 ppm	

	Start	Finish		Start	Finish
Calibration slope:	0.998830	0.998453	Backgd or Offset:	-0.002	-0.175
Calibration intercept:	-5.420000	-5.780000	Coeff or Slope:	1.051	1.062

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-1.1	----
as found span	2920	80.0	1605.9	1601.1	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	4.0	----
high point	2920	80.0	1605.9	1610.0	0.997
second point	2960	40.0	802.9	768.5	1.045
third point	2980	20.0	401.5	400.3	1.003
as left zero	3000	0.0	0.0	5.5	----
as left span	2930	80.0	1600.5	1614.5	0.991
Average Correction Factor					1.015

Baseline Corr As found:	1602.20	Prev response:	1598.57	*% change:	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Instrument installation calibration, linearity adjustment performed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	October 5, 2023	Previous Calibration	September 29, 2023
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:17	End Time (MST)	14:45
Analyzer make	API T360	Analyzer serial #	489

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	4.0	----	Correlation Coefficient	0.999269	≥0.995
1605.9	1610.0	0.9974			
802.9	768.5	1.0448	Slope	0.998453	0.90 - 1.10
401.5	400.3	1.0029			
			Intercept	-5.780000	+/-10



CO₂ Calibration Plot

Date: October 5, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	October 25, 2023	Last Cal Date:	October 5, 2023
Start time (MST):	10:34	End time (MST):	14:52
Reason:	Maintenance		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2026
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
N2 Gen Make/Model:	Peak Scientific		Serial Number:	771048317

Analyzer Information

Analyzer make:	API T360	Analyzer serial #:	489
Analyzer Range	0 - 2,000 ppm		

	Start	Finish		Start	Finish
Calibration slope:	0.998453	1.001286	Backgd or Offset:	-0.175	-0.037
Calibration intercept:	-5.780000	-6.620000	Coeff or Slope:	1.062	0.938

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	95.8	----
as found span	2920	80.0	1605.9	1715.6	0.936
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	-0.2	----
high point	2920	80.0	1605.9	1601.9	1.002
second point	2960	40.0	802.9	802.0	1.001
third point	2980	20.0	401.5	383.7	1.046
as left zero	3000	0.0	0.0	-0.2	----
as left span	2930	80.0	1600.5	1598.0	1.002
Average Correction Factor					1.017

Baseline Corr As found:	1619.80	Prev response:	1597.60	*% change:	1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Nitrogen generator swapped out after as founds. linearity adjustment performed on the midpoint.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	October 25, 2023	Previous Calibration	October 5, 2023
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:34	End Time (MST)	14:52
Analyzer make	API T360	Analyzer serial #	489

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.2	----	Correlation Coefficient	0.999859	≥0.995
1605.9	1601.9	1.0025	Slope	1.001286	0.90 - 1.10
802.9	802.0	1.0012	Intercept	-6.620000	+/-10
401.5	383.7	1.0463			



CO₂ Calibration Plot

Date: October 25, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

PLUVIO CALIBRATION

Version-07-2017

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	October 11, 2023	Last Cal Date:	April 20, 2023
Start time (MST):	11:50	End time (MST):	13:00
Make/Model#:	OTT Pluvio 2	Pluvio Serial #:	363526

Annual Weight Test

Date of Check:	October 11, 2023	Previous Check Date:	April 20, 2023
Weights Serial #:	59781	Passed/Failed:	Pass

Spring Drip Test

Date of Check:		April 20, 2022	Previous Check Date:		October 22, 2021
Start Time:	11:45		Stop Time:		12:25
Nozzle Size Used:	100				
Amount of Water Used:	1 L		Data Logger Total:		24.08 mm
Converted Amount of Water used:	25 mm				

Rainfall (mm) = 1 L / 0.04 m² = 25 mm is what we should be reading.

Fall Drip Test

Date of Check:		October 11, 2023	Previous Check Date:		April 20, 2023
Start Time:	12:30		Stop Time:		13:00
Nozzle Size Used:	100				
Amount of Water Used:	1 L		Data Logger Total:		24.3 mm
Converted Amount of Water used:	25 mm				

Rainfall (mm) = 1 L / 0.04 m² = 25 mm is what we should be reading.

Notes:

Performed By: _____ Tyler Tracksell




OTT Pluvio2 - Guided accuracy test



See test details below:

Date:	2023-10-10 13:07:05
OTT Pluvio2 type:	OTT Pluvio2
Bucket type:	400 cm ²
Serial number / HW Index:	363523 /f3
Firmware version:	V1.50.3
Intensity units:	mm/min
Measured total weight:	4339.92 gram (108.50 mm;4.272 inch)
Information of applied weight:	500.00 gram (12.50 mm;0.492 inch)
Measured test weight:	502.24 gram (12.56 mm;0.494 inch)
Permitted deviation:	+/-4.00 gram (+/-0.10 mm;+/-0.004 inch)
Deviation:	2.24 gram (0.06 mm;0.002 inch)
base weight	3837.68 gram (95.94 mm;3.777 inch)

Test passed: Yes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS19 FIREBAG OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 27, 2023	Previous Calibration:	September 28, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:18	End Time (MST):	13:54
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	1.000000	
799.5	802.0	0.9968			≥0.995
400.3	401.8	0.9962	Slope	1.002865	
200.1	200.6	0.9975			0.90 - 1.10
			Intercept	0.237458	+/-30



SO2 Calibration Plot

Date: October 27, 2023

Location: Firebag





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Firebag Station number: AMS19
 Calibration Date: October 3, 2023 Last Cal Date: September 11, 2023
 Start time (MST): 9:39 End time (MST): 14:34
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.114 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517427
 Removed Cal Gas Conc: 5.114 ppm Rem Gas Exp Date: n/a
 Removed Gas Cyl #: n/a Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1607
 ZAG Make/Model: Teledyne API T701 Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680032
 Converter make: Global Converter serial #: 2022-222
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993765	0.999195	Backgd or Offset:	2.53
Calibration intercept:	-0.161463	-0.101544	Coeff or Slope:	1.122
				1.136

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4922	78.2	80.0	79.7	1.001
as found 2nd point	4961	39.1	40.0	39.8	1.000
as found 3rd point	4980	19.6	20.0	19.8	1.002
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4922	78.2	80.0	79.8	1.002
second point	4961	39.1	40.0	39.9	1.002
third point	4980	19.6	20.0	19.9	1.007
as left zero	5000	0.0	0.0	0.0	----
as left span	4922	78.2	80.0	78.6	1.018
SO2 Scrubber Check	4922	78.3	800.2	0.0	----
Date of last scrubber change:	January 18, 2023			Ave Corr Factor	1.004
Date of last converter efficiency test:	n/a			efficiency	

Baseline Corr As found: 79.9 Prev response: 79.32 *% change: 0.7%
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 0.999195 AF Intercept: -0.201544
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999999

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Ran SOx scrubber check after cal zero. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 11, 2023
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	9:39	End Time (MST):	14:34
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999999	≥0.995
80.0	79.8	1.0023			
40.0	39.9	1.0023	Slope	0.999195	0.90 - 1.10
20.0	19.9	1.0075			
			Intercept	-0.101544	+/-3



H₂S Calibration Plot

Date: October 3, 2023

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	October 27, 2023	Last Cal Date:	September 28, 2023
Start time (MST):	10:18	End time (MST):	13:54
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1607
ZAG Make/Model:	API T701	Serial Number:	1118

Analyzer Information

Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089
Analyzer Range:	0 - 20 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998771	0.998006	Background:	3.48	2.16
Calibration intercept:	-0.052126	-0.059333	Coefficient:	3.796	3.758

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.00	0.29	----
as found span	4919	81.1	17.31	17.73	0.976
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.00	-0.02	----
high point	4919	81.1	17.31	17.24	1.004
second point	4959	40.6	8.66	8.54	1.015
third point	4980	20.3	4.33	4.24	1.021
as left zero	5000	0.0	0.00	-0.02	----
as left span	4919	81.1	17.31	17.31	1.000
Average Correction Factor					1.013
Baseline Corr As found:	17.44	Previous response	17.23	*% change	1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Swapped Hydrogen cylinder after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 27, 2023	Previous Calibration:	September 28, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:18	End Time (MST):	13:54
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.02	----	Correlation Coefficient	0.999970	≥0.995
17.31	17.24	1.0038			
8.66	8.54	1.0145	Slope	0.998006	0.90 - 1.10
4.33	4.24	1.0206			
			Intercept	-0.059333	+/-1.5



THC Calibration Plot

Date: October 27, 2023

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Firebag
Calibration Date: October 26, 2023
Start time (MST): 10:08
Reason: Routine
Station number: AMS 19
Last Cal Date: September 29, 2023
End time (MST): 14:52

Calibration Standards

NO Gas Cylinder #: T2Y1K63
NOX Cal Gas Conc: 51.12 ppm
Removed Cylinder #: n/a
Removed Gas NOX Conc: 51.12 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 30, 2023
NO Cal Gas Conc: 49.40 ppm
Removed Gas Exp Date: n/a
Removed Gas NO Conc: 49.40 ppm
NO gas Diff:
Serial Number: 1607
Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1410661309

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.098	1.078	NO bkgnd or offset:	7.6	7.5
NOX coeff or slope:	0.993	0.995	NOX bkgnd or offset:	8.6	7.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	213.1	217.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998622	0.998388
NO _x Cal Offset:	0.176163	0.736136
NO Cal Slope:	1.000897	0.998826
NO Cal Offset:	0.249004	0.248718
NO ₂ Cal Slope:	0.998565	0.996831
NO ₂ Cal Offset:	-1.744715	-0.602383



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.0	0.0	-1.1	-0.1	-1.0	----	----
as found span	4919	81.0	828.1	800.3	27.9	842.0	809.0	33.9	0.9835	0.9892
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	4999	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
high point	4919	81.0	828.1	800.3	27.9	827.0	799.4	27.9	1.0014	1.0011
second point	4960	40.5	414.0	400.1	13.9	414.8	400.1	14.8	0.9981	1.0000
third point	4980	20.2	206.5	199.6	6.9	207.7	199.9	7.8	0.9943	0.9983
as left zero	4999	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as left span	4919	81.0	828.1	364.9	463.3	826.0	361.6	464.7	1.0026	1.0091
Average Correction Factor									0.9979	0.9998

Corrected As found	NO _x = 843.1 ppb	NO = 809.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.9%
Previous Response	NO _x = 827.2 ppb	NO = 801.2 ppb		*Percent Change	NO = 1.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.5	362.1	463.3	461.7	1.0034	99.7%
2nd GPT point (200 ppb O3)	797.5	578.6	246.8	244.5	1.0093	99.1%
3rd GPT point (100 ppb O3)	797.5	689.0	136.4	135.2	1.0086	99.1%
Average Correction Factor					1.0071	99.3%

Notes:

Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 26, 2023	Previous Calibration:	September 29, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:08	End Time (MST):	14:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
828.1	827.0	1.0014		
414.0	414.8	0.9981		
206.5	207.7	0.9943		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 26, 2023	Previous Calibration:	September 29, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:08	End Time (MST):	14:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
800.3	799.4	1.0011			
400.1	400.1	1.0000			
199.6	199.9	0.9983			
			Slope	0.998826	0.90 - 1.10
			Intercept	0.248718	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 26, 2023	Previous Calibration:	September 29, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:08	End Time (MST):	14:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
463.3	461.7	1.0034		
246.8	244.5	1.0093		
136.4	135.2	1.0086		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 26, 2023

Location: Firebag





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS20 MACKAY RIVER OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 12, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:37	End Time (MST):	10:10
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.4	----	Correlation Coefficient	0.999958	≥0.995
800.3	805.7	0.9933			
400.7	404.9	0.9896	Slope	1.005419	0.90 - 1.10
199.8	205.8	0.9709			
			Intercept	1.910989	+/-30



SO2 Calibration Plot

Date: October 12, 2023

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: MacKay River Station number: AMS20
 Calibration Date: October 5, 2023 Last Cal Date: September 11, 2023
 Start time (MST): 7:02 End time (MST): 11:00
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.12 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC515997
 Removed Cal Gas Conc: 5.12 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1220
 ZAG Make/Model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 43iQ TLE Analyzer serial #: 12124313139
 Converter make: Global Converter serial #: 2022-226
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995379	1.009229	Backgd or Offset: 0.9	0.91
Calibration intercept:	0.407078	0.247261	Coeff or Slope: 0.501	0.501

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4922	78.1	80.0	82.3	0.974
as found 2nd point	4961	39.0	39.9	41.3	0.972
as found 3rd point	4980	19.5	20.0	21.1	0.956
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4922	78.1	80.0	81.0	0.988
second point	4961	39.0	40.0	40.7	0.982
third point	4980	19.5	20.0	20.3	0.986
as left zero	5000	0.0	0.0	0.4	----
as left span	4922	78.1	80.0	80.5	0.994
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:	May 25, 2023			Ave Corr Factor	0.985
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 82.1 Prev response: 80.01 *% change: 2.5%
 Baseline Corr 2nd AF pt: 41.1 AF Slope: 1.025033 AF Intercept: 0.379896
 Baseline Corr 3rd AF pt: 20.9 AF Correlation: 0.999973

* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 11, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:02	End Time (MST):	11:00
Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	12124313139

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999989	
80.0	81.0	0.9881			≥0.995
40.0	40.7	0.9820	Slope	1.009229	
20.0	20.3	0.9860			0.90 - 1.10
			Intercept	0.247261	+/-3



H₂S Calibration Plot

Date: October 5, 2023

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	MacKay River	Station number:	AMS20
Calibration Date:	October 23, 2023	Last Cal Date:	October 5, 2023
Start time (MST):	10:20	End time (MST):	12:00
Reason:	Removal		

Calibration Standards

Cal Gas Concentration:	5.12	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC515997			
Removed Cal Gas Conc:	5.12	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1220
ZAG Make/Model:	Teledyne API 701		Serial Number:	4522

Analyzer Information

Analyzer make:	Thermo 43iQ TLE	Analyzer serial #:	12124313139
Converter make:	Global	Converter serial #:	2022-226
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009229		Backgd or Offset:	0.9	0.91
Calibration intercept:	0.247261		Coeff or Slope:	0.501	0.501

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4922	78.1	80.0	83.7	0.959
as found 2nd point	4961	39.0	39.9	42.0	0.958
as found 3rd point	4980	19.5	20.0	21.5	0.942
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

SO₂ Scrubber Check

Date of last scrubber change:	May 25, 2023	Ave Corr Factor	
Date of last converter efficiency test:		efficiency	

Baseline Corr As found:	83.4	Prev response:	80.96	*% change:	2.9%
Baseline Corr 2nd AF pt:	41.7	AF Slope:	1.041325	AF Intercept:	0.460195
Baseline Corr 3rd AF pt:	21.2	AF Correlation:	0.999977		

* = > +/-5% change initiates investigation


Notes: Removal Due to Lamp intensity increasing beyond limits.

Calibration Performed By: Melissa Lemay

H₂S Calibration Plot

Date: October 23, 2023

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	MacKay River	Station number:	AMS20
Calibration Date:	October 24, 2023	Last Cal Date:	
Start time (MST):	6:50	End time (MST):	9:31
Reason:	Install		

Calibration Standards

Cal Gas Concentration:	5.12	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC515997			
Removed Cal Gas Conc:	5.12	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1220
ZAG Make/Model:	Teledyne API 701		Serial Number:	4522

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117
Converter make:	Global	Converter serial #:	2022-226
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		1.001813	Backgd or Offset:		3.07
Calibration intercept:		0.239528	Coeff or Slope:		1.083

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	78.1	80.0	80.3	0.997
second point	4961	39.0	40.0	40.4	0.989
third point	4980	19.5	20.0	20.5	0.975
as left zero	5000	0.0	0.0	0.2	----
as left span	4922	78.1	80.0	79.3	1.009

SO₂ Scrubber Check

Date of last scrubber change:	May 25, 2023	Ave Corr Factor	0.987
Date of last converter efficiency test:		efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: Analyzer installed due to removed analyzer not working properly. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 24, 2023	Previous Calibration:	
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	6:50	End Time (MST):	9:31
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999959	
80.0	80.3	0.9967			≥0.995
40.0	40.4	0.9893	Slope	1.001813	
20.0	20.5	0.9749			0.90 - 1.10
			Intercept	0.239528	+/-3



H₂S Calibration Plot

Date: October 24, 2023

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 12, 2023	Previous Calibration:	September 12, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:36	End Time (MST):	10:09
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.03	----	Correlation Coefficient	0.999958	≥0.995
17.34	17.30	1.0023			
8.68	8.75	0.9922	Slope	0.994974	0.90 - 1.10
4.33	4.43	0.9773			
			Intercept	0.076785	+/-1.5



THC Calibration Plot

Date: October 12, 2023

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: MacKay River Station number: AMS20
Calibration Date: October 6, 2023 Last Cal Date: September 14, 2023
Start time (MST): 6:45 End time (MST): 10:58
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 1220
ZAG make/model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.536	1.486	NO bkgnd or offset:	4.3	4.1
NOX coeff or slope:	0.992	0.994	NOX bkgnd or offset:	4.3	4.1
NO2 coeff or slope:	0.995	0.995	Reaction cell Press:	189.4	189.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996495	0.995380
NO _x Cal Offset:	3.649884	3.749814
NO Cal Slope:	0.997991	0.994536
NO Cal Offset:	2.330879	2.591219
NO ₂ Cal Slope:	1.001543	1.007687
NO ₂ Cal Offset:	-0.862665	-0.458883



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	----	----
as found span	4917	83.3	819.5	800.3	19.2	849.0	828.0	21.0	0.9652	0.9665
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.1	----	----
high point	4917	83.3	819.5	800.3	19.2	818.0	797.6	20.4	1.0018	1.0034
second point	4956	41.7	410.4	400.8	9.6	413.0	401.4	11.6	0.9938	0.9986
third point	4979	20.8	204.6	199.9	4.8	211.7	204.7	7.0	0.9666	0.9763
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4917	83.3	819.5	442.7	376.8	817.6	437.1	380.4	1.0023	1.0128
Average Correction Factor									0.9874	0.9928

Corrected As found	NO _x = 849.0 ppb	NO = 828.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 3.4%
Previous Response	NO _x = 820.2 ppb	NO = 801.0 ppb		*Percent Change	NO = 3.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.9	437.3	376.8	379.4	0.9930	100.7%
2nd GPT point (200 ppb O3)	794.9	608.9	205.2	206.4	0.9940	100.6%
3rd GPT point (100 ppb O3)	794.9	698.4	115.7	115.2	1.0040	99.6%
Average Correction Factor					0.9970	100.3%

Notes:

No maintenance done. Span adjusted.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 6, 2023	Previous Calibration:	September 14, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	6:45	End Time (MST):	10:58
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
819.5	818.0	1.0018		
410.4	413.0	0.9938		
204.6	211.7	0.9666		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 6, 2023	Previous Calibration:	September 14, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	6:45	End Time (MST):	10:58
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
800.3	797.6	1.0034			
400.8	401.4	0.9986			
199.9	204.7	0.9763			
			Slope	0.994536	0.90 - 1.10
			Intercept	2.591219	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 6, 2023	Previous Calibration:	September 14, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	6:45	End Time (MST):	10:58
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
376.8	379.4	0.9930		
205.2	206.4	0.9940		
115.7	115.2	1.0040		



NO_x Calibration Plot

Date: October 6, 2023

Location: MacKay River





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS21
CONKLIN
OCTOBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	October 20, 2023	Last Cal Date:	September 26, 2023
Start time (MST):	11:21	End time (MST):	15:33
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.93	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	<u>CC259455</u>			
Removed Cal Gas Conc:	49.93	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3810
ZAG Make/Model:	Teledyne API 701		Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000644	0.997591	Backgd or Offset:	26.8	27.4
Calibration intercept:	0.775866	0.795745	Coeff or Slope:	0.883	0.892

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5005	0.0	0.0	0.2	----
as found span	4920	80.2	800.8	798.5	1.003
as found 2nd point	4960	40.1	400.4	400.4	1.000
as found 3rd point	4980	20.0	200.1	201.7	0.992
new cylinder response					
calibrator zero	5005	0.0	0.0	0.3	----
high point	4920	80.2	800.8	799.5	1.002
second point	4960	40.1	400.4	400.4	1.000
third point	4980	20.0	200.1	201.0	0.996
as left zero	5005	0.0	0.0	0.2	----
as left span	4920	80.2	800.8	799.5	1.002
Average Correction Factor					0.999

Baseline Corr As found:	798.30	Previous response	802.14	*% change	-0.5%
Baseline Corr 2nd AF pt:	400.20	AF Slope:	0.996106	AF Intercept:	1.215744
Baseline Corr 3rd AF pt:	201.50	AF Correlation:	0.999992		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 26, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:21	End Time (MST):	15:33
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999998	≥0.995
800.8	799.5	1.0017			
400.4	400.4	1.0001	Slope	0.997591	0.90 - 1.10
200.1	201.0	0.9956			
			Intercept	0.795745	+/-30



SO2 Calibration Plot

Date: October 20, 2023

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Conklin Station number: AMS21
 Calibration Date: October 17, 2023 Last Cal Date: September 14, 2023
 Start time (MST): 11:07 End time (MST): 15:56
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.00 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC501204
 Removed Cal Gas Conc: 5.00 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3810
 ZAG Make/Model: Teledyne API 701H Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1236656116
 Converter make: CD-Nova 101 Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007429	0.994857	Backgd or Offset:	2.4	2.3
Calibration intercept:	0.140000	0.180000	Coeff or Slope:	0.958	0.947

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.0	80.0	80.2	0.998
as found 2nd point	4960	40.0	40.0	40.6	0.985
as found 3rd point	4980	20.0	20.0	20.2	0.990
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.0	80.0	79.7	1.004
second point	4960	40.0	40.0	40.1	0.998
third point	4980	20.0	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.4	----
as left span	4920	80.0	80.0	78.4	1.020
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.999
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 80.2 Prev response: 80.73 *% change: -0.7%
 Baseline Corr 2nd AF pt: 40.6 AF Slope: 1.002571 AF Intercept: 0.160000
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999953

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 14, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:07	End Time (MST):	15:56
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999992	≥0.995
80.0	79.7	1.0038			
40.0	40.1	0.9975	Slope	0.994857	0.90 - 1.10
20.0	20.1	0.9950			
			Intercept	0.180000	+/-3



TRS Calibration Plot

Date: October 17, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	October 20, 2023	Last Cal Date:	September 26, 2023
Start time (MST):	11:21	End time (MST):	15:33
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.29E-04	2.32E-04	NMHC SP Ratio:	5.09E-05
CH ₄ Retention time:	12.00	12.00	NMHC Peak Area:	179681
				172466

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	16.68	1.027
as found 2nd point	4960	40.1	8.56	8.30	1.032
as found 3rd point	4980	20.0	4.28	4.17	1.026
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.13	1.000
second point	4960	40.1	8.56	8.53	1.003
third point	4980	20.0	4.27	4.29	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.11	1.001
Average Correction Factor					1.000
Baseline Corr AF:	16.68	Prev response	17.10	*% change	-2.5%
Baseline Corr 2nd AF:	8.3	AF Slope:	0.973575	AF Intercept:	-0.007854
Baseline Corr 3rd AF:	4.2	AF Correlation:	0.999991	* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0	
as found span	4920	80.2	9.14	8.79	1.039
as found 2nd point	4960	40.1	4.57	4.40	1.040
as found 3rd point	4980	20.0	2.28	2.21	1.031
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.14	0.999
second point	4960	40.1	4.57	4.58	0.998
third point	4980	20.0	2.28	2.31	0.987
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.14	1.000
Average Correction Factor					0.995
Baseline Corr AF:	8.79	Prev response	9.09	*% change	-3.4%
Baseline Corr 2nd AF:	4.4	AF Slope:	0.961354	AF Intercept:	0.006594
Baseline Corr 3rd AF:	2.2	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.88	1.013
as found 2nd point	4960	40.1	3.99	3.90	1.023
as found 3rd point	4980	20.0	2.00	1.96	1.021
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.99	0.999
second point	4960	40.1	3.99	3.95	1.010
third point	4980	20.0	1.99	1.98	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.98	1.001
Average Correction Factor					1.005
Baseline Corr AF:	7.88	Prev response	8.01	*% change	-1.6%
Baseline Corr 2nd AF:	3.90	AF Slope:	0.987302	AF Intercept:	-0.014048
Baseline Corr 3rd AF:	1.96	AF Correlation:	0.999968	* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998108	0.999923
THC Cal Offset:	0.004967	-0.000623
CH ₄ Cal Slope:	1.004174	1.000697
CH ₄ Cal Offset:	-0.009247	-0.013854
NMHC Cal Slope:	0.992770	0.999521
NMHC Cal Offset:	0.013615	0.012632

Notes: Changed the inlet filter and H2 cylinder after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 26, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:21	End Time (MST):	15:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	≥ 0.995			
17.13	17.13	0.9995						
8.56	8.53	1.0035				Slope	0.999923	0.90 - 1.10
4.27	4.29	0.9960						
			Intercept	-0.000623	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 26, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:21	End Time (MST):	15:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999965	≥0.995
7.99	7.99	0.9992			
3.99	3.95	1.0101			
1.99	1.98	1.0059			
			Slope	1.000697	0.90 - 1.10
			Intercept	-0.013854	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 26, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:21	End Time (MST):	15:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999989	≥ 0.995
9.14	9.14	0.9995			
4.57	4.58	0.9982			
2.28	2.31	0.9871			
			Slope	0.999521	0.90 - 1.10
			Intercept	0.012632	+/-0.5



NMHC Calibration Plot

Date: October 20, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	October 24, 2023	Last Cal Date:	NA
Start time (MST):	10:10	End time (MST):	15:22
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	NA	4.01E-04	NMHC SP Ratio:	NA	9.22E-05
CH ₄ Retention time:	NA	14.6	NMHC Peak Area:	NA	99089
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.06	1.004
second point	4960	40.1	8.56	8.56	1.000
third point	4980	20.0	4.28	4.32	0.991
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.06	1.004

				Average Correction Factor	0.998
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.11	1.004
second point	4960	40.1	4.57	4.57	1.000
third point	4980	20.0	2.28	2.31	0.989
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.08	1.006
Average Correction Factor					0.998
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.95	1.004
second point	4960	40.1	3.99	3.99	1.001
third point	4980	20.0	2.00	2.01	0.993
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.98	1.000
Average Correction Factor					0.999
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		0.995269
THC Cal Offset:		0.028364
CH ₄ Cal Slope:		0.995272
CH ₄ Cal Offset:		0.011356
NMHC Cal Slope:		0.995141
NMHC Cal Offset:		0.017008

Notes: Installing a new NMHC instrument. Optimized flame. Enabled "use zero chromatogram". Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 24, 2023	Previous Calibration:	NA
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:10	End Time (MST):	15:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999987	≥ 0.995			
17.13	17.06	1.0038						
8.56	8.56	1.0003				Slope	0.995269	0.90 - 1.10
4.28	4.32	0.9908						
			Intercept	0.028364	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 24, 2023	Previous Calibration:	NA
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:10	End Time (MST):	15:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999991	≥0.995
7.99	7.95	1.0040			
3.99	3.99	1.0005			
2.00	2.01	0.9933			
			Slope	0.995272	0.90 - 1.10
			Intercept	0.011356	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 24, 2023	Previous Calibration:	NA
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:10	End Time (MST):	15:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999983	≥ 0.995			
9.14	9.11	1.0038						
4.57	4.57	1.0004				Slope	0.995141	0.90 - 1.10
2.28	2.31	0.9886						
			Intercept	0.017008	± 0.5			



NMHC Calibration Plot

Date: October 24, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	October 25, 2023	Last Cal Date:	October 24, 2023
Start time (MST):	11:50	End time (MST):	15:50
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
C3H8 Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
Removed C3H8 Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	NA	4.01E-04	NMHC SP Ratio:	NA
CH4 Retention time:	NA	14.6	NMHC Peak Area:	NA
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				9.22E-05
				99089
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.04	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.03	1.006
second point	4960	40.1	8.56	8.51	1.006
third point	4980	20.0	4.28	4.31	0.993
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.06	1.004

Average Correction Factor				1.001
Baseline Corr AF:	17.04	Prev response	17.07	*% change -0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	8.93	1.024
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.07	1.007
second point	4960	40.1	4.57	4.54	1.007
third point	4980	20.0	2.28	2.30	0.991
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.11	1.003
Average Correction Factor					1.002
Baseline Corr AF:	8.93	Prev response	9.11	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	8.12	0.984
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.95	1.004
second point	4960	40.1	3.99	3.98	1.005
third point	4980	20.0	2.00	2.01	0.994
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.95	1.005
Average Correction Factor					1.001
Baseline Corr AF:	8.12	Prev response	7.96	*% change	2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.995269	0.993314
THC Cal Offset:	0.028364	0.021764
CH ₄ Cal Slope:	0.995272	0.995000
CH ₄ Cal Offset:	0.011356	0.007556
NMHC Cal Slope:	0.995141	0.991515
NMHC Cal Offset:	0.017008	0.015007

Notes: Some baseline dips occurred in last 24 hrs, found retention time moved up. Some maintenance items done, no change to chromatography, new zero chromatogram captured.

Calibration Performed By: Kelly Baragar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 25, 2023	Previous Calibration:	October 24, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:50	End Time (MST):	15:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995			
17.13	17.03	1.0056						
8.56	8.51	1.0058				Slope	0.993314	0.90 - 1.10
4.28	4.31	0.9926						
			Intercept	0.021764	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 25, 2023	Previous Calibration:	October 24, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:50	End Time (MST):	15:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	≥0.995			
7.99	7.95	1.0041						
3.99	3.98	1.0045				Slope	0.995000	0.90 - 1.10
2.00	2.01	0.9943						
			Intercept	0.007556	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 25, 2023	Previous Calibration:	October 24, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:50	End Time (MST):	15:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999980	≥ 0.995			
9.14	9.07	1.0072						
4.57	4.54	1.0068				Slope	0.991515	0.90 - 1.10
2.28	2.30	0.9912						
			Intercept	0.015007	± 0.5			



NMHC Calibration Plot

Date: October 25, 2023

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Conklin
Calibration Date: October 27, 2023
Start time (MST): 10:02
Reason: Routine
Station number: AMS21
Last Cal Date: September 21, 2023
End time (MST): 14:45

Calibration Standards

NO Gas Cylinder #: T2Y1P1H
NOX Cal Gas Conc: 51.09 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 51.09 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.39 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 50.39 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 691

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1501663731

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.104	1.094	NO bkgnd or offset:	11	10.9
NOX coeff or slope:	1.000	1.000	NOX bkgnd or offset:	11.1	11.0
NO2 coeff or slope:	1.000	0.999	Reaction cell Press:	167.3	167.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.993898	0.999869
NO _x Cal Offset:	2.303154	2.584550
NO Cal Slope:	0.994967	1.002036
NO Cal Offset:	1.360599	1.361881
NO ₂ Cal Slope:	0.989964	1.004241
NO ₂ Cal Offset:	-0.688924	-0.358712



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.5	-0.1	----	----
as found span	4921	79.4	811.2	800.1	11.1	821.5	808.0	13.5	0.9875	0.9902
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.4	-0.1	----	----
high point	4921	79.4	811.2	800.1	11.1	812.1	802.2	10.0	0.9989	0.9974
second point	4960	39.7	405.7	400.1	5.6	410.0	403.3	6.7	0.9895	0.9921
third point	4980	19.8	202.3	199.6	2.8	207.8	203.0	4.8	0.9737	0.9830
as left zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.4	-0.1	----	----
as left span	4921	79.4	811.2	389.0	422.2	813.8	391.0	422.7	0.9969	0.9950
Average Correction Factor									0.9874	0.9909

Corrected As found	NO _x = 822.2 ppb	NO = 808.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.7%
Previous Response	NO _x = 808.6 ppb	NO = 797.4 ppb		*Percent Change	NO = 1.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI: ;

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	803.1	392.0	422.2	423.5	0.9970	100.3%
2nd GPT point (200 ppb O3)	803.1	597.8	216.4	217.7	0.9941	100.6%
3rd GPT point (100 ppb O3)	803.1	700.7	113.5	112.8	1.0063	99.4%
Average Correction Factor					0.9991	100.1%

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 27, 2023	Previous Calibration:	September 21, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:02	End Time (MST):	14:45
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.5	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
811.2	812.1	0.9989		
405.7	410.0	0.9895		
202.3	207.8	0.9737		
			0.999934	
			0.999869	
			2.584550	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 27, 2023	Previous Calibration:	September 21, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:02	End Time (MST):	14:45
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.4	----	Correlation Coefficient	0.999978	≥0.995
800.1	802.2	0.9974			
400.1	403.3	0.9921	Slope	1.002036	0.90 - 1.10
199.6	203.0	0.9830			
			Intercept	1.361881	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 27, 2023	Previous Calibration:	September 21, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:02	End Time (MST):	14:45
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥0.995
422.2	423.5	0.9970		
216.4	217.7	0.9941	Slope	0.90 - 1.10
113.5	112.8	1.0063		
			Intercept	+/-20

NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 27, 2023

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Conklin Station number: AMS21
 Calibration Date: October 27, 2023 Last Cal Date: September 11, 2023
 Start time (MST): 10:11 End time (MST): 14:00
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T750 Serial Number: 72
 ZAG Make/Model: Teledyne API 751H Serial Number: 77

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1501663734
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001714	1.001314	Backgd or Offset:	-2.0	-0.3
Calibration intercept:	0.400000	1.520000	Coeff or Slope:	1.005	1.005

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	1.2	----
as found span	5000	1564.0	400.0	401.2	0.997
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	5000	950.9	400.0	400.6	0.999
second point	5000	804.0	200.0	204.5	0.978
third point	5000	703.6	100.0	102.0	0.980
as left zero	5000	0.0	0.0	0.2	----
as left span	5000	936.0	400.0	401.4	0.997
Average Correction Factor					0.986

Baseline Corr As found:	400.0	Previous response	401.1	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Used a portable calibrator and zero air for the calibration. Changed the inlet filter after as founds. Adjusted the zero only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 27, 2023	Previous Calibration:	September 11, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:11	End Time (MST):	14:00
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999861	≥0.995
400.0	400.6	0.9985			
200.0	204.5	0.9780	Slope	1.001314	0.90 - 1.10
100.0	102.0	0.9804			
			Intercept	1.520000	+/- 5



O₃ Calibration Plot

Date: October 27, 2023

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Conklin Station number: AMS 21
 Calibration Date: October 27, 2023 Last Cal Date: September 26, 2023
 Start time (MST): 14:03 End time (MST): 14:27

Analyzer Make: API T640 S/N: 326
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat S/N: 388755
 Temp/RH standard: Alicat S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-4.1	-4.7	-4.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	715.4	716.95	715.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.93	5.12	4.93	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: October 27, 2023 Last Cal Date: September 26, 2023
 PM w/o HEPA: 11.2 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>September 26, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 26, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: PMT check was completed in September. Leak check passed, no adjustments made.

Calibration by: Max Farrell



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER
OCTOBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	October 5, 2023	Last Cal Date:	September 28, 2023
Start time (MST):	9:12	End time (MST):	12:46
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.11	ppm	Cal Gas Exp Date:	January 18, 2029
Cal Gas Cylinder #:	CC281519			
Removed Cal Gas Conc:	50.11	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005095	0.996104	Backgd or Offset:	21.4	20.7
Calibration intercept:	2.463336	1.485197	Coeff or Slope:	0.989	0.997

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	1.2	----
as found span	4920	79.8	799.8	805.0	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	1.3	----
high point	4920	79.8	799.8	797.9	1.002
second point	4960	39.9	399.9	400.4	0.999
third point	4980	20.0	200.4	201.0	0.997
as left zero	5000	0.0	0.0	1.6	----
as left span	4920	79.8	799.8	796.5	1.004
Average Correction Factor					0.999

Baseline Corr As found:	803.80	Previous response	806.33	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 28, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:12	End Time (MST):	12:46
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.3	----	Correlation Coefficient	0.999999	≥0.995
799.8	797.9	1.0024			
399.9	400.4	0.9987	Slope	0.996104	0.90 - 1.10
200.4	201.0	0.9972			
			Intercept	1.485197	+/-30



SO2 Calibration Plot

Date: October 5, 2023

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Janvier Station number: AMS22
 Calibration Date: October 18, 2023 Last Cal Date: September 21, 2023
 Start time (MST): 9:18 End time (MST): 14:18
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.03 ppm Cal Gas Exp Date: April 16, 2022
 Cal Gas Cylinder #: DT0018680
 Removed Cal Gas Conc: 5.03 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680031
 Converter make: CDN-101 Converter serial #: 587
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003650	1.011516	Backgd or Offset: 3.04	3.04
Calibration intercept:	-0.039106	0.060524	Coeff or Slope: 1.161	1.161

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4920	79.5	80.0	73.7	1.087
as found 2nd point	4960	39.8	40.0	36.9	1.088
as found 3rd point	4980	19.9	20.0	18.5	1.088
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	79.5	80.0	80.7	0.991
second point	4960	39.8	40.0	41.2	0.972
third point	4980	19.9	20.0	20.1	0.996
as left zero	5000	0.0	0.0	0.3	----
as left span	4920	79.5	80.0	81.1	0.986
SO2 Scrubber Check	4920	79.8	798.0	0.3	----

Date of last scrubber change:	Ave Corr Factor	0.986
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 73.6 Prev response: 80.24 *% change: -9.0%
 Baseline Corr 2nd AF pt: 36.8 AF Slope: 0.920202 AF Intercept: 0.082530
 Baseline Corr 3rd AF pt: 18.4 AF Correlation: 1.000000

* = > +/-5% change initiates investigation

Notes: As found span is 9% low, completed an investigation and searched for leaks. Throughout the calibration the scrubber beads must have dried enough to reach the target span as there were no adjustments made and the span reached target. See docit for more info.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 21, 2023
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	9:18	End Time (MST):	14:18
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999847	≥0.995
80.0	80.7	0.9911			
40.0	41.2	0.9719	Slope	1.011516	0.90 - 1.10
20.0	20.1	0.9960			
			Intercept	0.060524	+/-3



TRS Calibration Plot

Date: October 18, 2023

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	October 5, 2023	Last Cal Date:	September 28, 2023
Start time (MST):	9:13	End time (MST):	12:47
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH ₄ Cal Gas Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
C ₃ H ₈ Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
Removed C ₃ H ₈ Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.240E-04	2.200E-04	NMHC SP Ratio:	4.50E-05
CH ₄ Retention time:	13.6	13.8	NMHC Peak Area:	203170
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				4.43E-05
				206448
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	17.70	0.970
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.21	0.998
second point	4960	39.9	8.59	8.48	1.012
third point	4980	20.0	4.30	4.23	1.018
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	16.93	1.015

			Average Correction Factor	1.009
Baseline Corr AF:	17.70	Prev response	17.18	*% change 3.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	9.58	0.955
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.15	9.13	1.002
second point	4960	39.9	4.57	4.53	1.011
third point	4980	20.0	2.29	2.25	1.021
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.15	9.05	1.011
Average Correction Factor					1.011
Baseline Corr AF:	9.58	Prev response	9.15	*% change	4.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	8.12	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.09	0.992
second point	4960	39.9	4.01	3.96	1.013
third point	4980	20.0	2.01	1.98	1.014
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	7.88	1.018
Average Correction Factor					1.006
Baseline Corr AF:	8.12	Prev response	8.03	*% change	1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000655	1.003343
THC Cal Offset:	-0.005800	-0.059000
CH ₄ Cal Slope:	1.000614	1.008361
CH ₄ Cal Offset:	0.000436	-0.033775
NMHC Cal Slope:	1.000704	0.998915
NMHC Cal Offset:	-0.007036	-0.024625

Notes: Changed the inlet filter and the N2 cylinder after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 28, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:13	End Time (MST):	12:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999929	≥ 0.995			
17.17	17.21	0.9976						
8.59	8.48	1.0120				Slope	1.003343	0.90 - 1.10
4.30	4.23	1.0179						
			Intercept	-0.059000	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 28, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:13	End Time (MST):	12:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999866	≥0.995
8.03	8.09	0.9923			
4.01	3.96	1.0132			
2.01	1.98	1.0137			
			Slope	1.008361	0.90 - 1.10
			Intercept	-0.033775	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 28, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:13	End Time (MST):	12:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999965	≥ 0.995			
9.15	9.13	1.0022						
4.57	4.53	1.0107				Slope	0.998915	0.90 - 1.10
2.29	2.25	1.0211						
			Intercept	-0.024625	± 0.5			



NMHC Calibration Plot

Date: October 5, 2023

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: October 17, 2023 Last Cal Date: September 19, 2023
Start time (MST): 10:00 End time (MST): 14:39
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC424183 Cal Gas Expiry Date: April 16, 2023
NOX Cal Gas Conc: 48.60 ppm NO Cal Gas Conc: 48.60 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.60 ppm Removed Gas NO Conc: 48.60 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3806
ZAG make/model: Teledyne API T701 Serial Number: 201

Analyzer Information

Analyzer make: API T200 Analyzer serial #: 833
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.826	0.826	NO bkgnd or offset:	-5.6	-5.6
NOX coeff or slope:	0.815	0.815	NOX bkgnd or offset:	-3.9	-3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	5.0	5.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998659	0.991915
NO _x Cal Offset:	2.728523	2.589941
NO Cal Slope:	0.998929	0.989600
NO Cal Offset:	2.308958	2.050558
NO ₂ Cal Slope:	0.997414	0.999193
NO ₂ Cal Offset:	-0.454360	0.162793



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.8	2.0	-0.3	----	----
as found span	4918	82.3	799.9	799.9	0.0	797.8	789.4	8.3	1.0026	1.0133
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	2.0	2.1	-0.1	----	----
high point	4918	82.3	799.9	799.9	0.0	795.6	793.4	2.2	1.0054	1.0082
second point	4959	41.2	400.4	400.4	0.0	400.5	399.1	1.4	0.9999	1.0034
third point	4980	20.6	200.2	200.2	0.0	201.5	199.6	1.9	0.9936	1.0030
as left zero	5000	0.0	0.0	0.0	0.0	1.9	2.2	-0.3	----	----
as left span	4918	82.3	799.9	406.7	393.2	792.3	400.2	392.2	1.0096	1.0163
Average Correction Factor									0.9996	1.0049

Corrected As found	NO _x = 796.0 ppb	NO = 787.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.7%	
Previous Response	NO _x = 801.6 ppb	NO = 801.4 ppb		*Percent Change	NO = -1.8%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	792.1	398.9	393.2	392.8	1.0010	99.9%
2nd GPT point (200 ppb O3)	792.1	599.2	192.9	193.4	0.9974	100.3%
3rd GPT point (100 ppb O3)	792.1	694.9	97.2	97.3	0.9990	100.1%
Average Correction Factor					0.9991	100.1%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 19, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:00	End Time (MST):	14:39
Analyzer make:	API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	2.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.9	795.6	1.0054		
400.4	400.5	0.9999		
200.2	201.5	0.9936		
			0.999997	
			0.991915	
			2.589941	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 19, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:00	End Time (MST):	14:39
Analyzer make:	API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	2.1	----	Correlation Coefficient 0.999997	≥0.995	
799.9	793.4	1.0082			
400.4	399.1	1.0034			
200.2	199.6	1.0030			
			Slope	0.989600	0.90 - 1.10
			Intercept	2.050558	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 19, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:00	End Time (MST):	14:39
Analyzer make:	API T200	Analyzer serial #:	833

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
393.2	392.8	1.0010		
192.9	193.4	0.9974		
97.2	97.3	0.9990		



NO_x Calibration Plot

Date: October 17, 2023

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	October 20, 2023	Last Cal Date:	September 7, 2023
Start time (MST):	9:18	End time (MST):	12:12
Reason:	Routine		

Calibration Standards

O ₃ generation mode:	Photometer		
Calibrator Make/Model:	Teledyne API T700	Serial Number:	3806
ZAG Make/Model:	Teledyne API T701H	Serial Number:	201

Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #:	7046
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000914	0.996857	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	0.840000	1.800000	Coeff or Slope:	1.021	1.021

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.6	----
as found span	4895	905.3	400.0	401.2	0.997
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	1.1	----
high point	4895	905.3	400.0	400.1	1.000
second point	4895	756.7	200.0	201.8	0.991
third point	4895	656.1	100.0	102.0	0.980
as left zero	5000	800.0	0.0	1.5	----
as left span	4895	904.3	400.0	403.3	0.992
Average Correction Factor					0.990

Baseline Corr As found:	400.6	Previous response	401.2	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 7, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:18	End Time (MST):	12:12
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.1	----	Correlation Coefficient	0.999985	
400.0	400.1	0.9998			≥0.995
200.0	201.8	0.9911	Slope	0.996857	
100.0	102.0	0.9804			0.90 - 1.10
			Intercept	1.800000	+/- 5



O₃ Calibration Plot

Date: October 20, 2023

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: October 20, 2023 Last Cal Date: September 28, 2023
 Start time (MST): 12:16 End time (MST): 13:13

Analyzer Make: Teledyne API T640 S/N: 325
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	14.1	13.5	14.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	713.8	714.4	713.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	5.01	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>October 20, 2023</u>	Last Cal Date: <u>September 28, 2023</u>			
	PM w/o HEPA: <u>9.8</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.1	11.1	11.1	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>16.2</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>October 20, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>October 20, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: July 26, 2023
 Date RH/T Sensor Cleaned: July 26, 2023

Notes:

No adjustments needed. Leak check passed.

Calibration by: Max Farrell



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23 FORT HILLS OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Fort Hills Station number: AMS23
 Calibration Date: October 17, 2023 Last Cal Date: September 7, 2023
 Start time (MST): 8:20 End time (MST): 11:16
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.76 ppm Cal Gas Exp Date: January 5, 2025
 Cal Gas Cylinder #: CC281425
 Removed Cal Gas Conc: 49.76 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 451
 ZAG Make/Model: API T701 Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 1160290012
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002497	0.999149	Backgd or Offset:	17.8	17.8
Calibration intercept:	-0.744078	-0.923240	Coeff or Slope:	1.031	1.040

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4920	80.3	799.1	790.1	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.3	799.1	797.9	1.002
second point	4960	40.2	400.1	398.6	1.004
third point	4980	20.1	200.0	197.7	1.012
as left zero	5000	0.0	0.0	-0.1	----
as left span	4920	80.3	799.1	799.6	0.999
Average Correction Factor					1.006

Baseline Corr As found: 790.20 Previous response 800.35 *% change -1.3%
 Baseline Corr 2nd AF pt: NA AF Slope: AF Intercept:
 Baseline Corr 3rd AF pt: NA AF Correlation:

* = > +/-5% change initiates investigation

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 7, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:20	End Time (MST):	11:16
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999992	≥0.995
799.1	797.9	1.0015			
400.1	398.6	1.0036	Slope	0.999149	0.90 - 1.10
200.0	197.7	1.0118			
			Intercept	-0.923240	+/-30



SO2 Calibration Plot

Date: October 17, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Hills Station number: AMS23
 Calibration Date: October 2, 2023 Last Cal Date: September 6, 2023
 Start time (MST): 8:30 End time (MST): 12:18
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.20 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517372
 Removed Cal Gas Conc: 5.20 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 451
 ZAG Make/Model: API T701 Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1300156232
 Converter make: CDN-101 Converter serial #: 594
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001459	1.004317	Backgd or Offset: 1.19	1.19
Calibration intercept:	0.061824	-0.038148	Coeff or Slope: 1.124	1.124

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4923	77.0	80.0	77.0	1.039
as found 2nd point	4962	38.5	40.0	38.7	1.034
as found 3rd point	4981	19.2	19.9	19.7	1.013
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.0	80.0	80.2	0.998
second point	4962	38.5	40.0	40.5	0.988
third point	4981	19.2	19.9	19.7	1.013
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.0	80.0	80.8	0.990
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.999
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 77.0 Prev response: 80.18 *% change: -4.1%
 Baseline Corr 2nd AF pt: 38.7 AF Slope: 0.960605 AF Intercept: 0.241194
 Baseline Corr 3rd AF pt: 19.7 AF Correlation: 0.999952

* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 2, 2023	Previous Calibration:	September 6, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:30	End Time (MST):	12:18
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999933	
80.0	80.2	0.9975			≥0.995
40.0	40.5	0.9876	Slope	1.004317	
19.9	19.7	1.0126			0.90 - 1.10
			Intercept	-0.038148	+/-3



TRS Calibration Plot

Date: October 2, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	October 17, 2023	Last Cal Date:	September 7, 2023
Start time (MST):	8:20	End time (MST):	11:13
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
Removed C ₃ H ₈ Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.34E-04	2.34E-04	NMHC SP Ratio:	5.10E-05	5.10E-05
CH ₄ Retention time:	13.0	13.0	NMHC Peak Area:	180460	180460
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	17.29	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.30	0.994
second point	4960	40.2	8.61	8.67	0.993
third point	4980	20.1	4.30	4.34	0.992
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.19	17.32	0.993
Average Correction Factor					0.993

Baseline Corr AF:	17.29	Prev response	17.27	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.20	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.21	0.994
second point	4960	40.2	4.59	4.63	0.990
third point	4980	20.1	2.29	2.34	0.980
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	9.16	9.21	0.994
Average Correction Factor					0.988
Baseline Corr AF:	9.20	Prev response	9.21	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	8.09	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.09	0.993
second point	4960	40.2	4.02	4.03	0.998
third point	4980	20.1	2.01	2.00	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.03	8.11	0.990
Average Correction Factor					0.999
Baseline Corr AF:	8.09	Prev response	8.05	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.005421	1.006154
THC Cal Offset:	-0.016398	0.005584
CH ₄ Cal Slope:	1.005916	1.007912
CH ₄ Cal Offset:	-0.032039	-0.014057
NMHC Cal Slope:	1.003864	1.004487
NMHC Cal Offset:	0.017642	0.017645

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 7, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:20	End Time (MST):	11:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
17.19	17.30	0.9938						
8.61	8.67	0.9927				Slope	1.006154	0.90 - 1.10
4.30	4.34	0.9916						
			Intercept	0.005584	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 7, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:20	End Time (MST):	11:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995
8.03	8.09	0.9929			
4.02	4.03	0.9979			
2.01	2.00	1.0054			
			Slope	1.007912	0.90 - 1.10
			Intercept	-0.014057	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 7, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:20	End Time (MST):	11:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999983	≥ 0.995			
9.16	9.21	0.9945						
4.59	4.63	0.9904				Slope	1.004487	0.90 - 1.10
2.29	2.34	0.9798						
			Intercept	0.017645	± 0.5			



NMHC Calibration Plot

Date: October 17, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Hills Station number: AMS23
Calibration Date: October 11, 2023 Last Cal Date: September 5, 2023
Start time (MST): 7:00 End time (MST): 11:19
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC332703 Cal Gas Expiry Date: January 28, 2024
NOX Cal Gas Conc: 49.7 ppm NO Cal Gas Conc: 49.7 ppm
Removed Cylinder #: N/A Removed Gas Exp Date: N/A
Removed Gas NOX Conc: 49.7 ppm Removed Gas NO Conc: 49.7 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 451
ZAG make/model: Teledyne API T701 Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1152430007
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.076	1.089	NO bkgnd or offset:	3.3	3.3
NOX coeff or slope:	0.991	0.993	NOX bkgnd or offset:	3.7	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	161.6	161.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996239	1.000322
NO _x Cal Offset:	-0.017022	0.004208
NO Cal Slope:	0.998440	0.999781
NO Cal Offset:	-1.437143	-0.956508
NO ₂ Cal Slope:	1.016048	1.002559
NO ₂ Cal Offset:	0.951555	-0.119059



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.3	-0.1	----	----
as found span	4920	80.5	800.2	800.2	0.0	790.0	788.9	0.9	1.013	1.014
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
high point	4920	80.5	800.2	800.2	0.0	799.9	799.1	0.8	1.000	1.001
second point	4960	40.2	399.6	399.6	0.0	400.9	399.0	1.9	0.997	1.001
third point	4980	20.1	199.8	199.8	0.0	199.6	197.6	1.9	1.001	1.011
as left zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.3	-0.2	----	----
as left span	4920	80.5	800.2	448.1	352.1	799.7	446.1	353.6	1.001	1.004
Average Correction Factor									0.999	1.005

Corrected As found	NO _x = 790.5 ppb	NO = 789.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.8%
Previous Response	NO _x = 797.1 ppb	NO = 797.5 ppb		*Percent Change	NO = -1.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.8	445.7	352.1	352.8	0.998	100.2%
2nd GPT point (200 ppb O3)	797.8	623.2	174.6	175.2	0.997	100.3%
3rd GPT point (100 ppb O3)	797.8	708.5	89.3	89.2	1.001	99.9%
Average Correction Factor					0.999	100.1%

Notes: No maintenance done. Span adjusted. Due to drifting during the GPT the 2nd NO ref point used.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 11, 2023	Previous Calibration:	September 5, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:00	End Time (MST):	11:19
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.4	----	Correlation Coefficient	≥0.995	
800.2	799.9	1.0003			
399.6	400.9	0.9967			
199.8	199.6	1.0010			
			Slope	1.000322	0.90 - 1.10
			Intercept	0.004208	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 11, 2023	Previous Calibration:	September 5, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:00	End Time (MST):	11:19
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
800.2	799.1	1.0013			
399.6	399.0	1.0014			
199.8	197.6	1.0111			
			Slope	0.999781	0.90 - 1.10
			Intercept	-0.956508	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 11, 2023	Previous Calibration:	September 5, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:00	End Time (MST):	11:19
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999998	≥0.995
352.1	352.8	0.9980			
174.6	175.2	0.9966			
89.3	89.2	1.0011			
			Slope	1.002559	0.90 - 1.10
			Intercept	-0.119059	+/-20



NO_x Calibration Plot

Date: October 11, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Hills Station number: AMS 23
 Calibration Date: October 17, 2023 Last Cal Date: September 7, 2023
 Start time (MST): 7:14 End time (MST): 8:31

Analyzer Make: API T640 S/N: 1546
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	6.3	6	6.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	730.1	726.6	730.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.08	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>October 17, 2023</u>		Last Cal Date: <u>September 7, 2023</u>		
	PM w/o HEPA: <u>18</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.9	10.9	10.9	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: <u>400</u>		w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:	<u>October 17, 2023</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>October 17, 2023</u>				

Annual Maintenance

Date Sample Tube Cleaned: October 17, 2023
 Date RH/T Sensor Cleaned: October 17, 2023

Leak check passed before and after cleaning. No adjustments done. Head cleaned.

Notes:

Calibration by: Melissa Lemay



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Fort Hills	Station Number:	AMS 23
Calibration Date:	October 2, 2023	Prev Cal Date:	July 20, 2022
Start Time (MST):	7:55	End Time (MST):	8:30
Tower Height (m):	10m	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	P19837
WS Calibrator:	MetOne 053	Serial Number:	R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.5	-0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥0.9995
Calculated slope		0.999473	0.90 - 1.10
Calculated intercept		0.026227	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	B14267
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	12:13:40	Calc Declination*:	13.76 Degrees
Deadband calc:	357.0 degrees (<i>Limit 4 deg</i>)	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0		---
90		
180		
270		
357		

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope			0.90 - 1.10
Calculated intercept			+/- 4

Notes: Installed new WS as old WS failed torque test.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Fort Hills	Station Number:	AMS 23
Calibration Date:	October 2, 2023	Prev Cal Date:	July 20, 2022
Start Time (MST):	7:55	End Time (MST):	8:30
Tower Height (m):	10m	Reason:	Removal

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	B17268
WS Calibrator:	MetOne 053	Serial Number:	R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.5	-0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999955	0.999999	≥0.9995
Calculated slope	1.030888	0.999473	0.90 - 1.10
Calculated intercept	-0.371642	0.026227	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	B14267
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	12:13:40	Calc Declination*:	13.76 Degrees
Deadband calc:	0.5 degrees (<i>Limit 4 deg</i>)	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.1	---
90	91.3	0.4%
180	179.9	0.0%
270	269.8	-0.1%
357	356.6	-0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999993	0.999993	≥0.9995
Calculated slope	0.990386	1.002788	0.90 - 1.10
Calculated intercept	0.352148	-0.640493	+/- 4

Notes: Checked cross arm with solarnoon before and after tower take down. WS removed due to torque test fail.

Calibration Performed By: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25 WASKŌW OHCI PIMÂTISIWIN OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number:	AMS25
Calibration Date:	October 19, 2023	Last Cal Date:	September 28, 2023
Start time (MST):	7:25	End time (MST):	10:43
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.54	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC437219			
Removed Cal Gas Conc:	50.54	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	747
ZAG Make/Model:	API T701		Serial Number:	4765

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.990940	0.998007	Backgd or Offset:	9.7	9.9
Calibration intercept:	0.604127	0.803956	Coeff or Slope:	0.988	1.001

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4921	79.2	800.5	787.6	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4921	79.2	800.5	799.1	1.002
second point	4960	39.6	400.3	401.8	0.996
third point	4980	19.8	200.1	200.1	1.000
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	79.2	800.5	805.5	0.994
Average Correction Factor					0.999

Baseline Corr As found:	787.50	Previous response	793.87	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 28, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	7:25	End Time (MST):	10:43
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999991	≥0.995
800.5	799.1	1.0018			
400.3	401.8	0.9963	Slope	0.998007	0.90 - 1.10
200.1	200.1	1.0002			
			Intercept	0.803956	+/-30



SO2 Calibration Plot

Date: October 19, 2023

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS25
 Calibration Date: October 20, 2023 Last Cal Date: September 26, 2023
 Start time (MST): 6:20 End time (MST): 10:52
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC517099
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 747
 ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
 Converter make: Global G-150 Converter serial #: 2022-219
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.980956	1.004258	Backgd or Offset: 3.25	3.30
Calibration intercept:	0.100000	0.140000	Coeff or Slope: 1.079	1.108

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4920	80.0	79.5	78.0	1.017
as found 2nd point	4960	40.0	39.7	39.0	1.016
as found 3rd point	4980	20.0	19.9	19.5	1.013
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.0	79.5	79.7	0.997
second point	4960	40.0	39.7	40.6	0.979
third point	4980	20.0	19.9	19.9	0.998
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	80.0	800.0	803.0	0.996
SO2 Scrubber Check	4921	79.2	800.0	0.0	----
Date of last scrubber change:	20-Jun-23			Ave Corr Factor	0.991
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 78.1 Prev response: 78.04 *% change: 0.1%
 Baseline Corr 2nd AF pt: 39.1 AF Slope: 0.982682 AF Intercept: -0.060000
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999999

* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 26, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	6:20	End Time (MST):	10:52
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999877	
79.5	79.7	0.9969			≥0.995
39.7	40.6	0.9785	Slope	1.004258	
19.9	19.9	0.9982			0.90 - 1.10
			Intercept	0.140000	+/-3



H₂S Calibration Plot

Date: October 20, 2023

Location: Waskow ohci Pimatisiwin





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS26 CHRISTINA LAKE OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 26, 2023	Previous Calibration:	September 19, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	12:14	End Time (MST):	16:06
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.0	----	Correlation Coefficient	0.999952	≥0.995
800.0	808.7	0.9892			
400.6	400.9	0.9992	Slope	1.010565	0.90 - 1.10
201.6	200.9	1.0036			
			Intercept	-1.377444	+/-30



SO2 Calibration Plot

Date: October 26, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Christina Lake	Station number:	AMS26
Calibration Date:	October 27, 2023	Last Cal Date:	September 20, 2023
Start time (MST):	8:08	End time (MST):	12:41
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.89	ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	<u>EY0002466</u>			
Removed Cal Gas Conc:	4.89	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	281
ZAG Make/Model:	API T701H		Serial Number:	832

Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032
Converter make:	NA	Converter serial #:	NA
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010898	0.992122	Backgd or Offset:	35.0
Calibration intercept:	0.179125	0.528176	Coeff or Slope:	1.093
				33.6
				1.060

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.6	----
as found span	4930	81.9	79.9	83.5	0.964
as found 2nd point	4972	41.0	40.0	41.8	0.971
as found 3rd point	4994	20.6	20.1	21.4	0.966
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.7	----
high point	4930	81.9	79.9	79.9	1.000
second point	4972	41.0	40.0	40.1	0.997
third point	4994	20.6	20.1	20.3	0.990
as left zero	5000	0.0	0.0	0.7	----
as left span	4926	81.9	80.0	81.4	0.982
SO2 Scrubber Check	4931	80.9	807.1	-0.1	----
Date of last scrubber change:	27-Feb-19			Ave Corr Factor	0.996
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	82.9	Prev response:	80.96	*% change:	2.3%
Baseline Corr 2nd AF pt:	41.2	AF Slope:	1.037198	AF Intercept:	0.525632
Baseline Corr 3rd AF pt:	20.8	AF Correlation:	0.999984		

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after MAF's. Ran scrubber check after calibrator zero. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 27, 2023	Previous Calibration:	September 20, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	8:08	End Time (MST):	12:41
Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient	0.999978	
79.9	79.9	1.0001			≥0.995
40.0	40.1	0.9974	Slope	0.992122	
20.1	20.3	0.9896			0.90 - 1.10
			Intercept	0.528176	+/-3



H₂S Calibration Plot

Date: October 27, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Christina Lake	Station number:	AMS26
Calibration Date:	October 26, 2023	Last Cal Date:	September 19, 2023
Start time (MST):	10:59	End time (MST):	16:55
Reason:	Routine		

Calibration Standards

NO Gas Cylinder #:	T2Y1P4C	Cal Gas Expiry Date:	November 12, 2023		
NOX Cal Gas Conc:	50.82	ppm	NO Cal Gas Conc:	50.02	ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA		
Removed Gas NOX Conc:	50.82	ppm	Removed Gas NO Conc:	50.02	ppm
NOX gas Diff:		NO gas Diff:			
Calibrator Model:	API T700	Serial Number:	3253		
ZAG make/model:	API T701H	Serial Number:	832		

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1173480006

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.370	1.300	NO bkgnd or offset:	2.6	2.5
NOX coeff or slope:	0.991	0.997	NOX bkgnd or offset:	3.4	3.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	161.0	162.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000604	1.001847
NO _x Cal Offset:	-2.180000	-0.048939
NO Cal Slope:	1.004213	1.001247
NO Cal Offset:	-2.740000	-0.399872
NO ₂ Cal Slope:	0.995671	1.005662
NO ₂ Cal Offset:	-0.892478	-0.810879



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.9	-0.1	-0.7	----	----
as found span	4928	80.1	812.8	800.0	12.8	855.4	841.9	13.5	0.9502	0.9503
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.6	0.0	-0.7	----	----
high point	4928	80.1	812.8	800.0	12.8	813.8	800.7	13.1	0.9988	0.9992
second point	4969	40.1	406.8	400.4	6.4	408.4	400.7	7.7	0.9962	0.9993
third point	4982	20.1	204.2	201.0	3.2	204.7	200.2	4.5	0.9976	1.0040
as left zero	5000	0.0	0.0	0.0	0.0	-0.6	0.0	-0.6	----	----
as left span	4822	80.1	830.4	422.0	408.4	814.7	420.3	394.3	1.0193	1.0040
Average Correction Factor									0.9975	1.0008

Corrected As found	NO _x = 856.3 ppb	NO = 842.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 5.3%
Previous Response	NO _x = 811.1 ppb	NO = 800.7 ppb		*Percent Change	NO = 4.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.7	402.1	408.4	409.8	0.9966	100.3%
2nd GPT point (200 ppb O3)	797.7	597.0	213.5	214.3	0.9962	100.4%
3rd GPT point (100 ppb O3)	797.7	698.1	112.4	111.8	1.0053	99.5%
Average Correction Factor					0.9994	100.1%

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 26, 2023	Previous Calibration:	September 19, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:59	End Time (MST):	16:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
812.8	813.8	0.9988		
406.8	408.4	0.9962		
204.2	204.7	0.9976		
			0.999996	
			1.001847	
			-0.048939	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 26, 2023	Previous Calibration:	September 19, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:59	End Time (MST):	16:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 $0.90 - 1.10$ ± 20
800.0	800.7	0.9992		
400.4	400.7	0.9993		
201.0	200.2	1.0040		
			0.999998	
			1.001247	
			-0.399872	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 26, 2023	Previous Calibration:	September 19, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:59	End Time (MST):	16:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.7	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
408.4	409.8	0.9966		
213.5	214.3	0.9962		
112.4	111.8	1.0053		
			0.999996	
			1.005662	
			-0.810879	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 26, 2023

Location: Christina Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27
JACKFISH 2/3
OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	October 18, 2023	Last Cal Date:	September 28, 2023
Start time (MST):	9:31	End time (MST):	12:20
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	SG9133974BAL			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3811
ZAG Make/Model:	API 701		Serial Number:	135

Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	12124313138
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.990232	0.999819	Backgd or Offset:	7.2	7.3
Calibration intercept:	-2.057655	-2.538114	Coeff or Slope:	0.891	0.900

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	79.1	800.2	788.8	1.014
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4921	79.1	800.2	798.8	1.002
second point	4961	39.5	399.5	395.9	1.009
third point	4980	19.8	200.3	194.5	1.030
as left zero	5000	0.0	0.0	0.3	----
as left span	4921	79.1	800.2	800.8	0.999
Average Correction Factor					1.014

Baseline Corr As found:	788.80	Previous response	790.29	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 28, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	9:31	End Time (MST):	12:20
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999937	≥0.995
800.2	798.8	1.0017			
399.5	395.9	1.0092	Slope	0.999819	0.90 - 1.10
200.3	194.5	1.0298			
			Intercept	-2.538114	+/-30



SO2 Calibration Plot

Date: October 18, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Jackfish 2/3 Station number: AMS27
 Calibration Date: October 17, 2023 Last Cal Date: September 27, 2023
 Start time (MST): 8:47 End time (MST): 12:34
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.41 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345023
 Removed Cal Gas Conc: 5.41 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3811
 ZAG Make/Model: API 701 Serial Number: 135

Analyzer Information

Analyzer make: API T101 Analyzer serial #: 621
 Converter make: Converter serial #:
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002337	1.010889	Backgd or Offset: 29.9	29.9
Calibration intercept:	-0.137809	-0.037634	Coeff or Slope: 0.928	0.928

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4926	74.1	80.2	82.0	0.980
as found 2nd point	4963	37.0	40.0	40.9	0.984
as found 3rd point	4982	18.5	20.0	19.9	1.016
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	74.1	80.2	81.0	0.990
second point	4963	37.0	40.0	40.6	0.986
third point	4982	18.5	20.0	19.9	1.006
as left zero	5000	0.0	0.0	0.2	----
as left span	4926	74.1	80.2	79.9	1.003
SO2 Scrubber Check	4921	79.1	791.0	0.2	----
Date of last scrubber change:				Ave Corr Factor	0.994
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 81.8 Prev response: 80.2 *% change: 1.9%
 Baseline Corr 2nd AF pt: 40.7 AF Slope: 1.023151 AF Intercept: -0.117470
 Baseline Corr 3rd AF pt: 19.7 AF Correlation: 0.999912

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments required.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 27, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	8:47	End Time (MST):	12:34
Analyzer make:	API T101	Analyzer serial #:	621

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	0.999963
80.2	81.0	0.9898		
40.0	40.6	0.9861	Slope	1.010889
20.0	19.9	1.0058		
			Intercept	-0.037634
				+/-3



H₂S Calibration Plot

Date: October 17, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as found span	4921	79.4	816.8	800.3	16.5	809.8	789.9	20.0	1.0086	1.0132
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.2	----	----
high point	4921	79.4	816.8	800.3	16.5	812.6	798.3	14.2	1.0052	1.0025
second point	4960	39.7	408.5	400.2	8.3	401.1	394.7	6.4	1.0183	1.0139
third point	4980	19.8	203.7	199.6	4.1	200.1	194.4	5.8	1.0180	1.0267
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	1.2	-1.2	----	----
as left span	4921	79.4	816.8	420.5	400.0	812.4	416.0	396.3	1.0054	1.0108
Average Correction Factor									1.0139	1.0144

Corrected As found	NO _x = 809.9 ppb	NO = 790.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.4%
Previous Response	NO _x = 813.2 ppb	NO = 800.3 ppb		*Percent Change	NO = -1.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.1	411.6	400.0	399.4	1.0015	99.8%
2nd GPT point (200 ppb O3)	795.1	620.3	191.3	190.2	1.0059	99.4%
3rd GPT point (100 ppb O3)	795.1	710.1	101.5	100.3	1.0121	98.8%
Average Correction Factor					1.0065	99.4%

Notes: Changed the sample inlet filter after as founds. No adjustments required.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 28, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:35	End Time (MST):	14:42
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
816.8	812.6	1.0052		
408.5	401.1	1.0183		
203.7	200.1	1.0180		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 28, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:35	End Time (MST):	14:42
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
800.3	798.3	1.0025		
400.2	394.7	1.0139		
199.6	194.4	1.0267		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	September 28, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:35	End Time (MST):	14:42
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
400.0	399.4	1.0015		
191.3	190.2	1.0059		
101.5	100.3	1.0121		
			0.999993	
			0.999566	
			-0.710230	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 19, 2023

Location: Jackfish 2/3





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS29
SURMONT 2
OCTOBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 23, 2023	Previous Calibration:	September 5, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:59	End Time (MST):	13:58
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.5	----	Correlation Coefficient	0.999987	≥0.995
800.1	800.0	1.0001			
400.6	399.3	1.0032	Slope	1.001516	0.90 - 1.10
199.8	196.7	1.0159			
			Intercept	-1.785913	+/-30



SO2 Calibration Plot

Date: October 23, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	October 24, 2023	Last Cal Date:	September 19, 2023
Start time (MST):	9:38	End time (MST):	14:34
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	<u>5.391</u>	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	<u>CC508338</u>			
Removed Cal Gas Conc:	<u>5.391</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>CC508338</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	5472
ZAG Make/Model:	Teledyne API T701		Serial Number:	4297

Analyzer Information

Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170
Converter make:	Global	Converter serial #:	2022-220
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005327	1.003047	Backgd or Offset:	1.19
Calibration intercept:	-0.202877	-0.022987	Coeff or Slope:	1.080
				0.92
				1.080

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	4926	74.2	80.0	81.1	0.982
as found 2nd point	4963	37.2	40.1	40.3	0.985
as found 3rd point	4982	18.6	20.1	19.9	0.988
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4926	74.2	80.0	80.2	0.998
second point	4963	37.2	40.1	40.3	0.995
third point	4982	18.6	20.1	20.0	1.003
as left zero	5000	0.0	0.0	0.0	----
as left span	4926	74.2	80.0	79.2	1.010
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.998
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	81.5	Prev response:	80.23	*% change:	1.6%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.019043	AF Intercept:	-0.483499
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999994		

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted zero. Scrubber check done after cal zero, passed.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 24, 2023	Previous Calibration:	September 19, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:38	End Time (MST):	14:34
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999995	
80.0	80.2	0.9975			≥0.995
40.1	40.3	0.9952	Slope	1.003047	
20.1	20.0	1.0027			0.90 - 1.10
			Intercept	-0.022987	+/-3



H₂S Calibration Plot

Date: October 24, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	October 23, 2023	Last Cal Date:	September 5, 2023
Start time (MST):	10:59	End time (MST):	13:58
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC356008	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
C3H8 Cal Gas Conc.	<u>205.7</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
Removed C3H8 Conc.	<u>205.7</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5472
ZAG Make/Model:	Teledyne API T701	Serial Number:	4297

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1170050149
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006607	1.007852	Background:	3.64	3.64
Calibration intercept:	-0.029039	-0.102215	Coefficient:	3.987	3.987

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.09	----
as found span	4918	81.3	17.31	17.33	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	-0.08	----
high point	4918	81.3	17.31	17.36	0.997
second point	4959	40.6	8.65	8.60	1.005
third point	4979	20.3	4.32	4.23	1.022
as left zero	5000	0.0	0.00	-0.10	----
as left span	4918	81.3	17.31	17.38	0.996
Average Correction Factor					1.008
Baseline Corr As found:	17.42	Previous response	17.40	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 23, 2023	Previous Calibration:	September 5, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:59	End Time (MST):	13:58
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.08	----	Correlation Coefficient	0.999990	≥0.995
17.31	17.36	0.9974			
8.65	8.60	1.0053	Slope	1.007852	0.90 - 1.10
4.32	4.23	1.0223			
			Intercept	-0.102215	+/-1.5



THC Calibration Plot

Date: October 23, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	October 25, 2023	Last Cal Date:	September 6, 2023
Start time (MST):	9:38	End time (MST):	14:35
Reason:	Routine		

Calibration Standards

NO Gas Cylinder #:	T12YYFE	Cal Gas Expiry Date:	October 30, 2024
NOX Cal Gas Conc:	47.46 ppm	NO Cal Gas Conc:	47.46 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	47.46 ppm	Removed Gas NO Conc:	47.46 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	Teledyne API T700	Serial Number:	5472
ZAG make/model:	Teledyne API T701	Serial Number:	4297

Analyzer Information

Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.386	1.386	NO bkgnd or offset:	1.4	1.4
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	1.4	1.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	178.0	175.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997951	1.006760
NO _x Cal Offset:	0.647509	-0.933669
NO Cal Slope:	0.996977	1.006372
NO Cal Offset:	-0.611801	-1.673031
NO ₂ Cal Slope:	0.993705	0.995839
NO ₂ Cal Offset:	1.353170	0.777688



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
as found span	4916	84.2	799.2	799.2	0.0	803.0	801.0	1.6	0.9953	0.9977
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4916	84.2	799.2	799.2	0.0	804.0	803.0	1.0	0.9940	0.9953
second point	4958	42.1	399.6	399.6	0.0	401.4	401.0	0.4	0.9955	0.9965
third point	4979	21.1	200.3	200.3	0.0	199.3	197.3	2.0	1.0049	1.0151
as left zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
as left span	4916	84.2	799.2	410.3	388.9	800.0	410.7	389.3	0.9990	0.9990
Average Correction Factor									0.9982	1.0023

Corrected As found	NO _x = 803.2 ppb	NO = 801.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.6%
Previous Response	NO _x = 798.2 ppb	NO = 796.2 ppb		*Percent Change	NO = 0.6%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.6	407.7	388.9	387.9	1.0026	99.7%
2nd GPT point (200 ppb O3)	796.6	604.9	191.7	191.5	1.0010	99.9%
3rd GPT point (100 ppb O3)	796.6	701.7	94.9	96.3	0.9855	101.5%
Average Correction Factor					0.9964	100.4%

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 25, 2023	Previous Calibration:	September 6, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:38	End Time (MST):	14:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.2	804.0	0.9940		
399.6	401.4	0.9955		
200.3	199.3	1.0049		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 25, 2023	Previous Calibration:	September 6, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:38	End Time (MST):	14:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.2	803.0	0.9953		
399.6	401.0	0.9965		
200.3	197.3	1.0151		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 25, 2023	Previous Calibration:	September 6, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:38	End Time (MST):	14:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
388.9	387.9	1.0026			
191.7	191.5	1.0010			
94.9	96.3	0.9855			
			Slope	0.995839	0.90 - 1.10
			Intercept	0.777688	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 25, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name:	Surmont 2	Station number:	AMS 29
Calibration Date:	October 25, 2023	Last Cal Date:	September 7, 2023
Start time (MST):	9:49	End time (MST):	11:50
Analyzer Make:	API T640	S/N:	253
Particulate Fraction:	PM2.5		
Flow Meter Make/Model:	Alicat FP-25BT	S/N:	388750
Temp/RH standard:	Alicat FP-25BT	S/N:	388750

Monthly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Measured</u>	<u>As left</u>	<u>Adjusted</u>	<i>(Limits)</i>
T (°C)	-5.8	-6.00	-5.8	<input type="checkbox"/>	<i>+/- 2 °C</i>
P (mmHg)	715.7	716.80	715.7	<input type="checkbox"/>	<i>+/- 10 mmHg</i>
flow (LPM)	4.93	5.092	4.93	<input type="checkbox"/>	<i>+/- 0.25 LPM</i>
Leak Test:	Date of check: <u>October 25, 2023</u>		Last Cal Date: <u>September 7, 2023</u>		
	PM w/o HEPA: <u>4.2</u>		PM w/ HEPA: <u>0</u>		<i><0.2 ug/m3</i>

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Post maintenance</u>	<u>As left</u>	<u>Adjusted</u>	<i>(Limits)</i>
PMT Peak Test	10.9	11.2	11.2	<input type="checkbox"/>	<i>11.3 +/- 0.5</i>
Post-maintenance leak check:	PM w/o HEPA: <u>2.9</u>		w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:	<u>October 25, 2023</u>				<i><0.2 ug/m3</i>
Disposable Filter Changed:	<u>October 25, 2023</u>				

Annual Maintenance

Date Sample Tube Cleaned: October 25, 2023
 Date RH/T Sensor Cleaned: October 25, 2023

Notes: No adjustments made. Completed annual and quarterly maintenance.

Calibration by: Braiden Boutilier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS30
ELLS RIVER
OCTOBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 1, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:26	End Time (MST):	12:25
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999947	≥0.995
800.4	799.3	1.0013			
400.2	395.8	1.0112	Slope	1.000474	0.90 - 1.10
200.1	194.8	1.0272			
			Intercept	-2.915894	+/-30



SO2 Calibration Plot

Date: October 3, 2023

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Ells River Station number: AMS30
 Calibration Date: October 26, 2023 Last Cal Date: September 7, 2023
 Start time (MST): 8:05 End time (MST): 11:54
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.08 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002443
 Removed Cal Gas Conc: 5.08 ppm Rem Gas Exp Date:
 Removed Gas Cyl #: Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3061
 ZAG Make/Model: API T701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1410661331
 Converter make: CDN - 101 Converter serial #: 562
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001492	0.998492	Backgd or Offset:	1.60 1.55
Calibration intercept:	-0.079138	-0.099119	Coeff or Slope:	1.136 1.100

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	78.7	80.0	82.3	0.972
as found 2nd point	4961	39.4	40.0	40.9	0.979
as found 3rd point	4980	19.7	20.0	20.1	0.996
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	78.7	80.0	79.8	1.002
second point	4961	39.4	40.0	39.8	1.006
third point	4980	19.7	20.0	19.8	1.011
as left zero	5000	0.0	0.0	0.1	----
as left span	4921	78.7	80.0	80.2	0.997
SO2 Scrubber Check	4921	79.2	800.4	0.1	----
Date of last scrubber change:	N/A			Ave Corr Factor	1.006
Date of last converter efficiency test:	N/A			efficiency	

Baseline Corr As found: 82.3 Prev response: 80.00 *% change: 2.8%
 Baseline Corr 2nd AF pt: 40.9 AF Slope: 1.030935 AF Intercept: -0.259676
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999954

* = > +/-5% change initiates investigation

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 26, 2023	Previous Calibration:	September 7, 2023
Station Name:	Ells River	Station Number:	AMS30
Start Time (MST):	8:05	End Time (MST):	11:54
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999993	≥0.995
80.0	79.8	1.0021			
40.0	39.8	1.0057	Slope	0.998492	0.90 - 1.10
20.0	19.8	1.0109			
			Intercept	-0.099119	+/-3



TRS Calibration Plot

Date: October 26, 2023

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	October 3, 2023	Last Cal Date:	September 1, 2023
Start time (MST):	9:26	End time (MST):	12:24
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm		
Diff between cyl (CH ₄):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701H	Serial Number:	3061
		Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.44E-04	2.45E-04	NMHC SP Ratio:	4.53E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	201064
				191712

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.59	1.027
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.01	1.001
second point	4960	39.6	8.51	8.42	1.012
third point	4980	19.8	4.26	4.14	1.028
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	16.98	1.003

				Average Correction Factor	1.014
Baseline Corr AF:	16.59	Prev response	16.98	*% change	-2.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	8.73	1.044
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.09	1.002
second point	4960	39.6	4.56	4.49	1.014
third point	4980	19.8	2.28	2.20	1.035
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.10	1.001
Average Correction Factor					1.017
Baseline Corr AF:	8.73	Prev response	9.07	*% change	-3.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.86	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.91	1.000
second point	4960	39.6	3.96	3.92	1.009
third point	4980	19.8	1.98	1.94	1.021
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.88	1.005
Average Correction Factor					1.010
Baseline Corr AF:	7.86	Prev response	7.91	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000637	1.000476
THC Cal Offset:	-0.057538	-0.062338
CH ₄ Cal Slope:	1.002936	1.000886
CH ₄ Cal Offset:	-0.025358	-0.022757
NMHC Cal Slope:	0.998741	0.999995
NMHC Cal Offset:	-0.032581	-0.039581

Notes: Inlet filter changed after As Finds, span adjusted.

Calibration Performed By: Ryan Power



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 1, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:26	End Time (MST):	12:24
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999938	≥ 0.995			
17.03	17.01	1.0012						
8.51	8.42	1.0116				Slope	1.000476	0.90 - 1.10
4.26	4.14	1.0283						
			Intercept	-0.062338	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 1, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:26	End Time (MST):	12:24
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999961	≥0.995
7.91	7.91	1.0004			
3.96	3.92	1.0089			
1.98	1.94	1.0211			
			Slope	1.000886	0.90 - 1.10
			Intercept	-0.022757	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 3, 2023	Previous Calibration:	September 1, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:26	End Time (MST):	12:24
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999913	≥ 0.995			
9.11	9.09	1.0020						
4.56	4.49	1.0142				Slope	0.999995	0.90 - 1.10
2.28	2.20	1.0346						
			Intercept	-0.039581	± 0.5			



NMHC Calibration Plot

Date: October 3, 2023

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Ells River
Calibration Date: October 20, 2023
Start time (MST): 9:12
Reason: Routine
Station number: AMS 30
Last Cal Date: September 6, 2023
End time (MST): 13:41

Calibration Standards

NO Gas Cylinder #: T2Y1P2R
NOX Cal Gas Conc: 50.83 ppm
Removed Cylinder #: T2Y1P2R
Removed Gas NOX Conc: 50.83 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 49.97 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 49.97 ppm
NO gas Diff:
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 710321429

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.129	1.147	NO bkgnd or offset:	13.8	14.0
NOX coeff or slope:	0.986	0.992	NOX bkgnd or offset:	13.8	14.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	189.7	190.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996459	1.001461
NO _x Cal Offset:	-1.100000	-1.680000
NO Cal Slope:	1.000500	1.000029
NO Cal Offset:	-2.240000	-2.600000
NO ₂ Cal Slope:	0.993934	1.004460
NO ₂ Cal Offset:	-0.960703	0.734060



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.7	0.0	----	----
as found span	4920	80.0	813.3	799.5	13.8	799.7	786.2	13.6	1.0170	1.0169
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.4	0.0	----	----
high point	4920	80.0	813.3	799.5	13.8	813.4	798.0	15.4	0.9999	1.0019
second point	4960	40.0	406.6	399.8	6.9	405.0	396.2	8.8	1.0040	1.0090
third point	4980	20.0	203.3	199.9	3.4	200.6	195.0	5.6	1.0136	1.0250
as left zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.4	0.0	----	----
as left span	4920	80.0	813.3	425.2	388.1	815.8	422.8	393.0	0.9969	1.0057
Average Correction Factor									1.0058	1.0120

Corrected As found	NO _x = 800.4 ppb	NO = 786.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.1%	
Previous Response	NO _x = 809.3 ppb	NO = 797.7 ppb		*Percent Change	NO = -1.4%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	795.2	420.9	388.1	390.2	0.9945	100.6%
2nd GPT point (200 ppb O ₃)	795.2	613.6	195.4	197.1	0.9912	100.9%
3rd GPT point (100 ppb O ₃)	795.2	701.0	108.0	110.1	0.9806	102.0%
Average Correction Factor					0.9888	101.1%

Notes: Adjusted the span only. Noticed the background shifted a bit from the previous calibration.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 6, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:12	End Time (MST):	13:41
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.3	813.4	0.9999		
406.6	405.0	1.0040		
203.3	200.6	1.0136		
			0.999989	
			1.001461	
			-1.680000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 6, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:12	End Time (MST):	13:41
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
799.5	798.0	1.0019		
399.8	396.2	1.0090		
199.9	195.0	1.0250		
			0.999965	
			1.000029	
			-2.600000	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 20, 2023	Previous Calibration:	September 6, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:12	End Time (MST):	13:41
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
388.1	390.2	0.9945		
195.4	197.1	0.9912		
108.0	110.1	0.9806		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 20, 2023

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Ells River Station number: AMS 30
 Calibration Date: October 27, 2023 Last Cal Date: September 8, 2023
 Start time (MST): 8:08 End time (MST): 9:12

Analyzer Make: API T640 S/N: 875
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat S/N: 388753
 Temp/RH standard: Alicat S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-7.6	-7.8	-7.6	<input type="checkbox"/>	
P (mmHg)	739.4	741	739.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	4.96	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>October 27, 2023</u>		Last Cal Date: <u>September 8, 2023</u>		
	PM w/o HEPA: <u>3.9</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.7	10.8	10.8	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: <u>8.8</u>		w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:	<u>October 27, 2023</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>October 27, 2023</u>				

Annual Maintenance

Date Sample Tube Cleaned: October 27, 2023
 Date RH/T Sensor Cleaned: _____

Notes: No adjustments done. Head cleaned. Flow, PMT and Leak checked before and after cleaning.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS501
LEISMER
OCTOBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	NA
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	12:45	End Time (MST):	14:58
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	≥0.995
800.2	801.0	0.9990		
400.2	401.4	0.9969	Slope	0.90 - 1.10
200.1	197.2	1.0145		
			Intercept	+/-30



SO2 Calibration Plot

Date: September 12, 2023

Location: Leismer





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 18, 2023	Previous Calibration:	September 12, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:01	End Time (MST):	11:58
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.3	----	Correlation Coefficient	≥0.995
800.2	802.1	0.9976		
400.2	402.0	0.9954		
200.1	197.0	1.0156		
			Slope	0.90 - 1.10
			Intercept	+/-30



SO2 Calibration Plot

Date: October 18, 2023

Location: Leismer





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Leismer	Station number:	AMS501
Calibration Date:	September 12, 2023	Last Cal Date:	NA
Start time (MST):	9:51	End time (MST):	12:48
Reason:	Install		

Calibration Standards

Cal Gas Concentration:	5.14	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511843			
Removed Cal Gas Conc:	5.14	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T750		Serial Number:	282
ZAG Make/Model:	API 751H		Serial Number:	321

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020
Converter make:	Global G150	Converter serial #:	2022-218
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.011146	Backgd or Offset:	NA	3.35
Calibration intercept:	NA	-0.518157	Coeff or Slope:	NA	1.066

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.5	----
high point	4922	77.8	80.0	80.3	0.996
second point	4961	38.9	40.0	40.1	0.997
third point	4981	19.4	19.9	19.5	1.023
as left zero	5000	0.0	0.0	-0.3	----
as left span	4922	77.8	80.0	80.9	0.989
SO2 Scrubber Check	4921	79.2	792.0	0.0	----
Date of last scrubber change:	24-Feb-23			Ave Corr Factor	1.005
Date of last converter efficiency test:	December 1, 2022			efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes:

Install Calibration.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	September 12, 2023	Previous Calibration:	NA
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:51	End Time (MST):	12:48
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.5	----	Correlation Coefficient	0.999984	
80.0	80.3	0.9960			≥0.995
40.0	40.1	0.9973	Slope	1.011146	
19.9	19.5	1.0226			0.90 - 1.10
			Intercept	-0.518157	+/-3



H₂S Calibration Plot

Date: September 12, 2023

Location: Leismer





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Leismer Station number: AMS501
 Calibration Date: October 17, 2023 Last Cal Date: September 12, 2023
 Start time (MST): 9:06 End time (MST): 13:19
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.14 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511843
 Removed Cal Gas Conc: 5.14 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 2659
 ZAG Make/Model: API 701 Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1180540020
 Converter make: Global G150 Converter serial #: 2022-218
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011146	1.006859	Backgd or Offset:	3.41	3.41
Calibration intercept:	-0.518157	-0.318205	Coeff or Slope:	1.066	1.082

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4922	77.8	80.0	78.8	1.011
as found 2nd point	4961	38.9	40.0	39.3	1.010
as found 3rd point	4981	19.4	19.9	19.2	1.023
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4922	77.8	80.0	80.2	0.997
second point	4961	38.9	40.0	40.0	1.000
third point	4981	19.4	19.9	19.7	1.012
as left zero	5000	0.0	0.0	-0.2	----
as left span	4922	77.8	80.0	80.1	0.999
SO2 Scrubber Check	4921	79.2	792.0	0.1	----
Date of last scrubber change:	24-Feb-23			Ave Corr Factor	1.003
Date of last converter efficiency test:	December 1, 2022			efficiency	

Baseline Corr As found: 79.1 Prev response: 80.35 *% change: -1.6%
 Baseline Corr 2nd AF pt: 39.6 AF Slope: 0.990003 AF Intercept: -0.378626
 Baseline Corr 3rd AF pt: 19.5 AF Correlation: 0.999988

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Scrubber test done after calibrator zero. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 17, 2023	Previous Calibration:	September 12, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:06	End Time (MST):	13:19
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999998	≥0.995
80.0	80.2	0.9973			
40.0	40.0	0.9997	Slope	1.006859	0.90 - 1.10
19.9	19.7	1.0123			
			Intercept	-0.318205	+/-3



H₂S Calibration Plot

Date: October 17, 2023

Location: Leismer





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Leismer
Calibration Date: September 14, 2023
Start time (MST): 9:13
Reason: Removal
Station number: AMS501
Last Cal Date: NA
End time (MST): 13:15

Calibration Standards

NO Gas Cylinder #: T26811M
NOX Cal Gas Conc: 47.46 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.46 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API 701
Cal Gas Expiry Date: October 30, 2024
NO Cal Gas Conc: 47.39 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.39 ppm
NO gas Diff:
Serial Number: 2659
Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1218153356

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	NA	1.408	NO bkgnd or offset:	NA	4.6
NOX coeff or slope:	NA	0.991	NOX bkgnd or offset:	NA	4.9
NO2 coeff or slope:	NA	1.000	Reaction cell Press:	NA	169.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	NA	0.999520
NO _x Cal Offset:	NA	-0.528014
NO Cal Slope:	NA	1.000654
NO Cal Offset:	NA	-1.507988
NO ₂ Cal Slope:	NA	0.991533
NO ₂ Cal Offset:	NA	0.027894



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.2	-0.2	----	----
high point	4916	84.4	801.1	799.9	1.2	799.8	799.1	0.8	1.0016	1.0010
second point	4958	42.2	400.5	400.0	0.6	401.0	399.4	1.6	0.9989	1.0014
third point	4979	21.1	200.3	200.0	0.3	198.7	196.4	2.3	1.0079	1.0182
as left zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.4	-0.2	----	----
as left span	4916	84.4	801.1	403.8	397.3	807.3	409.0	398.4	0.9923	0.9872
Average Correction Factor									1.0028	1.0069

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* => +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	804.7	408.6	397.3	393.9	1.0086	99.1%
2nd GPT point (200 ppb O3)	804.7	606.0	199.9	198.1	1.0090	99.1%
3rd GPT point (100 ppb O3)	804.7	699.6	106.3	105.8	1.0046	99.5%
Average Correction Factor					1.0074	99.3%

Notes: Install calibration. Adjusted span. Used 2nd NO reference point due to drift. Accidentally powered off the router that cause the communication drop shown in the graph.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	NA
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:13	End Time (MST):	13:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	≥0.995
801.1	799.8	1.0016		
400.5	401.0	0.9989	Slope	0.90 - 1.10
200.3	198.7	1.0079		
			Intercept	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	NA
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:13	End Time (MST):	13:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.9	799.1	1.0010		
400.0	399.4	1.0014		
200.0	196.4	1.0182		
			0.999980	
			1.000654	
			-1.507988	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	September 14, 2023	Previous Calibration:	NA
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:13	End Time (MST):	13:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
397.3	393.9	1.0086			
199.9	198.1	1.0090			
106.3	105.8	1.0046			
			Slope	0.991533	0.90 - 1.10
			Intercept	0.027894	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: September 14, 2023

Location: Leismer





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Leismer
Calibration Date: October 26, 2023
Start time (MST): 8:43
Reason: Removal
Station number: AMS501
Last Cal Date: September 14, 2023
End time (MST): 14:30

Calibration Standards

NO Gas Cylinder #: T26811M
NOX Cal Gas Conc: 47.46 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.46 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API 701
Cal Gas Expiry Date: October 30, 2024
NO Cal Gas Conc: 47.39 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.39 ppm
NO gas Diff:
Serial Number: 2659
Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1218153356

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.408	1.301	NO bkgnd or offset:	4.6	3.9
NOX coeff or slope:	0.991	0.995	NOX bkgnd or offset:	4.9	4.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	169.5	169.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999520	1.009935
NO _x Cal Offset:	-0.528014	-4.483299
NO Cal Slope:	1.000654	1.007430
NO Cal Offset:	-1.507988	-5.190972
NO ₂ Cal Slope:	0.991533	1.006440
NO ₂ Cal Offset:	0.027894	-0.613191



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.7	-0.1	----	----
as found span	4916	84.4	801.1	799.9	1.2	861.2	857.8	3.5	0.9302	0.9325
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
high point	4870	83.4	799.1	797.9	1.2	805.0	801.4	3.5	0.9926	0.9956
second point	4923	41.8	399.6	399.0	0.6	396.0	393.6	2.4	1.0090	1.0137
third point	4947	20.6	196.8	196.5	0.3	190.5	188.0	2.5	1.0331	1.0453
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as left span	4870	83.4	799.1	400.4	398.7	797.7	403.0	394.7	1.0017	0.9936
Average Correction Factor									1.0116	1.0182

Corrected As found	NO _x = 862.0 ppb	NO = 858.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 7.2%
Previous Response	NO _x = 800.1 ppb	NO = 798.9 ppb		*Percent Change	NO = 6.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.7	400.2	398.7	400.7	0.9950	100.5%
2nd GPT point (200 ppb O3)	797.7	600.2	198.7	199.7	0.9949	100.5%
3rd GPT point (100 ppb O3)	797.7	694.8	104.1	103.2	1.0085	99.2%
Average Correction Factor					0.9995	100.1%

Notes: Changed inlet filter after as founds. Adjusted zero and span. 3rd point was failing, used the Dilution and Source flows from the flow meters (Alicat).

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 26, 2023	Previous Calibration:	September 14, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	8:43	End Time (MST):	14:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
799.1	805.0	0.9926			
399.6	396.0	1.0090			
196.8	190.5	1.0331			
			Slope	1.009935	0.90 - 1.10
			Intercept	-4.483299	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 26, 2023	Previous Calibration:	September 14, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	8:43	End Time (MST):	14:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
797.9	801.4	0.9956		
399.0	393.6	1.0137		
196.5	188.0	1.0453		
			0.999809	
			1.007430	
			-5.190972	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 26, 2023	Previous Calibration:	September 14, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	8:43	End Time (MST):	14:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
398.7	400.7	0.9950			
198.7	199.7	0.9949			
104.1	103.2	1.0085			
			Slope	1.006440	0.90 - 1.10
			Intercept	-0.613191	+/-20



NO_x Calibration Plot

Date: October 26, 2023

Location: Leismer





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Leismer	Station Number:	AMS 501
Calibration Date:	September 11, 2023	Prev Cal Date:	N/A
Start Time (MST):	11:45	End Time (MST):	12:45
Tower Height (m):	9.8	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	Y18362
WS Calibrator:	MetOne 053	Serial Number:	CA 03845

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.2	0.2%
400	39.4	39.4	0.1%
600	58.6	58.7	0.2%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥0.9995
Calculated slope		0.998949	0.90 - 1.10
Calculated intercept		-0.013169	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	R14654
As Found Declination (deg east of True North):		As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:22	Calc Declination*:	13.38 Degrees
Deadband calc:	1.0 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.3	---
90	88.8	-0.3%
180	182.4	0.7%
270	271.5	0.4%
357	356.3	-0.2%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999944	≥0.9995
Calculated slope		0.999058	0.90 - 1.10
Calculated intercept		-0.290512	+/- 4

Notes: Install calibrations. Bearings are in good condition, used compass to line up the crossarm.

Calibration Performed By: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS505
SAWBONES BAY
OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 6, 2023	Previous Calibration:	September 22, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:27	End Time (MST):	11:21
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321323

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.5	----	Correlation Coefficient	0.999996	≥0.995
799.8	800.4	0.9993			
399.9	399.7	1.0005	Slope	1.001967	0.90 - 1.10
200.4	198.7	1.0088			
			Intercept	-1.152684	+/-30



SO2 Calibration Plot

Date: October 6, 2023

Location: Sawbones Bay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Sawbones Bay Station number: AMS505
 Calibration Date: October 5, 2023 Last Cal Date: September 21, 2023
 Start time (MST): 8:25 End time (MST): 12:33
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517397
 Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 5112
 ZAG Make/Model: Teledyne API T701 Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43iQ Analyzer serial #: 1228021057
 Converter make: Global 150 Converter serial #: 2022-224
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003503	1.008501	Backgd or Offset: 2.21	2.21
Calibration intercept:	0.081954	0.082063	Coeff or Slope: 1.021	1.021

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4922	77.7	80.0	81.6	0.981
as found 2nd point	4961	38.8	40.0	40.2	0.994
as found 3rd point	4981	19.4	20.0	20.1	0.994
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4922	77.7	80.0	80.8	0.991
second point	4961	38.8	40.0	40.4	0.989
third point	4981	19.4	20.0	20.2	0.989
as left zero	5000	0.0	0.0	0.3	----
as left span	4922	77.7	80.0	80.3	0.997
SO2 Scrubber Check	4922	77.8	778.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.990
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 81.6 Prev response: 80.40 *% change: 1.5%
 Baseline Corr 2nd AF pt: 40.2 AF Slope: 1.019929 AF Intercept: -0.217881
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999944

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after calibrator zero. Scrubber check and passed after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 21, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:25	End Time (MST):	12:33
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1228021057

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	1.000000	≥0.995
80.0	80.8	0.9905			
40.0	40.4	0.9892	Slope	1.008501	0.90 - 1.10
20.0	20.2	0.9891			
			Intercept	0.082063	+/-3



H₂S Calibration Plot

Date: October 5, 2023

Location: Sawbones Bay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Sawbones Bay	Station number:	AMS505
Calibration Date:	October 4, 2023	Last Cal Date:	September 20, 2023
Start time (MST):	8:35	End time (MST):	12:41
Reason:	Routine		

Calibration Standards

NO Gas Cylinder #:	T1FY3PK	Cal Gas Expiry Date:	January 14, 2024
NOX Cal Gas Conc:	47.94 ppm	NO Cal Gas Conc:	47.94 ppm
Removed Cylinder #:	T1FY3PK	Removed Gas Exp Date:	January 14, 2024
Removed Gas NOX Conc:	47.94 ppm	Removed Gas NO Conc:	47.94 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	5112
ZAG make/model:	API T701H	Serial Number:	690

Analyzer Information

Analyzer make:	API T200	Analyzer serial #:	4260
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.055	1.074	NO bkgnd or offset:	0.4
NOX coeff or slope:	1.052	1.066	NOX bkgnd or offset:	1.3
NO2 coeff or slope:	NA	NA	Reaction cell Press:	7.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001507	1.002452
NO _x Cal Offset:	-1.530484	-0.711135
NO Cal Slope:	1.001793	1.005224
NO Cal Offset:	-1.830393	-1.131231
NO ₂ Cal Slope:	0.999010	0.995052
NO ₂ Cal Offset:	0.464663	-1.022252



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.1	-0.3	1.5	----	----
as found span	4917	83.4	799.6	799.6	0.0	786.2	781.5	4.6	1.0170	1.0231
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
high point	4917	83.4	799.6	799.6	0.0	801.0	802.8	-1.7	0.9982	0.9960
second point	4958	41.7	399.8	399.8	0.0	400.1	401.3	-1.1	0.9994	0.9964
third point	4979	20.9	200.4	200.4	0.0	199.5	198.6	0.8	1.0045	1.0090
as left zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
as left span	4916	83.4	799.7	352.0	447.7	800.8	358.9	441.8	0.9987	0.9809
Average Correction Factor									1.0007	1.0005

Corrected As found	NO _x = 785.1 ppb	NO = 781.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.8%	
Previous Response	NO _x = 799.2 ppb	NO = 799.2 ppb		*Percent Change	NO = -2.2%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	804.3	356.6	447.7	445.0	1.0061	99.4%
2nd GPT point (200 ppb O3)	804.3	549.9	254.4	251.9	1.0099	99.0%
3rd GPT point (100 ppb O3)	804.3	655.3	149.0	146.0	1.0205	98.0%
Average Correction Factor					1.0122	98.8%

Notes: Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 20, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:35	End Time (MST):	12:41
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
799.6	801.0	0.9982			
399.8	400.1	0.9994			
200.4	199.5	1.0045			
			Slope	1.002452	0.90 - 1.10
			Intercept	-0.711135	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 20, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:35	End Time (MST):	12:41
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
799.6	802.8	0.9960			
399.8	401.3	0.9964			
200.4	198.6	1.0090			
			Slope	1.005224	0.90 - 1.10
			Intercept	-1.131231	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 20, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:35	End Time (MST):	12:41
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
447.7	445.0	1.0061			
254.4	251.9	1.0099			
149.0	146.0	1.0205			
			Slope	0.995052	0.90 - 1.10
			Intercept	-1.022252	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 4, 2023

Location: Sawbones Bay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS507
KIRBY SOUTH
OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Kirby South	Station number:	AMS 507
Calibration Date:	October 4, 2023	Last Cal Date:	September 14, 2023
Start time (MST):	14:20	End time (MST):	18:17
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.18	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC303554</u>			
Removed Cal Gas Conc:	49.18	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3804
ZAG Make/Model:	API T701H		Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004608	1.004435	Backgd or Offset:	20.1	20.0
Calibration intercept:	-0.669160	-1.308487	Coeff or Slope:	1.135	1.135

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4919	81.3	799.6	803.0	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4919	81.3	799.6	803.0	0.996
second point	4959	40.7	400.3	398.6	1.004
third point	4980	20.3	199.7	199.1	1.003
as left zero	5000	0.0	0.0	-0.1	----
as left span	4919	81.3	799.6	798.9	1.001
Average Correction Factor					1.001

Baseline Corr As found:	803.10	Previous response	802.63	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 14, 2023
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	14:20	End Time (MST):	18:17
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	≥0.995
799.6	803.0	0.9958		
400.3	398.6	1.0044	Slope	0.90 - 1.10
199.7	199.1	1.0028		
			Intercept	+/-30



SO2 Calibration Plot

Date: October 4, 2023

Location: Kirby South





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Kirby South Station number: AMS507
 Calibration Date: October 4, 2023 Last Cal Date: September 14, 2023
 Start time (MST): 9:44 End time (MST): 14:34
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.167 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517378
 Removed Cal Gas Conc: 5.167 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3804
 ZAG Make/Model: API T701H Serial Number: 880

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1150840012
 Converter make: Global Converter serial #: 2022-197
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992888	1.003178	Backgd or Offset:	1.54	1.75
Calibration intercept:	0.019018	-0.041177	Coeff or Slope:	1.048	1.068

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4923	77.4	80.0	79.1	1.012
as found 2nd point	4961	38.8	40.1	39.4	1.020
as found 3rd point	4981	19.3	19.9	19.6	1.023
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.4	80.0	80.2	0.997
second point	4961	38.8	40.1	40.2	0.997
third point	4981	19.3	19.9	19.9	1.002
as left zero	5000	0.0	0.0	0.1	----
as left span	4923	77.4	80.0	80.4	0.995
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:	25-Jul-23			Ave Corr Factor	0.999
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.0 Prev response: 79.43 *% change: -0.5%
 Baseline Corr 2nd AF pt: 39.3 AF Slope: 0.988174 AF Intercept: -0.040945
 Baseline Corr 3rd AF pt: 19.5 AF Correlation: 0.999980

** = > +/-5% change initiates investigation*

Notes: Changed sample inlet filter after as founds. Scrubber check done after MPAF's, passed. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 14, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	9:44	End Time (MST):	14:34
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999998	
80.0	80.2	0.9972			≥0.995
40.1	40.2	0.9975	Slope	1.003178	
19.9	19.9	1.0022			0.90 - 1.10
			Intercept	-0.041177	+/-3



H₂S Calibration Plot

Date: October 4, 2023

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 14, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	14:20	End Time (MST):	18:17
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	-0.08	----	Correlation Coefficient	0.999992	≥0.995
17.26	17.24	1.0013			
8.64	8.55	1.0109	Slope	1.002940	0.90 - 1.10
4.31	4.24	1.0159			
			Intercept	-0.087189	+/-1.5



THC Calibration Plot

Date: October 4, 2023

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 19, 2023	Previous Calibration:	October 4, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	10:50	End Time (MST):	15:05
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.02	----	Correlation Coefficient	0.999996	
17.26	17.23	1.0019			≥0.995
8.64	8.61	1.0038	Slope	0.997581	
4.31	4.29	1.0057			0.90 - 1.10
			Intercept	-0.000204	+/-1.5



THC Calibration Plot

Date: October 19, 2023

Location: Kirby South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	October 4, 2023	Last Cal Date:	September 15, 2023
Start time (MST):	10:00	End time (MST):	17:50
Reason:	Routine		

Calibration Standards

NO Gas Cylinder #:	T34ULGL	Cal Gas Expiry Date:	March 8, 2025
NOX Cal Gas Conc:	49.39 ppm	NO Cal Gas Conc:	49.02 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	49.39 ppm	Removed Gas NO Conc:	49.02 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T750	Serial Number:	282
ZAG make/model:	API 751H	Serial Number:	321

Analyzer Information

Analyzer make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1182340006

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.485	1.424	NO bkgnd or offset:	1.5	1.5
NOX coeff or slope:	0.994	0.997	NOX bkgnd or offset:	1.6	1.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	167.97	168.83

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.009702	1.001586
NO _x Cal Offset:	-4.611056	-1.471182
NO Cal Slope:	1.011497	1.001191
NO Cal Offset:	-5.412656	-1.933293
NO ₂ Cal Slope:	1.004706	0.995145
NO ₂ Cal Offset:	1.319502	0.363035



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
as found span	4919	81.0	800.1	794.1	6.0	835.0	830.0	4.8	0.9582	0.9568
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
high point	4919	81.0	800.1	794.1	6.0	801.0	794.2	6.4	0.9989	0.9999
second point	4960	40.5	400.0	397.0	3.0	397.2	394.1	3.1	1.0071	1.0074
third point	4980	20.2	199.5	198.0	1.5	198.0	195.0	3.0	1.0077	1.0156
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
as left span	4919	81.0	800.1	389.6	410.5	782.5	363.9	418.7	1.0225	1.0707
Average Correction Factor									1.0046	1.0076

Corrected As found	NO _x = 835.2 ppb	NO = 830.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 3.8%
Previous Response	NO _x = 803.3 ppb	NO = 797.8 ppb		*Percent Change	NO = 3.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.8	390.3	410.5	409.5	1.0024	99.8%
2nd GPT point (200 ppb O3)	794.8	590.2	210.6	207.7	1.0139	98.6%
3rd GPT point (100 ppb O3)	794.8	694.0	106.8	108.6	0.9834	101.7%
Average Correction Factor					0.9999	100.0%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 15, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	10:00	End Time (MST):	17:50
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.1	801.0	0.9989		
400.0	397.2	1.0071		
199.5	198.0	1.0077		
			0.999980	
			1.001586	
			-1.471182	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 15, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	10:00	End Time (MST):	17:50
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
794.1	794.2	0.9999		
397.0	394.1	1.0074		
198.0	195.0	1.0156		
			0.999977	
			1.001191	
			-1.933293	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 4, 2023	Previous Calibration:	September 15, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	10:00	End Time (MST):	17:50
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
410.5	409.5	1.0024			
210.6	207.7	1.0139			
106.8	108.6	0.9834			
			Slope	0.995145	0.90 - 1.10
			Intercept	0.363035	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 4, 2023

Location: Kirby South





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS511
BLACKGOLD
OCTOBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 29, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Blackgold	Station number:	AMS511
Calibration Date:	September 26, 2023	Last Cal Date:	NA
Start time (MST):	14:55	End time (MST):	17:35
Reason:	Install		

Calibration Standards

Cal Gas Concentration:	50.06	ppm	Cal Gas Exp Date:	January 5, 2029
Cal Gas Cylinder #:	CC147416			
Removed Cal Gas Conc:	50.06	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	5258
ZAG Make/Model:	T701		Serial Number:	138

Analyzer Information

Analyzer make: Thermo scientific Analyzer serial #: 1160290014
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.009285	Backgd or Offset:	NA	31.2
Calibration intercept:	NA	-1.252082	Coeff or Slope:	NA	1.097

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	80.0	800.0	806.5	0.992
second point	4968	40.1	400.8	403.7	0.993
third point	4987	20.2	202.0	200.5	1.007
as left zero	5000	0.0	0.0	0.5	----
as left span	4926	80.0	800.0	811.9	0.985
Average Correction Factor					0.997

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Install calibration. Changed sample inlet filter before calibrator zero. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	September 26, 2023	Previous Calibration:	NA
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	14:55	End Time (MST):	17:35
Analyzer make:	Thermo scientific	Analyzer serial #:	1160290014

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999982	≥0.995
800.0	806.5	0.9919			
400.8	403.7	0.9929	Slope	1.009285	0.90 - 1.10
202.0	200.5	1.0072			
			Intercept	-1.252082	+/-30



SO2 Calibration Plot

Date: September 26, 2023

Location: Blackgold





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Blackgold	Station number:	AMS511
Calibration Date:	October 5, 2023	Last Cal Date:	September 26, 2023
Start time (MST):	11:30	End time (MST):	16:20
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.06	ppm	Cal Gas Exp Date:	January 5, 2029
Cal Gas Cylinder #:	CC147416			
Removed Cal Gas Conc:	50.06	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	5258
ZAG Make/Model:	Teledyne API 701		Serial Number:	138

Analyzer Information

Analyzer make: Thermo scientific Analyzer serial #: 1160290014
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009285	1.001162	Backgd or Offset:	31.2	31.8
Calibration intercept:	-1.252082	-1.503208	Coeff or Slope:	1.097	1.107

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	1.0	----
as found span	4926	80.0	800.0	791.0	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.3	----
high point	4926	80.0	800.0	799.9	1.000
second point	4968	40.1	400.8	400.2	1.002
third point	4987	20.2	202.0	198.0	1.020
as left zero	5000	0.0	0.0	0.0	----
as left span	4926	80.0	800.0	798.3	1.002
Average Correction Factor					1.007

Baseline Corr As found:	790.00	Previous response	806.18	*% change	-2.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 26, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	11:30	End Time (MST):	16:20
Analyzer make:	Thermo scientific	Analyzer serial #:	1160290014

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999969	
800.0	799.9	1.0001			≥0.995
400.8	400.2	1.0016	Slope	1.001162	
202.0	198.0	1.0200			0.90 - 1.10
			Intercept	-1.503208	+/-30



SO2 Calibration Plot

Date: October 5, 2023

Location: Blackgold





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Blackgold	Station number:	AMS511
Calibration Date:	September 28, 2023	Last Cal Date:	NA
Start time (MST):	12:05	End time (MST):	14:20
Reason:	Install		

Calibration Standards

Cal Gas Concentration:	5.139	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC511397			
Removed Cal Gas Conc:	5.139	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	5258
ZAG Make/Model:	API T701		Serial Number:	60

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090
Converter make:	Global G150	Converter serial #:	2022-227
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.003061	Backgd or Offset:	NA	3.11
Calibration intercept:	NA	-0.179226	Coeff or Slope:	NA	1.137

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	77.8	80.0	80.1	0.998
second point	4961	38.9	40.0	39.9	1.002
third point	4981	19.5	20.0	19.7	1.017
as left zero	5000	0.0	0.0	0.1	----
as left span	4922	77.9	80.0	80.3	0.997
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.006
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: Installation calibration, no issues. SOX scrubber tested post calibration but before as left zero/span.

Calibration Performed By: Kelly Baragar



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	September 28, 2023	Previous Calibration:	NA
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	12:05	End Time (MST):	14:20
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999975	
80.0	80.1	0.9983			≥0.995
40.0	39.9	1.0021	Slope	1.003061	
20.0	19.7	1.0173			0.90 - 1.10
			Intercept	-0.179226	+/-3



H₂S Calibration Plot

Date: September 28, 2023

Location: Blackgold





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Blackgold Station number: AMS511
 Calibration Date: October 5, 2023 Last Cal Date: September 28, 2023
 Start time (MST): 6:57 End time (MST): 11:50
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.139 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC511397
 Removed Cal Gas Conc: 5.139 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 5258
 ZAG Make/Model: API T701 Serial Number: 138

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1336160090
 Converter make: Global G150 Converter serial #: 2022-227
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003061	0.999059	Backgd or Offset:	3.11
Calibration intercept:	-0.179226	-0.039168	Coeff or Slope:	1.137

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4922	77.8	80.0	80.0	1.000
as found 2nd point	4961	38.9	40.0	40.0	1.000
as found 3rd point	4981	19.5	20.0	19.6	1.022
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4922	77.8	80.0	79.9	1.001
second point	4961	38.9	40.0	39.9	1.002
third point	4981	19.5	20.0	19.8	1.012
as left zero	5000	0.0	0.0	0.3	----
as left span	4922	77.9	80.0	79.8	1.003
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.005
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 80.0 Prev response: 80.03 *% change: 0.0%
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 1.002345 AF Intercept: -0.179169
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999961

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Scrubber check done after cal zero. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 28, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	6:57	End Time (MST):	11:50
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999984	
80.0	79.9	1.0008			≥0.995
40.0	39.9	1.0021	Slope	0.999059	
20.0	19.8	1.0121			0.90 - 1.10
			Intercept	-0.039168	+/-3



H₂S Calibration Plot

Date: October 5, 2023

Location: Blackgold





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Harvest Blackgold	Station number:	AMS511
Calibration Date:	September 30, 2023	Last Cal Date:	NA
Start time (MST):	10:13	End time (MST):	13:40
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC147416	Cal Gas Expiry Date:	January 5, 2029
CH4 Cal Gas Conc.	<u>498.90</u> ppm	CH4 Equiv Conc.	1070.90 ppm
C3H8 Cal Gas Conc.	<u>208.00</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>498.90</u> ppm	CH4 Equiv Conc.	1070.90 ppm
Removed C3H8 Conc.	<u>208.00</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5258
ZAG Make/Model:	Teledyne API 701	Serial Number:	138

Analyzer Information

Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1317958295
Analyzer Range:	0 - 20 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.000628	Background:	NA	0.960
Calibration intercept:	NA	-0.054987	Coefficient:	NA	0.562

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.9	17.11	17.10	1.001
second point	4960	40.0	8.57	8.48	1.010
third point	4980	20.0	4.28	4.18	1.026
as left zero	5000	0.0	0.00	-0.02	----
as left span	4926	80.0	17.11	17.17	0.997
Average Correction Factor					1.012
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Install calibration. New sample inlet filter and ne pump installed. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	September 30, 2023	Previous Calibration:	NA
Station Name:	Harvest Blackgold	Station Number:	AMS511
Start Time (MST):	10:13	End Time (MST):	13:40
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1317958295

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.00	----	Correlation Coefficient	0.999947	≥0.995
17.11	17.10	1.0008			
8.57	8.48	1.0099	Slope	1.000628	0.90 - 1.10
4.28	4.18	1.0255			
			Intercept	-0.054987	+/-1.5



THC Calibration Plot

Date: September 30, 2023

Location: Harvest Blackgold





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Blackgold	Station number:	AMS511
Calibration Date:	October 5, 2023	Last Cal Date:	September 30, 2023
Start time (MST):	11:30	End time (MST):	16:20
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC147416	Cal Gas Expiry Date:	January 5, 2029
CH4 Cal Gas Conc.	<u>498.90</u> ppm	CH4 Equiv Conc.	1070.90 ppm
C3H8 Cal Gas Conc.	<u>208.00</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>498.90</u> ppm	CH4 Equiv Conc.	1070.90 ppm
Removed C3H8 Conc.	<u>208.00</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5258
ZAG Make/Model:	Teledyne API 701	Serial Number:	138

Analyzer Information

Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1317958295
Analyzer Range:	0 - 20 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000628	0.997401	Background:	0.96	0.94
Calibration intercept:	-0.054987	-0.038565	Coefficient:	0.562	0.550

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.04	----
as found span	4920	79.9	17.11	17.43	0.982
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.03	----
high point	4920	79.9	17.11	17.07	1.003
second point	4960	40.0	8.57	8.46	1.013
third point	4980	20.0	4.28	4.17	1.028
as left zero	5000	0.0	0.00	0.01	----
as left span	4926	80.0	17.11	17.13	0.999
Average Correction Factor					1.014
Baseline Corr As found:	17.39	Previous response	17.07	*% change	1.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Changed the Hydrogen cylinder after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 30, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	11:30	End Time (MST):	16:20
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1317958295

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.03	----	Correlation Coefficient	0.999916	
17.11	17.07	1.0025			≥0.995
8.57	8.46	1.0127	Slope	0.997401	
4.28	4.17	1.0277			0.90 - 1.10
			Intercept	-0.038565	+/-1.5



THC Calibration Plot

Date: October 5, 2023

Location: Blackgold





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Blackgold
Calibration Date: September 26, 2023
Start time (MST): 10:26
Reason: Install
Station number: AMS511
Last Cal Date: NA
End time (MST): 14:57

Calibration Standards

NO Gas Cylinder #: T2XX7A5
NOX Cal Gas Conc: 51.0 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 51.0 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 14, 2024
NO Cal Gas Conc: 49.1 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 49.1 ppm
NO gas Diff:
Serial Number: 5258
Serial Number: 138

Analyzer Information

Analyzer make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 7029

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	NA	1.024	NO bkgnd or offset:	NA	-0.2
NOX coeff or slope:	NA	1.019	NOX bkgnd or offset:	NA	2.5
NO2 coeff or slope:	NA	1.000	Reaction cell Press:	NA	4.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	NA	1.001353
NO _x Cal Offset:	NA	-2.877465
NO Cal Slope:	NA	1.001728
NO Cal Offset:	NA	-2.897845
NO ₂ Cal Slope:	NA	0.999779
NO ₂ Cal Offset:	NA	-1.154126



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-1.2	-0.1	-1.1	----	----
high point	4924	81.6	831.4	799.9	31.5	830.2	799.4	30.7	1.0014	1.0007
second point	4966	40.9	416.6	400.8	15.8	414.3	398.5	15.8	1.0056	1.0059
third point	4986	20.6	209.8	201.9	7.9	205.0	195.7	9.3	1.0236	1.0317
as left zero	5000	0.0	0.0	0.0	0.0	-0.8	0.4	-1.1	----	----
as left span	4924	81.6	831.4	405.2	426.2	818.5	394.8	423.6	1.0157	1.0264
Average Correction Factor									1.0102	1.0127

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	787.4	392.7	426.2	424.7	1.0034	99.7%
2nd GPT point (200 ppb O ₃)	787.4	598.3	220.6	220.1	1.0021	99.8%
3rd GPT point (100 ppb O ₃)	787.4	694.6	124.3	122.5	1.0144	98.6%
Average Correction Factor					1.0066	99.3%

Notes: Install calibration. Changed sample inlet filter before calibrator zero. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	September 26, 2023	Previous Calibration:	NA
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	10:26	End Time (MST):	14:57
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-1.2	----	Correlation Coefficient 0.999978	≥0.995	
831.4	830.2	1.0014			
416.6	414.3	1.0056			
209.8	205.0	1.0236			
			Slope	1.001353	0.90 - 1.10
			Intercept	-2.877465	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	September 26, 2023	Previous Calibration:	NA
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	10:26	End Time (MST):	14:57
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient 0.999937	≥0.995	
799.9	799.4	1.0007			
400.8	398.5	1.0059			
201.9	195.7	1.0317			
			Slope	1.001728	0.90 - 1.10
			Intercept	-2.897845	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	September 26, 2023	Previous Calibration:	NA
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	10:26	End Time (MST):	14:57
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-1.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
426.2	424.7	1.0034		
220.6	220.1	1.0021		
124.3	122.5	1.0144		
			0.999990	
			0.999779	
			-1.154126	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: September 26, 2023

Location: Blackgold





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Blackgold
Calibration Date: October 5, 2023
Start time (MST): 7:06
Reason: Routine
Station number: AMS511
Last Cal Date: September 26, 2023
End time (MST): 16:02

Calibration Standards

NO Gas Cylinder #: T2XX7A5
NOX Cal Gas Conc: 51.0 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 51.0 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T750
ZAG make/model: Teledyne API T751H
Cal Gas Expiry Date: January 14, 2024
NO Cal Gas Conc: 49.1 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 49.1 ppm
NO gas Diff:
Serial Number: 282
Serial Number: 321

Analyzer Information

Analyzer make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 7029

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.024	1.029	NO bkgnd or offset:	-0.2	0.2
NOX coeff or slope:	1.019	1.029	NOX bkgnd or offset:	2.5	1.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.2	4.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001353	0.994686
NO _x Cal Offset:	-2.877465	1.102483
NO Cal Slope:	1.001728	0.983189
NO Cal Offset:	-2.897845	2.228032
NO ₂ Cal Slope:	0.999779	1.005880
NO ₂ Cal Offset:	-1.154126	0.450694



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.4	-0.5	----	----
as found span	4924	81.6	831.4	799.9	31.5	819.0	789.9	28.8	1.0151	1.0127
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.3	0.2	----	----
high point	4924	81.6	831.4	799.9	31.5	827.0	786.2	40.7	1.0053	1.0175
second point	4966	40.9	416.6	400.8	15.8	417.5	401.3	16.3	0.9979	0.9989
third point	4986	20.6	209.8	201.9	7.9	210.1	200.8	9.3	0.9988	1.0055
as left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.2	0.2	----	----
as left span	4924	81.6	831.4	404.4	427.0	826.0	81.4	744.6	1.0065	4.9684
Average Correction Factor									1.0006	1.0073

Corrected As found	NO _x = 819.1 ppb	NO = 789.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.3%	
Previous Response	NO _x = 829.6 ppb	NO = 798.4 ppb		*Percent Change	NO = -1.1%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	795.9	400.4	427.0	429.6	0.9939	100.6%
2nd GPT point (200 ppb O ₃)	795.9	599.0	228.4	230.8	0.9894	101.1%
3rd GPT point (100 ppb O ₃)	795.9	698.6	128.8	129.9	0.9912	100.9%
Average Correction Factor					0.9915	100.9%

Notes: Portable setup used for calibration. Adjusted zero and span. As left span was out, the current calibrator is outputting inconsistent GPT points, will look to replace the calibrator this month.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 26, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	7:06	End Time (MST):	16:02
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
831.4	827.0	1.0053			
416.6	417.5	0.9979			
209.8	210.1	0.9988			
			Slope	0.994686	0.90 - 1.10
			Intercept	1.102483	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 26, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	7:06	End Time (MST):	16:02
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
799.9	786.2	1.0175			
400.8	401.3	0.9989			
201.9	200.8	1.0055			
			Slope	0.983189	0.90 - 1.10
			Intercept	2.228032	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	October 5, 2023	Previous Calibration:	September 26, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	7:06	End Time (MST):	16:02
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
427.0	429.6	0.9939		
228.4	230.8	0.9894		
128.8	129.9	0.9912		
			0.999994	
			1.005880	
			0.450694	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: October 5, 2023

Location: Blackgold





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Blackgold
Calibration Date: October 5, 2023
Start time (MST): 7:06
Reason: Routine
Station number: AMS511
Last Cal Date: September 26, 2023
End time (MST): 9:25

Calibration Standards

NO Gas Cylinder #: T0F8P52
NOX Cal Gas Conc: 47.4 ppm
Removed Cylinder #: T2XX7A5
Removed Gas NOX Conc: 51.0 ppm
NOX gas Diff: 2.5%
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: August 16, 2026
NO Cal Gas Conc: 47.4 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 49.1 ppm
NO gas Diff: 2.5%
Serial Number: 2445
Serial Number: 138

Analyzer Information

Analyzer make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 7029

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.024	1.029	NO bkgnd or offset:	-0.2	0.2
NOX coeff or slope:	1.019	1.029	NOX bkgnd or offset:	2.5	1.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.2	4.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.994686	
NO _x Cal Offset:	1.102483	
NO Cal Slope:	0.983189	
NO Cal Offset:	2.228032	
NO ₂ Cal Slope:	1.005880	
NO ₂ Cal Offset:	0.450694	



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span	4924	81.6	831.4	799.9	31.5	803.0	769.2	34.1	1.0354	1.0399
as found 2nd										
as found 3rd										
new cyl resp	4916	84.4	800.6	800.6	0.0	792.4	788.4	4.0	1.0103	1.0154
calibrator zero										
high point										
second point										
third point										
as left zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.2	-0.3	----	----
as left span	4916	84.4	800.6			785.3	396.1	389.2	1.0194	

Average Correction Factor

Corrected As found	NO _x = NA ppb	NO = NA ppb	<i>* => +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = 828.1 ppb	NO = 788.7 ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)						
2nd GPT point (200 ppb O3)						
3rd GPT point (100 ppb O3)						
Average Correction Factor						


Notes: Changed NO cylinder, cylinder#: T2XX7A5 does not meet internal audit standards. Replaced with cylinder#: T0F8P52.

Calibration Performed By: Braiden Boutilier

NO_x Calibration Plot

Date: October 5, 2023

Location: Blackgold





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Blackgold	Station Number:	AMS 511
Calibration Date:	September 26 & 28th, 2023	Prev Cal Date:	NA
Start Time (MST):	11:52	End Time (MST):	13:58
Tower Height (m):	10.0	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	B10017
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥ 0.9995
Calculated slope		0.998909	<i>0.90 - 1.10</i>
Calculated intercept		0.030357	<i>+/- 2</i>

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	D13603
As Found Declination (deg east of True North):	<u>NA</u>	As Left Declination (deg east of True North):	<u>13</u>
Solar noon time (MST):	13:18	Calc Declination*:	13.08 Degrees
Deadband calc:	1.1 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.6	---
90	91.2	0.3%
180	181.9	0.5%
270	271.7	0.5%
357	356.5	-0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999978	≥ 0.9995
Calculated slope		1.001818	<i>0.90 - 1.10</i>
Calculated intercept		-1.308016	<i>+/- 4</i>

Notes: Install calibration, WS completed Sep 26th and WD was done on Sep 28th. Set to true North as measured by solar noon and compass.

Calibration Performed By: Kelly Baragar/Devin Russell



End of Report