



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

NOVEMBER 2023 MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

December 22, 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION


CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01
BERTHA GANTER - FORT MCKAY

NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	November 1, 2023	Last Cal Date:	October 4, 2023
Start time (MST):	9:15	End time (MST):	14:07
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC418809			
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998709	0.999167	Backgd or Offset:	19.2	19.8
Calibration intercept:	0.106791	-0.353461	Coeff or Slope:	0.887	0.887

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.7	----
as found span	4918	81.3	800.3	796.7	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4918	81.3	800.3	799.6	1.001
second point	4959	40.7	400.6	399.4	1.003
third point	4979	20.3	199.8	198.9	1.005
as left zero	5000	0.0	0.0	0.5	----
as left span	4918	81.3	800.3	798.3	1.002
Average Correction Factor					1.003

Baseline Corr As found:	796.00	Previous response	799.34	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted zero.

Calibration Performed By: Rene Chamberland & Jan Castro



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 4, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:15	End Time (MST):	14:07
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	≥0.995
800.3	799.6	1.0008		
400.6	399.4	1.0030	Slope	0.90 - 1.10
199.8	198.9	1.0046		
			Intercept	+/-30



SO2 Calibration Plot

Date: October 4, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: November 6, 2023 Last Cal Date: October 16, 2023
 Start time (MST): 10:22 End time (MST): 16:48
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153461
 Converter make: CD Nova Converter serial #: 470
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993649	0.997792	Backgd or Offset: 2.25	2.28
Calibration intercept:	0.200001	0.180000	Coeff or Slope: 0.906	0.914

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4921	78.4	80.0	79.1	1.013
as found 2nd point	4960	39.2	40.0	40.1	1.000
as found 3rd point	4980	19.6	20.0	20.1	1.000
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	78.4	80.0	80.0	1.000
second point	4960	39.2	40.0	40.1	0.997
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.3	----
as left span	4921	78.4	80.0	78.8	1.015
SO2 Scrubber Check	4919	81.3	813.0	0.0	----

Date of last scrubber change:	December 17, 2021	Ave Corr Factor	0.997
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 79.0 Prev response: 79.69 *% change: -0.9%
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 0.987220 AF Intercept: 0.299995
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999950

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Adjuste span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 16, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:22	End Time (MST):	16:48
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	1.000000	≥0.995
80.0	80.0	0.9999			
40.0	40.1	0.9975	Slope	0.997792	0.90 - 1.10
20.0	20.1	0.9949			
			Intercept	0.180000	+/-3



TRS Calibration Plot

Date: November 6, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: November 6, 2023 Last Cal Date: October 16, 2023
 Start time (MST): 10:22 End time (MST): 16:48
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 5609

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1200326167
 Converter make: Thermo Converter Converter serial #: N/A
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.989711	0.995140	Backgd or Offset: 1.71	1.73
Calibration intercept:	0.316799	0.276804	Coeff or Slope: 0.997	1.009

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4922	78.4	80.0	79.4	1.008
as found 2nd point	4960	39.2	40.0	40.0	1.002
as found 3rd point	4980	19.6	20.0	20.0	1.005
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4922	78.4	80.0	79.8	1.002
second point	4960	39.2	40.0	40.2	0.995
third point	4980	19.6	20.0	20.2	0.990
as left zero	5000	0.0	0.0	1.0	----
as left span	4922	78.4	80.0	77.8	1.028
SO2 Scrubber Check	4919	81.3	813.0	0.0	----

Date of last scrubber change:	March 21, 2022	Ave Corr Factor	0.996
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 79.3 Prev response: 79.46 *% change: -0.2%
 Baseline Corr 2nd AF pt: 39.9 AF Slope: 0.991762 AF Intercept: 0.176812
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999989

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 16, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:22	End Time (MST):	16:48
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999993	≥0.995
80.0	79.8	1.0022			
40.0	40.2	0.9950	Slope	0.995140	0.90 - 1.10
20.0	20.2	0.9900			
			Intercept	0.276804	+/-3



H₂S Calibration Plot

Date: November 6, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	November 1, 2023	Last Cal Date:	October 4, 2023
Start time (MST):	9:15	End time (MST):	14:07
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.98E-04	3.01E-04	NMHC SP Ratio:	6.57E-05
CH ₄ Retention time:	14.6	14.6	NMHC Peak Area:	139930
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.27	16.99	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	17.27	17.26	1.001
second point	4959	40.7	8.64	8.46	1.022
third point	4980	20.3	4.31	4.21	1.023
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.27	17.31	0.998
Average Correction Factor					1.015

Baseline Corr AF:	16.99	Prev response	17.24	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4918	81.3	9.18	8.97	1.024
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	9.18	9.17	1.002
second point	4959	40.7	4.60	4.53	1.014
third point	4980	20.3	2.29	2.28	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	9.18	9.22	0.996
Average Correction Factor					1.008
Baseline Corr AF:	8.97	Prev response	9.21	*% change	-2.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	8.03	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	8.09	8.09	0.999
second point	4959	40.7	4.05	3.93	1.031
third point	4980	20.3	2.02	1.94	1.042
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.09	0.999
Average Correction Factor					1.024
Baseline Corr AF:	8.03	Prev response	8.03	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002996	1.000035
THC Cal Offset:	-0.080272	-0.073775
CH ₄ Cal Slope:	1.000153	1.002522
CH ₄ Cal Offset:	-0.055151	-0.057596
NMHC Cal Slope:	1.005501	0.998093
NMHC Cal Offset:	-0.025121	-0.016179

Notes: Changed out the inlet filters and N2 cylinder after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland & Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 4, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:15	End Time (MST):	14:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999861	≥ 0.995
17.27	17.26	1.0006			
8.64	8.46	1.0220			
4.31	4.21	1.0230			
			Slope	1.000035	0.90 - 1.10
			Intercept	-0.073775	+/-0.5

THC Calibration Curve





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 4, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:15	End Time (MST):	14:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999684	≥0.995
8.09	8.09	0.9992			
4.05	3.93	1.0307			
2.02	1.94	1.0416			
			Slope	1.002522	0.90 - 1.10
			Intercept	-0.057596	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 4, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:15	End Time (MST):	14:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999953	≥0.995
9.18	9.17	1.0017			
4.60	4.53	1.0141	Slope	0.998093	0.90 - 1.10
2.29	2.28	1.0071			
			Intercept	-0.016179	+/-0.5



NMHC Calibration Plot

Date: November 1, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	November 10, 2023	Last Cal Date:	November 1, 2023
Start time (MST):	11:58	End time (MST):	13:35
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.01E-04	3.01E-04	NMHC SP Ratio:	6.84E-05	6.84E-05
CH ₄ Retention time:	14.6	14.6	NMHC Peak Area:	134289	134289
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.27	17.27	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.27	17.15	1.007
Average Correction Factor					

Baseline Corr AF:	17.27	Prev response	17.19	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4918	81.3	9.18	9.17	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	9.18	9.07	1.012
Average Correction Factor					
Baseline Corr AF:	9.17	Prev response	9.15	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	8.10	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.08	1.001
Average Correction Factor					
Baseline Corr AF:	8.10	Prev response	8.05	*% change	0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000035	
THC Cal Offset:	-0.073775	
CH ₄ Cal Slope:	1.002522	
CH ₄ Cal Offset:	-0.057596	
NMHC Cal Slope:	0.998093	
NMHC Cal Offset:	-0.016179	


Notes: Changed out H2 cylinder

Calibration Performed By: Rene Chamberland

NMHC Calibration Plot

Date: November 10, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
Calibration Date: November 3, 2023 Last Cal Date: October 3, 2023
Start time (MST): 9:00 End time (MST): 13:35
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1P9L Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.84 ppm NO Cal Gas Conc: 50.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.84 ppm Removed Gas NO Conc: 50.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3565
ZAG make/model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1218153357
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.378	1.411	NO bkgnd or offset:	7.0	7.2
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	7.8	7.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	191.0	195.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999045	0.999719
NO _x Cal Offset:	-0.740000	-0.780000
NO Cal Slope:	0.998016	0.998701
NO Cal Offset:	-1.060000	-1.300000
NO ₂ Cal Slope:	1.000561	1.003579
NO ₂ Cal Offset:	1.020945	0.884749



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	0.0	-0.3	----	----
as found span	4920	80.0	813.4	800.6	12.8	795.7	780.2	15.5	1.0223	1.0262
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	-0.1	----	----
high point	4920	80.0	813.4	800.6	12.8	812.8	799.0	13.8	1.0008	1.0021
second point	4960	40.0	406.7	400.3	6.4	405.5	397.7	7.8	1.0030	1.0066
third point	4980	20.0	203.4	200.2	3.2	201.7	197.4	4.3	1.0082	1.0140
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.2	----	----
as left span	4920	80.0	813.4	395.8	417.6	811.4	392.4	418.9	1.0025	1.0088
Average Correction Factor									1.0040	1.0075

Corrected As found	NO _x = 796.0 ppb	NO = 780.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -2.0%	
Previous Response	NO _x = 811.9 ppb	NO = 798.0 ppb		*Percent Change	NO = -2.3%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.8	390.0	417.6	419.5	0.9955	100.5%
2nd GPT point (200 ppb O3)	794.8	594.2	213.4	215.4	0.9907	100.9%
3rd GPT point (100 ppb O3)	794.8	695.1	112.5	114.9	0.9791	102.1%
Average Correction Factor					0.9884	101.2%

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland & Jan Castro



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 3, 2023	Previous Calibration:	October 3, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:00	End Time (MST):	13:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
813.4	812.8	1.0008			
406.7	405.5	1.0030			
203.4	201.7	1.0082			
			Slope	0.999719	0.90 - 1.10
			Intercept	-0.780000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 3, 2023	Previous Calibration:	October 3, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:00	End Time (MST):	13:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.6	799.0	1.0021		
400.3	397.7	1.0066		
200.2	197.4	1.0140		
			0.999988	
			0.998701	
			-1.300000	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 3, 2023	Previous Calibration:	October 3, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:00	End Time (MST):	13:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥0.995
417.6	419.5	0.9955		
213.4	215.4	0.9907	Slope	0.90 - 1.10
112.5	114.9	0.9791		
			Intercept	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 3, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: November 2, 2023 Last Cal Date: October 2, 2023
 Start time (MST): 9:57 End time (MST): 13:00
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 1107
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003114	1.000714	Backgd or Offset:	3.2	3.8
Calibration intercept:	0.080000	0.200000	Coeff or Slope:	1.010	1.010

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	5000	863.1	400.0	401.5	0.996
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.0	----
high point	5000	863.1	400.0	400.3	0.999
second point	5000	742.5	200.0	200.7	0.997
third point	5000	651.7	100.0	100.3	0.997
as left zero	5000	0.0	0.0	-0.1	----
as left span	5000	863.1	400.0	403.3	0.992
Average Correction Factor					0.998

Baseline Corr As found:	401.1	Previous response	401.3	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Adjusted zero only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 2, 2023	Previous Calibration:	October 2, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:57	End Time (MST):	13:00
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	≥0.995
400.0	400.3	0.9993		
200.0	200.7	0.9965	Slope	0.90 - 1.10
100.0	100.3	0.9970		
			Intercept	+/- 5



O₃ Calibration Plot

Date: November 2, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
 Calibration Date: November 14, 2023 Last Cal Date: October 19, 2023
 Start time (MST): 12:59 End time (MST): 14:31

Analyzer Make: Teledyne API T640 S/N: 324
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	1.1	0.4	1.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	725.7	723.2	725.7	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.10	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 14, 2023</u>		Last Cal Date: <u>October 19, 2023</u>		
	PM w/o HEPA: <u>6.3</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>October 19, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>October 19, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: September 14, 2023
 Date RH/T Sensor Cleaned: October 19, 2023

Notes: Flow, temperature, and pressure all within limits. Leak check passed.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	November 8, 2023	Last Cal Date:	October 23, 2023
Start time (MST):	10:50	End time (MST):	14:38
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	<u>3040</u>	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>3040</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000806	1.001848	Backgd or Offset:	-0.012	-0.013
Calibration intercept:	0.185827	0.149849	Coeff or Slope:	0.990	0.990

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.6	40.9	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.6	40.6	0.998
second point	4966	33.3	20.2	20.7	0.977
third point	4983	16.7	10.2	10.3	0.986
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.5	40.2	1.009
Average Correction Factor					0.987

Baseline Corr As found:	40.81	Prev response:	40.77	*% change:	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 23, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:50	End Time (MST):	14:38
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999863	
40.6	40.6	0.9979			≥0.995
20.2	20.7	0.9768	Slope	1.001848	
10.2	10.3	0.9858			0.90 - 1.10
			Intercept	0.149849	+/-1.5



CO Calibration Plot

Date: November 8, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	November 7, 2023	Last Cal Date:	October 13, 2023
Start time (MST):	11:06	End time (MST):	14:18
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>60,200</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
N2 Gen Make/Model:	Peak Sci NG5000		Serial Number:	7220900034

Analyzer Information

Analyzer make: Teledyne API 360	Analyzer serial #: 442
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001376	1.001056	Backgd or Offset:	0.045	0.045
Calibration intercept:	-5.300000	-6.100000	Coeff or Slope:	0.876	0.874

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-0.3	----
as found span	2920	80.0	1605.3	1618.0	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.1	----
high point	2920	80.0	1605.3	1604.0	1.001
second point	2960	40.0	802.7	794.4	1.010
third point	2980	20.0	401.3	389.4	1.031
as left zero	3000	0.0	0.0	-0.1	----
as left span	2960	40.0	802.7	781.7	1.027
Average Correction Factor					1.014

Baseline Corr As found:	1618.30	Prev response:	1602.24	*% change:	1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	November 7, 2023	Previous Calibration	October 13, 2023
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS01
Start Time (MST)	11:06	End Time (MST)	14:18
Analyzer make	Teledyne API 360	Analyzer serial #	442

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999932	≥0.995
1605.3	1604.0	1.0008			
802.7	794.4	1.0104	Slope	1.001056	0.90 - 1.10
401.3	389.4	1.0306			
			Intercept	-6.100000	+/-10



CO₂ Calibration Plot

Date: November 7, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	November 14, 2023	Last Cal Date:	October 18, 2023
Start time (MST):	10:53	End time (MST):	15:56
NH3 Cal Date:	November 15, 2023	Last Cal Date:	October 19, 2023
Start time (MST):	11:05	End time (MST):	14:50
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	50.84	ppm	NO Gas Cylinder #:	T2Y1P9L
NO Cal Gas Conc:	50.04	ppm	NO Cal Gas Expiry:	March 3, 2028
Removed NOX Conc:	50.84	ppm	Removed Cylinder #:	NA
Removed NO Conc:	50.04	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.58	ppm	NH3 Gas Cylinder #:	CC743587
			NH3 Cal Gas Expiry:	August 22, 2024
Removed NH3 Conc:	76.58	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	Teledyne API T700		Serial Number:	3565
ZAG make/model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	475
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	5.10
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	506

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.991	0.987	TN coefficient:	0.996	0.989
NOX coefficient:	0.993	0.989	NO bkgnd:	-0.9	-0.9
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-0.3	-0.3
NH3 coefficient:	0.943	0.932	TN bkgnd:	1.2	1.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000492	0.999888
NO _x Cal Offset:	-1.780000	-1.540000
NO Cal Slope:	1.000500	0.999686
NO Cal Offset:	-1.880000	-2.120000
NO ₂ Cal Slope:	1.000741	0.998244
NO ₂ Cal Offset:	-0.538093	-0.546307
NH3 Cal Slope:	0.995734	1.000230
NH3 Cal Offset:	1.216520	-2.765159
TN Cal Slope:	0.998191	1.002817
TN Cal Offset:	1.291202	-2.497861



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.8	0.0	----	----
as found NO	4920	80.0	813.4	813.4	----	815.7	815.6	0.3	0.997	----
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
high NO point	4920	80.0	813.4	813.4	----	812.5	813.0	-0.6	1.001	----
NO/O3 point	4920	80.0	813.4	813.4	----	811.6	810.7	1.1	1.002	----
as found NH3	3416	84.1	1840.1	----	1840.1	1784.0	----	1779.1	1.031	1.034
new NH3 cyl rp							----			
first NH3	3418	82.2	1798.5	----	1798.5	1804.0	----	1799.2	0.997	1.000
second NH3	3454	45.7	1000.0	----	1000.0	995.3	----	992.3	1.005	1.008
third NH3	3477	22.8	498.9	----	498.9	497.4	----	495.7	1.003	1.006
Average Correction Factor									1.0017	1.0046

Corrected As found TN = 816.5 ppb NO_x = 816.4 ppb NH3 = 1779.1 ppb

Previous Response TN = 813.3 ppb NO_x = 812.1 ppb NH3 = 1833.5 ppb

NH3 Previous Converter Efficiency = 94.3%

NH3 Current Converter Efficiency = 93.2%

*Percent Change TN = 0.4%

*Percent Change NO_x = 0.5%

*Percent Change NH3 = -3.1%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.6	-0.6	----	----
as found span	4920	80.0	813.4	800.6	813.4	820.7	804.3	821.5	0.9912	0.9954
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.3	0.0	----	----
high point	4920	80.0	813.4	800.6	813.4	813.0	799.9	812.5	1.0005	1.0009
second point	4960	40.0	406.7	400.3	406.7	403.2	395.6	403.5	1.0087	1.0119
third point	4980	20.0	203.4	200.2	203.4	200.9	196.4	199.9	1.0122	1.0191
Average Correction Factor									1.0072	1.0107

Baseline Corr As fnd	TN = 822.1 ppb	NO _x = 821.3 ppb	NO = 804.9 ppb	*Percent Change	TN = 1.1%
Previous Response	TN = 813.3 ppb	NO _x = 812.1 ppb	NO = 799.2 ppb	*Percent Change	NO _x = 1.1%
				*Percent Change	NO = 0.7%

* = > +/-5% change initiates investigation

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	0.0	----	----
calibration zero	----	----	0.0	-0.2	----	----
1st GPT point (400 ppb O3)	795.6	387.3	421.1	420.0	1.0026	99.7%
2nd GPT point (200 ppb O3)	795.6	593.8	214.6	213.5	1.0052	99.5%
3rd GPT point (100 ppb O3)	795.6	695.2	113.2	112.1	1.0098	99.0%
Average Correction Factor					1.0059	99.4%

Notes: Changed the inlet filter after as founds. Adjusted NO_x/NO/TN span. Adjusted NH₃ span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TN Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 15, 2023	Previous Calibration:	October 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:53	End Time (MST):	15:56
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1798.5	1804.0	0.9970		
1000.0	995.3	1.0047		
498.9	497.4	1.0029		
			0.999977	
			1.002817	
			-2.497861	

TN Calibration Curve





Wood Buffalo Environmental Association

NH₃ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 15, 2023	Previous Calibration:	October 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:53	End Time (MST):	15:56
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1798.5	1799.2	0.9996		
1000.0	992.3	1.0078		
498.9	495.7	1.0064		
			0.999976	
			1.000230	
			-2.765159	

NH₃ Calibration Curve





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	November 14, 2023	Previous Calibration:	October 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:53	End Time (MST):	15:56
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.4	813.0	1.0005		
406.7	403.2	1.0087		
203.4	200.9	1.0122		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 14, 2023	Previous Calibration:	October 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:53	End Time (MST):	15:56
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.3	----	Correlation Coefficient	0.999951	≥0.995
800.6	799.9	1.0009			
400.3	395.6	1.0119	Slope	0.999686	0.90 - 1.10
200.2	196.4	1.0191			
			Intercept	-2.120000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 14, 2023	Previous Calibration:	October 18, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:53	End Time (MST):	15:56
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
421.1	420.0	1.0026			
214.6	213.5	1.0052	Slope	0.90 - 1.10	
113.2	112.1	1.0098			
			Intercept	-0.546307	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 14, 2023


Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: November 15, 2023

Location: Bertha Ganter-Fort McKay






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	November 22, 2023	Last Cal Date:	October 20, 2023
Start time (MST):	10:25	End time (MST):	11:45
Reason:	As Found		

Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1185
ZAG Make/Model:	API T701		Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999174		Backgd or Offset:	17.5	17.5
Calibration intercept:	0.714964		Coeff or Slope:	0.774	0.774

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4920	80.2	801.6	799.3	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

Average Correction Factor

Baseline Corr As found:	798.80	Previous response	801.70	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation


Notes: Generated as founds for a calibration. Flow dropped to near zero, suspecting a failed pump. Will return November 23 with a replacement.

Calibration Performed By: Jan Castro/ Braiden Boutilier

SO2 Calibration Plot

Date: November 22, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	November 23, 2023	Last Cal Date:	October 20, 2023
Start time (MST):	13:17	End time (MST):	15:57
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1185
ZAG Make/Model:	API T701		Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999174	0.996481	Backgd or Offset:	17.5	20.6
Calibration intercept:	0.714964	-0.365997	Coeff or Slope:	0.774	0.763

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.2	801.6	798.5	1.004
second point	4960	40.1	400.8	399.4	1.004
third point	4980	20.0	199.9	198.0	1.010
as left zero	5000	0.0	0.0	-0.3	----
as left span	4920	80.2	801.6	798.7	1.004
Average Correction Factor					1.006

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed pump. Adjusted zero and span.

Calibration Performed By: Jan Castro/ Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 23, 2023	Previous Calibration:	October 20, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	13:17	End Time (MST):	15:57
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999997	≥0.995
801.6	798.5	1.0039			
400.8	399.4	1.0036	Slope	0.996481	0.90 - 1.10
199.9	198.0	1.0097			
			Intercept	-0.365997	+/-30



SO2 Calibration Plot

Date: November 23, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mildred Lake Station number: AMS02
 Calibration Date: November 24, 2023 Last Cal Date: October 25, 2023
 Start time (MST): 11:14 End time (MST): 15:13
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.29 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345191
 Removed Cal Gas Conc: 5.29 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1185
 ZAG Make/Model: API T701 Serial Number: 5608

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 12113311966
 Converter make: Global G150 Converter serial #: 2022-198
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002536	1.010680	Backgd or Offset: 1.71	1.72
Calibration intercept:	-0.059198	-0.019191	Coeff or Slope: 0.754	0.754

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4924	75.6	80.0	81.1	0.988
as found 2nd point	4962	37.8	40.0	40.4	0.992
as found 3rd point	4981	18.9	20.0	20.1	1.000
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4924	75.6	80.0	80.8	0.990
second point	4962	37.8	40.0	40.5	0.988
third point	4981	18.9	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.1	----
as left span	4924	75.6	80.0	81.1	0.986
SO2 Scrubber Check	4920	80.2	802.0	0.2	----
Date of last scrubber change:	20-Sep-23			Ave Corr Factor	0.991
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 81.0 Prev response: 80.13 *% change: 1.1%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 1.013394 AF Intercept: -0.039192
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999986

* = > +/-5% change initiates investigation

Notes: SO2 scrubber checked passed after the calibrated zero. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 24, 2023	Previous Calibration:	October 25, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	11:14	End Time (MST):	15:13
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999995	≥0.995
80.0	80.8	0.9900			
40.0	40.5	0.9875	Slope	1.010680	0.90 - 1.10
20.0	20.1	0.9949			
			Intercept	-0.019191	+/-3



H₂S Calibration Plot

Date: November 24, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	November 22, 2023	Last Cal Date:	October 20, 2023
Start time (MST):	10:25	End time (MST):	15:41
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
C ₃ H ₈ Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
Removed C ₃ H ₈ Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.04E-04	3.07E-04	NMHC SP Ratio:	4.59E-05	4.52E-05
CH ₄ Retention time:	14.8	15.0	NMHC Peak Area:	191592	194477
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	16.80	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.79	1.002
second point	4960	40.1	8.41	8.36	1.006
third point	4980	20.0	4.19	4.17	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.81	1.000
Average Correction Factor					1.005

Baseline Corr AF:	16.80	Prev response	16.75	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	8.90	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.79	1.001
second point	4960	40.1	4.40	4.40	1.000
third point	4980	20.0	2.19	2.20	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.76	1.004
Average Correction Factor					0.999
Baseline Corr AF:	8.90	Prev response	8.80	*% change	1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	7.91	1.015
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.00	1.003
second point	4960	40.1	4.01	3.96	1.014
third point	4980	20.0	2.00	1.96	1.019
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	8.04	0.998
Average Correction Factor					1.012
Baseline Corr AF:	7.91	Prev response	7.95	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.996697	0.998545
THC Cal Offset:	-0.011323	-0.015916
CH ₄ Cal Slope:	0.994304	0.997280
CH ₄ Cal Offset:	-0.023065	-0.020258
NMHC Cal Slope:	0.998828	0.999088
NMHC Cal Offset:	0.011943	0.003941

Notes: Changed filters after as founds. Change hydrogen and nitrogen cylinder. Span adjusted only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 22, 2023	Previous Calibration:	October 20, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:25	End Time (MST):	15:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995
16.82	16.79	1.0017			
8.41	8.36	1.0059			
4.19	4.17	1.0070			
			Slope	0.998545	0.90 - 1.10
			Intercept	-0.015916	+/-0.5

THC Calibration Curve





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 22, 2023	Previous Calibration:	October 20, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:25	End Time (MST):	15:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999964	≥0.995			
8.02	8.00	1.0035						
4.01	3.96	1.0135				Slope	0.997280	0.90 - 1.10
2.00	1.96	1.0193						
			Intercept	-0.020258	+/-0.5			

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 22, 2023	Previous Calibration:	October 20, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:25	End Time (MST):	15:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995
8.80	8.79	1.0006			
4.40	4.40	1.0002	Slope	0.999088	0.90 - 1.10
2.19	2.20	0.9965			
			Intercept	0.003941	± 0.5



NMHC Calibration Plot

Date: November 22, 2023

Location: Mildred Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023







Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	November 28, 2023	Last Cal Date:	October 10, 2023
Start time (MST):	8:45	End time (MST):	11:36
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.87	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC446753			
Removed Cal Gas Conc:	50.87	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701		Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003929	1.001971	Backgd or Offset:	24.1	24.1
Calibration intercept:	-0.144528	1.315277	Coeff or Slope:	0.873	0.873

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4921	78.6	799.7	804.0	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.5	----
high point	4921	78.6	799.7	801.9	0.997
second point	4961	39.3	399.8	403.3	0.991
third point	4980	19.6	199.4	201.3	0.991
as left zero	5000	0.0	0.0	0.5	----
as left span	4921	78.6	799.7	808.7	0.989
Average Correction Factor					0.993

Baseline Corr As found:	803.80	Previous response	802.74	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments and maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 10, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:45	End Time (MST):	11:36
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999991	≥0.995
799.7	801.9	0.9973			
399.8	403.3	0.9914	Slope	1.001971	0.90 - 1.10
199.4	201.3	0.9907			
			Intercept	1.315277	+/-30



SO2 Calibration Plot

Date: November 28, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	November 1, 2023	Last Cal Date:	October 13, 2023
Start time (MST):	6:40	End time (MST):	11:02
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.42	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC345266			
Removed Cal Gas Conc:	5.42	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701H		Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400
Converter make:	Global	Converter serial #:	2022-200
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994226	0.998493	Backgd or Offset:	1.8
Calibration intercept:	0.082074	0.082213	Coeff or Slope:	1.095
				1.115

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.1	80.3	79.3	1.013
as found 2nd point	4963	37.0	40.1	39.8	1.008
as found 3rd point	4982	18.5	20.1	19.6	1.023
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	74.1	80.3	80.2	1.002
second point	4963	37.0	40.1	40.4	0.993
third point	4982	18.5	20.1	19.9	1.008
as left zero	5000	0.0	0.0	0.1	----
as left span	4926	74.1	80.3	79.5	1.010
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	1.001
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	79.3	Prev response:	79.94	*% change:	-0.8%
Baseline Corr 2nd AF pt:	39.8	AF Slope:	0.988393	AF Intercept:	-0.038044
Baseline Corr 3rd AF pt:	19.6	AF Correlation:	0.999978		

* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 13, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:40	End Time (MST):	11:02
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
80.3	80.2	1.0015		
40.1	40.4	0.9928	Slope	0.90 - 1.10
20.1	19.9	1.0076		
			Intercept	+/-3



H₂S Calibration Plot

Date: November 1, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	November 28, 2023	Last Cal Date:	October 10, 2023
Start time (MST):	8:45	End time (MST):	11:35
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	NA	Cal Gas Expiry Date:	March 10, 1931
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	1.95E-04	2.00E-04	NMHC SP Ratio:	4.19E-05
CH ₄ Retention time:	11.8	11.8	NMHC Peak Area:	210347
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON
				ON

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	16.64	16.06	1.036
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	16.64	16.61	1.002
second point	4961	39.3	8.32	8.22	1.012
third point	4980	19.6	4.15	4.06	1.022
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	16.64	16.60	1.002
Average Correction Factor					1.012

Baseline Corr AF:	16.06	Prev response	16.61	*% change	-3.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	8.82	8.44	1.045
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	8.82	8.80	1.002
second point	4961	39.3	4.41	4.37	1.009
third point	4980	19.6	2.20	2.16	1.018
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	8.82	8.80	1.002
Average Correction Factor					1.010
Baseline Corr AF:	8.44	Prev response	8.80	*% change	-4.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	7.82	7.62	1.026
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	7.82	7.81	1.001
second point	4961	39.3	3.91	3.85	1.015
third point	4980	19.6	1.95	1.90	1.026
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	7.82	7.80	1.002
Average Correction Factor					1.014
Baseline Corr AF:	7.62	Prev response	7.82	*% change	-2.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000279	0.999538
THC Cal Offset:	-0.027945	-0.049557
CH ₄ Cal Slope:	1.002623	1.000446
CH ₄ Cal Offset:	-0.020104	-0.029913
NMHC Cal Slope:	0.998214	0.998967
NMHC Cal Offset:	-0.007641	-0.020044

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 10, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:45	End Time (MST):	11:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999956	≥ 0.995			
16.64	16.61	1.0016						
8.32	8.22	1.0118				Slope	0.999538	0.90 - 1.10
4.15	4.06	1.0218						
			Intercept	-0.049557	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 10, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:45	End Time (MST):	11:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999923	≥0.995
7.82	7.81	1.0008			
3.91	3.85	1.0150			
1.95	1.90	1.0259			
			Slope	1.000446	0.90 - 1.10
			Intercept	-0.029913	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 10, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:45	End Time (MST):	11:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999976	≥ 0.995			
8.82	8.80	1.0020						
4.41	4.37	1.0090				Slope	0.998967	0.90 - 1.10
2.20	2.16	1.0182						
			Intercept	-0.020044	+/-0.5			



NMHC Calibration Plot

Date: November 28, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: November 6, 2023
Start time (MST): 7:10
Reason: Routine
Station number: AMS04
Last Cal Date: October 18, 2023
End time (MST): 12:08

Calibration Standards

NO Gas Cylinder #: T36RH1F
NOX Cal Gas Conc: 51.16 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 51.16 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: August 18, 2023
NO Cal Gas Conc: 50.91 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.91 ppm
NO gas Diff:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 721

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.180	1.185	NO bkgnd or offset:	-0.6	-0.6
NOX coeff or slope:	1.178	1.181	NOX bkgnd or offset:	-0.3	-0.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.4	4.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000587	1.001677
NO _x Cal Offset:	-0.272971	-0.753866
NO Cal Slope:	0.998332	0.998936
NO Cal Offset:	-0.653356	-1.613678
NO ₂ Cal Slope:	1.005579	0.999002
NO ₂ Cal Offset:	-0.636037	0.854730



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.5	0.4	0.1	----	----
as found span	4922	78.1	799.1	795.2	3.9	791.1	784.2	6.9	1.0101	1.0140
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.5	0.5	0.0	----	----
high point	4922	78.1	799.1	795.2	3.9	800.2	793.8	6.5	0.9986	1.0018
second point	4961	39.1	400.1	398.1	2.0	399.7	395.0	4.7	1.0009	1.0079
third point	4981	19.5	199.5	198.5	1.0	197.6	194.6	3.0	1.0096	1.0202
as left zero	5000	0.0	0.0	0.0	0.0	0.5	0.9	-0.4	----	----
as left span	4922	78.1	799.1	381.6	417.5	793.2	378.2	414.9	1.0074	1.0090
Average Correction Factor									1.0031	1.0099

Corrected As found	NO _x = 790.6 ppb	NO = 783.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.1%	
Previous Response	NO _x = 799.3 ppb	NO = 793.2 ppb		*Percent Change	NO = -1.2%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	791.7	378.1	417.5	417.5	1.0000	100.0%
2nd GPT point (200 ppb O3)	791.7	583.6	212.0	213.1	0.9949	100.5%
3rd GPT point (100 ppb O3)	791.7	690.3	105.3	106.9	0.9851	101.5%
Average Correction Factor					0.9933	100.7%

Notes:

No Maintenance Done. Span adjusted.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 18, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:10	End Time (MST):	12:08
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	≥0.995	
799.1	800.2	0.9986			
400.1	399.7	1.0009			
199.5	197.6	1.0096			
			Slope	1.001677	0.90 - 1.10
			Intercept	-0.753866	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 18, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:10	End Time (MST):	12:08
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999968	≥0.995
795.2	793.8	1.0018			
398.1	395.0	1.0079	Slope	0.998936	0.90 - 1.10
198.5	194.6	1.0202			
			Intercept	-1.613678	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 18, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:10	End Time (MST):	12:08
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
417.5	417.5	1.0000		
212.0	213.1	0.9949		
105.3	106.9	0.9851		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 6, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: November 22, 2023 Last Cal Date: October 10, 2023
 Start time (MST): 7:15 End time (MST): 8:40
 Reason: Removal

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 2961
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995171		Backgd or Offset:	-2.6	-2.6
Calibration intercept:	0.820000		Coeff or Slope:	1.008	1.008

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.6	----
as found span	5000	978.0	400.0	396.3	1.009
as found 2nd point	5000	811.0	200.0	197.9	1.011
as found 3rd point	5000	700.1	100.0	99.3	1.007
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

Average Correction Factor					
Baseline Corr As found:	396.9	Previous response	398.9	*% change	-0.5%
Baseline Corr 2nd AF pt:	-198.4	AF Slope:	0.991686	AF Intercept:	-0.320000
Baseline Corr 3rd AF pt:	-98.6	AF Correlation:	0.999997		

* = > +/-5% change initiates investigation


Notes: Removal Due to unstable Zero

Calibration Performed By: Melissa Lemay

O₃ Calibration Plot

Date: November 22, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: November 22, 2023 Last Cal Date:
 Start time (MST): 7:15 End time (MST): 11:39
 Reason: Install

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 7045
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		1.005543	Backgd or Offset:		-1.4
Calibration intercept:		0.680000	Coeff or Slope:		1.022

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.3	----
high point	5000	980.5	400.0	402.8	0.993
second point	5000	812.0	200.0	201.7	0.992
third point	5000	701.5	100.0	101.8	0.982
as left zero	5000	0.0	0.0	1.1	----
as left span	5000	979.7	400.0	405.4	0.987
Average Correction Factor					0.989

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Installed due to Old analyzer having unstable zeroes. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 22, 2023	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:15	End Time (MST):	11:39
Analyzer make:	API T400	Analyzer serial #:	7045

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999995	≥0.995
400.0	402.8	0.9930			
200.0	201.7	0.9916	Slope	1.005543	0.90 - 1.10
100.0	101.8	0.9823			
			Intercept	0.680000	+/- 5



O₃ Calibration Plot

Date: November 22, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: November 28, 2023 Last Cal Date: October 18, 2023
 Start time (MST): 7:38 End time (MST): 8:44

Analyzer Make: API T640 S/N: 322
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Deltacal S/N: 1451
 Temp/RH standard: Deltacal S/N: 1451

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	1.1	0.8	1.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	722	723.7	722	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.03	4.92	5.03	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 28, 2023</u>	Last Cal Date: <u>October 18, 2023</u>			
	PM w/o HEPA: <u>4.4</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.5	10.9	10.9	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>4</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>November 28, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>November 28, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: November 28, 2023
 Date RH/T Sensor Cleaned: May 23, 2023

Notes: No adjustments done. PM10 leak check at 0.5ug/m3 before cleaning, after cleaning 0.0ug/m3. Leak check, PMT check and Flow checked before and after cleaning.

Calibration by: Melissa Lemay




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS05
MANNIX
NOVEMBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Mannix	Station number:	AMS05
Calibration Date:	November 6, 2023	Last Cal Date:	October 4, 2023
Start time (MST):	10:20	End time (MST):	13:43
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.02	ppm	Cal Gas Exp Date:	January 12, 2029
Cal Gas Cylinder #:	XC026809B			
Removed Cal Gas Conc:	50.02	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	621
ZAG Make/Model:	API T701H		Serial Number:	832

Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 1008841399
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994288	1.001714	Backgd or Offset:	8.9	9.1
Calibration intercept:	1.060000	-0.140000	Coeff or Slope:	0.930	0.944

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4920	80.0	800.3	790.6	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.7	----
high point	4920	80.0	800.3	801.9	0.998
second point	4960	40.0	400.2	400.5	0.999
third point	4980	20.0	200.1	199.3	1.004
as left zero	5000	0.0	0.0	0.7	----
as left span	4920	80.0	800.3	806.6	0.992
Average Correction Factor					1.000

Baseline Corr As found:	790.10	Previous response	796.81	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 4, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	10:20	End Time (MST):	13:43
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient	0.999995	≥0.995
800.3	801.9	0.9980			
400.2	400.5	0.9992	Slope	1.001714	0.90 - 1.10
200.1	199.3	1.0039			
			Intercept	-0.140000	+/-30



SO2 Calibration Plot

Date: November 6, 2023

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mannix Station number: AMS05
 Calibration Date: November 14, 2023 Last Cal Date: October 16, 2023
 Start time (MST): 9:53 End time (MST): 14:44
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.92 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002433
 Removed Cal Gas Conc: 4.92 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1845
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1203169745
 Converter make: Global Converter serial #: 2022-196
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.970189	0.993603	Backgd or Offset: 2.18	2.18
Calibration intercept:	0.340566	0.221019	Coeff or Slope: 0.866	0.866

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4919	81.3	80.0	80.7	0.994
as found 2nd point	4960	40.7	40.0	40.6	0.991
as found 3rd point	4980	20.3	20.0	20.2	0.999
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4919	81.3	80.0	80.0	1.000
second point	4960	40.7	40.0	39.3	1.019
third point	4980	20.3	20.0	20.3	0.984
as left zero	5000	0.0	0.0	0.9	----
as left span	4919	81.3	80.0	77.1	1.038
SO2 Scrubber Check	4920	80.0	800.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	1.001
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 80.5 Prev response: 77.95 *% change: 3.2%
 Baseline Corr 2nd AF pt: 40.4 AF Slope: 1.006900 AF Intercept: 0.180576
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999994

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 14, 2023	Previous Calibration:	October 16, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:53	End Time (MST):	14:44
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999805	≥0.995
80.0	80.0	0.9999			
40.0	39.3	1.0189	Slope	0.993603	0.90 - 1.10
20.0	20.3	0.9839			
			Intercept	0.221019	+/-3



H₂S Calibration Plot

Date: November 14, 2023

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	November 6, 2023	Last Cal Date:	October 4, 2023
Start time (MST):	10:20	End time (MST):	13:43
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.64E-05	2.61E-05	NMHC SP Ratio:	4.50E-05
CH ₄ Retention time:	15.00	15.00	NMHC Peak Area:	203435
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				4.48E-05
				204322
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.40	0.990
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.25	0.999
second point	4960	40.0	8.61	8.62	0.999
third point	4980	20.0	4.31	4.30	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.36	0.992
Average Correction Factor					1.000

Baseline Corr AF:	17.40	Prev response	17.23	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	9.15	9.22	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.17	0.998
second point	4960	40.0	4.57	4.59	0.996
third point	4980	20.0	2.29	2.30	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.15	9.23	0.991
Average Correction Factor					0.996
Baseline Corr AF:	9.22	Prev response	9.15	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	8.18	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.08	1.000
second point	4960	40.0	4.04	4.03	1.003
third point	4980	20.0	2.02	2.00	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.13	0.994
Average Correction Factor					1.004
Baseline Corr AF:	8.18	Prev response	8.07	*% change	1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999944	1.001390
THC Cal Offset:	0.001800	-0.005600
CH ₄ Cal Slope:	1.001075	1.000679
CH ₄ Cal Offset:	-0.012600	-0.010200
NMHC Cal Slope:	0.998844	1.001993
NMHC Cal Offset:	0.014800	0.004200

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 4, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:20	End Time (MST):	13:43
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
17.23	17.25	0.9988						
8.61	8.62	0.9992				Slope	1.001390	0.90 - 1.10
4.31	4.30	1.0017						
			Intercept	-0.005600	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 4, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:20	End Time (MST):	13:43
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999993	≥0.995
8.08	8.08	0.9999			
4.04	4.03	1.0033	Slope	1.000679	0.90 - 1.10
2.02	2.00	1.0093			
			Intercept	-0.010200	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 4, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:20	End Time (MST):	13:43
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
9.15	9.17	0.9979						
4.57	4.59	0.9960				Slope	1.001993	0.90 - 1.10
2.29	2.30	0.9952				Intercept	0.004200	+/-0.5


NMHC Calibration Curve



NMHC Calibration Plot

Date: November 6, 2023

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	November 22, 2023	Last Cal Date:	November 6, 2023
Start time (MST):	10:23	End time (MST):	12:45
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.61E-05	2.61E-05	NMHC SP Ratio:	4.48E-05	4.48E-05
CH ₄ Retention time:	15.00	15.00	NMHC Peak Area:	204322	204322
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.52	0.983
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.44	0.988
second point					
third point					
as left zero					
as left span					
Average Correction Factor					0.988

Baseline Corr AF:	17.52	Prev response	17.24	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	9.15	9.32	0.982
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.27	0.987
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.987
Baseline Corr AF:	9.32	Prev response	9.17	*% change	1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	8.19	0.986
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.17	0.989
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.989
Baseline Corr AF:	8.19	Prev response	8.07	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001390	1.012539
THC Cal Offset:	-0.005600	0.000000
CH ₄ Cal Slope:	1.000679	1.011586
CH ₄ Cal Offset:	-0.010200	0.000000
NMHC Cal Slope:	1.001993	1.013381
NMHC Cal Offset:	0.004200	0.000000


Notes: Changed the N2 and H2 cylinders after as founds.

Calibration Performed By: Max Farrell

NMHC Calibration Plot

Date: November 22, 2023

Location: Mannix






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	November 17, 2023	Last Cal Date:	October 19, 2023
Start time (MST):	9:40	End time (MST):	13:07
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.78	ppm	Cal Gas Exp Date:	September 9, 2024
Cal Gas Cylinder #:	AAL070632			
Removed Cal Gas Conc:	49.78	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	5608

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998819	0.990369	Backgd or Offset:	17.4
Calibration intercept:	2.459939	2.041893	Coeff or Slope:	0.911

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4920	80.3	799.5	793.0	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.3	----
high point	4920	80.3	799.5	792.9	1.008
second point	4960	40.2	400.2	399.4	1.002
third point	4980	20.1	200.1	201.9	0.991
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	80.3	799.5	795.5	1.005
Average Correction Factor					1.001

Baseline Corr As found:	793.20	Previous response	800.98	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 17, 2023	Previous Calibration:	October 19, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:40	End Time (MST):	13:07
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	≥0.995
799.5	792.9	1.0083		
400.2	399.4	1.0021	Slope	0.90 - 1.10
200.1	201.9	0.9912		
			Intercept	+/-30



SO2 Calibration Plot

Date: November 17, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	November 15, 2023	Last Cal Date:	October 13, 2023
Start time (MST):	9:19	End time (MST):	13:35
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.328	ppm	Cal Gas Exp Date:	February 14, 2025
Cal Gas Cylinder #:	CC506659			
Removed Cal Gas Conc:	5.328	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995918	1.002314	Backgd or Offset:	2.05	2.01
Calibration intercept:	0.297166	0.340228	Coeff or Slope:	1.189	1.168

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4925	75.1	80.0	80.3	0.997
as found 2nd point	4963	37.5	40.0	40.7	0.982
as found 3rd point	4981	18.8	20.0	20.7	0.968
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4925	75.1	80.0	80.4	0.995
second point	4963	37.5	40.0	40.6	0.984
third point	4981	18.8	20.0	20.6	0.973
as left zero	5000	0.0	0.0	0.1	----
as left span	4925	75.1	80.0	80.3	0.997
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	0.984
Date of last converter efficiency test:	efficiency				

Baseline Corr As found:	80.3	Prev response:	80.00	*% change:	0.4%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.001742	AF Intercept:	0.360261
Baseline Corr 3rd AF pt:	20.7	AF Correlation:	0.999900		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 15, 2023	Previous Calibration:	October 13, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:19	End Time (MST):	13:35
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999955	≥0.995
80.0	80.4	0.9953			
40.0	40.6	0.9841	Slope	1.002314	0.90 - 1.10
20.0	20.6	0.9725			
			Intercept	0.340228	+/-3



TRS Calibration Plot

Date: November 15, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	November 17, 2023	Last Cal Date:	October 19, 2023
Start time (MST):	9:40	End time (MST):	13:06
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH ₄ Cal Gas Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.15E-04	2.15E-04	NMHC SP Ratio:	4.87E-05
CH ₄ Retention time:	14.0	14.0	NMHC Peak Area:	186260
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4920	80.3	17.12	17.11	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	17.22	0.994
second point	4960	40.2	8.57	8.62	0.994
third point	4980	20.1	4.29	4.36	0.983
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.12	17.15	0.999
Average Correction Factor					0.991

Baseline Corr AF:	17.11	Prev response	17.16	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919.7	80.3	9.07	9.07	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	9.07	9.14	0.992
second point	4960	40.2	4.54	4.59	0.990
third point	4980	20.1	2.27	2.33	0.974
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	9.07	9.08	0.999
Average Correction Factor					0.985
Baseline Corr AF:	9.06	Prev response	9.10	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919.7	80.3	8.06	8.05	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	8.06	8.09	0.996
second point	4960	40.2	4.03	4.04	0.999
third point	4980	20.1	2.02	2.03	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	8.06	8.06	0.999
Average Correction Factor					0.997
Baseline Corr AF:	8.05	Prev response	8.06	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000103	1.004841
THC Cal Offset:	0.036627	0.018616
CH ₄ Cal Slope:	0.999903	1.003960
CH ₄ Cal Offset:	0.001645	-0.002161
NMHC Cal Slope:	1.000180	1.006556
NMHC Cal Offset:	0.035381	0.019576

Notes: Changed the inlet filter and the N2 cylinder after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 17, 2023	Previous Calibration:	October 19, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:40	End Time (MST):	13:06
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999990	≥ 0.995			
17.12	17.22	0.9942						
8.57	8.62	0.9944				Slope	1.004841	0.90 - 1.10
4.29	4.36	0.9835						
			Intercept	0.018616	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 17, 2023	Previous Calibration:	October 19, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:40	End Time (MST):	13:06
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	≥0.995
8.06	8.09	0.9958			
4.03	4.04	0.9994			
2.02	2.03	0.9948			
			Slope	1.003960	0.90 - 1.10
			Intercept	-0.002161	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 17, 2023	Previous Calibration:	October 19, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:40	End Time (MST):	13:06
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999975	≥ 0.995			
9.07	9.14	0.9920						
4.54	4.59	0.9895				Slope	1.006556	0.90 - 1.10
2.27	2.33	0.9736						
			Intercept	0.019576	± 0.5			



NMHC Calibration Plot

Date: November 17, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Patricia McInnes
Calibration Date: November 7, 2023
Start time (MST): 9:31
Reason: Routine
Station number: AMS06
Last Cal Date: October 12, 2023
End time (MST): 14:13

Calibration Standards

NO Gas Cylinder #: T30YCWN
NOX Cal Gas Conc: 47.94 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 47.94 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: April 11, 2025
NO Cal Gas Conc: 46.39 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 46.39 ppm
NO gas Diff:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1172750022

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.825	0.825	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.987	0.987	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.2	157.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000493	0.988764
NO _x Cal Offset:	2.095627	2.638018
NO Cal Slope:	1.001945	0.990325
NO Cal Offset:	0.962538	1.304636
NO ₂ Cal Slope:	1.003982	0.998490
NO ₂ Cal Offset:	-0.085539	0.166835



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.9	0.5	0.4	----	----
as found span	4914	86.2	826.5	799.7	26.7	820.3	790.9	29.3	1.0075	1.0112
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.8	0.5	0.3	----	----
high point	4914	86.2	826.5	799.7	26.7	818.9	793.0	25.9	1.0092	1.0085
second point	4957	43.1	413.2	399.9	13.4	412.1	397.4	14.7	1.0028	1.0062
third point	4978	21.6	207.1	200.4	6.7	209.3	200.8	8.6	0.9896	0.9981
as left zero	5000	0.0	0.0	0.0	0.0	0.5	0.5	0.0	----	----
as left span	4914	86.2	826.5	409.4	417.0	817.7	402.9	414.8	1.0107	1.0162
Average Correction Factor									1.0005	1.0043

Corrected As found	NO _x = 819.4 ppb	NO = 790.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.2%	
Previous Response	NO _x = 829.0 ppb	NO = 802.2 ppb		*Percent Change	NO = -1.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	791.0	400.7	417.0	416.6	1.0010	99.9%
2nd GPT point (200 ppb O3)	791.0	604.7	213.0	212.9	1.0006	99.9%
3rd GPT point (100 ppb O3)	791.0	699.4	118.3	118.1	1.0019	99.8%
Average Correction Factor					1.0011	99.9%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 12, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:31	End Time (MST):	14:13
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.8	----	Correlation Coefficient	≥0.995	
826.5	818.9	1.0092			
413.2	412.1	1.0028			
207.1	209.3	0.9896			
			Slope	0.988764	0.90 - 1.10
			Intercept	2.638018	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 12, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:31	End Time (MST):	14:13
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	≥0.995	
799.7	793.0	1.0085			
399.9	397.4	1.0062			
200.4	200.8	0.9981			
			Slope	0.990325	0.90 - 1.10
			Intercept	1.304636	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 12, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:31	End Time (MST):	14:13
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
417.0	416.6	1.0010		
213.0	212.9	1.0006		
118.3	118.1	1.0019		
			0.999999	
			0.998490	
			0.166835	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 7, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Patricia McInnes Station number: AMS06
 Calibration Date: November 10, 2023 Last Cal Date: October 2, 2023
 Start time (MST): 9:48 End time (MST): 12:45
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3566
 ZAG Make/Model: API T701 Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1300156234
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003714	1.005971	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	-2.700000	-1.720000	Coeff or Slope:	1.039	1.039

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	-0.4	----
as found span	5000	1303.0	400.0	401.6	0.996
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.0	----
high point	5000	1303.0	400.0	401.5	0.996
second point	5000	966.5	200.0	198.7	1.007
third point	5000	794.3	100.0	97.1	1.030
as left zero	5000	800.0	0.0	-0.1	----
as left span	5000	1303.0	400.0	405.2	0.987
Average Correction Factor					1.011

Baseline Corr As found:	402.0	Previous response	398.8	*% change	0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 10, 2023	Previous Calibration:	October 2, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:48	End Time (MST):	12:45
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	0.999916	
400.0	401.5	0.9963			≥0.995
200.0	198.7	1.0065	Slope	1.005971	
100.0	97.1	1.0299			0.90 - 1.10
			Intercept	-1.720000	+/- 5



O₃ Calibration Plot

Date: November 10, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: November 17, 2023 Last Cal Date: October 19, 2023
 Start time (MST): 13:13 End time (MST): 13:45

Analyzer Make: API T640 S/N: 766
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT FP-25 S/N: 388755
 Temp/RH standard: ALICAT FP-25 S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	5.2	4.3	5.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	718.1	718.5	718.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	4.81	5.00	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: November 17, 2023 Last Cal Date: October 19, 2023
 PM w/o HEPA: 3.9 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: _____ w/ HEPA: _____
 Date Optical Chamber Cleaned: September 22, 2023 <0.2 ug/m3
 Disposable Filter Changed: September 22, 2023

Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023
 Date RH/T Sensor Cleaned: April 13, 2023

Quarterly calibration was completed in September. Leak check passed, no adjustments made.

Notes:

Calibration by: Max Farrell



Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	November 1, 2023	Last Cal Date:	October 3, 2023
Start time (MST):	8:44	End time (MST):	12:30
NH3 Cal Date:	November 1, 2023	Last Cal Date:	October 3, 2023
Start time (MST):	12:31	End time (MST):	15:43
Reason:	Removal and calibration cylinder change		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	NA
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.29	ppm	NH3 Gas Cylinder #:	EB0108520
Removed NH3 Conc:	77.80	ppm	NH3 Cal Gas Expiry:	August 22, 2024
NH3 gas Diff:	-5.3%		Removed Cylinder #:	CC710812
Calibrator Model:	API T700		Removed cyl Expiry:	March 30, 2023
ZAG make/model:	API T701		Serial Number:	3566
			Serial Number:	4602

Analyzer Information

Analyzer model: Teledyne API T201	Analyzer serial #: 152
Converter model: Teledyne API T501	Converter serial #: 147
NH3 Range (ppb): 0 - 2000 ppb	Reaction cell Press: 6.50
NOX Range (ppb): 0 - 1000 ppb	Sample Flow: 542

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.788	0.788	TN coefficient:	0.789	0.789
NOX coefficient:	0.787	0.787	NO bkgnd:	-0.1	-0.1
NO2 coefficient:	1.000	1.000	NOX bkgnd:	0.0	0.0
NH3 coefficient:	0.908	0.908	TN bkgnd:	0.0	0.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.004061	1.031267
NO _x Cal Offset:	0.254959	0.789511
NO Cal Slope:	1.000231	1.030718
NO Cal Offset:	1.162694	1.216970
NO ₂ Cal Slope:	0.998161	1.002440
NO ₂ Cal Offset:	-0.986846	1.111820
NH3 Cal Slope:	0.995038	1.034784
NH3 Cal Offset:	8.227208	10.702884
TN Cal Slope:	1.001184	1.040414
TN Cal Offset:	7.279855	10.906453



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.5	-0.1	1.6	----	----
as found NO	4914	86.2	826.5	826.5	----	858.3	853.0	5.5	0.963	----
calibrator zero	5000	0.0	0.0	0.0	0.0	1.5	-0.1	1.6	----	----
high NO point	4914	86.2	826.5	826.5	----	858.3	853.0	5.5	0.963	----
NO/O3 point	4914	86.2	826.5	826.5	----	853.4	846.9	6.6	0.968	----
as found NH3	3419	81.0	1800.5	----	1800.5	1876.5	----	1866.3	0.960	0.965
new NH3 cyl rp	3417	82.6	1800.6	----	1800.6	1779.5	----	1769.7	1.012	1.017
first NH3	3419	81.0	1800.5	----	1800.5	1876.5	----	1866.3	0.960	0.965
second NH3	3455	45.0	1000.3	----	1000.3	1062.2	----	1056.3	0.942	0.947
third NH3	3478	22.5	500.1	----	500.1	537.7	----	534.3	0.930	0.936
Average Correction Factor									0.9657	0.9492

Corrected As found TN = 856.8 ppb NO_x = 853.1 ppb NH3 = 1864.7 ppb

Previous Response TN = 834.7 ppb NO_x = 830.1 ppb NH3 = 1799.8 ppb

NH3 Previous Converter Efficiency = 90.8%

NH3 Current Converter Efficiency = 90.8%

*Percent Change TN = 2.6%

*Percent Change NO_x = 2.7%

*Percent Change NH3 = 3.5%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	1.5	----	----
as found span	4914	86.2	826.5	799.7	826.5	853.0	824.4	858.3	0.9689	0.9701
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	1.5	----	----
high point	4914	86.2	826.5	799.7	826.5	853.0	824.4	858.3	0.9689	0.9701
second point	4957	43.1	413.2	399.9	413.2	426.3	415.4	431.7	0.9694	0.9626
third point	4978	21.6	207.1	200.4	207.1	216.0	208.2	219.2	0.9589	0.9626
Average Correction Factor									0.9657	0.9651

Baseline Corr As fnd	TN = 856.8 ppb	NO _x = 853.1 ppb	NO = 824.5 ppb	*Percent Change	TN = 2.6%
Previous Response	TN = 834.7 ppb	NO _x = 830.1 ppb	NO = 801.1 ppb	*Percent Change	NO _x = 2.7%
				*Percent Change	NO = 2.8%
				<i>* = > +/-5% change initiates investigation</i>	

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	0.0	----	----
calibration zero	----	----	0.0	0.0	----	----
1st GPT point (400 ppb O3)	818.4	414.9	430.2	432.0	0.9959	100.4%
2nd GPT point (200 ppb O3)	818.4	622.7	222.4	223.9	0.9934	100.7%
3rd GPT point (100 ppb O3)	818.4	723.2	121.9	125.0	0.9754	102.5%
Average Correction Factor					0.9882	101.2%

Notes: Completing removal calibration due to how slow the instrument reaches baseline after daily spans. Completed the removal calibration and changed the NH3 cal gas at the end and ran a span with the new cal gas. Swapped out the instruments, install calibration will be completed tomorrow.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TN Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 3, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:44	End Time (MST):	12:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	1.5	----	Correlation Coefficient 0.999842	≥0.995	
1800.5	1876.5	0.9595			
1000.3	1062.2	0.9417			
500.1	537.7	0.9300			
			Slope	1.040414	0.90 - 1.10
			Intercept	10.906453	+/-20

TN Calibration Curve





Wood Buffalo Environmental Association

NH₃ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 3, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:44	End Time (MST):	12:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	1.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.5	1866.3	0.9648		
1000.3	1056.3	0.9470		
500.1	534.3	0.9359		
			0.999848	
			1.034784	
			10.702884	

NH₃ Calibration Curve





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 3, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:44	End Time (MST):	12:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
826.5	853.0	0.9689			
413.2	426.3	0.9694			
207.1	216.0	0.9589			
			Slope	1.031267	0.90 - 1.10
			Intercept	0.789511	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 3, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:44	End Time (MST):	12:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999980	≥0.995
799.7	824.4	0.9701			
399.9	415.4	0.9626	Slope	1.030718	0.90 - 1.10
200.4	208.2	0.9626			
			Intercept	1.216970	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 3, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:44	End Time (MST):	12:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.0	----	Correlation Coefficient	0.999958	≥0.995
430.2	432.0	0.9959			
222.4	223.9	0.9934			
121.9	125.0	0.9754			
			Slope	1.002440	0.90 - 1.10
			Intercept	1.111820	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 1, 2023


Location: Patricia McInnes



NH₃ Calibration Plot

Date: November 1, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	November 2, 2023	Last Cal Date:	N/A
Start time (MST):	8:41	End time (MST):	13:00
NH3 Cal Date:	November 2, 2023	Last Cal Date:	N/A
Start time (MST):	13:00	End time (MST):	15:02
Reason:	Install		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	NA
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.29	ppm	NH3 Gas Cylinder #:	EB0108520
			NH3 Cal Gas Expiry:	August 22, 2024
Removed NH3 Conc:	76.29	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	API T700		Serial Number:	3566
ZAG make/model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	808
Converter model:	Teledyne API T501	Converter serial #:	484
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	4.80
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	26.6

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	N/A	0.870	TN coefficient:	N/A	0.858
NOX coefficient:	N/A	0.860	NO bkgrnd:	N/A	-0.985
NO2 coefficient:	N/A	1.000	NOX bkgrnd:	N/A	-0.562
NH3 coefficient:	N/A	0.871	TN bkgrnd:	N/A	5.018

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	N/A	0.996853
NO _x Cal Offset:	N/A	-0.762741
NO Cal Slope:	N/A	1.000156
NO Cal Offset:	N/A	-0.936161
NO ₂ Cal Slope:	N/A	1.003217
NO ₂ Cal Offset:	N/A	-0.787714
NH3 Cal Slope:	N/A	1.025836
NH3 Cal Offset:	N/A	3.642287
TN Cal Slope:	N/A	1.030917
TN Cal Offset:	N/A	3.830269



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found NO										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high NO point	4914	86.2	826.5	826.5	----	824.9	823.9	0.9	1.002	----
NO/O3 point										
as found NH3										
new NH3 cyl rp							----			
first NH3	3419	81.0	1765.6	----	1765.6	1821.6	----	1812.6	0.969	0.974
second NH3	3455	45.0	980.9	----	980.9	1017.4	----	1012.0	0.964	0.969
third NH3	3478	22.5	490.4	----	490.4	513.2	----	510.4	0.956	0.961
Average Correction Factor									1.0019	0.9680

Corrected As found TN = NA ppb NO_x = NA ppb NH3 = NA ppb

Previous Response TN = NA ppb NO_x = NA ppb NH3 = NA ppb

NH3 Previous Converter Efficiency = N/A

NH3 Current Converter Efficiency = 87.1%

*Percent Change TN = NA

*Percent Change NO_x = NA

*Percent Change NH3 = NA

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>		
as found zero												
as found span												
new NO cyl rp												
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----		
high point	4914	86.2	826.5	799.7	826.5	823.9	799.2	824.9	1.0031	1.0007		
second point	4957	43.1	413.2	399.9	413.2	409.5	399.1	410.6	1.0091	1.0019		
third point	4978	21.6	207.1	200.4	207.1	205.8	198.2	205.5	1.0064	1.0112		
Average Correction Factor									1.0062	1.0046		
Baseline Corr As fnd	TN =	NA	ppb	NO _x =	NA	ppb	NO =	NA	ppb	*Percent Change	TN =	NA
Previous Response	TN =	NA	ppb	NO _x =	NA	ppb	NO =	NA	ppb	*Percent Change	NO _x =	NA
										*Percent Change	NO =	NA

** = > +/-5% change initiates investigation*

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero						
calibration zero	----	----	0.0	0.0	----	----
1st GPT point (400 ppb O3)	786.9	399.3	414.3	415.4	0.9974	100.3%
2nd GPT point (200 ppb O3)	786.9	603.5	210.1	209.3	1.0039	99.6%
3rd GPT point (100 ppb O3)	786.9	695.8	117.8	116.8	1.0087	99.1%
Average Correction Factor					1.0034	99.7%

Notes:

Changed the inlet filter prior to the install calibration. Adjusted zero and both NOX and NH3 spans.

Calibration Performed By:

Max Farrell



Wood Buffalo Environmental Association

TN Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 2, 2023	Previous Calibration:	N/A
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:41	End Time (MST):	13:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient 0.999977	≥0.995	
1765.6	1821.6	0.9692			
980.9	1017.4	0.9641			
490.4	513.2	0.9555			
			Slope	1.030917	0.90 - 1.10
			Intercept	3.830269	+/-20

TN Calibration Curve





Wood Buffalo Environmental Association

NH₃ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 2, 2023	Previous Calibration:	N/A
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:41	End Time (MST):	13:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
1765.6	1812.6	0.9741			
980.9	1012.0	0.9692			
490.4	510.4	0.9607			
			Slope	1.025836	0.90 - 1.10
			Intercept	3.642287	+/-20

NH₃ Calibration Curve





Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 2, 2023	Previous Calibration:	N/A
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:41	End Time (MST):	13:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
826.5	823.9	1.0031			
413.2	409.5	1.0091			
207.1	205.8	1.0064			
			Slope	0.996853	0.90 - 1.10
			Intercept	-0.762741	+/-20

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 2, 2023	Previous Calibration:	N/A
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:41	End Time (MST):	13:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
799.7	799.2	1.0007		
399.9	399.1	1.0019		
200.4	198.2	1.0112		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	November 2, 2023	Previous Calibration:	N/A
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:41	End Time (MST):	13:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
414.3	415.4	0.9974		
210.1	209.3	1.0039		
117.8	116.8	1.0087		
			0.999981	
			1.003217	
			-0.787714	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 2, 2023


Location: Patricia McInnes



NH₃ Calibration Plot

Date: November 2, 2023

Location: Patricia McInnes






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	November 20, 2023	Last Cal Date:	October 19, 2023
Start time (MST):	11:29	End time (MST):	15:12
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.52	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC282115			
Removed Cal Gas Conc:	50.52	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 701H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001786	1.001046	Backgd or Offset:	2.61	2.59
Calibration intercept:	2.644970	2.304446	Coeff or Slope:	0.834	0.834

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	79.2	800.2	801.3	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	79.3	801.2	802.8	0.998
second point	4960	39.6	400.2	405.4	0.987
third point	4980	19.8	200.1	203.7	0.982
as left zero	5000	0.0	0.0	0.2	----
as left span	4921	79.2	800.2	802.5	0.997
Average Correction Factor					0.989

Baseline Corr As found:	801.40	Previous response	804.28	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filters, Hydrogen and Nitrogen cylinders after as founds. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar/ Jan Castro



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 20, 2023	Previous Calibration:	October 19, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:29	End Time (MST):	15:12
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999959	≥0.995
801.2	802.8	0.9980			
400.2	405.4	0.9871	Slope	1.001046	0.90 - 1.10
200.1	203.7	0.9822			
			Intercept	2.304446	+/-30



SO2 Calibration Plot

Date: November 20, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: November 15, 2023 Last Cal Date: October 10, 2023
 Start time (MST): 10:09 End time (MST): 15:18
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.25 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC504080
 Removed Cal Gas Conc: 5.25 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3805
 ZAG Make/Model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE Analyzer serial #: 1180540018
 Converter make: CDN-101 Converter serial #: 551
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.014626	1.017229	Backgd or Offset: 2.33	2.40
Calibration intercept:	-0.002205	0.057461	Coeff or Slope: 0.899	0.893

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4925	75.5	79.3	80.3	0.987
as found 2nd point	4962	37.7	39.6	40.3	0.982
as found 3rd point	4981	18.9	19.8	20.1	0.987
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4925	75.5	79.3	80.9	0.980
second point	4962	37.7	39.6	39.9	0.993
third point	4981	18.9	19.9	20.5	0.969
as left zero	5000	0.0	0.0	0.6	----
as left span	4925	75.5	79.3	79.5	0.998
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	0.981
Date of last converter efficiency test:	April 22, 2022			92.6% efficiency	

Baseline Corr As found: 80.3 Prev response: 80.42 *% change: -0.2%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 1.013330 AF Intercept: 0.037834
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999992

* = > +/-5% change initiates investigation

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 15, 2023	Previous Calibration:	October 10, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:09	End Time (MST):	15:18
Analyzer make:	CDN-101	Analyzer serial #:	551

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999918	
79.3	80.9	0.9804			≥0.995
39.6	39.9	0.9927	Slope	1.017229	
19.9	20.5	0.9686			0.90 - 1.10
			Intercept	0.057461	+/-3



TRS Calibration Plot

Date: November 15, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	9.09	9.18	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	9.09	9.09	1.000
second point	4960	39.6	4.55	4.61	0.986
third point	4980	19.8	2.27	2.33	0.974
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	9.09	9.13	0.996
Average Correction Factor					0.987
Baseline Corr AF:	9.18	Prev response	9.10	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.94	8.03	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.94	7.94	1.000
second point	4960	39.6	3.97	3.93	1.011
third point	4980	19.8	1.98	1.94	1.024
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.94	7.95	0.999
Average Correction Factor					1.011
Baseline Corr AF:	8.03	Prev response	7.93	*% change	1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999306	0.999689
THC Cal Offset:	0.011259	0.010659
CH ₄ Cal Slope:	1.001244	1.001158
CH ₄ Cal Offset:	-0.019158	-0.026358
NMHC Cal Slope:	0.997363	0.998306
NMHC Cal Offset:	0.030417	0.037416

Notes: Changed inlet filters, Hydrogen cylinder and Nitrogen cylinder after as founds. Adjusted span only.

Calibration Performed By: Aswin Sasi Kumar/ Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 20, 2023	Previous Calibration:	October 19, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:29	End Time (MST):	15:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥0.995
17.03	17.03	1.0001			
8.52	8.54	0.9976	Slope	0.999689	0.90 - 1.10
4.26	4.27	0.9964			
			Intercept	0.010659	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 20, 2023	Previous Calibration:	October 19, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:29	End Time (MST):	15:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999948	≥0.995			
7.94	7.94	1.0002						
3.97	3.93	1.0106				Slope	1.001158	0.90 - 1.10
1.98	1.94	1.0236						
			Intercept	-0.026358	+/-0.5			

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 20, 2023	Previous Calibration:	October 19, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:29	End Time (MST):	15:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999915	≥ 0.995
9.09	9.09	1.0002			
4.55	4.61	0.9863			
2.27	2.33	0.9738			
			Slope	0.998306	0.90 - 1.10
			Intercept	0.037416	+/-0.5



NMHC Calibration Plot

Date: November 20, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Athabasca Valley
 Calibration Date: November 1, 2023
 Start time (MST): 9:07
 Reason: Routine
 Station number: AMS07
 Last Cal Date: October 6, 2023
 End time (MST): 14:21

Calibration Standards

NO Gas Cylinder #: T2Y1KA4
 NOX Cal Gas Conc: 50.92 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 50.92 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701H
 Cal Gas Expiry Date: November 30, 2023
 NO Cal Gas Conc: 49.92 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 49.92 ppm
 NO gas Diff:
 Serial Number: 3805
 Serial Number: 198

Analyzer Information

Analyzer make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb
 Analyzer serial #: 1160120024

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.077	1.083	NO bkgnd or offset:	7.6	7.6
NOX coeff or slope:	0.994	0.993	NOX bkgnd or offset:	7.8	7.8
NO ₂ coeff or slope:	1.000	1.000	Reaction cell Press:	210.8	212.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000077	0.998915
NO _x Cal Offset:	1.278990	1.318809
NO Cal Slope:	1.001202	1.001602
NO Cal Offset:	1.075147	1.035305
NO ₂ Cal Slope:	1.000884	0.997138
NO ₂ Cal Offset:	-0.671765	0.710095



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
as found span	4920	80.2	816.7	800.7	16.0	814.2	796.7	17.4	1.0031	1.0050
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
high point	4920	80.2	816.7	800.7	16.0	816.3	802.1	14.2	1.0005	0.9982
second point	4960	40.1	408.4	400.4	8.0	410.7	403.7	7.1	0.9943	0.9917
third point	4980	20.0	203.7	199.7	4.0	205.3	201.3	4.0	0.9921	0.9920
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as left span	4920	80.2	816.7	399.3	417.4	822.1	402.5	419.6	0.9935	0.9920
Average Correction Factor									0.9957	0.9940

Corrected As found	NO _x = 814.1 ppb	NO = 796.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.5%	
Previous Response	NO _x = 818.1 ppb	NO = 802.7 ppb		*Percent Change	NO = -0.7%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	800.4	399.0	417.4	416.9	1.0013	99.9%
2nd GPT point (200 ppb O ₃)	800.4	596.7	219.7	219.4	1.0015	99.8%
3rd GPT point (100 ppb O ₃)	800.4	699.5	116.9	118.3	0.9885	101.2%
Average Correction Factor					0.9971	100.3%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 6, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:07	End Time (MST):	14:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
816.7	816.3	1.0005			
408.4	410.7	0.9943			
203.7	205.3	0.9921			
			Slope	0.998915	0.90 - 1.10
			Intercept	1.318809	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 6, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:07	End Time (MST):	14:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
800.7	802.1	0.9982		
400.4	403.7	0.9917		
199.7	201.3	0.9920		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 6, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:07	End Time (MST):	14:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	≥0.995	
417.4	416.9	1.0013			
219.7	219.4	1.0015			
116.9	118.3	0.9885			
			Slope	0.997138	0.90 - 1.10
			Intercept	0.710095	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 1, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: November 14, 2023 Last Cal Date: October 3, 2023
 Start time (MST): 10:55 End time (MST): 14:42
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: T700 Serial Number: 3805
 ZAG Make/Model: T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1152220023
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004229	1.003886	Backgd or Offset:	-2.6	-2.7
Calibration intercept:	0.060000	0.320000	Coeff or Slope:	1.535	1.549

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-1.8	----
as found span	5000	1414.8	400.0	396.0	1.010
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.7	----
high point	5000	1415.7	400.0	401.4	0.997
second point	5000	1039.9	200.0	201.5	0.993
third point	5000	856.2	100.0	101.8	0.982
as left zero	5000	0.0	0.0	-0.1	----
as left span	5000	1416.0	400.0	403.6	0.991
Average Correction Factor					0.990

Baseline Corr As found:	397.8	Previous response	401.8	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 14, 2023	Previous Calibration:	October 3, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:55	End Time (MST):	14:42
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.7	----	Correlation Coefficient	0.999970	≥0.995
400.0	401.4	0.9965			
200.0	201.5	0.9926	Slope	1.003886	0.90 - 1.10
100.0	101.8	0.9823			
			Intercept	0.320000	+/- 5



O₃ Calibration Plot

Date: November 14, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Athabasca Valley Station number: AMS 07
 Calibration Date: November 10, 2023 Last Cal Date: October 19, 2023
 Start time (MST): 11:10 End time (MST): 12:35

Analyzer Make: API T640 S/N: 871
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	0.2	-0.1	0.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	729.4	731.6	729.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.08	5.20	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 10, 2023</u>		Last Cal Date: <u>October 19, 2023</u>		
	PM w/o HEPA: <u>5.5</u>		PM w/ HEPA: <u>0.0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	12.9	12.9	10.8	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: <u>3.8</u>		w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:	<u>November 10, 2023</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>November 10, 2023</u>				

Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022
 Date RH/T Sensor Cleaned: December 5, 2022

Notes: Temp, flow and pressure checked. Leak check passed. Changed sample pump as pump power at 100%. Quarterly maintenance carried out, PMT adjusted. Flow confirmed as 5.1 L/min

Calibration by: Ryan Power



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	November 24, 2023	Last Cal Date:	October 13, 2023
Start time (MST):	11:15	End time (MST):	14:48
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date:	December 12, 2026
Cal Gas Cylinder #:	LL66942			
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 700H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998346	1.000228	Backgd or Offset:	4.336	4.331
Calibration intercept:	0.098508	0.080550	Coeff or Slope:	1.093	1.087

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.0	40.1	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.0	40.0	1.000
second point	4967	33.3	20.0	20.2	0.990
third point	4983	16.7	10.0	10.1	0.988
as left zero	5000	0.0	0.0	0.1	----
as left span	4933	66.7	40.0	40.0	1.002
Average Correction Factor					0.992

Baseline Corr As found:	40.06	Prev response:	40.05	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 24, 2023	Previous Calibration:	October 13, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:15	End Time (MST):	14:48
Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999965	≥0.995
40.0	40.0	0.9996			
20.0	20.2	0.9895	Slope	1.000228	0.90 - 1.10
10.0	10.1	0.9882			
			Intercept	0.080550	+/-1.5



CO Calibration Plot

Date: November 24, 2023

Location: Athabasca Valley






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	November 8, 2023	Last Cal Date:	October 11, 2023
Start time (MST):	12:35	End time (MST):	15:17
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC196697			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3060
ZAG Make/Model:	Teledyne API T701		Serial Number:	260

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003114	1.000960	Backgd or Offset:	4.65	4.61
Calibration intercept:	-2.624243	-1.244570	Coeff or Slope:	0.955	0.965

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-3.0	----
as found span	4920	80.3	800.4	788.8	1.015
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-2.6	----
high point	4920	80.3	800.4	799.7	1.001
second point	4960	40.2	400.7	399.1	1.004
third point	4980	20.1	200.4	201.6	0.994
as left zero	5000	0.0	0.0	-2.8	----
as left span	4920	80.3	800.4	799.4	1.001
				Average Correction Factor	1.000

Baseline Corr As found:	791.80	Previous response	800.25	*% change	-1.1%
-------------------------	--------	-------------------	--------	-----------	-------

** = > +/-5% change initiates investigation*

Notes: changed inlet filters after as founds. Adjustment made to span.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 11, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	12:35	End Time (MST):	15:17
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-2.6	----	Correlation Coefficient	0.999978	≥0.995
800.4	799.7	1.0009			
400.7	399.1	1.0040	Slope	1.000960	0.90 - 1.10
200.4	201.6	0.9938			
			Intercept	-1.244570	+/-30



SO2 Calibration Plot

Date: November 8, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	November 8, 2023	Last Cal Date:	October 12, 2023
Start time (MST):	8:49	End time (MST):	12:23
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.97	ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	EY0002276			
Removed Cal Gas Conc:	4.97	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
ZAG Make/Model:	Teledyne API T701		Serial Number:	260

Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744
Converter make:	CDN-101	Converter serial #:	14639
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.016423	0.995282	Backgd or Offset:	0.99	0.98
Calibration intercept:	0.539126	0.578775	Coeff or Slope:	0.752	0.741

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4920	80.5	80.0	80.7	0.996
as found 2nd point	4960	40.2	40.0	40.6	0.994
as found 3rd point	4980	20.1	20.0	20.8	0.979
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4920	80.5	80.0	80.1	0.999
second point	4960	40.2	40.0	40.5	0.987
third point	4980	20.1	20.0	20.6	0.970
as left zero	5000	0.0	0.0	0.6	----
as left span	4920	80.5	80.0	80.1	0.999
SO2 Scrubber Check	4919.7	80.3	803.0	0.2	----
Date of last scrubber change:	March 7, 2022			Ave Corr Factor	0.985
Date of last converter efficiency test:	March 15, 2022			100.7% efficiency	

Baseline Corr As found:	80.3	Prev response:	81.86	*% change:	-1.9%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	1.002282	AF Intercept:	0.558849
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999979		

* = > +/-5% change initiates investigation

Notes: Sampled inlet filter changed after as founds. Scrubber check passed. Adjusted span.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 12, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:49	End Time (MST):	12:23
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999975	≥0.995
80.0	80.1	0.9989			
40.0	40.5	0.9866	Slope	0.995282	0.90 - 1.10
20.0	20.6	0.9699			
			Intercept	0.578775	+/-3



TRS Calibration Plot

Date: November 8, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
Calibration Date: November 7, 2023 Last Cal Date: October 12, 2023
Start time (MST): 11:40 End time (MST): 15:59
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC363447 Cal Gas Expiry Date: February 2, 2024
NOX Cal Gas Conc: 48.80 ppm NO Cal Gas Conc: 48.80 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.80 ppm Removed Gas NO Conc: 48.80 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3060
ZAG make/model: Teledyne API T701H Serial Number: 260

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1426262592
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.101	1.101	NO bkgnd or offset:	10	10.4
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	11.0	11.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	145.7	141.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.993517	0.999072
NO _x Cal Offset:	-0.020000	1.260000
NO Cal Slope:	0.991532	0.997544
NO Cal Offset:	-0.300000	0.920000
NO ₂ Cal Slope:	1.008121	0.999622
NO ₂ Cal Offset:	1.625692	-0.784149



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.9	-0.6	-0.3	----	----
as found span	4918	82.0	800.3	800.3	0.0	783.0	778.2	4.6	1.0221	1.0284
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-1.0	-0.6	-0.4	----	----
high point	4918	82.0	800.3	800.3	0.0	800.4	799.1	1.2	0.9999	1.0015
second point	4959	41.0	400.2	400.2	0.0	400.1	399.1	1.0	1.0001	1.0027
third point	4980	20.5	200.1	200.1	0.0	204.8	203.2	1.6	0.9770	0.9846
as left zero	5000	0.0	0.0	0.0	0.0	-1.2	-0.4	-0.8	----	----
as left span	4918	82.0	800.3	398.8	401.5	806.4	407.6	398.4	0.9925	0.9785
Average Correction Factor									0.9923	0.9963

Corrected As found	NO _x = 783.9 ppb	NO = 778.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.4%	
Previous Response	NO _x = 795.1 ppb	NO = 793.2 ppb		*Percent Change	NO = -1.9%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	807.1	405.6	401.5	400.8	1.0017	99.8%
2nd GPT point (200 ppb O3)	807.1	609.9	197.2	196.0	1.0061	99.4%
3rd GPT point (100 ppb O3)	807.1	708.8	98.3	97.2	1.0113	98.9%
Average Correction Factor					1.0064	99.4%

Notes: sampled inlet filter changed after as founds. Adjusted Span.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 12, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	11:40	End Time (MST):	15:59
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-1.0	----	Correlation Coefficient Slope Intercept	0.999944 0.999072 1.260000 ≥0.995 0.90 - 1.10 +/-20
800.3	800.4	0.9999		
400.2	400.1	1.0001		
200.1	204.8	0.9770		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 12, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	11:40	End Time (MST):	15:59
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.6	----	Correlation Coefficient	≥0.995
800.3	799.1	1.0015		
400.2	399.1	1.0027	Slope	0.90 - 1.10
200.1	203.2	0.9846		
			Intercept	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 12, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	11:40	End Time (MST):	15:59
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.4	----	Correlation Coefficient	≥0.995	
401.5	400.8	1.0017			
197.2	196.0	1.0061			
98.3	97.2	1.0113			
			Slope	0.999622	0.90 - 1.10
			Intercept	-0.784149	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 7, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: November 7, 2023 Last Cal Date: October 11, 2023
 Start time (MST): 9:04 End time (MST): 11:28
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 260

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3872
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002143	0.996657	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	-0.200000	-0.240000	Coeff or Slope:	1.036	1.036

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	1.0	----
as found span	5000	913.0	400.0	400.4	0.999
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.8	----
high point	5000	914.7	400.0	399.1	1.002
second point	5000	786.4	200.0	198.1	1.010
third point	5000	701.3	100.0	98.7	1.013
as left zero	5000	0.0	0.0	0.9	----
as left span	5000	963.3	400.0	399.0	1.003
Average Correction Factor					1.008

Baseline Corr As found:	399.4	Previous response	400.7	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed out inlet filter after as found. No adjustments

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 11, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	9:04	End Time (MST):	11:28
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.8	----	Correlation Coefficient	0.999965	≥0.995
400.0	399.1	1.0023			
200.0	198.1	1.0096	Slope	0.996657	0.90 - 1.10
100.0	98.7	1.0132			
			Intercept	-0.240000	+/- 5



O₃ Calibration Plot

Date: November 7, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
 Calibration Date: November 24, 2023 Last Cal Date: October 11, 2023
 Start time (MST): 14:19 End time (MST): 14:56

Analyzer Make: Teledyne API T640 S/N: 216
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388747
 Temp/RH standard: Alicat FP-25 S/N: 388747

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	0.3	0.3	0.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	736.5	738.9	736.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	5.02	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 24, 2023</u>	Last Cal Date: <u>October 11, 2023</u>			
	PM w/o HEPA: <u>4.98</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.1	11.0	11.1	<input type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>24.5</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>July 25, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 25, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: July 25, 2023
 Date RH/T Sensor Cleaned: July 25, 2023

Notes: No adjustment made

Calibration by: Matthew C



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	November 10, 2023	Last Cal Date:	October 17, 2023
Start time (MST):	8:30	End time (MST):	11:04
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API T701H		Serial Number:	260

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001383	1.003222	Backgd or Offset:	-0.014	-0.014
Calibration intercept:	0.066908	0.146903	Coeff or Slope:	1.006	1.006

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.06	----
as found span	4933	66.7	40.4	40.7	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4934	66.7	40.4	40.6	0.994
second point	4967	33.3	20.2	20.5	0.985
third point	4983	16.7	10.1	10.3	0.980
as left zero	5000	0.0	0.0	0.1	----
as left span	2960	40.0	40.4	40.4	1.000
Average Correction Factor					0.986

Baseline Corr As found:	40.65	Prev response:	40.55	*% change:	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustments made.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 10, 2023	Previous Calibration:	October 17, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:30	End Time (MST):	11:04
Analyzer make:	API T300	Analyzer serial #:	3505

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999982	≥ 0.995
40.4	40.6	0.9945	Slope	1.003222	0.90 - 1.10
20.2	20.5	0.9853	Intercept	0.146903	+/-1.5
10.1	10.3	0.9797			



CO Calibration Plot

Date: November 10, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	November 10, 2023	Last Cal Date:	October 17, 2023
Start time (MST):	11:12	End time (MST):	14:05
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
N2 Gen Make/Model:	NG 5000		Serial Number:	771048318

Analyzer Information

Analyzer make: Teledyne API T360 Analyzer serial #: 289
 Analyzer Range 0 - 2,000 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.986227	0.982170	Backgd or Offset:	0.008	0.008
Calibration intercept:	1.960000	4.560000	Coeff or Slope:	1.019	1.019

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	2.3	----
as found span	2920	80.0	1605.9	1583.1	1.014
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	2.7	----
high point	2920	80.0	1605.9	1584.7	1.013
second point	2960	40.0	802.9	782.6	1.026
third point	2980	20.0	401.5	408.4	0.983
as left zero	3000	0.0	0.0	3.4	----
as left span	2960	40.0	802.9	779.4	1.030
Average Correction Factor					1.007

Baseline Corr As found:	1580.80	Prev response:	1585.71	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as found. No Adjustments made.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	November 10, 2023	Previous Calibration	October 17, 2023
Station Name	Fort Chipewyan	Station Number	AMS08
Start Time (MST)	11:12	End Time (MST)	14:56
Analyzer make	Teledyne API T360	Analyzer serial #	289

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	2.7	----	Correlation Coefficient	0.999842	≥0.995
1605.9	1584.7	1.0134			
802.9	782.6	1.0260	Slope	0.982170	0.90 - 1.10
401.5	408.4	0.9830			
			Intercept	4.560000	+/-20



CO₂ Calibration Plot

Date: November 10, 2023

Location: Fort Chipewyan





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	November 10, 2023	Last Cal Date:	October 2, 2023
Start time (MST):	9:47	End time (MST):	12:50
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.96	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC151285			
Removed Cal Gas Conc:	49.96	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3812
ZAG Make/Model:	API T701		Serial Number:	4888

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range 0 - 1000 ppb
Analyzer serial #: 1118148498

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995922	1.002594	Backgd or Offset:	10.3	10.3
Calibration intercept:	-0.090033	-0.728669	Coeff or Slope:	0.973	0.970

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4919	80.2	801.5	804.0	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4919	80.2	801.5	803.2	0.998
second point	4959	40.1	400.8	400.6	1.000
third point	4980	20.0	199.8	199.1	1.004
as left zero	5000	0.0	0.0	-0.2	----
as left span	4919	80.2	801.5	801.7	1.000
Average Correction Factor					1.001

Baseline Corr As found:	804.30	Previous response	798.13	*% change	0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 10, 2023	Previous Calibration:	October 2, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:47	End Time (MST):	12:50
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	≥0.995
801.5	803.2	0.9979		
400.8	400.6	1.0004	Slope	0.90 - 1.10
199.8	199.1	1.0037		
			Intercept	+/-30



SO2 Calibration Plot

Date: November 10, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Barge Landing Station number: AMS09
 Calibration Date: November 1, 2023 Last Cal Date: October 11, 2023
 Start time (MST): 9:14 End time (MST): 13:11
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.171 ppm Cal Gas Exp Date: August 22, 2026
 Cal Gas Cylinder #: CC511415
 Removed Cal Gas Conc: 5.171 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3812
 ZAG Make/Model: API T701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1331259320
 Converter make: CDN-101 Converter serial #: 519
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997327	1.009543	Backgd or Offset:	3.06	2.91
Calibration intercept:	0.165639	0.059575	Coeff or Slope:	1.252	1.200

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4923	77.4	80.0	80.9	0.989
as found 2nd point	4961	38.7	40.0	40.6	0.986
as found 3rd point	4981	18.8	19.4	20.1	0.967
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.4	80.0	80.8	0.991
second point	4961	38.7	40.0	40.6	0.986
third point	4981	19.3	20.0	20.2	0.988
as left zero	5000	0.0	0.0	0.1	----
as left span	4923	77.4	80.0	80.8	0.991
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:	28-Feb-23			Ave Corr Factor	0.988
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	80.9	Prev response:	79.99	*% change:	1.1%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.008985	AF Intercept:	0.209020
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999966		

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SOx scrubber check done after calibrator zero.
 Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 11, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:14	End Time (MST):	13:11
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999993	
80.0	80.8	0.9906			≥0.995
40.0	40.6	0.9859	Slope	1.009543	
20.0	20.2	0.9881			0.90 - 1.10
			Intercept	0.059575	+/-3



TRS Calibration Plot

Date: November 1, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	November 10, 2023	Last Cal Date:	October 2, 2023
Start time (MST):	9:47	End time (MST):	12:50
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.58E-04	2.63E-04	NMHC SP Ratio:	4.85E-05	4.93E-05
CH ₄ Retention time:	15.80	15.80	NMHC Peak Area:	188320	185351
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	16.83	1.017
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.13	1.000
second point	4960	40.1	8.56	8.57	0.999
third point	4980	20.0	4.27	4.29	0.994
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.17	0.997
				CF Correction Factor	0.998

Baseline Corr AF:	16.83	Prev response	17.10	*% change	-1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	9.14	8.98	1.017
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.13	1.000
second point	4960	40.1	4.57	4.57	0.999
third point	4980	20.0	2.28	2.30	0.992
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	9.14	9.15	0.998
Average Correction Factor					0.997
Baseline Corr AF:	8.98	Prev response	9.13	*% change	-1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	7.85	1.017
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	7.99	0.999
second point	4960	40.1	3.99	3.99	0.999
third point	4980	20.0	1.99	2.00	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	8.02	0.995
Average Correction Factor					0.998
Baseline Corr AF:	7.85	Prev response	7.97	*% change	-1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997687	0.999996
THC Cal Offset:	0.018043	0.010258
CH ₄ Cal Slope:	0.997459	1.000951
CH ₄ Cal Offset:	0.004658	0.001466
NMHC Cal Slope:	0.998099	0.998936
NMHC Cal Offset:	0.012785	0.009192

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 10, 2023	Previous Calibration:	October 2, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:47	End Time (MST):	12:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥0.995
17.12	17.13	0.9996			
8.56	8.57	0.9991	Slope	0.999996	0.90 - 1.10
4.27	4.29	0.9941			
			Intercept	0.010258	+/-0.5

THC Calibration Curve





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 10, 2023	Previous Calibration:	October 2, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:47	End Time (MST):	12:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995
7.98	7.99	0.9988			
3.99	3.99	0.9992			
1.99	2.00	0.9967			
			Slope	1.000951	0.90 - 1.10
			Intercept	0.001466	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 10, 2023	Previous Calibration:	October 2, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:47	End Time (MST):	12:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995
9.14	9.13	1.0004			
4.57	4.57	0.9990	Slope	0.998936	0.90 - 1.10
2.28	2.30	0.9918			
			Intercept	0.009192	± 0.5



NMHC Calibration Plot

Date: November 10, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	November 15, 2023	Last Cal Date:	November 10, 2023
Start time (MST):	10:12	End time (MST):	13:42
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.63E-04	2.67E-04	NMHC SP Ratio:	4.93E-05	4.84E-05
CH ₄ Retention time:	15.80	16.20	NMHC Peak Area:	185351	188752
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	17.07	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.01	1.006
second point	4960	40.1	8.56	8.51	1.006
third point	4980	20.0	4.27	4.25	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.02	1.006
				AF Correction Factor	1.006

Baseline Corr AF:	17.07	Prev response	17.13	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	9.14	9.27	0.986
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.09	1.005
second point	4960	40.1	4.57	4.55	1.004
third point	4980	20.0	2.28	2.27	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	9.14	9.11	1.003
Average Correction Factor					1.004
Baseline Corr AF:	9.27	Prev response	9.14	*% change	1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	7.80	1.023
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	7.92	1.007
second point	4960	40.1	3.99	3.96	1.009
third point	4980	20.0	1.99	1.98	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	7.91	1.009
Average Correction Factor					1.007
Baseline Corr AF:	7.80	Prev response	7.99	*% change	-2.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999996	0.993496
THC Cal Offset:	0.010258	0.004421
CH ₄ Cal Slope:	1.000951	0.992406
CH ₄ Cal Offset:	0.001466	0.000046
NMHC Cal Slope:	0.998936	0.994848
NMHC Cal Offset:	0.009192	0.002776

Notes: Baseline dipping seen during daily checks. Calibrated zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 15, 2023	Previous Calibration:	November 10, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:12	End Time (MST):	13:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>		
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995		
17.12	17.01	1.0064		Slope	0.993496	0.90 - 1.10	
8.56	8.51	1.0057			Intercept	0.004421	+/-0.5
4.27	4.25	1.0046					

THC Calibration Curve





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 15, 2023	Previous Calibration:	November 10, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:12	End Time (MST):	13:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥0.995
7.98	7.92	1.0074			
3.99	3.96	1.0090			
1.99	1.98	1.0058			
			Slope	0.992406	0.90 - 1.10
			Intercept	0.000046	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 15, 2023	Previous Calibration:	November 10, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:12	End Time (MST):	13:42
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995
9.14	9.09	1.0051			
4.57	4.55	1.0036			
2.28	2.27	1.0036			
			Slope	0.994848	0.90 - 1.10
			Intercept	0.002776	+/-0.5



NMHC Calibration Plot

Date: November 15, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	November 17, 2023	Last Cal Date:	November 15, 2023
Start time (MST):	10:45	End time (MST):	N/A
Reason:	Removal		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.67E-04	N/A	NMHC SP Ratio:	4.84E-05
CH4 Retention time:	16.20	N/A	NMHC Peak Area:	188752
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	13.71	1.249
as found 2nd point	4960	40.1	8.56	7.96	1.075
as found 3rd point	4980	20.0	4.27	4.00	1.068
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

			AF Correction Factor	
Baseline Corr AF:	13.71	Prev response	17.01	*% change -24.1%
Baseline Corr 2nd AF:	8.0	AF Slope:	0.796633	AF Intercept: 0.453990
Baseline Corr 3rd AF:	4.0	AF Correlation:	0.991698	* = +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	9.14	9.03	1.011
as found 2nd point	4960	40.1	4.57	4.50	1.014
as found 3rd point	4980	20.0	2.28	2.26	1.010
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	9.03	Prev response	9.09	*% change	-0.6%
Baseline Corr 2nd AF:	4.5	AF Slope:	0.988482	AF Intercept:	-0.001039
Baseline Corr 3rd AF:	2.3	AF Correlation:	0.999997	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	4.67	1.708
as found 2nd point	4960	40.1	3.99	3.46	1.153
as found 3rd point	4980	20.0	1.99	1.74	1.143
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	4.67	Prev response	7.92	*% change	-69.5%
Baseline Corr 2nd AF:	3.46	AF Slope:	0.576666	AF Intercept:	0.455628
Baseline Corr 3rd AF:	1.74	AF Correlation:	0.930188	* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.993496	
THC Cal Offset:	0.004421	
CH ₄ Cal Slope:	0.992406	
CH ₄ Cal Offset:	0.000046	
NMHC Cal Slope:	0.994848	
NMHC Cal Offset:	0.002776	


Notes: Baseline dipping seen during daily checks. Removing instrument for further maintenance at the shop.

Calibration Performed By: Aswin Sasi Kumar

NMHC Calibration Plot

Date: November 17, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	November 17, 2023	Last Cal Date:	N/A
Start time (MST):	13:00	End time (MST):	15:45
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	N/A	2.53E-04	NMHC SP Ratio:	N/A	4.34E-05
CH ₄ Retention time:	N/A	15.00	NMHC Peak Area:	N/A	210777
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.17	0.997
second point	4960	40.1	8.56	8.58	0.998
third point	4980	20.0	4.27	4.28	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.19	0.996
				CF Correction Factor	0.997

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.18	0.995
second point	4960	40.1	4.57	4.59	0.995
third point	4980	20.0	2.28	2.29	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	9.14	9.20	0.993
Average Correction Factor					0.995
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	7.99	0.999
second point	4960	40.1	3.99	3.99	1.001
third point	4980	20.0	1.99	1.99	0.998
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	7.99	0.999
Average Correction Factor					0.999
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	N/A	1.003147
THC Cal Offset:	N/A	-0.002334
CH ₄ Cal Slope:	N/A	1.001123
CH ₄ Cal Offset:	N/A	-0.001135
NMHC Cal Slope:	N/A	1.005040
NMHC Cal Offset:	N/A	-0.001199

Notes:

Install calibration. Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 17, 2023	Previous Calibration:	N/A
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	13:00	End Time (MST):	15:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995
17.12	17.17	0.9968			
8.56	8.58	0.9979			
4.27	4.28	0.9971			
			Slope	1.003147	0.90 - 1.10
			Intercept	-0.002334	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 17, 2023	Previous Calibration:	N/A
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	13:00	End Time (MST):	15:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥0.995
7.98	7.99	0.9987			
3.99	3.99	1.0007			
1.99	1.99	0.9982			
			Slope	1.001123	0.90 - 1.10
			Intercept	-0.001135	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 17, 2023	Previous Calibration:	N/A
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	13:00	End Time (MST):	15:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995
9.14	9.18	0.9951			
4.57	4.59	0.9953			
2.28	2.29	0.9961			
			Slope	1.005040	0.90 - 1.10
			Intercept	-0.001199	+/-0.5



NMHC Calibration Plot

Date: November 17, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Barge Landing
Calibration Date: November 22, 2023
Start time (MST): 10:13
Reason: Routine
Station number: AMS09
Last Cal Date: October 13, 2023
End time (MST): 14:46

Calibration Standards

NO Gas Cylinder #: DT0036634
NOX Cal Gas Conc: 50.00 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 50.00 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: January 28, 2024
NO Cal Gas Conc: 49.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 49.70 ppm
NO gas Diff:
Serial Number: 3812
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1426262593

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.162	1.128	NO bkgnd or offset:	10.6	10.3
NOX coeff or slope:	0.994	0.992	NOX bkgnd or offset:	10.9	10.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	177.7	178.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001860	0.998965
NO _x Cal Offset:	0.949799	0.368958
NO Cal Slope:	1.002626	0.999327
NO Cal Offset:	-0.151784	-0.472648
NO ₂ Cal Slope:	1.002309	1.001499
NO ₂ Cal Offset:	0.845161	0.067768



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.2	-0.3	----	----
as found span	4919	80.5	805.1	800.3	4.8	834.0	823.9	10.1	0.965	0.971
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	0.0	-0.3	----	----
high point	4919	80.5	805.1	800.3	4.8	804.2	799.4	4.8	1.001	1.001
second point	4959	40.2	402.1	399.7	2.4	402.6	398.9	3.8	0.999	1.002
third point	4979	20.1	201.0	199.8	1.2	201.7	198.6	3.1	0.997	1.006
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
as left span	4919	80.5	805.1	452.2	352.9	805.3	451.0	354.3	1.000	1.003
Average Correction Factor									0.999	1.003

Corrected As found	NO _x = 834.5 ppb	NO = 824.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 3.2%
Previous Response	NO _x = 807.5 ppb	NO = 802.2 ppb		*Percent Change	NO = 2.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.3	447.2	352.9	353.3	0.999	100.1%
2nd GPT point (200 ppb O3)	795.3	665.9	134.2	135.1	0.994	100.6%
3rd GPT point (100 ppb O3)	795.3	728.6	71.5	71.7	0.998	100.2%
Average Correction Factor					0.997	100.3%

Notes:

Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 22, 2023	Previous Calibration:	October 13, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:13	End Time (MST):	14:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
805.1	804.2	1.0011			
402.1	402.6	0.9987			
201.0	201.7	0.9967			
			Slope	0.998965	0.90 - 1.10
			Intercept	0.368958	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 22, 2023	Previous Calibration:	October 13, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:13	End Time (MST):	14:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
800.3	799.4	1.0011			
399.7	398.9	1.0019			
199.8	198.6	1.0062			
			Slope	0.999327	0.90 - 1.10
			Intercept	-0.472648	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 22, 2023	Previous Calibration:	October 13, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:13	End Time (MST):	14:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999992	≥0.995
352.9	353.3	0.9990			
134.2	135.1	0.9936			
71.5	71.7	0.9976			
			Slope	1.001499	0.90 - 1.10
			Intercept	0.067768	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 22, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Barge Landing Station number: AMS 09
 Calibration Date: November 22, 2023 Last Cal Date: October 19, 2023
 Start time (MST): 10:39 End time (MST): 11:02

Analyzer Make: API T640 S/N: 844
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388746
 Temp/RH standard: Alicat FP-25 S/N: 388746

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-13.5	-14.0	-13.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	738.7	745.1	738.7	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	4.94	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 22, 2023</u>	Last Cal Date: <u>October 19, 2023</u>	PM w/o HEPA: <u>4.4</u>	PM w/ HEPA: <u>0.0</u>	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		<0.2 ug/m3
Date Optical Chamber Cleaned:		<u>October 19, 2023</u>			
Disposable Filter Changed:		<u>October 19, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: August 23, 2023
 Date RH/T Sensor Cleaned: August 23, 2023

Notes:

Inlet head looks good. No adjustments made. Leak check passed.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION


CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS11
LOWER CAMP**

NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	November 22, 2023	Last Cal Date:	October 12, 2023
Start time (MST):	11:51	End time (MST):	15:31
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.25	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC2216			
Removed Cal Gas Conc:	49.25	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3807
ZAG Make/Model:	Teledyne API T701		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	100841398
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003950	0.997987	Backgd or Offset:	14.6	14.5
Calibration intercept:	0.611030	0.750571	Coeff or Slope:	1.034	1.034

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.7	----
as found span	4919	81.3	800.8	801.0	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	1.0	----
high point	4919	81.3	800.8	800.1	1.001
second point	4959	40.7	400.9	400.5	1.001
third point	4980	20.3	199.9	200.2	0.999
as left zero	5000	0.0	0.0	0.8	----
as left span	4919	81.3	800.8	802.4	0.998
Average Correction Factor					1.000

Baseline Corr As found:	800.30	Previous response	804.53	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 22, 2023	Previous Calibration:	October 12, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	11:51	End Time (MST):	15:31
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	1.0	----	Correlation Coefficient	≥0.995
800.8	800.1	1.0008		
400.9	400.5	1.0010	Slope	0.90 - 1.10
199.9	200.2	0.9987		
			Intercept	+/-30



SO2 Calibration Plot

Date: November 22, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Lower Camp Station number: AMS11
 Calibration Date: November 30, 2023 Last Cal Date: October 11, 2023
 Start time (MST): 11:13 End time (MST): 15:13
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.429 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC501097
 Removed Cal Gas Conc: 5.429 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3807
 ZAG Make/Model: API T701H Serial Number: 196

Analyzer Information

Analyzer make: Thermo 450iQ Analyzer serial #: CM20080003
 Converter make: NA Converter serial #: NA
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.022511	1.012622	Backgd or Offset: 14.1	14.2
Calibration intercept:	-0.206756	0.214550	Coeff or Slope: 1.001	1.001

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4926	73.6	79.9	81.3	0.987
as found 2nd point	4963	36.8	40.0	40.6	0.992
as found 3rd point	4982	18.6	20.2	20.7	0.990
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.5	----
high point	4926	73.6	79.9	81.3	0.983
second point	4963	36.8	40.0	40.5	0.987
third point	4982	18.6	20.2	20.4	0.990
as left zero	5000	0.0	0.0	0.5	----
as left span	4926	73.6	79.9	81.0	0.987
SO2 Scrubber Check	4919	81.1	811.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.987
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 81.0 Prev response: 81.51 *% change: -0.6%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 1.013497 AF Intercept: 0.233916
 Baseline Corr 3rd AF pt: 20.4 AF Correlation: 0.999993

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Completed SO2 scrubber check after calibrator zero.
 No adjustments required.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 30, 2023	Previous Calibration:	October 11, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	11:13	End Time (MST):	15:13
Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999942	≥0.995
79.9	81.3	0.9830			
40.0	40.5	0.9866	Slope	1.012622	0.90 - 1.10
20.2	20.4	0.9899			
			Intercept	0.214550	+/-3



H₂S Calibration Plot

Date: November 30, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	November 22, 2023	Last Cal Date:	October 12, 2023
Start time (MST):	11:51	End time (MST):	15:31
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.5 ppm		
Diff between cyl (CH ₄):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701	Serial Number:	3807
		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.99E-04	2.99E-04	NMHC SP Ratio:	5.79E-05
CH ₄ Retention time:	14.0	14.0	NMHC Peak Area:	158468
				158468

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	17.35	17.38	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	17.35	17.38	0.998
second point	4959	40.7	8.69	8.67	1.002
third point	4980	20.3	4.33	4.33	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	17.35	17.47	0.993

Average Correction Factor				1.001
Baseline Corr AF:	17.38	Prev response	17.21	*% change 1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	9.19	9.22	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	9.19	9.23	0.996
second point	4959	40.7	4.60	4.60	0.999
third point	4980	20.3	2.29	2.30	0.998
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	9.19	9.27	0.991
Average Correction Factor					0.998
Baseline Corr AF:	9.22	Prev response	9.09	*% change	1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	8.16	8.16	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	8.16	8.15	1.001
second point	4959	40.7	4.09	4.07	1.004
third point	4980	20.3	2.04	2.03	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	8.16	8.19	0.996
Average Correction Factor					1.003
Baseline Corr AF:	8.16	Prev response	8.12	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.994388	1.001857
THC Cal Offset:	-0.039793	-0.011999
CH ₄ Cal Slope:	0.996052	0.998980
CH ₄ Cal Offset:	-0.009288	-0.005496
NMHC Cal Slope:	0.992885	1.004102
NMHC Cal Offset:	-0.030904	-0.005505

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 22, 2023	Previous Calibration:	October 12, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	11:51	End Time (MST):	15:31
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
17.35	17.38	0.9982						
8.69	8.67	1.0017				Slope	1.001857	0.90 - 1.10
4.33	4.33	1.0017						
			Intercept	-0.011999	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-01-2020


Station Information

Calibration Date:	November 22, 2023	Previous Calibration:	October 12, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	11:51	End Time (MST):	15:31
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥0.995
8.16	8.15	1.0012			
4.09	4.07	1.0038			
2.04	2.03	1.0054			
			Slope	0.998980	0.90 - 1.10
			Intercept	-0.005496	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary

Version-01-2020


Station Information

Calibration Date:	November 22, 2023	Previous Calibration:	October 12, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	11:51	End Time (MST):	15:31
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995
9.19	9.23	0.9959			
4.60	4.60	0.9994			
2.29	2.30	0.9984			
			Slope	1.004102	0.90 - 1.10
			Intercept	-0.005505	+/-0.5


NMHC Calibration Curve



NMHC Calibration Plot

Date: November 22, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	November 30, 2023	Last Cal Date:	November 22, 2023
Start time (MST):	9:49	End time (MST):	11:16
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
ZAG make/model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.99E-04	2.99E-04	NMHC SP Ratio:	5.79E-05
CH ₄ Retention time:	14.0	14.0	NMHC Peak Area:	158468

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	17.35	17.32	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	17.35	17.33	1.001
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	1.001
Baseline Corr AF:	17.32	Prev response	17.37	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	9.19	9.17	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	9.19	9.16	1.003
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.003
Baseline Corr AF:	9.17	Prev response	9.22	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	8.16	8.15	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	8.16	8.16	1.000
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.000
Baseline Corr AF:	8.15	Prev response	8.15	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001857	0.998535
THC Cal Offset:	-0.011999	0.000000
CH ₄ Cal Slope:	0.998980	0.999874
CH ₄ Cal Offset:	-0.005496	0.000000
NMHC Cal Slope:	1.004102	0.997020
NMHC Cal Offset:	-0.005505	0.000000

Notes:

Swapped Nitrogen cylinder.


Calibration Performed By:

Mohammed Kashif

NMHC Calibration Plot

Date: November 30, 2023

Location: Lower Camp





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13 FORT MCKAY SOUTH NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	November 6, 2023	Last Cal Date:	October 18, 2023
Start time (MST):	10:17	End time (MST):	13:35
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.55	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC260812			
Removed Cal Gas Conc:	50.55	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2448
ZAG Make/Model:	API 701		Serial Number:	1117

Analyzer Information

Analyzer make:	API T100	Analyzer serial #:	599
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004043	1.002128	Backgd or Offset:	91.1	91.1
Calibration intercept:	-3.158122	-3.438386	Coeff or Slope:	0.709	0.709

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.5	----
as found span	4921	79.1	799.7	798.3	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.5	----
high point	4921	79.1	799.7	799.4	1.000
second point	4961	39.5	399.3	395.3	1.010
third point	4980	19.8	200.2	194.2	1.031
as left zero	5000	0.0	0.0	-0.5	----
as left span	4921	79.1	799.7	798.2	1.002
Average Correction Factor					1.014

Baseline Corr As found:	798.80	Previous response	799.76	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustment.

Calibration Performed By: Jan Castro and Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 18, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:17	End Time (MST):	13:35
Analyzer make:	API T100	Analyzer serial #:	599

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.5	----	Correlation Coefficient	0.999939	≥0.995
799.7	799.4	1.0004			
399.3	395.3	1.0101	Slope	1.002128	0.90 - 1.10
200.2	194.2	1.0308			
			Intercept	-3.438386	+/-30



SO2 Calibration Plot

Date: November 6, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	November 23, 2023	Last Cal Date:	October 12, 2023
Start time (MST):	9:51	End time (MST):	13:55
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.34	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC500241			
Removed Cal Gas Conc:	5.34	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2448
ZAG Make/Model:	Teledyne API 701		Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017
Converter make:	CDN-101	Converter serial #:	521
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006178	1.003766	Backgd or Offset:	3.77	3.77
Calibration intercept:	-0.262271	-0.102198	Coeff or Slope:	1.130	1.130

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4925	75.5	80.6	81.3	0.992
as found 2nd point	4962	37.7	40.3	40.4	0.997
as found 3rd point	4981	18.9	20.2	19.9	1.014
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4925	75.5	80.6	80.9	0.997
second point	4962	37.7	40.3	40.3	0.999
third point	4981	18.9	20.2	19.9	1.014
as left zero	5000	0.0	0.0	0.2	----
as left span	4925	75.5	80.6	79.5	1.014
SO2 Scrubber Check	4921	79.1	791.0	-0.1	----
Date of last scrubber change:	20-Mar-20		Ave Corr Factor		1.003
Date of last converter efficiency test:	NA		efficiency		

Baseline Corr As found:	81.3	Prev response:	80.86	*% change:	0.5%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.010004	AF Intercept:	-0.222225
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999964		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 23, 2023	Previous Calibration:	October 12, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	9:51	End Time (MST):	13:55
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
80.6	80.9	0.9966		
40.3	40.3	0.9992	Slope	0.90 - 1.10
20.2	19.9	1.0144		
			Intercept	+/-3



TRS Calibration Plot

Date: November 23, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	November 6, 2023	Last Cal Date:	October 18, 2023
Start time (MST):	10:17	End time (MST):	13:35
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH4 Cal Gas Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
C3H8 Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
Removed C3H8 Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API701	Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.31E-04	2.36E-04	NMHC SP Ratio:	5.16E-05
CH4 Retention time:	13.40	13.60	NMHC Peak Area:	1756954
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				5.05E-05
				179781
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	17.05	17.20	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.09	0.998
second point	4961	39.5	8.51	8.44	1.008
third point	4980	19.8	4.27	4.14	1.030
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	17.05	17.15	0.994
Average Correction Factor					1.012

Baseline Corr AF:	17.20	Prev response	17.01	*% change	1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	9.08	9.36	0.970
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	9.08	9.10	0.997
second point	4961	39.5	4.53	4.51	1.005
third point	4980	19.8	2.27	2.21	1.027
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	9.08	9.14	0.993
Average Correction Factor					1.010
Baseline Corr AF:	9.36	Prev response	9.06	*% change	3.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	7.97	7.84	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.99	0.998
second point	4961	39.5	3.98	3.93	1.011
third point	4980	19.8	1.99	1.93	1.034
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	7.97	8.01	0.994
Average Correction Factor					1.014
Baseline Corr AF:	7.84	Prev response	7.95	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002214	1.004439
THC Cal Offset:	-0.072976	-0.071566
CH ₄ Cal Slope:	1.003429	1.004647
CH ₄ Cal Offset:	-0.043993	-0.038988
NMHC Cal Slope:	1.001224	1.004722
NMHC Cal Offset:	-0.029785	-0.033179

Notes: Changed inlet filter after as founds. Adjusted span.

Calibration Performed By: Jan Castro and Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 18, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:17	End Time (MST):	13:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999920	≥ 0.995			
17.05	17.09	0.9977						
8.51	8.44	1.0082				Slope	1.004439	0.90 - 1.10
4.27	4.14	1.0300						
			Intercept	-0.071566	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 18, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:17	End Time (MST):	13:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999890	≥0.995
7.97	7.99	0.9976			
3.98	3.93	1.0115			
1.99	1.93	1.0339			
			Slope	1.004647	0.90 - 1.10
			Intercept	-0.038988	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 18, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:17	End Time (MST):	13:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999939	≥ 0.995
9.08	9.10	0.9973			
4.53	4.51	1.0052			
2.27	2.21	1.0266			
			Slope	1.004722	0.90 - 1.10
			Intercept	-0.033179	+/-0.5



NMHC Calibration Plot

Date: November 6, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	November 21, 2023	Last Cal Date:	November 6, 2023
Start time (MST):	10:38	End time (MST):	13:36
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH4 Cal Gas Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
C3H8 Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
Removed C3H8 Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API701	Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	NA	2.76E-04	NMHC SP Ratio:	NA
CH4 Retention time:	NA	15.00	NMHC Peak Area:	NA
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				4.61E-05
				197092
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.15	0.994
second point	4961	39.5	8.51	8.48	1.004
third point	4980	19.8	4.27	4.16	1.026
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	17.05	16.95	1.006
Average Correction Factor					1.008

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* => +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	9.08	9.06	1.002
second point	4961	39.5	4.53	4.50	1.007
third point	4980	19.8	2.27	2.20	1.035
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	9.08	8.87	1.023
Average Correction Factor					1.015
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	8.09	0.985
second point	4961	39.5	3.98	3.98	1.000
third point	4980	19.8	1.99	1.96	1.016
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	7.97	8.08	0.986
Average Correction Factor					1.000
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	NA	1.008301
THC Cal Offset:	NA	-0.071364
CH ₄ Cal Slope:	NA	1.016730
CH ₄ Cal Offset:	NA	-0.036593
NMHC Cal Slope:	NA	1.000565
NMHC Cal Offset:	NA	-0.034171

Notes: Install calibration. Changed inlet filter. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 21, 2023	Previous Calibration:	November 6, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:38	End Time (MST):	13:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999921	≥ 0.995
17.05	17.15	0.9939			
8.51	8.48	1.0040			
4.27	4.16	1.0263			
			Slope	1.008301	0.90 - 1.10
			Intercept	-0.071364	+/-0.5

THC Calibration Curve





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 21, 2023	Previous Calibration:	November 6, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:38	End Time (MST):	13:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999900	≥ 0.995
7.97	8.09	0.9853			
3.98	3.98	1.0003			
1.99	1.96	1.0160			
			Slope	1.016730	0.90 - 1.10
			Intercept	-0.036593	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 21, 2023	Previous Calibration:	November 6, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:38	End Time (MST):	13:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999926	≥ 0.995
9.08	9.06	1.0019			
4.53	4.50	1.0072			
2.27	2.20	1.0355			
			Slope	1.000565	0.90 - 1.10
			Intercept	-0.034171	+/-0.5



NMHC Calibration Plot

Date: November 21, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort McKay South Station number: AMS 13
Calibration Date: November 24, 2023 Last Cal Date: October 24, 2023
Start time (MST): 8:45 End time (MST): 13:04
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1P76 Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.98 ppm NO Cal Gas Conc: 49.32 ppm
Removed Cylinder #: N/A Removed Gas Exp Date: N/A
Removed Gas NOX Conc: 50.98 ppm Removed Gas NO Conc: 49.32 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: API T700 Serial Number: 2448
ZAG make/model: API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1410661329
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.124	1.129	NO bkgnd or offset:	10.4	10.4
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	10.4	10.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.8	160.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999331	0.999580
NO _x Cal Offset:	-2.071318	-1.911243
NO Cal Slope:	1.002262	1.001676
NO Cal Offset:	-2.945200	-2.864919
NO ₂ Cal Slope:	1.001663	1.000024
NO ₂ Cal Offset:	-0.500859	-0.523265



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	-0.1	----	----
as found span	4919	81.1	826.9	800.0	26.9	824.5	796.5	28.1	1.0029	1.0043
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.0	----	----
high point	4919	81.1	826.9	800.0	26.9	825.7	800.0	25.7	1.0014	0.9999
second point	4960	40.6	413.9	400.4	13.5	410.7	396.6	14.1	1.0078	1.0097
third point	4980	20.3	207.0	200.2	6.7	202.9	194.7	8.2	1.0200	1.0284
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4919	81.1	826.9	383.0	443.9	830.9	383.0	448.0	0.9952	0.9999
Average Correction Factor									1.0098	1.0127

Corrected As found	NO _x = 824.7 ppb	NO = 796.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.1%	
Previous Response	NO _x = 824.3 ppb	NO = 798.8 ppb		*Percent Change	NO = -0.3%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	797.6	380.6	443.9	443.5	1.0010	99.9%
2nd GPT point (200 ppb O ₃)	797.6	589.5	235.0	234.9	1.0005	99.9%
3rd GPT point (100 ppb O ₃)	797.6	693.6	130.9	129.4	1.0118	98.8%
Average Correction Factor					1.0044	99.6%

Notes: Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 24, 2023	Previous Calibration:	October 24, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:45	End Time (MST):	13:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
826.9	825.7	1.0014			
413.9	410.7	1.0078			
207.0	202.9	1.0200			
			Slope	0.999580	0.90 - 1.10
			Intercept	-1.911243	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 24, 2023	Previous Calibration:	October 24, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:45	End Time (MST):	13:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999933	≥0.995
800.0	800.0	0.9999			
400.4	396.6	1.0097	Slope	1.001676	0.90 - 1.10
200.2	194.7	1.0284			
			Intercept	-2.864919	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 24, 2023	Previous Calibration:	October 24, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:45	End Time (MST):	13:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
443.9	443.5	1.0010		
235.0	234.9	1.0005		
130.9	129.4	1.0118		
			0.999986	
			1.000024	
			-0.523265	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 24, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: November 20, 2023 Last Cal Date: October 6, 2023
 Start time (MST): 10:08 End time (MST): 13:23
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3871
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998314	0.997971	Backgd or Offset:	2.4	2.4
Calibration intercept:	0.820000	0.780000	Coeff or Slope:	0.967	0.967

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	5000	977.0	400.0	400.6	0.999
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	5000	977.0	400.0	399.5	1.001
second point	5000	838.0	200.0	200.9	0.996
third point	5000	735.9	100.0	101.4	0.986
as left zero	5000	0.0	0.0	0.3	----
as left span	5000	977.0	400.0	402.9	0.993
Average Correction Factor					0.994

Baseline Corr As found:	400.7	Previous response	400.1	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 20, 2023	Previous Calibration:	October 6, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:08	End Time (MST):	13:23
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	≥0.995
400.0	399.5	1.0013		
200.0	200.9	0.9955	Slope	0.90 - 1.10
100.0	101.4	0.9862		
			Intercept	+/- 5



O₃ Calibration Plot

Date: November 20, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort McKay South Station number: AMS 13
 Calibration Date: November 24, 2023 Last Cal Date: October 24, 2023
 Start time (MST): 10:11 End time (MST): 10:37

Analyzer Make: API T640 S/N: 1335
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388746
 Temp/RH standard: Alicat FP-25 S/N: 388746

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2.00	2.07	2.00	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	735.50	737.54	735.50	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	4.91	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 24, 2023</u>	Last Cal Date: <u>October 24, 2023</u>			
	PM w/o HEPA: <u>3.6</u>	PM w/ HEPA: <u>0.0</u>		<0.2 ug/m3	

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>October 24, 2023</u>		<0.2 ug/m3	
Disposable Filter Changed:		<u>October 24, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: June 29, 2023
 Date RH/T Sensor Cleaned: June 29, 2023

Leak check passed.

Notes:

Calibration by: Sean Bala




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC
NOVEMBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	November 7, 2023	Last Cal Date:	October 3, 2023
Start time (MST):	10:54	End time (MST):	14:23
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.95	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC279389			
Removed Cal Gas Conc:	49.95	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API T701H		Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006812	0.994980	Backgd or Offset:	25.2	25.3
Calibration intercept:	-2.184734	-1.917407	Coeff or Slope:	0.798	0.798

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.7	----
as found span	4938	80.3	799.3	793.4	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	1.1	----
high point	4938	80.3	799.3	795.1	1.005
second point	4979	40.2	400.1	393.9	1.016
third point	4998	20.2	201.1	195.6	1.028
as left zero	5000	0.0	0.0	1.0	----
as left span	4938	80.3	799.3	794.9	1.005
Average Correction Factor					1.016

Baseline Corr As found:	792.70	Previous response	802.53	*% change	-1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 3, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:54	End Time (MST):	14:23
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	1.1	----	Correlation Coefficient	≥0.995
799.3	795.1	1.0052		
400.1	393.9	1.0156	Slope	0.90 - 1.10
201.1	195.6	1.0279		
			Intercept	+/-30



SO2 Calibration Plot

Date: November 7, 2023

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	November 1, 2023	Last Cal Date:	October 5, 2023
Start time (MST):	9:56	End time (MST):	14:07
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.15	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC510379			
Removed Cal Gas Conc:	5.15	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API 701H		Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582
Converter make:	CD Nova CDN-101	Converter serial #:	503
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997157	1.008310	Backgd or Offset:	2.28
Calibration intercept:	-0.105310	-0.045461	Coeff or Slope:	0.992
				2.31
				0.992

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4938	77.9	80.0	81.6	0.978
as found 2nd point	4973	38.9	40.0	40.6	0.980
as found 3rd point	4997	19.5	20.0	20.1	0.986
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4938	77.9	80.0	80.6	0.992
second point	4973	38.9	40.0	40.2	0.994
third point	4997	19.5	20.0	20.1	0.996
as left zero	5000	0.0	0.0	0.1	----
as left span	4938	77.9	80.0	79.7	1.003
SO2 Scrubber Check	4936	80.3	800.4	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.994
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	81.8	Prev response:	79.65	*% change:	2.6%
Baseline Corr 2nd AF pt:	40.8	AF Slope:	1.023352	AF Intercept:	-0.285524
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999995		

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 5, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	9:56	End Time (MST):	14:07
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999998	
80.0	80.6	0.9920			≥0.995
40.0	40.2	0.9939	Slope	1.008310	
20.0	20.1	0.9956			0.90 - 1.10
			Intercept	-0.045461	+/-3



TRS Calibration Plot

Date: November 1, 2023

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	November 7, 2023	Last Cal Date:	October 3, 2023
Start time (MST):	10:54	End time (MST):	14:23
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.75E-04	3.72E-04	NMHC SP Ratio:	4.49E-05
CH ₄ Retention time:	12.20	12.20	NMHC Peak Area:	203038
				198214

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	17.02	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.13	0.998
second point	4979	40.2	8.56	8.53	1.004
third point	4998	20.2	4.30	4.24	1.014
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	17.10	17.18	0.996

Average Correction Factor				1.005
Baseline Corr AF:	17.02	Prev response	17.07	*% change -0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	8.96	1.017
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.13	0.998
second point	4979	40.2	4.56	4.56	1.001
third point	4998	20.2	2.29	2.26	1.016
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	9.11	9.18	0.992
Average Correction Factor					1.005
Baseline Corr AF:	8.96	Prev response	9.03	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	8.06	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	8.00	0.999
second point	4979	40.2	4.00	3.97	1.007
third point	4998	20.2	2.01	1.99	1.013
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	7.99	7.99	1.000
Average Correction Factor					1.006
Baseline Corr AF:	8.06	Prev response	8.04	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000528	1.002761
THC Cal Offset:	-0.040856	-0.036088
CH ₄ Cal Slope:	1.008350	1.001781
CH ₄ Cal Offset:	-0.017069	-0.016831
NMHC Cal Slope:	0.994097	1.003445
NMHC Cal Offset:	-0.024989	-0.019055

Notes: Sample inlet filter changed after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 3, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:54	End Time (MST):	14:23
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999980	≥ 0.995			
17.10	17.13	0.9983						
8.56	8.53	1.0036				Slope	1.002761	0.90 - 1.10
4.30	4.24	1.0142						
			Intercept	-0.036088	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 3, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:54	End Time (MST):	14:23
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999977	≥ 0.995			
7.99	8.00	0.9989						
4.00	3.97	1.0065				Slope	1.001781	0.90 - 1.10
2.01	1.99	1.0125						
			Intercept	-0.016831	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 3, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:54	End Time (MST):	14:23
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999977	≥ 0.995			
9.11	9.13	0.9979						
4.56	4.56	1.0008				Slope	1.003445	0.90 - 1.10
2.29	2.26	1.0162						
			Intercept	-0.019055	± 0.5			



NMHC Calibration Plot

Date: November 7, 2023

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	November 23, 2023	Last Cal Date:	November 7, 2023
Start time (MST):	12:17	End time (MST):	14:01
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.72E-04	3.72E-04	NMHC SP Ratio:	4.60E-05
CH ₄ Retention time:	12.20	12.20	NMHC Peak Area:	198214

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	17.09	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.09	1.001
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	1.001
Baseline Corr AF:	17.09	Prev response	17.11	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	9.19	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.20	0.991
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.991
Baseline Corr AF:	9.19	Prev response	9.13	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	7.90	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	7.89	1.013
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.013
Baseline Corr AF:	7.90	Prev response	7.99	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002761	0.999080
THC Cal Offset:	-0.036088	0.000000
CH ₄ Cal Slope:	1.001781	0.987419
CH ₄ Cal Offset:	-0.016831	0.000000
NMHC Cal Slope:	1.003445	1.008974
NMHC Cal Offset:	-0.019055	0.000000

Notes:

Swapped Nitrogen cylinder.


Calibration Performed By:

Mohammed Kashif

NMHC Calibration Plot

Date: November 23, 2023

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Anzac
Calibration Date: November 2, 2023
Start time (MST): 9:58
Reason: Routine
Station number: AMS 14
Last Cal Date: October 4, 2023
End time (MST): 14:29

Calibration Standards

NO Gas Cylinder #: T2Y1P8D
NOX Cal Gas Conc: 50.92 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 50.92 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API 701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.05 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.05 ppm
NO gas Diff:
Serial Number: 3060
Serial Number: 357

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1426262592

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.411	1.411	NO bkgnd or offset:	3.8	3.8
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.8	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	159.4	160.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996975	1.002002
NO _x Cal Offset:	-0.888234	-0.728718
NO Cal Slope:	0.996206	1.002134
NO Cal Offset:	-2.526016	-2.326785
NO ₂ Cal Slope:	1.005379	1.004193
NO ₂ Cal Offset:	1.537643	0.821549



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.2	0.3	----	----
as found span	4936	80.2	814.1	800.2	13.9	819.2	800.1	19.2	0.9938	1.0001
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.0	0.3	----	----
high point	4936	80.2	814.1	800.2	13.9	815.5	800.9	14.7	0.9983	0.9991
second point	4979	40.1	406.8	399.9	7.0	406.5	396.8	9.7	1.0008	1.0077
third point	4999	20.1	203.9	200.4	3.5	202.5	196.5	6.0	1.0070	1.0200
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.0	0.3	----	----
as left span	4936	80.2	814.1	415.8	398.3	816.2	418.7	397.5	0.9975	0.9931
Average Correction Factor									1.0020	1.0090

Corrected As found	NO _x = 819.1 ppb	NO = 800.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.0%
Previous Response	NO _x = 810.8 ppb	NO = 794.6 ppb		*Percent Change	NO = 0.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.3	413.9	398.3	400.5	0.9945	100.5%
2nd GPT point (200 ppb O3)	798.3	610.9	201.3	203.1	0.9912	100.9%
3rd GPT point (100 ppb O3)	798.3	676.2	136.0	138.1	0.9849	101.5%
Average Correction Factor					0.9902	101.0%

Notes:

Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 2, 2023	Previous Calibration:	October 4, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:58	End Time (MST):	14:29
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	≥0.995	
814.1	815.5	0.9983			
406.8	406.5	1.0008			
203.9	202.5	1.0070			
			Slope	1.002002	0.90 - 1.10
			Intercept	-0.728718	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 2, 2023	Previous Calibration:	October 4, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:58	End Time (MST):	14:29
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
800.2	800.9	0.9991			
399.9	396.8	1.0077			
200.4	196.5	1.0200			
			Slope	1.002134	0.90 - 1.10
			Intercept	-2.326785	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 2, 2023	Previous Calibration:	October 4, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:58	End Time (MST):	14:29
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
398.3	400.5	0.9945		
201.3	203.1	0.9912		
136.0	138.1	0.9849		
			0.999990	
			1.004193	
			0.821549	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 2, 2023

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 13, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:48	End Time (MST):	15:53
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220026

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999999	≥0.995
400.0	400.9	0.9978			
200.0	201.0	0.9950	Slope	1.001200	0.90 - 1.10
100.0	100.7	0.9930			
			Intercept	0.540000	+/- 5



O₃ Calibration Plot

Date: November 6, 2023

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	November 21, 2023	Last Cal Date:	NA
Start time (MST):	13:28	End time (MST):	15:43
Reason:	Install		

Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	API T700	Serial Number:	3060
ZAG Make/Model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595
Analyzer Range:	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.002171	Backgd or Offset:	NA	1.4
Calibration intercept:	NA	-0.580000	Coeff or Slope:	NA	1.620

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.3	----
high point	5000	1080.7	400.0	400.7	0.998
second point	5000	822.6	200.0	199.5	1.003
third point	5000	691.7	100.0	98.7	1.013
as left zero	5000	0.0	0.0	0.1	----
as left span	5000	919.5	400.0	403.1	0.992
Average Correction Factor					1.005

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Install calibration. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 21, 2023	Previous Calibration:	NA
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	13:28	End Time (MST):	15:43
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	≥0.995
400.0	400.7	0.9983		
200.0	199.5	1.0025	Slope	0.90 - 1.10
100.0	98.7	1.0132		
			Intercept	+/- 5



O₃ Calibration Plot

Date: November 21, 2023

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	November 21, 2023	Last Cal Date:	November 6, 2023
Start time (MST):	10:58	End time (MST):	12:48
Reason:	Removal		

Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	API T700	Serial Number:	3060
ZAG Make/Model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1152220026
Analyzer Range:	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001200		Backgd or Offset:	-1.6	NA
Calibration intercept:	0.540000		Coeff or Slope:	1.013	NA

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	5000	918.1	400.0	386.1	1.036
as found 2nd point	5000	805.1	200.0	192.9	1.037
as found 3rd point	5000	712.5	100.0	96.2	1.040
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

Average Correction Factor					
Baseline Corr As found:	385.9	Previous response	401.0	*% change	-3.9%
Baseline Corr 2nd AF pt:	192.7	AF Slope:	0.965086	AF Intercept:	-0.040000
Baseline Corr 3rd AF pt:	96.0	AF Correlation:	0.999998		

* = > +/-5% change initiates investigation

Notes:

Removal calibration.


Calibration Performed By:

Mohammed Kashif

O₃ Calibration Plot

Date: November 21, 2023

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Anzac Station number: AMS 14
 Calibration Date: November 7, 2023 Last Cal Date: October 6, 2023
 Start time (MST): 13:23 End time (MST): 13:38

Analyzer Make: API T640 S/N: 825
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388749
 Temp/RH standard: Alicat FP-25 S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2.7	3.2	2.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	705.7	706.5	705.7	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	5.00	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 7, 2023</u>	Last Cal Date: <u>October 6, 2023</u>			
	PM w/o HEPA: <u>3.4</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>October 6, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>October 6, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: July 6, 2023
 Date RH/T Sensor Cleaned: July 6, 2023

Notes: No adjustments required.

Calibration by: Mohammed Kashif




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS17
WAPASU
NOVEMBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 11, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:01	End Time (MST):	13:52
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient	0.999966	
800.0	800.4	0.9995			≥0.995
400.0	396.7	1.0084	Slope	1.000367	
199.5	197.2	1.0117			0.90 - 1.10
			Intercept	-1.259622	+/-30



SO2 Calibration Plot

Date: November 6, 2023

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	November 28, 2023	Last Cal Date:	October 16, 2023
Start time (MST):	10:35	End time (MST):	15:40
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.076	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511852			
Removed Cal Gas Conc:	5.076	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2449
ZAG Make/Model:	API T701H		Serial Number:	359

Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583
Converter make:	n/a	Converter serial #:	n/a
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001854	1.001426	Backgd or Offset:	12.2
Calibration intercept:	0.060776	0.000779	Coeff or Slope:	1.114

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	78.8	80.0	82.3	0.971
as found 2nd point	4961	39.4	40.0	40.9	0.976
as found 3rd point	4980	19.7	20.0	20.5	0.971
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	78.8	80.0	80.3	0.996
second point	4961	39.4	40.0	39.6	1.010
third point	4980	19.7	20.0	20.2	0.990
as left zero	5000	0.0	0.0	0.3	----
as left span	4921	78.8	80.0	79.1	1.011
SO2 Scrubber Check	4921	79.4	800.0	0.2	----
Date of last scrubber change:		n/a		Ave Corr Factor	0.999
Date of last converter efficiency test:		n/a			efficiency

Baseline Corr As found:	82.4	Prev response:	80.21	*% change:	2.7%
Baseline Corr 2nd AF pt:	41.0	AF Slope:	1.029711	AF Intercept:	-0.139183
Baseline Corr 3rd AF pt:	20.6	AF Correlation:	0.999992		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 16, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:35	End Time (MST):	15:40
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
80.0	80.3	0.9963		
40.0	39.6	1.0100	Slope	0.90 - 1.10
20.0	20.2	0.9901		
			Intercept	+/-3



H₂S Calibration Plot

Date: November 28, 2023

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 11, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:01	End Time (MST):	13:52
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.08	----	Correlation Coefficient	0.999940	≥0.995
17.09	17.08	1.0009			
8.55	8.38	1.0201	Slope	1.004169	0.90 - 1.10
4.26	4.15	1.0276			
			Intercept	-0.126321	+/-1.5



THC Calibration Plot

Date: November 6, 2023

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Wapasu
Calibration Date: November 21, 2023
Start time (MST): 11:15
Reason: Routine
Station number: AMS17
Last Cal Date: October 18, 2023
End time (MST): 16:07

Calibration Standards

NO Gas Cylinder #: T375YK8
NOX Cal Gas Conc: 49.11 ppm
Removed Cylinder #: T375YK8
Removed Gas NOX Conc: 49.11 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: April 13, 2025
NO Cal Gas Conc: 48.07 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 48.07 ppm
NO gas Diff:
Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: Thermo Scientific 42iQ
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 12300522720

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.294	1.280	NO bkgnd or offset:	9.2	8.4
NOX coeff or slope:	0.990	0.986	NOX bkgnd or offset:	9.3	8.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	363.7	341.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002375	1.003061
NO _x Cal Offset:	-3.420000	-4.540000
NO Cal Slope:	1.001687	1.006988
NO Cal Offset:	-3.940000	-5.320000
NO ₂ Cal Slope:	1.008551	0.997809
NO ₂ Cal Offset:	-0.167241	0.732274



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.5	0.0	----	----
as found span	4917	83.2	817.2	799.9	17.3	827.7	809.6	18.2	0.9873	0.9880
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
high point	4917	83.2	817.2	799.9	17.3	817.8	803.1	14.8	0.9993	0.9960
second point	4958	41.6	408.6	399.9	8.7	401.6	393.6	8.0	1.0174	1.0161
third point	4979	20.8	204.3	200.0	4.3	197.2	191.9	5.3	1.0360	1.0421
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
as left span	4917	83.2	817.2	395.7	421.5	816.2	396.7	419.6	1.0012	0.9974
Average Correction Factor									1.0176	1.0181

Corrected As found	NO _x = 828.2 ppb	NO = 810.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.5%
Previous Response	NO _x = 815.7 ppb	NO = 797.3 ppb		*Percent Change	NO = 1.6%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	801.1	396.9	421.5	420.7	1.0019	99.8%
2nd GPT point (200 ppb O3)	801.1	602.5	215.9	217.2	0.9940	100.6%
3rd GPT point (100 ppb O3)	801.1	704.7	113.7	114.5	0.9931	100.7%
Average Correction Factor					0.9963	100.4%

Notes: Change filters after as founds. Manifold cleaned. Adjusted span only.

Calibration Performed By: Aswin Sasi Kumar/ Jan Castro



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 21, 2023	Previous Calibration:	October 18, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:15	End Time (MST):	16:07
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999867	≥0.995
817.2	817.8	0.9993			
408.6	401.6	1.0174	Slope	1.003061	0.90 - 1.10
204.3	197.2	1.0360			
			Intercept	-4.540000	+/-20

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 21, 2023	Previous Calibration:	October 18, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:15	End Time (MST):	16:07
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999815	≥0.995
799.9	803.1	0.9960			
399.9	393.6	1.0161	Slope	1.006988	0.90 - 1.10
200.0	191.9	1.0421			
			Intercept	-5.320000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 21, 2023	Previous Calibration:	October 18, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:15	End Time (MST):	16:07
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
421.5	420.7	1.0019		
215.9	217.2	0.9940		
113.7	114.5	0.9931		
			0.999978	
			0.997809	
			0.732274	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 21, 2023

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Wapasu Station number: AMS17
 Calibration Date: November 2, 2023 Last Cal Date: October 4, 2023
 Start time (MST): 10:32 End time (MST): 13:36
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 2449
 ZAG Make/Model: API T701H Serial Number: 359

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 3870
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002371	1.003143	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.440000	-0.400000	Coeff or Slope:	1.014	1.014

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	5000	1077.3	400.0	400.6	0.999
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.0	----
high point	5000	1077.3	400.0	401.1	0.997
second point	5000	900.3	200.0	199.9	1.001
third point	5000	789.5	100.0	99.6	1.004
as left zero	5000	0.0	0.0	0.5	----
as left span	5000	1077.3	400.0	405.7	0.986
Average Correction Factor					1.001

Baseline Corr As found:	400.5	Previous response	400.5	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 2, 2023	Previous Calibration:	October 4, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:32	End Time (MST):	13:36
Analyzer make:	API T400	Analyzer serial #:	3870

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999995	
400.0	401.1	0.9973			≥0.995
200.0	199.9	1.0005	Slope	1.003143	
100.0	99.6	1.0040			0.90 - 1.10
			Intercept	-0.400000	+/- 5



O₃ Calibration Plot

Date: November 2, 2023

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Wapasu Station number: AMS 17
 Calibration Date: November 28, 2023 Last Cal Date: October 18, 2023
 Start time (MST): 14:25 End time (MST): 15:09

Analyzer Make: API T640 S/N: 1183
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	3.0	2.4	3.0	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	706.6	704.2	706.6	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	4.92	4.99	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 28, 2023</u>	Last Cal Date: <u>September 22, 2023</u>			
	PM w/o HEPA: <u>2.7</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>October 18, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 22, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: Temp, pressure and flow checked. Leak check passed.

Calibration by: Aswin Sasi Kumar




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	November 7, 2023	Last Cal Date:	October 17, 2023
Start time (MST):	11:32	End time (MST):	15:10
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.40	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC463851			
Removed Cal Gas Conc:	49.40	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API 701H		Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998676	1.003346	Backgd or Offset:	22.2	22.6
Calibration intercept:	-0.203578	-0.163010	Coeff or Slope:	0.795	0.803

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5009	0.0	0.0	0.2	----
as found span	4919	81.0	800.3	793.0	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.5	----
high point	4919	81.0	800.3	803.0	0.997
second point	4959	40.5	400.2	401.4	0.997
third point	4979	20.2	199.6	199.2	1.002
as left zero	5000	0.0	0.0	0.3	----
as left span	4919	81.0	800.3	799.7	1.001
Average Correction Factor					0.999

Baseline Corr As found:	792.80	Previous response	799.02	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 17, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:32	End Time (MST):	15:10
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	≥0.995
800.3	803.0	0.9966		
400.2	401.4	0.9970	Slope	0.90 - 1.10
199.6	199.2	1.0020		
			Intercept	+/-30



SO2 Calibration Plot

Date: November 7, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: November 29, 2023 Last Cal Date: October 26, 2023
 Start time (MST): 10:34 End time (MST): 15:15
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.479 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC500395
 Removed Cal Gas Conc: 5.479 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
 ZAG Make/Model: Teledyne API T701 Serial Number: 360

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359
 Converter make: CD Nova CDN-101 Converter serial #: 555
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002584	1.000731	Backgd or Offset:	2.66	2.66
Calibration intercept:	0.401023	0.240905	Coeff or Slope:	1.157	1.157

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4927	73.0	80.0	82.2	0.974
as found 2nd point	4964	36.5	40.0	41.5	0.966
as found 3rd point	4983	18.3	20.0	20.5	0.983
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4927	73.0	80.0	80.3	0.996
second point	4964	36.5	40.0	40.2	0.995
third point	4983	18.3	20.0	20.4	0.983
as left zero	5000	0.0	0.0	0.4	----
as left span	4927	73.0	80.0	79.0	1.013
SO2 Scrubber Check	4923	77.1	771.0	0.2	----
Date of last scrubber change:	17-Dec-21			Ave Corr Factor	0.991
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 82.1 Prev response: 80.60 *% change: 1.8%
 Baseline Corr 2nd AF pt: 41.4 AF Slope: 1.027591 AF Intercept: 0.100556
 Baseline Corr 3rd AF pt: 20.4 AF Correlation: 0.999961

* = > +/-5% change initiates investigation

Notes: Scrubber check completed after calibrator zero. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 29, 2023	Previous Calibration:	October 26, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:34	End Time (MST):	15:15
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999996	≥0.995
80.0	80.3	0.9962			
40.0	40.2	0.9948	Slope	1.000731	0.90 - 1.10
20.0	20.4	0.9827			
			Intercept	0.240905	+/-3



TRS Calibration Plot

Date: November 29, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	November 7, 2023	Last Cal Date:	October 17, 2023
Start time (MST):	11:32	End time (MST):	15:10
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.24E-04	2.25E-04	NMHC SP Ratio:	5.04E-05	5.10E-05
CH ₄ Retention time:	12.7	12.7	NMHC Peak Area:	182001	179849
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	17.28	17.14	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.25	1.002
second point	4959	40.5	8.64	8.59	1.007
third point	4979	20.2	4.31	4.23	1.019
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.44	0.991
Average Correction Factor					1.009

Baseline Corr AF:	17.14	Prev response	17.24	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	9.09	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.16	1.001
second point	4959	40.5	4.58	4.58	1.000
third point	4979	20.2	2.29	2.28	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	9.17	9.27	0.989
Average Correction Factor					1.002
Baseline Corr AF:	9.09	Prev response	9.16	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	8.11	8.06	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	8.09	1.003
second point	4959	40.5	4.06	4.00	1.013
third point	4979	20.2	2.02	1.95	1.036
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.17	0.993
Average Correction Factor					1.017
Baseline Corr AF:	8.06	Prev response	8.08	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998344	0.999178
THC Cal Offset:	-0.008797	-0.036603
CH ₄ Cal Slope:	0.998456	0.999274
CH ₄ Cal Offset:	-0.020622	-0.034024
NMHC Cal Slope:	0.998233	0.999119
NMHC Cal Offset:	0.011624	-0.002179

Notes:

N2 Cylinder swapped out. Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 17, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:32	End Time (MST):	15:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999979	≥ 0.995
17.28	17.25	1.0020			
8.64	8.59	1.0066			
4.31	4.23	1.0190			
			Slope	0.999178	0.90 - 1.10
			Intercept	-0.036603	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 17, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:32	End Time (MST):	15:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999920	≥0.995
8.11	8.09	1.0028			
4.06	4.00	1.0135			
2.02	1.95	1.0356			
			Slope	0.999274	0.90 - 1.10
			Intercept	-0.034024	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 17, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:32	End Time (MST):	15:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
9.17	9.16	1.0012						
4.58	4.58	1.0001				Slope	0.999119	0.90 - 1.10
2.29	2.28	1.0047						
			Intercept	-0.002179	± 0.5			



NMHC Calibration Plot

Date: November 7, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: November 30, 2023 Last Cal Date: October 27, 2023
Start time (MST): 11:33 End time (MST): 16:25
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2XX7ME Cal Gas Expiry Date: January 14, 2024
NOX Cal Gas Conc: 50.48 ppm NO Cal Gas Conc: 49.22 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.48 ppm Removed Gas NO Conc: 49.22 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 2658
ZAG make/model: Teledyne API 701H Serial Number: 360

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1336160088
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.072	1.087	NO bkgnd or offset:	3.0	3.1
NOX coeff or slope:	0.985	0.985	NOX bkgnd or offset:	3.0	3.1
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	250.3	248.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999718	1.001069
NO _x Cal Offset:	-0.329761	-0.390046
NO Cal Slope:	1.000010	1.001210
NO Cal Offset:	-1.269777	-1.489996
NO ₂ Cal Slope:	1.001575	1.001477
NO ₂ Cal Offset:	0.299687	1.056533



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as found span	4919	81.3	820.8	800.3	20.5	811.4	787.6	23.8	1.0115	1.0161
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4919	81.3	820.8	800.3	20.5	821.4	800.3	21.0	0.9992	1.0000
second point	4959	40.7	410.9	400.7	10.3	411.0	399.5	11.5	0.9998	1.0029
third point	4980	20.3	204.9	199.8	5.1	204.1	196.7	7.4	1.0041	1.0159
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
as left span	4919	81.3	820.8	351.3	469.5	825.9	349.8	476.2	0.9938	1.0042
Average Correction Factor									1.0010	1.0063

Corrected As found	NO _x = 811.5 ppb	NO = 787.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.1%
Previous Response	NO _x = 820.2 ppb	NO = 799.0 ppb		*Percent Change	NO = -1.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.8	350.8	469.5	470.8	0.9972	100.3%
2nd GPT point (200 ppb O3)	799.8	590.0	230.3	232.0	0.9926	100.7%
3rd GPT point (100 ppb O3)	799.8	697.7	122.6	124.9	0.9815	101.9%
Average Correction Factor					0.9904	101.0%

Notes: Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 30, 2023	Previous Calibration:	October 27, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:33	End Time (MST):	16:25
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
820.8	821.4	0.9992		
410.9	411.0	0.9998		
204.9	204.1	1.0041		
			0.999998	
			1.001069	
			-0.390046	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 30, 2023	Previous Calibration:	October 27, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:33	End Time (MST):	16:25
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
800.3	800.3	1.0000			0.999983
400.7	399.5	1.0029			1.001210
199.8	196.7	1.0159			-1.489996

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 30, 2023	Previous Calibration:	October 27, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:33	End Time (MST):	16:25
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
469.5	470.8	0.9972		
230.3	232.0	0.9926		
122.6	124.9	0.9815		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 30, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: November 23, 2023 Last Cal Date: October 23, 2023
 Start time (MST): 11:40 End time (MST): 15:01
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
 ZAG Make/Model: Teledyne API 701H Serial Number: 360

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 825
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999771	1.001400	Backgd or Offset:	2.500	2.400
Calibration intercept:	0.240000	-0.220000	Coeff or Slope:	1.007	0.978

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	-0.6	----
as found span	4888	1128.7	400.0	412.8	0.969
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	-0.3	----
high point	4888	1101.7	400.0	400.3	0.999
second point	4888	863.9	200.0	200.1	1.000
third point	4888	741.4	100.0	100.0	1.000
as left zero	5000	800.0	0.0	-0.2	----
as left span	4812	1097.9	400.0	402.5	0.994
Average Correction Factor					1.000

Baseline Corr As found:	413.4	Previous response	400.1	*% change	3.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Adjusted span.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 23, 2023	Previous Calibration:	October 23, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	11:40	End Time (MST):	15:01
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	1.000000	≥0.995
400.0	400.3	0.9993			
200.0	200.1	0.9995	Slope	1.001400	0.90 - 1.10
100.0	100.0	1.0000			
			Intercept	-0.220000	+/- 5



O₃ Calibration Plot

Date: November 23, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Stony Mountain Station number: AMS 18
 Calibration Date: November 30, 2023 Last Cal Date: October 27, 2023
 Start time (MST): 15:05 End time (MST): 15:55

Analyzer Make: API T640 S/N: 1162
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-5.5	-6.1	-5.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	697.6	698.3	697.6	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.05	4.98	5.05	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 30, 2023</u>	Last Cal Date: <u>October 27, 2023</u>			
	PM w/o HEPA: <u>4.1</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	NA		NA	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>September 25, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 25, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022
 Date RH/T Sensor Cleaned: August 30, 2022

Notes: No adjustments needed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	November 8, 2023	Last Cal Date:	October 12, 2023
Start time (MST):	11:32	End time (MST):	14:36
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,050	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	3,050	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701		Serial Number:	360

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	Start	Finish		Start	Finish
Calibration slope:	0.991149	0.998397	Backgd or Offset:	-0.010	-0.010
Calibration intercept:	0.191815	0.047809	Coeff or Slope:	0.901	0.906

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.7	40.6	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.7	40.6	1.002
second point	4966	33.3	20.3	20.5	0.990
third point	4983	16.7	10.2	10.2	1.004
as left zero	5000	0.0	0.0	0.1	----
as left span	4933	66.7	40.7	41.0	0.993
Average Correction Factor					0.999

Baseline Corr As found:	40.45	Prev response:	40.52	*% change:	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

** = > +/-5% change initiates investigation*

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 12, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:32	End Time (MST):	14:36
Analyzer make:	API T300	Analyzer serial #:	3504

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999941
40.7	40.6	1.0024		
20.3	20.5	0.9896	Slope	0.998397
10.2	10.2	1.0037		
			Intercept	0.047809
				<i>+/-1.5</i>



CO Calibration Plot

Date: November 8, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	November 16, 2023	Last Cal Date:	October 25, 2023
Start time (MST):	11:45	End time (MST):	15:48
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2026
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
N2 Gen Make/Model:	Peak Scientific		Serial Number:	771048317

Analyzer Information

Analyzer make:	API T360	Analyzer serial #:	489
Analyzer Range	0 - 2,000 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001286	1.002631	Backgd or Offset:	-0.037	-0.037
Calibration intercept:	-6.620000	-3.040000	Coeff or Slope:	0.938	0.938

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	0.0	----
as found span	2920	80.0	1605.9	1609.0	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.8	----
high point	2920	80.0	1605.9	1611.0	0.997
second point	2960	40.0	802.9	794.0	1.011
third point	2980	20.0	401.5	399.7	1.004
as left zero	3000	0.0	0.0	0.1	----
as left span	2930	80.0	1600.5	1609.7	0.994
Average Correction Factor					1.004

Baseline Corr As found:	1609.00	Prev response:	1601.31	*% change:	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Linearity adjustment performed on the third-point. All points verified after linearity adjustment.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	November 16, 2023	Previous Calibration	October 25, 2023
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	11:45	End Time (MST)	15:48
Analyzer make	API T360	Analyzer serial #	489

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.8	----	Correlation Coefficient	0.999933	≥0.995
1605.9	1611.0	0.9968	Slope	1.002631	0.90 - 1.10
802.9	794.0	1.0113	Intercept	-3.040000	+/-10
401.5	399.7	1.0044			



CO₂ Calibration Plot

Date: November 16, 2023

Location: Stony Mountain






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS19 FIREBAG NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023







Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	November 29, 2023	Last Cal Date:	October 27, 2023
Start time (MST):	11:22	End time (MST):	14:58
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.29	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC716618			
Removed Cal Gas Conc:	49.29	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1607
ZAG Make/Model:	API T701		Serial Number:	1118

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002865	1.000607	Backgd or Offset:	9.8	9.9
Calibration intercept:	0.237458	1.477490	Coeff or Slope:	0.975	0.975

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.2	----
as found span	4919	81.1	799.5	801.0	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.0	0.5	----
high point	4919	81.1	799.5	801.0	0.998
second point	4959	40.6	400.3	402.3	0.995
third point	4980	20.3	200.1	202.8	0.987
as left zero	4999	0.0	0.0	0.5	----
as left span	4919	81.1	799.5	804.0	0.994
Average Correction Factor					0.993

Baseline Corr As found:	800.80	Previous response	802.00	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 29, 2023	Previous Calibration:	October 27, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:22	End Time (MST):	14:58
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999993	≥0.995
799.5	801.0	0.9981			
400.3	402.3	0.9949	Slope	1.000607	0.90 - 1.10
200.1	202.8	0.9867			
			Intercept	1.477490	+/-30



SO2 Calibration Plot

Date: November 29, 2023

Location: Firebag





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Firebag Station number: AMS19
 Calibration Date: November 21, 2023 Last Cal Date: October 3, 2023
 Start time (MST): 10:57 End time (MST): 17:04
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.114 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517427
 Removed Cal Gas Conc: 5.114 ppm Rem Gas Exp Date: n/a
 Removed Gas Cyl #: n/a Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1607
 ZAG Make/Model: Teledyne API T701 Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680032
 Converter make: Global Converter serial #: 2022-222
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999195	0.995764	Backgd or Offset: 2.57	2.62
Calibration intercept:	-0.101544	0.018558	Coeff or Slope: 1.136	1.164

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4922	78.2	80.0	78.7	1.016
as found 2nd point	4961	39.1	40.0	38.9	1.028
as found 3rd point	4980	19.6	20.0	19.8	1.013
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4922	78.2	80.0	79.7	1.004
second point	4961	39.1	40.0	39.8	1.005
third point	4980	19.6	20.0	19.9	1.007
as left zero	5000	0.0	0.0	0.1	----
as left span	4922	78.2	80.0	80.5	0.994
SO2 Scrubber Check	4922	78.3	800.2	0.0	----
Date of last scrubber change:	January 18, 2023			Ave Corr Factor	1.005
Date of last converter efficiency test:	n/a			efficiency	

Baseline Corr As found: 78.7 Prev response: 79.81 *% change: -1.4%
 Baseline Corr 2nd AF pt: 38.9 AF Slope: 0.983047 AF Intercept: -0.061291
 Baseline Corr 3rd AF pt: 19.8 AF Correlation: 0.999950

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Ran SOx scrubber check after cal zero. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 21, 2023	Previous Calibration:	October 3, 2023
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	10:57	End Time (MST):	17:04
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999995	≥0.995
80.0	79.7	1.0035			
40.0	39.8	1.0048	Slope	0.995764	0.90 - 1.10
20.0	19.9	1.0075			
			Intercept	0.018558	+/-3



H₂S Calibration Plot

Date: November 21, 2023

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	November 29, 2023	Last Cal Date:	October 27, 2023
Start time (MST):	11:22	End time (MST):	14:58
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1607
ZAG Make/Model:	API T701	Serial Number:	1118

Analyzer Information

Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089
Analyzer Range:	0 - 20 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998006	1.002039	Background:	2.16	2.04
Calibration intercept:	-0.059333	-0.092137	Coefficient:	3.758	3.793

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.00	-0.20	----
as found span	4919	81.1	17.31	17.04	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.00	-0.03	----
high point	4919	81.1	17.31	17.29	1.001
second point	4959	40.6	8.66	8.53	1.016
third point	4980	20.3	4.33	4.20	1.031
as left zero	5000	0.0	0.00	0.05	----
as left span	4919	81.1	17.31	17.36	0.997
Average Correction Factor					1.016
Baseline Corr As found:	17.24	Previous response	17.21	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filters after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 29, 2023	Previous Calibration:	October 27, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:22	End Time (MST):	14:58
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.03	----	Correlation Coefficient	0.999928	≥0.995
17.31	17.29	1.0009			
8.66	8.53	1.0157	Slope	1.002039	0.90 - 1.10
4.33	4.20	1.0311			
			Intercept	-0.092137	+/-1.5



THC Calibration Plot

Date: November 29, 2023

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Firebag
Calibration Date: November 17, 2023
Start time (MST): 11:07
Reason: Routine
Station number: AMS 19
Last Cal Date: October 26, 2023
End time (MST): 15:51

Calibration Standards

NO Gas Cylinder #: T2Y1K63
NOX Cal Gas Conc: 51.12 ppm
Removed Cylinder #: n/a
Removed Gas NOX Conc: 51.12 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 30, 2023
NO Cal Gas Conc: 49.40 ppm
Removed Gas Exp Date: n/a
Removed Gas NO Conc: 49.40 ppm
NO gas Diff:
Serial Number: 1607
Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1410661309

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.078	1.091	NO bkgnd or offset:	7.5	7.6
NOX coeff or slope:	0.995	0.992	NOX bkgnd or offset:	7.6	7.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	217.4	212.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998388	0.998263
NO _x Cal Offset:	0.736136	0.656194
NO Cal Slope:	0.998826	1.002153
NO Cal Offset:	0.248718	0.109347
NO ₂ Cal Slope:	0.996831	1.003810
NO ₂ Cal Offset:	-0.602383	-1.268713



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
as found span	4919	81.0	828.1	800.3	27.9	820.0	790.4	29.7	1.0099	1.0125
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	4999	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
high point	4919	81.0	828.1	800.3	27.9	827.0	802.0	25.3	1.0014	0.9979
second point	4960	40.5	414.0	400.1	13.9	414.2	401.2	13.0	0.9996	0.9973
third point	4980	20.2	206.5	199.6	6.9	207.8	200.3	7.5	0.9938	0.9963
as left zero	4999	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
as left span	4919	81.0	828.1	369.7	458.5	827.0	366.9	460.2	1.0014	1.0076
Average Correction Factor									0.9983	0.9972

Corrected As found	NO _x = 820.3 ppb	NO = 790.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.9%
Previous Response	NO _x = 827.5 ppb	NO = 799.6 ppb		*Percent Change	NO = -1.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.9	367.3	458.5	459.7	0.9973	100.3%
2nd GPT point (200 ppb O3)	797.9	583.3	242.5	241.3	1.0048	99.5%
3rd GPT point (100 ppb O3)	797.9	691.4	134.4	132.5	1.0141	98.6%
Average Correction Factor					1.0054	99.5%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 17, 2023	Previous Calibration:	October 26, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:07	End Time (MST):	15:51
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
828.1	827.0	1.0014			
414.0	414.2	0.9996			
206.5	207.8	0.9938			
			Slope	0.998263	0.90 - 1.10
			Intercept	0.656194	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 17, 2023	Previous Calibration:	October 26, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:07	End Time (MST):	15:51
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	1.000000	≥0.995
800.3	802.0	0.9979			
400.1	401.2	0.9973	Slope	1.002153	0.90 - 1.10
199.6	200.3	0.9963			
			Intercept	0.109347	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 17, 2023	Previous Calibration:	October 26, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:07	End Time (MST):	15:51
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
458.5	459.7	0.9973		
242.5	241.3	1.0048		
134.4	132.5	1.0141		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 17, 2023

Location: Firebag






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS20 MACKAY RIVER NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 17, 2023	Previous Calibration:	October 12, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:00	End Time (MST):	10:52
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999954	≥0.995
800.3	794.6	1.0071			
400.7	399.7	1.0024	Slope	0.990612	0.90 - 1.10
199.8	203.7	0.9810			
			Intercept	2.671171	+/-30



SO2 Calibration Plot

Date: November 17, 2023

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: MacKay River Station number: AMS20
 Calibration Date: November 9, 2023 Last Cal Date: October 24, 2023
 Start time (MST): 8:20 End time (MST): 12:28
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.12 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC515997
 Removed Cal Gas Conc: 5.12 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1220
 ZAG Make/Model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1236656117
 Converter make: Global Converter serial #: 2022-226
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001813	0.980821	Backgd or Offset:	3.1	3.07
Calibration intercept:	0.239528	0.499202	Coeff or Slope:	1.083	1.083

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4922	78.1	80.0	79.9	1.002
as found 2nd point	4961	39.0	39.9	40.1	0.998
as found 3rd point	4980	19.5	20.0	20.3	0.989
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4922	78.1	80.0	78.8	1.016
second point	4961	39.0	40.0	40.0	0.999
third point	4980	19.5	20.0	20.3	0.985
as left zero	5000	0.0	0.0	0.3	----
as left span	4922	78.1	80.0	78.2	1.023
SO2 Scrubber Check	4982	81.3	802.8	-0.1	----
Date of last scrubber change:	May 25, 2023			Ave Corr Factor	1.000
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.8 Prev response: 80.36 *% change: -0.7%
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 0.996881 AF Intercept: 0.239361
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999986

* = > +/-5% change initiates investigation

Notes: No adjustments or maintenance done. Sox scrubber checked after the calibrator zero.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 9, 2023	Previous Calibration:	October 24, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:20	End Time (MST):	12:28
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999923	≥0.995
80.0	78.8	1.0157			
40.0	40.0	0.9992	Slope	0.980821	0.90 - 1.10
20.0	20.3	0.9845			
			Intercept	0.499202	+/-3



H₂S Calibration Plot

Date: November 9, 2023

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	MacKay River	Station number:	AMS20
Calibration Date:	November 17, 2023	Last Cal Date:	October 12, 2023
Start time (MST):	8:00	End time (MST):	10:50
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC306868	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.40</u> ppm	CH4 Equiv Conc.	1066.45 ppm
C3H8 Cal Gas Conc.	<u>206.20</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.40</u> ppm	CH4 Equiv Conc.	1066.45 ppm
Removed C3H8 Conc.	<u>206.20</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	1220
ZAG Make/Model:	Teledyne API 701	Serial Number:	4522

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1501663727
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994974	0.998693	Background:	3.520	2.970
Calibration intercept:	0.076785	0.059818	Coefficient:	5.636	5.605

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.46	----
as found span	4919	81.3	17.34	17.05	1.017
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	-0.02	----
high point	4919	81.3	17.34	17.35	0.999
second point	4959	40.7	8.68	8.73	0.994
third point	4980	20.3	4.33	4.49	0.964
as left zero	5000	0.0	0.00	0.17	----
as left span	4919	81.3	17.34	17.16	1.010
Average Correction Factor					0.986
Baseline Corr As found:	17.51	Previous response	17.33	*% change	1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No maintenance done. Zero and span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 17, 2023	Previous Calibration:	October 12, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:00	End Time (MST):	10:50
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.02	----	Correlation Coefficient	0.999888	≥0.995
17.34	17.35	0.9994			
8.68	8.73	0.9944	Slope	0.998693	0.90 - 1.10
4.33	4.49	0.9643			
			Intercept	0.059818	+/-1.5



THC Calibration Plot

Date: November 17, 2023

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: MacKay River Station number: AMS20
Calibration Date: November 8, 2023 Last Cal Date: October 6, 2023
Start time (MST): 8:20 End time (MST): 12:31
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 1220
ZAG make/model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.486	1.539	NO bkgnd or offset:	4.1	4.3
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	4.1	4.3
NO2 coeff or slope:	0.995	0.995	Reaction cell Press:	190.0	189.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995380	0.996982
NO _x Cal Offset:	3.749814	2.850185
NO Cal Slope:	0.994536	0.997804
NO Cal Offset:	2.591219	1.471555
NO ₂ Cal Slope:	1.007687	1.001029
NO ₂ Cal Offset:	-0.458883	-0.990724



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
as found span	4917	83.3	819.5	800.3	19.2	800.7	778.9	21.7	1.0234	1.0275
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.1	----	----
high point	4917	83.3	819.5	800.3	19.2	819.0	799.8	19.2	1.0006	1.0006
second point	4956	41.7	410.4	400.8	9.6	411.9	400.6	11.3	0.9964	1.0006
third point	4979	20.8	204.6	199.9	4.8	210.5	203.4	7.1	0.9722	0.9826
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.2	----	----
as left span	4917	83.3	819.5	459.9	359.6	814.7	453.0	361.7	1.0058	1.0152
Average Correction Factor									0.9897	0.9946

Corrected As found	NO _x = 800.7 ppb	NO = 779.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -2.3%
Previous Response	NO _x = 819.4 ppb	NO = 798.5 ppb		*Percent Change	NO = -2.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.6	454.2	359.6	359.7	0.9996	100.0%
2nd GPT point (200 ppb O3)	794.6	616.0	197.8	196.2	1.0079	99.2%
3rd GPT point (100 ppb O3)	794.6	701.2	112.6	110.6	1.0177	98.3%
Average Correction Factor					1.0084	99.2%

Notes:

No maintenance done. Span adjusted.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 6, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	8:20	End Time (MST):	12:31
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
819.5	819.0	1.0006			
410.4	411.9	0.9964			
204.6	210.5	0.9722			
			Slope	0.996982	0.90 - 1.10
			Intercept	2.850185	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 6, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	8:20	End Time (MST):	12:31
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
800.3	799.8	1.0006			
400.8	400.6	1.0006			
199.9	203.4	0.9826			
			Slope	0.997804	0.90 - 1.10
			Intercept	1.471555	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 6, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:20	End Time (MST):	12:31
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.0	0.1	----	Correlation Coefficient	0.999949	≥0.995			
359.6	359.7	0.9996						
197.8	196.2	1.0079				Slope	1.001029	0.90 - 1.10
112.6	110.6	1.0177						
			Intercept	-0.990724	+/-20			


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 8, 2023

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: MacKay River Station number: AMS20
Calibration Date: November 30, 2023 Last Cal Date: November 8, 2023
Start time (MST): 8:00 End time (MST): 10:49
Reason: As Found Pump and Charcoal Change

Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 1220
ZAG make/model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.539	1.539	NO bkgnd or offset:	4.3	
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	4.3	
NO2 coeff or slope:	0.995	0.995	Reaction cell Press:	194.7	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996982	
NO _x Cal Offset:	2.850185	
NO Cal Slope:	0.997804	
NO Cal Offset:	1.471555	
NO ₂ Cal Slope:	1.001029	
NO ₂ Cal Offset:	-0.990724	



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as found span	4917	83.3	819.5	800.3	19.2	830.7	808.7	22.0	0.9865	0.9896
as found 2nd	4956	41.7	410.4	400.8	9.6	417.0	404.9	12.1	0.9843	0.9900
as found 3rd	4979	20.8	204.6	199.9	4.8	212.7	205.2	7.5	0.9621	0.9739
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

Average Correction Factor

Corrected As found	NO _x = 830.8 ppb	NO = 808.8 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 1.3%
Previous Response	NO _x = 819.8 ppb	NO = 800.0 ppb			*Percent Change	NO = 1.1%
Baseline Corr 2nd pt	NO _x = 417.1 ppb	NO = 405.0 ppb	As found	NO _x r ² : 0.999953	Nx SI: 1.011748	Nx Int: 2.230
Baseline Corr 3rd pt	NO _x = 212.8 ppb	NO = 205.3 ppb	As found	NO r ² : 0.999978	NO SI: 1.009183	NO Int: 1.211
			As found	NO ₂ r ² : 0.999944	NO ₂ SI: 1.004982	NO ₂ Int: -1.156

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	0.0	----	----
as found GPT point (400 ppb NO ₂)	796.8	454.7	361.3	362.7	0.9960	100.4%
as found GPT point (200 ppb NO ₂)	796.8	617.5	198.5	197.5	1.0048	99.5%
as found GPT point (100 ppb NO ₂)	796.8	702.9	113.1	111.3	1.0158	98.4%
1st GPT point (400 ppb O ₃)						
2nd GPT point (200 ppb O ₃)						
3rd GPT point (100 ppb O ₃)						

Average Correction Factor

Notes:

As founds to change the Pump and Charcoal.


Calibration Performed By:

Melissa Lemay

NO_x Calibration Plot

Date: November 30, 2023

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: MacKay River Station number: AMS20
Calibration Date: December 1, 2023 Last Cal Date: November 30, 2023
Start time (MST): 8:02 End time (MST): 12:14
Reason: Routine Pump and Charcoal Change, PMT adjusted

Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 1220
ZAG make/model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.005	0.981	NO bkgnd or offset:	4.4	4.0
NOX coeff or slope:	0.994	0.997	NOX bkgnd or offset:	4.5	4.1
NO2 coeff or slope:	1.006	0.995	Reaction cell Press:	190.6	190.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996982	1.007887
NO _x Cal Offset:	2.850185	2.589374
NO Cal Slope:	0.997804	0.995492
NO Cal Offset:	1.471555	1.631214
NO ₂ Cal Slope:	1.001029	1.005354
NO ₂ Cal Offset:	-0.990724	0.716865



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
high point	4917	83.3	819.5	800.3	19.2	827.6	797.9	29.7	0.9902	1.0030
second point	4956	41.7	410.4	400.8	9.6	416.4	400.4	16.0	0.9857	1.0011
third point	4979	20.8	204.6	199.9	4.8	212.2	202.8	9.4	0.9644	0.9855
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.0	----	----
as left span	4917	83.3	819.5	456.5	363.0	819.7	456.2	363.4	0.9997	1.0007
Average Correction Factor									0.9801	0.9965

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* => +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.1	451.3	363.0	365.6	0.9928	100.7%
2nd GPT point (200 ppb O3)	795.1	611.0	203.3	204.0	0.9964	100.4%
3rd GPT point (100 ppb O3)	795.1	697.0	117.3	120.5	0.9731	102.8%
Average Correction Factor					0.9874	101.3%

Notes: Calibration from Pump and Charcoal changed, and PMT adjusted November 30,2023. Zero and Span adjusted. Due to drifting during the GPT the 2nd NO ref point used.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 30, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	8:02	End Time (MST):	12:14
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
819.5	827.6	0.9902			
410.4	416.4	0.9857			
204.6	212.2	0.9644			
			Slope	1.007887	0.90 - 1.10
			Intercept	2.589374	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 30, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	8:02	End Time (MST):	12:14
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999978	≥0.995
800.3	797.9	1.0030			
400.8	400.4	1.0011	Slope	0.995492	0.90 - 1.10
199.9	202.8	0.9855			
			Intercept	1.631214	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 30, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:02	End Time (MST):	12:14
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
363.0	365.6	0.9928		
203.3	204.0	0.9964		
117.3	120.5	0.9731		



NO_x Calibration Plot

Date: December 1, 2023

Location: MacKay River






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS21
CONKLIN
NOVEMBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	November 30, 2023	Last Cal Date:	October 20, 2023
Start time (MST):	12:41	End time (MST):	15:52
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.93	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	<u>CC259455</u>			
Removed Cal Gas Conc:	49.93	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3810
ZAG Make/Model:	Teledyne API 701		Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997591	0.998475	Backgd or Offset:	27.4	27.7
Calibration intercept:	0.795745	1.835846	Coeff or Slope:	0.892	0.904

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5005	0.0	0.0	0.5	----
as found span	4920	80.2	800.8	788.8	1.015
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5005	0.0	0.0	0.5	----
high point	4920	80.2	800.8	800.8	1.000
second point	4960	40.1	400.4	402.3	0.995
third point	4980	20.0	200.1	203.0	0.986
as left zero	5005	0.0	0.0	0.6	----
as left span	4920	80.2	800.8	802.9	0.997
Average Correction Factor					0.994

Baseline Corr As found:	788.30	Previous response	799.71	*% change	-1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 30, 2023	Previous Calibration:	October 20, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	12:41	End Time (MST):	15:52
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.5	----	Correlation Coefficient	0.999987	
800.8	800.8	1.0001			≥0.995
400.4	402.3	0.9954	Slope	0.998475	
200.1	203.0	0.9858			0.90 - 1.10
			Intercept	1.835846	+/-30



SO2 Calibration Plot

Date: November 30, 2023

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Conklin Station number: AMS21
 Calibration Date: November 28, 2023 Last Cal Date: October 17, 2023
 Start time (MST): 11:50 End time (MST): 17:15
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.00 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC501204
 Removed Cal Gas Conc: 5.00 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3810
 ZAG Make/Model: Teledyne API 701H Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1236656116
 Converter make: CD-Nova 101 Converter serial #: NA
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994857	0.997143	Backgd or Offset: 2.3	2.38
Calibration intercept:	0.180000	0.200000	Coeff or Slope: 0.947	0.963

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.0	80.0	79.0	1.013
as found 2nd point	4960	40.0	40.0	39.6	1.010
as found 3rd point	4980	20.0	20.0	20.0	1.000
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.0	80.0	79.9	1.001
second point	4960	40.0	40.0	40.2	0.995
third point	4980	20.0	20.0	20.2	0.990
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	80.0	80.0	79.0	1.013
SO2 Scrubber Check	4920	80.2	802.0	0.1	----

Date of last scrubber change:	Ave Corr Factor	0.995
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 79.0 Prev response: 79.77 *% change: -1.0%
 Baseline Corr 2nd AF pt: 39.6 AF Slope: 0.986571 AF Intercept: 0.120000
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999989

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 17, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:50	End Time (MST):	17:15
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999991	≥0.995
80.0	79.9	1.0013			
40.0	40.2	0.9950	Slope	0.997143	0.90 - 1.10
20.0	20.2	0.9901			
			Intercept	0.200000	+/-3



TRS Calibration Plot

Date: November 28, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	November 30, 2023	Last Cal Date:	October 25, 2023
Start time (MST):	12:41	End time (MST):	15:52
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.01E-04	3.82E-04	NMHC SP Ratio:	9.22E-05
CH ₄ Retention time:	14.6	14.2	NMHC Peak Area:	99089
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				9.44E-05
				96823
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.07	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.11	1.001
second point	4960	40.1	8.56	8.57	0.999
third point	4980	20.0	4.28	4.34	0.985
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.15	0.998

Average Correction Factor				0.995
Baseline Corr AF:	17.07	Prev response	17.03	*% change 0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.13	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.14	1.000
second point	4960	40.1	4.57	4.57	1.001
third point	4980	20.0	2.28	2.30	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.16	0.998
Average Correction Factor					0.998
Baseline Corr AF:	9.13	Prev response	9.08	*% change	0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.94	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.96	1.003
second point	4960	40.1	3.99	4.00	0.998
third point	4980	20.0	2.00	2.05	0.974
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	8.00	0.999
Average Correction Factor					0.992
Baseline Corr AF:	7.94	Prev response	7.95	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.993314	0.997585
THC Cal Offset:	0.021764	0.030767
CH ₄ Cal Slope:	0.995000	0.995028
CH ₄ Cal Offset:	0.007556	0.026958
NMHC Cal Slope:	0.991515	0.999806
NMHC Cal Offset:	0.015007	0.003609

Notes: Changed the inlet filter and H2 cylinder after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 30, 2023	Previous Calibration:	October 25, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	12:41	End Time (MST):	15:52
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999981	≥ 0.995			
17.13	17.11	1.0011						
8.56	8.57	0.9994				Slope	0.997585	0.90 - 1.10
4.28	4.34	0.9851						
			Intercept	0.030767	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 30, 2023	Previous Calibration:	October 25, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	12:41	End Time (MST):	15:52
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999940	≥0.995
7.99	7.96	1.0028			
3.99	4.00	0.9980			
2.00	2.05	0.9744			
			Slope	0.995028	0.90 - 1.10
			Intercept	0.026958	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 30, 2023	Previous Calibration:	October 25, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	12:41	End Time (MST):	15:52
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
9.14	9.14	0.9997						
4.57	4.57	1.0008				Slope	0.999806	0.90 - 1.10
2.28	2.30	0.9947						
			Intercept	0.003609	± 0.5			



NMHC Calibration Plot

Date: November 30, 2023

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Conklin
Calibration Date: November 29, 2023
Start time (MST): 12:21
Reason: Routine
Station number: AMS21
Last Cal Date: October 27, 2023
End time (MST): 16:37

Calibration Standards

NO Gas Cylinder #: T2Y1P1H
NOX Cal Gas Conc: 51.09 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 51.09 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.39 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 50.39 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 691

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1501663731

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.094	1.094	NO bkgnd or offset:	10.9	10.9
NOX coeff or slope:	1.000	1.000	NOX bkgnd or offset:	11.0	11.1
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	153.0	167.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999869	1.003588
NO _x Cal Offset:	2.584550	2.465024
NO Cal Slope:	1.002036	1.004807
NO Cal Offset:	1.361881	1.242275
NO ₂ Cal Slope:	1.004241	1.006340
NO ₂ Cal Offset:	-0.358712	-0.474448



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.5	-0.1	----	----
as found span	4921	79.4	811.2	800.1	11.1	817.4	801.8	15.6	0.9925	0.9979
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
high point	4921	79.4	811.2	800.1	11.1	815.1	804.4	10.8	0.9953	0.9947
second point	4960	39.7	405.7	400.1	5.6	411.4	404.2	7.2	0.9861	0.9899
third point	4980	19.8	202.3	199.6	2.8	208.0	203.2	4.8	0.9727	0.9820
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.4	0.1	----	----
as left span	4921	79.4	811.2	385.1	426.1	810.5	381.9	428.6	1.0009	1.0085
Average Correction Factor									0.9847	0.9889

Corrected As found	NO _x = 818.0 ppb	NO = 802.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.5%
Previous Response	NO _x = 813.7 ppb	NO = 803.1 ppb		*Percent Change	NO = -0.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI: ;

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.3	382.3	426.1	428.8	0.9937	100.6%
2nd GPT point (200 ppb O3)	797.3	594.9	213.5	213.5	1.0001	100.0%
3rd GPT point (100 ppb O3)	797.3	698.4	110.0	110.2	0.9983	100.2%
Average Correction Factor					0.9974	100.3%

Notes: Changed the inlet filter after as founds. No adjustments made. Used the 2nd GPT reference point because of drift.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 29, 2023	Previous Calibration:	October 27, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	12:21	End Time (MST):	16:37
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
811.2	815.1	0.9953		
405.7	411.4	0.9861		
202.3	208.0	0.9727		
			0.999947	
			1.003588	
			2.465024	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 29, 2023	Previous Calibration:	October 27, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	12:21	End Time (MST):	16:37
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999983	≥0.995
800.1	804.4	0.9947			
400.1	404.2	0.9899	Slope	1.004807	0.90 - 1.10
199.6	203.2	0.9820			
			Intercept	1.242275	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 29, 2023	Previous Calibration:	October 27, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	12:21	End Time (MST):	16:37
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
426.1	428.8	0.9937		
213.5	213.5	1.0001		
110.0	110.2	0.9983		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 29, 2023

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	November 22, 2023	Last Cal Date:	October 27, 2023
Start time (MST):	11:50	End time (MST):	15:52
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	Teledyne API T700	Serial Number:	3810
ZAG Make/Model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734
Analyzer Range:	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001314	1.000543	Backgd or Offset:	-0.3	-1.1
Calibration intercept:	1.520000	0.280000	Coeff or Slope:	1.005	1.005

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-1.5	----
as found span	5000	1564.0	400.0	397.7	1.006
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.2	----
high point	5000	950.9	400.0	400.5	0.999
second point	5000	804.0	200.0	200.3	0.999
third point	5000	703.6	100.0	100.5	0.995
as left zero	5000	0.0	0.0	0.2	----
as left span	5000	936.0	400.0	402.4	0.994
Average Correction Factor					0.997

Baseline Corr As found:	399.2	Previous response	402.0	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 22, 2023	Previous Calibration:	October 27, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:50	End Time (MST):	15:52
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	1.000000	≥0.995
400.0	400.5	0.9988			
200.0	200.3	0.9985	Slope	1.000543	0.90 - 1.10
100.0	100.5	0.9950			
			Intercept	0.280000	+/- 5



O₃ Calibration Plot

Date: November 22, 2023

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Conklin Station number: AMS 21
 Calibration Date: November 30, 2023 Last Cal Date: October 27, 2023
 Start time (MST): 12:49 End time (MST): 13:40

Analyzer Make: API T640 S/N: 326
 Particulate Fraction: PM2.5

Flow Meter Make/Model: DeltaCal S/N: 1450
 Temp/RH standard: DeltaCal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-3	-3.4	-3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	706	705.2	706	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	5.05	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 30, 2023</u>	Last Cal Date: <u>October 27, 2023</u>			
	PM w/o HEPA: <u>4.2</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>September 26, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 26, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: Verified flow, temperature, and pressure. Leak check passed. No adjustments made. Firmware updated

Calibration by: Rene Chamberland




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER
NOVEMBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 3, 2023	Previous Calibration:	October 5, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	8:58	End Time (MST):	13:03
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	1.5	----	Correlation Coefficient	≥0.995
799.8	802.5	0.9966		
399.9	402.6	0.9933	Slope	0.90 - 1.10
200.4	202.8	0.9884		
			Intercept	+/-30



SO2 Calibration Plot

Date: November 3, 2023

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Janvier	Station number:	AMS22
Calibration Date:	November 9, 2023	Last Cal Date:	October 18, 2023
Start time (MST):	10:44	End time (MST):	14:58
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.03	ppm	Cal Gas Exp Date:	April 16, 2022
Cal Gas Cylinder #:	DT0018680			
Removed Cal Gas Conc:	5.03	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031
Converter make:	CDN-101	Converter serial #:	587
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011516	0.986936	Backgd or Offset:	3.04	3.04
Calibration intercept:	0.060524	0.621094	Coeff or Slope:	1.161	1.161

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4920	79.5	80.0	78.3	1.027
as found 2nd point	4960	39.8	40.0	40.0	1.011
as found 3rd point	4980	19.9	20.0	20.3	1.006
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4920	79.5	80.0	79.4	1.007
second point	4960	39.8	40.0	40.4	0.991
third point	4980	19.9	20.0	20.5	0.977
as left zero	5000	0.0	0.0	0.5	----
as left span	4920	79.5	80.0	79.7	1.004
SO2 Scrubber Check	4920	79.8	798.0	0.0	----

Date of last scrubber change:	Ave Corr Factor	0.992
Date of last converter efficiency test:	efficiency	

Baseline Corr As found:	77.9	Prev response:	80.97	*% change:	-3.9%
Baseline Corr 2nd AF pt:	39.6	AF Slope:	0.973077	AF Intercept:	0.681308
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999918		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 9, 2023	Previous Calibration:	October 18, 2023
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	10:44	End Time (MST):	14:58
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999954	
80.0	79.4	1.0074			≥0.995
40.0	40.4	0.9911	Slope	0.986936	
20.0	20.5	0.9766			0.90 - 1.10
			Intercept	0.621094	+/-3



TRS Calibration Plot

Date: November 9, 2023

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	November 3, 2023	Last Cal Date:	October 5, 2023
Start time (MST):	8:58	End time (MST):	10:20
Reason:	Removal		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH ₄ Cal Gas Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
C ₃ H ₈ Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
Removed C ₃ H ₈ Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.200E-04	2.200E-04	NMHC SP Ratio:	4.43E-05	4.43E-05
CH ₄ Retention time:	13.8	13.8	NMHC Peak Area:	206448	206448
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	16.86	1.019
as found 2nd point	4960	39.9	8.59	8.37	1.026
as found 3rd point	4980	20.0	4.30	4.21	1.023
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

			Average Correction Factor		
Baseline Corr AF:	16.86	Prev response	17.17	*% change	-1.8%
Baseline Corr 2nd AF:	8.4	AF Slope:	0.981757	AF Intercept:	-0.019524
Baseline Corr 3rd AF:	4.2	AF Correlation:	0.999985	<i>* = +/-5% change initiates investigation</i>	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	8.96	1.021
as found 2nd point	4960	39.9	4.57	4.46	1.025
as found 3rd point	4980	20.0	2.29	2.22	1.035
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	
Baseline Corr AF:	8.96	Prev response	9.11	*% change	-1.7%
Baseline Corr 2nd AF:	4.5	AF Slope:	0.980695	AF Intercept:	-0.016186
Baseline Corr 3rd AF:	2.2	AF Correlation:	0.999985	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	7.90	1.016
as found 2nd point	4960	39.9	4.01	3.91	1.027
as found 3rd point	4980	20.0	2.01	1.99	1.011
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	
Baseline Corr AF:	7.90	Prev response	8.06	*% change	-2.1%
Baseline Corr 2nd AF:	3.91	AF Slope:	0.982868	AF Intercept:	-0.003738
Baseline Corr 3rd AF:	1.99	AF Correlation:	0.999955	* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.003343	
THC Cal Offset:	-0.059000	
CH ₄ Cal Slope:	1.008361	
CH ₄ Cal Offset:	-0.033775	
NMHC Cal Slope:	0.998915	
NMHC Cal Offset:	-0.024625	


Notes: Changed the inlet filter and the N2 cylinder after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell

NMHC Calibration Plot

Date: November 3, 2023

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	November 3, 2023	Last Cal Date:	N/A
Start time (MST):	10:50	End time (MST):	13:03
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH ₄ Cal Gas Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
C ₃ H ₈ Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
Removed C ₃ H ₈ Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	N/A	2.470E-04	NMHC SP Ratio:	N/A	4.94E-05
CH ₄ Retention time:	N/A	15.0	NMHC Peak Area:	N/A	185124
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.09	1.005
second point	4960	39.9	8.59	8.61	0.997
third point	4980	20.0	4.30	4.40	0.979
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.01	1.010
Average Correction Factor					0.993

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.15	9.07	1.008
second point	4960	39.9	4.57	4.60	0.994
third point	4980	20.0	2.29	2.37	0.966
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.15	8.97	1.020
Average Correction Factor					0.989
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.02	1.001
second point	4960	39.9	4.01	4.01	1.000
third point	4980	20.0	2.01	2.02	0.994
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.04	0.998
Average Correction Factor					0.998
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	N/A	0.993524
THC Cal Offset:	N/A	0.058796
CH ₄ Cal Slope:	N/A	0.998422
CH ₄ Cal Offset:	N/A	0.006635
NMHC Cal Slope:	N/A	0.989089
NMHC Cal Offset:	N/A	0.052961

Notes: Install calibration. Swapped out the H2 cylinder and the inlet filter prior to the calibration. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 3, 2023	Previous Calibration:	N/A
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:50	End Time (MST):	13:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999945	≥ 0.995			
17.17	17.09	1.0047						
8.59	8.61	0.9968				Slope	0.993524	0.90 - 1.10
4.30	4.40	0.9788						
			Intercept	0.058796	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 3, 2023	Previous Calibration:	N/A
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:50	End Time (MST):	13:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999996	≥0.995
8.03	8.02	1.0010			
4.01	4.01	1.0001			
2.01	2.02	0.9937			
			Slope	0.998422	0.90 - 1.10
			Intercept	0.006635	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 3, 2023	Previous Calibration:	N/A
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:50	End Time (MST):	13:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999843	≥ 0.995			
9.15	9.07	1.0080						
4.57	4.60	0.9938				Slope	0.989089	0.90 - 1.10
2.29	2.37	0.9656						
			Intercept	0.052961	± 0.5			



NMHC Calibration Plot

Date: November 3, 2023

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: November 8, 2023 Last Cal Date: October 17, 2023
Start time (MST): 10:25 End time (MST): 14:56
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC424183 Cal Gas Expiry Date: April 16, 2023
NOX Cal Gas Conc: 48.60 ppm NO Cal Gas Conc: 48.60 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.60 ppm Removed Gas NO Conc: 48.60 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3806
ZAG make/model: Teledyne API T701 Serial Number: 201

Analyzer Information

Analyzer make: API T200 Analyzer serial #: 833
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.826	0.826	NO bkgnd or offset:	-5.6	-5.6
NOX coeff or slope:	0.815	0.815	NOX bkgnd or offset:	-3.9	-3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	5.0	5.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.991915	0.990857
NO _x Cal Offset:	2.589941	2.110476
NO Cal Slope:	0.989600	0.989570
NO Cal Offset:	2.050558	1.211114
NO ₂ Cal Slope:	0.999193	1.001094
NO ₂ Cal Offset:	0.162793	0.263799



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.9	2.1	-0.2	----	----
as found span	4918	82.3	799.9	799.9	0.0	792.3	786.9	5.5	1.0096	1.0165
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	1.5	2.0	-0.5	----	----
high point	4918	82.3	799.9	799.9	0.0	794.6	793.1	1.5	1.0067	1.0086
second point	4959	41.2	400.4	400.4	0.0	398.6	397.3	1.3	1.0046	1.0079
third point	4980	20.6	200.2	200.2	0.0	201.5	198.4	3.1	0.9936	1.0091
as left zero	5000	0.0	0.0	0.0	0.0	1.7	2.3	-0.6	----	----
as left span	4918	82.3	799.9	409.4	390.5	789.9	399.5	390.4	1.0127	1.0248
Average Correction Factor									1.0016	1.0085

Corrected As found	NO _x = 790.4 ppb	NO = 784.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.7%
Previous Response	NO _x = 796.0 ppb	NO = 793.6 ppb		*Percent Change	NO = -1.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	789.9	399.4	390.5	390.7	0.9995	100.1%
2nd GPT point (200 ppb O3)	789.9	596.5	193.4	194.6	0.9938	100.6%
3rd GPT point (100 ppb O3)	789.9	692.7	97.2	98.1	0.9908	100.9%
Average Correction Factor					0.9947	100.5%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 17, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:25	End Time (MST):	14:56
Analyzer make:	API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	1.5	----	Correlation Coefficient	≥0.995	
799.9	794.6	1.0067			
400.4	398.6	1.0046			
200.2	201.5	0.9936			
			Slope	0.990857	0.90 - 1.10
			Intercept	2.110476	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 17, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:25	End Time (MST):	14:56
Analyzer make:	API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	2.0	----	Correlation Coefficient	≥0.995	
799.9	793.1	1.0086			
400.4	397.3	1.0079			
200.2	198.4	1.0091			
			Slope	0.989570	0.90 - 1.10
			Intercept	1.211114	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 17, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:25	End Time (MST):	14:56
Analyzer make:	API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.5	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
390.5	390.7	0.9995		
193.4	194.6	0.9938		
97.2	98.1	0.9908		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 8, 2023

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: November 23, 2023 Last Cal Date: October 20, 2023
 Start time (MST): 10:19 End time (MST): 13:41
 Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701H Serial Number: 691

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 7046
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996857	1.006114	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	1.800000	0.780000	Coeff or Slope:	1.021	1.021

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	1.8	----
as found span	4895	905.3	400.0	402.8	0.993
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.8	----
high point	4895	905.3	400.0	403.1	0.992
second point	4895	756.7	200.0	202.4	0.988
third point	4895	656.1	100.0	101.1	0.989
as left zero	5000	800.0	0.0	0.8	----
as left span	4895	904.3	400.0	403.2	0.992
Average Correction Factor					0.990

Baseline Corr As found:	401.0	Previous response	400.5	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 23, 2023	Previous Calibration:	October 20, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:19	End Time (MST):	13:41
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.8	----	Correlation Coefficient	≥0.995
400.0	403.1	0.9923		
200.0	202.4	0.9881	Slope	0.90 - 1.10
100.0	101.1	0.9891		
			Intercept	+/- 5



O₃ Calibration Plot

Date: November 23, 2023

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: November 23, 2023 Last Cal Date: October 20, 2023
 Start time (MST): 13:43 End time (MST): 14:01

Analyzer Make: Teledyne API T640 S/N: 325
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-3.3	-3.7	-3.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	726.5	727.1	726.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.10	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 23, 2023</u>	Last Cal Date: <u>October 20, 2023</u>			
	PM w/o HEPA: <u>7.3</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>October 20, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>October 20, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: July 26, 2023
 Date RH/T Sensor Cleaned: July 26, 2023

Notes:

Quarterly calibration was completed in October. No adjustments needed. Leak check passed.

Calibration by: Max Farrell




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23 FORT HILLS NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	November 20, 2023	Last Cal Date:	October 17, 2023
Start time (MST):	9:45	End time (MST):	12:20
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.76	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC281425		Rem Gas Exp Date:	N/A
Removed Cal Gas Conc:	49.76	ppm	Diff between cyl:	
Removed Gas Cyl #:	N/A		Serial Number:	451
Calibrator Make/Model:	API T700		Serial Number:	5611
ZAG Make/Model:	API T701			

Analyzer Information

Analyzer make: Thermo 43i
 Analyzer Range 0 - 1000 ppb
 Analyzer serial #: 1160290012

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999149	0.993772	Backgd or Offset:	17.8	17.8
Calibration intercept:	-0.923240	-0.642432	Coeff or Slope:	1.040	1.040

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4920	80.3	799.1	793.2	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.3	799.1	793.6	1.007
second point	4960	40.2	400.1	397.2	1.007
third point	4980	20.1	200.0	197.1	1.015
as left zero	5000	0.0	0.0	0.4	----
as left span	4920	80.3	799.1	794.1	1.006
Average Correction Factor					1.010

Baseline Corr As found:	792.90	Previous response	797.49	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments or maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 20, 2023	Previous Calibration:	October 17, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:45	End Time (MST):	12:20
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999995	
799.1	793.6	1.0069			≥0.995
400.1	397.2	1.0072	Slope	0.993772	
200.0	197.1	1.0149			0.90 - 1.10
			Intercept	-0.642432	+/-30



SO2 Calibration Plot

Date: November 20, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	November 15, 2023	Last Cal Date:	October 2, 2023
Start time (MST):	8:58	End time (MST):	12:49
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.20	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517372			
Removed Cal Gas Conc:	5.20	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	451
ZAG Make/Model:	API T701		Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232
Converter make:	CDN-101	Converter serial #:	594
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004317	1.000174	Backgd or Offset:	1.19
Calibration intercept:	-0.038148	-0.118216	Coeff or Slope:	1.124

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4923	77.0	80.0	79.0	1.011
as found 2nd point	4962	38.5	40.0	39.6	1.007
as found 3rd point	4981	19.2	19.9	19.6	1.013
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.0	80.0	79.9	1.001
second point	4962	38.5	40.0	40.0	1.000
third point	4981	19.2	19.9	19.6	1.018
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.0	80.0	80.2	0.998
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.006
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	79.1	Prev response:	80.31	*% change:	-1.5%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.989031	AF Intercept:	-0.078363
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999994		

* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 15, 2023	Previous Calibration:	October 2, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:58	End Time (MST):	12:49
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999977	
80.0	79.9	1.0013			≥0.995
40.0	40.0	0.9999	Slope	1.000174	
19.9	19.6	1.0178			0.90 - 1.10
			Intercept	-0.118216	+/-3



TRS Calibration Plot

Date: November 15, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	November 20, 2023	Last Cal Date:	October 17, 2023
Start time (MST):	9:45	End time (MST):	12:19
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
Removed C ₃ H ₈ Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.34E-04	2.34E-04	NMHC SP Ratio:	5.10E-05	5.01E-05
CH ₄ Retention time:	13.0	13.0	NMHC Peak Area:	180460	183497
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	17.48	0.984
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.30	0.994
second point	4960	40.2	8.61	8.65	0.995
third point	4980	20.1	4.30	4.33	0.994
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.19	17.31	0.993
Average Correction Factor					0.994

Baseline Corr AF:	17.48	Prev response	17.30	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.38	0.976
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.23	0.992
second point	4960	40.2	4.59	4.65	0.986
third point	4980	20.1	2.29	2.34	0.980
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	9.16	9.24	0.991
Average Correction Factor					0.986
Baseline Corr AF:	9.38	Prev response	9.22	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	8.10	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.07	0.995
second point	4960	40.2	4.02	4.00	1.005
third point	4980	20.1	2.01	1.99	1.010
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.03	8.07	0.995
Average Correction Factor					1.004
Baseline Corr AF:	8.10	Prev response	8.08	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.006154	1.006220
THC Cal Offset:	0.005584	-0.002409
CH ₄ Cal Slope:	1.007912	1.005350
CH ₄ Cal Offset:	-0.014057	-0.020045
NMHC Cal Slope:	1.004487	1.006983
NMHC Cal Offset:	0.017645	0.017636

Notes:

Span adjusted. No Maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 20, 2023	Previous Calibration:	October 17, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:45	End Time (MST):	12:19
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995
17.19	17.30	0.9938			
8.61	8.65	0.9950			
4.30	4.33	0.9939			
			Slope	1.006220	0.90 - 1.10
			Intercept	-0.002409	+/-0.5

THC Calibration Curve





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	November 20, 2023	Previous Calibration:	October 17, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:45	End Time (MST):	12:19
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999965	≥0.995
8.03	8.07	0.9954			
4.02	4.00	1.0054			
2.01	1.99	1.0104			
			Slope	1.005350	0.90 - 1.10
			Intercept	-0.020045	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	November 20, 2023	Previous Calibration:	October 17, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:45	End Time (MST):	12:19
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999982	≥ 0.995			
9.16	9.23	0.9923						
4.59	4.65	0.9861				Slope	1.006983	0.90 - 1.10
2.29	2.34	0.9798						
			Intercept	0.017636	± 0.5			



NMHC Calibration Plot

Date: November 20, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	November 28, 2023	Last Cal Date:	November 20, 2023
Start time (MST):	8:35	End time (MST):	10:05
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
Removed C ₃ H ₈ Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.34E-04	2.34E-04	NMHC SP Ratio:	5.01E-05	5.01E-05
CH ₄ Retention time:	13.0	13.0	NMHC Peak Area:	183497	183497
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	17.06	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.30	0.994
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	0.994
Baseline Corr AF:	17.06	Prev response	17.30	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.12	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.23	0.992
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.992
Baseline Corr AF:	9.12	Prev response	9.24	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	7.94	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.07	0.995
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.995
Baseline Corr AF:	7.94	Prev response	8.06	*% change	-1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.006220	1.006282
THC Cal Offset:	-0.002409	0.000000
CH ₄ Cal Slope:	1.005350	1.004640
CH ₄ Cal Offset:	-0.020045	0.000000
NMHC Cal Slope:	1.006983	1.007722
NMHC Cal Offset:	0.017636	0.000000

Notes:


Nitrogen Cylinder Changed.

Calibration Performed By: Melissa Lemay

NMHC Calibration Plot

Date: November 28, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Hills
Calibration Date: November 7, 2023
Start time (MST): 7:35
Reason: Routine
Station number: AMS23
Last Cal Date: October 11, 2023
End time (MST): 12:04

Calibration Standards

NO Gas Cylinder #: CC332703
NOX Cal Gas Conc: 49.7 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 49.7 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 28, 2024
NO Cal Gas Conc: 49.7 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 49.7 ppm
NO gas Diff:
Serial Number: 451
Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1152430007

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.089	1.089	NO bkgnd or offset:	3.3	3
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	3.8	3.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	161.6	161.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000322	1.003266
NO _x Cal Offset:	0.004208	0.024268
NO Cal Slope:	0.999781	1.001225
NO Cal Offset:	-0.956508	-1.036547
NO ₂ Cal Slope:	1.002559	1.003745
NO ₂ Cal Offset:	-0.119059	0.698395



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.5	-0.1	----	----
as found span	4920	80.5	800.2	800.2	0.0	802.7	794.4	8.3	0.997	1.007
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4920	80.5	800.2	800.2	0.0	802.8	800.5	2.2	0.997	1.000
second point	4960	40.2	399.6	399.6	0.0	401.0	398.9	2.1	0.996	1.002
third point	4980	20.1	199.8	199.8	0.0	200.3	197.7	2.6	0.997	1.011
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as left span	4920	80.5	800.2	447.8	352.4	802.1	447.9	354.4	0.998	1.000
Average Correction Factor									0.997	1.004

Corrected As found	NO _x = 803.3 ppb	NO = 794.9 ppb	<i>* => +/-5% change initiates investigation</i>		*Percent Change	NO _x = 0.4%
Previous Response	NO _x = 800.4 ppb	NO = 799.0 ppb			*Percent Change	NO = -0.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.9	447.5	352.4	354.3	0.995	100.5%
2nd GPT point (200 ppb O3)	799.9	625.1	174.8	175.9	0.994	100.6%
3rd GPT point (100 ppb O3)	799.9	711.3	88.6	90.6	0.978	102.3%
Average Correction Factor					0.989	101.1%

Notes:

No maintenance done. Zero adjusted.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 11, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:35	End Time (MST):	12:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.2	802.8	0.9967			
399.6	401.0	0.9964			
199.8	200.3	0.9975			
			Slope	1.003266	0.90 - 1.10
			Intercept	0.024268	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 11, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:35	End Time (MST):	12:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
800.2	800.5	0.9996			0.999992
399.6	398.9	1.0017			1.001225
199.8	197.7	1.0106			-1.036547

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 11, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:35	End Time (MST):	12:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
352.4	354.3	0.9946			
174.8	175.9	0.9937			
88.6	90.6	0.9779			
			Slope	1.003745	0.90 - 1.10
			Intercept	0.698395	+/-20



NO_x Calibration Plot

Date: November 7, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Hills Station number: AMS 23
 Calibration Date: November 20, 2023 Last Cal Date: October 17, 2023
 Start time (MST): 8:42 End time (MST): 9:44

Analyzer Make: API T640 S/N: 1546
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-7.1	-7.06	-7.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	740.1	739.97	740.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.13	5.00	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: November 20, 2023 Last Cal Date: October 17, 2023
 PM w/o HEPA: 6.4 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.8	10.7	10.7	<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: 15.6 w/ HEPA: 0.0
 Date Optical Chamber Cleaned: November 20, 2023 <0.2 ug/m3
 Disposable Filter Changed: November 20, 2023

Annual Maintenance

Date Sample Tube Cleaned: October 17, 2023
 Date RH/T Sensor Cleaned: October 17, 2023

Leak check passed before and after cleaning. No adjustments done. Head cleaned. Low memory Warning before cleaning, disappeared after cleaning.

Notes:

Calibration by: Melissa Lemay




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25 WASKŌW OHCI PIMÂTISIWIN NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 21, 2023	Previous Calibration:	October 19, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	10:35	End Time (MST):	13:35
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient	1.000000	≥0.995
800.5	797.6	1.0037			
400.3	399.1	1.0030	Slope	0.995565	0.90 - 1.10
200.1	199.7	1.0022			
			Intercept	0.584063	+/-30



SO2 Calibration Plot

Date: November 21, 2023

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS25
 Calibration Date: November 24, 2023 Last Cal Date: October 20, 2023
 Start time (MST): 7:08 End time (MST): 11:52
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC517099
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 747
 ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
 Converter make: Global G-150 Converter serial #: 2022-219
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004258	0.998073	Backgd or Offset:	3.30	3.30
Calibration intercept:	0.140000	0.180000	Coeff or Slope:	1.108	1.108

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.0	79.5	79.7	0.997
as found 2nd point	4960	40.0	39.7	40.0	0.993
as found 3rd point	4980	20.0	19.9	20.0	0.993
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.0	79.5	79.4	1.001
second point	4960	40.0	39.7	40.0	0.993
third point	4980	20.0	19.9	20.0	0.993
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	80.0	800.0	795.1	1.006
SO2 Scrubber Check	4921	79.2	800.0	0.0	----
Date of last scrubber change:	20-Jun-23			Ave Corr Factor	0.996
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.7 Prev response: 79.93 *% change: -0.3%
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 1.002963 AF Intercept: 0.060000
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999996

* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 24, 2023	Previous Calibration:	October 20, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	7:08	End Time (MST):	11:52
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999988	
79.5	79.4	1.0007			≥0.995
39.7	40.0	0.9932	Slope	0.998073	
19.9	20.0	0.9932			0.90 - 1.10
			Intercept	0.180000	+/-3



H₂S Calibration Plot

Date: November 24, 2023

Location: Waskow ohci Pimatisiwin






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS26 CHRISTINA LAKE NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 29, 2023	Previous Calibration:	October 26, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	9:12	End Time (MST):	12:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.2	----	Correlation Coefficient	0.999994	≥0.995
800.0	800.1	0.9998			
400.6	400.1	1.0012	Slope	0.999235	0.90 - 1.10
201.6	200.8	1.0041			
			Intercept	0.269171	+/-30



SO2 Calibration Plot

Date: November 29, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Christina Lake Station number: AMS26
 Calibration Date: November 29, 2023 Last Cal Date: October 27, 2023
 Start time (MST): 12:06 End time (MST): 15:56
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.89 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002466
 Removed Cal Gas Conc: 4.89 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 281
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 450i Analyzer serial #: 1180030032
 Converter make: NA Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992122	1.002002	Backgd or Offset: 33.6	33.9
Calibration intercept:	0.528176	0.607400	Coeff or Slope: 1.060	1.060

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4930	81.9	79.9	81.2	0.988
as found 2nd point	4972	41.0	40.0	41.1	0.980
as found 3rd point	4994	20.6	20.1	20.8	0.980
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.6	----
high point	4930	81.9	79.9	80.7	0.990
second point	4972	41.0	40.0	40.6	0.985
third point	4994	20.6	20.1	20.8	0.966
as left zero	5000	0.0	0.0	0.7	----
as left span	4926	81.9	80.0	81.5	0.981
SO2 Scrubber Check	4931	80.9	807.1	0.1	----
Date of last scrubber change:	27-Feb-19			Ave Corr Factor	0.980
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 80.9 Prev response: 79.81 *% change: 1.4%
 Baseline Corr 2nd AF pt: 40.8 AF Slope: 1.012163 AF Intercept: 0.426783
 Baseline Corr 3rd AF pt: 20.5 AF Correlation: 0.999981

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after MAF's. Ran scrubber check after calibrator zero passed. No adjustment made.

Calibration Performed By: Mohammed Kashif and Jan Castro



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 29, 2023	Previous Calibration:	October 27, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	12:06	End Time (MST):	15:56
Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.6	----	Correlation Coefficient	≥0.995
79.9	80.7	0.9902		
40.0	40.6	0.9851	Slope	0.90 - 1.10
20.1	20.8	0.9658		
			Intercept	+/-3



H₂S Calibration Plot

Date: November 29, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.1	-0.7	----	----
as found span	4920	80.0	813.1	800.3	12.8	809.0	794.7	14.8	1.0051	1.0071
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.6	0.0	-0.7	----	----
high point	4920	80.0	813.1	800.3	12.8	808.0	795.0	12.6	1.0063	1.0067
second point	4960	40.0	406.6	400.2	6.4	405.5	397.8	7.7	1.0026	1.0059
third point	4980	20.0	203.3	200.1	3.2	203.1	199.3	3.8	1.0009	1.0039
as left zero	5000	0.0	0.0	0.0	0.0	-0.7	0.0	-0.6	----	----
as left span	4920	80.0	813.1	411.6	401.5	807.0	408.3	398.3	1.0076	1.0081
Average Correction Factor									1.0033	1.0055

Corrected As found	NO _x = 809.8 ppb	NO = 794.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.6%	
Previous Response	NO _x = 814.6 ppb	NO = 800.9 ppb		*Percent Change	NO = -0.8%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	791.8	403.1	401.5	403.1	0.9960	100.4%
2nd GPT point (200 ppb O3)	791.8	596.4	208.2	208.9	0.9966	100.3%
3rd GPT point (100 ppb O3)	791.8	695.2	109.4	110.9	0.9865	101.4%
Average Correction Factor					0.9931	100.7%

Notes: Changed sample inlet filter after as founds. No adjustment made.

Calibration Performed By: Mohammed Kashif and Jan Castro



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 26, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	13:26	End Time (MST):	18:03
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.1	808.0	1.0063		
406.6	405.5	1.0026		
203.3	203.1	1.0009		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 26, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	13:26	End Time (MST):	18:03
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	795.0	1.0067		
400.2	397.8	1.0059		
200.1	199.3	1.0039		
			0.999999	
			0.993160	
			0.280000	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 26, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	13:26	End Time (MST):	18:03
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.7	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
401.5	403.1	0.9960		
208.2	208.9	0.9966		
109.4	110.9	0.9865		
			0.999982	
			1.004527	
			-0.038830	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 28, 2023

Location: Christina Lake






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27
JACKFISH 2/3
NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 14, 2023	Previous Calibration:	October 18, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	12:25	End Time (MST):	16:05
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999961	≥0.995
800.2	798.8	1.0017			
399.5	395.9	1.0092	Slope	0.999320	0.90 - 1.10
200.3	195.9	1.0225			
			Intercept	-2.038541	+/-30



SO2 Calibration Plot

Date: November 14, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Jackfish 2/3 Station number: AMS27
 Calibration Date: November 15, 2023 Last Cal Date: October 17, 2023
 Start time (MST): 9:05 End time (MST): 14:24
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.41 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345023
 Removed Cal Gas Conc: 5.41 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3811
 ZAG Make/Model: API 701 Serial Number: 135

Analyzer Information

Analyzer make: API T101 Analyzer serial #: 621
 Converter make: Converter serial #:
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010889	0.996209	Backgd or Offset: 29.9	29.9
Calibration intercept:	-0.037634	-0.098000	Coeff or Slope: 0.928	0.928

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4926	74.1	80.2	80.9	0.995
as found 2nd point	4963	37.0	40.0	40.3	1.001
as found 3rd point	4982	18.5	20.0	19.9	1.021
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	74.1	80.2	79.8	1.005
second point	4963	37.0	40.0	39.9	1.003
third point	4982	18.5	20.0	19.5	1.026
as left zero	5000	0.0	0.0	0.2	----
as left span	4926	74.1	80.2	80.5	0.996
SO2 Scrubber Check	4921	79.1	791.0	0.1	----

Date of last scrubber change:	Ave Corr Factor	1.011
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 80.6 Prev response: 81.01 *% change: -0.5%
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 1.007187 AF Intercept: 0.042159
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999952

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments required.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 15, 2023	Previous Calibration:	October 17, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:05	End Time (MST):	14:24
Analyzer make:	API T101	Analyzer serial #:	621

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999951	
80.2	79.8	1.0047			≥0.995
40.0	39.9	1.0034	Slope	0.996209	
20.0	19.5	1.0264			0.90 - 1.10
			Intercept	-0.098000	+/-3



H₂S Calibration Plot

Date: November 15, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Jackfish 2/3
Calibration Date: November 9, 2023
Start time (MST): 10:33
Reason: Routine
Station number: AMS27
Last Cal Date: October 19, 2023
End time (MST): 15:39

Calibration Standards

NO Gas Cylinder #: T2Y1P35
NOX Cal Gas Conc: 51.44 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 51.44 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.40 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.40 ppm
NO gas Diff:
Serial Number: 3811
Serial Number: 135

Analyzer Information

Analyzer make: API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 722

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.123	1.123	NO bkgnd or offset:	0.3	0.3
NOX coeff or slope:	1.110	1.110	NOX bkgnd or offset:	1.2	1.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	7.6	9.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995490	1.005005
NO _x Cal Offset:	-2.256328	-2.855689
NO Cal Slope:	0.998987	1.007782
NO Cal Offset:	-2.840289	-2.818947
NO ₂ Cal Slope:	0.999566	0.986277
NO ₂ Cal Offset:	-0.710230	0.637074



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.4	-0.1	----	----
as found span	4921	79.4	816.8	800.3	16.5	817.6	801.1	16.5	0.9990	0.9990
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.2	----	----
high point	4921	79.4	816.8	800.3	16.5	819.5	805.3	14.2	0.9967	0.9938
second point	4960	39.7	408.5	400.2	8.3	406.0	398.6	7.4	1.0061	1.0040
third point	4980	19.8	203.7	199.6	4.1	199.3	195.7	3.6	1.0221	1.0199
as left zero	5000	0.0	0.0	0.0	0.0	-0.5	0.1	-0.6	----	----
as left span	4921	79.4	816.8	424.1	405.2	816.1	423.3	392.7	1.0009	1.0019
Average Correction Factor									1.0083	1.0059

Corrected As found	NO _x = 818.1 ppb	NO = 801.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.9%
Previous Response	NO _x = 810.9 ppb	NO = 796.6 ppb		*Percent Change	NO = 0.6%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	805.7	417.0	405.2	399.8	1.0135	98.7%
2nd GPT point (200 ppb O3)	805.7	628.7	193.5	192.2	1.0068	99.3%
3rd GPT point (100 ppb O3)	805.7	722.1	100.1	100.0	1.0011	99.9%
Average Correction Factor					1.0072	99.3%

Notes:

Changed the sample inlet filter after as founds. No adjustments required.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 9, 2023	Previous Calibration:	October 19, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	10:33	End Time (MST):	15:39
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
816.8	819.5	0.9967		
408.5	406.0	1.0061		
203.7	199.3	1.0221		
			0.999948	
			1.005005	
			-2.855689	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 9, 2023	Previous Calibration:	October 19, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	10:33	End Time (MST):	15:39
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	805.3	0.9938		
400.2	398.6	1.0040		
199.6	195.7	1.0199		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 9, 2023	Previous Calibration:	October 19, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	10:33	End Time (MST):	15:39
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
405.2	399.8	1.0135		
193.5	192.2	1.0068		
100.1	100.0	1.0011		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 9, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
high point	4921	79.4	816.8	800.3	16.5	816.0	797.2	18.8	1.0010	1.0039
second point	4960	39.7	408.5	400.2	8.3	404.0	395.9	8.1	1.0110	1.0109
third point	4980	19.8	203.7	199.6	4.1	198.4	192.2	6.3	1.0268	1.0385
as left zero	5000	0.0	0.0	0.0	0.0	1.5	1.3	0.3	----	----
as left span	4921	79.4	816.8	412.3	399.1	808.0	403.3	404.5	1.0109	1.0223
Average Correction Factor									1.0129	1.0177

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* => +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	802.0	419.4	399.1	395.9	1.0081	99.2%
2nd GPT point (200 ppb O3)	802.0	623.9	194.6	192.2	1.0126	98.8%
3rd GPT point (100 ppb O3)	802.0	714.0	104.5	100.5	1.0399	96.2%
Average Correction Factor					1.0202	98.0%

Notes: Swapped external pump. As finds not done due to low span response. Adjusted span only.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 20, 2023	Previous Calibration:	November 9, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	12:17	End Time (MST):	16:59
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
816.8	816.0	1.0010		
408.5	404.0	1.0110		
203.7	198.4	1.0268		
			0.999947	
			1.000794	
			-2.976420	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 20, 2023	Previous Calibration:	November 9, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	12:17	End Time (MST):	16:59
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999924	≥0.995
800.3	797.2	1.0039			
400.2	395.9	1.0109	Slope	0.998689	0.90 - 1.10
199.6	192.2	1.0385			
			Intercept	-3.261265	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 20, 2023	Previous Calibration:	November 9, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	12:17	End Time (MST):	16:59
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
399.1	395.9	1.0081			
194.6	192.2	1.0126			
104.5	100.5	1.0399			
			Slope	0.994531	0.90 - 1.10
			Intercept	-1.480680	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 20, 2023

Location: Jackfish 2/3





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS29
SURMONT 2
NOVEMBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 16, 2023	Previous Calibration:	October 23, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:55	End Time (MST):	17:02
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
800.1	804.0	0.9952		
400.6	401.8	0.9970	Slope	0.90 - 1.10
199.8	198.3	1.0077		
			Intercept	+/-30



SO2 Calibration Plot

Date: November 16, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	November 1, 2023	Last Cal Date:	October 24, 2023
Start time (MST):	9:44	End time (MST):	14:25
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	<u>5.391</u>	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	<u>CC508338</u>			
Removed Cal Gas Conc:	<u>5.391</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>CC508338</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	5472
ZAG Make/Model:	Teledyne API T701		Serial Number:	4297

Analyzer Information

Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170
Converter make:	Global	Converter serial #:	2022-220
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003047	1.001467	Backgd or Offset:	0.92	0.92
Calibration intercept:	-0.022987	-0.142608	Coeff or Slope:	1.080	1.080

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4926	74.2	80.0	80.7	0.989
as found 2nd point	4963	37.2	40.1	40.1	0.995
as found 3rd point	4982	18.6	20.1	20.0	0.993
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4926	74.2	80.0	80.1	0.999
second point	4963	37.2	40.1	39.8	1.008
third point	4982	18.6	20.1	19.9	1.008
as left zero	5000	0.0	0.0	0.0	----
as left span	4926	74.2	80.0	79.5	1.006
SO2 Scrubber Check	4919	81.3	813.0	-0.1	----

Date of last scrubber change:	Ave Corr Factor	1.005
Date of last converter efficiency test:	efficiency	

Baseline Corr As found:	80.9	Prev response:	80.22	*% change:	0.8%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.011184	AF Intercept:	-0.283103
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999988		

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made. Scrubber check done after cal zero, passed.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 1, 2023	Previous Calibration:	October 24, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	9:44	End Time (MST):	14:25
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	≥0.995
80.0	80.1	0.9988		
40.1	39.8	1.0077	Slope	0.90 - 1.10
20.1	19.9	1.0077		
			Intercept	+/-3



H₂S Calibration Plot

Date: November 1, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	November 16, 2023	Last Cal Date:	October 23, 2023
Start time (MST):	10:55	End time (MST):	17:02
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC356008	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
C3H8 Cal Gas Conc.	<u>205.7</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
Removed C3H8 Conc.	<u>205.7</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5472
ZAG Make/Model:	Teledyne API T701	Serial Number:	4297

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1170050149
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007852	0.998269	Background:	3.64	3.53
Calibration intercept:	-0.102215	-0.044666	Coefficient:	3.987	3.986

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.12	----
as found span	4918	81.3	17.31	17.12	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	-0.03	----
high point	4918	81.3	17.31	17.26	1.003
second point	4959	40.6	8.65	8.54	1.012
third point	4979	20.3	4.32	4.28	1.010
as left zero	5000	0.0	0.00	-0.04	----
as left span	4918	81.3	17.31	17.42	0.994
Average Correction Factor					1.008
Baseline Corr As found:	17.24	Previous response	17.35	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed out Hydrogen cylinder. Flame extinguished, causing drift. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 16, 2023	Previous Calibration:	October 23, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:55	End Time (MST):	17:02
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	-0.03	----	Correlation Coefficient	0.999982	≥0.995
17.31	17.26	1.0031			
8.65	8.54	1.0124	Slope	0.998269	0.90 - 1.10
4.32	4.28	1.0099			
			Intercept	-0.044666	+/-1.5



THC Calibration Plot

Date: November 16, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
as found span	4916	84.2	799.2	799.2	0.0	812.0	809.0	2.2	0.9842	0.9879
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
high point	4916	84.2	799.2	799.2	0.0	797.9	801.0	-3.4	1.0016	0.9977
second point	4958	42.1	399.6	399.6	0.0	401.0	399.6	1.3	0.9965	1.0000
third point	4979	21.1	200.3	200.3	0.0	199.2	197.4	1.8	1.0054	1.0146
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as left span	4916	84.2	799.2	409.3	389.9	797.5	406.4	391.1	1.0021	1.0071
Average Correction Factor									1.0012	1.0041

Corrected As found	NO _x = 812.2 ppb	NO = 809.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.1%
Previous Response	NO _x = 803.7 ppb	NO = 802.6 ppb		*Percent Change	NO = 0.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.1	406.2	389.9	388.2	1.0044	99.6%
2nd GPT point (200 ppb O3)	796.1	609.3	186.8	186.6	1.0011	99.9%
3rd GPT point (100 ppb O3)	796.1	701.2	94.9	95.2	0.9968	100.3%
Average Correction Factor					1.0008	99.9%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 14, 2023	Previous Calibration:	October 25, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:46	End Time (MST):	15:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
799.2	797.9	1.0016			
399.6	401.0	0.9965			
200.3	199.2	1.0054			
			Slope	0.999094	0.90 - 1.10
			Intercept	0.047580	+/-20

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 14, 2023	Previous Calibration:	October 25, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:46	End Time (MST):	15:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999982	≥0.995
799.2	801.0	0.9977			
399.6	399.6	1.0000	Slope	1.003756	0.90 - 1.10
200.3	197.4	1.0146			
			Intercept	-1.632802	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 14, 2023	Previous Calibration:	October 25, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:46	End Time (MST):	15:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
389.9	388.2	1.0044		
186.8	186.6	1.0011		
94.9	95.2	0.9968		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 14, 2023

Location: Surmont 2






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS30 ELLS RIVER NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 3, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:49	End Time (MST):	14:53
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999908	
800.4	800.6	0.9997			≥0.995
400.2	395.1	1.0130	Slope	1.002373	
200.1	193.8	1.0325			0.90 - 1.10
			Intercept	-3.655880	+/-30



SO2 Calibration Plot

Date: November 8, 2023

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Ells River	Station number:	AMS30
Calibration Date:	November 30, 2023	Last Cal Date:	October 26, 2023
Start time (MST):	11:43	End time (MST):	15:15
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.08	ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	EY0002443			
Removed Cal Gas Conc:	5.08	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3061
ZAG Make/Model:	API T701H		Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331
Converter make:	CDN - 101	Converter serial #:	562
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998492	1.001780	Backgd or Offset:	1.55	1.57
Calibration intercept:	-0.099119	-0.139199	Coeff or Slope:	1.100	1.100

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	78.7	80.0	79.8	1.002
as found 2nd point	4961	39.4	40.0	39.5	1.013
as found 3rd point	4980	19.7	20.0	19.7	1.016
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	78.7	80.0	80.0	1.000
second point	4961	39.4	40.0	40.0	1.001
third point	4980	19.7	20.0	19.7	1.016
as left zero	5000	0.0	0.0	0.2	----
as left span	4921	78.7	80.0	80.3	0.996
SO2 Scrubber Check	4921	79.2	800.4	0.1	----
Date of last scrubber change:	N/A			Ave Corr Factor	1.005
Date of last converter efficiency test:	N/A			efficiency	

Baseline Corr As found:	79.8	Prev response:	79.74	*% change:	0.1%
Baseline Corr 2nd AF pt:	39.5	AF Slope:	0.998489	AF Intercept:	-0.199028
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999959		

* = > +/-5% change initiates investigation

Notes: SOx scrubber check done after calibrator zero, passed. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 30, 2023	Previous Calibration:	October 26, 2023
Station Name:	Ells River	Station Number:	AMS30
Start Time (MST):	11:43	End Time (MST):	15:15
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	≥0.995
80.0	80.0	0.9995		
40.0	40.0	1.0007	Slope	0.90 - 1.10
20.0	19.7	1.0161		
			Intercept	+/-3



TRS Calibration Plot

Date: November 30, 2023

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	November 8, 2023	Last Cal Date:	October 3, 2023
Start time (MST):	10:49	End time (MST):	14:53
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm		
Diff between cyl (CH ₄):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701H	Serial Number:	3061
		Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.45E-04	2.46E-04	NMHC SP Ratio:	4.75E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	191712
				166643

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	15.77	1.080
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.01	1.001
second point	4960	39.6	8.51	8.43	1.010
third point	4980	19.8	4.26	4.17	1.022
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.07	0.998

Average Correction Factor				1.011
Baseline Corr AF:	15.77	Prev response	16.97	*% change -7.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	7.93	1.150
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.09	1.002
second point	4960	39.6	4.56	4.53	1.006
third point	4980	19.8	2.28	2.24	1.016
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.15	0.996
Average Correction Factor					1.008
Baseline Corr AF:	7.93	Prev response	9.07	*% change	-14.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.84	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.91	1.000
second point	4960	39.6	3.96	3.90	1.014
third point	4980	19.8	1.98	1.92	1.029
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.92	0.999
Average Correction Factor					1.015
Baseline Corr AF:	7.84	Prev response	7.90	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000476	1.000046
THC Cal Offset:	-0.062338	-0.048138
CH ₄ Cal Slope:	1.000886	1.001246
CH ₄ Cal Offset:	-0.022757	-0.032757
NMHC Cal Slope:	0.999995	0.998929
NMHC Cal Offset:	-0.039581	-0.015582

Notes: Inlet filter and N2 cylinder changed after As Finds, adjusted span only.

Calibration Performed By: Mohammed Kashif and Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 3, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:49	End Time (MST):	14:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999963	≥ 0.995			
17.03	17.01	1.0012						
8.51	8.43	1.0096				Slope	1.000046	0.90 - 1.10
4.26	4.17	1.0216						
			Intercept	-0.048138	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-01-2020


Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 3, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:49	End Time (MST):	14:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999917	≥0.995
7.91	7.91	1.0004			
3.96	3.90	1.0141			
1.98	1.92	1.0291			
			Slope	1.001246	0.90 - 1.10
			Intercept	-0.032757	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 3, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:49	End Time (MST):	14:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995			
9.11	9.09	1.0020						
4.56	4.53	1.0057				Slope	0.998929	0.90 - 1.10
2.28	2.24	1.0157						
			Intercept	-0.015582	± 0.5			



NMHC Calibration Plot

Date: November 8, 2023

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Ells River Station number: AMS 30
Calibration Date: November 28, 2023 Last Cal Date: October 20, 2023
Start time (MST): 9:46 End time (MST): 14:34
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1P2R Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.83 ppm NO Cal Gas Conc: 49.97 ppm
Removed Cylinder #: Removed Gas Exp Date:
Removed Gas NOX Conc: 50.83 ppm Removed Gas NO Conc: 49.97 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: API T700 Serial Number: 3061
ZAG make/model: API T701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 710321429
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.147	1.147	NO bkgnd or offset:	14.0	14.0
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	14.0	14.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	190.3	190.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001461	0.995096
NO _x Cal Offset:	-1.680000	-1.240000
NO Cal Slope:	1.000029	0.993053
NO Cal Offset:	-2.600000	-2.360000
NO ₂ Cal Slope:	1.004460	1.003095
NO ₂ Cal Offset:	0.734060	0.737995



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.7	-0.1	----	----
as found span	4920	80.0	813.3	799.5	13.8	812.9	794.0	18.9	1.0005	1.0070
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.4	0.0	----	----
high point	4920	80.0	813.3	799.5	13.8	808.4	792.7	15.7	1.0060	1.0086
second point	4960	40.0	406.6	399.8	6.9	403.1	393.3	9.8	1.0088	1.0164
third point	4980	20.0	203.3	199.9	3.4	200.3	194.4	5.9	1.0151	1.0282
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
as left span	4920	80.0	813.3	436.7	376.6	808.6	429.2	379.3	1.0058	1.0175
Average Correction Factor									1.0100	1.0177

Corrected As found	NO _x = 813.7 ppb	NO = 794.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.1%	
Previous Response	NO _x = 812.8 ppb	NO = 796.9 ppb		*Percent Change	NO = -0.3%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	788.4	425.6	376.6	377.9	0.9965	100.4%
2nd GPT point (200 ppb O ₃)	788.4	615.3	186.9	189.1	0.9882	101.2%
3rd GPT point (100 ppb O ₃)	788.4	697.4	104.8	106.2	0.9864	101.4%
Average Correction Factor					0.9903	101.0%

Notes:

Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 20, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:46	End Time (MST):	14:34
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.5	----	Correlation Coefficient	≥0.995	
813.3	808.4	1.0060			
406.6	403.1	1.0088			
203.3	200.3	1.0151			
			Slope	0.995096	0.90 - 1.10
			Intercept	-1.240000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 20, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:46	End Time (MST):	14:34
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.5	792.7	1.0086		
399.8	393.3	1.0164		
199.9	194.4	1.0282		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 20, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:46	End Time (MST):	14:34
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
376.6	377.9	0.9965			
186.9	189.1	0.9882			
104.8	106.2	0.9864			
			Correlation Coefficient	0.999976	≥0.995
			Slope	1.003095	0.90 - 1.10
			Intercept	0.737995	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 28, 2023

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Ells River Station number: AMS 30
 Calibration Date: November 30, 2023 Last Cal Date: October 27, 2023
 Start time (MST): 10:59 End time (MST): 12:37

Analyzer Make: API T640 S/N: 875
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388750
 Temp/RH standard: Alicat FP-25BT S/N: 388750

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-5.1	-5.23	-5.1	<input type="checkbox"/>	
P (mmHg)	729.7	731.6	729.7	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.126	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 30, 2023</u>	Last Cal Date: <u>October 27, 2023</u>			
	PM w/o HEPA: <u>2.1</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>October 27, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>October 27, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: October 27, 2023
 Date RH/T Sensor Cleaned: November 30, 2023

No adjustments made. Cleaned RH/T sensor.

Notes:

Calibration by: Braiden Boutilier




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS501
LEISMER
NOVEMBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 18, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	13:38	End Time (MST):	16:23
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient	≥0.995
800.2	802.5	0.9971		
400.2	399.7	1.0011	Slope	0.90 - 1.10
200.1	197.4	1.0135		
			Intercept	+/-30



SO2 Calibration Plot

Date: November 28, 2023

Location: Leismer





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Leismer Station number: AMS501
 Calibration Date: November 29, 2023 Last Cal Date: October 17, 2023
 Start time (MST): 8:15 End time (MST): 12:23
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.14 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511843
 Removed Cal Gas Conc: 5.14 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 2659
 ZAG Make/Model: API 701 Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1180540020
 Converter make: Global G150 Converter serial #: 2022-218
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006859	1.002431	Backgd or Offset:	3.41
Calibration intercept:	-0.318205	-0.138350	Coeff or Slope:	1.082
				3.47
				1.105

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4922	77.8	80.0	78.4	1.018
as found 2nd point	4961	38.9	40.0	39.1	1.018
as found 3rd point	4981	19.4	19.9	19.2	1.028
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4922	77.8	80.0	80.0	1.000
second point	4961	38.9	40.0	40.1	0.997
third point	4981	19.4	19.9	19.7	1.012
as left zero	5000	0.0	0.0	-0.1	----
as left span	4922	77.8	80.0	80.3	0.996
SO2 Scrubber Check	4921	79.2	792.0	0.0	----

Date of last scrubber change:	24-Feb-23	Ave Corr Factor	1.003
Date of last converter efficiency test:	December 1, 2022	efficiency	

Baseline Corr As found: 78.6 Prev response: 80.21 *% change: -2.1%
 Baseline Corr 2nd AF pt: 39.3 AF Slope: 0.983574 AF Intercept: -0.278750
 Baseline Corr 3rd AF pt: 19.4 AF Correlation: 0.999992

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Scrubber test done after calibrator zero. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 29, 2023	Previous Calibration:	October 17, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	8:15	End Time (MST):	12:23
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	≥0.995
80.0	80.0	0.9998		
40.0	40.1	0.9973	Slope	0.90 - 1.10
19.9	19.7	1.0123		
			Intercept	+/-3



H₂S Calibration Plot

Date: November 29, 2023

Location: Leismer





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Leismer
Calibration Date: November 28, 2023
Start time (MST): 9:23
Reason: Removal
Station number: AMS501
Last Cal Date: October 26, 2023
End time (MST): 13:42

Calibration Standards

NO Gas Cylinder #: T26811M
NOX Cal Gas Conc: 47.46 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.46 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API 701
Cal Gas Expiry Date: October 30, 2024
NO Cal Gas Conc: 47.39 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.39 ppm
NO gas Diff:
Serial Number: 2659
Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1218153356

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.301	1.225	NO bkgnd or offset:	3.9	3.6
NOX coeff or slope:	0.995	0.992	NOX bkgnd or offset:	4.1	3.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	169.2	165.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.009935	1.007139
NO _x Cal Offset:	-4.483299	-3.748007
NO Cal Slope:	1.007430	1.006069
NO Cal Offset:	-5.190972	-4.627968
NO ₂ Cal Slope:	1.006440	0.997452
NO ₂ Cal Offset:	-0.613191	0.289298



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
as found span	4916	84.4	801.1	799.9	1.2	840.3	833.5	6.8	0.9533	0.9597
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
high point	4916	84.4	801.1	799.9	1.2	804.3	801.9	2.6	0.9960	0.9975
second point	4958	42.2	400.5	400.0	0.6	399.6	396.9	2.6	1.0024	1.0077
third point	4979	21.1	200.3	200.0	0.3	193.0	191.1	2.0	1.0377	1.0465
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as left span	4916	84.4	801.1	378.2	422.9	806.6	390.4	416.2	0.9931	0.9687
Average Correction Factor									1.0120	1.0172

Corrected As found	NO _x = 840.7 ppb	NO = 833.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 4.3%
Previous Response	NO _x = 804.5 ppb	NO = 800.6 ppb		*Percent Change	NO = 4.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	808.3	386.6	422.9	422.1	1.0019	99.8%
2nd GPT point (200 ppb O3)	808.3	597.5	212.0	211.5	1.0023	99.8%
3rd GPT point (100 ppb O3)	808.3	697.9	111.6	112.0	0.9963	100.4%
Average Correction Factor					1.0001	100.0%

Notes: Changed inlet filter after as founds. Adjusted zero and span. Used 2nd NO reference point due to drift.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 26, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:23	End Time (MST):	13:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
801.1	804.3	0.9960		
400.5	399.6	1.0024		
200.3	193.0	1.0377		
			0.999887	
			1.007139	
			-3.748007	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 26, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:23	End Time (MST):	13:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999847	≥0.995
799.9	801.9	0.9975			
400.0	396.9	1.0077	Slope	1.006069	0.90 - 1.10
200.0	191.1	1.0465			
			Intercept	-4.627968	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 28, 2023	Previous Calibration:	October 26, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:23	End Time (MST):	13:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
422.9	422.1	1.0019		
212.0	211.5	1.0023		
111.6	112.0	0.9963		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 28, 2023

Location: Leismer






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS505
SAWBONES BAY
NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 6, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	13:47	End Time (MST):	16:26
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321323

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥0.995
799.8	798.7	1.0014		
399.9	398.5	1.0035	Slope	0.90 - 1.10
200.4	197.3	1.0159		
			Intercept	+/-30



SO2 Calibration Plot

Date: November 8, 2023

Location: Sawbones Bay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Sawbones Bay Station number: AMS505
 Calibration Date: November 9, 2023 Last Cal Date: October 5, 2023
 Start time (MST): 8:22 End time (MST): 12:31
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517397
 Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 5112
 ZAG Make/Model: Teledyne API T701 Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43iQ Analyzer serial #: 1228021057
 Converter make: Global 150 Converter serial #: 2022-224
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008501	0.994794	Backgd or Offset: 2.21	2.21
Calibration intercept:	0.082063	0.261745	Coeff or Slope: 1.021	1.021

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4922	77.7	80.0	80.0	1.002
as found 2nd point	4961	38.8	40.0	40.4	0.992
as found 3rd point	4981	19.4	20.0	20.2	0.994
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4922	77.7	80.0	79.9	1.002
second point	4961	38.8	40.0	40.0	0.999
third point	4981	19.4	20.0	20.1	0.994
as left zero	5000	0.0	0.0	0.4	----
as left span	4922	77.7	80.0	79.6	1.005
SO2 Scrubber Check	4922	77.8	778.0	-0.1	----

Date of last scrubber change:	Ave Corr Factor	0.998
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 79.9 Prev response: 80.80 *% change: -1.1%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 0.998216 AF Intercept: 0.241983
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999969

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after calibrator zero. Scrubber check and passed after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 9, 2023	Previous Calibration:	October 5, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:22	End Time (MST):	12:31
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1228021057

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995
80.0	79.9	1.0017			
40.0	40.0	0.9991	Slope	0.994794	0.90 - 1.10
20.0	20.1	0.9940			
			Intercept	0.261745	+/-3



H₂S Calibration Plot

Date: November 9, 2023

Location: Sawbones Bay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.3	-0.3	0.5	----	----
as found span	4917	83.4	799.6	799.6	0.0	802.3	803.7	-1.3	0.9966	0.9949
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.3	----	----
high point	4917	83.4	799.6	799.6	0.0	801.2	801.4	-0.4	0.9980	0.9977
second point	4958	41.7	399.8	399.8	0.0	399.3	399.7	-0.4	1.0014	1.0004
third point	4979	20.9	200.4	200.4	0.0	197.8	197.0	0.8	1.0131	1.0172
as left zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
as left span	4916	83.4	799.7	332.4	467.3	795.9	332.5	463.4	1.0048	0.9998
Average Correction Factor									1.0041	1.0051

Corrected As found	NO _x = 802.0 ppb	NO = 804.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.1%
Previous Response	NO _x = 800.8 ppb	NO = 802.6 ppb		*Percent Change	NO = 0.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.5	330.2	467.3	467.3	1.0000	100.0%
2nd GPT point (200 ppb O3)	797.5	543.5	254.0	253.5	1.0020	99.8%
3rd GPT point (100 ppb O3)	797.5	647.7	149.8	148.8	1.0067	99.3%
Average Correction Factor					1.0029	99.7%

Notes:

Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 4, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:53	End Time (MST):	13:50
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
799.6	801.2	0.9980			
399.8	399.3	1.0014			
200.4	197.8	1.0131			
			Slope	1.003423	0.90 - 1.10
			Intercept	-1.650848	+/-20

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 4, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:53	End Time (MST):	13:50
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999975	≥0.995
799.6	801.4	0.9977			
399.8	399.7	1.0004	Slope	1.003879	0.90 - 1.10
200.4	197.0	1.0172			
			Intercept	-1.810600	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 4, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:53	End Time (MST):	13:50
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
467.3	467.3	1.0000			
254.0	253.5	1.0020			
149.8	148.8	1.0067			
			Slope	1.000996	0.90 - 1.10
			Intercept	-0.666972	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 8, 2023

Location: Sawbones Bay






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS507
KIRBY SOUTH
NOVEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 4, 2023
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	13:24	End Time (MST):	16:48
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.5	----	Correlation Coefficient	0.999977	≥0.995
799.6	804.0	0.9946			
400.3	398.9	1.0036	Slope	1.005777	0.90 - 1.10
199.7	200.0	0.9983			
			Intercept	-1.328222	+/-30



SO2 Calibration Plot

Date: November 6, 2023

Location: Kirby South





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Kirby South Station number: AMS507
 Calibration Date: November 8, 2023 Last Cal Date: October 4, 2023
 Start time (MST): 11:17 End time (MST): 16:37
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.167 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517378
 Removed Cal Gas Conc: 5.167 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T750 Serial Number: 282
 ZAG Make/Model: API T751H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1150840012
 Converter make: Global Converter serial #: 2022-197
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003178	1.004602	Backgd or Offset: 1.75	1.98
Calibration intercept:	-0.041177	0.058972	Coeff or Slope: 1.068	1.058

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4923	77.4	80.0	81.4	0.984
as found 2nd point	4961	38.8	40.1	40.4	0.995
as found 3rd point	4981	19.3	19.9	20.3	0.987
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.4	80.0	80.4	0.995
second point	4961	38.8	40.1	40.3	0.995
third point	4981	19.3	19.9	20.2	0.987
as left zero	5000	0.0	0.0	0.1	----
as left span	4923	77.4	80.0	78.4	1.020
SO2 Scrubber Check	4919	80.0	800.2	0.1	----
Date of last scrubber change:	25-Jul-23			Ave Corr Factor	0.992
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 81.3 Prev response: 80.19 *% change: 1.4%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 1.016164 AF Intercept: -0.020760
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999960

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Scrubber check done after MPAF's, passed. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 4, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	11:17	End Time (MST):	16:37
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	≥0.995
80.0	80.4	0.9948		
40.1	40.3	0.9950	Slope	0.90 - 1.10
19.9	20.2	0.9873		
			Intercept	+/-3



H₂S Calibration Plot

Date: November 8, 2023

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	November 6, 2023	Last Cal Date:	October 19, 2023
Start time (MST):	13:24	End time (MST):	16:48
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC303554	Cal Gas Expiry Date:	March 23, 2025
CH4 Cal Gas Conc.	496.6 ppm	CH4 Equiv Conc.	1061.7 ppm
C3H8 Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	496.6 ppm	CH4 Equiv Conc.	1061.7 ppm
Removed C3H8 Conc.	205.5 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	3804
ZAG Make/Model:	API T701H	Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005
Analyzer Range:	0 - 20 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997581	1.004599	Background:	3.12	3.15
Calibration intercept:	-0.000204	-0.117967	Coefficient:	3.720	3.755

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.07	----
as found span	4919	81.3	17.26	17.10	1.010
as found 2nd point	4959	40.7	8.64		
as found 3rd point	4980	20.3	4.31		
new cylinder response					
calibrator zero	5000	0.0	0.00	-0.09	----
high point	4919	81.3	17.26	17.26	1.000
second point	4959	40.7	8.64	8.49	1.018
third point	4980	20.3	4.31	4.22	1.022
as left zero	5000	0.0	0.00	-0.12	----
as left span	4919	81.3	17.26	17.23	1.002
Average Correction Factor					1.013
Baseline Corr As found:	17.17	Previous response	17.22	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 6, 2023	Previous Calibration:	October 19, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	13:24	End Time (MST):	16:48
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.09	----	Correlation Coefficient	0.999951	≥0.995
17.26	17.26	1.0002			
8.64	8.49	1.0180	Slope	1.004599	0.90 - 1.10
4.31	4.22	1.0219			
			Intercept	-0.117967	+/-1.5



THC Calibration Plot

Date: November 6, 2023

Location: Kirby South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	November 8, 2023	Last Cal Date:	October 4, 2023
Start time (MST):	11:17	End time (MST):	16:48
Reason:	Routine		

Calibration Standards

NO Gas Cylinder #:	T34ULGL	Cal Gas Expiry Date:	March 8, 2025
NOX Cal Gas Conc:	49.39 ppm	NO Cal Gas Conc:	49.02 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	49.39 ppm	Removed Gas NO Conc:	49.02 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	3804
ZAG make/model:	API 701H	Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.424	1.424	NO bkgnd or offset:	1.5 1.7
NOX coeff or slope:	0.997	0.997	NOX bkgnd or offset:	1.6 1.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	168.83 202.82

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001586	1.007732
NO _x Cal Offset:	-1.471182	-4.271621
NO Cal Slope:	1.001191	1.008462
NO Cal Offset:	-1.933293	-5.833654
NO ₂ Cal Slope:	0.995145	1.005705
NO ₂ Cal Offset:	0.363035	0.483774



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.5	0.0	----	----
as found span	4919	81.0	800.1	794.1	6.0	810.0	797.8	12.0	0.9878	0.9954
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
high point	4919	81.0	800.1	794.1	6.0	804.0	798.1	5.7	0.9952	0.9950
second point	4960	40.5	400.0	397.0	3.0	396.8	390.8	6.0	1.0081	1.0159
third point	4980	20.2	199.5	198.0	1.5	193.0	189.0	4.1	1.0338	1.0478
as left zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
as left span	4919	81.0	800.1	417.6	382.5	801.0	411.4	389.1	0.9989	1.0151
Average Correction Factor									1.0124	1.0196

Corrected As found	NO _x = 810.5 ppb	NO = 798.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.3%
Previous Response	NO _x = 799.9 ppb	NO = 793.1 ppb		*Percent Change	NO = 0.6%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	794.1	417.6	382.5	384.7	0.9943	100.6%
2nd GPT point (200 ppb O ₃)	794.1	634.3	165.8	168.3	0.9851	101.5%
3rd GPT point (100 ppb O ₃)	794.1	720.3	79.8	80.7	0.9888	101.1%
Average Correction Factor					0.9894	101.1%

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 4, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	11:17	End Time (MST):	16:48
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.4	----	Correlation Coefficient	≥0.995	
800.1	804.0	0.9952			
400.0	396.8	1.0081			
199.5	193.0	1.0338			
			Slope	1.007732	0.90 - 1.10
			Intercept	-4.271621	+/-20

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 4, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	11:17	End Time (MST):	16:48
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
794.1	798.1	0.9950		
397.0	390.8	1.0159		
198.0	189.0	1.0478		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 8, 2023	Previous Calibration:	October 4, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	11:17	End Time (MST):	16:48
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	0.999979	≥0.995
382.5	384.7	0.9943			
165.8	168.3	0.9851	Slope	1.005705	0.90 - 1.10
79.8	80.7	0.9888			
			Intercept	0.483774	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 8, 2023

Location: Kirby South






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS511
BLACKGOLD
NOVEMBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2023





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Blackgold	Station number:	AMS511
Calibration Date:	November 7, 2023	Last Cal Date:	October 5, 2023
Start time (MST):	9:54	End time (MST):	13:22
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.06	ppm	Cal Gas Exp Date:	January 5, 2029
Cal Gas Cylinder #:	CC147416			
Removed Cal Gas Conc:	50.06	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	5258
ZAG Make/Model:	Teledyne API 701		Serial Number:	138

Analyzer Information

Analyzer make: Thermo scientific Analyzer serial #: 1160290014
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001162	1.002724	Backgd or Offset:	31.8	34.9
Calibration intercept:	-1.503208	-1.426148	Coeff or Slope:	1.107	1.156

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.9	----
as found span	4926	80.0	800.0	766.1	1.044
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4926	80.0	800.0	801.0	0.999
second point	4968	40.1	400.8	401.0	1.000
third point	4987	20.2	202.0	199.1	1.014
as left zero	5000	0.0	0.0	0.1	----
as left span	4926	80.0	800.0	802.0	0.998
Average Correction Factor					1.004

Baseline Corr As found:	765.20	Previous response	799.43	*% change	-4.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 5, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	9:54	End Time (MST):	13:22
Analyzer make:	Thermo scientific	Analyzer serial #:	1160290014

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient	≥0.995
800.0	801.0	0.9988		
400.8	401.0	0.9996	Slope	0.90 - 1.10
202.0	199.1	1.0143		
			Intercept	+/-30



SO2 Calibration Plot

Date: November 7, 2023

Location: Blackgold





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Blackgold Station number: AMS511
 Calibration Date: November 9, 2023 Last Cal Date: October 5, 2023
 Start time (MST): 8:13 End time (MST): 14:17
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.139 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC511397
 Removed Cal Gas Conc: 5.139 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T750 Serial Number: 282
 ZAG Make/Model: API T751H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1336160090
 Converter make: Global G150 Converter serial #: 2022-227
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999059	1.006353	Backgd or Offset:	3.08
Calibration intercept:	-0.039168	-0.019423	Coeff or Slope:	1.137
				3.37
				1.159

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4922	77.8	80.0	80.2	1.001
as found 2nd point	4961	38.9	40.0	39.9	1.010
as found 3rd point	4981	19.5	20.0	20.3	1.002
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	77.8	80.0	80.5	0.993
second point	4961	38.9	40.0	40.1	0.997
third point	4981	19.5	20.0	20.2	0.992
as left zero	5000	0.0	0.0	0.0	----
as left span	4922	77.9	80.0	79.3	1.009
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.994
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 79.9 Prev response: 79.85 *% change: 0.1%
 Baseline Corr 2nd AF pt: 39.6 AF Slope: 0.998777 AF Intercept: 0.220716
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999975

* = > +/-5% change initiates investigation

Notes: Scrubber check done after cal zero. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	November 9, 2023	Previous Calibration:	October 5, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:13	End Time (MST):	14:17
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	≥0.995
80.0	80.5	0.9934		
40.0	40.1	0.9971	Slope	0.90 - 1.10
20.0	20.2	0.9921		
			Intercept	+/-3



H₂S Calibration Plot

Date: November 9, 2023

Location: Blackgold





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	November 7, 2023	Previous Calibration:	October 5, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	9:54	End Time (MST):	13:22
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1317958295

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.01	----	Correlation Coefficient	0.999918	
17.11	17.09	1.0014			≥0.995
8.57	8.43	1.0163	Slope	1.000266	
4.28	4.17	1.0282			0.90 - 1.10
			Intercept	-0.072772	+/-1.5



THC Calibration Plot

Date: November 7, 2023

Location: Blackgold





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Blackgold
Calibration Date: November 9, 2023
Start time (MST): 8:13
Reason: Routine
Station number: AMS511
Last Cal Date: October 5, 2023
End time (MST): 14:33

Calibration Standards

NO Gas Cylinder #: T0F8P52
NOX Cal Gas Conc: 47.4 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.4 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: August 16, 2026
NO Cal Gas Conc: 47.4 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.4 ppm
NO gas Diff:
Serial Number: 2445
Serial Number: 138

Analyzer Information

Analyzer make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 7029

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.029	1.041	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.029	1.038	NOX bkgnd or offset:	1.1	0.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.1	4.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.994686	1.004049
NO _x Cal Offset:	1.102483	-1.772798
NO Cal Slope:	0.983189	1.003178
NO Cal Offset:	2.228032	-2.292773
NO ₂ Cal Slope:	1.005880	1.000729
NO ₂ Cal Offset:	0.450694	0.157575



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.3	-0.3	----	----
as found span	4916	84.4	800.6	800.6	0.0	790.6	781.0	9.6	1.0126	1.0250
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4916	84.4	800.6	800.6	0.0	803.0	802.0	1.4	0.9970	0.9982
second point	4958	42.2	400.3	400.3	0.0	398.9	397.9	1.0	1.0035	1.0061
third point	4979	21.1	200.2	200.2	0.0	197.8	196.5	1.3	1.0119	1.0186
as left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	----	----
as left span	4916	84.4	800.6	404.3	396.3	798.8	403.5	395.3	1.0022	1.0019
Average Correction Factor									1.0041	1.0076

Corrected As found	NO _x = 791.1 ppb	NO = 781.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.8%	
Previous Response	NO _x = 797.4 ppb	NO = 789.3 ppb		*Percent Change	NO = -1.0%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.0	401.7	396.3	396.7	0.9990	100.1%
2nd GPT point (200 ppb O3)	798.0	631.2	166.8	167.1	0.9982	100.2%
3rd GPT point (100 ppb O3)	798.0	716.6	81.4	81.8	0.9951	100.5%
Average Correction Factor					0.9974	100.3%

Notes:

Adjusted zero and span. No other maintenance done.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	November 9, 2023	Previous Calibration:	October 5, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:13	End Time (MST):	14:33
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	0.999979	≥0.995
800.6	803.0	0.9970			
400.3	398.9	1.0035	Slope	1.004049	0.90 - 1.10
200.2	197.8	1.0119			
			Intercept	-1.772798	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 9, 2023	Previous Calibration:	October 5, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:13	End Time (MST):	14:33
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	0.999966	≥0.995
800.6	802.0	0.9982			
400.3	397.9	1.0061	Slope	1.003178	0.90 - 1.10
200.2	196.5	1.0186			
			Intercept	-2.292773	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	November 9, 2023	Previous Calibration:	October 5, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:13	End Time (MST):	14:33
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
396.3	396.7	0.9990			
166.8	167.1	0.9982			
81.4	81.8	0.9951			
			Correlation Coefficient	0.999999	≥0.995
			Slope	1.000729	0.90 - 1.10
			Intercept	0.157575	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: November 9, 2023

Location: Blackgold





End of Report