



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

DECEMBER 2023 MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

January 31, 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01
BERTHA GANTER - FORT MCKAY

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	December 13, 2023	Last Cal Date:	November 1, 2023
Start time (MST):	10:15	End time (MST):	13:16
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC418809			
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999167	0.998738	Backgd or Offset:	19.8	20.0
Calibration intercept:	-0.353461	-0.153291	Coeff or Slope:	0.887	0.887

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4918	81.3	800.3	794.7	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.5	----
high point	4918	81.3	800.3	799.6	1.001
second point	4959	40.7	400.6	399.1	1.004
third point	4979	20.3	199.8	199.1	1.004
as left zero	5000	0.0	0.0	0.5	----
as left span	4918	81.3	800.3	799.0	1.002
Average Correction Factor					1.003

Baseline Corr As found:	794.50	Previous response	799.25	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:15	End Time (MST):	13:16
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999996	≥0.995
800.3	799.6	1.0008			
400.6	399.1	1.0037	Slope	0.998738	0.90 - 1.10
199.8	199.1	1.0036			
			Intercept	-0.153291	+/-30



SO2 Calibration Plot

Date: November 1, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: December 12, 2023 Last Cal Date: November 6, 2023
 Start time (MST): 11:23 End time (MST): 17:24
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153461
 Converter make: CD Nova Converter serial #: 470
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997792	0.999507	Backgd or Offset: 2.28	2.25
Calibration intercept:	0.180000	0.219996	Coeff or Slope: 0.914	0.904

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4921	78.4	80.0	81.4	0.985
as found 2nd point	4960	39.2	40.0	40.9	0.983
as found 3rd point	4980	19.6	20.0	20.4	0.990
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	78.4	80.0	80.1	0.999
second point	4960	39.2	40.0	40.4	0.990
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	78.4	80.0	80.2	0.997
SO2 Scrubber Check	4919	81.3	813.0	0.1	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	0.995
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 81.2 Prev response: 80.00 *% change: 1.5%
 Baseline Corr 2nd AF pt: 40.7 AF Slope: 1.015651 AF Intercept: 0.179998
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999995

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 6, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:23	End Time (MST):	17:24
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999983	
80.0	80.1	0.9987			≥0.995
40.0	40.4	0.9901	Slope	0.999507	
20.0	20.1	0.9949			0.90 - 1.10
			Intercept	0.219996	+/-3



TRS Calibration Plot

Date: December 12, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: December 12, 2023 Last Cal Date: November 6, 2023
 Start time (MST): 11:23 End time (MST): 17:24
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1200326167
 Converter make: Thermo Converter Converter serial #: N/A
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995140	0.985566	Backgd or Offset:	1.73	1.72
Calibration intercept:	0.276804	0.536834	Coeff or Slope:	1.009	1.009

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4922	78.4	80.0	78.7	1.019
as found 2nd point	4960	39.2	40.0	40.0	1.005
as found 3rd point	4980	19.6	20.0	20.1	1.005
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.5	----
high point	4922	78.4	80.0	79.3	1.009
second point	4960	39.2	40.0	40.1	0.997
third point	4980	19.6	20.0	20.2	0.990
as left zero	5000	0.0	0.0	1.1	----
as left span	4922	78.4	80.0	76.7	1.043
SO2 Scrubber Check	4919	81.3	813.0	-0.1	----
Date of last scrubber change:	March 21, 2022			Ave Corr Factor	0.999
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 78.5 Prev response: 79.85 *% change: -1.7%
 Baseline Corr 2nd AF pt: 39.8 AF Slope: 0.981329 AF Intercept: 0.416826
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999940

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 6, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:23	End Time (MST):	17:24
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999992	≥0.995
80.0	79.3	1.0085			
40.0	40.1	0.9975	Slope	0.985566	0.90 - 1.10
20.0	20.2	0.9900			
			Intercept	0.536834	+/-3



H₂S Calibration Plot

Date: December 12, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	December 13, 2023	Last Cal Date:	November 1, 2023
Start time (MST):	10:15	End time (MST):	13:16
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.01E-04	3.04E-04	NMHC SP Ratio:	6.84E-05	7.01E-05
CH ₄ Retention time:	14.6	14.6	NMHC Peak Area:	134289	131151
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.27	17.14	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	17.27	17.22	1.003
second point	4959	40.7	8.64	8.46	1.021
third point	4980	20.3	4.31	4.21	1.024
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.27	17.24	1.001
Average Correction Factor					1.016

Baseline Corr AF:	17.14	Prev response	17.19	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4918	81.3	9.18	9.08	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	9.18	9.13	1.006
second point	4959	40.7	4.60	4.55	1.011
third point	4980	20.3	2.29	2.28	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	9.18	9.17	1.001
Average Correction Factor					1.007
Baseline Corr AF:	9.08	Prev response	9.15	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	8.05	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	8.09	8.09	1.000
second point	4959	40.7	4.05	3.91	1.034
third point	4980	20.3	2.02	1.93	1.045
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.07	1.002
Average Correction Factor					1.026
Baseline Corr AF:	8.05	Prev response	8.05	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000035	0.997839
THC Cal Offset:	-0.073775	-0.066185
CH ₄ Cal Slope:	1.002522	1.002253
CH ₄ Cal Offset:	-0.057596	-0.062393
NMHC Cal Slope:	0.998093	0.994150
NMHC Cal Offset:	-0.016179	-0.004590

Notes: Changed out the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:15	End Time (MST):	13:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999896	≥ 0.995			
17.27	17.22	1.0029						
8.64	8.46	1.0212				Slope	0.997839	0.90 - 1.10
4.31	4.21	1.0239						
			Intercept	-0.066185	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:15	End Time (MST):	13:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999625	≥ 0.995			
8.09	8.09	0.9996						
4.05	3.91	1.0341				Slope	1.002253	0.90 - 1.10
2.02	1.93	1.0454						
			Intercept	-0.062393	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:15	End Time (MST):	13:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	≥ 0.995			
9.18	9.13	1.0056						
4.60	4.55	1.0105				Slope	0.994150	0.90 - 1.10
2.29	2.28	1.0058						
			Intercept	-0.004590	± 0.5			



NMHC Calibration Plot

Date: December 13, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
Calibration Date: December 4, 2023 Last Cal Date: November 3, 2023
Start time (MST): 10:26 End time (MST): 14:50
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1P9L Cal Gas Expiry Date: December 11, 2023
NOX Cal Gas Conc: 50.84 ppm NO Cal Gas Conc: 50.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.84 ppm Removed Gas NO Conc: 50.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3565
ZAG make/model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1218153357
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.411	1.431	NO bkgnd or offset:	7.2	7.3
NOX coeff or slope:	0.991	0.989	NOX bkgnd or offset:	7.9	8.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	194.2	194.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999719	1.000323
NO _x Cal Offset:	-0.780000	-0.720000
NO Cal Slope:	0.998701	1.001285
NO Cal Offset:	-1.300000	-1.280000
NO ₂ Cal Slope:	1.003579	1.001111
NO ₂ Cal Offset:	0.884749	-0.531962



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
as found span	4920	80.0	813.4	800.6	12.8	804.7	788.7	16.1	1.0109	1.0151
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
high point	4920	80.0	813.4	800.6	12.8	813.4	801.1	12.3	1.0000	0.9994
second point	4960	40.0	406.7	400.3	6.4	405.7	398.8	6.9	1.0025	1.0038
third point	4980	20.0	203.4	200.2	3.2	201.9	197.8	4.1	1.0072	1.0119
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as left span	4920	80.0	813.4	396.5	416.9	810.9	393.8	417.0	1.0031	1.0070
Average Correction Factor									1.0033	1.0051

Corrected As found	NO _x = 804.6 ppb	NO = 788.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.0%	
Previous Response	NO _x = 812.4 ppb	NO = 798.3 ppb		*Percent Change	NO = -1.2%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.2	393.1	416.9	417.1	0.9995	100.0%
2nd GPT point (200 ppb O3)	797.2	596.1	213.9	213.4	1.0023	99.8%
3rd GPT point (100 ppb O3)	797.2	695.6	114.4	113.4	1.0088	99.1%
Average Correction Factor					1.0036	99.6%

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 3, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:26	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.4	813.4	1.0000		
406.7	405.7	1.0025		
203.4	201.9	1.0072		
			0.999995	
			1.000323	
			-0.720000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 3, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:26	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.6	801.1	0.9994			
400.3	398.8	1.0038			
200.2	197.8	1.0119			
			Slope	1.001285	0.90 - 1.10
			Intercept	-1.280000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 3, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:26	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
416.9	417.1	0.9995			
213.9	213.4	1.0023			
114.4	113.4	1.0088			
			Slope	1.001111	0.90 - 1.10
			Intercept	-0.531962	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 4, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: December 1, 2023 Last Cal Date: November 2, 2023
 Start time (MST): 11:07 End time (MST): 13:41
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 1107
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000714	1.004914	Backgd or Offset:	3.8	3.8
Calibration intercept:	0.200000	0.040000	Coeff or Slope:	1.010	1.012

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	5000	863.1	400.0	400.5	0.999
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	5000	863.1	400.0	401.9	0.995
second point	5000	742.5	200.0	201.2	0.994
third point	5000	651.7	100.0	100.6	0.994
as left zero	5000	0.0	0.0	0.0	----
as left span	5000	863.1	400.0	405.3	0.987
Average Correction Factor					0.994

Baseline Corr As found:	400.7	Previous response	400.5	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 2, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:07	End Time (MST):	13:41
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999999	≥0.995
400.0	401.9	0.9953			
200.0	201.2	0.9940	Slope	1.004914	0.90 - 1.10
100.0	100.6	0.9940			
			Intercept	0.040000	+/- 5



O₃ Calibration Plot

Date: December 1, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
 Calibration Date: December 12, 2023 Last Cal Date: November 14, 2023
 Start time (MST): 13:14 End time (MST): 14:09

Analyzer Make: Teledyne API T640 S/N: 324
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-7.7	-8.4	-7.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	734.1	729.2	734.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	4.96	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 12, 2023</u>		Last Cal Date: <u>November 14, 2023</u>		
	PM w/o HEPA: <u>7.8</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>October 19, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>October 19, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: September 14, 2023
 Date RH/T Sensor Cleaned: October 19, 2023

Notes: Flow, temperature, and pressure all within limits. Leak check passed.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	December 11, 2023	Last Cal Date:	November 8, 2023
Start time (MST):	10:16	End time (MST):	13:07
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	<u>3040</u>	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>3040</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001848	1.001201	Backgd or Offset:	-0.013	-0.013
Calibration intercept:	0.149849	0.203827	Coeff or Slope:	0.990	0.989

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.6	40.8	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4933	66.7	40.6	40.7	0.997
second point	4966	33.3	20.2	20.8	0.976
third point	4983	16.7	10.2	10.4	0.978
as left zero	5000	0.0	0.0	0.1	----
as left span	2960	40.0	40.5	40.2	1.008
Average Correction Factor					0.984

Baseline Corr As found:	40.73	Prev response:	40.78	*% change:	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 11, 2023	Previous Calibration:	November 8, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:16	End Time (MST):	13:07
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999873	
40.6	40.7	0.9972			≥0.995
20.2	20.8	0.9759	Slope	1.001201	
10.2	10.4	0.9782			0.90 - 1.10
			Intercept	0.203827	+/-1.5



CO Calibration Plot

Date: December 11, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	December 8, 2023	Last Cal Date:	November 7, 2023
Start time (MST):	10:50	End time (MST):	14:09
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>60,200</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
N2 Gen Make/Model:	Peak Sci NG5000		Serial Number:	7220900034

Analyzer Information

Analyzer make: Teledyne API 360	Analyzer serial #: 442
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001056	1.000878	Backgd or Offset:	0.045	0.045
Calibration intercept:	-6.100000	-5.800000	Coeff or Slope:	0.874	0.875

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-0.2	----
as found span	2920	80.0	1605.3	1593.9	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.0	----
high point	2920	80.0	1605.3	1604.0	1.001
second point	2960	40.0	802.7	794.2	1.011
third point	2980	20.0	401.3	390.4	1.028
as left zero	3000	0.0	0.0	-0.1	----
as left span	2960	40.0	802.7	783.6	1.024
Average Correction Factor					1.013

Baseline Corr As found:	1594.10	Prev response:	1600.93	*% change:	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	December 8, 2023	Previous Calibration	November 7, 2023
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS01
Start Time (MST)	10:50	End Time (MST)	14:09
Analyzer make	Teledyne API 360	Analyzer serial #	442

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999940	≥0.995
1605.3	1604.0	1.0008			
802.7	794.2	1.0107	Slope	1.000878	0.90 - 1.10
401.3	390.4	1.0280			
			Intercept	-5.800000	+/-10



CO₂ Calibration Plot

Date: December 8, 2023

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	December 6, 2023	Last Cal Date:	November 14, 2023
Start time (MST):	11:12	End time (MST):	15:35
NH3 Cal Date:	December 7, 2023	Last Cal Date:	November 15, 2023
Start time (MST):	11:30	End time (MST):	14:36
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	50.84	ppm	NO Gas Cylinder #:	T2Y1P9L
NO Cal Gas Conc:	50.04	ppm	NO Cal Gas Expiry:	December 11, 2023
Removed NOX Conc:	50.84	ppm	Removed Cylinder #:	NA
Removed NO Conc:	50.04	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.58	ppm	NH3 Gas Cylinder #:	CC743587
			NH3 Cal Gas Expiry:	August 22, 2024
Removed NH3 Conc:	76.58	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	Teledyne API T700		Serial Number:	3565
ZAG make/model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	475
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	5.20
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	534

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.987	0.987	TN coefficient:	0.989	0.989
NOX coefficient:	0.989	0.989	NO bkgnd:	-0.9	-0.9
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-0.3	-0.3
NH3 coefficient:	0.932	0.932	TN bkgnd:	1.2	1.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999888	1.004707
NO _x Cal Offset:	-1.540000	-2.480000
NO Cal Slope:	0.999686	1.004525
NO Cal Offset:	-2.120000	-3.440000
NO ₂ Cal Slope:	0.998244	0.995838
NO ₂ Cal Offset:	-0.546307	-1.454892
NH3 Cal Slope:	1.000230	1.001746
NH3 Cal Offset:	-2.765159	-2.214411
TN Cal Slope:	1.002817	1.004330
TN Cal Offset:	-2.497861	-1.994595



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
as found NO	4920	80.0	813.4	813.4	----	815.9	818.0	-2.1	0.997	----
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
high NO point	4920	80.0	813.4	813.4	----	816.9	816.7	0.4	0.996	----
NO/O3 point	4920	80.0	813.4	813.4	----	813.1	812.2	1.0	1.000	----
as found NH3	3416	84.1	1840.1	----	1840.1	1808.0	----	1803.1	1.018	1.021
new NH3 cyl rp							----			
first NH3	3418	82.2	1798.5	----	1798.5	1808.0	----	1803.1	0.995	0.997
second NH3	3454	45.7	1000.0	----	1000.0	994.9	----	992.3	1.005	1.008
third NH3	3477	22.8	498.9	----	498.9	500.7	----	498.9	0.996	1.000
Average Correction Factor									0.9981	1.0017

Corrected As found TN = 816.2 ppb NO_x = 818.3 ppb NH3 = 1803.1 ppb

Previous Response TN = 813.2 ppb NO_x = 811.8 ppb NH3 = 1837.8 ppb

NH3 Previous Converter Efficiency = 93.2%

NH3 Current Converter Efficiency = 93.2%

*Percent Change TN = 0.4%

*Percent Change NO_x = 0.8%

*Percent Change NH3 = -1.9%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.8	----	----
as found span	4920	80.0	813.4	800.6	813.4	816.4	799.4	816.4	0.9964	1.0016
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.4	0.1	----	----
high point	4920	80.0	813.4	800.6	813.4	816.7	803.4	816.9	0.9960	0.9966
second point	4960	40.0	406.7	400.3	406.7	403.0	394.8	403.7	1.0092	1.0140
third point	4980	20.0	203.4	200.2	203.4	200.5	195.1	200.5	1.0143	1.0259
Average Correction Factor									1.0065	1.0122

Baseline Corr As fnd	TN = 817.2 ppb	NO _x = 816.6 ppb	NO = 799.4 ppb	*Percent Change	TN = 0.5%
Previous Response	TN = 813.2 ppb	NO _x = 811.8 ppb	NO = 798.3 ppb	*Percent Change	NO _x = 0.6%
				*Percent Change	NO = 0.1%
				<i>* = > +/-5% change initiates investigation</i>	

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	-0.2	----	----
calibration zero	----	----	0.0	-0.3	----	----
1st GPT point (400 ppb O3)	798.6	386.3	425.1	422.6	1.0059	99.4%
2nd GPT point (200 ppb O3)	798.6	591.9	219.5	216.3	1.0148	98.5%
3rd GPT point (100 ppb O3)	798.6	691.8	119.6	116.6	1.0257	97.5%
Average Correction Factor					1.0155	98.5%

Notes:

Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By:

Rene Chamberland



Wood Buffalo Environmental Association

TN Calibration Summary


Version-05-2023

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 14, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:12	End Time (MST):	15:35
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
1798.5	1808.0	0.9948			
1000.0	994.9	1.0051			
498.9	500.7	0.9963			
			Slope	1.004330	0.90 - 1.10
			Intercept	-1.994595	+/-20





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 14, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:12	End Time (MST):	15:35
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
1798.5	1803.1	0.9975			
1000.0	992.3	1.0078			
498.9	498.9	0.9999			
			Slope	1.001746	0.90 - 1.10
			Intercept	-2.214411	+/-20





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 14, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:12	End Time (MST):	15:35
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.4	816.7	0.9960		
406.7	403.0	1.0092		
203.4	200.5	1.0143		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 14, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:12	End Time (MST):	15:35
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.4	----	Correlation Coefficient	0.999879	≥0.995
800.6	803.4	0.9966			
400.3	394.8	1.0140	Slope	1.004525	0.90 - 1.10
200.2	195.1	1.0259			
			Intercept	-3.440000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 14, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:12	End Time (MST):	15:35
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
425.1	422.6	1.0059		
219.5	216.3	1.0148		
119.6	116.6	1.0257		



NO_x Calibration Plot

Date: December 6, 2023


Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: December 7, 2023

Location: Bertha Ganter-Fort McKay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	December 8, 2023	Last Cal Date:	November 23, 2023
Start time (MST):	10:14	End time (MST):	14:47
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1185
ZAG Make/Model:	API T701		Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996481	1.000671	Backgd or Offset:	20.6	18.6
Calibration intercept:	-0.365997	-0.784833	Coeff or Slope:	0.763	0.775

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-2.4	----
as found span	4920	80.2	801.6	779.6	1.028
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	80.2	801.6	802.0	1.000
second point	4960	40.1	400.8	399.2	1.004
third point	4980	20.0	199.9	199.1	1.004
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	80.2	801.6	800.0	1.002
Average Correction Factor					1.003

Baseline Corr As found:	782.00	Previous response	798.46	*% change	-2.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 23, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:14	End Time (MST):	14:47
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999994	≥0.995
801.6	802.0	0.9996			
400.8	399.2	1.0041	Slope	1.000671	0.90 - 1.10
199.9	199.1	1.0041			
			Intercept	-0.784833	+/-30



SO2 Calibration Plot

Date: December 8, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	December 18, 2023	Last Cal Date:	November 24, 2023
Start time (MST):	10:53	End time (MST):	15:46
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.29	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC345191			
Removed Cal Gas Conc:	5.29	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1185
ZAG Make/Model:	API T701		Serial Number:	5608

Analyzer Information

Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966
Converter make:	Global G150	Converter serial #:	2022-198
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010680	1.008965	Backgd or Offset:	1.72	1.70
Calibration intercept:	-0.019191	-0.059195	Coeff or Slope:	0.754	0.740

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4924	75.6	80.0	82.2	0.974
as found 2nd point	4962	37.8	40.0	40.9	0.980
as found 3rd point	4981	18.9	20.0	20.3	0.990
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4924	75.6	80.0	80.7	0.991
second point	4962	37.8	40.0	40.3	0.992
third point	4981	18.9	20.0	19.9	1.005
as left zero	5000	0.0	0.0	0.1	----
as left span	4924	75.6	80.0	80.3	0.996
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:	20-Sep-23			Ave Corr Factor	0.996
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	82.1	Prev response:	80.83	*% change:	1.6%
Baseline Corr 2nd AF pt:	40.8	AF Slope:	1.027396	AF Intercept:	-0.079182
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999978		

* = > +/-5% change initiates investigation

Notes: Scrubber check done after cal zero, passed. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 18, 2023	Previous Calibration:	November 24, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:53	End Time (MST):	15:46
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999979	≥0.995
80.0	80.7	0.9912			
40.0	40.3	0.9924	Slope	1.008965	0.90 - 1.10
20.0	19.9	1.0049			
			Intercept	-0.059195	+/-3



H₂S Calibration Plot

Date: December 18, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	December 8, 2023	Last Cal Date:	November 22, 2023
Start time (MST):	10:14	End time (MST):	14:47
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
C ₃ H ₈ Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
Removed C ₃ H ₈ Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.07E-04	3.11E-04	NMHC SP Ratio:	4.52E-05	4.57E-05
CH ₄ Retention time:	15.0	15.2	NMHC Peak Area:	194477	192457
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	16.65	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.81	1.000
second point	4960	40.1	8.41	8.38	1.003
third point	4980	20.0	4.19	4.17	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.83	0.999
Average Correction Factor					1.003

Baseline Corr AF:	16.65	Prev response	16.78	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	8.73	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.80	0.999
second point	4960	40.1	4.40	4.39	1.001
third point	4980	20.0	2.19	2.19	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.81	0.998
Average Correction Factor					1.000
Baseline Corr AF:	8.73	Prev response	8.79	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	7.92	1.014
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.01	1.002
second point	4960	40.1	4.01	3.98	1.007
third point	4980	20.0	2.00	1.98	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	8.02	1.000
Average Correction Factor					1.007
Baseline Corr AF:	7.92	Prev response	7.98	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998545	0.999782
THC Cal Offset:	-0.015916	-0.013512
CH ₄ Cal Slope:	0.997280	0.998946
CH ₄ Cal Offset:	-0.020258	-0.012854
NMHC Cal Slope:	0.999088	1.000492
NMHC Cal Offset:	0.003941	-0.001458

Notes: Changed inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 22, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:14	End Time (MST):	14:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995
16.82	16.81	1.0005			
8.41	8.38	1.0035			
4.19	4.17	1.0056			
			Slope	0.999782	0.90 - 1.10
			Intercept	-0.013512	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 22, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:14	End Time (MST):	14:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999987	≥ 0.995			
8.02	8.01	1.0016						
4.01	3.98	1.0074				Slope	0.998946	0.90 - 1.10
2.00	1.98	1.0120						
			Intercept	-0.012854	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 22, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:14	End Time (MST):	14:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
8.80	8.80	0.9995						
4.40	4.39	1.0008				Slope	1.000492	0.90 - 1.10
2.19	2.19	0.9997						
			Intercept	-0.001458	± 0.5			



NMHC Calibration Plot

Date: December 8, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	December 21, 2023	Last Cal Date:	December 8, 2023
Start time (MST):	13:11	End time (MST):	15:46
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
C ₃ H ₈ Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
Removed C ₃ H ₈ Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:		2.69E-04	NMHC SP Ratio:	6.67E-05
CH ₄ Retention time:		15.7	NMHC Peak Area:	131964
Zero Chromatogram:		OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.81	1.001
second point	4960	40.1	8.41	8.44	0.997
third point	4980	20.0	4.19	4.29	0.977
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.98	0.990
Average Correction Factor					0.992

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.78	1.001
second point	4960	40.1	4.40	4.42	0.995
third point	4980	20.0	2.19	2.27	0.968
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	9.00	0.978
Average Correction Factor					0.988
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.02	1.000
second point	4960	40.1	4.01	4.02	0.998
third point	4980	20.0	2.00	2.03	0.988
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	7.99	1.005
Average Correction Factor					0.995
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		0.997589
THC Cal Offset:		0.046113
CH ₄ Cal Slope:		0.998643
CH ₄ Cal Offset:		0.011958
NMHC Cal Slope:		0.996030
NMHC Cal Offset:		0.035954

Notes:

Install Calibration. Adjusted span.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 21, 2023	Previous Calibration:	December 8, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	13:11	End Time (MST):	15:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999961	≥ 0.995
16.82	16.81	1.0006			
8.41	8.44	0.9968			
4.19	4.29	0.9774			
			Slope	0.997589	0.90 - 1.10
			Intercept	0.046113	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 21, 2023	Previous Calibration:	December 8, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	13:11	End Time (MST):	15:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999988	≥ 0.995
8.02	8.02	1.0004			
4.01	4.02	0.9984			
2.00	2.03	0.9876			
			Slope	0.998643	0.90 - 1.10
			Intercept	0.011958	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 21, 2023	Previous Calibration:	December 8, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	13:11	End Time (MST):	15:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999916	≥ 0.995			
8.80	8.78	1.0014						
4.40	4.42	0.9950				Slope	0.996030	0.90 - 1.10
2.19	2.27	0.9680						
			Intercept	0.035954	± 0.5			



NMHC Calibration Plot

Date: December 21, 2023

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	December 23, 2023	Last Cal Date:	December 21, 2023
Start time (MST):	12:26	End time (MST):	15:13
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
C ₃ H ₈ Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
Removed C ₃ H ₈ Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.69E-04	2.63E-04	NMHC SP Ratio:	6.67E-05	6.75E-05
CH ₄ Retention time:	15.7	15.5	NMHC Peak Area:	131964	130400
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	17.08	0.985
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.82	1.000
second point	4960	40.1	8.41	8.46	0.994
third point	4980	20.0	4.19	4.22	0.994
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	17.07	0.985
Average Correction Factor					0.996

Baseline Corr AF:	17.08	Prev response	16.82	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	8.91	0.987
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.81	0.998
second point	4960	40.1	4.40	4.46	0.985
third point	4980	20.0	2.19	2.23	0.983
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	9.08	0.969
Average Correction Factor					0.989
Baseline Corr AF:	8.91	Prev response	8.80	*% change	1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	8.16	0.983
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.01	1.002
second point	4960	40.1	4.01	4.00	1.003
third point	4980	20.0	2.00	1.99	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	7.99	1.004
Average Correction Factor					1.004
Baseline Corr AF:	8.16	Prev response	8.02	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997589	1.000046
THC Cal Offset:	0.046113	0.020295
CH ₄ Cal Slope:	0.998643	0.998148
CH ₄ Cal Offset:	0.011958	-0.004054
NMHC Cal Slope:	0.996030	1.001244
NMHC Cal Offset:	0.035954	0.025148

Notes: Changed sample inlet filters after as founds. Adjusted span.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 23, 2023	Previous Calibration:	December 21, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	12:26	End Time (MST):	15:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999988	≥ 0.995			
16.82	16.82	0.9998						
8.41	8.46	0.9935				Slope	1.000046	0.90 - 1.10
4.19	4.22	0.9941						
			Intercept	0.020295	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 23, 2023	Previous Calibration:	December 21, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	12:26	End Time (MST):	15:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
8.02	8.01	1.0022						
4.01	4.00	1.0026				Slope	0.998148	0.90 - 1.10
2.00	1.99	1.0069				Intercept	-0.004054	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 23, 2023	Previous Calibration:	December 21, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	12:26	End Time (MST):	15:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999943	≥ 0.995			
8.80	8.81	0.9982						
4.40	4.46	0.9851				Slope	1.001244	0.90 - 1.10
2.19	2.23	0.9831						
			Intercept	0.025148	± 0.5			



NMHC Calibration Plot

Date: December 23, 2023

Location: Mildred Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 14, 2023	Previous Calibration:	November 28, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:35	End Time (MST):	10:28
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999992	≥0.995
799.7	798.5	1.0016			
399.8	401.1	0.9968	Slope	0.998629	0.90 - 1.10
199.4	199.1	1.0016			
			Intercept	0.434228	+/-30



SO2 Calibration Plot

Date: December 14, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: December 8, 2023 Last Cal Date: November 1, 2023
 Start time (MST): 8:50 End time (MST): 12:40
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.42 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345266
 Removed Cal Gas Conc: 5.42 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701H Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1008841400
 Converter make: Global Converter serial #: 2022-200
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998493	1.000346	Backgd or Offset:	1.8	1.8
Calibration intercept:	0.082213	0.042159	Coeff or Slope:	1.115	1.115

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4926	74.1	80.3	80.9	0.993
as found 2nd point	4963	37.0	40.1	40.5	0.990
as found 3rd point	4982	18.5	20.1	19.9	1.008
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	74.1	80.3	80.4	0.999
second point	4963	37.0	40.1	40.2	0.998
third point	4982	18.5	20.1	20.0	1.003
as left zero	5000	0.0	0.0	0.1	----
as left span	4926	74.1	80.3	79.9	1.005
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	1.000
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 80.9 Prev response: 80.28 *% change: 0.8%
 Baseline Corr 2nd AF pt: 40.5 AF Slope: 1.008597 AF Intercept: -0.097636
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999977

* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrator zero. No adjustments or maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 1, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:50	End Time (MST):	12:40
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999996	≥0.995
80.3	80.4	0.9990			
40.1	40.2	0.9977	Slope	1.000346	0.90 - 1.10
20.1	20.0	1.0026			
			Intercept	0.042159	+/-3



H₂S Calibration Plot

Date: December 8, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: December 8, 2023 Last Cal Date: November 28, 2023
 Start time (MST): 7:35 End time (MST): 8:51
 Reason: Cylinder Change Hydrogen and Nitrogen Cylinder Change

Calibration Standards

Gas Cert Reference: NA Cal Gas Expiry Date: March 10, 1931
 CH₄ Cal Gas Conc. 497.2 ppm CH₄ Equiv Conc. 1058.2 ppm
 C₃H₈ Cal Gas Conc. 204.0 ppm
 Removed Gas Cert: NA Removed Gas Expiry:
 Removed CH₄ Conc. 497.2 ppm CH₄ Equiv Conc. 1058.2 ppm
 Removed C₃H₈ Conc. 204.0 ppm Diff between cyl (THC):
 Diff between cyl (CH₄): Diff between cyl (NM):
 Calibrator Model: API T700 Serial Number: 3808
 ZAG make/model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 1222762077
 THC Range (ppm): 0 - 20 ppm
 NMHC Range (ppm): 0 - 10 ppm CH₄ Range (ppm): 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.00E-04	2.00E-04	NMHC SP Ratio:	4.38E-05
CH ₄ Retention time:	11.8	11.8	NMHC Peak Area:	201200
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	16.64	16.46	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	16.64	16.34	1.018
second point					
third point					
as left zero					
as left span					
Average Correction Factor					1.018

Baseline Corr AF: 16.46 Prev response: 16.58 *% change: -0.7%
 Baseline Corr 2nd AF: NA AF Slope: AF Intercept:
 Baseline Corr 3rd AF: NA AF Correlation: * = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	8.82	8.76	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	8.82	8.71	1.013
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.013
Baseline Corr AF:	8.76	Prev response	8.79	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	7.82	7.69	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	7.82	7.63	1.024
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.024
Baseline Corr AF:	7.69	Prev response	7.79	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999538	0.982193
THC Cal Offset:	-0.049557	0.000000
CH ₄ Cal Slope:	1.000446	0.976127
CH ₄ Cal Offset:	-0.029913	0.000000
NMHC Cal Slope:	0.998967	0.987570
NMHC Cal Offset:	-0.020044	0.000000


Notes: Hydrogen and Nitrogen Cylinder Changed.

Calibration Performed By: Melissa Lemay

NMHC Calibration Plot

Date: December 8, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	December 14, 2023	Last Cal Date:	November 28, 2023
Start time (MST):	7:35	End time (MST):	10:27
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	NA	Cal Gas Expiry Date:	March 10, 1931
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.00E-04	2.03E-04	NMHC SP Ratio:	4.38E-05
CH ₄ Retention time:	11.8	11.8	NMHC Peak Area:	201200
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON
				198327
				ON

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	16.64	16.37	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	16.64	16.63	1.000
second point	4961	39.3	8.32	8.23	1.011
third point	4980	19.6	4.15	4.05	1.024
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	16.64	16.61	1.002
Average Correction Factor					1.012

Baseline Corr AF:	16.37	Prev response	16.58	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	8.82	8.68	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	8.82	8.81	1.001
second point	4961	39.3	4.41	4.38	1.007
third point	4980	19.6	2.20	2.16	1.018
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	8.82	8.79	1.003
Average Correction Factor					1.009
Baseline Corr AF:	8.68	Prev response	8.79	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	7.82	7.69	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	7.82	7.82	1.000
second point	4961	39.3	3.91	3.85	1.015
third point	4980	19.6	1.95	1.89	1.031
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	7.82	7.82	1.000
Average Correction Factor					1.015
Baseline Corr AF:	7.69	Prev response	7.79	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999538	1.001050
THC Cal Offset:	-0.049557	-0.055556
CH ₄ Cal Slope:	1.000446	1.002202
CH ₄ Cal Offset:	-0.029913	-0.035913
NMHC Cal Slope:	0.998967	1.000029
NMHC Cal Offset:	-0.020044	-0.019643

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 14, 2023	Previous Calibration:	November 28, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:35	End Time (MST):	10:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999948	≥ 0.995			
16.64	16.63	1.0004						
8.32	8.23	1.0106				Slope	1.001050	0.90 - 1.10
4.15	4.05	1.0243						
			Intercept	-0.055556	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 14, 2023	Previous Calibration:	November 28, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:35	End Time (MST):	10:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999897	≥ 0.995			
7.82	7.82	0.9996						
3.91	3.85	1.0150				Slope	1.002202	0.90 - 1.10
1.95	1.89	1.0313						
			Intercept	-0.035913	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 14, 2023	Previous Calibration:	November 28, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:35	End Time (MST):	10:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999977	≥ 0.995			
8.82	8.81	1.0011						
4.41	4.38	1.0067				Slope	1.000029	0.90 - 1.10
2.20	2.16	1.0182						
			Intercept	-0.019643	± 0.5			



NMHC Calibration Plot

Date: December 14, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: December 11, 2023
Start time (MST): 8:07
Reason: Routine
Station number: AMS04
Last Cal Date: November 6, 2023
End time (MST): 13:11

Calibration Standards

NO Gas Cylinder #: CC324979
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: T36RH1F
Removed Gas NOX Conc: 51.16 ppm
NOX gas Diff: 6.3%
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: November 3, 2032
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: August 18, 2023
Removed Gas NO Conc: 50.91 ppm
NO gas Diff: 5.7%
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 721

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.185	1.163	NO bkgnd or offset:	-0.6	-0.6
NOX coeff or slope:	1.181	1.148	NOX bkgnd or offset:	-0.3	-0.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.4	4.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001677	1.000893
NO _x Cal Offset:	-0.753866	-0.273094
NO Cal Slope:	0.998936	1.000600
NO Cal Offset:	-1.613678	-1.154426
NO ₂ Cal Slope:	0.999002	0.986885
NO ₂ Cal Offset:	0.854730	-0.475922



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.3	0.2	0.1	----	----
as found span	4922	78.1	799.1	795.2	3.9	778.3	769.4	9.1	1.0267	1.0335
as found 2nd										
as found 3rd										
new cyl resp	4918	81.8	800.0	798.4	1.6	830.0	817.5	12.5	0.9639	0.9766
calibrator zero	5000	0.0	0.0	0.0	0.0	0.5	0.9	-0.3	----	----
high point	4918	81.8	800.0	798.4	1.6	801.2	798.9	2.4	0.9985	0.9994
second point	4959	40.9	400.0	399.2	0.8	398.7	396.8	1.9	1.0033	1.0060
third point	4980	20.4	199.5	199.1	0.4	199.3	196.3	3.0	1.0010	1.0142
as left zero	5000	0.0	0.0	0.0	0.0	0.7	1.1	-0.4	----	----
as left span	4918	81.8	800.0	398.7	401.3	797.8	396.7	400.9	1.0028	1.0050
Average Correction Factor									1.0009	1.0065

Corrected As found	NO _x = 778.0 ppb	NO = 769.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -2.8%
Previous Response	NO _x = 799.7 ppb	NO = 792.7 ppb		*Percent Change	NO = -3.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.6	396.9	401.3	396.1	1.0132	98.7%
2nd GPT point (200 ppb O3)	796.6	597.1	201.1	196.7	1.0226	97.8%
3rd GPT point (100 ppb O3)	796.6	698.9	99.3	98.2	1.0116	98.9%
Average Correction Factor					1.0158	98.4%

Notes:

Nox calibration Gas cylinder Changed. Span adjusted.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 11, 2023	Previous Calibration:	November 6, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:07	End Time (MST):	13:11
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.5	----	Correlation Coefficient	≥0.995
800.0	801.2	0.9985		
400.0	398.7	1.0033	Slope	0.90 - 1.10
199.5	199.3	1.0010		
			Intercept	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 11, 2023	Previous Calibration:	November 6, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:07	End Time (MST):	13:11
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.9	----	Correlation Coefficient	≥0.995	
798.4	798.9	0.9994			
399.2	396.8	1.0060			
199.1	196.3	1.0142			
			Slope	1.000600	0.90 - 1.10
			Intercept	-1.154426	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 11, 2023	Previous Calibration:	November 6, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:07	End Time (MST):	13:11
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
401.3	396.1	1.0132			
201.1	196.7	1.0226			
99.3	98.2	1.0116			
			Slope	0.986885	0.90 - 1.10
			Intercept	-0.475922	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 11, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: December 12, 2023 Last Cal Date: November 22, 2023
 Start time (MST): 9:00 End time (MST): 10:21
 Reason: Removal

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 7045
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005543		Backgd or Offset:	-1.4	-1.4
Calibration intercept:	0.680000		Coeff or Slope:	1.022	1.022

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.8	----
as found span	5000	982.6	400.0	408.7	0.979
as found 2nd point	5000	813.3	200.0	204.9	0.976
as found 3rd point	5000	701.4	100.0	104.2	0.960
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

Average Correction Factor					
Baseline Corr As found:	407.9	Previous response	402.9	*% change	1.2%
Baseline Corr 2nd AF pt:	-203.8	AF Slope:	1.018571	AF Intercept:	1.400000
Baseline Corr 3rd AF pt:	-100.7	AF Correlation:	0.999986		

* = > +/-5% change initiates investigation


Notes: Removed to put original analyzer back in.

Calibration Performed By: Melissa Lemay

O₃ Calibration Plot

Date: December 12, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: December 12, 2023 Last Cal Date:
 Start time (MST): 10:25 End time (MST): 12:27
 Reason: Install

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 2961
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		1.001229	Backgd or Offset:		-3.5
Calibration intercept:		0.760000	Coeff or Slope:		1.011

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.2	----
high point	5000	985.4	400.0	401.0	0.998
second point	5000	813.9	200.0	201.2	0.994
third point	5000	703.2	100.0	101.5	0.985
as left zero	5000	0.0	0.0	1.4	----
as left span	5000	985.1	400.0	402.0	0.995
Average Correction Factor					0.992

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Install of original O3 analyzer back into station. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	10:25	End Time (MST):	12:27
Analyzer make:	API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999991	
400.0	401.0	0.9975			≥0.995
200.0	201.2	0.9940	Slope	1.001229	
100.0	101.5	0.9852			0.90 - 1.10
			Intercept	0.760000	+/- 5



O₃ Calibration Plot

Date: December 12, 2023

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: December 14, 2023 Last Cal Date: November 28, 2023
 Start time (MST): 10:29 End time (MST): 10:53

Analyzer Make: API T640 S/N: 322
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Deltacal S/N: 1451
 Temp/RH standard: Deltacal S/N: 1451

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	1.3	1	1.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	730.1	731.5	730.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	4.93	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 14, 2023</u>	Last Cal Date: <u>November 28, 2023</u>			
	PM w/o HEPA: <u>8.6</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>November 28, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>November 28, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: November 28, 2023
 Date RH/T Sensor Cleaned: May 23, 2023

Leak check Passed. No adjustments done. Head Cleaned.

Notes:

Calibration by: Melissa Lemay



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: December 21, 2023 Last Cal Date: December 14, 2023
 Start time (MST): 10:00 End time (MST): 10:25

Analyzer Make: API T640 S/N: 322
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Deltacal S/N: 1451
 Temp/RH standard: Deltacal S/N: 1451

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-3.1	-3.3	-3.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	725.1	726.4	725.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	4.90	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 21, 2023</u>		Last Cal Date: <u>December 14, 2023</u>		
	PM w/o HEPA: <u>5</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11	11		<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: _____		w/ HEPA: _____		
Date Optical Chamber Cleaned:	<u>November 28, 2023</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>November 28, 2023</u>				

Annual Maintenance

Date Sample Tube Cleaned: November 28, 2023
 Date RH/T Sensor Cleaned: May 23, 2023

Removal due to not being able to update software

Notes:

Calibration by: Melissa Lemay



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: December 21, 2023 Last Cal Date:
 Start time (MST): 10:25 End time (MST): 10:50

Analyzer Make: API T640 S/N: 321
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Deltacal S/N: 1451
 Temp/RH standard: Deltacal S/N: 1451

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-3	-3.3	-3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	725.2	726.4	725.2	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.10	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 21, 2023</u>		Last Cal Date: _____		
	PM w/o HEPA: <u>1.5</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.2	11.2	11.2	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: _____		w/ HEPA: _____		
Date Optical Chamber Cleaned:	_____				<0.2 ug/m3
Disposable Filter Changed:	_____				

Annual Maintenance

Date Sample Tube Cleaned: November 28, 2023
 Date RH/T Sensor Cleaned: May 23, 2023

Installed due to old T640 not updating software. No adjustments done. Leak check passed.

Notes:

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS05
MANNIX
DECEMBER 2023**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 6, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	10:09	End Time (MST):	13:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.8	----	Correlation Coefficient	0.999996	≥0.995
800.3	806.6	0.9922			
400.2	403.8	0.9910	Slope	1.007440	0.90 - 1.10
200.1	200.9	0.9959			
			Intercept	0.280000	+/-30



SO2 Calibration Plot

Date: December 8, 2023

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mannix Station number: AMS05
 Calibration Date: December 12, 2023 Last Cal Date: November 14, 2023
 Start time (MST): 9:50 End time (MST): 15:47
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.92 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002433
 Removed Cal Gas Conc: 4.92 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1845
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1203169745
 Converter make: Global Converter serial #: 2022-196
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993603	0.988046	Backgd or Offset: 2.18	2.18
Calibration intercept:	0.221019	0.440521	Coeff or Slope: 0.866	0.866

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4919	81.3	80.0	80.6	0.995
as found 2nd point	4960	40.7	40.0	40.5	0.994
as found 3rd point	4980	20.3	20.0	20.2	0.999
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.6	----
high point	4919	81.3	80.0	79.5	1.006
second point	4960	40.7	40.0	40.1	0.999
third point	4980	20.3	20.0	19.9	1.004
as left zero	5000	0.0	0.0	0.8	----
as left span	4919	81.3	80.0	77.5	1.032
SO2 Scrubber Check	4920	80.0	800.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	1.003
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 80.4 Prev response: 79.70 *% change: 0.9%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 1.005471 AF Intercept: 0.180602
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999998

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made. Random spikes during points, diagnostics are all good. Problem could be due to the H2S gas as its running low. Further info on docit note.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 14, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:50	End Time (MST):	15:47
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999968	≥0.995
80.0	79.5	1.0062			
40.0	40.1	0.9986	Slope	0.988046	0.90 - 1.10
20.0	19.9	1.0037			
			Intercept	0.440521	+/-3



H₂S Calibration Plot

Date: December 12, 2023

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	December 8, 2023	Last Cal Date:	November 6, 2023
Start time (MST):	10:08	End time (MST):	13:01
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.61E-05	2.61E-05	NMHC SP Ratio:	4.48E-05
CH ₄ Retention time:	15.00	15.00	NMHC Peak Area:	204322
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.26	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.21	1.001
second point	4960	40.0	8.61	8.60	1.002
third point	4980	20.0	4.31	4.30	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.29	0.997
Average Correction Factor					1.001

Baseline Corr AF:	17.26	Prev response	17.24	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	9.15	9.20	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.19	0.996
second point	4960	40.0	4.57	4.59	0.996
third point	4980	20.0	2.29	2.30	0.993
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.15	9.24	0.990
Average Correction Factor					0.995
Baseline Corr AF:	9.20	Prev response	9.17	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	8.07	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.03	1.007
second point	4960	40.0	4.04	4.00	1.009
third point	4980	20.0	2.02	2.00	1.011
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.05	1.004
Average Correction Factor					1.009
Baseline Corr AF:	8.07	Prev response	8.07	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001390	0.999274
THC Cal Offset:	-0.005600	-0.003400
CH ₄ Cal Slope:	1.000679	0.993789
CH ₄ Cal Offset:	-0.010200	-0.005600
NMHC Cal Slope:	1.001993	1.003992
NMHC Cal Offset:	0.004200	0.002200

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 6, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:08	End Time (MST):	13:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995
17.23	17.21	1.0008			
8.61	8.60	1.0017			
4.31	4.30	1.0017			
			Slope	0.999274	0.90 - 1.10
			Intercept	-0.003400	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 6, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:08	End Time (MST):	13:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥0.995
8.08	8.03	1.0065			
4.04	4.00	1.0088			
2.02	2.00	1.0113			
			Slope	0.993789	0.90 - 1.10
			Intercept	-0.005600	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 6, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:08	End Time (MST):	13:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995			
9.15	9.19	0.9958						
4.57	4.59	0.9958				Slope	1.003992	0.90 - 1.10
2.29	2.30	0.9934						
			Intercept	0.002200	± 0.5			



NMHC Calibration Plot

Date: December 8, 2023

Location: Mannix





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 17, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:07	End Time (MST):	13:25
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999982	≥0.995
799.5	803.0	0.9956			
400.2	403.4	0.9921	Slope	1.003106	0.90 - 1.10
200.1	204.3	0.9795			
			Intercept	1.659655	+/-30



SO2 Calibration Plot

Date: December 13, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	December 7, 2023	Last Cal Date:	November 15, 2023
Start time (MST):	9:44	End time (MST):	13:42
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.328	ppm	Cal Gas Exp Date:	February 14, 2025
Cal Gas Cylinder #:	CC506659			
Removed Cal Gas Conc:	5.328	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002314	0.991885	Backgd or Offset:	2.01	2.01
Calibration intercept:	0.340228	0.580307	Coeff or Slope:	1.168	1.168

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4925	75.1	80.0	79.4	1.009
as found 2nd point	4963	37.5	40.0	40.0	1.001
as found 3rd point	4981	18.8	20.0	20.3	0.992
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4925	75.1	80.0	79.7	1.004
second point	4963	37.5	40.0	40.6	0.984
third point	4981	18.8	20.0	20.7	0.968
as left zero	5000	0.0	0.0	0.2	----
as left span	4925	75.1	80.0	81.2	0.986
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	0.985
Date of last converter efficiency test:	efficiency				

Baseline Corr As found:	79.3	Prev response:	80.55	*% change:	-1.6%
Baseline Corr 2nd AF pt:	39.9	AF Slope:	0.989888	AF Intercept:	0.300193
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999970		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 15, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:44	End Time (MST):	13:42
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999877	≥0.995
80.0	79.7	1.0041			
40.0	40.6	0.9841	Slope	0.991885	0.90 - 1.10
20.0	20.7	0.9678			
			Intercept	0.580307	+/-3



TRS Calibration Plot

Date: December 7, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	December 13, 2023	Last Cal Date:	November 17, 2023
Start time (MST):	10:07	End time (MST):	13:23
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH ₄ Cal Gas Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.15E-04	2.17E-04	NMHC SP Ratio:	4.87E-05
CH ₄ Retention time:	14.0	14.0	NMHC Peak Area:	186260
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				184991
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4920	80.3	17.12	17.09	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	17.17	0.997
second point	4960	40.2	8.57	8.62	0.994
third point	4980	20.1	4.29	4.37	0.981
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.12	17.18	0.997
Average Correction Factor					0.991

Baseline Corr AF:	17.09	Prev response	17.22	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919.7	80.3	9.07	9.08	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	9.07	9.12	0.995
second point	4960	40.2	4.54	4.59	0.988
third point	4980	20.1	2.27	2.34	0.971
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	9.07	9.12	0.994
Average Correction Factor					0.984
Baseline Corr AF:	9.08	Prev response	9.15	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919.7	80.3	8.06	8.01	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	8.06	8.06	1.000
second point	4960	40.2	4.03	4.03	1.002
third point	4980	20.1	2.02	2.03	0.993
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	8.06	8.06	1.000
Average Correction Factor					0.998
Baseline Corr AF:	8.01	Prev response	8.09	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.004841	1.001451
THC Cal Offset:	0.018616	0.034020
CH ₄ Cal Slope:	1.003960	0.999322
CH ₄ Cal Offset:	-0.002161	0.004445
NMHC Cal Slope:	1.006556	1.003545
NMHC Cal Offset:	0.019576	0.028776

Notes: Changed the inlet filter and the N2 cylinder after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 17, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:07	End Time (MST):	13:23
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999980	≥ 0.995			
17.12	17.17	0.9973						
8.57	8.62	0.9942				Slope	1.001451	0.90 - 1.10
4.29	4.37	0.9810						
			Intercept	0.034020	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 17, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:07	End Time (MST):	13:23
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995
8.06	8.06	1.0000			
4.03	4.03	1.0017			
2.02	2.03	0.9928			
			Slope	0.999322	0.90 - 1.10
			Intercept	0.004445	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 17, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:07	End Time (MST):	13:23
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999954	≥ 0.995
9.07	9.12	0.9947			
4.54	4.59	0.9880			
2.27	2.34	0.9707			
			Slope	1.003545	0.90 - 1.10
			Intercept	0.028776	+/-0.5



NMHC Calibration Plot

Date: December 13, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Patricia McInnes
Calibration Date: December 4, 2023
Start time (MST): 9:56
Reason: Routine
Station number: AMS06
Last Cal Date: November 7, 2023
End time (MST): 14:47

Calibration Standards

NO Gas Cylinder #: T30YCWN
NOX Cal Gas Conc: 47.94 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 47.94 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: April 11, 2025
NO Cal Gas Conc: 46.39 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 46.39 ppm
NO gas Diff:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1172750022

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.825	0.825	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.987	0.987	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.2	157.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.988764	0.992028
NO _x Cal Offset:	2.638018	2.457352
NO Cal Slope:	0.990325	0.993713
NO Cal Offset:	1.304636	1.244073
NO ₂ Cal Slope:	0.998490	0.998330
NO ₂ Cal Offset:	0.166835	-0.223314



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.6	0.4	0.2	----	----
as found span	4914	86.2	826.5	799.7	26.7	821.1	793.7	27.6	1.0065	1.0076
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.6	0.4	0.2	----	----
high point	4914	86.2	826.5	799.7	26.7	821.4	795.5	26.0	1.0062	1.0053
second point	4957	43.1	413.2	399.9	13.4	413.3	399.1	14.2	0.9998	1.0019
third point	4978	21.6	207.1	200.4	6.7	209.8	201.2	8.6	0.9872	0.9961
as left zero	5000	0.0	0.0	0.0	0.0	0.3	0.4	0.0	----	----
as left span	4914	86.2	826.5	403.9	422.5	819.2	398.9	420.3	1.0089	1.0126
Average Correction Factor									0.9977	1.0011

Corrected As found	NO _x = 820.5 ppb	NO = 793.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.1%
Previous Response	NO _x = 819.8 ppb	NO = 793.3 ppb		*Percent Change	NO = 0.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	793.6	397.8	422.5	421.9	1.0015	99.9%
2nd GPT point (200 ppb O3)	793.6	602.6	217.7	216.7	1.0047	99.5%
3rd GPT point (100 ppb O3)	793.6	700.4	119.9	119.2	1.0060	99.4%
Average Correction Factor					1.0041	99.6%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 7, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:56	End Time (MST):	14:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient 0.999977	≥0.995	
826.5	821.4	1.0062			
413.2	413.3	0.9998			
207.1	209.8	0.9872			
			Slope	0.992028	0.90 - 1.10
			Intercept	2.457352	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 7, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:56	End Time (MST):	14:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	≥0.995	
799.7	795.5	1.0053			
399.9	399.1	1.0019			
200.4	201.2	0.9961			
			Slope	0.993713	0.90 - 1.10
			Intercept	1.244073	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 7, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:56	End Time (MST):	14:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
422.5	421.9	1.0015		
217.7	216.7	1.0047		
119.9	119.2	1.0060		
			0.999994	
			0.998330	
			-0.223314	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 4, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Patricia McInnes Station number: AMS06
 Calibration Date: December 18, 2023 Last Cal Date: November 10, 2023
 Start time (MST): 10:38 End time (MST): 13:35
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3566
 ZAG Make/Model: API T701 Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1300156234
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005971	1.007657	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	-1.720000	-1.340000	Coeff or Slope:	1.039	1.039

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.0	----
as found span	5000	1303.0	400.0	402.2	0.995
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	-0.1	----
high point	5000	1303.0	400.0	402.2	0.995
second point	5000	966.5	200.0	200.0	1.000
third point	5000	794.3	100.0	97.9	1.021
as left zero	5000	800.0	0.0	-0.1	----
as left span	5000	1303.0	400.0	406.3	0.984
Average Correction Factor					1.005

Baseline Corr As found:	402.2	Previous response	400.7	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 18, 2023	Previous Calibration:	November 10, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:38	End Time (MST):	13:35
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999954	≥0.995
400.0	402.2	0.9945			
200.0	200.0	1.0000	Slope	1.007657	0.90 - 1.10
100.0	97.9	1.0215			
			Intercept	-1.340000	+/- 5



O₃ Calibration Plot

Date: December 18, 2023

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: December 1, 2023 Last Cal Date: November 17, 2023
 Start time (MST): 9:23 End time (MST): 10:43

Analyzer Make: API T640 S/N: 766
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT FP-25 S/N: 388755
 Temp/RH standard: ALICAT FP-25 S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-13.2	-13.0	-13.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	717.9	717.7	717.9	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.97	4.79	4.97	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: December 1, 2023 Last Cal Date: November 17, 2023
 PM w/o HEPA: 7.7 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.0	10.0	11.0	<input checked="" type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: 15.6 w/ HEPA: 0
 Date Optical Chamber Cleaned: December 1, 2023 <0.2 ug/m3
 Disposable Filter Changed: December 1, 2023

Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023
 Date RH/T Sensor Cleaned: April 13, 2023

Quarterly calibrations completed . Leak check passed, adjusted the PMT peak.

Notes:

Calibration by: Max Farrell



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: December 14, 2023 Last Cal Date: December 1, 2023
 Start time (MST): 11:11 End time (MST): 11:30

Analyzer Make: API T640 S/N: 766
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT FP-25 S/N: 388755
 Temp/RH standard: ALICAT FP-25 S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2.7	2.2	N/A	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	728.8	729.3	N/A	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.15	N/A	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: December 14, 2023 Last Cal Date: December 1, 2023
 PM w/o HEPA: 7.7 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: _____ w/ HEPA: _____
 Date Optical Chamber Cleaned: December 1, 2023 <0.2 ug/m3
 Disposable Filter Changed: December 1, 2023

Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023
 Date RH/T Sensor Cleaned: April 13, 2023

Removal calibration, replacing the instrument because cant turn the data alignment factor on.

Notes:

Calibration by: Max Farrell



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: December 14, 2023 Last Cal Date: N/A
 Start time (MST): 12:10 End time (MST): 12:56

Analyzer Make: API T640 S/N: 871
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT FP-25 S/N: 388755
 Temp/RH standard: ALICAT FP-25 S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	N/A	2.8	3.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	N/A	729	728	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	N/A	5.12	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 14, 2023</u>	Last Cal Date: <u>N/A</u>			
	PM w/o HEPA: <u>3.1</u>	PM w/ HEPA: <u>0</u>		<0.2 ug/m3	

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	N/A	11.1	11.1	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: _____		w/ HEPA: _____		
Date Optical Chamber Cleaned:	<u>December 14, 2023</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>December 14, 2023</u>				

Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023
 Date RH/T Sensor Cleaned: April 13, 2023

Install calibration. No adjustments made, completed quarterly maintenance. Leak check passed.

Notes:

Calibration by: Max Farrell



Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	December 6, 2023	Last Cal Date:	November 2, 2023
Start time (MST):	9:33	End time (MST):	14:26
NH3 Cal Date:	December 6, 2023	Last Cal Date:	November 2, 2023
Start time (MST):	14:35	End time (MST):	16:34
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	NA
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.29	ppm	NH3 Gas Cylinder #:	EB0108520
Removed NH3 Conc:	76.29	ppm	NH3 Cal Gas Expiry:	August 22, 2024
NH3 gas Diff:			Removed Cylinder #:	NA
Calibrator Model:	API T700		Removed cyl Expiry:	NA
ZAG make/model:	API T701		Serial Number:	3566
			Serial Number:	4602

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	808
Converter model:	Teledyne API T501	Converter serial #:	484
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	4.80
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	26.6

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.870	0.854	TN coefficient:	0.858	0.849
NOX coefficient:	0.860	0.848	NO bkgnd:	-0.985	-0.985
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-0.562	-0.562
NH3 coefficient:	0.871	0.891	TN bkgnd:	5.018	5.018

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996853	0.997270
NO _x Cal Offset:	-0.762741	0.936426
NO Cal Slope:	1.000156	0.997858
NO Cal Offset:	-0.936161	0.443186
NO ₂ Cal Slope:	1.003217	1.001189
NO ₂ Cal Offset:	-0.787714	-1.443145
NH3 Cal Slope:	1.025836	1.005013
NH3 Cal Offset:	3.642287	3.755107
TN Cal Slope:	1.030917	1.011495
TN Cal Offset:	3.830269	4.256407



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-1.1	0.9	-2.0	----	----
as found NO	4914	86.2	826.5	826.5	----	833.2	833.6	-0.6	0.992	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.9	0.0	-0.9	----	----
high NO point	4914	86.2	826.5	826.5	----	827.0	825.0	1.9	0.999	----
NO/O3 point	4914	86.2	826.5	826.5	----	825.0	822.6	2.3	1.002	----
as found NH3	3417	82.6	1800.6	----	1800.6	1849.7	----	1837.6	0.973	0.980
new NH3 cyl rp	3417	82.6	1800.6	----	1800.6	----	----	----	----	----
first NH3	3417	82.6	1800.6	----	1800.6	1822.0	----	1810.1	0.988	0.995
second NH3	3454	45.9	1000.5	----	1000.5	1020.5	----	1013.3	0.980	0.987
third NH3	3477	22.9	499.2	----	499.2	513.7	----	509.4	0.972	0.980
Average Correction Factor									1.0005	0.9874

Corrected As found TN = 834.3 ppb NO_x = 832.7 ppb NH3 = 1839.6 ppb

Previous Response TN = 855.8 ppb NO_x = 823.1 ppb NH3 = 1850.8 ppb

NH3 Previous Converter Efficiency = 87.1%

NH3 Current Converter Efficiency = 89.1%

*Percent Change TN = -2.6%

*Percent Change NO_x = 1.2%

*Percent Change NH3 = -0.6%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.9	0.2	-1.1	----	----
as found span	4914	86.2	826.5	799.7	826.5	833.6	806.3	833.2	0.9914	0.9919
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.2	-0.9	----	----
high point	4914	86.2	826.5	799.7	826.5	825.0	798.7	827.0	1.0018	1.0013
second point	4957	43.1	413.2	399.9	413.2	412.5	398.5	415.2	1.0018	1.0034
third point	4978	21.6	207.1	200.4	207.1	209.1	201.4	209.2	0.9905	0.9951
Average Correction Factor									0.9980	1.0000

Baseline Corr As fnd	TN = 834.3 ppb	NO _x = 832.7 ppb	NO = 806.1 ppb	*Percent Change	TN = -2.6%
Previous Response	TN = 855.8 ppb	NO _x = 823.1 ppb	NO = 798.9 ppb	*Percent Change	NO _x = 1.2%
				*Percent Change	NO = 0.9%
				<i>* = > +/-5% change initiates investigation</i>	

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	0.7	----	----
calibration zero	----	----	0.0	-0.2	----	----
1st GPT point (400 ppb O3)	796.0	401.3	421.4	421.3	1.0003	100.0%
2nd GPT point (200 ppb O3)	796.0	602.8	219.9	217.8	1.0097	99.0%
3rd GPT point (100 ppb O3)	796.0	698.7	124.0	121.6	1.0199	98.0%
Average Correction Factor					1.0100	99.0%

Notes: Changed the inlet filter after as founds. Adjusted both NOX and NH3 spans.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TN Calibration Summary

Version-05-2023


Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 2, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:33	End Time (MST):	14:26
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.9	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.6	1822.0	0.9883		
1000.5	1020.5	0.9804		
499.2	513.7	0.9717		
			0.999957	
			1.011495	
			4.256407	

TN Calibration Curve





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 2, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:33	End Time (MST):	14:26
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.9	----	Correlation Coefficient 0.999964	≥0.995	
1800.6	1810.1	0.9948			
1000.5	1013.3	0.9874			
499.2	509.4	0.9799			
			Slope	1.005013	0.90 - 1.10
			Intercept	3.755107	+/-20





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 2, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:33	End Time (MST):	14:26
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.5	825.0	1.0018		
413.2	412.5	1.0018		
207.1	209.1	0.9905		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 2, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:33	End Time (MST):	14:26
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.7	798.7	1.0013		
399.9	398.5	1.0034		
200.4	201.4	0.9951		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 2, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:33	End Time (MST):	14:26
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥ 0.995 $0.90 - 1.10$ ± 20
421.4	421.3	1.0003		
219.9	217.8	1.0097		
124.0	121.6	1.0199		
			0.999954	
			1.001189	
			-1.443145	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 6, 2023


Location: Patricia McInnes



NH₃ Calibration Plot

Date: December 6, 2023

Location: Patricia McInnes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	December 15, 2023	Last Cal Date:	November 20, 2023
Start time (MST):	10:14	End time (MST):	13:40
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.52	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC282115			
Removed Cal Gas Conc:	50.52	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 701H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001046	0.998820	Backgd or Offset:	2.59	2.59
Calibration intercept:	2.304446	2.284124	Coeff or Slope:	0.834	0.834

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	79.2	800.2	799.2	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	79.3	801.2	800.9	1.000
second point	4960	39.6	400.2	404.8	0.989
third point	4980	19.8	200.1	203.0	0.986
as left zero	5000	0.0	0.0	0.2	----
as left span	4921	79.2	800.2	802.1	0.998
Average Correction Factor					0.991

Baseline Corr As found:	799.20	Previous response	803.35	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 15, 2023	Previous Calibration:	November 20, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:14	End Time (MST):	13:40
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999955	≥0.995
801.2	800.9	1.0004			
400.2	404.8	0.9885	Slope	0.998820	0.90 - 1.10
200.1	203.0	0.9856			
			Intercept	2.284124	+/-30



SO2 Calibration Plot

Date: December 15, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: December 19, 2023 Last Cal Date: November 15, 2023
 Start time (MST): 10:43 End time (MST): 17:04
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.25 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC504080
 Removed Cal Gas Conc: 5.25 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3805
 ZAG Make/Model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE Analyzer serial #: 1180540018
 Converter make: CDN-101 Converter serial #: 551
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.017229	1.003385	Backgd or Offset:	2.40	2.60
Calibration intercept:	0.057461	0.237769	Coeff or Slope:	0.893	0.983

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4925	75.5	79.3	79.5	0.997
as found 2nd point	4962	37.7	39.6	39.8	0.995
as found 3rd point	4981	18.9	19.8	20.0	0.992
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4925	75.5	79.3	79.8	0.994
second point	4962	37.7	39.6	40.0	0.990
third point	4981	18.9	19.9	20.2	0.983
as left zero	5000	0.0	0.0	0.5	----
as left span	4925	75.5	79.3	80.0	0.991
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	0.989
Date of last converter efficiency test:	April 22, 2022			92.6% efficiency	

Baseline Corr As found: 79.5 Prev response: 80.69 *% change: -1.5%
 Baseline Corr 2nd AF pt: 39.8 AF Slope: 1.002661 AF Intercept: 0.057785
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999997

* = > +/-5% change initiates investigation

Notes: Scrubber failed. Scrubber changed out, span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:43	End Time (MST):	17:04
Analyzer make:	CDN-101	Analyzer serial #:	551

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999999	≥0.995
79.3	79.8	0.9939			
39.6	40.0	0.9902	Slope	1.003385	0.90 - 1.10
19.9	20.2	0.9830			
			Intercept	0.237769	+/-3



TRS Calibration Plot

Date: December 19, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: December 31, 2023 Last Cal Date: December 19, 2023
 Start time (MST): 10:26 End time (MST): 14:53
 Reason: Maintenance

Calibration Standards

Cal Gas Concentration: 5.25 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC504080
 Removed Cal Gas Conc: 5.25 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3805
 ZAG Make/Model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE Analyzer serial #: 1180540018
 Converter make: CDN-101 Converter serial #: 551
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003385	1.012033	Backgd or Offset:	2.60 2.35
Calibration intercept:	0.237769	-0.062267	Coeff or Slope:	0.983 0.893

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4925	75.5	79.3	88.4	0.897
as found 2nd point	4962	37.7	39.6	44.0	0.900
as found 3rd point	4981	18.9	19.8	21.9	0.906
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4925	75.5	79.3	80.3	0.988
second point	4962	37.7	39.6	39.9	0.993
third point	4981	18.9	19.9	19.9	0.998
as left zero	5000	0.0	0.0	0.2	----
as left span	4925	75.5	79.3	79.4	0.999

SO2 Scrubber Check

Date of last scrubber change:	25-Feb-22	Ave Corr Factor	0.993
Date of last converter efficiency test:	April 22, 2022	92.6% efficiency	

Baseline Corr As found: 88.4 Prev response: 79.77 *% change: 9.8%
 Baseline Corr 2nd AF pt: 44.0 AF Slope: 1.116006 AF Intercept: -0.122467
 Baseline Corr 3rd AF pt: 21.9 AF Correlation: 0.999991

* = > +/-5% change initiates investigation

Notes: Calibrating due to drift.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 31, 2023	Previous Calibration:	December 19, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:26	End Time (MST):	14:53
Analyzer make:	CDN-101	Analyzer serial #:	551

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999980	
79.3	80.3	0.9877			≥0.995
39.6	39.9	0.9927	Slope	1.012033	
19.9	19.9	0.9978			0.90 - 1.10
			Intercept	-0.062267	+/-3



TRS Calibration Plot

Date: December 31, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	December 15, 2023	Last Cal Date:	November 20, 2023
Start time (MST):	10:14	End time (MST):	13:40
Reason:			

Calibration Standards

Gas Cert Reference:	CC282115	Cal Gas Expiry Date:	
CH ₄ Cal Gas Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	501.2 ppm	CH ₄ Equiv Conc.	1075.1 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>	
CH ₄ SP Ratio:	0.000245	0.000251	NMHC SP Ratio:	4.27E-05	4.42E-05
CH ₄ Retention time:	13.8	13.8	NMHC Peak Area:	213790	205554
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.56	1.028
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.14	0.993
second point	4960	39.6	8.52	8.57	0.994
third point	4980	19.8	4.26	4.29	0.992
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.08	0.997

				Average Correction Factor	0.993
Baseline Corr AF:	16.56	Prev response	17.03	*% change	-2.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	9.09	8.81	1.032
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	9.09	9.17	0.992
second point	4960	39.6	4.55	4.61	0.986
third point	4980	19.8	2.27	2.34	0.973
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	9.09	9.14	0.995
Average Correction Factor					0.983
Baseline Corr AF:	8.81	Prev response	9.11	*% change	-3.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.94	7.76	1.023
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.94	7.98	0.995
second point	4960	39.6	3.97	3.95	1.005
third point	4980	19.8	1.98	1.95	1.016
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.94	7.94	1.000
Average Correction Factor					1.005
Baseline Corr AF:	7.76	Prev response	7.92	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999689	1.006595
THC Cal Offset:	0.010659	0.000457
CH ₄ Cal Slope:	1.001158	1.005894
CH ₄ Cal Offset:	-0.026358	-0.023559
NMHC Cal Slope:	0.998306	1.006942
NMHC Cal Offset:	0.037416	0.023816

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 15, 2023	Previous Calibration:	November 20, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:14	End Time (MST):	13:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995
17.03	17.14	0.9933			
8.52	8.57	0.9941			
4.26	4.29	0.9922			
			Slope	1.006595	0.90 - 1.10
			Intercept	0.000457	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 15, 2023	Previous Calibration:	November 20, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:14	End Time (MST):	13:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999958	≥0.995
7.94	7.98	0.9953			
3.97	3.95	1.0048			
1.98	1.95	1.0158			
			Slope	1.005894	0.90 - 1.10
			Intercept	-0.023559	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 15, 2023	Previous Calibration:	November 20, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:14	End Time (MST):	13:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999969	≥ 0.995
9.09	9.17	0.9918			
4.55	4.61	0.9857			
2.27	2.34	0.9725			
			Slope	1.006942	0.90 - 1.10
			Intercept	0.023816	+/-0.5



NMHC Calibration Plot

Date: December 15, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Athabasca Valley Station number: AMS07
Calibration Date: December 13, 2023 Last Cal Date: November 1, 2023
Start time (MST): 10:04 End time (MST): 15:34
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1KA4 Cal Gas Expiry Date: November 30, 2023
NOX Cal Gas Conc: 50.92 ppm NO Cal Gas Conc: 49.92 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.92 ppm Removed Gas NO Conc: 49.92 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: API T700 Serial Number: 3805
ZAG make/model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1160120024
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.083	1.095	NO bkgnd or offset:	7.6	7.7
NOX coeff or slope:	0.993	0.994	NOX bkgnd or offset:	7.8	7.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	212.3	212.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998915	0.999335
NO _x Cal Offset:	1.318809	1.718942
NO Cal Slope:	1.001602	1.000317
NO Cal Offset:	1.035305	1.435177
NO ₂ Cal Slope:	0.997138	0.998685
NO ₂ Cal Offset:	0.710095	0.269234



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
as found span	4920	80.2	816.7	800.7	16.0	808.0	792.1	16.1	1.0108	1.0108
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4920	80.2	816.7	800.7	16.0	816.6	801.2	15.3	1.0002	0.9994
second point	4960	40.1	408.4	400.4	8.0	412.1	404.0	8.1	0.9909	0.9910
third point	4980	20.0	203.7	199.7	4.0	205.9	201.7	4.2	0.9892	0.9900
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
as left span	4920	80.2	816.7	396.3	420.4	821.2	399.3	421.9	0.9945	0.9924
Average Correction Factor									0.9934	0.9934

Corrected As found	NO _x = 808.0 ppb	NO = 792.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.1%
Previous Response	NO _x = 817.2 ppb	NO = 803.0 ppb		*Percent Change	NO = -1.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	800.2	395.8	420.4	419.8	1.0015	99.8%
2nd GPT point (200 ppb O3)	800.2	598.3	217.9	218.8	0.9961	100.4%
3rd GPT point (100 ppb O3)	800.2	697.9	118.3	118.1	1.0020	99.8%
Average Correction Factor					0.9999	100.0%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 1, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:04	End Time (MST):	15:34
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
816.7	816.6	1.0002			
408.4	412.1	0.9909			
203.7	205.9	0.9892			
			Slope	0.999335	0.90 - 1.10
			Intercept	1.718942	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 1, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:04	End Time (MST):	15:34
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
800.7	801.2	0.9994		
400.4	404.0	0.9910		
199.7	201.7	0.9900		





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 1, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:04	End Time (MST):	15:34
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
420.4	419.8	1.0015			
217.9	218.8	0.9961			
118.3	118.1	1.0020			
			Slope	0.998685	0.90 - 1.10
			Intercept	0.269234	+/-20



NO_x Calibration Plot

Date: December 13, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: December 14, 2023 Last Cal Date: November 14, 2023
 Start time (MST): 10:41 End time (MST): 14:27
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: T700 Serial Number: 3805
 ZAG Make/Model: T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1152220023
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003886	0.994743	Backgd or Offset:	-2.7	-2.7
Calibration intercept:	0.320000	0.420000	Coeff or Slope:	1.549	1.549

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-1.8	----
as found span	5000	1414.8	400.0	396.5	1.009
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.7	----
high point	5000	1415.7	400.0	397.9	1.005
second point	5000	1039.9	200.0	199.5	1.003
third point	5000	856.2	100.0	101.3	0.987
as left zero	5000	0.0	0.0	-0.2	----
as left span	5000	1416.0	400.0	401.6	0.996
Average Correction Factor					0.998

Baseline Corr As found:	398.3	Previous response	401.9	*% change	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 14, 2023	Previous Calibration:	November 14, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:41	End Time (MST):	14:27
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.7	----	Correlation Coefficient	0.999960	≥0.995
400.0	397.9	1.0053			
200.0	199.5	1.0025	Slope	0.994743	0.90 - 1.10
100.0	101.3	0.9872			
			Intercept	0.420000	+/- 5



O₃ Calibration Plot

Date: December 14, 2023

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Athabasca Valley Station number: AMS 07
 Calibration Date: December 13, 2023 Last Cal Date: November 10, 2023
 Start time (MST): 10:35 End time (MST): NA

Analyzer Make: API T640 S/N: 871
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-2.4	-2.9	N/A	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	725.6	728.45	N/A	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.98	5.50	N/A	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 13, 2023</u>	Last Cal Date: <u>November 10, 2023</u>			
	PM w/o HEPA: <u>4.1</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	N/A	N/A	N/A	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>November 10, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>November 10, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022
 Date RH/T Sensor Cleaned: December 5, 2022

Notes: Temp, flow and pressure checked. Leak check done before removal of the instrument. This instrument is being removed to be replaced by an AMU unit.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Athabasca Valley Station number: AMS 07
 Calibration Date: December 13, 2023 Last Cal Date: N/A
 Start time (MST): 12:48 End time (MST): 14:33

Analyzer Make: API T640 S/N: 645
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-1.6	-1.1	-1.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	728.8	728.45	728.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.94	5.48	4.99	<input checked="" type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 13, 2023</u>	Last Cal Date: <u>N/A</u>			
	PM w/o HEPA: <u>3.9</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	N/A	N/A	N/A	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>N/A</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>N/A</u>			

Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022
 Date RH/T Sensor Cleaned: December 5, 2022

Notes: Install calibration. Temp, flow and pressure checked. Leak check done.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	December 8, 2023	Last Cal Date:	November 24, 2023
Start time (MST):	10:00	End time (MST):	13:00
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date:	December 12, 2026
Cal Gas Cylinder #:	LL66942			
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 700H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	Start	Finish		Start	Finish
Calibration slope:	1.000228	1.000200	Backgd or Offset:	4.331	4.331
Calibration intercept:	0.080550	0.158544	Coeff or Slope:	1.087	1.087

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.0	40.1	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4933	66.7	40.0	40.1	0.997
second point	4967	33.3	20.0	20.3	0.987
third point	4983	16.7	10.0	10.2	0.982
as left zero	5000	0.0	0.0	0.1	----
as left span	4933	66.7	40.0	40.1	0.998
Average Correction Factor					0.989

Baseline Corr As found:	40.02	Prev response:	40.11	*% change:	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 24, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:00	End Time (MST):	13:00
Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999977	
40.0	40.1	0.9973			≥0.995
20.0	20.3	0.9866	Slope	1.000200	
10.0	10.2	0.9824			0.90 - 1.10
			Intercept	0.158544	+/-1.5



CO Calibration Plot

Date: December 8, 2023

Location: Athabasca Valley





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	December 13, 2023	Last Cal Date:	November 8, 2023
Start time (MST):	10:03	End time (MST):	12:56
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC196697			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
ZAG Make/Model:	Teledyne API T701		Serial Number:	135

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000960	0.988680	Backgd or Offset:	4.61	4.62
Calibration intercept:	-1.244570	-0.942140	Coeff or Slope:	0.965	0.965

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-3.1	----
as found span	4920	80.3	800.4	790.6	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-2.7	----
high point	4920	80.3	800.4	790.1	1.013
second point	4960	40.2	400.7	394.3	1.016
third point	4980	20.1	200.4	200.1	1.001
as left zero	5000	0.0	0.0	-2.7	----
as left span	4920	80.3	800.4	792.4	1.010
Average Correction Factor					1.010

Baseline Corr As found:	793.70	Previous response	799.91	*% change	-0.8%
-------------------------	--------	-------------------	--------	-----------	-------

** = > +/-5% change initiates investigation*

Notes: changed inlet filters after as founds.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 8, 2023
Station Name:	December 13, 2023	Station Number:	Fort Chipewyan
Start Time (MST):	10:03	End Time (MST):	12:56
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-2.7	----	Correlation Coefficient	0.999963	≥0.995
800.4	790.1	1.0130			
400.7	394.3	1.0162	Slope	0.988680	0.90 - 1.10
200.4	200.1	1.0013			
			Intercept	-0.942140	+/-30



SO2 Calibration Plot

Date: December 13, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	December 15, 2023	Last Cal Date:	November 8, 2023
Start time (MST):	13:38	End time (MST):	17:30
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.97	ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	EY0002276			
Removed Cal Gas Conc:	4.97	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
ZAG Make/Model:	Teledyne API T701		Serial Number:	135

Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744
Converter make:	CDN-101	Converter serial #:	14639
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995282	0.988139	Backgd or Offset:	0.98	0.99
Calibration intercept:	0.578775	0.678668	Coeff or Slope:	0.741	0.741

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4920	80.5	80.0	79.4	1.014
as found 2nd point	4960	40.2	40.0	40.2	1.006
as found 3rd point	4980	20.1	20.0	20.6	0.994
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.5	----
high point	4920	80.5	80.0	79.6	1.005
second point	4960	40.2	40.0	40.4	0.989
third point	4980	20.1	20.0	20.5	0.975
as left zero	5000	0.0	0.0	0.6	----
as left span	4920	80.5	80.0	79.8	1.003
SO2 Scrubber Check	4919.7	80.3	803.0	0.0	----
Date of last scrubber change:	March 7, 2022			Ave Corr Factor	0.990
Date of last converter efficiency test:	March 15, 2022			100.7% efficiency	

Baseline Corr As found:	78.9	Prev response:	80.21	*% change:	-1.7%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.984855	AF Intercept:	0.718584
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999965		

* = > +/-5% change initiates investigation

Notes: Sampled inlet filter changed after as founds. Scrubber check passed.

Calibration Performed By: Matthew Courtoreille



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 15, 2023	Previous Calibration:	November 8, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	13:38	End Time (MST):	17:30
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999967	≥0.995
80.0	79.6	1.0051			
40.0	40.4	0.9890	Slope	0.988139	0.90 - 1.10
20.0	20.5	0.9746			
			Intercept	0.678668	+/-3



TRS Calibration Plot

Date: December 15, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Chipewyan
Calibration Date: December 7, 2023
Start time (MST): 13:14
Reason: Routine
Station number: AMS08
Last Cal Date: November 7, 2023
End time (MST): 17:16

Calibration Standards

NO Gas Cylinder #: CC363447
NOX Cal Gas Conc: 48.80 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.80 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: February 2, 2024
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3060
Serial Number: 260

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1426262592

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.101	1.138	NO bkgnd or offset:	10.4	10.4
NOX coeff or slope:	0.994	0.992	NOX bkgnd or offset:	11.3	11.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	141.2	141.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999072	0.993574
NO _x Cal Offset:	1.260000	-0.140000
NO Cal Slope:	0.997544	0.992303
NO Cal Offset:	0.920000	-0.220000
NO ₂ Cal Slope:	0.999622	0.994878
NO ₂ Cal Offset:	-0.784149	0.734900



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-1.6	-0.7	-0.9	----	----
as found span	4918	82.0	800.3	800.3	0.0	798.2	794.6	3.6	1.0027	1.0072
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-1.4	-0.5	-0.9	----	----
high point	4918	82.0	800.3	800.3	0.0	794.5	793.8	0.7	1.0073	1.0082
second point	4959	41.0	400.2	400.2	0.0	397.8	397.0	0.4	1.0059	1.0080
third point	4980	20.5	200.1	200.1	0.0	200.1	198.6	1.5	0.9999	1.0075
as left zero	5000	0.0	0.0	0.0	0.0	-1.4	-0.5	-0.9	----	----
as left span	4918	82.0	800.3	400.6	399.7	789.3	397.9	391.4	1.0140	1.0068
Average Correction Factor									1.0044	1.0079

Corrected As found	NO _x = 799.8 ppb	NO = 795.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.1%
Previous Response	NO _x = 800.8 ppb	NO = 799.3 ppb		*Percent Change	NO = -0.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.9	396.2	399.7	398.1	1.0040	99.6%
2nd GPT point (200 ppb O3)	795.9	598.1	197.8	196.9	1.0046	99.5%
3rd GPT point (100 ppb O3)	795.9	702.4	93.5	96.3	0.9709	103.0%
Average Correction Factor					0.9932	100.7%

Notes: sampled inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 7, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	13:14	End Time (MST):	17:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-1.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	794.5	1.0073		
400.2	397.8	1.0059		
200.1	200.1	0.9999		
			0.999988	
			0.993574	
			-0.140000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 7, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	13:14	End Time (MST):	17:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.5	----	Correlation Coefficient	≥0.995	
800.3	793.8	1.0082			
400.2	397.0	1.0080			
200.1	198.6	1.0075			
			Slope	0.992303	0.90 - 1.10
			Intercept	-0.220000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 7, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	13:14	End Time (MST):	17:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.9	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
399.7	398.1	1.0040		
197.8	196.9	1.0046		
93.5	96.3	0.9709		



NO_x Calibration Plot

Date: December 7, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: December 7, 2023 Last Cal Date: November 7, 2023
 Start time (MST): 10:36 End time (MST): 13:09
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 260

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3872
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996657	0.969000	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	-0.240000	0.700000	Coeff or Slope:	1.036	1.036

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.9	----
as found span	5000	913.0	400.0	391.6	1.021
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	1.4	----
high point	5000	914.7	400.0	388.5	1.030
second point	5000	786.4	200.0	194.6	1.028
third point	5000	701.3	100.0	96.6	1.035
as left zero	5000	0.0	0.0	1.3	----
as left span	5000	963.3	400.0	388.3	1.030
Average Correction Factor					1.031

Baseline Corr As found:	390.7	Previous response	398.4	*% change	-2.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed out inlet filter after as found. No adjustments

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 7, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	10:36	End Time (MST):	13:09
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.4	----	Correlation Coefficient	0.999981	≥0.995
400.0	388.5	1.0296			
200.0	194.6	1.0277	Slope	0.969000	0.90 - 1.10
100.0	96.6	1.0352			
			Intercept	0.700000	+/- 5



O₃ Calibration Plot

Date: December 7, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
 Calibration Date: December 15, 2023 Last Cal Date: November 24, 2023
 Start time (MST): 11:31 End time (MST): 11:54

Analyzer Make: Teledyne API T640 S/N: 216
 Particulate Fraction: PM2.5

Flow Meter Make/Model: FP-25 S/N: 388747
 Temp/RH standard: FP-25 S/N: 388747

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	0.8	1.0		<input type="checkbox"/>	+/- 2 °C
P (mmHg)	733.3	735.6		<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.01		<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>November 24, 2023</u>	Last Cal Date: <u>November 24, 2023</u>			
	PM w/o HEPA: <u>4.98</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.1	11.0	11.2	<input type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>24.5</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>July 25, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>July 25, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: July 25, 2023
 Date RH/T Sensor Cleaned: July 25, 2023

Notes: Analyzer removed from service.

Calibration by: Matthew C



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
 Calibration Date: December 15, 2023 Last Cal Date: November 15, 2023
 Start time (MST): 13:05 End time (MST): 13:35

Analyzer Make: Teledyne API T640 S/N: 319
 Particulate Fraction: PM2.5

Flow Meter Make/Model: FP-25 S/N: 388747
 Temp/RH standard: FP-25 S/N: 388747

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2.8	2.7	2.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	733.8	734.1	733.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	4.99	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 15, 2023</u>	Last Cal Date: <u>December 15, 2023</u>			
	PM w/o HEPA: <u>4.99</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.0	11.0	11.0	<input type="checkbox"/>	11.3 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>4.99</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>December 15, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>December 15, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: July 25, 2023
 Date RH/T Sensor Cleaned: July 25, 2023

Notes: Install audit

Calibration by: Matthew C



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	December 26, 2023	Last Cal Date:	November 10, 2023
Start time (MST):	13:14	End time (MST):	16:20
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3252
ZAG Make/Model:	API T701H		Serial Number:	135

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003222	0.994764	Backgd or Offset:	-0.014	-0.014
Calibration intercept:	0.146903	0.168931	Coeff or Slope:	1.006	0.998

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.33	----
as found span	4933	66.7	40.4	41.4	0.977
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4934	66.7	40.4	40.3	1.003
second point	4967	33.3	20.2	20.4	0.990
third point	4983	16.7	10.1	10.3	0.983
as left zero	5000	0.0	0.0	0.1	----
as left span	2960	40.0	40.4	40.1	1.007
Average Correction Factor					0.992

Baseline Corr As found:	41.06	Prev response:	40.70	*% change:	0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Adjustments made to Zero and span point.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 26, 2023	Previous Calibration:	November 10, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	13:14	End Time (MST):	16:20
Analyzer make:	API T300	Analyzer serial #:	3505

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999954	≥0.995
40.4	40.3	1.0031			
20.2	20.4	0.9902	Slope	0.994764	0.90 - 1.10
10.1	10.3	0.9835			
			Intercept	0.168931	+/-1.5



CO Calibration Plot

Date: December 26, 2023

Location: Fort Chipewyan





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	December 26, 2023	Last Cal Date:	November 10, 2023
Start time (MST):	16:20	End time (MST):	20:56
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
N2 Gen Make/Model:	NG 5000		Serial Number:	135

Analyzer Information

Analyzer make: Teledyne API T360	Analyzer serial #: 289
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.982170	1.001204	Backgd or Offset:	0.008	-0.063
Calibration intercept:	4.560000	-16.666667	Coeff or Slope:	1.019	1.087

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	3.8	----
as found span	2920	80.0	1605.9	1616.8	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	-2.4	----
high point	2920	80.0	1605.9	1605.4	1.000
second point	2960	40.0	802.9	758.7	1.058
third point					
as left zero					
as left span					

Average Correction Factor				1.029
Baseline Corr As found:	1613.00	Prev response:	1581.79	*% change: 1.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF pt:	NA	AF Correlation:		

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as found. Made adjustments to linearity on second and third points.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	December 26, 2023	Previous Calibration	November 10, 2023
Station Name	Fort Chipewyan	Station Number	AMS08
Start Time (MST)	16:20	End Time (MST)	20:56
Analyzer make	Teledyne API T360	Analyzer serial #	289

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-2.4	----	Correlation Coefficient	≥0.995	
1605.9	1605.4	1.0003			
802.9	758.7	1.0583			
			Slope	1.001204	0.90 - 1.10
			Intercept	-16.666667	+/-20



CO₂ Calibration Plot

Date: December 26, 2023

Location: Fort Chipewyan





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	December 8, 2023	Last Cal Date:	November 10, 2023
Start time (MST):	10:06	End time (MST):	13:10
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.96	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC151285			
Removed Cal Gas Conc:	49.96	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3812
ZAG Make/Model:	API T701		Serial Number:	4888

Analyzer Information

Analyzer make: Thermo 43i	Analyzer serial #: 1118148498
Analyzer Range 0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002594	0.995778	Backgd or Offset:	10.3	10.2
Calibration intercept:	-0.728669	0.410488	Coeff or Slope:	0.970	0.962

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4919	80.2	801.5	806.3	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4919	80.2	801.5	798.6	1.004
second point	4959	40.1	400.8	399.0	1.004
third point	4980	20.0	199.8	200.0	0.999
as left zero	5000	0.0	0.0	0.2	----
as left span	4919	80.2	801.5	801.0	1.001
Average Correction Factor					1.002

Baseline Corr As found:	806.40	Previous response	802.84	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 10, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:06	End Time (MST):	13:10
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	0.999998	≥0.995
801.5	798.6	1.0036			
400.8	399.0	1.0044	Slope	0.995778	0.90 - 1.10
199.8	200.0	0.9992			
			Intercept	0.410488	+/-30



SO2 Calibration Plot

Date: December 8, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Barge Landing Station number: AMS09
 Calibration Date: December 4, 2023 Last Cal Date: November 1, 2023
 Start time (MST): 10:15 End time (MST): 14:12
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.171 ppm Cal Gas Exp Date: August 22, 2026
 Cal Gas Cylinder #: CC511415
 Removed Cal Gas Conc: 5.171 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3812
 ZAG Make/Model: API T701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1331259320
 Converter make: CDN-101 Converter serial #: 519
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009543	1.006260	Backgd or Offset:	2.91	2.88
Calibration intercept:	0.059575	0.099474	Coeff or Slope:	1.200	1.176

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4923	77.4	80.0	81.2	0.986
as found 2nd point	4961	38.7	40.0	40.6	0.986
as found 3rd point	4981	18.8	19.4	20.2	0.963
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4923	77.4	80.0	80.6	0.993
second point	4961	38.7	40.0	40.5	0.988
third point	4981	19.3	20.0	20.1	0.993
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.4	80.0	80.7	0.992
SO2 Scrubber Check	4920	80.2	802.0	-0.1	----
Date of last scrubber change:	28-Feb-23			Ave Corr Factor	0.991
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 81.2 Prev response: 80.86 *% change: 0.4%
 Baseline Corr 2nd AF pt: 40.6 AF Slope: 1.012395 AF Intercept: 0.190094
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999957

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SOx scrubber check done after calibrator zero.
 Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 1, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:15	End Time (MST):	14:12
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999993	
80.0	80.6	0.9931			≥0.995
40.0	40.5	0.9883	Slope	1.006260	
20.0	20.1	0.9930			0.90 - 1.10
			Intercept	0.099474	+/-3



TRS Calibration Plot

Date: December 4, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	December 8, 2023	Last Cal Date:	November 17, 2023
Start time (MST):	10:06	End time (MST):	13:10
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.53E-04	2.53E-04	NMHC SP Ratio:	4.34E-05
CH ₄ Retention time:	15.00	15.00	NMHC Peak Area:	210777
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	17.11	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.11	1.000
second point	4960	40.1	8.56	8.55	1.001
third point	4980	20.0	4.27	4.27	0.999
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.12	1.000
				Response Correction Factor	1.000

Baseline Corr AF:	17.11	Prev response	17.17	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	9.14	9.18	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.18	0.995
second point	4960	40.1	4.57	4.58	0.997
third point	4980	20.0	2.28	2.29	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	9.14	9.19	0.994
Average Correction Factor					0.996
Baseline Corr AF:	9.18	Prev response	9.18	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	7.93	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	7.93	1.007
second point	4960	40.1	3.99	3.96	1.007
third point	4980	20.0	1.99	1.98	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	7.93	1.006
Average Correction Factor					1.006
Baseline Corr AF:	7.93	Prev response	7.99	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.003147	0.999422
THC Cal Offset:	-0.002334	0.000050
CH ₄ Cal Slope:	1.001123	0.993236
CH ₄ Cal Offset:	-0.001135	0.001649
NMHC Cal Slope:	1.005040	1.004440
NMHC Cal Offset:	-0.001199	-0.000799

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 17, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:06	End Time (MST):	13:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
17.12	17.11	1.0004						
8.56	8.55	1.0014				Slope	0.999422	0.90 - 1.10
4.27	4.27	0.9994						
			Intercept	0.000050	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 17, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:06	End Time (MST):	13:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
7.98	7.93	1.0065						
3.99	3.96	1.0072				Slope	0.993236	0.90 - 1.10
1.99	1.98	1.0037						
			Intercept	0.001649	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	November 17, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:06	End Time (MST):	13:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995			
9.14	9.18	0.9955						
4.57	4.58	0.9966				Slope	1.004440	0.90 - 1.10
2.28	2.29	0.9952						
			Intercept	-0.000799	± 0.5			



NMHC Calibration Plot

Date: December 8, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	December 5, 2023	Last Cal Date:	November 22, 2023
Start time (MST):	11:14	End time (MST):	15:42
Reason:	Routine		

Calibration Standards

NO Gas Cylinder #:	DT0036634	Cal Gas Expiry Date:	January 28, 2024
NOX Cal Gas Conc:	50.00 ppm	NO Cal Gas Conc:	49.70 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	50.00 ppm	Removed Gas NO Conc:	49.70 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.128	1.151	NO bkgnd or offset:	10.3	10.5
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	10.6	10.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	173.1	173.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998965	0.999604
NO _x Cal Offset:	0.368958	0.568998
NO Cal Slope:	0.999327	1.000613
NO Cal Offset:	-0.472648	-0.472474
NO ₂ Cal Slope:	1.001499	1.001041
NO ₂ Cal Offset:	0.067768	-0.216477



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
as found span	4919	80.5	805.1	800.3	4.8	791.9	783.0	8.9	1.017	1.022
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.1	----	----
high point	4919	80.5	805.1	800.3	4.8	805.0	800.5	4.6	1.000	1.000
second point	4959	40.2	402.1	399.7	2.4	402.9	399.2	3.8	0.998	1.001
third point	4979	20.1	201.0	199.8	1.2	202.0	199.0	2.9	0.995	1.004
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as left span	4919	80.5	805.1	458.0	347.1	803.0	455.6	347.4	1.003	1.005
Average Correction Factor									0.998	1.002

Corrected As found	NO _x = 792.1 ppb	NO = 783.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.6%
Previous Response	NO _x = 804.6 ppb	NO = 799.2 ppb		*Percent Change	NO = -2.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.2	454.9	347.1	347.3	1.000	100.0%
2nd GPT point (200 ppb O3)	797.2	668.2	133.8	134.0	0.999	100.1%
3rd GPT point (100 ppb O3)	797.2	731.4	70.6	69.9	1.010	99.0%
Average Correction Factor					1.003	99.7%

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 22, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	11:14	End Time (MST):	15:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
805.1	805.0	1.0001		
402.1	402.9	0.9979		
201.0	202.0	0.9952		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 22, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	11:14	End Time (MST):	15:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
800.3	800.5	0.9997			
399.7	399.2	1.0011			
199.8	199.0	1.0042			
			Slope	1.000613	0.90 - 1.10
			Intercept	-0.472474	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 22, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	11:14	End Time (MST):	15:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
347.1	347.3	0.9995			
133.8	134.0	0.9987			
70.6	69.9	1.0105			
			Correlation Coefficient	0.999992	≥0.995
			Slope	1.001041	0.90 - 1.10
			Intercept	-0.216477	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 5, 2023

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Barge Landing Station number: AMS 09
 Calibration Date: December 8, 2023 Last Cal Date: November 22, 2023
 Start time (MST): 13:12 End time (MST): 13:26

Analyzer Make: API T640 S/N: 844
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388746
 Temp/RH standard: Alicat FP-25 S/N: 388746

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-4.40	-4.71	-4.40	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	727.30	734.42	727.30	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	4.90	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 8, 2023</u>	Last Cal Date: <u>November 22, 2023</u>			
	PM w/o HEPA: <u>3.0</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>October 19, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>October 19, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: August 23, 2023
 Date RH/T Sensor Cleaned: August 23, 2023

Notes: Inlet head looks good. No adjustments made. Leak check passed.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS11
LOWER CAMP**

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Lower Camp Station number: AMS11
Calibration Date: December 19, 2023 Last Cal Date: November 22, 2023
Start time (MST): 10:24 End time (MST): 14:04
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.25 ppm Cal Gas Exp Date: February 23, 2025
Cal Gas Cylinder #: CC2216
Removed Cal Gas Conc: 49.25 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3807
ZAG Make/Model: Teledyne API T701 Serial Number: 196

Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 100841398
Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997987	0.998200	Backgd or Offset:	14.5	14.6
Calibration intercept:	0.750571	0.350846	Coeff or Slope:	1.034	1.034

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4919	81.3	800.8	796.5	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.8	----
high point	4919	81.3	800.8	800.1	1.001
second point	4959	40.7	400.9	399.7	1.003
third point	4980	20.3	199.9	199.9	1.000
as left zero	5000	0.0	0.0	0.9	----
as left span	4919	81.3	800.8	800.2	1.001
Average Correction Factor					1.001

Baseline Corr As found: 796.00 Previous response 799.90 *% change -0.5%
Baseline Corr 2nd AF pt: NA AF Slope: AF Intercept:
Baseline Corr 3rd AF pt: NA AF Correlation:

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 22, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:24	End Time (MST):	14:04
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.8	----	Correlation Coefficient	0.999997	≥0.995
800.8	800.1	1.0008			
400.9	399.7	1.0030	Slope	0.998200	0.90 - 1.10
199.9	199.9	1.0002			
			Intercept	0.350846	+/-30



SO2 Calibration Plot

Date: December 19, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Lower Camp Station number: AMS11
 Calibration Date: December 20, 2023 Last Cal Date: November 30, 2023
 Start time (MST): 10:48 End time (MST): 14:49
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.429 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC501097
 Removed Cal Gas Conc: 5.429 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3807
 ZAG Make/Model: API T701H Serial Number: 196

Analyzer Information

Analyzer make: Thermo 450iQ Analyzer serial #: CM20080003
 Converter make: NA Converter serial #: NA
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.012622	1.009917	Backgd or Offset: 14.2	14.8
Calibration intercept:	0.214550	0.234284	Coeff or Slope: 1.001	1.001

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.7	----
as found span	4926	73.6	79.9	80.8	0.998
as found 2nd point	4963	36.8	40.0	40.9	0.994
as found 3rd point	4982	18.6	20.2	20.8	1.005
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4926	73.6	79.9	80.9	0.988
second point	4963	36.8	40.0	40.7	0.982
third point	4982	18.6	20.2	20.6	0.980
as left zero	5000	0.0	0.0	0.4	----
as left span	4926	73.6	79.9	80.8	0.989
SO2 Scrubber Check	4919	81.1	811.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.983
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 80.1 Prev response: 81.14 *% change: -1.3%
 Baseline Corr 2nd AF pt: 40.2 AF Slope: 1.003040 AF Intercept: 0.675086
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999989

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter and performed a power cycle on the analyzer after as founds.
 Completed SO2 scrubber check after calibrator zero. Adjusted zero only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 20, 2023	Previous Calibration:	November 30, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:48	End Time (MST):	14:49
Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999995	
79.9	80.9	0.9879			≥0.995
40.0	40.7	0.9818	Slope	1.009917	
20.2	20.6	0.9803			0.90 - 1.10
			Intercept	0.234284	+/-3



H₂S Calibration Plot

Date: December 20, 2023

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	December 19, 2023	Last Cal Date:	November 22, 2023
Start time (MST):	10:24	End time (MST):	14:04
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
ZAG make/model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.99E-04	3.00E-04	NMHC SP Ratio:	5.79E-05
CH ₄ Retention time:	14.0	14.2	NMHC Peak Area:	158468
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	17.35	17.28	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	17.35	17.38	0.998
second point	4959	40.7	8.69	8.63	1.006
third point	4980	20.3	4.33	4.31	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	17.35	17.41	0.997
Average Correction Factor					1.003

Baseline Corr AF:	17.28	Prev response	17.37	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	9.19	9.18	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	9.19	9.23	0.996
second point	4959	40.7	4.60	4.57	1.007
third point	4980	20.3	2.29	2.29	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	9.19	9.22	0.997
Average Correction Factor					1.002
Baseline Corr AF:	9.18	Prev response	9.22	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	8.16	8.11	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	8.16	8.15	1.001
second point	4959	40.7	4.09	4.06	1.006
third point	4980	20.3	2.04	2.03	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	8.16	8.19	0.996
Average Correction Factor					1.004
Baseline Corr AF:	8.11	Prev response	8.15	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001857	1.001730
THC Cal Offset:	-0.011999	-0.022785
CH ₄ Cal Slope:	0.998980	0.998587
CH ₄ Cal Offset:	-0.005496	-0.005091
NMHC Cal Slope:	1.004102	1.004174
NMHC Cal Offset:	-0.005505	-0.017294

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 22, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:24	End Time (MST):	14:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999981	≥ 0.995			
17.35	17.38	0.9982						
8.69	8.63	1.0061				Slope	1.001730	0.90 - 1.10
4.33	4.31	1.0042						
			Intercept	-0.022785	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 22, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:24	End Time (MST):	14:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999993	≥ 0.995
8.16	8.15	1.0012			
4.09	4.06	1.0058			
2.04	2.03	1.0044			
			Slope	0.998587	0.90 - 1.10
			Intercept	-0.005091	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 22, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:24	End Time (MST):	14:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999965	≥ 0.995			
9.19	9.23	0.9959						
4.60	4.57	1.0067				Slope	1.004174	0.90 - 1.10
2.29	2.29	1.0041						
			Intercept	-0.017294	± 0.5			



NMHC Calibration Plot

Date: #####

Location: Lower Camp





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13
FORT MCKAY SOUTH

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	December 18, 2023	Last Cal Date:	November 6, 2023
Start time (MST):	10:31	End time (MST):	13:52
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.55	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC260812			
Removed Cal Gas Conc:	50.55	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2448
ZAG Make/Model:	API 701		Serial Number:	1117

Analyzer Information

Analyzer make:	API T100	Analyzer serial #:	599
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002128	1.000755	Backgd or Offset:	91.1	91.1
Calibration intercept:	-3.438386	-2.857974	Coeff or Slope:	0.709	0.709

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	4921	79.1	799.7	796.4	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4921	79.1	799.7	798.5	1.001
second point	4961	39.5	399.3	396.1	1.008
third point	4980	19.8	200.2	194.5	1.029
as left zero	5000	0.0	0.0	-0.3	----
as left span	4921	79.1	799.7	797.1	1.003
Average Correction Factor					1.013

Baseline Corr As found:	796.80	Previous response	797.95	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustment.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 18, 2023	Previous Calibration:	November 6, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:31	End Time (MST):	13:52
Analyzer make:	API T100	Analyzer serial #:	599

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999951	≥0.995
799.7	798.5	1.0015			
399.3	396.1	1.0081	Slope	1.000755	0.90 - 1.10
200.2	194.5	1.0292			
			Intercept	-2.857974	+/-30



SO2 Calibration Plot

Date: December 18, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: December 5, 2023 Last Cal Date: November 23, 2023
 Start time (MST): 9:45 End time (MST): 13:44
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.34 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC500241
 Removed Cal Gas Conc: 5.34 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API 701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1180540017
 Converter make: CDN-101 Converter serial #: 521
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003766	0.997103	Backgd or Offset: 3.77	3.77
Calibration intercept:	-0.102198	-0.142201	Coeff or Slope: 1.130	1.130

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4925	75.5	80.6	81.0	0.995
as found 2nd point	4962	37.7	40.3	40.4	0.997
as found 3rd point	4981	18.9	20.2	19.7	1.025
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4925	75.5	80.6	80.3	1.004
second point	4962	37.7	40.3	40.0	1.007
third point	4981	18.9	20.2	19.8	1.019
as left zero	5000	0.0	0.0	0.2	----
as left span	4925	75.5	80.6	79.9	1.009
SO2 Scrubber Check	4921	79.1	791.0	0.1	----
Date of last scrubber change:	20-Mar-20		Ave Corr Factor		1.010
Date of last converter efficiency test:	NA		efficiency		

Baseline Corr As found:	81.0	Prev response:	80.83	*% change:	0.2%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.007024	AF Intercept:	-0.242118
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999939		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 23, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	9:45	End Time (MST):	13:44
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999984	
80.6	80.3	1.0041			≥0.995
40.3	40.0	1.0067	Slope	0.997103	
20.2	19.8	1.0195			0.90 - 1.10
			Intercept	-0.142201	+/-3



TRS Calibration Plot

Date: December 5, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	December 18, 2023	Last Cal Date:	November 21, 2023
Start time (MST):	10:37	End time (MST):	13:52
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH4 Cal Gas Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
C3H8 Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
Removed C3H8 Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API701	Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.76E-04	2.80E-04	NMHC SP Ratio:	4.61E-05	4.70E-05
CH4 Retention time:	15.00	14.80	NMHC Peak Area:	197092	193260
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	17.05	16.82	1.014
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.08	0.998
second point	4961	39.5	8.51	8.45	1.008
third point	4980	19.8	4.27	4.14	1.030
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	17.05	16.81	1.014

Average Correction Factor	1.012
---------------------------	-------

Baseline Corr AF:	16.82	Prev response	17.12	*% change	-1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	9.08	8.90	1.020
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	9.08	9.10	0.998
second point	4961	39.5	4.53	4.50	1.007
third point	4980	19.8	2.27	2.20	1.033
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	9.08	8.93	1.017
Average Correction Factor					1.013
Baseline Corr AF:	8.90	Prev response	9.05	*% change	-1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	7.97	7.92	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.98	0.998
second point	4961	39.5	3.98	3.94	1.009
third point	4980	19.8	1.99	1.94	1.027
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	7.97	7.88	1.011
Average Correction Factor					1.011
Baseline Corr AF:	7.92	Prev response	8.06	*% change	-1.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.008301	1.003916
THC Cal Offset:	-0.071364	-0.069164
CH ₄ Cal Slope:	1.016730	1.003456
CH ₄ Cal Offset:	-0.036593	-0.030586
NMHC Cal Slope:	1.000565	1.004193
NMHC Cal Offset:	-0.034171	-0.038578

Notes: Changed inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 18, 2023	Previous Calibration:	November 21, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:37	End Time (MST):	13:52
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999925	≥0.995
17.05	17.08	0.9982			
8.51	8.45	1.0079	Slope	1.003916	0.90 - 1.10
4.27	4.14	1.0300			
			Intercept	-0.069164	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 18, 2023	Previous Calibration:	November 21, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:37	End Time (MST):	13:52
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999933	≥0.995			
7.97	7.98	0.9983						
3.98	3.94	1.0089				Slope	1.003456	0.90 - 1.10
1.99	1.94	1.0269						
			Intercept	-0.030586	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 18, 2023	Previous Calibration:	November 21, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:37	End Time (MST):	13:52
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999917	≥ 0.995
9.08	9.10	0.9982			
4.53	4.50	1.0072			
2.27	2.20	1.0326			
			Slope	1.004193	0.90 - 1.10
			Intercept	-0.038578	+/-0.5



NMHC Calibration Plot

Date: December 18, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.1	1.2	-0.1	----	----
as found span	4919	81.1	826.9	800.0	26.9	827.1	798.9	28.2	0.9997	1.0013
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.1	----	----
high point	4919	81.1	826.9	800.0	26.9	825.7	801.6	23.9	1.0014	0.9979
second point	4960	40.6	413.9	400.4	13.5	411.1	397.8	13.3	1.0068	1.0066
third point	4980	20.3	207.0	200.2	6.7	202.3	194.7	7.7	1.0231	1.0284
as left zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.5	0.0	----	----
as left span	4919	81.1	826.9	383.9	443.0	829.4	382.0	447.4	0.9970	1.0049
Average Correction Factor									1.0104	1.0110

Corrected As found	NO _x = 826.0 ppb	NO = 797.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.2%	
Previous Response	NO _x = 824.6 ppb	NO = 798.4 ppb		*Percent Change	NO = -0.1%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	795.7	379.6	443.0	443.4	0.9992	100.1%
2nd GPT point (200 ppb O ₃)	795.7	582.2	240.4	240.1	1.0014	99.9%
3rd GPT point (100 ppb O ₃)	795.7	689.3	133.3	131.7	1.0123	98.8%
Average Correction Factor					1.0043	99.6%

Notes: Changed inlet filter after as founds. Adjusted zero and span. Used 2nd NO reference point due to drift.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 24, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:35	End Time (MST):	14:10
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.9	825.7	1.0014		
413.9	411.1	1.0068		
207.0	202.3	1.0231		
			0.999968	
			1.000174	
			-2.251490	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 24, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:35	End Time (MST):	14:10
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
800.0	801.6	0.9979		
400.4	397.8	1.0066		
200.2	194.7	1.0284		
			0.999935	
			1.004205	
			-3.125534	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 24, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:35	End Time (MST):	14:10
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
443.0	443.4	0.9992			
240.4	240.1	1.0014			
133.3	131.7	1.0123			
			Slope	1.001643	0.90 - 1.10
			Intercept	-0.703923	+/-20



NO_x Calibration Plot

Date: December 19, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: December 11, 2023 Last Cal Date: November 20, 2023
 Start time (MST): 9:53 End time (MST): 12:17
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3871
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997971		Backgd or Offset:	2.4	2.4
Calibration intercept:	0.780000		Coeff or Slope:	0.967	0.967

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	5000	977.0	400.0	402.3	0.994
as found 2nd point	5000	838.0	200.0	202.2	0.989
as found 3rd point	5000	735.9	100.0	102.2	0.978
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

Average Correction Factor				
---------------------------	--	--	--	--

Baseline Corr As found:	402.4	Previous response	400.0	*% change	0.6%
Baseline Corr 2nd AF pt:	-200.1	AF Slope:	1.004857	AF Intercept:	0.800000
Baseline Corr 3rd AF pt:	-100.0	AF Correlation:	0.999977		

* = > +/-5% change initiates investigation


Notes: Changed inlet filter after multi-points as founds and UV source lamp. Troubleshooting will be continue for tomorrow.

Calibration Performed By: Sean Bala

O₃ Calibration Plot

Date: December 11, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: December 12, 2023 Last Cal Date: December 11, 2023
 Start time (MST): 11:30 End time (MST): 13:25
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3871
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002057	1.002029	Backgd or Offset:	2.4	3.9
Calibration intercept:	0.840000	0.420000	Coeff or Slope:	0.967	0.971

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.0	----
high point	5000	989.8	400.0	401.0	0.998
second point	5000	849.9	200.0	201.1	0.995
third point	5000	745.1	100.0	101.0	0.990
as left zero	5000	0.0	0.0	-0.1	----
as left span	5000	977.0	400.0	402.2	0.995
Average Correction Factor					0.994

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Multi point As Finds carried out yesterday. Lamp changed and optimized, left to warm up for 30 minutes.

Calibration Performed By: Ryan Power



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	December 11, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	11:30	End Time (MST):	13:25
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999995	
400.0	401.0	0.9975			≥0.995
200.0	201.1	0.9945	Slope	1.002029	
100.0	101.0	0.9901			0.90 - 1.10
			Intercept	0.420000	+/- 5



O₃ Calibration Plot

Date: December 12, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: December 13, 2023 Last Cal Date: December 12, 2023
 Start time (MST): 10:15 End time (MST): 11:28
 Reason: Maintenance

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3871
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002029	1.007857	Backgd or Offset:	3.9	3.9
Calibration intercept:	0.420000	0.000000	Coeff or Slope:	0.971	0.971

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.9	----
high point	5000	989.8	400.0	402.7	0.993
second point	5000	849.9	200.0	202.0	0.990
third point	5000	745.1	100.0	101.7	0.983
as left zero					
as left span					

Average Correction Factor	0.989
---------------------------	-------

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Lamp change from yesterday was not settled. O3 ref raised to 5100 mv yesterday evening and readings halted. Adjusted lamp pot to 4600 mv and confirmed linearity.

Calibration Performed By: Ryan Power



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	December 12, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:15	End Time (MST):	11:28
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.9	----	Correlation Coefficient	0.999977	≥0.995
400.0	402.7	0.9933			
200.0	202.0	0.9901	Slope	1.007857	0.90 - 1.10
100.0	101.7	0.9833			
			Intercept	0.000000	+/- 5



O₃ Calibration Plot

Date: December 13, 2023

Location: Fort McKay South





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort McKay South Station number: AMS 13
Calibration Date: December 19, 2023 Last Cal Date: November 24, 2023
Start time (MST): 11:06 End time (MST): 11:28

Analyzer Make: API T640 S/N: 1335
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388746
Temp/RH standard: Alicat FP-25 S/N: 388746

Monthly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Measured</u>	<u>As left</u>	<u>Adjusted</u>	(Limits)
T (°C)	-11.30	-11.67	-11.30	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	737.00	738.99	737.00	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.92	4.80	4.92	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 19, 2023</u>	Last Cal Date: <u>November 24, 2023</u>			
	PM w/o HEPA: <u>4.5</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Post maintenance</u>	<u>As left</u>	<u>Adjusted</u>	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>October 24, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>October 24, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: June 29, 2023
Date RH/T Sensor Cleaned: June 29, 2023

No adjustment made. Leak check passed.

Notes:

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC**

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 7, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:14	End Time (MST):	14:35
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	1.0	----	Correlation Coefficient	0.999965	
799.3	786.9	1.0157			≥0.995
400.1	391.0	1.0232	Slope	0.984326	
201.1	195.1	1.0306			0.90 - 1.10
			Intercept	-1.112464	+/-30



SO2 Calibration Plot

Date: December 7, 2023

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Anzac Station number: AMS14
 Calibration Date: December 5, 2023 Last Cal Date: November 1, 2023
 Start time (MST): 10:13 End time (MST): 15:03
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC510379
 Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3060
 ZAG Make/Model: API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153582
 Converter make: CD Nova CDN-101 Converter serial #: 503
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008310	1.011740	Backgd or Offset: 2.31	2.33
Calibration intercept:	-0.045461	-0.265448	Coeff or Slope: 0.992	0.992

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4938	77.9	80.0	81.1	0.984
as found 2nd point	4973	38.9	40.0	40.5	0.982
as found 3rd point	4997	19.5	20.0	20.2	0.981
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4938	77.9	80.0	80.7	0.991
second point	4973	38.9	40.0	40.1	0.996
third point	4997	19.5	20.0	19.8	1.011
as left zero	5000	0.0	0.0	-0.1	----
as left span	4938	77.9	80.0	78.9	1.013
SO2 Scrubber Check	4936	80.3	800.4	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.999
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 81.3 Prev response: 80.60 *% change: 0.9%
 Baseline Corr 2nd AF pt: 40.7 AF Slope: 1.016350 AF Intercept: -0.165503
 Baseline Corr 3rd AF pt: 20.4 AF Correlation: 0.999999

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 1, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:13	End Time (MST):	15:03
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999981	
80.0	80.7	0.9907			≥0.995
40.0	40.1	0.9964	Slope	1.011740	
20.0	19.8	1.0107			0.90 - 1.10
			Intercept	-0.265448	+/-3



TRS Calibration Plot

Date: December 5, 2023

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	December 7, 2023	Last Cal Date:	November 7, 2023
Start time (MST):	11:14	End time (MST):	14:35
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.72E-04	3.72E-04	NMHC SP Ratio:	4.60E-05
CH ₄ Retention time:	12.20	12.20	NMHC Peak Area:	198214

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	17.21	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.20	0.994
second point	4979	40.2	8.56	8.56	1.001
third point	4998	20.2	4.30	4.25	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	17.10	17.16	0.997

				Average Correction Factor	1.002
Baseline Corr AF:	17.21	Prev response	17.11	*% change	0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	9.29	0.981
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.28	0.982
second point	4979	40.2	4.56	4.61	0.989
third point	4998	20.2	2.29	2.28	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	9.11	9.26	0.984
Average Correction Factor					0.991
Baseline Corr AF:	9.29	Prev response	9.13	*% change	1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	7.92	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	7.92	1.008
second point	4979	40.2	4.00	3.94	1.015
third point	4998	20.2	2.01	1.97	1.022
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	7.99	7.90	1.011
Average Correction Factor					1.015
Baseline Corr AF:	7.92	Prev response	7.99	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002761	1.007099
THC Cal Offset:	-0.036088	-0.042334
CH ₄ Cal Slope:	1.001781	0.992465
CH ₄ Cal Offset:	-0.016831	-0.015977
NMHC Cal Slope:	1.003445	1.019691
NMHC Cal Offset:	-0.019055	-0.026156

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 7, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:14	End Time (MST):	14:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999972	≥ 0.995			
17.10	17.20	0.9941						
8.56	8.56	1.0006				Slope	1.007099	0.90 - 1.10
4.30	4.25	1.0123						
			Intercept	-0.042334	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 7, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:14	End Time (MST):	14:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999980	≥0.995			
7.99	7.92	1.0084						
4.00	3.94	1.0150				Slope	0.992465	0.90 - 1.10
2.01	1.97	1.0223						
			Intercept	-0.015977	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 7, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:14	End Time (MST):	14:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999964	≥ 0.995			
9.11	9.28	0.9821						
4.56	4.61	0.9886				Slope	1.019691	0.90 - 1.10
2.29	2.28	1.0037						
			Intercept	-0.026156	± 0.5			



NMHC Calibration Plot

Date: December 7, 2023

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	December 21, 2023	Last Cal Date:	December 7, 2023
Start time (MST):	10:02	End time (MST):	11:41
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	NA
Removed CH ₄ Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.72E-04	3.72E-04	NMHC SP Ratio:	4.60E-05
CH ₄ Retention time:	12.2	12.2	NMHC Peak Area:	198214
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	17.18	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.15	0.997
second point					
third point					
as left zero					
as left span					
Average Correction Factor					0.997

Baseline Corr AF:	17.18	Prev response	17.18	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	9.35	0.975
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.33	0.976
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.976
Baseline Corr AF:	9.35	Prev response	9.27	*% change	0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	7.83	1.021
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	7.82	1.022
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.022
Baseline Corr AF:	7.83	Prev response	7.91	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.007099	1.002822
THC Cal Offset:	-0.042334	0.000000
CH ₄ Cal Slope:	0.992465	0.978407
CH ₄ Cal Offset:	-0.015977	0.000000
NMHC Cal Slope:	1.019691	1.024117
NMHC Cal Offset:	-0.026156	0.000000


Notes: Swapped Hydrogen cylinder.

Calibration Performed By: Mohammed Kashif

NMHC Calibration Plot

Date: December 21, 2023

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Anzac
Calibration Date: December 4, 2023
Start time (MST): 10:31
Reason: Routine
Station number: AMS 14
Last Cal Date: November 2, 2023
End time (MST): 14:54

Calibration Standards

NO Gas Cylinder #: T2Y1P8D
NOX Cal Gas Conc: 50.92 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 50.92 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API 701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.05 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 50.05 ppm
NO gas Diff:
Serial Number: 3060
Serial Number: 357

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1426262592

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.411	1.411	NO bkgnd or offset:	3.8	3.8
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.8	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.3	158.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002002	1.001131
NO _x Cal Offset:	-0.728718	-0.568562
NO Cal Slope:	1.002134	1.001791
NO Cal Offset:	-2.326785	-2.406556
NO ₂ Cal Slope:	1.004193	1.001950
NO ₂ Cal Offset:	0.821549	0.975434



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.2	-0.2	0.4	----	----
as found span	4936	80.2	814.1	800.2	13.9	817.6	799.2	18.5	0.9957	1.0013
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	-0.1	0.3	----	----
high point	4936	80.2	814.1	800.2	13.9	814.8	800.4	14.2	0.9992	0.9998
second point	4979	40.1	406.8	399.9	7.0	406.6	397.0	9.7	1.0006	1.0072
third point	4999	20.1	203.9	200.4	3.5	202.5	196.1	6.4	1.0070	1.0221
as left zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.3	----	----
as left span	4936	80.2	814.1	419.7	394.4	814.3	417.4	396.9	0.9998	1.0055
Average Correction Factor									1.0022	1.0097

Corrected As found	NO _x = 817.4 ppb	NO = 799.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.3%
Previous Response	NO _x = 815.0 ppb	NO = 799.6 ppb		*Percent Change	NO = 0.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.1	415.6	394.4	395.7	0.9967	100.3%
2nd GPT point (200 ppb O3)	796.1	609.6	200.4	202.4	0.9902	101.0%
3rd GPT point (100 ppb O3)	796.1	702.9	107.1	108.8	0.9845	101.6%
Average Correction Factor					0.9905	101.0%

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 2, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:31	End Time (MST):	14:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
814.1	814.8	0.9992		
406.8	406.6	1.0006		
203.9	202.5	1.0070		
			0.999994	
			1.001131	
			-0.568562	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 2, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:31	End Time (MST):	14:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
800.2	800.4	0.9998			
399.9	397.0	1.0072			
200.4	196.1	1.0221			
			Slope	1.001791	0.90 - 1.10
			Intercept	-2.406556	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 2, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:31	End Time (MST):	14:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
394.4	395.7	0.9967		
200.4	202.4	0.9902		
107.1	108.8	0.9845		



NO_x Calibration Plot

Date: December 4, 2023

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Anzac Station number: AMS14
 Calibration Date: December 6, 2023 Last Cal Date: November 21, 2023
 Start time (MST): 11:05 End time (MST): 14:20
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3060
 ZAG Make/Model: API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1426262595
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002171	1.002086	Backgd or Offset:	1.4	1.4
Calibration intercept:	-0.580000	-0.040000	Coeff or Slope:	1.620	1.620

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	5000	918.8	400.0	401.5	0.996
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.3	----
high point	5000	918.8	400.0	400.9	0.998
second point	5000	803.8	200.0	200.4	0.998
third point	5000	709.8	100.0	99.7	1.003
as left zero	5000	0.0	0.0	0.4	----
as left span	5000	918.8	400.0	406.5	0.984
Average Correction Factor					1.000

Baseline Corr As found:	401.1	Previous response	400.3	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments required.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 21, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:05	End Time (MST):	14:20
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999996	≥0.995
400.0	400.9	0.9978			
200.0	200.4	0.9980	Slope	1.002086	0.90 - 1.10
100.0	99.7	1.0030			
			Intercept	-0.040000	+/- 5



O₃ Calibration Plot

Date: December 6, 2023

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Anzac Station number: AMS 14
 Calibration Date: December 6, 2023 Last Cal Date: November 7, 2023
 Start time (MST): 14:29 End time (MST): 15:28

Analyzer Make: API T640 S/N: 825
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388749
 Temp/RH standard: Alicat FP-25 S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	6.1	5.4	6.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	706.6	707.6	706.6	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	4.97	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 6, 2023</u>	Last Cal Date: <u>November 7, 2023</u>			
	PM w/o HEPA: <u>2.9</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.8	10.9	10.8	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>2.1</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>December 6, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>December 6, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: July 6, 2023
 Date RH/T Sensor Cleaned: July 6, 2023

Notes: Replaced sample pump due to 80% drive alarm and completed quarterly maintenance (chamber cleaning and PMT check).

Calibration by: Mohammed Kashif



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS17
WAPASU**

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 6, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:11	End Time (MST):	15:11
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	≥0.995
800.0	802.4	0.9970		
400.0	397.3	1.0069	Slope	0.90 - 1.10
199.5	197.3	1.0112		
			Intercept	+/-30



SO2 Calibration Plot

Date: December 4, 2023

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	December 5, 2023	Last Cal Date:	November 28, 2023
Start time (MST):	10:38	End time (MST):	15:30
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.076	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511852			
Removed Cal Gas Conc:	5.076	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2449
ZAG Make/Model:	API T701H		Serial Number:	359

Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583
Converter make:	n/a	Converter serial #:	n/a
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001426	1.005283	Backgd or Offset:	12.2
Calibration intercept:	0.000779	0.040787	Coeff or Slope:	1.114

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	78.8	80.0	82.4	0.971
as found 2nd point	4961	39.4	40.0	40.9	0.978
as found 3rd point	4980	19.7	20.0	20.2	0.990
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	78.8	80.0	80.6	0.993
second point	4961	39.4	40.0	39.9	1.002
third point	4980	19.7	20.0	20.3	0.985
as left zero	5000	0.0	0.0	0.7	----
as left span	4921	78.8	80.0	79.1	1.011
SO2 Scrubber Check	4921	79.4	800.0	0.4	----
Date of last scrubber change:		n/a		Ave Corr Factor	0.993
Date of last converter efficiency test:		n/a			efficiency

Baseline Corr As found:	82.4	Prev response:	80.12	*% change:	2.8%
Baseline Corr 2nd AF pt:	40.9	AF Slope:	1.031282	AF Intercept:	-0.219180
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999968		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 28, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:38	End Time (MST):	15:30
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999953	≥0.995
80.0	80.6	0.9926			
40.0	39.9	1.0024	Slope	1.005283	0.90 - 1.10
20.0	20.3	0.9853			
			Intercept	0.040787	+/-3



H₂S Calibration Plot

Date: December 5, 2023

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	December 4, 2023	Last Cal Date:	November 6, 2023
Start time (MST):	11:11	End time (MST):	13:11
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	ALM066507	Cal Gas Expiry Date:	January 12, 2029
CH4 Cal Gas Conc.	<u>503.5</u> ppm	CH4 Equiv Conc.	1076.3 ppm
C3H8 Cal Gas Conc.	<u>208.3</u> ppm		
Removed Gas Cert:	n/a	Removed Gas Expiry:	n/a
Removed CH4 Conc.	<u>503.5</u> ppm	CH4 Equiv Conc.	1076.3 ppm
Removed C3H8 Conc.	<u>208.3</u> ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	2449
ZAG Make/Model:	API 701H	Serial Number:	359

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1218153352
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004169	1.000963	Background:	3.300	3.300
Calibration intercept:	-0.126321	-0.028353	Coefficient:	4.460	4.460

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.04	----
as found span	4921	79.4	17.09	17.11	0.999
as found 2nd point	4960	39.7	8.55	8.46	1.010
as found 3rd point	4980	19.8	4.26	4.20	1.016
new cylinder response					
calibrator zero	5000	0.0	0.00	0.04	----
high point	4921	79.4	17.09	17.12	0.998
second point	4960	39.7	8.55	8.47	1.010
third point	4980	19.8	4.26	4.19	1.016
as left zero	5000	0.0	0.00	-0.01	----
as left span	4920	79.4	17.09	17.20	0.994
Average Correction Factor					1.008
Baseline Corr As found:	17.15	Previous response	17.04	*% change	0.6%
Baseline Corr 2nd AF pt:	8.49	AF Slope:	1.003541	AF Intercept:	-0.069128
Baseline Corr 3rd AF pt:	4.23	AF Correlation:	0.999972		

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 6, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:11	End Time (MST):	13:11
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.04	----	Correlation Coefficient	0.999929	≥0.995
17.09	17.12	0.9983			
8.55	8.47	1.0095	Slope	1.000963	0.90 - 1.10
4.26	4.19	1.0163			
			Intercept	-0.028353	+/-1.5



THC Calibration Plot

Date: December 4, 2023

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	December 21, 2023	Last Cal Date:	December 4, 2023
Start time (MST):	11:36	End time (MST):	15:00
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	ALM066507	Cal Gas Expiry Date:	January 12, 2029
CH4 Cal Gas Conc.	<u>503.5</u> ppm	CH4 Equiv Conc.	1076.3 ppm
C3H8 Cal Gas Conc.	<u>208.3</u> ppm		
Removed Gas Cert:	n/a	Removed Gas Expiry:	n/a
Removed CH4 Conc.	<u>503.5</u> ppm	CH4 Equiv Conc.	1076.3 ppm
Removed C3H8 Conc.	<u>208.3</u> ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	2449
ZAG Make/Model:	API 701H	Serial Number:	359

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1218153352
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000963	1.011452	Background:	3.300	3.300
Calibration intercept:	-0.028353	-0.239261	Coefficient:	4.460	4.451

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.12	----
as found span	4921	79.4	17.09	17.34	0.985
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	-0.24	----
high point	4921	79.4	17.09	17.08	1.001
second point	4960	39.7	8.55	8.31	1.028
third point	4980	19.8	4.26	4.13	1.032
as left zero	5000	0.0	0.00	-0.14	----
as left span	4921	79.4	17.09	17.07	1.001
Average Correction Factor					1.020
Baseline Corr As found:	17.22	Previous response	17.08	*% change	0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filters and Hydrogen cylinder after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 21, 2023	Previous Calibration:	December 4, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:36	End Time (MST):	15:00
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<i>Limits</i>	
0.00	-0.24	----	Correlation Coefficient	0.999917	
17.09	17.08	1.0006			<i>≥0.995</i>
8.55	8.31	1.0285	Slope	1.011452	
4.26	4.13	1.0318			<i>0.90 - 1.10</i>
			Intercept	-0.239261	<i>+/-1.5</i>



THC Calibration Plot

Date: December 21, 2023

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
as found span	4917	83.2	817.2	799.9	17.3	823.9	809.5	14.5	0.9919	0.9881
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4917	83.2	817.2	799.9	17.3	816.7	800.4	16.3	1.0006	0.9994
second point	4958	41.6	408.6	399.9	8.7	399.8	391.0	8.8	1.0220	1.0229
third point	4979	20.8	204.3	200.0	4.3	200.9	198.0	2.9	1.0169	1.0100
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.0	----	----
as left span	4917	83.2	817.2	393.8	423.4	815.9	388.6	427.3	1.0015	1.0133
Average Correction Factor									1.0132	1.0107

Corrected As found	NO _x = 824.1 ppb	NO = 809.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.1%
Previous Response	NO _x = 815.2 ppb	NO = 800.2 ppb		*Percent Change	NO = 1.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.5	392.4	423.4	425.0	0.9962	100.4%
2nd GPT point (200 ppb O3)	798.5	595.3	220.5	221.5	0.9955	100.5%
3rd GPT point (100 ppb O3)	798.5	700.4	115.4	115.6	0.9983	100.2%
Average Correction Factor					0.9967	100.3%

Notes:

Change filters after as founds. Adjusted span only.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 21, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:55	End Time (MST):	15:53
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
817.2	816.7	1.0006		
408.6	399.8	1.0220		
204.3	200.9	1.0169		
			0.999868	
			0.999676	
			-3.080000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 21, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:55	End Time (MST):	15:53
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999839	≥0.995
799.9	800.4	0.9994			
399.9	391.0	1.0229	Slope	1.000330	0.90 - 1.10
200.0	198.0	1.0100			
			Intercept	-2.740000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 21, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:55	End Time (MST):	15:53
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
423.4	425.0	0.9962			0.999999
220.5	221.5	0.9955			1.004002
115.4	115.6	0.9983			-0.063828


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 7, 2023

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Wapasu Station number: AMS17
 Calibration Date: December 6, 2023 Last Cal Date: November 2, 2023
 Start time (MST): 10:30 End time (MST): 14:10
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 2449
 ZAG Make/Model: API T701H Serial Number: 359

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 3870
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003143	1.002943	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.400000	-0.540000	Coeff or Slope:	1.014	1.014

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	5000	1077.3	400.0	399.2	1.002
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.0	----
high point	5000	1077.3	400.0	400.9	0.998
second point	5000	900.3	200.0	199.8	1.001
third point	5000	789.5	100.0	99.2	1.008
as left zero	5000	0.0	0.0	0.6	----
as left span	5000	1077.3	400.0	405.6	0.986
Average Correction Factor					1.002

Baseline Corr As found:	398.9	Previous response	400.9	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 2, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:30	End Time (MST):	14:10
Analyzer make:	API T400	Analyzer serial #:	3870

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999992	≥0.995
400.0	400.9	0.9978			
200.0	199.8	1.0010	Slope	1.002943	0.90 - 1.10
100.0	99.2	1.0081			
			Intercept	-0.540000	+/- 5



O₃ Calibration Plot

Date: December 6, 2023

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Wapasu Station number: AMS 17
 Calibration Date: December 7, 2023 Last Cal Date: November 28, 2023
 Start time (MST): 14:41 End time (MST): 15:20

Analyzer Make: API T640 S/N: 1183
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-1.2	-1.8	-1.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	705.3	704.2	705.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.93	4.92	4.93	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 7, 2023</u>	Last Cal Date: <u>November 28, 2023</u>			
	PM w/o HEPA: <u>5.4</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>October 18, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 22, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: Temp, pressure and flow checked. Leak check passed.

Calibration by: Aswin Sasi Kumar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 7, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:50	End Time (MST):	14:55
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.6	----	Correlation Coefficient	0.999994	
800.3	799.1	1.0015			≥0.995
400.2	399.8	1.0010	Slope	0.998520	
199.6	198.0	1.0081			0.90 - 1.10
			Intercept	-0.124003	+/-30



SO2 Calibration Plot

Date: December 19, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: December 21, 2023 Last Cal Date: November 29, 2023
 Start time (MST): 11:30 End time (MST): 17:07
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.479 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC500395
 Removed Cal Gas Conc: 5.479 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
 ZAG Make/Model: Teledyne API T701 Serial Number: 360

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359
 Converter make: CD Nova CDN-101 Converter serial #: 555
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000731	1.000015	Backgd or Offset: 2.66	2.66
Calibration intercept:	0.240905	0.540982	Coeff or Slope: 1.157	1.157

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4927	73.0	80.0	82.3	0.974
as found 2nd point	4964	36.5	40.0	41.3	0.973
as found 3rd point	4983	18.3	20.0	20.5	0.988
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4927	73.0	80.0	80.3	0.996
second point	4964	36.5	40.0	41.0	0.975
third point	4983	18.3	20.0	20.6	0.973
as left zero	5000	0.0	0.0	0.3	----
as left span	4927	73.0	80.0	79.9	1.001
SO2 Scrubber Check	4923	77.1	771.0	0.0	----
Date of last scrubber change:	17-Dec-21			Ave Corr Factor	0.982
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 82.1 Prev response: 80.29 *% change: 2.2%
 Baseline Corr 2nd AF pt: 41.1 AF Slope: 1.027592 AF Intercept: 0.100548
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999983

* = > +/-5% change initiates investigation

Notes: Scrubber check completed after calibrator zero. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 21, 2023	Previous Calibration:	November 29, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	11:30	End Time (MST):	17:07
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999906	≥0.995
80.0	80.3	0.9962			
40.0	41.0	0.9754	Slope	1.000015	0.90 - 1.10
20.0	20.6	0.9732			
			Intercept	0.540982	+/-3



TRS Calibration Plot

Date: December 21, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	December 19, 2023	Last Cal Date:	November 7, 2023
Start time (MST):	10:50	End time (MST):	14:55
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.25E-04	2.24E-04	NMHC SP Ratio:	5.10E-05
CH ₄ Retention time:	12.7	12.7	NMHC Peak Area:	179849
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	17.28	17.39	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.34	0.997
second point	4959	40.5	8.64	8.64	1.000
third point	4979	20.2	4.31	4.27	1.010
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.44	0.991
Average Correction Factor					1.002

Baseline Corr AF:	17.39	Prev response	17.23	*% change	0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	9.23	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.16	1.001
second point	4959	40.5	4.58	4.60	0.996
third point	4979	20.2	2.29	2.30	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	9.17	9.24	0.992
Average Correction Factor					0.997
Baseline Corr AF:	9.23	Prev response	9.16	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	8.11	8.15	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	8.18	0.992
second point	4959	40.5	4.06	4.04	1.003
third point	4979	20.2	2.02	1.97	1.027
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.20	0.990
Average Correction Factor					1.007
Baseline Corr AF:	8.15	Prev response	8.07	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999178	1.004329
THC Cal Offset:	-0.036603	-0.027780
CH ₄ Cal Slope:	0.999274	1.010541
CH ₄ Cal Offset:	-0.034024	-0.037005
NMHC Cal Slope:	0.999119	0.998956
NMHC Cal Offset:	-0.002179	0.009225

Notes: Inlet filter and H2 cylinder changed after as founds. Span adjusted.

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 7, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:50	End Time (MST):	14:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999988	≥ 0.995			
17.28	17.34	0.9966						
8.64	8.64	0.9997				Slope	1.004329	0.90 - 1.10
4.31	4.27	1.0097						
			Intercept	-0.027780	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 7, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:50	End Time (MST):	14:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999907	≥0.995			
8.11	8.18	0.9918						
4.06	4.04	1.0032				Slope	1.010541	0.90 - 1.10
2.02	1.97	1.0267						
			Intercept	-0.037005	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 7, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:50	End Time (MST):	14:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995			
9.17	9.16	1.0008						
4.58	4.60	0.9964				Slope	0.998956	0.90 - 1.10
2.29	2.30	0.9951						
			Intercept	0.009225	± 0.5			



NMHC Calibration Plot

Date: December 19, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: December 21, 2023 Last Cal Date: November 30, 2023
Start time (MST): 11:45 End time (MST): 17:21
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2XX7ME Cal Gas Expiry Date: January 14, 2024
NOX Cal Gas Conc: 50.48 ppm NO Cal Gas Conc: 49.22 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 50.48 ppm Removed Gas NO Conc: 49.22 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T750 Serial Number: 282
ZAG make/model: Teledyne API T751H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1336160088
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.087	1.076	NO bkgnd or offset:	3.1	3.0
NOX coeff or slope:	0.985	0.988	NOX bkgnd or offset:	3.1	3.0
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	248.8	249.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001069	0.997071
NO _x Cal Offset:	-0.390046	1.071123
NO Cal Slope:	1.001210	0.998008
NO Cal Offset:	-1.489996	0.231259
NO ₂ Cal Slope:	1.001477	1.005966
NO ₂ Cal Offset:	1.056533	0.956250



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
as found span	4919	81.3	820.8	800.3	20.5	832.0	810.0	21.7	0.9865	0.9880
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.0	0.3	----	----
high point	4919	81.3	820.8	800.3	20.5	819.5	799.2	20.3	1.0015	1.0013
second point	4959	40.7	410.9	400.7	10.3	409.8	399.0	10.8	1.0028	1.0042
third point	4980	20.3	204.9	199.8	5.1	207.1	200.7	6.4	0.9896	0.9956
as left zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
as left span	4919	81.3	820.8	420.6	400.2	812.0	352.6	459.2	1.0108	1.1928
Average Correction Factor									0.9979	1.0004

Corrected As found	NO _x = 832.0 ppb	NO = 810.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.3%
Previous Response	NO _x = 821.2 ppb	NO = 799.7 ppb		*Percent Change	NO = 1.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.0	416.3	400.2	403.0	0.9930	100.7%
2nd GPT point (200 ppb O3)	796.0	606.8	209.7	212.7	0.9858	101.4%
3rd GPT point (100 ppb O3)	796.0	703.0	113.5	115.5	0.9826	101.8%
Average Correction Factor					0.9871	101.3%

Notes: Portable calibrator used for calibration. Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 21, 2023	Previous Calibration:	November 30, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:45	End Time (MST):	17:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
820.8	819.5	1.0015		
410.9	409.8	1.0028		
204.9	207.1	0.9896		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 21, 2023	Previous Calibration:	November 30, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:45	End Time (MST):	17:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
800.3	799.2	1.0013			
400.7	399.0	1.0042			
199.8	200.7	0.9956			
			Correlation Coefficient	0.999993	≥0.995
			Slope	0.998008	0.90 - 1.10
			Intercept	0.231259	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 21, 2023	Previous Calibration:	November 30, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:45	End Time (MST):	17:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	≥0.995	
400.2	403.0	0.9930			
209.7	212.7	0.9858			
113.5	115.5	0.9826			
			Slope	1.005966	0.90 - 1.10
			Intercept	0.956250	+/-20



NO_x Calibration Plot

Date: December 21, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: December 12, 2023 Last Cal Date: November 23, 2023
 Start time (MST): 11:50 End time (MST): 14:40
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
 ZAG Make/Model: Teledyne API 701H Serial Number: 360

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 825
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001400	1.002257	Backgd or Offset:	2.4	2.4
Calibration intercept:	-0.220000	-1.420000	Coeff or Slope:	0.978	0.978

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	NA	0.0	-1.3	----
as found span	4888	1141.9	400.0	399.7	1.001
as found 2nd point					
as found 3rd point					
calibrator zero	5000	NA	0.0	-1.0	----
high point	4888	1138.1	400.0	399.5	1.001
second point	4888	884.5	200.0	199.4	1.003
third point	4888	741.4	100.0	98.0	1.020
as left zero	5000	NA	0.0	-1.4	----
as left span	4812	1097.9	400.0	402.2	0.995
Average Correction Factor					1.008

Baseline Corr As found:	401.0	Previous response	400.3	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments performed.

Calibration Performed By: Kelly Baragar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 23, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	11:50	End Time (MST):	14:40
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-1.0	----	Correlation Coefficient	≥0.995
400.0	399.5	1.0013		
200.0	199.4	1.0030	Slope	0.90 - 1.10
100.0	98.0	1.0204		
			Intercept	+/- 5



O₃ Calibration Plot

Date: December 12, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Stony Mountain Station number: AMS 18
 Calibration Date: December 20, 2023 Last Cal Date: November 30, 2023
 Start time (MST): 12:30 End time (MST): 13:25

Analyzer Make: API T640 S/N: 1162
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-5.1	-5.5	-5.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	694.1	698.3	694.1	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	0.0	4.99	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 20, 2023</u>	Last Cal Date: <u>November 30, 2023</u>			
	PM w/o HEPA: <u>10.4</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	NA		NA	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>September 25, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 25, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022
 Date RH/T Sensor Cleaned: August 30, 2022

Notes: Pump was DOA, pump swapped. No adjustments needed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	December 12, 2023	Last Cal Date:	November 8, 2023
Start time (MST):	11:50	End time (MST):	15:06
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,050	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	3,050	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T750		Serial Number:	281
ZAG Make/Model:	Teledyne API T751		Serial Number:	321

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998397	0.992948	Backgd or Offset:	-0.010	-0.010
Calibration intercept:	0.047809	0.219785	Coeff or Slope:	0.906	0.906

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.7	40.6	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4933	66.7	40.7	40.5	1.004
second point	4966	33.3	20.3	20.6	0.986
third point	4983	16.7	10.2	10.4	0.982
as left zero	5000	0.0	0.0	0.1	----
as left span	4933	66.7	40.7	40.9	0.995
Average Correction Factor					0.991

Baseline Corr As found:	40.50	Prev response:	40.67	*% change:	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments performed.

Calibration Performed By: Kelly Baragar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 8, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:50	End Time (MST):	15:06
Analyzer make:	API T300	Analyzer serial #:	3504

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999917	≥0.995
40.7	40.5	1.0044			
20.3	20.6	0.9862	Slope	0.992948	0.90 - 1.10
10.2	10.4	0.9824			
			Intercept	0.219785	+/-1.5



CO Calibration Plot

Date: December 12, 2023

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	December 20, 2023	Last Cal Date:	November 16, 2023
Start time (MST):	10:46	End time (MST):	13:40
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2026
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
N2 Gen Make/Model:	Peak Scientific		Serial Number:	771048317

Analyzer Information

Analyzer make: API T360	Analyzer serial #: 489
Analyzer Range 0 - 2,000 ppm	

	Start	Finish		Start	Finish
Calibration slope:	1.002631	0.999080	Backgd or Offset:	-0.037	-0.037
Calibration intercept:	-3.040000	-2.620000	Coeff or Slope:	0.938	0.938

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	0.2	----
as found span	2920	80.0	1605.9	1604.6	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.6	----
high point	2920	80.0	1605.9	1605.0	1.001
second point	2960	40.0	802.9	793.1	1.012
third point	2980	20.0	401.5	398.5	1.007
as left zero	3000	0.0	0.0	0.3	----
as left span	2930	80.0	1600.5	1601.6	0.999
Average Correction Factor					1.007

Baseline Corr As found:	1604.40	Prev response:	1607.05	*% change:	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	December 20, 2023	Previous Calibration	November 16, 2023
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:46	End Time (MST)	13:40
Analyzer make	API T360	Analyzer serial #	489

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999955	≥0.995
1605.9	1605.0	1.0005			
802.9	793.1	1.0124	Slope	0.999080	0.90 - 1.10
401.5	398.5	1.0074			
			Intercept	-2.620000	+/-10



CO₂ Calibration Plot

Date: December 20, 2023

Location: Stony Mountain





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS19
FIREBAG**

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 29, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:29	End Time (MST):	15:26
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	1.000000	≥0.995
799.5	804.0	0.9944			
400.3	402.2	0.9952	Slope	1.005909	0.90 - 1.10
200.1	200.8	0.9965			
			Intercept	-0.303040	+/-30



SO2 Calibration Plot

Date: December 13, 2023

Location: Firebag





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Firebag	Station number:	AMS19
Calibration Date:	December 11, 2023	Last Cal Date:	November 21, 2023
Start time (MST):	10:48	End time (MST):	16:40
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.114	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517427			
Removed Cal Gas Conc:	5.114	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1607
ZAG Make/Model:	Teledyne API T701		Serial Number:	1118

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032
Converter make:	Global	Converter serial #:	2022-222
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995764	1.000196	Backgd or Offset:	2.62
Calibration intercept:	0.018558	-0.061587	Coeff or Slope:	1.164
				2.82
				1.176

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4922	78.2	80.0	80.3	0.997
as found 2nd point	4961	39.1	40.0	40.3	0.995
as found 3rd point	4980	19.6	20.0	20.2	0.997
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4922	78.2	80.0	79.9	1.001
second point	4961	39.1	40.0	40.0	1.000
third point	4980	19.6	20.0	20.0	1.002
as left zero	5000	0.0	0.0	0.1	----
as left span	4922	78.2	80.0	78.9	1.014
SO2 Scrubber Check	4922	78.3	800.2	0.1	----
Date of last scrubber change:	January 18, 2023			Ave Corr Factor	1.001
Date of last converter efficiency test:	n/a			efficiency	

Baseline Corr As found:	80.2	Prev response:	79.66	*% change:	0.7%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	1.002911	AF Intercept:	0.118373
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999998		

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Ran SOx scrubber check after cal zero. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 11, 2023	Previous Calibration:	November 21, 2023
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	10:48	End Time (MST):	16:40
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999998	≥0.995
80.0	79.9	1.0010			
40.0	40.0	0.9998	Slope	1.000196	0.90 - 1.10
20.0	20.0	1.0024			
			Intercept	-0.061587	+/-3



H₂S Calibration Plot

Date: December 11, 2023

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 29, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:29	End Time (MST):	15:26
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<i>Limits</i>	
0.00	0.08	----	Correlation Coefficient	0.999952	≥0.995
17.31	17.35	0.9974			
8.66	8.62	1.0051	Slope	0.998011	0.90 - 1.10
4.33	4.36	0.9935			
			Intercept	0.042876	+/-1.5



THC Calibration Plot

Date: December 13, 2023

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Firebag
Calibration Date: December 12, 2023
Start time (MST): 10:42
Reason: Routine
Station number: AMS 19
Last Cal Date: November 17, 2023
End time (MST): 16:23

Calibration Standards

NO Gas Cylinder #: DT0044018
NOX Cal Gas Conc: 48.9 ppm
Removed Cylinder #: T2Y1K63
Removed Gas NOX Conc: 51.12 ppm
NOX gas Diff: 0.5%
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.7 ppm
Removed Gas Exp Date: November 30, 2023
Removed Gas NO Conc: 49.40 ppm
NO gas Diff: 0.6%
Serial Number: 1607
Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1410661309

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.091	1.089	NO bkgnd or offset:	7.6	7.6
NOX coeff or slope:	0.992	0.993	NOX bkgnd or offset:	7.7	7.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	212.4	215.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998263	1.005184
NO _x Cal Offset:	0.656194	0.260458
NO Cal Slope:	1.002153	1.007037
NO Cal Offset:	0.109347	-0.099613
NO ₂ Cal Slope:	1.003810	0.999988
NO ₂ Cal Offset:	-1.268713	-1.617748



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
as found span	4919	81.0	828.1	800.3	27.9	828.0	801.0	26.5	1.0002	0.9991
as found 2nd										
as found 3rd										
new cyl resp	4919	81.5	797.1	793.8	3.3	800.8	799.3	1.6	0.9953	0.9931
calibrator zero	4999	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
high point	4919	81.5	797.1	793.8	3.3	801.0	799.1	1.7	0.9951	0.9934
second point	4959	40.7	398.0	396.4	1.6	401.3	399.7	1.5	0.9919	0.9918
third point	4980	20.4	199.5	198.7	0.8	200.8	199.6	1.1	0.9936	0.9955
as left zero	4999	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as left span	4919	81.0	792.2	360.5	431.7	798.2	366.5	431.7	0.9925	0.9837
Average Correction Factor									0.9935	0.9935

Corrected As found	NO _x = 828.4 ppb	NO = 801.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.1%
Previous Response	NO _x = 827.4 ppb	NO = 802.1 ppb		*Percent Change	NO = -0.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	793.9	365.5	431.7	430.6	1.0025	99.8%
2nd GPT point (200 ppb O3)	793.9	578.7	218.5	216.7	1.0081	99.2%
3rd GPT point (100 ppb O3)	793.9	687.7	109.5	105.9	1.0336	96.7%
Average Correction Factor					1.0147	98.6%

Notes: Replaced NO cylinder #T2Y1K63 with #DT0044018. Adjusted span. Second high GPT point used for GPT calibration formulas.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 17, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:42	End Time (MST):	16:23
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
797.1	801.0	0.9951		
398.0	401.3	0.9919		
199.5	200.8	0.9936		
			0.999996	
			1.005184	
			0.260458	





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 17, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:42	End Time (MST):	16:23
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
793.8	799.1	0.9934		
396.4	399.7	0.9918		
198.7	199.6	0.9955		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 17, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:42	End Time (MST):	16:23
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
431.7	430.6	1.0025			
218.5	216.7	1.0081			
109.5	105.9	1.0336			
			Slope	0.999988	0.90 - 1.10
			Intercept	-1.617748	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 12, 2023

Location: Firebag





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS20
MACKAY RIVER**

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 17, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:55	End Time (MST):	11:41
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999945	≥0.995
800.3	795.8	1.0056			
400.7	400.9	0.9994	Slope	0.991883	0.90 - 1.10
199.8	204.5	0.9771			
			Intercept	3.051032	+/-30



SO2 Calibration Plot

Date: December 4, 2023

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: MacKay River Station number: AMS20
 Calibration Date: December 5, 2023 Last Cal Date: November 9, 2023
 Start time (MST): 8:35 End time (MST): 12:54
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.12 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC515997
 Removed Cal Gas Conc: 5.12 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1220
 ZAG Make/Model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1236656117
 Converter make: Global Converter serial #: 2022-226
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.980821	0.996103	Backgd or Offset: 3.07	3.19
Calibration intercept:	0.499202	0.439357	Coeff or Slope: 1.083	1.113

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4922	78.1	80.0	78.7	1.017
as found 2nd point	4961	39.0	39.9	39.5	1.014
as found 3rd point	4980	19.5	20.0	20.1	0.998
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4922	78.1	80.0	80.1	0.999
second point	4961	39.0	40.0	40.2	0.994
third point	4980	19.5	20.0	20.7	0.965
as left zero	5000	0.0	0.0	0.2	----
as left span	4922	78.1	80.0	78.9	1.014
SO2 Scrubber Check	4982	81.3	802.8	-0.1	----
Date of last scrubber change:	May 25, 2023			Ave Corr Factor	0.986
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 78.6 Prev response: 78.94 *% change: -0.4%
 Baseline Corr 2nd AF pt: 39.4 AF Slope: 0.981448 AF Intercept: 0.279067
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999974

* = > +/-5% change initiates investigation

Notes: No maintenance done. Span adjusted. Sox scrubber checked after the calibrator zero.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 9, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:35	End Time (MST):	12:54
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999946	≥0.995
80.0	80.1	0.9992			
40.0	40.2	0.9942	Slope	0.996103	0.90 - 1.10
20.0	20.7	0.9655			
			Intercept	0.439357	+/-3



H₂S Calibration Plot

Date: December 5, 2023

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 17, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:55	End Time (MST):	11:41
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.00	----	Correlation Coefficient	0.999987	≥0.995
17.34	17.31	1.0017			
8.68	8.71	0.9967	Slope	0.997735	0.90 - 1.10
4.33	4.37	0.9907			
			Intercept	0.026586	+/-1.5



THC Calibration Plot

Date: December 4, 2023

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: MacKay River Station number: AMS20
Calibration Date: November 30, 2023 Last Cal Date: November 8, 2023
Start time (MST): 8:00 End time (MST): 10:49
Reason: As Found Pump and Charcoal Change

Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 1220
ZAG make/model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.539	1.539	NO bkgnd or offset:	4.3	
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	4.3	
NO2 coeff or slope:	0.995	0.995	Reaction cell Press:	194.7	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996982	
NO _x Cal Offset:	2.850185	
NO Cal Slope:	0.997804	
NO Cal Offset:	1.471555	
NO ₂ Cal Slope:	1.001029	
NO ₂ Cal Offset:	-0.990724	



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as found span	4917	83.3	819.5	800.3	19.2	830.7	808.7	22.0	0.9865	0.9896
as found 2nd	4956	41.7	410.4	400.8	9.6	417.0	404.9	12.1	0.9843	0.9900
as found 3rd	4979	20.8	204.6	199.9	4.8	212.7	205.2	7.5	0.9621	0.9739
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

Average Correction Factor

Corrected As found	NO _x = 830.8 ppb	NO = 808.8 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 1.3%
Previous Response	NO _x = 819.8 ppb	NO = 800.0 ppb			*Percent Change	NO = 1.1%
Baseline Corr 2nd pt	NO _x = 417.1 ppb	NO = 405.0 ppb	As found	NO _x r ² : 0.999953	Nx SI: 1.011748	Nx Int: 2.230
Baseline Corr 3rd pt	NO _x = 212.8 ppb	NO = 205.3 ppb	As found	NO r ² : 0.999978	NO SI: 1.009183	NO Int: 1.211
			As found	NO ₂ r ² : 0.999944	NO ₂ SI: 1.004982	NO ₂ Int: -1.156

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	0.0	----	----
as found GPT point (400 ppb NO ₂)	796.8	454.7	361.3	362.7	0.9960	100.4%
as found GPT point (200 ppb NO ₂)	796.8	617.5	198.5	197.5	1.0048	99.5%
as found GPT point (100 ppb NO ₂)	796.8	702.9	113.1	111.3	1.0158	98.4%
1st GPT point (400 ppb O ₃)						
2nd GPT point (200 ppb O ₃)						
3rd GPT point (100 ppb O ₃)						

Average Correction Factor

Notes:

As founds to change the Pump and Charcoal.


Calibration Performed By:

Melissa Lemay

NO_x Calibration Plot

Date: November 30, 2023

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: MacKay River Station number: AMS20
Calibration Date: December 1, 2023 Last Cal Date: November 30, 2023
Start time (MST): 8:02 End time (MST): 12:14
Reason: Routine Pump and Charcoal Change, PMT adjusted

Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 1220
ZAG make/model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.005	0.981	NO bkgnd or offset:	4.4	4.0
NOX coeff or slope:	0.994	0.997	NOX bkgnd or offset:	4.5	4.1
NO2 coeff or slope:	1.006	0.995	Reaction cell Press:	190.6	190.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996982	1.007887
NO _x Cal Offset:	2.850185	2.589374
NO Cal Slope:	0.997804	0.995492
NO Cal Offset:	1.471555	1.631214
NO ₂ Cal Slope:	1.001029	1.005354
NO ₂ Cal Offset:	-0.990724	0.716865



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
high point	4917	83.3	819.5	800.3	19.2	827.6	797.9	29.7	0.9902	1.0030
second point	4956	41.7	410.4	400.8	9.6	416.4	400.4	16.0	0.9857	1.0011
third point	4979	20.8	204.6	199.9	4.8	212.2	202.8	9.4	0.9644	0.9855
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.0	----	----
as left span	4917	83.3	819.5	456.5	363.0	819.7	456.2	363.4	0.9997	1.0007
Average Correction Factor									0.9801	0.9965

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* => +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.1	451.3	363.0	365.6	0.9928	100.7%
2nd GPT point (200 ppb O3)	795.1	611.0	203.3	204.0	0.9964	100.4%
3rd GPT point (100 ppb O3)	795.1	697.0	117.3	120.5	0.9731	102.8%
Average Correction Factor					0.9874	101.3%

Notes: Calibration from Pump and Charcoal changed, and PMT adjusted November 30,2023. Zero and Span adjusted. Due to drifting during the GPT the 2nd NO ref point used.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 30, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	8:02	End Time (MST):	12:14
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
819.5	827.6	0.9902		
410.4	416.4	0.9857		
204.6	212.2	0.9644		
			0.999950	
			1.007887	
			2.589374	





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 30, 2023
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	8:02	End Time (MST):	12:14
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
800.3	797.9	1.0030		
400.8	400.4	1.0011		
199.9	202.8	0.9855		





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 30, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:02	End Time (MST):	12:14
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
363.0	365.6	0.9928		
203.3	204.0	0.9964		
117.3	120.5	0.9731		



NO_x Calibration Plot

Date: December 1, 2023

Location: MacKay River





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS21
CONKLIN**

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 30, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:30	End Time (MST):	14:46
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999990	≥0.995
800.8	798.2	1.0033			
400.4	399.8	1.0016	Slope	0.995935	0.90 - 1.10
200.1	201.7	0.9922			
			Intercept	0.975766	+/-30



SO2 Calibration Plot

Date: December 4, 2023

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Conklin Station number: AMS21
 Calibration Date: December 7, 2023 Last Cal Date: November 28, 2023
 Start time (MST): 11:15 End time (MST): 15:27
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.00 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CCS01204
 Removed Cal Gas Conc: 5.00 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3810
 ZAG Make/Model: Teledyne API 701H Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1236656116
 Converter make: CD-Nova 101 Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997143	0.995286	Backgd or Offset:	2.4	2.38
Calibration intercept:	0.200000	0.240000	Coeff or Slope:	0.963	0.963

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.0	80.0	80.0	1.000
as found 2nd point	4960	40.0	40.0	40.1	0.998
as found 3rd point	4980	20.0	20.0	20.3	0.985
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.0	80.0	79.8	1.003
second point	4960	40.0	40.0	40.1	0.998
third point	4980	20.0	20.0	20.3	0.985
as left zero	5000	0.0	0.0	0.3	----
as left span	4920	80.0	80.0	79.8	1.003
SO2 Scrubber Check	4920	80.2	802.0	0.1	----

Date of last scrubber change:	Ave Corr Factor	0.995
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 80.0 Prev response: 79.97 *% change: 0.0%
 Baseline Corr 2nd AF pt: 40.1 AF Slope: 0.998857 AF Intercept: 0.140000
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999984

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Scrubber check done after calibrator zero and it passed. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 28, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:15	End Time (MST):	15:27
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999986	≥0.995
80.0	79.8	1.0025			
40.0	40.1	0.9975	Slope	0.995286	0.90 - 1.10
20.0	20.3	0.9852			
			Intercept	0.240000	+/-3



TRS Calibration Plot

Date: December 7, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	December 4, 2023	Last Cal Date:	November 30, 2023
Start time (MST):	11:30	End time (MST):	14:46
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.82E-04	3.82E-04	NMHC SP Ratio:	9.44E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	96823
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.23	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.16	0.998
second point	4960	40.1	8.56	8.59	0.996
third point	4980	20.0	4.28	4.38	0.976
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.10	1.002
Average Correction Factor					0.990
Baseline Corr AF:	17.23	Prev response	17.11	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.24	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.19	0.994
second point	4960	40.1	4.57	4.59	0.995
third point	4980	20.0	2.28	2.33	0.978
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.17	0.997
Average Correction Factor					0.989
Baseline Corr AF:	9.24	Prev response	9.14	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.99	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.97	1.003
second point	4960	40.1	3.99	4.00	0.999
third point	4980	20.0	2.00	2.05	0.974
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.93	1.007
Average Correction Factor					0.992
Baseline Corr AF:	7.99	Prev response	7.97	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997585	1.000220
THC Cal Offset:	0.030767	0.040771
CH ₄ Cal Slope:	0.995028	0.995229
CH ₄ Cal Offset:	0.026958	0.025758
NMHC Cal Slope:	0.999806	1.004132
NMHC Cal Offset:	0.003609	0.014813

Notes: Changed the inlet filter and N2 cylinder after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 30, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:30	End Time (MST):	14:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999964	≥ 0.995			
17.13	17.16	0.9980						
8.56	8.59	0.9964				Slope	1.000220	0.90 - 1.10
4.28	4.38	0.9763						
			Intercept	0.040771	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 30, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:30	End Time (MST):	14:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999941	≥0.995			
7.99	7.97	1.0025						
3.99	4.00	0.9990				Slope	0.995229	0.90 - 1.10
2.00	2.05	0.9744						
			Intercept	0.025758	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 30, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:30	End Time (MST):	14:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999979	≥ 0.995			
9.14	9.19	0.9945						
4.57	4.59	0.9947				Slope	1.004132	0.90 - 1.10
2.28	2.33	0.9785						
			Intercept	0.014813	± 0.5			



NMHC Calibration Plot

Date: December 4, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	December 8, 2023	Last Cal Date:	December 4, 2023
Start time (MST):	10:36	End time (MST):	13:44
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.82E-04	4.00E-03	NMHC SP Ratio:	9.44E-05
CH ₄ Retention time:	14.2	14.6	NMHC Peak Area:	96823
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				97339
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	16.61	1.031
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.03	1.006
second point	4960	40.1	8.56	8.51	1.007
third point	4980	20.0	4.28	4.35	0.984
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.04	1.005
Average Correction Factor					0.999
Baseline Corr AF:	16.61	Prev response	17.17	*% change	-3.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.15	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.08	1.007
second point	4960	40.1	4.57	4.52	1.012
third point	4980	20.0	2.28	2.28	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.10	1.004
Average Correction Factor					1.007
Baseline Corr AF:	9.15	Prev response	9.19	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.46	1.070
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.95	1.005
second point	4960	40.1	3.99	3.99	1.002
third point	4980	20.0	2.00	2.07	0.963
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.94	1.006
Average Correction Factor					0.990
Baseline Corr AF:	7.46	Prev response	7.97	*% change	-6.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000220	0.992179
THC Cal Offset:	0.040771	0.036766
CH ₄ Cal Slope:	0.995229	0.991837
CH ₄ Cal Offset:	0.025758	0.036359
NMHC Cal Slope:	1.004132	0.992553
NMHC Cal Offset:	0.014813	-0.000393

Notes: Did not change filter; filter changed last December 4, 2023. Span adjusted.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	December 4, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:36	End Time (MST):	13:44
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999958	≥ 0.995			
17.13	17.03	1.0059						
8.56	8.51	1.0068				Slope	0.992179	0.90 - 1.10
4.28	4.35	0.9837						
			Intercept	0.036766	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	December 4, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:36	End Time (MST):	13:44
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999863	≥0.995			
7.99	7.95	1.0047						
3.99	3.99	1.0020				Slope	0.991837	0.90 - 1.10
2.00	2.07	0.9631						
			Intercept	0.036359	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 8, 2023	Previous Calibration:	December 4, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:36	End Time (MST):	13:44
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999989	≥ 0.995			
9.14	9.08	1.0069						
4.57	4.52	1.0117				Slope	0.992553	0.90 - 1.10
2.28	2.28	1.0025						
			Intercept	-0.000393	± 0.5			



NMHC Calibration Plot

Date: December 8, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	December 11, 2023	Last Cal Date:	December 4, 2023
Start time (MST):	11:28	End time (MST):	12:53
Reason:	Removal		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.82E-04	N/A	NMHC SP Ratio:	9.44E-05
CH ₄ Retention time:	14.2	N/A	NMHC Peak Area:	96823
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	16.79	1.020
as found 2nd point	4960	40.1	8.56	8.43	1.016
as found 3rd point	4980	20.0	4.27	4.32	0.988
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

			Average Correction Factor		
Baseline Corr AF:	16.79	Prev response	17.17	*% change	-2.3%
Baseline Corr 2nd AF:	8.4	AF Slope:	0.977974	AF Intercept:	0.060517
Baseline Corr 3rd AF:	4.3	AF Correlation:	0.999928	* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.11	1.003
as found 2nd point	4960	40.1	4.57	4.53	1.009
as found 3rd point	4980	20.0	2.28	2.28	1.001
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	9.11	Prev response	9.19	*% change	-0.9%
Baseline Corr 2nd AF:	4.5	AF Slope:	0.996321	AF Intercept:	-0.003579
Baseline Corr 3rd AF:	2.3	AF Correlation:	0.999988	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.68	1.040
as found 2nd point	4960	40.1	3.99	3.90	1.025
as found 3rd point	4980	20.0	1.99	2.05	0.974
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	7.68	Prev response	7.97	*% change	-3.8%
Baseline Corr 2nd AF:	3.90	AF Slope:	0.956920	AF Intercept:	0.063297
Baseline Corr 3rd AF:	2.05	AF Correlation:	0.999669	* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000220	N/A
THC Cal Offset:	0.040771	N/A
CH ₄ Cal Slope:	0.995229	N/A
CH ₄ Cal Offset:	0.025758	N/A
NMHC Cal Slope:	1.004132	N/A
NMHC Cal Offset:	0.014813	N/A


Notes: Baseline dipping during daily system checks. Removing instrument for further troubleshooting in the shop.

Calibration Performed By: Jan Castro

NMHC Calibration Plot

Date: December 11, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	December 11, 2023	Last Cal Date:	N/A
Start time (MST):	13:20	End time (MST):	16:00
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	N/A	4.00E-03	NMHC SP Ratio:	N/A	9.39E-05
CH ₄ Retention time:	N/A	14.6	NMHC Peak Area:	N/A	97339
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.12	1.000
second point	4960	40.1	8.56	8.50	1.007
third point	4980	20.0	4.28	4.30	0.996
as left zero	5000	0.0	0.00	0.02	----
as left span	4920	80.2	17.13	16.71	1.025

				Average Correction Factor	1.001
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.13	1.001
second point	4960	40.1	4.57	4.57	1.001
third point	4980	20.0	2.28	2.30	0.992
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	8.96	1.020
Average Correction Factor					0.998
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.99	1.000
second point	4960	40.1	3.99	3.93	1.016
third point	4980	20.0	2.00	2.00	1.000
as left zero	5000	0.0	0.00	0.02	----
as left span	4920	80.2	7.99	7.75	1.031
Average Correction Factor					1.005
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	N/A	0.998913
THC Cal Offset:	N/A	-0.004433
CH ₄ Cal Slope:	N/A	0.999250
CH ₄ Cal Offset:	N/A	-0.013042
NMHC Cal Slope:	N/A	0.998205
NMHC Cal Offset:	N/A	0.009009

Notes: Installation Calibration. Span adjusted.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 11, 2023	Previous Calibration:	N/A
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	13:20	End Time (MST):	16:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999978	≥0.995
17.13	17.12	1.0003			
8.56	8.50	1.0074	Slope	0.998913	0.90 - 1.10
4.28	4.30	0.9959			
			Intercept	-0.004433	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 11, 2023	Previous Calibration:	N/A
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	13:20	End Time (MST):	16:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999914	≥0.995			
7.99	7.99	0.9999						
3.99	3.93	1.0161				Slope	0.999250	0.90 - 1.10
2.00	2.00	1.0003						
			Intercept	-0.013042	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 11, 2023	Previous Calibration:	N/A
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	13:20	End Time (MST):	16:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999993	≥ 0.995			
9.14	9.13	1.0010						
4.57	4.57	1.0006				Slope	0.998205	0.90 - 1.10
2.28	2.30	0.9916						
			Intercept	0.009009	± 0.5			



NMHC Calibration Plot

Date: December 11, 2023

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	December 14, 2023	Last Cal Date:	December 11, 2023
Start time (MST):	10:56	End time (MST):	13:45
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.00E-03	3.69E-04	NMHC SP Ratio:	9.39E-05
CH ₄ Retention time:	14.6	17.2	NMHC Peak Area:	97339
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				122336
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.04	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.10	1.001
second point	4960	40.1	8.56	8.57	0.999
third point	4980	20.0	4.28	4.34	0.986
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.16	0.998

Average Correction Factor				0.996
Baseline Corr AF:	17.04	Prev response	17.10	*% change -0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	8.89	1.028
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.14	1.000
second point	4960	40.1	4.57	4.59	0.995
third point	4980	20.0	2.28	2.32	0.984
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.15	0.999
Average Correction Factor					0.993
Baseline Corr AF:	8.89	Prev response	9.13	*% change	-2.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	8.15	0.980
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.96	1.003
second point	4960	40.1	3.99	3.98	1.003
third point	4980	20.0	2.00	2.02	0.989
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	8.02	0.996
Average Correction Factor					0.998
Baseline Corr AF:	8.15	Prev response	7.97	*% change	2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998913	0.997358
THC Cal Offset:	-0.004433	0.029967
CH ₄ Cal Slope:	0.999250	0.995501
CH ₄ Cal Offset:	-0.013042	0.011557
NMHC Cal Slope:	0.998205	0.998993
NMHC Cal Offset:	0.009009	0.018610

Notes: Changed inlet filters after as founds. Span adjusted.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 14, 2023	Previous Calibration:	December 11, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:56	End Time (MST):	13:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999984	≥ 0.995			
17.13	17.10	1.0015						
8.56	8.57	0.9992				Slope	0.997358	0.90 - 1.10
4.28	4.34	0.9865						
			Intercept	0.029967	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 14, 2023	Previous Calibration:	December 11, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:56	End Time (MST):	13:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999983	≥0.995			
7.99	7.96	1.0033						
3.99	3.98	1.0033				Slope	0.995501	0.90 - 1.10
2.00	2.02	0.9889						
			Intercept	0.011557	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 14, 2023	Previous Calibration:	December 11, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:56	End Time (MST):	13:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999981	≥ 0.995
9.14	9.14	0.9999			
4.57	4.59	0.9954			
2.28	2.32	0.9844			
			Slope	0.998993	0.90 - 1.10
			Intercept	0.018610	+/-0.5



NMHC Calibration Plot

Date: December 14, 2023

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Conklin
Calibration Date: December 6, 2023
Start time (MST): 11:35
Reason: Routine
Station number: AMS21
Last Cal Date: November 29, 2023
End time (MST): 16:02

Calibration Standards

NO Gas Cylinder #: T2Y1P1H
NOX Cal Gas Conc: 51.09 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 51.09 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 50.39 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 50.39 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 691

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1501663731

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.094	1.077	NO bkgnd or offset:	10.9	10.8
NOX coeff or slope:	1.000	0.999	NOX bkgnd or offset:	11.1	10.9
NO2 coeff or slope:	0.999	1.000	Reaction cell Press:	167.3	151.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003588	0.992390
NO _x Cal Offset:	2.465024	2.363089
NO Cal Slope:	1.004807	0.994482
NO Cal Offset:	1.242275	1.080383
NO ₂ Cal Slope:	1.006340	0.991295
NO ₂ Cal Offset:	-0.474448	1.480919



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.4	0.1	----	----
as found span	4921	79.4	811.2	800.1	11.1	822.6	811.4	11.2	0.9862	0.9861
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.2	----	----
high point	4921	79.4	811.2	800.1	11.1	806.2	796.1	10.1	1.0063	1.0051
second point	4960	39.7	405.7	400.1	5.6	406.1	399.7	6.3	0.9990	1.0011
third point	4980	19.8	202.3	199.6	2.8	205.8	200.9	4.9	0.9831	0.9933
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.4	0.1	----	----
as left span	4921	79.4	811.2	379.0	432.2	807.1	378.1	429.0	1.0051	1.0025
Average Correction Factor									0.9961	0.9998

Corrected As found	NO _x = 822.9 ppb	NO = 811.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.8%
Previous Response	NO _x = 816.6 ppb	NO = 805.2 ppb		*Percent Change	NO = 0.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI: ;

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	793.5	372.4	432.2	429.6	1.0061	99.4%
2nd GPT point (200 ppb O3)	793.5	585.1	219.5	218.7	1.0037	99.6%
3rd GPT point (100 ppb O3)	793.5	690.2	114.4	116.9	0.9787	102.2%
Average Correction Factor					0.9962	100.4%

Notes: Changed the inlet filter after as founds. Adjusted Span only. Used the 2nd reference point because of drift.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 29, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:35	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
811.2	806.2	1.0063		
405.7	406.1	0.9990		
202.3	205.8	0.9831		
			0.999954	
			0.992390	
			2.363089	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 29, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:35	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
800.1	796.1	1.0051			
400.1	399.7	1.0011			
199.6	200.9	0.9933			
			Slope	0.994482	0.90 - 1.10
			Intercept	1.080383	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 29, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:35	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
432.2	429.6	1.0061		
219.5	218.7	1.0037		
114.4	116.9	0.9787		



NO_x Calibration Plot

Date: December 6, 2023

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 22, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:49	End Time (MST):	15:48
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999992	≥0.995
400.0	400.4	0.9990			
200.0	201.2	0.9940	Slope	1.001229	0.90 - 1.10
100.0	100.4	0.9960			
			Intercept	0.260000	+/- 5



O₃ Calibration Plot

Date: December 5, 2023

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Conklin Station number: AMS 21
 Calibration Date: December 7, 2023 Last Cal Date: November 30, 2023
 Start time (MST): 12:00 End time (MST): 13:54

Analyzer Make: API T640 S/N: 326
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT S/N: 388751
 Temp/RH standard: ALICAT S/N: 388751

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2	2.48	2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	700.7	702.65	700.7	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	5.05	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 7, 2023</u>	Last Cal Date: <u>November 30, 2023</u>			
	PM w/o HEPA: <u>2.6</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.7	10.1	10.7	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>3.3</u>	w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:		<u>December 7, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>December 7, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: December 7, 2023
 Date RH/T Sensor Cleaned: December 7, 2023

Notes: Verified flow, temperature, and pressure. Leak check passed. PMT adjusted after maintenance.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER**

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 3, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:44	End Time (MST):	15:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999998	≥0.995
799.8	798.4	1.0017			
399.9	399.9	1.0000	Slope	0.998449	0.90 - 1.10
200.4	200.6	0.9992			
			Intercept	0.164515	+/-30



SO2 Calibration Plot

Date: December 5, 2023

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Janvier Station number: AMS22
 Calibration Date: December 19, 2023 Last Cal Date: November 9, 2023
 Start time (MST): 10:45 End time (MST): 14:53
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.03 ppm Cal Gas Exp Date: April 16, 2022
 Cal Gas Cylinder #: DT0018680
 Removed Cal Gas Conc: 5.03 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680031
 Converter make: CDN-101 Converter serial #: 587
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.986936	0.997936	Backgd or Offset:	3.04	3.59
Calibration intercept:	0.621094	0.260962	Coeff or Slope:	1.161	1.179

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4920	79.5	80.0	79.0	1.019
as found 2nd point					
as found 3rd point					
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	79.5	80.0	80.0	1.000
second point	4960	39.8	40.0	40.3	0.994
third point	4980	19.9	20.0	20.4	0.981
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	79.5	80.0	79.5	1.006
SO2 Scrubber Check	4920	79.8	798.0	0.0	----

Date of last scrubber change:	Ave Corr Factor	0.992
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 78.5 Prev response: 79.56 *% change: -1.4%
 Baseline Corr 2nd AF pt: NA AF Slope: NA AF Intercept: NA
 Baseline Corr 3rd AF pt: NA AF Correlation: NA

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Brought a new cal gas cylinder to replace but the new cylinder is leaking, will continue using the old cylinder for now. Ran a SO2 scrubber check after calibrator zero. Adjusted zero and span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 9, 2023
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	10:45	End Time (MST):	14:53
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999981	≥0.995
80.0	80.0	0.9998			
40.0	40.3	0.9936	Slope	0.997936	0.90 - 1.10
20.0	20.4	0.9814			
			Intercept	0.260962	+/-3



TRS Calibration Plot

Date: December 19, 2023

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	December 5, 2023	Last Cal Date:	November 3, 2023
Start time (MST):	10:43	End time (MST):	15:00
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH ₄ Cal Gas Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
C ₃ H ₈ Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
Removed C ₃ H ₈ Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.470E-04	2.560E-04	NMHC SP Ratio:	4.94E-05
CH ₄ Retention time:	15.0	15.2	NMHC Peak Area:	185124
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	16.46	1.043
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.16	1.001
second point	4960	39.9	8.59	8.60	0.999
third point	4980	20.0	4.30	4.33	0.994
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.18	1.000
Average Correction Factor					0.998

Baseline Corr AF:	16.46	Prev response	17.12	*% change	-4.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	8.76	1.045
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.15	9.15	1.000
second point	4960	39.9	4.57	4.58	0.999
third point	4980	20.0	2.29	2.31	0.992
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.15	9.15	1.000
Average Correction Factor					0.997
Baseline Corr AF:	8.76	Prev response	9.10	*% change	-3.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	7.71	1.041
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.02	1.001
second point	4960	39.9	4.01	4.02	0.999
third point	4980	20.0	2.01	2.02	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.03	1.000
Average Correction Factor					0.998
Baseline Corr AF:	7.71	Prev response	8.02	*% change	-4.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.993524	0.998979
THC Cal Offset:	0.058796	0.014795
CH ₄ Cal Slope:	0.998422	0.998920
CH ₄ Cal Offset:	0.006635	0.005636
NMHC Cal Slope:	0.989089	0.999269
NMHC Cal Offset:	0.052961	0.008959

Notes: Changed the inlet filter after as founds. Adjusted the span only. Multipoint as found completed for SO₂ maintenance.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 3, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:43	End Time (MST):	15:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
17.17	17.16	1.0005						
8.59	8.60	0.9988				Slope	0.998979	0.90 - 1.10
4.30	4.33	0.9937						
			Intercept	0.014795	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary

Version-06-2022


Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 3, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:43	End Time (MST):	15:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥0.995
8.03	8.02	1.0007			
4.01	4.02	0.9989			
2.01	2.02	0.9956			
			Slope	0.998920	0.90 - 1.10
			Intercept	0.005636	+/-0.5

CH₄ Calibration Curve





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 3, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:43	End Time (MST):	15:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995
9.15	9.15	1.0001			
4.57	4.58	0.9986			
2.29	2.31	0.9920			
			Slope	0.999269	0.90 - 1.10
			Intercept	0.008959	+/-0.5



NMHC Calibration Plot

Date: December 5, 2023

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: December 19, 2023 Last Cal Date: December 5, 2023
 Start time (MST): 9:20 End time (MST): 10:43
 Reason: Cylinder Change N2 cylinder change

Calibration Standards

Gas Cert Reference: CC281519 Cal Gas Expiry Date: January 18, 2029
 CH₄ Cal Gas Conc. 502.8 ppm CH₄ Equiv Conc. 1075.9 ppm
 C₃H₈ Cal Gas Conc. 208.4 ppm
 Removed Gas Cert: Removed Gas Expiry:
 Removed CH₄ Conc. 502.8 ppm CH₄ Equiv Conc. 1075.9 ppm
 Removed C₃H₈ Conc. 208.4 ppm Diff between cyl (THC):
 Diff between cyl (CH₄): Diff between cyl (NM):
 Calibrator Model: Teledyne API 700 Serial Number: 3806
 ZAG make/model: Teledyne API 701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 1331259520
 THC Range (ppm): 0 - 20 ppm
 NMHC Range (ppm): 0 - 10 ppm CH₄ Range (ppm): 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.560E-04	2.560E-04	NMHC SP Ratio:	5.17E-05	5.17E-05
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	176965	176965
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	17.12	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.54	0.979
second point					
third point					
as left zero					
as left span					
Average Correction Factor					0.979

Baseline Corr AF: 17.12 Prev response 17.17 *% change -0.3%
 Baseline Corr 2nd AF: NA AF Slope: AF Intercept:
 Baseline Corr 3rd AF: NA AF Correlation: * = +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	9.19	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.15	9.22	0.992
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.992
Baseline Corr AF:	9.19	Prev response	9.15	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	7.93	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.31	0.966
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.966
Baseline Corr AF:	7.93	Prev response	8.02	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998979	1.021427
THC Cal Offset:	0.014795	0.000000
CH ₄ Cal Slope:	0.998920	1.035513
CH ₄ Cal Offset:	0.005636	0.000000
NMHC Cal Slope:	0.999269	1.007976
NMHC Cal Offset:	0.008959	0.000000


Notes: Changed the N2 cylinder after as founds.

Calibration Performed By: Max Farrell

NMHC Calibration Plot

Date: December 19, 2023

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: December 15, 2023 Last Cal Date: November 8, 2023
Start time (MST): 10:03 End time (MST): 15:20
Reason: Cylinder Change

Calibration Standards

NO Gas Cylinder #:	DT0047765	Cal Gas Expiry Date:	March 11, 2031
NOX Cal Gas Conc:	48.90 ppm	NO Cal Gas Conc:	48.80 ppm
Removed Cylinder #:	CC424183	Removed Gas Exp Date:	April 16, 2023
Removed Gas NOX Conc:	48.60 ppm	Removed Gas NO Conc:	48.60 ppm
NOX gas Diff:	-0.8%	NO gas Diff:	0.7%
Calibrator Model:	Teledyne API T700	Serial Number:	3806
ZAG make/model:	Teledyne API T701	Serial Number:	201

Analyzer Information

Analyzer make:	Analyzer serial #:			
NOX Range (ppb):	0 - 1000 ppb			
	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.826	0.834	NO bkgnd or offset:	-5.6
NOX coeff or slope:	0.815	0.827	NOX bkgnd or offset:	-3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	5.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.990857	0.994409
NO _x Cal Offset:	2.110476	2.604040
NO Cal Slope:	0.989570	0.998646
NO Cal Offset:	1.211114	1.684010
NO ₂ Cal Slope:	1.001094	0.995534
NO ₂ Cal Offset:	0.263799	-1.435399



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.6	2.1	-0.5	----	----
as found span	4918	82.3	799.9	799.9	0.0	789.5	783.3	6.2	1.0132	1.0212
as found 2nd										
as found 3rd										
new cyl resp	4918	82.0	802.0	800.3	1.6	785.5	789.5	-4.0	1.0210	1.0137
calibrator zero	5000	0.0	0.0	0.0	0.0	1.7	2.0	-0.4	----	----
high point	4918	82.0	802.0	800.3	1.6	799.3	800.8	-1.5	1.0033	0.9994
second point	4960	41.0	400.9	400.1	0.8	402.7	401.9	0.8	0.9955	0.9955
third point	4980	20.5	200.5	200.1	0.4	202.2	200.6	1.7	0.9914	0.9973
as left zero	5000	0.0	0.0	0.0	0.0	1.8	2.6	-0.8	----	----
as left span	4918	82.0	802.0	406.8	395.1	797.1	406.4	390.7	1.0061	1.0010
Average Correction Factor									0.9968	0.9974

Corrected As found	NO _x = 787.9 ppb	NO = 781.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.9%
Previous Response	NO _x = 794.7 ppb	NO = 792.8 ppb		*Percent Change	NO = -1.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.3	404.8	395.1	392.3	1.0072	99.3%
2nd GPT point (200 ppb O3)	798.3	605.3	194.6	192.3	1.0122	98.8%
3rd GPT point (100 ppb O3)	798.3	700.4	99.5	96.3	1.0336	96.7%
Average Correction Factor					1.0177	98.3%

Notes:

Changed the cal gas cylinder. Adjusted the span.

Calibration Performed By:

Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 15, 2023	Previous Calibration:	November 8, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:03	End Time (MST):	15:20
Analyzer make:		Analyzer serial #:	

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	1.7	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
802.0	799.3	1.0033		
400.9	402.7	0.9955		
200.5	202.2	0.9914		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 15, 2023	Previous Calibration:	November 8, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:03	End Time (MST):	15:20
Analyzer make:		Analyzer serial #:	

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	2.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
800.3	800.8	0.9994			0.999996
400.1	401.9	0.9955			0.998646
200.1	200.6	0.9973			1.684010

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 15, 2023	Previous Calibration:	November 8, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:03	End Time (MST):	15:20
Analyzer make:		Analyzer serial #:	

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.4	----	Correlation Coefficient	≥0.995	
395.1	392.3	1.0072			
194.6	192.3	1.0122	Slope	0.90 - 1.10	
99.5	96.3	1.0336			
			Intercept	-1.435399	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 15, 2023

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: December 11, 2023 Last Cal Date: November 23, 2023
 Start time (MST): 10:48 End time (MST): 13:47
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701H Serial Number: 691

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 7046
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006114	1.003171	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	0.780000	0.820000	Coeff or Slope:	1.021	1.021

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	1.0	----
as found span	4895	905.3	400.0	403.3	0.992
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.6	----
high point	4895	905.3	400.0	401.9	0.995
second point	4895	756.7	200.0	201.8	0.991
third point	4895	656.1	100.0	101.2	0.988
as left zero	5000	800.0	0.0	0.8	----
as left span	4895	904.3	400.0	403.8	0.991
Average Correction Factor					0.991

Baseline Corr As found:	402.3	Previous response	403.2	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 11, 2023	Previous Calibration:	November 23, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:48	End Time (MST):	13:47
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.6	----	Correlation Coefficient	≥0.995
400.0	401.9	0.9953		
200.0	201.8	0.9911	Slope	0.90 - 1.10
100.0	101.2	0.9881		
			Intercept	+/- 5



O₃ Calibration Plot

Date: December 11, 2023

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: December 15, 2023 Last Cal Date: November 23, 2023
 Start time (MST): 13:58 End time (MST): 14:18

Analyzer Make: Teledyne API T640 S/N: 325
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450
 Temp/RH standard: Delta Cal S/N: 1450

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	3.70	3.75	3.70	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	718.6	719.4	718.6	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	5.10	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>December 15, 2023</u>	Last Cal Date: <u>November 23, 2023</u>			
	PM w/o HEPA: <u>3.9</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>October 20, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>October 20, 2023</u>			

Annual Maintenance

Date Sample Tube Cleaned: July 26, 2023
 Date RH/T Sensor Cleaned: July 26, 2023

Notes:

Quarterly calibration was completed in October. No adjustments needed. Leak check passed.

Calibration by: Max Farrell



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23
FORT HILLS

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 20, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:45	End Time (MST):	11:28
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.4	----	Correlation Coefficient	0.999992	≥0.995
799.1	800.5	0.9982			
400.1	399.5	1.0014	Slope	1.001895	0.90 - 1.10
200.0	198.7	1.0067			
			Intercept	-0.683633	+/-30



SO2 Calibration Plot

Date: December 13, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Hills Station number: AMS23
 Calibration Date: December 6, 2023 Last Cal Date: November 15, 2023
 Start time (MST): 8:20 End time (MST): 12:18
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.20 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517372
 Removed Cal Gas Conc: 5.20 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 451
 ZAG Make/Model: API T701 Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1300156232
 Converter make: CDN-101 Converter serial #: 594
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000174	0.991033	Backgd or Offset: 1.19	1.19
Calibration intercept:	-0.118216	-0.198385	Coeff or Slope: 1.124	1.124

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4923	77.0	80.0	78.0	1.024
as found 2nd point	4962	38.5	40.0	39.4	1.013
as found 3rd point	4981	19.2	19.9	19.6	1.013
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4923	77.0	80.0	79.1	1.011
second point	4962	38.5	40.0	39.5	1.013
third point	4981	19.2	19.9	19.4	1.028
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.0	80.0	79.6	1.005
SO2 Scrubber Check	4920	80.3	803.0	0.2	----
Date of last scrubber change:				Ave Corr Factor	1.017
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 78.1 Prev response: 79.90 *% change: -2.3%
 Baseline Corr 2nd AF pt: 39.5 AF Slope: 0.975890 AF Intercept: 0.081410
 Baseline Corr 3rd AF pt: 19.7 AF Correlation: 0.999958

* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 15, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:20	End Time (MST):	12:18
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999988	≥0.995
80.0	79.1	1.0114			
40.0	39.5	1.0126	Slope	0.991033	0.90 - 1.10
19.9	19.4	1.0282			
			Intercept	-0.198385	+/-3



TRS Calibration Plot

Date: December 6, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	December 13, 2023	Last Cal Date:	November 20, 2023
Start time (MST):	8:45	End time (MST):	11:27
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
Removed C ₃ H ₈ Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.34E-04	2.36E-04	NMHC SP Ratio:	5.01E-05	5.09E-05
CH ₄ Retention time:	13.0	13.0	NMHC Peak Area:	183497	180787
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	17.05	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.23	0.998
second point	4960	40.2	8.61	8.64	0.996
third point	4980	20.1	4.30	4.35	0.989
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.19	17.21	0.999
Average Correction Factor					0.994

Baseline Corr AF:	17.05	Prev response	17.30	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.07	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.18	0.998
second point	4960	40.2	4.59	4.63	0.990
third point	4980	20.1	2.29	2.36	0.972
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	9.16	9.17	0.999
Average Correction Factor					0.987
Baseline Corr AF:	9.07	Prev response	9.24	*% change	-1.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	7.98	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.06	0.997
second point	4960	40.2	4.02	4.01	1.003
third point	4980	20.1	2.01	1.98	1.016
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.03	8.05	0.998
Average Correction Factor					1.005
Baseline Corr AF:	7.98	Prev response	8.06	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.006220	1.001568
THC Cal Offset:	-0.002409	0.017601
CH ₄ Cal Slope:	1.005350	1.004639
CH ₄ Cal Offset:	-0.020045	-0.020048
NMHC Cal Slope:	1.006983	1.000371
NMHC Cal Offset:	0.017636	0.031648

Notes:

Span adjusted. No Maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 20, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:45	End Time (MST):	11:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995			
17.19	17.23	0.9978						
8.61	8.64	0.9962				Slope	1.001568	0.90 - 1.10
4.30	4.35	0.9893						
			Intercept	0.017601	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 20, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:45	End Time (MST):	11:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999972	≥0.995
8.03	8.06	0.9966			
4.02	4.01	1.0029			
2.01	1.98	1.0155			
			Slope	1.004639	0.90 - 1.10
			Intercept	-0.020048	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 20, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:45	End Time (MST):	11:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999945	≥ 0.995			
9.16	9.18	0.9977						
4.59	4.63	0.9904				Slope	1.000371	0.90 - 1.10
2.29	2.36	0.9715						
			Intercept	0.031648	± 0.5			



NMHC Calibration Plot

Date: December 13, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Hills Station number: AMS23
Calibration Date: December 7, 2023 Last Cal Date: November 7, 2023
Start time (MST): 7:35 End time (MST): 12:04
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC332703 Cal Gas Expiry Date: January 28, 2024
NOX Cal Gas Conc: 49.7 ppm NO Cal Gas Conc: 49.7 ppm
Removed Cylinder #: N/A Removed Gas Exp Date: N/A
Removed Gas NOX Conc: 49.7 ppm Removed Gas NO Conc: 49.7 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 451
ZAG make/model: Teledyne API T701 Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1152430007
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.089	1.081	NO bkgnd or offset:	3	3
NOX coeff or slope:	0.993	0.989	NOX bkgnd or offset:	3.2	3.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	162.0	162

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003266	0.999822
NO _x Cal Offset:	0.024268	0.304106
NO Cal Slope:	1.001225	1.001566
NO Cal Offset:	-1.036547	-0.455910
NO ₂ Cal Slope:	1.003745	1.002916
NO ₂ Cal Offset:	0.698395	-0.422416



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
as found span	4920	80.5	800.2	800.2	0.0	812.9	808.3	4.5	0.984	0.990
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
high point	4920	80.5	800.2	800.2	0.0	799.8	800.8	-1.0	1.000	0.999
second point	4960	40.2	399.6	399.6	0.0	401.0	400.7	0.4	0.996	0.997
third point	4980	20.1	199.8	199.8	0.0	199.8	198.4	1.4	1.000	1.007
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	-0.1	----	----
as left span	4920	80.5	800.2	448.5	351.7	799.3	447.3	352.1	1.001	1.003
Average Correction Factor									0.999	1.001

Corrected As found	NO _x = 813.1 ppb	NO = 808.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.3%
Previous Response	NO _x = 802.8 ppb	NO = 800.1 ppb		*Percent Change	NO = 1.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.1	447.4	351.7	352.8	0.997	100.3%
2nd GPT point (200 ppb O3)	799.1	624.7	174.4	173.3	1.006	99.4%
3rd GPT point (100 ppb O3)	799.1	711.5	87.6	87.8	0.998	100.2%
Average Correction Factor					1.000	100.0%

Notes:

No maintenance done. Span adjusted.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 7, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:35	End Time (MST):	12:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.2	799.8	1.0005		
399.6	401.0	0.9964		
199.8	199.8	0.9999		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 7, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:35	End Time (MST):	12:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.2	800.8	0.9992		
399.6	400.7	0.9972		
199.8	198.4	1.0070		
			0.999992	
			1.001566	
			-0.455910	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 7, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:35	End Time (MST):	12:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
351.7	352.8	0.9969			
174.4	173.3	1.0063			
87.6	87.8	0.9977			
			Correlation Coefficient	0.999972	≥0.995
			Slope	1.002916	0.90 - 1.10
			Intercept	-0.422416	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 7, 2023

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Fort Hills Station number: AMS 23
 Calibration Date: December 13, 2023 Last Cal Date: November 20, 2023
 Start time (MST): 8:15 End time (MST): 8:44

Analyzer Make: API T640 S/N: 1546
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2.4	2.2	2.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	724.8	724.3	724.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.05	5.14	5.05	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: December 13, 2023 Last Cal Date: November 20, 2023
 PM w/o HEPA: 2.2 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: _____ w/ HEPA: _____
 Date Optical Chamber Cleaned: November 20, 2023 <0.2 ug/m3
 Disposable Filter Changed: November 20, 2023

Annual Maintenance

Date Sample Tube Cleaned: October 17, 2023
 Date RH/T Sensor Cleaned: October 17, 2023

No adjustments done. Leak check passed. Head cleaned.

Notes:

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25
WASKŌW OHCI PIMÂTISIWIN

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number:	AMS25
Calibration Date:	December 19, 2023	Last Cal Date:	November 20, 2023
Start time (MST):	7:45	End time (MST):	10:52
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.54	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC437219			
Removed Cal Gas Conc:	50.54	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	747
ZAG Make/Model:	API T701		Serial Number:	4765

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995565	0.999891	Backgd or Offset:	9.9	11.0
Calibration intercept:	0.584063	-0.656026	Coeff or Slope:	1.011	1.020

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.7	----
as found span	4921	79.2	800.5	794.1	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4921	79.2	800.5	799.7	1.001
second point	4960	39.6	400.3	400.3	1.000
third point	4980	19.8	200.1	198.4	1.009
as left zero	5000	0.0	0.0	-0.2	----
as left span	4921	79.2	800.5	798.8	1.002
Average Correction Factor					1.003

Baseline Corr As found:	793.40	Previous response	797.56	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No maintenance done. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	November 20, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	7:45	End Time (MST):	10:52
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999995	≥0.995
800.5	799.7	1.0010			
400.3	400.3	1.0000	Slope	0.999891	0.90 - 1.10
200.1	198.4	1.0088			
			Intercept	-0.656026	+/-30



SO2 Calibration Plot

Date: December 19, 2023

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS25
 Calibration Date: December 15, 2023 Last Cal Date: November 24, 2023
 Start time (MST): 7:20 End time (MST): 11:43
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC517099
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 747
 ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
 Converter make: Global G-150 Converter serial #: 2022-219
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998073	1.003251	Backgd or Offset: 3.30	3.30
Calibration intercept:	0.180000	0.200000	Coeff or Slope: 1.108	1.108

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4920	80.0	79.5	79.8	0.997
as found 2nd point	4960	40.0	39.7	40.1	0.993
as found 3rd point	4980	20.0	19.9	20.1	0.993
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.0	79.5	79.8	0.996
second point	4960	40.0	39.7	40.3	0.986
third point	4980	20.0	19.9	20.1	0.988
as left zero	5000	0.0	0.0	0.3	----
as left span	4920	80.0	800.0	801.7	0.998
SO2 Scrubber Check	4921	79.2	800.0	0.1	----
Date of last scrubber change:	20-Jun-23			Ave Corr Factor	0.990
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.7 Prev response: 79.48 *% change: 0.3%
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 1.002963 AF Intercept: 0.160000
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999996

* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 15, 2023	Previous Calibration:	November 24, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	7:20	End Time (MST):	11:43
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999976	
79.5	79.8	0.9957			≥0.995
39.7	40.3	0.9858	Slope	1.003251	
19.9	20.1	0.9883			0.90 - 1.10
			Intercept	0.200000	+/-3



H₂S Calibration Plot

Date: December 15, 2023

Location: Waskow ohci Pimatisiwin





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS26
CHRISTINA LAKE**

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 29, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	14:47	End Time (MST):	17:27
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.1	----	Correlation Coefficient	0.999996	≥0.995
800.0	798.3	1.0021			
400.6	399.3	1.0032	Slope	0.996949	0.90 - 1.10
201.6	200.7	1.0046			
			Intercept	0.370507	+/-30



SO2 Calibration Plot

Date: December 12, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Christina Lake Station number: AMS26
 Calibration Date: December 12, 2023 Last Cal Date: November 29, 2023
 Start time (MST): 10:57 End time (MST): 14:48
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.89 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002466
 Removed Cal Gas Conc: 4.89 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 281
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 450i Analyzer serial #: 1180030032
 Converter make: NA Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002002	1.001571	Backgd or Offset: 33.9	33.9
Calibration intercept:	0.607400	0.647458	Coeff or Slope: 1.060	1.060

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4930	81.9	79.9	81.7	0.980
as found 2nd point	4972	41.0	40.0	41.4	0.971
as found 3rd point	4994	20.6	20.1	21.1	0.961
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.6	----
high point	4930	81.9	79.9	80.6	0.991
second point	4972	41.0	40.0	40.9	0.978
third point	4994	20.6	20.1	20.7	0.970
as left zero	5000	0.0	0.0	0.8	----
as left span	4930	81.9	79.9	81.3	0.983
SO2 Scrubber Check	4931	80.9	807.1	0.1	----
Date of last scrubber change:	27-Feb-19			Ave Corr Factor	0.980
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 81.5 Prev response: 80.68 *% change: 1.0%
 Baseline Corr 2nd AF pt: 41.2 AF Slope: 1.018750 AF Intercept: 0.446266
 Baseline Corr 3rd AF pt: 20.9 AF Correlation: 0.999955

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after MAF's. Ran scrubber check after calibrator zero passed. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 29, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:57	End Time (MST):	14:48
Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999985	
79.9	80.6	0.9914			≥0.995
40.0	40.9	0.9778	Slope	1.001571	
20.1	20.7	0.9704			0.90 - 1.10
			Intercept	0.647458	+/-3



H₂S Calibration Plot

Date: December 12, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 19, 2023	Previous Calibration:	December 12, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	11:06	End Time (MST):	15:32
Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999969	≥0.995
80.0	80.1	0.9986			
40.0	39.9	1.0025	Slope	1.003426	0.90 - 1.10
20.0	20.2	0.9900			
			Intercept	-0.141605	+/-3



H₂S Calibration Plot

Date: December 19, 2023

Location: Christina Lake





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.1	-0.7	----	----
as found span	4920	80.0	813.1	800.3	12.8	811.5	797.1	14.5	1.0020	1.0040
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.7	0.0	-0.7	----	----
high point	4920	80.0	813.1	800.3	12.8	809.8	796.2	13.6	1.0041	1.0052
second point	4960	40.0	406.6	400.2	6.4	406.1	399.0	7.1	1.0011	1.0029
third point	4980	20.0	203.3	200.1	3.2	203.3	198.8	4.5	0.9999	1.0064
as left zero	5000	0.0	0.0	0.0	0.0	-0.7	0.0	-0.7	----	----
as left span	4920	80.0	813.1	410.6	402.5	808.0	404.2	403.8	1.0063	1.0159
Average Correction Factor									1.0017	1.0048

Corrected As found	NO _x = 812.3 ppb	NO = 797.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.4%
Previous Response	NO _x = 808.7 ppb	NO = 795.1 ppb		*Percent Change	NO = 0.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.1	404.4	402.5	403.0	0.9988	100.1%
2nd GPT point (200 ppb O3)	794.1	597.6	209.3	210.2	0.9957	100.4%
3rd GPT point (100 ppb O3)	794.1	693.4	113.5	114.4	0.9921	100.8%
Average Correction Factor					0.9955	100.4%

Notes: Changed sample inlet filter after as founds. No adjustment made. Used the 2nd reference point.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 28, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	8:27	End Time (MST):	12:48
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.7	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.1	809.8	1.0041		
406.6	406.1	1.0011		
203.3	203.3	0.9999		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 28, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	8:27	End Time (MST):	12:48
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
800.3	796.2	1.0052		
400.2	399.0	1.0029		
200.1	198.8	1.0064		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 28, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	8:27	End Time (MST):	12:48
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.7	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
402.5	403.0	0.9988		
209.3	210.2	0.9957		
113.5	114.4	0.9921		
			0.999986	
			1.002313	
			-0.019410	



NO_x Calibration Plot

Date: December 13, 2023

Location: Christina Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27
JACKFISH 2/3

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	December 12, 2023	Last Cal Date:	November 14, 2023
Start time (MST):	10:02	End time (MST):	12:50
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	<u>SG9133974BAL</u>			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3811
ZAG Make/Model:	API 701		Serial Number:	268

Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	12124313138
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999320	0.989063	Backgd or Offset:	7.7	7.6
Calibration intercept:	-2.038541	-2.298638	Coeff or Slope:	0.950	0.950

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4921	79.1	800.2	787.9	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.3	----
high point	4921	79.1	800.2	790.5	1.012
second point	4961	39.5	399.5	391.3	1.021
third point	4980	19.8	200.3	193.4	1.036
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	79.1	800.2	792.5	1.010
Average Correction Factor					1.023

Baseline Corr As found:	787.70	Previous response	797.58	*% change	-1.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 14, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	10:02	End Time (MST):	12:50
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999950	≥0.995
800.2	790.5	1.0122			
399.5	391.3	1.0211	Slope	0.989063	0.90 - 1.10
200.3	193.4	1.0357			
			Intercept	-2.298638	+/-30



SO2 Calibration Plot

Date: December 12, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Jackfish 2/3 Station number: AMS27
 Calibration Date: December 11, 2023 Last Cal Date: November 15, 2023
 Start time (MST): 12:47 End time (MST): 17:22
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.41 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345023
 Removed Cal Gas Conc: 5.41 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3811
 ZAG Make/Model: API 701 Serial Number: 135

Analyzer Information

Analyzer make: API T101 Analyzer serial #: 621
 Converter make: Converter serial #:
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996209	1.007470	Backgd or Offset: 29.9	29.9
Calibration intercept:	-0.098000	-0.217773	Coeff or Slope: 0.928	0.953

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4926	74.1	80.2	79.3	1.008
as found 2nd point	4963	37.0	40.0	39.4	1.011
as found 3rd point	4982	18.5	20.0	19.0	1.042
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	74.1	80.2	80.6	0.995
second point	4963	37.0	40.0	40.3	0.993
third point	4982	18.5	20.0	19.4	1.032
as left zero	5000	0.0	0.0	0.1	----
as left span	4926	74.1	80.2	81.1	0.989
SO2 Scrubber Check	4921	79.1	791.0	0.1	----

Date of last scrubber change:	Ave Corr Factor	1.007
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 79.5 Prev response: 79.77 *% change: -0.3%
 Baseline Corr 2nd AF pt: 39.6 AF Slope: 0.994215 AF Intercept: -0.478088
 Baseline Corr 3rd AF pt: 19.2 AF Correlation: 0.999924

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 11, 2023	Previous Calibration:	November 15, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	12:47	End Time (MST):	17:22
Analyzer make:	API T101	Analyzer serial #:	621

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
80.2	80.6	0.9947			
40.0	40.3	0.9934			
20.0	19.4	1.0317			
			Slope	1.007470	0.90 - 1.10
			Intercept	-0.217773	+/-3



H₂S Calibration Plot

Date: December 11, 2023

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.0	----	----
as found span	4921	79.4	816.8	800.3	16.5	811.4	788.8	22.5	1.0067	1.0146
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	0.3	-0.6	----	----
high point	4921	79.4	816.8	800.3	16.5	813.1	796.3	16.7	1.0046	1.0050
second point	4960	39.7	408.5	400.2	8.3	402.5	394.0	8.4	1.0148	1.0157
third point	4980	19.8	203.7	199.6	4.1	196.2	192.7	3.5	1.0383	1.0358
as left zero	5000	0.0	0.0	0.0	0.0	0.6	1.6	-1.0	----	----
as left span	4921	79.4	816.8	425.4	398.1	810.5	419.2	391.4	1.0078	1.0148
Average Correction Factor									1.0192	1.0188

Corrected As found	NO _x = 811.5 ppb	NO = 789.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.8%
Previous Response	NO _x = 818.0 ppb	NO = 803.7 ppb		*Percent Change	NO = -1.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.5	416.9	398.1	392.9	1.0133	98.7%
2nd GPT point (200 ppb O3)	798.5	627.1	187.9	186.8	1.0060	99.4%
3rd GPT point (100 ppb O3)	798.5	714.4	100.6	96.7	1.0405	96.1%
Average Correction Factor					1.0199	98.1%

Notes: Changed the sample inlet filter after as founds. No adjustments required.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 9, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	12:48	End Time (MST):	17:59
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
816.8	813.1	1.0046		
408.5	402.5	1.0148		
203.7	196.2	1.0383		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 9, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	12:48	End Time (MST):	17:59
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	796.3	1.0050		
400.2	394.0	1.0157		
199.6	192.7	1.0358		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 9, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	12:48	End Time (MST):	17:59
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
398.1	392.9	1.0133		
187.9	186.8	1.0060		
100.6	96.7	1.0405		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 12, 2023

Location: Jackfish 2/3





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS29
SURMONT 2**

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 16, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:27	End Time (MST):	14:04
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999968	≥0.995
800.1	805.0	0.9939			
400.6	399.3	1.0032	Slope	1.006697	0.90 - 1.10
199.8	198.2	1.0082			
			Intercept	-1.824842	+/-30



SO2 Calibration Plot

Date: December 1, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Surmont 2 Station number: AMS29
 Calibration Date: December 4, 2023 Last Cal Date: November 1, 2023
 Start time (MST): 10:43 End time (MST): 15:25
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.391 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC508338
 Removed Cal Gas Conc: 5.391 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: CC508338 Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 5472
 ZAG Make/Model: Teledyne API T701 Serial Number: 4297

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 1200326170
 Converter make: Global Converter serial #: 2022-220
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001467	1.003331	Backgd or Offset:	0.92	0.92
Calibration intercept:	-0.142608	-0.082920	Coeff or Slope:	1.080	1.074

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4926	74.2	80.0	80.7	0.990
as found 2nd point	4963	37.2	40.1	40.7	0.983
as found 3rd point	4982	18.6	20.1	20.3	0.983
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4926	74.2	80.0	80.2	0.998
second point	4963	37.2	40.1	40.1	1.000
third point	4982	18.6	20.1	20.1	0.998
as left zero	5000	0.0	0.0	0.0	----
as left span	4926	74.2	80.0	79.9	1.001
SO2 Scrubber Check	4919	81.3	813.0	0.0	----

Date of last scrubber change:	Ave Corr Factor	0.998
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 80.8 Prev response: 79.98 *% change: 1.0%
 Baseline Corr 2nd AF pt: 40.8 AF Slope: 1.009767 AF Intercept: 0.016561
 Baseline Corr 3rd AF pt: 20.4 AF Correlation: 0.999984

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span. Scrubber check done after cal zero, passed.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 4, 2023	Previous Calibration:	November 1, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:43	End Time (MST):	15:25
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥0.995
80.0	80.2	0.9975		
40.1	40.1	1.0002	Slope	0.90 - 1.10
20.1	20.1	0.9977		
			Intercept	+/-3



H₂S Calibration Plot

Date: December 4, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	December 1, 2023	Last Cal Date:	November 16, 2023
Start time (MST):	10:27	End time (MST):	14:04
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC356008	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
C3H8 Cal Gas Conc.	<u>205.7</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
Removed C3H8 Conc.	<u>205.7</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5472
ZAG Make/Model:	Teledyne API T701	Serial Number:	4297

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1170050149
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998269	1.007452	Background:	3.53	3.45
Calibration intercept:	-0.044666	-0.027439	Coefficient:	3.986	4.017

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.09	----
as found span	4918	81.3	17.31	17.23	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.03	----
high point	4918	81.3	17.31	17.45	0.992
second point	4959	40.6	8.65	8.64	1.001
third point	4979	20.3	4.32	4.28	1.011
as left zero	5000	0.0	0.00	0.02	----
as left span	4918	81.3	17.31	17.44	0.993
Average Correction Factor					1.001
Baseline Corr As found:	17.32	Previous response	17.24	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 16, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:27	End Time (MST):	14:04
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.03	----	Correlation Coefficient	0.999944	≥0.995
17.31	17.45	0.9922			
8.65	8.64	1.0007	Slope	1.007452	0.90 - 1.10
4.32	4.28	1.0110			
			Intercept	-0.027439	+/-1.5



THC Calibration Plot

Date: December 1, 2023

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
as found span	4916	84.2	799.2	799.2	0.0	795.3	793.9	1.4	1.0049	1.0067
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
high point	4916	84.2	799.2	799.2	0.0	797.9	798.4	-0.5	1.0016	1.0010
second point	4958	42.1	399.6	399.6	0.0	401.1	398.7	2.5	0.9963	1.0023
third point	4979	21.1	200.3	200.3	0.0	199.1	197.3	1.8	1.0059	1.0151
as left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
as left span	4916	84.2	799.2	415.8	383.4	797.8	411.0	386.8	1.0017	1.0117
Average Correction Factor									1.0013	1.0061

Corrected As found	NO _x = 795.5 ppb	NO = 794.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.4%
Previous Response	NO _x = 798.5 ppb	NO = 800.6 ppb		*Percent Change	NO = -0.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.5	411.1	383.4	384.2	0.9979	100.2%
2nd GPT point (200 ppb O3)	794.5	607.9	186.6	187.1	0.9973	100.3%
3rd GPT point (100 ppb O3)	794.5	699.9	94.6	95.8	0.9875	101.3%
Average Correction Factor					0.9942	100.6%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 14, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:55	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
799.2	797.9	1.0016			
399.6	401.1	0.9963			
200.3	199.1	1.0059			
			Slope	0.999051	0.90 - 1.10
			Intercept	0.087651	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 14, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:55	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
799.2	798.4	1.0010		
399.6	398.7	1.0023		
200.3	197.3	1.0151		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 5, 2023	Previous Calibration:	November 14, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:55	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
383.4	384.2	0.9979			
186.6	187.1	0.9973			
94.6	95.8	0.9875			
			Slope	1.001015	0.90 - 1.10
			Intercept	0.481350	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 5, 2023

Location: Surmont 2





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS30
ELLS RIVER**

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 8, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:14	End Time (MST):	13:18
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999960	≥0.995
800.4	798.5	1.0023			
400.2	395.7	1.0114	Slope	0.999175	0.90 - 1.10
200.1	195.4	1.0241			
			Intercept	-2.535883	+/-30



SO2 Calibration Plot

Date: December 1, 2023

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Ells River Station number: AMS30
 Calibration Date: December 15, 2023 Last Cal Date: November 30, 2023
 Start time (MST): 11:19 End time (MST): 15:14
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.08 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002443
 Removed Cal Gas Conc: 5.08 ppm Rem Gas Exp Date:
 Removed Gas Cyl #: Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3061
 ZAG Make/Model: API T701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1410661331
 Converter make: CDN - 101 Converter serial #: 562
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001780	1.000065	Backgd or Offset:	1.57	1.57
Calibration intercept:	-0.139199	0.020811	Coeff or Slope:	1.100	1.100

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	78.7	80.0	79.8	1.002
as found 2nd point	4961	39.4	40.0	39.9	1.003
as found 3rd point	4980	19.7	20.0	19.9	1.006
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	78.7	80.0	80.0	1.000
second point	4961	39.4	40.0	40.1	0.998
third point	4980	19.7	20.0	19.9	1.006
as left zero	5000	0.0	0.0	0.2	----
as left span	4921	78.7	80.0	80.1	0.998
SO2 Scrubber Check	4921	79.2	800.4	0.1	----
Date of last scrubber change:	N/A		Ave Corr Factor		1.001
Date of last converter efficiency test:	N/A		efficiency		

Baseline Corr As found: 79.8 Prev response: 79.97 *% change: -0.2%
 Baseline Corr 2nd AF pt: 39.9 AF Slope: 0.998207 AF Intercept: -0.039148
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999999

* = > +/-5% change initiates investigation

Notes: Change inlet filters after multipoint as founds. SOx scrubber check done after calibrator zero, passed. No adjustments made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 15, 2023	Previous Calibration:	November 30, 2023
Station Name:	Ells River	Station Number:	AMS30
Start Time (MST):	11:19	End Time (MST):	15:14
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999992	
80.0	80.0	0.9995			≥0.995
40.0	40.1	0.9982	Slope	1.000065	
20.0	19.9	1.0058			0.90 - 1.10
			Intercept	0.020811	+/-3



TRS Calibration Plot

Date: December 15, 2023

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	December 1, 2023	Last Cal Date:	November 8, 2023
Start time (MST):	10:14	End time (MST):	13:18
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm		
Diff between cyl (CH ₄):		Diff between cyl (THC):	
Diff between cyl (NMHC):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
ZAG make/model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.46E-04	2.48E-04	NMHC SP Ratio:	5.47E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	166643
				157070

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.54	1.029
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	16.95	1.005
second point	4960	39.6	8.51	8.42	1.011
third point	4980	19.8	4.26	4.17	1.020
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	16.90	1.008

				Average Correction Factor	1.012
Baseline Corr AF:	16.54	Prev response	16.98	*% change	-2.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	8.72	1.045
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.10	1.001
second point	4960	39.6	4.56	4.52	1.009
third point	4980	19.8	2.28	2.24	1.016
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.05	1.007
Average Correction Factor					1.009
Baseline Corr AF:	8.72	Prev response	9.09	*% change	-4.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.82	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.85	1.009
second point	4960	39.6	3.96	3.90	1.014
third point	4980	19.8	1.98	1.93	1.025
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.86	1.008
Average Correction Factor					1.016
Baseline Corr AF:	7.82	Prev response	7.89	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000046	0.996248
THC Cal Offset:	-0.048138	-0.036338
CH ₄ Cal Slope:	1.001246	0.992583
CH ₄ Cal Offset:	-0.032757	-0.016756
NMHC Cal Slope:	0.998929	0.999405
NMHC Cal Offset:	-0.015582	-0.019981

Notes: Inlet filter changed after As Finds, adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 8, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:14	End Time (MST):	13:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999978	≥ 0.995			
17.03	16.95	1.0047						
8.51	8.42	1.0112				Slope	0.996248	0.90 - 1.10
4.26	4.17	1.0199						
			Intercept	-0.036338	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 8, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:14	End Time (MST):	13:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999979	≥0.995
7.91	7.85	1.0085			
3.96	3.90	1.0141			
1.98	1.93	1.0248			
			Slope	0.992583	0.90 - 1.10
			Intercept	-0.016756	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 1, 2023	Previous Calibration:	November 8, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:14	End Time (MST):	13:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999975	≥ 0.995
9.11	9.10	1.0014			
4.56	4.52	1.0092			
2.28	2.24	1.0157			
			Slope	0.999405	0.90 - 1.10
			Intercept	-0.019981	+/-0.5



NMHC Calibration Plot

Date: December 1, 2023

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	December 15, 2023	Last Cal Date:	December 1, 2023
Start time (MST):	9:49	End time (MST):	11:16
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
ZAG make/model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.46E-04	2.48E-04	NMHC SP Ratio:	5.47E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	166643
				157070

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.86	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	16.85	1.010
second point					
third point					
as left zero					
as left span					
Average Correction Factor					1.010
Baseline Corr AF:	16.86	Prev response	16.98	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	8.96	1.017
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	8.94	1.019
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.019
Baseline Corr AF:	8.96	Prev response	9.09	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.90	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.91	1.000
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.000
Baseline Corr AF:	7.90	Prev response	7.89	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000046	0.989645
THC Cal Offset:	-0.048138	0.000000
CH ₄ Cal Slope:	1.001246	0.999630
CH ₄ Cal Offset:	-0.032757	0.000000
NMHC Cal Slope:	0.998929	0.981192
NMHC Cal Offset:	-0.015582	0.000000

Notes:


Cylinder changed after as founds.

Calibration Performed By: Jan Castro

NMHC Calibration Plot

Date: December 15, 2023

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Ells River
Calibration Date: December 18, 2023
Start time (MST): 10:59
Reason: Routine
Station number: AMS 30
Last Cal Date: November 28, 2023
End time (MST): 15:31

Calibration Standards

NO Gas Cylinder #: T2Y1P2R
NOX Cal Gas Conc: 50.83 ppm
Removed Cylinder #: T2Y1P2R
Removed Gas NOX Conc: 50.83 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: December 11, 2023
NO Cal Gas Conc: 49.97 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 49.97 ppm
NO gas Diff:
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 710321429

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.147	1.153	NO bkgnd or offset:	14.0	14.1
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	14.0	14.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	190.3	189.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995096	1.000506
NO _x Cal Offset:	-1.240000	-1.440000
NO Cal Slope:	0.993053	0.998642
NO Cal Offset:	-2.360000	-2.440000
NO ₂ Cal Slope:	1.003095	1.005022
NO ₂ Cal Offset:	0.737995	0.180492



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.7	-0.1	----	----
as found span	4920	80.0	813.3	799.5	13.8	812.9	794.3	18.5	1.0005	1.0066
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.6	-0.1	----	----
high point	4920	80.0	813.3	799.5	13.8	812.6	796.7	15.8	1.0008	1.0035
second point	4960	40.0	406.6	399.8	6.9	405.2	396.5	8.7	1.0036	1.0082
third point	4980	20.0	203.3	199.9	3.4	201.0	194.9	6.1	1.0115	1.0256
as left zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.6	0.0	----	----
as left span	4920	80.0	813.3	425.1	388.2	813.0	423.4	389.5	1.0003	1.0041
Average Correction Factor									1.0053	1.0124

Corrected As found	NO _x = 813.7 ppb	NO = 795.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.7%
Previous Response	NO _x = 808.1 ppb	NO = 791.6 ppb		*Percent Change	NO = 0.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.0	420.6	388.2	390.0	0.9953	100.5%
2nd GPT point (200 ppb O3)	795.0	610.8	198.0	199.7	0.9913	100.9%
3rd GPT point (100 ppb O3)	795.0	698.3	110.5	111.2	0.9933	100.7%
Average Correction Factor					0.9933	100.7%

Notes:

Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 18, 2023	Previous Calibration:	November 28, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:59	End Time (MST):	15:31
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
813.3	812.6	1.0008		
406.6	405.2	1.0036		
203.3	201.0	1.0115		
			0.999995	
			1.000506	
			-1.440000	





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 18, 2023	Previous Calibration:	November 28, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:59	End Time (MST):	15:31
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.5	796.7	1.0035		
399.8	396.5	1.0082		
199.9	194.9	1.0256		





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 18, 2023	Previous Calibration:	November 28, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:59	End Time (MST):	15:31
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
388.2	390.0	0.9953		
198.0	199.7	0.9913		
110.5	111.2	0.9933		



NO_x Calibration Plot

Date: December 18, 2023

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Ells River Station number: AMS 30
 Calibration Date: December 18, 2023 Last Cal Date: November 30, 2023
 Start time (MST): 13:10 End time (MST): 14:17

Analyzer Make: API T640 S/N: 875
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388751
 Temp/RH standard: Alicat FP-25BT S/N: 388751

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	0.4	-0.28	0.4	<input type="checkbox"/>	+/- 2 deg C
P (mmHg)	722.3	723.97	722.3	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.98	4.95	4.98	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: December 18, 2023 Last Cal Date: November 30, 2023
 PM w/o HEPA: 1.1 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: _____ w/ HEPA: _____
 Date Optical Chamber Cleaned: October 27, 2023 <0.2 ug/m3
 Disposable Filter Changed: October 27, 2023

Annual Maintenance

Date Sample Tube Cleaned: October 27, 2023
 Date RH/T Sensor Cleaned: November 30, 2023

No adjustments made.

Notes:

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS501
LEISMER**

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 28, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	13:13	End Time (MST):	15:55
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999985	
800.2	798.7	1.0019			≥0.995
400.2	398.8	1.0034	Slope	0.999030	
200.1	196.9	1.0161			0.90 - 1.10
			Intercept	-1.115967	+/-30



SO2 Calibration Plot

Date: December 12, 2023

Location: Leismer





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Leismer	Station number:	AMS501
Calibration Date:	December 12, 2023	Last Cal Date:	November 29, 2023
Start time (MST):	9:32	End time (MST):	13:19
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.14	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511843			
Removed Cal Gas Conc:	5.14	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2659
ZAG Make/Model:	API 701		Serial Number:	4427

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020
Converter make:	Global G150	Converter serial #:	2022-218
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002431	0.995716	Backgd or Offset:	3.47
Calibration intercept:	-0.138350	-0.078455	Coeff or Slope:	1.105

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4922	77.8	80.0	79.1	1.009
as found 2nd point	4961	38.9	40.0	39.6	1.005
as found 3rd point	4981	19.4	19.9	19.6	1.007
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4922	77.8	80.0	79.5	1.006
second point	4961	38.9	40.0	39.9	1.002
third point	4981	19.4	19.9	19.7	1.012
as left zero	5000	0.0	0.0	0.0	----
as left span	4922	77.8	80.0	79.5	1.006
SO2 Scrubber Check	4921	79.2	792.0	0.1	----
Date of last scrubber change:	24-Feb-23			Ave Corr Factor	1.007
Date of last converter efficiency test:	December 1, 2022			efficiency	

Baseline Corr As found:	79.3	Prev response:	80.04	*% change:	-0.9%
Baseline Corr 2nd AF pt:	39.8	AF Slope:	0.991572	AF Intercept:	-0.158512
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999995		

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Scrubber test done after calibrator zero. No adjustment.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 12, 2023	Previous Calibration:	November 29, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:32	End Time (MST):	13:19
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999990	≥0.995
80.0	79.5	1.0061			
40.0	39.9	1.0023	Slope	0.995716	0.90 - 1.10
19.9	19.7	1.0123			
			Intercept	-0.078455	+/-3



H₂S Calibration Plot

Date: December 12, 2023

Location: Leismer





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Leismer
Calibration Date: December 13, 2023
Start time (MST): 8:21
Reason: Removal

Station number: AMS501
Last Cal Date: November 28, 2023
End time (MST): 13:46

Calibration Standards

NO Gas Cylinder #: T26811M
NOX Cal Gas Conc: 47.46 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.46 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API 701

Cal Gas Expiry Date: October 30, 2024
NO Cal Gas Conc: 47.39 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.39 ppm
NO gas Diff:
Serial Number: 2659
Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1218153356

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.225	1.238	NO bkgnd or offset:	3.6	3.7
NOX coeff or slope:	0.992	0.990	NOX bkgnd or offset:	3.6	3.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	165.0	166.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.007139	1.001988
NO _x Cal Offset:	-3.748007	-4.067947
NO Cal Slope:	1.006069	1.004511
NO Cal Offset:	-4.627968	-4.907940
NO ₂ Cal Slope:	0.997452	0.999993
NO ₂ Cal Offset:	0.289298	0.065389



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.3	0.2	----	----
as found span	4916	84.4	801.1	799.9	1.2	796.5	792.4	4.1	1.0057	1.0094
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.2	----	----
high point	4916	84.4	801.1	799.9	1.2	800.5	800.9	-0.3	1.0007	0.9987
second point	4958	42.2	400.5	400.0	0.6	395.7	394.8	0.8	1.0122	1.0131
third point	4979	21.1	200.3	200.0	0.3	192.1	190.8	1.3	1.0426	1.0481
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4916	84.4	801.1	410.8	390.3	798.7	411.7	387.0	1.0030	0.9978
Average Correction Factor									1.0185	1.0200

Corrected As found	NO _x = 796.6 ppb	NO = 792.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.8%
Previous Response	NO _x = 803.0 ppb	NO = 800.1 ppb		*Percent Change	NO = -0.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.5	409.4	390.3	390.4	0.9997	100.0%
2nd GPT point (200 ppb O3)	798.5	487.8	311.9	312.0	0.9996	100.0%
3rd GPT point (100 ppb O3)	798.5	640.8	158.9	158.7	1.0011	99.9%
Average Correction Factor					1.0002	100.0%

Notes: Changed inlet filter after as founds. Adjusted span. 3rd point was failing, back to calibrator zero to check any drift. Used 2nd NO reference point due to drift.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 28, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	8:21	End Time (MST):	13:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
801.1	800.5	1.0007			
400.5	395.7	1.0122			
200.3	192.1	1.0426			
			Slope	1.001988	0.90 - 1.10
			Intercept	-4.067947	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 28, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	8:21	End Time (MST):	13:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
799.9	800.9	0.9987			
400.0	394.8	1.0131			
200.0	190.8	1.0481			
			Slope	1.004511	0.90 - 1.10
			Intercept	-4.907940	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 28, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	8:21	End Time (MST):	13:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
390.3	390.4	0.9997			
311.9	312.0	0.9996			
158.9	158.7	1.0011			
			Slope	0.999993	0.90 - 1.10
			Intercept	0.065389	+/-20



NO_x Calibration Plot

Date: December 13, 2023

Location: Leismer





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS505
SAWBONES BAY

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 8, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	13:35	End Time (MST):	16:19
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321323

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999991	≥0.995
799.8	799.5	1.0004			
399.9	400.3	0.9990	Slope	1.000723	0.90 - 1.10
200.4	198.4	1.0103			
			Intercept	-0.792239	+/-30



SO2 Calibration Plot

Date: December 6, 2023

Location: Sawbones Bay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Sawbones Bay Station number: AMS505
 Calibration Date: December 6, 2023 Last Cal Date: November 9, 2023
 Start time (MST): 9:42 End time (MST): 13:42
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517397
 Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 5112
 ZAG Make/Model: Teledyne API T701 Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43iQ Analyzer serial #: 1228021057
 Converter make: Global 150 Converter serial #: 2022-224
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994794	0.987083	Backgd or Offset: 2.21	2.21
Calibration intercept:	0.261745	0.281580	Coeff or Slope: 1.021	1.021

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4922	77.7	80.0	79.6	1.009
as found 2nd point	4961	38.8	40.0	39.8	1.009
as found 3rd point	4981	19.4	20.0	20.1	1.004
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4922	77.7	80.0	79.3	1.009
second point	4961	38.8	40.0	39.7	1.007
third point	4981	19.4	20.0	20.0	0.999
as left zero	5000	0.0	0.0	0.4	----
as left span	4922	77.7	80.0	79.0	1.013
SO2 Scrubber Check	4922	77.8	778.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.005
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 79.4 Prev response: 79.88 *% change: -0.7%
 Baseline Corr 2nd AF pt: 39.6 AF Slope: 0.991138 AF Intercept: 0.229684
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999998

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after calibrator zero. Scrubber check and passed after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 9, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	9:42	End Time (MST):	13:42
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1228021057

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient	1.000000
80.0	79.3	1.0093		
40.0	39.7	1.0067	Slope	0.987083
20.0	20.0	0.9990		
			Intercept	0.281580
				+/-3



H₂S Calibration Plot

Date: December 6, 2023

Location: Sawbones Bay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.3	0.2	----	----
as found span	4917	83.4	799.6	799.6	0.0	803.0	801.6	1.4	0.9957	0.9975
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	0.0	-0.3	----	----
high point	4917	83.4	799.6	799.6	0.0	801.1	801.8	-0.6	0.9981	0.9972
second point	4958	41.7	399.8	399.8	0.0	400.0	399.9	0.1	0.9996	0.9999
third point	4979	20.9	200.4	200.4	0.0	198.3	197.5	0.7	1.0106	1.0146
as left zero	5000	0.0	0.0	0.0	0.0	-0.4	0.0	-0.4	----	----
as left span	4916	83.4	799.7	332.5	467.2	797.3	333.3	463.9	1.0031	0.9977
Average Correction Factor									1.0028	1.0039

Corrected As found	NO _x = 803.1 ppb	NO = 801.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.3%
Previous Response	NO _x = 800.7 ppb	NO = 800.9 ppb		*Percent Change	NO = 0.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.9	331.7	467.2	465.8	1.0030	99.7%
2nd GPT point (200 ppb O3)	798.9	547.2	251.7	250.7	1.0040	99.6%
3rd GPT point (100 ppb O3)	798.9	650.1	148.8	148.0	1.0054	99.5%
Average Correction Factor					1.0041	99.6%

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 8, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:26	End Time (MST):	12:42
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
799.6	801.1	0.9981			
399.8	400.0	0.9996			
200.4	198.3	1.0106			
			Slope	1.003180	0.90 - 1.10
			Intercept	-1.290916	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 8, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:26	End Time (MST):	12:42
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.6	801.8	0.9972		
399.8	399.9	0.9999		
200.4	197.5	1.0146		
			0.999979	
			1.004108	
			-1.590759	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 8, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:26	End Time (MST):	12:42
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
467.2	465.8	1.0030		
251.7	250.7	1.0040		
148.8	148.0	1.0054		
			1.000000	
			0.997701	
			-0.376391	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 7, 2023

Location: Sawbones Bay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS507
KIRBY SOUTH

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 6, 2023
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	13:10	End Time (MST):	16:13
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999997	
799.6	799.0	1.0008			≥0.995
400.3	398.9	1.0036	Slope	0.998878	
199.7	199.7	0.9998			0.90 - 1.10
			Intercept	-0.089018	+/-30



SO2 Calibration Plot

Date: December 7, 2023

Location: Kirby South





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 24, 2023	Previous Calibration:	December 7, 2023
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	14:15	End Time (MST):	15:30
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.7	----	Correlation Coefficient	0.999989	
799.6	813.2	0.9833			≥0.995
400.3	405.1	0.9883	Slope	1.016010	
199.7	203.0	0.9835			0.90 - 1.10
			Intercept	-0.008833	+/-30



SO2 Calibration Plot

Date: December 24, 2023

Location: Kirby South





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Kirby South Station number: AMS507
 Calibration Date: December 7, 2023 Last Cal Date: November 8, 2023
 Start time (MST): 8:35 End time (MST): 13:17
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.167 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517378
 Removed Cal Gas Conc: 5.167 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3804
 ZAG Make/Model: API T701H Serial Number: 880

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1150840012
 Converter make: Global Converter serial #: 2022-197
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004602	1.002178	Backgd or Offset:	1.98 1.81
Calibration intercept:	0.058972	-0.081174	Coeff or Slope:	1.058 1.079

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4923	77.4	80.0	79.0	1.009
as found 2nd point	4961	38.8	40.1	39.3	1.013
as found 3rd point	4981	19.3	19.9	19.3	1.018
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.4	80.0	80.1	0.998
second point	4961	38.8	40.1	40.1	1.000
third point	4981	19.3	19.9	19.8	1.007
as left zero	5000	0.0	0.0	0.2	----
as left span	4923	77.4	80.0	79.7	1.003
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:	25-Jul-23			Ave Corr Factor	1.002
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.3 Prev response: 80.41 *% change: -1.4%
 Baseline Corr 2nd AF pt: 39.6 AF Slope: 0.992034 AF Intercept: -0.401072
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999992

** = > +/-5% change initiates investigation*

Notes: Changed sample inlet filter after as founds. Scrubber check done after MPAF's, passed. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 8, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	8:35	End Time (MST):	13:17
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999995	
80.0	80.1	0.9985			≥0.995
40.1	40.1	0.9999	Slope	1.002178	
19.9	19.8	1.0072			0.90 - 1.10
			Intercept	-0.081174	+/-3



H₂S Calibration Plot

Date: December 7, 2023

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 7, 2023	Previous Calibration:	November 6, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	13:10	End Time (MST):	16:13
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.00	----	Correlation Coefficient	0.999974	
17.26	17.26	1.0002			≥0.995
8.64	8.59	1.0062	Slope	1.001294	
4.31	4.23	1.0192			0.90 - 1.10
			Intercept	-0.045008	+/-1.5



THC Calibration Plot

Date: December 7, 2023

Location: Kirby South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	December 6, 2023	Last Cal Date:	November 8, 2023
Start time (MST):	12:10	End time (MST):	16:56
Reason:	Routine		

Calibration Standards

NO Gas Cylinder #:	T34ULGL	Cal Gas Expiry Date:	March 8, 2025
NOX Cal Gas Conc:	49.39 ppm	NO Cal Gas Conc:	49.02 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	49.39 ppm	Removed Gas NO Conc:	49.02 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	3804
ZAG make/model:	API 701H	Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>	
NO coeff or slope:	1.424	1.397	NO bkgnd or offset:	1.7	1.7
NOX coeff or slope:	0.997	0.995	NOX bkgnd or offset:	1.8	1.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	202.82	202.87

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.007732	1.002591
NO _x Cal Offset:	-4.271621	-4.472891
NO Cal Slope:	1.008462	1.001411
NO Cal Offset:	-5.833654	-4.934867
NO ₂ Cal Slope:	1.005705	0.999171
NO ₂ Cal Offset:	0.483774	0.864517



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.4	-0.2	----	----
as found span	4919	81.0	800.1	794.1	6.0	822.0	811.0	11.0	0.9734	0.9792
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
high point	4919	81.0	800.1	794.1	6.0	799.9	792.7	7.2	1.0003	1.0018
second point	4960	40.5	400.0	397.0	3.0	394.3	390.1	4.2	1.0145	1.0177
third point	4980	20.2	199.5	198.0	1.5	191.4	188.9	2.5	1.0425	1.0483
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
as left span	4919	81.0	800.1	410.2	389.9	797.4	407.7	389.6	1.0034	1.0062
Average Correction Factor									1.0191	1.0226

Corrected As found	NO _x = 822.6 ppb	NO = 811.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 2.5%
Previous Response	NO _x = 802.0 ppb	NO = 795.0 ppb		*Percent Change	NO = 2.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	792.3	408.4	389.9	389.9	1.0000	100.0%
2nd GPT point (200 ppb O3)	792.3	623.2	175.1	176.9	0.9898	101.0%
3rd GPT point (100 ppb O3)	792.3	711.6	86.7	87.7	0.9885	101.2%
Average Correction Factor					0.9928	100.7%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 8, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	12:10	End Time (MST):	16:56
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.1	799.9	1.0003		
400.0	394.3	1.0145		
199.5	191.4	1.0425		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 8, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	12:10	End Time (MST):	16:56
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
794.1	792.7	1.0018		
397.0	390.1	1.0177		
198.0	188.9	1.0483		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 6, 2023	Previous Calibration:	November 8, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	12:10	End Time (MST):	16:56
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
389.9	389.9	1.0000			
175.1	176.9	0.9898			
86.7	87.7	0.9885			
			Slope	0.999171	0.90 - 1.10
			Intercept	0.864517	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 6, 2023

Location: Kirby South





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS511
BLACKGOLD**

DECEMBER 2023

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 14, 2023	Previous Calibration:	November 7, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	12:13	End Time (MST):	15:31
Analyzer make:	Thermo scientific	Analyzer serial #:	1160290014

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999996	≥0.995
799.0	801.2	0.9973			
399.8	402.6	0.9931	Slope	1.002260	0.90 - 1.10
200.0	201.0	0.9948			
			Intercept	0.830491	+/-30



SO2 Calibration Plot

Date: December 14, 2023

Location: Blackgold





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Blackgold Station number: AMS511
 Calibration Date: December 13, 2023 Last Cal Date: November 9, 2023
 Start time (MST): 9:57 End time (MST): 14:02
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.139 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC511397
 Removed Cal Gas Conc: 5.139 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 2445
 ZAG Make/Model: API T701 Serial Number: 138

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1336160090
 Converter make: Global G150 Converter serial #: 2022-227
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006353	0.982906	Backgd or Offset:	3.37
Calibration intercept:	-0.019423	0.001152	Coeff or Slope:	1.159

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4922	77.8	80.0	78.2	1.023
as found 2nd point	4961	38.9	40.0	39.2	1.020
as found 3rd point	4981	19.5	20.0	19.6	1.022
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4922	77.8	80.0	78.6	1.017
second point	4961	38.9	40.0	39.4	1.015
third point	4981	19.5	20.0	19.5	1.028
as left zero	5000	0.0	0.0	0.3	----
as left span	4922	77.9	80.0	77.8	1.029
SO2 Scrubber Check	4920	80.0	800.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	1.020
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 78.2 Prev response: 80.45 *% change: -2.9%
 Baseline Corr 2nd AF pt: 39.2 AF Slope: 0.978048 AF Intercept: 0.021159
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999998

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	December 13, 2023	Previous Calibration:	November 9, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	9:57	End Time (MST):	14:02
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999982	≥0.995
80.0	78.6	1.0174			
40.0	39.4	1.0148	Slope	0.982906	0.90 - 1.10
20.0	19.5	1.0277			
			Intercept	0.001152	+/-3



H₂S Calibration Plot

Date: December 13, 2023

Location: Blackgold





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	December 14, 2023	Previous Calibration:	November 7, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	12:13	End Time (MST):	15:31
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1317958295

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.06	----	Correlation Coefficient	0.999972	≥0.995
17.11	17.07	1.0027			
8.57	8.50	1.0078	Slope	0.994672	0.90 - 1.10
4.28	4.26	1.0065			
			Intercept	0.019633	+/-1.5



THC Calibration Plot

Date: December 14, 2023

Location: Blackgold





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Blackgold Station number: AMS511
Calibration Date: December 22, 2023 Last Cal Date: November 9, 2023
Start time (MST): 10:16 End time (MST): 14:43
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T0F8P52 Cal Gas Expiry Date: August 16, 2026
NOX Cal Gas Conc: 47.43 ppm NO Cal Gas Conc: 47.43 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 47.43 ppm Removed Gas NO Conc: 47.43 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 2445
ZAG make/model: Teledyne API T701 Serial Number: 138

Analyzer Information

Analyzer make: Teledyne API T200 Analyzer serial #: 7029
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.041	1.041	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.038	1.038	NOX bkgnd or offset:	0.4	0.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.2	4.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.004049	1.012843
NO _x Cal Offset:	-1.772798	0.247028
NO Cal Slope:	1.003178	1.011487
NO Cal Offset:	-2.292773	-0.052956
NO ₂ Cal Slope:	1.000729	0.998436
NO ₂ Cal Offset:	0.157575	1.931124



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.8	-0.1	0.9	----	----
as found span	4916	84.4	800.6	800.6	0.0	810.9	806.4	4.4	0.9872	0.9928
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.5	0.0	0.5	----	----
high point	4916	84.4	800.6	800.6	0.0	811.1	809.7	1.4	0.9870	0.9887
second point	4958	42.2	400.3	400.3	0.0	405.9	404.9	1.0	0.9862	0.9887
third point	4979	21.1	200.2	200.2	0.0	202.5	202.3	0.2	0.9884	0.9894
as left zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
as left span	4916	84.4	800.6	391.8	408.8	808.3	407.3	401.1	0.9904	0.9618
Average Correction Factor									0.9872	0.9889

Corrected As found	NO _x = 810.1 ppb	NO = 806.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.0%
Previous Response	NO _x = 802.0 ppb	NO = 800.8 ppb		*Percent Change	NO = 0.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	809.0	400.2	408.8	409.5	0.9983	100.2%
2nd GPT point (200 ppb O3)	809.0	642.8	166.2	168.8	0.9846	101.6%
3rd GPT point (100 ppb O3)	809.0	729.0	80.0	82.9	0.9650	103.6%
Average Correction Factor					0.9826	101.8%

Notes: Changed the sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	December 22, 2023	Previous Calibration:	November 9, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	10:16	End Time (MST):	14:43
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.5	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.6	811.1	0.9870		
400.3	405.9	0.9862		
200.2	202.5	0.9884		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 22, 2023	Previous Calibration:	November 9, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	10:16	End Time (MST):	14:43
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
800.6	809.7	0.9887		
400.3	404.9	0.9887		
200.2	202.3	0.9894		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	December 22, 2023	Previous Calibration:	November 9, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	10:16	End Time (MST):	14:43
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.5	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
408.8	409.5	0.9983		
166.2	168.8	0.9846		
80.0	82.9	0.9650		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: December 22, 2023

Location: Blackgold





End of Report