



**WOOD BUFFALO  
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive  
Fort McMurray, AB T9K 0K4  
P: 780.799.4420 E: info@wbea.org  
[wbea.org](http://wbea.org)

Wood Buffalo Environmental Association

# JANUARY 2024 MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

February 29, 2024

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS01**  
**BERTHA GANTER - FORT MCKAY**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	January 11, 2024	Last Cal Date:	December 13, 2023
Start time (MST):	11:21	End time (MST):	14:54
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC418809			
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998738	0.999624	Backgd or Offset:	20.0	19.6
Calibration intercept:	-0.153291	-0.113430	Coeff or Slope:	0.887	0.891

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4918	81.3	800.3	794.3	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4918	81.3	800.3	800.2	1.000
second point	4959	40.7	400.6	399.8	1.002
third point	4979	20.3	199.8	199.3	1.003
as left zero	5000	0.0	0.0	0.4	----
as left span	4918	81.3	800.3	801.2	0.999
Average Correction Factor					1.002

Baseline Corr As found:	794.20	Previous response	799.10	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 13, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:21	End Time (MST):	14:54
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999998	
800.3	800.2	1.0001			≥0.995
400.6	399.8	1.0020	Slope	0.999624	
199.8	199.3	1.0026			0.90 - 1.10
			Intercept	-0.113430	+/-30



SO2 Calibration Plot

Date: December 13, 2023

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01  
 Calibration Date: January 25, 2024 Last Cal Date: December 12, 2023  
 Start time (MST): 11:01 End time (MST): 16:52  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024  
 Cal Gas Cylinder #: CC511749  
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A  
 Removed Gas Cyl #: N/A Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565  
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

### Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153461  
 Converter make: CD Nova Converter serial #: 470  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999507	0.994364	Backgd or Offset:	2.25 2.32
Calibration intercept:	0.219996	0.300000	Coeff or Slope:	0.904 0.937

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4921	78.4	80.0	77.6	1.032
as found 2nd point	4960	39.2	40.0	38.8	1.034
as found 3rd point	4980	19.6	20.0	19.4	1.036
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	78.4	80.0	79.8	1.002
second point	4960	39.2	40.0	40.1	0.997
third point	4980	19.6	20.0	20.3	0.985
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	78.4	80.0	79.1	1.011
SO2 Scrubber Check	4919	81.3	813.0	0.1	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	0.995
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 77.5 Prev response: 80.17 \*% change: -3.5%  
 Baseline Corr 2nd AF pt: 38.7 AF Slope: 0.969076 AF Intercept: 0.060000  
 Baseline Corr 3rd AF pt: 19.3 AF Correlation: 0.999999

\* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	December 12, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:01	End Time (MST):	16:52
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999993	≥0.995
80.0	79.8	1.0024			
40.0	40.1	0.9975	Slope	0.994364	0.90 - 1.10
20.0	20.3	0.9851			
			Intercept	0.300000	+/-3



TRS Calibration Plot

Date: January 25, 2024

Location: Bertha Ganter-Fort McKay







# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	January 25, 2024	Last Cal Date:	December 12, 2023
Start time (MST):	11:01	End time (MST):	16:52
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.10	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	<u>CC511749</u>			
Removed Cal Gas Conc:	5.10	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	<u>N/A</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167
Converter make:	Global	Converter serial #:	2022-221
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.985566	0.998712	Backgd or Offset:	1.72
Calibration intercept:	0.536834	0.176812	Coeff or Slope:	1.009
				2.45
				0.992

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4922	78.4	80.0	78.2	1.026
as found 2nd point	4960	39.2	40.0	39.7	1.015
as found 3rd point	4980	19.6	20.0	20.1	1.010
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4922	78.4	80.0	80.1	0.998
second point	4960	39.2	40.0	40.1	0.997
third point	4980	19.6	20.0	20.0	1.000
as left zero	5000	0.0	0.0	0.0	----
as left span	4922	78.4	80.0	79.9	1.001
SO2 Scrubber Check	4919	81.3	813.0	0.1	----
Date of last scrubber change:	January 25, 2024			Ave Corr Factor	0.999
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	77.9	Prev response:	79.34	*% change:	-1.9%
Baseline Corr 2nd AF pt:	39.4	AF Slope:	0.973468	AF Intercept:	0.516858
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999955		

\* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Replaced the converter. Changed out the scrubber beads. Removed the hydrator. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	December 12, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:01	End Time (MST):	16:52
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999988	
80.0	80.1	0.9985			≥0.995
40.0	40.1	0.9975	Slope	0.998712	
20.0	20.0	0.9999			0.90 - 1.10
			Intercept	0.176812	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 25, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	January 11, 2024	Last Cal Date:	December 13, 2023
Start time (MST):	11:21	End time (MST):	14:54
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH <sub>4</sub> Cal Gas Conc.	497.2 ppm	CH <sub>4</sub> Equiv Conc.	1061.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	497.2 ppm	CH <sub>4</sub> Equiv Conc.	1061.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.04E-04	4.37E-04	NMHC SP Ratio:	7.01E-05	7.11E-05
CH <sub>4</sub> Retention time:	14.6	16.7	NMHC Peak Area:	131151	129269
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.27	17.46	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	17.27	17.21	1.003
second point	4959	40.7	8.64	8.53	1.013
third point	4980	20.3	4.31	4.27	1.011
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.27	17.43	0.990
Average Correction Factor					1.009

Baseline Corr AF:	17.46	Prev response	17.16	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4918	81.3	9.18	9.38	0.978
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	9.18	9.15	1.003
second point	4959	40.7	4.60	4.57	1.007
third point	4980	20.3	2.29	2.29	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	9.18	9.29	0.989
Average Correction Factor					1.004
Baseline Corr AF:	9.38	Prev response	9.12	*% change	2.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	8.07	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	8.09	8.07	1.002
second point	4959	40.7	4.05	3.97	1.020
third point	4980	20.3	2.02	1.98	1.021
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.15	0.993
Average Correction Factor					1.014
Baseline Corr AF:	8.07	Prev response	8.04	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997839	0.996788
THC Cal Offset:	-0.066185	-0.028998
CH <sub>4</sub> Cal Slope:	1.002253	0.998653
CH <sub>4</sub> Cal Offset:	-0.062393	-0.029405
NMHC Cal Slope:	0.994150	0.996266
NMHC Cal Offset:	-0.004590	-0.001592

Notes: Changed out the inlet filter after as founds. Lowered the carrier run pressure and adjusted the window timings. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 13, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:21	End Time (MST):	14:54
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999972	$\geq 0.995$
17.27	17.21	1.0033			
8.64	8.53	1.0129			
4.31	4.27	1.0107			
			Slope	0.996788	0.90 - 1.10
			Intercept	-0.028998	+/-0.5





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 13, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:21	End Time (MST):	14:54
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999900	≥0.995
8.09	8.07	1.0019			
4.05	3.97	1.0200			
2.02	1.98	1.0211			
			Slope	0.998653	0.90 - 1.10
			Intercept	-0.029405	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 13, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:21	End Time (MST):	14:54
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	$\geq 0.995$
9.18	9.15	1.0034			
4.60	4.57	1.0068			
2.29	2.29	1.0018			
			Slope	0.996266	0.90 - 1.10
			Intercept	-0.001592	+/-0.5





NMHC Calibration Plot

Date: January 11, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	January 22, 2024	Last Cal Date:	January 11, 2024
Start time (MST):	11:08	End time (MST):	15:30
Reason:	Cylinder Change		

### Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH <sub>4</sub> Cal Gas Conc.	497.2 ppm	CH <sub>4</sub> Equiv Conc.	1061.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	497.2 ppm	CH <sub>4</sub> Equiv Conc.	1061.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	4.37E-04	4.37E-04	NMHC SP Ratio:	7.11E-05	7.11E-05
CH <sub>4</sub> Retention time:	16.7	16.7	NMHC Peak Area:	129269	129269
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.27	17.61	0.980
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.27	19.66	0.878
Average Correction Factor					

Baseline Corr AF:	17.61	Prev response	17.18	*% change	2.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4918	81.3	9.18	9.33	0.985
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	9.18	10.61	<b>0.865</b>
Average Correction Factor					
Baseline Corr AF:	9.33	Prev response	9.15	*% change	1.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<b>* = &gt; +/-5% change initiates investigation</b>	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	8.29	0.976
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	9.05	<b>0.893</b>
Average Correction Factor					
Baseline Corr AF:	8.29	Prev response	8.05	*% change	2.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<b>* = &gt; +/-5% change initiates investigation</b>	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.996788	
THC Cal Offset:	-0.028998	
CH <sub>4</sub> Cal Slope:	0.998653	
CH <sub>4</sub> Cal Offset:	-0.029405	
NMHC Cal Slope:	0.996266	
NMHC Cal Offset:	-0.001592	


Notes: Changed out the H2 and N2 cylinders.

Calibration Performed By: Rene Chamberland

NMHC Calibration Plot

Date: January 22, 2024

Location: Bertha Ganter-Fort McKay



— THC\_Avg — NMHC\_Avg — CH4\_Avg



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Bertha Ganter-Fort McKay      Station number: AMS01  
Calibration Date: January 4, 2024      Last Cal Date: December 4, 2023  
Start time (MST): 11:08      End time (MST): 16:01  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T2Y1P9L      Cal Gas Expiry Date: December 11, 2023  
NOX Cal Gas Conc: 50.84 ppm      NO Cal Gas Conc: 50.04 ppm  
Removed Cylinder #: NA      Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 50.84 ppm      Removed Gas NO Conc: 50.04 ppm  
NOX gas Diff:      NO gas Diff:  
Calibrator Model: Teledyne API T700      Serial Number: 3565  
ZAG make/model: Teledyne API T701      Serial Number: 4890

### Analyzer Information

Analyzer make: Thermo 42i      Analyzer serial #: 1218153357  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.431	1.442	NO bkgnd or offset:	7.3	7.1
NOX coeff or slope:	0.989	0.992	NOX bkgnd or offset:	8.0	7.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	199.2	199.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000323	1.000562
NO <sub>x</sub> Cal Offset:	-0.720000	0.220000
NO Cal Slope:	1.001285	0.999101
NO Cal Offset:	-1.280000	-0.740000
NO <sub>2</sub> Cal Slope:	1.001111	1.003895
NO <sub>2</sub> Cal Offset:	-0.531962	0.627156



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-1.0	-0.3	-0.7	----	----
as found span	4920	80.0	813.4	800.6	12.8	808.3	792.1	16.1	1.0064	1.0108
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.1	----	----
high point	4920	80.0	813.4	800.6	12.8	813.7	799.3	14.5	0.9997	1.0017
second point	4960	40.0	406.7	400.3	6.4	408.2	399.6	8.5	0.9964	1.0018
third point	4980	20.0	203.4	200.2	3.2	203.3	198.0	5.4	1.0003	1.0109
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
as left span	4920	80.0	813.4	413.0	400.4	809.8	414.2	395.6	1.0045	0.9972
Average Correction Factor									0.9988	1.0048

Corrected As found	NO <sub>x</sub> = 809.3 ppb	NO = 792.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.5%	
Previous Response	NO <sub>x</sub> = 813.0 ppb	NO = 800.4 ppb		*Percent Change	NO = -1.0%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.1	407.5	400.4	402.4	0.9950	100.5%
2nd GPT point (200 ppb O3)	795.1	601.0	206.9	208.3	0.9933	100.7%
3rd GPT point (100 ppb O3)	795.1	698.6	109.3	111.1	0.9838	101.6%
Average Correction Factor					0.9907	100.9%

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 4, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:08	End Time (MST):	16:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
813.4	813.7	0.9997			
406.7	408.2	0.9964			
203.4	203.3	1.0003			
			Slope	1.000562	0.90 - 1.10
			Intercept	0.220000	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 4, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:08	End Time (MST):	16:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	$\geq 0.995$	
800.6	799.3	1.0017			
400.3	399.6	1.0018			
200.2	198.0	1.0109			
			Slope	0.999101	0.90 - 1.10
			Intercept	-0.740000	+/-20

**NO Calibration Curve**







# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 4, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:08	End Time (MST):	16:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
400.4	402.4	0.9950			
206.9	208.3	0.9933			
109.3	111.1	0.9838			
			Slope	1.003895	0.90 - 1.10
			Intercept	0.627156	+/-20


**NO<sub>2</sub> Calibration Curve**



**NO<sub>x</sub> Calibration Plot**

Date: January 4, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Bertha Ganter-Fort McKay      Station number: AMS01  
 Calibration Date: January 3, 2024      Last Cal Date: December 1, 2023  
 Start time (MST): 10:48      End time (MST): 13:46  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 3565  
 ZAG Make/Model: Teledyne API T701      Serial Number: 4890

### Analyzer Information

Analyzer make: Teledyne API T400      Analyzer serial #: 1107  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004914	0.999800	Backgd or Offset:	3.8	4.8
Calibration intercept:	0.040000	0.360000	Coeff or Slope:	1.012	1.014

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	5000	863.1	400.0	400.3	0.999
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.2	----
high point	5000	863.1	400.0	400.0	1.000
second point	5000	742.5	200.0	201.0	0.995
third point	5000	651.7	100.0	100.1	0.999
as left zero	5000	0.0	0.0	0.0	----
as left span	5000	863.1	400.0	403.6	0.991
Average Correction Factor					0.998

Baseline Corr As found:	399.8	Previous response	402.0	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted zero only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 3, 2024	Previous Calibration:	December 1, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:48	End Time (MST):	13:46
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999993	≥0.995
400.0	400.0	1.0000			
200.0	201.0	0.9950	Slope	0.999800	0.90 - 1.10
100.0	100.1	0.9990			
			Intercept	0.360000	+/- 5



O<sub>3</sub> Calibration Plot

Date: January 3, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Fort McKay - Bertha Ganter      Station number: AMS 01  
 Calibration Date: January 26, 2024      Last Cal Date: December 12, 2023  
 Start time (MST): 11:37      End time (MST): 13:49

Analyzer Make: Teledyne API T640      S/N: 324  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat      S/N: 388752  
 Temp/RH standard: Alicat      S/N: 388752

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-5.7	-6.3	-5.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	736.2	735.4	736.2	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.15	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 26, 2024</u>		Last Cal Date: <u>December 12, 2023</u>		
	PM w/o HEPA: <u>34.2</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning :      Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.5	11	11	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: <u>13.5</u>		w/ HEPA: <u>0</u>		
Date Optical Chamber Cleaned:	<u>January 26, 2024</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>January 26, 2024</u>				

### Annual Maintenance

Date Sample Tube Cleaned: September 14, 2023  
 Date RH/T Sensor Cleaned: January 26, 2024

Notes:      Flow, temperature, and pressure all within limits. Leak check passed. PMT peak test within limits. Optical chamber and RH/T sensor cleaned. Disposable filter changed.

Calibration by:      Rene Chamberland



# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	January 8, 2024	Last Cal Date:	December 11, 2023
Start time (MST):	10:38	End time (MST):	14:00
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	<u>3040</u>	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>3040</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

### Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001201	1.000186	Backgd or Offset:	-0.013	-0.013
Calibration intercept:	0.203827	0.241822	Coeff or Slope:	0.989	0.983

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4933	66.7	40.6	41.0	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4933	66.7	40.6	40.7	0.998
second point	4966	33.3	20.2	20.8	0.974
third point	4983	16.7	10.2	10.5	0.972
as left zero	5000	0.0	0.0	0.1	----
as left span	2960	40.0	40.5	40.2	1.010
Average Correction Factor					0.981

Baseline Corr As found:	40.79	Prev response:	40.81	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 8, 2024	Previous Calibration:	December 11, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:38	End Time (MST):	14:00
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999843	$\geq 0.995$
40.6	40.7	0.9977	Slope	1.000186	0.90 - 1.10
20.2	20.8	0.9745	Intercept	0.241822	+/-1.5
10.2	10.5	0.9717			






CO Calibration Plot

Date: January 8, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Bertha Ganter-Fort McKay      Station number: AMS01  
 Calibration Date: January 5, 2024      Last Cal Date: December 8, 2023  
 Start time (MST): 11:19      End time (MST): 14:30  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 60,200 ppm      Cal Gas Exp Date: December 1, 2028  
 Cal Gas Cylinder #: ALM042207  
 Removed Cal Gas Conc: 60,200 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 3565  
 N2 Gen Make/Model: Peak Sci NG5000      Serial Number: 7220900034

### Analyzer Information

Analyzer make: Teledyne API 360      Analyzer serial #: 442  
 Analyzer Range 0 - 2,000 ppm

	<b><u>Start</u></b>	<b><u>Finish</u></b>		<b><u>Start</u></b>	<b><u>Finish</u></b>
Calibration slope:	1.000878	1.001967	Backgd or Offset:	0.045	0.045
Calibration intercept:	-5.800000	-5.940000	Coeff or Slope:	0.875	0.876

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-0.2	----
as found span	2920	80.0	1605.3	1597.8	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.2	----
high point	2920	80.0	1605.3	1605.4	1.000
second point	2960	40.0	802.7	795.9	1.009
third point	2980	20.0	401.3	389.6	1.030
as left zero	3000	0.0	0.0	-0.1	----
as left span	2960	40.0	802.7	786.7	1.020
<b>Average Correction Factor</b>					<b>1.013</b>

Baseline Corr As found: 1598.00      Prev response: 1600.94      \*% change: -0.2%  
 Baseline Corr 2nd AF pt: NA      AF Slope:      AF Intercept:  
 Baseline Corr 3rd AF pt: NA      AF Correlation:

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date	January 5, 2024	Previous Calibration	December 8, 2023
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS01
Start Time (MST)	11:19	End Time (MST)	14:30
Analyzer make	Teledyne API 360	Analyzer serial #	442

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999933	≥0.995
1605.3	1605.4	1.0000			
802.7	795.9	1.0085	Slope	1.001967	0.90 - 1.10
401.3	389.6	1.0301			
			Intercept	-5.940000	+/-10



CO<sub>2</sub> Calibration Plot

Date: January 5, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## TN - NO<sub>x</sub> - NH<sub>3</sub> Calibration Report

Version-05-2023

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	January 9, 2024	Last Cal Date:	December 6, 2023
Start time (MST):	12:23	End time (MST):	16:55
NH3 Cal Date:	January 10, 2024	Last Cal Date:	December 7, 2023
Start time (MST):	11:10	End time (MST):	14:15
Reason:	Routine		

### Calibration Standards

NOX Cal Gas Conc:	50.84	ppm	NO Gas Cylinder #:	T2Y1P9L
NO Cal Gas Conc:	50.04	ppm	NO Cal Gas Expiry:	December 11, 2023
Removed NOX Conc:	50.84	ppm	Removed Cylinder #:	NA
Removed NO Conc:	50.04	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.58	ppm	NH3 Gas Cylinder #:	CC743587
			NH3 Cal Gas Expiry:	August 22, 2024
Removed NH3 Conc:	76.58	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	Teledyne API T700		Serial Number:	3565
ZAG make/model:	Teledyne API T701		Serial Number:	4890

### Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	475
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	5.50
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	514

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.987	0.964	TN coefficient:	0.989	0.968
NOX coefficient:	0.989	0.967	NO bkgnd:	-0.9	-0.9
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-0.3	-0.3
NH3 coefficient:	0.932	0.932	TN bkgnd:	1.2	1.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.004707	0.998960
NO <sub>x</sub> Cal Offset:	-2.480000	-2.460000
NO Cal Slope:	1.004525	0.998530
NO Cal Offset:	-3.440000	-2.840000
NO <sub>2</sub> Cal Slope:	0.995838	0.997727
NO <sub>2</sub> Cal Offset:	-1.454892	-0.655687
NH3 Cal Slope:	1.001746	1.004832
NH3 Cal Offset:	-2.214411	-8.183521
TN Cal Slope:	1.004330	1.007188
TN Cal Offset:	-1.994595	-8.051061



# Wood Buffalo Environmental Association

## TN - NOX - NH<sub>3</sub> Calibration Report

Version-05-2023

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-1.2	-1.2	0.0	----	----
as found NO	4920	80.0	813.4	813.4	----	812.1	813.0	-0.6	1.002	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.2	----	----
high NO point	4920	80.0	813.4	813.4	----	812.4	811.8	0.9	1.001	----
NO/O3 point	4920	80.0	813.4	813.4	----	808.8	806.7	1.8	1.006	----
as found NH3	3416	84.1	1840.1	----	1840.1	1808.1	----	1803.9	1.018	1.020
new NH3 cyl rp							----			
first NH3	3418	82.2	1798.5	----	1798.5	1808.1	----	1803.9	0.995	0.997
second NH3	3454	45.7	1000.0	----	1000.0	994.7	----	992.0	1.005	1.008
third NH3	3477	22.8	498.9	----	498.9	486.4	----	484.9	1.026	1.029
<b>Average Correction Factor</b>									<b>1.0035</b>	<b>1.0113</b>

Corrected As found    TN = 813.3 ppb    NO<sub>x</sub> = 814.2 ppb    NH3 = 1803.9 ppb

Previous Response    TN = 815 ppb    NO<sub>x</sub> = 814.8 ppb    NH3 = 1841.1 ppb

NH3 Previous Converter Efficiency = 93.2%

NH3 Current Converter Efficiency = 93.2%

\*Percent Change    TN = -0.2%

\*Percent Change    NO<sub>x</sub> = -0.1%

\*Percent Change    NH3 = -2.1%

\* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## NO<sub>x</sub> - NO - NO<sub>2</sub> Calibration Report

Version-05-2023

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as found span	4920	80.0	813.4	800.6	813.4	835.2	819.5	835.4	0.9739	0.9770
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.3	----	----
high point	4920	80.0	813.4	800.6	813.4	811.8	798.5	812.4	1.0020	1.0027
second point	4960	40.0	406.7	400.3	406.7	401.2	394.2	401.3	1.0138	1.0155
third point	4980	20.0	203.4	200.2	203.4	199.3	194.9	200.0	1.0204	1.0270
<b>Average Correction Factor</b>									<b>1.0121</b>	<b>1.0151</b>

Baseline Corr As fnd TN = 835.3 ppb NO<sub>x</sub> = 835.1 ppb NO = 819.5 ppb \*Percent Change TN = 2.4%  
 Previous Response TN = 815 ppb NO<sub>x</sub> = 814.8 ppb NO = 800.8 ppb \*Percent Change NO<sub>x</sub> = 2.4%  
 \*Percent Change NO = 2.3%  
 \* = > +/-5% change initiates investigation

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	0.0	----	----
calibration zero	----	----	0.0	-0.2	----	----
1st GPT point (400 ppb O3)	794.2	387.9	419.1	417.5	1.0038	99.6%
2nd GPT point (200 ppb O3)	794.2	600.9	206.1	205.5	1.0029	99.7%
3rd GPT point (100 ppb O3)	794.2	694.2	112.8	110.9	1.0171	98.3%
<b>Average Correction Factor</b>					<b>1.0080</b>	<b>99.2%</b>

Notes: Changed the inlet filter after as founds. Adjusted the NO<sub>x</sub>/TN span.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association

## TN Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 6, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:10	End Time (MST):	14:15
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999907	≥0.995
1798.5	1808.1	0.9947			
1000.0	994.7	1.0053	Slope	1.007188	0.90 - 1.10
498.9	486.4	1.0256			
			Intercept	-8.051061	+/-20







# Wood Buffalo Environmental Association

## NH<sub>3</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 6, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:10	End Time (MST):	14:15
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1798.5	1803.9	0.9970		
1000.0	992.0	1.0081		
498.9	484.9	1.0288		





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	January 9, 2024	Previous Calibration:	December 6, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	12:23	End Time (MST):	16:55
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
813.4	811.8	1.0020			
406.7	401.2	1.0138			
203.4	199.3	1.0204			
			Slope	0.998960	0.90 - 1.10
			Intercept	-2.460000	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-05-2023


### Station Information

Calibration Date:	January 9, 2024	Previous Calibration:	December 6, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	12:23	End Time (MST):	16:55
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.6	798.5	1.0027			
400.3	394.2	1.0155			
200.2	194.9	1.0270			
			Slope	0.998530	0.90 - 1.10
			Intercept	-2.840000	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	January 9, 2024	Previous Calibration:	December 6, 2023
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	12:23	End Time (MST):	16:55
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
419.1	417.5	1.0038			
206.1	205.5	1.0029			
112.8	110.9	1.0171			
			Slope	0.997727	0.90 - 1.10
			Intercept	-0.655687	+/-20



# NO<sub>x</sub> Calibration Plot

Date: January 9, 2024


Location: Bertha Ganter-Fort McKay



# NH<sub>3</sub> Calibration Plot

Date: January 10, 2024

Location: Bertha Ganter-Fort McKay





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS02 MILDRED LAKE JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	January 1, 2024	Last Cal Date:	December 8, 2023
Start time (MST):	11:18	End time (MST):	14:47
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1185
ZAG Make/Model:	API T701		Serial Number:	4891

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000671	0.994869	Backgd or Offset:	18.6	18.4
Calibration intercept:	-0.784833	0.074375	Coeff or Slope:	0.775	0.795

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.8	----
as found span	4920	80.2	801.6	784.2	1.022
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.2	801.6	797.8	1.005
second point	4960	40.1	400.8	398.2	1.007
third point	4980	20.0	199.9	199.5	1.002
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	80.2	801.6	810.0	0.990
Average Correction Factor					1.005

Baseline Corr As found:	785.00	Previous response	801.40	*% change	-2.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 1, 2024	Previous Calibration:	December 8, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	11:18	End Time (MST):	14:47
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999998	
801.6	797.8	1.0048			≥0.995
400.8	398.2	1.0066	Slope	0.994869	
199.9	199.5	1.0021			0.90 - 1.10
			Intercept	0.074375	+/-30



SO2 Calibration Plot

Date: January 1, 2024

Location: Mildred Lake





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Mildred Lake      Station number: AMS02  
 Calibration Date: January 10, 2024      Last Cal Date: December 18, 2023  
 Start time (MST): 10:25      End time (MST): 15:50  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.29 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC345191  
 Removed Cal Gas Conc: 5.29 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 1185  
 ZAG Make/Model: API T701      Serial Number: 5608

### Analyzer Information

Analyzer make: Thermo 43iQTL      Analyzer serial #: 12113311966  
 Converter make: Global G150      Converter serial #: 2022-198  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008965	1.001107	Backgd or Offset: 1.70	1.68
Calibration intercept:	-0.059195	0.040801	Coeff or Slope: 0.740	0.740

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4924	75.6	80.0	79.3	1.009
as found 2nd point	4962	37.8	40.0	39.8	1.005
as found 3rd point	4981	18.9	20.0	19.9	1.005
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4924	75.6	80.0	80.1	0.999
second point	4962	37.8	40.0	40.2	0.995
third point	4981	18.9	20.0	19.9	1.005
as left zero	5000	0.0	0.0	0.1	----
as left span	4924	75.6	80.0	79.7	1.004
SO2 Scrubber Check	4920	80.2	802.0	0.1	----
Date of last scrubber change:	20-Sep-23			Ave Corr Factor	0.999
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.3      Prev response: 80.65      \*% change: -1.7%  
 Baseline Corr 2nd AF pt: 39.8      AF Slope: 0.991249      AF Intercept: 0.060795  
 Baseline Corr 3rd AF pt: 19.9      AF Correlation: 0.999995

\* = > +/-5% change initiates investigation

Notes: Scrubber check done after cal zero, passed. No adjustments made.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 18, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:25	End Time (MST):	15:50
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
80.0	80.1	0.9986		
40.0	40.2	0.9949	Slope	0.90 - 1.10
20.0	19.9	1.0049		
			Intercept	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 10, 2024

Location: Mildred Lake





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	January 1, 2024	Last Cal Date:	December 23, 2023
Start time (MST):	11:18	End time (MST):	14:47
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH <sub>4</sub> Cal Gas Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1048.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1048.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.63E-04	2.62E-04	NMHC SP Ratio:	6.75E-05	6.18E-05
CH <sub>4</sub> Retention time:	15.5	15.3	NMHC Peak Area:	130400	142482
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	17.66	0.952
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.79	1.002
second point	4960	40.1	8.41	8.35	1.007
third point	4980	20.0	4.19	4.15	1.011
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.64	1.011
Average Correction Factor					1.007

Baseline Corr AF:	17.66	Prev response	16.84	*% change	4.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	9.60	0.916
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.78	1.002
second point	4960	40.1	4.40	4.41	0.998
third point	4980	20.0	2.19	2.19	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.65	1.017
Average Correction Factor					1.000
Baseline Corr AF:	9.60	Prev response	8.83	*% change	8.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	8.06	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.00	1.003
second point	4960	40.1	4.01	3.95	1.016
third point	4980	20.0	2.00	1.96	1.022
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	7.99	1.005
Average Correction Factor					1.014
Baseline Corr AF:	8.06	Prev response	8.00	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000046	0.998824
THC Cal Offset:	0.020295	-0.024719
CH <sub>4</sub> Cal Slope:	0.998148	0.998021
CH <sub>4</sub> Cal Offset:	-0.004054	-0.025858
NMHC Cal Slope:	1.001244	0.998479
NMHC Cal Offset:	0.025148	0.003536

Notes: Changed sample inlet filters after as founds. Adjusted span.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 1, 2024	Previous Calibration:	December 23, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	11:18	End Time (MST):	14:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999989	$\geq 0.995$			
16.82	16.79	1.0017						
8.41	8.35	1.0071				Slope	0.998824	0.90 - 1.10
4.19	4.15	1.0111						
			Intercept	-0.024719	$\pm 0.5$			







# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 1, 2024	Previous Calibration:	December 23, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	11:18	End Time (MST):	14:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999938	$\geq 0.995$
8.02	8.00	1.0029			
4.01	3.95	1.0163			
2.00	1.96	1.0224			
			Slope	0.998021	0.90 - 1.10
			Intercept	-0.025858	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 1, 2024	Previous Calibration:	December 23, 2023
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	11:18	End Time (MST):	14:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	$\geq 0.995$			
8.80	8.78	1.0017						
4.40	4.41	0.9979				Slope	0.998479	0.90 - 1.10
2.19	2.19	1.0016						
			Intercept	0.003536	$\pm 0.5$			



NMHC Calibration Plot

Date: January 1, 2024

Location: Mildred Lake





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	January 25, 2024	Last Cal Date:	January 1, 2024
Start time (MST):	10:26	End time (MST):	13:46
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH <sub>4</sub> Cal Gas Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1048.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1048.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.62E-04	2.62E-04	NMHC SP Ratio:	6.18E-05	6.08E-05
CH <sub>4</sub> Retention time:	15.3	15.5	NMHC Peak Area:	142482	144769
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	16.95	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.81	1.000
second point	4960	40.1	8.41	8.43	0.998
third point	4980	20.0	4.19	4.19	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.76	1.003
Average Correction Factor					0.999

Baseline Corr AF:	16.95	Prev response	16.77	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	9.00	0.977
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.79	1.001
second point	4960	40.1	4.40	4.41	0.997
third point	4980	20.0	2.19	2.20	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.79	1.001
Average Correction Factor					0.998
Baseline Corr AF:	9.00	Prev response	8.79	*% change	2.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	7.95	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.02	1.000
second point	4960	40.1	4.01	4.02	0.998
third point	4980	20.0	2.00	1.99	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	7.97	1.006
Average Correction Factor					1.001
Baseline Corr AF:	7.95	Prev response	7.98	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998824	0.999652
THC Cal Offset:	-0.024719	0.005690
CH <sub>4</sub> Cal Slope:	0.998021	1.000099
CH <sub>4</sub> Cal Offset:	-0.025858	-0.001651
NMHC Cal Slope:	0.998479	0.999219
NMHC Cal Offset:	0.003536	0.006941

Notes: Swapped Hydrogen cylinder. Adjusted span.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	January 1, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:26	End Time (MST):	13:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	$\geq 0.995$			
16.82	16.81	1.0005						
8.41	8.43	0.9975				Slope	0.999652	0.90 - 1.10
4.19	4.19	1.0000						
			Intercept	0.005690	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	January 1, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:26	End Time (MST):	13:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999997	≥0.995
8.02	8.02	1.0004			
4.01	4.02	0.9984			
2.00	1.99	1.0044			
			Slope	1.000099	0.90 - 1.10
			Intercept	-0.001651	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	January 1, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:26	End Time (MST):	13:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999996	$\geq 0.995$
8.80	8.79	1.0006			
4.40	4.41	0.9972			
2.19	2.20	0.9961			
			Slope	0.999219	0.90 - 1.10
			Intercept	0.006941	+/-0.5






NMHC Calibration Plot

Date: January 25, 2024

Location: Mildred Lake





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS04 BUFFALO VIEWPOINT JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	January 17, 2024	Last Cal Date:	December 14, 2023
Start time (MST):	7:34	End time (MST):	10:35
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.87	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC446753			
Removed Cal Gas Conc:	50.87	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701		Serial Number:	362

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998629	0.986256	Backgd or Offset:	24.1      23.8
Calibration intercept:	0.434228	0.211572	Coeff or Slope:	0.873      0.864

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	78.6	799.7	804.5	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.6	----
high point	4921	78.6	799.7	789.2	1.013
second point	4961	39.3	399.8	394.2	1.014
third point	4980	19.6	199.4	196.6	1.014
as left zero	5000	0.0	0.0	0.3	----
as left span	4921	78.6	799.7	786.9	1.016
Average Correction Factor					1.014

Baseline Corr As found:	804.60	Previous response	799.08	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Span adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	December 14, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:34	End Time (MST):	10:35
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999999	≥0.995
799.7	789.2	1.0134			
399.8	394.2	1.0142	Slope	0.986256	0.90 - 1.10
199.4	196.6	1.0144			
			Intercept	0.211572	+/-30



SO2 Calibration Plot

Date: January 17, 2024

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Buffalo Viewpoint Station number: AMS04  
 Calibration Date: January 11, 2024 Last Cal Date: December 8, 2023  
 Start time (MST): 7:34 End time (MST): 12:10  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.42 ppm Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC345266  
 Removed Cal Gas Conc: 5.42 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 3808  
 ZAG Make/Model: API T701H Serial Number: 362

### Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1008841400  
 Converter make: Global Converter serial #: 2022-200  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000346	0.996076	Backgd or Offset:	1.83	1.87
Calibration intercept:	0.042159	0.242122	Coeff or Slope:	1.115	1.143

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4926	74.1	80.3	78.0	1.031
as found 2nd point	4963	37.0	40.1	39.5	1.018
as found 3rd point	4982	18.5	20.1	19.6	1.028
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4926	74.1	80.3	80.2	1.002
second point	4963	37.0	40.1	40.4	0.993
third point	4982	18.5	20.1	20.0	1.003
as left zero	5000	0.0	0.0	0.4	----
as left span	4926	74.1	80.3	80.2	1.002
SO2 Scrubber Check	4920	80.0	800.0	-0.1	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	0.999
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 77.9 Prev response: 80.39 \*% change: -3.2%  
 Baseline Corr 2nd AF pt: 39.4 AF Slope: 0.970320 AF Intercept: 0.221672  
 Baseline Corr 3rd AF pt: 19.5 AF Correlation: 0.999947

\* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 8, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:34	End Time (MST):	12:10
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999973	≥0.995
80.3	80.2	1.0015			
40.1	40.4	0.9928	Slope	0.996076	0.90 - 1.10
20.1	20.0	1.0026			
			Intercept	0.242122	+/-3






# H<sub>2</sub>S Calibration Plot

Date: January 11, 2024

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	January 17, 2024	Last Cal Date:	December 14, 2023
Start time (MST):	7:34	End time (MST):	10:33
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	NA	Cal Gas Expiry Date:	March 10, 1931
CH <sub>4</sub> Cal Gas Conc.	497.2 ppm	CH <sub>4</sub> Equiv Conc.	1058.2 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	497.2 ppm	CH <sub>4</sub> Equiv Conc.	1058.2 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.03E-04	2.10E-04	NMHC SP Ratio:	4.45E-05
CH <sub>4</sub> Retention time:	11.8	11.8	NMHC Peak Area:	198327
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON
				ON

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	16.64	16.11	1.033
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	16.64	16.62	1.001
second point	4961	39.3	8.32	8.20	1.014
third point	4980	19.6	4.15	4.03	1.029
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	16.64	16.60	1.002
Average Correction Factor					1.015

Baseline Corr AF:	16.11	Prev response	16.60	*% change	-3.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	8.82	8.52	1.035
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	8.82	8.80	1.002
second point	4961	39.3	4.41	4.37	1.009
third point	4980	19.6	2.20	2.15	1.023
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	8.82	8.80	1.002
Average Correction Factor					1.011
Baseline Corr AF:	8.52	Prev response	8.80	*% change	-3.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	7.82	7.58	1.031
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	7.82	7.82	1.000
second point	4961	39.3	3.91	3.83	1.020
third point	4980	19.6	1.95	1.88	1.037
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	7.82	7.80	1.002
Average Correction Factor					1.019
Baseline Corr AF:	7.58	Prev response	7.80	*% change	-2.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001050	1.000639
THC Cal Offset:	-0.055556	-0.067561
CH <sub>4</sub> Cal Slope:	1.002202	1.002348
CH <sub>4</sub> Cal Offset:	-0.035913	-0.043914
NMHC Cal Slope:	1.000029	0.999123
NMHC Cal Offset:	-0.019643	-0.023647

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	December 14, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:34	End Time (MST):	10:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999920	$\geq 0.995$
16.64	16.62	1.0010			
8.32	8.20	1.0143			
4.15	4.03	1.0294			
			Slope	1.000639	0.90 - 1.10
			Intercept	-0.067561	+/-0.5





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	December 14, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:34	End Time (MST):	10:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999836	≥0.995
7.82	7.82	0.9996			
3.91	3.83	1.0203			
1.95	1.88	1.0368			
			Slope	1.002348	0.90 - 1.10
			Intercept	-0.043914	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	December 14, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:34	End Time (MST):	10:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999967	$\geq 0.995$			
8.82	8.80	1.0022						
4.41	4.37	1.0090				Slope	0.999123	0.90 - 1.10
2.20	2.15	1.0229						
			Intercept	-0.023647	$\pm 0.5$			



NMHC Calibration Plot

Date: January 17, 2024

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	January 26, 2024	Last Cal Date:	January 17, 2024
Start time (MST):	9:20	End time (MST):	10:46
Reason:	Cylinder Change		

### Calibration Standards

Gas Cert Reference:	NA	Cal Gas Expiry Date:	March 10, 1931
CH <sub>4</sub> Cal Gas Conc.	497.2 ppm	CH <sub>4</sub> Equiv Conc.	1058.2 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	497.2 ppm	CH <sub>4</sub> Equiv Conc.	1058.2 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.10E-04	2.10E-04	NMHC SP Ratio:	4.60E-05
CH <sub>4</sub> Retention time:	11.8	11.8	NMHC Peak Area:	191608
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	16.64	16.39	1.015
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	16.64	16.38	1.016
second point					
third point					
as left zero					
as left span					
Average Correction Factor					1.016

Baseline Corr AF:	16.39	Prev response	16.58	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	8.82	8.70	1.014
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	8.82	8.65	1.020
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.020
Baseline Corr AF:	8.70	Prev response	8.79	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	7.82	7.69	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	7.82	7.72	1.013
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.013
Baseline Corr AF:	7.69	Prev response	7.79	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000639	0.984598
THC Cal Offset:	-0.067561	0.000000
CH <sub>4</sub> Cal Slope:	1.002348	0.987641
CH <sub>4</sub> Cal Offset:	-0.043914	0.000000
NMHC Cal Slope:	0.999123	0.980767
NMHC Cal Offset:	-0.023647	0.000000


Notes: Hydrogen Cylinder Changed.

Calibration Performed By: Melissa Lemay

NMHC Calibration Plot

Date: January 26, 2024

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Buffalo Viewpoint Station number: AMS04  
Calibration Date: January 12, 2024 Last Cal Date: December 11, 2023  
Start time (MST): 6:43 End time (MST): 11:43  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: CC324979 Cal Gas Expiry Date: November 3, 1932  
NO<sub>x</sub> Cal Gas Conc: 48.90 ppm NO Cal Gas Conc: 48.80 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NO<sub>x</sub> Conc: 48.90 ppm Removed Gas NO Conc: 48.80 ppm  
NO<sub>x</sub> gas Diff: NO gas Diff:  
Calibrator Model: API T700 Serial Number: 3808  
ZAG make/model: API T701 Serial Number: 362

### Analyzer Information

Analyzer make: API T200 Analyzer serial #: 721  
NO<sub>x</sub> Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.163	1.182	NO bkgnd or offset:	-0.6	-0.6
NO <sub>x</sub> coeff or slope:	1.148	1.169	NO <sub>x</sub> bkgnd or offset:	-0.3	-0.3
NO <sub>2</sub> coeff or slope:	1.000	1.000	Reaction cell Press:	4.4	4.4

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000893	1.000181
NO <sub>x</sub> Cal Offset:	-0.273094	-0.473674
NO Cal Slope:	1.000600	1.002304
NO Cal Offset:	-1.154426	-1.574438
NO <sub>2</sub> Cal Slope:	0.986885	0.987754
NO <sub>2</sub> Cal Offset:	-0.475922	-0.437615



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	0.4	-0.5	----	----
as found span	4918	81.8	800.0	798.4	1.6	787.8	788.2	-0.4	1.0155	1.0129
as found 2nd										
as found 3rd										
new cyl resp	4918	81.8	800.0	798.4	1.6					
calibrator zero	5000	0.0	0.0	0.0	0.0	0.5	0.7	-0.1	----	----
high point	4918	81.8	800.0	798.4	1.6	800.2	799.7	0.4	0.9998	0.9984
second point	4959	40.9	400.0	399.2	0.8	399.1	397.7	1.4	1.0023	1.0038
third point	4980	20.4	199.5	199.1	0.4	198.1	195.5	2.6	1.0070	1.0184
as left zero	5000	0.0	0.0	0.0	0.0	0.6	0.9	-0.4	----	----
as left span	4918	81.8	800.0	398.5	401.5	793.1	402.8	390.3	1.0087	0.9893
Average Correction Factor									1.0030	1.0068

Corrected As found	NO <sub>x</sub> = 787.8 ppb	NO = 787.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -1.6%
Previous Response	NO <sub>x</sub> = 800.5 ppb	NO = 797.7 ppb		*Percent Change	NO = -1.3%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.8	396.9	401.5	396.3	1.0132	98.7%
2nd GPT point (200 ppb O3)	796.8	595.2	203.2	200.3	1.0147	98.6%
3rd GPT point (100 ppb O3)	796.8	696.2	102.2	100.1	1.0213	97.9%
Average Correction Factor					1.0164	98.4%

Notes:

No Maintenance done. Span adjusted.

Calibration Performed By:

Melissa Lemay



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 12, 2024	Previous Calibration:	December 11, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:43	End Time (MST):	11:43
Analyzer make:	API T200	Analyzer serial #:	721

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.5	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.0	800.2	0.9998		
400.0	399.1	1.0023		
199.5	198.1	1.0070		
			0.999993	
			1.000181	
			-0.473674	

**NO<sub>x</sub> Calibration Curve**





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 12, 2024	Previous Calibration:	December 11, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:43	End Time (MST):	11:43
Analyzer make:	API T200	Analyzer serial #:	721

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.7	----	Correlation Coefficient	≥0.995	
798.4	799.7	0.9984			
399.2	397.7	1.0038			
199.1	195.5	1.0184			
			Slope	1.002304	0.90 - 1.10
			Intercept	-1.574438	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 12, 2024	Previous Calibration:	December 11, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:43	End Time (MST):	11:43
Analyzer make:	API T200	Analyzer serial #:	721

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
401.5	396.3	1.0132		
203.2	200.3	1.0147		
102.2	100.1	1.0213		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: January 12, 2024

Location: Buffalo Viewpoint







# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Buffalo Viewpoint      Station number: AMS04  
 Calibration Date: January 8, 2024      Last Cal Date: December 12, 2023  
 Start time (MST): 10:10      End time (MST): 12:52  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: API T700      Serial Number: 3808  
 ZAG Make/Model: API T701      Serial Number: 362

### Analyzer Information

Analyzer make: API T400      Analyzer serial #: 2961  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001229	1.002571	Backgd or Offset:	-3.5	-3.5
Calibration intercept:	0.760000	0.800000	Coeff or Slope:	1.011	1.011

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	5000	982.6	400.0	401.8	0.996
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.9	----
high point	5000	984.0	400.0	401.8	0.996
second point	5000	812.4	200.0	201.5	0.993
third point	5000	702.1	100.0	100.8	0.992
as left zero	5000	0.0	0.0	1.1	----
as left span	5000	984.8	400.0	401.7	0.996
Average Correction Factor					0.993

Baseline Corr As found:	401.6	Previous response	401.3	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No adjustments and maintenance done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 8, 2024	Previous Calibration:	December 12, 2023
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	10:10	End Time (MST):	12:52
Analyzer make:	API T400	Analyzer serial #:	2961

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.9	----	Correlation Coefficient	≥0.995
400.0	401.8	0.9955		
200.0	201.5	0.9926	Slope	0.90 - 1.10
100.0	100.8	0.9921		
			Intercept	+/- 5



O<sub>3</sub> Calibration Plot

Date: January 8, 2024

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04  
 Calibration Date: January 17, 2024 Last Cal Date: December 21, 2023  
 Start time (MST): 10:35 End time (MST): 11:10

Analyzer Make: API T640 S/N: 321  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Deltacal S/N: 1451  
 Temp/RH standard: Deltacal S/N: 1451

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-19.2	-20.2	-19.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	731.8	733.4	731.8	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.13	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 17, 2024</u>		Last Cal Date: <u>December 21, 2023</u>		
	PM w/o HEPA: <u>2.9</u>		PM w/ HEPA: <u>0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		_____			<0.2 ug/m3
Disposable Filter Changed:		_____			

### Annual Maintenance

Date Sample Tube Cleaned: November 28, 2023  
 Date RH/T Sensor Cleaned: May 23, 2023

No adjustments done. Leak check passed. Head cleaned.

Notes:

Calibration by: Melissa Lemay



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS05 MANNIX JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Mannix	Station number:	AMS05
Calibration Date:	January 19, 2024	Last Cal Date:	December 8, 2023
Start time (MST):	10:15	End time (MST):	13:45
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.02	ppm	Cal Gas Exp Date:	January 12, 2029
Cal Gas Cylinder #:	XC026809B			
Removed Cal Gas Conc:	50.02	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	621
ZAG Make/Model:	API T701H		Serial Number:	832

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007440	1.003770	Backgd or Offset:	9.1	9.1
Calibration intercept:	0.280000	0.040000	Coeff or Slope:	0.944	0.944

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4920	80.0	800.3	804.3	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.7	----
high point	4920	80.0	800.3	803.7	0.996
second point	4960	40.0	400.2	401.4	0.997
third point	4980	20.0	200.1	200.2	0.999
as left zero	5000	0.0	0.0	0.7	----
as left span	4920	80.0	800.3	805.3	0.994
Average Correction Factor					0.997

Baseline Corr As found:	803.90	Previous response	806.55	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 19, 2024	Previous Calibration:	December 8, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	10:15	End Time (MST):	13:45
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.7	----	Correlation Coefficient	0.999997	
800.3	803.7	0.9958			≥0.995
400.2	401.4	0.9969	Slope	1.003770	
200.1	200.2	0.9994			0.90 - 1.10
			Intercept	0.040000	+/-30



SO2 Calibration Plot

Date: January 19, 2024

Location: Mannix







# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Mannix Station number: AMS05  
 Calibration Date: January 4, 2024 Last Cal Date: December 12, 2023  
 Start time (MST): 10:07 End time (MST): 16:36  
 Reason: Cylinder Change and Maintenance

### Calibration Standards

Cal Gas Concentration: 4.96 ppm Cal Gas Exp Date: November 15, 2026  
 Cal Gas Cylinder #: DT0037363  
 Removed Cal Gas Conc: 4.92 ppm Rem Gas Exp Date: February 9, 2024  
 Removed Gas Cyl #: EY0002433 Diff between cyl: -1.5%  
 Calibrator Make/Model: API T700 Serial Number: 1845  
 ZAG Make/Model: API T701H Serial Number: 832

### Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1203169745  
 Converter make: Global Converter serial #: 2022-196  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.988046	0.971240	Backgd or Offset: 2.18	2.18
Calibration intercept:	0.440521	0.762815	Coeff or Slope: 0.866	0.866

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4919	81.3	80.0	78.5	1.022
as found 2nd point	4960	40.7	40.0	40.5	0.994
as found 3rd point	4980	20.3	20.0	20.0	1.009
new cylinder response	4919	80.6	80.0	77.3	1.034

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.8	----
high point	4919	80.6	80.0	78.4	1.020
second point	4960	40.3	40.0	39.7	1.007
third point	4980	20.2	20.0	20.1	0.997
as left zero	5000	0.0	0.0	1.0	----
as left span	4919	80.6	80.0	76.8	1.041
SO2 Scrubber Check	4920	80.0	800.0	-0.1	----

Date of last scrubber change:	Ave Corr Factor	1.008
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 78.3 Prev response: 79.48 \*% change: -1.5%  
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 0.979340 AF Intercept: 0.520256  
 Baseline Corr 3rd AF pt: 19.8 AF Correlation: 0.999755

\* = > +/-5% change initiates investigation

Notes: Completed multipoint as founds. Changed the external pump as a preventative maintenance.  
 Changed the cal gas cylinder because its getting low. Instrument has a very slow response, could be the new gas. Will monitor and swap it out if needed. No adjustments made.  
 Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 12, 2023
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	10:07	End Time (MST):	16:36
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1203169745

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.8	----	Correlation Coefficient	0.999991	≥0.995
80.0	78.4	1.0199			
40.0	39.7	1.0069	Slope	0.971240	0.90 - 1.10
20.0	20.1	0.9969			
			Intercept	0.762815	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 4, 2024

Location: Mannix





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Mannix	Station number:	AMS05
Calibration Date:	January 9, 2024	Last Cal Date:	January 4, 2024
Start time (MST):	9:50	End time (MST):	11:55
Reason:	Removal		

### Calibration Standards

Cal Gas Concentration:	4.96	ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	DT0037363			
Removed Cal Gas Conc:	4.96	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1845
ZAG Make/Model:	API T701H		Serial Number:	832

### Analyzer Information

Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1203169745
Converter make:	Global	Converter serial #:	2022-196
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.971240		Backgd or Offset:	2.18	
Calibration intercept:	0.762815		Coeff or Slope:	0.866	

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4919	81.3	80.6	76.8	1.054
as found 2nd point	4960	40.7	40.4	39.6	1.027
as found 3rd point	4980	20.3	20.1	20.0	1.022
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

### SO<sub>2</sub> Scrubber Check

Date of last scrubber change:	Ave Corr Factor
Date of last converter efficiency test:	efficiency

Baseline Corr As found:	76.5	Prev response:	79.09	*% change:	-3.4%
Baseline Corr 2nd AF pt:	39.3	AF Slope:	0.947493	AF Intercept:	0.740376
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999774		

\* = > +/-5% change initiates investigation


Notes: Spans are very slow to reach target. Removal calibration to swap out the instrument.

Calibration Performed By: Max Farrell

H<sub>2</sub>S Calibration Plot

Date: January 9, 2024

Location: Mannix





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Mannix Station number: AMS05  
 Calibration Date: January 10, 2024 Last Cal Date:  
 Start time (MST): 10:23 End time (MST): 13:30  
 Reason: Install

### Calibration Standards

Cal Gas Concentration: 4.96 ppm Cal Gas Exp Date: November 15, 2026  
 Cal Gas Cylinder #: DT0037363  
 Removed Cal Gas Conc: 4.96 ppm Rem Gas Exp Date:  
 Removed Gas Cyl #: Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 1845  
 ZAG Make/Model: API T701H Serial Number: 832

### Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1200326169  
 Converter make: Global Converter serial #: 2022225  
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:		0.999261	Backgd or Offset:	1.27
Calibration intercept:		0.182265	Coeff or Slope:	1.022

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4919	80.6	80.0	80.0	1.000
second point	4960	40.3	40.0	40.3	0.992
third point	4980	20.2	20.0	20.2	0.992
as left zero	5000	0.0	0.0	0.6	----
as left span	4919	80.6	80.0	79.9	1.001
SO2 Scrubber Check	4920	80.0	800.0	0.6	----
Date of last scrubber change:				Ave Corr Factor	0.994
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: NA Prev response: NA \*% change: NA  
 Baseline Corr 2nd AF pt: NA AF Slope: NA AF Intercept: NA  
 Baseline Corr 3rd AF pt: NA AF Correlation: NA

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter. Ran a SO2 scrubber test after calibrator zero. Adjusted the span.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	10:23	End Time (MST):	13:30
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326169

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999988	
80.0	80.0	0.9995			≥0.995
40.0	40.3	0.9919	Slope	0.999261	
20.0	20.2	0.9920			0.90 - 1.10
			Intercept	0.182265	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 10, 2024

Location: Mannix







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	January 19, 2024	Last Cal Date:	December 8, 2023
Start time (MST):	10:15	End time (MST):	13:45
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH <sub>4</sub> Cal Gas Conc.	504.9 ppm	CH <sub>4</sub> Equiv Conc.	1076.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	504.9 ppm	CH <sub>4</sub> Equiv Conc.	1076.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.61E-05	2.64E-05	NMHC SP Ratio:	4.48E-05	4.43E-05
CH <sub>4</sub> Retention time:	15.00	15.20	NMHC Peak Area:	204322	206470
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.28	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.23	1.000
second point	4960	40.0	8.61	8.61	1.001
third point	4980	20.0	4.31	4.29	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.22	1.000
Average Correction Factor					1.001

Baseline Corr AF:	17.28	Prev response	17.21	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	9.15	9.28	0.986
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.15	1.000
second point	4960	40.0	4.57	4.58	0.999
third point	4980	20.0	2.29	2.28	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.15	9.17	0.998
Average Correction Factor					1.001
Baseline Corr AF:	9.28	Prev response	9.19	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	8.00	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.09	0.999
second point	4960	40.0	4.04	4.03	1.003
third point	4980	20.0	2.02	2.01	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.05	1.003
Average Correction Factor					1.002
Baseline Corr AF:	8.00	Prev response	8.02	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999274	1.000521
THC Cal Offset:	-0.003400	-0.007800
CH <sub>4</sub> Cal Slope:	0.993789	1.001401
CH <sub>4</sub> Cal Offset:	-0.005600	-0.007000
NMHC Cal Slope:	1.003992	0.999969
NMHC Cal Offset:	0.002200	-0.001200

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 19, 2024	Previous Calibration:	December 8, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:15	End Time (MST):	13:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥0.995
17.23	17.23	0.9997			
8.61	8.61	1.0008	Slope	1.000521	0.90 - 1.10
4.31	4.29	1.0031			
			Intercept	-0.007800	+/-0.5





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 19, 2024	Previous Calibration:	December 8, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:15	End Time (MST):	13:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	≥0.995
8.08	8.09	0.9988			
4.04	4.03	1.0025			
2.02	2.01	1.0038			
			Slope	1.001401	0.90 - 1.10
			Intercept	-0.007000	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 19, 2024	Previous Calibration:	December 8, 2023
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:15	End Time (MST):	13:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	$\geq 0.995$
9.15	9.15	1.0003			
4.57	4.58	0.9993			
2.29	2.28	1.0026			
			Slope	0.999969	0.90 - 1.10
			Intercept	-0.001200	+/-0.5



NMHC Calibration Plot

Date: January 19, 2024

Location: Mannix





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS06 PATRICIA MCINNES JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	January 5, 2024	Last Cal Date:	December 13, 2023
Start time (MST):	10:32	End time (MST):	14:50
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.78	ppm	Cal Gas Exp Date:	September 9, 2024
Cal Gas Cylinder #:	AAL070632			
Removed Cal Gas Conc:	49.78	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	5608

### Analyzer Information

Analyzer make: Thermo 43i  
 Analyzer Range 0 - 1000 ppb

Analyzer serial #: 1160290013

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003106	0.999660	Backgd or Offset:	17.7	17.7
Calibration intercept:	1.659655	1.240714	Coeff or Slope:	0.922	0.922

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4920	80.3	799.5	797.8	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.3	799.5	800.2	0.999
second point	4960	40.2	400.2	400.8	0.999
third point	4980	20.1	200.1	203.3	0.984
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	80.3	799.5	799.3	1.000
<b>Average Correction Factor</b>					<b>0.994</b>

Baseline Corr As found:	797.90	Previous response	803.61	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made. Maintenance completed to the H2 gen for the THC instrument after as founds.

Calibration Performed By: Max Farrell





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 13, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:32	End Time (MST):	14:50
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999983	
799.5	800.2	0.9991			≥0.995
400.2	400.8	0.9986	Slope	0.999660	
200.1	203.3	0.9843			0.90 - 1.10
			Intercept	1.240714	+/-30



SO2 Calibration Plot

Date: January 5, 2024

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	January 18, 2024	Last Cal Date:	December 7, 2023
Start time (MST):	9:31	End time (MST):	13:47
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.328	ppm	Cal Gas Exp Date:	February 14, 2025
Cal Gas Cylinder #:	CC506659			
Removed Cal Gas Conc:	5.328	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

### Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.991885	1.002314	Backgd or Offset:	2.01	1.95
Calibration intercept:	0.580307	0.440228	Coeff or Slope:	1.168	1.150

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4925	75.1	80.0	81.2	0.987
as found 2nd point	4963	37.5	40.0	41.2	0.972
as found 3rd point	4981	18.8	20.0	21.0	0.959
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4925	75.1	80.0	80.5	0.994
second point	4963	37.5	40.0	40.7	0.982
third point	4981	18.8	20.0	20.7	0.968
as left zero	5000	0.0	0.0	0.3	----
as left span	4925	75.1	80.0	80.7	0.992
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	0.981
Date of last converter efficiency test:	efficiency				

Baseline Corr As found:	81.1	Prev response:	79.96	*% change:	1.4%
Baseline Corr 2nd AF pt:	41.1	AF Slope:	1.011741	AF Intercept:	0.460263
Baseline Corr 3rd AF pt:	20.9	AF Correlation:	0.999902		

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 7, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:31	End Time (MST):	13:47
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient	$\geq 0.995$
80.0	80.5	0.9941		
40.0	40.7	0.9817	Slope	0.90 - 1.10
20.0	20.7	0.9678		
			Intercept	+/-3



TRS Calibration Plot

Date: January 18, 2024

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	January 5, 2024	Last Cal Date:	December 13, 2023
Start time (MST):	10:32	End time (MST):	14:50
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH <sub>4</sub> Cal Gas Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	501.6 ppm	CH <sub>4</sub> Equiv Conc.	1066.2 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	4602

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.17E-04	2.17E-04	NMHC SP Ratio:	4.91E-05
CH <sub>4</sub> Retention time:	14.0	14.0	NMHC Peak Area:	184991
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4920	80.3	17.12	17.37	0.986
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	17.16	0.998
second point	4960	40.2	8.57	8.68	0.988
third point	4980	20.1	4.29	4.44	0.965
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.12	17.21	0.995
Average Correction Factor					0.984

Baseline Corr AF:	17.36	Prev response	17.18	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4919.7	80.3	9.07	9.36	0.969
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	9.07	9.09	0.997
second point	4960	40.2	4.54	4.60	0.988
third point	4980	20.1	2.27	2.34	0.969
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	9.07	9.17	0.989
Average Correction Factor					0.985
Baseline Corr AF:	9.35	Prev response	9.13	*% change	2.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919.7	80.3	8.06	8.01	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	8.06	8.07	0.998
second point	4960	40.2	4.03	4.08	0.988
third point	4980	20.1	2.02	2.10	0.961
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	8.06	8.04	1.001
Average Correction Factor					0.982
Baseline Corr AF:	8.01	Prev response	8.05	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001451	1.000106
THC Cal Offset:	0.034020	0.074604
CH <sub>4</sub> Cal Slope:	0.999322	0.999043
CH <sub>4</sub> Cal Offset:	0.004445	0.039427
NMHC Cal Slope:	1.003545	1.001025
NMHC Cal Offset:	0.028776	0.034777

Notes: Changed the inlet filter and completed maintenance on the H2 generator after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 13, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:32	End Time (MST):	14:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999913	$\geq 0.995$			
17.12	17.16	0.9976						
8.57	8.68	0.9878				Slope	1.000106	0.90 - 1.10
4.29	4.44	0.9653						
			Intercept	0.074604	$\pm 0.5$			







# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 13, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:32	End Time (MST):	14:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999888	≥0.995			
8.06	8.07	0.9982						
4.03	4.08	0.9882				Slope	0.999043	0.90 - 1.10
2.02	2.10	0.9611						
			Intercept	0.039427	+/-0.5			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 13, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:32	End Time (MST):	14:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999933	$\geq 0.995$			
9.07	9.09	0.9970						
4.54	4.60	0.9878				Slope	1.001025	0.90 - 1.10
2.27	2.34	0.9691						
			Intercept	0.034777	$\pm 0.5$			



NMHC Calibration Plot

Date: January 5, 2024

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Patricia McInnes  
Calibration Date: January 8, 2024  
Start time (MST): 9:54  
Reason: Routine  
Station number: AMS06  
Last Cal Date: December 4, 2023  
End time (MST): 14:46

### Calibration Standards

NO Gas Cylinder #: T30YCWN  
NOX Cal Gas Conc: 47.94 ppm  
Removed Cylinder #: N/A  
Removed Gas NOX Conc: 47.94 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701  
Cal Gas Expiry Date: April 11, 2025  
NO Cal Gas Conc: 46.39 ppm  
Removed Gas Exp Date: N/A  
Removed Gas NO Conc: 46.39 ppm  
NO gas Diff:  
Serial Number: 3566  
Serial Number: 4602

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1172750022

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.825	0.825	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.987	0.987	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.2	157.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.992028	0.998293
NO <sub>x</sub> Cal Offset:	2.457352	1.916153
NO Cal Slope:	0.993713	0.999944
NO Cal Offset:	1.244073	1.062930
NO <sub>2</sub> Cal Slope:	0.998330	0.999549
NO <sub>2</sub> Cal Offset:	-0.223314	-1.654321



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	0.3	-0.5	----	----
as found span	4914	86.2	826.5	799.7	26.7	828.3	800.3	27.9	0.9978	0.9993
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.4	-0.2	----	----
high point	4914	86.2	826.5	799.7	26.7	826.0	800.4	25.6	1.0005	0.9992
second point	4957	43.1	413.2	399.9	13.4	415.5	401.3	14.2	0.9945	0.9964
third point	4978	21.6	207.1	200.4	6.7	210.4	202.1	8.4	0.9844	0.9917
as left zero	5000	0.0	0.0	0.0	0.0	0.4	0.7	-0.3	----	----
as left span	4914	86.2	826.5	799.7	26.7	819.5	794.6	24.8	1.0085	1.0065
Average Correction Factor									0.9932	0.9958

Corrected As found	NO <sub>x</sub> = 828.5 ppb	NO = 800.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.7%
Previous Response	NO <sub>x</sub> = 822.3 ppb	NO = 795.9 ppb		*Percent Change	NO = 0.5%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.5	402.3	420.9	419.7	1.0029	99.7%
2nd GPT point (200 ppb O3)	796.5	601.5	221.7	219.9	1.0083	99.2%
3rd GPT point (100 ppb O3)	796.5	700.1	123.1	119.4	1.0312	97.0%
Average Correction Factor					1.0141	98.6%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 8, 2024	Previous Calibration:	December 4, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:54	End Time (MST):	14:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
826.5	826.0	1.0005			
413.2	415.5	0.9945			
207.1	210.4	0.9844			
			Slope	0.998293	0.90 - 1.10
			Intercept	1.916153	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 8, 2024	Previous Calibration:	December 4, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:54	End Time (MST):	14:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.7	800.4	0.9992		
399.9	401.3	0.9964		
200.4	202.1	0.9917		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 8, 2024	Previous Calibration:	December 4, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:54	End Time (MST):	14:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
420.9	419.7	1.0029			
221.7	219.9	1.0083			
123.1	119.4	1.0312			
			Slope	0.999549	0.90 - 1.10
			Intercept	-1.654321	+/-20

**NO<sub>2</sub> Calibration Curve**






NO<sub>x</sub> Calibration Plot

Date: January 8, 2024

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Patricia McInnes      Station number: AMS06  
 Calibration Date: January 2, 2024      Last Cal Date: December 18, 2023  
 Start time (MST): 11:52      End time (MST): 15:41  
 Reason: Maintenance

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: API T700      Serial Number: 3566  
 ZAG Make/Model: API T701      Serial Number: 4602

### Analyzer Information

Analyzer make: Thermo 49i      Analyzer serial #: 1300156234  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007657	1.003886	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	-1.340000	-1.480000	Coeff or Slope:	1.039	1.026

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	-0.3	----
as found span	5000	1303.0	400.0	402.4	0.994
as found 2nd point	5000	966.5	200.0	199.2	1.004
as found 3rd point	5000	794.3	100.0	98.0	1.020
calibrator zero	5000	800.0	0.0	0.1	----
high point	5000	1303.0	400.0	400.9	0.998
second point	5000	966.5	200.0	198.4	1.008
third point	5000	794.3	100.0	97.4	1.027
as left zero	5000	800.0	0.0	0.2	----
as left span	5000	1303.0	400.0	402.8	0.993
Average Correction Factor					1.011

Baseline Corr As found:	402.7	Previous response	401.7	*% change	0.2%
Baseline Corr 2nd AF pt:	-203.2	AF Slope:	1.008257	AF Intercept:	-1.620000
Baseline Corr 3rd AF pt:	-101.2	AF Correlation:	0.999951		

\* = > +/-5% change initiates investigation

Notes: Changed the external pump and the inlet filter after multipoint as founds. Adjusted the span.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 2, 2024	Previous Calibration:	December 18, 2023
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	11:52	End Time (MST):	15:41
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999929	≥0.995
400.0	400.9	0.9978			
200.0	198.4	1.0081	Slope	1.003886	0.90 - 1.10
100.0	97.4	1.0267			
			Intercept	-1.480000	+/- 5



O<sub>3</sub> Calibration Plot

Date: January 2, 2024

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Patricia McInnes      Station number: AMS 06  
 Calibration Date: January 18, 2024      Last Cal Date: December 1, 2023  
 Start time (MST): 14:03      End time (MST): 15:01

Analyzer Make: API T640      S/N: 766  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT FP-25      S/N: 388755  
 Temp/RH standard: ALICAT FP-25      S/N: 388755

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-15.5	-15.1	-15.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	741.7	743.8	741.7	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	5.15	5.02	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test:      Date of check: January 18, 2024      Last Cal Date: December 1, 2023  
 PM w/o HEPA: 7.9      PM w/ HEPA: 0      <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning :      Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check:      PM w/o HEPA: \_\_\_\_\_ w/ HEPA: \_\_\_\_\_  
 Date Optical Chamber Cleaned:      December 1, 2023      <0.2 ug/m3  
 Disposable Filter Changed:      December 1, 2023

### Annual Maintenance

Date Sample Tube Cleaned:      April 13, 2023  
 Date RH/T Sensor Cleaned:      April 13, 2023

Quarterly calibrations completed last month. Leak check passed, no adjustments made.

Notes:

Calibration by:      Max Farrell



# Wood Buffalo Environmental Association

## TN - NO<sub>x</sub> - NH<sub>3</sub> Calibration Report

Version-05-2023

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	January 16, 2024	Last Cal Date:	December 6, 2023
Start time (MST):	9:28	End time (MST):	13:31
NH3 Cal Date:	January 16, 2024	Last Cal Date:	December 6, 2023
Start time (MST):	13:32	End time (MST):	15:10
Reason:	Routine		

### Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	NA
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.29	ppm	NH3 Gas Cylinder #:	EB0108520
Removed NH3 Conc:	76.29	ppm	NH3 Cal Gas Expiry:	August 22, 2024
NH3 gas Diff:			Removed Cylinder #:	NA
Calibrator Model:	API T700		Removed cyl Expiry:	NA
ZAG make/model:	API T701		Serial Number:	3566
			Serial Number:	4602

### Analyzer Information

Analyzer model: Teledyne API T201	Analyzer serial #: 808
Converter model: Teledyne API T501	Converter serial #: 484
NH3 Range (ppb): 0 - 2000 ppb	Reaction cell Press: 4.80
NOX Range (ppb): 0 - 1000 ppb	Sample Flow: 26.6

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.854	0.854	TN coefficient:	0.849	0.849
NOX coefficient:	0.848	0.848	NO bkgrnd:	-0.985	-0.985
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	-0.562	-0.562
NH3 coefficient:	0.891	0.891	TN bkgrnd:	5.018	5.018

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.997270	1.000699
NO <sub>x</sub> Cal Offset:	0.936426	1.395952
NO Cal Slope:	0.997858	1.003275
NO Cal Offset:	0.443186	0.522150
NO <sub>2</sub> Cal Slope:	1.001189	0.994054
NO <sub>2</sub> Cal Offset:	-1.443145	-1.155779
NH3 Cal Slope:	1.005013	1.000855
NH3 Cal Offset:	3.755107	2.210125
TN Cal Slope:	1.011495	1.008599
TN Cal Offset:	4.256407	2.771030



# Wood Buffalo Environmental Association

## TN - NOX - NH<sub>3</sub> Calibration Report

Version-05-2023

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-1.9	0.2	-2.0	----	----
as found NO	4914	86.2	826.5	826.5	----	827.9	826.0	1.7	0.998	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-1.2	0.4	-1.6	----	----
high NO point	4914	86.2	826.5	826.5	----	828.0	828.0	0.0	0.998	----
NO/O3 point	4914	86.2	826.5	826.5	----	827.0	824.2	2.8	0.999	----
as found NH3	3417	82.6	1800.6	----	1800.6	1816.3	----	1801.7	0.991	0.999
new NH3 cyl rp	3417	82.6	1800.6	----	1800.6	----	----	----	----	----
first NH3	3417	82.6	1800.6	----	1800.6	1816.3	----	1801.7	0.991	0.999
second NH3	3454	45.9	1000.5	----	1000.5	1014.7	----	1006.8	0.986	0.994
third NH3	3477	22.9	499.2	----	499.2	510.0	----	505.1	0.979	0.988
<b>Average Correction Factor</b>									<b>0.9987</b>	<b>0.9938</b>

Corrected As found      TN = 829.8 ppb      NO<sub>x</sub> = 825.8 ppb      NH3 = 1803.7 ppb

Previous Response      TN = 840.2 ppb      NO<sub>x</sub> = 825.1 ppb      NH3 = 1813.4 ppb

NH3 Previous Converter Efficiency = 89.1%

NH3 Current Converter Efficiency = 89.1%

\*Percent Change      TN = -1.3%

\*Percent Change      NO<sub>x</sub> = 0.1%

\*Percent Change      NH3 = -0.5%

\* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## NO<sub>x</sub> - NO - NO<sub>2</sub> Calibration Report

Version-05-2023

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	-1.9	----	----
as found span	4914	86.2	826.5	799.7	826.5	826.0	794.7	827.9	1.0005	1.0063
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.4	0.2	-1.2	----	----
high point	4914	86.2	826.5	799.7	826.5	828.0	803.0	828.0	0.9981	0.9959
second point	4957	43.1	413.2	399.9	413.2	415.2	401.0	416.5	0.9953	0.9972
third point	4978	21.6	207.1	200.4	207.1	209.8	202.5	211.2	0.9872	0.9897
Average Correction Factor									0.9935	0.9943

Baseline Corr As fnd	TN = 829.8 ppb	NO <sub>x</sub> = 825.8 ppb	NO = 794.7 ppb	*Percent Change	TN = -1.3%
Previous Response	TN = 840.2 ppb	NO <sub>x</sub> = 825.1 ppb	NO = 798.5 ppb	*Percent Change	NO <sub>x</sub> = 0.1%
				*Percent Change	NO = -0.5%
				<i>* = &gt; +/-5% change initiates investigation</i>	

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	0.1	----	----
calibration zero	----	----	0.0	0.2	----	----
1st GPT point (400 ppb O3)	800.3	399.8	427.2	424.2	1.0071	99.3%
2nd GPT point (200 ppb O3)	800.3	600.2	226.8	224.0	1.0126	98.8%
3rd GPT point (100 ppb O3)	800.3	700.7	126.3	122.7	1.0295	97.1%
Average Correction Factor					1.0164	98.4%

Notes:

Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By:

Max Farrell





# Wood Buffalo Environmental Association

## TN Calibration Summary

Version-05-2023


### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	December 6, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:28	End Time (MST):	13:31
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-1.2	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ 0.90 - 1.10 +/-20
1800.6	1816.3	0.9914		
1000.5	1014.7	0.9860		
499.2	510.0	0.9788		
			0.999975	
			1.008599	
			2.771030	

TN Calibration Curve





# Wood Buffalo Environmental Association

## NH<sub>3</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	December 6, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:28	End Time (MST):	13:31
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-1.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.6	1801.7	0.9994		
1000.5	1006.8	0.9938		
499.2	505.1	0.9883		
			0.999976	
			1.000855	
			2.210125	





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	December 6, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:28	End Time (MST):	13:31
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	≥0.995	
826.5	828.0	0.9981			
413.2	415.2	0.9953			
207.1	209.8	0.9872			
			Slope	1.000699	0.90 - 1.10
			Intercept	1.395952	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	December 6, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:28	End Time (MST):	13:31
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ $0.90 - 1.10$ $\pm 20$
799.7	803.0	0.9959		
399.9	401.0	0.9972		
200.4	202.5	0.9897		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-05-2023

### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	December 6, 2023
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:28	End Time (MST):	13:31
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
427.2	424.2	1.0071		
226.8	224.0	1.0126		
126.3	122.7	1.0295		



NO<sub>x</sub> Calibration Plot

Date: January 16, 2024


Location: Patricia McInnes



# NH<sub>3</sub> Calibration Plot

Date: January 16, 2024

Location: Patricia McInnes





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS07 ATHABASCA VALLEY JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	January 18, 2024	Last Cal Date:	December 15, 2023
Start time (MST):	10:20	End time (MST):	16:00
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.06	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC320556			
Removed Cal Gas Conc:	50.52	ppm	Rem Gas Exp Date:	December 29, 2028
Removed Gas Cyl #:	CC282115		Diff between cyl:	-0.3%
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 701H		Serial Number:	198

### Analyzer Information

Analyzer make: Thermo 43i-LTE                      Analyzer serial #: 1507864683  
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998820	1.003653	Backgd or Offset:	2.59	2.62
Calibration intercept:	2.284124	2.143576	Coeff or Slope:	0.834	0.846

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	79.2	800.2	791.9	1.010
as found 2nd point					
as found 3rd point					
new cylinder response	4920	79.8	799.0	788.2	1.014
calibrator zero	5000	0.0	0.0	0.2	----
high point	4920	79.8	799.0	802.4	0.996
second point	4960	39.9	399.5	406.1	0.984
third point	4980	20.0	200.2	203.7	0.983
as left zero	5000	0.0	0.0	0.0	----
as left span	4920	79.8	799.0	800.1	0.999
Average Correction Factor					0.987

Baseline Corr As found:	791.90	Previous response	801.54	*% change	-1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Calibration gas swapped out after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:20	End Time (MST):	16:00
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999955	
799.0	802.4	0.9957			≥0.995
399.5	406.1	0.9837	Slope	1.003653	
200.2	203.7	0.9830			0.90 - 1.10
			Intercept	2.143576	+/-30



SO2 Calibration Plot

Date: January 18, 2024

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Athabasca Valley      Station number: AMS07  
 Calibration Date: January 9, 2024      Last Cal Date: December 31, 2023  
 Start time (MST): 10:00      End time (MST): 14:00  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.25 ppm      Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC504080  
 Removed Cal Gas Conc: 5.25 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3805  
 ZAG Make/Model: API T701H      Serial Number: 198

### Analyzer Information

Analyzer make: Thermo 43i LTE      Analyzer serial #: 1180540018  
 Converter make: CDN-101      Converter serial #: 551  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.012033	1.004826	Backgd or Offset: 2.35	2.40
Calibration intercept:	-0.062267	0.237795	Coeff or Slope: 0.893	0.906

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4925	75.5	79.3	77.8	1.020
as found 2nd point	4962	37.7	39.6	39.0	1.018
as found 3rd point	4981	18.9	19.8	19.4	1.028
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4925	75.5	79.3	79.9	0.993
second point	4962	37.7	39.6	40.1	0.988
third point	4981	18.9	19.9	20.2	0.983
as left zero	5000	0.0	0.0	0.6	----
as left span	4925	75.5	79.3	79.4	0.999
SO2 Scrubber Check	4920	79.2	800.4	0.4	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	0.988
Date of last converter efficiency test:	April 22, 2022			92.6% efficiency	

Baseline Corr As found: 77.7      Prev response: 80.16      \*% change: -3.2%  
 Baseline Corr 2nd AF pt: 38.9      AF Slope: 0.981028      AF Intercept: 0.057909  
 Baseline Corr 3rd AF pt: 19.3      AF Correlation: 0.999991

\* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 9, 2024	Previous Calibration:	December 31, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:00	End Time (MST):	14:00
Analyzer make:	CDN-101	Analyzer serial #:	551

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999998	≥0.995
79.3	79.9	0.9926			
39.6	40.1	0.9878	Slope	1.004826	0.90 - 1.10
19.9	20.2	0.9830			
			Intercept	0.237795	+/-3



TRS Calibration Plot

Date: January 9, 2024

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	January 18, 2024	Last Cal Date:	December 15, 2023
Start time (MST):	10:20	End time (MST):	16:00
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	March 10, 2031
CH <sub>4</sub> Cal Gas Conc.	496.0 ppm	CH <sub>4</sub> Equiv Conc.	1059.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	CC282115	Removed Gas Expiry:	December 29, 2028
Removed CH <sub>4</sub> Conc.	501.2 ppm	CH <sub>4</sub> Equiv Conc.	1075.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.7 ppm	Diff between cyl (THC):	1.9%
Diff between cyl (CH <sub>4</sub> ):	0.5%	Diff between cyl (NM):	3.2%
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	0.000251	N/A	NMHC SP Ratio:	4.42E-05
CH <sub>4</sub> Retention time:	13.8	N/A	NMHC Peak Area:	205554
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	17.25	0.987
as found 2nd point					
as found 3rd point					
new cylinder response	4920	79.8	16.91	17.47	0.968
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	16.91	17.02	0.994
second point	4960	39.9	8.46	8.51	0.994
third point	4980	20.0	4.24	4.26	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	16.91	17.03	0.993

Average Correction Factor				0.994
Baseline Corr AF:	17.25	Prev response	17.14	*% change 0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	9.09	9.16	0.993
as found 2nd point					
as found 3rd point					
new cylinder response	4920	79.8	9.00	9.37	0.960
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.00	9.10	0.989
second point	4960	39.9	4.50	4.59	0.980
third point	4980	20.0	2.26	2.33	0.967
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.00	9.12	0.987
Average Correction Factor					0.979
Baseline Corr AF:	9.16	Prev response	9.18	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.94	8.09	0.982
as found 2nd point					
as found 3rd point					
new cylinder response	4920	79.8	7.92	8.10	0.977
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	7.92	7.92	0.999
second point	4960	39.9	3.96	3.92	1.009
third point	4980	20.0	1.98	1.93	1.028
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	7.92	7.92	1.000
Average Correction Factor					1.012
Baseline Corr AF:	8.09	Prev response	7.96	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.006595	1.006214
THC Cal Offset:	0.000457	-0.000562
CH <sub>4</sub> Cal Slope:	1.005894	1.002531
CH <sub>4</sub> Cal Offset:	-0.023559	-0.029927
NMHC Cal Slope:	1.006942	1.009365
NMHC Cal Offset:	0.023816	0.028966

Notes: Calibration cylinder swapped out after as founds. Span adjusted. Could not get as left diagnostics due to instrument going offline during as lefts.

Calibration Performed By: Aswin Sasi Kumar





# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:20	End Time (MST):	16:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	$\geq 0.995$			
16.91	17.02	0.9939						
8.46	8.51	0.9935				Slope	1.006214	0.90 - 1.10
4.24	4.26	0.9946						
			Intercept	-0.000562	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:20	End Time (MST):	16:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999935	$\geq 0.995$			
7.92	7.92	0.9993						
3.96	3.92	1.0092				Slope	1.002531	0.90 - 1.10
1.98	1.93	1.0280						
			Intercept	-0.029927	$\pm 0.5$			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 15, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:20	End Time (MST):	16:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999953	≥0.995
9.00	9.10	0.9893			
4.50	4.59	0.9803	Slope	1.009365	0.90 - 1.10
2.26	2.33	0.9674			
			Intercept	0.028966	+/-0.5



NMHC Calibration Plot

Date: January 18, 2024

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Athabasca Valley  
 Station number: AMS07  
 Calibration Date: January 5, 2024  
 Last Cal Date: December 13, 2023  
 Start time (MST): 10:20  
 End time (MST): 16:00  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T2Y1KA4  
 Cal Gas Expiry Date: November 30, 2023  
 NOX Cal Gas Conc: 50.92 ppm  
 NO Cal Gas Conc: 49.92 ppm  
 Removed Cylinder #: NA  
 Removed Gas Exp Date: NA  
 Removed Gas NOX Conc: 50.92 ppm  
 Removed Gas NO Conc: 49.92 ppm  
 NOX gas Diff: NO gas Diff:  
 Calibrator Model: API T700  
 Serial Number: 3805  
 ZAG make/model: API T701H  
 Serial Number: 198

### Analyzer Information

Analyzer make: Thermo 42i  
 Analyzer serial #: 1160120024  
 NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.095	1.092	NO bkgnd or offset:	7.7	7.7
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	7.9	7.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	212.7	216.3

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.999335	0.998118
NO <sub>x</sub> Cal Offset:	1.718942	1.478599
NO Cal Slope:	1.000317	1.000317
NO Cal Offset:	1.435177	1.135182
NO <sub>2</sub> Cal Slope:	0.998685	1.000671
NO <sub>2</sub> Cal Offset:	0.269234	0.394845



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as found span	4920	80.2	816.7	800.7	16.0	821.5	804.0	17.6	0.9942	0.9959
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
high point	4920	80.2	816.7	800.7	16.0	815.6	801.2	14.4	1.0014	0.9994
second point	4960	40.1	408.4	400.4	8.0	411.0	403.1	8.0	0.9936	0.9932
third point	4980	20.0	203.7	199.7	4.0	205.2	201.4	3.8	0.9926	0.9915
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
as left span	4920	80.2	816.7	398.9	417.8	820.7	399.2	421.7	0.9952	0.9992
Average Correction Factor									0.9959	0.9947

Corrected As found	NO <sub>x</sub> = 821.6 ppb	NO = 804.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.5%
Previous Response	NO <sub>x</sub> = 817.9 ppb	NO = 802.4 ppb		*Percent Change	NO = 0.2%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	798.4	396.6	417.8	418.4	0.9987	100.1%
2nd GPT point (200 ppb O <sub>3</sub> )	798.4	599.7	214.7	215.4	0.9969	100.3%
3rd GPT point (100 ppb O <sub>3</sub> )	798.4	698.4	116.0	116.7	0.9943	100.6%
Average Correction Factor					0.9966	100.3%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 13, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:20	End Time (MST):	16:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	≥0.995	
816.7	815.6	1.0014			
408.4	411.0	0.9936			
203.7	205.2	0.9926			
			Slope	0.998118	0.90 - 1.10
			Intercept	1.478599	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 13, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:20	End Time (MST):	16:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
800.7	801.2	0.9994			0.999987
400.4	403.1	0.9932			1.000317
199.7	201.4	0.9915			1.135182







# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 13, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:20	End Time (MST):	16:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
417.8	418.4	0.9987		
214.7	215.4	0.9969		
116.0	116.7	0.9943		



NO<sub>x</sub> Calibration Plot

Date: January 5, 2024

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Athabasca Valley      Station number: AMS07  
 Calibration Date: January 12, 2024      Last Cal Date: December 14, 2023  
 Start time (MST): 12:45      End time (MST): 15:52  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: T700      Serial Number: 3805  
 ZAG Make/Model: T701H      Serial Number: 198

### Analyzer Information

Analyzer make: Thermo 49i      Analyzer serial #: 1152220023  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994743	0.995143	Backgd or Offset:	-2.7	-1.6
Calibration intercept:	0.420000	0.500000	Coeff or Slope:	1.549	1.549

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	5000	1414.8	400.0	397.1	1.007
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.3	----
high point	5000	1415.7	400.0	398.5	1.004
second point	5000	1039.9	200.0	199.5	1.003
third point	5000	856.2	100.0	100.3	0.997
as left zero	5000	0.0	0.0	0.4	----
as left span	5000	1416.0	400.0	399.7	1.001
Average Correction Factor					1.001

Baseline Corr As found:	396.7	Previous response	398.3	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 12, 2024	Previous Calibration:	December 14, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	12:45	End Time (MST):	15:52
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999999	
400.0	398.5	1.0038			≥0.995
200.0	199.5	1.0025	Slope	0.995143	
100.0	100.3	0.9970			0.90 - 1.10
			Intercept	0.500000	+/- 5



O<sub>3</sub> Calibration Plot

Date: January 12, 2024

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Athabasca Valley Station number: AMS 07  
 Calibration Date: January 18, 2024 Last Cal Date: December 13, 2023  
 Start time (MST): 14:03 End time (MST): 15:49

Analyzer Make: API T640 S/N: 645  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754  
 Temp/RH standard: Alicat FP-25BT S/N: 388754

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-14.8	-15.7	-14.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	756.4	754.2	756.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.01	5.12	5.01	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 18, 2024</u>	Last Cal Date: <u>December 13, 2023</u>			
	PM w/o HEPA: <u>3.5</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	N/A	N/A	N/A	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>N/A</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>N/A</u>			

### Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022  
 Date RH/T Sensor Cleaned: December 5, 2022

Notes: Inlet head trap jar was found shattered, replaced jar. Temp, flow and pressure checked. Leak check done.

Calibration by: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	January 2, 2024	Last Cal Date:	December 8, 2023
Start time (MST):	11:50	End time (MST):	15:24
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date:	December 12, 2026
Cal Gas Cylinder #:	LL66942			
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 700H		Serial Number:	198

### Analyzer Information

Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	<b><u>Start</u></b>	<b><u>Finish</u></b>		<b><u>Start</u></b>	<b><u>Finish</u></b>
Calibration slope:	1.000200	0.993944	Backgd or Offset:	4.331	4.425
Calibration intercept:	0.158544	0.080556	Coeff or Slope:	1.087	1.087

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.0	40.1	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.0	39.8	1.006
second point	4967	33.3	20.0	20.1	0.995
third point	4983	16.7	10.0	10.1	0.996
as left zero	5000	0.0	0.0	0.0	----
as left span	4933	66.7	40.0	39.8	1.006
<b>Average Correction Factor</b>					<b>0.999</b>

Baseline Corr As found:	39.97	Prev response:	40.19	*% change:	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 2, 2024	Previous Calibration:	December 8, 2023
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:50	End Time (MST):	15:24
Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999966	≥0.995
40.0	39.8	1.0058			
20.0	20.1	0.9955	Slope	0.993944	0.90 - 1.10
10.0	10.1	0.9961			
			Intercept	0.080556	+/-1.5






CO Calibration Plot

Date: January 2, 2024

Location: Athabasca Valley





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS08 FORT CHIPEWYAN JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	January 18, 2024	Last Cal Date:	December 13, 2023
Start time (MST):	15:22	End time (MST):	18:13
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC196697			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
ZAG Make/Model:	Teledyne API T701		Serial Number:	135

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.988680	1.002560	Backgd or Offset:	4.62	4.56
Calibration intercept:	-0.942140	-1.305017	Coeff or Slope:	0.965	0.989

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-3.0	----
as found span	4920	80.3	800.4	778.7	1.028
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-2.7	----
high point	4920	80.3	800.4	800.8	0.999
second point	4960	40.2	400.7	400.0	1.002
third point	4980	20.1	200.4	201.7	0.993
as left zero	5000	0.0	0.0	-2.5	----
as left span	4920	80.3	800.4	803.0	0.997
Average Correction Factor					0.998

Baseline Corr As found:	781.70	Previous response	790.38	*% change	-1.1%
-------------------------	--------	-------------------	--------	-----------	-------

\* = > +/-5% change initiates investigation

Notes: changed inlet filters after as founds. Made adjustment to span.

Calibration Performed By: Morgan Voyaguer



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 13, 2023
Station Name:	December 13, 2023	Station Number:	Fort Chipewyan
Start Time (MST):	15:22	End Time (MST):	18:13
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-2.7	----	Correlation Coefficient	0.999981	≥0.995
800.4	800.8	0.9995			
400.7	400.0	1.0017	Slope	1.002560	0.90 - 1.10
200.4	201.7	0.9933			
			Intercept	-1.305017	+/-30



SO2 Calibration Plot

Date: January 18, 2024

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	January 31, 2024	Last Cal Date:	January 18, 2024
Start time (MST):	7:20	End time (MST):	22:30
Reason:	Maintenance		

### Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC196697			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
ZAG Make/Model:	Teledyne API T701		Serial Number:	135

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002560	0.995420	Backgd or Offset:	4.56	1.86
Calibration intercept:	-1.305017	1.396452	Coeff or Slope:	0.989	0.989

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-3.1	----
as found span	4920	80.3	800.4	796.5	1.005
as found 2nd point	4960	40.2	400.7	396.2	1.011
as found 3rd point	4980	20.1	200.4	200.6	0.999
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4920	80.3	800.4	797.8	1.003
second point	4960	40.2	400.7	399.6	1.003
third point	4980	20.1	200.4	203.4	0.985
as left zero					
as left span					

Average Correction Factor				0.997
---------------------------	--	--	--	-------

Baseline Corr As found:	799.60	Previous response	801.13	*% change	-0.2%
<i>* = &gt; +/-5% change initiates investigation</i>					

Notes: Pump replaced after multipoint as founds. Zero adjusted only.

Calibration Performed By: Devin Russell



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 31, 2024	Previous Calibration:	January 18, 2024
Station Name:	December 13, 2023	Station Number:	Fort Chipewyan
Start Time (MST):	7:20	End Time (MST):	22:30
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999972	≥0.995
800.4	797.8	1.0032			
400.7	399.6	1.0027	Slope	0.995420	0.90 - 1.10
200.4	203.4	0.9850			
			Intercept	1.396452	+/-30



SO2 Calibration Plot

Date: January 31, 2024

Location: Fort Chipewyan







# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Fort Chipewyan Station number: AMS08  
 Calibration Date: January 26, 2024 Last Cal Date: December 15, 2023  
 Start time (MST): 14:46 End time (MST): 18:28  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: February 9, 2024  
 Cal Gas Cylinder #: EY0002276  
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252  
 ZAG Make/Model: Teledyne API T701 Serial Number: 135

### Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 1203169744  
 Converter make: CDN-101 Converter serial #: 14639  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.988139	0.988284	Backgd or Offset:	0.99	0.99
Calibration intercept:	0.678668	0.598606	Coeff or Slope:	0.741	0.741

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4920	80.5	80.0	79.4	1.014
as found 2nd point	4960	40.2	40.0	40.1	1.009
as found 3rd point	4980	20.1	20.0	20.5	0.999
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.5	----
high point	4920	80.5	80.0	79.6	1.005
second point	4960	40.2	40.0	40.2	0.994
third point	4980	20.1	20.0	20.4	0.979
as left zero	5000	0.0	0.0	0.6	----
as left span	4920	80.5	80.0	79.8	1.003
SO2 Scrubber Check	4919.7	80.3	803.0	0.1	----
Date of last scrubber change:	March 7, 2022			Ave Corr Factor	0.993
Date of last converter efficiency test:	March 15, 2022			100.7% efficiency	

Baseline Corr As found: 78.9 Prev response: 79.74 \*% change: -1.1%  
 Baseline Corr 2nd AF pt: 39.6 AF Slope: 0.985141 AF Intercept: 0.658554  
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999982

\* = > +/-5% change initiates investigation

Notes: Sampled inlet filter changed after as founds. Scrubber check passed.

Calibration Performed By: Matthew Courtoreille



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 26, 2024	Previous Calibration:	December 15, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	14:46	End Time (MST):	18:28
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999991	≥0.995
80.0	79.6	1.0051			
40.0	40.2	0.9940	Slope	0.988284	0.90 - 1.10
20.0	20.4	0.9794			
			Intercept	0.598606	+/-3



TRS Calibration Plot

Date: January 26, 2024

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Fort Chipewyan Station number: AMS08  
Calibration Date: January 10, 2024 Last Cal Date: December 7, 2023  
Start time (MST): 9:04 End time (MST): 13:32  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: CC363447 Cal Gas Expiry Date: February 2, 2024  
NOX Cal Gas Conc: 48.80 ppm NO Cal Gas Conc: 48.80 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 48.80 ppm Removed Gas NO Conc: 48.80 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 3252  
ZAG make/model: Teledyne API T701H Serial Number: 135

### Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1426262592  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.138	1.138	NO bkgnd or offset:	10.4	10.3
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	11.3	11.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	141.2	141.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.993574	1.009182
NO <sub>x</sub> Cal Offset:	-0.140000	1.020000
NO Cal Slope:	0.992303	1.007925
NO Cal Offset:	-0.220000	1.260000
NO <sub>2</sub> Cal Slope:	0.994878	1.010781
NO <sub>2</sub> Cal Offset:	0.734900	-3.206174



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-1.4	-0.4	-1.0	----	----
as found span	4918	82.0	800.3	800.3	0.0	808.3	806.6	1.8	0.9901	0.9922
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-1.3	-0.2	-1.0	----	----
high point	4918	82.0	800.3	800.3	0.0	808.4	808.1	-0.1	0.9900	0.9904
second point	4959	41.0	400.2	400.2	0.0	403.4	402.6	0.8	0.9920	0.9939
third point	4980	20.5	200.1	200.1	0.0	207.0	206.2	0.9	0.9666	0.9703
as left zero	5000	0.0	0.0	0.0	0.0	-1.3	-0.3	-1.0	----	----
as left span	4918	82.0	800.3	394.3	406.0	802.6	404.5	397.8	0.9972	0.9748
Average Correction Factor									0.9828	0.9849

Corrected As found	NO <sub>x</sub> = 809.7 ppb	NO = 807.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 1.8%
Previous Response	NO <sub>x</sub> = 795.0 ppb	NO = 793.9 ppb		*Percent Change	NO = 1.6%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	821.9	415.9	406.0	409.9	0.9905	101.0%
2nd GPT point (200 ppb O3)	821.9	612.9	209.0	202.4	1.0326	96.8%
3rd GPT point (100 ppb O3)	821.9	720.4	101.5	100.1	1.0140	98.6%
Average Correction Factor					1.0124	98.8%

Notes:

sampled inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Matthew C



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 7, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	9:04	End Time (MST):	13:32
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-1.3	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ $0.90 - 1.10$ $\pm 20$
800.3	808.4	0.9900		
400.2	403.4	0.9920		
200.1	207.0	0.9666		





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 7, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	9:04	End Time (MST):	13:32
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
800.3	808.1	0.9904			
400.2	402.6	0.9939			
200.1	206.2	0.9703			
			Slope	1.007925	0.90 - 1.10
			Intercept	1.260000	+/-20

NO Calibration Curve





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 7, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	9:04	End Time (MST):	13:32
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-1.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
406.0	409.9	0.9905		
209.0	202.4	1.0326		
101.5	100.1	1.0140		






NO<sub>x</sub> Calibration Plot

Date: January 10, 2024

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Fort Chipewyan      Station number: AMS08  
 Calibration Date: January 8, 2024      Last Cal Date: December 7, 2023  
 Start time (MST): 10:21      End time (MST): 12:53  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 3252  
 ZAG Make/Model: Teledyne API T701      Serial Number: 135

### Analyzer Information

Analyzer make: Teledyne API T400      Analyzer serial #: 3872  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.969000	0.976343	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	0.700000	0.340000	Coeff or Slope:	1.036	1.036

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	5000	913.0	400.0	393.7	1.016
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.3	----
high point	5000	914.7	400.0	390.6	1.024
second point	5000	786.4	200.0	196.4	1.018
third point	5000	701.3	100.0	97.5	1.026
as left zero	5000	0.0	0.0	0.7	----
as left span	5000	963.3	400.0	393.3	1.017
Average Correction Factor					1.023

Baseline Corr As found:	393.4	Previous response	388.3	*% change	1.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed out inlet filter after as found. No adjustments

Calibration Performed By: Matthew C



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 8, 2024	Previous Calibration:	December 7, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	10:21	End Time (MST):	12:53
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999989	≥0.995
400.0	390.6	1.0241			
200.0	196.4	1.0183	Slope	0.976343	0.90 - 1.10
100.0	97.5	1.0256			
			Intercept	0.340000	+/- 5



O<sub>3</sub> Calibration Plot

Date: January 8, 2024

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Fort Chipewyan Station number: AMS 08  
 Calibration Date: January 31, 2024 Last Cal Date: December 15, 2023  
 Start time (MST): 18:40 End time (MST): 19:10

Analyzer Make: Teledyne API T640 S/N: 319  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388747  
 Temp/RH standard: Alicat FP-25BT S/N: 388747

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	1	-0.6	1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	730.1	731.1	730.1	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.08	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39	----	39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	6.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: **11.3** Expiry Date: 04-2019  
 Lot No.: 100128-050-004

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: December 15, 2023  
 Date Disposable Filter Changed: December 15, 2023

Post- maintenance Zero Verification: PM w/ HEPA: \_\_\_\_\_ <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: July 25, 2023  
 Date RH/T Sensor Cleaned: July 25, 2023

Notes: Leak check passed. Flow, temperature, and pressure checked. No adjustments needed.

Calibration by: Devin Russell



# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name: Fort Chipewyan Station number: AMS08  
 Calibration Date: January 18, 2024 Last Cal Date: December 26, 2023  
 Start time (MST): 13:02 End time (MST): 15:24  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 3,030 ppm Cal Gas Exp Date: December 1, 2028  
 Cal Gas Cylinder #: ALM014846  
 Removed Cal Gas Conc: 3,030 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 3252  
 ZAG Make/Model: API T701H Serial Number: 135

### Analyzer Information

Analyzer make: API T300 Analyzer serial #: 3505  
 Analyzer Range: 0 - 50 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994764	0.989053	Backgd or Offset:	-0.014	-0.014
Calibration intercept:	0.168931	0.314884	Coeff or Slope:	0.998	0.998

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.19	----
as found span	4933	66.7	40.4	40.3	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4934	66.7	40.4	40.2	1.005
second point	4967	33.3	20.2	20.4	0.992
third point	4983	16.7	10.1	10.4	0.969
as left zero	5000	0.0	0.0	0.1	----
as left span	2960	40.0	40.4	40.4	1.000
Average Correction Factor					0.989

Baseline Corr As found: 40.07 Prev response: 40.38 \*% change: -0.8%  
 Baseline Corr 2nd AF pt: NA AF Slope: AF Intercept:  
 Baseline Corr 3rd AF pt: NA AF Correlation:

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. no adjustments made.

Calibration Performed By: Morgan Voyageur



# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 26, 2023
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	13:02	End Time (MST):	15:24
Analyzer make:	API T300	Analyzer serial #:	3505

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999955	≥0.995
40.4	40.2	1.0048			
20.2	20.4	0.9916	Slope	0.989053	0.90 - 1.10
10.1	10.4	0.9694			
			Intercept	0.314884	+/-1.5



CO Calibration Plot

Date: January 18, 2024

Location: Fort Chipewyan







# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	January 18, 2024	Last Cal Date:	December 26, 2023
Start time (MST):	8:55	End time (MST):	12:57
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
N2 Gen Make/Model:	NG 5000		Serial Number:	135

### Analyzer Information

Analyzer make: Teledyne API T360	Analyzer serial #: 289
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001165	0.994632	Backgd or Offset:	-0.063	-0.063
Calibration intercept:	-2.460000	-6.520000	Coeff or Slope:	1.087	1.085

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-0.8	----
as found span	2920	80.0	1605.9	1623.7	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	-0.1	----
high point	2920	80.0	1605.9	1599.9	1.004
second point	2960	40.0	802.9	771.0	1.041
third point	2980	20.0	401.5	398.3	1.008
as left zero	3000	0.0	0.0	1.9	----
as left span	2960	40.0	802.9	770.5	1.042
Average Correction Factor					1.018

Baseline Corr As found:	1624.50	Prev response:	1605.28	*% change:	1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as found. Made adjustments to span.

Calibration Performed By: Morgan Voyageur



# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date	January 18, 2024	Previous Calibration	December 26, 2023
Station Name	Fort Chipewyan	Station Number	AMS08
Start Time (MST)	8:55	End Time (MST)	12:57
Analyzer make	Teledyne API T360	Analyzer serial #	289

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999569	<b>≥0.995</b>
1605.9	1599.9	1.0037	Slope	0.994632	<b>0.90 - 1.10</b>
802.9	771.0	1.0414	Intercept	-6.520000	<b>+/-20</b>
401.5	398.3	1.0080			



CO<sub>2</sub> Calibration Plot

Date: January 18, 2024

Location: Fort Chipewyan





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS09 BARGE LANDING JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Barge Landing Station number: AMS09  
Calibration Date: January 4, 2024 Last Cal Date: December 8, 2023  
Start time (MST): 10:13 End time (MST): 13:36  
Reason: Routine

### Calibration Standards

Cal Gas Concentration: 49.96 ppm Cal Gas Exp Date: January 5, 2025  
Cal Gas Cylinder #: CC151285  
Removed Cal Gas Conc: 49.96 ppm Rem Gas Exp Date: NA  
Removed Gas Cyl #: NA Diff between cyl:  
Calibrator Make/Model: API T700 Serial Number: 3812  
ZAG Make/Model: API T701 Serial Number: 4888

### Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 1118148498  
Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995778	0.997105	Backgd or Offset: 10.2	10.2
Calibration intercept:	0.410488	0.270160	Coeff or Slope: 0.962	0.955

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4919	80.2	801.5	808.3	0.992
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4919	80.2	801.5	799.1	1.003
second point	4959	40.1	400.8	400.8	1.000
third point	4980	20.0	199.8	199.0	1.004
as left zero	5000	0.0	0.0	0.2	----
as left span	4919	80.2	801.5	803.1	0.998
Average Correction Factor:					1.002

Baseline Corr As found: 808.20 Previous response 798.51 \*% change 1.2%  
Baseline Corr 2nd AF pt: NA AF Slope: AF Intercept:  
Baseline Corr 3rd AF pt: NA AF Correlation:

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:13	End Time (MST):	13:36
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999996	≥0.995
801.5	799.1	1.0030			
400.8	400.8	0.9999	Slope	0.997105	0.90 - 1.10
199.8	199.0	1.0042			
			Intercept	0.270160	+/-30



SO2 Calibration Plot

Date: January 4, 2024

Location: Barge Landing





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Barge Landing      Station number: AMS09  
 Calibration Date: January 9, 2024      Last Cal Date: December 4, 2023  
 Start time (MST): 9:57      End time (MST): 14:11  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.171 ppm      Cal Gas Exp Date: August 22, 2026  
 Cal Gas Cylinder #: CC511415  
 Removed Cal Gas Conc: 5.171 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3812  
 ZAG Make/Model: API T701      Serial Number: 4888

### Analyzer Information

Analyzer make: Thermo 43i-TLE      Analyzer serial #: 1331259320  
 Converter make: CDN-101      Converter serial #: 519  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006260	1.008544	Backgd or Offset:	2.88	2.88
Calibration intercept:	0.099474	0.119536	Coeff or Slope:	1.176	1.176

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4923	77.4	80.0	80.1	0.998
as found 2nd point	4961	38.7	40.0	40.3	0.991
as found 3rd point	4981	19.3	20.0	20.0	0.993
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4923	77.4	80.0	80.8	0.991
second point	4961	38.7	40.0	40.6	0.986
third point	4981	19.3	20.0	20.2	0.988
as left zero	5000	0.0	0.0	0.1	----
as left span	4923	77.4	80.0	81.3	0.985
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:	28-Feb-23			Ave Corr Factor	0.988
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 80.2      Prev response: 80.64      \*% change: -0.6%  
 Baseline Corr 2nd AF pt: 40.4      AF Slope: 1.001977      AF Intercept: -0.000569  
 Baseline Corr 3rd AF pt: 20.1      AF Correlation: 0.999984

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SOx scrubber check done after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala





# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 9, 2024	Previous Calibration:	December 4, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:57	End Time (MST):	14:11
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999995	
80.0	80.8	0.9906			≥0.995
40.0	40.6	0.9859	Slope	1.008544	
20.0	20.2	0.9881			0.90 - 1.10
			Intercept	0.119536	+/-3



TRS Calibration Plot

Date: January 9, 2024

Location: Barge Landing





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	January 4, 2024	Last Cal Date:	December 8, 2023
Start time (MST):	10:13	End time (MST):	13:36
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	497.6 ppm	CH <sub>4</sub> Equiv Conc.	1067.1 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	497.6 ppm	CH <sub>4</sub> Equiv Conc.	1067.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.53E-04	2.58E-04	NMHC SP Ratio:	4.34E-05
CH <sub>4</sub> Retention time:	15.00	15.00	NMHC Peak Area:	210777
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				209298
				OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	16.97	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.14	0.999
second point	4960	40.1	8.56	8.57	0.999
third point	4980	20.0	4.27	4.29	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.17	0.997
				Response Correction Factor	0.998

Baseline Corr AF:	16.97	Prev response	17.11	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	9.14	9.08	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.13	1.000
second point	4960	40.1	4.57	4.56	1.001
third point	4980	20.0	2.28	2.28	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	9.14	9.15	0.998
Average Correction Factor					1.000
Baseline Corr AF:	9.08	Prev response	9.18	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	7.89	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	8.00	0.998
second point	4960	40.1	3.99	4.01	0.996
third point	4980	20.0	1.99	2.01	0.992
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	8.03	0.995
Average Correction Factor					0.995
Baseline Corr AF:	7.89	Prev response	7.93	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999422	1.000804
THC Cal Offset:	0.000050	0.005458
CH <sub>4</sub> Cal Slope:	0.993236	1.001867
CH <sub>4</sub> Cal Offset:	0.001649	0.006270
NMHC Cal Slope:	1.004440	0.999587
NMHC Cal Offset:	-0.000799	-0.000411

Notes: Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:13	End Time (MST):	13:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	$\geq 0.995$			
17.12	17.14	0.9990						
8.56	8.57	0.9986				Slope	1.000804	0.90 - 1.10
4.27	4.29	0.9961						
			Intercept	0.005458	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:13	End Time (MST):	13:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999997	≥0.995
7.98	8.00	0.9977			
3.99	4.01	0.9959			
1.99	2.01	0.9917			
			Slope	1.001867	0.90 - 1.10
			Intercept	0.006270	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 8, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:13	End Time (MST):	13:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	$\geq 0.995$
9.14	9.13	1.0003			
4.57	4.56	1.0014			
2.28	2.28	0.9996			
			Slope	0.999587	0.90 - 1.10
			Intercept	-0.000411	+/-0.5



NMHC Calibration Plot

Date: January 4, 2024

Location: Barge Landing







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Barge Landing  
Calibration Date: January 18, 2024  
Start time (MST): 10:31  
Reason: Routine  
Station number: AMS09  
Last Cal Date: December 5, 2023  
End time (MST): 15:28

### Calibration Standards

NO Gas Cylinder #: T2Y1KDH  
NOX Cal Gas Conc: 47.38 ppm  
Removed Cylinder #: DT0036634  
Removed Gas NOX Conc: 50.00 ppm  
NOX gas Diff: 0.7%  
Calibrator Model: API T700  
ZAG make/model: API T701  
Cal Gas Expiry Date: November 17, 2026  
NO Cal Gas Conc: 46.94 ppm  
Removed Gas Exp Date: January 28, 2024  
Removed Gas NO Conc: 49.70 ppm  
NO gas Diff: 1.4%  
Serial Number: 3812  
Serial Number: 4888

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1426262593

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.151	1.092	NO bkgnd or offset:	10.5	10.0
NOX coeff or slope:	0.992	0.998	NOX bkgnd or offset:	10.8	10.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	173.4	180.1

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.999604	0.999176
NO <sub>x</sub> Cal Offset:	0.568998	0.418407
NO Cal Slope:	1.000613	0.999039
NO Cal Offset:	-0.472474	-0.303738
NO <sub>2</sub> Cal Slope:	1.001041	1.005229
NO <sub>2</sub> Cal Offset:	-0.216477	0.853830



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
as found span	4919	80.5	805.1	800.3	4.8	841.6	833.9	7.7	0.957	0.960
as found 2nd										
as found 3rd										
new cyl resp	4915	85.3	808.3	800.7	7.5	851.4	847.2	4.2	0.949	0.945
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	0.1	-0.2	----	----
high point	4915	85.3	808.3	800.7	7.5	807.6	799.8	7.7	1.001	1.001
second point	4957	42.6	403.7	400.0	3.7	404.4	399.3	5.1	0.998	1.002
third point	4979	21.3	201.8	200.0	1.9	202.5	198.9	3.6	0.997	1.005
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
as left span	4915	85.3	808.3	455.5	352.7	804.8	451.5	353.3	1.004	1.009
Average Correction Factor									0.999	1.003

Corrected As found	NO <sub>x</sub> = 842.0 ppb	NO = 834.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 4.4%
Previous Response	NO <sub>x</sub> = 805.3 ppb	NO = 800.3 ppb		*Percent Change	NO = 4.0%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.4	449.2	352.7	355.0	0.994	100.7%
2nd GPT point (200 ppb O3)	794.4	662.4	139.5	141.6	0.985	101.5%
3rd GPT point (100 ppb O3)	794.4	726.7	75.2	77.4	0.972	102.9%
Average Correction Factor					0.983	101.7%

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 5, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:31	End Time (MST):	15:28
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
808.3	807.6	1.0008			
403.7	404.4	0.9983			
201.8	202.5	0.9967			
			Slope	0.999176	0.90 - 1.10
			Intercept	0.418407	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 5, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:31	End Time (MST):	15:28
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.7	799.8	1.0012			
400.0	399.3	1.0017			
200.0	198.9	1.0053			
			Slope	0.999039	0.90 - 1.10
			Intercept	-0.303738	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 5, 2023
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:31	End Time (MST):	15:28
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999965	≥0.995
352.7	355.0	0.9935			
139.5	141.6	0.9852			
75.2	77.4	0.9717			
			Slope	1.005229	0.90 - 1.10
			Intercept	0.853830	+/-20



NO<sub>x</sub> Calibration Plot

Date: January 18, 2024

Location: Barge Landing





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Barge Landing Station number: AMS 09  
 Calibration Date: January 22, 2024 Last Cal Date: December 8, 2023  
 Start time (MST): 10:29 End time (MST): 11:46

Analyzer Make: API T640 S/N: 844  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388746  
 Temp/RH standard: Alicat FP-25 S/N: 388746

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-22.20	-22.80	-22.20	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	731.20	738.38	731.20	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.97	4.90	4.97	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 22, 2024</u>	Last Cal Date: <u>December 8, 2023</u>			
	PM w/o HEPA: <u>6.2</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.0	11.0	11.0	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>13.9</u>	w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:		<u>January 22, 2024</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>January 22, 2024</u>			

### Annual Maintenance

Date Sample Tube Cleaned: August 23, 2023  
 Date RH/T Sensor Cleaned: August 23, 2023

Notes: Inlet head looks good. After maintenance PMT peak test was within limits. No adjustments made.  
 Leak check passed.

Calibration by: Sean Bala



# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Barge Landing	Station Number:	AMS 09
Calibration Date:	December 1, 2023	Prev Cal Date:	N/A
Start Time (MST):	12:00	End Time (MST):	14:10
Tower Height (m):	10.0	Reason:	Install

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	B4128
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.6	0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999999	<i>≥0.9995</i>
Calculated slope		0.998550	<i>0.90 - 1.10</i>
Calculated intercept		0.034415	<i>+/- 2</i>

### Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	D14057
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:29	Calc Declination*:	13.72 Degrees
Deadband calc:	1.0 degrees ( <i>Limit 4 deg</i> )		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.4	---
90	87.4	-0.7%
180	179.1	-0.3%
270	270.0	0.0%
357	356.4	-0.2%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999967	<i>≥0.9995</i>
Calculated slope		0.999287	<i>0.90 - 1.10</i>
Calculated intercept		0.863397	<i>+/- 4</i>

Notes: Installing new 10m WS/WD sensors.

Calibration Performed By: Max Farrell & Sean Bala





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS11  
LOWER CAMP**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	January 3, 2024	Last Cal Date:	December 19, 2023
Start time (MST):	10:17	End time (MST):	13:33
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.25	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC2216			
Removed Cal Gas Conc:	49.25	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3807
ZAG Make/Model:	Teledyne API T701		Serial Number:	196

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	100841398
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998200	1.001184	Backgd or Offset:	14.6	14.6
Calibration intercept:	0.350846	-0.069544	Coeff or Slope:	1.034	1.034

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.6	----
as found span	4919	81.3	800.8	799.9	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	1.0	----
high point	4919	81.3	800.8	802.2	0.998
second point	4959	40.7	400.9	400.7	1.001
third point	4980	20.3	199.9	199.1	1.004
as left zero	5000	0.0	0.0	0.9	----
as left span	4919	81.3	800.8	800.2	1.001
Average Correction Factor					1.001

Baseline Corr As found:	799.30	Previous response	799.67	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 3, 2024	Previous Calibration:	December 19, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:17	End Time (MST):	13:33
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.0	----	Correlation Coefficient	0.999992	≥0.995
800.8	802.2	0.9982			
400.9	400.7	1.0005	Slope	1.001184	0.90 - 1.10
199.9	199.1	1.0042			
			Intercept	-0.069544	+/-30



SO2 Calibration Plot

Date: January 3, 2024

Location: Lower Camp





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Lower Camp      Station number: AMS11  
 Calibration Date: January 22, 2024      Last Cal Date: December 20, 2023  
 Start time (MST): 11:12      End time (MST): 17:17  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.429 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC501097  
 Removed Cal Gas Conc: 5.429 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3807  
 ZAG Make/Model: API T701H      Serial Number: 196

### Analyzer Information

Analyzer make: Thermo 450iQ      Analyzer serial #: CM20080003  
 Converter make: NA      Converter serial #: NA  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009917	0.992455	Backgd or Offset: 14.8	15.9
Calibration intercept:	0.234284	-0.204215	Coeff or Slope: 1.001	1.015

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4926	73.6	79.9	79.4	1.012
as found 2nd point	4963	36.8	40.0	39.9	1.012
as found 3rd point	4982	18.6	20.2	20.3	1.015
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4926	73.6	79.9	79.1	1.010
second point	4963	36.8	40.0	39.5	1.012
third point	4982	18.6	20.2	19.8	1.020
as left zero	5000	0.0	0.0	-0.4	----
as left span	4926	73.6	79.9	79.7	1.003
SO2 Scrubber Check	4919	81.1	811.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	1.014
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 79.0      Prev response: 80.95      \*% change: -2.5%  
 Baseline Corr 2nd AF pt: 39.5      AF Slope: 0.988733      AF Intercept: 0.376111  
 Baseline Corr 3rd AF pt: 19.9      AF Correlation: 0.999999

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted both zero and span. Ran scrubber check after cal zero. Sox scrubber check passed after hydrating the scrubber beads.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 22, 2024	Previous Calibration:	December 20, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	11:12	End Time (MST):	17:17
Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999999	≥0.995
79.9	79.1	1.0104			
40.0	39.5	1.0116	Slope	0.992455	0.90 - 1.10
20.2	19.8	1.0199			
			Intercept	-0.204215	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 22, 2024

Location: Lower Camp





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	January 3, 2024	Last Cal Date:	December 19, 2023
Start time (MST):	10:17	End time (MST):	13:33
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH <sub>4</sub> Cal Gas Conc.	502.0 ppm	CH <sub>4</sub> Equiv Conc.	1067.1 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	502.0 ppm	CH <sub>4</sub> Equiv Conc.	1067.1 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
ZAG make/model:	API T701	Serial Number:	196

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.00E-04	3.00E-04	NMHC SP Ratio:	5.77E-05
CH <sub>4</sub> Retention time:	14.2	14.2	NMHC Peak Area:	158880
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	17.35	17.29	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	17.35	17.33	1.001
second point	4959	40.7	8.69	8.64	1.005
third point	4980	20.3	4.33	4.32	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	17.35	17.41	0.997
Average Correction Factor					1.003

Baseline Corr AF:	17.28	Prev response	17.36	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	9.19	9.21	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	9.19	9.22	0.997
second point	4959	40.7	4.60	4.60	1.001
third point	4980	20.3	2.29	2.30	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	9.19	9.22	0.997
Average Correction Factor					0.998
Baseline Corr AF:	9.21	Prev response	9.21	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.3	8.16	8.07	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.3	8.16	8.11	1.006
second point	4959	40.7	4.09	4.05	1.010
third point	4980	20.3	2.04	2.02	1.010
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.3	8.16	8.19	0.996
Average Correction Factor					1.009
Baseline Corr AF:	8.07	Prev response	8.15	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001730	0.998445
THC Cal Offset:	-0.022785	-0.009595
CH <sub>4</sub> Cal Slope:	0.998587	0.993827
CH <sub>4</sub> Cal Offset:	-0.005091	-0.006092
NMHC Cal Slope:	1.004174	1.002671
NMHC Cal Offset:	-0.017294	-0.003503

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 3, 2024	Previous Calibration:	December 19, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:17	End Time (MST):	13:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	$\geq 0.995$			
17.35	17.33	1.0015						
8.69	8.64	1.0054				Slope	0.998445	0.90 - 1.10
4.33	4.32	1.0031						
			Intercept	-0.009595	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 3, 2024	Previous Calibration:	December 19, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:17	End Time (MST):	13:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	<b>≥0.995</b>
8.16	8.11	1.0063			
4.09	4.05	1.0103			
2.04	2.02	1.0099			
			Slope	0.993827	<b>0.90 - 1.10</b>
			Intercept	-0.006092	<b>+/-0.5</b>





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 3, 2024	Previous Calibration:	December 19, 2023
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:17	End Time (MST):	13:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	$\geq 0.995$			
9.19	9.22	0.9971						
4.60	4.60	1.0010				Slope	1.002671	0.90 - 1.10
2.29	2.30	0.9971						
			Intercept	-0.003503	$\pm 0.5$			



NMHC Calibration Plot

Date: January 3, 2024

Location: Lower Camp





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS13 FORT MCKAY SOUTH**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	January 15, 2024	Last Cal Date:	December 18, 2023
Start time (MST):	10:26	End time (MST):	13:55
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.55	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC260812			
Removed Cal Gas Conc:	50.55	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2448
ZAG Make/Model:	API 701		Serial Number:	1117

### Analyzer Information

Analyzer make:	API T100	Analyzer serial #:	599
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000755	1.003443	Backgd or Offset:	91.1	87.3
Calibration intercept:	-2.857974	-2.798230	Coeff or Slope:	0.709	0.713

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-1.4	----
as found span	4921	79.1	799.7	795.6	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	79.1	799.7	801.0	0.998
second point	4961	39.5	399.3	396.7	1.007
third point	4980	19.8	200.2	195.0	1.027
as left zero	5000	0.0	0.0	0.1	----
as left span	4921	79.1	799.7	799.9	1.000
<b>Average Correction Factor</b>					<b>1.011</b>

Baseline Corr As found:	797.00	Previous response	797.43	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 18, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:26	End Time (MST):	13:55
Analyzer make:	API T100	Analyzer serial #:	599

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999940	
799.7	801.0	0.9984			≥0.995
399.3	396.7	1.0066	Slope	1.003443	
200.2	195.0	1.0266			0.90 - 1.10
			Intercept	-2.798230	+/-30






SO2 Calibration Plot

Date: January 15, 2024

Location: Fort McKay South





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	January 17, 2024	Last Cal Date:	December 5, 2023
Start time (MST):	10:24	End time (MST):	15:16
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.34	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC500241			
Removed Cal Gas Conc:	5.34	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2448
ZAG Make/Model:	Teledyne API 701		Serial Number:	1117

### Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017
Converter make:	CDN-101	Converter serial #:	521
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997103	0.997527	Backgd or Offset:	3.77	4.15
Calibration intercept:	-0.142201	-0.182182	Coeff or Slope:	1.130	1.157

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4925	75.5	80.6	78.6	1.027
as found 2nd point	4962	37.7	40.3	39.1	1.032
as found 3rd point	4981	18.9	20.2	19.4	1.046
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4925	75.5	80.6	80.2	1.005
second point	4962	37.7	40.3	40.1	1.004
third point	4981	18.9	20.2	19.9	1.014
as left zero	5000	0.0	0.0	0.4	----
as left span	4925	75.5	80.6	79.6	1.013
SO2 Scrubber Check	4921	79.1	791.0	0.2	----
Date of last scrubber change:	20-Mar-20		Ave Corr Factor		1.008
Date of last converter efficiency test:	NA		efficiency		

Baseline Corr As found:	78.5	Prev response:	80.25	*% change:	-2.2%
Baseline Corr 2nd AF pt:	39.0	AF Slope:	0.974845	AF Intercept:	-0.082177
Baseline Corr 3rd AF pt:	19.3	AF Correlation:	0.999975		

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. Adjusted zero and span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	December 5, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:24	End Time (MST):	15:16
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999995	
80.6	80.2	1.0053			≥0.995
40.3	40.1	1.0041	Slope	0.997527	
20.2	19.9	1.0144			0.90 - 1.10
			Intercept	-0.182182	+/-3



TRS Calibration Plot

Date: January 17, 2024

Location: Fort McKay South





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	January 15, 2024	Last Cal Date:	December 18, 2023
Start time (MST):	10:26	End time (MST):	13:55
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH4 Cal Gas Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
C3H8 Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
Removed C3H8 Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API701	Serial Number:	1117

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.80E-04	2.94E-04	NMHC SP Ratio:	4.70E-05	4.74E-05
CH4 Retention time:	14.80	15.00	NMHC Peak Area:	193260	191512
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4921	79.1	17.05	16.37	1.041
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.05	1.000
second point	4961	39.5	8.51	8.45	1.008
third point	4980	19.8	4.27	4.12	1.035
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	17.05	16.89	1.009

				Average Correction Factor	1.014
Baseline Corr AF:	16.37	Prev response	17.04	*% change	-4.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	9.08	8.99	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	9.08	9.08	1.000
second point	4961	39.5	4.53	4.48	1.013
third point	4980	19.8	2.27	2.18	1.042
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	9.08	8.88	1.022
Average Correction Factor					1.018
Baseline Corr AF:	8.99	Prev response	9.08	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4921	79.1	7.97	7.39	1.078
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.97	0.999
second point	4961	39.5	3.98	3.97	1.002
third point	4980	19.8	1.99	1.94	1.027
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	7.97	8.01	0.995
Average Correction Factor					1.009
Baseline Corr AF:	7.38	Prev response	7.96	*% change	-7.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.003916	1.002747
THC Cal Offset:	-0.069164	-0.071954
CH <sub>4</sub> Cal Slope:	1.003456	1.003122
CH <sub>4</sub> Cal Offset:	-0.030586	-0.024172
NMHC Cal Slope:	1.004193	1.002658
NMHC Cal Offset:	-0.038578	-0.047981

Notes: Changed inlet filter after as founds. Adjusted span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 18, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:26	End Time (MST):	13:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999916	$\geq 0.995$			
17.05	17.05	0.9997						
8.51	8.45	1.0079				Slope	1.002747	0.90 - 1.10
4.27	4.12	1.0350						
			Intercept	-0.071954	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 18, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:26	End Time (MST):	13:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999946	≥0.995			
7.97	7.97	0.9991						
3.98	3.97	1.0018				Slope	1.003122	0.90 - 1.10
1.99	1.94	1.0275						
			Intercept	-0.024172	+/-0.5			







# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 18, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:26	End Time (MST):	13:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999873	≥0.995
9.08	9.08	1.0000			
4.53	4.48	1.0131	Slope	1.002658	0.90 - 1.10
2.27	2.18	1.0416			
			Intercept	-0.047981	+/-0.5



NMHC Calibration Plot

Date: January 15, 2024

Location: Fort McKay South





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Fort McKay South  
Calibration Date: January 25, 2024  
Start time (MST): 9:37  
Reason: Routine  
Station number: AMS 13  
Last Cal Date: December 19, 2023  
End time (MST): 14:42

### Calibration Standards

NO Gas Cylinder #: T2UP1RP  
NOX Cal Gas Conc: 48.25 ppm  
Removed Cylinder #: T2Y1P76  
Removed Gas NOX Conc: 50.98 ppm  
NOX gas Diff: -1.4%  
Calibrator Model: API T700  
ZAG make/model: API T701  
Cal Gas Expiry Date: November 17, 2026  
NO Cal Gas Conc: 47.88 ppm  
Removed Gas Exp Date: December 11, 2023  
Removed Gas NO Conc: 49.32 ppm  
NO gas Diff: -0.1%  
Serial Number: 2448  
Serial Number: 1117

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1410661329

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.136	1.133	NO bkgnd or offset:	12	10.9
NOX coeff or slope:	0.989	1.001	NOX bkgnd or offset:	12.0	11.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.5	160.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000174	1.001306
NO <sub>x</sub> Cal Offset:	-2.251490	-2.253060
NO Cal Slope:	1.004205	1.003396
NO Cal Offset:	-3.125534	-3.151422
NO <sub>2</sub> Cal Slope:	1.001643	0.998110
NO <sub>2</sub> Cal Offset:	-0.703923	-0.700632



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-1.3	-1.3	0.0	----	----
as found span	4919	81.1	826.9	800.0	26.9	830.7	802.5	28.2	0.9954	0.9968
as found 2nd										
as found 3rd										
new cyl resp	4917	83.5	805.7	799.5	6.2	798.1	801.4	-3.4	1.0095	0.9976
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4917	83.5	805.7	799.5	6.2	805.6	800.6	4.9	1.0001	0.9986
second point	4958	41.8	403.4	400.3	3.1	400.5	397.0	3.5	1.0072	1.0083
third point	4979	20.9	201.7	200.1	1.5	197.6	194.6	3.0	1.0207	1.0285
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.4	0.1	----	----
as left span	4917	83.5	805.7	376.1	429.6	810.0	378.2	431.8	0.9947	0.9945
Average Correction Factor									1.0093	1.0118

Corrected As found	NO <sub>x</sub> = 832.0 ppb	NO = 803.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.9%
Previous Response	NO <sub>x</sub> = 824.8 ppb	NO = 800.2 ppb		*Percent Change	NO = 0.4%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:
					NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	797.2	373.8	429.6	428.5	1.0025	99.7%
2nd GPT point (200 ppb O <sub>3</sub> )	797.2	586.4	217.0	215.3	1.0078	99.2%
3rd GPT point (100 ppb O <sub>3</sub> )	797.2	691.8	111.6	110.1	1.0134	98.7%
Average Correction Factor					1.0079	99.2%

Notes: Changed inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	December 19, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:37	End Time (MST):	14:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
805.7	805.6	1.0001		
403.4	400.5	1.0072		
201.7	197.6	1.0207		
			0.999967	
			1.001306	
			-2.253060	

NO<sub>x</sub> Calibration Curve





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	December 19, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:37	End Time (MST):	14:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.5	800.6	0.9986		
400.3	397.0	1.0083		
200.1	194.6	1.0285		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	December 19, 2023
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:37	End Time (MST):	14:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
429.6	428.5	1.0025		
217.0	215.3	1.0078		
111.6	110.1	1.0134		
			0.999987	
			0.998110	
			-0.700632	



NO<sub>x</sub> Calibration Plot

Date: January 25, 2024

Location: Fort McKay South







# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Fort McKay South      Station number: AMS13  
 Calibration Date: January 5, 2024      Last Cal Date: December 13, 2023  
 Start time (MST): 10:02      End time (MST): 13:20  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 2448  
 ZAG Make/Model: Teledyne API T701      Serial Number: 1117

### Analyzer Information

Analyzer make: Teledyne API T400      Analyzer serial #: 3871  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007857	1.000914	Backgd or Offset:	3.9	3.9
Calibration intercept:	0.000000	0.240000	Coeff or Slope:	0.971	0.971

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.7	----
as found span	5000	989.8	400.0	400.6	0.999
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.4	----
high point	5000	989.8	400.0	400.2	1.000
second point	5000	849.9	200.0	201.0	0.995
third point	5000	745.1	100.0	100.8	0.992
as left zero	5000	0.0	0.0	-0.6	----
as left span	5000	989.8	400.0	401.4	0.997
Average Correction Factor					0.996

Baseline Corr As found:	401.3	Previous response	403.1	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 13, 2023
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:02	End Time (MST):	13:20
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.4	----	Correlation Coefficient	0.999987	≥0.995
400.0	400.2	0.9995			
200.0	201.0	0.9950	Slope	1.000914	0.90 - 1.10
100.0	100.8	0.9921			
			Intercept	0.240000	+/- 5



O<sub>3</sub> Calibration Plot

Date: January 5, 2024

Location: Fort McKay South





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Fort McKay South Station number: AMS 13  
 Calibration Date: January 22, 2024 Last Cal Date: December 19, 2023  
 Start time (MST): 13:42 End time (MST): 14:43

Analyzer Make: API T640 S/N: 1335  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388746  
 Temp/RH standard: Alicat FP-25 S/N: 388746

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-17.70	-18.54	-17.70	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	735.20	737.20	735.20	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.92	4.81	4.92	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 22, 2024</u>		Last Cal Date: <u>December 19, 2023</u>		
	PM w/o HEPA: <u>5.6</u>		PM w/ HEPA: <u>0.0</u>		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	8.9	10.9	10.9	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:	PM w/o HEPA: <u>5.1</u>		w/ HEPA: <u>0.0</u>		
Date Optical Chamber Cleaned:	<u>January 22, 2024</u>				<0.2 ug/m3
Disposable Filter Changed:	<u>January 22, 2024</u>				

### Annual Maintenance

Date Sample Tube Cleaned: June 29, 2023  
 Date RH/T Sensor Cleaned: June 29, 2023

PMT peak test adjusted. Leak check passed.

Notes:

Calibration by: Sean Bala



## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS14  
ANZAC**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 7, 2023
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:10	End Time (MST):	14:30
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.2	----	Correlation Coefficient	0.999926	≥0.995
799.3	800.3	0.9987			
400.1	395.9	1.0105	Slope	1.001175	0.90 - 1.10
201.1	197.2	1.0196			
			Intercept	-1.861173	+/-30



SO2 Calibration Plot

Date: January 4, 2024

Location: Anzac







# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Anzac Station number: AMS14  
 Calibration Date: January 18, 2024 Last Cal Date: December 5, 2023  
 Start time (MST): 10:50 End time (MST): 15:13  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC510379  
 Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 3060  
 ZAG Make/Model: API 701H Serial Number: 357

### Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153582  
 Converter make: CD Nova CDN-101 Converter serial #: 503  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011740	1.012453	Backgd or Offset: 2.33	2.40
Calibration intercept:	-0.265448	-0.065392	Coeff or Slope: 0.992	1.043

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4938	77.9	80.0	77.4	1.032
as found 2nd point	4973	38.9	40.0	38.3	1.041
as found 3rd point	4997	19.5	20.0	19.0	1.048
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4938	77.9	80.0	80.9	0.988
second point	4973	38.9	40.0	40.5	0.987
third point	4997	19.5	20.0	19.9	1.006
as left zero	5000	0.0	0.0	0.3	----
as left span	4938	77.9	80.0	80.4	0.994
SO2 Scrubber Check	4936	80.3	800.4	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.993
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 77.5 Prev response: 80.66 \*% change: -4.1%  
 Baseline Corr 2nd AF pt: 38.4 AF Slope: 0.969763 AF Intercept: -0.285258  
 Baseline Corr 3rd AF pt: 19.1 AF Correlation: 0.999970

\* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. Adjusted span only.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 5, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:50	End Time (MST):	15:13
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	0.999965
80.0	80.9	0.9883		
40.0	40.5	0.9866	Slope	1.012453
20.0	19.9	1.0056		
			Intercept	-0.065392
				+/-3



TRS Calibration Plot

Date: January 18, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	January 4, 2024	Last Cal Date:	December 7, 2023
Start time (MST):	11:10	End time (MST):	14:30
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.72E-04	3.89E-04	NMHC SP Ratio:	4.60E-05
CH <sub>4</sub> Retention time:	12.20	12.60	NMHC Peak Area:	198214
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	16.83	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.20	0.994
second point	4979	40.2	8.56	8.49	1.008
third point	4998	20.2	4.30	4.23	1.018
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	17.10	17.15	0.997
Average Correction Factor					1.007

Baseline Corr AF:	16.83	Prev response	17.18	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	9.20	0.990
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.28	0.982
second point	4979	40.2	4.56	4.54	1.006
third point	4998	20.2	2.29	2.24	1.023
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	9.11	9.12	0.999
Average Correction Factor					1.004
Baseline Corr AF:	9.20	Prev response	9.27	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	7.63	1.047
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	7.92	1.008
second point	4979	40.2	4.00	3.96	1.010
third point	4998	20.2	2.01	1.99	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	7.99	8.02	0.996
Average Correction Factor					1.010
Baseline Corr AF:	7.63	Prev response	7.91	*% change	-3.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.007099	1.007150
THC Cal Offset:	-0.042334	-0.063715
CH <sub>4</sub> Cal Slope:	0.992465	0.991854
CH <sub>4</sub> Cal Offset:	-0.015977	-0.003590
NMHC Cal Slope:	1.019691	1.020347
NMHC Cal Offset:	-0.026156	-0.059525

Notes: Sample inlet filter and Nitrogen cylinder were changed after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 7, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:10	End Time (MST):	14:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999927	$\geq 0.995$			
17.10	17.20	0.9941						
8.56	8.49	1.0078				Slope	1.007150	0.90 - 1.10
4.30	4.23	1.0178						
			Intercept	-0.063715	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 7, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:10	End Time (MST):	14:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	$\geq 0.995$			
7.99	7.92	1.0084						
4.00	3.96	1.0099				Slope	0.991854	0.90 - 1.10
2.01	1.99	1.0115						
			Intercept	-0.003590	$\pm 0.5$			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 7, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:10	End Time (MST):	14:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999783	$\geq 0.995$			
9.11	9.28	0.9821						
4.56	4.54	1.0058				Slope	1.020347	0.90 - 1.10
2.29	2.24	1.0235						
			Intercept	-0.059525	$\pm 0.5$			






NMHC Calibration Plot

Date: January 4, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	January 16, 2024	Last Cal Date:	January 4, 2024
Start time (MST):	11:10	End time (MST):	14:54
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.89E-04	4.20E-04	NMHC SP Ratio:	4.60E-05
CH <sub>4</sub> Retention time:	12.60	13.00	NMHC Peak Area:	198370
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	16.56	1.033
as found 2nd point	4979	40.2	8.56	8.19	1.045
as found 3rd point	4998	20.2	4.30	4.08	1.055
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.05	1.003
second point	4979	40.2	8.56	8.50	1.007
third point	4998	20.2	4.30	4.25	1.014
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	17.10	17.14	0.998
Average Correction Factor					1.008

Baseline Corr AF:	16.56	Prev response	17.16	*% change	-3.6%
Baseline Corr 2nd AF:	8.2	AF Slope:	0.969546	AF Intercept:	-0.054256
Baseline Corr 3rd AF:	4.1	AF Correlation:	0.999945	* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	9.17	0.993
as found 2nd point	4979	40.2	4.56	4.52	1.009
as found 3rd point	4998	20.2	2.29	2.24	1.022
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.08	1.004
second point	4979	40.2	4.56	4.54	1.005
third point	4998	20.2	2.29	2.27	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	9.11	9.20	0.991
Average Correction Factor					1.007
Baseline Corr AF:	9.17	Prev response	9.24	*% change	-0.7%
Baseline Corr 2nd AF:	4.5	AF Slope:	1.008095	AF Intercept:	-0.039867
Baseline Corr 3rd AF:	2.2	AF Correlation:	0.999903	* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	7.39	1.081
as found 2nd point	4979	40.2	4.00	3.67	1.089
as found 3rd point	4998	20.2	2.01	1.84	1.094
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	7.97	1.002
second point	4979	40.2	4.00	3.96	1.010
third point	4998	20.2	2.01	1.98	1.015
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	7.99	7.94	1.006
Average Correction Factor					1.009
Baseline Corr AF:	7.39	Prev response	7.92	*% change	-7.2%
Baseline Corr 2nd AF:	3.67	AF Slope:	0.925403	AF Intercept:	-0.013789
Baseline Corr 3rd AF:	1.84	AF Correlation:	0.999981	* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.007150	0.997643
THC Cal Offset:	-0.063715	-0.024242
CH <sub>4</sub> Cal Slope:	0.991854	0.998476
CH <sub>4</sub> Cal Offset:	-0.003590	-0.016013
NMHC Cal Slope:	1.020347	0.996574
NMHC Cal Offset:	-0.059525	-0.007627

Notes: Adjusted the span only.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	January 4, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:10	End Time (MST):	14:54
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999991	$\geq 0.995$			
17.10	17.05	1.0030						
8.56	8.50	1.0069				Slope	0.997643	0.90 - 1.10
4.30	4.25	1.0135						
			Intercept	-0.024242	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	January 4, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:10	End Time (MST):	14:54
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999979	≥0.995			
7.99	7.97	1.0022						
4.00	3.96	1.0096				Slope	0.998476	0.90 - 1.10
2.01	1.98	1.0151						
			Intercept	-0.016013	+/-0.5			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	January 4, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:10	End Time (MST):	14:54
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	$\geq 0.995$			
9.11	9.08	1.0041						
4.56	4.54	1.0045				Slope	0.996574	0.90 - 1.10
2.29	2.27	1.0122						
			Intercept	-0.007627	$\pm 0.5$			



NMHC Calibration Plot

Date: January 16, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	January 24, 2024	Last Cal Date:	January 16, 2024
Start time (MST):	10:14	End time (MST):	14:30
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	4.20E-04	4.72E-04	NMHC SP Ratio:	4.60E-05
CH <sub>4</sub> Retention time:	13.00	13.40	NMHC Peak Area:	198389
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	15.87	1.078
as found 2nd point	4979	40.2	8.56	7.90	1.083
as found 3rd point	4998	20.2	4.30	3.95	1.091
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.13	0.998
second point	4979	40.2	8.56	8.51	1.007
third point	4998	20.2	4.30	4.26	1.010
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	17.10	17.14	0.998
Average Correction Factor					1.005

Baseline Corr AF:	15.87	Prev response	17.04	*% change	-7.4%
Baseline Corr 2nd AF:	7.9	AF Slope:	0.928535	AF Intercept:	-0.026776
Baseline Corr 3rd AF:	3.9	AF Correlation:	0.999987	<i>* = &gt; +/-5% change initiates investigation</i>	





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	8.76	1.040
as found 2nd point	4979	40.2	4.56	4.33	1.055
as found 3rd point	4998	20.2	2.29	2.15	1.064
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.15	0.996
second point	4979	40.2	4.56	4.54	1.006
third point	4998	20.2	2.29	2.26	1.013
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	9.11	9.20	0.990
Average Correction Factor					1.005
Baseline Corr AF:	8.76	Prev response	9.07	*% change	-3.6%
Baseline Corr 2nd AF:	4.3	AF Slope:	0.962578	AF Intercept:	-0.032175
Baseline Corr 3rd AF:	2.2	AF Correlation:	0.999927	* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	7.11	1.124
as found 2nd point	4979	40.2	4.00	3.58	1.118
as found 3rd point	4998	20.2	2.01	1.79	1.122
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	7.98	1.001
second point	4979	40.2	4.00	3.97	1.008
third point	4998	20.2	2.01	2.00	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	7.99	7.94	1.006
Average Correction Factor					1.005
Baseline Corr AF:	7.11	Prev response	7.96	*% change	-12.0%
Baseline Corr 2nd AF:	3.58	AF Slope:	0.889917	AF Intercept:	0.005397
Baseline Corr 3rd AF:	1.79	AF Correlation:	0.999990	* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997643	1.002215
THC Cal Offset:	-0.024242	-0.033495
CH <sub>4</sub> Cal Slope:	0.998476	0.999209
CH <sub>4</sub> Cal Offset:	-0.016013	-0.009829
NMHC Cal Slope:	0.996574	1.004825
NMHC Cal Offset:	-0.007627	-0.024066

Notes: Adjusted the span only.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 24, 2024	Previous Calibration:	January 16, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:14	End Time (MST):	14:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999977	$\geq 0.995$			
17.10	17.13	0.9983						
8.56	8.51	1.0065				Slope	1.002215	0.90 - 1.10
4.30	4.26	1.0097						
			Intercept	-0.033495	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 24, 2024	Previous Calibration:	January 16, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:14	End Time (MST):	14:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999984	≥0.995
7.99	7.98	1.0008			
4.00	3.97	1.0081			
2.01	2.00	1.0059			
			Slope	0.999209	0.90 - 1.10
			Intercept	-0.009829	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 24, 2024	Previous Calibration:	January 16, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:14	End Time (MST):	14:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999964	$\geq 0.995$			
9.11	9.15	0.9961						
4.56	4.54	1.0056				Slope	1.004825	0.90 - 1.10
2.29	2.26	1.0131						
			Intercept	-0.024066	$\pm 0.5$			



NMHC Calibration Plot

Date: January 24, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	January 30, 2024	Last Cal Date:	January 24, 2024
Start time (MST):	9:56	End time (MST):	11:20
Reason:	Removal		

### Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.72E-04	NA	NMHC SP Ratio:	4.80E-05
CH4 Retention time:	13.40	NA	NMHC Peak Area:	188364
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	16.08	1.063
as found 2nd point	4979	40.2	8.56	8.08	1.060
as found 3rd point	4998	20.2	4.30	4.02	1.071
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

			Average Correction Factor		
Baseline Corr AF:	16.08	Prev response	17.11	*% change	-6.4%
Baseline Corr 2nd AF:	8.1	AF Slope:	0.941139	AF Intercept:	-0.005949
Baseline Corr 3rd AF:	4.0	AF Correlation:	0.999989	* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	8.80	1.036
as found 2nd point	4979	40.2	4.56	4.41	1.035
as found 3rd point	4998	20.2	2.29	2.18	1.050
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	8.80	Prev response	9.13	*% change	-3.8%
Baseline Corr 2nd AF:	4.4	AF Slope:	0.966685	AF Intercept:	-0.011819
Baseline Corr 3rd AF:	2.2	AF Correlation:	0.999984	* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	7.28	1.097
as found 2nd point	4979	40.2	4.00	3.67	1.089
as found 3rd point	4998	20.2	2.01	1.83	1.096
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	7.28	Prev response	7.97	*% change	-9.5%
Baseline Corr 2nd AF:	3.67	AF Slope:	0.911756	AF Intercept:	0.006471
Baseline Corr 3rd AF:	1.83	AF Correlation:	0.999981	* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002215	
THC Cal Offset:	-0.033495	
CH <sub>4</sub> Cal Slope:	0.999209	
CH <sub>4</sub> Cal Offset:	-0.009829	
NMHC Cal Slope:	1.004825	
NMHC Cal Offset:	-0.024066	


Notes: Removal only. Observed no changes in the analyzer's diagnostics, and there were no alarms, but internal maintenance may be required.

Calibration Performed By: Mohammed Kashif

NMHC Calibration Plot

Date: January 30, 2024

Location: Anzac







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	January 31, 2024	Last Cal Date:	NA
Start time (MST):	10:32	End time (MST):	13:28
Reason:	Install		

### Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	499.3 ppm	CH <sub>4</sub> Equiv Conc.	1068.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	12227620776
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	NA	2.53E-04	NMHC SP Ratio:	NA	5.94E-05
CH <sub>4</sub> Retention time:	NA	14.20	NMHC Peak Area:	NA	153300
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.07	1.002
second point	4979	40.2	8.56	8.42	1.017
third point	4998	20.2	4.30	4.16	1.034
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	17.10	17.03	1.005
Average Correction Factor					1.017

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.09	1.003
second point	4979	40.2	4.56	4.49	1.016
third point	4998	20.2	2.29	2.22	1.031
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	9.11	9.05	1.007
Average Correction Factor					1.017
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	7.99	1.000
second point	4979	40.2	4.00	3.93	1.017
third point	4998	20.2	2.01	1.94	1.037
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	7.99	7.97	1.002
Average Correction Factor					1.018
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		1.000080
THC Cal Offset:		-0.078002
CH <sub>4</sub> Cal Slope:		1.001814
CH <sub>4</sub> Cal Offset:		-0.041197
NMHC Cal Slope:		0.998836
NMHC Cal Offset:		-0.037406

Notes: Installation. Adjusted span only.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 31, 2024	Previous Calibration:	NA
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:32	End Time (MST):	13:28
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620776

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999899	$\geq 0.995$			
17.10	17.07	1.0017						
8.56	8.42	1.0166				Slope	1.000080	0.90 - 1.10
4.30	4.16	1.0337						
			Intercept	-0.078002	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 31, 2024	Previous Calibration:	NA
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:32	End Time (MST):	13:28
Analyser make:	Thermo 55i	Analyser serial #:	12227620776

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999874	≥0.995
7.99	7.99	1.0003			
4.00	3.93	1.0165	Slope	1.001814	0.90 - 1.10
2.01	1.94	1.0371			
			Intercept	-0.041197	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 31, 2024	Previous Calibration:	NA
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:32	End Time (MST):	13:28
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620776

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999918	$\geq 0.995$
9.11	9.09	1.0028			
4.56	4.49	1.0164			
2.29	2.22	1.0313			
			Slope	0.998836	0.90 - 1.10
			Intercept	-0.037406	+/-0.5



NMHC Calibration Plot

Date: January 31, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Anzac  
Calibration Date: January 5, 2024  
Start time (MST): 10:50  
Reason: Cylinder Change  
Station number: AMS14  
Last Cal Date: December 4, 2023  
End time (MST): 16:45

### Calibration Standards

NO Gas Cylinder #: DT0037092  
NOX Cal Gas Conc: 60.7 ppm  
Removed Cylinder #: T2Y1P8D  
Removed Gas NOX Conc: 50.9 ppm  
NOX gas Diff: -1.7%  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API 701H  
Cal Gas Expiry Date: May 16, 2031  
NO Cal Gas Conc: 60.4 ppm  
Removed Gas Exp Date: December 11, 2023  
Removed Gas NO Conc: 50.1 ppm  
NO gas Diff: -0.7%  
Serial Number: 3060  
Serial Number: 357

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1426262592

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.411	1.411	NO bkgnd or offset:	3.8	3.8
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.8	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	158.2	159.1

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.001131	0.995638
NO <sub>x</sub> Cal Offset:	-0.568562	-0.669728
NO Cal Slope:	1.001791	1.002181
NO Cal Offset:	-2.406556	-2.129361
NO <sub>2</sub> Cal Slope:	1.001950	0.992183
NO <sub>2</sub> Cal Offset:	0.975434	-0.766639



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.2	----	----
as found span	4936	80.2	814.1	800.2	13.9	825.8	807.6	18.1	0.9859	0.9908
as found 2nd										
as found 3rd										
new cyl resp	4934	66.3	804.8	800.9	4.0	802.8	802.6	0.3	1.0025	0.9978
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.1	----	----
high point	4934	66.3	804.8	800.9	4.0	800.8	801.2	-0.5	1.0050	0.9996
second point	4985	33.2	401.6	399.6	2.0	399.5	398.2	1.3	1.0052	1.0035
third point	5004	16.7	201.9	200.9	1.0	199.1	196.6	2.5	1.0141	1.0219
as left zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.1	----	----
as left span	4934	66.3	804.8	416.3	388.6	804.8	414.2	390.7	1.0000	1.0050
Average Correction Factor									1.0081	1.0083

Corrected As found	NO <sub>x</sub> =	825.9 ppb	NO =	807.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> =	1.4%
Previous Response	NO <sub>x</sub> =	814.5 ppb	NO =	799.2 ppb		*Percent Change	NO =	1.1%
Baseline Corr 2nd pt	NO <sub>x</sub> =	NA ppb	NO =	NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> =	NA ppb	NO =	NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
					As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO <sub>2</sub> )						
as found GPT point (200 ppb NO <sub>2</sub> )						
as found GPT point (100 ppb NO <sub>2</sub> )						
1st GPT point (400 ppb O <sub>3</sub> )	800.0	415.4	388.6	385.4	1.0082	99.2%
2nd GPT point (200 ppb O <sub>3</sub> )	800.0	605.2	198.8	195.5	1.0168	98.4%
3rd GPT point (100 ppb O <sub>3</sub> )	800.0	704.8	99.2	97.1	1.0214	97.9%
Average Correction Factor					1.0155	98.5%

Notes:

Sample inlet filter and cal gas cylinder changed after as founds. No adjustments made.

Calibration Performed By:

Mohammed Kashif





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 4, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:50	End Time (MST):	16:45
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
804.8	800.8	1.0050		
401.6	399.5	1.0052		
201.9	199.1	1.0141		

**NO<sub>x</sub> Calibration Curve**





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 4, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:50	End Time (MST):	16:45
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999967	≥0.995
800.9	801.2	0.9996			
399.6	398.2	1.0035	Slope	1.002181	0.90 - 1.10
200.9	196.6	1.0219			
			Intercept	-2.129361	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 4, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:50	End Time (MST):	16:45
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
388.6	385.4	1.0082			
198.8	195.5	1.0168			
99.2	97.1	1.0214			
			Slope	0.992183	0.90 - 1.10
			Intercept	-0.766639	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: January 5, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	January 8, 2024	Last Cal Date:	December 6, 2023
Start time (MST):	10:48	End time (MST):	13:46
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	API T700	Serial Number:	3060
ZAG Make/Model:	API 701H	Serial Number:	357

### Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595
Analyzer Range:	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002086	1.001343	Backgd or Offset:	1.4	1.4
Calibration intercept:	-0.040000	-0.260000	Coeff or Slope:	1.620	1.620

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	5000	918.8	400.0	401.6	0.996
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.2	----
high point	5000	918.8	400.0	400.4	0.999
second point	5000	803.8	200.0	200.1	1.000
third point	5000	709.8	100.0	99.2	1.008
as left zero	5000	0.0	0.0	-0.2	----
as left span	5000	918.8	400.0	406.0	0.985
Average Correction Factor					1.002

Baseline Corr As found:	401.3	Previous response	400.8	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments required.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 8, 2024	Previous Calibration:	December 6, 2023
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:48	End Time (MST):	13:46
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999992	≥0.995
400.0	400.4	0.9990			
200.0	200.1	0.9995	Slope	1.001343	0.90 - 1.10
100.0	99.2	1.0081			
			Intercept	-0.260000	+/- 5



O<sub>3</sub> Calibration Plot

Date: January 8, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Anzac Station number: AMS 14  
 Calibration Date: January 24, 2024 Last Cal Date: December 6, 2023  
 Start time (MST): 12:49 End time (MST): 13:05

Analyzer Make: API T640 S/N: 825  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388749  
 Temp/RH standard: Alicat FP-25 S/N: 388749

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-7.6	-7.3	-7.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	710.0	710.8	710.0	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	4.98	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 24, 2024</u>	Last Cal Date: <u>December 6, 2023</u>			
	PM w/o HEPA: <u>13.0</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>December 6, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>December 6, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: July 6, 2023  
 Date RH/T Sensor Cleaned: July 6, 2023

Notes: No adjustments made. Leak check passed. Head cleaned.

Calibration by: Mohammed Kashif





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS17  
WAPASU**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 8, 2024	Previous Calibration:	December 4, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:20	End Time (MST):	14:38
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999952	≥0.995
800.0	803.0	0.9962			
400.0	397.2	1.0072	Slope	1.004353	0.90 - 1.10
199.5	196.8	1.0138			
			Intercept	-2.079066	+/-30



SO2 Calibration Plot

Date: January 8, 2024

Location: Wapasu





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Wapasu Station number: AMS17  
 Calibration Date: January 29, 2024 Last Cal Date: December 5, 2023  
 Start time (MST): 11:17 End time (MST): 15:49  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.076 ppm Cal Gas Exp Date: September 16, 2024  
 Cal Gas Cylinder #: CC511852  
 Removed Cal Gas Conc: 5.076 ppm Rem Gas Exp Date: n/a  
 Removed Gas Cyl #: n/a Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 2449  
 ZAG Make/Model: API T701H Serial Number: 359

### Analyzer Information

Analyzer make: Thermo 450i Analyzer serial #: 1218153583  
 Converter make: n/a Converter serial #: n/a  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005283	1.009996	Backgd or Offset:	12.2
Calibration intercept:	0.040787	-0.299186	Coeff or Slope:	1.114

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4921	78.8	80.0	80.6	0.995
as found 2nd point	4961	39.4	40.0	40.9	0.983
as found 3rd point	4980	19.7	20.0	20.6	0.980
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4921	78.8	80.0	80.5	0.994
second point	4961	39.4	40.0	40.2	0.995
third point	4980	19.7	20.0	19.7	1.015
as left zero	5000	0.0	0.0	0.0	----
as left span	4921	78.8	80.0	80.0	1.000
SO2 Scrubber Check	4921	79.4	800.0	-0.1	----
Date of last scrubber change:	n/a			Ave Corr Factor	1.001
Date of last converter efficiency test:	n/a			efficiency	

Baseline Corr As found: 80.4 Prev response: 80.46 \*% change: -0.1%  
 Baseline Corr 2nd AF pt: 40.7 AF Slope: 1.004425 AF Intercept: 0.420818  
 Baseline Corr 3rd AF pt: 20.4 AF Correlation: 0.999949

\* = > +/-5% change initiates investigation

Notes: Pump changed out after as founds. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 29, 2024	Previous Calibration:	December 5, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:17	End Time (MST):	15:49
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient	≥0.995
80.0	80.5	0.9938		
40.0	40.2	0.9949	Slope	0.90 - 1.10
20.0	19.7	1.0153		
			Intercept	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 29, 2024

Location: Wapasu





# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	January 8, 2024	Last Cal Date:	December 21, 2023
Start time (MST):	11:20	End time (MST):	14:38
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	ALM066507	Cal Gas Expiry Date:	January 12, 2029
CH4 Cal Gas Conc.	<u>503.5</u> ppm	CH4 Equiv Conc.	1076.3 ppm
C3H8 Cal Gas Conc.	<u>208.3</u> ppm		
Removed Gas Cert:	n/a	Removed Gas Expiry:	n/a
Removed CH4 Conc.	<u>503.5</u> ppm	CH4 Equiv Conc.	1076.3 ppm
Removed C3H8 Conc.	<u>208.3</u> ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	2449
ZAG Make/Model:	API 701H	Serial Number:	359

### Analyzer Information

Analyzer make: Thermo 51i-LT      Analyzer serial #: 1218153352  
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011452	0.993554	Background:	3.300	3.020
Calibration intercept:	-0.239261	0.002025	Coefficient:	4.451	4.390

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.24	----
as found span	4921	79.4	17.09	17.01	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.06	----
high point	4921	79.4	17.09	17.03	1.004
second point	4960	39.7	8.55	8.42	1.015
third point	4980	19.8	4.26	4.21	1.013
as left zero	5000	0.0	0.00	0.01	----
as left span	4921	79.4	17.09	17.02	1.004
Average Correction Factor					1.011
Baseline Corr As found:	17.25	Previous response	17.05	*% change	1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = +/-5% change initiates investigation

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar





# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 8, 2024	Previous Calibration:	December 21, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:20	End Time (MST):	14:38
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.06	----	Correlation Coefficient	0.999931	≥0.995
17.09	17.03	1.0039			
8.55	8.42	1.0145	Slope	0.993554	0.90 - 1.10
4.26	4.21	1.0134			
			Intercept	0.002025	+/-1.5



THC Calibration Plot

Date: January 8, 2024

Location: Wapasu





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Wapasu  
Calibration Date: January 10, 2024  
Start time (MST): 11:15  
Reason: Routine  
Station number: AMS17  
Last Cal Date: December 7, 2023  
End time (MST): 15:35

### Calibration Standards

NO Gas Cylinder #: T375YK8  
NOX Cal Gas Conc: 49.11 ppm  
Removed Cylinder #: T375YK8  
Removed Gas NOX Conc: 49.11 ppm  
NOX gas Diff: T375YK8  
Calibrator Model: API T700  
ZAG make/model: API T701H  
Cal Gas Expiry Date: April 13, 2025  
NO Cal Gas Conc: 48.07 ppm  
Removed Gas Exp Date: T375YK8  
Removed Gas NO Conc: 48.07 ppm  
NO gas Diff: T375YK8  
Serial Number: 2449  
Serial Number: 359

### Analyzer Information

Analyzer make: Thermo Scientific 42iQ  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 12300522720

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.264	1.250	NO bkgnd or offset:	8.1	8.1
NOX coeff or slope:	0.989	0.986	NOX bkgnd or offset:	8.1	8.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	326.0	326.0

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.999676	1.003354
NO <sub>x</sub> Cal Offset:	-3.080000	-4.920000
NO Cal Slope:	1.000330	1.006788
NO Cal Offset:	-2.740000	-5.500000
NO <sub>2</sub> Cal Slope:	1.004002	0.997030
NO <sub>2</sub> Cal Offset:	-0.063828	-0.662092



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.6	-0.1	----	----
as found span	4917	83.2	817.2	799.9	17.3	827.8	809.3	18.5	0.9872	0.9884
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.4	-0.1	----	----
high point	4917	83.2	817.2	799.9	17.3	817.7	802.8	14.8	0.9994	0.9964
second point	4958	41.6	408.6	399.9	8.7	401.4	393.3	8.1	1.0179	1.0169
third point	4979	20.8	204.3	200.0	4.3	196.6	191.6	4.9	1.0392	1.0437
as left zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.5	-0.1	----	----
as left span	4917	83.2	817.2	393.8	423.4	816.8	392.6	424.1	1.0004	1.0029
Average Correction Factor									1.0188	1.0190

Corrected As found	NO <sub>x</sub> = 828.5 ppb	NO = 809.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 1.8%
Previous Response	NO <sub>x</sub> = 813.8 ppb	NO = 797.4 ppb		*Percent Change	NO = 1.5%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.3	393.2	423.4	421.8	1.0038	99.6%
2nd GPT point (200 ppb O3)	799.3	594.4	222.2	220.6	1.0073	99.3%
3rd GPT point (100 ppb O3)	799.3	698.5	118.1	116.5	1.0138	98.6%
Average Correction Factor					1.0083	99.2%

Notes:

Change filters after as founds. Adjusted span only.

Calibration Performed By:

Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 7, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:15	End Time (MST):	15:35
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.5	----	Correlation Coefficient	0.999858	≥0.995
817.2	817.7	0.9994			
408.6	401.4	1.0179	Slope	1.003354	0.90 - 1.10
204.3	196.6	1.0392			
			Intercept	-4.920000	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 7, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:15	End Time (MST):	15:35
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.4	----	Correlation Coefficient	0.999809	≥0.995
799.9	802.8	0.9964			
399.9	393.3	1.0169	Slope	1.006788	0.90 - 1.10
200.0	191.6	1.0437			
			Intercept	-5.500000	+/-20

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 7, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:15	End Time (MST):	15:35
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
423.4	421.8	1.0038		
222.2	220.6	1.0073		
118.1	116.5	1.0138		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: January 10, 2024

Location: Wapasu







# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	January 4, 2024	Last Cal Date:	December 6, 2023
Start time (MST):	11:30	End time (MST):	14:33
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	API T700	Serial Number:	2449
ZAG Make/Model:	API T701H	Serial Number:	359

### Analyzer Information

Analyzer make:	API T400	Analyzer serial #:	3870
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002943	1.001714	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.540000	-0.500000	Coeff or Slope:	1.014	1.014

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.7	----
as found span	5000	1077.3	400.0	399.3	1.002
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.3	----
high point	5000	1077.3	400.0	400.6	0.999
second point	5000	900.3	200.0	199.4	1.003
third point	5000	789.5	100.0	98.9	1.011
as left zero	5000	0.0	0.0	0.6	----
as left span	5000	1077.3	400.0	404.8	0.988
Average Correction Factor					1.004

Baseline Corr As found:	398.6	Previous response	400.6	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 6, 2023
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:30	End Time (MST):	14:33
Analyzer make:	API T400	Analyzer serial #:	3870

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999982
400.0	400.6	0.9985		
200.0	199.4	1.0030	Slope	1.001714
100.0	98.9	1.0111		
			Intercept	-0.500000
				+/- 5



O<sub>3</sub> Calibration Plot

Date: January 4, 2024

Location: Wapasu





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Wapasu Station number: AMS 17  
 Calibration Date: January 29, 2024 Last Cal Date: December 7, 2023  
 Start time (MST): 15:10 End time (MST): 15:46

Analyzer Make: API T640 S/N: 1183  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388751  
 Temp/RH standard: Alicat FP-25BT S/N: 388751

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	8.6	8.1	8.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	707.5	704.2	707.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.97	4.95	4.97	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 29, 2024</u>	Last Cal Date: <u>December 7, 2023</u>			
	PM w/o HEPA: <u>0.9</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>October 18, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 22, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_  
 Date RH/T Sensor Cleaned: \_\_\_\_\_

Notes: Temp, pressure and flow checked. Leak check passed.

Calibration by: Aswin Sasi Kumar



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS18 STONY MOUNTAIN**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	January 11, 2024	Last Cal Date:	December 19, 2023
Start time (MST):	11:45	End time (MST):	14:53
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.40	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC463851			
Removed Cal Gas Conc:	49.40	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API 701H		Serial Number:	360

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998520	1.005444	Backgd or Offset:	22.4	22.5
Calibration intercept:	-0.124003	-0.422589	Coeff or Slope:	0.803	0.813

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5009	0.0	0.0	0.5	----
as found span	4919	81.0	800.3	792.2	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.6	----
high point	4919	81.0	800.3	804.8	0.994
second point	4959	40.5	400.2	401.2	0.997
third point	4979	20.2	199.6	199.4	1.001
as left zero	5000	0.0	0.0	0.5	----
as left span	4919	81.0	800.3	804.4	0.995
Average Correction Factor					0.998

Baseline Corr As found:	791.70	Previous response	798.97	*% change	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 19, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:45	End Time (MST):	14:53
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999992	≥0.995
800.3	804.8	0.9944			
400.2	401.2	0.9975	Slope	1.005444	0.90 - 1.10
199.6	199.4	1.0010			
			Intercept	-0.422589	+/-30



SO2 Calibration Plot

Date: January 11, 2024

Location: Stony Mountain







# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Stony Mountain Station number: AMS18  
 Calibration Date: January 30, 2024 Last Cal Date: December 21, 2023  
 Start time (MST): 11:15 End time (MST): 15:36  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.479 ppm Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC500395  
 Removed Cal Gas Conc: 5.479 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658  
 ZAG Make/Model: Teledyne API T701 Serial Number: 360

### Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359  
 Converter make: CD Nova CDN-101 Converter serial #: 555  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000015	1.000728	Backgd or Offset: 2.66	2.66
Calibration intercept:	0.540982	0.341016	Coeff or Slope: 1.157	1.157

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4927	73.0	80.0	82.1	0.977
as found 2nd point	4964	36.5	40.0	41.3	0.973
as found 3rd point	4983	18.3	20.0	20.6	0.983
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4927	73.0	80.0	80.3	0.996
second point	4964	36.5	40.0	40.6	0.985
third point	4983	18.3	20.0	20.3	0.988
as left zero	5000	0.0	0.0	0.4	----
as left span	4927	73.0	80.0	79.0	1.013
SO2 Scrubber Check	4923	77.1	771.0	0.0	----
Date of last scrubber change:	17-Dec-21			Ave Corr Factor	0.990
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 81.9 Prev response: 80.54 \*% change: 1.7%  
 Baseline Corr 2nd AF pt: 41.1 AF Slope: 1.024592 AF Intercept: 0.180567  
 Baseline Corr 3rd AF pt: 20.4 AF Correlation: 0.999990

\* = > +/-5% change initiates investigation

Notes: Scrubber check completed after calibrator zero. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 30, 2024	Previous Calibration:	December 21, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	11:15	End Time (MST):	15:36
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999978	
80.0	80.3	0.9962			≥0.995
40.0	40.6	0.9850	Slope	1.000728	
20.0	20.3	0.9876			0.90 - 1.10
			Intercept	0.341016	+/-3



TRS Calibration Plot

Date: January 30, 2024

Location: Stony Mountain





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	January 11, 2024	Last Cal Date:	December 19, 2023
Start time (MST):	11:45	End time (MST):	14:53
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH <sub>4</sub> Cal Gas Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.24E-04	2.33E-04	NMHC SP Ratio:	5.04E-05
CH <sub>4</sub> Retention time:	12.7	13.1	NMHC Peak Area:	181790
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				5.08E-05
				180289
				OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	17.28	17.13	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.32	0.998
second point	4959	40.5	8.64	8.64	1.000
third point	4979	20.2	4.31	4.26	1.011
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.40	0.993
Average Correction Factor					1.003

Baseline Corr AF:	17.13	Prev response	17.33	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	9.05	1.013
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.17	0.999
second point	4959	40.5	4.58	4.61	0.995
third point	4979	20.2	2.29	2.28	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	9.17	9.24	0.992
Average Correction Factor					0.999
Baseline Corr AF:	9.05	Prev response	9.17	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	8.11	8.09	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	8.14	0.996
second point	4959	40.5	4.06	4.03	1.006
third point	4979	20.2	2.02	1.98	1.022
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.17	0.994
Average Correction Factor					1.008
Baseline Corr AF:	8.09	Prev response	8.16	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.004329	1.003258
THC Cal Offset:	-0.027780	-0.027185
CH <sub>4</sub> Cal Slope:	1.010541	1.005442
CH <sub>4</sub> Cal Offset:	-0.037005	-0.028409
NMHC Cal Slope:	0.998956	1.001076
NMHC Cal Offset:	0.009225	0.002223

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 19, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:45	End Time (MST):	14:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999987	$\geq 0.995$			
17.28	17.32	0.9978						
8.64	8.64	1.0001				Slope	1.003258	0.90 - 1.10
4.31	4.26	1.0113						
			Intercept	-0.027185	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 19, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:45	End Time (MST):	14:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999944	≥0.995			
8.11	8.14	0.9962						
4.06	4.03	1.0059				Slope	1.005442	0.90 - 1.10
2.02	1.98	1.0220						
			Intercept	-0.028409	+/-0.5			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 19, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:45	End Time (MST):	14:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999991	$\geq 0.995$
9.17	9.17	0.9994			
4.58	4.61	0.9947			
2.29	2.28	1.0017			
			Slope	1.001076	0.90 - 1.10
			Intercept	0.002223	+/-0.5






NMHC Calibration Plot

Date: January 11, 2024

Location: Stony Mountain





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	January 22, 2024	Last Cal Date:	January 11, 2024
Start time (MST):	11:40	End time (MST):	15:08
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH <sub>4</sub> Cal Gas Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.33E-04	2.27E-04	NMHC SP Ratio:	5.08E-05
CH <sub>4</sub> Retention time:	13.1	12.7	NMHC Peak Area:	180289
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				8.92E-05
				102774
				OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	17.28	16.35	1.057
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.44	0.991
second point	4959	40.5	8.64	8.62	1.002
third point	4979	20.2	4.31	4.22	1.020
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.22	1.003
Average Correction Factor					1.004

Baseline Corr AF:	16.35	Prev response	17.31	*% change	-5.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	8.89	1.031
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.32	0.984
second point	4959	40.5	4.58	4.61	0.995
third point	4979	20.2	2.29	2.26	1.014
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	9.17	9.12	1.006
Average Correction Factor					0.998
Baseline Corr AF:	8.89	Prev response	9.18	*% change	-3.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	8.11	7.46	1.088
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	8.12	0.999
second point	4959	40.5	4.06	4.01	1.011
third point	4979	20.2	2.02	1.97	1.027
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.11	1.001
Average Correction Factor					1.012
Baseline Corr AF:	7.46	Prev response	8.13	*% change	-9.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.003258	1.011154
THC Cal Offset:	-0.027185	-0.070367
CH <sub>4</sub> Cal Slope:	1.005442	1.002808
CH <sub>4</sub> Cal Offset:	-0.028409	-0.031814
NMHC Cal Slope:	1.001076	1.018713
NMHC Cal Offset:	0.002223	-0.038752

Notes: Dealing with dips on CH<sub>4</sub>. N<sub>2</sub> changed after As Finds. Low counts on chromatogram maybe leading to 9% low span. Input board calibrated. Carrier pressure increased 5psi, window adjusted.

Calibration Performed By: Ryan Power



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 22, 2024	Previous Calibration:	January 11, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:40	End Time (MST):	15:08
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999925	$\geq 0.995$			
17.28	17.44	0.9908						
8.64	8.62	1.0021				Slope	1.011154	0.90 - 1.10
4.31	4.22	1.0204						
			Intercept	-0.070367	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 22, 2024	Previous Calibration:	January 11, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:40	End Time (MST):	15:08
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999928	≥0.995			
8.11	8.12	0.9989						
4.06	4.01	1.0107				Slope	1.002808	0.90 - 1.10
2.02	1.97	1.0272						
			Intercept	-0.031814	+/-0.5			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 22, 2024	Previous Calibration:	January 11, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:40	End Time (MST):	15:08
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999921	$\geq 0.995$			
9.17	9.32	0.9835						
4.58	4.61	0.9949				Slope	1.018713	0.90 - 1.10
2.29	2.26	1.0141						
			Intercept	-0.038752	$\pm 0.5$			



NMHC Calibration Plot

Date: January 22, 2024

Location: Stony Mountain





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	January 23, 2024	Last Cal Date:	January 22, 2024
Start time (MST):	12:00	End time (MST):	N/A
Reason:	Removal		

### Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH <sub>4</sub> Cal Gas Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.27E-04	N/A	NMHC SP Ratio:	8.92E-05
CH <sub>4</sub> Retention time:	12.7	N/A	NMHC Peak Area:	102774
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	17.28	17.01	1.016
as found 2nd point	4959	40.5	8.64	8.43	1.025
as found 3rd point	4979	20.2	4.31	4.16	1.037
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

			Average Correction Factor		
Baseline Corr AF:	17.01	Prev response	17.40	*% change	-2.3%
Baseline Corr 2nd AF:	8.4	AF Slope:	0.985558	AF Intercept:	-0.049656
Baseline Corr 3rd AF:	4.2	AF Correlation:	0.999960	* = > +/-5% change initiates investigation	





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	8.92	1.028
as found 2nd point	4959	40.5	4.58	4.46	1.029
as found 3rd point	4979	20.2	2.29	2.18	1.048
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	8.92	Prev response	9.30	*% change	-4.3%
Baseline Corr 2nd AF:	4.5	AF Slope:	0.974470	AF Intercept:	-0.017839
Baseline Corr 3rd AF:	2.2	AF Correlation:	0.999973	* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	8.11	8.09	1.003
as found 2nd point	4959	40.5	4.06	3.98	1.020
as found 3rd point	4979	20.2	2.02	1.97	1.026
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	8.09	Prev response	8.10	*% change	-0.2%
Baseline Corr 2nd AF:	3.98	AF Slope:	0.998088	AF Intercept:	-0.031816
Baseline Corr 3rd AF:	1.97	AF Correlation:	0.999901	* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.011154	
THC Cal Offset:	-0.070367	
CH <sub>4</sub> Cal Slope:	1.002808	
CH <sub>4</sub> Cal Offset:	-0.031814	
NMHC Cal Slope:	1.018713	
NMHC Cal Offset:	-0.038752	


Notes: Dips still present after maintenance performed on January 22, 2024. Removing instrument for further maintenance back at the shop

Calibration Performed By: Aswin Sasi Kumar

NMHC Calibration Plot

Date: January 23, 2024

Location: Stony Mountain





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	January 23, 2024	Last Cal Date:	
Start time (MST):	14:40	End time (MST):	16:49
Reason:	Install		

### Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH <sub>4</sub> Cal Gas Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	500.8 ppm	CH <sub>4</sub> Equiv Conc.	1066.8 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:		3.13E-04	NMHC SP Ratio:	6.14E-05
CH <sub>4</sub> Retention time:		16.2	NMHC Peak Area:	149456
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.40	0.993
second point	4959	40.5	8.64	8.74	0.989
third point	4979	20.2	4.31	4.37	0.987
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.59	0.982
Average Correction Factor					0.990

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.27	0.989
second point	4959	40.5	4.58	4.64	0.987
third point	4979	20.2	2.29	2.34	0.976
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	9.17	9.42	0.973
Average Correction Factor					0.984
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	8.13	0.998
second point	4959	40.5	4.06	4.09	0.992
third point	4979	20.2	2.02	2.02	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.18	0.992
Average Correction Factor					0.997
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		1.006554
THC Cal Offset:		0.018647
CH <sub>4</sub> Cal Slope:		1.002919
CH <sub>4</sub> Cal Offset:		0.002793
NMHC Cal Slope:		1.009996
NMHC Cal Offset:		0.014455

Notes: Install calibration. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 23, 2024	Previous Calibration:	
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	14:40	End Time (MST):	16:49
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999993	$\geq 0.995$			
17.28	17.40	0.9932						
8.64	8.74	0.9890				Slope	1.006554	0.90 - 1.10
4.31	4.37	0.9868						
			Intercept	0.018647	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 23, 2024	Previous Calibration:	
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	14:40	End Time (MST):	16:49
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999985	$\geq 0.995$
8.11	8.13	0.9977			
4.06	4.09	0.9917			
2.02	2.02	1.0003			
			Slope	1.002919	0.90 - 1.10
			Intercept	0.002793	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 23, 2024	Previous Calibration:	
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	14:40	End Time (MST):	16:49
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999987	$\geq 0.995$			
9.17	9.27	0.9890						
4.58	4.64	0.9872				Slope	1.009996	0.90 - 1.10
2.29	2.34	0.9756						
			Intercept	0.014455	$\pm 0.5$			



NMHC Calibration Plot

Date: January 23, 2024

Location: Stony Mountain







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Stony Mountain Station number: AMS 18  
Calibration Date: January 25, 2024 Last Cal Date: December 21, 2023  
Start time (MST): 11:25 End time (MST): 16:02  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T26DHGA Cal Gas Expiry Date: November 17, 2026  
NOX Cal Gas Conc: 48.28 ppm NO Cal Gas Conc: 47.58 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 48.28 ppm Removed Gas NO Conc: 47.58 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 2658  
ZAG make/model: Teledyne API T701 Serial Number: 13779

### Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1336160088  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.076	1.102	NO bkgnd or offset:	3.0	3.1
NOX coeff or slope:	0.988	0.988	NOX bkgnd or offset:	3.0	3.1
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	249.1	252.3

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.001069	1.000344
NO <sub>x</sub> Cal Offset:	-0.390046	-0.080000
NO Cal Slope:	1.001210	1.001207
NO Cal Offset:	-1.489996	-0.760000
NO <sub>2</sub> Cal Slope:	1.001477	0.998708
NO <sub>2</sub> Cal Offset:	1.056533	0.435337



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.3	-0.1	0.5	----	----
as found span	4916	84.0	811.1	799.3	11.8	790.7	778.5	12.2	1.0258	1.0268
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	-0.1	0.3	----	----
high point	4916	84.0	811.1	799.3	11.8	811.1	799.6	11.6	1.0000	0.9997
second point	4958	42.0	405.6	399.7	5.9	406.5	399.9	6.6	0.9977	0.9994
third point	4979	21.0	202.8	199.8	2.9	201.8	198.1	3.7	1.0048	1.0088
as left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
as left span	4916	84.0	811.1	361.3	449.8	815.0	357.0	458.0	0.9952	1.0122
Average Correction Factor									1.0008	1.0026

Corrected As found	NO <sub>x</sub> = 790.4 ppb	NO = 778.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -2.7%	
Previous Response	NO <sub>x</sub> = 811.6 ppb	NO = 798.8 ppb		*Percent Change	NO = -2.6%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	800.6	362.6	449.8	449.7	1.0001	100.0%
2nd GPT point (200 ppb O3)	800.6	587.2	225.2	224.9	1.0012	99.9%
3rd GPT point (100 ppb O3)	800.6	696.7	115.7	116.4	0.9936	100.6%
Average Correction Factor					0.9983	100.2%

Notes:

Calibration gas cylinder T2XX7ME removed on January 23, 2024. Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	December 21, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:25	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
811.1	811.1	1.0000		
405.6	406.5	0.9977		
202.8	201.8	1.0048		
			0.999995	
			1.000344	
			-0.080000	





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	December 21, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:25	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
799.3	799.6	0.9997			
399.7	399.9	0.9994			
199.8	198.1	1.0088			
			Slope	1.001207	0.90 - 1.10
			Intercept	-0.760000	+/-20





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	December 21, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:25	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
449.8	449.7	1.0001		
225.2	224.9	1.0012		
115.7	116.4	0.9936		



NO<sub>x</sub> Calibration Plot

Date: January 25, 2024

Location: Stony Mountain





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Stony Mountain Station number: AMS18  
 Calibration Date: January 24, 2024 Last Cal Date: December 12, 2023  
 Start time (MST): 11:30 End time (MST): 15:30  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658  
 ZAG Make/Model: Teledyne API 701H Serial Number: 360

### Analyzer Information

Analyzer make: API T400 Analyzer serial #: 825  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002257	1.002743	Backgd or Offset:	2.4	-0.7
Calibration intercept:	-1.420000	0.020000	Coeff or Slope:	0.978	0.971

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	NA	0.0	-1.8	----
as found span	4804	1141.9	400.0	400.3	0.999
as found 2nd point	4804	884.5	200.0	201.1	0.995
as found 3rd point	4888	741.4	100.0	99.9	1.001
calibrator zero	5000	NA	0.0	0.0	----
high point	4888	1138.1	400.0	401.5	0.996
second point	4888	884.5	200.0	199.4	1.003
third point	4888	741.4	100.0	101.1	0.989
as left zero	5000	NA	0.0	0.0	----
as left span	4812	1097.9	400.0	404.0	0.990
Average Correction Factor					0.996

Baseline Corr As found:	402.1	Previous response	399.5	*% change	0.7%
Baseline Corr 2nd AF pt:	202.9	AF Slope:	1.004771	AF Intercept:	-0.960000
Baseline Corr 3rd AF pt:	101.7	AF Correlation:	0.999972		

*\* = > +/-5% change initiates investigation*

Notes: Pump changed out after multi-point as founds. Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 24, 2024	Previous Calibration:	December 12, 2023
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	11:30	End Time (MST):	15:30
Analyzer make:	API T400	Analyzer serial #:	825

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999975	≥0.995
400.0	401.5	0.9963			
200.0	199.4	1.0030	Slope	1.002743	0.90 - 1.10
100.0	101.1	0.9891			
			Intercept	0.020000	+/- 5






O<sub>3</sub> Calibration Plot

Date: January 24, 2024

Location: Stony Mountain





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Stony Mountain Station number: AMS 18  
 Calibration Date: January 30, 2024 Last Cal Date: December 20, 2023  
 Start time (MST): 14:20 End time (MST): 15:27

Analyzer Make: API T640 S/N: 1162  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748  
 Temp/RH standard: Alicat FP-25BT S/N: 388748

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	11.9	11.64	11.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	691.2	691.09	691.2	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.97	4.72	5.02	<input checked="" type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 30, 2024</u>	Last Cal Date: <u>December 20, 2023</u>			
	PM w/o HEPA: <u>0.8</u>	PM w/ HEPA: <u>0.0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	NA		NA	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>September 25, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>September 25, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022  
 Date RH/T Sensor Cleaned: August 30, 2022

Notes: Flow adjusted.

Calibration by: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Report

Version-01-2020

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	January 17, 2024	Last Cal Date:	December 12, 2023
Start time (MST):	11:15	End time (MST):	14:15
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,050	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	3,050	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701H		Serial Number:	355

### Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	<b><u>Start</u></b>	<b><u>Finish</u></b>		<b><u>Start</u></b>	<b><u>Finish</u></b>
Calibration slope:	0.992948	0.999887	Backgd or Offset:	-0.010	-0.010
Calibration intercept:	0.219785	0.043782	Coeff or Slope:	0.906	0.908

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4933	66.7	40.7	41.1	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.7	40.7	1.001
second point	4966	33.3	20.3	20.5	0.990
third point	4983	16.7	10.2	10.2	1.001
as left zero	5000	0.0	0.0	0.0	----
as left span	4933	66.7	40.7	40.9	0.995
Average Correction Factor					0.997

Baseline Corr As found:	40.89	Prev response:	40.62	*% change:	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	December 12, 2023
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:15	End Time (MST):	14:15
Analyzer make:	API T300	Analyzer serial #:	3504

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999961	
40.7	40.7	1.0007			$\geq 0.995$
20.3	20.5	0.9901	Slope	0.999887	
10.2	10.2	1.0007			$0.90 - 1.10$
			Intercept	0.043782	$\pm 1.5$



CO Calibration Plot

Date: January 17, 2024

Location: Stony Mountain





# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	January 15, 2024	Last Cal Date:	December 20, 2023
Start time (MST):	11:37	End time (MST):	14:36
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2026
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
N2 Gen Make/Model:	Peak Scientific		Serial Number:	771048317

### Analyzer Information

Analyzer make: API T360	Analyzer serial #: 489
Analyzer Range 0 - 2,000 ppm	

	<b>Start</b>	<b>Finish</b>		<b>Start</b>	<b>Finish</b>
Calibration slope:	0.999080	0.996375	Backgd or Offset:	-0.037	-0.037
Calibration intercept:	-2.620000	-3.620000	Coeff or Slope:	0.938	0.938

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	0.2	----
as found span	2920	80.0	1605.9	1597.1	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.4	----
high point	2920	80.0	1605.9	1600.0	1.004
second point	2960	40.0	802.9	789.7	1.017
third point	2980	20.0	401.5	395.5	1.015
as left zero	3000	0.0	0.0	0.1	----
as left span	2930	80.0	1600.5	1594.9	1.004
<b>Average Correction Factor</b>					<b>1.012</b>

Baseline Corr As found:	1596.90	Prev response:	1601.77	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date	January 15, 2024	Previous Calibration	December 20, 2023
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	11:37	End Time (MST)	14:36
Analyzer make	API T360	Analyzer serial #	489

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.4	----	Correlation Coefficient	0.999947	<b>≥0.995</b>
1605.9	1600.0	1.0037	Slope	0.996375	<b>0.90 - 1.10</b>
802.9	789.7	1.0168	Intercept	-3.620000	<b>+/-10</b>
401.5	395.5	1.0151			



CO<sub>2</sub> Calibration Plot

Date: January 15, 2024

Location: Stony Mountain







## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS19  
FIREBAG**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	January 8, 2024	Last Cal Date:	December 13, 2023
Start time (MST):	12:01	End time (MST):	15:55
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.29	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC716618			
Removed Cal Gas Conc:	49.29	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1607
ZAG Make/Model:	API T701		Serial Number:	1118

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005909	0.996990	Backgd or Offset:	10.6	10.5
Calibration intercept:	-0.303040	0.518257	Coeff or Slope:	0.993	0.984

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	-0.3	----
as found span	4919	81.1	799.5	805.0	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.0	-0.1	----
high point	4919	81.1	799.5	797.2	1.003
second point	4959	40.6	400.3	400.1	1.000
third point	4980	20.3	200.1	200.5	0.998
as left zero	4999	0.0	0.0	-0.3	----
as left span	4919	81.1	799.5	800.0	0.999
Average Correction Factor					1.000

Baseline Corr As found:	805.30	Previous response	803.89	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 8, 2024	Previous Calibration:	December 13, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	12:01	End Time (MST):	15:55
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999997	≥0.995
799.5	797.2	1.0028			
400.3	400.1	1.0004	Slope	0.996990	0.90 - 1.10
200.1	200.5	0.9980			
			Intercept	0.518257	+/-30



SO2 Calibration Plot

Date: January 8, 2024

Location: Firebag





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Firebag Station number: AMS19  
 Calibration Date: January 22, 2024 Last Cal Date: December 11, 2023  
 Start time (MST): 11:22 End time (MST): 16:49  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.114 ppm Cal Gas Exp Date: February 5, 2024  
 Cal Gas Cylinder #: CC517427  
 Removed Cal Gas Conc: 5.114 ppm Rem Gas Exp Date: n/a  
 Removed Gas Cyl #: n/a Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1607  
 ZAG Make/Model: Teledyne API T701 Serial Number: 1118

### Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680032  
 Converter make: Global Converter serial #: 2022-222  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000196	1.004771	Backgd or Offset:	2.82
Calibration intercept:	-0.061587	-0.021722	Coeff or Slope:	1.176
				3.38
				1.225

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4922	78.2	80.0	77.7	1.032
as found 2nd point	4961	39.1	40.0	39.1	1.028
as found 3rd point	4980	19.6	20.0	19.8	1.023
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4922	78.2	80.0	80.3	0.996
second point	4961	39.1	40.0	40.2	0.995
third point	4980	19.6	20.0	20.2	0.992
as left zero	5000	0.0	0.0	0.0	----
as left span	4922	78.2	80.0	79.8	1.002
SO2 Scrubber Check	4922	78.3	800.2	0.0	----
Date of last scrubber change:	January 18, 2023			Ave Corr Factor	0.994
Date of last converter efficiency test:	n/a			efficiency	

Baseline Corr As found: 77.5 Prev response: 79.93 \*% change: -3.1%  
 Baseline Corr 2nd AF pt: 38.9 AF Slope: 0.968467 AF Intercept: 0.299057  
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999992

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Ran SOx scrubber check after cal zero. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 22, 2024	Previous Calibration:	December 11, 2023
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	11:22	End Time (MST):	16:49
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient	≥0.995
80.0	80.3	0.9960		
40.0	40.2	0.9948	Slope	0.90 - 1.10
20.0	20.2	0.9925		
			Intercept	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 22, 2024

Location: Firebag







# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	January 8, 2024	Last Cal Date:	December 13, 2023
Start time (MST):	12:01	End time (MST):	15:55
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1607
ZAG Make/Model:	API T701	Serial Number:	1118

### Analyzer Information

Analyzer make: Thermo 51i-LT                      Analyzer serial #: 1336160089  
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998011	0.995342	Background:	2.04	2.03
Calibration intercept:	0.042876	0.045092	Coefficient:	3.793	3.764

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.00	0.12	----
as found span	4919	81.1	17.31	17.50	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.00	0.08	----
high point	4919	81.1	17.31	17.31	1.000
second point	4959	40.6	8.66	8.58	1.010
third point	4980	20.3	4.33	4.37	0.990
as left zero	5000	0.0	0.00	0.02	----
as left span	4919	81.1	17.31	17.38	0.996
Average Correction Factor					1.000
Baseline Corr As found:	17.38	Previous response	17.31	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filters after as founds. Adjusted span.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 8, 2024	Previous Calibration:	December 13, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	12:01	End Time (MST):	15:55
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.08	----	Correlation Coefficient	0.999933	
17.31	17.31	0.9997			≥0.995
8.66	8.58	1.0098	Slope	0.995342	
4.33	4.37	0.9905			0.90 - 1.10
			Intercept	0.045092	+/-1.5



THC Calibration Plot

Date: January 8, 2024

Location: Firebag





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Firebag  
Calibration Date: January 23, 2024  
Start time (MST): 11:17  
Reason: Routine  
Station number: AMS 19  
Last Cal Date: December 12, 2023  
End time (MST): 15:38

### Calibration Standards

NO Gas Cylinder #: DT0044018  
NOX Cal Gas Conc: 48.9 ppm  
Removed Cylinder #: n/a  
Removed Gas NOX Conc: 48.9 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701  
Cal Gas Expiry Date: November 3, 2031  
NO Cal Gas Conc: 48.7 ppm  
Removed Gas Exp Date: n/a  
Removed Gas NO Conc: 48.7 ppm  
NO gas Diff:  
Serial Number: 1607  
Serial Number: 1118

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1410661309

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.089	1.089	NO bkgnd or offset:	7.6	7.6
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	7.7	7.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	215.5	217.1

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.005184	1.011351
NO <sub>x</sub> Cal Offset:	0.260458	0.260390
NO Cal Slope:	1.007037	1.014424
NO Cal Offset:	-0.099613	-0.039625
NO <sub>2</sub> Cal Slope:	0.999988	1.000829
NO <sub>2</sub> Cal Offset:	-1.617748	-0.524220



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
as found span	4919	81.0	792.2	788.9	3.2	810.0	807.0	2.6	0.9780	0.9776
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	4999	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4919	81.5	797.1	793.8	3.3	806.0	805.0	0.5	0.9889	0.9861
second point	4959	40.7	398.0	396.4	1.6	403.6	402.7	0.9	0.9862	0.9844
third point	4980	20.4	199.5	198.7	0.8	202.0	201.2	0.8	0.9877	0.9876
as left zero	4999	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as left span	4919	81.0	792.2	367.3	424.9	799.6	374.5	425.1	0.9907	0.9808
Average Correction Factor									0.9876	0.9860

Corrected As found	NO <sub>x</sub> = 810.3 ppb	NO = 807.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 1.7%
Previous Response	NO <sub>x</sub> = 796.5 ppb	NO = 794.4 ppb		*Percent Change	NO = 1.6%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.5	375.9	424.9	424.8	1.0001	100.0%
2nd GPT point (200 ppb O3)	797.5	587.8	213.0	212.8	1.0008	99.9%
3rd GPT point (100 ppb O3)	797.5	694.5	106.3	105.0	1.0120	98.8%
Average Correction Factor					1.0043	99.6%

Notes: Changed sample inlet filters after as founds. No adjustments made. Second high NO reference point used in GPT formulas.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 23, 2024	Previous Calibration:	December 12, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:17	End Time (MST):	15:38
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
797.1	806.0	0.9889			
398.0	403.6	0.9862			
199.5	202.0	0.9877			
			Slope	1.011351	0.90 - 1.10
			Intercept	0.260390	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 23, 2024	Previous Calibration:	December 12, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:17	End Time (MST):	15:38
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
793.8	805.0	0.9861		
396.4	402.7	0.9844		
198.7	201.2	0.9876		

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 23, 2024	Previous Calibration:	December 12, 2023
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:17	End Time (MST):	15:38
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
424.9	424.8	1.0001		
213.0	212.8	1.0008		
106.3	105.0	1.0120		
			0.999990	
			1.000829	
			-0.524220	

**NO<sub>2</sub> Calibration Curve**






NO<sub>x</sub> Calibration Plot

Date: January 23, 2024

Location: Firebag





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS20  
MACKAY RIVER**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	December 4, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:25	End Time (MST):	11:55
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999942	
800.3	798.2	1.0026			≥0.995
400.7	401.7	0.9974	Slope	0.995124	
199.8	205.1	0.9743			0.90 - 1.10
			Intercept	2.791182	+/-30



SO2 Calibration Plot

Date: January 16, 2024

Location: MacKay River





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: MacKay River      Station number: AMS20  
 Calibration Date: January 3, 2024      Last Cal Date: December 5, 2023  
 Start time (MST): 8:44      End time (MST): 12:49  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.12 ppm      Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC515997  
 Removed Cal Gas Conc: 5.12 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700      Serial Number: 1220  
 ZAG Make/Model: Teledyne API 701      Serial Number: 4522

### Analyzer Information

Analyzer make: Thermo 43i TLE      Analyzer serial #: 1236656117  
 Converter make: Global      Converter serial #: 2022-226  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996103	0.991531	Backgd or Offset: 3.19	3.19
Calibration intercept:	0.439357	0.499365	Coeff or Slope: 1.113	1.113

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4922	78.1	80.0	80.4	0.996
as found 2nd point	4961	39.0	39.9	40.6	0.986
as found 3rd point	4980	19.5	20.0	20.6	0.974
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4922	78.1	80.0	79.7	1.004
second point	4961	39.0	40.0	40.3	0.992
third point	4980	19.5	20.0	20.6	0.970
as left zero	5000	0.0	0.0	0.3	----
as left span	4922	78.1	80.0	79.0	1.013
SO2 Scrubber Check	4982	81.3	802.8	0.0	----
Date of last scrubber change:	May 25, 2023			Ave Corr Factor	0.989
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 80.3      Prev response: 80.10      \*% change: 0.2%  
 Baseline Corr 2nd AF pt: 40.5      AF Slope: 1.002738      AF Intercept: 0.359563  
 Baseline Corr 3rd AF pt: 20.5      AF Correlation: 0.999950

\* = > +/-5% change initiates investigation

Notes: No maintenance and adjustments done. Sox scrubber checked after the calibrator zero.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 3, 2024	Previous Calibration:	December 5, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:44	End Time (MST):	12:49
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999935	≥0.995
80.0	79.7	1.0042			
40.0	40.3	0.9917	Slope	0.991531	0.90 - 1.10
20.0	20.6	0.9702			
			Intercept	0.499365	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 3, 2024

Location: MacKay River









# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	December 4, 2023
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:25	End Time (MST):	11:54
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.01	----	Correlation Coefficient	0.999910	≥0.995
17.34	17.19	1.0089			
8.68	8.64	1.0044	Slope	0.987892	0.90 - 1.10
4.33	4.45	0.9729			
			Intercept	0.076023	+/-1.5



THC Calibration Plot

Date: January 16, 2024

Location: MacKay River







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	0.0	----	----
as found span	4917	83.3	819.5	800.3	19.2	841.7	817.8	23.9	0.9736	0.9786
as found 2nd	4956	41.7	410.4	400.8	9.6	423.3	410.0	13.3	0.9696	0.9777
as found 3rd	4979	20.8	204.6	199.9	4.8	216.2	207.8	8.4	0.9465	0.9618
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

### Average Correction Factor

Corrected As found	NO <sub>x</sub> = 841.7 ppb	NO = 817.7 ppb	<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = 1.6%
Previous Response	NO <sub>x</sub> = 828.5 ppb	NO = 798.3 ppb			*Percent Change	NO = 2.4%
Baseline Corr 2nd pt	NO <sub>x</sub> = 423.3 ppb	NO = 409.9 ppb	As found	NO <sub>x</sub> r <sup>2</sup> : 0.999942	Nx SI: 1.024870	Nx Int: 2.749
Baseline Corr 3rd pt	NO <sub>x</sub> = 216.2 ppb	NO = 207.7 ppb	As found	NO r <sup>2</sup> : 0.999978	NO SI: 1.020292	NO Int: 1.570
			As found	NO <sub>2</sub> r <sup>2</sup> : 0.999993	NO <sub>2</sub> SI: 1.012757	NO <sub>2</sub> Int: -0.346

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	0.0	----	----
as found GPT point (400 ppb NO <sub>2</sub> )	808.4	461.5	366.1	370.5	0.9880	101.2%
as found GPT point (200 ppb NO <sub>2</sub> )	808.4	626.2	201.4	203.7	0.9885	101.2%
as found GPT point (100 ppb NO <sub>2</sub> )	808.4	712.1	115.5	116.0	0.9953	100.5%
1st GPT point (400 ppb O <sub>3</sub> )						
2nd GPT point (200 ppb O <sub>3</sub> )						
3rd GPT point (100 ppb O <sub>3</sub> )						

### Average Correction Factor

Notes:

As founds to clean the reaction cell. NO and Nox spans still slowly dropping during calibration.


Calibration Performed By:

Melissa Lemay

NO<sub>x</sub> Calibration Plot

Date: January 4, 2024

Location: MacKay River





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: MacKay River Station number: AMS20  
Calibration Date: January 5, 2024 Last Cal Date: January 4, 2024  
Start time (MST): 8:10 End time (MST): 11:58  
Reason: Routine Calibration after Cleaning reaction cell

### Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025  
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 1220  
ZAG make/model: Teledyne API 701 Serial Number: 4522

### Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.981	0.904	NO bkgnd or offset:	4.0	3.7
NOX coeff or slope:	0.997	0.992	NOX bkgnd or offset:	4.1	3.7
NO2 coeff or slope:	0.995	0.995	Reaction cell Press:	190.9	190.9

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.007887	0.991609
NO <sub>x</sub> Cal Offset:	2.589374	3.242799
NO Cal Slope:	0.995492	0.992474
NO Cal Offset:	1.631214	2.102999
NO <sub>2</sub> Cal Slope:	1.005354	1.000326
NO <sub>2</sub> Cal Offset:	0.716865	-2.097451



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.4	0.4	0.0	----	----
high point	4917	83.3	819.5	800.3	19.2	814.8	795.8	19.1	1.0057	1.0057
second point	4958	41.7	410.3	400.7	9.6	410.3	399.8	10.5	0.9999	1.0022
third point	4979	20.8	204.6	199.9	4.8	209.8	202.7	7.1	0.9754	0.9860
as left zero	5000	0.0	0.0	0.0	0.0	0.5	0.4	0.1	----	----
as left span	4917	83.3	819.5	464.3	355.2	801.8	447.5	354.4	1.0220	1.0375
Average Correction Factor									0.9937	0.9979

Corrected As found	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	<i>* =&gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = NA
Previous Response	NO <sub>x</sub> = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
					As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	786.2	450.2	355.2	354.5	1.0019	99.8%
2nd GPT point (200 ppb O3)	786.2	609.5	195.9	192.8	1.0159	98.4%
3rd GPT point (100 ppb O3)	786.2	692.9	112.5	108.0	1.0413	96.0%
Average Correction Factor					1.0197	98.1%

Notes: Calibration after cleaning the reaction cell. Span adjusted.

Calibration Performed By: Melissa Lemay





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	January 4, 2024
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	8:10	End Time (MST):	11:58
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
819.5	814.8	1.0057		
410.3	410.3	0.9999		
204.6	209.8	0.9754		
			0.999938	
			0.991609	
			3.242799	





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	January 4, 2024
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	8:10	End Time (MST):	11:58
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	795.8	1.0057		
400.7	399.8	1.0022		
199.9	202.7	0.9860		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	January 4, 2024
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:10	End Time (MST):	11:58
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>			
0.0	0.0	----	Correlation Coefficient	0.999807	≥0.995			
355.2	354.5	1.0019						
195.9	192.8	1.0159				Slope	1.000326	0.90 - 1.10
112.5	108.0	1.0413						
			Intercept	-2.097451	+/-20			



NO<sub>x</sub> Calibration Plot

Date: January 5, 2024

Location: MacKay River





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: MacKay River Station number: AMS20  
Calibration Date: January 24, 2024 Last Cal Date: January 5, 2024  
Start time (MST): 8:05 End time (MST): 11:04  
Reason: As Found Remove Internal Zero/Span Valve, replace O-ring

### Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025  
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 1220  
ZAG make/model: Teledyne API 701 Serial Number: 4522

### Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.904	0.904	NO bkgnd or offset:	3.7	3.7
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	3.7	3.7
NO2 coeff or slope:	0.995	0.995	Reaction cell Press:	188.2	188.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.991609	
NO <sub>x</sub> Cal Offset:	3.242799	
NO Cal Slope:	0.992474	
NO Cal Offset:	2.102999	
NO <sub>2</sub> Cal Slope:	1.000326	
NO <sub>2</sub> Cal Offset:	-2.097451	



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as found span	4917	83.3	819.5	800.3	19.2	806.5	786.5	20.0	1.0161	1.0175
as found 2nd	4958	41.7	410.3	400.7	9.6	405.0	394.1	10.9	1.0130	1.0167
as found 3rd	4979	20.8	204.6	199.9	4.8	206.1	199.4	6.6	0.9929	1.0023
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

### Average Correction Factor

Corrected As found	NO <sub>x</sub> = 806.4 ppb	NO = 786.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -1.2%
Previous Response	NO <sub>x</sub> = 815.8 ppb	NO = 796.4 ppb		*Percent Change	NO = -1.3%
Baseline Corr 2nd pt	NO <sub>x</sub> = 404.9 ppb	NO = 394.1 ppb	As found	NO <sub>x</sub> r <sup>2</sup> : 0.999963	Nx SI: 0.982294
Baseline Corr 3rd pt	NO <sub>x</sub> = 206.0 ppb	NO = 199.4 ppb	As found	NO r <sup>2</sup> : 0.999983	NO SI: 0.981522
			As found	NO <sub>2</sub> r <sup>2</sup> : 0.999972	NO <sub>2</sub> SI: 1.012861
					NO <sub>2</sub> Int: -0.728

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	0.1	----	----
as found GPT point (400 ppb NO <sub>2</sub> )	769.9	429.6	359.5	363.9	0.9878	101.2%
as found GPT point (200 ppb NO <sub>2</sub> )	769.9	593.8	195.3	196.5	0.9937	100.6%
as found GPT point (100 ppb NO <sub>2</sub> )	769.9	680.0	109.1	108.9	1.0014	99.9%
1st GPT point (400 ppb O <sub>3</sub> )						
2nd GPT point (200 ppb O <sub>3</sub> )						
3rd GPT point (100 ppb O <sub>3</sub> )						

### Average Correction Factor


Notes: As founds to do maintenance on the analyzer. Due to drifting during the GPT the 2nd NO ref point used for as found GPT. Internal Valve removed, Reaction cell O ring replaced. Adjusted the PMT.

Calibration Performed By: Melissa Lemay

NO<sub>x</sub> Calibration Plot

Date: January 24, 2024

Location: MacKay River





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: MacKay River Station number: AMS20  
Calibration Date: January 25, 2024 Last Cal Date: January 24, 2024  
Start time (MST): 7:46 End time (MST): 11:37  
Reason: Routine Remove Internal Zero/Span Valve, replace O-ring, Adjusted the PMT

### Calibration Standards

NO Gas Cylinder #: T376265 Cal Gas Expiry Date: April 13, 2025  
NOX Cal Gas Conc: 49.19 ppm NO Cal Gas Conc: 48.04 ppm  
Removed Cylinder #: NA Removed Gas Exp Date: NA  
Removed Gas NOX Conc: 49.19 ppm Removed Gas NO Conc: 48.04 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 1220  
ZAG make/model: Teledyne API 701 Serial Number: 4522

### Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1505164379  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.904	0.990	NO bkgnd or offset:	3.7	2.8
NOX coeff or slope:	0.992	0.994	NOX bkgnd or offset:	3.7	3.0
NO2 coeff or slope:	0.995	1.000	Reaction cell Press:	164.5	164.5

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.991609	0.997760
NO <sub>x</sub> Cal Offset:	3.242799	3.562350
NO Cal Slope:	0.992474	0.996944
NO Cal Offset:	2.102999	2.762658
NO <sub>2</sub> Cal Slope:	1.000326	1.002101
NO <sub>2</sub> Cal Offset:	-2.097451	-0.557879





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
high point	4917	83.3	819.5	800.3	19.2	819.6	799.4	20.3	0.9998	1.0011
second point	4958	41.7	410.3	400.7	9.6	414.0	403.1	10.9	0.9910	0.9940
third point	4979	20.8	204.6	199.9	4.8	211.9	205.1	6.8	0.9657	0.9744
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
as left span	4917	83.3	819.5	444.6	374.9	818.3	441.7	376.7	1.0014	1.0066
Average Correction Factor									0.9855	0.9898

Corrected As found	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = NA
Previous Response	NO <sub>x</sub> = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
					As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.8	441.1	374.9	375.3	0.9988	100.1%
2nd GPT point (200 ppb O3)	796.8	611.9	204.1	203.9	1.0008	99.9%
3rd GPT point (100 ppb O3)	796.8	701.0	115.0	114.0	1.0084	99.2%
Average Correction Factor					1.0027	99.7%

Notes:

PMT adjusted, Zero/Span Valve removed, O ring replaced.

Calibration Performed By:

Melissa Lemay



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	January 24, 2024
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	7:46	End Time (MST):	11:37
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
819.5	819.6	0.9998			
410.3	414.0	0.9910			
204.6	211.9	0.9657			
			Slope	0.997760	0.90 - 1.10
			Intercept	3.562350	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	January 24, 2024
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	7:46	End Time (MST):	11:37
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
800.3	799.4	1.0011			
400.7	403.1	0.9940			
199.9	205.1	0.9744			
			Slope	0.996944	0.90 - 1.10
			Intercept	2.762658	+/-20

NO Calibration Curve





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	January 24, 2024
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:46	End Time (MST):	11:37
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.0	-0.1	----	Correlation Coefficient	0.999991	≥0.995			
374.9	375.3	0.9988						
204.1	203.9	1.0008				Slope	1.002101	0.90 - 1.10
115.0	114.0	1.0084						
			Intercept	-0.557879	+/-20			



NO<sub>x</sub> Calibration Plot

Date: January 25, 2024

Location: MacKay River





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS21  
CONKLIN**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	January 15, 2024	Last Cal Date:	December 4, 2023
Start time (MST):	11:59	End time (MST):	15:03
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.93	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	<u>CC259455</u>			
Removed Cal Gas Conc:	49.93	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3810
ZAG Make/Model:	Teledyne API 701		Serial Number:	691

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995935	0.997790	Backgd or Offset:	28.6	28.3
Calibration intercept:	0.975766	1.075857	Coeff or Slope:	0.904	0.911

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5005	0.0	0.0	-0.1	----
as found span	4920	80.2	800.8	791.1	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5005	0.0	0.0	0.0	----
high point	4920	80.2	800.8	799.9	1.001
second point	4960	40.1	400.4	400.3	1.000
third point	4980	20.0	200.1	202.4	0.989
as left zero	5005	0.0	0.0	-0.1	----
as left span	4920	80.2	800.8	800.8	1.000
Average Correction Factor					0.997

Baseline Corr As found:	791.20	Previous response	798.57	*% change	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 4, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:59	End Time (MST):	15:03
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	0.999988	
800.8	799.9	1.0012			≥0.995
400.4	400.3	1.0003	Slope	0.997790	
200.1	202.4	0.9887			0.90 - 1.10
			Intercept	1.075857	+/-30






SO2 Calibration Plot

Date: January 15, 2024

Location: Conklin





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Conklin Station number: AMS21  
 Calibration Date: January 24, 2024 Last Cal Date: December 7, 2023  
 Start time (MST): 10:56 End time (MST): 15:54  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.00 ppm Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC501204  
 Removed Cal Gas Conc: 5.00 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3810  
 ZAG Make/Model: Teledyne API 701H Serial Number: 691

### Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1236656116  
 Converter make: CD-Nova 101 Converter serial #: NA  
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995286	1.011429	Backgd or Offset:	2.4	2.44
Calibration intercept:	0.240000	0.300000	Coeff or Slope:	0.963	0.992

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.0	80.0	79.9	1.001
as found 2nd point	4960	40.0	40.0	40.0	1.000
as found 3rd point	4980	20.0	20.0	20.3	0.985
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4920	80.0	80.0	81.1	0.986
second point	4960	40.0	40.0	41.0	0.976
third point	4980	20.0	20.0	20.5	0.976
as left zero	5000	0.0	0.0	0.4	----
as left span	4920	80.0	80.0	80.9	0.989
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.979
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 79.9 Prev response: 79.86 \*% change: 0.0%  
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 0.997429 AF Intercept: 0.140000  
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999981

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Scrubber check done after calibrator zero and it passed.  
 Adjusted span only.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 24, 2024	Previous Calibration:	December 7, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:56	End Time (MST):	15:54
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999977	
80.0	81.1	0.9864			≥0.995
40.0	41.0	0.9756	Slope	1.011429	
20.0	20.5	0.9756			0.90 - 1.10
			Intercept	0.300000	+/-3



TRS Calibration Plot

Date: January 24, 2024

Location: Conklin





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	January 15, 2024	Last Cal Date:	December 14, 2023
Start time (MST):	11:59	End time (MST):	15:03
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	497.9 ppm	CH <sub>4</sub> Equiv Conc.	1067.7 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	497.9 ppm	CH <sub>4</sub> Equiv Conc.	1067.7 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.69E-04	3.86E-04	NMHC SP Ratio:	7.47E-05
CH <sub>4</sub> Retention time:	17.2	17.4	NMHC Peak Area:	122336
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	16.88	1.015
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.02	1.006
second point	4960	40.1	8.56	8.56	1.000
third point	4980	20.0	4.28	4.34	0.987
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.00	1.007

Average Correction Factor				0.998
Baseline Corr AF:	16.88	Prev response	17.11	*% change -1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.11	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.11	1.003
second point	4960	40.1	4.57	4.58	0.998
third point	4980	20.0	2.28	2.32	0.984
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.10	1.004
Average Correction Factor					0.995
Baseline Corr AF:	9.11	Prev response	9.15	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.77	1.028
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.90	1.011
second point	4960	40.1	3.99	3.98	1.003
third point	4980	20.0	2.00	2.02	0.990
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.89	1.012
Average Correction Factor					1.001
Baseline Corr AF:	7.77	Prev response	7.96	*% change	-2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997358	0.992346
THC Cal Offset:	0.029967	0.043763
CH <sub>4</sub> Cal Slope:	0.995501	0.988303
CH <sub>4</sub> Cal Offset:	0.011557	0.021954
NMHC Cal Slope:	0.998993	0.995866
NMHC Cal Offset:	0.018610	0.021609

Notes: Changed sample inlet filters after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 14, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:59	End Time (MST):	15:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999970	$\geq 0.995$			
17.13	17.02	1.0064						
8.56	8.56	1.0003				Slope	0.992346	0.90 - 1.10
4.28	4.34	0.9869						
			Intercept	0.043763	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 14, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:59	End Time (MST):	15:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999965	<b>≥0.995</b>
7.99	7.90	1.0105			
3.99	3.98	1.0030			
2.00	2.02	0.9904			
			Slope	0.988303	<b>0.90 - 1.10</b>
			Intercept	0.021954	<b>+/-0.5</b>







# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 14, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:59	End Time (MST):	15:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999973	$\geq 0.995$			
9.14	9.11	1.0028						
4.57	4.58	0.9982				Slope	0.995866	0.90 - 1.10
2.28	2.32	0.9839						
			Intercept	0.021609	$\pm 0.5$			



NMHC Calibration Plot

Date: January 15, 2024

Location: Conklin





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	January 25, 2024	Last Cal Date:	January 15, 2024
Start time (MST):	10:31	End time (MST):	12:58
Reason:	Cylinder Change		

### Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
C3H8 Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
Removed C3H8 Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.69E-04	3.86E-04	NMHC SP Ratio:	7.47E-05
CH4 Retention time:	17.2	17.4	NMHC Peak Area:	122336
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.33	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.29	0.990
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	0.990
Baseline Corr AF:	17.33	Prev response	17.11	*% change	1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.13	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.05	1.010
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.010
Baseline Corr AF:	9.13	Prev response	9.15	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	8.20	0.974
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	8.24	0.969
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.969
Baseline Corr AF:	8.20	Prev response	7.96	*% change	2.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997358	1.009622
THC Cal Offset:	0.029967	0.000000
CH <sub>4</sub> Cal Slope:	0.995501	1.031806
CH <sub>4</sub> Cal Offset:	0.011557	0.000000
NMHC Cal Slope:	0.998993	0.990237
NMHC Cal Offset:	0.018610	0.000000


Notes: Changed H2 cylinder after as founds. No adjustment made.

Calibration Performed By: Jan Castro

NMHC Calibration Plot

Date: January 25, 2024

Location: Conklin





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Conklin  
Calibration Date: January 19, 2024  
Start time (MST): 9:41  
Reason: Routine  
Station number: AMS21  
Last Cal Date: December 6, 2023  
End time (MST): 17:29

### Calibration Standards

NO Gas Cylinder #: SA18828  
NOX Cal Gas Conc: 48.80 ppm  
Removed Cylinder #: T2Y1P1H  
Removed Gas NOX Conc: 51.09 ppm  
NOX gas Diff: -0.1%  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701H  
Cal Gas Expiry Date: November 3, 2031  
NO Cal Gas Conc: 48.80 ppm  
Removed Gas Exp Date: December 11, 2023  
Removed Gas NO Conc: 50.39 ppm  
NO gas Diff: 0.0%  
Serial Number: 3810  
Serial Number: 364

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1501663731

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.077	1.062	NO bkgnd or offset:	10.8	10.6
NOX coeff or slope:	0.999	0.983	NOX bkgnd or offset:	10.9	10.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	151.8	152.1

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.992390	0.997935
NO <sub>x</sub> Cal Offset:	2.363089	2.088043
NO Cal Slope:	0.994482	0.999591
NO Cal Offset:	1.080383	1.508039
NO <sub>2</sub> Cal Slope:	0.991295	0.998906
NO <sub>2</sub> Cal Offset:	1.480919	0.745906



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.0	----	----
as found span	4921	79.4	811.2	800.1	11.1	833.1	816.8	16.4	0.9738	0.9796
as found 2nd										
as found 3rd										
new cyl resp	4918	82.0	800.3	800.3	0.0	821.0	817.0	4.3	0.9748	0.9796
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
high point	4918	82.0	800.3	800.3	0.0	800.0	801.0	-1.0	1.0004	0.9992
second point	4959	41.0	400.2	400.2	0.0	401.5	401.5	0.0	0.9967	0.9967
third point	4980	20.5	200.1	200.1	0.0	204.6	203.5	1.1	0.9778	0.9831
as left zero										
as left span										

Average Correction Factor | 0.9916 | 0.9930

Corrected As found	NO <sub>x</sub> = 833.3 ppb	NO = 817.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 3.1%
Previous Response	NO <sub>x</sub> = 807.4 ppb	NO = 796.8 ppb		*Percent Change	NO = 2.5%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI: ;

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.5	386.7	412.8	413.0	0.9995	100.0%
2nd GPT point (200 ppb O3)	799.5	595.7	203.8	203.8	1.0000	100.0%
3rd GPT point (100 ppb O3)	799.5	700.1	99.4	101.5	0.9793	102.1%
Average Correction Factor					0.9929	100.7%

Notes: Changed the inlet filter and NO<sub>x</sub> cylinder after as founds. Adjusted Span only. Used the 2nd reference point because of drift. Did not do as left because it will be late to finish.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 19, 2024	Previous Calibration:	December 6, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:41	End Time (MST):	17:29
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	800.0	1.0004		
400.2	401.5	0.9967		
200.1	204.6	0.9778		







# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 19, 2024	Previous Calibration:	December 6, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:41	End Time (MST):	17:29
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ $0.90 - 1.10$ $\pm 20$
800.3	801.0	0.9992		
400.2	401.5	0.9967		
200.1	203.5	0.9831		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 19, 2024	Previous Calibration:	December 6, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:41	End Time (MST):	17:29
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
412.8	413.0	0.9995		
203.8	203.8	1.0000		
99.4	101.5	0.9793		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: January 19, 2024

Location: Conklin





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Conklin  
Calibration Date: January 12, 2024  
Start time (MST): 11:25  
Reason: Routine  
Station number: AMS21  
Last Cal Date: December 5, 2023  
End time (MST): 14:35

### Calibration Standards

O3 generation mode: Photometer  
Calibrator Make/Model: Teledyne API T700  
ZAG Make/Model: Teledyne API 701H  
Serial Number: 3810  
Serial Number: 691

### Analyzer Information

Analyzer make: Thermo 49i  
Analyzer Range 0 - 500 ppb  
Analyzer serial #: 1501663734

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001229	1.000771	Backgd or Offset:	-1.2	-1.1
Calibration intercept:	0.260000	0.740000	Coeff or Slope:	0.999	1.005

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.2	----
high point	5000	950.9	400.0	400.8	0.998
second point	5000	804.0	200.0	201.1	0.995
third point	5000	703.6	100.0	101.4	0.986
as left zero	5000	0.0	0.0	0.8	----
as left span	5000	936.0	400.0	402.7	0.993
Average Correction Factor					0.993

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Pump DOA, replaced. Adjusted span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 12, 2024	Previous Calibration:	December 5, 2023
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:25	End Time (MST):	14:35
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999992	≥0.995
400.0	400.8	0.9980			
200.0	201.1	0.9945	Slope	1.000771	0.90 - 1.10
100.0	101.4	0.9862			
			Intercept	0.740000	+/- 5



O<sub>3</sub> Calibration Plot

Date: January 12, 2024

Location: Conklin





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Conklin Station number: AMS 21  
 Calibration Date: January 25, 2024 Last Cal Date: December 7, 2023  
 Start time (MST): 11:36 End time (MST): 11:54

Analyzer Make: API T640 S/N: 326  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: ALICAT S/N: 388754  
 Temp/RH standard: ALICAT S/N: 388754

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-8.20	-8.23	-8.20	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	705.00	706.53	705.00	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	5.08	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 25, 2024</u>	Last Cal Date: <u>December 7, 2023</u>			
	PM w/o HEPA: <u>37.30</u>	PM w/ HEPA: <u>0.00</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>December 7, 2023</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>December 7, 2023</u>			

### Annual Maintenance

Date Sample Tube Cleaned: December 7, 2023  
 Date RH/T Sensor Cleaned: December 7, 2023

Notes: Verified flow, temperature, and pressure. Leak check passed. No adjustment required.

Calibration by: Jan Castro



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS22  
JANVIER**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024







# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 5, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:50	End Time (MST):	13:14
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999999	
799.8	802.1	0.9971			≥0.995
399.9	401.7	0.9955	Slope	1.002679	
200.4	201.5	0.9947			0.90 - 1.10
			Intercept	0.383769	+/-30



SO2 Calibration Plot

Date: January 10, 2024

Location: Janvier





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name:	Janvier	Station number:	AMS22
Calibration Date:	January 24, 2024	Last Cal Date:	December 19, 2023
Start time (MST):	11:05	End time (MST):	15:22
Reason:	Cylinder Change		

### Calibration Standards

Cal Gas Concentration:	5.02	ppm	Cal Gas Exp Date:	April 16, 2022
Cal Gas Cylinder #:	CC424047			
Removed Cal Gas Conc:	5.03	ppm	Rem Gas Exp Date:	April 16, 2022
Removed Gas Cyl #:	DT0018680		Diff between cyl:	-1.3%
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031
Converter make:	CDN-101	Converter serial #:	587
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997936	0.999522	Backgd or Offset:	3.59	3.71
Calibration intercept:	0.260962	0.500639	Coeff or Slope:	1.179	1.205

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4920	79.5	80.0	79.3	1.012
as found 2nd point					
as found 3rd point					
new cylinder response	4920	79.7	80.0	78.3	1.022

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4920	79.7	80.0	80.4	0.995
second point	4960	39.8	40.0	40.6	0.984
third point	4980	19.9	20.0	20.5	0.975
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	79.7	80.0	79.6	1.005
SO2 Scrubber Check	4920	79.8	798.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.985
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	79.0	Prev response:	80.08	*% change:	-1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

\* = > +/-5% change initiates investigation

Notes: Swapped out the calibration gas after as founds. Changed the inlet filter after cylinder change. Ran a SO2 scrubber check after calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 24, 2024	Previous Calibration:	December 19, 2023
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	11:05	End Time (MST):	15:22
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999988	≥0.995
80.0	80.4	0.9953			
40.0	40.6	0.9843	Slope	0.999522	0.90 - 1.10
20.0	20.5	0.9746			
			Intercept	0.500639	+/-3



TRS Calibration Plot

Date: January 24, 2024

Location: Janvier





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name: Janvier Station number: AMS 22  
 Calibration Date: January 10, 2024 Last Cal Date: December 5, 2023  
 Start time (MST): 9:50 End time (MST): 13:14  
 Reason: Routine Here to address dipping baseline

### Calibration Standards

Gas Cert Reference: CC281519 Cal Gas Expiry Date: January 18, 2029  
 CH<sub>4</sub> Cal Gas Conc. 502.8 ppm CH<sub>4</sub> Equiv Conc. 1075.9 ppm  
 C<sub>3</sub>H<sub>8</sub> Cal Gas Conc. 208.4 ppm  
 Removed Gas Cert: Removed Gas Expiry:  
 Removed CH<sub>4</sub> Conc. 502.8 ppm CH<sub>4</sub> Equiv Conc. 1075.9 ppm  
 Removed C<sub>3</sub>H<sub>8</sub> Conc. 208.4 ppm Diff between cyl (THC):  
 Diff between cyl (CH<sub>4</sub>): Diff between cyl (NM):  
 Calibrator Model: Teledyne API 700 Serial Number: 3806  
 ZAG make/model: Teledyne API 701 Serial Number: 4890

### Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 1331259520  
 THC Range (ppm): 0 - 20 ppm  
 NMHC Range (ppm): 0 - 10 ppm CH<sub>4</sub> Range (ppm): 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.560E-04	3.030E-04	NMHC SP Ratio:	5.17E-05
CH <sub>4</sub> Retention time:	15.2	15.6	NMHC Peak Area:	176965
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4920	79.8	17.17	15.32	<b>1.121</b>
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.16	1.000
second point	4960	39.9	8.59	8.62	0.996
third point	4980	20.0	4.30	4.36	0.988
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.11	1.004
Average Correction Factor					0.995

Baseline Corr AF: 15.31 Prev response: 17.17 \*% change: **-12.2%**  
 Baseline Corr 2nd AF: NA AF Slope: AF Intercept:  
 Baseline Corr 3rd AF: NA AF Correlation: \* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	8.54	1.071
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.15	9.15	1.000
second point	4960	39.9	4.57	4.60	0.995
third point	4980	20.0	2.29	2.32	0.987
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.15	9.08	1.007
Average Correction Factor					0.994
Baseline Corr AF:	8.54	Prev response	9.15	*% change	-7.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4920	79.8	8.03	6.78	1.184
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.01	1.002
second point	4960	39.9	4.01	4.02	0.997
third point	4980	20.0	2.01	2.04	0.988
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.03	0.999
Average Correction Factor					0.996
Baseline Corr AF:	6.77	Prev response	8.02	*% change	-18.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998979	0.998667
THC Cal Offset:	0.014795	0.030390
CH <sub>4</sub> Cal Slope:	0.998920	0.997468
CH <sub>4</sub> Cal Offset:	0.005636	0.014235
NMHC Cal Slope:	0.999269	0.999032
NMHC Cal Offset:	0.008959	0.017157

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Jan Castro





# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 5, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:50	End Time (MST):	13:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999986	$\geq 0.995$			
17.17	17.16	1.0005						
8.59	8.62	0.9959				Slope	0.998667	0.90 - 1.10
4.30	4.36	0.9875						
			Intercept	0.030390	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 5, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:50	End Time (MST):	13:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999986	≥0.995			
8.03	8.01	1.0016						
4.01	4.02	0.9974				Slope	0.997468	0.90 - 1.10
2.01	2.04	0.9883						
			Intercept	0.014235	+/-0.5			





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 5, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:50	End Time (MST):	13:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999984	$\geq 0.995$
9.15	9.15	1.0001			
4.57	4.60	0.9949			
2.29	2.32	0.9868			
			Slope	0.999032	0.90 - 1.10
			Intercept	0.017157	+/-0.5



NMHC Calibration Plot

Date: January 10, 2024

Location: Janvier





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	January 17, 2024	Last Cal Date:	January 10, 2024
Start time (MST):	10:52	End time (MST):	12:11
Reason:	Removal		

### Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH <sub>4</sub> Cal Gas Conc.	502.8 ppm	CH <sub>4</sub> Equiv Conc.	1075.9 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	502.8 ppm	CH <sub>4</sub> Equiv Conc.	1075.9 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.030E-04	3.030E-04	NMHC SP Ratio:	5.65E-05	5.65E-05
CH <sub>4</sub> Retention time:	15.6	15.6	NMHC Peak Area:	164596	164596
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	18.45	0.931
as found 2nd point	4960	39.9	8.59	9.25	0.928
as found 3rd point	4980	20.0	4.30	4.67	0.922
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

			Average Correction Factor		
Baseline Corr AF:	18.45	Prev response	17.18	*% change	6.9%
Baseline Corr 2nd AF:	9.3	AF Slope:	1.073669	AF Intercept:	0.023477
Baseline Corr 3rd AF:	4.7	AF Correlation:	0.999992	* = +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	9.55	0.958
as found 2nd point	4960	39.9	4.57	4.79	0.954
as found 3rd point	4980	20.0	2.29	2.42	0.949
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	9.55	Prev response	9.16	*% change	4.1%
Baseline Corr 2nd AF:	4.8	AF Slope:	1.043421	AF Intercept:	0.012460
Baseline Corr 3rd AF:	2.4	AF Correlation:	0.999992	* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	8.89	0.903
as found 2nd point	4960	39.9	4.01	4.46	0.900
as found 3rd point	4980	20.0	2.01	2.25	<b>0.892</b>
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	8.89	Prev response	8.02	*% change	<b>9.8%</b>
Baseline Corr 2nd AF:	4.46	AF Slope:	1.106864	AF Intercept:	0.013020
Baseline Corr 3rd AF:	2.25	AF Correlation:	0.999990	* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998667	
THC Cal Offset:	0.030390	
CH <sub>4</sub> Cal Slope:	0.997468	
CH <sub>4</sub> Cal Offset:	0.014235	
NMHC Cal Slope:	0.999032	
NMHC Cal Offset:	0.017157	


Notes: Swapping out the instrument due to dips.

Calibration Performed By: Max Farrell

NMHC Calibration Plot

Date: January 17, 2024

Location: Janvier





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	January 17, 2024	Last Cal Date:	
Start time (MST):	12:28	End time (MST):	14:41
Reason:	Install		

### Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH <sub>4</sub> Cal Gas Conc.	502.8 ppm	CH <sub>4</sub> Equiv Conc.	1075.9 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	502.8 ppm	CH <sub>4</sub> Equiv Conc.	1075.9 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:		2.940E-04	NMHC SP Ratio:	6.33E-05
CH <sub>4</sub> Retention time:		13.2	NMHC Peak Area:	144559
Zero Chromatogram:		OFF	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.11	1.004
second point	4960	39.9	8.59	8.55	1.004
third point	4980	20.0	4.30	4.29	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.11	1.004
Average Correction Factor					1.003

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.15	9.09	1.006
second point	4960	39.9	4.57	4.55	1.005
third point	4980	20.0	2.29	2.30	0.999
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.15	9.10	1.005
Average Correction Factor					1.003
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.02	1.001
second point	4960	39.9	4.01	4.00	1.003
third point	4980	20.0	2.01	2.00	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.01	1.002
Average Correction Factor					1.003
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		0.996236
THC Cal Offset:		0.001413
CH <sub>4</sub> Cal Slope:		0.999459
CH <sub>4</sub> Cal Offset:		-0.005759
NMHC Cal Slope:		0.993433
NMHC Cal Offset:		0.007572

Notes: Swapped out the instrument due to dips and the H<sub>2</sub> cylinder as it was getting low. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:28	End Time (MST):	14:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	$\geq 0.995$
17.17	17.11	1.0036			
8.59	8.55	1.0042			
4.30	4.29	1.0022			
			Slope	0.996236	0.90 - 1.10
			Intercept	0.001413	+/-0.5





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:28	End Time (MST):	14:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	<b>≥0.995</b>
8.03	8.02	1.0009			
4.01	4.00	1.0029			
2.01	2.00	1.0061			
			Slope	0.999459	<b>0.90 - 1.10</b>
			Intercept	-0.005759	<b>+/-0.5</b>





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:28	End Time (MST):	14:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999996	$\geq 0.995$
9.15	9.09	1.0061			
4.57	4.55	1.0049			
2.29	2.30	0.9989			
			Slope	0.993433	0.90 - 1.10
			Intercept	0.007572	+/-0.5



NMHC Calibration Plot

Date: January 17, 2024

Location: Janvier





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Janvier Station number: AMS 22  
Calibration Date: January 15, 2024 Last Cal Date: December 15, 2023  
Start time (MST): 11:16 End time (MST): 16:10  
Reason: Routine

### Calibration Standards

NO Gas Cylinder #: DT0047765 Cal Gas Expiry Date: March 11, 2031  
NOX Cal Gas Conc: 48.90 ppm NO Cal Gas Conc: 48.80 ppm  
Removed Cylinder #: Removed Gas Exp Date:  
Removed Gas NOX Conc: 48.90 ppm Removed Gas NO Conc: 48.80 ppm  
NOX gas Diff: NO gas Diff:  
Calibrator Model: Teledyne API T700 Serial Number: 3806  
ZAG make/model: Teledyne API T701 Serial Number: 201

### Analyzer Information

Analyzer make: Analyzer serial #:  
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.834	0.854	NO bkgnd or offset:	-5.6	0.0
NOX coeff or slope:	0.827	0.849	NOX bkgnd or offset:	-3.9	-0.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	5.7	5.9

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.994409	1.000095
NO <sub>x</sub> Cal Offset:	2.604040	1.384206
NO Cal Slope:	0.998646	1.000674
NO Cal Offset:	1.684010	0.024014
NO <sub>2</sub> Cal Slope:	0.995534	1.008342
NO <sub>2</sub> Cal Offset:	-1.435399	1.498437



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.6	2.2	-0.5	----	----
as found span	4918	82.0	802.0	800.3	1.6	773.7	767.4	6.3	1.0365	1.0429
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	----	----
high point	4918	82.0	802.0	800.3	1.6	802.4	800.9	1.5	0.9995	0.9993
second point	4960	41.0	400.9	400.1	0.8	404.0	400.2	3.8	0.9923	0.9997
third point	4980	20.5	200.5	200.1	0.4	202.6	200.5	2.1	0.9895	0.9978
as left zero	5000	0.0	0.0	0.0	0.0	0.3	0.0	0.3	----	----
as left span	4918	82.0	802.0	424.6	377.3	797.1	406.4	390.7	1.0061	1.0448
Average Correction Factor									0.9938	0.9989

Corrected As found	NO <sub>x</sub> = 772.1 ppb	NO = 765.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -3.6%
Previous Response	NO <sub>x</sub> = 800.1 ppb	NO = 800.9 ppb		*Percent Change	NO = -4.7%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	802.4	426.7	377.3	381.0	0.9904	101.0%
2nd GPT point (200 ppb O3)	802.4	609.1	194.9	199.4	0.9776	102.3%
3rd GPT point (100 ppb O3)	802.4	706.7	97.3	100.8	0.9657	103.6%
Average Correction Factor					0.9779	102.3%

Notes: Changed the inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 15, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:16	End Time (MST):	16:10
Analyzer make:		Analyzer serial #:	

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
802.0	802.4	0.9995		
400.9	404.0	0.9923		
200.5	202.6	0.9895		







# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 15, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:16	End Time (MST):	16:10
Analyzer make:		Analyzer serial #:	

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ 0.90 - 1.10 +/-20
800.3	800.9	0.9993		
400.1	400.2	0.9997		
200.1	200.5	0.9978		

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 15, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:16	End Time (MST):	16:10
Analyzer make:		Analyzer serial #:	

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
377.3	381.0	0.9904		
194.9	199.4	0.9776		
97.3	100.8	0.9657		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: January 15, 2024

Location: Janvier





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

Version-01-2020

### Station Information

Station Name: Janvier Station number: AMS 22  
 Calibration Date: January 3, 2024 Last Cal Date: December 11, 2023  
 Start time (MST): 11:32 End time (MST): 14:21  
 Reason: Routine

### Calibration Standards

O3 generation mode: Photometer  
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806  
 ZAG Make/Model: Teledyne API T701H Serial Number: 691

### Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 7046  
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003171	1.005200	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	0.820000	0.740000	Coeff or Slope:	1.021	1.021

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.1	----
as found span	4895	905.3	400.0	403.6	0.991
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.5	----
high point	4895	905.3	400.0	402.8	0.993
second point	4895	756.7	200.0	201.6	0.992
third point	4895	656.1	100.0	101.7	0.983
as left zero	5000	800.0	0.0	0.6	----
as left span	4895	904.3	400.0	403.0	0.993
Average Correction Factor					0.989

Baseline Corr As found:	403.5	Previous response	402.1	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 3, 2024	Previous Calibration:	December 11, 2023
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:32	End Time (MST):	14:21
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999997	≥0.995
400.0	402.8	0.9930			
200.0	201.6	0.9921	Slope	1.005200	0.90 - 1.10
100.0	101.7	0.9833			
			Intercept	0.740000	+/- 5



O<sub>3</sub> Calibration Plot

Date: January 3, 2024

Location: Janvier





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Janvier Station number: AMS 22  
 Calibration Date: January 24, 2024 Last Cal Date: December 15, 2023  
 Start time (MST): 13:00 End time (MST): 14:00

Analyzer Make: Teledyne API T640 S/N: 325  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450  
 Temp/RH standard: Delta Cal S/N: 1450

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-4.40	-4.70	-4.40	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	712.5	713.0	712.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	5.10	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 24, 2024</u>	Last Cal Date: <u>December 15, 2023</u>			
	PM w/o HEPA: <u>3.9</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>January 24, 2024</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>January 24, 2024</u>			

### Annual Maintenance

Date Sample Tube Cleaned: July 26, 2023  
 Date RH/T Sensor Cleaned: July 26, 2023

Notes: Monthly and quarterly calibration completed. After cleaning the optical chamber pump did not turn back on. Will be back tomorrow to complete maintenance.

Calibration by: Max Farrell



# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Janvier Station number: AMS 22  
 Calibration Date: January 25, 2024 Last Cal Date: January 24, 2024  
 Start time (MST): 9:15 End time (MST): 9:42

Analyzer Make: Teledyne API T640 S/N: 325  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Delta Cal S/N: 1450  
 Temp/RH standard: Delta Cal S/N: 1450

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-4.40	-4.70	-4.40	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	712.50	713.00	712.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.02	5.10	5.02	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 25, 2024</u>	Last Cal Date: <u>January 24, 2024</u>			
	PM w/o HEPA: <u>3.90</u>	PM w/ HEPA: <u>0.00</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.10	10.10	10.90	<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>38.90</u>	w/ HEPA: <u>0.00</u>		
Date Optical Chamber Cleaned:		<u>January 24, 2024</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>January 24, 2024</u>			

### Annual Maintenance

Date Sample Tube Cleaned: July 26, 2023  
 Date RH/T Sensor Cleaned: July 26, 2023

#### Notes:

As found checks, optical chamber cleaning and filter change was completed on the 24th. Electrical connection was left disconnected by mistake which caused the pump to turn off, reconnected the line and the instrument is operating normal now. PMT adjustment completed and leak check passed.

Calibration by: Jan Castro





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Janvier Station number: AMS 22  
 Calibration Date: January 31, 2024 Last Cal Date: January 25, 2024  
 Start time (MST): 11:35 End time (MST): 12:11

Analyzer Make: Teledyne API T640 S/N: 325  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat S/N: 388752  
 Temp/RH standard: Alicat S/N: 388752

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	6.50	5.96	6.50	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	713.4	714.5	713.4	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.04	5.09	5.04	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 31, 2024</u>	Last Cal Date: <u>January 25, 2024</u>			
	PM w/o HEPA: <u>0.9</u>	PM w/ HEPA: <u>0</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input checked="" type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		<u>January 24, 2024</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>January 24, 2024</u>			

### Annual Maintenance

Date Sample Tube Cleaned: July 26, 2023  
 Date RH/T Sensor Cleaned: July 26, 2023

### Notes:

Verified flow, temperature, and pressure. Leak check passed.

Calibration by: Rene Chamberland



# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name: Janvier Station Number: AMS 22  
Calibration Date: November 16, 2023 Prev Cal Date: N/A  
Start Time (MST): 13:00 End Time (MST): 14:10  
Tower Height (m): 10.0 Reason: Install

### Wind Speed Information

Sensor make/model: Met One 010C-1 Serial Number: D16121  
WS Calibrator: MetOne 053 Serial Number: P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.5	-0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999999	≥0.9995
Calculated slope		0.999473	0.90 - 1.10
Calculated intercept		0.026227	+/- 2

### Wind Direction Information

Sensor make/model: Met One 020C-1 Serial Number: D14528  
As Found Declination (deg east of True North): N/A As Left Declination (deg east of True North): 14  
Solar noon time (MST): 13:25 Calc Declination\*: 13.02 Degrees  
Deadband calc: 4.1 degrees (*Limit 4 deg*) \* - calculated declination as per NOAA website

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	2.0	---
90	86.7	-0.9%
180	178.0	-0.6%
270	266.9	-0.9%
357	354.9	-0.6%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999924	≥0.9995
Calculated slope		1.008921	0.90 - 1.10
Calculated intercept		0.114807	+/- 4

Notes: Installing new WS/WD sensors for the 10 meter tower, no issues to note. Used a compass to line up the crossarm.

Calibration Performed By: Max Farrell, Jan Castro



## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS23  
FORT HILLS**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 13, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:08	End Time (MST):	10:58
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	0.999993	
799.1	798.5	1.0007			≥0.995
400.1	398.6	1.0036	Slope	0.999893	
200.0	198.0	1.0103			0.90 - 1.10
			Intercept	-0.983326	+/-30



SO2 Calibration Plot

Date: January 15, 2024

Location: Fort Hills





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Fort Hills Station number: AMS23  
 Calibration Date: January 9, 2024 Last Cal Date: December 6, 2023  
 Start time (MST): 8:55 End time (MST): 12:56  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.20 ppm Cal Gas Exp Date: February 5, 2024  
 Cal Gas Cylinder #: CC517372  
 Removed Cal Gas Conc: 5.20 ppm Rem Gas Exp Date: N/A  
 Removed Gas Cyl #: N/A Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 451  
 ZAG Make/Model: API T701 Serial Number: 5611

### Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1300156232  
 Converter make: CDN-101 Converter serial #: 594  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.991033	0.995462	Backgd or Offset: 1.19	1.19
Calibration intercept:	-0.198385	-0.178345	Coeff or Slope: 1.124	1.124

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4923	77.0	80.0	79.9	1.001
as found 2nd point	4962	38.5	40.0	39.8	1.005
as found 3rd point	4981	19.2	19.9	19.6	1.018
new cylinder response					

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.0	80.0	79.5	1.006
second point	4962	38.5	40.0	39.7	1.007
third point	4981	19.2	19.9	19.4	1.028
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.0	80.0	79.2	1.010
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.014
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 79.9 Prev response: 79.09 \*% change: 1.0%  
 Baseline Corr 2nd AF pt: 39.8 AF Slope: 0.999888 AF Intercept: -0.158204  
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999981

\* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information


Calibration Date:	January 9, 2024	Previous Calibration:	December 6, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:55	End Time (MST):	12:56
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999967	
80.0	79.5	1.0063			≥0.995
40.0	39.7	1.0075	Slope	0.995462	
19.9	19.4	1.0282			0.90 - 1.10
			Intercept	-0.178345	+/-3









# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	January 15, 2024	Last Cal Date:	December 13, 2023
Start time (MST):	8:08	End time (MST):	10:56
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH <sub>4</sub> Cal Gas Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1070.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH <sub>4</sub> Conc.	500.2 ppm	CH <sub>4</sub> Equiv Conc.	1070.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.36E-04	2.31E-04	NMHC SP Ratio:	5.09E-05
CH <sub>4</sub> Retention time:	13.0	13.0	NMHC Peak Area:	180787
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	17.69	0.972
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.33	0.992
second point	4960	40.2	8.61	8.64	0.996
third point	4980	20.1	4.30	4.34	0.992
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.19	17.32	0.993
Average Correction Factor					0.993

Baseline Corr AF:	17.69	Prev response	17.24	*% change	2.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-06-2022

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.50	0.964
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.22	0.993
second point	4960	40.2	4.59	4.62	0.993
third point	4980	20.1	2.29	2.35	0.975
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	9.16	9.19	0.997
Average Correction Factor					0.987
Baseline Corr AF:	9.50	Prev response	9.19	*% change	3.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	8.20	0.980
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.12	0.989
second point	4960	40.2	4.02	4.02	1.000
third point	4980	20.1	2.01	1.99	1.012
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.03	8.13	0.988
Average Correction Factor					1.001
Baseline Corr AF:	8.20	Prev response	8.05	*% change	1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001568	1.007788
THC Cal Offset:	0.017601	-0.007208
CH <sub>4</sub> Cal Slope:	1.004639	1.012164
CH <sub>4</sub> Cal Offset:	-0.020048	-0.027257
NMHC Cal Slope:	1.000371	1.005072
NMHC Cal Offset:	0.031648	0.018048

Notes: Span adjusted. No Maintenance done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 13, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:08	End Time (MST):	10:56
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999994	$\geq 0.995$			
17.19	17.33	0.9920						
8.61	8.64	0.9962				Slope	1.007788	0.90 - 1.10
4.30	4.34	0.9921						
			Intercept	-0.007208	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 13, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:08	End Time (MST):	10:56
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999945	≥0.995
8.03	8.12	0.9893			
4.02	4.02	1.0004	Slope	1.012164	0.90 - 1.10
2.01	1.99	1.0120			
			Intercept	-0.027257	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-06-2022

### Station Information

Calibration Date:	January 15, 2024	Previous Calibration:	December 13, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:08	End Time (MST):	10:56
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999974	$\geq 0.995$			
9.16	9.22	0.9934						
4.59	4.62	0.9925				Slope	1.005072	0.90 - 1.10
2.29	2.35	0.9752						
			Intercept	0.018048	$\pm 0.5$			



NMHC Calibration Plot

Date: January 15, 2024

Location: Fort Hills





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Fort Hills  
Calibration Date: January 10, 2024  
Start time (MST): 7:35  
Reason: Routine  
Station number: AMS23  
Last Cal Date: December 7, 2023  
End time (MST): 12:04

### Calibration Standards

NO Gas Cylinder #: CC332703  
NOX Cal Gas Conc: 49.7 ppm  
Removed Cylinder #: N/A  
Removed Gas NOX Conc: 49.7 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701  
Cal Gas Expiry Date: January 28, 2024  
NO Cal Gas Conc: 49.7 ppm  
Removed Gas Exp Date: N/A  
Removed Gas NO Conc: 49.7 ppm  
NO gas Diff:  
Serial Number: 451  
Serial Number: 5611

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1152430007

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.081	1.081	NO bkgnd or offset:	3	3
NOX coeff or slope:	0.989	0.989	NOX bkgnd or offset:	3.2	3.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	164.7	162

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.999822	0.998780
NO <sub>x</sub> Cal Offset:	0.304106	0.244003
NO Cal Slope:	1.001566	1.000209
NO Cal Offset:	-0.455910	-0.556171
NO <sub>2</sub> Cal Slope:	1.002916	0.999293
NO <sub>2</sub> Cal Offset:	-0.422416	-0.815351





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.2	----	----
as found span	4920	80.5	800.2	800.2	0.0	802.7	800.8	1.8	0.997	0.999
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
high point	4920	80.5	800.2	800.2	0.0	798.8	799.6	-0.8	1.002	1.001
second point	4960	40.2	399.6	399.6	0.0	400.8	400.1	0.8	0.997	0.999
third point	4980	20.1	199.8	199.8	0.0	199.4	198.0	1.5	1.002	1.009
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as left span	4920	80.5	800.2	445.5	354.7	796.4	441.9	354.4	1.005	1.008
Average Correction Factor									1.000	1.003

Corrected As found	NO <sub>x</sub> = 803.0 ppb	NO = 800.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.3%
Previous Response	NO <sub>x</sub> = 800.3 ppb	NO = 801.0 ppb		*Percent Change	NO = 0.0%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.0	442.3	354.7	353.9	1.002	99.8%
2nd GPT point (200 ppb O3)	797.0	621.8	175.2	174.2	1.006	99.4%
3rd GPT point (100 ppb O3)	797.0	706.1	90.9	89.1	1.020	98.0%
Average Correction Factor					1.009	99.1%

Notes:

No maintenance done. Span adjusted.

Calibration Performed By:

Melissa Lemay



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 7, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:35	End Time (MST):	12:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.2	798.8	1.0017		
399.6	400.8	0.9969		
199.8	199.4	1.0020		





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 7, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:35	End Time (MST):	12:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.2	799.6	1.0007		
399.6	400.1	0.9987		
199.8	198.0	1.0090		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 7, 2023
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:35	End Time (MST):	12:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
354.7	353.9	1.0023		
175.2	174.2	1.0057		
90.9	89.1	1.0202		



NO<sub>x</sub> Calibration Plot

Date: January 10, 2024

Location: Fort Hills





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Fort Hills Station number: AMS 23  
 Calibration Date: January 15, 2024 Last Cal Date: December 13, 2023  
 Start time (MST): 10:59 End time (MST): 11:25

Analyzer Make: API T640 S/N: 1546  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753  
 Temp/RH standard: Alicat FP-25BT S/N: 388753

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-29.4	-29.7	-29.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	748.5	747.9	748.5	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.06	5.00	<input type="checkbox"/>	+/- 0.25 LPM

Leak Test: Date of check: January 15, 2024 Last Cal Date: December 13, 2023  
 PM w/o HEPA: 6.8 PM w/ HEPA: 0 <0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Post-maintenance leak check: PM w/o HEPA: \_\_\_\_\_ w/ HEPA: \_\_\_\_\_  
 Date Optical Chamber Cleaned: November 20, 2023 <0.2 ug/m3  
 Disposable Filter Changed: November 20, 2023

### Annual Maintenance

Date Sample Tube Cleaned: October 17, 2023  
 Date RH/T Sensor Cleaned: October 17, 2023

No adjustments done. Leak check passed. Head cleaned.

Notes:

Calibration by: Melissa Lemay



## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS25**  
**WASKŌW OHCI PIMÂTISIWIN**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024







# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 22, 2024	Previous Calibration:	December 19, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	8:59	End Time (MST):	12:20
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	1.000000	
800.5	799.9	1.0008			≥0.995
400.3	400.2	1.0003	Slope	0.999234	
200.1	199.8	1.0017			0.90 - 1.10
			Intercept	0.023996	+/-30



SO2 Calibration Plot

Date: January 22, 2024

Location: Waskow ohci Pimatisiwin





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Waskow ohci Pimatisiwin      Station number: AMS25  
 Calibration Date: January 18, 2024      Last Cal Date: December 15, 2023  
 Start time (MST): 7:14      End time (MST): 12:04  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 4.97 ppm      Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC517099  
 Removed Cal Gas Conc: 4.97 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 747  
 ZAG Make/Model: API T701      Serial Number: 261

### Analyzer Information

Analyzer make: Thermo 43i-LTE      Analyzer serial #: 1170050146  
 Converter make: Global G-150      Converter serial #: 2022-219  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003251	1.002531	Backgd or Offset:	3.30	3.30
Calibration intercept:	0.200000	0.300000	Coeff or Slope:	1.108	1.125

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4920	80.0	79.5	78.8	1.010
as found 2nd point	4960	40.0	39.7	39.6	1.006
as found 3rd point	4980	20.0	19.9	19.8	1.008
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4920	80.0	79.5	79.9	0.994
second point	4960	40.0	39.7	40.3	0.986
third point	4980	20.0	19.9	20.1	0.988
as left zero	5000	0.0	0.0	0.3	----
as left span	4920	80.0	800.0	809.7	0.988
SO2 Scrubber Check	4921	79.2	800.0	0.0	----
Date of last scrubber change:	20-Jun-23			Ave Corr Factor	0.990
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 78.7      Prev response: 79.91      \*% change: -1.5%  
 Baseline Corr 2nd AF pt: 39.5      AF Slope: 0.990593      AF Intercept: 0.140000  
 Baseline Corr 3rd AF pt: 19.7      AF Correlation: 0.999995

\* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 15, 2023
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	7:14	End Time (MST):	12:04
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient	≥0.995
79.5	79.9	0.9944		
39.7	40.3	0.9858	Slope	0.90 - 1.10
19.9	20.1	0.9883		
			Intercept	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 18, 2024

Location: Waskow ohci Pimatisiwin





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS26  
CHRISTINA LAKE**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 23, 2024	Previous Calibration:	December 12, 2023
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	10:28	End Time (MST):	13:21
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999999	
800.9	799.3	1.0020			≥0.995
400.4	399.1	1.0033	Slope	0.997965	
200.2	199.9	1.0016			0.90 - 1.10
			Intercept	-0.124030	+/-30






SO2 Calibration Plot

Date: January 23, 2024

Location: Christina Lake





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Christina Lake      Station number: AMS26  
 Calibration Date: January 17, 2024      Last Cal Date: December 19, 2023  
 Start time (MST): 9:15      End time (MST): 14:15  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.05 ppm      Cal Gas Exp Date: November 15, 2026  
 Cal Gas Cylinder #: DT0014831  
 Removed Cal Gas Conc: 5.05 ppm      Rem Gas Exp Date:  
 Removed Gas Cyl #:      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3253  
 ZAG Make/Model: API T701H      Serial Number: 832

### Analyzer Information

Analyzer make: Thermo 450i      Analyzer serial #: 1180030032  
 Converter make: NA      Converter serial #: NA  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003426	1.001140	Backgd or Offset:	36.2      36.0
Calibration intercept:	-0.141605	0.238387	Coeff or Slope:	1.095      1.115

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.9	----
as found span	4921	79.2	80.0	78.9	1.002
as found 2nd point	4960	39.6	40.0	38.9	1.005
as found 3rd point	4980	19.8	20.0	19.4	0.985
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	79.2	80.0	80.2	0.997
second point	4960	39.6	40.0	40.5	0.988
third point	4980	19.8	20.0	20.3	0.985
as left zero	5000	0.0	0.0	0.1	----
as left span	4921	79.2	80.0	80.9	0.989
SO2 Scrubber Check	4931	80.9	807.1	-0.5	----
Date of last scrubber change:	27-Feb-19			Ave Corr Factor	0.990
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.8      Prev response: 80.12      \*% change: -0.4%  
 Baseline Corr 2nd AF pt: 39.8      AF Slope: 0.995996      AF Intercept: -0.781590  
 Baseline Corr 3rd AF pt: 20.3      AF Correlation: 0.999969

\* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SO2 scrubber check after calibrator zero done and passed. Adjusted zero and span.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	December 19, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	9:15	End Time (MST):	14:15
Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999977	≥0.995
80.0	80.2	0.9974			
40.0	40.5	0.9876	Slope	1.001140	0.90 - 1.10
20.0	20.3	0.9852			
			Intercept	0.238387	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 17, 2024

Location: Christina Lake







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.1	-0.6	----	----
as found span	4920	80.0	813.1	800.3	12.8	812.0	796.6	15.1	1.0014	1.0047
as found 2nd										
as found 3rd										
new cyl resp	4918	82.1	802.9	799.6	3.3	806.0	798.9	6.7	0.9962	1.0009
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.6	0.1	-0.7	----	----
high point	4918	82.1	802.9	799.6	3.3	803.0	801.0	2.0	0.9999	0.9983
second point	4959	41.1	401.9	400.3	1.6	402.7	401.6	1.1	0.9981	0.9968
third point	4980	20.5	200.5	199.7	0.8	201.8	200.6	1.1	0.9934	0.9953
as left zero	5000	0.0	0.0	0.0	0.0	-0.7	0.0	-0.7	----	----
as left span	4918	82.1	802.9	400.7	402.2	796.2	413.9	382.4	1.0084	0.9682
Average Correction Factor									0.9971	0.9968

Corrected As found	NO <sub>x</sub> = 812.6 ppb	NO = 796.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.3%
Previous Response	NO <sub>x</sub> = 810.4 ppb	NO = 796.5 ppb		*Percent Change	NO = 0.0%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	800.5	401.6	402.2	400.4	1.0045	99.6%
2nd GPT point (200 ppb O3)	800.5	606.9	196.9	195.6	1.0066	99.3%
3rd GPT point (100 ppb O3)	800.5	704.9	98.9	97.9	1.0101	99.0%
Average Correction Factor					1.0070	99.3%

Notes: Changed sample inlet filter and NO<sub>x</sub> cylinder after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	December 13, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	11:17	End Time (MST):	16:56
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
802.9	803.0	0.9999		
401.9	402.7	0.9981		
200.5	201.8	0.9934		
			0.999994	
			1.000236	
			0.306500	





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	December 13, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	11:17	End Time (MST):	16:56
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
799.6	801.0	0.9983			
400.3	401.6	0.9968			
199.7	200.6	0.9953			
			Slope	1.001429	0.90 - 1.10
			Intercept	0.426316	+/-20

NO Calibration Curve







# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	December 13, 2023
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	11:17	End Time (MST):	16:56
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.7	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
402.2	400.4	1.0045		
196.9	195.6	1.0066		
98.9	97.9	1.0101		


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: January 16, 2024

Location: Christina Lake





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS27  
JACKFISH 2/3**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	January 12, 2024	Last Cal Date:	December 12, 2023
Start time (MST):	14:46	End time (MST):	17:35
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	<u>SG9133974BAL</u>			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3811
ZAG Make/Model:	API 701		Serial Number:	268

### Analyzer Information

Analyzer make: Thermo 43iQ  
 Analyzer Range 0 - 1000 ppb  
 Analyzer serial #: 12124313138

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.989063	1.003874	Backgd or Offset:	7.6	7.8
Calibration intercept:	-2.298638	-2.957498	Coeff or Slope:	0.950	0.977

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.7	----
as found span	4921	79.1	800.2	780.7	1.025
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.9	----
high point	4921	79.1	800.2	801.8	0.998
second point	4961	39.5	399.5	397.5	1.005
third point	4980	19.8	200.3	193.4	1.036
as left zero	5000	0.0	0.0	0.9	----
as left span	4921	79.1	800.2	804.6	0.994
Average Correction Factor					1.013

Baseline Corr As found:	780.00	Previous response	789.11	*% change	-1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 12, 2024	Previous Calibration:	December 12, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	14:46	End Time (MST):	17:35
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12124313138

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.9	----	Correlation Coefficient	0.999887	≥0.995
800.2	801.8	0.9980			
399.5	397.5	1.0051	Slope	1.003874	0.90 - 1.10
200.3	193.4	1.0357			
			Intercept	-2.957498	+/-30



SO2 Calibration Plot

Date: January 12, 2024

Location: Jackfish 2/3





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	January 27, 2024	Last Cal Date:	NA
Start time (MST):	12:18	End time (MST):	14:27
Reason:	Install		

### Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	<u>SG9133974BAL</u>			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3811
ZAG Make/Model:	API 701		Serial Number:	268

### Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169745
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.001831	Backgd or Offset:	NA	2.4
Calibration intercept:	NA	-0.817522	Coeff or Slope:	NA	0.774

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4921	79.1	800.2	801.0	0.999
second point	4961	39.5	399.5	400.1	0.999
third point	4980	19.8	200.3	197.8	1.013
as left zero	5000	0.0	0.0	0.2	----
as left span	4921	79.1	800.2	798.1	1.003
Average Correction Factor					1.003

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Installation calibrations. Adjusted span only.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 27, 2024	Previous Calibration:	NA
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	12:18	End Time (MST):	14:27
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169745

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient	≥0.995
800.2	801.0	0.9990		
399.5	400.1	0.9986	Slope	0.90 - 1.10
200.3	197.8	1.0127		
			Intercept	+/-30






SO2 Calibration Plot

Date: January 27, 2024

Location: Jackfish 2/3





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Jackfish 2/3      Station number: AMS27  
 Calibration Date: January 12, 2024      Last Cal Date: December 11, 2023  
 Start time (MST): 8:21      End time (MST): 14:45  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.41 ppm      Cal Gas Exp Date: January 4, 2025  
 Cal Gas Cylinder #: CC345023  
 Removed Cal Gas Conc: 5.41 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA      Diff between cyl:  
 Calibrator Make/Model: API T700      Serial Number: 3811  
 ZAG Make/Model: API 701      Serial Number: 268

### Analyzer Information

Analyzer make: API T101      Analyzer serial #: 621  
 Converter make:      Converter serial #:  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007470	1.010462	Backgd or Offset: 29.9	29.9
Calibration intercept:	-0.217773	0.002360	Coeff or Slope: 0.953	1.013

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	4926	74.1	80.2	76.0	1.049
as found 2nd point	4963	37.0	40.0	37.9	1.045
as found 3rd point	4982	18.5	20.0	18.7	1.048
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4926	74.1	80.2	81.0	0.990
second point	4963	37.0	40.0	40.7	0.984
third point	4982	18.5	20.0	19.8	1.011
as left zero	5000	0.0	0.0	0.1	----
as left span	4926	74.1	80.2	80.6	0.995
SO2 Scrubber Check	4921	79.1	791.0	-0.1	----

Date of last scrubber change:	Ave Corr Factor	0.995
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 76.4      Prev response: 80.56      \*% change: -5.4%  
 Baseline Corr 2nd AF pt: 38.3      AF Slope: 0.953017      AF Intercept: -0.358868  
 Baseline Corr 3rd AF pt: 19.1      AF Correlation: 0.999995

\* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. Adjusted span only.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 12, 2024	Previous Calibration:	December 11, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	8:21	End Time (MST):	14:45
Analyzer make:	API T101	Analyzer serial #:	621

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999922	≥0.995
80.2	81.0	0.9898			
40.0	40.7	0.9836	Slope	1.010462	0.90 - 1.10
20.0	19.8	1.0109			
			Intercept	0.002360	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 12, 2024

Location: Jackfish 2/3







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.3	-0.4	----	----
as found span	4921	79.4	816.8	800.3	16.5	809.0	793.9	15.1	1.0096	1.0080
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.4	-0.6	----	----
high point	4921	79.4	816.8	800.3	16.5	808.9	798.3	10.6	1.0098	1.0025
second point	4960	39.7	408.5	400.2	8.3	399.8	392.2	7.6	1.0217	1.0204
third point	4980	19.8	203.7	199.6	4.1	195.4	192.8	2.6	1.0425	1.0352
as left zero	5000	0.0	0.0	0.0	0.0	0.9	0.9	0.0	----	----
as left span	4921	79.4	816.8	430.6	402.5	800.1	414.0	386.2	1.0209	1.0401
Average Correction Factor									1.0247	1.0194

Corrected As found	NO <sub>x</sub> = 809.7 ppb	NO = 794.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.2%
Previous Response	NO <sub>x</sub> = 811.5 ppb	NO = 794.6 ppb		*Percent Change	NO = 0.0%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.7	410.7	402.5	393.4	1.0232	97.7%
2nd GPT point (200 ppb O3)	796.7	620.8	192.4	185.7	1.0362	96.5%
3rd GPT point (100 ppb O3)	796.7	711.2	102.0	97.9	1.0420	96.0%
Average Correction Factor					1.0338	96.7%

Notes: Changed sample inlet filter after as founds. No adjustments made. Third GPT point barely passed intending to investigate further.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 12, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	12:17	End Time (MST):	17:58
Analyzer make:	API T200	Analyzer serial #:	722

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
816.8	808.9	1.0098		
408.5	399.8	1.0217		
203.7	195.4	1.0425		
			0.999917	
			0.992430	
			-3.538303	





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 12, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	12:17	End Time (MST):	17:58
Analyzer make:	API T200	Analyzer serial #:	722

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	≥0.995	
800.3	798.3	1.0025			
400.2	392.2	1.0204			
199.6	192.8	1.0352			
			Slope	0.998816	0.90 - 1.10
			Intercept	-3.680579	+/-20

NO Calibration Curve







# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 12, 2023
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	12:17	End Time (MST):	17:58
Analyzer make:	API T200	Analyzer serial #:	722

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.6	----	Correlation Coefficient	≥0.995	
402.5	393.4	1.0232			
192.4	185.7	1.0362	Slope	0.90 - 1.10	
102.0	97.9	1.0420			
			Intercept	-1.570350	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: January 11, 2024

Location: Jackfish 2/3





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS29  
SURMONT 2**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 1, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	11:06	End Time (MST):	15:58
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999986	≥0.995
800.1	800.2	0.9999			
400.6	399.4	1.0030	Slope	1.001159	0.90 - 1.10
199.8	197.0	1.0143			
			Intercept	-1.385825	+/-30



SO2 Calibration Plot

Date: January 5, 2024

Location: Surmont 2







# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 3, 2024	Previous Calibration:	December 4, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:42	End Time (MST):	15:02
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999991	≥0.995
80.0	80.3	0.9963			
40.1	40.2	0.9977	Slope	1.004616	0.90 - 1.10
20.1	19.9	1.0077			
			Intercept	-0.102938	+/-3






H<sub>2</sub>S Calibration Plot

Date: January 3, 2024

Location: Surmont 2





# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	January 5, 2024	Last Cal Date:	December 1, 2023
Start time (MST):	11:06	End time (MST):	15:58
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC356008	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
C3H8 Cal Gas Conc.	<u>205.7</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
Removed C3H8 Conc.	<u>205.7</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5472
ZAG Make/Model:	Teledyne API T701	Serial Number:	4297

### Analyzer Information

Analyzer make: Thermo 51i-LT                      Analyzer serial #: 1170050149  
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007452	1.001735	Background:	3.45	3.62
Calibration intercept:	-0.027439	-0.005651	Coefficient:	4.017	4.021

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.12	----
as found span	4918	81.3	17.31	17.60	0.984
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.02	----
high point	4918	81.3	17.31	17.36	0.997
second point	4959	40.6	8.65	8.62	1.003
third point	4979	20.3	4.32	4.31	1.003
as left zero	5000	0.0	0.00	0.13	----
as left span	4918	81.3	17.31	17.66	0.980
Average Correction Factor					1.001
Baseline Corr As found:	17.48	Previous response	17.42	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Adjusted zero and span. Excessive noise fixed after reconnecting ethernet cable.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 5, 2024	Previous Calibration:	December 1, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	11:06	End Time (MST):	15:58
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.02	----	Correlation Coefficient	0.999983	
17.31	17.36	0.9974			≥0.995
8.65	8.62	1.0030	Slope	1.001735	
4.32	4.31	1.0033			0.90 - 1.10
			Intercept	-0.005651	+/-1.5



THC Calibration Plot

Date: January 5, 2024

Location: Surmont 2







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
as found span	4916	84.2	799.2	799.2	0.0	793.5	790.0	3.5	1.0072	1.0116
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4916	84.2	799.2	799.2	0.0	797.0	796.0	0.9	1.0028	1.0040
second point	4958	42.1	399.6	399.6	0.0	400.7	399.6	1.1	0.9973	1.0000
third point	4979	21.1	200.3	200.3	0.0	198.1	196.8	1.4	1.0110	1.0177
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as left span	4916	84.2	799.2	407.2	392.0	792.6	401.0	391.6	1.0083	1.0154
Average Correction Factor									1.0037	1.0072

Corrected As found	NO <sub>x</sub> = 793.7 ppb	NO = 790.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -0.6%	
Previous Response	NO <sub>x</sub> = 798.5 ppb	NO = 798.1 ppb		*Percent Change	NO = -1.0%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	792.4	400.4	392.0	394.3	0.9942	100.6%
2nd GPT point (200 ppb O3)	792.4	602.1	190.3	190.0	1.0016	99.8%
3rd GPT point (100 ppb O3)	792.4	698.6	93.8	95.4	0.9832	101.7%
Average Correction Factor					0.9930	100.7%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 5, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:46	End Time (MST):	16:24
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.2	797.0	1.0028		
399.6	400.7	0.9973		
200.3	198.1	1.0110		





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 5, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:46	End Time (MST):	16:24
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
799.2	796.0	1.0040			
399.6	399.6	1.0000			
200.3	196.8	1.0177			
			Slope	0.997476	0.90 - 1.10
			Intercept	-0.811525	+/-20

**NO Calibration Curve**







# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 4, 2024	Previous Calibration:	December 5, 2023
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:46	End Time (MST):	16:24
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
392.0	394.3	0.9942			
190.3	190.0	1.0016			
93.8	95.4	0.9832			
			Slope	1.004575	0.90 - 1.10
			Intercept	0.126759	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: January 4, 2024

Location: Surmont 2







## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS30  
ELLS RIVER**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	January 3, 2024	Last Cal Date:	December 1, 2023
Start time (MST):	12:27	End time (MST):	15:33
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.53	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC494126			
Removed Cal Gas Conc:	50.53	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3061
ZAG Make/Model:	API T701H		Serial Number:	358

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999175	1.001174	Backgd or Offset:	9.5	9.5
Calibration intercept:	-2.535883	-1.736008	Coeff or Slope:	0.982	0.982

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4921	79.2	800.4	800.1	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	79.2	800.4	800.0	1.000
second point	4960	39.6	400.2	399.5	1.002
third point	4980	19.8	200.1	195.8	1.022
as left zero	5000	0.0	0.0	-0.1	----
as left span	4921	79.2	800.4	801.7	0.998
Average Correction Factor					1.008

Baseline Corr As found:	800.40	Previous response	797.17	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Inlet filter changed after As Found, no adjustments made.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 3, 2024	Previous Calibration:	December 1, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	12:27	End Time (MST):	15:33
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999967	
800.4	800.0	1.0005			≥0.995
400.2	399.5	1.0018	Slope	1.001174	
200.1	195.8	1.0220			0.90 - 1.10
			Intercept	-1.736008	+/-30



SO2 Calibration Plot

Date: January 3, 2024

Location: Ells River





# Wood Buffalo Environmental Association

## TRS Calibration Report

Version-11-2021

### Station Information

Station Name: Ells River Station number: AMS30  
 Calibration Date: January 18, 2024 Last Cal Date: December 15, 2023  
 Start time (MST): 10:40 End time (MST): 15:31  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 4.99 ppm Cal Gas Exp Date: November 15, 2026  
 Cal Gas Cylinder #: CC505806  
 Removed Cal Gas Conc: 5.08 ppm Rem Gas Exp Date: February 9, 2024  
 Removed Gas Cyl #: EY0002443 Diff between cyl: 2.3%  
 Calibrator Make/Model: API T700 Serial Number: 3061  
 ZAG Make/Model: API T701H Serial Number: 358

### Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1410661331  
 Converter make: CDN - 101 Converter serial #: 562  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000065	1.005472	Backgd or Offset:	1.57	2.05
Calibration intercept:	0.020811	-0.120388	Coeff or Slope:	1.100	1.072

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	78.7	80.0	80.3	0.996
as found 2nd point					
as found 3rd point					
new cylinder response	4920	80.2	80.0	82.3	0.972

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	80.2	80.0	80.4	0.995
second point	4960	40.1	40.0	40.0	1.000
third point	4980	20.0	20.0	20.0	0.998
as left zero	5000	0.0	0.0	0.4	----
as left span	4920	80.2	80.0	80.0	1.000
SO2 Scrubber Check	4921	79.2	800.4	-0.1	----
Date of last scrubber change:	N/A			Ave Corr Factor	0.998
Date of last converter efficiency test:	N/A			efficiency	

Baseline Corr As found:	80.3	Prev response:	79.99	*% change:	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

\* = > +/-5% change initiates investigation

Notes: Change inlet filters and H2S cylinder after as founds. SOx scrubber check done after calibrator zero, passed. Adjusted zero and span.

Calibration Performed By: Jan Castro





# Wood Buffalo Environmental Association

## TRS Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 15, 2023
Station Name:	Ells River	Station Number:	AMS30
Start Time (MST):	10:40	End Time (MST):	15:31
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999995	≥0.995
80.0	80.4	0.9955			
40.0	40.0	1.0005	Slope	1.005472	0.90 - 1.10
20.0	20.0	0.9980			
			Intercept	-0.120388	+/-3



TRS Calibration Plot

Date: January 18, 2024

Location: Ells River





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	January 3, 2024	Last Cal Date:	December 1, 2023
Start time (MST):	12:27	End time (MST):	15:33
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH <sub>4</sub> Cal Gas Conc.	499.7 ppm	CH <sub>4</sub> Equiv Conc.	1075.0 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	499.7 ppm	CH <sub>4</sub> Equiv Conc.	1075.0 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	209.2 ppm		
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701H	Serial Number:	3061
		Serial Number:	358

### Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH <sub>4</sub> Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.48E-04	2.51E-04	NMHC SP Ratio:	5.80E-05
CH <sub>4</sub> Retention time:	14.2	14.2	NMHC Peak Area:	157070
				151854

### THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.73	1.018
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.07	0.997
second point	4960	39.6	8.51	8.49	1.003
third point	4980	19.8	4.26	4.19	1.015
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.05	0.999

Average Correction Factor				1.005
Baseline Corr AF:	16.73	Prev response	16.93	*% change -1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

Version-01-2020

### NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	8.90	1.024
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.17	0.994
second point	4960	39.6	4.56	4.58	0.995
third point	4980	19.8	2.28	2.27	1.006
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.13	0.998
Average Correction Factor					0.998
Baseline Corr AF:	8.90	Prev response	9.09	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.83	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.90	1.002
second point	4960	39.6	3.96	3.91	1.012
third point	4980	19.8	1.98	1.93	1.026
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.92	1.000
Average Correction Factor					1.013
Baseline Corr AF:	7.83	Prev response	7.84	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.996248	1.003892
THC Cal Offset:	-0.036338	-0.039540
CH <sub>4</sub> Cal Slope:	0.992583	0.999470
CH <sub>4</sub> Cal Offset:	-0.016756	-0.026357
NMHC Cal Slope:	0.999405	1.007395
NMHC Cal Offset:	-0.019981	-0.012583

Notes: Inlet filter changed after As Finds, adjusted span only.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 3, 2024	Previous Calibration:	December 1, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	12:27	End Time (MST):	15:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999975	$\geq 0.995$			
17.03	17.07	0.9973						
8.51	8.49	1.0028				Slope	1.003892	0.90 - 1.10
4.26	4.19	1.0153						
			Intercept	-0.039540	$\pm 0.5$			





# Wood Buffalo Environmental Association

## CH<sub>4</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 3, 2024	Previous Calibration:	December 1, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	12:27	End Time (MST):	15:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999948	≥0.995
7.91	7.90	1.0020			
3.96	3.91	1.0117			
1.98	1.93	1.0264			
			Slope	0.999470	0.90 - 1.10
			Intercept	-0.026357	+/-0.5





# Wood Buffalo Environmental Association

## NMHC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 3, 2024	Previous Calibration:	December 1, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	12:27	End Time (MST):	15:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999990	$\geq 0.995$			
9.11	9.17	0.9936						
4.56	4.58	0.9951				Slope	1.007395	0.90 - 1.10
2.28	2.27	1.0059						
			Intercept	-0.012583	$\pm 0.5$			



NMHC Calibration Plot

Date: January 3, 2024

Location: Ells River







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Ells River  
Calibration Date: January 11, 2024  
Start time (MST): 10:56  
Reason: Routine  
Station number: AMS 30  
Last Cal Date: December 18, 2023  
End time (MST): 16:37

### Calibration Standards

NO Gas Cylinder #: DT0045703  
NOX Cal Gas Conc: 60.4 ppm  
Removed Cylinder #: T2Y1P2R  
Removed Gas NOX Conc: 50.83 ppm  
NOX gas Diff: -1.0%  
Calibrator Model: API T700  
ZAG make/model: API T701H  
Cal Gas Expiry Date: May 16, 2031  
NO Cal Gas Conc: 60.1 ppm  
Removed Gas Exp Date: December 11, 2023  
Removed Gas NO Conc: 49.97 ppm  
NO gas Diff: -0.4%  
Serial Number: 3061  
Serial Number: 358

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 710321429

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.153	1.165	NO bkgnd or offset:	14.1	13.6
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	14.1	13.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	189.7	194.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000506	1.002653
NO <sub>x</sub> Cal Offset:	-1.440000	-1.638433
NO Cal Slope:	0.998642	1.007289
NO Cal Offset:	-2.440000	-2.359365
NO <sub>2</sub> Cal Slope:	1.005022	0.981718
NO <sub>2</sub> Cal Offset:	0.180492	0.452475



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.9	-0.9	0.0	----	----
as found span	4920	80.0	813.3	799.5	13.8	812.0	795.9	16.5	1.0016	1.0045
as found 2nd										
as found 3rd										
new cyl resp	4933	66.6	804.5	800.5	4.0	795.6	793.7	1.8	1.0112	1.0086
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4933	66.6	804.5	800.5	4.0	806.0	805.0	1.2	0.9982	0.9944
second point	4967	33.3	402.3	400.3	2.0	400.3	400.1	0.2	1.0049	1.0004
third point	4983	16.6	200.5	199.5	1.0	198.3	196.1	2.1	1.0112	1.0175
as left zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
as left span	4933	66.6	804.5	425.4	379.1	801.0	427.6	373.6	1.0044	0.9949
Average Correction Factor									1.0048	1.0041

Corrected As found	NO <sub>x</sub> = 812.9 ppb	NO = 796.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.1%
Previous Response	NO <sub>x</sub> = 812.3 ppb	NO = 796.0 ppb		*Percent Change	NO = 0.1%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.9	423.8	379.1	371.8	1.0196	98.1%
2nd GPT point (200 ppb O3)	798.9	612.1	190.8	189.8	1.0052	99.5%
3rd GPT point (100 ppb O3)	798.9	700.1	102.8	100.6	1.0218	97.9%
Average Correction Factor					1.0156	98.5%

Notes: Swapped NO gas cylinder#: T2Y1P2R with cylinder#: DT0045703. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 18, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:56	End Time (MST):	16:37
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
804.5	806.0	0.9982		
402.3	400.3	1.0049		
200.5	198.3	1.0112		





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 18, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:56	End Time (MST):	16:37
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ <i>0.90 - 1.10</i> <i>+/-20</i>
800.5	805.0	0.9944		
400.3	400.1	1.0004		
199.5	196.1	1.0175		
			0.999963	
			1.007289	
			-2.359365	

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 18, 2023
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:56	End Time (MST):	16:37
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
379.1	371.8	1.0196			
190.8	189.8	1.0052			
102.8	100.6	1.0218			
			Slope	0.981718	0.90 - 1.10
			Intercept	0.452475	+/-20



NO<sub>x</sub> Calibration Plot

Date: January 11, 2024

Location: Ells River





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2023

### Station Information

Station Name: Ells River Station number: AMS 30  
 Calibration Date: January 22, 2024 Last Cal Date: December 18, 2023  
 Start time (MST): 10:36 End time (MST): 12:14

Analyzer Make: API T640 S/N: 875  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388751  
 Temp/RH standard: Alicat FP-25BT S/N: 388751

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-22.30	-23.40	-22.30	<input type="checkbox"/>	+/- 2 deg C
P (mmHg)	735.20	736.60	735.20	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	5.00	5.00	5.00	<input type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: <u>January 22, 2024</u>	Last Cal Date: <u>December 18, 2023</u>			
	PM w/o HEPA: <u>5.30</u>	PM w/ HEPA: <u>0.00</u>			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

### Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.00	11.00	11.00	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: <u>5.30</u>	w/ HEPA: <u>0.00</u>		
Date Optical Chamber Cleaned:		<u>January 22, 2024</u>			<0.2 ug/m3
Disposable Filter Changed:		<u>January 22, 2024</u>			

### Annual Maintenance

Date Sample Tube Cleaned: October 27, 2023  
 Date RH/T Sensor Cleaned: November 30, 2023

Notes:

Verified flow, temperature and pressure. Leak check passed. No adjustment made.

Calibration by: Jan Castro



## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS501  
LEISMER**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024







# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 16, 2024	Previous Calibration:	December 12, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	10:02	End Time (MST):	13:10
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999977	≥0.995
800.2	799.2	1.0013			
400.2	399.2	1.0024	Slope	1.000572	0.90 - 1.10
200.1	196.1	1.0202			
			Intercept	-1.756005	+/-30



SO2 Calibration Plot

Date: January 16, 2024

Location: Leismer





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Leismer Station number: AMS501  
 Calibration Date: January 23, 2024 Last Cal Date: December 12, 2023  
 Start time (MST): 9:43 End time (MST): 13:48  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.14 ppm Cal Gas Exp Date: September 16, 2024  
 Cal Gas Cylinder #: CC511843  
 Removed Cal Gas Conc: 5.14 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 2659  
 ZAG Make/Model: API 701 Serial Number: 4427

### Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1180540020  
 Converter make: Global G150 Converter serial #: 2022-218  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995716	0.999147	Backgd or Offset:	3.47
Calibration intercept:	-0.078455	0.001539	Coeff or Slope:	1.105
				1.128

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4922	77.8	80.0	77.5	1.031
as found 2nd point	4961	38.9	40.0	38.9	1.025
as found 3rd point	4981	19.4	19.9	19.3	1.028
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4922	77.8	80.0	79.9	1.001
second point	4961	38.9	40.0	40.1	0.997
third point	4981	19.4	19.9	19.7	1.012
as left zero	5000	0.0	0.0	0.2	----
as left span	4922	77.8	80.0	79.9	1.001
SO2 Scrubber Check	4921	79.2	792.0	0.0	----
Date of last scrubber change:	24-Feb-23			Ave Corr Factor	1.004
Date of last converter efficiency test:	December 1, 2022			efficiency	

Baseline Corr As found: 77.6 Prev response: 79.56 \*% change: -2.5%  
 Baseline Corr 2nd AF pt: 39.0 AF Slope: 0.970286 AF Intercept: -0.038970  
 Baseline Corr 3rd AF pt: 19.4 AF Correlation: 0.999992

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Scrubber test done after calibrator zero. Adjusted span only.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 23, 2024	Previous Calibration:	December 12, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:43	End Time (MST):	13:48
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999977	≥0.995
80.0	79.9	1.0010			
40.0	40.1	0.9973	Slope	0.999147	0.90 - 1.10
19.9	19.7	1.0123			
			Intercept	0.001539	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 23, 2024

Location: Leismer





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Leismer  
Calibration Date: January 24, 2024  
Start time (MST): 9:41  
Reason: Removal  
Station number: AMS501  
Last Cal Date: December 13, 2023  
End time (MST): 14:15

### Calibration Standards

NO Gas Cylinder #: T26811M  
NOX Cal Gas Conc: 47.46 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 47.46 ppm  
NOX gas Diff:  
Calibrator Model: API T700  
ZAG make/model: API 701  
Cal Gas Expiry Date: October 30, 2024  
NO Cal Gas Conc: 47.39 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 47.39 ppm  
NO gas Diff:  
Serial Number: 2659  
Serial Number: 4427

### Analyzer Information

Analyzer make: Thermo 42i  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 1218153356

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.238	1.186	NO bkgnd or offset:	3.7	3.5
NOX coeff or slope:	0.990	0.991	NOX bkgnd or offset:	3.6	3.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	166.2	166.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.001988	1.001018
NO <sub>x</sub> Cal Offset:	-4.067947	-1.427986
NO Cal Slope:	1.004511	1.002011
NO Cal Offset:	-4.907940	-2.907962
NO <sub>2</sub> Cal Slope:	0.999993	1.000174
NO <sub>2</sub> Cal Offset:	0.065389	0.184957



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.6	-0.1	0.7	----	----
as found span	4916	84.4	801.1	799.9	1.2	839.5	836.1	3.7	0.9542	0.9567
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.4	0.0	0.4	----	----
high point	4916	84.4	801.1	799.9	1.2	801.2	799.8	1.5	0.9998	1.0001
second point	4958	42.2	400.5	400.0	0.6	399.1	397.1	2.0	1.0036	1.0072
third point	4979	21.1	200.3	200.0	0.3	196.9	194.1	2.8	1.0172	1.0303
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
as left span	4916	84.4	801.1	412.8	388.3	797.1	412.1	384.9	1.0050	1.0016
Average Correction Factor									1.0069	1.0125

Corrected As found	NO <sub>x</sub> = 838.9 ppb	NO = 836.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 4.8%
Previous Response	NO <sub>x</sub> = 798.6 ppb	NO = 798.6 ppb		*Percent Change	NO = 4.5%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.6	409.5	388.3	388.6	0.9992	100.1%
2nd GPT point (200 ppb O3)	796.6	612.2	185.6	185.8	0.9988	100.1%
3rd GPT point (100 ppb O3)	796.6	706.1	91.7	91.6	1.0009	99.9%
Average Correction Factor					0.9996	100.0%

Notes: Changed inlet filter after as founds. Adjusted span. 3rd point was triggered after high point to check linearity.

Calibration Performed By: Sean Bala





# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 24, 2024	Previous Calibration:	December 13, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:41	End Time (MST):	14:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
801.1	801.2	0.9998		
400.5	399.1	1.0036		
200.3	196.9	1.0172		
			0.999975	
			1.001018	
			-1.427986	





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 24, 2024	Previous Calibration:	December 13, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:41	End Time (MST):	14:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
799.9	799.8	1.0001			0.999938
400.0	397.1	1.0072			1.002011
200.0	194.1	1.0303			-2.907962

**NO Calibration Curve**





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 24, 2024	Previous Calibration:	December 13, 2023
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:41	End Time (MST):	14:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.4	----	Correlation Coefficient	≥0.995	
388.3	388.6	0.9992			
185.6	185.8	0.9988			
91.7	91.6	1.0009			
			Slope	1.000174	0.90 - 1.10
			Intercept	0.184957	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: January 24, 2024

Location: Leismer





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS505**  
**SAWBONES BAY**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Sawbones Bay	Station number:	AMS505
Calibration Date:	January 11, 2024	Last Cal Date:	December 6, 2023
Start time (MST):	8:21	End time (MST):	11:12
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	51.4	ppm	Cal Gas Exp Date:	February 15, 2029
Cal Gas Cylinder #:	EY0000672			
Removed Cal Gas Conc:	51.40	ppm	Rem Gas Exp Date:	February 15, 2029
Removed Gas Cyl #:	EY0000672		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	5112
ZAG Make/Model:	Teledyne API T701H		Serial Number:	690

### Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	0710321323
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000723	1.001437	Backgd or Offset:	20.5	20.5
Calibration intercept:	-0.792239	-0.892107	Coeff or Slope:	0.994	0.994

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.5	----
as found span	4922	77.8	799.8	802.8	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4922	77.8	799.8	799.8	1.000
second point	4961	38.9	399.9	401.1	0.997
third point	4981	19.5	200.4	197.9	1.013
as left zero	5000	0.0	0.0	-0.5	----
as left span	4922	77.8	799.8	800.9	0.999
Average Correction Factor					1.003

Baseline Corr As found:	803.30	Previous response	799.60	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed inlet filter. No adjustment made.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 11, 2024	Previous Calibration:	December 6, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:21	End Time (MST):	11:12
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321323

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.2	----	Correlation Coefficient	0.999981	
799.8	799.8	1.0000			≥0.995
399.9	401.1	0.9970	Slope	1.001437	
200.4	197.9	1.0128			0.90 - 1.10
			Intercept	-0.892107	+/-30



SO2 Calibration Plot

Date: January 11, 2024

Location: Sawbones Bay







# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name:	Sawbones Bay	Station number:	AMS505
Calibration Date:	January 8, 2024	Last Cal Date:	NA
Start time (MST):	13:45	End time (MST):	15:59
Reason:	Install		

### Calibration Standards

Cal Gas Concentration:	5.15	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517397			
Removed Cal Gas Conc:	5.15	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledybe API T700		Serial Number:	5112
ZAG Make/Model:	Teledybe API T701		Serial Number:	690

### Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	12113311965
Converter make:	Global 150	Converter serial #:	2022-224
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.006361	Backgd or Offset:	NA	1.01
Calibration intercept:	NA	-0.218059	Coeff or Slope:	NA	1.119

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4922	77.7	80.0	80.4	0.995
second point	4961	38.8	40.0	39.9	1.002
third point	4981	19.4	20.0	19.8	1.009
as left zero	5000	0.0	0.0	0.0	----
as left span	4922	77.7	80.0	79.9	1.002
SO2 Scrubber Check	4922	77.8	778.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.002
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

\* = > +/-5% change initiates investigation

Notes: Install Calibration. Adjusted span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 8, 2024	Previous Calibration:	NA
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	13:45	End Time (MST):	15:59
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12113311965

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999989	≥0.995
80.0	80.4	0.9955			
40.0	39.9	1.0016	Slope	1.006361	0.90 - 1.10
20.0	19.8	1.0091			
			Intercept	-0.218059	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 8, 2024

Location: Sawbones Bay







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.2	-0.4	----	----
as found span	4917	83.4	799.6	799.6	0.0	804.3	803.7	0.6	0.9941	0.9949
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
high point	4917	83.4	799.6	799.6	0.0	799.4	799.0	0.3	1.0002	1.0007
second point	4958	41.7	399.8	399.8	0.0	398.1	397.2	0.9	1.0044	1.0067
third point	4979	20.9	200.4	200.4	0.0	197.0	196.3	0.7	1.0172	1.0209
as left zero	5000	0.0	0.0	0.0	0.0	-0.4	0.0	-0.4	----	----
as left span	4916	83.4	799.7	334.7	465.0	791.9	331.8	460.1	1.0099	1.0088
Average Correction Factor									1.0073	1.0094

Corrected As found	NO <sub>x</sub> = 805.0 ppb	NO = 803.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.5%
Previous Response	NO <sub>x</sub> = 800.8 ppb	NO = 801.3 ppb		*Percent Change	NO = 0.3%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.4	329.4	465.0	464.8	1.0004	100.0%
2nd GPT point (200 ppb O3)	794.4	546.5	247.9	248.2	0.9988	100.1%
3rd GPT point (100 ppb O3)	794.4	647.2	147.2	148.3	0.9926	100.7%
Average Correction Factor					0.9973	100.3%

Notes:

Changed inlet filter after as founds. Adjusted span.

Calibration Performed By:

Sean Bala



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 7, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	10:27	End Time (MST):	14:44
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.6	799.4	1.0002		
399.8	398.1	1.0044		
200.4	197.0	1.0172		
			0.999983	
			1.001378	
			-1.910345	





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 7, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	10:27	End Time (MST):	14:44
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
799.6	799.0	1.0007			
399.8	397.2	1.0067			
200.4	196.3	1.0209			
			Slope	1.000734	0.90 - 1.10
			Intercept	-2.109999	+/-20





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 10, 2024	Previous Calibration:	December 7, 2023
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	10:27	End Time (MST):	14:44
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
465.0	464.8	1.0004		
247.9	248.2	0.9988		
147.2	148.3	0.9926		






NO<sub>x</sub> Calibration Plot

Date: January 10, 2024

Location: Sawbones Bay





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS507  
KIRBY SOUTH**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Kirby South	Station number:	AMS 507
Calibration Date:	January 16, 2024	Last Cal Date:	December 24, 2023
Start time (MST):	12:24	End time (MST):	15:03
Reason:	Maintenance		

### Calibration Standards

Cal Gas Concentration:	49.18	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC303554</u>			
Removed Cal Gas Conc:	49.18	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3804
ZAG Make/Model:	API T701H		Serial Number:	880

### Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.016010		Backgd or Offset:	20.2      20.3
Calibration intercept:	-0.008833		Coeff or Slope:	1.146      1.146

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.8	----
as found span	4919	81.3	799.6	811.0	0.986
as found 2nd point	4959	40.7	400.3	403.6	0.992
as found 3rd point	4980	20.3	199.7	203.3	0.982
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

Average Correction Factor				
Baseline Corr As found:	810.20	Previous response	812.41	*% change      -0.3%
Baseline Corr 2nd AF pt:	402.80	AF Slope:	1.012737	AF Intercept:      0.311630
Baseline Corr 3rd AF pt:	202.50	AF Correlation:	0.999982	

\* = > +/-5% change initiates investigation


Notes: MPAFs done, swapped out UV lamp. Instrument will be left flagged overnight and will be calibrated Jan 17.

Calibration Performed By: Braiden Boutilier

SO2 Calibration Plot

Date: January 16, 2024

Location: Kirby South





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Kirby South	Station number:	AMS 507
Calibration Date:	January 17, 2024	Last Cal Date:	January 16, 2024
Start time (MST):	12:32	End time (MST):	16:52
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.18	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC303554</u>			
Removed Cal Gas Conc:	49.18	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3804
ZAG Make/Model:	API T701H		Serial Number:	880

### Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.016010	1.005405	Backgd or Offset:	20.2	24.6
Calibration intercept:	-0.008833	-1.048100	Coeff or Slope:	1.146	1.100

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.8	----
as found span	4919	81.3	799.6	811.0	0.986
as found 2nd point	4959	40.7	400.3	403.6	0.992
as found 3rd point	4980	20.3	199.7	203.3	0.982
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4919	81.3	799.6	804.0	0.995
second point	4959	40.7	400.3	398.9	1.004
third point	4980	20.3	199.7	200.4	0.996
as left zero	5000	0.0	0.0	-0.2	----
as left span	4919	81.3	799.6	796.6	1.004
Average Correction Factor					0.998

Baseline Corr As found:	810.20	Previous response	812.41	*% change	-0.3%
Baseline Corr 2nd AF pt:	402.80	AF Slope:	1.012737	AF Intercept:	0.311630
Baseline Corr 3rd AF pt:	202.50	AF Correlation:	0.999982		

\* = > +/-5% change initiates investigation

Notes: As founds done Jan 16 during maintenance. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	January 16, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	12:32	End Time (MST):	16:52
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999975	≥0.995
799.6	804.0	0.9946	Slope	1.005405	0.90 - 1.10
400.3	398.9	1.0036	Intercept	-1.048100	+/-30
199.7	200.4	0.9963			



SO2 Calibration Plot

Date: January 17, 2024

Location: Kirby South





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Kirby South      Station number: AMS507  
 Calibration Date: January 17, 2024      Last Cal Date: December 7, 2023  
 Start time (MST): 12:29      End time (MST): 19:05  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.05 ppm      Cal Gas Exp Date: November 15, 2026  
 Cal Gas Cylinder #: DT0019762  
 Removed Cal Gas Conc: 5.167 ppm      Rem Gas Exp Date: NA  
 Removed Gas Cyl #: CC517378      Diff between cyl: 2.4%  
 Calibrator Make/Model: API T750      Serial Number: 281  
 ZAG Make/Model: API T751H      Serial Number: 321

### Analyzer Information

Analyzer make: Thermo 43i TLE      Analyzer serial #: 1150840012  
 Converter make: Global      Converter serial #: 2022-197  
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002178	0.997813	Backgd or Offset:	1.81	1.74
Calibration intercept:	-0.081174	0.179050	Coeff or Slope:	1.079	1.052

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4923	77.4	80.0	79.6	1.005
as found 2nd point	4961	38.8	40.1	39.9	1.005
as found 3rd point	4981	19.3	19.9	20.2	0.987
new cylinder response	4921	79.2	80.0	81.6	0.980

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	79.2	80.0	80.1	0.999
second point	4960	39.6	40.0	39.8	1.005
third point	4980	19.8	20.0	20.3	0.985
as left zero	5000	0.0	0.0	0.1	----
as left span	4921	79.2	80.0	78.7	1.016
SO2 Scrubber Check	4919	80.0	800.2	0.1	----
Date of last scrubber change:	25-Jul-23			Ave Corr Factor	0.996
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 79.6      Prev response: 80.07      \*% change: -0.6%  
 Baseline Corr 2nd AF pt: 39.9      AF Slope: 0.993743      AF Intercept: 0.139102  
 Baseline Corr 3rd AF pt: 20.2      AF Correlation: 0.999975

*\* = > +/-5% change initiates investigation*

Notes: Changed cyl#: CC517378, with cyl#: DT0019762. Ran SOx scrubber check after cal zero. Adjusted span.

Calibration Performed By: Braiden Boutillier





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	December 7, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	12:29	End Time (MST):	19:05
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999965	≥0.995
80.0	80.1	0.9987			
40.0	39.8	1.0050	Slope	0.997813	0.90 - 1.10
20.0	20.3	0.9852			
			Intercept	0.179050	+/-3



H<sub>2</sub>S Calibration Plot

Date: January 17, 2024

Location: Kirby South





THC Calibration Plot

Date: January 16, 2024

Location: Kirby South







# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	January 16, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	12:24	End Time (MST):	16:52
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

### Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.06	----	Correlation Coefficient	0.999994	≥0.995
17.26	17.22	1.0025			
8.64	8.64	1.0003	Slope	0.993376	0.90 - 1.10
4.31	4.38	0.9852			
			Intercept	0.069808	+/-1.5



THC Calibration Plot

Date: January 17, 2024

Location: Kirby South









# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.6	0.0	----	----
as found span	4919	81.0	800.1	794.1	6.0	803.0	797.7	5.3	0.9964	0.9955
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
high point	4919	81.0	800.1	794.1	6.0	801.0	796.0	5.1	0.9989	0.9976
second point	4960	40.5	400.0	397.0	3.0	394.4	390.6	3.9	1.0142	1.0164
third point	4980	20.2	199.5	198.0	1.5	192.0	189.7	2.3	1.0392	1.0439
as left zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.4	-0.1	----	----
as left span	4919	81.0	800.1	413.9	386.2	792.8	408.9	383.9	1.0092	1.0123
Average Correction Factor									1.0175	1.0193

Corrected As found	NO <sub>x</sub> = 803.7 ppb	NO = 798.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = 0.7%
Previous Response	NO <sub>x</sub> = 797.7 ppb	NO = 790.3 ppb		*Percent Change	NO = 1.0%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	791.4	411.2	386.2	386.1	1.0002	100.0%
2nd GPT point (200 ppb O3)	791.4	623.5	173.9	173.5	1.0023	99.8%
3rd GPT point (100 ppb O3)	791.4	714.8	82.6	84.5	0.9774	102.3%
Average Correction Factor					0.9933	100.7%

Notes:

Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By:

Braiden Boutilier



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 6, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	8:51	End Time (MST):	14:24
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.1	801.0	0.9989		
400.0	394.4	1.0142		
199.5	192.0	1.0392		
			0.999874	
			1.003862	
			-4.492458	





# Wood Buffalo Environmental Association

## NO Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 6, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	8:51	End Time (MST):	14:24
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.3	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ 0.90 - 1.10 +/-20
794.1	796.0	0.9976		
397.0	390.6	1.0164		
198.0	189.7	1.0439		





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 18, 2024	Previous Calibration:	December 6, 2023
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	8:51	End Time (MST):	14:24
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
386.2	386.1	1.0002			
173.9	173.5	1.0023			
82.6	84.5	0.9774			
			Slope	0.997888	0.90 - 1.10
			Intercept	0.693913	+/-20


**NO<sub>2</sub> Calibration Curve**



NO<sub>x</sub> Calibration Plot

Date: January 18, 2024

Location: Kirby South





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS511  
BLACKGOLD**

**JANUARY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

February 29, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

Version-01-2020

### Station Information

Station Name:	Blackgold	Station number:	AMS511
Calibration Date:	January 17, 2024	Last Cal Date:	December 14, 2023
Start time (MST):	12:22	End time (MST):	15:33
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.06	ppm	Cal Gas Exp Date:	January 5, 2029
Cal Gas Cylinder #:	CC147416			
Removed Cal Gas Conc:	50.06	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	5258
ZAG Make/Model:	Teledyne API 701		Serial Number:	138

### Analyzer Information

Analyzer make: Thermo scientific                      Analyzer serial #: 1160290014  
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002260	1.000172	Backgd or Offset:	35.4	36.5
Calibration intercept:	0.830491	2.710765	Coeff or Slope:	1.187	1.228

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	1.6	----
as found span	4926	79.9	799.0	774.7	1.031
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	2.0	----
high point	4926	79.9	799.0	801.4	0.997
second point	4968	40.0	399.8	403.3	0.991
third point	4987	20.0	200.0	203.2	0.984
as left zero	5000	0.0	0.0	1.9	----
as left span	4926	80.0	800.0	801.8	0.998
Average Correction Factor					0.991

Baseline Corr As found:	773.10	Previous response	801.65	*% change	-3.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes:    Changed the sample inlet filter after as founds. Adjusted span only.

Calibration Performed By:                      Mohammed Kashif



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	December 14, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	12:22	End Time (MST):	15:33
Analyzer make:	Thermo scientific	Analyzer serial #:	1160290014

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	2.0	----	Correlation Coefficient	0.999996	
799.0	801.4	0.9970			≥0.995
399.8	403.3	0.9914	Slope	1.000172	
200.0	203.2	0.9841			0.90 - 1.10
			Intercept	2.710765	+/-30






SO2 Calibration Plot

Date: January 17, 2024

Location: Blackgold





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

Version-11-2021

### Station Information

Station Name: Blackgold Station number: AMS511  
 Calibration Date: January 25, 2024 Last Cal Date: December 13, 2023  
 Start time (MST): 11:19 End time (MST): 15:16  
 Reason: Routine

### Calibration Standards

Cal Gas Concentration: 5.139 ppm Cal Gas Exp Date: January 3, 2026  
 Cal Gas Cylinder #: CC511397  
 Removed Cal Gas Conc: 5.139 ppm Rem Gas Exp Date: NA  
 Removed Gas Cyl #: NA Diff between cyl:  
 Calibrator Make/Model: API T700 Serial Number: 2445  
 ZAG Make/Model: API T701 Serial Number: 138

### Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1336160090  
 Converter make: Global G150 Converter serial #: 2022-227  
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.982906	0.977047	Backgd or Offset: 3.41	3.39
Calibration intercept:	0.001152	0.181198	Coeff or Slope: 1.159	1.159

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4922	77.8	80.0	78.4	1.021
as found 2nd point	4961	38.9	40.0	39.3	1.020
as found 3rd point	4981	19.5	20.0	19.9	1.012
new cylinder response					

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4922	77.8	80.0	78.3	1.021
second point	4961	38.9	40.0	39.3	1.017
third point	4981	19.5	20.0	19.7	1.017
as left zero	5000	0.0	0.0	0.2	----
as left span	4922	77.9	80.0	77.9	1.027
SO2 Scrubber Check	4920	80.0	800.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	1.019
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 78.3 Prev response: 78.60 \*% change: -0.4%  
 Baseline Corr 2nd AF pt: 39.2 AF Slope: 0.978479 AF Intercept: 0.181087  
 Baseline Corr 3rd AF pt: 19.8 AF Correlation: 0.999994

\* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Summary


Version-11-2021

### Station Information

Calibration Date:	January 25, 2024	Previous Calibration:	December 13, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	11:19	End Time (MST):	15:16
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999998	≥0.995
80.0	78.3	1.0213			
40.0	39.3	1.0174	Slope	0.977047	0.90 - 1.10
20.0	19.7	1.0173			
			Intercept	0.181198	+/-3



# H<sub>2</sub>S Calibration Plot

Date: January 25, 2024

Location: Blackgold





# Wood Buffalo Environmental Association

## THC Calibration Report

Version-01-2020

### Station Information

Station Name:	Blackgold	Station number:	AMS511
Calibration Date:	January 17, 2024	Last Cal Date:	December 14, 2023
Start time (MST):	12:22	End time (MST):	15:33
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC147416	Cal Gas Expiry Date:	January 5, 2029
CH4 Cal Gas Conc.	<u>498.90</u> ppm	CH4 Equiv Conc.	1070.90 ppm
C3H8 Cal Gas Conc.	<u>208.00</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>498.90</u> ppm	CH4 Equiv Conc.	1070.90 ppm
Removed C3H8 Conc.	<u>208.00</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5258
ZAG Make/Model:	Teledyne API 701	Serial Number:	138

### Analyzer Information

Analyzer make: Thermo 51i-LT      Analyzer serial #: 1317958295  
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994672	0.996004	Background:	0.90	0.89
Calibration intercept:	0.019633	-0.004347	Coefficient:	0.575	0.567

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.05	----
as found span	4920	79.9	17.11	17.35	0.986
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.07	----
high point	4920	79.9	17.11	17.09	1.001
second point	4960	40.0	8.57	8.45	1.014
third point	4980	20.0	4.28	4.22	1.016
as left zero	5000	0.0	0.00	0.08	----
as left span	4926	80.0	17.11	17.16	0.997
Average Correction Factor					1.010
Baseline Corr As found:	17.30	Previous response	17.04	*% change	1.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

\* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## THC Calibration Summary


Version-01-2020

### Station Information

Calibration Date:	January 17, 2024	Previous Calibration:	December 14, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	12:22	End Time (MST):	15:33
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1317958295

### Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.07	----	Correlation Coefficient	0.999893	
17.11	17.09	1.0012			≥0.995
8.57	8.45	1.0143	Slope	0.996004	
4.28	4.22	1.0156			0.90 - 1.10
			Intercept	-0.004347	+/-1.5



THC Calibration Plot

Date: January 17, 2024

Location: Blackgold





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Station Information

Station Name: Blackgold  
Calibration Date: January 9, 2024  
Start time (MST): 11:11  
Reason: Routine  
Station number: AMS511  
Last Cal Date: December 22, 2023  
End time (MST): 16:14

### Calibration Standards

NO Gas Cylinder #: T0F8P52  
NOX Cal Gas Conc: 47.43 ppm  
Removed Cylinder #: NA  
Removed Gas NOX Conc: 47.43 ppm  
NOX gas Diff:  
Calibrator Model: Teledyne API T700  
ZAG make/model: Teledyne API T701  
Cal Gas Expiry Date: August 16, 2026  
NO Cal Gas Conc: 47.43 ppm  
Removed Gas Exp Date: NA  
Removed Gas NO Conc: 47.43 ppm  
NO gas Diff:  
Serial Number: 2445  
Serial Number: 138

### Analyzer Information

Analyzer make: Teledyne API T200  
NOX Range (ppb): 0 - 1000 ppb  
Analyzer serial #: 7029

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.041	1.041	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.038	1.038	NOX bkgnd or offset:	0.4	0.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.2	4.1

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.012843	1.002964
NO <sub>x</sub> Cal Offset:	0.247028	-1.992779
NO Cal Slope:	1.011487	1.002150
NO Cal Offset:	-0.052956	-2.932730
NO <sub>2</sub> Cal Slope:	0.998436	1.004437
NO <sub>2</sub> Cal Offset:	1.931124	1.342612





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

Version-04-2020

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
as found span	4916	84.4	800.6	800.6	0.0	792.4	789.6	2.8	1.0103	1.0139
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4916	84.4	800.6	800.6	0.0	802.0	801.0	1.1	0.9982	0.9994
second point	4958	42.2	400.3	400.3	0.0	398.2	396.1	2.1	1.0053	1.0106
third point	4979	21.1	200.2	200.2	0.0	197.1	195.3	1.7	1.0155	1.0249
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as left span	4916	84.4	800.6	429.3	371.3	798.0	433.6	364.4	1.0032	0.9900
Average Correction Factor									1.0063	1.0116

Corrected As found	NO <sub>x</sub> = 792.7 ppb	NO = 789.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO <sub>x</sub> = -2.3%
Previous Response	NO <sub>x</sub> = 811.1 ppb	NO = 809.7 ppb		*Percent Change	NO = -2.5%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.0	425.7	371.3	373.5	0.9941	100.6%
2nd GPT point (200 ppb O3)	797.0	620.2	176.8	180.1	0.9817	101.9%
3rd GPT point (100 ppb O3)	797.0	707.4	89.6	92.3	0.9707	103.0%
Average Correction Factor					0.9822	101.8%

Notes: Changed the sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association

## NO<sub>x</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 9, 2024	Previous Calibration:	December 22, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	11:11	End Time (MST):	16:14
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999974	≥0.995
800.6	802.0	0.9982			
400.3	398.2	1.0053	Slope	1.002964	0.90 - 1.10
200.2	197.1	1.0155			
			Intercept	-1.992779	+/-20





# Wood Buffalo Environmental Association

## NO Calibration Summary

Version-04-2020


### Station Information

Calibration Date:	January 9, 2024	Previous Calibration:	December 22, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	11:11	End Time (MST):	16:14
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	$\geq 0.995$ 0.90 - 1.10 +/-20
800.6	801.0	0.9994		
400.3	396.1	1.0106		
200.2	195.3	1.0249		
			0.999941	
			1.002150	
			-2.932730	

NO Calibration Curve





# Wood Buffalo Environmental Association

## NO<sub>2</sub> Calibration Summary


Version-04-2020

### Station Information

Calibration Date:	January 9, 2024	Previous Calibration:	December 22, 2023
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	11:11	End Time (MST):	16:14
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
371.3	373.5	0.9941		
176.8	180.1	0.9817		
89.6	92.3	0.9707		



NO<sub>x</sub> Calibration Plot

Date: January 9, 2024

Location: Blackgold





End of Report