



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

FEBRUARY 2024 MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

March 28, 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01
BERTHA GANTER - FORT MCKAY

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	February 5, 2024	Last Cal Date:	January 11, 2024
Start time (MST):	11:04	End time (MST):	14:07
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC418809			
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999624	1.001365	Backgd or Offset:	19.6	19.7
Calibration intercept:	-0.113430	-0.173166	Coeff or Slope:	0.891	0.891

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4918	81.3	800.3	799.9	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4918	81.3	800.3	801.7	0.998
second point	4959	40.7	400.6	400.0	1.001
third point	4979	20.3	199.8	199.8	1.000
as left zero	5000	0.0	0.0	0.5	----
as left span	4918	81.3	800.3	802.8	0.997
Average Correction Factor					1.000

Baseline Corr As found:	799.60	Previous response	799.85	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 11, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:04	End Time (MST):	14:07
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999996	≥0.995
800.3	801.7	0.9982			
400.6	400.0	1.0015	Slope	1.001365	0.90 - 1.10
199.8	199.8	1.0001			
			Intercept	-0.173166	+/-30



SO2 Calibration Plot

Date: January 11, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: February 7, 2024 Last Cal Date: January 25, 2024
 Start time (MST): 10:02 End time (MST): 14:36
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153461
 Converter make: CD Nova Converter serial #: 470
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994364	0.997793	Backgd or Offset:	2.32	2.45
Calibration intercept:	0.300000	0.179996	Coeff or Slope:	0.937	0.937

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4921	78.4	80.0	80.4	0.997
as found 2nd point	4960	39.2	40.0	40.3	0.997
as found 3rd point	4980	19.6	20.0	20.3	0.995
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4921	78.4	80.0	79.9	1.001
second point	4960	39.2	40.0	40.3	0.993
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.2	----
as left span	4921	78.4	80.0	79.5	1.006
SO2 Scrubber Check	4919	81.3	813.0	0.2	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	0.996
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 80.2 Prev response: 79.84 *% change: 0.4%
 Baseline Corr 2nd AF pt: 40.1 AF Slope: 1.002364 AF Intercept: 0.220000
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999999

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted zero only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 7, 2024	Previous Calibration:	January 25, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:02	End Time (MST):	14:36
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999982	≥0.995
80.0	79.9	1.0012			
40.0	40.3	0.9925	Slope	0.997793	0.90 - 1.10
20.0	20.1	0.9949			
			Intercept	0.179996	+/-3



TRS Calibration Plot

Date: February 7, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: February 7, 2024 Last Cal Date: January 25, 2024
 Start time (MST): 10:02 End time (MST): 14:36
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1200326167
 Converter make: Global Converter serial #: 2022-221
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998712	1.002571	Backgd or Offset: 2.45	2.04
Calibration intercept:	0.176812	0.116774	Coeff or Slope: 0.992	0.992

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	4922	78.4	80.0	80.4	0.990
as found 2nd point	4960	39.2	40.0	40.0	0.990
as found 3rd point	4980	19.6	20.0	19.7	0.995
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	78.4	80.0	80.2	0.997
second point	4960	39.2	40.0	40.4	0.990
third point	4980	19.6	20.0	20.2	0.990
as left zero	5000	0.0	0.0	0.1	----
as left span	4922	78.4	80.0	80.2	0.997
SO2 Scrubber Check	4919	81.3	813.0	0.4	----

Date of last scrubber change:	January 25, 2024	Ave Corr Factor	0.992
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 80.8 Prev response: 80.04 *% change: 0.9%
 Baseline Corr 2nd AF pt: 40.4 AF Slope: 1.010914 AF Intercept: -0.443233
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999998

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted zero only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 7, 2024	Previous Calibration:	January 25, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:02	End Time (MST):	14:36
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999984	≥0.995
80.0	80.2	0.9972			
40.0	40.4	0.9901	Slope	1.002571	0.90 - 1.10
20.0	20.2	0.9900			
			Intercept	0.116774	+/-3



H₂S Calibration Plot

Date: February 7, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	February 5, 2024	Last Cal Date:	January 11, 2024
Start time (MST):	11:04	End time (MST):	14:07
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.37E-04	4.30E-04	NMHC SP Ratio:	7.11E-05
CH ₄ Retention time:	16.7	16.7	NMHC Peak Area:	129269
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.27	17.38	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	17.27	17.28	0.999
second point	4959	40.7	8.64	8.50	1.017
third point	4980	20.3	4.31	4.28	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.27	17.28	0.999
Average Correction Factor					1.007

Baseline Corr AF:	17.38	Prev response	17.18	*% change	1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4918	81.3	9.18	9.12	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	9.18	9.17	1.001
second point	4959	40.7	4.60	4.55	1.011
third point	4980	20.3	2.29	2.30	0.998
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	9.18	9.21	0.997
Average Correction Factor					1.003
Baseline Corr AF:	9.12	Prev response	9.15	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	8.26	0.978
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	8.09	8.11	0.997
second point	4959	40.7	4.05	3.96	1.023
third point	4980	20.3	2.02	1.99	1.016
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.07	1.002
Average Correction Factor					1.012
Baseline Corr AF:	8.26	Prev response	8.05	*% change	2.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.996788	1.000371
THC Cal Offset:	-0.028998	-0.041569
CH ₄ Cal Slope:	0.998653	1.002874
CH ₄ Cal Offset:	-0.029405	-0.035591
NMHC Cal Slope:	0.996266	0.998167
NMHC Cal Offset:	-0.001592	-0.005978

Notes: Changed out the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 11, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:04	End Time (MST):	14:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999909	≥ 0.995			
17.27	17.28	0.9992						
8.64	8.50	1.0166				Slope	1.000371	0.90 - 1.10
4.31	4.28	1.0063						
			Intercept	-0.041569	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 11, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:04	End Time (MST):	14:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999801	≥0.995
8.09	8.11	0.9972			
4.05	3.96	1.0234			
2.02	1.99	1.0159			
			Slope	1.002874	0.90 - 1.10
			Intercept	-0.035591	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 11, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:04	End Time (MST):	14:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999964	≥ 0.995
9.18	9.17	1.0010			
4.60	4.55	1.0108			
2.29	2.30	0.9979			
			Slope	0.998167	0.90 - 1.10
			Intercept	-0.005978	+/-0.5



NMHC Calibration Plot

Date: February 5, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
Calibration Date: February 26, 2024 Last Cal Date: January 4, 2024
Start time (MST): 10:29 End time (MST): 16:46
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC335700 Cal Gas Expiry Date: September 1, 2032
NOX Cal Gas Conc: 59.40 ppm NO Cal Gas Conc: 59.20 ppm
Removed Cylinder #: T2Y1P9L Removed Gas Exp Date: December 11, 2023
Removed Gas NOX Conc: 50.84 ppm Removed Gas NO Conc: 50.04 ppm
NOX gas Diff: -2.2% NO gas Diff: -0.7%
Calibrator Model: Teledyne API T700 Serial Number: 3565
ZAG make/model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1218153357
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.442	1.556	NO bkgnd or offset:	7.1	7.6
NOX coeff or slope:	0.992	0.999	NOX bkgnd or offset:	7.2	7.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	199.2	208.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000562	1.000310
NO _x Cal Offset:	0.220000	-0.160000
NO Cal Slope:	0.999101	1.001719
NO Cal Offset:	-0.740000	-0.920000
NO ₂ Cal Slope:	1.003895	0.995800
NO ₂ Cal Offset:	0.627156	-0.397873



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	0.0	----	----
as found span	4920	80.0	813.4	800.6	12.8	761.9	744.1	17.8	1.0676	1.0760
as found 2nd										
as found 3rd										
new cyl resp	4932	67.6	803.1	800.4	2.7	737.2	739.2	-2.0	1.0894	1.0828
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.1	----	----
high point	4932	67.6	803.1	800.4	2.7	803.8	801.8	2.1	0.9991	0.9982
second point	4966	33.8	401.5	400.2	1.4	400.0	398.2	1.8	1.0039	1.0050
third point	4983	16.9	200.8	200.1	0.7	201.2	199.2	2.1	0.9979	1.0045
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
as left span	4932	67.6	803.1	393.2	409.9	797.3	390.7	406.6	1.0073	1.0064
Average Correction Factor									1.0003	1.0026

Corrected As found	NO _x = 762.0 ppb	NO = 744.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -6.8%
Previous Response	NO _x = 814.1 ppb	NO = 799.2 ppb		*Percent Change	NO = -7.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.9	390.7	409.9	407.9	1.0049	99.5%
2nd GPT point (200 ppb O3)	797.9	594.8	205.8	204.7	1.0054	99.5%
3rd GPT point (100 ppb O3)	797.9	697.1	103.5	101.9	1.0157	98.5%
Average Correction Factor					1.0087	99.1%

Notes: Changed the NO calibration gas cylinder after as founds. Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 26, 2024	Previous Calibration:	January 4, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:29	End Time (MST):	16:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
803.1	803.8	0.9991		
401.5	400.0	1.0039		
200.8	201.2	0.9979		
			0.999991	
			1.000310	
			-0.160000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 26, 2024	Previous Calibration:	January 4, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:29	End Time (MST):	16:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.4	801.8	0.9982		
400.2	398.2	1.0050		
200.1	199.2	1.0045		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 26, 2024	Previous Calibration:	January 4, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:29	End Time (MST):	16:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
409.9	407.9	1.0049			
205.8	204.7	1.0054			
103.5	101.9	1.0157			
			Slope	0.995800	0.90 - 1.10
			Intercept	-0.397873	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 26, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: February 2, 2024 Last Cal Date: January 3, 2024
 Start time (MST): 11:06 End time (MST): 14:10
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 1107
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999800	0.999143	Backgd or Offset:	4.8	5.3
Calibration intercept:	0.360000	1.000000	Coeff or Slope:	1.014	1.015

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.6	----
as found span	5000	863.1	400.0	400.5	0.999
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.4	----
high point	5000	863.1	400.0	400.3	0.999
second point	5000	742.5	200.0	201.3	0.994
third point	5000	651.7	100.0	101.4	0.986
as left zero	5000	0.0	0.0	0.4	----
as left span	5000	863.1	400.0	405.6	0.986
Average Correction Factor					0.993

Baseline Corr As found:	399.9	Previous response	400.3	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted zero only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 3, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:06	End Time (MST):	14:10
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999989	≥0.995
400.0	400.3	0.9993			
200.0	201.3	0.9935	Slope	0.999143	0.90 - 1.10
100.0	101.4	0.9862			
			Intercept	1.000000	+/- 5



O₃ Calibration Plot

Date: February 2, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
 Calibration Date: February 27, 2024 Last Cal Date: January 26, 2024
 Start time (MST): 12:47 End time (MST): 13:39

Analyzer Make: Teledyne API T640 S/N: 324
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752
 Temp/RH standard: Alicat FP-25BT S/N: 388752

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-15.5	-16.21	-15.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	742.2	741.64	742.2	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	5.137	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	43	----	43	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.1	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: _____ January 26, 2024
 Date Disposable Filter Changed: _____ January 26, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ September 14, 2023
 Date RH/T Sensor Cleaned: _____ January 26, 2024

Notes: Verified flow, temperature, and pressure. Leak check passed.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	February 21, 2024	Last Cal Date:	January 8, 2024
Start time (MST):	10:42	End time (MST):	13:39
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	<u>3040</u>	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>3040</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000186	1.001370	Backgd or Offset:	-0.013	-0.013
Calibration intercept:	0.241822	0.145823	Coeff or Slope:	0.983	0.989

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4933	66.7	40.6	41.0	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.6	40.6	0.998
second point	4966	33.3	20.2	20.7	0.979
third point	4983	16.7	10.2	10.3	0.985
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.5	40.2	1.010
Average Correction Factor					0.987

Baseline Corr As found:	40.70	Prev response:	40.81	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 8, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:42	End Time (MST):	13:39
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999891	
40.6	40.6	0.9982			≥0.995
20.2	20.7	0.9792	Slope	1.001370	
10.2	10.3	0.9849			0.90 - 1.10
			Intercept	0.145823	+/-1.5



CO Calibration Plot

Date: February 21, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	February 6, 2024	Last Cal Date:	January 5, 2024
Start time (MST):	9:59	End time (MST):	12:52
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>60,200</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
N2 Gen Make/Model:	Peak Sci NG5000		Serial Number:	7220900034

Analyzer Information

Analyzer make: Teledyne API 360	Analyzer serial #: 442
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001967	0.998372	Backgd or Offset:	0.045	0.045
Calibration intercept:	-5.940000	-5.840000	Coeff or Slope:	0.876	0.876

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-0.2	----
as found span	2920	80.0	1605.3	1596.8	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.1	----
high point	2920	80.0	1605.3	1599.2	1.004
second point	2960	40.0	802.7	794.5	1.010
third point	2980	20.0	401.3	387.6	1.035
as left zero	3000	0.0	0.0	0.0	----
as left span	2960	40.0	802.7	784.0	1.024

Average Correction Factor	1.017
---------------------------	-------

Baseline Corr As found:	1597.00	Prev response:	1602.55	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	February 6, 2024	Previous Calibration	January 5, 2024
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS01
Start Time (MST)	9:59	End Time (MST)	12:52
Analyzer make	Teledyne API 360	Analyzer serial #	442

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999933	≥0.995
1605.3	1599.2	1.0038			
802.7	794.5	1.0103	Slope	0.998372	0.90 - 1.10
401.3	387.6	1.0354			
			Intercept	-5.840000	+/-10



CO₂ Calibration Plot

Date: February 6, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	February 27, 2024	Last Cal Date:	January 9, 2024
Start time (MST):	11:26	End time (MST):	16:30
NH3 Cal Date:	February 27, 2024	Last Cal Date:	January 10, 2024
Start time (MST):	16:50	End time (MST):	18:26
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	59.40	ppm	NO Gas Cylinder #:	CC335700
NO Cal Gas Conc:	59.20	ppm	NO Cal Gas Expiry:	September 1, 2032
Removed NOX Conc:	59.40	ppm	Removed Cylinder #:	NA
Removed NO Conc:	59.20	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.58	ppm	NH3 Gas Cylinder #:	CC743587
			NH3 Cal Gas Expiry:	August 22, 2024
Removed NH3 Conc:	76.58	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	Teledyne API T700		Serial Number:	3565
ZAG make/model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	475
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.10
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	519

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.964	0.956	TN coefficient:	0.968	0.968
NOX coefficient:	0.967	0.962	NO bkgnd:	-0.9	-0.9
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-0.3	-0.3
NH3 coefficient:	0.932	0.946	TN bkgnd:	1.2	1.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998960	0.999528
NO _x Cal Offset:	-2.460000	-1.960000
NO Cal Slope:	0.998530	0.999963
NO Cal Offset:	-2.840000	-2.580000
NO ₂ Cal Slope:	0.997727	0.992493
NO ₂ Cal Offset:	-0.655687	-0.912249
NH3 Cal Slope:	1.004832	1.004564
NH3 Cal Offset:	-8.183521	-0.612859
TN Cal Slope:	1.007188	1.007097
TN Cal Offset:	-8.051061	-0.451064



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
as found NO	4932	67.6	803.1	803.1	----	804.6	806.0	-1.4	0.998	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.1	----	----
high NO point	4932	67.6	803.1	803.1	----	803.0	801.7	1.5	1.000	----
NO/O3 point	4932	67.6	803.1	803.1	----	798.0	794.6	3.3	1.006	----
as found NH3	3416	84.1	1840.1	----	1840.1	1812.4	----	1807.6	1.015	1.018
new NH3 cyl rp							----			
first NH3	3418	82.2	1798.5	----	1798.5	1815.9	----	1811.4	0.990	0.993
second NH3	3454	45.7	1000.0	----	1000.0	993.8	----	991.0	1.006	1.009
third NH3	3477	22.8	498.9	----	498.9	509.5	----	507.5	0.979	0.983
Average Correction Factor									1.0032	0.9950

Corrected As found TN = 805 ppb NO_x = 806.3 ppb NH3 = 1807.7 ppb

Previous Response TN = 800.8 ppb NO_x = 799.8 ppb NH3 = 1840.8 ppb

NH3 Previous Converter Efficiency = 93.2%

NH3 Current Converter Efficiency = 94.6%

*Percent Change TN = 0.5%

*Percent Change NO_x = 0.8%

*Percent Change NH3 = -1.8%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.4	----	----
as found span	4932	67.6	803.1	800.4	803.1	806.0	797.9	804.6	0.9964	1.0031
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.2	----	----
high point	4932	67.6	803.1	800.4	803.1	801.7	799.4	803.0	1.0017	1.0012
second point	4966	33.8	401.5	400.2	401.5	398.3	395.2	397.7	1.0081	1.0126
third point	4983	16.9	200.8	200.1	200.8	197.1	195.8	197.1	1.0186	1.0219
Average Correction Factor									1.0095	1.0119

Baseline Corr As fnd	TN = 805 ppb	NO _x = 806.3 ppb	NO = 798.1 ppb	*Percent Change	TN = 0.5%
Previous Response	TN = 800.8 ppb	NO _x = 799.8 ppb	NO = 796.4 ppb	*Percent Change	NO _x = 0.8%
				*Percent Change	NO = 0.2%
				<i>* = > +/-5% change initiates investigation</i>	

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	-0.1	----	----
calibration zero	----	----	0.0	-0.2	----	----
1st GPT point (400 ppb O3)	795.6	386.8	411.5	407.8	1.0091	99.1%
2nd GPT point (200 ppb O3)	795.6	590.4	207.9	205.3	1.0127	98.7%
3rd GPT point (100 ppb O3)	795.6	690.0	108.3	105.7	1.0246	97.6%
Average Correction Factor					1.0155	98.5%

Notes: Changed the inlet filter after as founds. Adjusted the NO_x/TN span. Adjusted the NH₃ span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TN Calibration Summary


Version-05-2023

Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	January 9, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	16:50	End Time (MST):	18:26
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.2	----	Correlation Coefficient	0.999862	≥0.995
1798.5	1815.9	0.9904			
1000.0	993.8	1.0062	Slope	1.007097	0.90 - 1.10
498.9	509.5	0.9791			
			Intercept	-0.451064	+/-20





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	January 9, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	16:50	End Time (MST):	18:26
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
1798.5	1811.4	0.9929			
1000.0	991.0	1.0091			
498.9	507.5	0.9830			
			Slope	1.004564	0.90 - 1.10
			Intercept	-0.612859	+/-20





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	January 9, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:26	End Time (MST):	16:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
803.1	801.7	1.0017			
401.5	398.3	1.0081	Slope	0.90 - 1.10	
200.8	197.1	1.0186			
			Intercept	-1.960000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	January 9, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:26	End Time (MST):	16:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
800.4	799.4	1.0012			
400.2	395.2	1.0126	Slope	0.90 - 1.10	
200.1	195.8	1.0219			
			Intercept	-2.580000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-05-2023


Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	January 9, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:26	End Time (MST):	16:30
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
411.5	407.8	1.0091		
207.9	205.3	1.0127		
108.3	105.7	1.0246		
			0.999985	
			0.992493	
			-0.912249	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 27, 2024

Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: February 27, 2024

Location: Bertha Ganter-Fort McKay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	February 15, 2024	Last Cal Date:	January 1, 2024
Start time (MST):	10:35	End time (MST):	13:02
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1185
ZAG Make/Model:	API T701		Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994869	0.998019	Backgd or Offset:	18.4	18.4
Calibration intercept:	0.074375	-0.505025	Coeff or Slope:	0.795	0.795

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4920	80.2	801.6	806.8	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.2	801.6	800.4	1.002
second point	4960	40.1	400.8	397.6	1.008
third point	4980	20.0	199.9	199.5	1.002
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	80.2	801.6	796.5	1.006
Average Correction Factor					1.004

Baseline Corr As found:	806.60	Previous response	797.61	*% change	1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments or maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 1, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:35	End Time (MST):	13:02
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999986	
801.6	800.4	1.0016			≥0.995
400.8	397.6	1.0081	Slope	0.998019	
199.9	199.5	1.0021			0.90 - 1.10
			Intercept	-0.505025	+/-30



SO2 Calibration Plot

Date: February 15, 2024

Location: Mildred Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mildred Lake Station number: AMS02
 Calibration Date: February 2, 2024 Last Cal Date: January 10, 2024
 Start time (MST): 10:56 End time (MST): 16:00
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.29 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345191
 Removed Cal Gas Conc: 5.29 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1185
 ZAG Make/Model: API T701 Serial Number: 5608

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 12113311966
 Converter make: Global G150 Converter serial #: 2022-198
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001107	0.998393	Backgd or Offset:	1.68	1.68
Calibration intercept:	0.040801	-0.039204	Coeff or Slope:	0.740	0.731

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4924	75.6	80.0	81.4	0.983
as found 2nd point	4962	37.8	40.0	40.5	0.988
as found 3rd point	4981	18.9	20.0	19.9	1.005
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4924	75.6	80.0	79.9	1.001
second point	4962	37.8	40.0	39.8	1.005
third point	4981	18.9	20.0	19.8	1.010
as left zero	5000	0.0	0.0	0.1	----
as left span	4924	75.6	80.0	79.9	1.001
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:	20-Sep-23			Ave Corr Factor	1.005
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 81.4 Prev response: 80.12 *% change: 1.6%
 Baseline Corr 2nd AF pt: 40.5 AF Slope: 1.019252 AF Intercept: -0.219189
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999965

* = > +/-5% change initiates investigation

Notes: Reset analyzer after MPAF's due to screen freeze after station power reset. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 10, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:56	End Time (MST):	16:00
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
80.0	79.9	1.0011		
40.0	39.8	1.0049	Slope	0.90 - 1.10
20.0	19.8	1.0099		
			Intercept	+/-3



H₂S Calibration Plot

Date: February 2, 2024

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	February 15, 2024	Last Cal Date:	January 25, 2024
Start time (MST):	7:47	End time (MST):	10:41
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
C ₃ H ₈ Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
Removed C ₃ H ₈ Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.62E-04	2.73E-04	NMHC SP Ratio:	6.08E-05	6.17E-05
CH ₄ Retention time:	15.5	15.9	NMHC Peak Area:	144769	142620
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	16.48	1.021
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.79	1.002
second point	4960	40.1	8.41	8.35	1.007
third point	4980	20.0	4.19	4.15	1.010
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.80	1.001
Average Correction Factor					1.006

Baseline Corr AF:	16.48	Prev response	16.82	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	8.75	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.76	1.004
second point	4960	40.1	4.40	4.37	1.006
third point	4980	20.0	2.19	2.17	1.011
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.76	1.004
Average Correction Factor					1.007
Baseline Corr AF:	8.75	Prev response	8.80	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	7.73	1.038
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.02	1.000
second point	4960	40.1	4.01	3.98	1.007
third point	4980	20.0	2.00	1.98	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	8.03	0.999
Average Correction Factor					1.005
Baseline Corr AF:	7.73	Prev response	8.02	*% change	-3.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999652	0.998749
THC Cal Offset:	0.005690	-0.021919
CH ₄ Cal Slope:	1.000099	1.000498
CH ₄ Cal Offset:	-0.001651	-0.013049
NMHC Cal Slope:	0.999219	0.996921
NMHC Cal Offset:	0.006941	-0.008470

Notes: Zero chromatogram turned on. Zero and span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 25, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	7:47	End Time (MST):	10:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999991	≥ 0.995			
16.82	16.79	1.0017						
8.41	8.35	1.0066				Slope	0.998749	0.90 - 1.10
4.19	4.15	1.0099						
			Intercept	-0.021919	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 25, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	7:47	End Time (MST):	10:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999983	≥ 0.995			
8.02	8.02	0.9999						
4.01	3.98	1.0072				Slope	1.000498	0.90 - 1.10
2.00	1.98	1.0090						
			Intercept	-0.013049	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 25, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	7:47	End Time (MST):	10:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
8.80	8.76	1.0036						
4.40	4.37	1.0061				Slope	0.996921	0.90 - 1.10
2.19	2.17	1.0108						
			Intercept	-0.008470	± 0.5			



NMHC Calibration Plot

Date: February 15, 2024

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	February 28, 2024	Last Cal Date:	February 15, 2024
Start time (MST):	12:00	End time (MST):	13:46
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
C ₃ H ₈ Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
Removed C ₃ H ₈ Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.73E-04	2.73E-04	NMHC SP Ratio:	6.17E-05	6.17E-05
CH ₄ Retention time:	15.9	15.9	NMHC Peak Area:	142620	142620
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	16.64	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.78	1.002
Average Correction Factor					

Baseline Corr AF:	16.64	Prev response	16.78	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	8.71	1.010
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.86	0.993
Average Correction Factor					
Baseline Corr AF:	8.71	Prev response	8.76	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	7.93	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	7.93	1.012
Average Correction Factor					
Baseline Corr AF:	7.93	Prev response	8.01	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998749	
THC Cal Offset:	-0.021919	
CH ₄ Cal Slope:	1.000498	
CH ₄ Cal Offset:	-0.013049	
NMHC Cal Slope:	0.996921	
NMHC Cal Offset:	-0.008470	


Notes: Changed nitrogen cylinder.

Calibration Performed By: Braiden Boutillier

NMHC Calibration Plot

Date: February 28, 2024

Location: Mildred Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	February 9, 2024	Last Cal Date:	January 17, 2024
Start time (MST):	7:17	End time (MST):	10:03
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.87	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC446753			
Removed Cal Gas Conc:	50.87	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701		Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.986256	0.999358	Backgd or Offset:	23.8	24.1
Calibration intercept:	0.211572	0.554368	Coeff or Slope:	0.864	0.875

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4921	78.6	799.7	790.2	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.5	----
high point	4921	78.6	799.7	799.5	1.000
second point	4961	39.3	399.8	400.9	0.997
third point	4980	19.6	199.4	199.4	1.000
as left zero	5000	0.0	0.0	0.3	----
as left span	4921	78.6	799.7	802.2	0.997
Average Correction Factor					0.999

Baseline Corr As found:	790.00	Previous response	788.96	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Span adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 9, 2024	Previous Calibration:	January 17, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:17	End Time (MST):	10:03
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999997	≥0.995
799.7	799.5	1.0003			
399.8	400.9	0.9973	Slope	0.999358	0.90 - 1.10
199.4	199.4	1.0001			
			Intercept	0.554368	+/-30



SO2 Calibration Plot

Date: February 9, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	February 23, 2024	Last Cal Date:	January 11, 2024
Start time (MST):	7:20	End time (MST):	11:29
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.42	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC345266			
Removed Cal Gas Conc:	5.42	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701H		Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400
Converter make:	Global	Converter serial #:	2022-200
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996076	0.994653	Backgd or Offset:	1.87
Calibration intercept:	0.242122	0.142076	Coeff or Slope:	1.143
				1.98
				1.182

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4926	74.1	80.3	78.0	1.032
as found 2nd point	4963	37.0	40.1	39.2	1.028
as found 3rd point	4982	18.5	20.1	19.5	1.039
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4926	74.1	80.3	80.0	1.004
second point	4963	37.0	40.1	40.2	0.998
third point	4982	18.5	20.1	19.9	1.008
as left zero	5000	0.0	0.0	0.2	----
as left span	4926	74.1	80.3	82.9	0.969
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	1.003
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	77.8	Prev response:	80.25	*% change:	-3.1%
Baseline Corr 2nd AF pt:	39.0	AF Slope:	0.969328	AF Intercept:	0.181522
Baseline Corr 3rd AF pt:	19.3	AF Correlation:	0.999989		

* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 23, 2024	Previous Calibration:	January 11, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:20	End Time (MST):	11:29
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	≥0.995
80.3	80.0	1.0040		
40.1	40.2	0.9977	Slope	0.90 - 1.10
20.1	19.9	1.0076		
			Intercept	+/-3



H₂S Calibration Plot

Date: February 23, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: February 29, 2024 Last Cal Date: February 23, 2024
 Start time (MST): 7:15 End time (MST): 11:59
 Reason: Maintenance

Calibration Standards

Cal Gas Concentration: 5.42 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345266
 Removed Cal Gas Conc: 5.42 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701H Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1008841400
 Converter make: Global Converter serial #: 2022-200
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994653	0.997925	Backgd or Offset: 1.98	1.09
Calibration intercept:	0.142076	0.202175	Coeff or Slope: 1.182	1.130

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4926	74.1	80.3	76.8	1.047
as found 2nd point	4963	37.0	40.1	38.5	1.044
as found 3rd point	4982	18.5	20.1	19.1	1.055
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4926	74.1	80.3	80.3	1.000
second point	4963	37.0	40.1	40.4	0.993
third point	4982	18.5	20.1	20.1	0.998
as left zero	5000	0.0	0.0	0.2	----
as left span	4926	74.1	80.3	80.0	1.004

SO₂ Scrubber Check

Date of last scrubber change:	16-May-23	Ave Corr Factor	0.997
Date of last converter efficiency test:		efficiency	

Baseline Corr As found: 76.7 Prev response: 80.04 *% change: -4.3%
 Baseline Corr 2nd AF pt: 38.4 AF Slope: 0.955670 AF Intercept: 0.061201
 Baseline Corr 3rd AF pt: 19.0 AF Correlation: 0.999991

** = > +/-5% change initiates investigation*

Notes: Nightly span higher than the span through the filter. Filter holder and External Valve replaced.
 Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 29, 2024	Previous Calibration:	February 23, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:15	End Time (MST):	11:59
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999987	≥0.995
80.3	80.3	1.0003			
40.1	40.4	0.9928	Slope	0.997925	0.90 - 1.10
20.1	20.1	0.9976			
			Intercept	0.202175	+/-3



H₂S Calibration Plot

Date: February 29, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	February 6, 2024	Last Cal Date:	January 17, 2024
Start time (MST):	7:29	End time (MST):	8:37
Reason:	Removal		

Calibration Standards

Gas Cert Reference:	NA	Cal Gas Expiry Date:	March 10, 1931
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1222762077
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.10E-04	2.10E-04	NMHC SP Ratio:	4.60E-05
CH ₄ Retention time:	11.8	11.8	NMHC Peak Area:	191608
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	16.64	16.30	1.020
as found 2nd point	4961	39.3	8.32	8.00	1.040
as found 3rd point	4980	19.6	4.15	3.94	1.053
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

			Average Correction Factor		
Baseline Corr AF:	16.30	Prev response	16.58	*% change	-1.7%
Baseline Corr 2nd AF:	8.0	AF Slope:	0.981627	AF Intercept:	-0.081248
Baseline Corr 3rd AF:	3.9	AF Correlation:	0.999868	* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	8.82	8.67	1.017
as found 2nd point	4961	39.3	4.41	4.29	1.027
as found 3rd point	4980	19.6	2.20	2.12	1.040
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	
Baseline Corr AF:	8.67	Prev response	8.79	*% change	-1.3%
Baseline Corr 2nd AF:	4.3	AF Slope:	0.984795	AF Intercept:	-0.027881
Baseline Corr 3rd AF:	2.1	AF Correlation:	0.999951	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	7.82	7.63	1.024
as found 2nd point	4961	39.3	3.91	3.71	1.054
as found 3rd point	4980	19.6	1.95	1.82	1.070
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	
Baseline Corr AF:	7.63	Prev response	7.79	*% change	-2.1%
Baseline Corr 2nd AF:	3.71	AF Slope:	0.978507	AF Intercept:	-0.054166
Baseline Corr 3rd AF:	1.82	AF Correlation:	0.999713	* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000639	
THC Cal Offset:	-0.067561	
CH ₄ Cal Slope:	1.002348	
CH ₄ Cal Offset:	-0.043914	
NMHC Cal Slope:	0.999123	
NMHC Cal Offset:	-0.023647	


Notes: Removal due to pump not coming out to be replaced and ignitor connector on board breaking.

Calibration Performed By: Melissa Lemay

NMHC Calibration Plot

Date: February 6, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	February 6, 2024	Last Cal Date:	
Start time (MST):	11:00	End time (MST):	12:54
Reason:	Install		

Calibration Standards

Gas Cert Reference:	NA	Cal Gas Expiry Date:	March 10, 1931
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:		4.32E-04	NMHC SP Ratio:	1.12E-04
CH ₄ Retention time:		13.7	NMHC Peak Area:	78857
Zero Chromatogram:		OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	16.64	16.61	1.002
second point	4961	39.3	8.32	8.35	0.997
third point	4980	19.6	4.15	4.17	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	16.64	16.84	0.988
Average Correction Factor					0.998

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	8.82	8.79	1.003
second point	4961	39.3	4.41	4.44	0.992
third point	4980	19.6	2.20	2.23	0.988
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	8.82	8.98	0.983
Average Correction Factor					0.995
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	7.82	7.81	1.000
second point	4961	39.3	3.91	3.90	1.001
third point	4980	19.6	1.95	1.94	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	7.82	7.87	0.994
Average Correction Factor					1.001
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		0.997969
THC Cal Offset:		0.019863
CH ₄ Cal Slope:		0.999698
CH ₄ Cal Offset:		-0.002104
NMHC Cal Slope:		0.996112
NMHC Cal Offset:		0.022966

Notes: Install due to other analyzer not working.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	11:00	End Time (MST):	12:54
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999991	≥ 0.995			
16.64	16.61	1.0018						
8.32	8.35	0.9965				Slope	0.997969	0.90 - 1.10
4.15	4.17	0.9951						
			Intercept	0.019863	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	11:00	End Time (MST):	12:54
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	≥0.995
7.82	7.81	1.0005			
3.91	3.90	1.0010			
1.95	1.94	1.0027			
			Slope	0.999698	0.90 - 1.10
			Intercept	-0.002104	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	11:00	End Time (MST):	12:54
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999958	≥ 0.995			
8.82	8.79	1.0033						
4.41	4.44	0.9922				Slope	0.996112	0.90 - 1.10
2.20	2.23	0.9884						
			Intercept	0.022966	± 0.5			



NMHC Calibration Plot

Date: February 6, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: February 9, 2024 Last Cal Date: February 6, 2024
 Start time (MST): 7:17 End time (MST): 10:02
 Reason: Maintenance recalibrated after having a few days to settle

Calibration Standards

Gas Cert Reference: NA Cal Gas Expiry Date: March 10, 1931
 CH₄ Cal Gas Conc. 497.2 ppm CH₄ Equiv Conc. 1058.2 ppm
 C₃H₈ Cal Gas Conc. 204.0 ppm
 Removed Gas Cert: NA Removed Gas Expiry:
 Removed CH₄ Conc. 497.2 ppm CH₄ Equiv Conc. 1058.2 ppm
 Removed C₃H₈ Conc. 204.0 ppm Diff between cyl (THC):
 Diff between cyl (CH₄): Diff between cyl (NM):
 Calibrator Model: API T700 Serial Number: 3808
 ZAG make/model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 1426262594
 THC Range (ppm): 0 - 20 ppm
 NMHC Range (ppm): 0 - 10 ppm CH₄ Range (ppm): 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.32E-04	4.21E-04	NMHC SP Ratio:	1.12E-04
CH ₄ Retention time:	13.7	13.7	NMHC Peak Area:	78857
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	16.64	17.18	0.969
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	16.64	16.60	1.002
second point	4961	39.3	8.32	8.32	0.999
third point	4980	19.6	4.15	4.16	0.998
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	16.64	16.64	1.000
Average Correction Factor					1.000

Baseline Corr AF: 17.18 Prev response: 16.62 *% change: 3.2%
 Baseline Corr 2nd AF: NA AF Slope: AF Intercept:
 Baseline Corr 3rd AF: NA AF Correlation: * = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	8.82	9.13	0.966
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	8.82	8.80	1.003
second point	4961	39.3	4.41	4.42	0.997
third point	4980	19.6	2.20	2.22	0.992
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	8.82	8.81	1.001
Average Correction Factor					0.997
Baseline Corr AF:	9.13	Prev response	8.81	*% change	3.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	7.82	8.05	0.971
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	7.82	7.81	1.001
second point	4961	39.3	3.91	3.90	1.001
third point	4980	19.6	1.95	1.94	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	7.82	7.82	0.999
Average Correction Factor					1.003
Baseline Corr AF:	8.05	Prev response	7.81	*% change	3.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997969	0.997921
THC Cal Offset:	0.019863	0.010460
CH ₄ Cal Slope:	0.999698	0.999537
CH ₄ Cal Offset:	-0.002104	-0.004306
NMHC Cal Slope:	0.996112	0.996864
NMHC Cal Offset:	0.022966	0.013566

Notes: Recalibrated due to having a few days to settle. Span Adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 9, 2024	Previous Calibration:	February 6, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:17	End Time (MST):	10:02
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
16.64	16.60	1.0019						
8.32	8.32	0.9992				Slope	0.997921	0.90 - 1.10
4.15	4.16	0.9984						
			Intercept	0.010460	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 9, 2024	Previous Calibration:	February 6, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:17	End Time (MST):	10:02
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995
7.82	7.81	1.0008			
3.91	3.90	1.0015			
1.95	1.94	1.0058			
			Slope	0.999537	0.90 - 1.10
			Intercept	-0.004306	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 9, 2024	Previous Calibration:	February 6, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:17	End Time (MST):	10:02
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999988	≥ 0.995			
8.82	8.80	1.0026						
4.41	4.42	0.9973				Slope	0.996864	0.90 - 1.10
2.20	2.22	0.9925						
			Intercept	0.013566	± 0.5			



NMHC Calibration Plot

Date: February 9, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: February 12, 2024
Start time (MST): 7:45
Reason: Routine
Station number: AMS04
Last Cal Date: January 12, 2024
End time (MST): 12:52

Calibration Standards

NO Gas Cylinder #: CC324979
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: November 3, 1932
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 721

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.182	1.162	NO bkgnd or offset:	-0.6	-0.6
NOX coeff or slope:	1.169	1.154	NOX bkgnd or offset:	-0.3	-0.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.3	4.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000181	1.001294
NO _x Cal Offset:	-0.473674	-0.213331
NO Cal Slope:	1.002304	0.999626
NO Cal Offset:	-1.574438	-0.914344
NO ₂ Cal Slope:	0.987754	0.997058
NO ₂ Cal Offset:	-0.437615	-0.192985



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.0	----	----
as found span	4918	81.8	800.0	798.4	1.6	808.6	805.6	3.0	0.9894	0.9911
as found 2nd										
as found 3rd										
new cyl resp	4918	81.8	800.0	798.4	1.6					
calibrator zero	5000	0.0	0.0	0.0	0.0	0.5	0.8	-0.3	----	----
high point	4918	81.8	800.0	798.4	1.6	801.2	798.3	2.9	0.9985	1.0001
second point	4959	40.9	400.0	399.2	0.8	400.0	396.5	3.4	1.0000	1.0068
third point	4980	20.4	199.5	199.1	0.4	198.8	196.9	1.8	1.0035	1.0111
as left zero	5000	0.0	0.0	0.0	0.0	0.8	1.2	-0.4	----	----
as left span	4918	81.8	800.0	404.7	395.3	792.3	398.9	393.4	1.0098	1.0145
Average Correction Factor									1.0007	1.0060

Corrected As found	NO _x = 808.3 ppb	NO = 805.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.1%
Previous Response	NO _x = 799.7 ppb	NO = 798.7 ppb		*Percent Change	NO = 0.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.4	401.7	395.3	393.9	1.0036	99.6%
2nd GPT point (200 ppb O3)	795.4	600.9	196.1	195.5	1.0033	99.7%
3rd GPT point (100 ppb O3)	795.4	696.4	100.6	100.2	1.0044	99.6%
Average Correction Factor					1.0038	99.6%

Notes:

Span adjusted. No maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 12, 2024	Previous Calibration:	January 12, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:45	End Time (MST):	12:52
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.5	----	Correlation Coefficient	≥0.995	
800.0	801.2	0.9985			
400.0	400.0	1.0000			
199.5	198.8	1.0035			
			Slope	1.001294	0.90 - 1.10
			Intercept	-0.213331	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 12, 2024	Previous Calibration:	January 12, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:45	End Time (MST):	12:52
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.8	----	Correlation Coefficient	≥0.995	
798.4	798.3	1.0001			
399.2	396.5	1.0068			
199.1	196.9	1.0111			
			Slope	0.999626	0.90 - 1.10
			Intercept	-0.914344	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 12, 2024	Previous Calibration:	January 12, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	7:45	End Time (MST):	12:52
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
395.3	393.9	1.0036			
196.1	195.5	1.0033			
100.6	100.2	1.0044			
			Slope	0.997058	0.90 - 1.10
			Intercept	-0.192985	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 12, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: February 9, 2024 Last Cal Date: January 8, 2024
 Start time (MST): 10:01 End time (MST): 12:47
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 2961
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002571	0.999029	Backgd or Offset:	-3.5	-2.2
Calibration intercept:	0.800000	0.420000	Coeff or Slope:	1.011	1.011

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	1.0	----
as found span	5000	986.1	400.0	399.1	1.002
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.1	----
high point	5000	986.7	400.0	399.9	1.000
second point	5000	816.9	200.0	200.3	0.999
third point	5000	707.0	100.0	100.7	0.993
as left zero	5000	0.0	0.0	-0.1	----
as left span	5000	987.9	400.0	401.9	0.995
Average Correction Factor					0.997

Baseline Corr As found:	398.1	Previous response	401.8	*% change	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No Maintenance done. Zero adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 9, 2024	Previous Calibration:	January 8, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	10:01	End Time (MST):	12:47
Analyzer make:	API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999997	≥0.995
400.0	399.9	1.0003			
200.0	200.3	0.9985	Slope	0.999029	0.90 - 1.10
100.0	100.7	0.9930			
			Intercept	0.420000	+/- 5



O₃ Calibration Plot

Date: February 9, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: February 27, 2024 Last Cal Date: January 17, 2024
 Start time (MST): 9:18 End time (MST): 10:21

Analyzer Make: API T640 S/N: 321
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 1451
 Temp/RH standard: Alicat FP-25BT S/N: 1451

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-22	-22.9	-22	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	734.4	736.3	734.4	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.2	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	43	----	43	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	22.0	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 6-10-2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11	11	11	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 27, 2024
 Date Disposable Filter Changed: February 27, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: February 27, 2024
 Date RH/T Sensor Cleaned: February 27, 2024

Notes: Flow, Leak and PMT checked before and after cleaning. No adjustments done.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS05 MANNIX FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Mannix Station number: AMS05
 Calibration Date: February 2, 2024 Last Cal Date: January 19, 2024
 Start time (MST): 9:52 End time (MST): 13:45
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.02 ppm Cal Gas Exp Date: January 12, 2029
 Cal Gas Cylinder #: XC026809B
 Removed Cal Gas Conc: 50.02 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 621
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 1008841399
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003770	1.004227	Backgd or Offset:	9.1	9.1
Calibration intercept:	0.040000	-0.020000	Coeff or Slope:	0.944	0.944

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4920	80.0	800.3	800.1	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.6	----
high point	4920	80.0	800.3	804.2	0.995
second point	4960	40.0	400.2	400.9	0.998
third point	4980	20.0	200.1	200.7	0.997
as left zero	5000	0.0	0.0	0.9	----
as left span	4920	80.0	800.3	806.7	0.992
Average Correction Factor					0.997

Baseline Corr As found: 799.60 Previous response 803.38 *% change -0.5%
 Baseline Corr 2nd AF pt: NA AF Slope: AF Intercept:
 Baseline Corr 3rd AF pt: NA AF Correlation:

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 19, 2024
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:52	End Time (MST):	13:45
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999996	≥0.995
800.3	804.2	0.9952			
400.2	400.9	0.9982	Slope	1.004227	0.90 - 1.10
200.1	200.7	0.9969			
			Intercept	-0.020000	+/-30



SO2 Calibration Plot

Date: February 2, 2024

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mannix Station number: AMS05
 Calibration Date: February 15, 2024 Last Cal Date: January 10, 2024
 Start time (MST): 9:48 End time (MST): 14:13
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.96 ppm Cal Gas Exp Date: November 15, 2026
 Cal Gas Cylinder #: DT0037363
 Removed Cal Gas Conc: 4.96 ppm Rem Gas Exp Date:
 Removed Gas Cyl #: Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1845
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1200326169
 Converter make: Global Converter serial #: 2022225
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999261	1.008837	Backgd or Offset: 1.27	1.24
Calibration intercept:	0.182265	0.022150	Coeff or Slope: 1.022	0.998

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4919	80.6	80.0	82.7	0.967
as found 2nd point	4960	40.3	40.0	40.9	0.977
as found 3rd point	4980	20.2	20.0	20.7	0.968
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4919	80.6	80.0	80.7	0.991
second point	4960	40.3	40.0	40.4	0.989
third point	4980	20.2	20.0	20.1	0.997
as left zero	5000	0.0	0.0	0.1	----
as left span	4919	80.6	80.0	80.4	0.995
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.992
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 82.7 Prev response: 80.08 *% change: 3.2%
 Baseline Corr 2nd AF pt: 40.9 AF Slope: 1.033715 AF Intercept: -0.098418
 Baseline Corr 3rd AF pt: 20.7 AF Correlation: 0.999962

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber test after calibrator zero. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 10, 2024
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:48	End Time (MST):	14:13
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326169

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999992	≥0.995
80.0	80.7	0.9909			
40.0	40.4	0.9895	Slope	1.008837	0.90 - 1.10
20.0	20.1	0.9969			
			Intercept	0.022150	+/-3



H₂S Calibration Plot

Date: February 15, 2024

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	February 2, 2024	Last Cal Date:	January 19, 2024
Start time (MST):	9:52	End time (MST):	13:45
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.64E-05	2.71E-05	NMHC SP Ratio:	4.43E-05
CH ₄ Retention time:	15.2	15.4	NMHC Peak Area:	206470
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	16.96	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.02	----
high point	4920	80.0	17.23	17.19	1.002
second point	4960	40.0	8.61	8.60	1.001
third point	4980	20.0	4.31	4.31	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.26	0.998
Average Correction Factor					1.001

Baseline Corr AF:	16.96	Prev response	17.23	*% change	-1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	9.15	9.10	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.18	0.997
second point	4960	40.0	4.57	4.59	0.996
third point	4980	20.0	2.29	2.31	0.992
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.15	9.22	0.993
Average Correction Factor					0.995
Baseline Corr AF:	9.10	Prev response	9.15	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	7.86	1.028
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.02	----
high point	4920	80.0	8.08	8.01	1.009
second point	4960	40.0	4.04	4.01	1.008
third point	4980	20.0	2.02	2.00	1.010
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.04	1.004
Average Correction Factor					1.009
Baseline Corr AF:	7.86	Prev response	8.08	*% change	-2.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000521	0.996805
THC Cal Offset:	-0.007800	0.014200
CH ₄ Cal Slope:	1.001401	0.989588
CH ₄ Cal Offset:	-0.007000	0.009000
NMHC Cal Slope:	0.999969	1.002817
NMHC Cal Offset:	-0.001200	0.005400

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 19, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:52	End Time (MST):	13:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.02	----	Correlation Coefficient	1.000000	≥ 0.995			
17.23	17.19	1.0024						
8.61	8.60	1.0014				Slope	0.996805	0.90 - 1.10
4.31	4.31	1.0003						
			Intercept	0.014200	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 19, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:52	End Time (MST):	13:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.02	----	Correlation Coefficient	0.999997	≥0.995			
8.08	8.01	1.0092						
4.04	4.01	1.0080				Slope	0.989588	0.90 - 1.10
2.02	2.00	1.0103						
			Intercept	0.009000	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 19, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:52	End Time (MST):	13:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
9.15	9.18	0.9968						
4.57	4.59	0.9960				Slope	1.002817	0.90 - 1.10
2.29	2.31	0.9917						
			Intercept	0.005400	± 0.5			



NMHC Calibration Plot

Date: February 2, 2024

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	February 8, 2024	Last Cal Date:	February 2, 2024
Start time (MST):	11:05	End time (MST):	14:05
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.71E-05	2.70E-05	NMHC SP Ratio:	4.49E-05
CH ₄ Retention time:	15.4	15.6	NMHC Peak Area:	203563
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				208017
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.22	1.000
second point	4960	40.0	8.61	8.60	1.002
third point	4980	20.0	4.31	4.28	1.006
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	16.93	1.017
Average Correction Factor					1.003

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.13	1.002
second point	4960	40.0	4.57	4.57	1.002
third point	4980	20.0	2.29	2.28	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.15	8.98	1.019
Average Correction Factor					1.003
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.09	0.999
second point	4960	40.0	4.04	4.03	1.002
third point	4980	20.0	2.02	2.00	1.008
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	7.94	1.017
Average Correction Factor					1.003
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.996805	1.000056
THC Cal Offset:	0.014200	-0.011800
CH ₄ Cal Slope:	0.989588	1.001500
CH ₄ Cal Offset:	0.009000	-0.009600
NMHC Cal Slope:	1.002817	0.998520
NMHC Cal Offset:	0.005400	-0.002400

Notes: Pump was dead upon arrival. Changed the pump and adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 8, 2024	Previous Calibration:	February 2, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	11:05	End Time (MST):	14:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995
17.23	17.22	1.0003			
8.61	8.60	1.0016			
4.31	4.28	1.0060			
			Slope	1.000056	$0.90 - 1.10$
			Intercept	-0.011800	± 0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 8, 2024	Previous Calibration:	February 2, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	11:05	End Time (MST):	14:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999994	≥0.995
8.08	8.09	0.9991			
4.04	4.03	1.0023			
2.02	2.00	1.0078			
			Slope	1.001500	0.90 - 1.10
			Intercept	-0.009600	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 8, 2024	Previous Calibration:	February 2, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	11:05	End Time (MST):	14:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
9.15	9.13	1.0017						
4.57	4.57	1.0017				Slope	0.998520	0.90 - 1.10
2.29	2.28	1.0043						
			Intercept	-0.002400	± 0.5			



NMHC Calibration Plot

Date: February 8, 2024

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	February 27, 2024	Last Cal Date:	February 8, 2024
Start time (MST):	10:01	End time (MST):	15:13
Reason:	Maintenance and cylinder change		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.70E-05	2.98E-05	NMHC SP Ratio:	4.40E-05
CH ₄ Retention time:	15.6	16.4	NMHC Peak Area:	208017
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	16.21	1.063
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.24	0.999
second point	4960	40.0	8.61	8.63	0.998
third point	4980	20.0	4.31	4.32	0.996
as left zero	5000	0.0	0.00	0.02	----
as left span	4920	80.0	17.23	17.27	0.997
Average Correction Factor					0.998

Baseline Corr AF:	16.21	Prev response	17.22	*% change	-6.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	9.15	9.00	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.16	0.999
second point	4960	40.0	4.57	4.56	1.003
third point	4980	20.0	2.29	2.28	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.15	9.12	1.003
Average Correction Factor					1.001
Baseline Corr AF:	9.00	Prev response	9.13	*% change	-1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	7.21	1.120
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.08	1.000
second point	4960	40.0	4.04	4.07	0.991
third point	4980	20.0	2.02	2.04	0.990
as left zero	5000	0.0	0.00	0.02	----
as left span	4920	80.0	8.08	8.16	0.990
Average Correction Factor					0.994
Baseline Corr AF:	7.21	Prev response	8.08	*% change	-12.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000056	1.000620
THC Cal Offset:	-0.011800	0.007200
CH ₄ Cal Slope:	1.001500	0.999873
CH ₄ Cal Offset:	-0.009600	0.014400
NMHC Cal Slope:	0.998520	1.001368
NMHC Cal Offset:	-0.002400	-0.006800

Notes: CH₄ channel drifted down and the instrument started dipping. Completed an investigation. No multipoint as founds since as found span is outside of limits. Changed the N₂ cylinder, increased the zag pressure, adjusted window timings and the span. See docit note for details.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	February 8, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:01	End Time (MST):	15:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995
17.23	17.24	0.9992			
8.61	8.63	0.9980			
4.31	4.32	0.9962			
			Slope	1.000620	0.90 - 1.10
			Intercept	0.007200	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	February 8, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:01	End Time (MST):	15:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999977	≥0.995
8.08	8.08	0.9998			
4.04	4.07	0.9915			
2.02	2.04	0.9905			
			Slope	0.999873	0.90 - 1.10
			Intercept	0.014400	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	February 8, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:01	End Time (MST):	15:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995			
9.15	9.16	0.9986						
4.57	4.56	1.0030				Slope	1.001368	0.90 - 1.10
2.29	2.28	1.0017						
			Intercept	-0.006800	± 0.5			



NMHC Calibration Plot

Date: February 27, 2024

Location: Mannix





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	February 6, 2024	Last Cal Date:	January 5, 2024
Start time (MST):	10:30	End time (MST):	14:16
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.78	ppm	Cal Gas Exp Date:	September 9, 2024
Cal Gas Cylinder #:	AAL070632			
Removed Cal Gas Conc:	49.78	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	5608

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999660	0.996386	Backgd or Offset:	17.7	17.7
Calibration intercept:	1.240714	1.161262	Coeff or Slope:	0.922	0.922

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.3	799.5	798.2	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.3	799.5	797.4	1.003
second point	4960	40.2	400.2	399.8	1.001
third point	4980	20.1	200.1	202.2	0.990
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	80.3	799.5	798.2	1.002
Average Correction Factor					0.998

Baseline Corr As found:	798.20	Previous response	800.44	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 5, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:30	End Time (MST):	14:16
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	0.999988	
799.5	797.4	1.0026			≥0.995
400.2	399.8	1.0011	Slope	0.996386	
200.1	202.2	0.9897			0.90 - 1.10
			Intercept	1.161262	+/-30



SO2 Calibration Plot

Date: February 6, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	February 7, 2024	Last Cal Date:	January 18, 2024
Start time (MST):	9:56	End time (MST):	14:18
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.328	ppm	Cal Gas Exp Date:	February 14, 2025
Cal Gas Cylinder #:	CC506659			
Removed Cal Gas Conc:	5.328	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002314	0.994029	Backgd or Offset:	1.95	1.95
Calibration intercept:	0.440228	0.280242	Coeff or Slope:	1.150	1.150

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4925	75.1	80.0	79.6	1.007
as found 2nd point	4963	37.5	40.0	40.3	0.994
as found 3rd point	4981	18.8	20.0	20.4	0.987
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4925	75.1	80.0	79.7	1.004
second point	4963	37.5	40.0	40.2	0.994
third point	4981	18.8	20.0	20.3	0.987
as left zero	5000	0.0	0.0	0.2	----
as left span	4925	75.1	80.0	79.9	1.002
SO2 Scrubber Check	4920	80.3	803.0	0.1	----
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	0.995
Date of last converter efficiency test:	efficiency				

Baseline Corr As found:	79.5	Prev response:	80.65	*% change:	-1.4%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	0.992457	AF Intercept:	0.360275
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999940		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 7, 2024	Previous Calibration:	January 18, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:56	End Time (MST):	14:18
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999971	
80.0	79.7	1.0041			≥0.995
40.0	40.2	0.9939	Slope	0.994029	
20.0	20.3	0.9869			0.90 - 1.10
			Intercept	0.280242	+/-3



TRS Calibration Plot

Date: February 7, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	February 6, 2024	Last Cal Date:	January 5, 2024
Start time (MST):	10:30	End time (MST):	14:15
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH ₄ Cal Gas Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.17E-04	2.15E-04	NMHC SP Ratio:	4.81E-05
CH ₄ Retention time:	14.0	14.2	NMHC Peak Area:	188492
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.12	17.09	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	17.05	1.004
second point	4960	40.2	8.57	8.52	1.006
third point	4980	20.1	4.29	4.31	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.12	16.88	1.014
Average Correction Factor					1.002

Baseline Corr AF:	17.08	Prev response	17.20	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919.7	80.3	9.07	9.08	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	9.07	9.04	1.003
second point	4960	40.2	4.54	4.53	1.002
third point	4980	20.1	2.27	2.30	0.987
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	9.07	8.93	1.015
Average Correction Factor					0.997
Baseline Corr AF:	9.08	Prev response	9.11	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919.7	80.3	8.06	8.00	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	8.06	8.01	1.006
second point	4960	40.2	4.03	3.99	1.011
third point	4980	20.1	2.02	2.01	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	8.06	7.95	1.013
Average Correction Factor					1.007
Baseline Corr AF:	8.00	Prev response	8.09	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000106	0.994828
THC Cal Offset:	0.074604	0.012664
CH ₄ Cal Slope:	0.999043	0.993816
CH ₄ Cal Offset:	0.039427	-0.003140
NMHC Cal Slope:	1.001025	0.995576
NMHC Cal Offset:	0.034777	0.016404

Notes: Changed the inlet filter and N2 cylinder after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 5, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:30	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999990	≥ 0.995			
17.12	17.05	1.0042						
8.57	8.52	1.0064				Slope	0.994828	0.90 - 1.10
4.29	4.31	0.9953						
			Intercept	0.012664	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 5, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:30	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999991	≥0.995
8.06	8.01	1.0058			
4.03	3.99	1.0110			
2.02	2.01	1.0047			
			Slope	0.993816	0.90 - 1.10
			Intercept	-0.003140	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 5, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:30	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999981	≥ 0.995			
9.07	9.04	1.0029						
4.54	4.53	1.0024				Slope	0.995576	0.90 - 1.10
2.27	2.30	0.9868						
			Intercept	0.016404	± 0.5			



NMHC Calibration Plot

Date: February 6, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Patricia McInnes
Calibration Date: February 5, 2024
Start time (MST): 10:25
Reason: Routine
Station number: AMS06
Last Cal Date: January 8, 2024
End time (MST): 15:05

Calibration Standards

NO Gas Cylinder #: T30YCWN
NOX Cal Gas Conc: 47.94 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 47.94 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: April 11, 2025
NO Cal Gas Conc: 46.39 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 46.39 ppm
NO gas Diff:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1172750022

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.825	0.825	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.987	0.987	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.2	157.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998293	1.000023
NO _x Cal Offset:	1.916153	2.315635
NO Cal Slope:	0.999944	1.002417
NO Cal Offset:	1.062930	1.022423
NO ₂ Cal Slope:	0.999549	0.999140
NO ₂ Cal Offset:	-1.654321	-0.524585



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	0.2	-0.4	----	----
as found span	4914	86.2	826.5	799.7	26.7	829.1	799.5	29.7	0.9968	1.0003
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.4	-0.2	----	----
high point	4914	86.2	826.5	799.7	26.7	827.8	802.4	25.5	0.9984	0.9967
second point	4957	43.1	413.2	399.9	13.4	416.5	402.1	14.4	0.9922	0.9945
third point	4978	21.6	207.1	200.4	6.7	211.5	202.6	8.9	0.9793	0.9892
as left zero	5000	0.0	0.0	0.0	0.0	0.3	0.5	-0.3	----	----
as left span	4914	86.2	826.5	401.9	424.5	826.1	404.2	422.0	1.0004	0.9944
Average Correction Factor									0.9899	0.9935

Corrected As found	NO _x = 829.3 ppb	NO = 799.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.3%
Previous Response	NO _x = 827.0 ppb	NO = 800.8 ppb		*Percent Change	NO = -0.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	800.8	403.0	424.5	423.8	1.0017	99.8%
2nd GPT point (200 ppb O3)	800.8	603.0	224.5	223.7	1.0037	99.6%
3rd GPT point (100 ppb O3)	800.8	703.5	124.0	123.0	1.0083	99.2%
Average Correction Factor					1.0046	99.5%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 8, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:25	End Time (MST):	15:05
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	≥0.995	
826.5	827.8	0.9984			
413.2	416.5	0.9922			
207.1	211.5	0.9793			
			Slope	1.000023	0.90 - 1.10
			Intercept	2.315635	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 8, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:25	End Time (MST):	15:05
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	≥0.995	
799.7	802.4	0.9967			
399.9	402.1	0.9945			
200.4	202.6	0.9892			
			Slope	1.002417	0.90 - 1.10
			Intercept	1.022423	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 8, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:25	End Time (MST):	15:05
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
424.5	423.8	1.0017		
224.5	223.7	1.0037		
124.0	123.0	1.0083		
			0.999997	
			0.999140	
			-0.524585	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 5, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Patricia McInnes Station number: AMS06
 Calibration Date: February 16, 2024 Last Cal Date: January 2, 2024
 Start time (MST): 9:41 End time (MST): 13:00
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3566
 ZAG Make/Model: API T701 Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1300156234
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003886	1.003629	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	-1.480000	-0.860000	Coeff or Slope:	1.026	1.026

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	-0.1	----
as found span	5000	1303.0	400.0	401.1	0.997
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.0	----
high point	5000	1303.0	400.0	401.1	0.997
second point	5000	966.5	200.0	199.2	1.004
third point	5000	794.3	100.0	98.8	1.012
as left zero	5000	800.0	0.0	-0.2	----
as left span	5000	1303.0	400.0	403.2	0.992
Average Correction Factor					1.004

Baseline Corr As found:	401.2	Previous response	400.1	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 16, 2024	Previous Calibration:	January 2, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:41	End Time (MST):	13:00
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999978	
400.0	401.1	0.9973			≥0.995
200.0	199.2	1.0040	Slope	1.003629	
100.0	98.8	1.0121			0.90 - 1.10
			Intercept	-0.860000	+/- 5



O₃ Calibration Plot

Date: February 16, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: February 16, 2024 Last Cal Date: January 18, 2024
 Start time (MST): 14:17 End time (MST): 14:39

Analyzer Make: API T640 S/N: 766
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388755
 Temp/RH standard: Alicat FP-25BT S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	3.6	3.1	3.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	728.1	730.7	728.1	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	5.11	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39	----	39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	4.8	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date:
Lot No.:

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: December 1, 2023
Date Disposable Filter Changed: December 1, 2023

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023
Date RH/T Sensor Cleaned: April 13, 2023

Notes: Quarterly calibrations completed in December. Leak check passed, no adjustments made.

Calibration by: Max Farrell



Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	February 21, 2024	Last Cal Date:	January 16, 2024
Start time (MST):	8:55	End time (MST):	13:13
NH3 Cal Date:	February 21, 2024	Last Cal Date:	January 16, 2024
Start time (MST):	13:30	End time (MST):	15:15
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	NA
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.29	ppm	NH3 Gas Cylinder #:	EB0108520
Removed NH3 Conc:	76.29	ppm	NH3 Cal Gas Expiry:	August 22, 2024
NH3 gas Diff:			Removed Cylinder #:	NA
Calibrator Model:	API T700		Removed cyl Expiry:	NA
ZAG make/model:	API T701		Serial Number:	3566
			Serial Number:	4602

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	808
Converter model:	Teledyne API T501	Converter serial #:	484
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	4.80
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	26.6

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.854	0.854	TN coefficient:	0.849	0.849
NOX coefficient:	0.848	0.848	NO bkgrnd:	-0.985	-0.985
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	-0.562	-0.562
NH3 coefficient:	0.891	0.891	TN bkgrnd:	5.018	5.018

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000699	0.995140
NO _x Cal Offset:	1.395952	2.256598
NO Cal Slope:	1.003275	0.997671
NO Cal Offset:	0.522150	1.183452
NO ₂ Cal Slope:	0.994054	0.997218
NO ₂ Cal Offset:	-1.155779	0.194612
NH3 Cal Slope:	1.000855	1.010551
NH3 Cal Offset:	2.210125	1.110139
TN Cal Slope:	1.008599	1.017335
TN Cal Offset:	2.771030	2.438080



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-1.8	0.1	-1.8	----	----
as found NO	4914	86.2	826.5	826.5	----	825.5	826.0	-0.7	1.001	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-1.1	0.5	-1.6	----	----
high NO point	4914	86.2	826.5	826.5	----	826.4	824.0	2.2	1.000	----
NO/O3 point	4914	86.2	826.5	826.5	----	822.6	819.6	2.8	1.005	----
as found NH3	3417	82.6	1800.6	----	1800.6	1833.6	----	1820.6	0.982	0.989
new NH3 cyl rp	3417	82.6	1800.6	----	1800.6	----	----	----	----	----
first NH3	3417	82.6	1800.6	----	1800.6	1833.6	----	1820.6	0.982	0.989
second NH3	3454	45.9	1000.5	----	1000.5	1018.6	----	1009.9	0.982	0.991
third NH3	3477	22.9	499.2	----	499.2	516.2	----	510.7	0.967	0.977
Average Correction Factor									1.0024	0.9857

Corrected As found TN = 827.3 ppb NO_x = 825.9 ppb NH3 = 1822.4 ppb

Previous Response TN = 836.3 ppb NO_x = 828.4 ppb NH3 = 1804.4 ppb

NH3 Previous Converter Efficiency = 89.1%

NH3 Current Converter Efficiency = 89.1%

*Percent Change TN = -1.1%

*Percent Change NO_x = -0.3%

*Percent Change NH3 = 1.0%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	-1.8	----	----
as found span	4914	86.2	826.5	799.7	826.5	826.0	798.7	825.5	1.0005	1.0013
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.5	0.3	-1.1	----	----
high point	4914	86.2	826.5	799.7	826.5	824.0	798.4	826.4	1.0030	1.0017
second point	4957	43.1	413.2	399.9	413.2	413.8	401.2	416.5	0.9986	0.9967
third point	4978	21.6	207.1	200.4	207.1	210.5	201.6	209.6	0.9839	0.9942
Average Correction Factor									0.9952	0.9975

Baseline Corr As fnd	TN = 827.3 ppb	NO _x = 825.9 ppb	NO = 798.6 ppb	*Percent Change	TN = -1.1%
Previous Response	TN = 836.3 ppb	NO _x = 828.4 ppb	NO = 802.9 ppb	*Percent Change	NO _x = -0.3%
				*Percent Change	NO = -0.5%

* = > +/-5% change initiates investigation

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	0.0	----	----
calibration zero	----	----	0.0	0.2	----	----
1st GPT point (400 ppb O3)	794.2	396.0	424.9	423.8	1.0026	99.7%
2nd GPT point (200 ppb O3)	794.2	597.9	223.0	223.0	1.0001	100.0%
3rd GPT point (100 ppb O3)	794.2	699.3	121.6	121.2	1.0035	99.7%
Average Correction Factor					1.0021	99.8%

Notes:

Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By:

Max Farrell



Wood Buffalo Environmental Association

TN Calibration Summary


Version-05-2023

Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 16, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:55	End Time (MST):	13:13
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-1.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
1800.6	1833.6	0.9820		
1000.5	1018.6	0.9822		
499.2	516.2	0.9670		
			0.999972	
			1.017335	
			2.438080	





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 16, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:55	End Time (MST):	13:13
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-1.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.6	1820.6	0.9890		
1000.5	1009.9	0.9907		
499.2	510.7	0.9774		
			0.999978	
			1.010551	
			1.110139	





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 16, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:55	End Time (MST):	13:13
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	≥0.995
826.5	824.0	1.0030		
413.2	413.8	0.9986	Slope	0.90 - 1.10
207.1	210.5	0.9839		
			Intercept	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 16, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:55	End Time (MST):	13:13
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
799.7	798.4	1.0017		
399.9	401.2	0.9967		
200.4	201.6	0.9942		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 16, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:55	End Time (MST):	13:13
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
424.9	423.8	1.0026			
223.0	223.0	1.0001			
121.6	121.2	1.0035			
			Slope	0.997218	0.90 - 1.10
			Intercept	0.194612	+/-20



NO_x Calibration Plot

Date: February 21, 2024


Location: Patricia McInnes



NH₃ Calibration Plot

Date: February 21, 2024

Location: Patricia McInnes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	February 2, 2024	Last Cal Date:	January 18, 2024
Start time (MST):	10:13	End time (MST):	14:36
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.06	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC320556			
Removed Cal Gas Conc:	50.06	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 701H		Serial Number:	198

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1507864683
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003653	1.002036	Backgd or Offset:	2.62	2.70
Calibration intercept:	2.143576	1.783896	Coeff or Slope:	0.846	0.852

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	79.8	799.0	806.9	0.990
as found 2nd point	4960	39.9	399.5	407.4	0.981
as found 3rd point	4980	20.0	200.2	205.1	0.976
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	79.8	799.0	801.0	0.997
second point	4960	39.9	399.5	404.6	0.987
third point	4980	20.0	200.2	203.0	0.986
as left zero	5000	0.0	0.0	0.0	----
as left span	4920	79.8	799.0	801.8	0.996
Average Correction Factor					0.990

Baseline Corr As found:	806.90	Previous response	804.05	*% change	0.4%
Baseline Corr 2nd AF pt:	407.40	AF Slope:	1.009233	AF Intercept:	1.942240
Baseline Corr 3rd AF pt:	205.10	AF Correlation:	0.999966		

* = > +/-5% change initiates investigation

Notes: Pump swapped out after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 18, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:13	End Time (MST):	14:36
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999967	≥0.995
799.0	801.0	0.9975			
399.5	404.6	0.9874	Slope	1.002036	0.90 - 1.10
200.2	203.0	0.9864			
			Intercept	1.783896	+/-30



SO2 Calibration Plot

Date: February 2, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: February 22, 2024 Last Cal Date: January 9, 2024
 Start time (MST): 11:02 End time (MST): 15:40
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.25 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC504080
 Removed Cal Gas Conc: 5.25 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3805
 ZAG Make/Model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE Analyzer serial #: 1180540018
 Converter make: CDN-101 Converter serial #: 551
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004826	1.017941	Backgd or Offset: 2.40	2.40
Calibration intercept:	0.237795	0.157770	Coeff or Slope: 0.906	0.901

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4925	75.5	79.3	81.2	0.977
as found 2nd point	4962	37.7	39.6	40.4	0.982
as found 3rd point	4981	18.9	19.8	20.3	0.982
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4925	75.5	79.3	80.9	0.980
second point	4962	37.7	39.6	40.5	0.978
third point	4981	18.9	19.9	20.3	0.978
as left zero	5000	0.0	0.0	0.4	----
as left span	4925	75.5	79.3	80.1	0.990
SO2 Scrubber Check	4920	79.2	800.4	0.1	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	0.979
Date of last converter efficiency test:	April 22, 2022			92.6% efficiency	

Baseline Corr As found: 81.1 Prev response: 79.89 *% change: 1.5%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 1.023284 AF Intercept: 0.017677
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999992

* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 22, 2024	Previous Calibration:	January 9, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:02	End Time (MST):	15:40
Analyzer make:	CDN-101	Analyzer serial #:	551

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999998	≥0.995
79.3	80.9	0.9804			
39.6	40.5	0.9780	Slope	1.017941	0.90 - 1.10
19.9	20.3	0.9782			
			Intercept	0.157770	+/-3



TRS Calibration Plot

Date: February 22, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	February 2, 2024	Last Cal Date:	January 18, 2024
Start time (MST):	10:13	End time (MST):	14:36
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	N/A	2.49E-04	NMHC SP Ratio:	N/A	4.43E-05
CH ₄ Retention time:	N/A	13.8	NMHC Peak Area:	N/A	203131
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	16.91	16.59	1.019
as found 2nd point	4960	39.9	8.46	8.30	1.019
as found 3rd point	4980	20.0	4.24	4.13	1.026
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	16.91	17.03	0.993
second point	4960	39.9	8.46	8.51	0.993
third point	4980	20.0	4.24	4.24	0.999
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	16.91	17.03	0.993

Average Correction Factor				0.995
Baseline Corr AF:	16.59	Prev response	17.02	*% change -2.6%
Baseline Corr 2nd AF:	8.3	AF Slope:	0.981619	AF Intercept: -0.009252
Baseline Corr 3rd AF:	4.1	AF Correlation:	0.999996	* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.00	8.80	1.022
as found 2nd point	4960	39.9	4.50	4.44	1.014
as found 3rd point	4980	20.0	2.26	2.24	1.008
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.00	9.11	0.988
second point	4960	39.9	4.50	4.59	0.981
third point	4980	20.0	2.26	2.31	0.975
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.00	9.11	0.988
Average Correction Factor					0.981
Baseline Corr AF:	8.80	Prev response	9.11	*% change	-3.5%
Baseline Corr 2nd AF:	4.4	AF Slope:	0.977272	AF Intercept:	0.019843
Baseline Corr 3rd AF:	2.2	AF Correlation:	0.999973	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	7.92	7.79	1.016
as found 2nd point	4960	39.9	3.96	3.86	1.025
as found 3rd point	4980	20.0	1.98	1.90	1.046
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	7.92	7.92	0.999
second point	4960	39.9	3.96	3.93	1.008
third point	4980	20.0	1.98	1.93	1.028
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	7.92	7.92	1.000
Average Correction Factor					1.012
Baseline Corr AF:	7.79	Prev response	7.91	*% change	-1.5%
Baseline Corr 2nd AF:	3.86	AF Slope:	0.986387	AF Intercept:	-0.029495
Baseline Corr 3rd AF:	1.90	AF Correlation:	0.999934	* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.006214	1.007382
THC Cal Offset:	-0.000562	-0.009960
CH ₄ Cal Slope:	1.002531	1.002574
CH ₄ Cal Offset:	-0.029927	-0.029327
NMHC Cal Slope:	1.009365	1.011409
NMHC Cal Offset:	0.028966	0.019168

Notes: H2 swapped out after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 18, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:13	End Time (MST):	14:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
16.91	17.03	0.9931						
8.46	8.51	0.9933				Slope	1.007382	0.90 - 1.10
4.24	4.24	0.9988						
			Intercept	-0.009960	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 18, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:13	End Time (MST):	14:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999937	≥0.995
7.92	7.92	0.9993			
3.96	3.93	1.0084			
1.98	1.93	1.0280			
			Slope	1.002574	0.90 - 1.10
			Intercept	-0.029327	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 18, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:13	End Time (MST):	14:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999977	≥ 0.995			
9.00	9.11	0.9880						
4.50	4.59	0.9806				Slope	1.011409	0.90 - 1.10
2.26	2.31	0.9749						
			Intercept	0.019168	± 0.5			



NMHC Calibration Plot

Date: February 2, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	February 20, 2024	Last Cal Date:	NA
Start time (MST):	11:15	End time (MST):	13:25
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	496.0 ppm	CH4 Equiv Conc.	1059.8 ppm
C3H8 Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	496.0 ppm	CH4 Equiv Conc.	1059.8 ppm
Removed C3H8 Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	N/A	2.45E-04	NMHC SP Ratio:	N/A	4.93E-05
CH4 Retention time:	N/A	13.4	NMHC Peak Area:	N/A	182375
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	16.91	16.95	0.998
second point	4960	39.9	8.46	8.50	0.995
third point	4980	20.0	4.24	4.28	0.991
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	16.91	16.91	1.000
Average Correction Factor					0.995

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.00	9.04	0.995
second point	4960	39.9	4.50	4.55	0.990
third point	4980	20.0	2.26	2.29	0.985
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.00	9.02	0.998
Average Correction Factor					0.990
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	7.92	7.91	1.001
second point	4960	39.9	3.96	3.95	1.002
third point	4980	20.0	1.98	1.99	0.998
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	7.92	7.89	1.003
Average Correction Factor					1.000
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	NA	1.001667
THC Cal Offset:	NA	0.015848
CH ₄ Cal Slope:	NA	0.999012
CH ₄ Cal Offset:	NA	0.001263
NMHC Cal Slope:	NA	1.004116
NMHC Cal Offset:	NA	0.014385

Notes:

Installation calibration

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 20, 2024	Previous Calibration:	NA
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:15	End Time (MST):	13:25
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
16.91	16.95	0.9979						
8.46	8.50	0.9954				Slope	1.001667	0.90 - 1.10
4.24	4.28	0.9911						
			Intercept	0.015848	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 20, 2024	Previous Calibration:	NA
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:15	End Time (MST):	13:25
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥0.995
7.92	7.91	1.0007			
3.96	3.95	1.0018			
1.98	1.99	0.9980			
			Slope	0.999012	0.90 - 1.10
			Intercept	0.001263	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 20, 2024	Previous Calibration:	NA
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:15	End Time (MST):	13:25
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999987	≥ 0.995			
9.00	9.04	0.9953						
4.50	4.55	0.9898				Slope	1.004116	0.90 - 1.10
2.26	2.29	0.9851						
			Intercept	0.014385	± 0.5			



NMHC Calibration Plot

Date: February 20, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Athabasca Valley Station number: AMS07
Calibration Date: February 27, 2024 Last Cal Date: January 5, 2024
Start time (MST): 11:15 End time (MST): 16:55
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0033919 Cal Gas Expiry Date: January 9, 2032
NOX Cal Gas Conc: 59.90 ppm NO Cal Gas Conc: 59.90 ppm
Removed Cylinder #: T2Y1KA4 Removed Gas Exp Date: November 30, 2023
Removed Gas NOX Conc: 50.92 ppm Removed Gas NO Conc: 49.92 ppm
NOX gas Diff: 1.1% NO gas Diff: 0.4%
Calibrator Model: API T700 Serial Number: 3805
ZAG make/model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1160120024
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.092	1.079	NO bkgnd or offset:	7.7	7.6
NOX coeff or slope:	0.994	0.990	NOX bkgnd or offset:	7.9	7.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	216.3	214.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998118	0.995415
NO _x Cal Offset:	1.478599	2.191943
NO Cal Slope:	1.000317	0.999970
NO Cal Offset:	1.135182	2.371891
NO ₂ Cal Slope:	1.000671	0.995591
NO ₂ Cal Offset:	0.394845	-1.458446



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
as found span	4920	80.2	816.7	800.7	16.0	824.1	806.0	18.2	0.9910	0.9934
as found 2nd										
as found 3rd										
new cyl resp	4933	66.8	800.3	800.3	0.0	816.3	808.6	7.6	0.9804	0.9897
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4933	66.8	800.3	800.3	0.0	797.2	800.8	-3.8	1.0039	0.9994
second point	4966	33.4	400.2	400.2	0.0	403.2	405.7	-2.4	0.9925	0.9864
third point	4983	16.7	200.1	200.1	0.0	202.5	203.5	-1.0	0.9880	0.9832
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
as left span	4933	66.8	800.3	399.5	400.8	795.7	400.2	395.6	1.0058	0.9982
Average Correction Factor									0.9948	0.9896

Corrected As found	NO _x = 824.3 ppb	NO = 806.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.9%
Previous Response	NO _x = 816.7 ppb	NO = 802.1 ppb		*Percent Change	NO = 0.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	800.5	399.7	400.8	398.1	1.0068	99.3%
2nd GPT point (200 ppb O3)	800.5	603.3	197.2	194.7	1.0128	98.7%
3rd GPT point (100 ppb O3)	800.5	703.0	97.5	93.8	1.0394	96.2%
Average Correction Factor					1.0197	98.1%

Notes: Calibration gas cylinder changed out as part of QA audit follow-up. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	January 5, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:15	End Time (MST):	16:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
800.3	797.2	1.0039			
400.2	403.2	0.9925			
200.1	202.5	0.9880			
			Correlation Coefficient	0.999954	≥0.995
			Slope	0.995415	0.90 - 1.10
			Intercept	2.191943	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	January 5, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:15	End Time (MST):	16:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
800.3	800.8	0.9994			
400.2	405.7	0.9864			
200.1	203.5	0.9832			
			Correlation Coefficient	0.999943	≥0.995
			Slope	0.999970	0.90 - 1.10
			Intercept	2.371891	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	January 5, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:15	End Time (MST):	16:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
400.8	398.1	1.0068			
197.2	194.7	1.0128			
97.5	93.8	1.0394			
			Slope	0.995591	0.90 - 1.10
			Intercept	-1.458446	+/-20



NO_x Calibration Plot

Date: February 27, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: February 16, 2024 Last Cal Date: January 12, 2024
 Start time (MST): 10:24 End time (MST): 14:22
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: T700 Serial Number: 3805
 ZAG Make/Model: T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1152220023
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995143	0.993286	Backgd or Offset:	-1.6	-1.6
Calibration intercept:	0.500000	0.100000	Coeff or Slope:	1.549	1.549

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	5000	1414.8	400.0	395.8	1.011
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.5	----
high point	5000	1415.7	400.0	397.0	1.008
second point	5000	1039.9	200.0	199.4	1.003
third point	5000	856.2	100.0	99.8	1.002
as left zero	5000	0.0	0.0	-0.7	----
as left span	5000	1416.0	400.0	398.8	1.003
Average Correction Factor					1.004

Baseline Corr As found:	396.1	Previous response	398.6	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 16, 2024	Previous Calibration:	January 12, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:24	End Time (MST):	14:22
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.5	----	Correlation Coefficient	0.999987	≥0.995
400.0	397.0	1.0076			
200.0	199.4	1.0030	Slope	0.993286	0.90 - 1.10
100.0	99.8	1.0020			
			Intercept	0.100000	+/- 5



O₃ Calibration Plot

Date: February 16, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Athabasca Valley Station number: AMS 07
 Calibration Date: February 27, 2024 Last Cal Date: January 18, 2024
 Start time (MST): 14:40 End time (MST): 16:49

Analyzer Make: S/N: 645
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-13	-14.2	-13	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	743.1	742.3	743.1	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.94	4.95	4.94	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	34	----	34	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: October 6, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	13.0	11.0	11.0	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 27, 2024
 Date Disposable Filter Changed: February 27, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022
 Date RH/T Sensor Cleaned: December 5, 2022

Notes: Temp, flow and pressure check. Quarterly maintenance carried out. PMT adjusted.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	February 23, 2024	Last Cal Date:	January 2, 2024
Start time (MST):	12:41	End time (MST):	15:55
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date:	December 12, 2026
Cal Gas Cylinder #:	LL66942			
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 700H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993944	0.996688	Backgd or Offset:	4.425	4.651
Calibration intercept:	0.080556	0.062533	Coeff or Slope:	1.087	1.087

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.0	40.2	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.0	39.9	1.003
second point	4967	33.3	20.0	20.1	0.995
third point	4983	16.7	10.0	10.1	0.995
as left zero	5000	0.0	0.0	0.0	----
as left span	4933	66.7	40.0	39.8	1.005
Average Correction Factor					0.998

Baseline Corr As found:	40.08	Prev response:	39.86	*% change:	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 23, 2024	Previous Calibration:	January 2, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	12:41	End Time (MST):	15:55
Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999982	≥0.995
40.0	39.9	1.0031			
20.0	20.1	0.9955	Slope	0.996688	0.90 - 1.10
10.0	10.1	0.9951			
			Intercept	0.062533	+/-1.5



CO Calibration Plot

Date: February 23, 2024

Location: Athabasca Valley





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	February 15, 2024	Last Cal Date:	January 31, 2024
Start time (MST):	13:18	End time (MST):	15:41
Reason:	Maintenance		

Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC196697			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
ZAG Make/Model:	Teledyne API T701		Serial Number:	135

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995420	0.994207	Backgd or Offset:	1.86	1.83
Calibration intercept:	1.396452	1.596267	Coeff or Slope:	0.989	0.989

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	4920	80.3	800.4	795.7	1.006
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4920	80.3	800.4	796.6	1.005
second point	4960	40.2	400.7	400.3	1.001
third point	4980	20.1	200.4	203.1	0.986
as left zero	5000	0.0	0.0	-0.3	----
as left span	4920	80.3	800.4	799.2	1.001

Average Correction Factor	0.997
---------------------------	-------

Baseline Corr As found:	796.10	Previous response	798.11	*% change	-0.3%
-------------------------	--------	-------------------	--------	-----------	-------

** = > +/-5% change initiates investigation*

Notes: Changed out inlet filter after as found. No adjustment made.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 31, 2024
Station Name:	December 13, 2023	Station Number:	Fort Chipewyan
Start Time (MST):	13:18	End Time (MST):	15:41
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999972	≥0.995
800.4	796.6	1.0047			
400.7	400.3	1.0010	Slope	0.994207	0.90 - 1.10
200.4	203.1	0.9865			
			Intercept	1.596267	+/-30



SO2 Calibration Plot

Date: February 15, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: February 13, 2024 Last Cal Date: January 26, 2024
 Start time (MST): 8:59 End time (MST): 12:46
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002276
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 1203169744
 Converter make: CDN-101 Converter serial #: 14639
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.988284	0.991000	Backgd or Offset:	0.99	0.98
Calibration intercept:	0.598606	0.578560	Coeff or Slope:	0.741	0.741

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4920	80.5	80.0	79.2	1.017
as found 2nd point	4960	40.2	40.0	40.1	1.009
as found 3rd point	4980	20.1	20.0	20.4	1.004
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.6	----
high point	4920	80.5	80.0	79.9	1.001
second point	4960	40.2	40.0	40.1	0.996
third point	4980	20.1	20.0	20.4	0.979
as left zero	5000	0.0	0.0	0.6	----
as left span	4920	80.5	80.0	79.8	1.003
SO2 Scrubber Check	4919.7	80.3	803.0	0.0	----
Date of last scrubber change:	March 7, 2022			Ave Corr Factor	0.992
Date of last converter efficiency test:	March 15, 2022			100.7% efficiency	

Baseline Corr As found: 78.7 Prev response: 79.67 *% change: -1.2%
 Baseline Corr 2nd AF pt: 39.6 AF Slope: 0.982998 AF Intercept: 0.658541
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999978

* = > +/-5% change initiates investigation

Notes: Sampled inlet filter changed after as founds. Scrubber check passed.

Calibration Performed By: Matthew Courtoreille



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 26, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	8:59	End Time (MST):	12:46
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999998	≥0.995
80.0	79.9	1.0014			
40.0	40.1	0.9964	Slope	0.991000	0.90 - 1.10
20.0	20.4	0.9794			
			Intercept	0.578560	+/-3



TRS Calibration Plot

Date: February 13, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
Calibration Date: February 20, 2024 Last Cal Date: NA
Start time (MST): 14:02 End time (MST): 16:34
Reason: Removal

Calibration Standards

NO Gas Cylinder #: CC363447 Cal Gas Expiry Date: February 2, 2024
NOX Cal Gas Conc: 48.80 ppm NO Cal Gas Conc: 48.80 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.80 ppm Removed Gas NO Conc: 48.80 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3252
ZAG make/model: Teledyne API T701H Serial Number: 135

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1426262592
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.138	NA	NO bkgnd or offset:	10.3	NA
NOX coeff or slope:	0.992	NA	NOX bkgnd or offset:	11.3	NA
NO2 coeff or slope:	1.000	NA	Reaction cell Press:	141.2	NA

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.009182	NA
NO _x Cal Offset:	1.020000	NA
NO Cal Slope:	1.007925	NA
NO Cal Offset:	1.260000	NA
NO ₂ Cal Slope:	1.010781	NA
NO ₂ Cal Offset:	-3.206174	NA



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.3	1.3	-1.0	----	----
as found span	4918	82.0	800.3	800.3	0.0	741.1	737.4	3.7	1.0799	1.0853
as found 2nd	4959	41.0	400.2	400.2	0.0	380.5	378.3	2.2	1.0517	1.0578
as found 3rd	4980	20.5	200.1	200.1	0.0	193.5	192.3	1.3	1.0340	1.0405
new cyl resp										
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

Average Correction Factor

Corrected As found	NO _x = 740.8 ppb	NO = 736.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -9.2%
Previous Response	NO _x = 808.7 ppb	NO = 807.9 ppb		*Percent Change	NO = -9.8%
Baseline Corr 2nd pt	NO _x = 380.2 ppb	NO = 377.0 ppb	As found	NO _x r ² : 0.999734	Nx SI: 0.923602
Baseline Corr 3rd pt	NO _x = 193.2 ppb	NO = 191.0 ppb	As found	NO r ² : 0.999782	NO SI: 0.918047
			As found	NO ₂ r ² : 1.000000	NO ₂ SI: 0.993120
					NO ₂ Int: -1.000

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero	----	----	0.0	-1.0	----	----
as found GPT point (400 ppb NO2)	749.2	371.3	377.9	374.3	1.0096	99.0%
as found GPT point (200 ppb NO2)	749.2					
as found GPT point (100 ppb NO2)	749.2					
1st GPT point (400 ppb O3)						
2nd GPT point (200 ppb O3)						
3rd GPT point (100 ppb O3)						

Average Correction Factor

Notes:


Removal multipoint as founds done.

Calibration Performed By: Jan Castro and Braiden Boutilier

NO_x Calibration Plot

Date: February 20, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
Calibration Date: February 21, 2024 Last Cal Date: NA
Start time (MST): 7:45 End time (MST): 12:10
Reason: Install

Calibration Standards

NO Gas Cylinder #: DT0046831 Cal Gas Expiry Date: January 9, 2032
NOX Cal Gas Conc: 60.20 ppm NO Cal Gas Conc: 60.00 ppm
Removed Cylinder #: CC363447 Removed Gas Exp Date: February 2, 2024
Removed Gas NOX Conc: 48.80 ppm Removed Gas NO Conc: 48.80 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3252
ZAG make/model: Teledyne API T701H Serial Number: 135

Analyzer Information

Analyzer make: API T200 Analyzer serial #: 4460
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	NA	1.154	NO bkgnd or offset:	NA	-2.5
NOX coeff or slope:	NA	1.152	NOX bkgnd or offset:	NA	-2.2
NO2 coeff or slope:	NA	1.000	Reaction cell Press:	NA	2.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	NA	0.997032
NO _x Cal Offset:	NA	2.294675
NO Cal Slope:	NA	0.998455
NO Cal Offset:	NA	1.134987
NO ₂ Cal Slope:	NA	1.001043
NO ₂ Cal Offset:	NA	0.109543



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.4	0.3	0.1	----	----
high point	4933	66.7	803.1	800.4	2.7	802.0	799.7	2.4	1.0014	1.0009
second point	4967	33.3	400.9	399.6	1.3	403.2	401.2	2.0	0.9943	0.9960
third point	4983	16.7	201.1	200.4	0.7	204.5	201.6	2.9	0.9832	0.9940
as left zero										
as left span										

								Average Correction Factor	
								0.9930	0.9970
Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x =	NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO =	NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:	
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.0	393.0	407.7	407.5	1.0004	100.0%
2nd GPT point (200 ppb O3)	798.0	588.8	211.9	214.2	0.9891	101.1%
3rd GPT point (100 ppb O3)	798.0	694.0	106.7	105.6	1.0101	99.0%
					Average Correction Factor	100.0%

Notes: Install calibrations. Sample inlet filters changed before calibrator zero. Adjusted zero and span.

Calibration Performed By: Jan Castro and Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	NA
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	7:45	End Time (MST):	12:10
Analyzer make:	API T200	Analyzer serial #:	13806

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.4	----	Correlation Coefficient	≥0.995	
803.1	802.0	1.0014			
400.9	403.2	0.9943			
201.1	204.5	0.9832			
			Slope	0.997032	0.90 - 1.10
			Intercept	2.294675	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	NA
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	7:45	End Time (MST):	12:10
Analyzer make:	API T200	Analyzer serial #:	13806

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient	≥0.995
800.4	799.7	1.0009		
399.6	401.2	0.9960	Slope	0.90 - 1.10
200.4	201.6	0.9940		
			Intercept	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	NA
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	7:45	End Time (MST):	12:10
Analyzer make:	API T200	Analyzer serial #:	13806

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
407.7	407.5	1.0004		
211.9	214.2	0.9891		
106.7	105.6	1.0101		
			0.999932	
			1.001043	
			0.109543	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 21, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: February 6, 2024 Last Cal Date: January 8, 2024
 Start time (MST): 15:27 End time (MST): 17:38
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3872
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.976343	1.027914	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	0.340000	0.240000	Coeff or Slope:	1.036	1.036

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	5000	913.0	400.0	408.1	0.980
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.6	----
high point	5000	914.7	400.0	411.6	0.972
second point	5000	786.4	200.0	205.6	0.973
third point	5000	701.3	100.0	102.7	0.974
as left zero	5000	0.0	0.0	0.8	----
as left span	5000	963.3	400.0	408.1	0.980
Average Correction Factor					0.973

Baseline Corr As found:	407.7	Previous response	390.9	*% change	4.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed out inlet filter after as found. No adjustments

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 8, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	15:27	End Time (MST):	17:38
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999996	≥0.995
400.0	411.6	0.9718			
200.0	205.6	0.9728	Slope	1.027914	0.90 - 1.10
100.0	102.7	0.9737			
			Intercept	0.240000	+/- 5



O₃ Calibration Plot

Date: February 6, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
 Calibration Date: February 20, 2024 Last Cal Date: January 31, 2024
 Start time (MST): 15:25 End time (MST): 16:10

Analyzer Make: Teledyne API T640 S/N: 319
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-11.80	-12.05	-11.80	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	735.20	736.58	735.20	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.02	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	42.00	----	42.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.40	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: 10-Jun-24
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.10	10.90	10.90	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 20, 2024
 Date Disposable Filter Changed: February 20, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 25, 2023
 Date RH/T Sensor Cleaned: July 25, 2023

Notes: Leak check passed. Flow, temperature, and pressure checked. No adjustments needed.

Calibration by: Jan Castro and Braiden Boutilier



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	February 6, 2024	Last Cal Date:	January 18, 2024
Start time (MST):	13:07	End time (MST):	15:21
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3252
ZAG Make/Model:	API T701H		Serial Number:	135

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.989053	0.994000	Backgd or Offset:	-0.014	-0.014
Calibration intercept:	0.314884	0.374934	Coeff or Slope:	0.998	0.998

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.20	----
as found span	4933	66.7	40.4	40.4	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4934	66.7	40.4	40.4	0.999
second point	4967	33.3	20.2	20.6	0.980
third point	4983	16.7	10.1	10.5	0.961
as left zero	5000	0.0	0.0	0.3	----
as left span	2960	40.0	40.4	40.3	1.004
Average Correction Factor					0.980

Baseline Corr As found:	40.24	Prev response:	40.30	*% change:	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. no adjustments made.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 18, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	13:07	End Time (MST):	15:21
Analyzer make:	API T300	Analyzer serial #:	3505

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999918	≥0.995
40.4	40.4	0.9994			
20.2	20.6	0.9796	Slope	0.994000	0.90 - 1.10
10.1	10.5	0.9611			
			Intercept	0.374934	+/-1.5



CO Calibration Plot

Date: February 6, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	February 6, 2024	Last Cal Date:	January 18, 2024
Start time (MST):	9:02	End time (MST):	12:57
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
N2 Gen Make/Model:	NG 5000		Serial Number:	135

Analyzer Information

Analyzer make: Teledyne API T360	Analyzer serial #: 289
Analyzer Range 0 - 2,000 ppm	

	Start	Finish		Start	Finish
Calibration slope:	0.994632	1.000467	Backgd or Offset:	-0.063	-0.063
Calibration intercept:	-6.520000	-10.320000	Coeff or Slope:	1.085	1.094

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-3.3	----
as found span	2920	80.0	1605.9	1536.9	1.045
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	-1.7	----
high point	2920	80.0	1605.9	1606.6	1.000
second point	2960	40.0	802.9	770.7	1.042
third point	2980	20.0	401.5	394.7	1.017
as left zero	3000	0.0	0.0	-0.1	----
as left span	2960	40.0	802.9	768.5	1.045
Average Correction Factor					1.020

Baseline Corr As found:	1540.20	Prev response:	1590.73	*% change:	-3.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as found. Made adjustments to span. Operator error well adjusting span fix problem restarted back to zero cal.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	February 6, 2024	Previous Calibration	January 18, 2024
Station Name	Fort Chipewyan	Station Number	AMS08
Start Time (MST)	9:02	End Time (MST)	12:57
Analyzer make	Teledyne API T360	Analyzer serial #	289

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-1.7	----	Correlation Coefficient	0.999512	≥0.995
1605.9	1606.6	0.9995			
802.9	770.7	1.0418	Slope	1.000467	0.90 - 1.10
401.5	394.7	1.0171			
			Intercept	-10.320000	+/-20



CO₂ Calibration Plot

Date: February 6, 2024

Location: Fort Chipewyan





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 16, 2024	Previous Calibration:	January 4, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:23	End Time (MST):	13:33
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999998	≥0.995
801.5	802.0	0.9994			
400.8	402.0	0.9969	Slope	0.999855	0.90 - 1.10
199.8	201.2	0.9932			
			Intercept	0.931270	+/-30



SO2 Calibration Plot

Date: February 16, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Barge Landing Station number: AMS09
 Calibration Date: February 1, 2024 Last Cal Date: January 9, 2024
 Start time (MST): 9:46 End time (MST): 14:02
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.171 ppm Cal Gas Exp Date: August 22, 2026
 Cal Gas Cylinder #: CC511415
 Removed Cal Gas Conc: 5.171 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3812
 ZAG Make/Model: API T701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1331259320
 Converter make: CDN-101 Converter serial #: 519
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008544	0.999264	Backgd or Offset: 2.88	2.93
Calibration intercept:	0.119536	0.119396	Coeff or Slope: 1.176	1.201

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4923	77.4	80.0	77.8	1.029
as found 2nd point	4961	38.7	40.0	38.9	1.029
as found 3rd point	4981	19.3	20.0	19.4	1.029
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.4	80.0	80.0	1.001
second point	4961	38.7	40.0	40.3	0.993
third point	4981	19.3	20.0	20.1	0.993
as left zero	5000	0.0	0.0	0.1	----
as left span	4923	77.4	80.0	79.7	1.004
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:	28-Feb-23			Ave Corr Factor	0.996
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 77.8 Prev response: 80.84 *% change: -3.9%
 Baseline Corr 2nd AF pt: 38.9 AF Slope: 0.971999 AF Intercept: -0.001140
 Baseline Corr 3rd AF pt: 19.4 AF Correlation: 1.000000

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SOx scrubber check done after calibrator zero.
 Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 1, 2024	Previous Calibration:	January 9, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	9:46	End Time (MST):	14:02
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999983	
80.0	80.0	1.0005			≥0.995
40.0	40.3	0.9932	Slope	0.999264	
20.0	20.1	0.9930			0.90 - 1.10
			Intercept	0.119396	+/-3



TRS Calibration Plot

Date: February 1, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	February 16, 2024	Last Cal Date:	January 4, 2024
Start time (MST):	10:23	End time (MST):	13:33
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.58E-04	2.58E-04	NMHC SP Ratio:	4.37E-05
CH ₄ Retention time:	15.00	15.00	NMHC Peak Area:	209298
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	17.18	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.11	1.000
second point	4960	40.1	8.56	8.57	0.999
third point	4980	20.0	4.27	4.29	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.17	0.997

				Response Correction Factor	0.998
Baseline Corr AF:	17.18	Prev response	17.14	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	9.14	9.14	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.11	1.003
second point	4960	40.1	4.57	4.56	1.002
third point	4980	20.0	2.28	2.28	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	9.14	9.13	1.001
Average Correction Factor					1.002
Baseline Corr AF:	9.14	Prev response	9.13	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	8.03	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	8.00	0.998
second point	4960	40.1	3.99	4.01	0.995
third point	4980	20.0	1.99	2.01	0.990
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	8.04	0.993
Average Correction Factor					0.995
Baseline Corr AF:	8.03	Prev response	8.00	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000804	0.999215
THC Cal Offset:	0.005458	0.010853
CH ₄ Cal Slope:	1.001867	1.001638
CH ₄ Cal Offset:	0.006270	0.008070
NMHC Cal Slope:	0.999587	0.997023
NMHC Cal Offset:	-0.000411	0.003583

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 16, 2024	Previous Calibration:	January 4, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:23	End Time (MST):	13:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
17.12	17.11	1.0004						
8.56	8.57	0.9991				Slope	0.999215	0.90 - 1.10
4.27	4.29	0.9955						
			Intercept	0.010853	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 16, 2024	Previous Calibration:	January 4, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:23	End Time (MST):	13:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	≥0.995			
7.98	8.00	0.9978						
3.99	4.01	0.9954				Slope	1.001638	0.90 - 1.10
1.99	2.01	0.9902						
			Intercept	0.008070	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 16, 2024	Previous Calibration:	January 4, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:23	End Time (MST):	13:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
9.14	9.11	1.0028						
4.57	4.56	1.0016				Slope	0.997023	0.90 - 1.10
2.28	2.28	1.0000						
			Intercept	0.003583	± 0.5			



NMHC Calibration Plot

Date: February 16, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Barge Landing
Calibration Date: February 13, 2024
Start time (MST): 10:07
Reason: Routine
Station number: AMS09
Last Cal Date: January 18, 2024
End time (MST): 14:26

Calibration Standards

NO Gas Cylinder #: T2Y1KDH
NOX Cal Gas Conc: 47.38 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.38 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 46.94 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 46.94 ppm
NO gas Diff:
Serial Number: 3812
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1426262593

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.092	1.120	NO bkgnd or offset:	10.0	10.3
NOX coeff or slope:	0.998	0.998	NOX bkgnd or offset:	10.3	10.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	180.1	176.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999176	0.999459
NO _x Cal Offset:	0.418407	0.518402
NO Cal Slope:	0.999039	0.998611
NO Cal Offset:	-0.303738	-0.704018
NO ₂ Cal Slope:	1.005229	1.004419
NO ₂ Cal Offset:	0.853830	0.978163



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.2	-0.2	----	----
as found span	4915	85.3	808.3	800.7	7.5	791.0	780.7	10.3	1.022	1.026
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
high point	4915	85.3	808.3	800.7	7.5	808.0	799.3	8.8	1.000	1.002
second point	4957	42.6	403.7	400.0	3.7	404.4	398.3	6.1	0.998	1.004
third point	4979	21.3	201.8	200.0	1.9	202.8	198.3	4.5	0.995	1.008
as left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.1	----	----
as left span	4915	85.3	808.3	414.9	393.3	804.6	410.7	393.9	1.005	1.010
Average Correction Factor									0.998	1.005

Corrected As found	NO _x = 791.4 ppb	NO = 780.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -2.1%	
Previous Response	NO _x = 808.0 ppb	NO = 799.7 ppb		*Percent Change	NO = -2.4%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	793.8	408.0	393.3	395.6	0.994	100.6%
2nd GPT point (200 ppb O3)	793.8	601.9	199.4	201.4	0.990	101.0%
3rd GPT point (100 ppb O3)	793.8	699.4	101.9	104.7	0.973	102.7%
Average Correction Factor					0.986	101.4%

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 18, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:07	End Time (MST):	14:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
808.3	808.0	1.0003		
403.7	404.4	0.9983		
201.8	202.8	0.9952		
			0.999997	
			0.999459	
			0.518402	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 18, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:07	End Time (MST):	14:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
800.7	799.3	1.0018			
400.0	398.3	1.0042			
200.0	198.3	1.0083			
			Slope	0.998611	0.90 - 1.10
			Intercept	-0.704018	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 18, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:07	End Time (MST):	14:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
393.3	395.6	0.9942			
199.4	201.4	0.9901			
101.9	104.7	0.9733			
			Slope	1.004419	0.90 - 1.10
			Intercept	0.978163	+/-20



NO_x Calibration Plot

Date: February 13, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Barge Landing Station number: AMS 09
 Calibration Date: February 16, 2024 Last Cal Date: January 22, 2024
 Start time (MST): 12:04 End time (MST): 12:39

Analyzer Make: API T640 S/N: 844
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388746
 Temp/RH standard: Alicat FP-25BT S/N: 388746

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-0.10	-0.11	-0.10	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	731.10	738.88	731.10	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	4.92	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37.00	----	37.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	5.60	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: January 22, 2024
 Date Disposable Filter Changed: January 22, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 23, 2023
 Date RH/T Sensor Cleaned: August 23, 2023

Notes: Inlet head looks good. No adjustments. Leak check passed.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS11
LOWER CAMP**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 3, 2024
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	7:35	End Time (MST):	11:09
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.0	----	Correlation Coefficient	0.999989	
799.6	798.3	1.0017			≥0.995
400.9	399.3	1.0041	Slope	0.997880	
200.9	198.9	1.0100			0.90 - 1.10
			Intercept	-0.244782	+/-30



SO2 Calibration Plot

Date: February 15, 2024

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	February 28, 2024	Last Cal Date:	January 22, 2024
Start time (MST):	11:15	End time (MST):	14:30
Reason:	As Found		

Calibration Standards

Cal Gas Concentration:	5.429	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC501097			
Removed Cal Gas Conc:	5.429	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3807
ZAG Make/Model:	API T701H		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003
Converter make:	NA	Converter serial #:	NA
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992455		Backgd or Offset:	15.9
Calibration intercept:	-0.204215		Coeff or Slope:	1.015
				NA
				NA

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.5	----
as found span	4926	73.6	79.9	79.5	0.999
as found 2nd point	4963	36.8	40.0	39.9	0.989
as found 3rd point	4982	18.6	20.2	19.7	1.000
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

SO₂ Scrubber Check

Date of last scrubber change:	Ave Corr Factor
Date of last converter efficiency test:	efficiency

Baseline Corr As found:	80.0	Prev response:	79.11	*% change:	1.1%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.001611	AF Intercept:	-0.424852
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999965		

** = > +/-5% change initiates investigation*


Notes: Changed sample inlet filter after as founds. Performed a scrubber check after calibrator zero, and it failed. Halted the calibration and opted to replace the scrubber beads. We'll resume and complete the calibration tomorrow.

Calibration Performed By: Mohammed Kashif

H₂S Calibration Plot

Date: February 28, 2024

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	February 29, 2024	Last Cal Date:	January 22, 2024
Start time (MST):	10:15	End time (MST):	13:01
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.429	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC501097			
Removed Cal Gas Conc:	5.429	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3807
ZAG Make/Model:	API T701H		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003
Converter make:	NA	Converter serial #:	NA
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992455	1.011205	Backgd or Offset:	15.9
Calibration intercept:	-0.204215	-0.085817	Coeff or Slope:	1.015
				15.8
				1.015

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4926	73.6	79.9	80.8	0.989
second point	4963	36.8	40.0	40.2	0.994
third point	4982	18.6	20.2	20.3	0.995
as left zero	5000	0.0	0.0	0.2	----
as left span	4926	73.6	79.9	80.3	0.995
SO2 Scrubber Check	4935	81.5	812.3	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.993
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: Yesterday, the scrubber beads and sample filter were replaced. Performed a scrubber check after calibrator zero, and it passed. Adjusted zero only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 29, 2024	Previous Calibration:	January 22, 2024
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:15	End Time (MST):	13:01
Analyzer make:	Thermo 450iQ	Analyzer serial #:	CM20080003

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999992	≥0.995
79.9	80.8	0.9891			
40.0	40.2	0.9940	Slope	1.011205	0.90 - 1.10
20.2	20.3	0.9948			
			Intercept	-0.085817	+/-3



H₂S Calibration Plot

Date: February 29, 2024

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	February 15, 2024	Last Cal Date:	January 3, 2024
Start time (MST):	7:35	End time (MST):	11:09
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
ZAG make/model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.00E-04	2.88E-04	NMHC SP Ratio:	5.77E-05
CH ₄ Retention time:	14.2	14.6	NMHC Peak Area:	158880
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				5.30E-05
				173476
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4932	81.4	17.33	18.43	0.940
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4932	81.4	17.33	17.35	0.999
second point	4959	40.7	8.69	8.63	1.006
third point	4981	20.4	4.35	4.31	1.010
as left zero	5000	0.0	0.00	0.00	----
as left span	4932	81.4	17.33	17.49	0.990
Average Correction Factor					1.005

Baseline Corr AF:	18.43	Prev response	17.29	*% change	6.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4932	81.4	9.18	9.97	0.921
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4932	81.4	9.18	9.19	0.998
second point	4959	40.7	4.60	4.57	1.006
third point	4981	20.4	2.31	2.28	1.011
as left zero	5000	0.0	0.00	0.00	----
as left span	4932	81.4	9.18	9.23	0.995
Average Correction Factor					1.005
Baseline Corr AF:	9.97	Prev response	9.20	*% change	7.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4932	81.4	8.15	8.47	0.963
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4932	81.4	8.15	8.16	0.999
second point	4959	40.7	4.09	4.06	1.006
third point	4981	20.4	2.05	2.03	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4932	81.4	8.15	8.27	0.986
Average Correction Factor					1.005
Baseline Corr AF:	8.47	Prev response	8.09	*% change	4.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.998445	1.001990
THC Cal Offset:	-0.009595	-0.033339
CH ₄ Cal Slope:	0.993827	1.001462
CH ₄ Cal Offset:	-0.006092	-0.014179
NMHC Cal Slope:	1.002671	1.002471
NMHC Cal Offset:	-0.003503	-0.018960

Notes: Changed sample inlet filter and Nitrogen cylinder after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 3, 2024
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	7:35	End Time (MST):	11:09
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999979	≥ 0.995			
17.33	17.35	0.9986						
8.69	8.63	1.0061				Slope	1.001990	0.90 - 1.10
4.35	4.31	1.0104						
			Intercept	-0.033339	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 3, 2024
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	7:35	End Time (MST):	11:09
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999982	≥0.995			
8.15	8.16	0.9990						
4.09	4.06	1.0063				Slope	1.001462	0.90 - 1.10
2.05	2.03	1.0092						
			Intercept	-0.014179	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 3, 2024
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	7:35	End Time (MST):	11:09
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999978	≥ 0.995			
9.18	9.19	0.9982						
4.60	4.57	1.0058				Slope	1.002471	0.90 - 1.10
2.31	2.28	1.0114						
			Intercept	-0.018960	± 0.5			



NMHC Calibration Plot

Date: February 15, 2024

Location: Lower Camp





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13
FORT MCKAY SOUTH

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 12, 2024	Previous Calibration:	January 15, 2024
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:25	End Time (MST):	13:58
Analyzer make:	API T100	Analyzer serial #:	599

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	0.999954	≥0.995
799.7	801.0	0.9984			
399.3	397.2	1.0053	Slope	1.003642	0.90 - 1.10
200.2	195.4	1.0245			
			Intercept	-2.718121	+/-30



SO2 Calibration Plot

Date: February 12, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: February 6, 2024 Last Cal Date: January 17, 2024
 Start time (MST): 9:48 End time (MST): 14:19
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.34 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC500241
 Removed Cal Gas Conc: 5.34 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API 701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1180540017
 Converter make: CDN-101 Converter serial #: 521
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997527	1.001638	Backgd or Offset: 4.15	4.15
Calibration intercept:	-0.182182	-0.302148	Coeff or Slope: 1.157	1.157

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	4925	75.5	80.6	80.5	0.997
as found 2nd point	4962	37.7	40.3	39.9	0.999
as found 3rd point	4981	18.9	20.2	19.4	1.019
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4925	75.5	80.6	80.4	1.003
second point	4962	37.7	40.3	40.2	1.002
third point	4981	18.9	20.2	19.8	1.019
as left zero	5000	0.0	0.0	0.0	----
as left span	4925	75.5	80.6	79.9	1.009
SO2 Scrubber Check	4921	79.1	791.0	0.1	----
Date of last scrubber change:	20-Mar-20		Ave Corr Factor		1.008
Date of last converter efficiency test:	NA		efficiency		

Baseline Corr As found:	80.9	Prev response:	80.24	*% change:	0.8%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.005182	AF Intercept:	-0.602173
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999964		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 17, 2024
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	9:48	End Time (MST):	14:19
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999987	
80.6	80.4	1.0028			≥0.995
40.3	40.2	1.0016	Slope	1.001638	
20.2	19.8	1.0195			0.90 - 1.10
			Intercept	-0.302148	+/-3



TRS Calibration Plot

Date: February 6, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	February 12, 2024	Last Cal Date:	January 15, 2024
Start time (MST):	10:25	End time (MST):	13:58
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH4 Cal Gas Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
C3H8 Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
Removed C3H8 Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API701	Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.94E-04	2.94E-04	NMHC SP Ratio:	4.74E-05	4.83E-05
CH4 Retention time:	15.00	15.20	NMHC Peak Area:	191512	1879773
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	ON

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	17.05	16.90	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.03	1.001
second point	4961	39.5	8.51	8.40	1.013
third point	4980	19.8	4.27	4.15	1.028
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	17.05	16.86	1.011

Average Correction Factor	1.014
---------------------------	-------

Baseline Corr AF:	16.90	Prev response	17.02	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	9.08	8.91	1.019
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	9.08	9.08	1.000
second point	4961	39.5	4.53	4.51	1.006
third point	4980	19.8	2.27	2.23	1.021
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	9.08	8.95	1.015
Average Correction Factor					1.009
Baseline Corr AF:	8.91	Prev response	9.06	*% change	-1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	7.97	8.00	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.95	1.002
second point	4961	39.5	3.98	3.90	1.020
third point	4980	19.8	1.99	1.93	1.036
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	7.97	7.91	1.007
Average Correction Factor					1.019
Baseline Corr AF:	8.00	Prev response	7.97	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002747	1.000652
THC Cal Offset:	-0.071954	-0.065578
CH ₄ Cal Slope:	1.003122	1.000073
CH ₄ Cal Offset:	-0.024172	-0.040797
NMHC Cal Slope:	1.002658	1.001197
NMHC Cal Offset:	-0.047981	-0.024180

Notes: Changed inlet filter after as founds. Do zero chromatogram and use flatbaseline. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 12, 2024	Previous Calibration:	January 15, 2024
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:25	End Time (MST):	13:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999930	≥ 0.995			
17.05	17.03	1.0009						
8.51	8.40	1.0130				Slope	1.000652	0.90 - 1.10
4.27	4.15	1.0282						
			Intercept	-0.065578	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 12, 2024	Previous Calibration:	January 15, 2024
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:25	End Time (MST):	13:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999865	≥ 0.995
7.97	7.95	1.0017			
3.98	3.90	1.0203			
1.99	1.93	1.0360			
			Slope	1.000073	0.90 - 1.10
			Intercept	-0.040797	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 12, 2024	Previous Calibration:	January 15, 2024
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:25	End Time (MST):	13:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999967	≥ 0.995			
9.08	9.08	1.0003						
4.53	4.51	1.0061				Slope	1.001197	0.90 - 1.10
2.27	2.23	1.0215						
			Intercept	-0.024180	± 0.5			



NMHC Calibration Plot

Date: February 12, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort McKay South Station number: AMS 13
Calibration Date: February 20, 2024 Last Cal Date: January 25, 2024
Start time (MST): 10:01 End time (MST): 14:39
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2UP1RP Cal Gas Expiry Date: November 17, 2026
NOX Cal Gas Conc: 48.25 ppm NO Cal Gas Conc: 47.88 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.25 ppm Removed Gas NO Conc: 47.88 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: API T700 Serial Number: 2448
ZAG make/model: API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1410661329
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.133	1.128	NO bkgnd or offset:	10.9	10.1
NOX coeff or slope:	1.001	0.999	NOX bkgnd or offset:	11.0	10.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.2	156.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001306	0.999462
NO _x Cal Offset:	-2.253060	-1.952788
NO Cal Slope:	1.003396	1.002954
NO Cal Offset:	-3.151422	-2.771571
NO ₂ Cal Slope:	0.998110	1.000409
NO ₂ Cal Offset:	-0.700632	-0.685310



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.6	0.6	0.0	----	----
as found span	4917	83.5	805.7	799.5	6.2	810.9	803.3	7.6	0.9936	0.9953
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
high point	4917	83.5	805.7	799.5	6.2	804.3	800.4	3.8	1.0017	0.9989
second point	4958	41.8	403.4	400.3	3.1	400.3	397.7	2.6	1.0077	1.0065
third point	4979	20.9	201.7	200.1	1.5	197.5	194.8	2.7	1.0212	1.0274
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.0	----	----
as left span	4917	83.5	805.7	381.9	423.8	806.7	379.0	427.6	0.9988	1.0077
Average Correction Factor									1.0102	1.0109

Corrected As found	NO _x = 810.3 ppb	NO = 802.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.7%
Previous Response	NO _x = 804.5 ppb	NO = 799.1 ppb		*Percent Change	NO = 0.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	794.5	376.9	423.8	423.6	1.0004	100.0%
2nd GPT point (200 ppb O ₃)	794.5	586.1	214.6	213.7	1.0041	99.6%
3rd GPT point (100 ppb O ₃)	794.5	690.8	109.9	108.5	1.0127	98.7%
Average Correction Factor					1.0057	99.4%

Notes: Changed inlet filter after as founds. Adjusted zero and span. Used 2nd NO reference point due to drift.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 20, 2024	Previous Calibration:	January 25, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:01	End Time (MST):	14:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
805.7	804.3	1.0017			
403.4	400.3	1.0077			
201.7	197.5	1.0212			
			Slope	0.999462	0.90 - 1.10
			Intercept	-1.952788	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 20, 2024	Previous Calibration:	January 25, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:01	End Time (MST):	14:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999940	≥0.995
799.5	800.4	0.9989			
400.3	397.7	1.0065	Slope	1.002954	0.90 - 1.10
200.1	194.8	1.0274			
			Intercept	-2.771571	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 20, 2024	Previous Calibration:	January 25, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:01	End Time (MST):	14:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
423.8	423.6	1.0004		
214.6	213.7	1.0041		
109.9	108.5	1.0127		
			0.999988	
			1.000409	
			-0.685310	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 20, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: February 2, 2024 Last Cal Date: January 5, 2024
 Start time (MST): 10:12 End time (MST): 13:30
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3871
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000914	1.000914	Backgd or Offset:	3.9	3.7
Calibration intercept:	0.240000	0.740000	Coeff or Slope:	0.971	0.970

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.8	----
as found span	5000	989.8	400.0	400.7	0.998
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.0	----
high point	5000	989.8	400.0	400.7	0.998
second point	5000	849.9	200.0	201.4	0.993
third point	5000	745.1	100.0	101.5	0.985
as left zero	5000	0.0	0.0	0.4	----
as left span	5000	989.8	400.0	402.6	0.994
Average Correction Factor					0.992

Baseline Corr As found:	401.5	Previous response	400.6	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Adjusted zero.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 5, 2024
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:12	End Time (MST):	13:30
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999984	≥0.995
400.0	400.7	0.9983			
200.0	201.4	0.9930	Slope	1.000914	0.90 - 1.10
100.0	101.5	0.9852			
			Intercept	0.740000	+/- 5



O₃ Calibration Plot

Date: February 2, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay South Station number: AMS 13
 Calibration Date: February 20, 2024 Last Cal Date: January 22, 2024
 Start time (MST): 11:48 End time (MST): 12:05

Analyzer Make: API T640 S/N: 1335
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388746
 Temp/RH standard: Alicat FP-25BT S/N: 388746

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-7.80	-8.16	-7.80	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	731.20	733.35	731.20	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.03	4.91	5.03	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	46.00	----	46.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	16.60	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: January 22, 2024
 Date Disposable Filter Changed: January 22, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: June 29, 2023
 Date RH/T Sensor Cleaned: June 29, 2023

Notes: Inlet head cleaned. No adjustment made. Leak check passed.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	February 7, 2024	Last Cal Date:	January 4, 2024
Start time (MST):	10:47	End time (MST):	13:49
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.95	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC279389			
Removed Cal Gas Conc:	49.95	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API T701H		Serial Number:	357

Analyzer Information

Analyzer make: Thermo 43i
 Analyzer Range 0 - 1000 ppb
 Analyzer serial #: 0710321322

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001175	0.994566	Backgd or Offset:	25.6	25.6
Calibration intercept:	-1.861173	-0.897563	Coeff or Slope:	0.812	0.812

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	1.2	----
as found span	4938	80.3	799.3	794.9	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	1.7	----
high point	4938	80.3	799.3	795.4	1.005
second point	4979	40.2	400.1	395.5	1.012
third point	4998	20.2	201.1	196.6	1.023
as left zero	5000	0.0	0.0	1.6	----
as left span	4938	80.3	799.3	797.2	1.003
Average Correction Factor					1.013

Baseline Corr As found:	793.70	Previous response	798.35	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments needed.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 7, 2024	Previous Calibration:	January 4, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:47	End Time (MST):	13:49
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	1.7	----	Correlation Coefficient	0.999951	≥0.995
799.3	795.4	1.0049			
400.1	395.5	1.0115	Slope	0.994566	0.90 - 1.10
201.1	196.6	1.0227			
			Intercept	-0.897563	+/-30



SO2 Calibration Plot

Date: February 7, 2024

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Anzac Station number: AMS14
 Calibration Date: February 12, 2024 Last Cal Date: January 18, 2024
 Start time (MST): 10:33 End time (MST): 14:59
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC510379
 Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3060
 ZAG Make/Model: API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153582
 Converter make: CD Nova CDN-101 Converter serial #: 503
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.012453	1.003735	Backgd or Offset: 2.40	2.30
Calibration intercept:	-0.065392	-0.085416	Coeff or Slope: 1.043	0.984

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.2	----
as found span	4938	77.9	80.0	84.7	0.942
as found 2nd point	4973	38.9	40.0	42.7	0.932
as found 3rd point	4997	19.5	20.0	21.0	0.944
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4938	77.9	80.0	80.2	0.997
second point	4973	38.9	40.0	40.0	0.999
third point	4997	19.5	20.0	19.9	1.006
as left zero	5000	0.0	0.0	0.0	----
as left span	4938	77.9	80.0	78.6	1.017
SO2 Scrubber Check	4936	80.3	800.4	-0.1	----

Date of last scrubber change:	Ave Corr Factor	1.000
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 84.9 Prev response: 80.91 *% change: 4.7%
 Baseline Corr 2nd AF pt: 42.9 AF Slope: 1.062363 AF Intercept: -0.125652
 Baseline Corr 3rd AF pt: 21.2 AF Correlation: 0.999955

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 12, 2024	Previous Calibration:	January 18, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:33	End Time (MST):	14:59
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999995	
80.0	80.2	0.9969			≥0.995
40.0	40.0	0.9989	Slope	1.003735	
20.0	19.9	1.0056			0.90 - 1.10
			Intercept	-0.085416	+/-3



TRS Calibration Plot

Date: February 12, 2024

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	February 3, 2024	Last Cal Date:	January 24, 2024
Start time (MST):	11:35	End time (MST):	16:32
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.72E-04	2.25E-04	NMHC SP Ratio:	4.80E-05
CH ₄ Retention time:	13.40	13.30	NMHC Peak Area:	188364
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.12	0.999
second point	4979	40.2	8.56	8.48	1.009
third point	4998	20.2	4.30	4.23	1.018
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	17.10	17.18	0.996
Average Correction Factor					1.009

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.13	0.998
second point	4979	40.2	4.56	4.54	1.005
third point	4998	20.2	2.29	2.28	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	9.11	9.16	0.995
Average Correction Factor					1.002
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	7.99	1.000
second point	4979	40.2	4.00	3.94	1.014
third point	4998	20.2	2.01	1.94	1.035
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	7.99	8.01	0.997
Average Correction Factor					1.016
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002215	1.002044
THC Cal Offset:	-0.033495	-0.050467
CH ₄ Cal Slope:	0.999209	1.002402
CH ₄ Cal Offset:	-0.009829	-0.039003
NMHC Cal Slope:	1.004825	1.001944
NMHC Cal Offset:	-0.024066	-0.011066

Notes: Calibrated due to significant chromatogram changes. Updated window timings. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 3, 2024	Previous Calibration:	January 24, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:35	End Time (MST):	16:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999955	≥0.995
17.10	17.12	0.9990			
8.56	8.48	1.0095	Slope	1.002044	0.90 - 1.10
4.30	4.23	1.0183			
			Intercept	-0.050467	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 3, 2024	Previous Calibration:	January 24, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:35	End Time (MST):	16:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999888	≥0.995
7.99	7.99	0.9997			
4.00	3.94	1.0145	Slope	1.002402	0.90 - 1.10
2.01	1.94	1.0349			
			Intercept	-0.039003	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 3, 2024	Previous Calibration:	January 24, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:35	End Time (MST):	16:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999987	≥ 0.995			
9.11	9.13	0.9982						
4.56	4.54	1.0047				Slope	1.001944	0.90 - 1.10
2.29	2.28	1.0037						
			Intercept	-0.011066	± 0.5			



NMHC Calibration Plot

Date: February 3, 2024

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	February 14, 2024	Last Cal Date:	February 3, 2024
Start time (MST):	10:31	End time (MST):	12:50
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.25E-04	2.25E-04	NMHC SP Ratio:	4.11E-05
CH ₄ Retention time:	13.30	13.30	NMHC Peak Area:	221451
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	17.21	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.21	0.994
second point					
third point					
as left zero					
as left span					
Average Correction Factor					0.994

Baseline Corr AF:	17.21	Prev response	17.09	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	9.20	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.19	0.991
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.991
Baseline Corr AF:	9.20	Prev response	9.12	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	8.02	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	8.02	0.997
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.997
Baseline Corr AF:	8.02	Prev response	7.97	*% change	0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002044	1.006155
THC Cal Offset:	-0.050467	0.000000
CH ₄ Cal Slope:	1.002402	1.003190
CH ₄ Cal Offset:	-0.039003	0.000000
NMHC Cal Slope:	1.001944	1.008755
NMHC Cal Offset:	-0.011066	0.000000

Notes: Swapped out the N2 and H2 cylinders.

Calibration Performed By: Max Farrell

NMHC Calibration Plot

Date: February 14, 2024

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Anzac Station number: AMS14
Calibration Date: February 6, 2024 Last Cal Date: January 5, 2024
Start time (MST): 9:50 End time (MST): 14:39
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0037092 Cal Gas Expiry Date: May 16, 2031
NOX Cal Gas Conc: 60.7 ppm NO Cal Gas Conc: 60.4 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 60.7 ppm Removed Gas NO Conc: 60.4 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3060
ZAG make/model: Teledyne API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1426262592
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.411	1.411	NO bkgnd or offset:	3.8	3.8
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.8	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	159.1	160.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995638	1.001051
NO _x Cal Offset:	-0.669728	-0.650468
NO Cal Slope:	1.002181	1.006866
NO Cal Offset:	-2.129361	-1.970746
NO ₂ Cal Slope:	0.992183	0.997816
NO ₂ Cal Offset:	-0.766639	-1.058817



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as found span	4934	66.3	804.8	800.9	4.0	806.7	803.4	3.4	0.9977	0.9968
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4934	66.3	804.8	800.9	4.0	805.2	805.3	-0.1	0.9995	0.9945
second point	4985	33.2	401.6	399.6	2.0	401.6	399.6	2.0	1.0000	1.0000
third point	5004	16.7	201.9	200.9	1.0	200.3	198.2	2.1	1.0080	1.0136
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4934	66.3	804.8	415.7	389.2	804.8	421.4	383.4	1.0000	0.9864
Average Correction Factor									1.0025	1.0027

Corrected As found	NO _x =	806.8 ppb	NO =	803.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x =	0.8%
Previous Response	NO _x =	800.7 ppb	NO =	800.5 ppb		*Percent Change	NO =	0.4%
Baseline Corr 2nd pt	NO _x =	NA ppb	NO =	NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x =	NA ppb	NO =	NA ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	801.3	416.1	389.2	387.9	1.0033	99.7%
2nd GPT point (200 ppb O ₃)	801.3	608.6	196.7	194.5	1.0112	98.9%
3rd GPT point (100 ppb O ₃)	801.3	705.3	100.0	97.6	1.0244	97.6%
Average Correction Factor					1.0130	98.7%

Notes:

Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By:

Mohammed Kashif

CALS_287



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 5, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	9:50	End Time (MST):	14:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
804.8	805.2	0.9995		
401.6	401.6	1.0000		
201.9	200.3	1.0080		
			0.999994	
			1.001051	
			-0.650468	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 5, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	9:50	End Time (MST):	14:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
800.9	805.3	0.9945			
399.6	399.6	1.0000			
200.9	198.2	1.0136			
			Slope	1.006866	0.90 - 1.10
			Intercept	-1.970746	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 5, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	9:50	End Time (MST):	14:39
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	<i>≥0.995</i> <i>0.90 - 1.10</i> <i>+/-20</i>
389.2	387.9	1.0033		
196.7	194.5	1.0112		
100.0	97.6	1.0244		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 6, 2024

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	February 5, 2024	Last Cal Date:	January 8, 2024
Start time (MST):	11:31	End time (MST):	14:16
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	API T700	Serial Number:	3060
ZAG Make/Model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595
Analyzer Range:	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001343	0.999057	Backgd or Offset:	1.4	1.4
Calibration intercept:	-0.260000	-0.260000	Coeff or Slope:	1.620	1.620

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	5000	918.8	400.0	399.8	1.001
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.3	----
high point	5000	918.8	400.0	399.6	1.001
second point	5000	803.8	200.0	199.4	1.003
third point	5000	709.8	100.0	99.0	1.010
as left zero	5000	0.0	0.0	0.1	----
as left span	5000	918.8	400.0	403.5	0.991
Average Correction Factor					1.005

Baseline Corr As found:	399.4	Previous response	400.3	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filter changed after as founds. No adjustments required.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 8, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:31	End Time (MST):	14:16
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999991	≥0.995
400.0	399.6	1.0010			
200.0	199.4	1.0030	Slope	0.999057	0.90 - 1.10
100.0	99.0	1.0101			
			Intercept	-0.260000	+/- 5



O₃ Calibration Plot

Date: February 5, 2024

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Anzac Station number: AMS 14
 Calibration Date: February 26, 2024 Last Cal Date: January 24, 2024
 Start time (MST): 11:36 End time (MST): 13:40

Analyzer Make: AP T640 S/N: 825
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
 Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-21.4	-20.58	-21.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	717.8	719.24	717.8	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.036	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	43	----	43	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	34.9	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: _____ Expiry Date: _____
Lot No.: _____

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ December 6, 2023
Date Disposable Filter Changed: _____ December 6, 2023

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ July 6, 2023
Date RH/T Sensor Cleaned: _____ July 6, 2023

Notes: No adjustments made. Leak check passed. Head cleaned.

Calibration by: Mohammed Kashif



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS17
WAPASU**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 8, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:04	End Time (MST):	14:02
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.5	----	Correlation Coefficient	≥0.995
800.0	801.1	0.9986		
400.0	397.1	1.0074	Slope	0.90 - 1.10
199.5	197.4	1.0107		
			Intercept	+/-30



SO2 Calibration Plot

Date: February 6, 2024

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	February 12, 2024	Last Cal Date:	January 29, 2024
Start time (MST):	11:00	End time (MST):	15:25
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.076	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511852			
Removed Cal Gas Conc:	5.076	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2449
ZAG Make/Model:	API T701H		Serial Number:	359

Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583
Converter make:	n/a	Converter serial #:	n/a
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009996	1.010139	Backgd or Offset:	12.2
Calibration intercept:	-0.299186	0.220812	Coeff or Slope:	1.114

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	78.8	80.0	81.7	0.978
as found 2nd point	4961	39.4	40.0	40.3	0.990
as found 3rd point	4980	19.7	20.0	20.4	0.976
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	78.8	80.0	81.0	0.988
second point	4961	39.4	40.0	40.7	0.983
third point	4980	19.7	20.0	20.4	0.980
as left zero	5000	0.0	0.0	0.5	----
as left span	4921	78.8	80.0	79.8	1.003
SO2 Scrubber Check	4921	79.4	800.0	0.0	----
Date of last scrubber change:		n/a		Ave Corr Factor	0.984
Date of last converter efficiency test:		n/a			efficiency

Baseline Corr As found:	81.8	Prev response:	80.50	*% change:	1.6%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.021568	AF Intercept:	-0.179202
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999947		

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 12, 2024	Previous Calibration:	January 29, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:00	End Time (MST):	15:25
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999998	≥0.995
80.0	81.0	0.9877			
40.0	40.7	0.9827	Slope	1.010139	0.90 - 1.10
20.0	20.4	0.9804			
			Intercept	0.220812	+/-3



H₂S Calibration Plot

Date: February 12, 2024

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	February 6, 2024	Last Cal Date:	January 8, 2024
Start time (MST):	11:04	End time (MST):	14:02
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	ALM066507	Cal Gas Expiry Date:	January 12, 2029
CH4 Cal Gas Conc.	<u>503.5</u> ppm	CH4 Equiv Conc.	1076.3 ppm
C3H8 Cal Gas Conc.	<u>208.3</u> ppm		
Removed Gas Cert:	n/a	Removed Gas Expiry:	n/a
Removed CH4 Conc.	<u>503.5</u> ppm	CH4 Equiv Conc.	1076.3 ppm
Removed C3H8 Conc.	<u>208.3</u> ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	2449
ZAG Make/Model:	API 701H	Serial Number:	359

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1218153352
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993554	0.992396	Background:	3.020	3.020
Calibration intercept:	0.002025	0.034435	Coefficient:	4.390	4.390

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.03	----
as found span	4921	79.4	17.09	17.09	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.09	----
high point	4921	79.4	17.09	17.05	1.003
second point	4960	39.7	8.55	8.42	1.015
third point	4980	19.8	4.26	4.26	1.001
as left zero	5000	0.0	0.00	-0.02	----
as left span	4921	79.4	17.09	17.03	1.004
Average Correction Factor					1.006
Baseline Corr As found:	17.06	Previous response	16.98	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 8, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:04	End Time (MST):	14:02
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.09	----	Correlation Coefficient	0.999907	≥0.995
17.09	17.05	1.0027			
8.55	8.42	1.0151	Slope	0.992396	0.90 - 1.10
4.26	4.26	1.0008			
			Intercept	0.034435	+/-1.5



THC Calibration Plot

Date: February 6, 2024

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Wapasu
Calibration Date: February 13, 2024
Start time (MST): 10:59
Reason: Routine
Station number: AMS17
Last Cal Date: January 10, 2024
End time (MST): 16:05

Calibration Standards

NO Gas Cylinder #: T375YK8
NOX Cal Gas Conc: 49.11 ppm
Removed Cylinder #: T375YK8
Removed Gas NOX Conc: 49.11 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: April 13, 2025
NO Cal Gas Conc: 48.07 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 48.07 ppm
NO gas Diff:
Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: Thermo Scientific 42iQ
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 12300522720

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.250	1.264	NO bkgnd or offset:	8.1	8.1
NOX coeff or slope:	0.986	0.986	NOX bkgnd or offset:	8.1	8.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	326.0	335.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003354	0.998991
NO _x Cal Offset:	-4.920000	-5.260000
NO Cal Slope:	1.006788	1.001959
NO Cal Offset:	-5.500000	-5.660000
NO ₂ Cal Slope:	0.997030	0.995634
NO ₂ Cal Offset:	-0.662092	-1.245592



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	0.0	----	----
as found span	4917	83.2	817.2	799.9	17.3	807.8	792.4	15.5	1.0116	1.0094
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.1	----	----
high point	4917	83.2	817.2	799.9	17.3	814.3	799.3	14.9	1.0035	1.0007
second point	4958	41.6	408.6	399.9	8.7	398.3	389.9	8.4	1.0258	1.0258
third point	4979	20.8	204.3	200.0	4.3	195.3	191.0	4.3	1.0461	1.0470
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as left span	4917	83.2	817.2	393.2	424.0	813.3	391.6	421.7	1.0047	1.0040
Average Correction Factor									1.0252	1.0245

Corrected As found	NO _x = 808.2 ppb	NO = 792.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.8%	
Previous Response	NO _x = 815.0 ppb	NO = 799.8 ppb		*Percent Change	NO = -0.9%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	801.4	394.7	424.0	421.7	1.0055	99.5%
2nd GPT point (200 ppb O3)	801.4	598.2	220.5	217.5	1.0138	98.6%
3rd GPT point (100 ppb O3)	801.4	699.5	119.2	116.1	1.0267	97.4%
Average Correction Factor					1.0153	98.5%

Notes:

Change filters after as founds. Adjusted span only.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 10, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:59	End Time (MST):	16:05
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
817.2	814.3	1.0035			
408.6	398.3	1.0258			
204.3	195.3	1.0461			
			Slope	0.998991	0.90 - 1.10
			Intercept	-5.260000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 10, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:59	End Time (MST):	16:05
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999768	≥0.995
799.9	799.3	1.0007			
399.9	389.9	1.0258	Slope	1.001959	0.90 - 1.10
200.0	191.0	1.0470			
			Intercept	-5.660000	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 10, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:59	End Time (MST):	16:05
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
424.0	421.7	1.0055		
220.5	217.5	1.0138		
119.2	116.1	1.0267		
			0.999949	
			0.995634	
			-1.245592	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 13, 2024

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Wapasu Station number: AMS17
 Calibration Date: February 9, 2024 Last Cal Date: January 4, 2024
 Start time (MST): 10:45 End time (MST): 14:00
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 2449
 ZAG Make/Model: API T701H Serial Number: 359

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 3870
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001714	1.006914	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.500000	-0.560000	Coeff or Slope:	1.014	1.027

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	5000	1077.3	400.0	394.0	1.015
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.1	----
high point	5000	1077.3	400.0	402.5	0.994
second point	5000	900.3	200.0	200.6	0.997
third point	5000	789.5	100.0	99.4	1.006
as left zero	5000	0.0	0.0	0.2	----
as left span	5000	1077.3	400.0	409.8	0.976
Average Correction Factor					0.999

Baseline Corr As found:	393.7	Previous response	400.2	*% change	-1.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 9, 2024	Previous Calibration:	January 4, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:45	End Time (MST):	14:00
Analyzer make:	API T400	Analyzer serial #:	3870

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	≥0.995
400.0	402.5	0.9938		
200.0	200.6	0.9970	Slope	0.90 - 1.10
100.0	99.4	1.0060		
			Intercept	+/- 5



O₃ Calibration Plot

Date: February 9, 2024

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Wapasu Station number: AMS 17
 Calibration Date: February 15, 2024 Last Cal Date: February 13, 2024
 Start time (MST): 15:30 End time (MST): 16:30

Analyzer Make: Teledyne API T640 S/N: 1183
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388744
 Temp/RH standard: Alicat FP-25BT S/N: 388744

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-8.3	-8.5	-8.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	721.1	723.9	721.1	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	4.66	5.00	<input checked="" type="checkbox"/>	+/- 0.25 LPM
PW% (pump)		----		<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	5.1	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: _____ Expiry Date: _____
Lot No.: _____

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ February 13, 2024
Date Disposable Filter Changed: _____ February 13, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____
Date RH/T Sensor Cleaned: _____

Notes: Flow was observed low during audit, calibrated flow to volumetric standard.

Calibration by: Kelly Baragar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 1, 2024	Previous Calibration:	January 11, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:13	End Time (MST):	15:25
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999997	≥0.995
800.3	802.5	0.9972			
400.2	401.0	0.9980	Slope	1.002503	0.90 - 1.10
199.6	199.3	1.0015			
			Intercept	-0.043083	+/-30



SO2 Calibration Plot

Date: February 1, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: February 15, 2024 Last Cal Date: January 30, 2024
 Start time (MST): 11:15 End time (MST): 15:45
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.479 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC500395
 Removed Cal Gas Conc: 5.479 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
 ZAG Make/Model: Teledyne API T701 Serial Number: 360

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359
 Converter make: CD Nova CDN-101 Converter serial #: 555
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000728	0.995441	Backgd or Offset:	2.66	2.66
Calibration intercept:	0.341016	0.301088	Coeff or Slope:	1.157	1.157

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4927	73.0	80.0	81.0	0.989
as found 2nd point	4964	36.5	40.0	40.9	0.980
as found 3rd point	4983	18.3	20.0	20.4	0.988
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4927	73.0	80.0	79.9	1.001
second point	4964	36.5	40.0	40.2	0.995
third point	4983	18.3	20.0	20.2	0.992
as left zero	5000	0.0	0.0	0.3	----
as left span	4927	73.0	80.0	79.5	1.006
SO2 Scrubber Check	4923	77.1	771.0	-0.1	----
Date of last scrubber change:	17-Dec-21			Ave Corr Factor	0.996
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 80.9 Prev response: 80.39 *% change: 0.6%
 Baseline Corr 2nd AF pt: 40.8 AF Slope: 1.011732 AF Intercept: 0.180788
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999975

* = > +/-5% change initiates investigation

Notes: Scrubber check completed after calibrator zero. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 30, 2024
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	11:15	End Time (MST):	15:45
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999997	≥0.995
80.0	79.9	1.0012			
40.0	40.2	0.9948	Slope	0.995441	0.90 - 1.10
20.0	20.2	0.9925			
			Intercept	0.301088	+/-3



TRS Calibration Plot

Date: February 15, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	February 1, 2024	Last Cal Date:	January 23, 2024
Start time (MST):	11:13	End time (MST):	15:25
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.13E-04	2.88E-04	NMHC SP Ratio:	6.14E-05	5.97E-05
CH ₄ Retention time:	16.2	15.8	NMHC Peak Area:	149456	153566
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.42	----
as found span	4919	81.0	17.28	14.36	1.204
as found 2nd point	4959	40.5	8.64	7.45	1.159
as found 3rd point	4979	20.2	4.31	3.95	1.091
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.16	1.007
second point	4959	40.5	8.64	8.62	1.002
third point	4979	20.2	4.31	4.30	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.38	0.995
Average Correction Factor					1.004

Baseline Corr AF:	13.94	Prev response	17.41	*% change	-24.9%
Baseline Corr 2nd AF:	7.0	AF Slope:	0.805889	AF Intercept:	0.454086
Baseline Corr 3rd AF:	3.5	AF Correlation:	0.999968	* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	7.48	1.226
as found 2nd point	4959	40.5	4.58	3.80	1.207
as found 3rd point	4979	20.2	2.29	1.91	1.199
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.07	1.010
second point	4959	40.5	4.58	4.60	0.998
third point	4979	20.2	2.29	2.28	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	9.17	9.26	0.990
Average Correction Factor					1.003
Baseline Corr AF:	7.48	Prev response	9.27	*% change	-24.0%
Baseline Corr 2nd AF:	3.8	AF Slope:	0.815019	AF Intercept:	0.028057
Baseline Corr 3rd AF:	1.9	AF Correlation:	0.999917	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.42	----
as found span	4919	81.0	8.11	6.88	1.179
as found 2nd point	4959	40.5	4.06	3.66	1.110
as found 3rd point	4979	20.2	2.02	2.04	0.991
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	8.09	1.003
second point	4959	40.5	4.06	4.03	1.007
third point	4979	20.2	2.02	2.02	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.12	1.000
Average Correction Factor					1.004
Baseline Corr AF:	6.46	Prev response	8.14	*% change	-26.0%
Baseline Corr 2nd AF:	3.24	AF Slope:	0.795742	AF Intercept:	0.425429
Baseline Corr 3rd AF:	1.62	AF Correlation:	0.999997	* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.006554	0.993093
THC Cal Offset:	0.018647	0.015393
CH ₄ Cal Slope:	1.002919	0.997044
CH ₄ Cal Offset:	0.002793	-0.003610
NMHC Cal Slope:	1.009996	0.989722
NMHC Cal Offset:	0.014455	0.019003

Notes: Zero chromatogram turned on, carrier pressure increased from 27 psi to 28 psi. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 1, 2024	Previous Calibration:	January 23, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:13	End Time (MST):	15:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995
17.28	17.16	1.0068			
8.64	8.62	1.0023			
4.31	4.30	1.0024			
			Slope	0.993093	0.90 - 1.10
			Intercept	0.015393	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 1, 2024	Previous Calibration:	January 23, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:13	End Time (MST):	15:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999993	≥0.995
8.11	8.09	1.0027			
4.06	4.03	1.0072			
2.02	2.02	1.0028			
			Slope	0.997044	0.90 - 1.10
			Intercept	-0.003610	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 1, 2024	Previous Calibration:	January 23, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:13	End Time (MST):	15:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999951	≥ 0.995
9.17	9.07	1.0104			
4.58	4.60	0.9977			
2.29	2.28	1.0021			
			Slope	0.989722	0.90 - 1.10
			Intercept	0.019003	+/-0.5



NMHC Calibration Plot

Date: February 1, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: February 21, 2024 Last Cal Date: January 25, 2024
Start time (MST): 11:12 End time (MST): 16:16
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T26DHGA Cal Gas Expiry Date: November 17, 2026
NOX Cal Gas Conc: 48.28 ppm NO Cal Gas Conc: 47.58 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.28 ppm Removed Gas NO Conc: 47.58 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 2658
ZAG make/model: Teledyne API T701 Serial Number: 13779

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1336160088
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.102	1.104	NO bkgnd or offset:	3.1	3.1
NOX coeff or slope:	0.988	0.995	NOX bkgnd or offset:	3.1	3.1
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	252.3	253.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000344	1.001950
NO _x Cal Offset:	-0.080000	-1.100000
NO Cal Slope:	1.001207	1.001421
NO Cal Offset:	-0.760000	-1.260000
NO ₂ Cal Slope:	0.998708	1.009656
NO ₂ Cal Offset:	0.435337	1.192503



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as found span	4916	84.0	811.1	799.3	11.8	807.4	799.9	7.5	1.0046	0.9993
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
high point	4916	84.0	811.1	799.3	11.8	812.1	799.7	12.5	0.9988	0.9996
second point	4958	42.0	405.6	399.7	5.9	404.8	398.7	6.1	1.0019	1.0024
third point	4979	21.0	202.8	199.8	2.9	200.9	197.5	3.4	1.0093	1.0118
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as left span	4916	84.0	811.1	360.3	450.8	817.0	359.0	457.8	0.9928	1.0037
Average Correction Factor									1.0033	1.0046

Corrected As found	NO _x = 807.5 ppb	NO = 800.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.5%
Previous Response	NO _x = 811.3 ppb	NO = 799.5 ppb		*Percent Change	NO = 0.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.0	357.0	450.8	455.4	0.9898	101.0%
2nd GPT point (200 ppb O3)	796.0	591.2	216.6	221.6	0.9773	102.3%
3rd GPT point (100 ppb O3)	796.0	693.2	114.6	117.1	0.9783	102.2%
Average Correction Factor					0.9818	101.9%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 25, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:12	End Time (MST):	16:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
811.1	812.1	0.9988			
405.6	404.8	1.0019			
202.8	200.9	1.0093			
			Slope	1.001950	0.90 - 1.10
			Intercept	-1.100000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 25, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:12	End Time (MST):	16:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.3	799.7	0.9996		
399.7	398.7	1.0024		
199.8	197.5	1.0118		
			0.999990	
			1.001421	
			-1.260000	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 25, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:12	End Time (MST):	16:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
450.8	455.4	0.9898		
216.6	221.6	0.9773		
114.6	117.1	0.9783		



NO_x Calibration Plot

Date: February 21, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 14, 2024	Previous Calibration:	January 24, 2024
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	11:50	End Time (MST):	15:07
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.6	----	Correlation Coefficient	0.999998	
400.0	401.3	0.9968			≥0.995
200.0	200.6	0.9970	Slope	1.002057	
100.0	100.3	0.9970			0.90 - 1.10
			Intercept	0.340000	+/- 5



O₃ Calibration Plot

Date: February 14, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2023

Station Information

Station Name: Stony Mountain Station number: AMS 18
 Calibration Date: February 21, 2024 Last Cal Date: January 30, 2024
 Start time (MST): 14:43 End time (MST): 15:58

Analyzer Make: API T640 S/N: 1162
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	7.3	6.8	7.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	695.7	694.6	695.7	<input type="checkbox"/>	+/- 10 mmHg
flow (LPM)	4.99	6.06	4.99	<input checked="" type="checkbox"/>	+/- 0.25 LPM
Leak Test:	Date of check: February 21, 2024	Last Cal Date: January 30, 2024			
	PM w/o HEPA: 0.7	PM w/ HEPA: 0.0			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

Inlet cleaning : Inlet Head

Quarterly Calibration Test

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.9	10.9	10.9	<input type="checkbox"/>	10.9 +/- 0.5
Post-maintenance leak check:		PM w/o HEPA: _____	w/ HEPA: _____		
Date Optical Chamber Cleaned:		February 21, 2024			<0.2 ug/m3
Disposable Filter Changed:		February 21, 2024			

Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022
 Date RH/T Sensor Cleaned: August 30, 2022

Notes: Flow adjusted. Quarterly checks done. Filter changed out.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	February 5, 2024	Last Cal Date:	January 17, 2024
Start time (MST):	12:24	End time (MST):	15:36
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,050	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM063503			
Removed Cal Gas Conc:	3,050	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701H		Serial Number:	355

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999887	0.999634	Backgd or Offset:	-0.010	-0.010
Calibration intercept:	0.043782	0.085783	Coeff or Slope:	0.908	0.901

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4933	66.7	40.7	41.1	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.7	40.7	1.001
second point	4966	33.3	20.3	20.6	0.987
third point	4983	16.7	10.2	10.3	0.991
as left zero	5000	0.0	0.0	0.0	----
as left span	4933	66.7	40.7	40.7	0.999
Average Correction Factor					0.993

Baseline Corr As found:	41.15	Prev response:	40.73	*% change:	1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 17, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	12:24	End Time (MST):	15:36
Analyzer make:	API T300	Analyzer serial #:	3504

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999934	
40.7	40.7	1.0007			≥ 0.995
20.3	20.6	0.9867	Slope	0.999634	
10.2	10.3	0.9910			$0.90 - 1.10$
			Intercept	0.085783	± 1.5



CO Calibration Plot

Date: February 5, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	February 7, 2024	Last Cal Date:	January 15, 2024
Start time (MST):	10:41	End time (MST):	14:04
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	59100 ppm	Cal Gas Exp Date:	November 4, 2028
Cal Gas Cylinder #:	EB0065608		
Removed Cal Gas Conc:	59,100 ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	2658
N2 Gen Make/Model:	Peak Scientific	Serial Number:	771048317

Analyzer Information

Analyzer make: API T360	Analyzer serial #: 489
Analyzer Range 0 - 2,000 ppm	

	Start	Finish		Start	Finish
Calibration slope:	0.996375	1.030319	Backgd or Offset:	-0.037	-0.037
Calibration intercept:	-3.620000	-5.680000	Coeff or Slope:	0.938	0.945

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-0.1	----
as found span	2920	80.0	1576.0	1569.4	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	0.4	----
high point	2920	80.0	1576.0	1623.3	0.971
second point	2960	40.0	788.0	796.7	0.989
third point	2980	20.0	394.0	398.5	0.989
as left zero	3000	0.0	0.0	0.1	----
as left span	2930	80.0	1570.8	1620.3	0.969

Average Correction Factor	0.983
---------------------------	-------

Baseline Corr As found:	1569.50	Prev response:	1566.67	*% change:	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	February 7, 2024	Previous Calibration	January 15, 2024
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:41	End Time (MST)	14:04
Analyzer make	API T360	Analyzer serial #	489

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.4	----	Correlation Coefficient	0.999891	≥0.995
1576.0	1623.3	0.9709			
788.0	796.7	0.9891	Slope	1.030319	0.90 - 1.10
394.0	398.5	0.9887			
			Intercept	-5.680000	+/-10



CO₂ Calibration Plot

Date: February 7, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Stony Mountain	Station Number:	AMS 18
Calibration Date:	February 21, 2024	Prev Cal Date:	NA
Start Time (MST):	12:51	End Time (MST):	14:55
Tower Height (m):	10.0	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C	Serial Number:	D16124
WS Calibrator:	RM Young 053-120	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.6	0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥0.9995
Calculated slope		0.998960	0.90 - 1.10
Calculated intercept		0.026359	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	D14060
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	12:37	Calc Declination*:	14 Degrees
Deadband calc:	1.2 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.4	---
90	89.2	-0.2%
180	180.4	0.1%
270	269.8	-0.1%
357	356.2	-0.2%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999993	≥0.9995
Calculated slope		1.001927	0.90 - 1.10
Calculated intercept		-0.139279	+/- 4

Notes: Installing new WD sensors on the 10m tower. Crossarm aligned with true north. No issues to note.

Calibration Performed By: Aswin Sasi Kumar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS19
FIREBAG**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	February 27, 2024	Last Cal Date:	January 8, 2024
Start time (MST):	11:45	End time (MST):	14:51
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.29	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC716618			
Removed Cal Gas Conc:	49.29	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1607
ZAG Make/Model:	API T701		Serial Number:	1118

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308
Analyzer Range	0 - 1000 ppb		

	Start	Finish	Start	Finish
Calibration slope:	0.996990	1.003793	Backgd or Offset: 10.5	10.5
Calibration intercept:	0.518257	0.237576	Coeff or Slope: 0.984	0.984

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	-0.1	----
as found span	4919	81.1	799.5	801.0	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.0	0.1	----
high point	4919	81.1	799.5	803.0	0.996
second point	4959	40.6	400.3	401.1	0.998
third point	4980	20.3	200.1	201.9	0.991
as left zero	4999	0.0	0.0	0.0	----
as left span	4919	81.1	799.5	801.0	0.998
Average Correction Factor					0.995

Baseline Corr As found:	801.10	Previous response	797.58	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	January 8, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:45	End Time (MST):	14:51
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999996	≥0.995
799.5	803.0	0.9956			
400.3	401.1	0.9979	Slope	1.003793	0.90 - 1.10
200.1	201.9	0.9911			
			Intercept	0.237576	+/-30



SO2 Calibration Plot

Date: February 27, 2024

Location: Firebag





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Firebag	Station number:	AMS19
Calibration Date:	February 1, 2024	Last Cal Date:	January 22, 2024
Start time (MST):	11:28	End time (MST):	16:23
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.114	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517427			
Removed Cal Gas Conc:	5.114	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1607
ZAG Make/Model:	Teledyne API T701		Serial Number:	1118

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032
Converter make:	Global	Converter serial #:	2022-222
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004771	1.000909	Backgd or Offset:	3.38
Calibration intercept:	-0.021722	-0.061541	Coeff or Slope:	1.225
				1.166

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.6	----
as found span	4922	78.2	80.0	84.3	0.942
as found 2nd point	4961	39.1	40.0	41.9	0.941
as found 3rd point	4980	19.6	20.0	20.7	0.941
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	78.2	80.0	80.0	1.000
second point	4961	39.1	40.0	40.0	1.000
third point	4980	19.6	20.0	19.9	1.007
as left zero	5000	0.0	0.0	0.1	----
as left span	4922	78.2	80.0	79.4	1.007
SO2 Scrubber Check	4922	78.3	800.2	0.0	----
Date of last scrubber change:	January 18, 2023			Ave Corr Factor	1.002
Date of last converter efficiency test:	n/a			efficiency	

Baseline Corr As found:	84.9	Prev response:	80.34	*% change:	5.4%
Baseline Corr 2nd AF pt:	42.5	AF Slope:	1.061511	AF Intercept:	-0.582906
Baseline Corr 3rd AF pt:	21.3	AF Correlation:	1.000000		

* = > +/-5% change initiates investigation

Notes: Ran SOx scrubber check after cal zero. Adjusted zero and span. Large adjustment was due to changes in temperature.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 1, 2024	Previous Calibration:	January 22, 2024
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	11:28	End Time (MST):	16:23
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999995	≥0.995
80.0	80.0	0.9997			
40.0	40.0	0.9998	Slope	1.000909	0.90 - 1.10
20.0	19.9	1.0075			
			Intercept	-0.061541	+/-3



H₂S Calibration Plot

Date: February 1, 2024

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 27, 2024	Previous Calibration:	January 8, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:45	End Time (MST):	14:51
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.03	----	Correlation Coefficient	0.999987	≥0.995
17.31	17.31	0.9997			
8.66	8.61	1.0063	Slope	1.002921	0.90 - 1.10
4.33	4.26	1.0180			
			Intercept	-0.061568	+/-1.5



THC Calibration Plot

Date: February 27, 2024

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Firebag
Calibration Date: February 13, 2024
Start time (MST): 11:30
Reason: Routine
Station number: AMS 19
Last Cal Date: January 23, 2024
End time (MST): 16:55

Calibration Standards

NO Gas Cylinder #: DT0044018
NOX Cal Gas Conc: 48.9 ppm
Removed Cylinder #: n/a
Removed Gas NOX Conc: 48.9 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.7 ppm
Removed Gas Exp Date: n/a
Removed Gas NO Conc: 48.7 ppm
NO gas Diff:
Serial Number: 1607
Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1410661309

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.089	1.081	NO bkgnd or offset:	7.6	7.6
NOX coeff or slope:	0.993	0.996	NOX bkgnd or offset:	7.7	7.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	217.1	216.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.011351	1.006289
NO _x Cal Offset:	0.260390	0.400360
NO Cal Slope:	1.014424	1.006923
NO Cal Offset:	-0.039625	0.340247
NO ₂ Cal Slope:	1.000829	0.998051
NO ₂ Cal Offset:	-0.524220	-0.640112



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
as found span	4919	81.0	792.2	788.9	3.2	813.0	811.0	1.3	0.9744	0.9728
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	4999	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
high point	4919	81.5	797.1	793.8	3.3	802.0	799.3	3.1	0.9939	0.9931
second point	4959	40.7	398.0	396.4	1.6	401.8	400.1	1.7	0.9907	0.9908
third point	4980	20.4	199.5	198.7	0.8	201.4	200.6	0.8	0.9906	0.9905
as left zero	4999	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as left span	4919	81.0	792.2	377.4	414.8	807.0	387.6	418.9	0.9816	0.9737
Average Correction Factor									0.9917	0.9915

Corrected As found	NO _x = 813.3 ppb	NO = 811.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.5%
Previous Response	NO _x = 801.4 ppb	NO = 800.3 ppb		*Percent Change	NO = 1.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.8	385.3	414.8	413.5	1.0030	99.7%
2nd GPT point (200 ppb O3)	796.8	592.0	208.1	207.0	1.0051	99.5%
3rd GPT point (100 ppb O3)	796.8	694.7	105.4	103.8	1.0150	98.5%
Average Correction Factor					1.0077	99.2%

Notes: Changed sample inlet filters after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 23, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:30	End Time (MST):	16:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
797.1	802.0	0.9939			
398.0	401.8	0.9907			
199.5	201.4	0.9906			
			Slope	1.006289	0.90 - 1.10
			Intercept	0.400360	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 23, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:30	End Time (MST):	16:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
793.8	799.3	0.9931		
396.4	400.1	0.9908		
198.7	200.6	0.9905		
			0.999998	
			1.006923	
			0.340247	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 23, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:30	End Time (MST):	16:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
414.8	413.5	1.0030			
208.1	207.0	1.0051			
105.4	103.8	1.0150			
			Slope	0.998051	0.90 - 1.10
			Intercept	-0.640112	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 13, 2024

Location: Firebag





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS20
MACKAY RIVER**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 20, 2024	Previous Calibration:	January 16, 2024
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	9:00	End Time (MST):	11:51
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999941	
800.3	805.6	0.9934			≥0.995
400.7	405.7	0.9876	Slope	1.004391	
199.8	207.0	0.9653			0.90 - 1.10
			Intercept	2.870973	+/-30



SO2 Calibration Plot

Date: February 20, 2024

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: MacKay River Station number: AMS20
 Calibration Date: February 1, 2024 Last Cal Date: January 3, 2024
 Start time (MST): 7:48 End time (MST): 12:00
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.12 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC515997
 Removed Cal Gas Conc: 5.12 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1220
 ZAG Make/Model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1236656117
 Converter make: Global Converter serial #: 2022-226
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.991531	0.987249	Backgd or Offset: 3.19	3.19
Calibration intercept:	0.499365	0.499237	Coeff or Slope: 1.113	1.113

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4922	78.1	80.0	80.4	0.997
as found 2nd point	4961	39.0	39.9	40.5	0.991
as found 3rd point	4980	19.5	20.0	20.7	0.974
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4922	78.1	80.0	79.4	1.008
second point	4961	39.0	40.0	40.0	0.999
third point	4980	19.5	20.0	20.6	0.970
as left zero	5000	0.0	0.0	0.3	----
as left span	4922	78.1	80.0	79.1	1.012
SO2 Scrubber Check	4982	81.3	802.8	0.0	----
Date of last scrubber change:	May 25, 2023			Ave Corr Factor	0.992
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 80.2 Prev response: 79.79 *% change: 0.5%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 1.001167 AF Intercept: 0.439485
 Baseline Corr 3rd AF pt: 20.5 AF Correlation: 0.999958

* = > +/-5% change initiates investigation

Notes: No maintenance and adjustments done. Sox scrubber checked after the calibrator zero.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 1, 2024	Previous Calibration:	January 3, 2024
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:48	End Time (MST):	12:00
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999929	≥0.995
80.0	79.4	1.0080			
40.0	40.0	0.9992	Slope	0.987249	0.90 - 1.10
20.0	20.6	0.9702			
			Intercept	0.499237	+/-3



H₂S Calibration Plot

Date: February 1, 2024

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 20, 2024	Previous Calibration:	January 16, 2024
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	9:00	End Time (MST):	11:50
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.01	----	Correlation Coefficient	0.999975	≥0.995
17.34	17.34	1.0000			
8.68	8.74	0.9932	Slope	0.999403	0.90 - 1.10
4.33	4.39	0.9856			
			Intercept	0.034183	+/-1.5



THC Calibration Plot

Date: February 20, 2024

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as found span	4917	83.3	819.5	800.3	19.2	819.0	797.9	21.1	1.0006	1.0030
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4917	83.3	819.5	800.3	19.2	817.0	797.8	19.2	1.0030	1.0031
second point	4958	41.7	410.3	400.7	9.6	412.9	401.8	11.1	0.9936	0.9972
third point	4979	20.8	204.6	199.9	4.8	211.1	204.3	6.8	0.9694	0.9782
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.0	----	----
as left span	4917	83.3	819.5	449.1	370.4	811.8	439.2	372.5	1.0094	1.0225
Average Correction Factor									0.9887	0.9929

Corrected As found	NO _x = 819.1 ppb	NO = 797.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.3%
Previous Response	NO _x = 821.2 ppb	NO = 800.6 ppb		*Percent Change	NO = -0.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	792.9	441.7	370.4	370.1	1.0007	99.9%
2nd GPT point (200 ppb O3)	792.9	608.7	203.4	202.0	1.0067	99.3%
3rd GPT point (100 ppb O3)	792.9	696.6	115.5	112.6	1.0254	97.5%
Average Correction Factor					1.0109	98.9%

Notes:

No adjustments or maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 8, 2024	Previous Calibration:	January 25, 2024
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	8:00	End Time (MST):	12:17
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
819.5	817.0	1.0030			
410.3	412.9	0.9936			
204.6	211.1	0.9694			
			Slope	0.994581	0.90 - 1.10
			Intercept	3.602279	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 8, 2024	Previous Calibration:	January 25, 2024
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	8:00	End Time (MST):	12:17
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
800.3	797.8	1.0031			
400.7	401.8	0.9972			
199.9	204.3	0.9782			
			Correlation Coefficient	0.999953	≥0.995
			Slope	0.995045	0.90 - 1.10
			Intercept	2.502804	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 8, 2024	Previous Calibration:	January 25, 2024
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:00	End Time (MST):	12:17
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.0	----	Correlation Coefficient	0.999932	≥0.995
370.4	370.1	1.0007			
203.4	202.0	1.0067			
115.5	112.6	1.0254			
			Slope	1.000946	0.90 - 1.10
			Intercept	-1.281410	+/-20



NO_x Calibration Plot

Date: February 8, 2024

Location: MacKay River





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS21
CONKLIN**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 8, 2024	Previous Calibration:	January 15, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:07	End Time (MST):	13:04
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999981	
800.8	800.1	1.0009			≥0.995
400.4	401.4	0.9976	Slope	0.997719	
200.1	203.3	0.9843			0.90 - 1.10
			Intercept	1.675879	+/-30



SO2 Calibration Plot

Date: February 8, 2024

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Conklin Station number: AMS21
 Calibration Date: February 12, 2024 Last Cal Date: January 24, 2024
 Start time (MST): 11:32 End time (MST): 16:13
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.00 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC501204
 Removed Cal Gas Conc: 5.00 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3810
 ZAG Make/Model: Teledyne API 701H Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1236656116
 Converter make: CD-Nova 101 Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011429	1.001571	Backgd or Offset:	2.4	2.42
Calibration intercept:	0.300000	0.420000	Coeff or Slope:	0.992	0.980

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4920	80.0	80.0	81.7	0.980
as found 2nd point	4960	40.0	40.0	41.1	0.976
as found 3rd point	4980	20.0	20.0	20.7	0.971
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4920	80.0	80.0	80.4	0.995
second point	4960	40.0	40.0	40.7	0.983
third point	4980	20.0	20.0	20.6	0.971
as left zero	5000	0.0	0.0	0.8	----
as left span	4920	80.0	80.0	81.0	0.988
SO2 Scrubber Check	4920	80.2	802.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	0.983
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 81.6 Prev response: 81.21 *% change: 0.5%
 Baseline Corr 2nd AF pt: 41.0 AF Slope: 1.019429 AF Intercept: 0.220000
 Baseline Corr 3rd AF pt: 20.6 AF Correlation: 0.999989

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after multiple as founds. SO2 scrubber check done after calibrator zero and it passed. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 12, 2024	Previous Calibration:	January 24, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:32	End Time (MST):	16:13
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999960	≥0.995
80.0	80.4	0.9950			
40.0	40.7	0.9828	Slope	1.001571	0.90 - 1.10
20.0	20.6	0.9709			
			Intercept	0.420000	+/-3



TRS Calibration Plot

Date: February 12, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	February 8, 2024	Last Cal Date:	January 15, 2024
Start time (MST):	10:07	End time (MST):	13:04
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.86E-04	3.71E-04	NMHC SP Ratio:	7.53E-05
CH ₄ Retention time:	17.4	17.4	NMHC Peak Area:	122333
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				120807
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.45	0.982
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.15	0.999
second point	4960	40.1	8.56	8.61	0.995
third point	4980	20.0	4.28	4.35	0.983
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.05	1.004

Average Correction Factor				0.992
Baseline Corr AF:	17.45	Prev response	17.04	*% change 2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.13	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.17	0.997
second point	4960	40.1	4.57	4.61	0.991
third point	4980	20.0	2.28	2.32	0.983
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.14	1.000
Average Correction Factor					0.990
Baseline Corr AF:	9.13	Prev response	9.12	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	8.32	0.960
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.98	1.001
second point	4960	40.1	3.99	4.00	0.999
third point	4980	20.0	2.00	2.03	0.984
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.91	1.010
Average Correction Factor					0.995
Baseline Corr AF:	8.32	Prev response	7.91	*% change	4.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.992346	1.000154
THC Cal Offset:	0.043763	0.033769
CH ₄ Cal Slope:	0.988303	0.997762
CH ₄ Cal Offset:	0.021954	0.015158
NMHC Cal Slope:	0.995866	1.002131
NMHC Cal Offset:	0.021609	0.018811

Notes: Changed sample inlet filters after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 8, 2024	Previous Calibration:	January 15, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:07	End Time (MST):	13:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999982	≥ 0.995			
17.13	17.15	0.9987						
8.56	8.61	0.9947				Slope	1.000154	0.90 - 1.10
4.28	4.35	0.9833						
			Intercept	0.033769	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 8, 2024	Previous Calibration:	January 15, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:07	End Time (MST):	13:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999979	≥0.995
7.99	7.98	1.0009			
3.99	4.00	0.9990			
2.00	2.03	0.9840			
			Slope	0.997762	0.90 - 1.10
			Intercept	0.015158	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 8, 2024	Previous Calibration:	January 15, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:07	End Time (MST):	13:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999980	≥ 0.995			
9.14	9.17	0.9970						
4.57	4.61	0.9910				Slope	1.002131	0.90 - 1.10
2.28	2.32	0.9827						
			Intercept	0.018811	± 0.5			



NMHC Calibration Plot

Date: February 8, 2024

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Conklin
Calibration Date: February 14, 2024
Start time (MST): 9:21
Reason: Routine
Station number: AMS21
Last Cal Date: January 19, 2024
End time (MST): 13:54

Calibration Standards

NO Gas Cylinder #: SA18828
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 364

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1501663731

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.062	1.050	NO bkgnd or offset:	10.6	10.5
NOX coeff or slope:	0.983	0.993	NOX bkgnd or offset:	10.6	10.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	152.1	150.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997935	0.991077
NO _x Cal Offset:	2.088043	2.428003
NO Cal Slope:	0.999591	0.994636
NO Cal Offset:	1.508039	1.768002
NO ₂ Cal Slope:	0.998906	1.001397
NO ₂ Cal Offset:	0.745906	-0.933646



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
as found span	4918	82.0	802.0	800.1	1.9	810.8	808.8	2.0	0.9891	0.9892
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
high point	4918	82.0	802.0	800.3	1.6	796.3	797.2	-1.0	1.0071	1.0039
second point	4959	41.0	401.0	400.2	0.8	400.3	399.9	0.4	1.0017	1.0007
third point	4980	20.5	200.5	200.1	0.4	203.9	202.9	1.0	0.9832	0.9860
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	0.0	----	----
as left span	4918	82.0	802.0	387.7	414.2	791.2	381.7	409.5	1.0136	1.0158
Average Correction Factor									0.9973	0.9969

Corrected As found	NO _x = 811.1 ppb	NO = 809.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.1%
Previous Response	NO _x = 802.4 ppb	NO = 801.3 ppb		*Percent Change	NO = 1.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI: ;

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	793.0	380.4	414.2	414.7	0.9989	100.1%
2nd GPT point (200 ppb O ₃)	793.0	586.6	208.0	205.9	1.0104	99.0%
3rd GPT point (100 ppb O ₃)	793.0	690.4	104.2	103.2	1.0101	99.0%
Average Correction Factor					1.0065	99.4%

Notes: Changed sample inlet filter after as founds. Adjusted Span only. Used 2nd NO reference point because of drift.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 14, 2024	Previous Calibration:	January 19, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:21	End Time (MST):	13:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
802.0	796.3	1.0071			
401.0	400.3	1.0017			
200.5	203.9	0.9832			
			Correlation Coefficient	0.999959	≥0.995
			Slope	0.991077	0.90 - 1.10
			Intercept	2.428003	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 14, 2024	Previous Calibration:	January 19, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:21	End Time (MST):	13:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.3	797.2	1.0039			
400.2	399.9	1.0007			
200.1	202.9	0.9860			
			Slope	0.994636	0.90 - 1.10
			Intercept	1.768002	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 14, 2024	Previous Calibration:	January 19, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:21	End Time (MST):	13:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
414.2	414.7	0.9989		
208.0	205.9	1.0104		
104.2	103.2	1.0101		
			0.999959	
			1.001397	
			-0.933646	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 14, 2024

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Conklin
 Calibration Date: February 1, 2024
 Start time (MST): 10:10
 Reason: Routine
 Station number: AMS21
 Last Cal Date: January 12, 2024
 End time (MST): 13:46

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700
 ZAG Make/Model: Teledyne API 701H
 Serial Number: 3810
 Serial Number: 691

Analyzer Information

Analyzer make: Thermo 49i
 Analyzer Range 0 - 500 ppb
 Analyzer serial #: 1501663734

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000771	1.001543	Backgd or Offset:	-1.1	-1.2
Calibration intercept:	0.740000	0.180000	Coeff or Slope:	1.005	0.998

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	5000	952.6	400.0	402.6	0.994
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.1	----
high point	5000	952.6	400.0	400.8	0.998
second point	5000	801.3	200.0	200.4	0.998
third point	5000	807.8	100.0	100.5	0.995
as left zero	5000	800.0	0.0	0.0	----
as left span	5000	952.8	400.0	401.1	0.997
Average Correction Factor					0.997

Baseline Corr As found:	402.7	Previous response	401.0	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filters changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 1, 2024	Previous Calibration:	January 12, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:10	End Time (MST):	13:46
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	1.000000	≥0.995
400.0	400.8	0.9980			
200.0	200.4	0.9980	Slope	1.001543	0.90 - 1.10
100.0	100.5	0.9950			
			Intercept	0.180000	+/- 5



O₃ Calibration Plot

Date: February 1, 2024

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Conklin Station number: AMS 21
 Calibration Date: February 12, 2024 Last Cal Date: January 25, 2024
 Start time (MST): 12:45 End time (MST): 13:06

Analyzer Make: API T640 S/N: 326
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388751
 Temp/RH standard: Alicat FP-25BT S/N: 388751

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	1.50	1.03	1.50	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	704.00	705.70	704.00	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.98	5.05	4.98	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	40.00	----	40.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.30	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.90** Expiry Date: September 29, 2024
 Lot No.:

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: December 7, 2023
 Date Disposable Filter Changed: December 7, 2023

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: December 7, 2023
 Date RH/T Sensor Cleaned: December 7, 2023

Notes: Verified flow, pressure, temperature and pump power. Leak check passed. No adjustment made.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	February 23, 2024	Last Cal Date:	January 10, 2024
Start time (MST):	11:10	End time (MST):	14:02
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.11	ppm	Cal Gas Exp Date:	January 18, 2029
Cal Gas Cylinder #:	CC281519			
Removed Cal Gas Conc:	50.11	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002679	1.000307	Backgd or Offset:	23.7	24.0
Calibration intercept:	0.383769	1.464248	Coeff or Slope:	1.035	1.035

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4920	79.8	799.8	799.6	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.4	----
high point	4920	79.8	799.8	800.5	0.999
second point	4960	39.9	399.9	403.4	0.991
third point	4980	20.0	200.4	202.1	0.992
as left zero	5000	0.0	0.0	0.4	----
as left span	4920	79.8	799.8	803.4	0.996
Average Correction Factor					0.994

Baseline Corr As found:	799.50	Previous response	802.31	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 23, 2024	Previous Calibration:	January 10, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:10	End Time (MST):	14:02
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.4	----	Correlation Coefficient	0.999983	
799.8	800.5	0.9991			≥0.995
399.9	403.4	0.9913	Slope	1.000307	
200.4	202.1	0.9918			0.90 - 1.10
			Intercept	1.464248	+/-30



SO2 Calibration Plot

Date: February 23, 2024

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Janvier	Station number:	AMS22
Calibration Date:	February 22, 2024	Last Cal Date:	January 24, 2024
Start time (MST):	11:28	End time (MST):	15:20
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.02	ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	CC424047			
Removed Cal Gas Conc:	5.02	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031
Converter make:	CDN-101	Converter serial #:	587
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999522	0.996235	Backgd or Offset:	3.71
Calibration intercept:	0.500639	0.440650	Coeff or Slope:	1.205
				1.188

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4920	79.7	80.0	80.7	0.993
as found 2nd point	4960	39.8	40.0	40.9	0.979
as found 3rd point	4980	19.9	20.0	20.6	0.975
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4920	79.7	80.0	80.0	1.000
second point	4960	39.8	40.0	40.5	0.987
third point	4980	19.9	20.0	20.5	0.975
as left zero	5000	0.0	0.0	0.4	----
as left span	4920	79.7	80.0	80.6	0.993
SO2 Scrubber Check	4920	79.8	798.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.987
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	80.6	Prev response:	80.49	*% change:	0.1%
Baseline Corr 2nd AF pt:	40.8	AF Slope:	1.006373	AF Intercept:	0.360891
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999936		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran an SO2 scrubber check after the calibrator zero. Adjusted the span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 22, 2024	Previous Calibration:	January 24, 2024
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	11:28	End Time (MST):	15:20
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999951	
80.0	80.0	1.0003			≥0.995
40.0	40.5	0.9867	Slope	0.996235	
20.0	20.5	0.9746			0.90 - 1.10
			Intercept	0.440650	+/-3



TRS Calibration Plot

Date: February 22, 2024

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	February 23, 2024	Last Cal Date:	January 17, 2024
Start time (MST):	11:10	End time (MST):	14:03
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH ₄ Cal Gas Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
C ₃ H ₈ Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
Removed C ₃ H ₈ Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.940E-04	3.010E-04	NMHC SP Ratio:	6.33E-05	6.46E-05
CH ₄ Retention time:	13.2	13.4	NMHC Peak Area:	144559	141681
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	16.97	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.19	0.999
second point	4960	39.9	8.59	8.59	1.000
third point	4980	20.0	4.30	4.30	1.001
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.21	0.998
Average Correction Factor					1.000

Baseline Corr AF:	16.97	Prev response	17.11	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	9.04	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.15	9.17	0.998
second point	4960	39.9	4.57	4.58	0.999
third point	4980	20.0	2.29	2.29	1.001
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.15	9.17	0.998
Average Correction Factor					0.999
Baseline Corr AF:	9.04	Prev response	9.09	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	7.93	1.012
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.02	1.000
second point	4960	39.9	4.01	4.01	1.001
third point	4980	20.0	2.01	2.01	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.04	0.998
Average Correction Factor					1.000
Baseline Corr AF:	7.93	Prev response	8.01	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.996236	1.001221
THC Cal Offset:	0.001413	-0.004804
CH ₄ Cal Slope:	0.999459	0.999574
CH ₄ Cal Offset:	-0.005759	-0.000163
NMHC Cal Slope:	0.993433	1.002654
NMHC Cal Offset:	0.007572	-0.004841

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 23, 2024	Previous Calibration:	January 17, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:10	End Time (MST):	14:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995			
17.17	17.19	0.9989						
8.59	8.59	0.9998				Slope	1.001221	0.90 - 1.10
4.30	4.30	1.0008						
			Intercept	-0.004804	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 23, 2024	Previous Calibration:	January 17, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:10	End Time (MST):	14:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	≥0.995			
8.03	8.02	1.0004						
4.01	4.01	1.0009				Slope	0.999574	0.90 - 1.10
2.01	2.01	1.0001						
			Intercept	-0.000163	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 23, 2024	Previous Calibration:	January 17, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:10	End Time (MST):	14:03
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
9.15	9.17	0.9976						
4.57	4.58	0.9990				Slope	1.002654	0.90 - 1.10
2.29	2.29	1.0015						
			Intercept	-0.004841	± 0.5			



NMHC Calibration Plot

Date: February 23, 2024

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	February 29, 2024	Last Cal Date:	February 23, 2024
Start time (MST):	11:50	End time (MST):	15:51
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH ₄ Cal Gas Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
C ₃ H ₈ Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
Removed C ₃ H ₈ Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.010E-04	2.890E-04	NMHC SP Ratio:	6.46E-05	5.52E-05
CH ₄ Retention time:	13.4	13.1	NMHC Peak Area:	141681	165825
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	17.47	0.983
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.17	1.000
second point	4960	39.9	8.59	8.59	1.000
third point	4980	20.0	4.30	4.32	0.996
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.18	0.999
Average Correction Factor					0.999

Baseline Corr AF:	17.47	Prev response	17.19	*% change	1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	9.34	0.979
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.15	9.14	1.000
second point	4960	39.9	4.57	4.58	1.000
third point	4980	20.0	2.29	2.31	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.15	9.17	0.998
Average Correction Factor					0.998
Baseline Corr AF:	9.34	Prev response	9.17	*% change	1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	8.13	0.987
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.02	1.000
second point	4960	39.9	4.01	4.01	1.001
third point	4980	20.0	2.01	2.02	0.998
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.01	1.001
Average Correction Factor					1.000
Baseline Corr AF:	8.13	Prev response	8.02	*% change	1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001221	0.999432
THC Cal Offset:	-0.004804	0.007396
CH ₄ Cal Slope:	0.999574	0.999247
CH ₄ Cal Offset:	-0.000163	0.002236
NMHC Cal Slope:	1.002654	0.999206
NMHC Cal Offset:	-0.004841	0.005960

Notes: Adjusted the window timings. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 29, 2024	Previous Calibration:	February 23, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:50	End Time (MST):	15:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995
17.17	17.17	1.0002			
8.59	8.59	1.0000			
4.30	4.32	0.9962			
			Slope	0.999432	0.90 - 1.10
			Intercept	0.007396	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 29, 2024	Previous Calibration:	February 23, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:50	End Time (MST):	15:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
8.03	8.02	1.0005						
4.01	4.01	1.0006				Slope	0.999247	0.90 - 1.10
2.01	2.02	0.9976						
			Intercept	0.002236	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 29, 2024	Previous Calibration:	February 23, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:50	End Time (MST):	15:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
9.15	9.14	1.0003						
4.57	4.58	0.9997				Slope	0.999206	0.90 - 1.10
2.29	2.31	0.9945						
			Intercept	0.005960	± 0.5			



NMHC Calibration Plot

Date: February 29, 2024

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Janvier
Calibration Date: February 28, 2024
Start time (MST): 10:38
Reason: Routine
Station number: AMS 22
Last Cal Date: January 15, 2024
End time (MST): 15:00

Calibration Standards

NO Gas Cylinder #: DT0047765
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: March 11, 2031
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3806
Serial Number: 691

Analyzer Information

Analyzer make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 833

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.854	0.840	NO bkgnd or offset:	0.0	0.0
NOX coeff or slope:	0.849	0.835	NOX bkgnd or offset:	-0.2	-0.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	6.1	6.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000095	1.000594
NO _x Cal Offset:	1.384206	0.484102
NO Cal Slope:	1.000674	0.999460
NO Cal Offset:	0.024014	-0.576026
NO ₂ Cal Slope:	1.008342	1.004914
NO ₂ Cal Offset:	1.498437	0.223696



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.4	0.3	----	----
as found span	4918	82.0	802.0	800.3	1.6	813.9	804.1	9.9	0.9853	0.9953
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	-0.1	0.3	----	----
high point	4918	82.0	802.0	800.3	1.6	802.7	799.5	3.1	0.9991	1.0010
second point	4960	41.0	400.9	400.1	0.8	402.0	399.2	2.8	0.9973	1.0022
third point	4980	20.5	200.5	200.1	0.4	201.2	198.8	2.4	0.9964	1.0063
as left zero	5000	0.0	0.0	0.0	0.0	0.6	0.2	0.4	----	----
as left span	4918	82.0	802.0	408.6	393.3	797.4	405.5	391.9	1.0057	1.0077
Average Correction Factor									0.9976	1.0032

Corrected As found	NO _x = 813.9 ppb	NO = 804.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.3%
Previous Response	NO _x = 803.4 ppb	NO = 800.9 ppb		*Percent Change	NO = 0.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.5	403.8	393.3	395.4	0.9948	100.5%
2nd GPT point (200 ppb O3)	795.5	601.3	195.8	197.4	0.9921	100.8%
3rd GPT point (100 ppb O3)	795.5	697.4	99.7	100.1	0.9964	100.4%
Average Correction Factor					0.9944	100.6%

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 28, 2024	Previous Calibration:	January 15, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:38	End Time (MST):	15:00
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	≥0.995
802.0	802.7	0.9991		
400.9	402.0	0.9973	Slope	0.90 - 1.10
200.5	201.2	0.9964		
			Intercept	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 28, 2024	Previous Calibration:	January 15, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:38	End Time (MST):	15:00
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
800.3	799.5	1.0010			
400.1	399.2	1.0022			
200.1	198.8	1.0063			
			Slope	0.999460	0.90 - 1.10
			Intercept	-0.576026	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 28, 2024	Previous Calibration:	January 15, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:38	End Time (MST):	15:00
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
393.3	395.4	0.9948		
195.8	197.4	0.9921		
99.7	100.1	0.9964		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 28, 2024

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 20, 2024	Previous Calibration:	January 3, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:23	End Time (MST):	14:46
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999996	≥0.995
400.0	401.7	0.9958			
200.0	201.3	0.9935	Slope	1.003229	0.90 - 1.10
100.0	101.3	0.9872			
			Intercept	0.560000	+/- 5



O₃ Calibration Plot

Date: February 20, 2024

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: February 28, 2024 Last Cal Date: January 31, 2024
 Start time (MST): 11:01 End time (MST): 12:29

Analyzer Make: Teledyne API T640 S/N: 325
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752
 Temp/RH standard: Alicat FP-25BT S/N: 388752

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-14.9	-15.5	-14.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	713.3	714.4	713.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	5.106	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	42	----	42	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.4	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: January 24, 2024
 Date Disposable Filter Changed: January 24, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 26, 2023
 Date RH/T Sensor Cleaned: July 26, 2023

Notes: Verified flow, temperature, and pressure. Leak check passed.

Calibration by: Rene Chamberland



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS23
FORT HILLS**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	February 7, 2024	Last Cal Date:	January 15, 2024
Start time (MST):	9:17	End time (MST):	12:05
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.76	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC281425			
Removed Cal Gas Conc:	49.76	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	451
ZAG Make/Model:	API T701		Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999893	1.000522	Backgd or Offset:	18.5	18.5
Calibration intercept:	-0.983326	-0.903438	Coeff or Slope:	1.063	1.063

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4920	80.3	799.1	797.7	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	80.3	799.1	799.1	1.000
second point	4960	40.2	400.1	398.7	1.003
third point	4980	20.1	200.0	198.6	1.007
as left zero	5000	0.0	0.0	-0.1	----
as left span	4920	80.3	799.1	800.7	0.998
Average Correction Factor					1.004

Baseline Corr As found:	798.00	Previous response	798.03	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No Adjustments and Maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 7, 2024	Previous Calibration:	January 15, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:17	End Time (MST):	12:05
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999995	
799.1	799.1	1.0000			≥0.995
400.1	398.7	1.0034	Slope	1.000522	
200.0	198.6	1.0072			0.90 - 1.10
			Intercept	-0.903438	+/-30



SO2 Calibration Plot

Date: February 7, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Hills Station number: AMS23
 Calibration Date: February 13, 2024 Last Cal Date: January 9, 2024
 Start time (MST): 7:50 End time (MST): 11:40
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.20 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517372
 Removed Cal Gas Conc: 5.20 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 451
 ZAG Make/Model: API T701 Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1300156232
 Converter make: CDN-101 Converter serial #: 594
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995462	1.010172	Backgd or Offset:	1.19 1.96
Calibration intercept:	-0.178345	-0.218012	Coeff or Slope:	1.124 1.160

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4923	77.0	80.0	76.3	1.049
as found 2nd point	4962	38.5	40.0	38.2	1.047
as found 3rd point	4981	19.2	19.9	19.1	1.044
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4923	77.0	80.0	80.8	0.990
second point	4962	38.5	40.0	39.9	1.002
third point	4981	19.2	19.9	19.7	1.013
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.0	80.0	81.3	0.984
SO2 Scrubber Check	4920	80.3	803.0	0.0	----

Date of last scrubber change:	Ave Corr Factor	1.002
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 76.3 Prev response: 79.46 *% change: -4.1%
 Baseline Corr 2nd AF pt: 38.2 AF Slope: 0.953468 AF Intercept: 0.040902
 Baseline Corr 3rd AF pt: 19.1 AF Correlation: 0.999999

* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the third point. Span adjusted. Dianogsitcs similar to last months calibration.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021


Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 9, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:50	End Time (MST):	11:40
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999922	
80.0	80.8	0.9901			≥0.995
40.0	39.9	1.0024	Slope	1.010172	
19.9	19.7	1.0126			0.90 - 1.10
			Intercept	-0.218012	+/-3







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	February 7, 2024	Last Cal Date:	January 15, 2024
Start time (MST):	9:17	End time (MST):	12:04
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
Removed C ₃ H ₈ Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.31E-04	2.31E-04	NMHC SP Ratio:	4.93E-05	4.93E-05
CH ₄ Retention time:	13.0	13.0	NMHC Peak Area:	186451	186451
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	16.96	1.014
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.20	1.000
second point	4960	40.2	8.61	8.63	0.997
third point	4980	20.1	4.30	4.36	0.987
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.19	17.27	0.995
Average Correction Factor					0.995

Baseline Corr AF:	16.96	Prev response	17.32	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.02	1.015
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.19	0.997
second point	4960	40.2	4.59	4.65	0.986
third point	4980	20.1	2.29	2.38	0.963
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	9.16	9.21	0.994
Average Correction Factor					0.982
Baseline Corr AF:	9.02	Prev response	9.22	*% change	-2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	7.94	1.011
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.01	1.003
second point	4960	40.2	4.02	3.99	1.008
third point	4980	20.1	2.01	1.98	1.016
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.03	8.05	0.998
Average Correction Factor					1.009
Baseline Corr AF:	7.94	Prev response	8.10	*% change	-2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.007788	0.999507
THC Cal Offset:	-0.007208	0.025608
CH ₄ Cal Slope:	1.012164	0.997953
CH ₄ Cal Offset:	-0.027257	-0.014037
NMHC Cal Slope:	1.005072	1.000996
NMHC Cal Offset:	0.018048	0.041641

Notes: Span adjusted. Nitrogen cylinder changed.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 7, 2024	Previous Calibration:	January 15, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:17	End Time (MST):	12:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999989	≥ 0.995
17.19	17.20	0.9995			
8.61	8.63	0.9973			
4.30	4.36	0.9870			
			Slope	0.999507	0.90 - 1.10
			Intercept	0.025608	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 7, 2024	Previous Calibration:	January 15, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:17	End Time (MST):	12:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999986	≥0.995
8.03	8.01	1.0028			
4.02	3.99	1.0079			
2.01	1.98	1.0155			
			Slope	0.997953	0.90 - 1.10
			Intercept	-0.014037	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	February 7, 2024	Previous Calibration:	January 15, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	9:17	End Time (MST):	12:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999905	≥ 0.995			
9.16	9.19	0.9967						
4.59	4.65	0.9861				Slope	1.000996	0.90 - 1.10
2.29	2.38	0.9633						
			Intercept	0.041641	± 0.5			



NMHC Calibration Plot

Date: February 7, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Hills
Calibration Date: February 2, 2024
Start time (MST): 7:43
Reason: Routine
Station number: AMS23
Last Cal Date: January 10, 2024
End time (MST): 12:02

Calibration Standards

NO Gas Cylinder #: CC332703
NOX Cal Gas Conc: 49.7 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 49.7 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 28, 2024
NO Cal Gas Conc: 49.7 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 49.7 ppm
NO gas Diff:
Serial Number: 451
Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1152430007

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.081	1.102	NO bkgnd or offset:	3	3
NOX coeff or slope:	0.989	0.990	NOX bkgnd or offset:	3.2	3.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	162.9	162.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998780	1.002680
NO _x Cal Offset:	0.244003	0.104366
NO Cal Slope:	1.000209	1.001767
NO Cal Offset:	-0.556171	-0.976340
NO ₂ Cal Slope:	0.999293	1.002767
NO ₂ Cal Offset:	-0.815351	0.602578



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as found span	4920	80.5	800.2	800.2	0.0	784.3	781.7	2.6	1.020	1.024
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.1	----	----
high point	4920	80.5	800.2	800.2	0.0	802.2	800.9	1.4	0.997	0.999
second point	4960	40.2	399.6	399.6	0.0	401.3	399.4	1.9	0.996	1.000
third point	4980	20.1	199.8	199.8	0.0	200.2	197.8	2.3	0.998	1.010
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as left span	4920	80.5	800.2	470.9	329.3	802.0	470.7	331.4	0.998	1.000
Average Correction Factor									0.997	1.003

Corrected As found	NO _x = 784.4 ppb	NO = 781.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.9%	
Previous Response	NO _x = 799.4 ppb	NO = 799.8 ppb		*Percent Change	NO = -2.3%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	798.0	468.7	329.3	330.5	0.996	100.4%
2nd GPT point (200 ppb O ₃)	798.0	635.5	162.5	164.0	0.991	100.9%
3rd GPT point (100 ppb O ₃)	798.0	715.2	82.8	84.0	0.986	101.4%
Average Correction Factor					0.991	100.9%

Notes:

No maintenance done. Span adjusted.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 10, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:43	End Time (MST):	12:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.2	802.2	0.9975		
399.6	401.3	0.9957		
199.8	200.2	0.9980		
			0.999999	
			1.002680	
			0.104366	





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 10, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:43	End Time (MST):	12:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999992	≥0.995
800.2	800.9	0.9991			
399.6	399.4	1.0004	Slope	1.001767	0.90 - 1.10
199.8	197.8	1.0101			
			Intercept	-0.976340	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 2, 2024	Previous Calibration:	January 10, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:43	End Time (MST):	12:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits			
0.0	0.1	----	Correlation Coefficient	0.999988	≥0.995			
329.3	330.5	0.9964						
162.5	164.0	0.9909				Slope	1.002767	0.90 - 1.10
82.8	84.0	0.9857						
			Intercept	0.602578	+/-20			



NO_x Calibration Plot

Date: February 2, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Hills Station number: AMS 23
 Calibration Date: February 7, 2024 Last Cal Date: January 15, 2024
 Start time (MST): 8:04 End time (MST): 9:23

Analyzer Make: API T640 S/N: 1546
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-10.4	-10.7	-10.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	734.1	733.4	734.1	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5	5.09	5	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	40	----	40	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	11.5	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: 10-Jun-24
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.2	10.9	10.9	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 7, 2024
 Date Disposable Filter Changed: February 7, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 17, 2023
 Date RH/T Sensor Cleaned: October 17, 2023

Notes: No adjustments done. Leak check passed before and after cleaning.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25
WASKŌW OHCI PIMÂTISIWIN

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 16, 2024	Previous Calibration:	January 22, 2024
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	7:45	End Time (MST):	10:56
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999999	≥0.995
800.5	803.9	0.9958			
400.3	402.7	0.9941	Slope	1.004302	0.90 - 1.10
200.1	200.8	0.9967			
			Intercept	0.123899	+/-30



SO2 Calibration Plot

Date: February 16, 2024

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS25
 Calibration Date: February 14, 2024 Last Cal Date: January 18, 2024
 Start time (MST): 7:30 End time (MST): 12:00
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC517099
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 747
 ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
 Converter make: Global G-150 Converter serial #: 2022-219
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002531	1.005840	Backgd or Offset: 3.30	3.30
Calibration intercept:	0.300000	0.060000	Coeff or Slope: 1.125	1.113

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.0	79.5	80.9	0.982
as found 2nd point	4960	40.0	39.7	40.6	0.979
as found 3rd point	4980	20.0	19.9	20.3	0.979
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.0	79.5	80.0	0.993
second point	4960	40.0	39.7	40.0	0.993
third point	4980	20.0	19.9	20.0	0.993
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	80.0	800.0	805.2	0.994
SO2 Scrubber Check	4921	79.2	800.0	0.0	----
Date of last scrubber change:	20-Jun-23			Ave Corr Factor	0.993
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 80.9 Prev response: 79.96 *% change: 1.2%
 Baseline Corr 2nd AF pt: 40.6 AF Slope: 1.018066 AF Intercept: 0.060000
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999996

* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 14, 2024	Previous Calibration:	January 18, 2024
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	7:30	End Time (MST):	12:00
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999999	≥0.995
79.5	80.0	0.9932			
39.7	40.0	0.9932	Slope	1.005840	0.90 - 1.10
19.9	20.0	0.9932			
			Intercept	0.060000	+/-3



H₂S Calibration Plot

Date: February 14, 2024

Location: Waskow ohci Pimatisiwin





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS26
CHRISTINA LAKE**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name: Christina Lake Station number: AMS 26
Calibration Date: February 3, 2024 Last Cal Date: NA
Start time (MST): 13:02 End time (MST): 15:08
Reason: Install

Calibration Standards

Cal Gas Concentration: 49.56 ppm Cal Gas Exp Date: February 23, 2025
Cal Gas Cylinder #: CC362134
Removed Cal Gas Conc: 49.56 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 281
ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43i Analyzer serial #: 1152430005
Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	0.998578	Backgd or Offset:	NA	25.8
Calibration intercept:	NA	0.236006	Coeff or Slope:	NA	0.944

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4919	80.8	800.9	799.9	1.001
second point	4960	40.4	400.4	400.3	1.000
third point	4980	20.2	200.2	200.2	1.000
as left zero	5000	0.0	0.0	0.1	----
as left span	4919	80.8	800.9	799.8	1.001
Average Correction Factor					1.001

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Install calibrations. Sample inlet filters changed before calibrated zero. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 3, 2024	Previous Calibration:	NA
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	13:02	End Time (MST):	15:08
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430005

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	1.000000	
800.9	799.9	1.0013			≥0.995
400.4	400.3	1.0003	Slope	0.998578	
200.2	200.2	1.0001			0.90 - 1.10
			Intercept	0.236006	+/-30



SO2 Calibration Plot

Date: February 3, 2024

Location: Christina Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Christina Lake Station number: AMS26
 Calibration Date: February 7, 2024 Last Cal Date: January 17, 2024
 Start time (MST): 10:14 End time (MST): 14:21
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.05 ppm Cal Gas Exp Date: November 15, 2026
 Cal Gas Cylinder #: DT0014831
 Removed Cal Gas Conc: 5.05 ppm Rem Gas Exp Date:
 Removed Gas Cyl #: Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3253
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 450i Analyzer serial #: 1180030032
 Converter make: NA Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001140	1.008427	Backgd or Offset:	36.0 35.0
Calibration intercept:	0.238387	0.258384	Coeff or Slope:	1.115 1.086

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4921	79.2	80.0	82.4	0.974
as found 2nd point	4960	39.6	40.0	41.3	0.976
as found 3rd point	4980	19.8	20.0	21.0	0.966
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4921	79.2	80.0	80.9	0.989
second point	4960	39.6	40.0	40.7	0.983
third point	4980	19.8	20.0	20.3	0.985
as left zero	5000	0.0	0.0	0.4	----
as left span	4921	79.2	80.0	80.0	1.000
SO2 Scrubber Check	4931	80.9	807.1	0.0	----
Date of last scrubber change:	27-Feb-19			Ave Corr Factor	0.986
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	82.1	Prev response:	80.32	*% change:	2.2%
Baseline Corr 2nd AF pt:	41.0	AF Slope:	1.025572	AF Intercept:	0.358361
Baseline Corr 3rd AF pt:	20.7	AF Correlation:	0.999993		

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filters after multipoint as founds. SO2 scrubber check after calibrator zero done and passed. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 7, 2024	Previous Calibration:	January 17, 2024
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:14	End Time (MST):	14:21
Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999992	≥0.995
80.0	80.9	0.9887			
40.0	40.7	0.9828	Slope	1.008427	0.90 - 1.10
20.0	20.3	0.9852			
			Intercept	0.258384	+/-3



H₂S Calibration Plot

Date: February 7, 2024

Location: Christina Lake





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.9	-0.2	-0.7	----	----
as found span	4918	82.1	802.9	799.6	3.3	802.8	799.6	3.3	1.0002	1.0000
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.1	-0.6	----	----
high point	4918	82.1	802.9	799.6	3.3	802.7	798.7	4.1	1.0003	1.0012
second point	4959	41.1	401.9	400.3	1.6	402.5	399.8	2.7	0.9986	1.0013
third point	4980	20.5	200.5	199.7	0.8	201.8	200.6	1.1	0.9934	0.9953
as left zero	5000	0.0	0.0	0.0	0.0	-0.6	0.1	-0.7	----	----
as left span	4918	82.1	802.9	437.9	365.0	801.0	435.0	365.8	1.0024	1.0068
Average Correction Factor									0.9974	0.9992

Corrected As found	NO _x = 803.7 ppb	NO = 799.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.0%
Previous Response	NO _x = 803.4 ppb	NO = 801.2 ppb		*Percent Change	NO = -0.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.7	434.0	365.0	366.0	0.9972	100.3%
2nd GPT point (200 ppb O3)	795.7	613.5	185.5	187.2	0.9908	100.9%
3rd GPT point (100 ppb O3)	795.7	706.5	92.5	93.5	0.9891	101.1%
Average Correction Factor					0.9924	100.8%

Notes: Changed sample inlet filter after as founds. Adjusted span only. Used NO 2nd reference point because of drift.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 16, 2024
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:23	End Time (MST):	14:58
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.7	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
802.9	802.7	1.0003		
401.9	402.5	0.9986		
200.5	201.8	0.9934		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 16, 2024
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:23	End Time (MST):	14:58
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
799.6	798.7	1.0012			
400.3	399.8	1.0013			
199.7	200.6	0.9953			
			Slope	0.998413	0.90 - 1.10
			Intercept	0.406737	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 13, 2024	Previous Calibration:	January 16, 2024
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	10:23	End Time (MST):	14:58
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.6	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
365.0	366.0	0.9972		
185.5	187.2	0.9908		
92.5	93.5	0.9891		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 13, 2024

Location: Christina Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27
JACKFISH 2/3

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	February 13, 2024	Last Cal Date:	January 27, 2024
Start time (MST):	11:15	End time (MST):	14:32
Reason:	Removal		

Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	<u>SG9133974BAL</u>			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3811
ZAG Make/Model:	API 701		Serial Number:	268

Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169745
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001831		Backgd or Offset:	2.4	NA
Calibration intercept:	-0.817522		Coeff or Slope:	0.774	NA

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	79.1	800.2	797.5	1.003
as found 2nd point	4961	39.5	399.5	397.1	1.006
as found 3rd point	4980	19.8	200.3	194.6	1.029
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					

Average Correction Factor					
Baseline Corr As found:	797.50	Previous response	800.81	*% change	-0.4%
Baseline Corr 2nd AF pt:	397.10	AF Slope:	0.998674	AF Intercept:	-2.237394
Baseline Corr 3rd AF pt:	194.60	AF Correlation:	0.999955		

* = > +/-5% change initiates investigation


Notes: Removal calibration only.

Calibration Performed By: Mohammed Kashif

SO2 Calibration Plot

Date: February 13, 2024

Location: Jackfish 2/3





Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	February 14, 2024	Last Cal Date:	NA
Start time (MST):	11:45	End time (MST):	14:25
Reason:	Install		

Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	<u>SG9133974BAL</u>			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3811
ZAG Make/Model:	API 701		Serial Number:	268

Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	12124313138
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.006935	Backgd or Offset:	NA	8.4
Calibration intercept:	NA	-2.579017	Coeff or Slope:	NA	0.960

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	79.1	800.2	804.7	0.994
second point	4961	39.5	399.5	397.6	1.005
third point	4980	19.8	200.3	197.1	1.016
as left zero	5000	0.0	0.0	-0.1	----
as left span	4921	79.1	800.2	803.1	0.996
Average Correction Factor					1.005

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Installation calibration. Adjusted zero and span.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 14, 2024	Previous Calibration:	NA
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:45	End Time (MST):	14:25
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999950	≥0.995
800.2	804.7	0.9944			
399.5	397.6	1.0049	Slope	1.006935	0.90 - 1.10
200.3	197.1	1.0163			
			Intercept	-2.579017	+/-30



SO2 Calibration Plot

Date: February 14, 2024

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Jackfish 2/3 Station number: AMS27
 Calibration Date: February 22, 2024 Last Cal Date: January 12, 2024
 Start time (MST): 9:47 End time (MST): 14:32
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.41 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345023
 Removed Cal Gas Conc: 5.41 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3811
 ZAG Make/Model: API 701 Serial Number: 268

Analyzer Information

Analyzer make: API T101 Analyzer serial #: 621
 Converter make: Converter serial #:
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010462	1.013890	Backgd or Offset: 29.9	29.9
Calibration intercept:	0.002360	-0.217825	Coeff or Slope: 1.013	0.965

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4926	74.1	80.2	86.5	0.928
as found 2nd point	4963	37.0	40.0	43.1	0.931
as found 3rd point	4982	18.5	20.0	20.8	0.967
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4926	74.1	80.2	81.3	0.986
second point	4963	37.0	40.0	40.1	0.998
third point	4982	18.5	20.0	19.7	1.016
as left zero	5000	0.0	0.0	0.2	----
as left span	4926	74.1	80.2	81.3	0.986
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.000
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 86.4 Prev response: 81.02 *% change: 6.2%
 Baseline Corr 2nd AF pt: 43.0 AF Slope: 1.081165 AF Intercept: -0.276196
 Baseline Corr 3rd AF pt: 20.7 AF Correlation: 0.999884

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 22, 2024	Previous Calibration:	January 12, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	9:47	End Time (MST):	14:32
Analyzer make:	API T101	Analyzer serial #:	621

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999878	
80.2	81.3	0.9862			≥0.995
40.0	40.1	0.9984	Slope	1.013890	
20.0	19.7	1.0160			0.90 - 1.10
			Intercept	-0.217825	+/-3



H₂S Calibration Plot

Date: February 22, 2024

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as found span	4921	79.4	816.8	800.3	16.5	794.2	777.7	16.5	1.0285	1.0290
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.4	-0.4	----	----
high point	4921	79.4	816.8	800.3	16.5	819.0	802.6	16.4	0.9973	0.9971
second point	4960	39.7	408.5	400.2	8.3	407.7	397.5	10.2	1.0019	1.0068
third point	4980	19.8	203.7	199.6	4.1	195.4	192.8	2.6	1.0425	1.0352
as left zero	5000	0.0	0.0	0.0	0.0	2.5	0.7	1.9	----	----
as left span	4921	79.4	816.8	414.8	404.9	818.3	416.3	402.0	0.9982	0.9964
Average Correction Factor									1.0139	1.0130

Corrected As found	NO _x = 794.1 ppb	NO = 777.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.6%
Previous Response	NO _x = 807.1 ppb	NO = 795.7 ppb		*Percent Change	NO = -2.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	800.9	412.5	404.9	404.1	1.0020	99.8%
2nd GPT point (200 ppb O3)	800.9	625.4	192.0	192.6	0.9970	100.3%
3rd GPT point (100 ppb O3)	800.9	718.1	99.3	99.6	0.9971	100.3%
Average Correction Factor					0.9987	100.1%

Notes:

Changed sample inlet filter after as founds. Adjusted the span only.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 22, 2024	Previous Calibration:	January 11, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	14:32	End Time (MST):	19:34
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
816.8	819.0	0.9973		
408.5	407.7	1.0019		
203.7	195.4	1.0425		
			0.999868	
			1.006157	
			-3.917080	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 22, 2024	Previous Calibration:	January 11, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	14:32	End Time (MST):	19:34
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.4	----	Correlation Coefficient	≥0.995	
800.3	802.6	0.9971			
400.2	397.5	1.0068			
199.6	192.8	1.0352			
			Slope	1.005101	0.90 - 1.10
			Intercept	-3.480285	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 22, 2024	Previous Calibration:	January 11, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	14:32	End Time (MST):	19:34
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
404.9	404.1	1.0020		
192.0	192.6	0.9970		
99.3	99.6	0.9971		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 22, 2024

Location: Jackfish 2/3





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS29
SURMONT 2**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 5, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	11:39	End Time (MST):	15:12
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999979	≥0.995
800.1	798.0	1.0026			
400.6	397.7	1.0073	Slope	0.998273	0.90 - 1.10
199.8	196.1	1.0190			
			Intercept	-1.525570	+/-30



SO2 Calibration Plot

Date: February 5, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 12, 2024	Previous Calibration:	January 3, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:53	End Time (MST):	15:22
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999984	≥0.995
80.0	80.3	0.9963			
40.1	40.0	1.0027	Slope	1.004326	0.90 - 1.10
20.1	19.9	1.0077			
			Intercept	-0.142791	+/-3



H₂S Calibration Plot

Date: February 12, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	February 5, 2024	Last Cal Date:	January 5, 2024
Start time (MST):	11:39	End time (MST):	15:12
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC356008	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
C3H8 Cal Gas Conc.	<u>205.7</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
Removed C3H8 Conc.	<u>205.7</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5472
ZAG Make/Model:	Teledyne API T701	Serial Number:	4297

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1170050149
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001735	1.003467	Background:	3.62	3.62
Calibration intercept:	-0.005651	-0.070017	Coefficient:	4.021	4.021

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.03	----
as found span	4918	81.3	17.31	17.32	1.000
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	-0.07	----
high point	4918	81.3	17.31	17.30	1.001
second point	4959	40.6	8.65	8.62	1.003
third point	4979	20.3	4.32	4.26	1.016
as left zero	5000	0.0	0.00	-0.08	----
as left span	4918	81.3	17.31	17.37	0.997
Average Correction Factor					1.007
Baseline Corr As found:	17.35	Previous response	17.34	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 5, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	11:39	End Time (MST):	15:12
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	-0.07	----	Correlation Coefficient	0.999998	
17.31	17.30	1.0008			≥0.995
8.65	8.62	1.0030	Slope	1.003467	
4.32	4.26	1.0158			0.90 - 1.10
			Intercept	-0.070017	+/-1.5



THC Calibration Plot

Date: February 5, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 26, 2024	Previous Calibration:	February 5, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	11:48	End Time (MST):	15:14
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.11	----	Correlation Coefficient	0.999978	≥0.995
17.31	17.12	1.0113			
8.65	8.42	1.0268	Slope	0.995247	0.90 - 1.10
4.32	4.17	1.0380			
			Intercept	-0.136283	+/-1.5



THC Calibration Plot

Date: February 26, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as found span	4916	84.2	799.2	799.2	0.0	797.9	796.8	1.0	1.0016	1.0030
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4916	84.2	799.2	799.2	0.0	798.3	796.8	1.5	1.0011	1.0030
second point	4958	42.1	399.6	399.6	0.0	398.1	397.7	0.4	1.0038	1.0048
third point	4979	21.1	200.3	200.3	0.0	197.8	196.4	1.4	1.0125	1.0197
as left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
as left span	4916	84.2	799.2	412.6	386.6	793.5	406.8	386.7	1.0072	1.0142
Average Correction Factor									1.0058	1.0092

Corrected As found	NO _x = 798.0 ppb	NO = 796.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.0%
Previous Response	NO _x = 797.6 ppb	NO = 796.4 ppb		*Percent Change	NO = 0.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	793.7	407.1	386.6	385.9	1.0018	99.8%
2nd GPT point (200 ppb O3)	793.7	602.1	191.6	191.5	1.0005	99.9%
3rd GPT point (100 ppb O3)	793.7	695.2	98.5	97.8	1.0072	99.3%
Average Correction Factor					1.0032	99.7%

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 4, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:39	End Time (MST):	15:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
799.2	798.3	1.0011			
399.6	398.1	1.0038			
200.3	197.8	1.0125			
			Slope	0.999594	0.90 - 1.10
			Intercept	-1.052309	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 4, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:39	End Time (MST):	15:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.2	796.8	1.0030		
399.6	397.7	1.0048		
200.3	196.4	1.0197		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 4, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:39	End Time (MST):	15:21
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
386.6	385.9	1.0018		
191.6	191.5	1.0005		
98.5	97.8	1.0072		
			0.999996	
			0.998502	
			-0.096658	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 6, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Surmont Station number: AMS 29
 Calibration Date: February 12, 2024 Last Cal Date: January 4, 2024
 Start time (MST): 11:53 End time (MST): 13:43

Analyzer Make: API T640 S/N: 253
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	0.7	0.41	0.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	705.0	705.69	705.0	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	5.092	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	36	----	36	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	1.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.0	11.0	11.0	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 12, 2024
 Date Disposable Filter Changed: February 12, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 25, 2023
 Date RH/T Sensor Cleaned: October 25, 2023

Notes: No adjustments made. Quarterly maintenance completed. Leak checks passed.

Calibration by: Braiden Boutilier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS30
ELLS RIVER**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 3, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	11:19	End Time (MST):	14:21
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999961	≥0.995
800.4	800.0	1.0005			
400.2	397.4	1.0071	Slope	1.000960	0.90 - 1.10
200.1	195.6	1.0230			
			Intercept	-2.235937	+/-30



SO2 Calibration Plot

Date: February 5, 2024

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Ells River Station number: AMS30
 Calibration Date: February 6, 2024 Last Cal Date: January 18, 2024
 Start time (MST): 10:02 End time (MST): 14:52
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.99 ppm Cal Gas Exp Date: November 15, 2026
 Cal Gas Cylinder #: CC505806
 Removed Cal Gas Conc: 4.99 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3061
 ZAG Make/Model: API T701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1410661331
 Converter make: CDN - 101 Converter serial #: 562
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005472	1.002473	Backgd or Offset:	2.05 1.64
Calibration intercept:	-0.120388	0.159591	Coeff or Slope:	1.072 1.060

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.5	----
as found span	4920	80.2	80.0	80.1	0.993
as found 2nd point	4960	40.1	40.0	39.9	0.991
as found 3rd point	4980	20.0	20.0	19.1	1.018
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.2	80.0	80.3	0.997
second point	4960	40.1	40.0	40.4	0.991
third point	4980	20.0	20.0	20.3	0.983
as left zero	5000	0.0	0.0	0.8	----
as left span	4920	80.2	80.0	80.0	1.000
SO2 Scrubber Check	4921	79.2	800.4	0.1	----
Date of last scrubber change:	N/A		Ave Corr Factor		0.990
Date of last converter efficiency test:	N/A		efficiency		

Baseline Corr As found: 80.6 Prev response: 80.35 *% change: 0.3%
 Baseline Corr 2nd AF pt: 40.4 AF Slope: 1.009332 AF Intercept: -0.680519
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999944

* = > +/-5% change initiates investigation

Notes: Change inlet filters after multipoint as founds. SO2 scrubber check done after calibrator zero, passed. Adjusted zero and span.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 6, 2024	Previous Calibration:	January 18, 2024
Station Name:	Ells River	Station Number:	AMS30
Start Time (MST):	10:02	End Time (MST):	14:52
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999981	≥0.995
80.0	80.3	0.9967			
40.0	40.4	0.9906	Slope	1.002473	0.90 - 1.10
20.0	20.3	0.9833			
			Intercept	0.159591	+/-3



TRS Calibration Plot

Date: February 6, 2024

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	February 5, 2024	Last Cal Date:	January 3, 2024
Start time (MST):	11:19	End time (MST):	14:21
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm		
Diff between cyl (CH ₄):		Diff between cyl (THC):	
Calibrator Model:	API T700	Diff between cyl (NM):	
ZAG make/model:	API T701H	Serial Number:	3061
		Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.51E-04	2.53E-04	NMHC SP Ratio:	6.00E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	151854
				141799

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.56	1.028
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.03	1.000
second point	4960	39.6	8.51	8.47	1.005
third point	4980	19.8	4.26	4.15	1.026
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.24	0.988

				Average Correction Factor	1.010
Baseline Corr AF:	16.56	Prev response	17.05	*% change	-3.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	8.68	1.050
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.15	0.996
second point	4960	39.6	4.56	4.55	1.002
third point	4980	19.8	2.28	2.22	1.027
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.25	0.985
Average Correction Factor					1.009
Baseline Corr AF:	8.68	Prev response	9.17	*% change	-5.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.88	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.88	1.004
second point	4960	39.6	3.96	3.92	1.009
third point	4980	19.8	1.98	1.93	1.024
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.99	0.991
Average Correction Factor					1.012
Baseline Corr AF:	7.88	Prev response	7.88	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.003892	1.002141
THC Cal Offset:	-0.039540	-0.052740
CH ₄ Cal Slope:	0.999470	0.997247
CH ₄ Cal Offset:	-0.026357	-0.019157
NMHC Cal Slope:	1.007395	1.006266
NMHC Cal Offset:	-0.012583	-0.033582

Notes: Inlet filter changed after As Finds, adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 3, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	11:19	End Time (MST):	14:21
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999954	≥ 0.995			
17.03	17.03	0.9997						
8.51	8.47	1.0053				Slope	1.002141	0.90 - 1.10
4.26	4.15	1.0258						
			Intercept	-0.052740	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 3, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	11:19	End Time (MST):	14:21
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999972	≥0.995
7.91	7.88	1.0042			
3.96	3.92	1.0089			
1.98	1.93	1.0243			
			Slope	0.997247	0.90 - 1.10
			Intercept	-0.019157	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 5, 2024	Previous Calibration:	January 3, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	11:19	End Time (MST):	14:21
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999935	≥ 0.995			
9.11	9.15	0.9960						
4.56	4.55	1.0024				Slope	1.006266	0.90 - 1.10
2.28	2.22	1.0272						
			Intercept	-0.033582	± 0.5			



NMHC Calibration Plot

Date: February 5, 2024

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	February 27, 2024	Last Cal Date:	January 3, 2024
Start time (MST):	9:36	End time (MST):	11:07
Reason:	Cylinder Change N2 cylinder change		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH4 Cal Gas Conc.	499.7 ppm	CH4 Equiv Conc.	1075.0 ppm
C3H8 Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	499.7 ppm	CH4 Equiv Conc.	1075.0 ppm
Removed C3H8 Conc.	209.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
ZAG make/model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.51E-04	2.53E-04	NMHC SP Ratio:	6.00E-05
CH4 Retention time:	14.2	14.2	NMHC Peak Area:	151854
				141799

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	17.55	0.970
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	17.41	0.978
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	0.978
Baseline Corr AF:	17.55	Prev response	17.05	*% change	2.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	9.52	0.957
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.47	0.962
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.962
Baseline Corr AF:	9.52	Prev response	9.17	*% change	3.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	8.03	0.986
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.94	0.997
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.997
Baseline Corr AF:	8.03	Prev response	7.88	*% change	1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.003892	1.022298
THC Cal Offset:	-0.039540	0.000000
CH ₄ Cal Slope:	0.999470	1.002915
CH ₄ Cal Offset:	-0.026357	0.000000
NMHC Cal Slope:	1.007395	1.039135
NMHC Cal Offset:	-0.012583	0.000000


Notes: N2 cylinder changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro

NMHC Calibration Plot

Date: February 27, 2024

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Ells River
Calibration Date: February 4, 2024
Start time (MST): 11:20
Reason: Cylinder Change
Station number: AMS 30
Last Cal Date: January 11, 2024
End time (MST): 12:58

Calibration Standards

NO Gas Cylinder #: A0931384
NOX Cal Gas Conc: 51.12 ppm
Removed Cylinder #: DT0045703
Removed Gas NOX Conc: 60.4 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: November 30, 2023
NO Cal Gas Conc: 49.40 ppm
Removed Gas Exp Date: May 16, 2031
Removed Gas NO Conc: 60.1 ppm
NO gas Diff:
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 710321429

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.165	1.165	NO bkgnd or offset:	13.6	13.7
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	13.5	13.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	194.2	192.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.007658	
NO _x Cal Offset:	-1.638433	
NO Cal Slope:	1.007289	
NO Cal Offset:	-2.359365	
NO ₂ Cal Slope:	0.990195	
NO ₂ Cal Offset:	1.994477	



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp	4919	81.0	828.1	800.3	27.9	841.0	813.0	28.5	0.9847	0.9844
calibrator zero										
high point										
second point										
third point										
as left zero										
as left span										

Average Correction Factor									
Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* => +/-5% change initiates investigation</i>			*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb				*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:	
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)						
2nd GPT point (200 ppb O3)						
3rd GPT point (100 ppb O3)						
Average Correction Factor						


Notes: Cylinder#: DT0045703 empty due to leak. Replaced with cylinder#: A0931384. New cylinder response captured.

Calibration Performed By: Braiden Boutilier

NO_x Calibration Plot

Date: February 4, 2024

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Ells River
Calibration Date: February 23, 2024
Start time (MST): 10:00
Reason: Routine
Station number: AMS 30
Last Cal Date: January 11, 2024
End time (MST): 15:28

Calibration Standards

NO Gas Cylinder #: DT0027487
NOX Cal Gas Conc: 59.3 ppm
Removed Cylinder #: A0931384
Removed Gas NOX Conc: 51.12 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.1 ppm
Removed Gas Exp Date: November 30, 2023
Removed Gas NO Conc: 49.40 ppm
NO gas Diff:
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 710321429

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.165	1.182	NO bkgnd or offset:	13.6	14.0
NOX coeff or slope:	0.992	0.994	NOX bkgnd or offset:	13.5	13.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	194.2	188.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.007658	0.998634
NO _x Cal Offset:	-1.638433	-0.718485
NO Cal Slope:	1.007289	0.999588
NO Cal Offset:	-2.359365	-1.719843
NO ₂ Cal Slope:	0.990195	1.005754
NO ₂ Cal Offset:	1.994477	0.149457



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as found span	4932	67.7	803.0	800.3	2.7	826.2	788.6	37.6	0.9719	1.0148
as found 2nd										
as found 3rd										
new cyl resp	4932	67.7	803.0	800.3	2.7	792.0	790.1	1.9	1.0139	1.0129
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.1	----	----
high point	4932	67.7	803.0	800.3	2.7	801.3	799.0	2.4	1.0021	1.0016
second point	4966	33.8	400.9	399.5	1.4	400.0	397.1	2.9	1.0022	1.0061
third point	4983	16.9	200.4	199.8	0.7	198.1	195.9	2.2	1.0118	1.0197
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
as left span	4932	67.7	803.0	425.4	377.6	804.0	426.5	377.5	0.9987	0.9973
Average Correction Factor									1.0054	1.0091

Corrected As found	NO _x = 826.3 ppb	NO = 788.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 2.3%
Previous Response	NO _x = 807.5 ppb	NO = 803.7 ppb		*Percent Change	NO = -1.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	795.8	420.9	377.6	379.7	0.9945	100.6%
2nd GPT point (200 ppb O3)	795.8	615.3	183.2	185.1	0.9898	101.0%
3rd GPT point (100 ppb O3)	795.8	702.8	95.7	96.0	0.9970	100.3%
Average Correction Factor					0.9937	100.6%

Notes:

Changed sample inlet filters and NO_x cylinder after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 23, 2024	Previous Calibration:	January 11, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:00	End Time (MST):	15:28
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
803.0	801.3	1.0021			
400.9	400.0	1.0022			
200.4	198.1	1.0118			
			Slope	0.998634	0.90 - 1.10
			Intercept	-0.718485	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 23, 2024	Previous Calibration:	January 11, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:00	End Time (MST):	15:28
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.3	799.0	1.0016			
399.5	397.1	1.0061			
199.8	195.9	1.0197			
			Slope	0.999588	0.90 - 1.10
			Intercept	-1.719843	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 23, 2024	Previous Calibration:	January 11, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:00	End Time (MST):	15:28
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>			
0.0	0.1	----	Correlation Coefficient	0.999991	≥0.995			
377.6	379.7	0.9945						
183.2	185.1	0.9898				Slope	1.005754	0.90 - 1.10
95.7	96.0	0.9970						
			Intercept	0.149457	+/-20			



NO_x Calibration Plot

Date: February 23, 2024

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Ells River Station number: AMS 30
 Calibration Date: February 23, 2024 Last Cal Date: January 22, 2024
 Start time (MST): 12:27 End time (MST): 12:41

Analyzer Make: API T640 S/N: 875
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	6.90	6.49	6.90	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	718.00	719.67	718.00	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.98	5.05	4.98	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	45.00	----	45.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.50	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.90** Expiry Date: September 29, 2024
 Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: January 22, 2024
 Date Disposable Filter Changed: January 22, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 27, 2023
 Date RH/T Sensor Cleaned: February 23, 2024

Notes: Verified flow, temperature and pressure. Leak check passed. No adjustment made.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS501
LEISMER**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 8, 2024	Previous Calibration:	January 16, 2024
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	10:02	End Time (MST):	12:54
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	≥0.995
800.2	801.5	0.9984		
400.2	401.4	0.9969	Slope	0.90 - 1.10
200.1	197.4	1.0135		
			Intercept	+/-30



SO2 Calibration Plot

Date: February 8, 2024

Location: Leismer





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Leismer	Station number:	AMS501
Calibration Date:	February 7, 2024	Last Cal Date:	January 23, 2024
Start time (MST):	9:43	End time (MST):	13:51
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.14	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511843			
Removed Cal Gas Conc:	5.14	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2659
ZAG Make/Model:	API 701		Serial Number:	4427

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020
Converter make:	Global G150	Converter serial #:	2022-218
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999147	1.011575	Backgd or Offset:	3.57
Calibration intercept:	0.001539	-0.158186	Coeff or Slope:	1.128

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4922	77.8	80.0	80.4	0.994
as found 2nd point	4961	38.9	40.0	40.0	0.997
as found 3rd point	4981	19.4	19.9	19.8	1.002
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	77.8	80.0	80.8	0.990
second point	4961	38.9	40.0	40.3	0.992
third point	4981	19.4	19.9	19.8	1.007
as left zero	5000	0.0	0.0	-0.1	----
as left span	4922	77.8	80.0	81.2	0.985
SO2 Scrubber Check	4921	79.2	792.0	0.0	----
Date of last scrubber change:	24-Feb-23			Ave Corr Factor	0.996
Date of last converter efficiency test:	December 1, 2022			efficiency	

Baseline Corr As found:	80.5	Prev response:	79.91	*% change:	0.7%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	1.007002	AF Intercept:	-0.198215
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999993		

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Scrubber test done after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 7, 2024	Previous Calibration:	January 23, 2024
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:43	End Time (MST):	13:51
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999979	
80.0	80.8	0.9899			≥0.995
40.0	40.3	0.9923	Slope	1.011575	
19.9	19.8	1.0072			0.90 - 1.10
			Intercept	-0.158186	+/-3



H₂S Calibration Plot

Date: February 7, 2024

Location: Leismer





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Leismer
Calibration Date: February 21, 2024
Start time (MST): 9:44
Reason: Routine
Station number: AMS501
Last Cal Date: January 24, 2024
End time (MST): 13:55

Calibration Standards

NO Gas Cylinder #: T26811M
NOX Cal Gas Conc: 47.46 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.46 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API 701
Cal Gas Expiry Date: October 30, 2024
NO Cal Gas Conc: 47.39 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.39 ppm
NO gas Diff:
Serial Number: 2659
Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1218153356

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.186	1.186	NO bkgnd or offset:	3.5	3.5
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	3.5	3.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	166.2	166.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001018	1.006425
NO _x Cal Offset:	-1.427986	-3.447999
NO Cal Slope:	1.002011	1.006912
NO Cal Offset:	-2.907962	-4.347977
NO ₂ Cal Slope:	1.000174	0.999948
NO ₂ Cal Offset:	0.184957	0.272718



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.2	0.1	----	----
as found span	4916	84.4	801.1	799.9	1.2	805.1	802.0	3.2	0.9950	0.9974
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4916	84.4	801.1	799.9	1.2	804.2	802.8	1.3	0.9961	0.9964
second point	4958	42.2	400.5	400.0	0.6	398.9	397.5	1.5	1.0041	1.0062
third point	4979	21.1	200.3	200.0	0.3	193.9	191.8	2.1	1.0329	1.0427
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4916	84.4	801.1	404.9	396.2	796.7	401.9	394.8	1.0055	1.0074
Average Correction Factor									1.0110	1.0151

Corrected As found	NO _x = 805.1 ppb	NO = 802.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.6%
Previous Response	NO _x = 800.4 ppb	NO = 798.6 ppb		*Percent Change	NO = 0.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.3	401.3	396.2	396.3	0.9997	100.0%
2nd GPT point (200 ppb O3)	796.3	609.0	188.5	189.0	0.9973	100.3%
3rd GPT point (100 ppb O3)	796.3	704.1	93.4	93.7	0.9966	100.3%
Average Correction Factor					0.9979	100.2%

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 24, 2024
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:44	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
801.1	804.2	0.9961			
400.5	398.9	1.0041			
200.3	193.9	1.0329			
			Slope	1.006425	0.90 - 1.10
			Intercept	-3.447999	+/-20

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 24, 2024
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:44	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
799.9	802.8	0.9964			
400.0	397.5	1.0062			
200.0	191.8	1.0427			
			Slope	1.006912	0.90 - 1.10
			Intercept	-4.347977	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 24, 2024
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	9:44	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
396.2	396.3	0.9997			
188.5	189.0	0.9973			
93.4	93.7	0.9966			
			Slope	0.999948	0.90 - 1.10
			Intercept	0.272718	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 21, 2024

Location: Leismer





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS505
SAWBONES BAY

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 11, 2024
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	9:24	End Time (MST):	12:20
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321323

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient	≥0.995
799.8	802.1	0.9972		
399.9	400.7	0.9980	Slope	0.90 - 1.10
200.4	198.0	1.0123		
			Intercept	+/-30



SO2 Calibration Plot

Date: February 15, 2024

Location: Sawbones Bay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Sawbones Bay Station number: AMS505
 Calibration Date: February 22, 2024 Last Cal Date: January 8, 2024
 Start time (MST): 9:34 End time (MST): 13:23
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517397
 Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 5112
 ZAG Make/Model: Teledyne API T701 Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43iQ Analyzer serial #: 12113311965
 Converter make: Global 150 Converter serial #: 2022-224
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006361	1.014500	Backgd or Offset:	1.01
Calibration intercept:	-0.218059	-0.077888	Coeff or Slope:	1.119

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4922	77.7	80.0	81.5	0.981
as found 2nd point	4961	38.8	40.0	40.5	0.984
as found 3rd point	4981	19.4	20.0	20.0	0.994
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	77.7	80.0	81.2	0.986
second point	4961	38.8	40.0	40.3	0.992
third point	4981	19.4	20.0	20.2	0.989
as left zero	5000	0.0	0.0	0.1	----
as left span	4922	77.7	80.0	80.8	0.991
SO2 Scrubber Check	4922	77.8	778.0	-0.1	----

Date of last scrubber change:	Ave Corr Factor	0.989
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 81.6 Prev response: 80.33 *% change: 1.6%
 Baseline Corr 2nd AF pt: 40.6 AF Slope: 1.020497 AF Intercept: -0.237739
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999987

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 22, 2024	Previous Calibration:	January 8, 2024
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	9:34	End Time (MST):	13:23
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12113311965

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999989	
80.0	81.2	0.9857			≥0.995
40.0	40.3	0.9917	Slope	1.014500	
20.0	20.2	0.9891			0.90 - 1.10
			Intercept	-0.077888	+/-3



H₂S Calibration Plot

Date: February 22, 2024

Location: Sawbones Bay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.2	-0.2	----	----
as found span	4917	83.4	799.6	799.6	0.0	805.6	801.6	3.8	0.9925	0.9975
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.1	-0.4	----	----
high point	4917	83.4	799.6	799.6	0.0	799.7	799.9	-0.3	0.9998	0.9996
second point	4958	41.7	399.8	399.8	0.0	398.4	398.2	0.3	1.0036	1.0041
third point	4979	20.9	200.4	200.4	0.0	197.6	196.7	0.9	1.0141	1.0188
as left zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
as left span	4916	83.4	799.7	333.8	465.9	795.1	332.3	462.8	1.0058	1.0046
Average Correction Factor									1.0059	1.0075

Corrected As found	NO _x = 806.0 ppb	NO = 801.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.9%
Previous Response	NO _x = 798.8 ppb	NO = 798.1 ppb		*Percent Change	NO = 0.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.1	330.2	465.9	465.8	1.0002	100.0%
2nd GPT point (200 ppb O3)	796.1	545.1	251.0	250.9	1.0004	100.0%
3rd GPT point (100 ppb O3)	796.1	646.5	149.6	149.3	1.0020	99.8%
Average Correction Factor					1.0009	99.9%

Notes:

Changed inlet filter after as founds. Adjusted span.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 14, 2024	Previous Calibration:	January 10, 2024
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	9:30	End Time (MST):	13:42
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.5	----	Correlation Coefficient	≥0.995	
799.6	799.7	0.9998			
399.8	398.4	1.0036			
200.4	197.6	1.0141			
			Slope	1.001650	0.90 - 1.10
			Intercept	-1.730586	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 14, 2024	Previous Calibration:	January 10, 2024
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	9:30	End Time (MST):	13:42
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
799.6	799.9	0.9996			
399.8	398.2	1.0041			
200.4	196.7	1.0188			
			Slope	1.001864	0.90 - 1.10
			Intercept	-1.930254	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 14, 2024	Previous Calibration:	January 10, 2024
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	9:30	End Time (MST):	13:42
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.4	----	Correlation Coefficient	≥0.995	
465.9	465.8	1.0002			
251.0	250.9	1.0004			
149.6	149.3	1.0020			
			Slope	1.000683	0.90 - 1.10
			Intercept	-0.373004	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 14, 2024

Location: Sawbones Bay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS507
KIRBY SOUTH**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Kirby South	Station number:	AMS 507
Calibration Date:	February 14, 2024	Last Cal Date:	January 17, 2024
Start time (MST):	12:43	End time (MST):	15:54
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.18	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC303554</u>			
Removed Cal Gas Conc:	49.18	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3804
ZAG Make/Model:	API T701H		Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005405	0.997947	Backgd or Offset:	24.6	23.9
Calibration intercept:	-1.048100	-0.888482	Coeff or Slope:	1.100	1.072

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4919	81.3	799.6	816.0	0.980
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4919	81.3	799.6	798.0	1.002
second point	4959	40.7	400.3	396.7	1.009
third point	4980	20.3	199.7	198.6	1.005
as left zero	5000	0.0	0.0	0.1	----
as left span	4919	81.3	799.6	798.7	1.001
Average Correction Factor					1.006

Baseline Corr As found:	816.30	Previous response	802.89	*% change	1.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 14, 2024	Previous Calibration:	January 17, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	12:43	End Time (MST):	15:54
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	≥0.995
799.6	798.0	1.0020		
400.3	396.7	1.0092	Slope	0.90 - 1.10
199.7	198.6	1.0053		
			Intercept	+/-30



SO2 Calibration Plot

Date: February 14, 2024

Location: Kirby South





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	February 14, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	13:00	End Time (MST):	16:05
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999971	
799.6	799.0	1.0008			≥0.995
400.3	396.3	1.0102	Slope	0.999677	
199.7	197.2	1.0125			0.90 - 1.10
			Intercept	-1.668567	+/-30



SO2 Calibration Plot

Date: February 15, 2024

Location: Kirby South





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Kirby South Station number: AMS507
 Calibration Date: February 15, 2024 Last Cal Date: January 17, 2024
 Start time (MST): 7:35 End time (MST): 13:26
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.05 ppm Cal Gas Exp Date: November 15, 2026
 Cal Gas Cylinder #: DT0019762
 Removed Cal Gas Conc: 5.05 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: DT0019762 Diff between cyl:
 Calibrator Make/Model: API T750 Serial Number: 281
 ZAG Make/Model: API T751H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1150840012
 Converter make: Global Converter serial #: 2022-197
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997813	0.998813	Backgd or Offset:	1.74	1.74
Calibration intercept:	0.179050	0.219035	Coeff or Slope:	1.052	1.052

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4923	77.4	78.2	79.6	0.982
as found 2nd point	4961	38.8	39.2	39.9	0.982
as found 3rd point	4981	19.3	19.5	20.0	0.975
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	79.2	80.0	80.0	1.000
second point	4960	39.6	40.0	40.3	0.993
third point	4980	19.8	20.0	20.4	0.980
as left zero	5000	0.0	0.0	0.2	----
as left span	4921	79.2	80.0	80.2	0.997
SO2 Scrubber Check	4919	80.0	800.2	0.1	----
Date of last scrubber change:	25-Jul-23			Ave Corr Factor	0.991
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 79.6 Prev response: 78.18 *% change: 1.8%
 Baseline Corr 2nd AF pt: 39.9 AF Slope: 1.017647 AF Intercept: 0.058982
 Baseline Corr 3rd AF pt: 20.0 AF Correlation: 0.999995

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Ran SOx scrubber check after cal zero. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 17, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	7:35	End Time (MST):	13:26
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999965	≥0.995
80.0	80.0	0.9999			
40.0	40.3	0.9925	Slope	0.998813	0.90 - 1.10
20.0	20.4	0.9803			
			Intercept	0.219035	+/-3



H₂S Calibration Plot

Date: February 15, 2024

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 14, 2024	Previous Calibration:	January 16, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	12:43	End Time (MST):	15:54
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	-0.04	----	Correlation Coefficient	0.999969	
17.26	17.26	1.0002			≥0.995
8.64	8.57	1.0085	Slope	1.003790	
4.31	4.19	1.0287			0.90 - 1.10
			Intercept	-0.088611	+/-1.5



THC Calibration Plot

Date: February 14, 2024

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 16, 2024	Previous Calibration:	February 14, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	12:48	End Time (MST):	16:17
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.01	----	Correlation Coefficient	0.999831	≥0.995
17.26	17.24	1.0013			
8.64	8.47	1.0204	Slope	1.002093	0.90 - 1.10
4.31	4.11	1.0482			
			Intercept	-0.117790	+/-1.5



THC Calibration Plot

Date: February 16, 2024

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 22, 2024	Previous Calibration:	February 16, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	12:24	End Time (MST):	15:33
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.02	----	Correlation Coefficient	0.999999	≥0.995
17.26	17.26	1.0002			
8.64	8.67	0.9969	Slope	0.998323	0.90 - 1.10
4.31	4.34	0.9932			
			Intercept	0.032189	+/-1.5



THC Calibration Plot

Date: February 22, 2024

Location: Kirby South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	February 15, 2024	Last Cal Date:	January 18, 2024
Start time (MST):	7:45	End time (MST):	12:50
Reason:	Routine		

Calibration Standards

NO Gas Cylinder #:	T34ULGL	Cal Gas Expiry Date:	March 8, 2025
NOX Cal Gas Conc:	49.39 ppm	NO Cal Gas Conc:	49.02 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	49.39 ppm	Removed Gas NO Conc:	49.02 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	3804
ZAG make/model:	API 701H	Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>	
NO coeff or slope:	1.397	1.383	NO bkgnd or offset:	1.7	1.7
NOX coeff or slope:	0.995	0.998	NOX bkgnd or offset:	1.8	1.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	209.21	210.24

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003862	1.009303
NO _x Cal Offset:	-4.492458	-4.671552
NO Cal Slope:	1.005411	1.014663
NO Cal Offset:	-5.173949	-5.612277
NO ₂ Cal Slope:	0.997888	1.003572
NO ₂ Cal Offset:	0.693913	1.309741



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.6	0.2	----	----
as found span	4919	81.0	800.1	794.1	6.0	806.0	802.0	4.1	0.9927	0.9902
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.4	0.0	----	----
high point	4919	81.0	800.1	794.1	6.0	805.0	803.0	2.1	0.9939	0.9889
second point	4960	40.5	400.0	397.0	3.0	397.0	393.9	3.1	1.0076	1.0079
third point	4980	20.2	199.5	198.0	1.5	192.4	190.6	1.8	1.0370	1.0390
as left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
as left span	4919	81.0	800.1	399.9	400.2	808.0	402.3	406.0	0.9902	0.9941
Average Correction Factor									1.0129	1.0120

Corrected As found	NO _x = 806.4 ppb	NO = 802.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.0%
Previous Response	NO _x = 798.7 ppb	NO = 793.2 ppb		*Percent Change	NO = 1.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.0	404.8	400.2	402.2	0.9950	100.5%
2nd GPT point (200 ppb O3)	799.0	625.2	179.8	183.0	0.9825	101.8%
3rd GPT point (100 ppb O3)	799.0	718.9	86.1	88.5	0.9728	102.8%
Average Correction Factor					0.9834	101.7%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 18, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	7:45	End Time (MST):	12:50
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.1	805.0	0.9939		
400.0	397.0	1.0076		
199.5	192.4	1.0370		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 18, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	7:45	End Time (MST):	12:50
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
794.1	803.0	0.9889		
397.0	393.9	1.0079		
198.0	190.6	1.0390		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 15, 2024	Previous Calibration:	January 18, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	7:45	End Time (MST):	12:50
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
400.2	402.2	0.9950			
179.8	183.0	0.9825			
86.1	88.5	0.9728			
			Slope	1.003572	0.90 - 1.10
			Intercept	1.309741	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 15, 2024

Location: Kirby South





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS511
BLACKGOLD**

FEBRUARY 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

March 28, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Blackgold	Station number:	AMS511
Calibration Date:	February 21, 2024	Last Cal Date:	January 17, 2024
Start time (MST):	10:50	End time (MST):	14:30
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.06	ppm	Cal Gas Exp Date:	January 5, 2029
Cal Gas Cylinder #:	CC147416			
Removed Cal Gas Conc:	50.06	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	5258
ZAG Make/Model:	Teledyne API 701		Serial Number:	138

Analyzer Information

Analyzer make:	Thermo scientific	Analyzer serial #:	1160290014
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000172	1.003517	Backgd or Offset:	36.5	36.0
Calibration intercept:	2.710765	0.290852	Coeff or Slope:	1.228	1.202

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.7	----
as found span	4926	79.9	799.0	816.4	0.979
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.7	----
high point	4926	79.9	799.0	802.5	0.996
second point	4968	40.0	399.8	400.8	0.998
third point	4987	20.0	200.0	200.9	0.995
as left zero	5000	0.0	0.0	-0.6	----
as left span	4926	80.0	800.0	800.8	0.999
Average Correction Factor					0.996

Baseline Corr As found:	815.70	Previous response	801.86	*% change	1.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 17, 2024
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	10:50	End Time (MST):	14:30
Analyzer make:	Thermo scientific	Analyzer serial #:	1160290014

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient	0.999998	≥0.995
799.0	802.5	0.9957			
399.8	400.8	0.9976	Slope	1.003517	0.90 - 1.10
200.0	200.9	0.9953			
			Intercept	0.290852	+/-30



SO2 Calibration Plot

Date: February 21, 2024

Location: Blackgold





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Blackgold Station number: AMS511
 Calibration Date: February 8, 2024 Last Cal Date: January 25, 2024
 Start time (MST): 11:53 End time (MST): 16:12
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.139 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC511397
 Removed Cal Gas Conc: 5.139 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 2445
 ZAG Make/Model: API T701 Serial Number: 138

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1336160090
 Converter make: Global G150 Converter serial #: 2022-227
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.977047	0.998919	Backgd or Offset:	3.39
Calibration intercept:	0.181198	0.140727	Coeff or Slope:	1.159
				1.189

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4922	77.8	80.0	78.7	1.016
as found 2nd point	4961	38.9	40.0	39.3	1.017
as found 3rd point	4981	19.5	20.0	19.7	1.017
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4922	77.8	80.0	80.0	1.000
second point	4961	38.9	40.0	40.1	0.997
third point	4981	19.5	20.0	20.2	0.992
as left zero	5000	0.0	0.0	0.3	----
as left span	4922	77.9	80.0	79.8	1.003
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.996
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 78.7 Prev response: 78.31 *% change: 0.5%
 Baseline Corr 2nd AF pt: 39.3 AF Slope: 0.984195 AF Intercept: -0.018968
 Baseline Corr 3rd AF pt: 19.7 AF Correlation: 1.000000

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	February 8, 2024	Previous Calibration:	January 25, 2024
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	11:53	End Time (MST):	16:12
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999999	≥0.995
80.0	80.0	0.9996			
40.0	40.1	0.9971	Slope	0.998919	0.90 - 1.10
20.0	20.2	0.9921			
			Intercept	0.140727	+/-3



H₂S Calibration Plot

Date: February 8, 2024

Location: Blackgold





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	February 21, 2024	Previous Calibration:	January 17, 2024
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	10:50	End Time (MST):	14:30
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1317958295

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.06	----	Correlation Coefficient	0.999887	
17.11	16.85	1.0159			≥0.995
8.57	8.32	1.0301	Slope	0.982206	
4.28	4.14	1.0339			0.90 - 1.10
			Intercept	-0.016488	+/-1.5



THC Calibration Plot

Date: February 21, 2024

Location: Blackgold





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Blackgold
Calibration Date: February 23, 2024
Start time (MST): 8:58
Reason: Routine
Station number: AMS511
Last Cal Date: January 9, 2024
End time (MST): 13:36

Calibration Standards

NO Gas Cylinder #: T0F8P52
NOX Cal Gas Conc: 47.43 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.43 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: August 16, 2026
NO Cal Gas Conc: 47.43 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.43 ppm
NO gas Diff:
Serial Number: 2445
Serial Number: 138

Analyzer Information

Analyzer make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 7029

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.041	1.041	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.038	1.038	NOX bkgnd or offset:	0.4	0.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.1	4.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002964	1.016612
NO _x Cal Offset:	-1.992779	0.426960
NO Cal Slope:	1.002150	1.015955
NO Cal Offset:	-2.932730	-0.193015
NO ₂ Cal Slope:	1.004437	1.002724
NO ₂ Cal Offset:	1.342612	0.288096



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
as found span	4916	84.4	800.6	800.6	0.0	809.8	806.6	3.1	0.9886	0.9925
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4916	84.4	800.6	800.6	0.0	814.0	813.1	0.7	0.9835	0.9846
second point	4958	42.2	400.3	400.3	0.0	407.7	406.7	1.1	0.9819	0.9843
third point	4979	21.1	200.2	200.2	0.0	204.4	202.9	1.5	0.9792	0.9865
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as left span	4916	84.4	800.6	403.9	396.7	810.8	418.8	392.0	0.9874	0.9643
Average Correction Factor									0.9815	0.9851

Corrected As found	NO _x = 809.7 ppb	NO = 806.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.1%
Previous Response	NO _x = 800.9 ppb	NO = 799.3 ppb		*Percent Change	NO = 0.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	813.2	416.5	396.7	398.2	0.9962	100.4%
2nd GPT point (200 ppb O3)	813.2	647.2	166.0	166.0	1.0000	100.0%
3rd GPT point (100 ppb O3)	813.2	734.3	78.9	80.3	0.9826	101.8%
Average Correction Factor					0.9929	100.7%

Notes: Changed the sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	February 23, 2024	Previous Calibration:	January 9, 2024
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:58	End Time (MST):	13:36
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.6	814.0	0.9835		
400.3	407.7	0.9819		
200.2	204.4	0.9792		
			0.999998	
			1.016612	
			0.426960	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 23, 2024	Previous Calibration:	January 9, 2024
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:58	End Time (MST):	13:36
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	1.000000	
800.6	813.1	0.9846			≥0.995
400.3	406.7	0.9843	Slope	1.015955	
200.2	202.9	0.9865			0.90 - 1.10
			Intercept	-0.193015	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	February 23, 2024	Previous Calibration:	January 9, 2024
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:58	End Time (MST):	13:36
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
396.7	398.2	0.9962		
166.0	166.0	1.0000		
78.9	80.3	0.9826		
			0.999984	
			1.002724	
			0.288096	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: February 23, 2024

Location: Blackgold





End of Report