



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

MARCH 2024 MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

April 30, 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01 BERTHA GANTER - FORT MCKAY MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 5, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:11	End Time (MST):	14:41
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999991	≥0.995
800.3	798.5	1.0022			
400.6	397.9	1.0068	Slope	0.997052	0.90 - 1.10
199.8	199.1	1.0036			
			Intercept	-0.112969	+/-30



SO2 Calibration Plot

Date: February 5, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: March 18, 2024 Last Cal Date: February 7, 2024
 Start time (MST): 9:09 End time (MST): 13:18
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153461
 Converter make: CD Nova Converter serial #: 470
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997793	1.002936	Backgd or Offset: 2.45	2.43
Calibration intercept:	0.179996	0.199998	Coeff or Slope: 0.937	0.931

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4921	78.4	80.0	80.5	0.994
as found 2nd point	4960	39.2	40.0	40.5	0.988
as found 3rd point	4980	19.6	20.0	20.2	0.990
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4921	78.4	80.0	80.4	0.995
second point	4960	39.2	40.0	40.4	0.990
third point	4980	19.6	20.0	20.2	0.990
as left zero	5000	0.0	0.0	0.3	----
as left span	4921	78.4	80.0	80.5	0.994
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	0.992
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 80.5 Prev response: 80.00 *% change: 0.6%
 Baseline Corr 2nd AF pt: 40.5 AF Slope: 1.006365 AF Intercept: 0.079997
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999988

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 18, 2024	Previous Calibration:	February 7, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:09	End Time (MST):	13:18
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	≥ 0.995	
80.0	80.4	0.9949			
40.0	40.4	0.9901			
20.0	20.2	0.9900			
			Slope	1.002936	0.90 - 1.10
			Intercept	0.199998	+/-3



TRS Calibration Plot

Date: March 18, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: March 18, 2024 Last Cal Date: February 7, 2024
 Start time (MST): 9:09 End time (MST): 13:18
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.10 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511749
 Removed Cal Gas Conc: 5.10 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1200326167
 Converter make: Global Converter serial #: 2022-221
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002571	1.005285	Backgd or Offset: 2.04	2.02
Calibration intercept:	0.116774	-0.003216	Coeff or Slope: 0.992	0.985

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4922	78.4	80.0	80.9	0.988
as found 2nd point	4960	39.2	40.0	40.6	0.985
as found 3rd point	4980	19.6	20.0	20.2	0.990
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	78.4	80.0	80.4	0.995
second point	4960	39.2	40.0	40.2	0.995
third point	4980	19.6	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.0	----
as left span	4922	78.4	80.0	80.6	0.992
SO2 Scrubber Check	4919	81.3	813.0	0.1	----
Date of last scrubber change:	January 25, 2024			Ave Corr Factor	0.995
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 80.9 Prev response: 80.28 *% change: 0.8%
 Baseline Corr 2nd AF pt: 40.6 AF Slope: 1.012057 AF Intercept: 0.016754
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999995

* = > +/-5% change initiates investigation

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 18, 2024	Previous Calibration:	February 7, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:09	End Time (MST):	13:18
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	1.000000	≥0.995
80.0	80.4	0.9947			
40.0	40.2	0.9950	Slope	1.005285	0.90 - 1.10
20.0	20.1	0.9949			
			Intercept	-0.003216	+/-3



H₂S Calibration Plot

Date: March 18, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	March 5, 2024	Last Cal Date:	February 5, 2024
Start time (MST):	11:11	End time (MST):	14:41
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
ZAG make/model:	Teledyne API T701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.30E-04	4.30E-04	NMHC SP Ratio:	7.18E-05
CH ₄ Retention time:	16.7	16.7	NMHC Peak Area:	127929
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				7.20E-05
				127593
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	17.27	17.22	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	17.27	17.29	0.999
second point	4959	40.7	8.64	8.53	1.014
third point	4980	20.3	4.31	4.25	1.016
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	17.27	17.30	0.998
Average Correction Factor					1.009

Baseline Corr AF:	17.22	Prev response	17.23	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4918	81.3	9.18	9.10	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	9.18	9.14	1.005
second point	4959	40.7	4.60	4.55	1.011
third point	4980	20.3	2.29	2.26	1.014
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	9.18	9.19	1.000
Average Correction Factor					1.010
Baseline Corr AF:	9.10	Prev response	9.16	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4918	81.3	8.09	8.13	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4918	81.3	8.09	8.15	0.993
second point	4959	40.7	4.05	3.98	1.017
third point	4980	20.3	2.02	1.98	1.017
as left zero	5000	0.0	0.00	0.00	----
as left span	4918	81.3	8.09	8.12	0.996
Average Correction Factor					1.009
Baseline Corr AF:	8.13	Prev response	8.07	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000371	1.001618
THC Cal Offset:	-0.041569	-0.052990
CH ₄ Cal Slope:	1.002874	1.008062
CH ₄ Cal Offset:	-0.035591	-0.039196
NMHC Cal Slope:	0.998167	0.995931
NMHC Cal Offset:	-0.005978	-0.013994

Notes: Changed out the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 5, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:11	End Time (MST):	14:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999933	≥ 0.995			
17.27	17.29	0.9990						
8.64	8.53	1.0135				Slope	1.001618	0.90 - 1.10
4.31	4.25	1.0155						
			Intercept	-0.052990	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 5, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:11	End Time (MST):	14:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999821	≥ 0.995
8.09	8.15	0.9927			
4.05	3.98	1.0169			
2.02	1.98	1.0175			
			Slope	1.008062	0.90 - 1.10
			Intercept	-0.039196	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 5, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	11:11	End Time (MST):	14:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995
9.18	9.14	1.0045			
4.60	4.55	1.0108			
2.29	2.26	1.0138			
			Slope	0.995931	0.90 - 1.10
			Intercept	-0.013994	+/-0.5



NMHC Calibration Plot

Date: March 5, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
Calibration Date: March 14, 2024 Last Cal Date: February 26, 2024
Start time (MST): 10:32 End time (MST): 15:51
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC335700 Cal Gas Expiry Date: September 1, 2032
NOX Cal Gas Conc: 59.40 ppm NO Cal Gas Conc: 59.20 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 59.40 ppm Removed Gas NO Conc: 59.20 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3565
ZAG make/model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1218153357
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.556	1.496	NO bkgnd or offset:	7.6	7.4
NOX coeff or slope:	0.999	0.999	NOX bkgnd or offset:	7.6	7.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	196.0	200.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000310	0.999371
NO _x Cal Offset:	-0.160000	0.020000
NO Cal Slope:	1.001719	1.000862
NO Cal Offset:	-0.920000	-1.220000
NO ₂ Cal Slope:	0.995800	0.995953
NO ₂ Cal Offset:	-0.397873	0.117298



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.7	0.3	0.4	----	----
as found span	4932	67.6	803.1	800.4	2.7	868.0	860.8	7.3	0.9252	0.9298
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.6	0.3	0.3	----	----
high point	4932	67.6	803.1	800.4	2.7	802.8	800.6	2.1	1.0004	0.9997
second point	4966	33.8	401.5	400.2	1.4	401.3	398.6	2.7	1.0006	1.0040
third point	4983	16.9	200.8	200.1	0.7	199.9	197.5	2.4	1.0044	1.0131
as left zero	5000	0.0	0.0	0.0	0.0	0.3	0.0	0.3	----	----
as left span	4932	67.6	803.1	398.7	404.4	797.1	392.0	405.1	1.0075	1.0171
Average Correction Factor									1.0018	1.0056

Corrected As found	NO _x = 867.3 ppb	NO = 860.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 7.4%
Previous Response	NO _x = 803.2 ppb	NO = 800.8 ppb		*Percent Change	NO = 6.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.3	395.6	404.4	403.5	1.0022	99.8%
2nd GPT point (200 ppb O3)	797.3	593.0	207.0	204.7	1.0113	98.9%
3rd GPT point (100 ppb O3)	797.3	699.4	100.6	101.1	0.9951	100.5%
Average Correction Factor					1.0029	99.7%

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 14, 2024	Previous Calibration:	February 26, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:32	End Time (MST):	15:51
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.6	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
803.1	802.8	1.0004		
401.5	401.3	1.0006		
200.8	199.9	1.0044		
			0.999997	
			0.999371	
			0.020000	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 14, 2024	Previous Calibration:	February 26, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:32	End Time (MST):	15:51
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.4	800.6	0.9997		
400.2	398.6	1.0040		
200.1	197.5	1.0131		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 14, 2024	Previous Calibration:	February 26, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:32	End Time (MST):	15:51
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
404.4	403.5	1.0022		
207.0	204.7	1.0113		
100.6	101.1	0.9951		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 14, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: March 6, 2024 Last Cal Date: February 2, 2024
 Start time (MST): 10:42 End time (MST): 13:30
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 1107
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999143	1.001057	Backgd or Offset:	5.3	5.5
Calibration intercept:	1.000000	0.640000	Coeff or Slope:	1.015	1.044

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	5000	863.1	400.0	390.8	1.024
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.2	----
high point	5000	863.1	400.0	400.6	0.999
second point	5000	742.5	200.0	201.4	0.993
third point	5000	651.7	100.0	101.5	0.985
as left zero	5000	0.0	0.0	0.0	----
as left span	5000	863.1	400.0	400.7	0.998
Average Correction Factor					0.992

Baseline Corr As found:	390.9	Previous response	400.7	*% change	-2.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 2, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:42	End Time (MST):	13:30
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999980	≥0.995
400.0	400.6	0.9985			
200.0	201.4	0.9930	Slope	1.001057	0.90 - 1.10
100.0	101.5	0.9852			
			Intercept	0.640000	+/- 5



O₃ Calibration Plot

Date: March 6, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
 Calibration Date: March 18, 2024 Last Cal Date: February 27, 2024
 Start time (MST): 12:10 End time (MST): 13:04

Analyzer Make: Teledyne API T640 S/N: 324
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752
 Temp/RH standard: Alicat FP-25BT S/N: 388752

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-1.7	-2.41	-1.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	740.4	739.69	740.4	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.03	5.16	5.03	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	40	----	40	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	7.3	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: _____ January 26, 2024
 Date Disposable Filter Changed: _____ January 26, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ September 14, 2023
 Date RH/T Sensor Cleaned: _____ January 26, 2024

Notes: Verified flow, temperature, and pressure. Leak check passed.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS01
 Calibration Date: March 15, 2024 Last Cal Date: February 21, 2024
 Start time (MST): 10:02 End time (MST): 13:04
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 3040 ppm Cal Gas Exp Date: December 1, 2028
 Cal Gas Cylinder #: ALM042207
 Removed Cal Gas Conc: 3040 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Teledyne API T300 Analyzer serial #: 3520
 Analyzer Range: 0 - 50 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001370	1.000382	Backgd or Offset:	-0.013	-0.013
Calibration intercept:	0.145823	0.195846	Coeff or Slope:	0.989	0.989

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.6	40.7	0.998
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.1	----
high point	4933	66.7	40.6	40.6	0.998
second point	4966	33.3	20.2	20.7	0.976
third point	4983	16.7	10.2	10.3	0.983
as left zero	5000	0.0	0.0	0.1	----
as left span	2960	40.0	40.5	40.1	1.010
Average Correction Factor					0.986

Baseline Corr As found: 40.57 Prev response: 40.76 *% change: -0.5%
 Baseline Corr 2nd AF pt: NA AF Slope: AF Intercept:
 Baseline Corr 3rd AF pt: NA AF Correlation:

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 15, 2024	Previous Calibration:	February 21, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:02	End Time (MST):	13:04
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999869	≥0.995
40.6	40.6	0.9982			
20.2	20.7	0.9763	Slope	1.000382	0.90 - 1.10
10.2	10.3	0.9830			
			Intercept	0.195846	+/-1.5



CO Calibration Plot

Date: March 15, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
Calibration Date:	March 12, 2024	Last Cal Date:	February 6, 2024
Start time (MST):	10:13	End time (MST):	13:46
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	<u>ALM042207</u>			
Removed Cal Gas Conc:	<u>60,200</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
N2 Gen Make/Model:	Peak Sci NG5000		Serial Number:	7220900034

Analyzer Information

Analyzer make: Teledyne API 360	Analyzer serial #: 442
Analyzer Range 0 - 2,000 ppm	

	Start	Finish		Start	Finish
Calibration slope:	0.998372	1.000907	Backgd or Offset:	0.045	0.045
Calibration intercept:	-5.840000	-6.720000	Coeff or Slope:	0.876	0.876

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	0.0	----
as found span	2920	80.0	1605.3	1600.8	1.003
as found 2nd point	2960	40.0	802.7	793.9	1.011
as found 3rd point	2980	20.0	401.3	390.1	1.029
new cylinder response					
calibrator zero	3000	0.0	0.0	0.0	----
high point	2920	80.0	1605.3	1604.6	1.000
second point	2960	40.0	802.7	789.8	1.016
third point	2980	20.0	401.3	390.6	1.027
as left zero	3000	0.0	0.0	0.1	----
as left span	2960	40.0	802.7	778.7	1.031
Average Correction Factor					1.015

Baseline Corr As found:	1600.80	Prev response:	1596.88	*% change:	0.2%
Baseline Corr 2nd AF pt:	793.9	AF Slope:	0.998870	AF Intercept:	-5.340000
Baseline Corr 3rd AF pt:	390.1	AF Correlation:	0.999949		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	March 12, 2024	Previous Calibration	February 6, 2024
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS01
Start Time (MST)	10:13	End Time (MST)	13:46
Analyzer make	Teledyne API 360	Analyzer serial #	442

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999906	≥0.995
1605.3	1604.6	1.0005			
802.7	789.8	1.0163	Slope	1.000907	0.90 - 1.10
401.3	390.6	1.0275			
			Intercept	-6.720000	+/-10



CO₂ Calibration Plot

Date: March 12, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS01
NOX Cal Date:	March 19, 2024	Last Cal Date:	February 27, 2024
Start time (MST):	9:46	End time (MST):	14:14
NH3 Cal Date:	March 20, 2024	Last Cal Date:	February 27, 2024
Start time (MST):	10:00	End time (MST):	13:35
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	59.40	ppm	NO Gas Cylinder #:	CC335700
NO Cal Gas Conc:	59.20	ppm	NO Cal Gas Expiry:	September 1, 2032
Removed NOX Conc:	59.40	ppm	Removed Cylinder #:	NA
Removed NO Conc:	59.20	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.58	ppm	NH3 Gas Cylinder #:	CC743587
			NH3 Cal Gas Expiry:	August 22, 2024
Removed NH3 Conc:	76.58	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	Teledyne API T700		Serial Number:	3565
ZAG make/model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	475
Converter model:	Teledyne API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.00
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	552

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.956	0.960	TN coefficient:	0.968	0.969
NOX coefficient:	0.962	0.966	NO bkgnd:	-0.9	-0.9
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-0.3	-0.3
NH3 coefficient:	0.946	0.946	TN bkgnd:	1.2	1.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999528	1.001292
NO _x Cal Offset:	-1.960000	-2.380000
NO Cal Slope:	0.999963	1.000877
NO Cal Offset:	-2.580000	-2.600000
NO ₂ Cal Slope:	0.992493	0.999426
NO ₂ Cal Offset:	-0.912249	-0.898937
NH3 Cal Slope:	1.004564	1.002078
NH3 Cal Offset:	-0.612859	-0.363408
TN Cal Slope:	1.007097	1.004775
TN Cal Offset:	-0.451064	-0.436983



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	0.1	-0.2	----	----
as found NO	4932	67.6	803.1	803.1	----	804.7	806.3	-1.6	0.998	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
high NO point	4932	67.6	803.1	803.1	----	801.5	803.0	-1.3	1.002	----
NO/O3 point	4932	67.6	803.1	803.1	----	798.7	796.7	2.0	1.005	----
as found NH3	3418	82.2	1798.5	----	1798.5	1806.1	----	1801.4	0.996	0.998
new NH3 cyl rp							----			
first NH3	3418	82.2	1798.5	----	1798.5	1806.1	----	1801.4	0.996	0.998
second NH3	3454	45.7	1000.0	----	1000.0	1006.1	----	1003.5	0.994	0.997
third NH3	3477	22.8	498.9	----	498.9	499.4	----	497.8	0.999	1.002
Average Correction Factor									1.0037	0.9990

Corrected As found TN = 804.9 ppb NO_x = 806.2 ppb NH3 = 1801.6 ppb
 Previous Response TN = 808.3 ppb NO_x = 800.7 ppb NH3 = 1806.1 ppb

*Percent Change TN = -0.4%
 *Percent Change NO_x = 0.7%
 *Percent Change NH3 = -0.3%
 * = > +/-5% change initiates investigation

NH3 Previous Converter Efficiency = 94.6%
 NH3 Current Converter Efficiency = 94.6%



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated TN concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated TN concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	-0.7	----	----
as found span	4932	67.6	803.1	800.4	803.1	806.7	797.2	803.7	0.9955	1.0040
new NO cyl rp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.2	----	----
high point	4932	67.6	803.1	800.4	803.1	803.0	799.8	801.5	1.0001	1.0007
second point	4966	33.8	401.5	400.2	401.5	398.0	396.5	398.1	1.0089	1.0093
third point	4983	16.9	200.8	200.1	200.8	197.0	195.3	198.2	1.0191	1.0246
Average Correction Factor									1.0094	1.0115

Baseline Corr As fnd	TN = 804.4 ppb	NO _x = 806.9 ppb	NO = 797.4 ppb	*Percent Change	TN = -0.5%
Previous Response	TN = 808.3 ppb	NO _x = 800.7 ppb	NO = 797.8 ppb	*Percent Change	NO _x = 0.8%
				*Percent Change	NO = 0.0%

* = > +/-5% change initiates investigation

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found zero	----	----	0.0	-0.1	----	----
calibration zero	----	----	0.0	-0.2	----	----
1st GPT point (400 ppb O3)	795.8	385.0	413.5	412.9	1.0015	99.9%
2nd GPT point (200 ppb O3)	795.8	590.8	207.7	205.8	1.0093	99.1%
3rd GPT point (100 ppb O3)	795.8	694.1	104.4	103.1	1.0126	98.8%
Average Correction Factor					1.0078	99.2%

Notes: Changed the inlet filter after as founds. Adjusted the NO_x/TN span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TN Calibration Summary

Version-05-2023


Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 27, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:00	End Time (MST):	13:35
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥ 0.995 $0.90 - 1.10$ ± 20
1798.5	1806.1	0.9958		
1000.0	1006.1	0.9939		
498.9	499.4	0.9989		

TN Calibration Curve





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 27, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	10:00	End Time (MST):	13:35
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
1798.5	1801.4	0.9984			
1000.0	1003.5	0.9965			
498.9	497.8	1.0021			
			Slope	1.002078	0.90 - 1.10
			Intercept	-0.363408	+/-20





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	March 19, 2024	Previous Calibration:	February 27, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:46	End Time (MST):	14:14
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
803.1	803.0	1.0001		
401.5	398.0	1.0089		
200.8	197.0	1.0191		
			0.999968	
			1.001292	
			-2.380000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	March 19, 2024	Previous Calibration:	February 27, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:46	End Time (MST):	14:14
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
800.4	799.8	1.0007		
400.2	396.5	1.0093		
200.1	195.3	1.0246		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	March 19, 2024	Previous Calibration:	February 27, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS01
Start Time (MST):	9:46	End Time (MST):	14:14
Analyzer make:	Teledyne API T201	Analyzer serial #:	475

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
413.5	412.9	1.0015			
207.7	205.8	1.0093			
104.4	103.1	1.0126			
			Slope	0.999426	0.90 - 1.10
			Intercept	-0.898937	+/-20



NO_x Calibration Plot

Date: March 19, 2024


Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: March 20, 2024

Location: Bertha Ganter-Fort McKay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 19, 2024	Previous Calibration:	February 15, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:09	End Time (MST):	13:26
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999999	
801.6	802.0	0.9996			≥0.995
400.8	400.4	1.0011	Slope	1.000614	
199.9	199.4	1.0026			0.90 - 1.10
			Intercept	-0.364904	+/-30



SO2 Calibration Plot

Date: March 19, 2024

Location: Mildred Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mildred Lake Station number: AMS02
 Calibration Date: March 26, 2024 Last Cal Date: February 2, 2024
 Start time (MST): 10:09 End time (MST): 14:42
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.29 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345191
 Removed Cal Gas Conc: 5.29 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1185
 ZAG Make/Model: API T701 Serial Number: 4891

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 12113311966
 Converter make: Global G150 Converter serial #: 2022-198
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998393	1.000964	Backgd or Offset: 1.68	1.67
Calibration intercept:	-0.039204	0.020800	Coeff or Slope: 0.731	0.731

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4924	75.6	80.0	80.6	0.994
as found 2nd point	4962	37.8	40.0	40.3	0.995
as found 3rd point	4981	18.9	20.0	19.9	1.010
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4924	75.6	80.0	80.1	0.999
second point	4962	37.8	40.0	40.1	0.997
third point	4981	18.9	20.0	19.9	1.005
as left zero	5000	0.0	0.0	0.1	----
as left span	4924	75.6	80.0	80.2	0.997
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:	20-Sep-23			Ave Corr Factor	1.000
Date of last converter efficiency test:	March 26, 2024			efficiency	

Baseline Corr As found: 80.5 Prev response: 79.82 *% change: 0.8%
 Baseline Corr 2nd AF pt: 40.2 AF Slope: 1.007679 AF Intercept: -0.039196
 Baseline Corr 3rd AF pt: 19.8 AF Correlation: 0.999981

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Ran SOx scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 2, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:09	End Time (MST):	14:42
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999992	≥0.995
80.0	80.1	0.9986			
40.0	40.1	0.9974	Slope	1.000964	0.90 - 1.10
20.0	19.9	1.0049			
			Intercept	0.020800	+/-3



H₂S Calibration Plot

Date: March 26, 2024

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mildred Lake	Station number:	AMS02
Calibration Date:	March 7, 2024	Last Cal Date:	February 15, 2024
Start time (MST):	10:24	End time (MST):	13:21
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
C ₃ H ₈ Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1048.6 ppm
Removed C ₃ H ₈ Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
ZAG make/model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.73E-04	2.92E-04	NMHC SP Ratio:	6.17E-05
CH ₄ Retention time:	15.9	16.3	NMHC Peak Area:	142620
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF
				6.02E-05
				146130
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	16.82	16.56	1.016
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	16.82	16.84	0.999
second point	4960	40.1	8.41	8.38	1.004
third point	4980	20.0	4.19	4.18	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	16.82	16.76	1.004
Average Correction Factor					1.003

Baseline Corr AF:	16.56	Prev response	16.78	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.80	9.02	0.975
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.80	8.81	0.998
second point	4960	40.1	4.40	4.38	1.005
third point	4980	20.0	2.19	2.19	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.80	8.74	1.007
Average Correction Factor					1.002
Baseline Corr AF:	9.02	Prev response	8.76	*% change	2.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	8.02	7.54	1.064
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	8.02	8.02	1.000
second point	4960	40.1	4.01	4.00	1.003
third point	4980	20.0	2.00	1.99	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	8.02	8.02	1.000
Average Correction Factor					1.004
Baseline Corr AF:	7.54	Prev response	8.01	*% change	-6.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998749	1.001195
THC Cal Offset:	-0.021919	-0.017905
CH ₄ Cal Slope:	1.000498	1.000469
CH ₄ Cal Offset:	-0.013049	-0.008450
NMHC Cal Slope:	0.996921	1.001609
NMHC Cal Offset:	-0.008470	-0.009255

Notes: Sample inlet filters changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 15, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:24	End Time (MST):	13:21
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	≥ 0.995			
16.82	16.84	0.9990						
8.41	8.38	1.0041				Slope	1.001195	0.90 - 1.10
4.19	4.18	1.0046						
			Intercept	-0.017905	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 15, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:24	End Time (MST):	13:21
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995
8.02	8.02	1.0000			
4.01	4.00	1.0031			
2.00	1.99	1.0075			
			Slope	1.000469	0.90 - 1.10
			Intercept	-0.008450	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 15, 2024
Station Name:	Mildred Lake	Station Number:	AMS02
Start Time (MST):	10:24	End Time (MST):	13:21
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995			
8.80	8.81	0.9983						
4.40	4.38	1.0052				Slope	1.001609	0.90 - 1.10
2.19	2.19	1.0020						
			Intercept	-0.009255	± 0.5			



NMHC Calibration Plot

Date: March 7, 2024

Location: Mildred Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 21, 2024	Previous Calibration:	February 9, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:18	End Time (MST):	11:15
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999998	≥0.995
799.7	800.0	0.9997			
399.8	400.4	0.9985	Slope	1.000873	0.90 - 1.10
199.4	198.6	1.0042			
			Intercept	-0.325433	+/-30



SO2 Calibration Plot

Date: March 21, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	March 20, 2024	Last Cal Date:	February 29, 2024
Start time (MST):	6:00	End time (MST):	10:20
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.42	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC345266			
Removed Cal Gas Conc:	5.42	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3808
ZAG Make/Model:	API T701H		Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400
Converter make:	Global	Converter serial #:	2022-200
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997925	0.995649	Backgd or Offset:	1.09	1.09
Calibration intercept:	0.202175	0.182120	Coeff or Slope:	1.130	1.130

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4926	74.1	80.3	80.6	0.998
as found 2nd point	4963	37.0	40.1	40.3	0.998
as found 3rd point	4982	18.5	20.1	20.1	1.003
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4926	74.1	80.3	80.1	1.003
second point	4963	37.0	40.1	40.3	0.995
third point	4982	18.5	20.1	20.0	1.003
as left zero	5000	0.0	0.0	0.3	----
as left span	4926	74.1	80.3	80.0	1.004
SO2 Scrubber Check	4920	80.0	800.0	0.2	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	1.000
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	80.5	Prev response:	80.36	*% change:	0.2%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	1.002622	AF Intercept:	0.062213
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999998		

* = > +/-5% change initiates investigation

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 29, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:00	End Time (MST):	10:20
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999983	
80.3	80.1	1.0028			≥0.995
40.1	40.3	0.9952	Slope	0.995649	
20.1	20.0	1.0026			0.90 - 1.10
			Intercept	0.182120	+/-3



H₂S Calibration Plot

Date: March 20, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS04
Calibration Date:	March 21, 2024	Last Cal Date:	February 9, 2024
Start time (MST):	6:18	End time (MST):	11:14
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	NA	Cal Gas Expiry Date:	March 10, 1931
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
ZAG make/model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.21E-04	4.25E-04	NMHC SP Ratio:	1.08E-04
CH ₄ Retention time:	13.7	13.7	NMHC Peak Area:	81432
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	16.64	16.58	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	16.64	16.65	0.999
second point	4961	39.3	8.32	8.34	0.997
third point	4980	19.6	4.15	4.16	0.997
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	16.64	16.69	0.997
Average Correction Factor					0.998

Baseline Corr AF:	16.58	Prev response	16.61	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	8.82	8.78	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	8.82	8.82	1.000
second point	4961	39.3	4.41	4.43	0.995
third point	4980	19.6	2.20	2.22	0.990
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	8.82	8.86	0.995
Average Correction Factor					0.995
Baseline Corr AF:	8.78	Prev response	8.81	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	78.6	7.82	7.80	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	78.6	7.82	7.82	0.999
second point	4961	39.3	3.91	3.91	1.000
third point	4980	19.6	1.95	1.94	1.004
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	78.6	7.82	7.83	0.998
Average Correction Factor					1.001
Baseline Corr AF:	7.80	Prev response	7.81	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.997921	1.000586
THC Cal Offset:	0.010460	0.008071
CH ₄ Cal Slope:	0.999537	1.001043
CH ₄ Cal Offset:	-0.004306	-0.004702
NMHC Cal Slope:	0.996864	0.999701
NMHC Cal Offset:	0.013566	0.013373

Notes: Hydrogen Cylinder Replaced. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 21, 2024	Previous Calibration:	February 9, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:18	End Time (MST):	11:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
16.64	16.65	0.9993						
8.32	8.34	0.9972				Slope	1.000586	0.90 - 1.10
4.15	4.16	0.9965						
			Intercept	0.008071	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 21, 2024	Previous Calibration:	February 9, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:18	End Time (MST):	11:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥0.995
7.82	7.82	0.9993			
3.91	3.91	1.0004			
1.95	1.94	1.0042			
			Slope	1.001043	0.90 - 1.10
			Intercept	-0.004702	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 21, 2024	Previous Calibration:	February 9, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:18	End Time (MST):	11:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999989	≥ 0.995			
8.82	8.82	0.9997						
4.41	4.43	0.9946				Slope	0.999701	0.90 - 1.10
2.20	2.22	0.9898						
			Intercept	0.013373	± 0.5			



NMHC Calibration Plot

Date: March 21, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: March 1, 2024
Start time (MST): 6:50
Reason: Routine
Station number: AMS04
Last Cal Date: February 12, 2024
End time (MST): 11:56

Calibration Standards

NO Gas Cylinder #: CC324979
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: November 3, 1932
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 721

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.162	1.162	NO bkgnd or offset:	-0.6	-0.6
NOX coeff or slope:	1.154	1.154	NOX bkgnd or offset:	-0.3	-0.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.3	4.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001294	0.995753
NO _x Cal Offset:	-0.213331	-0.374543
NO Cal Slope:	0.999626	0.993186
NO Cal Offset:	-0.914344	-1.315833
NO ₂ Cal Slope:	0.997058	0.998315
NO ₂ Cal Offset:	-0.192985	1.614021



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
as found span	4918	81.8	800.0	798.4	1.6	792.8	789.1	3.6	1.0091	1.0118
as found 2nd										
as found 3rd										
new cyl resp	4918	81.8	800.0	798.4	1.6					
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.5	-0.2	----	----
high point	4918	81.8	800.0	798.4	1.6	796.5	792.7	3.7	1.0044	1.0072
second point	4959	40.9	400.0	399.2	0.8	397.9	393.8	4.1	1.0053	1.0137
third point	4980	20.4	199.5	199.1	0.4	197.4	194.9	2.5	1.0106	1.0215
as left zero	5000	0.0	0.0	0.0	0.0	0.4	1.0	-0.6	----	----
as left span	4918	81.8	800.0	402.7	397.3	787.1	393.9	393.2	1.0164	1.0223
Average Correction Factor									1.0068	1.0141

Corrected As found	NO _x = 792.7 ppb	NO = 788.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.0%
Previous Response	NO _x = 800.9 ppb	NO = 797.2 ppb		*Percent Change	NO = -1.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	789.8	394.1	397.3	397.6	0.9993	100.1%
2nd GPT point (200 ppb O3)	789.8	595.0	196.4	198.0	0.9921	100.8%
3rd GPT point (100 ppb O3)	789.8	694.2	97.2	100.9	0.9637	103.8%
Average Correction Factor					0.9850	101.5%

Notes:

No maintenance or adjustments done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 1, 2024	Previous Calibration:	February 12, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:50	End Time (MST):	11:56
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.3	----	Correlation Coefficient	≥0.995	
800.0	796.5	1.0044			
400.0	397.9	1.0053			
199.5	197.4	1.0106			
			Slope	0.995753	0.90 - 1.10
			Intercept	-0.374543	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 1, 2024	Previous Calibration:	February 12, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:50	End Time (MST):	11:56
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.5	----	Correlation Coefficient	≥0.995	
798.4	792.7	1.0072			
399.2	393.8	1.0137			
199.1	194.9	1.0215			
			Slope	0.993186	0.90 - 1.10
			Intercept	-1.315833	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 1, 2024	Previous Calibration:	February 12, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	6:50	End Time (MST):	11:56
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
397.3	397.6	0.9993			
196.4	198.0	0.9921			
97.2	100.9	0.9637			
			Slope	0.998315	0.90 - 1.10
			Intercept	1.614021	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 1, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Buffalo Viewpoint Station number: AMS04
 Calibration Date: March 15, 2024 Last Cal Date: February 9, 2024
 Start time (MST): 8:57 End time (MST): 11:15
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3808
 ZAG Make/Model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 2961
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999029	1.004971	Backgd or Offset:	-2.2	-2.2
Calibration intercept:	0.420000	-0.120000	Coeff or Slope:	1.011	1.011

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.6	----
as found span	5000	990.7	400.0	401.9	0.995
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.3	----
high point	5000	991.2	400.0	401.8	0.996
second point	5000	819.1	200.0	200.9	0.996
third point	5000	710.9	100.0	100.6	0.994
as left zero	5000	0.0	0.0	-0.5	----
as left span	5000	992.5	400.0	402.1	0.995
Average Correction Factor					0.995

Baseline Corr As found:	402.5	Previous response	400.0	*% change	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments or maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 15, 2024	Previous Calibration:	February 9, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS04
Start Time (MST):	8:57	End Time (MST):	11:15
Analyzer make:	API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999999	≥0.995
400.0	401.8	0.9955			
200.0	200.9	0.9955	Slope	1.004971	0.90 - 1.10
100.0	100.6	0.9940			
			Intercept	-0.120000	+/- 5



O₃ Calibration Plot

Date: March 15, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: March 20, 2024 Last Cal Date: February 27, 2024
 Start time (MST): 10:22 End time (MST): 11:01

Analyzer Make: API T640 S/N: 321
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 1451
 Temp/RH standard: Alicat FP-25BT S/N: 1451

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-13.6	-13.4	-13.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	741.9	743.6	741.9	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.14	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	41	----	41	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	5.5	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 6-10-2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 27, 2024
 Date Disposable Filter Changed: February 27, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: February 27, 2024
 Date RH/T Sensor Cleaned: February 27, 2024

Notes: No adjustments done. Leak check passed.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS05
MANNIX
MARCH 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 2, 2024
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	10:21	End Time (MST):	13:30
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999995	
800.3	804.2	0.9952			≥0.995
400.2	400.8	0.9984	Slope	1.005098	
200.1	199.8	1.0014			0.90 - 1.10
			Intercept	-0.700000	+/-30



SO2 Calibration Plot

Date: March 20, 2024

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Mannix Station number: AMS05
 Calibration Date: March 12, 2024 Last Cal Date: February 15, 2024
 Start time (MST): 9:28 End time (MST): 14:17
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.96 ppm Cal Gas Exp Date: November 15, 2026
 Cal Gas Cylinder #: DT0037363
 Removed Cal Gas Conc: 4.96 ppm Rem Gas Exp Date:
 Removed Gas Cyl #: Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 1845
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 1200326169
 Converter make: Global Converter serial #: 2022225
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008837	0.997116	Backgd or Offset:	1.24	1.23
Calibration intercept:	0.022150	-0.017688	Coeff or Slope:	0.998	0.978

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4919	80.6	80.0	81.4	0.982
as found 2nd point	4960	40.3	40.0	40.7	0.982
as found 3rd point	4980	20.2	20.0	20.1	0.997
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4919	80.6	80.0	79.8	1.002
second point	4960	40.3	40.0	39.7	1.007
third point	4980	20.2	20.0	19.9	1.007
as left zero	5000	0.0	0.0	0.1	----
as left span	4919	80.6	80.0	81.0	0.987
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.005
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 81.4 Prev response: 80.69 *% change: 0.9%
 Baseline Corr 2nd AF pt: 40.7 AF Slope: 1.019272 AF Intercept: -0.118001
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999983

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber test after calibrator zero. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 15, 2024
Station Name:	Mannix	Station Number:	AMS05
Start Time (MST):	9:28	End Time (MST):	14:17
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326169

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999987	
80.0	79.8	1.0020			≥0.995
40.0	39.7	1.0069	Slope	0.997116	
20.0	19.9	1.0069			0.90 - 1.10
			Intercept	-0.017688	+/-3



H₂S Calibration Plot

Date: March 12, 2024

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	March 10, 2024	Last Cal Date:	February 27, 2024
Start time (MST):	9:30	End time (MST):	12:08
Reason:	Maintenance	Baseline drop	

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.98E-04	2.74E-04	NMHC SP Ratio:	4.94E-05	4.75E-05
CH ₄ Retention time:	16.4	16.0	NMHC Peak Area:	185404	192745
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	16.62	1.036
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.20	1.002
second point	4960	40.0	8.61	8.61	1.001
third point	4980	20.0	4.31	4.30	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.16	1.004
Average Correction Factor					1.002

Baseline Corr AF:	16.62	Prev response	17.24	*% change	-3.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	9.15	9.26	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.13	1.002
second point	4960	40.0	4.57	4.58	0.999
third point	4980	20.0	2.29	2.29	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.15	9.07	1.009
Average Correction Factor					1.001
Baseline Corr AF:	9.26	Prev response	9.15	*% change	1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	7.37	1.097
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.07	1.001
second point	4960	40.0	4.04	4.03	1.002
third point	4980	20.0	2.02	2.01	1.005
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.09	0.999
Average Correction Factor					1.003
Baseline Corr AF:	7.37	Prev response	8.09	*% change	-9.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000620	0.998544
THC Cal Offset:	0.007200	-0.000400
CH ₄ Cal Slope:	0.999873	0.999590
CH ₄ Cal Offset:	0.014400	-0.004600
NMHC Cal Slope:	1.001368	0.998095
NMHC Cal Offset:	-0.006800	0.003800

Notes: CH₄ chromatogram displayed the peak edge on the back hitting the CH₄ window timing; adjusted carrier pressure from 28.0 to 30.0 psi and calibrated channels, no further response dips noted.

Calibration Performed By: Kelly Baragar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 10, 2024	Previous Calibration:	February 27, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:30	End Time (MST):	12:08
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995			
17.23	17.20	1.0016						
8.61	8.61	1.0009				Slope	0.998544	0.90 - 1.10
4.31	4.30	1.0024						
			Intercept	-0.000400	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 10, 2024	Previous Calibration:	February 27, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:30	End Time (MST):	12:08
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
8.08	8.07	1.0007						
4.04	4.03	1.0023				Slope	0.999590	0.90 - 1.10
2.02	2.01	1.0048						
			Intercept	-0.004600	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 10, 2024	Previous Calibration:	February 27, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:30	End Time (MST):	12:08
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
9.15	9.13	1.0019						
4.57	4.58	0.9993				Slope	0.998095	0.90 - 1.10
2.29	2.29	1.0004						
			Intercept	0.003800	± 0.5			



NMHC Calibration Plot

Date: March 10, 2024

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	March 13, 2024	Last Cal Date:	March 10, 2024
Start time (MST):	8:52	End time (MST):	10:12
Reason:	Removal		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.74E-04	NA	NMHC SP Ratio:	4.75E-05
CH ₄ Retention time:	16.0	NA	NMHC Peak Area:	192745
Zero Chromatogram:	ON	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	17.23	17.12	1.006
as found 2nd point	4960	40.0	8.61	8.59	1.003
as found 3rd point	4980	20.0	4.31	4.28	1.007
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span				17.16	

Average Correction Factor					
Baseline Corr AF:	17.12	Prev response	17.20	*% change	-0.5%
Baseline Corr 2nd AF:	8.6	AF Slope:	0.994185	AF Intercept:	0.003200
Baseline Corr 3rd AF:	4.3	AF Correlation:	0.999996	<i>* = +/-5% change initiates investigation</i>	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	9.15	9.52	0.961
as found 2nd point	4960	40.0	4.57	4.75	0.964
as found 3rd point	4980	20.0	2.29	2.37	0.966
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span				9.07	
Average Correction Factor					
Baseline Corr AF:	9.52	Prev response	9.13	*% change	4.0%
Baseline Corr 2nd AF:	4.7	AF Slope:	1.040898	AF Intercept:	-0.008000
Baseline Corr 3rd AF:	2.4	AF Correlation:	0.999996	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.0	8.08	7.60	1.063
as found 2nd point	4960	40.0	4.04	3.84	1.051
as found 3rd point	4980	20.0	2.02	1.91	1.058
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span				8.09	
Average Correction Factor					
Baseline Corr AF:	7.60	Prev response	8.07	*% change	-6.2%
Baseline Corr 2nd AF:	3.84	AF Slope:	0.941063	AF Intercept:	0.012000
Baseline Corr 3rd AF:	1.91	AF Correlation:	0.999958	* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998544	
THC Cal Offset:	-0.000400	
CH ₄ Cal Slope:	0.999590	
CH ₄ Cal Offset:	-0.004600	
NMHC Cal Slope:	0.998095	
NMHC Cal Offset:	0.003800	

Notes:

Removal Calibration.

Calibration Performed By:

Mohammed Kashif

NMHC Calibration Plot

Date: March 13, 2024

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	March 13, 2024	Last Cal Date:	NA
Start time (MST):	11:08	End time (MST):	13:25
Reason:	Install		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
ZAG make/model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	NA	2.76E-04	NMHC SP Ratio:	NA	5.97E-05
CH ₄ Retention time:	NA	14.4	NMHC Peak Area:	NA	153276
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	17.23	17.22	1.000
second point	4960	40.0	8.61	8.59	1.002
third point	4980	20.0	4.31	4.30	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	17.23	17.30	0.996
Average Correction Factor					1.002

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	9.15	9.15	1.000
second point	4960	40.0	4.57	4.60	0.994
third point	4980	20.0	2.29	2.32	0.987
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	9.15	9.23	0.991
Average Correction Factor					0.994
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.0	8.08	8.07	1.001
second point	4960	40.0	4.04	3.99	1.012
third point	4980	20.0	2.02	1.98	1.021
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.0	8.08	8.06	1.002
Average Correction Factor					1.011
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		0.999917
THC Cal Offset:		-0.008000
CH ₄ Cal Slope:		1.000113
CH ₄ Cal Offset:		-0.024200
NMHC Cal Slope:		0.999769
NMHC Cal Offset:		0.016600

Notes: Install calibration. Sample inlet filter changed. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 13, 2024	Previous Calibration:	NA
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	11:08	End Time (MST):	13:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
17.23	17.22	1.0002						
8.61	8.59	1.0023				Slope	0.999917	0.90 - 1.10
4.31	4.30	1.0027						
			Intercept	-0.008000	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 13, 2024	Previous Calibration:	NA
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	11:08	End Time (MST):	13:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999953	≥0.995
8.08	8.07	1.0009			
4.04	3.99	1.0118			
2.02	1.98	1.0205			
			Slope	1.000113	0.90 - 1.10
			Intercept	-0.024200	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 13, 2024	Previous Calibration:	NA
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	11:08	End Time (MST):	13:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999984	≥ 0.995
9.15	9.15	0.9995			
4.57	4.60	0.9937			
2.29	2.32	0.9874			
			Slope	0.999769	0.90 - 1.10
			Intercept	0.016600	+/-0.5



NMHC Calibration Plot

Date: March 13, 2024

Location: Mannix





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 6, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:20	End Time (MST):	14:15
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999987	≥0.995
799.5	798.0	1.0018			
400.2	401.2	0.9976	Slope	0.997030	0.90 - 1.10
200.1	202.5	0.9882			
			Intercept	1.560821	+/-30



SO2 Calibration Plot

Date: March 7, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	March 8, 2024	Last Cal Date:	February 7, 2024
Start time (MST):	9:29	End time (MST):	13:30
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.328	ppm	Cal Gas Exp Date:	February 14, 2025
Cal Gas Cylinder #:	CC506659			
Removed Cal Gas Conc:	5.328	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994029	0.993746	Backgd or Offset:	1.95	1.95
Calibration intercept:	0.280242	0.540147	Coeff or Slope:	1.150	1.150

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4925	75.1	80.0	80.1	1.000
as found 2nd point	4963	37.5	40.0	40.3	0.994
as found 3rd point	4981	18.8	20.0	20.6	0.977
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4925	75.1	80.0	79.9	1.002
second point	4963	37.5	40.0	40.4	0.989
third point	4981	18.8	20.0	20.8	0.963
as left zero	5000	0.0	0.0	0.2	----
as left span	4925	75.1	80.0	80.9	0.989
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	0.985
Date of last converter efficiency test:	efficiency				

Baseline Corr As found:	80.0	Prev response:	79.83	*% change:	0.2%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	0.998032	AF Intercept:	0.340131
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999958		

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 8, 2024	Previous Calibration:	February 7, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:29	End Time (MST):	13:30
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999916	≥0.995
80.0	79.9	1.0016			
40.0	40.4	0.9890	Slope	0.993746	0.90 - 1.10
20.0	20.8	0.9632			
			Intercept	0.540147	+/-3



TRS Calibration Plot

Date: March 8, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Patricia McInnes	Station number:	AMS06
Calibration Date:	March 7, 2024	Last Cal Date:	February 6, 2024
Start time (MST):	10:20	End time (MST):	14:15
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH ₄ Cal Gas Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
ZAG make/model:	API T701	Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.15E-04	2.21E-04	NMHC SP Ratio:	4.71E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	192619
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.12	16.81	1.018
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.12	17.12	1.000
second point	4960	40.2	8.57	8.58	0.999
third point	4980	20.1	4.29	4.35	0.986
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.12	17.12	1.000
Average Correction Factor					0.995

Baseline Corr AF:	16.81	Prev response	17.05	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919.7	80.3	9.07	8.94	1.015
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	9.07	9.07	1.000
second point	4960	40.2	4.54	4.56	0.995
third point	4980	20.1	2.27	2.32	0.979
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	9.07	9.08	0.999
Average Correction Factor					0.991
Baseline Corr AF:	8.93	Prev response	9.04	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919.7	80.3	8.06	7.88	1.023
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919.7	80.3	8.06	8.05	1.001
second point	4960	40.2	4.03	4.02	1.004
third point	4980	20.1	2.02	2.03	0.995
as left zero	5000	0.0	0.00	0.00	----
as left span	4919.7	80.3	8.06	8.04	1.001
Average Correction Factor					1.000
Baseline Corr AF:	7.88	Prev response	8.00	*% change	-1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.994828	0.998467
THC Cal Offset:	0.012664	0.026638
CH ₄ Cal Slope:	0.993816	0.998356
CH ₄ Cal Offset:	-0.003140	0.002848
NMHC Cal Slope:	0.995576	0.998464
NMHC Cal Offset:	0.016404	0.024190

Notes: Changed the inlet filter and reconnected the instrument to a H2 generator. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 6, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:20	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995			
17.12	17.12	1.0003						
8.57	8.58	0.9993				Slope	0.998467	0.90 - 1.10
4.29	4.35	0.9864						
			Intercept	0.026638	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 6, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:20	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999993	≥ 0.995			
8.06	8.05	1.0010						
4.03	4.02	1.0037				Slope	0.998356	0.90 - 1.10
2.02	2.03	0.9948						
			Intercept	0.002848	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 6, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	10:20	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999968	≥ 0.995
9.07	9.07	0.9999			
4.54	4.56	0.9952			
2.27	2.32	0.9791			
			Slope	0.998464	0.90 - 1.10
			Intercept	0.024190	+/-0.5



NMHC Calibration Plot

Date: March 7, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Patricia McInnes
Calibration Date: March 5, 2024
Start time (MST): 9:58
Reason: Routine
Station number: AMS06
Last Cal Date: February 5, 2024
End time (MST): 14:41

Calibration Standards

NO Gas Cylinder #: T30YCWN
NOX Cal Gas Conc: 47.94 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 47.94 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: April 11, 2025
NO Cal Gas Conc: 46.39 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 46.39 ppm
NO gas Diff:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1172750022

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.825	0.825	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.987	0.987	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.2	157.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000023	1.002416
NO _x Cal Offset:	2.315635	1.975128
NO Cal Slope:	1.002417	1.003032
NO Cal Offset:	1.022423	1.382192
NO ₂ Cal Slope:	0.999140	0.999602
NO ₂ Cal Offset:	-0.524585	-0.538439



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	0.2	-0.4	----	----
as found span	4914	86.2	826.5	799.7	26.7	832.5	804.7	27.8	0.9927	0.9938
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.4	-0.4	----	----
high point	4914	86.2	826.5	799.7	26.7	829.5	803.0	26.6	0.9963	0.9959
second point	4957	43.1	413.2	399.9	13.4	417.0	403.1	13.9	0.9910	0.9920
third point	4978	21.6	207.1	200.4	6.7	211.7	203.3	8.4	0.9784	0.9858
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.3	-0.4	----	----
as left span	4914	86.2	826.5	399.1	427.3	827.0	402.2	424.8	0.9993	0.9924
Average Correction Factor									0.9885	0.9913

Corrected As found	NO _x = 832.7 ppb	NO = 804.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.5%
Previous Response	NO _x = 828.8 ppb	NO = 802.7 ppb		*Percent Change	NO = 0.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	801.3	400.7	427.3	426.7	1.0015	99.9%
2nd GPT point (200 ppb O3)	801.3	604.0	224.0	223.3	1.0032	99.7%
3rd GPT point (100 ppb O3)	801.3	703.3	124.7	124.0	1.0058	99.4%
Average Correction Factor					1.0035	99.7%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 5, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:58	End Time (MST):	14:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
826.5	829.5	0.9963		
413.2	417.0	0.9910		
207.1	211.7	0.9784		





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 5, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:58	End Time (MST):	14:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.7	803.0	0.9959		
399.9	403.1	0.9920		
200.4	203.3	0.9858		





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 5, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:58	End Time (MST):	14:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
427.3	426.7	1.0015		
224.0	223.3	1.0032		
124.7	124.0	1.0058		



NO_x Calibration Plot

Date: March 5, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Patricia McInnes Station number: AMS06
 Calibration Date: March 22, 2024 Last Cal Date: February 16, 2024
 Start time (MST): 9:25 End time (MST): 12:32
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: API T700 Serial Number: 3566
 ZAG Make/Model: API T701 Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1300156234
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003629	1.004200	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	-0.860000	-0.660000	Coeff or Slope:	1.026	1.026

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.0	----
as found span	5000	1303.0	400.0	401.0	0.998
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.2	----
high point	5000	1303.0	400.0	401.4	0.997
second point	5000	966.5	200.0	199.9	1.001
third point	5000	794.3	100.0	98.8	1.012
as left zero	5000	800.0	0.0	0.2	----
as left span	5000	1303.0	400.0	401.8	0.996
Average Correction Factor					1.003

Baseline Corr As found:	401.0	Previous response	400.6	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 22, 2024	Previous Calibration:	February 16, 2024
Station Name:	Patricia McInnes	Station Number:	AMS06
Start Time (MST):	9:25	End Time (MST):	12:32
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999979	
400.0	401.4	0.9965			≥0.995
200.0	199.9	1.0005	Slope	1.004200	
100.0	98.8	1.0121			0.90 - 1.10
			Intercept	-0.660000	+/- 5



O₃ Calibration Plot

Date: March 22, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: March 22, 2024 Last Cal Date: February 16, 2024
 Start time (MST): 12:49 End time (MST): 13:55

Analyzer Make: API T640 S/N: 766
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388755
 Temp/RH standard: Alicat FP-25BT S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-7.9	-8.2	-7.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	737.9	740.4	737.9	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.12	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	42	----	42	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	7.3	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: 11-23-2023
 Lot No.: 100128-050-035

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.9	10.9	10.9	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ March 22, 2024
 Date Disposable Filter Changed: _____ March 22, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ April 13, 2023
 Date RH/T Sensor Cleaned: _____ April 13, 2023

Notes: Quarterly calibrations complete. Leak check passed, no adjustments made.

Calibration by: Max Farrell



Wood Buffalo Environmental Association

TN - NO_x - NH₃ Calibration Report

Version-05-2023

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	March 6, 2024	Last Cal Date:	February 21, 2024
Start time (MST):	9:40	End time (MST):	14:00
NH3 Cal Date:	March 6, 2024	Last Cal Date:	February 21, 2024
Start time (MST):	13:30	End time (MST):	16:00
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	NA
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.29	ppm	NH3 Gas Cylinder #:	EB0108520
Removed NH3 Conc:	76.29	ppm	NH3 Cal Gas Expiry:	August 22, 2024
NH3 gas Diff:			Removed Cylinder #:	NA
Calibrator Model:	API T700		Removed cyl Expiry:	NA
ZAG make/model:	API T701		Serial Number:	3566
			Serial Number:	4602

Analyzer Information

Analyzer model:	Teledyne API T201	Analyzer serial #:	808
Converter model:	Teledyne API T501	Converter serial #:	484
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	4.80
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	26.6

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.854	0.854	TN coefficient:	0.849	0.849
NOX coefficient:	0.848	0.848	NO bkgrnd:	-0.985	-0.985
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	-0.562	-0.562
NH3 coefficient:	0.891	0.896	TN bkgrnd:	5.018	5.018

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995140	0.992580
NO _x Cal Offset:	2.256598	1.257688
NO Cal Slope:	0.997671	0.996357
NO Cal Offset:	1.183452	-0.056315
NO ₂ Cal Slope:	0.997218	1.001744
NO ₂ Cal Offset:	0.194612	-0.225723
NH3 Cal Slope:	1.010551	1.000069
NH3 Cal Offset:	1.110139	3.359353
TN Cal Slope:	1.017335	1.006770
TN Cal Offset:	2.438080	4.605099



Wood Buffalo Environmental Association

TN - NOX - NH₃ Calibration Report

Version-05-2023

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated TN concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated TN concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	TN Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-2.1	0.0	-2.0	----	----
as found NO	4914	86.2	826.5	826.5	----	824.9	822.4	2.5	1.002	----
calibrator zero	5000	0.0	0.0	0.0	0.0	-1.0	0.2	-1.2	----	----
high NO point	4914	86.2	826.5	826.5	----	825.9	821.0	4.8	1.001	----
NO/O3 point	4914	86.2	826.5	826.5	----	824.0	820.3	3.5	1.003	----
as found NH3	3417	82.6	1800.6	----	1800.6	1824.0	----	1811.0	0.987	0.994
new NH3 cyl rp	3417	82.6	1800.6	----	1800.6	----	----	----	----	----
first NH3	3417	82.6	1800.6	----	1800.6	1814.1	----	1801.4	0.993	1.000
second NH3	3454	45.9	1000.5	----	1000.5	1015.0	----	1006.5	0.986	0.994
third NH3	3477	22.9	499.2	----	499.2	513.0	----	507.3	0.973	0.984
Average Correction Factor									1.0018	0.9925

Corrected As found TN = 827 ppb NO_x = 822.4 ppb NH3 = 1813.0 ppb

Previous Response TN = 843.2 ppb NO_x = 824.7 ppb NH3 = 1820.8 ppb

NH3 Previous Converter Efficiency = 89.1%

NH3 Current Converter Efficiency = 89.6%

*Percent Change TN = -2.0%

*Percent Change NO_x = -0.3%

*Percent Change NH3 = -0.4%

* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

TN Calibration Summary


Version-05-2023

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 21, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:40	End Time (MST):	14:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-1.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
1800.6	1814.1	0.9926		
1000.5	1015.0	0.9857		
499.2	513.0	0.9730		





Wood Buffalo Environmental Association

NH₃ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 21, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:40	End Time (MST):	14:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-1.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
1800.6	1801.4	0.9996		
1000.5	1006.5	0.9941		
499.2	507.3	0.9840		
			0.999968	
			1.000069	
			3.359353	





Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-05-2023

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 21, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:40	End Time (MST):	14:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
826.5	821.0	1.0066			
413.2	412.1	1.0028			
207.1	207.8	0.9967			
			Slope	0.992580	0.90 - 1.10
			Intercept	1.257688	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-05-2023


Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 21, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:40	End Time (MST):	14:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	≥0.995	
799.7	797.6	1.0027			
399.9	396.4	1.0088			
200.4	200.2	1.0011			
			Slope	0.996357	0.90 - 1.10
			Intercept	-0.056315	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-05-2023

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 21, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:40	End Time (MST):	14:00
Analyzer make:	Teledyne API T201	Analyzer serial #:	808

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
425.6	426.1	0.9989			
219.5	219.7	0.9992			
123.4	123.5	0.9994			
			Slope	1.001744	0.90 - 1.10
			Intercept	-0.225723	+/-20



NO_x Calibration Plot

Date: March 6, 2024


Location: Patricia McInnes



NH₃ Calibration Plot

Date: March 6, 2024

Location: Patricia McInnes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 22, 2024	Previous Calibration:	February 2, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:41	End Time (MST):	13:02
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999967	
799.0	799.8	0.9990			≥0.995
399.5	404.1	0.9886	Slope	1.000577	
200.2	202.4	0.9893			0.90 - 1.10
			Intercept	1.744314	+/-30



SO2 Calibration Plot

Date: March 22, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: March 7, 2024 Last Cal Date: February 22, 2024
 Start time (MST): 10:12 End time (MST): 15:05
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.25 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC504080
 Removed Cal Gas Conc: 5.25 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3805
 ZAG Make/Model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE Analyzer serial #: 1180540018
 Converter make: CDN-101 Converter serial #: 551
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.017941	0.998195	Backgd or Offset:	2.40	2.40
Calibration intercept:	0.157770	0.217844	Coeff or Slope:	0.901	0.901

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4925	75.5	79.3	79.3	1.001
as found 2nd point	4962	37.7	39.6	39.7	1.000
as found 3rd point	4981	18.9	19.8	20.3	0.982
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4925	75.5	79.3	79.4	0.999
second point	4962	37.7	39.6	39.8	0.995
third point	4981	18.9	19.9	19.9	0.998
as left zero	5000	0.0	0.0	0.5	----
as left span	4925	75.5	79.3	79.1	1.003
SO2 Scrubber Check	4920	79.2	800.4	0.4	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	0.997
Date of last converter efficiency test:	April 22, 2022			92.6% efficiency	

Baseline Corr As found: 79.2 Prev response: 80.85 *% change: -2.1%
 Baseline Corr 2nd AF pt: 39.6 AF Slope: 0.997616 AF Intercept: 0.257692
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999974

* = > +/-5% change initiates investigation

Notes: No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 22, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:12	End Time (MST):	15:05
Analyzer make:	CDN-101	Analyzer serial #:	551

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999992	≥0.995
79.3	79.4	0.9989			
39.6	39.8	0.9952	Slope	0.998195	0.90 - 1.10
19.9	19.9	0.9978			
			Intercept	0.217844	+/-3



TRS Calibration Plot

Date: March 7, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	March 22, 2024	Last Cal Date:	February 20, 2024
Start time (MST):	9:41	End time (MST):	13:02
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3805
ZAG make/model:	API 701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.45E-04	2.56E-04	NMHC SP Ratio:	4.93E-05
CH ₄ Retention time:	13.4	13.4	NMHC Peak Area:	182375
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	16.91	16.26	1.040
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	16.91	16.86	1.003
second point	4960	39.9	8.46	8.42	1.005
third point	4980	20.0	4.24	4.23	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	16.91	16.85	1.004

Average Correction Factor				1.004
Baseline Corr AF:	16.26	Prev response	16.96	*% change -4.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.00	8.68	1.037
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.00	8.96	1.005
second point	4960	39.9	4.50	4.49	1.003
third point	4980	20.0	2.26	2.26	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.00	8.94	1.006
Average Correction Factor					1.003
Baseline Corr AF:	8.68	Prev response	9.05	*% change	-4.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	7.92	7.58	1.044
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	7.92	7.90	1.002
second point	4960	39.9	3.96	3.93	1.007
third point	4980	20.0	1.98	1.97	1.007
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	7.92	7.91	1.001
Average Correction Factor					1.005
Baseline Corr AF:	7.58	Prev response	7.91	*% change	-4.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001667	0.996773
THC Cal Offset:	0.015848	-0.002926
CH ₄ Cal Slope:	0.999012	0.998463
CH ₄ Cal Offset:	0.001263	-0.008332
NMHC Cal Slope:	1.004116	0.995057
NMHC Cal Offset:	0.014385	0.005807

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 22, 2024	Previous Calibration:	February 20, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:41	End Time (MST):	13:02
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
16.91	16.86	1.0032						
8.46	8.42	1.0047				Slope	0.996773	0.90 - 1.10
4.24	4.23	1.0033						
			Intercept	-0.002926	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 22, 2024	Previous Calibration:	February 20, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:41	End Time (MST):	13:02
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999992	≥0.995			
7.92	7.90	1.0017						
3.96	3.93	1.0069				Slope	0.998463	0.90 - 1.10
1.98	1.97	1.0071						
			Intercept	-0.008332	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 22, 2024	Previous Calibration:	February 20, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:41	End Time (MST):	13:02
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995			
9.00	8.96	1.0047						
4.50	4.49	1.0029				Slope	0.995057	0.90 - 1.10
2.26	2.26	1.0000						
			Intercept	0.005807	± 0.5			



NMHC Calibration Plot

Date: March 22, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS 07
Calibration Date:	March 26, 2024	Last Cal Date:	March 22, 2024
Start time (MST):	9:00	End time (MST):	13:10
Reason:	Maintenance Investigation of intermittent baseline drop to zero		

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	496.0 ppm	CH4 Equiv Conc.	1059.8 ppm
C3H8 Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	496.0 ppm	CH4 Equiv Conc.	1059.8 ppm
Removed C3H8 Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3805
Zero Air Gen model:	Teledyne API T701H	Serial Number:	198

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 12227620777
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.56E-04	2.55E-04	NMHC SP Ratio: 5.14E-05	5.32E-05
CH4 Retention time:	13.4	13.4	NMHC Peak Area: 175223	169129
Zero Chromatogram:	OFF	OFF	Flat Baseline: OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.91	16.87	1.003
As found Mid point	4960	39.9	8.46	8.42	1.004
As found Low point	4980	20.0	4.24	4.25	0.998
New cylinder response					
Baseline Corr AF:	16.87	Prev response	16.86	*% change	0.1%
Baseline Corr 2nd AF:	8.42	AF Slope:	0.996733	AF Intercept:	0.005868
Baseline Corr 3rd AF:	4.25	AF Correlation:	0.999997	* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.91	16.96	0.997
Mid point	4960	39.9	8.46	8.50	0.995
Low point	4980	20.0	4.24	4.28	0.990
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.91	17.07	0.991
Average Correction Factor					0.994

Notes: Instrument was dipping from baseline readings to zero; replaced actuator.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.00	8.96	1.005
As found Mid point	4960	39.9	4.50	4.49	1.003
As found Low point	4980	20.0	2.26	2.27	0.992
New cylinder response					
Baseline Corr AF:	8.96	Prev response	8.96	*% change	0.0%
Baseline Corr 2nd AF:	4.49	AF Slope:	0.994360	AF Intercept:	0.012803
Baseline Corr 3rd AF:	2.27	AF Correlation:	0.999989	* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.00	9.04	0.995
Mid point	4960	39.9	4.50	4.56	0.986
Low point	4980	20.0	2.26	2.31	0.975
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.00	9.18	0.980
Average Correction Factor					0.985

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	7.92	7.91	1.001
As found Mid point	4960	39.9	3.96	3.94	1.005
As found Low point	4980	20.0	1.98	1.97	1.005
New cylinder response					
Baseline Corr AF:	7.91	Prev response	7.90	*% change	0.2%
Baseline Corr 2nd AF:	3.94	AF Slope:	0.999430	AF Intercept:	-0.006935
Baseline Corr 3rd AF:	1.97	AF Correlation:	0.999994	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	7.92	7.92	1.000
Mid point	4960	39.9	3.96	3.93	1.006
Low point	4980	20.0	1.98	1.97	1.008
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	7.92	7.89	1.003
Average Correction Factor					1.005

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.996773	1.002451
THC Cal Offset:	-0.002926	0.015044
CH ₄ Cal Slope:	0.998463	1.000657
CH ₄ Cal Offset:	-0.008332	-0.011935
NMHC Cal Slope:	0.995057	1.003673
NMHC Cal Offset:	0.005807	0.027380

Calibration Performed By: Kelly Baragar



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	March 22, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:00	End Time (MST):	13:10
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999996	≥0.995
16.91	16.96	0.9971	Slope	1.002451	0.90 - 1.10
8.46	8.50	0.9952	Intercept	0.015044	+/-0.5
4.24	4.28	0.9902			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	March 22, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:00	End Time (MST):	13:10
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999985 ≥0.995
7.92	7.92	0.9997	Slope	1.000657 0.90 - 1.10
3.96	3.93	1.0064	Intercept	-0.011935 +/-0.5
1.98	1.97	1.0081		





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	March 22, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:00	End Time (MST):	13:10
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999956 ≥0.995
9.00	9.04	0.9951	Slope	1.003673 0.90 - 1.10
4.50	4.56	0.9857	Intercept	0.027380 +/-0.5
2.26	2.31	0.9749		



NMHC Calibration Plot

Date: March 26, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
as found span	4933	66.8	803.0	800.3	2.7	795.6	799.0	-3.5	1.0093	1.0016
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
high point	4933	66.8	803.0	800.3	2.7	803.3	799.0	4.3	0.9996	1.0016
second point	4966	33.4	401.5	400.2	1.3	405.3	402.8	2.5	0.9907	0.9935
third point	4983	16.7	200.7	200.1	0.7	203.3	201.7	1.6	0.9874	0.9920
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as left span	4933	66.8	803.0	400.9	402.1	808.8	399.0	409.7	0.9928	1.0048
Average Correction Factor									0.9926	0.9957

Corrected As found	NO _x = 795.8 ppb	NO = 799.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.7%
Previous Response	NO _x = 801.5 ppb	NO = 802.6 ppb		*Percent Change	NO = -0.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	796.5	397.1	402.1	405.9	0.9906	101.0%
2nd GPT point (200 ppb O3)	796.5	599.0	200.2	204.2	0.9803	102.0%
3rd GPT point (100 ppb O3)	796.5	698.5	100.7	103.7	0.9708	103.0%
Average Correction Factor					0.9805	102.0%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 27, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:14	End Time (MST):	15:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
803.0	803.3	0.9996			
401.5	405.3	0.9907			
200.7	203.3	0.9874			
			Slope	0.999973	0.90 - 1.10
			Intercept	1.651897	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 27, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:14	End Time (MST):	15:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	799.0	1.0016		
400.2	402.8	0.9935		
200.1	201.7	0.9920		





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 27, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:14	End Time (MST):	15:26
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
402.1	405.9	0.9906		
200.2	204.2	0.9803		
100.7	103.7	0.9708		



NO_x Calibration Plot

Date: March 5, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Athabasca Valley Station number: AMS07
 Calibration Date: March 27, 2024 Last Cal Date: February 16, 2024
 Start time (MST): 6:05 End time (MST): 9:23
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: T700 Serial Number: 3805
 ZAG Make/Model: T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i Analyzer serial #: 1152220023
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993286	0.993629	Backgd or Offset:	-1.6	-1.6
Calibration intercept:	0.100000	1.040000	Coeff or Slope:	1.549	1.549

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-1.0	----
as found span	5000	1522.8	400.0	397.5	1.006
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.7	----
high point	5000	1523.6	400.0	398.4	1.004
second point	5000	1088.1	200.0	199.7	1.002
third point	5000	880.5	100.0	100.9	0.991
as left zero	5000	0.0	0.0	-1.2	----
as left span	5000	1522.4	400.0	398.7	1.003
Average Correction Factor					0.999

Baseline Corr As found:	398.5	Previous response	397.4	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments or maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 27, 2024	Previous Calibration:	February 16, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	6:05	End Time (MST):	9:23
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient	0.999996	≥0.995
400.0	398.4	1.0040			
200.0	199.7	1.0015	Slope	0.993629	0.90 - 1.10
100.0	100.9	0.9911			
			Intercept	1.040000	+/- 5



O₃ Calibration Plot

Date: March 27, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Athabasca Valley Station number: AMS 07
 Calibration Date: March 27, 2024 Last Cal Date: February 27, 2024
 Start time (MST): 5:27 End time (MST): 6:12

Analyzer Make: S/N: 645
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-5.4	-5.8	-5.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	735.9	735	735.9	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	4.8	5.02	<input checked="" type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	----	37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	4.0	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: October 6, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 27, 2024
 Date Disposable Filter Changed: February 27, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022
 Date RH/T Sensor Cleaned: December 5, 2022

Notes: Flow Adjusted. Leak check passed.

Calibration by: Melissa Lemay



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	March 27, 2024	Last Cal Date:	February 23, 2024
Start time (MST):	9:20	End time (MST):	12:19
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date:	December 12, 2026
Cal Gas Cylinder #:	LL66942			
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API 700H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996688	0.995632	Backgd or Offset:	4.425	4.642
Calibration intercept:	0.062533	0.048509	Coeff or Slope:	1.087	1.087

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4933	66.7	40.0	39.9	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4933	66.7	40.0	39.9	1.004
second point	4967	33.3	20.0	20.0	0.999
third point	4983	16.7	10.0	10.1	0.997
as left zero	5000	0.0	0.0	0.0	----
as left span	4933	66.7	40.0	39.7	1.008
Average Correction Factor					1.000

Baseline Corr As found:	39.82	Prev response:	39.95	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Zero adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 27, 2024	Previous Calibration:	February 23, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:20	End Time (MST):	12:19
Analyzer make:	Thermo 48i-LTE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999991	
40.0	39.9	1.0041			≥0.995
20.0	20.0	0.9989	Slope	0.995632	
10.0	10.1	0.9971			0.90 - 1.10
			Intercept	0.048509	+/-1.5



CO Calibration Plot

Date: March 27, 2024

Location: Athabasca Valley





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	March 20, 2024	Last Cal Date:	February 15, 2024
Start time (MST):	12:08	End time (MST):	14:47
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC196697			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
ZAG Make/Model:	Teledyne API T701		Serial Number:	135

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994207	1.003858	Backgd or Offset:	1.83	1.83
Calibration intercept:	1.596267	0.615265	Coeff or Slope:	0.989	0.989

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	4920	80.3	800.4	798.5	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	-0.2	----
high point	4920	80.3	800.4	804.3	0.995
second point	4960	40.2	400.7	401.4	0.998
third point	4980	20.1	200.4	203.8	0.983
as left zero	5000	0.0	0.0	-0.3	----
as left span	4920	80.3	800.4	802.6	0.997
Average Correction Factor					0.992

Baseline Corr As found:	798.90	Previous response	797.34	*% change	0.2%
-------------------------	--------	-------------------	--------	-----------	------

* = > +/-5% change initiates investigation

Notes: Changed out inlet filter after as found. No adjustment made.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 15, 2024
Station Name:	December 13, 2023	Station Number:	Fort Chipewyan
Start Time (MST):	12:08	End Time (MST):	14:47
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.2	----	Correlation Coefficient	0.999980	≥0.995
800.4	804.3	0.9951			
400.7	401.4	0.9983	Slope	1.003858	0.90 - 1.10
200.4	203.8	0.9831			
			Intercept	0.615265	+/-30



SO2 Calibration Plot

Date: March 20, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: March 20, 2024 Last Cal Date: February 13, 2024
 Start time (MST): 14:46 End time (MST): 18:33
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: February 9, 2024
 Cal Gas Cylinder #: EY0002276
 Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 1203169744
 Converter make: CDN-101 Converter serial #: 14639
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.991000	1.003284	Backgd or Offset:	0.98	0.99
Calibration intercept:	0.578560	0.498794	Coeff or Slope:	0.741	0.754

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.5	----
as found span	4920	80.5	80.0	78.5	1.026
as found 2nd point	4960	40.2	40.0	39.9	1.014
as found 3rd point	4980	20.1	20.0	19.9	1.030
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.6	----
high point	4920	80.5	80.0	80.8	0.990
second point	4960	40.2	40.0	40.6	0.984
third point	4980	20.1	20.0	20.4	0.979
as left zero	5000	0.0	0.0	0.7	----
as left span	4920	80.5	80.0	81.4	0.983
SO2 Scrubber Check	4919.7	80.3	803.0	0.1	----
Date of last scrubber change:	March 7, 2022			Ave Corr Factor	0.985
Date of last converter efficiency test:	March 15, 2022			100.7% efficiency	

Baseline Corr As found: 78.0 Prev response: 79.87 *% change: -2.4%
 Baseline Corr 2nd AF pt: 39.4 AF Slope: 0.975855 AF Intercept: 0.558435
 Baseline Corr 3rd AF pt: 19.4 AF Correlation: 0.999950

* = > +/-5% change initiates investigation

Notes: Sampled inlet filter changed after as founds. Scrubber check passed. Adjusted span.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 13, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	14:46	End Time (MST):	18:33
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999991	≥0.995
80.0	80.8	0.9902			
40.0	40.6	0.9842	Slope	1.003284	0.90 - 1.10
20.0	20.4	0.9794			
			Intercept	0.498794	+/-3



TRS Calibration Plot

Date: March 20, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
Calibration Date: March 20, 2024 Last Cal Date: February 21, 2024
Start time (MST): 7:54 End time (MST): 12:13
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0046831 Cal Gas Expiry Date: January 9, 2032
NOX Cal Gas Conc: 60.20 ppm NO Cal Gas Conc: 60.00 ppm
Removed Cylinder #: DT0046831 Removed Gas Exp Date: January 9, 2032
Removed Gas NOX Conc: 60.20 ppm Removed Gas NO Conc: 60.00 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 3252
ZAG make/model: Teledyne API T701H Serial Number: 135

Analyzer Information

Analyzer make: API T200 Analyzer serial #: 4460
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.154	1.200	NO bkgnd or offset:	-2.5	-2.5
NOX coeff or slope:	1.152	1.194	NOX bkgnd or offset:	-2.2	-2.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	2.9	2.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997032	0.995037
NO _x Cal Offset:	2.294675	3.395408
NO Cal Slope:	0.998455	0.998540
NO Cal Offset:	1.134987	1.155021
NO ₂ Cal Slope:	1.001043	0.993301
NO ₂ Cal Offset:	0.109543	0.722793



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	1.2	1.2	0.0	----	----
as found span	4933	66.7	803.1	800.4	2.7	774.0	767.4	6.3	1.0376	1.0431
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	1.3	1.4	-0.1	----	----
high point	4933	66.7	803.1	800.4	2.7	801.1	800.3	0.8	1.0025	1.0002
second point	4967	33.3	400.9	399.6	1.3	404.5	400.8	4.0	0.9911	0.9969
third point	4983	16.7	201.1	200.4	0.7	204.8	200.5	4.3	0.9818	0.9995
as left zero	5000	0.0	0.0	0.0	0.0	1.3	1.3	0.0	----	----
as left span	4933	66.7	803.1	391.1	412.0	785.6	377.7	407.9	1.0223	1.0356
Average Correction Factor									0.9918	0.9989

Corrected As found	NO _x = 772.8 ppb	NO = 766.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -3.9%
Previous Response	NO _x = 803.0 ppb	NO = 800.3 ppb		*Percent Change	NO = -4.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	790.9	381.6	412.0	409.2	1.0068	99.3%
2nd GPT point (200 ppb O3)	790.9	588.1	205.5	206.2	0.9965	100.4%
3rd GPT point (100 ppb O3)	790.9	695.1	98.5	98.7	0.9977	100.2%
Average Correction Factor					1.0003	100.0%

Notes: Sample inlet filters changed after as found. Adjusted span.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 21, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	7:54	End Time (MST):	12:13
Analyzer make:	API T200	Analyzer serial #:	12:13:00 PM

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	1.3	----	Correlation Coefficient	0.999963	≥0.995
803.1	801.1	1.0025			
400.9	404.5	0.9911			
201.1	204.8	0.9818			
			Slope	0.995037	0.90 - 1.10
			Intercept	3.395408	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 21, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	7:54	End Time (MST):	12:13
Analyzer make:	API T200	Analyzer serial #:	12:13:00 PM

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	1.4	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
800.4	800.3	1.0002		
399.6	400.8	0.9969		
200.4	200.5	0.9995		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 21, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	7:54	End Time (MST):	12:13
Analyzer make:	API T200	Analyzer serial #:	12:13:00 PM

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
412.0	409.2	1.0068		
205.5	206.2	0.9965		
98.5	98.7	0.9977		
			0.999966	
			0.993301	
			0.722793	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 20, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort Chipewyan Station number: AMS08
 Calibration Date: March 19, 2024 Last Cal Date: February 6, 2024
 Start time (MST): 7:14 End time (MST): 10:05
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3252
 ZAG Make/Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3872
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.027914	0.994571	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	0.240000	0.100000	Coeff or Slope:	1.036	1.036

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	5000	913.0	400.0	399.0	1.003
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	0.3	----
high point	5000	914.7	400.0	397.9	1.005
second point	5000	786.4	200.0	199.3	1.004
third point	5000	701.3	100.0	99.1	1.009
as left zero	5000	0.0	0.0	1.1	----
as left span	5000	963.3	400.0	398.0	1.005
Average Correction Factor					1.006

Baseline Corr As found:	398.8	Previous response	411.4	*% change	-3.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed out inlet filter after as found. No adjustments made. During as founds entered PPB point instead of GTPS point.

Calibration Performed By: Matthew C



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 19, 2024	Previous Calibration:	February 6, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	7:14	End Time (MST):	10:05
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999996	≥0.995
400.0	397.9	1.0053			
200.0	199.3	1.0035	Slope	0.994571	0.90 - 1.10
100.0	99.1	1.0091			
			Intercept	0.100000	+/- 5



O₃ Calibration Plot

Date: March 19, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
 Calibration Date: March 20, 2024 Last Cal Date: February 20, 2024
 Start time (MST): 13:10 End time (MST): 14:03

Analyzer Make: Teledyne API T640 S/N: 319
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-12.90	-12.45	-12.90	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	751.10	752.5	751.10	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.01	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	42%	----	42%	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.40	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: 10-Jun-24
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 20, 2024
 Date Disposable Filter Changed: February 20, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 25, 2023
 Date RH/T Sensor Cleaned: July 25, 2023

Notes: No adjustments needed.

Calibration by: Morgan Voyageur



Wood Buffalo Environmental Association

CO Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	March 19, 2024	Last Cal Date:	February 6, 2024
Start time (MST):	14:47	End time (MST):	17:32
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3252
ZAG Make/Model:	API T701H		Serial Number:	135

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994000	0.997054	Backgd or Offset:	-0.014	-0.015
Calibration intercept:	0.374934	0.100939	Coeff or Slope:	0.998	1.007

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.36	----
as found span	4933	66.7	40.4	40.8	0.991
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.0	----
high point	4934	66.7	40.4	40.3	1.002
second point	4967	33.3	20.2	20.4	0.992
third point	4983	16.7	10.1	10.2	0.992
as left zero	5000	0.0	0.0	0.0	----
as left span	2960	40.0	40.4	40.0	1.010
Average Correction Factor					0.995

Baseline Corr As found:	40.44	Prev response:	40.56	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Adjustments made to span and zero.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 19, 2024	Previous Calibration:	February 6, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	14:47	End Time (MST):	17:32
Analyzer make:	API T300	Analyzer serial #:	3505

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999971	
40.4	40.3	1.0021			≥ 0.995
20.2	20.4	0.9916	Slope	0.997054	
10.1	10.2	0.9922			$0.90 - 1.10$
			Intercept	0.100939	± 1.5



CO Calibration Plot

Date: March 19, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

CO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	March 19, 2024	Last Cal Date:	February 6, 2024
Start time (MST):	12:16	End time (MST):	14:46
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
N2 Gen Make/Model:	NG 5000		Serial Number:	135

Analyzer Information

Analyzer make: Teledyne API T360	Analyzer serial #: 289
Analyzer Range 0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000467	1.011498	Backgd or Offset:	-0.063
Calibration intercept:	-10.320000	-16.020000	Coeff or Slope:	1.094

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	3000	0.0	0.0	-4.9	----
as found span	2920	80.0	1605.9	1614.8	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	3000	0.0	0.0	-4.9	----
high point	2920	80.0	1605.9	1621.2	0.991
second point	2960	40.0	802.9	768.6	1.045
third point	2980	20.0	401.5	393.6	1.020
as left zero	3000	0.0	0.0	-4.8	----
as left span	2960	40.0	802.9	766.9	1.047
Average Correction Factor					1.018

Baseline Corr As found:	1619.70	Prev response:	1596.30	*% change:	1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as found. No adjustments made.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	March 19, 2024	Previous Calibration	February 6, 2024
Station Name	Fort Chipewyan	Station Number	AMS08
Start Time (MST)	12:16	End Time (MST)	14:46
Analyzer make	Teledyne API T360	Analyzer serial #	289

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-4.9	----	Correlation Coefficient	0.999265	≥0.995
1605.9	1621.2	0.9905			
802.9	768.6	1.0447	Slope	1.011498	0.90 - 1.10
401.5	393.6	1.0200			
			Intercept	-16.020000	+/-20



CO₂ Calibration Plot

Date: March 19, 2024

Location: Fort Chipewyan





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	March 6, 2024	Last Cal Date:	February 16, 2024
Start time (MST):	10:04	End time (MST):	13:14
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.96	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC151285			
Removed Cal Gas Conc:	49.96	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3812
ZAG Make/Model:	API T701		Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999855	1.001110	Backgd or Offset:	10.2	10.2
Calibration intercept:	0.931270	0.391536	Coeff or Slope:	0.963	0.963

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.3	----
as found span	4919	80.2	801.5	802.3	0.999
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.3	----
high point	4919	80.2	801.5	802.9	0.998
second point	4959	40.1	400.8	401.1	0.999
third point	4980	20.0	199.8	200.9	0.995
as left zero	5000	0.0	0.0	0.1	----
as left span	4919	80.2	801.5	803.2	0.998
Average Correction Factor					0.997

Baseline Corr As found:	802.60	Previous response	802.30	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 16, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:04	End Time (MST):	13:14
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995
801.5	802.9	0.9982			
400.8	401.1	0.9991	Slope	1.001110	0.90 - 1.10
199.8	200.9	0.9947			
			Intercept	0.391536	+/-30



SO2 Calibration Plot

Date: March 6, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Barge Landing Station number: AMS09
 Calibration Date: March 7, 2024 Last Cal Date: February 1, 2024
 Start time (MST): 10:14 End time (MST): 14:30
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.171 ppm Cal Gas Exp Date: August 22, 2026
 Cal Gas Cylinder #: CC511415
 Removed Cal Gas Conc: 5.171 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3812
 ZAG Make/Model: API T701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1331259320
 Converter make: CDN-101 Converter serial #: 519
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999264	1.000835	Backgd or Offset: 2.93	2.84
Calibration intercept:	0.119396	0.139405	Coeff or Slope: 1.201	1.170

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4923	77.4	80.0	81.5	0.982
as found 2nd point	4961	38.7	40.0	41.0	0.976
as found 3rd point	4981	19.3	20.0	20.3	0.983
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4923	77.4	80.0	80.2	0.998
second point	4961	38.7	40.0	40.3	0.993
third point	4981	19.3	20.0	20.1	0.993
as left zero	5000	0.0	0.0	0.1	----
as left span	4923	77.4	80.0	80.4	0.996
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:	28-Feb-23			Ave Corr Factor	0.995
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 81.5 Prev response: 80.10 *% change: 1.7%
 Baseline Corr 2nd AF pt: 41.0 AF Slope: 1.018680 AF Intercept: 0.039717
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999987

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. SOx scrubber check done after calibrator zero.
 Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 1, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:14	End Time (MST):	14:30
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999996	
80.0	80.2	0.9980			≥0.995
40.0	40.3	0.9932	Slope	1.000835	
20.0	20.1	0.9930			0.90 - 1.10
			Intercept	0.139405	+/-3



TRS Calibration Plot

Date: March 7, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	March 6, 2024	Last Cal Date:	February 16, 2024
Start time (MST):	10:04	End time (MST):	13:14
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.58E-04	2.58E-04	NMHC SP Ratio:	4.37E-05	4.37E-05
CH ₄ Retention time:	15.00	15.00	NMHC Peak Area:	209298	209298
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	17.09	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	17.04	1.004
second point	4960	40.1	8.56	8.50	1.007
third point	4980	20.0	4.27	4.26	1.002
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	17.12	17.05	1.004
				CF Correction Factor	1.004

Baseline Corr AF:	17.09	Prev response	17.12	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	9.14	9.12	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	9.08	1.006
second point	4960	40.1	4.57	4.53	1.009
third point	4980	20.0	2.28	2.27	1.003
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	9.14	9.09	1.005
Average Correction Factor					1.006
Baseline Corr AF:	9.12	Prev response	9.11	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	7.98	1.001
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	7.97	1.002
second point	4960	40.1	3.99	3.97	1.005
third point	4980	20.0	1.99	1.99	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	80.2	7.98	7.96	1.002
Average Correction Factor					1.002
Baseline Corr AF:	7.98	Prev response	8.00	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999215	0.995217
THC Cal Offset:	0.010853	0.001033
CH ₄ Cal Slope:	1.001638	0.997716
CH ₄ Cal Offset:	0.008070	-0.000741
NMHC Cal Slope:	0.997023	0.993146
NMHC Cal Offset:	0.003583	0.001575

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 16, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:04	End Time (MST):	13:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
17.12	17.04	1.0044						
8.56	8.50	1.0067				Slope	0.995217	0.90 - 1.10
4.27	4.26	1.0015						
			Intercept	0.001033	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 16, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:04	End Time (MST):	13:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995
7.98	7.97	1.0020			
3.99	3.97	1.0047			
1.99	1.99	1.0002			
			Slope	0.997716	0.90 - 1.10
			Intercept	-0.000741	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 16, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	10:04	End Time (MST):	13:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995			
9.14	9.08	1.0065						
4.57	4.53	1.0085				Slope	0.993146	0.90 - 1.10
2.28	2.27	1.0027						
			Intercept	0.001575	± 0.5			



NMHC Calibration Plot

Date: March 6, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Barge Landing	Station number:	AMS09
Calibration Date:	March 28, 2024	Last Cal Date:	March 6, 2024
Start time (MST):	9:43	End time (MST):	12:24
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	497.6 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
ZAG make/model:	API T701	Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.58E-04	2.58E-04	NMHC SP Ratio:	4.37E-05
CH ₄ Retention time:	15.00	15.00	NMHC Peak Area:	209298
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	17.12	16.88	1.014
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	17.12	16.66	1.028
second point					
third point					
as left zero					
as left span					

				AF Correction Factor	1.028
Baseline Corr AF:	16.88	Prev response	17.04	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	9.14	8.95	1.021
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	9.14	8.83	1.035
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.035
Baseline Corr AF:	8.95	Prev response	9.08	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	80.2	7.98	7.93	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	80.2	7.98	7.84	1.018
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.018
Baseline Corr AF:	7.93	Prev response	7.96	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.995217	0.973222
THC Cal Offset:	0.001033	0.000000
CH ₄ Cal Slope:	0.997716	0.981863
CH ₄ Cal Offset:	-0.000741	0.000000
NMHC Cal Slope:	0.993146	0.966000
NMHC Cal Offset:	0.001575	0.000000

Notes:


Nitrogen Cylinder Change.

Calibration Performed By: Melissa Lemay

NMHC Calibration Plot

Date: March 28, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Barge Landing
Calibration Date: March 22, 2024
Start time (MST): 8:51
Reason: Routine
Station number: AMS09
Last Cal Date: February 13, 2024
End time (MST): 13:14

Calibration Standards

NO Gas Cylinder #: T2Y1KDH
NOX Cal Gas Conc: 47.38 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.38 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 46.94 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 46.94 ppm
NO gas Diff:
Serial Number: 3812
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1426262593

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.120	1.094	NO bkgnd or offset:	10.3	10.0
NOX coeff or slope:	0.998	0.998	NOX bkgnd or offset:	10.6	10.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	176.5	181.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999459	1.000294
NO _x Cal Offset:	0.518402	0.198498
NO Cal Slope:	0.998611	0.999838
NO Cal Offset:	-0.704018	-0.783809
NO ₂ Cal Slope:	1.004419	0.999826
NO ₂ Cal Offset:	0.978163	-1.024158



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
as found span	4915	85.3	808.3	800.7	7.5	832.9	821.1	11.6	0.970	0.975
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.2	----	----
high point	4915	85.3	808.3	800.7	7.5	808.4	800.2	8.2	1.000	1.001
second point	4957	42.6	403.7	400.0	3.7	404.6	398.9	5.7	0.998	1.003
third point	4979	21.3	201.8	200.0	1.9	202.1	198.1	3.9	0.999	1.009
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.2	----	----
as left span	4915	85.3	808.3	418.0	390.2	799.3	408.8	390.4	1.011	1.023
Average Correction Factor									0.999	1.004

Corrected As found	NO _x = 833.3 ppb	NO = 821.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 3.0%
Previous Response	NO _x = 808.3 ppb	NO = 798.9 ppb		*Percent Change	NO = 2.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	790.4	407.7	390.2	389.5	1.002	99.8%
2nd GPT point (200 ppb O3)	790.4	597.3	200.6	199.3	1.007	99.3%
3rd GPT point (100 ppb O3)	790.4	692.9	105.0	103.0	1.019	98.1%
Average Correction Factor					1.009	99.1%

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 22, 2024	Previous Calibration:	February 13, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:51	End Time (MST):	13:14
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
808.3	808.4	0.9998		
403.7	404.6	0.9978		
201.8	202.1	0.9986		
			0.999999	
			1.000294	
			0.198498	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 22, 2024	Previous Calibration:	February 13, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:51	End Time (MST):	13:14
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.7	800.2	1.0007			
400.0	398.9	1.0027			
200.0	198.1	1.0094			
			Slope	0.999838	0.90 - 1.10
			Intercept	-0.783809	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 22, 2024	Previous Calibration:	February 13, 2024
Station Name:	Barge Landing	Station Number:	AMS09
Start Time (MST):	8:51	End Time (MST):	13:14
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
390.2	389.5	1.0018		
200.6	199.3	1.0066		
105.0	103.0	1.0195		



NO_x Calibration Plot

Date: March 22, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Barge Landing Station number: AMS 09
 Calibration Date: March 22, 2024 Last Cal Date: February 16, 2024
 Start time (MST): 9:43 End time (MST): 9:54

Analyzer Make: API T640 S/N: 844
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388746
 Temp/RH standard: Alicat FP-25BT S/N: 388746

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-12.60	-12.62	-12.60	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	741.80	749.69	741.80	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	4.97	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39.00	----	39.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	17.80	PM w/ HEPA: _____	0.2	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: January 22, 2024
 Date Disposable Filter Changed: January 22, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 23, 2023
 Date RH/T Sensor Cleaned: August 23, 2023

Notes: Inlet head looks good. No adjustments. Leak check passed.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS11 LOWER CAMP MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 15, 2024
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:47	End Time (MST):	14:43
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	1.2	----	Correlation Coefficient	0.999989	
799.6	796.1	1.0045			≥0.995
400.9	397.6	1.0083	Slope	0.994478	
200.9	199.2	1.0085			0.90 - 1.10
			Intercept	0.097406	+/-30



SO2 Calibration Plot

Date: March 26, 2024

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Lower Camp Station number: AMS11
 Calibration Date: March 25, 2024 Last Cal Date: February 29, 2024
 Start time (MST): 9:34 End time (MST): 11:22
 Reason: Removal

Calibration Standards

Cal Gas Concentration: 5.429 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC501097
 Removed Cal Gas Conc: 5.429 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3807
 ZAG Make/Model: API T701H Serial Number: 196

Analyzer Information

Analyzer make: Thermo 450iQ Analyzer serial #: CM20080003
 Converter make: NA Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011205		Backgd or Offset: 15.8	NA
Calibration intercept:	-0.085817		Coeff or Slope: 1.015	NA

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4926	73.6	79.9	80.2	0.995
as found 2nd point	4963	36.8	40.0	40.3	0.989
as found 3rd point	4982	18.6	20.2	20.2	0.995
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero				-0.1	
high point					
second point					
third point					
as left zero					
as left span					
SO2 Scrubber Check	4935	81.5	812.3	0.3	----
Date of last scrubber change:				Ave Corr Factor	
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 80.3 Prev response: 80.73 *% change: -0.5%
 Baseline Corr 2nd AF pt: 40.4 AF Slope: 1.005049 AF Intercept: -0.045255
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999987

* = > +/-5% change initiates investigation


Notes: Removal calibration. Ran SO2 scrubber check after as found span and it passed.

Calibration Performed By: Mohammed Kashif

H₂S Calibration Plot

Date: March 25, 2024

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	March 25, 2024	Last Cal Date:	NA
Start time (MST):	13:00	End time (MST):	15:29
Reason:	Install		

Calibration Standards

Cal Gas Concentration:	5.429	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC501097			
Removed Cal Gas Conc:	5.429	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3807
ZAG Make/Model:	API T701H		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	1203169745
Converter make:	Global G150	Converter serial #:	2022-223
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.012278	Backgd or Offset:	NA	3.01
Calibration intercept:	NA	-0.364532	Coeff or Slope:	NA	0.817

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4926	73.6	79.9	80.5	0.993
second point	4963	36.8	40.0	40.2	0.994
third point	4982	18.6	20.2	20.0	1.010
as left zero	5000	0.0	0.0	-0.2	----
as left span	4926	73.6	79.9	80.7	0.990
SO2 Scrubber Check	4935	81.5	812.3	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.999
Date of last converter efficiency test:				efficiency	

Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA		

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter. Adjusted both zero and span. Ran scrubber check after cal zero and it passed.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 25, 2024	Previous Calibration:	NA
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	13:00	End Time (MST):	15:29
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0		----	Correlation Coefficient	0.999989	
79.9	80.5	0.9928			≥0.995
40.0	40.2	0.9940	Slope	1.012278	
20.2	20.0	1.0097			0.90 - 1.10
			Intercept	-0.364532	+/-3



H₂S Calibration Plot

Date: March 25, 2024

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Lower Camp	Station number:	AMS11
Calibration Date:	March 26, 2024	Last Cal Date:	February 15, 2024
Start time (MST):	10:47	End time (MST):	14:43
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	502.0 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
ZAG make/model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.88E-04	2.68E-04	NMHC SP Ratio:	5.30E-05
CH ₄ Retention time:	14.6	14.6	NMHC Peak Area:	173476
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				4.81E-05
				190907
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4932	81.4	17.33	18.90	0.917
as found 2nd point	4959	40.7	8.69	9.37	0.927
as found 3rd point	4981	20.4	4.35	4.69	0.928
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4932	81.4	17.33	17.30	1.001
second point	4959	40.7	8.69	8.63	1.006
third point	4981	20.4	4.35	4.31	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4932	81.4	17.33	17.52	0.989
Average Correction Factor					1.005

Baseline Corr AF:	18.90	Prev response	17.33	*% change	8.3%
Baseline Corr 2nd AF:	9.4	AF Slope:	1.091445	AF Intercept:	-0.044688
Baseline Corr 3rd AF:	4.7	AF Correlation:	0.999957	* = +/-% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4932	81.4	9.18	10.10	0.909
as found 2nd point	4959	40.7	4.60	5.02	0.916
as found 3rd point	4981	20.4	2.31	2.51	0.917
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4932	81.4	9.18	9.16	1.001
second point	4959	40.7	4.60	4.58	1.004
third point	4981	20.4	2.31	2.29	1.008
as left zero	5000	0.0	0.00	0.00	----
as left span	4932	81.4	9.18	9.29	0.988
Average Correction Factor					1.005
Baseline Corr AF:	10.10	Prev response	9.18	*% change	9.1%
Baseline Corr 2nd AF:	5.0	AF Slope:	1.100657	AF Intercept:	-0.016446
Baseline Corr 3rd AF:	2.5	AF Correlation:	0.999982	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4932	81.4	8.15	8.81	0.926
as found 2nd point	4959	40.7	4.09	4.34	0.941
as found 3rd point	4981	20.4	2.05	2.18	0.940
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4932	81.4	8.15	8.14	1.002
second point	4959	40.7	4.09	4.06	1.008
third point	4981	20.4	2.05	2.03	1.010
as left zero	5000	0.0	0.00	0.00	----
as left span	4932	81.4	8.15	8.23	0.991
Average Correction Factor					1.006
Baseline Corr AF:	8.81	Prev response	8.15	*% change	7.5%
Baseline Corr 2nd AF:	4.34	AF Slope:	1.080823	AF Intercept:	-0.027841
Baseline Corr 3rd AF:	2.18	AF Correlation:	0.999919	* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001990	0.999083
THC Cal Offset:	-0.033339	-0.021522
CH ₄ Cal Slope:	1.001462	0.998771
CH ₄ Cal Offset:	-0.014179	-0.011567
NMHC Cal Slope:	1.002471	0.999149
NMHC Cal Offset:	-0.018960	-0.009354

Notes: Hydrogen generator was swapped with cylinder then performed as founds. Changed sample inlet filter after MAF's. Investigation was made and suspecting that the switch from the H₂ generator to an H₂ cylinder as the fuel source may have caused the off-limits readings, as the setup, diagnostics, and chromatograms appeared normal. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 15, 2024
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:47	End Time (MST):	14:43
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999991	≥ 0.995
17.33	17.30	1.0013			
8.69	8.63	1.0061			
4.35	4.31	1.0090			
			Slope	0.999083	0.90 - 1.10
			Intercept	-0.021522	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 15, 2024
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:47	End Time (MST):	14:43
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999987	≥0.995
8.15	8.14	1.0016			
4.09	4.06	1.0078			
2.05	2.03	1.0097			
			Slope	0.998771	0.90 - 1.10
			Intercept	-0.011567	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 15, 2024
Station Name:	Lower Camp	Station Number:	AMS11
Start Time (MST):	10:47	End Time (MST):	14:43
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995			
9.18	9.16	1.0013						
4.60	4.58	1.0045				Slope	0.999149	0.90 - 1.10
2.31	2.29	1.0083						
			Intercept	-0.009354	± 0.5			



NMHC Calibration Plot

Date: March 26, 2024

Location: Lower Camp





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13 FORT MCKAY SOUTH MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 12, 2024
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:23	End Time (MST):	13:35
Analyzer make:	API T100	Analyzer serial #:	599

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999954	
799.7	800.2	0.9994			≥0.995
399.3	396.9	1.0061	Slope	1.002127	
200.2	195.5	1.0240			0.90 - 1.10
			Intercept	-2.338170	+/-30



SO2 Calibration Plot

Date: March 12, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: March 21, 2024 Last Cal Date: February 6, 2024
 Start time (MST): 10:06 End time (MST): 14:19
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.34 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC500241
 Removed Cal Gas Conc: 5.34 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API 701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1180540017
 Converter make: CDN-101 Converter serial #: 521
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001638	1.007879	Backgd or Offset: 4.15	4.15
Calibration intercept:	-0.302148	-0.422268	Coeff or Slope: 1.157	1.157

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.4	----
as found span	4925	75.5	80.6	81.1	0.989
as found 2nd point	4962	37.7	40.3	40.0	0.997
as found 3rd point	4981	18.9	20.2	19.5	1.014
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.3	----
high point	4925	75.5	80.6	80.9	0.997
second point	4962	37.7	40.3	40.1	1.004
third point	4981	18.9	20.2	19.8	1.019
as left zero	5000	0.0	0.0	0.1	----
as left span	4925	75.5	80.6	80.1	1.007
SO2 Scrubber Check	4921	79.1	791.0	0.1	----
Date of last scrubber change:	20-Mar-20		Ave Corr Factor		1.007
Date of last converter efficiency test:	NA		efficiency		

Baseline Corr As found: 81.5 Prev response: 80.46 *% change: 1.3%
 Baseline Corr 2nd AF pt: 40.4 AF Slope: 1.012557 AF Intercept: -0.662257
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999953

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 21, 2024	Previous Calibration:	February 6, 2024
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:06	End Time (MST):	14:19
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	0.999990	≥0.995
80.6	80.9	0.9966			
40.3	40.1	1.0041	Slope	1.007879	0.90 - 1.10
20.2	19.8	1.0195			
			Intercept	-0.422268	+/-3



TRS Calibration Plot

Date: March 21, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort McKay South	Station number:	AMS13
Calibration Date:	March 12, 2024	Last Cal Date:	February 12, 2024
Start time (MST):	10:23	End time (MST):	13:35
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH4 Cal Gas Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
C3H8 Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
Removed C3H8 Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
ZAG make/model:	API701	Serial Number:	1117

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.94E-04	2.95E-04	NMHC SP Ratio:	4.83E-05	5.07E-05
CH4 Retention time:	15.20	15.20	NMHC Peak Area:	1879773	179256
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.02	----
as found span	4921	79.1	17.05	16.83	1.013
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	17.05	17.02	1.001
second point	4961	39.5	8.51	8.42	1.011
third point	4980	19.8	4.27	4.11	1.039
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	17.05	17.07	0.999

Average Correction Factor	1.017
---------------------------	-------

Baseline Corr AF:	16.81	Prev response	16.99	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.1	9.08	8.80	1.031
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	9.08	9.12	0.996
second point	4961	39.5	4.53	4.52	1.004
third point	4980	19.8	2.27	2.19	1.040
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	9.08	9.13	0.994
Average Correction Factor					1.013
Baseline Corr AF:	8.80	Prev response	9.07	*% change	-3.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.02	----
as found span	4921	79.1	7.97	8.03	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.1	7.97	7.91	1.007
second point	4961	39.5	3.98	3.90	1.020
third point	4980	19.8	1.99	1.92	1.039
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.1	7.97	7.94	1.004
Average Correction Factor					1.022
Baseline Corr AF:	8.01	Prev response	7.93	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.000652	1.001286
THC Cal Offset:	-0.065578	-0.078556
CH ₄ Cal Slope:	1.000073	0.994805
CH ₄ Cal Offset:	-0.040797	-0.034188
NMHC Cal Slope:	1.001197	1.007213
NMHC Cal Offset:	-0.024180	-0.045569

Notes: Changed inlet filter after as founds. Do zero chromatogram and Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 12, 2024
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:23	End Time (MST):	13:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999901	≥ 0.995
17.05	17.02	1.0013			
8.51	8.42	1.0110			
4.27	4.11	1.0393			
			Slope	1.001286	0.90 - 1.10
			Intercept	-0.078556	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 12, 2024
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:23	End Time (MST):	13:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999914	≥0.995			
7.97	7.91	1.0072						
3.98	3.90	1.0197				Slope	0.994805	0.90 - 1.10
1.99	1.92	1.0393						
			Intercept	-0.034188	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 12, 2024
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:23	End Time (MST):	13:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999875	≥ 0.995			
9.08	9.12	0.9960						
4.53	4.52	1.0039				Slope	1.007213	0.90 - 1.10
2.27	2.19	1.0397						
			Intercept	-0.045569	± 0.5			



NMHC Calibration Plot

Date: March 12, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort McKay South
Calibration Date: March 27, 2024
Start time (MST): 8:42
Reason: Routine
Station number: AMS 13
Last Cal Date: February 20, 2024
End time (MST): 12:47

Calibration Standards

NO Gas Cylinder #: T2UP1RP
NOX Cal Gas Conc: 48.25 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.25 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 47.88 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.88 ppm
NO gas Diff:
Serial Number: 2448
Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1410661329

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.128	1.128	NO bkgnd or offset:	10.1	10.1
NOX coeff or slope:	0.999	0.999	NOX bkgnd or offset:	10.2	10.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	156.4	157.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999462	0.994114
NO _x Cal Offset:	-1.952788	-1.991485
NO Cal Slope:	1.002954	0.997692
NO Cal Offset:	-2.771571	-2.930061
NO ₂ Cal Slope:	1.000409	0.999151
NO ₂ Cal Offset:	-0.685310	-1.075362



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as found span	4917	83.5	805.7	799.5	6.2	802.3	796.4	5.9	1.0042	1.0039
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
high point	4917	83.5	805.7	799.5	6.2	800.1	796.4	3.8	1.0070	1.0039
second point	4958	41.8	403.4	400.3	3.1	397.7	394.5	3.2	1.0143	1.0147
third point	4979	20.9	201.7	200.1	1.5	196.6	194.0	2.6	1.0259	1.0317
as left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
as left span	4917	83.5	805.7	383.6	422.1	802.2	378.3	423.9	1.0044	1.0141
Average Correction Factor									1.0157	1.0168

Corrected As found	NO _x = 802.4 ppb	NO = 796.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.1%	
Previous Response	NO _x = 803.3 ppb	NO = 799.1 ppb		*Percent Change	NO = -0.3%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	790.9	375.0	422.1	421.1	1.0023	99.8%
2nd GPT point (200 ppb O ₃)	790.9	583.6	213.5	212.0	1.0070	99.3%
3rd GPT point (100 ppb O ₃)	790.9	686.4	110.7	108.2	1.0229	97.8%
Average Correction Factor					1.0107	98.9%

Notes: Changed inlet filter after as founds. No adjustment made. Used 2nd NO reference point due to drift.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 27, 2024	Previous Calibration:	February 20, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:42	End Time (MST):	12:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient 0.999968	≥0.995	
805.7	800.1	1.0070			
403.4	397.7	1.0143			
201.7	196.6	1.0259			
			Slope	0.994114	0.90 - 1.10
			Intercept	-1.991485	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 27, 2024	Previous Calibration:	February 20, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:42	End Time (MST):	12:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.5	796.4	1.0039		
400.3	394.5	1.0147		
200.1	194.0	1.0317		





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 27, 2024	Previous Calibration:	February 20, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:42	End Time (MST):	12:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
422.1	421.1	1.0023		
213.5	212.0	1.0070		
110.7	108.2	1.0229		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 27, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Fort McKay South Station number: AMS13
 Calibration Date: March 5, 2024 Last Cal Date: February 2, 2024
 Start time (MST): 10:38 End time (MST): 14:11
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2448
 ZAG Make/Model: Teledyne API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 3871
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000914	0.999371	Backgd or Offset:	3.7	4.0
Calibration intercept:	0.740000	0.560000	Coeff or Slope:	0.970	0.984

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.5	----
as found span	5000	989.8	400.0	395.1	1.012
as found 2nd point					
as found 3rd point					
calibrator zero	5000	0.0	0.0	-0.3	----
high point	5000	989.8	400.0	399.7	1.001
second point	5000	849.9	200.0	201.4	0.993
third point	5000	745.1	100.0	101.0	0.990
as left zero	5000	0.0	0.0	-0.5	----
as left span	5000	989.8	400.0	400.1	1.000
Average Correction Factor					0.995

Baseline Corr As found:	395.6	Previous response	401.1	*% change	-1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 2, 2024
Station Name:	Fort McKay South	Station Number:	AMS13
Start Time (MST):	10:38	End Time (MST):	14:11
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999974	≥0.995
400.0	399.7	1.0008			
200.0	201.4	0.9930	Slope	0.999371	0.90 - 1.10
100.0	101.0	0.9901			
			Intercept	0.560000	+/- 5



O₃ Calibration Plot

Date: March 5, 2024

Location: Fort McKay South





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC
MARCH 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 7, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:15	End Time (MST):	14:37
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	1.1	----	Correlation Coefficient	0.999960	≥0.995
799.3	803.3	0.9950			
400.1	400.9	0.9979	Slope	1.005467	0.90 - 1.10
201.1	198.1	1.0150			
			Intercept	-1.163959	+/-30



SO2 Calibration Plot

Date: March 7, 2024

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Anzac Station number: AMS14
 Calibration Date: March 27, 2024 Last Cal Date: February 12, 2024
 Start time (MST): 9:47 End time (MST): 13:58
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC510379
 Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3060
 ZAG Make/Model: API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153582
 Converter make: CD Nova CDN-101 Converter serial #: 503
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003735	1.000445	Backgd or Offset: 2.30	2.29
Calibration intercept:	-0.085416	-0.045344	Coeff or Slope: 0.984	0.984

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4938	77.9	80.0	80.5	0.992
as found 2nd point	4973	38.9	40.0	40.3	0.989
as found 3rd point	4997	19.5	20.0	19.8	1.006
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4938	77.9	80.0	79.9	1.001
second point	4973	38.9	40.0	40.1	0.996
third point	4997	19.5	20.0	19.8	1.011
as left zero	5000	0.0	0.0	0.1	----
as left span	4938	77.9	80.0	78.6	1.017
SO2 Scrubber Check	4936	80.3	800.4	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.003
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 80.6 Prev response: 80.20 *% change: 0.5%
 Baseline Corr 2nd AF pt: 40.4 AF Slope: 1.009059 AF Intercept: -0.185387
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999978

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 27, 2024	Previous Calibration:	February 12, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	9:47	End Time (MST):	13:58
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999982	
80.0	79.9	1.0006			≥0.995
40.0	40.1	0.9964	Slope	1.000445	
20.0	19.8	1.0107			0.90 - 1.10
			Intercept	-0.045344	+/-3



TRS Calibration Plot

Date: March 27, 2024

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Anzac	Station number:	AMS14
Calibration Date:	March 7, 2024	Last Cal Date:	February 3, 2024
Start time (MST):	11:15	End time (MST):	14:37
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	499.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
ZAG make/model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.25E-04	2.25E-04	NMHC SP Ratio:	4.11E-05
CH ₄ Retention time:	13.30	13.30	NMHC Peak Area:	221451
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	17.10	17.05	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	17.10	17.01	1.006
second point	4979	40.2	8.56	8.41	1.018
third point	4998	20.2	4.30	4.17	1.032
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	17.10	17.00	1.006
Average Correction Factor					1.019

Baseline Corr AF:	17.05	Prev response	17.09	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	9.11	9.09	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	9.11	9.07	1.005
second point	4979	40.2	4.56	4.51	1.012
third point	4998	20.2	2.29	2.24	1.025
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	9.11	9.06	1.006
Average Correction Factor					1.014
Baseline Corr AF:	9.09	Prev response	9.12	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4938	80.3	7.99	7.96	1.004
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4938	80.3	7.99	7.94	1.007
second point	4979	40.2	4.00	3.91	1.024
third point	4998	20.2	2.01	1.93	1.042
as left zero	5000	0.0	0.00	0.00	----
as left span	4938	80.3	7.99	7.95	1.006
Average Correction Factor					1.024
Baseline Corr AF:	7.96	Prev response	7.97	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002044	0.995891
THC Cal Offset:	-0.050467	-0.064367
CH ₄ Cal Slope:	1.002402	0.995346
CH ₄ Cal Offset:	-0.039003	-0.040562
NMHC Cal Slope:	1.001944	0.995954
NMHC Cal Offset:	-0.011066	-0.023403

Notes: Sample inlet filter changed after as founds. No adjustments needed.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 3, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:15	End Time (MST):	14:37
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999932	≥ 0.995			
17.10	17.01	1.0057						
8.56	8.41	1.0177				Slope	0.995891	0.90 - 1.10
4.30	4.17	1.0325						
			Intercept	-0.064367	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 3, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:15	End Time (MST):	14:37
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999871	≥0.995			
7.99	7.94	1.0066						
4.00	3.91	1.0241				Slope	0.995346	0.90 - 1.10
2.01	1.93	1.0419				Intercept	-0.040562	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 3, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:15	End Time (MST):	14:37
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999969	≥ 0.995
9.11	9.07	1.0053			
4.56	4.51	1.0121			
2.29	2.24	1.0248			
			Slope	0.995954	0.90 - 1.10
			Intercept	-0.023403	+/-0.5



NMHC Calibration Plot

Date: March 7, 2024

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Anzac
Calibration Date: March 14, 2024
Start time (MST): 10:34
Reason: Routine
Station number: AMS14
Last Cal Date: February 6, 2024
End time (MST): 15:25

Calibration Standards

NO Gas Cylinder #: DT0037092
NOX Cal Gas Conc: 60.7 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.7 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API 701H
Cal Gas Expiry Date: May 16, 2031
NO Cal Gas Conc: 60.4 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.4 ppm
NO gas Diff:
Serial Number: 3060
Serial Number: 357

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 1426262592

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.411	1.411	NO bkgnd or offset:	3.8	3.8
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.8	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.0	160.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001051	0.993607
NO _x Cal Offset:	-0.650468	-0.629791
NO Cal Slope:	1.006866	0.999912
NO Cal Offset:	-1.970746	-2.109531
NO ₂ Cal Slope:	0.997816	0.994068
NO ₂ Cal Offset:	-1.058817	-1.527274



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as found span	4934	66.3	804.8	800.9	4.0	802.3	798.8	3.4	1.0032	1.0026
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
high point	4934	66.3	804.8	800.9	4.0	799.2	799.5	-0.2	1.0070	1.0017
second point	4985	33.2	401.6	399.6	2.0	398.6	397.0	1.7	1.0075	1.0066
third point	5004	16.7	201.9	200.9	1.0	199.0	196.4	2.6	1.0146	1.0229
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4934	66.3	804.8	416.8	388.1	800.0	414.3	385.8	1.0060	1.0059
Average Correction Factor									1.0097	1.0104

Corrected As found	NO _x =	802.4 ppb	NO =	799.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x =	-0.3%
Previous Response	NO _x =	805.0 ppb	NO =	804.4 ppb		*Percent Change	NO =	-0.7%
Baseline Corr 2nd pt	NO _x =	NA ppb	NO =	NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x =	NA ppb	NO =	NA ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	796.2	412.1	388.1	385.0	1.0080	99.2%
2nd GPT point (200 ppb O ₃)	796.2	608.4	191.8	188.5	1.0174	98.3%
3rd GPT point (100 ppb O ₃)	796.2	701.6	98.6	94.7	1.0409	96.1%
Average Correction Factor					1.0221	97.9%

Notes:

Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By:

Mohammed Kashif

CALS_266



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 14, 2024	Previous Calibration:	February 6, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:34	End Time (MST):	15:25
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
804.8	799.2	1.0070			
401.6	398.6	1.0075			
201.9	199.0	1.0146			
			Slope	0.993607	0.90 - 1.10
			Intercept	-0.629791	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 14, 2024	Previous Calibration:	February 6, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:34	End Time (MST):	15:25
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.1	----	Correlation Coefficient	≥0.995	
800.9	799.5	1.0017			
399.6	397.0	1.0066			
200.9	196.4	1.0229			
			Slope	0.999912	0.90 - 1.10
			Intercept	-2.109531	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 14, 2024	Previous Calibration:	February 6, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	10:34	End Time (MST):	15:25
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
388.1	385.0	1.0080		
191.8	188.5	1.0174		
98.6	94.7	1.0409		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 14, 2024

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 5, 2024
Station Name:	Anzac	Station Number:	AMS14
Start Time (MST):	11:54	End Time (MST):	14:34
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.7	----	Correlation Coefficient	0.999992	
400.0	402.4	0.9940			≥0.995
200.0	202.5	0.9877	Slope	1.004629	
100.0	101.0	0.9901			0.90 - 1.10
			Intercept	0.840000	+/- 5



O₃ Calibration Plot

Date: March 6, 2024

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Anzac Station number: AMS 14
 Calibration Date: March 27, 2024 Last Cal Date: February 26, 2024
 Start time (MST): 12:49 End time (MST): 13:08

Analyzer Make: AP T640 S/N: 825
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
 Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-5.8	-6.11	-5.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	712.1	713.18	712.1	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.975	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	40	----	40	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.1	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: _____ Expiry Date: _____
Lot No.: _____

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ December 6, 2023
Date Disposable Filter Changed: _____ December 6, 2023

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ July 6, 2023
Date RH/T Sensor Cleaned: _____ July 6, 2023

Notes: No adjustments made. Leak check passed. Head cleaned.

Calibration by: Mohammed Kashif



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Anzac	Station Number:	AMS 14
Calibration Date:	March 11, 2024	Prev Cal Date:	August 23, 2023
Start Time (MST):	10:00	End Time (MST):	13:16
Tower Height (m):	20.0	Reason:	Routine

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	E5132
WS Calibrator:	MetOne 053	Serial Number:	CA05231

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.2	0.2%
400	39.4	39.4	0.1%
600	58.6	58.6	0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		1.000000	≥0.9995
Calculated slope		0.999465	0.90 - 1.10
Calculated intercept		-0.013446	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	Z1048
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:34	Calc Declination*:	13.23 Degrees
Deadband calc:	-1.1 degrees (<i>Limit 4 deg</i>)		* - calculated declination as per NOAA website

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	1.0	---
90	89.9	0.0%
180	180.0	0.0%
270	270.6	0.2%
356	359.1	0.9%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999972	≥0.9995
Calculated slope		0.994477	0.90 - 1.10
Calculated intercept		0.072574	+/- 4

Notes: Performed WS and WD calibration. Noticed that the screw at the bottom or the 180-degree notch at the top of the WD sensor was not precisely facing south; instead, it was off by 20 degrees towards the SW. Realigned the WD so that the screw/180-degree notch is now facing south. The WS bearings seem to be in good condition.

Calibration Performed By: Mohammed Kashif and Max Farrell



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS17
WAPASU
MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 6, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:10	End Time (MST):	14:18
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient	≥0.995
800.0	802.5	0.9968		
400.0	396.9	1.0079	Slope	0.90 - 1.10
199.5	197.2	1.0117		
			Intercept	+/-30



SO2 Calibration Plot

Date: March 6, 2024

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Wapasu Station number: AMS17
 Calibration Date: March 12, 2024 Last Cal Date: February 12, 2024
 Start time (MST): 10:12 End time (MST): 15:15
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.076 ppm Cal Gas Exp Date: September 16, 2024
 Cal Gas Cylinder #: CC511852
 Removed Cal Gas Conc: 5.076 ppm Rem Gas Exp Date: n/a
 Removed Gas Cyl #: n/a Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 2449
 ZAG Make/Model: API T701H Serial Number: 359

Analyzer Information

Analyzer make: Thermo 450i Analyzer serial #: 1218153583
 Converter make: n/a Converter serial #: n/a
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010139	0.995282	Backgd or Offset:	12.2 11.7
Calibration intercept:	0.220812	0.340790	Coeff or Slope:	1.114 1.096

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4921	78.8	80.0	82.5	0.973
as found 2nd point	4961	39.4	40.0	40.8	0.988
as found 3rd point	4980	19.7	20.0	20.2	1.005
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.6	----
high point	4921	78.8	80.0	80.1	0.999
second point	4961	39.4	40.0	40.0	1.000
third point	4980	19.7	20.0	20.0	1.000
as left zero	5000	0.0	0.0	0.8	----
as left span	4921	78.8	80.0	79.5	1.006
SO2 Scrubber Check	4921	79.4	800.0	-0.2	----
Date of last scrubber change:	n/a			Ave Corr Factor	1.000
Date of last converter efficiency test:	n/a			efficiency	

Baseline Corr As found: 82.2 Prev response: 81.03 *% change: 1.4%
 Baseline Corr 2nd AF pt: 40.5 AF Slope: 1.029425 AF Intercept: -0.079189
 Baseline Corr 3rd AF pt: 19.9 AF Correlation: 0.999899

* = > +/-5% change initiates investigation

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 12, 2024
Station Name:	2024-03-12	Station Number:	Wapasu
Start Time (MST):	10:12	End Time (MST):	15:15
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.6	----	Correlation Coefficient	0.999951	≥0.995
80.0	80.1	0.9988			
40.0	40.0	0.9999	Slope	0.995282	0.90 - 1.10
20.0	20.0	1.0000			
			Intercept	0.340790	+/-3



H₂S Calibration Plot

Date: March 12, 2024

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 6, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:10	End Time (MST):	14:18
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	<u>Limits</u>	
0.00	0.09	----	Correlation Coefficient	0.999939	
17.09	17.04	1.0031			≥0.995
8.55	8.45	1.0109	Slope	0.992946	
4.26	4.23	1.0084			0.90 - 1.10
			Intercept	0.028821	+/-1.5



THC Calibration Plot

Date: March 6, 2024

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Wapasu
Calibration Date: March 26, 2024
Start time (MST): 6:47
Reason: Routine
Station number: AMS17
Last Cal Date: February 13, 2024
End time (MST): 11:06

Calibration Standards

NO Gas Cylinder #: T375YK8
NOX Cal Gas Conc: 49.11 ppm
Removed Cylinder #: T375YK8
Removed Gas NOX Conc: 49.11 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: April 13, 2025
NO Cal Gas Conc: 48.07 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 48.07 ppm
NO gas Diff:
Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: Thermo Scientific 42iQ
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 12300522720

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.264	1.345	NO bkgnd or offset:	8.1	9.7
NOX coeff or slope:	0.986	0.994	NOX bkgnd or offset:	8.1	9.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	386.5	386.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998991	1.004319
NO _x Cal Offset:	-5.260000	-4.540000
NO Cal Slope:	1.001959	1.001930
NO Cal Offset:	-5.660000	-5.200000
NO ₂ Cal Slope:	0.995634	1.007870
NO ₂ Cal Offset:	-1.245592	1.467173



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.8	0.0	----	----
as found span	4917	83.2	817.2	799.9	17.3	774.3	761.2	13.1	1.0554	1.0508
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
high point	4917	83.2	817.2	799.9	17.3	819.3	799.7	19.6	0.9974	1.0002
second point	4958	41.6	408.6	399.9	8.7	400.8	390.3	10.5	1.0194	1.0247
third point	4979	20.8	204.3	200.0	4.3	198.2	191.7	6.6	1.0308	1.0431
as left zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
as left span	4917	83.2	817.2	393.7	423.5	828.1	399.9	428.2	0.9868	0.9844
Average Correction Factor									1.0159	1.0227

Corrected As found	NO _x = 775.1 ppb	NO = 762.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -4.6%	
Previous Response	NO _x = 811.1 ppb	NO = 795.8 ppb		*Percent Change	NO = -4.4%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	807.1	400.9	423.5	427.5	0.9907	100.9%
2nd GPT point (200 ppb O3)	807.1	601.6	222.8	226.5	0.9837	101.7%
3rd GPT point (100 ppb O3)	807.1	706.5	117.9	122.3	0.9641	103.7%
Average Correction Factor					0.9795	102.1%

Notes: Diagnostics similar to last month. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 13, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	6:47	End Time (MST):	11:06
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
817.2	819.3	0.9974		
408.6	400.8	1.0194		
204.3	198.2	1.0308		
			0.999838	
			1.004319	
			-4.540000	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 13, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	6:47	End Time (MST):	11:06
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.9	799.7	1.0002		
399.9	390.3	1.0247		
200.0	191.7	1.0431		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 13, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	6:47	End Time (MST):	11:06
Analyzer make:	Thermo Scientific 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
423.5	427.5	0.9907		
222.8	226.5	0.9837		
117.9	122.3	0.9641		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 26, 2024

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 11, 2024	Previous Calibration:	February 9, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:43	End Time (MST):	14:02
Analyzer make:	API T400	Analyzer serial #:	3870

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999992	≥0.995
400.0	400.2	0.9995			
200.0	199.5	1.0025	Slope	1.000657	0.90 - 1.10
100.0	99.2	1.0081			
			Intercept	-0.340000	+/- 5



O₃ Calibration Plot

Date: March 11, 2024

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Wapasu Station number: AMS 17
 Calibration Date: March 26, 2024 Last Cal Date: February 15, 2024
 Start time (MST): 6:49 End time (MST): 7:12

Analyzer Make: Teledyne API T640 S/N: 1183
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-5.6	-5.4	-5.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	710.5	713.4	710.5	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.97	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)		----		<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	6.5	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: _____ Expiry Date: _____
Lot No.: _____

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ February 13, 2024
Date Disposable Filter Changed: _____ February 13, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____
Date RH/T Sensor Cleaned: _____

Notes: No adjustments done. Leak Check Passed.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	March 1, 2024	Last Cal Date:	February 1, 2024
Start time (MST):	12:27	End time (MST):	16:39
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.40	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC463851			
Removed Cal Gas Conc:	49.40	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API 701H		Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002503	1.000733	Backgd or Offset:	22.9	22.9
Calibration intercept:	-0.043083	0.276595	Coeff or Slope:	0.800	0.800

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5009	0.0	0.0	0.2	----
as found span	4919	81.0	800.3	797.4	1.004
as found 2nd point	4959	40.5	400.2	398.3	1.005
as found 3rd point	4979	20.2	199.6	197.4	1.011
new cylinder response					
calibrator zero	5000	0.0	0.0	0.6	----
high point	4919	81.0	800.3	801.1	0.999
second point	4959	40.5	400.2	401.2	0.997
third point	4979	20.2	199.6	199.3	1.002
as left zero	5000	0.0	0.0	0.7	----
as left span	4919	81.0	800.3	804.2	0.995
Average Correction Factor					0.999

Baseline Corr As found:	797.20	Previous response	802.24	*% change	-0.6%
Baseline Corr 2nd AF pt:	398.10	AF Slope:	0.996778	AF Intercept:	-0.564209
Baseline Corr 3rd AF pt:	197.20	AF Correlation:	0.999995		

* = > +/-5% change initiates investigation

Notes: Inlet filter changed after as founds. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 1, 2024	Previous Calibration:	February 1, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	12:27	End Time (MST):	16:39
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.6	----	Correlation Coefficient	0.999998	
800.3	801.1	0.9990			≥0.995
400.2	401.2	0.9975	Slope	1.000733	
199.6	199.3	1.0015			0.90 - 1.10
			Intercept	0.276595	+/-30



SO2 Calibration Plot

Date: March 1, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name:	Stony Mountain	Station number:	AMS18
Calibration Date:	March 27, 2024	Last Cal Date:	February 15, 2024
Start time (MST):	10:13	End time (MST):	14:50
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.479	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC500395			
Removed Cal Gas Conc:	5.479	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701		Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359
Converter make:	CD Nova CDN-101	Converter serial #:	555
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995441	1.002157	Backgd or Offset:	2.66
Calibration intercept:	0.301088	0.240997	Coeff or Slope:	1.157

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4927	73.0	80.0	82.0	0.978
as found 2nd point	4964	36.5	40.0	41.2	0.975
as found 3rd point	4983	18.3	20.0	20.4	0.992
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4927	73.0	80.0	80.4	0.995
second point	4964	36.5	40.0	40.4	0.990
third point	4983	18.3	20.0	20.2	0.992
as left zero	5000	0.0	0.0	0.4	----
as left span	4927	73.0	80.0	79.9	1.001
SO2 Scrubber Check	4923	77.1	771.0	0.1	----
Date of last scrubber change:	17-Dec-21			Ave Corr Factor	0.992
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	81.8	Prev response:	79.93	*% change:	2.3%
Baseline Corr 2nd AF pt:	41.0	AF Slope:	1.024019	AF Intercept:	0.100632
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999977		

* = > +/-5% change initiates investigation

Notes: Scrubber check completed after calibrator zero. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 27, 2024	Previous Calibration:	February 15, 2024
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:13	End Time (MST):	14:50
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999992	≥0.995
80.0	80.4	0.9949			
40.0	40.4	0.9899	Slope	1.002157	0.90 - 1.10
20.0	20.2	0.9925			
			Intercept	0.240997	+/-3



TRS Calibration Plot

Date: March 27, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	March 1, 2024	Last Cal Date:	February 1, 2024
Start time (MST):	12:27	End time (MST):	16:39
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.88E-04	3.10E-04	NMHC SP Ratio:	5.97E-05	5.89E-05
CH ₄ Retention time:	15.8	16.2	NMHC Peak Area:	153566	156523
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.09	----
as found span	4919	81.0	17.28	16.25	1.064
as found 2nd point	4959	40.5	8.64	8.28	1.044
as found 3rd point	4979	20.2	4.31	4.15	1.038
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	17.28	17.35	0.996
second point	4959	40.5	8.64	8.73	0.990
third point	4979	20.2	4.31	4.37	0.986
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	17.28	17.51	0.987
Average Correction Factor					0.991

Baseline Corr AF:	16.16	Prev response	17.18	*% change	-6.3%
Baseline Corr 2nd AF:	8.2	AF Slope:	0.935180	AF Intercept:	0.123368
Baseline Corr 3rd AF:	4.1	AF Correlation:	0.999947	* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	9.02	1.017
as found 2nd point	4959	40.5	4.58	4.50	1.018
as found 3rd point	4979	20.2	2.29	2.26	1.010
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	9.22	0.995
second point	4959	40.5	4.58	4.64	0.988
third point	4979	20.2	2.29	2.31	0.991
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	9.17	9.25	0.991
Average Correction Factor					0.991
Baseline Corr AF:	9.02	Prev response	9.09	*% change	-0.9%
Baseline Corr 2nd AF:	4.5	AF Slope:	0.982616	AF Intercept:	0.004996
Baseline Corr 3rd AF:	2.3	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.09	----
as found span	4919	81.0	8.11	7.23	1.122
as found 2nd point	4959	40.5	4.06	3.77	1.075
as found 3rd point	4979	20.2	2.02	1.89	1.071
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	8.11	8.14	0.997
second point	4959	40.5	4.06	4.10	0.991
third point	4979	20.2	2.02	2.06	0.981
as left zero	5000	0.0	0.00	0.00	----
as left span	4919	81.0	8.11	8.26	0.982
Average Correction Factor					0.989
Baseline Corr AF:	7.15	Prev response	8.09	*% change	-13.2%
Baseline Corr 2nd AF:	3.69	AF Slope:	0.881432	AF Intercept:	0.118371
Baseline Corr 3rd AF:	1.80	AF Correlation:	0.999693	* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.993093	1.003678
THC Cal Offset:	0.015393	0.027639
CH ₄ Cal Slope:	0.997044	1.002169
CH ₄ Cal Offset:	-0.003610	0.018204
NMHC Cal Slope:	0.989722	1.005151
NMHC Cal Offset:	0.019003	0.009635

Notes: Zero at 0.08 ppm. Span points had dips present and CH₄ span was 12% low. Re-did zero chromatogram to adjust zero. Calibrated span to adjust RT.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 1, 2024	Previous Calibration:	February 1, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	12:27	End Time (MST):	16:39
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995
17.28	17.35	0.9959			
8.64	8.73	0.9896			
4.31	4.37	0.9864			
			Slope	1.003678	0.90 - 1.10
			Intercept	0.027639	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 1, 2024	Previous Calibration:	February 1, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	12:27	End Time (MST):	16:39
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999977	≥0.995			
8.11	8.14	0.9968						
4.06	4.10	0.9907				Slope	1.002169	0.90 - 1.10
2.02	2.06	0.9809						
			Intercept	0.018204	+/-0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 1, 2024	Previous Calibration:	February 1, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	12:27	End Time (MST):	16:39
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995
9.17	9.22	0.9949			
4.58	4.64	0.9883			
2.29	2.31	0.9912			
			Slope	1.005151	0.90 - 1.10
			Intercept	0.009635	+/-0.5



NMHC Calibration Plot

Date: March 1, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	March 15, 2024	Last Cal Date:	March 1, 2024
Start time (MST):	11:18	End time (MST):	13:23
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH ₄ Cal Gas Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	500.8 ppm	CH ₄ Equiv Conc.	1066.8 ppm
Removed C ₃ H ₈ Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
ZAG make/model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.10E-04	3.10E-04	NMHC SP Ratio:	5.89E-05	5.89E-05
CH ₄ Retention time:	16.2	16.2	NMHC Peak Area:	156523	156523
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.04	----
as found span	4919	81.0	17.28	16.88	1.024
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.08	----
high point	4919	81.0	17.28	16.68	1.036
second point					
third point					
as left zero					
as left span					
Average Correction Factor					1.036

Baseline Corr AF:	16.84	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4919	81.0	9.17	9.05	1.013
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4919	81.0	9.17	8.94	1.025
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.025
Baseline Corr AF:	9.05	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.04	----
as found span	4919	81.0	8.11	7.83	1.036
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.08	----
high point	4919	81.0	8.11	7.74	1.049
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.049
Baseline Corr AF:	7.79	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	NA	NA
THC Cal Offset:	NA	NA
CH ₄ Cal Slope:	NA	NA
CH ₄ Cal Offset:	NA	NA
NMHC Cal Slope:	NA	NA
NMHC Cal Offset:	NA	NA


Notes: H2 cylinder change

Calibration Performed By: Moahmmed Kashif

NMHC Calibration Plot

Date: March 15, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: March 26, 2024 Last Cal Date: February 21, 2024
Start time (MST): 10:15 End time (MST): 15:06
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T26DHGA Cal Gas Expiry Date: November 17, 2026
NOX Cal Gas Conc: 48.28 ppm NO Cal Gas Conc: 47.58 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 48.28 ppm Removed Gas NO Conc: 47.58 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 2658
ZAG make/model: Teledyne API T701 Serial Number: 13779

Analyzer Information

Analyzer make: Thermo 42i Analyzer serial #: 1336160088
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.104	1.119	NO bkgnd or offset:	3.1	3.1
NOX coeff or slope:	0.995	0.994	NOX bkgnd or offset:	3.1	3.2
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	253.2	254.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001950	0.998146
NO _x Cal Offset:	-1.100000	-0.500000
NO Cal Slope:	1.001421	0.998819
NO Cal Offset:	-1.260000	-0.600000
NO ₂ Cal Slope:	1.009656	0.999802
NO ₂ Cal Offset:	1.192503	0.793514



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as found span	4916	84.0	811.1	799.3	11.8	800.8	787.8	12.9	1.0129	1.0147
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.1	----	----
high point	4916	84.0	811.1	799.3	11.8	809.1	797.8	11.3	1.0025	1.0019
second point	4958	42.0	405.6	399.7	5.9	404.8	399.2	5.5	1.0019	1.0012
third point	4979	21.0	202.8	199.8	2.9	200.9	197.8	3.2	1.0093	1.0103
as left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
as left span	4916	84.0	811.1	370.3	440.8	824.0	371.2	452.4	0.9843	0.9977
Average Correction Factor									1.0046	1.0045

Corrected As found	NO _x = 800.9 ppb	NO = 788.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.3%	
Previous Response	NO _x = 811.6 ppb	NO = 799.2 ppb		*Percent Change	NO = -1.4%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	796.8	367.8	440.8	440.9	0.9997	100.0%
2nd GPT point (200 ppb O ₃)	796.8	587.9	220.7	222.4	0.9922	100.8%
3rd GPT point (100 ppb O ₃)	796.8	693.5	115.1	116.1	0.9910	100.9%
Average Correction Factor					0.9943	100.6%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 21, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:15	End Time (MST):	15:06
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
811.1	809.1	1.0025			
405.6	404.8	1.0019			
202.8	200.9	1.0093			
			Slope	0.998146	0.90 - 1.10
			Intercept	-0.500000	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 21, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:15	End Time (MST):	15:06
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
799.3	797.8	1.0019		
399.7	399.2	1.0012		
199.8	197.8	1.0103		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 21, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:15	End Time (MST):	15:06
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
440.8	440.9	0.9997			
220.7	222.4	0.9922			
115.1	116.1	0.9910			
			Slope	0.999802	0.90 - 1.10
			Intercept	0.793514	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 26, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Stony Mountain Station number: AMS18
 Calibration Date: March 20, 2024 Last Cal Date: February 14, 2024
 Start time (MST): 10:39 End time (MST): 13:44
 Reason: Routine

Calibration Standards

O3 generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
 ZAG Make/Model: Teledyne API 701H Serial Number: 360

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 825
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002057	0.995743	Backgd or Offset:	0.3	0.3
Calibration intercept:	0.340000	0.220000	Coeff or Slope:	0.982	0.982

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	NA	0.0	0.1	----
as found span	4804	1141.9	400.0	400.0	1.000
as found 2nd point					
as found 3rd point					
calibrator zero	5000	NA	0.0	0.1	----
high point	4888	1138.1	400.0	398.3	1.004
second point	4888	884.5	200.0	199.9	1.001
third point	4888	741.4	100.0	99.6	1.004
as left zero	5000	NA	0.0	0.5	----
as left span	4812	1097.9	400.0	401.7	0.996
Average Correction Factor					1.003

Baseline Corr As found:	399.9	Previous response	401.2	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 14, 2024
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:39	End Time (MST):	13:44
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999996	≥0.995
400.0	398.3	1.0043			
200.0	199.9	1.0005	Slope	0.995743	0.90 - 1.10
100.0	99.6	1.0040			
			Intercept	0.220000	+/- 5



O₃ Calibration Plot

Date: March 20, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Stony Mountain Station number: AMS 18
 Calibration Date: March 27, 2024 Last Cal Date: February 21, 2024
 Start time (MST): 13:13 End time (MST): 14:30

Analyzer Make: API T640 S/N: 1162
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-6	-5.7	-6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	696.3	696.19	696.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.94	5.03	4.94	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39	----	39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	4.1	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: October 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	N/A		N/A	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 21, 2024
 Date Disposable Filter Changed: February 21, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022
 Date RH/T Sensor Cleaned: August 30, 2022

Notes: Flow, pressure and temp checked. Leak check passed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 14, 2024	Previous Calibration:	February 5, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:10	End Time (MST):	13:21
Analyzer make:	API T300	Analyzer serial #:	3504

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999947	≥0.995
41.1	40.9	1.0046			
20.5	20.7	0.9925	Slope	0.996462	0.90 - 1.10
10.3	10.2	1.0047			
			Intercept	0.037795	+/-1.5



CO Calibration Plot

Date: March 14, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date	March 21, 2024	Previous Calibration	February 7, 2024
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:40	End Time (MST)	13:53
Analyzer make	API T360	Analyzer serial #	489

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999923	≥0.995
1576.0	1576.5	0.9997			
788.0	775.8	1.0157	Slope	1.000660	0.90 - 1.10
394.0	388.4	1.0144			
			Intercept	-4.780000	+/-10



CO₂ Calibration Plot

Date: March 21, 2024

Location: Stony Mountain





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS19 FIREBAG MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 27, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:17	End Time (MST):	13:53
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999999	≥0.995
799.5	797.8	1.0021			
400.3	400.2	1.0002	Slope	0.997705	0.90 - 1.10
200.1	200.2	0.9995			
			Intercept	0.418240	+/-30



SO2 Calibration Plot

Date: March 12, 2024

Location: Firebag





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Firebag Station number: AMS19
 Calibration Date: March 5, 2024 Last Cal Date: February 1, 2024
 Start time (MST): 12:36 End time (MST): 18:10
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.114 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517427
 Removed Cal Gas Conc: 5.114 ppm Rem Gas Exp Date: n/a
 Removed Gas Cyl #: n/a Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1607
 ZAG Make/Model: Teledyne API T701 Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680032
 Converter make: Global Converter serial #: 2022-222
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000909	0.999337	Backgd or Offset: 2.77	3.22
Calibration intercept:	-0.061541	-0.081535	Coeff or Slope: 1.166	1.198

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.4	----
as found span	4922	78.2	80.0	77.8	1.033
as found 2nd point	4961	39.1	40.0	39.3	1.028
as found 3rd point	4980	19.6	20.0	19.8	1.033
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4922	78.2	80.0	79.8	1.002
second point	4961	39.1	40.0	40.0	1.000
third point	4980	19.6	20.0	19.9	1.007
as left zero	5000	0.0	0.0	0.0	----
as left span	4922	78.2	80.0	79.2	1.010
SO2 Scrubber Check	4922	78.3	800.2	0.0	----
Date of last scrubber change:	January 18, 2023			Ave Corr Factor	1.003
Date of last converter efficiency test:	n/a			efficiency	

Baseline Corr As found: 77.4 Prev response: 79.99 *% change: -3.3%
 Baseline Corr 2nd AF pt: 38.9 AF Slope: 0.968037 AF Intercept: 0.439136
 Baseline Corr 3rd AF pt: 19.4 AF Correlation: 0.999991

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Ran SOx scrubber check after cal zero. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 1, 2024
Station Name:	Firebag	Station Number:	AMS19
Start Time (MST):	12:36	End Time (MST):	18:10
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999995	≥0.995
80.0	79.8	1.0023			
40.0	40.0	0.9998	Slope	0.999337	0.90 - 1.10
20.0	19.9	1.0075			
			Intercept	-0.081535	+/-3



H₂S Calibration Plot

Date: March 5, 2024

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	March 12, 2024	Last Cal Date:	February 27, 2024
Start time (MST):	10:32	End time (MST):	13:53
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1607
ZAG Make/Model:	API T701	Serial Number:	1118

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1336160089
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002921	1.003849	Background:	2.03	2.15
Calibration intercept:	-0.061568	-0.040348	Coefficient:	3.764	3.788

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.00	0.04	----
as found span	4919	81.1	17.31	17.35	0.997
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	4999	0.0	0.00	0.02	----
high point	4919	81.1	17.31	17.37	0.996
second point	4959	40.6	8.66	8.60	1.007
third point	4980	20.3	4.33	4.27	1.015
as left zero	5000	0.0	0.00	-0.04	----
as left span	4919	81.1	17.31	17.49	0.989
Average Correction Factor					1.006
Baseline Corr As found:	17.31	Previous response	17.29	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Changed out hydrogen cylinder. Adjusted zero.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 27, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:32	End Time (MST):	13:53
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.02	----	Correlation Coefficient	0.999941	≥0.995
17.31	17.37	0.9963			
8.66	8.60	1.0075	Slope	1.003849	0.90 - 1.10
4.33	4.27	1.0149			
			Intercept	-0.040348	+/-1.5



THC Calibration Plot

Date: March 12, 2024


Location: Firebag



THC Calibration Plot

Date: March 20, 2024

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Firebag
Calibration Date: March 21, 2024
Start time (MST): 9:41
Reason: Routine
Station number: AMS 19
Last Cal Date: February 13, 2024
End time (MST): 14:49

Calibration Standards

NO Gas Cylinder #: DT0044018
NOX Cal Gas Conc: 48.9 ppm
Removed Cylinder #: n/a
Removed Gas NOX Conc: 48.9 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.7 ppm
Removed Gas Exp Date: n/a
Removed Gas NO Conc: 48.7 ppm
NO gas Diff:
Serial Number: 1607
Serial Number: 201

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1410661309

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.081	1.040	NO bkgnd or offset:	7.6	7.3
NOX coeff or slope:	0.996	0.995	NOX bkgnd or offset:	7.7	7.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	216.5	222.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.006289	0.993092
NO _x Cal Offset:	0.400360	-0.219382
NO Cal Slope:	1.006923	0.994756
NO Cal Offset:	0.340247	-0.699633
NO ₂ Cal Slope:	0.998051	1.011032
NO ₂ Cal Offset:	-0.640112	-1.734028



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	4999	0.0	0.0	0.0	0.0	-0.3	0.0	-0.3	----	----
as found span	4918	82.1	802.9	799.7	3.3	840.0	835.0	4.2	0.9559	0.9577
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	4999	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
high point	4918	82.1	802.9	799.7	3.3	797.6	795.3	2.3	1.0067	1.0055
second point	4959	41.1	402.0	400.3	1.6	397.9	396.6	1.3	1.0102	1.0094
third point	4980	20.5	200.5	199.7	0.8	199.3	197.6	1.7	1.0060	1.0105
as left zero	4999	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as left span	4918	82.1	802.9	395.7	407.3	785.5	373.4	412.1	1.0222	1.0596
Average Correction Factor									1.0076	1.0084

Corrected As found	NO _x = 840.3 ppb	NO = 835.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 3.8%
Previous Response	NO _x = 808.4 ppb	NO = 805.5 ppb		*Percent Change	NO = 3.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	777.7	373.7	407.3	411.1	0.9907	100.9%
2nd GPT point (200 ppb O3)	777.7	576.1	204.9	204.0	1.0043	99.6%
3rd GPT point (100 ppb O3)	777.7	678.5	102.5	100.5	1.0197	98.1%
Average Correction Factor					1.0049	99.5%

Notes: Changed sample inlet filters after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 21, 2024	Previous Calibration:	February 13, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:41	End Time (MST):	14:49
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
802.9	797.6	1.0067			
402.0	397.9	1.0102			
200.5	199.3	1.0060			
			Slope	0.993092	0.90 - 1.10
			Intercept	-0.219382	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 21, 2024	Previous Calibration:	February 13, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:41	End Time (MST):	14:49
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
799.7	795.3	1.0055			
400.3	396.6	1.0094			
199.7	197.6	1.0105			
			Slope	0.994756	0.90 - 1.10
			Intercept	-0.699633	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 21, 2024	Previous Calibration:	February 13, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:41	End Time (MST):	14:49
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
407.3	411.1	0.9907		
204.9	204.0	1.0043		
102.5	100.5	1.0197		
			0.999913	
			1.011032	
			-1.734028	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 21, 2024

Location: Firebag





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS20 MACKAY RIVER MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 20, 2024
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:24	End Time (MST):	11:17
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999942	≥0.995
800.3	805.9	0.9930			
400.7	403.9	0.9920	Slope	1.004703	0.90 - 1.10
199.8	206.8	0.9663			
			Intercept	2.311780	+/-30



SO2 Calibration Plot

Date: March 7, 2024

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: MacKay River Station number: AMS20
 Calibration Date: March 5, 2024 Last Cal Date: February 1, 2024
 Start time (MST): 8:10 End time (MST): 12:11
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.12 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC515997
 Removed Cal Gas Conc: 5.12 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 1220
 ZAG Make/Model: Teledyne API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1236656117
 Converter make: Global Converter serial #: 2022-226
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.987249	0.975111	Backgd or Offset: 3.19	3.19
Calibration intercept:	0.499237	0.699032	Coeff or Slope: 1.113	1.113

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4922	78.1	80.0	78.7	1.019
as found 2nd point	4961	39.0	39.9	39.8	1.008
as found 3rd point	4980	19.5	20.0	20.4	0.989
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4922	78.1	80.0	78.6	1.018
second point	4961	39.0	40.0	39.8	1.004
third point	4980	19.5	20.0	20.5	0.975
as left zero	5000	0.0	0.0	0.4	----
as left span	4922	78.1	80.0	78.4	1.021
SO2 Scrubber Check	4982	81.3	802.8	-0.1	----
Date of last scrubber change:	May 25, 2023			Ave Corr Factor	0.999
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 78.5 Prev response: 79.45 *% change: -1.2%
 Baseline Corr 2nd AF pt: 39.6 AF Slope: 0.979587 AF Intercept: 0.519121
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999924

* = > +/-5% change initiates investigation

Notes: No maintenance and adjustments done. Sox scrubber checked after the calibrator zero.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 1, 2024
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:10	End Time (MST):	12:11
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999933	≥0.995
80.0	78.6	1.0183			
40.0	39.8	1.0042	Slope	0.975111	0.90 - 1.10
20.0	20.5	0.9749			
			Intercept	0.699032	+/-3



H₂S Calibration Plot

Date: March 5, 2024

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 20, 2024
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	8:24	End Time (MST):	11:16
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.02	----	Correlation Coefficient	0.999962	≥0.995
17.34	17.31	1.0019			
8.68	8.68	1.0004	Slope	0.997159	0.90 - 1.10
4.33	4.41	0.9820			
			Intercept	0.028212	+/-1.5



THC Calibration Plot

Date: March 7, 2024

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
as found span	4917	83.3	819.5	800.3	19.2	826.8	805.1	21.5	0.9911	0.9940
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.1	----	----
high point	4917	83.3	819.5	800.3	19.2	822.3	802.9	19.4	0.9965	0.9968
second point	4958	41.7	410.3	400.7	9.6	415.6	405.0	10.6	0.9872	0.9893
third point	4979	20.8	204.6	199.9	4.8	212.7	206.2	6.6	0.9621	0.9692
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.1	----	----
as left span	4917	83.3	819.5	442.5	377.0	818.4	441.3	377.0	1.0013	1.0027
Average Correction Factor									0.9819	0.9851

Corrected As found	NO _x = 827.0 ppb	NO = 805.1 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.0%
Previous Response	NO _x = 818.6 ppb	NO = 798.8 ppb		*Percent Change	NO = 0.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.7	441.9	377.0	375.8	1.0031	99.7%
2nd GPT point (200 ppb O3)	799.7	613.5	205.4	204.1	1.0062	99.4%
3rd GPT point (100 ppb O3)	799.7	702.6	116.3	114.2	1.0180	98.2%
Average Correction Factor					1.0091	99.1%

Notes:

No adjustments or maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 8, 2024
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	7:55	End Time (MST):	12:13
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20	
819.5	822.3	0.9965			
410.3	415.6	0.9872			
204.6	212.7	0.9621			
			Correlation Coefficient	0.999902	≥0.995
			Slope	1.001037	0.90 - 1.10
			Intercept	3.662215	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 8, 2024
Station Name:	Mackay River	Station Number:	AMS20
Start Time (MST):	7:55	End Time (MST):	12:13
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.3	802.9	0.9968			
400.7	405.0	0.9893			
199.9	206.2	0.9692			
			Slope	1.001142	0.90 - 1.10
			Intercept	2.942582	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 8, 2024
Station Name:	MacKay River	Station Number:	AMS20
Start Time (MST):	7:55	End Time (MST):	12:13
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
377.0	375.8	1.0031		
205.4	204.1	1.0062		
116.3	114.2	1.0180		



NO_x Calibration Plot

Date: March 6, 2024

Location: MacKay River





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS21
CONKLIN
MARCH 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 18, 2024	Previous Calibration:	February 8, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:41	End Time (MST):	11:44
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.5	----	Correlation Coefficient	0.999987	
800.8	800.0	1.0011			<i>≥0.995</i>
400.4	401.8	0.9966	Slope	0.997419	
200.1	202.9	0.9863			<i>0.90 - 1.10</i>
			Intercept	1.855817	<i>+/-30</i>



SO2 Calibration Plot

Date: March 18, 2024

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Conklin Station number: AMS21
 Calibration Date: March 21, 2024 Last Cal Date: February 12, 2024
 Start time (MST): 9:23 End time (MST): 14:55
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.00 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC501204
 Removed Cal Gas Conc: 5.00 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3810
 ZAG Make/Model: Teledyne API 701H Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1236656116
 Converter make: CD-Nova 101 Converter serial #: NA
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001571	1.000286	Backgd or Offset: 2.4	2.77
Calibration intercept:	0.420000	0.140000	Coeff or Slope: 0.980	0.998

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4920	80.0	80.0	79.6	1.005
as found 2nd point	4960	40.0	40.0	40.0	1.000
as found 3rd point	4980	20.0	20.0	20.3	0.985
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4920	80.0	80.0	80.0	1.000
second point	4960	40.0	40.0	40.5	0.988
third point	4980	20.0	20.0	20.1	0.995
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	80.0	80.0	79.7	1.004
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.994
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 79.6 Prev response: 80.55 *% change: -1.2%
 Baseline Corr 2nd AF pt: 40.0 AF Slope: 0.993571 AF Intercept: 0.200000
 Baseline Corr 3rd AF pt: 20.3 AF Correlation: 0.999970

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after multiple as founds. SO2 scrubber check done after calibrator zero and it passed. Adjusted zero and span.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 21, 2024	Previous Calibration:	February 12, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	9:23	End Time (MST):	14:55
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999952	
80.0	80.0	1.0000			≥0.995
40.0	40.5	0.9877	Slope	1.000286	
20.0	20.1	0.9950			0.90 - 1.10
			Intercept	0.140000	+/-3



TRS Calibration Plot

Date: March 21, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	March 5, 2024	Last Cal Date:	February 8, 2024
Start time (MST):	12:58	End time (MST):	17:51
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.71E-04	3.60E-04	NMHC SP Ratio:	7.57E-05
CH ₄ Retention time:	17.4	17.4	NMHC Peak Area:	120807
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				122143
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4920	80.2	17.13	16.77	1.021
as found 2nd point	4960	40.1	8.56	8.44	1.014
as found 3rd point	4980	20.0	4.28	4.26	1.004
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4920	80.2	17.13	17.11	1.001
second point	4960	40.1	8.56	8.63	0.992
third point	4980	20.0	4.28	4.39	0.976
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.09	1.002
Average Correction Factor					0.990
Baseline Corr AF:	16.75	Prev response	17.16	*% change	-2.4%
Baseline Corr 2nd AF:	8.4	AF Slope:	0.977405	AF Intercept:	0.047951
Baseline Corr 3rd AF:	4.3	AF Correlation:	0.999977	<i>* = +/-5% change initiates investigation</i>	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.15	0.999
as found 2nd point	4960	40.1	4.57	4.61	0.991
as found 3rd point	4980	20.0	2.28	2.33	0.981
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.12	1.002
second point	4960	40.1	4.57	4.62	0.990
third point	4980	20.0	2.28	2.34	0.976
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.10	1.005
Average Correction Factor					0.989
Baseline Corr AF:	9.15	Prev response	9.18	*% change	-0.3%
Baseline Corr 2nd AF:	4.6	AF Slope:	0.999605	AF Intercept:	0.025410
Baseline Corr 3rd AF:	2.3	AF Correlation:	0.999963	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.01	----
as found span	4920	80.2	7.99	7.62	1.048
as found 2nd point	4960	40.1	3.99	3.83	1.042
as found 3rd point	4980	20.0	2.00	1.93	1.032
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4920	80.2	7.99	7.99	0.999
second point	4960	40.1	3.99	4.01	0.996
third point	4980	20.0	2.00	2.05	0.976
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	8.00	0.999
Average Correction Factor					0.990
Baseline Corr AF:	7.61	Prev response	7.98	*% change	-5.0%
Baseline Corr 2nd AF:	3.82	AF Slope:	0.951755	AF Intercept:	0.023140
Baseline Corr 3rd AF:	1.92	AF Correlation:	0.999988	* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000154	0.996984
THC Cal Offset:	0.033769	0.062768
CH ₄ Cal Slope:	0.997762	0.998048
CH ₄ Cal Offset:	0.015158	0.026159
NMHC Cal Slope:	1.002131	0.996179
NMHC Cal Offset:	0.018811	0.036609

Notes: Changed sample inlet filter and pump after multi point as founds. Adjusted zero and span.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 8, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	12:58	End Time (MST):	17:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.01	----	Correlation Coefficient	0.999949	≥ 0.995			
17.13	17.11	1.0010						
8.56	8.63	0.9924				Slope	0.996984	0.90 - 1.10
4.28	4.39	0.9757						
			Intercept	0.062768	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 8, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	12:58	End Time (MST):	17:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999966	≥ 0.995
7.99	7.99	0.9995			
3.99	4.01	0.9955			
2.00	2.05	0.9758			
			Slope	0.998048	0.90 - 1.10
			Intercept	0.026159	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 5, 2024	Previous Calibration:	February 8, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	12:58	End Time (MST):	17:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999922	≥ 0.995			
9.14	9.12	1.0022						
4.57	4.62	0.9895				Slope	0.996179	0.90 - 1.10
2.28	2.34	0.9755						
			Intercept	0.036609	± 0.5			



NMHC Calibration Plot

Date: March 5, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	March 9, 2024	Last Cal Date:	March 5, 2024
Start time (MST):	12:10	End time (MST):	13:54
Reason:	Removal		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.60E-04	NA	NMHC SP Ratio:	7.48E-05
CH ₄ Retention time:	17.4	NA	NMHC Peak Area:	122143
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.54	0.976
as found 2nd point	4960	40.1	8.56	8.79	0.974
as found 3rd point	4980	20.0	4.28	4.46	0.960

new cylinder response

calibrator zero

high point

second point

third point

as left zero

as left span

			Average Correction Factor		
Baseline Corr AF:	17.54	Prev response	17.14	*% change	2.3%
Baseline Corr 2nd AF:	8.8	AF Slope:	1.022850	AF Intercept:	0.033987
Baseline Corr 3rd AF:	4.5	AF Correlation:	0.999979	* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.24	0.989
as found 2nd point	4960	40.1	4.57	4.65	0.984
as found 3rd point	4980	20.0	2.28	2.36	0.966
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	9.24	Prev response	9.14	*% change	1.1%
Baseline Corr 2nd AF:	4.6	AF Slope:	1.009284	AF Intercept:	0.027215
Baseline Corr 3rd AF:	2.4	AF Correlation:	0.999959	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	8.30	0.962
as found 2nd point	4960	40.1	3.99	4.14	0.964
as found 3rd point	4980	20.0	2.00	2.09	0.953
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	8.30	Prev response	8.00	*% change	3.7%
Baseline Corr 2nd AF:	4.14	AF Slope:	1.038417	AF Intercept:	0.006372
Baseline Corr 3rd AF:	2.09	AF Correlation:	0.999991	* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.996984	NA
THC Cal Offset:	0.062768	NA
CH ₄ Cal Slope:	0.998048	NA
CH ₄ Cal Offset:	0.026159	NA
NMHC Cal Slope:	0.996179	NA
NMHC Cal Offset:	0.036609	NA

Notes: Removed due to dipping occurring after calibration and maintenance done on March 5. Installing analyzer SN: 1118148494.

Calibration Performed By: Braiden Boutillier

NMHC Calibration Plot

Date: March 9, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	March 9, 2024	Last Cal Date:	March 5, 2024
Start time (MST):	13:50	End time (MST):	18:25
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	NA	3.06E-04	NMHC SP Ratio:	NA
CH ₄ Retention time:	NA	13.8	NMHC Peak Area:	NA
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				4.91E-05
				186157
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.07	1.003
second point	4960	40.1	8.56	8.63	0.992
third point	4980	20.0	4.28	4.38	0.978
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.09	1.002

				Average Correction Factor	0.991
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.08	1.007
second point	4960	40.1	4.57	4.62	0.990
third point	4980	20.0	2.28	2.34	0.976
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.06	1.009
Average Correction Factor					0.991
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.99	0.999
second point	4960	40.1	3.99	4.02	0.994
third point	4980	20.0	2.00	2.04	0.980
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	8.03	0.995
Average Correction Factor					0.991
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	NA	0.995216
THC Cal Offset:	NA	0.062766
CH ₄ Cal Slope:	NA	0.999651
CH ₄ Cal Offset:	NA	0.019559
NMHC Cal Slope:	NA	0.991840
NMHC Cal Offset:	NA	0.043208

Notes: Install calibration. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 9, 2024	Previous Calibration:	March 5, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	13:50	End Time (MST):	18:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999936	≥ 0.995			
17.13	17.07	1.0032						
8.56	8.63	0.9922				Slope	0.995216	0.90 - 1.10
4.28	4.38	0.9781						
			Intercept	0.062766	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 9, 2024	Previous Calibration:	March 5, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	13:50	End Time (MST):	18:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999972	≥0.995
7.99	7.99	0.9990			
3.99	4.02	0.9938			
2.00	2.04	0.9801			
			Slope	0.999651	0.90 - 1.10
			Intercept	0.019559	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 9, 2024	Previous Calibration:	March 5, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	13:50	End Time (MST):	18:25
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999884	≥ 0.995			
9.14	9.08	1.0065						
4.57	4.62	0.9900				Slope	0.991840	0.90 - 1.10
2.28	2.34	0.9764						
			Intercept	0.043208	± 0.5			



NMHC Calibration Plot

Date: March 9, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	March 15, 2024	Last Cal Date:	March 5, 2024
Start time (MST):	8:37	End time (MST):	10:44
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	NA	3.06E-04	NMHC SP Ratio:	NA
CH ₄ Retention time:	NA	13.8	NMHC Peak Area:	NA
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				4.91E-05
				186157
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.07	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	
high point	4920	80.2	16.86	16.86	1.000
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	1.000
Baseline Corr AF:	17.07	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.24	0.989
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.16	0.998
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	0.998
Baseline Corr AF:	9.24	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.83	1.020
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.70	1.037
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.037
Baseline Corr AF:	7.83	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	NA	1.000264
THC Cal Offset:	NA	0.000000
CH ₄ Cal Slope:	NA	0.964313
CH ₄ Cal Offset:	NA	0.000000
NMHC Cal Slope:	NA	1.002054
NMHC Cal Offset:	NA	0.000000


Notes: H2 and N2 cylinder changed after as founds. Sample inlet filters changed before calibrator zero. No adjustment made.

Calibration Performed By: Jan Castro

NMHC Calibration Plot

Date: March 15, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	March 18, 2024	Last Cal Date:	March 9, 2024
Start time (MST):	8:41	End time (MST):	11:44
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.06E-04	3.04E-04	NMHC SP Ratio:	4.91E-05
CH ₄ Retention time:	13.8	13.8	NMHC Peak Area:	186157
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				4.82E-05
				189786
				OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.45	0.982
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	17.15	0.999
second point	4960	40.1	8.56	8.63	0.992
third point	4980	20.0	4.28	4.41	0.971
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.07	1.003

Average Correction Factor				0.987
Baseline Corr AF:	17.45	Prev response	17.11	*% change 2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.37	0.976
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.14	1.000
second point	4960	40.1	4.57	4.62	0.989
third point	4980	20.0	2.28	2.36	0.968
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.11	1.003
Average Correction Factor					0.986
Baseline Corr AF:	9.37	Prev response	9.11	*% change	2.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	8.08	0.988
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	8.01	0.997
second point	4960	40.1	3.99	4.01	0.996
third point	4980	20.0	2.00	2.05	0.976
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.96	1.003
Average Correction Factor					0.990
Baseline Corr AF:	8.08	Prev response	8.00	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.995216	0.999299
THC Cal Offset:	0.062766	0.059171
CH ₄ Cal Slope:	0.999651	1.000967
CH ₄ Cal Offset:	0.019559	0.018960
NMHC Cal Slope:	0.991840	0.997617
NMHC Cal Offset:	0.043208	0.040611

Notes: Maintenance calibration to address baseline dipping. Changed sample inlet filters after as founds.
Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 18, 2024	Previous Calibration:	March 9, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:41	End Time (MST):	11:44
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999943	≥ 0.995			
17.13	17.15	0.9987						
8.56	8.63	0.9922				Slope	0.999299	0.90 - 1.10
4.28	4.41	0.9715						
			Intercept	0.059171	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 18, 2024	Previous Calibration:	March 9, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:41	End Time (MST):	11:44
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999965	≥ 0.995			
7.99	8.01	0.9973						
3.99	4.01	0.9955				Slope	1.000967	0.90 - 1.10
2.00	2.05	0.9758						
			Intercept	0.018960	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 18, 2024	Previous Calibration:	March 9, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:41	End Time (MST):	11:44
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999909	≥ 0.995			
9.14	9.14	1.0001						
4.57	4.62	0.9893				Slope	0.997617	0.90 - 1.10
2.28	2.36	0.9677						
			Intercept	0.040611	± 0.5			



NMHC Calibration Plot

Date: March 18, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	March 19, 2024	Last Cal Date:	March 18, 2024
Start time (MST):	9:34	End time (MST):	11:37
Reason:	Maintenance	Change ZAG	

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.04E-04	3.04E-04	NMHC SP Ratio:	4.82E-05
CH ₄ Retention time:	13.8	13.8	NMHC Peak Area:	189786
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	16.74	1.023
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	17.13	16.74	1.023
second point					
third point					
as left zero					
as left span					

				Average Correction Factor	1.023
Baseline Corr AF:	16.74	Prev response	17.11	*% change	-2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.11	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.11	1.003
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.003
Baseline Corr AF:	9.11	Prev response	9.11	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	7.63	1.046
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	7.99	7.63	1.046
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	1.046
Baseline Corr AF:	7.63	Prev response	8.00	*% change	-4.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.995216	0.977506
THC Cal Offset:	0.062766	0.000000
CH ₄ Cal Slope:	0.999651	0.955923
CH ₄ Cal Offset:	0.019559	0.000000
NMHC Cal Slope:	0.991840	0.996802
NMHC Cal Offset:	0.043208	0.000000


Notes: ZAG was replaced after as founds. No adjustment made.

Calibration Performed By: Jan Castro

NMHC Calibration Plot

Date: March 19, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	March 20, 2024	Last Cal Date:	March 18, 2024
Start time (MST):	9:23	End time (MST):	10:47
Reason:	Removal		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.9 ppm	CH ₄ Equiv Conc.	1067.7 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.04E-04	NA	NMHC SP Ratio:	4.82E-05
CH ₄ Retention time:	13.8	NA	NMHC Peak Area:	189786
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	17.13	17.09	1.002
as found 2nd point	4960	40.1	8.56	8.62	0.994
as found 3rd point	4980	20.0	4.27	4.40	0.971

new cylinder response

calibrator zero

high point

second point

third point

as left zero

as left span

			Average Correction Factor		
Baseline Corr AF:	17.09	Prev response	17.11	*% change	-0.1%
Baseline Corr 2nd AF:	8.6	AF Slope:	0.995642	AF Intercept:	0.069188
Baseline Corr 3rd AF:	4.4	AF Correlation:	0.999923	* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	9.14	9.04	1.011
as found 2nd point	4960	40.1	4.57	4.54	1.006
as found 3rd point	4980	20.0	2.28	2.33	0.978
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	9.04	Prev response	9.11	*% change	-0.8%
Baseline Corr 2nd AF:	4.5	AF Slope:	0.986540	AF Intercept:	0.035018
Baseline Corr 3rd AF:	2.3	AF Correlation:	0.999919	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.2	7.99	8.04	0.993
as found 2nd point	4960	40.1	3.99	4.07	0.981
as found 3rd point	4980	20.0	1.99	2.07	0.963
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero					
as left span					
				Average Correction Factor	
Baseline Corr AF:	8.04	Prev response	8.00	*% change	0.5%
Baseline Corr 2nd AF:	4.07	AF Slope:	1.004757	AF Intercept:	0.035968
Baseline Corr 3rd AF:	2.07	AF Correlation:	0.999907	* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.995216	NA
THC Cal Offset:	0.062766	NA
CH ₄ Cal Slope:	0.999651	NA
CH ₄ Cal Offset:	0.019559	NA
NMHC Cal Slope:	0.991840	NA
NMHC Cal Offset:	0.043208	NA


Notes: Baseline dipping during daily system checks. Removing instrument for further troubleshooting in the shop.

Calibration Performed By: Jan Castro

NMHC Calibration Plot

Date: March 20, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	March 20, 2024	Last Cal Date:	NA
Start time (MST):	11:24	End time (MST):	13:35
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
C3H8 Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
Removed C3H8 Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
ZAG make/model:	Teledyne API 701H	Serial Number:	953

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH4 Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	NA	2.55E-04	NMHC SP Ratio:	5.96E-05
CH4 Retention time:	NA	15.0	NMHC Peak Area:	153262
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4920	80.2	17.13	17.04	1.005
second point	4960	40.1	8.56	8.59	0.997
third point	4980	20.0	4.28	4.36	0.982
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	17.13	17.00	1.008

				Average Correction Factor	0.995
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.2	9.14	9.11	1.003
second point	4960	40.1	4.57	4.59	0.995
third point	4980	20.0	2.28	2.33	0.980
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	9.14	9.11	1.003
Average Correction Factor					0.993
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero					
as found span					
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.01	----
high point	4920	80.2	7.99	7.93	1.007
second point	4960	40.1	3.99	3.99	1.000
third point	4980	20.0	2.00	2.03	0.984
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.2	7.99	7.88	1.013
Average Correction Factor					0.997
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	NA	0.993227
THC Cal Offset:	NA	0.057164
CH ₄ Cal Slope:	NA	0.991351
CH ₄ Cal Offset:	NA	0.026556
NMHC Cal Slope:	NA	0.995229
NMHC Cal Offset:	NA	0.029409

Notes: Install calibration. Adjusted zero and span. Use zero chromatogram

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	NA
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:24	End Time (MST):	13:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.01	----	Correlation Coefficient	0.999959	≥ 0.995			
17.13	17.04	1.0050						
8.56	8.59	0.9968				Slope	0.993227	0.90 - 1.10
4.28	4.36	0.9822						
			Intercept	0.057164	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	NA
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:24	End Time (MST):	13:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999969	≥0.995
7.99	7.93	1.0066			
3.99	3.99	1.0000			
2.00	2.03	0.9845			
			Slope	0.991351	0.90 - 1.10
			Intercept	0.026556	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	NA
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	11:24	End Time (MST):	13:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999952	≥ 0.995			
9.14	9.11	1.0032						
4.57	4.59	0.9949				Slope	0.995229	0.90 - 1.10
2.28	2.33	0.9797						
			Intercept	0.029409	± 0.5			



NMHC Calibration Plot

Date: March 20, 2024

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Conklin
Calibration Date: March 26, 2024
Start time (MST): 8:56
Reason: Routine
Station number: AMS21
Last Cal Date: February 14, 2024
End time (MST): 13:29

Calibration Standards

NO Gas Cylinder #: SA18828
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 364

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1501663731

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.050	1.004	NO bkgnd or offset:	10.5	10.1
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	10.6	10.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	150.6	147.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.993108	0.994854
NO _x Cal Offset:	2.428003	2.228020
NO Cal Slope:	0.994636	1.000020
NO Cal Offset:	1.768002	1.008029
NO ₂ Cal Slope:	1.004592	0.996734
NO ₂ Cal Offset:	-0.278421	-0.362609



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as found span	4918	82.0	802.0	800.1	1.9	837.8	836.6	1.1	0.9572	0.9564
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4918	82.0	802.0	800.3	1.6	799.0	801.0	-1.7	1.0037	0.9992
second point	4959	41.0	401.0	400.2	0.8	402.2	401.2	1.0	0.9970	0.9974
third point	4980	20.5	200.5	200.1	0.4	203.8	202.4	1.4	0.9837	0.9884
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
as left span	4918	82.0	802.0	385.6	416.3	796.6	382.9	413.7	1.0067	1.0071
Average Correction Factor									0.9948	0.9950

Corrected As found	NO _x = 837.9 ppb	NO = 836.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 4.7%
Previous Response	NO _x = 798.9 ppb	NO = 797.6 ppb		*Percent Change	NO = 4.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI: ;

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.7	383.0	416.3	415.1	1.0030	99.7%
2nd GPT point (200 ppb O3)	797.7	593.2	206.1	204.1	1.0100	99.0%
3rd GPT point (100 ppb O3)	797.7	696.4	102.9	102.3	1.0063	99.4%
Average Correction Factor					1.0064	99.4%

Notes: Changed sample inlet filter after as founds. Adjusted Span only. Used 2nd NO reference point because of drift.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 14, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:56	End Time (MST):	13:29
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
802.0	799.0	1.0037		
401.0	402.2	0.9970		
200.5	203.8	0.9837		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 14, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:56	End Time (MST):	13:29
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.3	801.0	0.9992		
400.2	401.2	0.9974		
200.1	202.4	0.9884		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 14, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	8:56	End Time (MST):	13:29
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
416.3	415.1	1.0030			
206.1	204.1	1.0100			
102.9	102.3	1.0063			
			Slope	0.996734	0.90 - 1.10
			Intercept	-0.362609	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 26, 2024

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name:	Conklin	Station number:	AMS21
Calibration Date:	March 15, 2024	Last Cal Date:	February 1, 2024
Start time (MST):	10:47	End time (MST):	14:15
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	Teledyne API T700	Serial Number:	3810
ZAG Make/Model:	Teledyne API 701H	Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734
Analyzer Range:	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001543	0.999171	Backgd or Offset:	-1.2	-1.1
Calibration intercept:	0.180000	0.320000	Coeff or Slope:	0.998	0.998

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.2	----
as found span	5000	954.0	400.0	399.7	1.001
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.1	----
high point	5000	951.2	400.0	399.9	1.000
second point	5000	807.8	200.0	200.2	0.999
third point	5000	703.5	100.0	100.5	0.995
as left zero	5000	800.0	0.0	0.0	----
as left span	5000	951.6	400.0	401.0	0.998
Average Correction Factor					0.998

Baseline Corr As found:	399.5	Previous response	400.8	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Sample inlet filters changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 15, 2024	Previous Calibration:	February 1, 2024
Station Name:	Conklin	Station Number:	AMS21
Start Time (MST):	10:47	End Time (MST):	14:15
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999999	
400.0	399.9	1.0003			≥0.995
200.0	200.2	0.9990	Slope	0.999171	
100.0	100.5	0.9950			0.90 - 1.10
			Intercept	0.320000	+/- 5



O₃ Calibration Plot

Date: March 15, 2024

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Conklin Station number: AMS 21
 Calibration Date: March 18, 2024 Last Cal Date: February 12, 2024
 Start time (MST): 11:24 End time (MST): 12:17

Analyzer Make: API T640 S/N: 326
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	7.80	7.26	7.80	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	708.90	710.90	708.90	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.05	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39.00	----	39.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	4.30	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.90** Expiry Date: September 29, 2024
 Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.10	11.10	11.10	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: March 18, 2024
 Date Disposable Filter Changed: March 18, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: December 7, 2023
 Date RH/T Sensor Cleaned: December 7, 2023

Notes: Verified flow, pressure, temperature and pump power. Leak check passed. No adjustment made.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER
MARCH 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 1, 2024	Previous Calibration:	February 23, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:47	End Time (MST):	16:54
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.4	----	Correlation Coefficient	≥0.995
799.8	801.4	0.9980		
399.9	402.0	0.9947	Slope	0.90 - 1.10
200.4	201.5	0.9947		
			Intercept	+/-30



SO2 Calibration Plot

Date: March 1, 2024

Location: Janvier





Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 11, 2024	Previous Calibration:	March 1, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:50	End Time (MST):	13:16
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999996	≥0.995
799.8	799.8	1.0000			
399.9	401.5	0.9960	Slope	0.999806	0.90 - 1.10
200.4	201.0	0.9972			
			Intercept	0.664437	+/-30



SO2 Calibration Plot

Date: March 11, 2024

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Janvier Station number: AMS22
 Calibration Date: March 26, 2024 Last Cal Date: February 22, 2024
 Start time (MST): 10:52 End time (MST): 14:47
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.02 ppm Cal Gas Exp Date: November 15, 2026
 Cal Gas Cylinder #: CC424047
 Removed Cal Gas Conc: 5.02 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680031
 Converter make: CDN-101 Converter serial #: 587
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996235	0.989096	Backgd or Offset:	3.64 3.65
Calibration intercept:	0.440650	0.440430	Coeff or Slope:	1.188 1.188

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4920	79.7	80.0	79.8	1.005
as found 2nd point	4960	39.8	40.0	40.3	0.997
as found 3rd point	4980	19.9	20.0	20.3	0.994
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.3	----
high point	4920	79.7	80.0	79.5	1.007
second point	4960	39.8	40.0	40.1	0.997
third point	4980	19.9	20.0	20.3	0.984
as left zero	5000	0.0	0.0	0.4	----
as left span	4920	79.7	80.0	79.5	1.007
SO2 Scrubber Check	4920	79.8	798.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.996
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 79.6 Prev response: 80.16 *% change: -0.7%
 Baseline Corr 2nd AF pt: 40.1 AF Slope: 0.994237 AF Intercept: 0.360568
 Baseline Corr 3rd AF pt: 20.1 AF Correlation: 0.999974

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. Ran an SO2 scrubber check after the calibrator zero. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 22, 2024
Station Name:	Janvier	Station Number:	AMS22
Start Time (MST):	10:52	End Time (MST):	14:47
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999984	≥0.995
80.0	79.5	1.0066			
40.0	40.1	0.9965	Slope	0.989096	0.90 - 1.10
20.0	20.3	0.9842			
			Intercept	0.440430	+/-3



TRS Calibration Plot

Date: March 26, 2024

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	March 1, 2024	Last Cal Date:	February 29, 2024
Start time (MST):	12:47	End time (MST):	16:54
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH ₄ Cal Gas Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
C ₃ H ₈ Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
Removed C ₃ H ₈ Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.890E-04	2.390E-04	NMHC SP Ratio:	5.52E-05	5.42E-05
CH ₄ Retention time:	13.1	11.7	NMHC Peak Area:	165825	168687
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	17.04	1.008
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.15	1.001
second point	4960	39.9	8.59	8.51	1.009
third point	4980	20.0	4.30	4.27	1.009
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.13	1.002
Average Correction Factor					1.006

Baseline Corr AF:	17.04	Prev response	17.17	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	9.19	0.995
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.15	9.12	1.003
second point	4960	39.9	4.57	4.57	1.002
third point	4980	20.0	2.29	2.29	1.000
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.15	9.14	1.000
Average Correction Factor					1.002
Baseline Corr AF:	9.19	Prev response	9.15	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	7.86	1.022
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.03	0.999
second point	4960	39.9	4.01	3.94	1.018
third point	4980	20.0	2.01	1.98	1.018
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	7.99	1.005
Average Correction Factor					1.012
Baseline Corr AF:	7.86	Prev response	8.02	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999432	0.999011
THC Cal Offset:	0.007396	-0.026194
CH ₄ Cal Slope:	0.999247	1.001182
CH ₄ Cal Offset:	0.002236	-0.029561
NMHC Cal Slope:	0.999206	0.996756
NMHC Cal Offset:	0.005960	0.003768

Notes: Changed the inlet filter and H₂ cylinder after as founds. Adjusted the carrier run pressure. Adjusted the window timings. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 1, 2024	Previous Calibration:	February 29, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:47	End Time (MST):	16:54
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999981	≥ 0.995
17.17	17.15	1.0012			
8.59	8.51	1.0092			
4.30	4.27	1.0086			
			Slope	0.999011	0.90 - 1.10
			Intercept	-0.026194	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 1, 2024	Previous Calibration:	February 29, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:47	End Time (MST):	16:54
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999896	≥ 0.995			
8.03	8.03	0.9994						
4.01	3.94	1.0179				Slope	1.001182	0.90 - 1.10
2.01	1.98	1.0183						
			Intercept	-0.029561	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 1, 2024	Previous Calibration:	February 29, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:47	End Time (MST):	16:54
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
9.15	9.12	1.0031						
4.57	4.57	1.0018				Slope	0.996756	0.90 - 1.10
2.29	2.29	1.0002						
			Intercept	0.003768	± 0.5			



NMHC Calibration Plot

Date: March 1, 2024

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	March 22, 2024	Last Cal Date:	March 1, 2024
Start time (MST):	11:21	End time (MST):	12:46
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH ₄ Cal Gas Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
C ₃ H ₈ Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
Removed C ₃ H ₈ Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.390E-04	2.390E-04	NMHC SP Ratio:	5.42E-05	5.42E-05
CH ₄ Retention time:	11.7	11.7	NMHC Peak Area:	168687	168687
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	17.43	0.985
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.58	0.977
Average Correction Factor					

Baseline Corr AF:	17.43	Prev response	17.13	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	9.13	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.15	9.23	0.991
Average Correction Factor					
Baseline Corr AF:	9.13	Prev response	9.12	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	8.29	0.968
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero					
high point					
second point					
third point					
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.35	0.961
Average Correction Factor					
Baseline Corr AF:	8.29	Prev response	8.00	*% change	3.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999011	
THC Cal Offset:	-0.026194	
CH ₄ Cal Slope:	1.001182	
CH ₄ Cal Offset:	-0.029561	
NMHC Cal Slope:	0.996756	
NMHC Cal Offset:	0.003768	


Notes: Changed the N2 cylinder after as founds.

Calibration Performed By: Rene Chamberland

NMHC Calibration Plot

Date: March 22, 2024

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	March 25, 2024	Last Cal Date:	March 1, 2024
Start time (MST):	10:19	End time (MST):	13:53
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH ₄ Cal Gas Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
C ₃ H ₈ Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	502.8 ppm	CH ₄ Equiv Conc.	1075.9 ppm
Removed C ₃ H ₈ Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
ZAG make/model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.390E-04	2.360E-04	NMHC SP Ratio:	5.42E-05	5.52E-05
CH ₄ Retention time:	11.7	11.6	NMHC Peak Area:	168687	165650
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	17.17	17.27	0.994
as found 2nd point	4960	39.9	8.59	8.57	1.002
as found 3rd point	4980	20.0	4.30	4.28	1.006
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	17.17	17.17	1.000
second point	4960	39.9	8.59	8.54	1.006
third point	4980	20.0	4.30	4.27	1.008
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	17.17	17.20	0.998
Average Correction Factor					1.005

Baseline Corr AF:	17.27	Prev response	17.13	*% change	0.8%
Baseline Corr 2nd AF:	8.6	AF Slope:	1.006545	AF Intercept:	-0.033818
Baseline Corr 3rd AF:	4.3	AF Correlation:	0.999979	* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	9.15	9.04	1.012
as found 2nd point	4960	39.9	4.57	4.50	1.016
as found 3rd point	4980	20.0	2.29	2.25	1.019
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	9.15	9.14	1.001
second point	4960	39.9	4.57	4.57	1.001
third point	4980	20.0	2.29	2.30	0.999
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	9.15	9.18	0.996
Average Correction Factor					1.000
Baseline Corr AF:	9.04	Prev response	9.12	*% change	-0.9%
Baseline Corr 2nd AF:	4.5	AF Slope:	0.989007	AF Intercept:	-0.010210
Baseline Corr 3rd AF:	2.3	AF Correlation:	0.999993	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	79.8	8.03	8.23	0.975
as found 2nd point	4960	39.9	4.01	4.07	0.986
as found 3rd point	4980	20.0	2.01	2.03	0.992
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	79.8	8.03	8.03	0.999
second point	4960	39.9	4.01	3.97	1.011
third point	4980	20.0	2.01	1.97	1.019
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	79.8	8.03	8.02	1.001
Average Correction Factor					1.010
Baseline Corr AF:	8.23	Prev response	8.00	*% change	2.7%
Baseline Corr 2nd AF:	4.07	AF Slope:	1.026422	AF Intercept:	-0.023208
Baseline Corr 3rd AF:	2.03	AF Correlation:	0.999956	* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999011	1.000249
THC Cal Offset:	-0.026194	-0.023996
CH ₄ Cal Slope:	1.001182	1.001651
CH ₄ Cal Offset:	-0.029561	-0.024958
NMHC Cal Slope:	0.996756	0.999256
NMHC Cal Offset:	0.003768	0.000762

Notes: Changed the pump after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 25, 2024	Previous Calibration:	March 1, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:19	End Time (MST):	13:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999988	≥ 0.995
17.17	17.17	1.0001			
8.59	8.54	1.0060			
4.30	4.27	1.0083			
			Slope	1.000249	0.90 - 1.10
			Intercept	-0.023996	+/-0.5





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 25, 2024	Previous Calibration:	March 1, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:19	End Time (MST):	13:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999949	≥ 0.995
8.03	8.03	0.9994			
4.01	3.97	1.0109			
2.01	1.97	1.0194			
			Slope	1.001651	0.90 - 1.10
			Intercept	-0.024958	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 25, 2024	Previous Calibration:	March 1, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:19	End Time (MST):	13:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995			
9.15	9.14	1.0006						
4.57	4.57	1.0014				Slope	0.999256	0.90 - 1.10
2.29	2.30	0.9989						
			Intercept	0.000762	± 0.5			



NMHC Calibration Plot

Date: March 25, 2024

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Janvier
Calibration Date: March 27, 2024
Start time (MST): 10:25
Reason: Routine
Station number: AMS 22
Last Cal Date: February 28, 2024
End time (MST): 14:26

Calibration Standards

NO Gas Cylinder #: DT0047765
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: March 11, 2031
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3806
Serial Number: 691

Analyzer Information

Analyzer make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 833

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.840	0.853	NO bkgnd or offset:	0.0	0.0
NOX coeff or slope:	0.835	0.845	NOX bkgnd or offset:	-0.2	-0.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	6.3	6.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000594	1.003829
NO _x Cal Offset:	0.484102	0.024109
NO Cal Slope:	0.999460	1.005929
NO Cal Offset:	-0.576026	-0.915862
NO ₂ Cal Slope:	1.004914	0.998553
NO ₂ Cal Offset:	0.223696	-0.654801



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
as found span	4918	82.0	802.0	800.3	1.6	791.5	782.9	8.6	1.0132	1.0223
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.2	----	----
high point	4918	82.0	802.0	800.3	1.6	805.3	804.4	0.9	0.9959	0.9949
second point	4960	41.0	400.9	400.1	0.8	402.0	401.8	0.3	0.9973	0.9957
third point	4980	20.5	200.5	200.1	0.4	201.2	198.8	2.4	0.9964	1.0063
as left zero	5000	0.0	0.0	0.0	0.0	0.5	0.1	0.4	----	----
as left span	4918	82.0	802.0	400.6	401.3	797.6	397.0	400.6	1.0055	1.0091
Average Correction Factor									0.9965	0.9990

Corrected As found	NO _x = 791.3 ppb	NO = 782.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.5%	
Previous Response	NO _x = 802.9 ppb	NO = 799.3 ppb		*Percent Change	NO = -2.1%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO ₂)						
as found GPT point (200 ppb NO ₂)						
as found GPT point (100 ppb NO ₂)						
1st GPT point (400 ppb O ₃)	799.8	400.1	401.3	400.6	1.0018	99.8%
2nd GPT point (200 ppb O ₃)	799.8	606.7	194.7	193.1	1.0085	99.2%
3rd GPT point (100 ppb O ₃)	799.8	705.8	95.6	94.2	1.0153	98.5%
Average Correction Factor					1.0085	99.2%

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 27, 2024	Previous Calibration:	February 28, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:25	End Time (MST):	14:26
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
802.0	805.3	0.9959		
400.9	402.0	0.9973		
200.5	201.2	0.9964		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 27, 2024	Previous Calibration:	February 28, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:25	End Time (MST):	14:26
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.3	804.4	0.9949			
400.1	401.8	0.9957			
200.1	198.8	1.0063			
			Slope	1.005929	0.90 - 1.10
			Intercept	-0.915862	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 27, 2024	Previous Calibration:	February 28, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:25	End Time (MST):	14:26
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
401.3	400.6	1.0018		
194.7	193.1	1.0085		
95.6	94.2	1.0153		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 27, 2024

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Version-01-2020

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: March 22, 2024 Last Cal Date: February 20, 2024
 Start time (MST): 12:45 End time (MST): 15:04
 Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
 Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
 ZAG Make/Model: Teledyne API T701H Serial Number: 691

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 7046
 Analyzer Range 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003229	1.003829	Backgd or Offset:	2.2	2.2
Calibration intercept:	0.560000	0.480000	Coeff or Slope:	1.027	1.027

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	800.0	0.0	0.1	----
as found span	4895	905.3	400.0	402.2	0.995
as found 2nd point					
as found 3rd point					
calibrator zero	5000	800.0	0.0	0.2	----
high point	4895	905.3	400.0	401.8	0.996
second point	4895	756.7	200.0	201.6	0.992
third point	4895	656.1	100.0	101.0	0.990
as left zero	5000	800.0	0.0	1.0	----
as left span	4895	904.3	400.0	404.5	0.989
Average Correction Factor					0.993

Baseline Corr As found:	402.1	Previous response	401.9	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

O₃ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 22, 2024	Previous Calibration:	February 20, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:45	End Time (MST):	15:04
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999997	
400.0	401.8	0.9955			≥0.995
200.0	201.6	0.9921	Slope	1.003829	
100.0	101.0	0.9901			0.90 - 1.10
			Intercept	0.480000	+/- 5



O₃ Calibration Plot

Date: March 22, 2024

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: March 26, 2024 Last Cal Date: February 28, 2024
 Start time (MST): 11:09 End time (MST): 12:13

Analyzer Make: Teledyne API T640 S/N: 325
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752
 Temp/RH standard: Alicat FP-25BT S/N: 388752

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-1.5	-1.7	-1.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	714.9	715.71	714.9	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.078	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39	----	39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.3	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: _____ January 24, 2024
 Date Disposable Filter Changed: _____ January 24, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ July 26, 2023
 Date RH/T Sensor Cleaned: _____ July 26, 2023

Notes: Verified flow, temperature, and pressure. Leak check passed.

Calibration by: Rene Chamberland



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23 FORT HILLS MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	March 8, 2024	Last Cal Date:	February 7, 2024
Start time (MST):	8:46	End time (MST):	11:34
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.76	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC281425			
Removed Cal Gas Conc:	49.76	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	451
ZAG Make/Model:	API T701		Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012
Analyzer Range	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000522	1.003512	Backgd or Offset:	18.5	18.5
Calibration intercept:	-0.903438	-0.524217	Coeff or Slope:	1.063	1.063

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4920	80.3	799.1	802.5	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.2	----
high point	4920	80.3	799.1	801.7	0.997
second point	4960	40.2	400.1	400.7	0.998
third point	4980	20.1	200.0	199.4	1.003
as left zero	5000	0.0	0.0	0.2	----
as left span	4920	80.3	799.1	804.9	0.993
Average Correction Factor					0.999

Baseline Corr As found:	802.60	Previous response	798.61	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 8, 2024	Previous Calibration:	February 7, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:46	End Time (MST):	11:34
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999996	≥0.995
799.1	801.7	0.9968			
400.1	400.7	0.9984	Slope	1.003512	0.90 - 1.10
200.0	199.4	1.0032			
			Intercept	-0.524217	+/-30



SO2 Calibration Plot

Date: March 8, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Fort Hills Station number: AMS23
 Calibration Date: March 13, 2024 Last Cal Date: February 13, 2024
 Start time (MST): 7:00 End time (MST): 11:26
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.20 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517372
 Removed Cal Gas Conc: 5.20 ppm Rem Gas Exp Date: N/A
 Removed Gas Cyl #: N/A Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 451
 ZAG Make/Model: API T701 Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1300156232
 Converter make: CDN-101 Converter serial #: 594
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010172	0.991459	Backgd or Offset:	1.96
Calibration intercept:	-0.218012	-0.038309	Coeff or Slope:	1.160

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.0	----
as found span	4923	77.0	80.0	77.9	1.027
as found 2nd point	4962	38.5	40.0	39.3	1.018
as found 3rd point	4981	19.2	19.9	19.7	1.013
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4923	77.0	80.0	79.3	1.009
second point	4962	38.5	40.0	39.6	1.010
third point	4981	19.2	19.9	19.7	1.013
as left zero	5000	0.0	0.0	0.0	----
as left span	4923	77.0	80.0	80.9	0.989
SO2 Scrubber Check	4920	80.3	803.0	0.0	----

Date of last scrubber change:		Ave Corr Factor	1.010
Date of last converter efficiency test:	March 13, 2024	102.7% efficiency	

Baseline Corr As found: 77.9 Prev response: 80.60 *% change: -3.5%
 Baseline Corr 2nd AF pt: 39.3 AF Slope: 0.973033 AF Intercept: 0.181374
 Baseline Corr 3rd AF pt: 19.7 AF Correlation: 0.999970

* = > +/-5% change initiates investigation

Notes: Converter efficiency test done after the third as found point at 102.7%. Pump changed out. SOx scrubber checked after the third point. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 13, 2024	Previous Calibration:	February 13, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	7:00	End Time (MST):	11:26
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999999	
80.0	79.3	1.0089			≥0.995
40.0	39.6	1.0100	Slope	0.991459	
19.9	19.7	1.0126			0.90 - 1.10
			Intercept	-0.038309	+/-3



TRS Calibration Plot

Date: March 13, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

Station Information

Station Name:	Fort Hills	Station number:	AMS23
Calibration Date:	March 8, 2024	Last Cal Date:	February 7, 2024
Start time (MST):	8:46	End time (MST):	11:33
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH ₄ Cal Gas Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH ₄ Conc.	500.2 ppm	CH ₄ Equiv Conc.	1070.6 ppm
Removed C ₃ H ₈ Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
ZAG make/model:	API T701	Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.31E-04	2.35E-04	NMHC SP Ratio:	4.93E-05
CH ₄ Retention time:	13.0	13.2	NMHC Peak Area:	186451
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	17.19	17.29	0.994
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	17.19	17.27	0.995
second point	4960	40.2	8.61	8.65	0.996
third point	4980	20.1	4.30	4.35	0.988
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	17.19	17.24	0.997
Average Correction Factor					0.993

Baseline Corr AF:	17.29	Prev response	17.21	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-06-2022

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	9.16	9.23	0.993
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	9.16	9.16	1.000
second point	4960	40.2	4.59	4.63	0.991
third point	4980	20.1	2.29	2.36	0.970
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	9.16	9.14	1.002
Average Correction Factor					0.987
Baseline Corr AF:	9.23	Prev response	9.21	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4920	80.3	8.03	8.07	0.996
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4920	80.3	8.03	8.12	0.989
second point	4960	40.2	4.02	4.02	1.000
third point	4980	20.1	2.01	1.99	1.010
as left zero	5000	0.0	0.00	0.00	----
as left span	4920	80.3	8.03	8.11	0.991
Average Correction Factor					1.000
Baseline Corr AF:	8.07	Prev response	8.00	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999507	1.004093
THC Cal Offset:	0.025608	0.011596
CH ₄ Cal Slope:	0.997953	1.012136
CH ₄ Cal Offset:	-0.014037	-0.025657
NMHC Cal Slope:	1.000996	0.997614
NMHC Cal Offset:	0.041641	0.036452

Notes: Span adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 8, 2024	Previous Calibration:	February 7, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:46	End Time (MST):	11:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995			
17.19	17.27	0.9953						
8.61	8.65	0.9956				Slope	1.004093	0.90 - 1.10
4.30	4.35	0.9884						
			Intercept	0.011596	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 8, 2024	Previous Calibration:	February 7, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:46	End Time (MST):	11:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999949	≥ 0.995			
8.03	8.12	0.9891						
4.02	4.02	1.0001				Slope	1.012136	0.90 - 1.10
2.01	1.99	1.0099						
			Intercept	-0.025657	± 0.5			





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-06-2022

Station Information

Calibration Date:	March 8, 2024	Previous Calibration:	February 7, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	8:46	End Time (MST):	11:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999927	≥ 0.995			
9.16	9.16	1.0002						
4.59	4.63	0.9914				Slope	0.997614	0.90 - 1.10
2.29	2.36	0.9703						
			Intercept	0.036452	± 0.5			



NMHC Calibration Plot

Date: March 8, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Fort Hills
Calibration Date: March 12, 2024
Start time (MST): 6:55
Reason: Routine
Station number: AMS23
Last Cal Date: February 2, 2024
End time (MST): 11:35

Calibration Standards

NO Gas Cylinder #: CC358149
NOX Cal Gas Conc: 60.3 ppm
Removed Cylinder #: CC332703
Removed Gas NOX Conc: 49.7 ppm
NOX gas Diff: 1.0%
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.1 ppm
Removed Gas Exp Date: January 28, 2024
Removed Gas NO Conc: 49.7 ppm
NO gas Diff: 2.1%
Serial Number: 451
Serial Number: 5611

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1152430007

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.102	1.132	NO bkgnd or offset:	3	3.1
NOX coeff or slope:	0.990	0.996	NOX bkgnd or offset:	3.2	3.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	170.3	170.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002680	0.998626
NO _x Cal Offset:	0.104366	-0.173464
NO Cal Slope:	1.001767	0.998303
NO Cal Offset:	-0.976340	-1.451679
NO ₂ Cal Slope:	1.002767	1.002264
NO ₂ Cal Offset:	0.602578	0.618019



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	-0.1	----	----
as found span	4920	80.5	800.2	800.2	0.0	767.0	762.9	4.1	1.043	1.049
as found 2nd										
as found 3rd										
new cyl resp	4934	66.3	799.5	796.9	2.7	774.1	775.2	-1.1	1.0329	1.0280
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.0	----	----
high point	4934	66.3	799.5	796.9	2.7	798.4	794.9	3.5	1.001	1.002
second point	4967	33.2	400.4	399.0	1.3	399.6	396.1	3.5	1.002	1.007
third point	4983	16.6	200.2	199.5	0.7	199.3	196.1	3.3	1.005	1.018
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
as left span	4934	66.3	799.5	479.3	320.3	801.0	477.0	324.1	0.998	1.005
Average Correction Factor									1.003	1.009

Corrected As found	NO _x = 767.0 ppb	NO = 762.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -4.6%	
Previous Response	NO _x = 802.4 ppb	NO = 800.6 ppb		*Percent Change	NO = -4.9%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	792.8	475.2	320.3	321.1	0.997	100.3%
2nd GPT point (200 ppb O3)	792.8	636.4	159.1	160.9	0.989	101.2%
3rd GPT point (100 ppb O3)	792.8	712.7	82.8	83.8	0.987	101.3%
Average Correction Factor					0.991	100.9%

Notes:

Calibration gas changed. Span adjusted.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 2, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	6:55	End Time (MST):	11:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	≥0.995	
799.5	798.4	1.0014			
400.4	399.6	1.0019			
200.2	199.3	1.0046			
			Slope	0.998626	0.90 - 1.10
			Intercept	-0.173464	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 2, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	6:55	End Time (MST):	11:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
796.9	794.9	1.0025		
399.0	396.1	1.0074		
199.5	196.1	1.0176		
			0.999980	
			0.998303	
			-1.451679	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 2, 2024
Station Name:	Fort Hills	Station Number:	AMS23
Start Time (MST):	6:55	End Time (MST):	11:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
320.3	321.1	0.9974		
159.1	160.9	0.9885		
82.8	83.8	0.9875		



NO_x Calibration Plot

Date: March 12, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Hills Station number: AMS 23
 Calibration Date: March 8, 2024 Last Cal Date: February 7, 2024
 Start time (MST): 8:04 End time (MST): 8:39

Analyzer Make: API T640 S/N: 1546
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-11.3	-11.5	-11.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	735.6	735	735.6	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5	4.86	5	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	44	----	44	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	6.8	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 10-Jun-24
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 7, 2024
 Date Disposable Filter Changed: February 7, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 17, 2023
 Date RH/T Sensor Cleaned: October 17, 2023

Notes: No adjustments done. Leak check passed.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25 WASKŌW OHCI PIMÂTISIWIN MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 11, 2024	Previous Calibration:	February 16, 2024
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	6:53	End Time (MST):	11:11
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999999	≥0.995
800.1	801.9	0.9977			
399.6	401.2	0.9959	Slope	1.002469	0.90 - 1.10
199.8	200.0	0.9990			
			Intercept	0.048260	+/-30



SO2 Calibration Plot

Date: March 11, 2024

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number:	AMS25
Calibration Date:	March 18, 2024	Last Cal Date:	February 14, 2024
Start time (MST):	6:20	End time (MST):	10:30
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.97	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC517099			
Removed Cal Gas Conc:	4.97	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	747
ZAG Make/Model:	API T701		Serial Number:	261

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146
Converter make:	Global G-150	Converter serial #:	2022-219
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005840	1.008429	Backgd or Offset:	3.30	3.30
Calibration intercept:	0.060000	0.120000	Coeff or Slope:	1.113	1.113

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4920	80.0	79.5	80.5	0.986
as found 2nd point	4960	40.0	39.7	40.3	0.983
as found 3rd point	4980	20.0	19.9	20.2	0.979
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4920	80.0	79.5	80.2	0.991
second point	4960	40.0	39.7	40.3	0.986
third point	4980	20.0	19.9	20.1	0.988
as left zero	5000	0.0	0.0	0.1	----
as left span	4920	80.0	800.0	809.5	0.988
SO2 Scrubber Check	4921	79.2	800.0	0.0	----
Date of last scrubber change:	20-Jun-23			Ave Corr Factor	0.988
Date of last converter efficiency test:					efficiency

Baseline Corr As found:	80.6	Prev response:	79.98	*% change:	0.8%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.013894	AF Intercept:	-0.020000
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999995		

* = > +/-5% change initiates investigation

Notes: SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 18, 2024	Previous Calibration:	February 14, 2024
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS25
Start Time (MST):	6:20	End Time (MST):	10:30
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999995	
79.5	80.2	0.9907			≥0.995
39.7	40.3	0.9858	Slope	1.008429	
19.9	20.1	0.9883			0.90 - 1.10
			Intercept	0.120000	+/-3



H₂S Calibration Plot

Date: March 18, 2024

Location: Waskow ohci Pimatisiwin





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS26 CHRISTINA LAKE MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 22, 2024	Previous Calibration:	February 3, 2024
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	9:42	End Time (MST):	12:28
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430005

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.7	----	Correlation Coefficient	0.999999	≥0.995
800.9	799.7	1.0015			
400.4	400.2	1.0005	Slope	0.997322	0.90 - 1.10
200.2	201.1	0.9956			
			Intercept	0.975976	+/-30



SO2 Calibration Plot

Date: March 22, 2024

Location: Christina Lake





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Christina Lake Station number: AMS26
 Calibration Date: March 12, 2024 Last Cal Date: February 7, 2024
 Start time (MST): 9:49 End time (MST): 14:18
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.05 ppm Cal Gas Exp Date: November 15, 2026
 Cal Gas Cylinder #: DT0014831
 Removed Cal Gas Conc: 5.05 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3253
 ZAG Make/Model: API T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 450i Analyzer serial #: 1180030032
 Converter make: NA Converter serial #: NA
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008427	0.996996	Backgd or Offset:	35.0 34.5
Calibration intercept:	0.258384	0.358415	Coeff or Slope:	1.086 1.064

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.8	----
as found span	4921	79.2	80.0	81.8	0.988
as found 2nd point	4960	39.6	40.0	41.7	0.978
as found 3rd point	4980	19.8	20.0	21.4	0.971
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.4	----
high point	4921	79.2	80.0	80.2	0.997
second point	4960	39.6	40.0	40.0	1.000
third point	4980	19.8	20.0	20.4	0.980
as left zero	5000	0.0	0.0	0.5	----
as left span	4921	79.2	80.0	80.0	1.000
SO2 Scrubber Check	4919	80.8	808.0	-0.1	----
Date of last scrubber change:	27-Feb-19			Ave Corr Factor	0.993
Date of last converter efficiency test:	efficiency				

Baseline Corr As found: 81.0 Prev response: 80.92 *% change: 0.1%
 Baseline Corr 2nd AF pt: 40.9 AF Slope: 1.011713 AF Intercept: 1.018369
 Baseline Corr 3rd AF pt: 20.6 AF Correlation: 0.999962

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filters after multipoint as founds. SO2 scrubber check after calibrator zero done and passed. Adjusted zero and span.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 7, 2024
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	9:49	End Time (MST):	14:18
Analyzer make:	Thermo 450i	Analyzer serial #:	1180030032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999978	≥0.995
80.0	80.2	0.9974			
40.0	40.0	1.0000	Slope	0.996996	0.90 - 1.10
20.0	20.4	0.9803			
			Intercept	0.358415	+/-3



H₂S Calibration Plot

Date: March 12, 2024

Location: Christina Lake





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.9	-0.2	-0.7	----	----
as found span	4918	82.1	802.9	799.6	3.3	811.1	804.8	6.3	0.9899	0.9936
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.7	0.0	-0.7	----	----
high point	4918	82.1	802.9	799.6	3.3	803.0	799.2	3.8	0.9999	1.0005
second point	4959	41.1	401.9	400.3	1.6	403.5	401.0	2.4	0.9962	0.9983
third point	4980	20.5	200.5	199.7	0.8	202.7	200.6	2.1	0.9890	0.9953
as left zero	5000	0.0	0.0	0.0	0.0	-0.6	0.1	-0.7	----	----
as left span	4918	82.1	802.9	421.0	381.9	800.4	423.2	377.1	1.0032	0.9949
Average Correction Factor									0.9950	0.9980

Corrected As found	NO _x = 812.0 ppb	NO = 805.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.1%
Previous Response	NO _x = 803.1 ppb	NO = 798.8 ppb		*Percent Change	NO = 0.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	797.5	418.9	381.9	383.4	0.9960	100.4%
2nd GPT point (200 ppb O3)	797.5	606.8	194.0	195.4	0.9928	100.7%
3rd GPT point (100 ppb O3)	797.5	703.2	97.6	99.1	0.9847	101.6%
Average Correction Factor					0.9912	100.9%

Notes:

Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By:

Jan Castro



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 13, 2024	Previous Calibration:	February 13, 2024
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	9:33	End Time (MST):	13:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.7	----	Correlation Coefficient	≥0.995	
802.9	803.0	0.9999			
401.9	403.5	0.9962			
200.5	202.7	0.9890			
			Slope	1.000066	0.90 - 1.10
			Intercept	0.766423	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 13, 2024	Previous Calibration:	February 13, 2024
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	9:33	End Time (MST):	13:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	≥0.995	
799.6	799.2	1.0005			
400.3	401.0	0.9983			
199.7	200.6	0.9953			
			Slope	0.999129	0.90 - 1.10
			Intercept	0.606305	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 13, 2024	Previous Calibration:	February 13, 2024
Station Name:	Christina Lake	Station Number:	AMS26
Start Time (MST):	9:33	End Time (MST):	13:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	-0.7	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
381.9	383.4	0.9960		
194.0	195.4	0.9928		
97.6	99.1	0.9847		



NO_x Calibration Plot

Date: March 13, 2024

Location: Christina Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27 JACKFISH 2/3 MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Version-01-2020

Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	March 8, 2024	Last Cal Date:	February 14, 2024
Start time (MST):	11:47	End time (MST):	14:46
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	<u>SG9133974BAL</u>			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3811
ZAG Make/Model:	API 701		Serial Number:	268

Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 12124313138
 Analyzer Range 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006935	1.003248	Backgd or Offset:	8.4	8.7
Calibration intercept:	-2.579017	-1.638441	Coeff or Slope:	0.960	0.983

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.2	----
as found span	4921	79.1	800.2	783.5	1.021
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.0	0.5	----
high point	4921	79.1	800.2	802.2	0.997
second point	4961	39.5	399.5	398.1	1.004
third point	4980	19.8	200.3	197.2	1.016
as left zero	5000	0.0	0.0	0.5	----
as left span	4921	79.1	800.2	807.8	0.991
Average Correction Factor					1.006

Baseline Corr As found:	783.30	Previous response	803.13	*% change	-2.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 8, 2024	Previous Calibration:	February 14, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:47	End Time (MST):	14:46
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.5	----	Correlation Coefficient	0.999968	≥0.995
800.2	802.2	0.9975			
399.5	398.1	1.0036	Slope	1.003248	0.90 - 1.10
200.3	197.2	1.0157			
			Intercept	-1.638441	+/-30



SO2 Calibration Plot

Date: March 8, 2024

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Jackfish 2/3 Station number: AMS27
 Calibration Date: March 12, 2024 Last Cal Date: February 22, 2024
 Start time (MST): 8:57 End time (MST): 13:18
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.41 ppm Cal Gas Exp Date: January 4, 2025
 Cal Gas Cylinder #: CC345023
 Removed Cal Gas Conc: 5.41 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3811
 ZAG Make/Model: API 701 Serial Number: 268

Analyzer Information

Analyzer make: API T101 Analyzer serial #: 621
 Converter make: Converter serial #:
 Analyzer Range 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.013890	1.014173	Backgd or Offset: 29.9	29.9
Calibration intercept:	-0.217825	-0.277753	Coeff or Slope: 0.965	0.965

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.3	----
as found span	4926	74.1	80.2	81.9	0.983
as found 2nd point	4963	37.0	40.0	40.6	0.993
as found 3rd point	4982	18.5	20.0	19.9	1.021
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.1	----
high point	4926	74.1	80.2	81.2	0.987
second point	4963	37.0	40.0	40.2	0.996
third point	4982	18.5	20.0	19.6	1.021
as left zero	5000	0.0	0.0	0.2	----
as left span	4926	74.1	80.2	81.1	0.989
SO2 Scrubber Check	4921	79.1	791.0	0.1	----

Date of last scrubber change:	Ave Corr Factor	1.001
Date of last converter efficiency test:	efficiency	

Baseline Corr As found: 81.6 Prev response: 81.07 *% change: 0.6%
 Baseline Corr 2nd AF pt: 40.3 AF Slope: 1.020445 AF Intercept: -0.097632
 Baseline Corr 3rd AF pt: 19.6 AF Correlation: 0.999891

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 12, 2024	Previous Calibration:	February 22, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	8:57	End Time (MST):	13:18
Analyzer make:	API T101	Analyzer serial #:	621

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999900	
80.2	81.2	0.9874			≥0.995
40.0	40.2	0.9959	Slope	1.014173	
20.0	19.6	1.0212			0.90 - 1.10
			Intercept	-0.277753	+/-3



H₂S Calibration Plot

Date: March 12, 2024

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
as found span	4921	79.4	816.8	800.3	16.5	806.4	787.3	19.2	1.0129	1.0165
as found 2nd										
as found 3rd										
new cyl resp	4942	66.5	800.6	799.3	1.3	793.6	793.2	0.4	1.0089	1.0077
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
high point	4942	66.5	800.6	799.3	1.3	797.5	798.7	-1.3	1.0039	1.0008
second point	4979	33.3	400.6	399.9	0.7	395.3	394.6	0.7	1.0134	1.0136
third point	4996	16.6	199.7	199.4	0.3	191.5	190.6	0.8	1.0428	1.0460
as left zero	5000	0.0	0.0	0.0	0.0	1.9	1.0	1.0	----	----
as left span	4942	66.5	800.6	417.2	390.6	796.2	412.8	383.4	1.0056	1.0107
Average Correction Factor									1.0200	1.0201

Corrected As found	NO _x = 806.6 ppb	NO = 787.4 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.4%	
Previous Response	NO _x = 817.9 ppb	NO = 800.9 ppb		*Percent Change	NO = -1.7%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.7	409.4	390.6	386.3	1.0112	98.9%
2nd GPT point (200 ppb O3)	798.7	622.9	177.1	175.5	1.0093	99.1%
3rd GPT point (100 ppb O3)	798.7	710.4	89.6	86.1	1.0410	96.1%
Average Correction Factor					1.0205	98.0%

Notes: Changed the NO calibration gas cylinder after as founds. Changed the inlet filter. Adjusted span only

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 19, 2024	Previous Calibration:	February 22, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	11:48	End Time (MST):	18:16
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.1	----	Correlation Coefficient	≥0.995	
800.6	797.5	1.0039			
400.6	395.3	1.0134			
199.7	191.5	1.0428			
			Slope	0.998635	0.90 - 1.10
			Intercept	-3.655643	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 19, 2024	Previous Calibration:	February 22, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	11:48	End Time (MST):	18:16
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999853	≥0.995
799.3	798.7	1.0008			
399.9	394.6	1.0136	Slope	1.002024	0.90 - 1.10
199.4	190.6	1.0460			
			Intercept	-4.335475	+/-20





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 19, 2024	Previous Calibration:	February 22, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS27
Start Time (MST):	11:48	End Time (MST):	18:16
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
390.6	386.3	1.0112		
177.1	175.5	1.0093		
89.6	86.1	1.0410		



NO_x Calibration Plot

Date: March 19, 2024

Location: Jackfish 2/3





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS29 SURMONT 2 MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 8, 2024	Previous Calibration:	February 5, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:55	End Time (MST):	14:36
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999977	≥0.995
800.1	805.0	0.9939			
400.6	400.5	1.0002	Slope	1.007141	0.90 - 1.10
199.8	197.8	1.0102			
			Intercept	-1.805499	+/-30



SO2 Calibration Plot

Date: March 8, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 7, 2024	Previous Calibration:	February 12, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	11:11	End Time (MST):	16:14
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.1	----	Correlation Coefficient	0.999992	≥0.995
80.0	79.9	1.0013			
40.1	39.9	1.0052	Slope	0.997613	0.90 - 1.10
20.1	20.0	1.0027			
			Intercept	0.017447	+/-3



H₂S Calibration Plot

Date: March 7, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Report

Version-01-2020

Station Information

Station Name:	Surmont 2	Station number:	AMS29
Calibration Date:	March 8, 2024	Last Cal Date:	February 26, 2024
Start time (MST):	10:55	End time (MST):	14:36
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC356008	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
C3H8 Cal Gas Conc.	<u>205.7</u> ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	<u>499.0</u> ppm	CH4 Equiv Conc.	1064.7 ppm
Removed C3H8 Conc.	<u>205.7</u> ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5472
ZAG Make/Model:	Teledyne API T701	Serial Number:	4698

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1170050149
 Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995247	1.004976	Background:	3.62	3.64
Calibration intercept:	-0.136283	-0.063439	Coefficient:	4.021	4.046

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.00	-0.09	----
as found span	4918	81.3	17.31	17.20	1.007
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	-0.02	----
high point	4918	81.3	17.31	17.36	0.997
second point	4959	40.6	8.65	8.60	1.005
third point	4979	20.3	4.32	4.24	1.020
as left zero	5000	0.0	0.00	-0.07	----
as left span	4918	81.3	17.31	17.44	0.993
Average Correction Factor					1.008
Baseline Corr As found:	17.29	Previous response	17.10	*% change	1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:			

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Small span adjustment made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 8, 2024	Previous Calibration:	February 26, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	10:55	End Time (MST):	14:36
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	-0.02	----	Correlation Coefficient	0.999970	≥0.995
17.31	17.36	0.9974			
8.65	8.60	1.0053	Slope	1.004976	0.90 - 1.10
4.32	4.24	1.0199			
			Intercept	-0.063439	+/-1.5



THC Calibration Plot

Date: March 8, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
as found span	4916	84.2	799.2	799.2	0.0	787.5	784.7	2.8	1.0149	1.0185
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
high point	4916	84.2	799.2	799.2	0.0	798.9	798.4	0.5	1.0004	1.0010
second point	4958	42.1	399.6	399.6	0.0	399.5	398.6	0.9	1.0003	1.0025
third point	4979	21.1	200.3	200.3	0.0	197.8	196.9	0.9	1.0125	1.0172
as left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as left span	4916	84.2	799.2	409.5	389.7	793.8	403.3	390.5	1.0068	1.0154
Average Correction Factor									1.0044	1.0069

Corrected As found	NO _x = 787.7 ppb	NO = 784.9 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.3%	
Previous Response	NO _x = 797.8 ppb	NO = 796.4 ppb		*Percent Change	NO = -1.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	793.2	403.5	389.7	389.1	1.0015	99.8%
2nd GPT point (200 ppb O3)	793.2	602.6	190.6	191.3	0.9963	100.4%
3rd GPT point (100 ppb O3)	793.2	696.5	96.7	96.4	1.0031	99.7%
Average Correction Factor					1.0003	100.0%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 6, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	11:05	End Time (MST):	16:48
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.2	798.9	1.0004		
399.6	399.5	1.0003		
200.3	197.8	1.0125		
			0.999989	
			1.000767	
			-1.012410	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 6, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	11:05	End Time (MST):	16:48
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥ 0.995 0.90 - 1.10 +/-20
799.2	798.4	1.0010		
399.6	398.6	1.0025		
200.3	196.9	1.0172		
			0.999983	
			1.000380	
			-1.452150	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 6, 2024
Station Name:	Surmont 2	Station Number:	AMS29
Start Time (MST):	11:05	End Time (MST):	16:48
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
389.7	389.1	1.0015		
190.6	191.3	0.9963		
96.7	96.4	1.0031		


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 6, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Surmont Station number: AMS 29
 Calibration Date: March 8, 2024 Last Cal Date: February 12, 2024
 Start time (MST): 10:59 End time (MST): 11:20

Analyzer Make: API T640 S/N: 253
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-0.9	-0.43	-0.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	710.0	711.07	710.0	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	5.128	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	36	----	36	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.1	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 12, 2024
 Date Disposable Filter Changed: February 12, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 25, 2023
 Date RH/T Sensor Cleaned: October 25, 2023

Notes: No adjustments made. Leak check passed.

Calibration by: Braiden Boutilier



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Surmont 2	Station Number:	AMS 29
Calibration Date:	March 7, 2024	Prev Cal Date:	June 8, 2023
Start Time (MST):	11:11	End Time (MST):	15:38
Tower Height (m):	10.0	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	W15275
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.2%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.6	-0.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999997	≥0.9995
Calculated slope		1.001844	0.90 - 1.10
Calculated intercept		-0.029522	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	A2005
As Found Declination (deg east of True North):	13° E	As Left Declination (deg east of True North):	13° E
Solar noon time (MST):	12:34	Calc Declination*:	13.12° E
Deadband calc:	NA degrees (<i>Limit 4 deg</i>)		Degrees

* - calculated declination as per NOAA website

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0		---
90		
180		
270		
357		

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope			0.90 - 1.10
Calculated intercept			+/- 4

Notes: WS install. Verified tower alignment using solar noon and a compass.

Calibration Performed By: Braiden Boutilier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS30 ELLS RIVER MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 5, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:27	End Time (MST):	13:30
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999939	
800.4	799.3	1.0013			≥0.995
400.2	396.0	1.0107	Slope	1.000331	
200.1	194.5	1.0288			0.90 - 1.10
			Intercept	-2.815894	+/-30



SO2 Calibration Plot

Date: March 6, 2024

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Version-11-2021

Station Information

Station Name: Ells River Station number: AMS30
 Calibration Date: March 11, 2024 Last Cal Date: February 6, 2024
 Start time (MST): 9:11 End time (MST): 13:40
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.99 ppm Cal Gas Exp Date: November 15, 2026
 Cal Gas Cylinder #: CC505806
 Removed Cal Gas Conc: 4.99 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 3061
 ZAG Make/Model: API T701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1410661331
 Converter make: CDN - 101 Converter serial #: 562
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002473	1.002902	Backgd or Offset: 1.64	1.64
Calibration intercept:	0.159591	-0.080433	Coeff or Slope: 1.060	1.060

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4920	80.2	80.0	79.4	1.007
as found 2nd point	4960	40.1	40.0	39.7	1.006
as found 3rd point	4980	20.0	20.0	19.4	1.024
new cylinder response					

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	-0.1	----
high point	4920	80.2	80.0	80.2	0.998
second point	4960	40.1	40.0	40.0	1.000
third point	4980	20.0	20.0	20.0	0.998
as left zero	5000	0.0	0.0	0.4	----
as left span	4920	80.2	80.0	80.0	1.000
SO2 Scrubber Check	4921	79.2	800.4	0.2	----
Date of last scrubber change:	N/A		Ave Corr Factor		0.999
Date of last converter efficiency test:	N/A		efficiency		

Baseline Corr As found: 79.5 Prev response: 80.39 *% change: -1.1%
 Baseline Corr 2nd AF pt: 39.8 AF Slope: 0.994769 AF Intercept: -0.220738
 Baseline Corr 3rd AF pt: 19.5 AF Correlation: 0.999976

* = > +/-5% change initiates investigation

Notes: Change inlet filters after multipoint as founds. SO2 scrubber check done after calibrator zero, passed. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

TRS Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 11, 2024	Previous Calibration:	February 6, 2024
Station Name:	Ells River	Station Number:	AMS30
Start Time (MST):	9:11	End Time (MST):	13:40
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	0.999998	≥0.995
80.0	80.2	0.9980			
40.0	40.0	1.0005	Slope	1.002902	0.90 - 1.10
20.0	20.0	0.9980			
			Intercept	-0.080433	+/-3



TRS Calibration Plot

Date: March 11, 2024

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	March 6, 2024	Last Cal Date:	February 5, 2024
Start time (MST):	10:27	End time (MST):	13:30
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm		
Diff between cyl (CH ₄):		Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
ZAG make/model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.53E-04	2.53E-04	NMHC SP Ratio:	6.42E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	141799

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	16.97	1.003
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	16.98	1.003
second point	4960	39.6	8.51	8.46	1.006
third point	4980	19.8	4.26	4.14	1.030
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.02	1.000

				Average Correction Factor	1.013
Baseline Corr AF:	16.97	Prev response	17.01	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	9.07	1.005
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.12	0.999
second point	4960	39.6	4.56	4.55	1.002
third point	4980	19.8	2.28	2.22	1.026
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.11	1.000
Average Correction Factor					1.009
Baseline Corr AF:	9.07	Prev response	9.14	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.90	1.002
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.86	1.007
second point	4960	39.6	3.96	3.91	1.013
third point	4980	19.8	1.98	1.91	1.034
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.91	1.000
Average Correction Factor					1.018
Baseline Corr AF:	7.90	Prev response	7.87	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.002141	0.999234
THC Cal Offset:	-0.052740	-0.050339
CH ₄ Cal Slope:	0.997247	0.994937
CH ₄ Cal Offset:	-0.019157	-0.025157
NMHC Cal Slope:	1.006266	1.002917
NMHC Cal Offset:	-0.033582	-0.025982

Notes: Inlet filter changed after As Finds, no adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 5, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:27	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999953	≥ 0.995			
17.03	16.98	1.0028						
8.51	8.46	1.0065				Slope	0.999234	0.90 - 1.10
4.26	4.14	1.0295						
			Intercept	-0.050339	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 5, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:27	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999951	≥0.995
7.91	7.86	1.0070			
3.96	3.91	1.0130			
1.98	1.91	1.0339			
			Slope	0.994937	0.90 - 1.10
			Intercept	-0.025157	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 6, 2024	Previous Calibration:	February 5, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:27	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999952	≥ 0.995			
9.11	9.12	0.9992						
4.56	4.55	1.0017				Slope	1.002917	0.90 - 1.10
2.28	2.22	1.0258						
			Intercept	-0.025982	± 0.5			



NMHC Calibration Plot

Date: March 6, 2024

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	March 23, 2024	Last Cal Date:	March 6, 2024
Start time (MST):	10:46	End time (MST):	14:13
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm		
Diff between cyl (CH ₄):		Diff between cyl (THC):	
		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
ZAG make/model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range (ppm):	0 - 20 ppm		
NMHC Range (ppm):	0 - 10 ppm	CH ₄ Range (ppm):	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.53E-04	2.55E-04	NMHC SP Ratio:	6.42E-05
CH ₄ Retention time:	14.2	14.6	NMHC Peak Area:	141799
				147488

THC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	17.03	17.24	0.987
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	17.03	16.99	1.002
second point	4960	39.6	8.51	8.44	1.009
third point	4980	19.8	4.26	4.17	1.020
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	17.03	17.09	0.997

Average Correction Factor				1.011
Baseline Corr AF:	17.24	Prev response	16.96	*% change 1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Version-01-2020

NMHC Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0	0.00	0.00	----
as found span	4921	79.2	9.11	9.40	0.970
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0	0.00	0.00	----
high point	4921	79.2	9.11	9.09	1.003
second point	4960	39.6	4.56	4.53	1.006
third point	4980	19.8	2.28	2.24	1.016
as left zero	5000	0	0.00	0.00	----
as left span	4921	79.2	9.11	9.18	0.993
Average Correction Factor					1.008
Baseline Corr AF:	9.40	Prev response	9.11	*% change	3.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dil air flow rate	Source gas flow rate	Calc conc (ppm) (Cc)	Ind conc (ppm) (Ic)	CF <i>Limit= 0.95-1.05</i>
as found zero	5000	0.0	0.00	0.00	----
as found span	4921	79.2	7.91	7.84	1.009
as found 2nd point					
as found 3rd point					
new cylinder response					
calibrator zero	5000	0.0	0.00	0.00	----
high point	4921	79.2	7.91	7.91	1.001
second point	4960	39.6	3.96	3.91	1.013
third point	4980	19.8	1.98	1.93	1.025
as left zero	5000	0.0	0.00	0.00	----
as left span	4921	79.2	7.91	7.91	1.001
Average Correction Factor					1.013
Baseline Corr AF:	7.84	Prev response	7.85	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999234	0.999013
THC Cal Offset:	-0.050339	-0.041938
CH ₄ Cal Slope:	0.994937	1.000221
CH ₄ Cal Offset:	-0.025157	-0.027957
NMHC Cal Slope:	1.002917	0.997951
NMHC Cal Offset:	-0.025982	-0.014182

Notes: Dips in instrument response. Chromatograms indicating peak has shifted. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 23, 2024	Previous Calibration:	March 6, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:46	End Time (MST):	14:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999972	≥ 0.995			
17.03	16.99	1.0021						
8.51	8.44	1.0090				Slope	0.999013	0.90 - 1.10
4.26	4.17	1.0204						
			Intercept	-0.041938	± 0.5			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 23, 2024	Previous Calibration:	March 6, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:46	End Time (MST):	14:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999939	≥ 0.995
7.91	7.91	1.0011			
3.96	3.91	1.0130			
1.98	1.93	1.0253			
			Slope	1.000221	0.90 - 1.10
			Intercept	-0.027957	+/-0.5





Wood Buffalo Environmental Association

NMHC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 23, 2024	Previous Calibration:	March 6, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:46	End Time (MST):	14:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>			
0.00	0.00	----	Correlation Coefficient	0.999988	≥ 0.995			
9.11	9.09	1.0030						
4.56	4.53	1.0057				Slope	0.997951	0.90 - 1.10
2.28	2.24	1.0162						
			Intercept	-0.014182	± 0.5			



NMHC Calibration Plot

Date: March 23, 2024

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	March 25, 2024	Last Cal Date:	March 23, 2024
Start time (MST):	10:45	End time (MST):	N/A
Reason:	Removal		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH4 Cal Gas Conc.	499.7 ppm	CH4 Equiv Conc.	1075.0 ppm
C3H8 Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	499.7 ppm	CH4 Equiv Conc.	1075.0 ppm
Removed C3H8 Conc.	209.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
Zero Air Gen model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1181490018
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.55E-04	N/A	NMHC SP Ratio:	6.42E-05	N/A
CH4 Retention time:	14.2	N/A	NMHC Peak Area:	141799	N/A
Zero Chromatogram:	OFF		Flat Baseline:	OFF	

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.2	17.03	17.02	1.000
As found Mid point	4960	39.6	8.51	8.39	1.015
As found Low point	4980	19.8	4.26	4.15	1.025
New cylinder response					
Baseline Corr AF:	17.02	Prev response	16.97	*% change	0.3%
Baseline Corr 2nd AF:	8.39	AF Slope:	1.000771	AF Intercept:	-0.065537
Baseline Corr 3rd AF:	4.15	AF Correlation:	0.999921	<i>* = > +/-5% change initiates investigation</i>	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Notes: Removal calibration, will complete maintenance at the WBEAC.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.2	9.11	9.02	1.010
As found Mid point	4960	39.6	4.56	4.46	1.022
As found Low point	4980	19.8	2.28	2.19	1.039
New cylinder response					
Baseline Corr AF:	9.02	Prev response	9.08	47116	
Baseline Corr 2nd AF:	4.46	AF Slope:	0.991705	AF Intercept:	-0.035779
Baseline Corr 3rd AF:	2.19	AF Correlation:	0.999927	<i>* = > +/-5% change initiates investigation</i>	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.2	7.91	8.00	0.990
As found Mid point	4960	39.6	3.96	3.93	1.007
As found Low point	4980	19.8	1.98	1.96	1.010
New cylinder response					
Baseline Corr AF:	8.00	Prev response	7.89	*% change	1.4%
Baseline Corr 2nd AF:	3.93	AF Slope:	1.011209	AF Intercept:	-0.029758
Baseline Corr 3rd AF:	1.96	AF Correlation:	0.999903	<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999013	
THC Cal Offset:	-0.041938	
CH ₄ Cal Slope:	1.000221	
CH ₄ Cal Offset:	-0.027957	
NMHC Cal Slope:	0.997951	
NMHC Cal Offset:	-0.014182	

Calibration Performed By: Aswin Sasi Kumar

NMHC Calibration Plot

Date: March 25, 2024

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	March 25, 2024	Last Cal Date:	N/A
Start time (MST):	12:28	End time (MST):	16:30
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH4 Cal Gas Conc.	499.7 ppm	CH4 Equiv Conc.	1075.0 ppm
C3H8 Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	499.7 ppm	CH4 Equiv Conc.	1075.0 ppm
Removed C3H8 Conc.	209.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
Zero Air Gen model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1193585649
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	N/A	3.64E-04	NMHC SP Ratio:	N/A
CH4 Retention time:	N/A	17.0	NMHC Peak Area:	N/A
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.2	17.03	17.04	0.999
Mid point	4960	39.6	8.51	8.47	1.005
Low point	4980	19.8	4.26	4.17	1.021
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.2	17.03	17.72	0.961
Average Correction Factor					1.008

Notes: Install calibration. As left CH₄ high.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero					Limit = 0.90-1.10
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	47116	
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.2	9.11	9.11	1.000
Mid point	4960	39.6	4.56	4.53	1.007
Low point	4980	19.8	2.28	2.23	1.020
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.2	9.11	9.19	0.992
Average Correction Factor					1.009

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero					Limit = 0.90-1.10
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.2	7.91	7.93	0.999
Mid point	4960	39.6	3.96	3.95	1.003
Low point	4980	19.8	1.98	1.94	1.022
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.2	7.91	8.53	0.928
Average Correction Factor					1.008

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999013	1.001973
THC Cal Offset:	-0.041938	-0.044740
CH ₄ Cal Slope:	1.000221	1.003138
CH ₄ Cal Offset:	-0.027957	-0.021558
NMHC Cal Slope:	0.997951	1.001236
NMHC Cal Offset:	-0.014182	-0.023782

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	March 25, 2024	Previous Calibration:	N/A
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	12:28	End Time (MST):	16:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999968	<i>≥0.995</i>
17.03	17.04	0.9995	Slope	1.001973	<i>0.90 - 1.10</i>
8.51	8.47	1.0050	Intercept	-0.044740	<i>+/-0.5</i>
4.26	4.17	1.0207			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	March 25, 2024	Previous Calibration:	N/A
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	12:28	End Time (MST):	16:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999963	<i>≥0.995</i>
7.91	7.93	0.9986	Slope	1.003138	<i>0.90 - 1.10</i>
3.96	3.95	1.0028	Intercept	-0.021558	<i>+/-0.5</i>
1.98	1.94	1.0222			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	March 25, 2024	Previous Calibration:	N/A
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	12:28	End Time (MST):	16:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999969	<i>≥0.995</i>
9.11	9.11	1.0000	Slope	1.001236	<i>0.90 - 1.10</i>
4.56	4.53	1.0068	Intercept	-0.023782	<i>+/-0.5</i>
2.28	2.23	1.0198			



NMHC Calibration Plot

Date: March 25, 2024

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	March 28, 2024	Last Cal Date:	N/A
Start time (MST):	13:43	End time (MST):	16:05
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
C ₃ H ₈ Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	499.7 ppm	CH ₄ Equiv Conc.	1075.0 ppm
Removed C ₃ H ₈ Conc.	209.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
Zero Air Gen model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1152430011
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	N/A	3.07E-04	NMHC SP Ratio:	N/A
CH ₄ Retention time:	N/A	17.4	NMHC Peak Area:	N/A
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
-----------	----------------------------------	--------------------------------	--	---------------------------------------	--

As found zero
 As found High point
 As found Mid point
 As found Low point
 New cylinder response

Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.2	17.03	17.05	0.999
Mid point	4960	39.6	8.51	8.50	1.002
Low point	4980	19.8	4.26	4.22	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.2	17.03	17.14	0.994
Average Correction Factor					1.003

Notes: Install calibration. Adjusted zero and span. Use zero chromatogram.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero					Limit = 0.90-1.10
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	47116	
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.2	9.11	9.15	0.996
Mid point	4960	39.6	4.56	4.60	0.991
Low point	4980	19.8	2.28	2.29	0.996
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.2	9.11	9.28	0.982
Average Correction Factor					0.994

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero					Limit = 0.90-1.10
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.2	7.91	7.90	1.001
Mid point	4960	39.6	3.96	3.90	1.014
Low point	4980	19.8	1.98	1.93	1.027
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.2	7.91	7.85	1.008
Average Correction Factor					1.014

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	NA	1.002121
THC Cal Offset:	NA	-0.024340
CH ₄ Cal Slope:	NA	1.000048
CH ₄ Cal Offset:	NA	-0.029357
NMHC Cal Slope:	NA	1.003946
NMHC Cal Offset:	NA	0.005416

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	March 28, 2024	Previous Calibration:	N/A
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	13:43	End Time (MST):	16:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999991	<i>≥0.995</i>
17.03	17.05	0.9987	Slope	1.002121	<i>0.90 - 1.10</i>
8.51	8.50	1.0017	Intercept	-0.024340	<i>+/-0.5</i>
4.26	4.22	1.0100			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	March 28, 2024	Previous Calibration:	N/A
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	13:43	End Time (MST):	16:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999932	<i>≥0.995</i>
7.91	7.90	1.0014	Slope	1.000048	<i>0.90 - 1.10</i>
3.96	3.90	1.0138	Intercept	-0.029357	<i>+/-0.5</i>
1.98	1.93	1.0269			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	March 28, 2024	Previous Calibration:	N/A
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	13:43	End Time (MST):	16:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999991	<i>≥0.995</i>
9.11	9.15	0.9963	Slope	1.003946	<i>0.90 - 1.10</i>
4.56	4.60	0.9910	Intercept	0.005416	<i>+/-0.5</i>
2.28	2.29	0.9958			



NMHC Calibration Plot

Date: March 28, 2024

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Ells River
Calibration Date: March 8, 2024
Start time (MST): 9:39
Reason: Routine
Station number: AMS 30
Last Cal Date: February 23, 2024
End time (MST): 14:08

Calibration Standards

NO Gas Cylinder #: DT0027487
NOX Cal Gas Conc: 59.3 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.3 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.1 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.1 ppm
NO gas Diff:
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Analyzer serial #: 710321429

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.182	1.182	NO bkgnd or offset:	14.0	13.9
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	13.9	13.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	188.7	192.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998634	0.999616
NO _x Cal Offset:	-0.718485	-0.638284
NO Cal Slope:	0.999588	1.002443
NO Cal Offset:	-1.719843	-1.418930
NO ₂ Cal Slope:	1.005754	0.996723
NO ₂ Cal Offset:	0.149457	-0.460536



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
as found span	4932	67.7	803.0	800.3	2.7	807.4	804.4	3.0	0.9945	0.9949
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
high point	4932	67.7	803.0	800.3	2.7	802.2	801.4	1.0	1.0010	0.9986
second point	4966	33.8	400.9	399.5	1.4	400.3	398.7	1.6	1.0015	1.0021
third point	4983	16.9	200.4	199.8	0.7	198.6	197.2	1.4	1.0093	1.0130
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4932	67.7	803.0	423.2	379.8	805.0	425.8	379.7	0.9975	0.9938
Average Correction Factor									1.0039	1.0046

Corrected As found	NO _x = 807.5 ppb	NO = 804.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.8%
Previous Response	NO _x = 801.2 ppb	NO = 798.2 ppb		*Percent Change	NO = 0.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	798.5	421.4	379.8	378.5	1.0035	99.7%
2nd GPT point (200 ppb O3)	798.5	613.5	187.7	186.0	1.0092	99.1%
3rd GPT point (100 ppb O3)	798.5	702.7	98.5	97.4	1.0114	98.9%
Average Correction Factor					1.0080	99.2%

Notes:

Changed sample inlet filters after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 8, 2024	Previous Calibration:	February 23, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:39	End Time (MST):	14:08
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
803.0	802.2	1.0010		
400.9	400.3	1.0015		
200.4	198.6	1.0093		





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 8, 2024	Previous Calibration:	February 23, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:39	End Time (MST):	14:08
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥ 0.995 <i>0.90 - 1.10</i> <i>+/-20</i>
800.3	801.4	0.9986		
399.5	398.7	1.0021		
199.8	197.2	1.0130		
			Correlation Coefficient	0.999985
			Slope	1.002443
			Intercept	-1.418930

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 8, 2024	Previous Calibration:	February 23, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:39	End Time (MST):	14:08
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999987	≥0.995
379.8	378.5	1.0035			
187.7	186.0	1.0092			
98.5	97.4	1.0114			
			Slope	0.996723	0.90 - 1.10
			Intercept	-0.460536	+/-20



NO_x Calibration Plot

Date: March 8, 2024

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Ells River Station number: AMS 30
 Calibration Date: March 8, 2024 Last Cal Date: February 23, 2024
 Start time (MST): 11:06 End time (MST): 11:18

Analyzer Make: API T640 S/N: 875
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-2.90	-3.02	-2.90	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	731.40	733.13	731.40	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.03	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	52.00	----	52.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	10.60	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.90** Expiry Date: September 29, 2024
 Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: January 22, 2024
 Date Disposable Filter Changed: January 22, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 27, 2023
 Date RH/T Sensor Cleaned: February 23, 2024

Notes: Verified flow, temperature and pressure. Leak check passed. No adjustment made.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS501
LEISMER
MARCH 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 15, 2024	Previous Calibration:	February 8, 2024
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	10:02	End Time (MST):	11:19
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999975	
800.2	805.2	0.9938			≥0.995
400.2	402.3	0.9947	Slope	1.007642	
200.1	197.7	1.0120			0.90 - 1.10
			Intercept	-1.456115	+/-30



SO2 Calibration Plot

Date: March 15, 2024

Location: Leismer





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Leismer	Station number:	AMS 501
Calibration Date:	April 8, 2024	Last Cal Date:	March 15, 2024
Start time (MST):	11:25	End time (MST):	12:55
Reason:	Removal		

Calibration Standards

Cal Gas Concentration:	50.52	ppm	Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #:	CC274266		
Removed Cal Gas Conc:	50.52	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	Teledyne API T700		Serial Number: 2659
Zero Air Gen Model:	Teledyne API T701		Serial Number: 4427

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1160290011
Analyzer Range:	0-1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007642	NA	Backgd or Offset:	19.8	19.9
Calibration intercept:	-1.456115	NA	Coeff or Slope:	0.981	0.981

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4921	79.2	800.2	804.0	0.995
As found Mid point	4960	39.6	400.2	400.7	0.998
As found Low point	4980	19.8	200.1	197.2	1.013
New cylinder response					
Baseline Corr As found:	804.3	Previous response	804.9	*% change	-0.1%
Baseline Corr 2nd AF pt:	401.0	AF Slope:	1.006485	AF Intercept:	-1.976071
Baseline Corr 3rd AF pt:	197.5	AF Correlation:	0.999978	* => +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor:


Notes: As founds completed for station removal planned for April 10.

Calibration Performed By: Braiden Boutilier

SO2 Calibration Plot

Date: April 8, 2024

Location: Leismer





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name:	Leismer	Station number:	AMS501
Calibration Date:	March 13, 2024	Last Cal Date:	February 7, 2024
Start time (MST):	8:31	End time (MST):	12:28
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.14	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511843			
Removed Cal Gas Conc:	5.14	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2659
ZAG Make/Model:	API 701		Serial Number:	4427

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020
Converter make:	Global G150	Converter serial #:	2022-218
Analyzer Range	0 - 100 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011575	1.010290	Backgd or Offset:	3.57
Calibration intercept:	-0.158186	-0.138211	Coeff or Slope:	1.128

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4922	77.8	80.0	80.1	0.997
as found 2nd point	4961	38.9	40.0	40.2	0.992
as found 3rd point	4981	19.4	19.9	19.7	1.007
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	77.8	80.0	80.7	0.991
second point	4961	38.9	40.0	40.3	0.992
third point	4981	19.4	19.9	19.8	1.007
as left zero	5000	0.0	0.0	-0.1	----
as left span	4922	77.8	80.0	80.8	0.990
SO2 Scrubber Check	4921	79.2	792.0	0.0	----
Date of last scrubber change:	24-Feb-23			Ave Corr Factor	0.997
Date of last converter efficiency test:	December 1, 2022			efficiency	

Baseline Corr As found:	80.2	Prev response:	80.75	*% change:	-0.7%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.003860	AF Intercept:	-0.138334
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999979		

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. Scrubber test done after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 13, 2024	Previous Calibration:	February 7, 2024
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	8:31	End Time (MST):	12:28
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999982	≥0.995
80.0	80.7	0.9911			
40.0	40.3	0.9923	Slope	1.010290	0.90 - 1.10
19.9	19.8	1.0072			
			Intercept	-0.138211	+/-3



H₂S Calibration Plot

Date: March 13, 2024

Location: Leismer





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name:	Leismer	Station number:	AMS 501
Calibration Date:	April 8, 2024	Last Cal Date:	March 13, 2024
Start time (MST):	12:53	End time (MST):	14:50
Reason:	Removal		

Calibration Standards

Cal Gas Concentration:	5.14 ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511843		
Removed Cal Gas Conc:	5.14 ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	2659
ZAG Make/Model:	Teledyne API T701	Serial Number:	4427

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020
Converter make:	Global G150	Converter serial #:	2022-218
Analyzer Range:	0 - 100 ppb	Converter Temp:	325.0 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010290	NA	Backgd or Offset:	3.57	3.61
Calibration intercept:	-0.138211	NA	Coeff or Slope:	1.128	1.128

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4922	77.8	80.0	80.4	0.992
As found Mid point	4961	38.9	40.0	40.4	0.985
As found Low point	4981	19.4	19.9	19.6	1.007
New cylinder response					
Baseline Corr As found:	80.6	Prev response:	80.67	*% change:	-0.1%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.009433	AF Intercept:	-0.258261
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999953	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:				Ave Corr Factor	
Date of last converter efficiency test:					


Notes: As founds completed for station removal planned for April 10. Scrubber check passed.

Calibration Performed By: Braiden Boutilier

H2S Calibration Plot

Date: April 8, 2024

Location: Leismer





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Leismer
Calibration Date: March 26, 2024
Start time (MST): 8:34
Reason: Routine
Station number: AMS501
Last Cal Date: February 21, 2024
End time (MST): 12:55

Calibration Standards

NO Gas Cylinder #: T26811M
NOX Cal Gas Conc: 47.46 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.46 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API 701
Cal Gas Expiry Date: October 30, 2024
NO Cal Gas Conc: 47.39 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.39 ppm
NO gas Diff:
Serial Number: 2659
Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb
Analyzer serial #: 1218153356

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.186	1.186	NO bkgnd or offset:	3.5	3.5
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	3.5	3.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	166.5	166.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.006425	1.003601
NO _x Cal Offset:	-3.447999	-2.208002
NO Cal Slope:	1.006912	1.004712
NO Cal Offset:	-4.347977	-3.327979
NO ₂ Cal Slope:	0.999948	1.002364
NO ₂ Cal Offset:	0.272718	0.612715



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.2	0.3	----	----
as found span	4916	84.4	801.1	799.9	1.2	812.4	809.5	2.8	0.9860	0.9881
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
high point	4916	84.4	801.1	799.9	1.2	802.5	801.6	0.9	0.9982	0.9979
second point	4958	42.2	400.5	400.0	0.6	399.9	398.0	1.9	1.0016	1.0049
third point	4979	21.1	200.3	200.0	0.3	195.5	193.5	2.0	1.0244	1.0335
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.2	----	----
as left span	4916	84.4	801.1	400.4	400.7	800.7	403.0	397.8	1.0005	0.9935
Average Correction Factor									1.0081	1.0121

Corrected As found	NO _x = 812.3 ppb	NO = 809.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.2%
Previous Response	NO _x = 802.8 ppb	NO = 801.1 ppb		*Percent Change	NO = 1.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	799.5	400.0	400.7	402.1	0.9965	100.4%
2nd GPT point (200 ppb O3)	799.5	614.0	186.7	187.8	0.9940	100.6%
3rd GPT point (100 ppb O3)	799.5	709.4	91.3	92.6	0.9858	101.4%
Average Correction Factor					0.9921	100.8%

Notes:

Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 21, 2024
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	8:34	End Time (MST):	12:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
801.1	802.5	0.9982		
400.5	399.9	1.0016		
200.3	195.5	1.0244		
			0.999951	
			1.003601	
			-2.208002	





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 21, 2024
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	8:34	End Time (MST):	12:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	0.0	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.9	801.6	0.9979		
400.0	398.0	1.0049		
200.0	193.5	1.0335		
			0.999916	
			1.004712	
			-3.327979	

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 26, 2024	Previous Calibration:	February 21, 2024
Station Name:	Leismer	Station Number:	AMS501
Start Time (MST):	8:34	End Time (MST):	12:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	0.2	----	Correlation Coefficient	≥0.995	
400.7	402.1	0.9965			
186.7	187.8	0.9940			
91.3	92.6	0.9858			
			Slope	1.002364	0.90 - 1.10
			Intercept	0.612715	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 26, 2024

Location: Leismer





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Leismer
 Station number: AMS 501
 Calibration Date: April 9, 2024
 Last Cal Date: March 26, 2024
 Start time (MST): 10:47
 End time (MST): 13:52
 Reason: Removal

Calibration Standards

NO Gas Cylinder #: T26811M
 NOX Cal Gas Conc: 47.46 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 47.46 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: October 30, 2024
 NO Cal Gas Conc: 47.39 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 47.39 ppm
 NO gas Diff:
 Serial Number: 2659
 Serial Number: 4427

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
AF High point	4916	84.4	801.1	799.9	1.2	794.3	791.8	2.5	1.0084	1.0099
AF Mid point	4958	42.2	400.5	400.0	0.6	393.1	391.5	1.6	1.0187	1.0211
AF Low point	4979	21.1	200.3	200.0	0.3	192.3	189.5	2.7	1.0409	1.0542
New cyl resp										
Previous Response	NO _x = 801.7 ppb	NO = 800.3 ppb				<i>* = > +/-5% change initiates investigation</i>		*Percent Change		NO _x = -0.9%
Baseline Corr 1st pt	NO _x = 794.4 ppb	NO = 792.0 ppb				<u>As Found Statistics</u>		*Percent Change		NO = -1.1%
Baseline Corr 2nd pt	NO _x = 393.2 ppb	NO = 391.7 ppb				As found	NO _x r ² : 0.999922	Nx SI: 0.993771	Nx Int: -3.388	
Baseline Corr 3rd pt	NO _x = 192.4 ppb	NO = 189.7 ppb				As found	NO r ² : 0.999870	NO SI: 0.993095	NO Int: -4.388	
						As found	NO ₂ r ² : 0.999917	NO ₂ SI: 0.997056	NO ₂ Int: 1.858	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	0.1	----	----
As found high GPT point	790.8	391.5	400.5	400.3	1.0005	100.0%
As found mid GPT point	790.8	610.0	182.0	184.5	0.9863	101.4%
As found low GPT point	790.8	702.7	89.3	92.3	0.9673	103.4%



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1170050148

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003601	NA
NO _x Cal Offset:	-2.208002	NA
NO Cal Slope:	1.004712	NA
NO Cal Offset:	-3.327979	NA
NO ₂ Cal Slope:	1.002364	NA
NO ₂ Cal Offset:	0.612715	NA

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.186	1.186	NO bkgnd or offset:	3.5	3.5
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	3.5	3.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	166.5	166.2

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero
 High point
 Mid point
 Low point
 As left zero
 As left span

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero
 High GPT point
 Mid GPT point
 Low GPT point

Average Correction Factor


Notes: As founds completed for station removal planned for April 10.

Calibration Performed By: Braiden Boutilier

NO_x Calibration Plot

Date: April 9, 2024

Location: Leismer





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Leismer	Station Number:	AMS 501
Calibration Date:	April 9, 2024	Prev Cal Date:	September 11, 2023
Start Time (MST):	12:37	End Time (MST):	13:30
Tower Height (m):	9.8	Reason:	Removal

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	Y18362
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.2%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥0.9995
Calculated slope		0.998868	0.90 - 1.10
Calculated intercept		0.028930	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	R14654
As Found Declination (deg east of True North):	<u>13.3</u>	As Left Declination (deg east of True North):	<u>NA</u>
Solar noon time (MST):	12:37	Calc Declination*:	13.3 Degrees
Deadband calc:	-0.6 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	2.4	---
90	91.6	0.4%
180	180.4	0.1%
270	270.2	0.1%
357	360.0	0.8%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999963	≥0.9995
Calculated slope		1.000158	0.90 - 1.10
Calculated intercept		-1.552811	+/- 4

Notes: As founds done for the station removal planned for April 10.

Calibration Performed By: Braiden Boutilier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS505 SAWBONES BAY MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 18, 2024	Previous Calibration:	February 15, 2024
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	9:40	End Time (MST):	12:37
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321323

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.3	----	Correlation Coefficient	0.999982	≥0.995
799.8	798.6	1.0015			
399.9	399.0	1.0023	Slope	0.999393	0.90 - 1.10
200.4	197.2	1.0164			
			Intercept	-1.051659	+/-30



SO2 Calibration Plot

Date: March 18, 2024

Location: Sawbones Bay





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Sawbones Bay Station number: AMS505
 Calibration Date: March 19, 2024 Last Cal Date: February 22, 2024
 Start time (MST): 8:37 End time (MST): 12:23
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: February 5, 2024
 Cal Gas Cylinder #: CC517397
 Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: Teledybe API T700 Serial Number: 5112
 ZAG Make/Model: Teledybe API T701 Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43iQ Analyzer serial #: 12113311965
 Converter make: Global 150 Converter serial #: 2022-224
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.014500	1.015068	Backgd or Offset:	1.01	1.01
Calibration intercept:	-0.077888	0.002254	Coeff or Slope:	1.119	1.119

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4922	77.7	80.0	81.7	0.978
as found 2nd point	4961	38.8	40.0	40.6	0.982
as found 3rd point	4981	19.4	20.0	20.1	0.989
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4922	77.7	80.0	81.2	0.986
second point	4961	38.8	40.0	40.7	0.982
third point	4981	19.4	20.0	20.2	0.989
as left zero	5000	0.0	0.0	0.1	----
as left span	4922	77.7	80.0	81.2	0.986
SO2 Scrubber Check	4922	77.8	778.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.986
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 81.8 Prev response: 81.12 *% change: 0.8%
 Baseline Corr 2nd AF pt: 40.7 AF Slope: 1.022781 AF Intercept: -0.217686
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999991

* = > +/-5% change initiates investigation

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 19, 2024	Previous Calibration:	February 22, 2024
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:37	End Time (MST):	12:23
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12113311965

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999993	
80.0	81.2	0.9857			≥0.995
40.0	40.7	0.9820	Slope	1.015068	
20.0	20.2	0.9891			0.90 - 1.10
			Intercept	0.002254	+/-3



H₂S Calibration Plot

Date: March 19, 2024

Location: Sawbones Bay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.2	-0.2	----	----
as found span	4917	83.4	799.6	799.6	0.0	806.0	803.1	2.9	0.9920	0.9956
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
high point	4917	83.4	799.6	799.6	0.0	799.7	798.9	0.8	0.9998	1.0008
second point	4958	41.7	399.8	399.8	0.0	399.0	398.7	0.4	1.0021	1.0029
third point	4979	20.9	200.4	200.4	0.0	197.8	197.4	0.3	1.0131	1.0152
as left zero	5000	0.0	0.0	0.0	0.0	0.3	0.7	-0.4	----	----
as left span	4916	83.4	799.7	338.3	461.4	795.3	336.6	458.6	1.0056	1.0052
Average Correction Factor									1.0050	1.0063

Corrected As found	NO _x = 806.4 ppb	NO = 803.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 0.9%	
Previous Response	NO _x = 799.2 ppb	NO = 799.1 ppb		*Percent Change	NO = 0.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	794.6	333.2	461.4	462.5	0.9976	100.2%
2nd GPT point (200 ppb O3)	794.6	546.9	247.7	248.9	0.9952	100.5%
3rd GPT point (100 ppb O3)	794.6	647.1	147.5	149.1	0.9893	101.1%
Average Correction Factor					0.9940	100.6%

Notes:

Changed inlet filter after as founds. Adjusted span.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 14, 2024
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:36	End Time (MST):	12:47
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.4	----	Correlation Coefficient	0.999991	≥0.995
799.6	799.7	0.9998			
399.8	399.0	1.0021	Slope	1.001550	0.90 - 1.10
200.4	197.8	1.0131			
			Intercept	-1.470566	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 14, 2024
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:36	End Time (MST):	12:47
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
799.6	798.9	1.0008		
399.8	398.7	1.0029		
200.4	197.4	1.0152		

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 14, 2024
Station Name:	Sawbones Bay	Station Number:	AMS505
Start Time (MST):	8:36	End Time (MST):	12:47
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
461.4	462.5	0.9976		
247.7	248.9	0.9952		
147.5	149.1	0.9893		
			0.999988	
			1.002397	
			0.386621	


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 20, 2024

Location: Sawbones Bay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS507 KIRBY SOUTH MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 14, 2024	Previous Calibration:	February 15, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	8:30	End Time (MST):	12:29
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.4	----	Correlation Coefficient	0.999986	≥0.995
799.6	797.2	1.0030			
400.3	397.2	1.0079	Slope	0.995804	0.90 - 1.10
199.7	200.2	0.9973			
			Intercept	0.311657	+/-30



SO2 Calibration Plot

Date: March 14, 2024

Location: Kirby South





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Kirby South Station number: AMS507
 Calibration Date: March 13, 2024 Last Cal Date: February 15, 2024
 Start time (MST): 11:15 End time (MST): 17:28
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.05 ppm Cal Gas Exp Date: November 15, 2026
 Cal Gas Cylinder #: DT0019762
 Removed Cal Gas Conc: 5.05 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: DT0019762 Diff between cyl:
 Calibrator Make/Model: API T750 Serial Number: 281
 ZAG Make/Model: API T751H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1150840012
 Converter make: Global Converter serial #: 2022-197
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998813	0.996813	Backgd or Offset: 1.74	1.73
Calibration intercept:	0.219035	0.139039	Coeff or Slope: 1.052	1.041

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	-0.1	----
as found span	4921	79.2	80.0	81.1	0.985
as found 2nd point	4960	39.6	40.0	40.6	0.983
as found 3rd point	4980	19.8	20.0	20.3	0.980
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.0	----
high point	4921	79.2	80.0	79.8	1.002
second point	4960	39.6	40.0	40.1	0.997
third point	4980	19.8	20.0	20.2	0.990
as left zero	5000	0.0	0.0	0.1	----
as left span	4921	79.2	80.0	79.7	1.004
SO2 Scrubber Check	4919	80.0	800.2	0.1	----
Date of last scrubber change:		25-Jul-23		Ave Corr Factor	0.997
Date of last converter efficiency test:					efficiency

Baseline Corr As found: 81.2 Prev response: 80.12 *% change: 1.3%
 Baseline Corr 2nd AF pt: 40.7 AF Slope: 1.014815 AF Intercept: -0.040976
 Baseline Corr 3rd AF pt: 20.4 AF Correlation: 0.999997

* = > +/-5% change initiates investigation

Notes: Changed sample inlet filter after as founds. Ran SO_x scrubber check after cal zero, passed.
 Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 13, 2024	Previous Calibration:	February 15, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	11:15	End Time (MST):	17:28
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.0	----	Correlation Coefficient	0.999986	≥0.995
80.0	79.8	1.0024			
40.0	40.1	0.9975	Slope	0.996813	0.90 - 1.10
20.0	20.2	0.9900			
			Intercept	0.139039	+/-3



H₂S Calibration Plot

Date: March 13, 2024

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 14, 2024	Previous Calibration:	February 22, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	8:30	End Time (MST):	12:29
Analyzer make:	Thermo 51i	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits
0.00	-0.02	----	Correlation Coefficient	≥ 0.995
17.26	17.22	1.0025		
8.64	8.60	1.0050	Slope	0.90 - 1.10
4.31	4.35	0.9913		
			Intercept	± 1.5



THC Calibration Plot

Date: March 14, 2024

Location: Kirby South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	March 13, 2024	Last Cal Date:	February 15, 2024
Start time (MST):	11:12	End time (MST):	13:20
Reason:	Removal		

Calibration Standards

NO Gas Cylinder #:	T34ULGL	Cal Gas Expiry Date:	March 8, 2025
NOX Cal Gas Conc:	49.39 ppm	NO Cal Gas Conc:	49.02 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	49.39 ppm	Removed Gas NO Conc:	49.02 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	3804
ZAG make/model:	API 701H	Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.383	1.383	1.7	1.7
NOX coeff or slope:	0.998	0.998	1.8	1.8
NO2 coeff or slope:	1.000	1.000	210.24	260.25

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.009303	0.999789
NO _x Cal Offset:	-4.671552	-5.004402
NO Cal Slope:	1.014663	1.001150
NO Cal Offset:	-5.612277	-5.705997
NO ₂ Cal Slope:	1.003572	
NO ₂ Cal Offset:	1.309741	



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.5	0.0	----	----
as found span	4919	81.0	800.1	794.1	6.0	828.0	812.0	15.6	0.9663	0.9780
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
high point	4919	81.0	800.1	794.1	6.0	797.7	792.5	5.2	1.0030	1.0020
second point	4960	40.5	400.1	397.1	3.0	391.4	387.8	3.6	1.0221	1.0239
third point	4980	20.2	199.5	198.0	1.5	190.6	187.9	2.7	1.0469	1.0540
as left zero										
as left span										

Average Correction Factor 1.0240 1.0266

Corrected As found	NO _x = 828.5 ppb	NO = 812.5 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 3.1%
Previous Response	NO _x = 802.9 ppb	NO = 800.2 ppb		*Percent Change	NO = 1.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)						
2nd GPT point (200 ppb O3)						
3rd GPT point (100 ppb O3)						

Average Correction Factor

Notes:

Analyzer does meet internal criteria. Maintenance planned for tomorrow.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 13, 2024	Previous Calibration:	February 15, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	11:12	End Time (MST):	13:20
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.3	----	Correlation Coefficient	≥0.995	
800.1	797.7	1.0030			
400.1	391.4	1.0221			
199.5	190.6	1.0469			
			Slope	0.999789	0.90 - 1.10
			Intercept	-5.004402	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 13, 2024	Previous Calibration:	February 15, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	11:12	End Time (MST):	13:20
Analyzer make:	Thermo 42iQ	Analyzer serial #:	1182340006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
794.1	792.5	1.0020		
397.1	387.8	1.0239		
198.0	187.9	1.0540		
			0.999774	
			1.001150	
			-5.705997	


NO Calibration Curve



NO_x Calibration Plot

Date: March 13, 2024

Location: Kirby South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name:	Kirby South	Station number:	AMS507
Calibration Date:	March 15, 2024	Last Cal Date:	March 13, 2024
Start time (MST):	10:30	End time (MST):	15:00
Reason:	Install		

Calibration Standards

NO Gas Cylinder #:	T34ULGL	Cal Gas Expiry Date:	March 8, 2025
NOX Cal Gas Conc:	49.39 ppm	NO Cal Gas Conc:	49.02 ppm
Removed Cylinder #:	NA	Removed Gas Exp Date:	NA
Removed Gas NOX Conc:	49.39 ppm	Removed Gas NO Conc:	49.02 ppm
NOX gas Diff:		NO gas Diff:	
Calibrator Model:	API T700	Serial Number:	3804
ZAG make/model:	API 701H	Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071
NOX Range (ppb):	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	NA	1.097	NO bkgnd or offset:	NA	1.8
NOX coeff or slope:	NA	0.996	NOX bkgnd or offset:	NA	2.5
NO2 coeff or slope:	NA	1.000	Reaction cell Press:	NA	196.86

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	NA	1.000646
NO _x Cal Offset:	NA	-4.504120
NO Cal Slope:	NA	1.002372
NO Cal Offset:	NA	-5.105630
NO ₂ Cal Slope:	NA	0.996340
NO ₂ Cal Offset:	NA	0.258912



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero										
as found span										
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.2	-0.2	----	----
high point	4919	81.0	800.1	794.1	6.0	798.4	793.6	4.8	1.0022	1.0007
second point	4960	40.5	400.1	397.1	3.0	393.1	389.7	3.4	1.0177	1.0189
third point	4980	20.2	199.5	198.0	1.5	191.5	189.0	2.5	1.0420	1.0478
as left zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.3	-0.5	----	----
as left span	4919	81.0	800.1	407.5	392.6	802.0	407.9	393.7	0.9977	0.9991
Average Correction Factor									1.0206	1.0225

Corrected As found	NO _x = NA	ppb	NO = NA	ppb	<i>* => +/-5% change initiates investigation</i>		*Percent Change	NO _x = NA
Previous Response	NO _x = NA	ppb	NO = NA	ppb			*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	791.2	404.6	392.6	391.1	1.0038	99.6%
2nd GPT point (200 ppb O3)	791.2	617.0	180.2	180.4	0.9989	100.1%
3rd GPT point (100 ppb O3)	791.2	709.4	87.8	87.9	0.9988	100.1%
Average Correction Factor					1.0005	100.0%

Notes:

Install calibration. Adjusted zero and span.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association

NO_x Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 15, 2024	Previous Calibration:	March 13, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	10:30	End Time (MST):	15:00
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
800.1	798.4	1.0022		
400.1	393.1	1.0177		
199.5	191.5	1.0420		
			0.999878	
			1.000646	
			-4.504120	

NO_x Calibration Curve





Wood Buffalo Environmental Association

NO Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 15, 2024	Previous Calibration:	March 13, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	10:30	End Time (MST):	15:00
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient Slope Intercept	≥0.995 0.90 - 1.10 +/-20
794.1	793.6	1.0007		
397.1	389.7	1.0189		
198.0	189.0	1.0478		





Wood Buffalo Environmental Association

NO₂ Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 15, 2024	Previous Calibration:	March 13, 2024
Station Name:	Kirby South	Station Number:	AMS507
Start Time (MST):	10:30	End Time (MST):	15:00
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.2	----	Correlation Coefficient	≥0.995	
392.6	391.1	1.0038			
180.2	180.4	0.9989			
87.8	87.9	0.9988			
			Slope	0.996340	0.90 - 1.10
			Intercept	0.258912	+/-20



NO_x Calibration Plot

Date: March 15, 2024

Location: Kirby South





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS511 BLACKGOLD MARCH 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

April 30, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 21, 2024	Previous Calibration:	February 21, 2024
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:12	End Time (MST):	11:16
Analyzer make:	Thermo scientific	Analyzer serial #:	1160290014

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.4	----	Correlation Coefficient	0.999999	≥0.995
799.0	805.5	0.9920			
399.8	403.5	0.9909	Slope	1.008482	0.90 - 1.10
200.0	201.8	0.9909			
			Intercept	-0.070301	+/-30



SO2 Calibration Plot

Date: March 21, 2024

Location: Blackgold





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Blackgold	Station number:	AMS 511
Calibration Date:	April 4, 2024	Last Cal Date:	March 21, 2024
Start time (MST):	12:31	End time (MST):	13:59
Reason:	Removal		

Calibration Standards

Cal Gas Concentration:	50.06	ppm	Cal Gas Exp Date: January 5, 2029
Cal Gas Cylinder #:	CC147416		
Removed Cal Gas Conc:	50.06	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	Teledyne API T700		Serial Number: 2445
Zero Air Gen Model:	Teledyne API 701		Serial Number: 138

Analyzer Information

Analyzer make:	Thermo scientific	Serial Number:	1160290014
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008482	NA	Backgd or Offset:	38.0
Calibration intercept:	-0.070301	NA	Coeff or Slope:	1.202

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-1.5	---
As found High point	4926	79.9	799.0	812.7	0.981
As found Mid point	4968	40.0	399.8	406.1	0.981
As found Low point	4987	20.0	200.0	201.4	0.986
New cylinder response					
Baseline Corr As found:	814.2	Previous response	805.7	*% change	1.0%
Baseline Corr 2nd AF pt:	407.6	AF Slope:	1.019395	AF Intercept:	-1.811745
Baseline Corr 3rd AF pt:	202.9	AF Correlation:	0.999998	<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor:


Notes: Removal Calibration.

Calibration Performed By: Mohammed Kashif

SO2 Calibration Plot

Date: April 4, 2024

Location: Blackgold





Wood Buffalo Environmental Association

H₂S Calibration Report

Version-11-2021

Station Information

Station Name: Blackgold Station number: AMS511
 Calibration Date: March 28, 2024 Last Cal Date: February 8, 2024
 Start time (MST): 8:14 End time (MST): 12:11
 Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.139 ppm Cal Gas Exp Date: January 3, 2026
 Cal Gas Cylinder #: CC511397
 Removed Cal Gas Conc: 5.139 ppm Rem Gas Exp Date: NA
 Removed Gas Cyl #: NA Diff between cyl:
 Calibrator Make/Model: API T700 Serial Number: 2445
 ZAG Make/Model: API T701 Serial Number: 138

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1336160090
 Converter make: Global G150 Converter serial #: 2022-227
 Analyzer Range: 0 - 100 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998919	0.999918	Backgd or Offset:	3.49 3.48
Calibration intercept:	0.140727	0.180786	Coeff or Slope:	1.189 1.189

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
as found zero	5000	0.0	0.0	0.1	----
as found span	4922	77.8	80.0	80.3	0.997
as found 2nd point	4961	38.9	40.0	40.2	0.997
as found 3rd point	4981	19.5	20.0	20.3	0.992
new cylinder response					

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
calibrator zero	5000	0.0	0.0	0.2	----
high point	4922	77.8	80.0	80.1	0.998
second point	4961	38.9	40.0	40.3	0.992
third point	4981	19.5	20.0	20.1	0.997
as left zero	5000	0.0	0.0	0.3	----
as left span	4922	77.9	80.0	80.1	0.999
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.996
Date of last converter efficiency test:				efficiency	

Baseline Corr As found: 80.2 Prev response: 80.02 *% change: 0.2%
 Baseline Corr 2nd AF pt: 40.1 AF Slope: 1.002494 AF Intercept: 0.140643
 Baseline Corr 3rd AF pt: 20.2 AF Correlation: 0.999998

* = > +/-5% change initiates investigation

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Version-11-2021

Station Information

Calibration Date:	March 28, 2024	Previous Calibration:	February 8, 2024
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:14	End Time (MST):	12:11
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	0.2	----	Correlation Coefficient	0.999990	≥0.995
80.0	80.1	0.9983			
40.0	40.3	0.9921	Slope	0.999918	0.90 - 1.10
20.0	20.1	0.9970			
			Intercept	0.180786	+/-3



H₂S Calibration Plot

Date: March 28, 2024

Location: Blackgold





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Blackgold	Station number: AMS 511
Calibration Date: April 4, 2024	Last Cal Date: March 28, 2024
Start time (MST): 10:32	End time (MST): 12:31
Reason: Removal	

Calibration Standards

Cal Gas Concentration: 5.14 ppm	Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC511397	
Removed Cal Gas Conc: 5.14 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 2445
ZAG Make/Model: API T701	Serial Number: 138

Analyzer Information

Analyzer make: Thermo 43i-TLE	Analyzer serial #: 1336160090
Converter make: Global G150	Converter serial #: 2022-227
Analyzer Range: 0 - 100 ppb	Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999918	NA	Backgd or Offset:	3.5	NA
Calibration intercept:	0.180786	NA	Coeff or Slope:	1.189	NA

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4922	77.8	80.0	80.8	0.991
As found Mid point	4961	38.9	40.0	40.6	0.987
As found Low point	4981	19.5	20.0	20.3	0.992
New cylinder response					
Baseline Corr As found:	80.7	Prev response:	80.14	*% change:	0.7%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	1.009496	AF Intercept:	0.120568
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999995	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	-------------------------------	-----------------------------	-------------------------------------	------------------------------------	---

Calibrator zero

High point

Mid point

Low point

As left zero

As left span

SO2 Scrubber Check

Date of last scrubber change:

Date of last converter efficiency test:

Ave Corr Factor


Notes: Removal Calibration. Ran SO2 scrubber check after as found span and it passed.

Calibration Performed By: Mohammed Kashif

H₂S Calibration Plot

Date: April 4, 2024

Location: Blackgold





Wood Buffalo Environmental Association

THC Calibration Summary


Version-01-2020

Station Information

Calibration Date:	March 21, 2024	Previous Calibration:	February 21, 2024
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:12	End Time (MST):	11:16
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1317958295

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (lc)	Correction factor (Cc/lc)	Statistical Evaluation	Limits	
0.00	0.07	----	Correlation Coefficient	0.999843	
17.11	17.18	0.9964			≥0.995
8.57	8.47	1.0118	Slope	1.001619	
4.28	4.20	1.0209			0.90 - 1.10
			Intercept	-0.026162	+/-1.5



THC Calibration Plot

Date: March 21, 2024

Location: Blackgold



THC Calibration Plot

Date: April 4, 2024

Location: Blackgold





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Station Information

Station Name: Blackgold Station number: AMS511
Calibration Date: March 20, 2024 Last Cal Date: February 23, 2024
Start time (MST): 8:47 End time (MST): 13:20
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T0F8P52 Cal Gas Expiry Date: August 16, 2026
NOX Cal Gas Conc: 47.43 ppm NO Cal Gas Conc: 47.43 ppm
Removed Cylinder #: NA Removed Gas Exp Date: NA
Removed Gas NOX Conc: 47.43 ppm Removed Gas NO Conc: 47.43 ppm
NOX gas Diff: NO gas Diff:
Calibrator Model: Teledyne API T700 Serial Number: 2445
ZAG make/model: Teledyne API T701 Serial Number: 138

Analyzer Information

Analyzer make: Teledyne API T200 Analyzer serial #: 7029
NOX Range (ppb): 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.041	1.041	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.038	1.038	NOX bkgnd or offset:	0.4	0.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.2	4.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.016612	1.015442
NO _x Cal Offset:	0.426960	0.386974
NO Cal Slope:	1.015955	1.015028
NO Cal Offset:	-0.193015	0.006993
NO ₂ Cal Slope:	1.002724	1.000776
NO ₂ Cal Offset:	0.288096	0.123095



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Version-04-2020

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
as found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
as found span	4916	84.4	800.6	800.6	0.0	810.9	809.5	1.6	0.9872	0.9889
as found 2nd										
as found 3rd										
new cyl resp										
calibrator zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	----	----
high point	4916	84.4	800.6	800.6	0.0	813.0	812.5	0.6	0.9847	0.9853
second point	4958	42.2	400.3	400.3	0.0	407.4	406.5	1.0	0.9826	0.9848
third point	4979	21.1	200.2	200.2	0.0	203.8	203.2	0.6	0.9821	0.9850
as left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
as left span	4916	84.4	800.6	399.9	400.7	807.6	410.6	396.8	0.9913	0.9738
Average Correction Factor									0.9831	0.9850

Corrected As found	NO _x = 811.0 ppb	NO = 809.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.4%
Previous Response	NO _x = 814.3 ppb	NO = 813.1 ppb		*Percent Change	NO = -0.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:
			As found	NO ₂ r ² :	NO ₂ SI:

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Calibration Limit = 0.95-1.05</i> <i>As Found Limit = 0.90-1.10</i>	Converter Efficiency <i>Calibration Limit = 96-104%</i>
as found GPT zero						
as found GPT point (400 ppb NO2)						
as found GPT point (200 ppb NO2)						
as found GPT point (100 ppb NO2)						
1st GPT point (400 ppb O3)	809.3	408.6	400.7	401.2	0.9988	100.1%
2nd GPT point (200 ppb O3)	809.3	638.7	170.6	170.5	1.0006	99.9%
3rd GPT point (100 ppb O3)	809.3	726.4	82.9	83.5	0.9928	100.7%
Average Correction Factor					0.9974	100.3%

Notes: Changed the sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

NO_x Calibration Summary


Version-04-2020

Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 23, 2024
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:47	End Time (MST):	13:20
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
800.6	813.0	0.9847			
400.3	407.4	0.9826			
200.2	203.8	0.9821			
			Slope	1.015442	0.90 - 1.10
			Intercept	0.386974	+/-20





Wood Buffalo Environmental Association

NO Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 23, 2024
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:47	End Time (MST):	13:20
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.0	-0.1	----	Correlation Coefficient	1.000000	≥0.995
800.6	812.5	0.9853			
400.3	406.5	0.9848	Slope	1.015028	0.90 - 1.10
200.2	203.2	0.9850			
			Intercept	0.006993	+/-20

NO Calibration Curve





Wood Buffalo Environmental Association

NO₂ Calibration Summary

Version-04-2020


Station Information

Calibration Date:	March 20, 2024	Previous Calibration:	February 23, 2024
Station Name:	Blackgold	Station Number:	AMS511
Start Time (MST):	8:47	End Time (MST):	13:20
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	≥0.995	
400.7	401.2	0.9988			
170.6	170.5	1.0006			
82.9	83.5	0.9928			
			Slope	1.000776	0.90 - 1.10
			Intercept	0.123095	+/-20


NO₂ Calibration Curve



NO_x Calibration Plot

Date: March 20, 2024

Location: Blackgold





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Blackgold
 Station number: AMS 511
 Calibration Date: April 5, 2024
 Last Cal Date: March 20, 2024
 Start time (MST): 10:55
 End time (MST): 14:16
 Reason: Removal

Calibration Standards

NO Gas Cylinder #: TOF8P52
 NOX Cal Gas Conc: 47.43 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 47.43 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: August 16, 2026
 NO Cal Gas Conc: 47.43 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 47.43 ppm
 NO gas Diff:
 Serial Number: 2445
 Serial Number: 138

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
AF High point	4916	84.4	800.6	800.6	0.0	812.7	809.4	3.4	0.9851	0.9889
AF Mid point	4958	42.2	400.3	400.3	0.0	407.3	404.7	2.6	0.9828	0.9889
AF Low point	4979	21.1	200.2	200.2	0.0	203.2	201.7	1.5	0.9850	0.9918

New cyl resp

Previous Response	NO _x = 813.3 ppb	NO = 812.6 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -0.1%
Baseline Corr 1st pt	NO _x = 812.7 ppb	NO = 809.5 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -0.4%
Baseline Corr 2nd pt	NO _x = 407.3 ppb	NO = 404.8 ppb	As found	NO _x r ² : 0.999998	Nx SI: 1.015299	Nx Int: 0.187
Baseline Corr 3rd pt	NO _x = 203.2 ppb	NO = 201.8 ppb	As found	NO r ² : 0.999999	NO SI: 1.011430	NO Int: -0.333
			As found	NO ₂ r ² : 0.999985	NO ₂ SI: 1.000073	NO ₂ Int: 0.838

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	0.1	----	----
As found high GPT point	809.6	411.4	398.2	398.7	0.9987	100.1%
As found mid GPT point	809.6	641.9	167.7	169.2	0.9911	100.9%
As found low GPT point	809.6	727.8	81.8	83.1	0.9844	101.6%



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 7029

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.015442	
NO _x Cal Offset:	0.386974	
NO Cal Slope:	1.015028	
NO Cal Offset:	0.006993	
NO ₂ Cal Slope:	1.000776	
NO ₂ Cal Offset:	0.123095	

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.041	NA	NO bkgnd or offset:	0.2	NA
NOX coeff or slope:	1.038	NA	NOX bkgnd or offset:	0.4	NA
NO2 coeff or slope:	1.000	NA	Reaction cell Press:	4.2	NA

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero
 High point
 Mid point
 Low point
 As left zero
 As left span

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero
 High GPT point
 Mid GPT point
 Low GPT point

Average Correction Factor

Notes:


Removal Calibration.

Calibration Performed By: Mohammed Kashif

NO_x Calibration Plot

Date: April 5, 2024

Location: Blackgold





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Black gold	Station Number:	AMS 511
Calibration Date:	April 8, 2024	Prev Cal Date:	September 28, 2023
Start Time (MST):	11:15	End Time (MST):	11:33
Tower Height (m):	10.0	Reason:	Removal

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	B10017
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.7	0.2%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999999	≥0.9995
Calculated slope		0.998137	0.90 - 1.10
Calculated intercept		0.028665	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	D13603
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>NA</u>
Solar noon time (MST):	13:21	Calc Declination*:	13.8 Degrees
Deadband calc:	-2.7 degrees (<i>Limit 4 deg</i>)	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.1	---
90	89.4	-0.2%
180	179.4	-0.2%
270	271.8	0.5%
358	359.8	0.5%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999984	≥0.9995
Calculated slope		0.993497	0.90 - 1.10
Calculated intercept		0.693651	+/- 4

Notes: Removal calibration.

Calibration Performed By: Mohammed Kashif and Devin Russel



End of Report