



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

APRIL 2024

MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

March 31, 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01 BERTHA GANTER - FORT MCKAY APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number: AMS 01
Calibration Date:	April 3, 2024	Last Cal Date: March 5, 2024
Start time (MST):	9:57	End time (MST): 14:14
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #:	CC418809		
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	Teledyne API T700		Serial Number: 3565
Zero Air Gen Model:	Teledyne API T701		Serial Number: 4890

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: JC1501301448
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997052	1.001305	Backgd or Offset:	19.5	20.8
Calibration intercept:	-0.112969	-0.794719	Coeff or Slope:	0.885	0.892

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	1.0	----
As found High point	4918	81.3	800.3	802.8	0.998
As found Mid point	4959	40.7	400.6	402.7	0.997
As found Low point	4979	20.3	199.8	199.8	1.005
New cylinder response					
Baseline Corr As found:	801.8	Previous response	797.8	*% change	0.5%
Baseline Corr 2nd AF pt:	401.7	AF Slope:	1.002567	AF Intercept:	0.505908
Baseline Corr 3rd AF pt:	198.8	AF Correlation:	0.999995	<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4918	81.3	800.3	801.5	0.998
Mid point	4959	40.7	400.6	398.4	1.006
Low point	4979	20.3	199.8	199.2	1.003
As left zero	5000	0.0	0.0	0.2	----
As left span	4918	81.3	800.3	802.3	0.997
Average Correction Factor:					1.002

Notes: Changed the inlet filter and pump after multi-point as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 5, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:57	End Time (MST):	14:14
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999984	≥0.995
800.3	801.5	0.9985	Slope	1.001305	0.90 - 1.10
400.6	398.4	1.0055	Intercept	-0.794719	+/-30
199.8	199.2	1.0031			



SO2 Calibration Plot

Date: April 3, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	April 9, 2024	Last Cal Date:	March 18, 2024
Start time (MST):	10:43	End time (MST):	14:58
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.10	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511749			
Removed Cal Gas Conc:	5.10	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461
Converter make:	CD Nova	Converter serial #:	470
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002936	1.001846	Backgd or Offset:	2.43	2.42
Calibration intercept:	0.199998	0.119998	Coeff or Slope:	0.931	0.918

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4921	78.4	80.0	80.7	0.991
As found Mid point	4960	39.2	40.0	40.5	0.987
As found Low point	4980	19.6	20.0	20.1	0.995
New cylinder response					
Baseline Corr As found:	80.7	Prev response:	80.41	*% change:	0.4%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	1.009563	AF Intercept:	-0.000003
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999993	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4921	78.4	80.0	80.2	0.997
Mid point	4960	39.2	40.0	40.3	0.992
Low point	4980	19.6	20.0	20.1	0.995
As left zero	5000	5000.0	2550.0	0.3	----
As left span	4921	78.4	80.0	80.3	0.996
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	0.995
Date of last converter efficiency test:					

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	April 9, 2024	Previous Calibration:	March 18, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:43	End Time (MST):	14:58
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999995	≥ 0.995
80.0	80.2	0.9972	Slope	1.001846	$0.90 - 1.10$
40.0	40.3	0.9923	Intercept	0.119998	± 3
20.0	20.1	0.9947			



TRS Calibration Plot

Date: April 9, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	April 9, 2024	Last Cal Date:	March 18, 2024
Start time (MST):	10:43	End time (MST):	14:58
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.10	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511749			
Removed Cal Gas Conc:	5.10	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1200326167
Converter make:	CD Nova	Converter serial #:	2022-221
Analyzer Range	0 - 100 ppb	Converter Temp:	350 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005285	1.003053	Backgd or Offset:	2.02	2.03
Calibration intercept:	-0.003216	0.056779	Coeff or Slope:	0.985	0.985

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4922	78.4	80.0	80.3	0.995
As found Mid point	4960	39.2	40.0	40.1	0.995
As found Low point	4980	19.6	20.0	19.9	1.000
New cylinder response					
Baseline Corr As found:	80.4	Prev response:	80.38	*% change:	0.0%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	1.005911	AF Intercept:	-0.143217
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999998	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4922	78.4	80.0	80.2	0.997
Mid point	4960	39.2	40.0	40.3	0.992
Low point	4980	19.6	20.0	20.1	0.995
As left zero	5000	0.0	0.0	0.1	----
As left span	4922	78.4	80.0	80.3	0.996
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	January 25, 2024			Ave Corr Factor	0.995
Date of last converter efficiency test:					

Notes: Inlet filter change and scrubber check completed after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 9, 2024	Previous Calibration:	March 18, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:43	End Time (MST):	14:58
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999993	≥0.995
80.0	80.2	0.9970	Slope	1.003053	0.90 - 1.10
40.0	40.3	0.9923	Intercept	0.056779	+/-3
20.0	20.1	0.9947			



H₂S Calibration Plot

Date: April 9, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	April 3, 2024	Last Cal Date:	March 5, 2024
Start time (MST):	9:57	End time (MST):	14:14
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	497.2 ppm	CH4 Equiv Conc.	1061.8 ppm
C3H8 Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.2 ppm	CH4 Equiv Conc.	1061.8 ppm
Removed C3H8 Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
Zero Air Gen model:	Teledyne API T701	Serial Number:	4890

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320040
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.30E-04	4.31E-04	NMHC SP Ratio:	7.20E-05	7.53E-05
CH4 Retention time:	16.7	16.7	NMHC Peak Area:	127593	122095
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	17.27	16.90	1.022
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.90	Prev response	17.24	*% change	-2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	17.27	17.28	0.999
Mid point	4959	40.7	8.64	8.56	1.010
Low point	4979	20.3	4.31	4.28	1.009
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	17.27	17.30	0.998
Average Correction Factor					1.006

Notes: Changed the inlet filter and H2 cylinder after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	9.18	8.78	1.046
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.78	Prev response	9.13	*% change	-4.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	9.18	9.21	0.997
Mid point	4959	40.7	4.60	4.60	0.998
Low point	4979	20.3	2.29	2.32	0.988
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	9.18	9.23	0.995
Average Correction Factor					0.995

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	8.09	8.12	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.12	Prev response	8.11	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	8.09	8.08	1.001
Mid point	4959	40.7	4.05	3.96	1.022
Low point	4979	20.3	2.02	1.96	1.033
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	8.09	8.07	1.001
Average Correction Factor					1.019

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001618	1.001141
THC Cal Offset:	-0.052990	-0.034037
CH ₄ Cal Slope:	1.008062	1.000742
CH ₄ Cal Offset:	-0.039196	-0.042118
NMHC Cal Slope:	0.995931	1.001953
NMHC Cal Offset:	-0.013994	0.007481

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 5, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:57	End Time (MST):	14:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999967	<i>≥0.995</i>
17.27	17.28	0.9991	Slope	1.001141	<i>0.90 - 1.10</i>
8.64	8.56	1.0095	Intercept	-0.034037	<i>+/-0.5</i>
4.31	4.28	1.0085			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 5, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:57	End Time (MST):	14:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999841 ≥0.995
8.09	8.08	1.0007	Slope	1.000742 0.90 - 1.10
4.05	3.96	1.0223	Intercept	-0.042118 +/-0.5
2.02	1.96	1.0327		





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 5, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:57	End Time (MST):	14:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999992 ≥0.995
9.18	9.21	0.9972	Slope	1.001953 0.90 - 1.10
4.60	4.60	0.9982	Intercept	0.007481 +/-0.5
2.29	2.32	0.9881		



NMHC Calibration Plot

Date: April 3, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay
 Station number: AMS 01
 Calibration Date: April 8, 2024
 Last Cal Date: March 14, 2024
 Start time (MST): 10:06
 End time (MST): 14:41
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC335700
 NOX Cal Gas Conc: 59.40 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 59.40 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: September 1, 2032
 NO Cal Gas Conc: 59.20 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 59.20 ppm
 NO gas Diff:
 Serial Number: 3565
 Serial Number: 4890

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.3	0.2	0.2	----	----
AF High point	4932	67.6	803.1	800.4	2.7	791.4	783.5	7.9	1.0152	1.0218
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 802.6 ppb		NO = 799.9 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -1.5%	
Baseline Corr 1st pt	NO _x = 791.1 ppb		NO = 783.3 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -2.1%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found NO _x r ² :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found NO r ² :		NO SI:	NO Int:	
						As found NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153357

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999371	0.999129
NO _x Cal Offset:	0.020000	0.380000
NO Cal Slope:	1.000862	0.999991
NO Cal Offset:	-1.220000	-1.040000
NO ₂ Cal Slope:	0.995953	0.995614
NO ₂ Cal Offset:	0.117298	1.058457

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.496	1.522	NO bkgnd or offset:	7.4	7.5
NOX coeff or slope:	0.999	0.997	NOX bkgnd or offset:	7.5	7.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	201.0	201.0

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.4	0.3	0.2	----	----
High point	4932	67.6	803.1	800.4	2.7	802.6	799.7	2.8	1.0006	1.0009
Mid point	4966	33.8	401.5	400.2	1.4	402.1	399.4	2.7	0.9986	1.0020
Low point	4983	16.9	200.8	200.1	0.7	200.6	197.1	3.5	1.0009	1.0152
As left zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.3	----	----
As left span	4932	67.6	803.1	390.6	412.5	802.3	390.6	411.7	1.0010	1.0000
Average Correction Factor									1.0000	1.0060

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	798.0	389.5	411.2	410.0	1.0029	99.7%
Mid GPT point	798.0	596.6	204.1	204.8	0.9966	100.3%
Low GPT point	798.0	698.7	102.0	103.4	0.9865	101.4%
Average Correction Factor					0.9953	100.5%

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 8, 2024	Previous Calibration:	March 14, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:06	End Time (MST):	14:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.4	----	Correlation Coefficient	0.999999	<i>≥0.995</i>
803.1	802.6	1.0006	Slope	0.999129	<i>0.90 - 1.10</i>
401.5	402.1	0.9986	Intercept	0.380000	<i>+/-20</i>
200.8	200.6	1.0009			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 8, 2024	Previous Calibration:	March 14, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:06	End Time (MST):	14:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999980	≥0.995
411.2	410.0	1.0029	Slope	0.995614	0.90 - 1.10
204.1	204.8	0.9966	Intercept	1.058457	+/-20
102.0	103.4	0.9865			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 8, 2024	Previous Calibration:	March 14, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:06	End Time (MST):	14:41
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999983	<i>≥0.995</i>
800.4	799.7	1.0009	Slope	0.999991	<i>0.90 - 1.10</i>
400.2	399.4	1.0020	Intercept	-1.040000	<i>+/-20</i>
200.1	197.1	1.0152			



NO_x Calibration Plot

Date: April 8, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number: AMS 01
Calibration Date:	April 2, 2024	Last Cal Date: March 6, 2024
Start time (MST):	9:21	End time (MST): 12:38
Reason:	Routine	

Calibration Standards

O3 generation mode:	Photometer	
Calibrator Make/Model:	Teledyne API T700	Serial Number: 3565
ZAG Make/Model:	Teledyne API T701	Serial Number: 4890

Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #: 1107
Analyzer Range	0 - 500 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001057	0.998829	Backgd or Offset:	5.5	4.6
Calibration intercept:	0.640000	0.180000	Coeff or Slope:	1.044	1.027

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-1.3	----
As found High point	5000	863.1	400.0	407.4	0.979
As found Mid point					
As found Low point					
Baseline Corr As found:	408.7	Previous response	401.1	*% change	1.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	5000	863.1	400.0	399.4	1.002
Mid point	5000	742.5	200.0	200.5	0.998
Low point	5000	651.7	100.0	100.2	0.998
As left zero	5000	0.0	0.0	-0.2	----
As left span	5000	863.1	400.0	400.1	1.000
Average Correction Factor					0.999

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 6, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:21	End Time (MST):	12:38
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999993	≥0.995
400.0	399.4	1.0015	Slope	0.998829	0.90 - 1.10
200.0	200.5	0.9975	Intercept	0.180000	+/- 5
100.0	100.2	0.9980			



O₃ Calibration Plot

Date: April 2, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
 Calibration Date: April 11, 2024 Last Cal Date: March 18, 2024
 Start time (MST): 12:29 End time (MST): 13:50

Analyzer Make: Teledyne API T640 S/N: 324
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752
 Temp/RH standard: Alicat FP-25BT S/N: 388752

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	13.9	12.94	13.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	736.2	735.46	736.2	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	5.122	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	38	----	38	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	7.3	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	8.8	10.9	10.9	<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: _____ April 11, 2024
 Date Disposable Filter Changed: _____ April 11, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ September 14, 2023
 Date RH/T Sensor Cleaned: _____ April 11, 2024

Notes: Verified flow, temperature, and pressure. Leak check passed. PMT peak test completed. Disposable filter changed. Optical chamber and RH/T sensor cleaned.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	April 10, 2024	Last Cal Date:	March 15, 2024
Start time (MST):	10:25	End time (MST):	13:40
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,040	ppm	Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #:	ALM042207		
Removed Cal Gas Conc:	3,040	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 3565
ZAG Make/Model:	Teledyne API T701		Serial Number: 4890

Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000382	1.000946	Backgd or Offset:	-0.013	-0.013
Calibration intercept:	0.195846	0.155852	Coeff or Slope:	0.989	0.991

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4933	66.7	40.6	40.8	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	40.66	Prev response:	40.77	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4933	66.7	40.6	40.6	0.999
Mid point	4966	33.3	20.2	20.7	0.977
Low point	4983	16.7	10.2	10.3	0.987
As left zero	5000	0.0	0.0	0.0	----
As left span	2960	40.0	40.5	40.1	1.010
Average Correction Factor					0.988

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


CO Calibration Summary

Station Information

Calibration Date:	April 10, 2024	Previous Calibration:	March 15, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:25	End Time (MST):	13:40
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999863	≥0.995
40.6	40.6	0.9987	Slope	1.000946	0.90 - 1.10
20.2	20.7	0.9773	Intercept	0.155852	+/-1.5
10.2	10.3	0.9868			



CO Calibration Plot

Date: April 10, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

CO₂ Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	April 4, 2024	Last Cal Date:	March 12, 2024
Start time (MST):	8:44	End time (MST):	11:55
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #:	ALM042207		
Removed Cal Gas Conc:	60,200	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 3565
N2 Gen Make/Model:	Peak Scientific		Serial Number: 7220900034

Analyzer Information

Analyzer make:	Teledyne API 360	Analyzer serial #: 442
Analyzer Range	0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000907	1.001476	Backgd or Offset:	0.045	0.045
Calibration intercept:	-6.720000	-5.720000	Coeff or Slope:	0.876	0.876

CO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	3000	0.0	0.0	-0.1	----
As found High Point	2920	80.0	1605.3	1594.9	1.006
As found Mid Point					
As found Low Point					
New cylinder response					
Baseline Corr As found:	1595.0	Prev response:	1600.1	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	3000	0.0	0.0	0.2	----
High point	2920	80.0	1605.3	1605.1	1.000
Mid point	2960	40.0	802.7	794.7	1.010
Low point	2980	20.0	401.3	390.6	1.027
As left zero	3000	0.0	0.0	0.0	----
As left span	2960	40.0	802.7	778.1	1.032
Average Correction Factor					1.013

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


CO₂ Calibration Summary

Station Information

Calibration Date	April 4, 2024	Previous Calibration	March 12, 2024
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS 01
Start Time (MST)	8:44	End Time (MST)	11:55
Analyzer make	Teledyne API 360	Analyzer serial #	442

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999938	≥0.995
1605.3	1605.1	1.0001	Slope	1.001476	0.90 - 1.10
802.7	794.7	1.0100	Intercept	-5.7	+/-20
401.3	390.6	1.0275			



CO₂ Calibration Plot

Date: April 4, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

Nt - NOX - NH3 Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
NOX Cal Date:	April 11, 2024	Last Cal Date:	March 19, 2024
Start time (MST):	10:12	End time (MST):	15:15
NH3 Cal Date:	April 11, 2024	Last Cal Date:	March 20, 2024
Start time (MST):	15:35	End time (MST):	18:00
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	59.40	ppm	NO Gas Cylinder #:	CC335700
NO Cal Gas Conc:	59.20	ppm	NO Cal Gas Expiry:	September 1, 2032
Removed NOX Conc:	59.40	ppm	Removed Cylinder #:	NA
Removed NO Conc:	59.20	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.58	ppm	NH3 Gas Cylinder #:	CC743587
Removed NH3 Conc:	76.58	ppm	NH3 Cal Gas Expiry:	August 22, 2024
NH3 gas Diff:			Removed Cylinder #:	NA
Calibrator Model:	API T700		Removed cyl Expiry:	NA
ZAG make/model:	API T701		Serial Number:	3565
			Serial Number:	4890

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	475
Converter model:	API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.20
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	540

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.960	0.952	Nt coefficient:	0.969	0.962
NOX coefficient:	0.966	0.959	NO bkgrnd:	-0.9	-0.9
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	-0.3	-0.3
NH3 coefficient:	0.946	0.946	Nt bkgrnd:	1.2	1.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001292	1.000623
NO _x Cal Offset:	-2.380000	-2.120000
NO Cal Slope:	0.998821	0.997950
NO Cal Offset:	-2.280000	-2.600000
NO ₂ Cal Slope:	0.999426	1.001154
NO ₂ Cal Offset:	-0.898937	0.886739
NH3 Cal Slope:	1.002078	0.997027
NH3 Cal Offset:	-0.363408	2.325094
Nt Cal Slope:	1.004775	1.000070
Nt Cal Offset:	-0.436983	2.166761



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

NO_x / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO _x Correction factor (Cc/Ic)	Baseline corr NO Correction factor (Cc/Ic)
									<i>Limit = 0.9 - 1.0</i>	<i>Limit = 0.9 - 1.0</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.2	-0.2	----	----
As found span	4932	67.6	803.1	800.4	803.1	813.1	802.3	813.1	0.9877	0.9976
AF GPT span										
new NO cyl rp										

Baseline Corr As Fd	Nt = 813.3 ppb	NO _x = 813.5 ppb	NO = 802.5 ppb		*Percent Change	Nt _(NO) = 0.8%
Previous Response	Nt = 806.49 ppb	NO _x = 801.7 ppb	NO = 797.2 ppb		*Percent Change	NO _x = 1.4%
**NO _x Δ (NO to GPT response) =					*Percent Change	NO = 0.7%
** = > +/-2% difference initiates investigation					* = > +/-5% change initiates investigation	

NO_x / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic)	NO Correction factor (Cc/Ic)
									<i>Limit = 0.95-1.05</i>	<i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.3	-0.1	----	----
High point	4932	67.6	803.1	800.4	803.1	803.0	798.1	800.3	1.0001	1.0029
Mid point	4966	33.8	401.5	400.2	401.5	396.7	393.2	397.8	1.0122	1.0178
Low point	4983	16.9	200.8	200.1	200.8	198.6	196.4	199.0	1.0109	1.0188
Average Correction Factor									1.0078	1.0132

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic)	Converter Efficiency
					<i>Limit = 0.95-1.05</i>	<i>Limit = 96-104%</i>
Calibration zero	----	----	0.0	-0.1	----	----
High GPT point (400 ppb O3)	794.8	384.6	412.9	414.0	0.9974	100.3%
Mid GPT point (200 ppb O3)	794.8	588.2	209.3	210.2	0.9957	100.4%
Low GPT point (100 ppb O3)	794.8	690.6	106.9	109.4	0.9772	102.3%
Average Correction Factor					0.9901	101.0%



Wood Buffalo Environmental Association NH₃ - N_T Calibration Report

NH₃ As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH3 Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.4	0.2	----	----
AF High point	3418	82.2	1798.5	----	1798.5	1798.2		1793.1	1.000	1.003
AF Mid point							----			
AF Low point							----			
new NH3 cyl rp										
Baseline Corr As Fd		Nt = 1798.4 ppb	NH3 = 1792.9 ppb						*Percent Change	Nt _(NH3) = -0.5%
Previous Response		Nt = 1806.7 ppb	NH3 = 1801.9 ppb			* = > +/-5% change initiates investigation			*Percent Change	NH3 = -0.5%

NH₃ Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.5	0.4	----	----
High point	3418	82.2	1798.5	----	1798.5	1798.2	----	1793.1	1.000	1.003
Mid point	3454	45.7	1000.0	----	1000.0	1006.8	----	1003.7	0.993	0.996
Low point	3477	22.8	498.9	----	498.9	501.4	----	499.7	0.995	0.998
Average Correction Factor									0.9961	0.9992
NH3 Previous Converter Efficiency =		90.8 %								
NH3 Current Converter Efficiency =		90.8 %								

Notes:

Changed the inlet filter after as founds. Adjusted the NOx/Nt span.

Calibration Performed By:

Rene Chamberland



Wood Buffalo Environmental Association


Nt Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 19, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:12	End Time (MST):	15:15
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999981	<i>≥0.995</i>
1798.5	1798.2	1.0002	Slope	1.000070	<i>0.90 - 1.10</i>
1000.0	1006.8	0.9932	Intercept	2.166761	<i>+/-20</i>
498.9	501.4	0.9949			





Wood Buffalo Environmental Association


NH₃ Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 19, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:12	End Time (MST):	15:15
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.4	----	Correlation Coefficient	0.999984	<i>≥0.995</i>
1798.5	1793.1	1.0030	Slope	0.997027	<i>0.90 - 1.10</i>
1000.0	1003.7	0.9963	Intercept	2.325094	<i>+/-20</i>
498.9	499.7	0.9983			





Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 19, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:12	End Time (MST):	15:15
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.5	----	Correlation Coefficient	0.999961	<i>≥0.995</i>
803.1	803.0	1.0001	Slope	1.000623	<i>0.90 - 1.10</i>
401.5	396.7	1.0122	Intercept	-2.120000	<i>+/-20</i>
200.8	198.6	1.0109			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 19, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:12	End Time (MST):	15:15
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.3	----	Correlation Coefficient	0.999936	<i>≥0.995</i>
800.4	798.1	1.0029	Slope	0.997950	<i>0.90 - 1.10</i>
400.2	393.2	1.0178	Intercept	-2.600000	<i>+/-20</i>
200.1	196.4	1.0188			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 19, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:12	End Time (MST):	15:15
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999964	<i>≥0.995</i>
412.9	414.0	0.9974	Slope	1.001154	<i>0.90 - 1.10</i>
209.3	210.2	0.9957	Intercept	0.886739	<i>+/-20</i>
106.9	109.4	0.9772			



NO_x Calibration Plot

Date: April 11, 2024

Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: April 11, 2024

Location: Bertha Ganter-Fort McKay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Mildred Lake	Station number:	AMS 02
Calibration Date:	April 5, 2024	Last Cal Date:	March 19, 2024
Start time (MST):	10:06	End time (MST):	15:35
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	1185
Zero Air Gen Model:	Teledyne API T701		Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	JC1404901075
Analyzer Range:	0-1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000614	0.993915	Backgd or Offset:	18.5	18.7
Calibration intercept:	-0.364904	-0.766439	Coeff or Slope:	0.787	0.787

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4920	80.2	801.6	797.6	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.5	Previous response	801.8	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4920	80.2	801.6	796.2	1.007
Mid point	4960	40.1	400.8	397.6	1.008
Low point	4980	20.0	199.9	197.2	1.014
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	80.2	801.6	794.9	1.008
Average Correction Factor:					1.010

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 5, 2024	Previous Calibration:	March 19, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:06	End Time (MST):	15:35
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.2	----	Correlation Coefficient	0.999997	≥0.995
801.6	796.2	1.0068	Slope	0.993915	0.90 - 1.10
400.8	397.6	1.0081	Intercept	-0.766439	+/-30
199.9	197.2	1.0138			



SO2 Calibration Plot

Date: April 5, 2024

Location: Mildred Lake





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Mildred Lake	Station number: AMS 02
Calibration Date: April 23, 2024	Last Cal Date: March 26, 2024
Start time (MST): 10:11	End time (MST): 16:00
Reason: Routine	

Calibration Standards

Cal Gas Concentration: 5.29 ppm	Cal Gas Exp Date: January 4, 2025
Cal Gas Cylinder #: CC345191	
Removed Cal Gas Conc: 5.29 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: Teledyne API T700	Serial Number: 1185
ZAG Make/Model: Teledyne API T701	Serial Number: 4891

Analyzer Information

Analyzer make: Thermo 43iQTL	Analyzer serial #: 12113311966
Converter make: Global G150	Converter serial #: 2022-198
Analyzer Range: 0 - 100 ppb	Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000964	1.008965	Backgd or Offset:	1.67	1.95
Calibration intercept:	0.020800	-0.259192	Coeff or Slope:	0.731	0.731

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4924	75.6	80.0	81.4	0.983
As found Mid point	4962	37.8	40.0	40.7	0.983
As found Low point	4981	18.9	20.0	19.9	1.005
New cylinder response					
Baseline Corr As found:	81.4	Prev response:	80.09	*% change:	1.6%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.019538	AF Intercept:	-0.179186
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999962	<i>* = > +/-5% change initiates investigation</i>	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4924	75.6	80.0	80.4	0.995
Mid point	4962	37.8	40.0	40.3	0.992
Low point	4981	18.9	20.0	19.7	1.015
As left zero	5000	0.0	0.0	-0.1	----
As left span	4924	75.6	80.0	80.5	0.994
SO2 Scrubber Check	4920	80.2	802.0	0.1	----
Date of last scrubber change:	September 20, 2023			Ave Corr Factor	1.001
Date of last converter efficiency test:	April 23, 2024			107.1% efficiency	

Notes: Changed inlet filter after running the converter efficiency test. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	April 23, 2024	Previous Calibration:	March 26, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:11	End Time (MST):	16:00
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999973	≥ 0.995
80.0	80.4	0.9949	Slope	1.008965	$0.90 - 1.10$
40.0	40.3	0.9924	Intercept	-0.259192	± 3
20.0	19.7	1.0151			



H2S Calibration Plot

Date: April 23, 2024

Location: Mildred Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Buffalo Viewpoint	Station number: AMS 04
Calibration Date:	April 15, 2024	Last Cal Date: March 21, 2024
Start time (MST):	8:48	End time (MST): 11:35
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	50.87	ppm	Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #:	CC446753		
Removed Cal Gas Conc:	50.87	ppm	Rem Gas Exp Date:
Removed Gas Cyl #:			Diff between cyl:
Calibrator Model:	API T700		Serial Number: 3808
Zero Air Gen Model:	API T701		Serial Number: 362

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: JC1327300932
Analyzer Range:	0-1000ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000873	1.006888	Backgd or Offset:	24.7	25.1
Calibration intercept:	-0.325433	0.395801	Coeff or Slope:	0.867	0.880

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4921	78.6	799.7	792.6	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	792.3	Previous response	800.1	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4921	78.6	799.7	805.1	0.993
Mid point	4961	39.3	399.8	404.5	0.988
Low point	4980	19.6	199.4	200.3	0.996
As left zero	5000	0.0	0.0	0.2	----
As left span	4921	78.6	799.7	804.7	0.994
Average Correction Factor:					0.992

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 15, 2024	Previous Calibration:	March 21, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	8:48	End Time (MST):	11:35
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999990	≥0.995
799.7	805.1	0.9933	Slope	1.006888	0.90 - 1.10
399.8	404.5	0.9884	Intercept	0.395801	+/-30
199.4	200.3	0.9956			



SO2 Calibration Plot

Date: April 15, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Buffalo Viewpoint	Station number: AMS 04
Calibration Date: April 5, 2024	Last Cal Date: March 20, 2024
Start time (MST): 6:40	End time (MST): 10:33
Reason: Routine	

Calibration Standards

Cal Gas Concentration: 5.42 ppm	Cal Gas Exp Date: January 4, 2025
Cal Gas Cylinder #: CC345266	
Removed Cal Gas Conc: 5.42 ppm	Rem Gas Exp Date:
Removed Gas Cyl #:	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 3808
ZAG Make/Model: API T701H	Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i-LTE	Analyzer serial #: 1008841400
Converter make: Global	Converter serial #: 2022-200
Analyzer Range: 0 - 100 ppb	Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995649	0.997925	Backgd or Offset:	1.09	1.09
Calibration intercept:	0.182120	0.102175	Coeff or Slope:	1.130	1.130

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4926	74.1	80.3	81.4	0.988
As found Mid point	4963	37.0	40.1	40.7	0.988
As found Low point	4982	18.5	20.1	20.1	1.003
New cylinder response					
Baseline Corr As found:	81.3	Prev response:	80.16	*% change:	1.4%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.013436	AF Intercept:	-0.017562
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999983	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4926	74.1	80.3	80.2	1.002
Mid point	4963	37.0	40.1	40.3	0.995
Low point	4982	18.5	20.1	20.0	1.003
As left zero	5000	0.0	0.0	0.2	----
As left span	4926	74.1	80.3	79.9	1.005
SO ₂ Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	1.000
Date of last converter efficiency test:					

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 5, 2024	Previous Calibration:	March 20, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	6:40	End Time (MST):	10:33
Analyzer make:	Global	Analyzer serial #:	2022-200

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999987	≥0.995
80.3	80.2	1.0015	Slope	0.997925	0.90 - 1.10
40.1	40.3	0.9952	Intercept	0.102175	+/-3
20.1	20.0	1.0026			



H₂S Calibration Plot

Date: April 5, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS 04
Calibration Date:	April 15, 2024	Last Cal Date:	March 21, 2024
Start time (MST):	8:47	End time (MST):	11:34
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	497.2 ppm	CH4 Equiv Conc.	1058.2 ppm
C3H8 Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	497.2 ppm	CH4 Equiv Conc.	1058.2 ppm
Removed C3H8 Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
Zero Air Gen model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1426262594
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.25E-04	4.22E-04	NMHC SP Ratio:	1.10E-04	1.07E-04
CH4 Retention time:	13.7	13.7	NMHC Peak Area:	80500	82687
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	---
As found High point	4921	78.6	16.64	16.95	0.982
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.95	Prev response	16.65	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	---
High point	4921	78.6	16.64	16.60	1.002
Mid point	4961	39.3	8.32	8.30	1.002
Low point	4980	19.6	4.15	4.14	1.001
As left zero	5000	0.0	0.00	0.00	---
As left span	4921	78.6	16.64	16.66	0.999
Average Correction Factor					1.002

Notes: No maintenance done. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	8.82	9.06	0.973
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.06	Prev response	8.83	*% change	2.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	8.82	8.80	1.003
Mid point	4961	39.3	4.41	4.41	1.000
Low point	4980	19.6	2.20	2.20	0.998
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	8.82	8.84	0.998
Average Correction Factor					1.000

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	7.82	7.89	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.89	Prev response	7.82	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	7.82	7.81	1.001
Mid point	4961	39.3	3.91	3.89	1.004
Low point	4980	19.6	1.95	1.94	1.004
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	7.82	7.82	0.999
Average Correction Factor					1.003

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000586	0.997908
THC Cal Offset:	0.008071	0.002057
CH ₄ Cal Slope:	1.001043	0.998894
CH ₄ Cal Offset:	-0.004702	-0.004105
NMHC Cal Slope:	0.999701	0.997215
NMHC Cal Offset:	0.013373	0.005963

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 15, 2024	Previous Calibration:	March 21, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	8:47	End Time (MST):	11:34
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	1.000000 ≥0.995
16.64	16.60	1.0021	Slope	0.997908 0.90 - 1.10
8.32	8.30	1.0016	Intercept	0.002057 +/-0.5
4.15	4.14	1.0013		





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 15, 2024	Previous Calibration:	March 21, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	8:47	End Time (MST):	11:34
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
7.82	7.81	1.0012	Slope	0.998894	<i>0.90 - 1.10</i>
3.91	3.89	1.0035	Intercept	-0.004105	<i>+/-0.5</i>
1.95	1.94	1.0042			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 15, 2024	Previous Calibration:	March 21, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	8:47	End Time (MST):	11:34
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
8.82	8.80	1.0026	Slope	0.997215	<i>0.90 - 1.10</i>
4.41	4.41	1.0000	Intercept	0.005963	<i>+/-0.5</i>
2.20	2.20	0.9983			



NMHC Calibration Plot

Date: April 15, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS 04
Calibration Date:	April 24, 2024	Last Cal Date:	April 15, 2024
Start time (MST):	7:36	End time (MST):	8:59
Reason:	Cylinder Change Nitrogen Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	497.2 ppm	CH4 Equiv Conc.	1058.2 ppm
C3H8 Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	497.2 ppm	CH4 Equiv Conc.	1058.2 ppm
Removed C3H8 Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
Zero Air Gen model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1426262594
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.25E-04	4.22E-04	NMHC SP Ratio:	1.10E-04	1.07E-04
CH4 Retention time:	13.7	13.7	NMHC Peak Area:	80500	82687
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	16.64	16.61	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.61	Prev response	16.65	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	16.64	16.55	1.005
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.005

Notes: Nitrogen Cylinder Change.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	8.82	8.78	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.78	Prev response	8.83	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	8.82	8.76	1.007
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.007

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	7.82	7.82	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.82	Prev response	7.82	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	7.82	7.80	1.002
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.002

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000586	0.994997
THC Cal Offset:	0.008071	0.000000
CH ₄ Cal Slope:	1.001043	0.998003
CH ₄ Cal Offset:	-0.004702	0.000000
NMHC Cal Slope:	0.999701	0.992673
NMHC Cal Offset:	0.013373	0.000000

Calibration Performed By: Melissa Lemay

NMHC Calibration Plot

Date: April 24, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Buffalo Viewpoint
 Station number: AMS 04
 Calibration Date: April 2, 2024
 Last Cal Date: March 1, 2024
 Start time (MST): 6:08
 End time (MST): 11:18
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC324979
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #:
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: APIT701
 Cal Gas Expiry Date: November 3, 2032
 NO Cal Gas Conc: 48.80 ppm
 Removed Gas Exp Date:
 Removed Gas NO Conc: 48.80 ppm
 NO gas Diff:
 Serial Number: 3808
 Serial Number: 362

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.4	0.3	0.1	----	----
AF High point	4918	81.8	800.0	798.4	1.6	784.3	777.1	7.2	1.0206	1.0278
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 796.3 ppb		NO = 791.6 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -1.6%	
Baseline Corr 1st pt	NO _x = 783.9 ppb		NO = 776.8 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -1.9%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: API T200
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 721

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995753	1.003822
NO _x Cal Offset:	-0.374543	-0.272631
NO Cal Slope:	0.993186	1.002903
NO Cal Offset:	-1.315833	-0.733480
NO ₂ Cal Slope:	0.998315	0.996313
NO ₂ Cal Offset:	1.614021	-1.280139

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.162	1.189	NO bkgnd or offset:	-0.6	-0.6
NOX coeff or slope:	1.154	1.177	NOX bkgnd or offset:	-0.3	-0.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.3	4.3

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.5	0.7	-0.2	----	----
High point	4918	81.8	800.0	798.4	1.6	803.4	801.0	2.4	0.9958	0.9968
Mid point	4959	40.9	400.0	399.2	0.8	400.3	398.0	2.2	0.9993	1.0030
Low point	4980	20.4	199.5	199.1	0.4	199.6	198.1	1.5	0.9995	1.0050
As left zero	5000	0.0	0.0	0.8	-0.8	0.7	0.8	-0.1	----	----
As left span	4918	81.8	800.0	401.8	800.0	792.6	401.8	390.8	1.0094	1.0000
Average Correction Factor									0.9982	1.0016

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	796.0	398.7	398.9	396.7	1.0056	99.4%
Mid GPT point	796.0	598.1	199.5	197.1	1.0124	98.8%
Low GPT point	796.0	694.2	103.4	100.6	1.0282	97.3%
Average Correction Factor					1.0154	98.5%

Notes: Span adjusted. Due to drifting during the GPT the 2nd NO ref point used.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 1, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	6:08	End Time (MST):	11:18
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.5	----	Correlation Coefficient	0.999994	<i>≥0.995</i>
800.0	803.4	0.9958	Slope	1.003822	<i>0.90 - 1.10</i>
400.0	400.3	0.9993	Intercept	-0.272631	<i>+/-20</i>
199.5	199.6	0.9995			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 1, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	6:08	End Time (MST):	11:18
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999965	≥0.995
398.9	396.7	1.0056	Slope	0.996313	0.90 - 1.10
199.5	197.1	1.0124	Intercept	-1.280139	+/-20
103.4	100.6	1.0282			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 1, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	6:08	End Time (MST):	11:18
Analyzer make:	API T200	Analyzer serial #:	721

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.7	----	Correlation Coefficient	0.999982	≥0.995
798.4	801.0	0.9968	Slope	1.002903	0.90 - 1.10
399.2	398.0	1.0030	Intercept	-0.733480	+/-20
199.1	198.1	1.0050			



NO_x Calibration Plot

Date: April 2, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS 04
Calibration Date:	April 15, 2024	Last Cal Date:	March 15, 2024
Start time (MST):	6:18	End time (MST):	8:48
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3808
Calibrator Make/Model:	APIP T700	Serial Number:	362
ZAG Make/Model:	API T701		

Analyzer Information

Analyzer make:	API T400	Analyzer serial #:	2961
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004971	1.003629	Backgd or Offset:	-2.2	-2.2
Calibration intercept:	-0.120000	0.540000	Coeff or Slope:	1.011	1.011

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	5000	995.7	400.0	401.1	0.996
As found Mid point					
As found Low point					
Baseline Corr As found:	401.5	Previous response	401.9	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	5000	995.3	400.0	401.8	0.996
Mid point	5000	820.9	200.0	201.5	0.993
Low point	5000	711.9	100.0	101.2	0.988
As left zero	5000	0.0	0.0	0.0	----
As left span	5000	995.7	400.0	403.5	0.991
Average Correction Factor					0.992

Notes: No adjustments and maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 15, 2024	Previous Calibration:	March 15, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	6:18	End Time (MST):	8:48
Analyzer make:	API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999997	≥0.995
400.0	401.8	0.9955	Slope	1.003629	0.90 - 1.10
200.0	201.5	0.9926	Intercept	0.540000	+/- 5
100.0	101.2	0.9881			



O₃ Calibration Plot

Date: April 15, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: April 24, 2024 Last Cal Date: March 20, 2024
 Start time (MST): 6:57 End time (MST): 7:22

Analyzer Make: API T640 S/N: 321
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	4.6	4.3	4.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	727.5	729.4	727.5	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.18	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	41	----	41	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	15.5	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: 6-10-2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 27, 2024
 Date Disposable Filter Changed: February 27, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: February 27, 2024
 Date RH/T Sensor Cleaned: February 27, 2024

Notes: No adjustments done. Leak check passed.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS05
MANNIX
APRIL 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Mannix	Station number: AMS 05
Calibration Date:	April 4, 2024	Last Cal Date: March 20, 2024
Start time (MST):	8:18	End time (MST): 11:55
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	50.02 ppm	Cal Gas Exp Date: January 12, 2029
Cal Gas Cylinder #:	XC026809B	
Removed Cal Gas Conc:	50.02 ppm	Rem Gas Exp Date: N/A
Removed Gas Cyl #:	N/A	Diff between cyl:
Calibrator Model:	API T700	Serial Number: 621
Zero Air Gen Model:	API T701H	Serial Number: 832

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 1008841399
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005098	0.999943	Backgd or Offset:	9.6	9.6
Calibration intercept:	-0.700000	0.380000	Coeff or Slope:	0.944	0.944

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	80.0	800.3	800.9	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	800.9	Previous response	803.7	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4920	80.0	800.3	800.4	1.000
Mid point	4960	40.0	400.2	401.0	0.998
Low point	4980	20.0	200.1	200.5	0.998
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	80.0	800.3	803.9	0.996
Average Correction Factor:					0.999

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 20, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	8:18	End Time (MST):	11:55
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999999	≥0.995
800.3	800.4	0.9999	Slope	0.999943	0.90 - 1.10
400.2	401.0	0.9979	Intercept	0.380000	+/-30
200.1	200.5	0.9979			



SO2 Calibration Plot

Date: April 4, 2024

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	April 19, 2024	Last Cal Date:	March 12, 2024
Start time (MST):	10:27	End time (MST):	13:45
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.96	ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	DT0037363			
Removed Cal Gas Conc:	4.96	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1845
ZAG Make/Model:	API T701H		Serial Number:	832

Analyzer Information

Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326169
Converter make:	Global	Converter serial #:	2022225
Analyzer Range	0 - 100 ppb	Converter Temp:	350 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997116	0.995684	Backgd or Offset:	1.23	1.23
Calibration intercept:	-0.017688	-0.017567	Coeff or Slope:	0.978	0.978

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4919	80.6	80.0	79.5	1.006
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	79.5	Prev response:	79.71	*% change:	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4919	80.6	80.0	79.6	1.005
Mid point	4960	40.3	40.0	39.9	1.002
Low point	4980	20.2	20.0	19.7	1.017
As left zero	5000	0.0	0.0	0.1	----
As left span	4919	80.6	80.0	79.6	1.005
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.008
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber test after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 19, 2024	Previous Calibration:	March 12, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:27	End Time (MST):	13:45
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326169

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999976	≥0.995
80.0	79.6	1.0045	Slope	0.995684	0.90 - 1.10
40.0	39.9	1.0019	Intercept	-0.017567	+/-3
20.0	19.7	1.0171			



H₂S Calibration Plot

Date: April 19, 2024

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	April 4, 2024	Last Cal Date:	March 13, 2024
Start time (MST):	8:18	End time (MST):	11:55
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	
CH4 Cal Gas Conc.	504.9 ppm	CH4 Equiv Conc.	1076.6 ppm
C3H8 Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	504.9 ppm	CH4 Equiv Conc.	1076.6 ppm
Removed C3H8 Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
Zero Air Gen model:	API T701	Serial Number:	5613

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1170050130
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.76E-04	2.73E-04	NMHC SP Ratio: 5.97E-05	5.89E-05
CH4 Retention time:	14.4	1.44E+01	NMHC Peak Area: 153276	155276
Zero Chromatogram:	OFF	OFF	Flat Baseline: OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	17.23	17.49	0.985
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.49	Prev response	17.22	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.0	17.23	17.21	1.001
Mid point	4960	40.0	8.61	8.61	1.001
Low point	4980	20.0	4.31	4.29	1.004
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.0	17.23	17.27	0.998
Average Correction Factor					1.002

Notes: Changed the inlet filter and the H2 cylinder after as founds. Adjusted the span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	9.15	9.30	0.984
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.30	Prev response	9.16	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.0	9.15	9.13	1.002
Mid point	4960	40.0	4.57	4.57	1.000
Low point	4980	20.0	2.29	2.28	1.002
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.0	9.15	9.16	0.998
Average Correction Factor					1.001

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	8.08	8.19	0.987
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.19	Prev response	8.06	*% change	1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.0	8.08	8.09	0.999
Mid point	4960	40.0	4.04	4.03	1.001
Low point	4980	20.0	2.02	2.01	1.006
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.0	8.08	8.10	0.997
Average Correction Factor					1.002

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999917	0.999340
THC Cal Offset:	-0.008000	-0.004400
CH ₄ Cal Slope:	1.000113	1.001641
CH ₄ Cal Offset:	-0.024200	-0.007600
NMHC Cal Slope:	0.999769	0.997707
NMHC Cal Offset:	0.016600	0.002600

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 13, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	8:18	End Time (MST):	11:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥0.995
17.23	17.21	1.0009	Slope	0.999340	0.90 - 1.10
8.61	8.61	1.0006	Intercept	-0.004400	+/-0.5
4.31	4.29	1.0038			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 13, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	8:18	End Time (MST):	11:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999996	<i>≥0.995</i>
8.08	8.09	0.9988	Slope	1.001641	<i>0.90 - 1.10</i>
4.04	4.03	1.0013	Intercept	-0.007600	<i>+/-0.5</i>
2.02	2.01	1.0058			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 13, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	8:18	End Time (MST):	11:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
9.15	9.13	1.0024	Slope	0.997707	<i>0.90 - 1.10</i>
4.57	4.57	1.0002	Intercept	0.002600	<i>+/-0.5</i>
2.29	2.28	1.0017			



NMHC Calibration Plot

Date: April 4, 2024

Location: Mannix





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	April 23, 2024	Last Cal Date:	March 7, 2024
Start time (MST):	8:26	End time (MST):	11:45
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.78	ppm	Cal Gas Exp Date:	September 9, 2024
Cal Gas Cylinder #:	AAL070632			
Removed Cal Gas Conc:	49.78	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	3566
Zero Air Gen Model:	API T701		Serial Number:	5608

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1160290013
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997030	1.004978	Backgd or Offset:	17.7	17.7
Calibration intercept:	1.560821	1.779365	Coeff or Slope:	0.922	0.922

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4919.7	80.3	799.5	800.9	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	800.9	Previous response	798.7	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.5	----
High point	4919.7	80.3	799.5	804.7	0.993
Mid point	4959.8	40.2	400.2	404.3	0.990
Low point	4979.9	20.1	200.1	204.4	0.979
As left zero	5000	0.0	0.0	0.3	----
As left span	4919.7	80.3	799.5	805.0	0.993
Average Correction Factor:					0.987

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 23, 2024	Previous Calibration:	March 7, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:26	End Time (MST):	11:45
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.5	----	Correlation Coefficient	0.999988	≥0.995
799.5	804.7	0.9935	Slope	1.004978	0.90 - 1.10
400.2	404.3	0.9899	Intercept	1.779365	+/-30
200.1	204.4	0.9790			



SO2 Calibration Plot

Date: April 23, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	April 18, 2024	Last Cal Date:	March 8, 2024
Start time (MST):	9:05 AM	End time (MST):	13:30
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.328	ppm	Cal Gas Exp Date:	February 14, 2025
Cal Gas Cylinder #:	CC506659			
Removed Cal Gas Conc:	5.328	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb	Converter Temp:	825 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993746	0.996746	Backgd or Offset:	1.95	1.98
Calibration intercept:	0.540147	0.460147	Coeff or Slope:	1.150	1.170

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4925	75.1	80.0	78.2	1.023
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	78.2	Prev response:	80.06	*% change:	-2.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	<i>* = +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4925	75.1	80.0	80.1	0.999
Mid point	4963	37.5	40.0	40.4	0.989
Low point	4981	18.8	20.0	20.7	0.968
As left zero	5000	0.0	0.0	0.9	----
As left span	4925	75.1	80.0	80.9	0.989
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:		December 20, 2021		Ave Corr Factor	0.985
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 8, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:05	End Time (MST):	13:30
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999951	≥ 0.995
80.0	80.1	0.9991	Slope	0.996746	$0.90 - 1.10$
40.0	40.4	0.9890	Intercept	0.460147	± 3
20.0	20.7	0.9678			



TRS Calibration Plot

Date: April 18, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	April 22, 2024	Last Cal Date:	March 7, 2024
Start time (MST):	9:43 AM	End time (MST):	11:15
Reason:	As Found		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH4 Cal Gas Conc.	501.6 ppm	CH4 Equiv Conc.	1066.2 ppm
C3H8 Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	501.6 ppm	CH4 Equiv Conc.	1066.2 ppm
Removed C3H8 Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
Zero Air Gen model:	API T701	Serial Number:	4602

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1118148495
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.21E-04	2.21E-04	4.84E-05	4.84E-05
CH4 Retention time:	14.2	14.2	187266	187266
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	17.12	16.78	1.021
As found Mid point	4960	40.2	8.57	8.41	1.019
As found Low point	4980	20.1	4.29	4.27	1.004
New cylinder response					
Baseline Corr AF:	16.78	Prev response	17.12	*% change	-2.1%
Baseline Corr 2nd AF:	8.41	AF Slope:	0.978449	AF Intercept:	0.030920
Baseline Corr 3rd AF:	4.27	AF Correlation:	0.999981	<i>* = > +/-5% change initiates investigation</i>	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Notes: Multipoint as founds completed to run a column bakeout. Purpose is to get perfect zero readings during zeros.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919.7	80.3	9.07	8.91	1.018
As found Mid point	4960	40.2	4.54	4.49	1.012
As found Low point	4980	20.1	2.27	2.29	0.990
New cylinder response					
Baseline Corr AF:	8.91	Prev response	9.08	*% change	-1.9%
Baseline Corr 2nd AF:	4.49	AF Slope:	0.980742	AF Intercept:	0.031027
Baseline Corr 3rd AF:	2.29	AF Correlation:	0.999944	* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919.7	80.3	8.06	7.86	1.025
As found Mid point	4960	40.2	4.03	3.92	1.028
As found Low point	4980	20.1	2.02	1.98	1.021
New cylinder response					
Baseline Corr AF:	7.86	Prev response	8.05	*% change	-2.3%
Baseline Corr 2nd AF:	3.92	AF Slope:	0.975585	AF Intercept:	-0.000106
Baseline Corr 3rd AF:	1.98	AF Correlation:	0.999993	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998467	
THC Cal Offset:	0.026638	
CH ₄ Cal Slope:	0.998356	
CH ₄ Cal Offset:	0.002848	
NMHC Cal Slope:	0.998464	
NMHC Cal Offset:	0.024190	

Calibration Performed By: Max Farrell

NMHC Calibration Plot

Date: April 22, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	April 23, 2024	Last Cal Date:	March 7, 2024
Start time (MST):	9:15 AM	End time (MST):	11:45
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH ₄ Cal Gas Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	501.6 ppm	CH ₄ Equiv Conc.	1066.2 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
Zero Air Gen model:	API T701	Serial Number:	4602

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1118148495
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.21E-04	2.26E-04	4.84E-05	4.92E-05
CH ₄ Retention time:	14.2	14.4	187266	184279
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919.7	80.3	17.12	17.14	0.999
Mid point	4960	40.2	8.57	8.58	0.999
Low point	4980	20.1	4.29	4.35	0.986
As left zero	5000	0.0	0.00	0.01	----
As left span	4919.7	80.3	17.12	17.14	0.999
Average Correction Factor					0.995

Notes: Completed multipoint as founds and ran a column bakeout yesterday. Adjusted the span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919.7	80.3	9.07	9.10	0.997
Mid point	4960	40.2	4.54	4.57	0.993
Low point	4980	20.1	2.27	2.33	0.975
As left zero	5000	0.0	0.00	0.01	----
As left span	4919.7	80.3	9.07	9.10	0.997
Average Correction Factor					0.988

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919.7	80.3	8.06	8.05	1.001
Mid point	4960	40.2	4.03	4.01	1.006
Low point	4980	20.1	2.02	2.02	0.999
As left zero	5000	0.0	0.00	0.00	----
As left span	4919.7	80.3	8.06	8.04	1.001
Average Correction Factor					1.002

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998467	0.999948
THC Cal Offset:	0.026638	0.023034
CH ₄ Cal Slope:	0.998356	0.998696
CH ₄ Cal Offset:	0.002848	-0.002350
NMHC Cal Slope:	0.998464	1.001388
NMHC Cal Offset:	0.024190	0.024584

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 23, 2024	Previous Calibration:	March 7, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:15	End Time (MST):	11:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999987 ≥0.995
17.12	17.14	0.9989	Slope	0.999948 0.90 - 1.10
8.57	8.58	0.9987	Intercept	0.023034 +/-0.5
4.29	4.35	0.9860		





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 23, 2024	Previous Calibration:	March 7, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:15	End Time (MST):	11:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999990	<i>≥0.995</i>
8.06	8.05	1.0008	Slope	0.998696	<i>0.90 - 1.10</i>
4.03	4.01	1.0059	Intercept	-0.002350	<i>+/-0.5</i>
2.02	2.02	0.9987			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 23, 2024	Previous Calibration:	March 7, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:15	End Time (MST):	11:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999964	<i>≥0.995</i>
9.07	9.10	0.9969	Slope	1.001388	<i>0.90 - 1.10</i>
4.54	4.57	0.9926	Intercept	0.024584	<i>+/-0.5</i>
2.27	2.33	0.9749			



NMHC Calibration Plot

Date: April 23, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Patricia McInnes
 Station number: AMS 06
 Calibration Date: April 10, 2024
 Last Cal Date: March 5, 2024
 Start time (MST): 9:26 AM
 End time (MST): 14:17
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T30YCWN
 NOX Cal Gas Conc: 47.94 ppm
 Removed Cylinder #: N/A
 Removed Gas NOX Conc: 47.94 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: April 11, 2025
 NO Cal Gas Conc: 46.39 ppm
 Removed Gas Exp Date: N/A
 Removed Gas NO Conc: 46.39 ppm
 NO gas Diff:
 Serial Number: 3566
 Serial Number: 4602

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.3	-0.4	----	----
AF High point	4914	86.2	826.5	799.7	26.7	817.6	788.7	28.9	1.0107	1.0144
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 830.4 ppb		NO = 803.5 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -1.6%	
Baseline Corr 1st pt	NO _x = 817.7 ppb		NO = 788.4 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -1.9%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1172750022

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002416	0.997990
NO _x Cal Offset:	1.975128	2.376066
NO Cal Slope:	1.003032	1.000731
NO Cal Offset:	1.382192	1.562605
NO ₂ Cal Slope:	0.999602	1.000336
NO ₂ Cal Offset:	-0.538439	-0.255207

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.825	0.836	NO bkgnd or offset:	3.2	3.3
NOX coeff or slope:	0.987	0.987	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.2	155.4

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.5	-0.2	----	----
High point	4914	86.2	826.5	799.7	26.7	826.1	801.3	25.0	1.0004	0.9980
Mid point	4957	43.1	413.2	399.9	13.4	415.9	402.4	13.5	0.9936	0.9937
Low point	4978	21.6	207.1	200.4	6.7	211.1	203.1	8.0	0.9811	0.9868
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.4	-0.2	----	----
As left span	4914	86.2	826.5	398.9	427.6	826.8	398.9	427.8	0.9996	1.0000
Average Correction Factor									0.9917	0.9929

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	799.3	399.8	426.2	426.5	0.9993	100.1%
Mid GPT point	799.3	605.7	220.3	219.0	1.0060	99.4%
Low GPT point	799.3	704.3	121.7	122.2	0.9961	100.4%
Average Correction Factor					1.0005	100.0%

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 10, 2024	Previous Calibration:	March 5, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:26	End Time (MST):	14:17
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999971	≥0.995
826.5	826.1	1.0004	Slope	0.997990	0.90 - 1.10
413.2	415.9	0.9936	Intercept	2.376066	+/-20
207.1	211.1	0.9811			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 10, 2024	Previous Calibration:	March 5, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:26	End Time (MST):	14:17
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999980	≥0.995
426.2	426.5	0.9993	Slope	1.000336	0.90 - 1.10
220.3	219.0	1.0060	Intercept	-0.255207	+/-20
121.7	122.2	0.9961			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 10, 2024	Previous Calibration:	March 5, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:26	End Time (MST):	14:17
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.5	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
799.7	801.3	0.9980	Slope	1.000731	<i>0.90 - 1.10</i>
399.9	402.4	0.9937	Intercept	1.562605	<i>+/-20</i>
200.4	203.1	0.9868			



NO_x Calibration Plot

Date: April 10, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	April 12, 2024	Last Cal Date:	March 22, 2024
Start time (MST):	10:02	End time (MST):	13:00
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3566
Calibrator Make/Model:	API T700	Serial Number:	4602
ZAG Make/Model:	API T701		

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004200	1.006514	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	-0.660000	-0.840000	Coeff or Slope:	1.026	1.026

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-0.3	----
As found High point	5000	1303.0	400.0	401.2	0.996
As found Mid point					
As found Low point					
Baseline Corr As found:	401.5	Previous response	401.0	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	-0.3	----
High point	5000	1303.0	400.0	402.0	0.995
Mid point	5000	966.5	200.0	200.3	0.999
Low point	5000	794.3	100.0	99.2	1.008
As left zero	5000	800.0	0.0	-0.1	----
As left span	5000	1303.0	400.0	402.4	0.994
Average Correction Factor					1.001

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 12, 2024	Previous Calibration:	March 22, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:02	End Time (MST):	13:00
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999992	≥0.995
400.0	402.0	0.9950	Slope	1.006514	0.90 - 1.10
200.0	200.3	0.9985	Intercept	-0.840000	+/- 5
100.0	99.2	1.0081			



O₃ Calibration Plot

Date: April 12, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: April 22, 2024 Last Cal Date: March 22, 2024
 Start time (MST): 12:56 End time (MST): 13:57

Analyzer Make: API T640 S/N: 766
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388755
 Temp/RH standard: Alicat FP-25BT S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	17.8	17.9	17.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	720.3	722.7	720.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.14	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	42	----	42	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	12.2	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: 11-23-2023
 Lot No.: 100128-050-035

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ March 22, 2024
 Date Disposable Filter Changed: _____ March 22, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ April 13, 2023
 Date RH/T Sensor Cleaned: _____ April 13, 2023

Notes: Quarterly calibrations completed last month. Leak check passed, no adjustments made.

Calibration by: Max Farrell



Wood Buffalo Environmental Association

Nt - NOX - NH3 Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	April 11, 2024	Last Cal Date:	March 6, 2024
Start time (MST):	7:01	End time (MST):	11:11
NH3 Cal Date:	April 11, 2024	Last Cal Date:	March 6, 2024
Start time (MST):	11:20	End time (MST):	13:00
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	N/A
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	N/A
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.3	ppm	NH3 Gas Cylinder #:	EB0108520
Removed NH3 Conc:	76.3	ppm	NH3 Cal Gas Expiry:	August 22, 2024
NH3 gas Diff:			Removed Cylinder #:	N/A
Calibrator Model:	API T700		Removed cyl Expiry:	N/A
ZAG make/model:	API T701		Serial Number:	3566
			Serial Number:	4602

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	808
Converter model:	API T501	Converter serial #:	484
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	4.80
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	26.6

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.854	0.854	Nt coefficient:	0.849	0.849
NOX coefficient:	0.848	0.848	NO bkgrnd:	-1.0	-1.0
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	-0.6	-0.6
NH3 coefficient:	0.896	0.896	Nt bkgrnd:	5.0	5.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.992580	0.993922
NO _x Cal Offset:	1.257688	1.997132
NO Cal Slope:	0.996357	0.997329
NO Cal Offset:	-0.056315	0.703361
NO ₂ Cal Slope:	1.001744	0.997014
NO ₂ Cal Offset:	-0.225723	-0.514270
NH3 Cal Slope:	1.000069	0.990318
NH3 Cal Offset:	3.359353	4.179726
Nt Cal Slope:	1.006770	0.997067
Nt Cal Offset:	4.605099	5.735735



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

NO_x / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO _x Correction factor (Cc/Ic)	Baseline corr NO Correction factor (Cc/Ic)
									<i>Limit = 0.9 - 1.0</i>	<i>Limit = 0.9 - 1.0</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	-1.9	----	----
As found span	4914	86.2	826.5	799.7	826.5	818.0	793.2	821.9	1.0103	1.0082
AF GPT span										
new NO cyl rp										

Baseline Corr As Fd	Nt = 823.8 ppb	NO _x = 817.7 ppb	NO = 792.9 ppb		*Percent Change	Nt _(NO) = -1.6%
Previous Response	Nt = 836.65 ppb	NO _x = 821.6 ppb	NO = 796.8 ppb		*Percent Change	NO _x = -0.5%
**NO _x Δ (NO to GPT response) =					*Percent Change	NO = -0.5%
<i>** = > +/-2% difference initiates investigation</i>					<i>* = > +/-5% change initiates investigation</i>	

NO_x / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic)	NO Correction factor (Cc/Ic)
									<i>Limit = 0.95-1.05</i>	<i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.6	0.3	-0.9	----	----
High point	4914	86.2	826.5	799.7	826.5	822.8	798.3	822.4	1.0044	1.0018
Mid point	4957	43.1	413.2	399.9	413.2	413.2	399.1	415.3	1.0001	1.0019
Low point	4978	21.6	207.1	200.4	207.1	209.4	201.4	209.9	0.9891	0.9951
Average Correction Factor									0.9979	0.9996

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic)	Converter Efficiency
					<i>Limit = 0.95-1.05</i>	<i>Limit = 96-104%</i>
Calibration zero	----	----	0.0	0.2	----	----
High GPT point (400 ppb O3)	795.0	394.7	427.0	425.7	1.0031	99.7%
Mid GPT point (200 ppb O3)	795.0	601.4	220.3	218.6	1.0079	99.2%
Low GPT point (100 ppb O3)	795.0	696.9	124.8	123.3	1.0123	98.8%
Average Correction Factor					1.0078	99.2%



Wood Buffalo Environmental Association NH₃ - N_T Calibration Report

NH₃ As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH3 Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.9	0.3	-2.2	----	----
AF High point	3417	82.6	1800.6	----	1800.6	1795.1		1782.1	1.002	1.009
AF Mid point							----			
AF Low point							----			
new NH3 cyl rp										
Baseline Corr As Fd		Nt = 1797.0 ppb	NH3 = 1784.3 ppb						*Percent Change	Nt _(NH3) = -1.1%
Previous Response		Nt = 1817.4 ppb	NH3 = 1804.1 ppb			* = > +/-5% change initiates investigation			*Percent Change	NH3 = -1.1%

NH₃ Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	-0.9	0.6	-1.4	----	----
High point	3417	82.6	1800.6	----	1800.6	1795.1	----	1782.1	1.003	1.010
Mid point	3454	45.9	1000.5	----	1000.5	1012.3	----	1003.2	0.988	0.997
Low point	3477	22.9	499.2	----	499.2	507.1	----	501.2	0.984	0.996
Average Correction Factor									0.9919	1.0012
NH3 Previous Converter Efficiency =		90.8 %								
NH3 Current Converter Efficiency =		90.8 %								

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


Nt Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 6, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:01	End Time (MST):	11:11
Analyzer make:	API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.9	----	Correlation Coefficient	0.999901	<i>≥0.995</i>
1800.6	1795.1	1.0031	Slope	0.997067	<i>0.90 - 1.10</i>
1000.5	1012.3	0.9884	Intercept	5.735735	<i>+/-20</i>
499.2	507.1	0.9844			





Wood Buffalo Environmental Association


NH₃ Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 6, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:01	End Time (MST):	11:11
Analyzer make:	API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-1.4	----	Correlation Coefficient	0.999923	<i>≥0.995</i>
1800.6	1782.1	1.0104	Slope	0.990318	<i>0.90 - 1.10</i>
1000.5	1003.2	0.9973	Intercept	4.179726	<i>+/-20</i>
499.2	501.2	0.9959			





Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 6, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:01	End Time (MST):	11:11
Analyzer make:	API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.6	----	Correlation Coefficient	0.999987	<i>≥0.995</i>
826.5	822.8	1.0044	Slope	0.993922	<i>0.90 - 1.10</i>
413.2	413.2	1.0001	Intercept	1.997132	<i>+/-20</i>
207.1	209.4	0.9891			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 6, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:01	End Time (MST):	11:11
Analyzer make:	API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999997	<i>≥0.995</i>
799.7	798.3	1.0018	Slope	0.997329	<i>0.90 - 1.10</i>
399.9	399.1	1.0019	Intercept	0.703361	<i>+/-20</i>
200.4	201.4	0.9951			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 6, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	7:01	End Time (MST):	11:11
Analyzer make:	API T201	Analyzer serial #:	808

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999985	<i>≥0.995</i>
427.0	425.7	1.0031	Slope	0.997014	<i>0.90 - 1.10</i>
220.3	218.6	1.0079	Intercept	-0.514270	<i>+/-20</i>
124.8	123.3	1.0123			



NO_x Calibration Plot

Date: April 11, 2024


Location: Patricia McInnes



NH₃ Calibration Plot

Date: April 11, 2024

Location: Patricia McInnes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	April 10, 2024	Last Cal Date:	March 22, 2024
Start time (MST):	10:45	End time (MST):	14:27
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.06	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC320556			
Removed Cal Gas Conc:	50.06	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	3805
Zero Air Gen Model:	API 701H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Serial Number:	1507864683
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000577	1.000104	Backgd or Offset:	2.6	2.6
Calibration intercept:	1.744314	1.984432	Coeff or Slope:	0.845	0.845

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	79.8	799.0	798.8	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	798.8	Previous response	801.2	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4920	79.8	799.0	799.5	0.999
Mid point	4960	39.9	399.5	404.5	0.988
Low point	4980	20.0	200.2	202.5	0.989
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	79.8	799.0	799.1	1.000
Average Correction Factor:					0.992

Notes: No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 10, 2024	Previous Calibration:	March 22, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:45	End Time (MST):	14:27
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999959	≥0.995
799.0	799.5	0.9994	Slope	1.000104	0.90 - 1.10
399.5	404.5	0.9876	Intercept	1.984432	+/-30
200.2	202.5	0.9888			



SO2 Calibration Plot

Date: April 10, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	April 15, 2024	Last Cal Date:	March 7, 2024
Start time (MST):	9:30	End time (MST):	14:27
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.25	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC504080			
Removed Cal Gas Conc:	5.25	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API T701H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 43i LTE	Analyzer serial #:	1180540018
Converter make:	CDN-101	Converter serial #:	551
Analyzer Range	0 - 100 ppb	Converter Temp:	840 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998195	1.013471	Backgd or Offset:	2.4	2.4
Calibration intercept:	0.217844	0.037852	Coeff or Slope:	0.901	0.901

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero)) Limit = 0.90-1.10
As found zero	5000	0.0	0.0	0.0	----
As found High point	4925	75.5	79.3	81.0	0.979
As found Mid point	4962	37.7	39.6	40.4	0.980
As found Low point	4981	18.9	19.8	20.0	0.992
New cylinder response					
Baseline Corr As found:	81.0	Prev response:	79.34	*% change:	2.0%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.022992	AF Intercept:	-0.122207
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999986	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4925	75.5	79.3	80.4	0.986
Mid point	4962	37.7	39.6	40.3	0.983
Low point	4981	18.9	19.9	20.0	0.993
As left zero	5000	0.0	0.0	0.3	----
As left span	4925	75.5	79.3	80.6	0.984
SO2 Scrubber Check	4920	79.2	792.1	0.1	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	0.987
Date of last converter efficiency test:	April 22, 2022				

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	April 15, 2024	Previous Calibration:	March 7, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:30	End Time (MST):	14:27
Analyzer make:	Thermo 43i LTE	Analyzer serial #:	1180540018

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999987	≥ 0.995
79.3	80.4	0.9865	Slope	1.013471	0.90 - 1.10
39.6	40.3	0.9829	Intercept	0.037852	+/-3
19.9	20.0	0.9928			



TRS Calibration Plot

Date: April 15, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS 07
Calibration Date:	April 10, 2024	Last Cal Date:	March 26, 2024
Start time (MST):	10:45	End time (MST):	14:27
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	496.0 ppm	CH4 Equiv Conc.	1059.8 ppm
C3H8 Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	496.0 ppm	CH4 Equiv Conc.	1059.8 ppm
Removed C3H8 Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3805
Zero Air Gen model:	Teledyne API T701H	Serial Number:	198

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 12227620777
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.55E-04	2.60E-05	NMHC SP Ratio:	5.32E-05	5.34E-05
CH4 Retention time:	13.4	13.4	NMHC Peak Area:	169129	168530
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.91	16.77	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.77	Prev response	16.97	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.91	16.88	1.002
Mid point	4960	39.9	8.46	8.44	1.003
Low point	4980	20.0	4.24	4.27	0.994
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.91	16.98	0.996
Average Correction Factor					1.000

Notes: Hydrogen cylinder swapped out after as founds. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.00	9.02	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.02	Prev response	9.06	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.00	8.96	1.004
Mid point	4960	39.9	4.50	4.50	1.000
Low point	4980	20.0	2.26	2.28	0.989
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.00	9.04	0.996
Average Correction Factor					0.998

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	7.92	7.75	1.021
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.75	Prev response	7.91	*% change	-2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	7.92	7.92	1.000
Mid point	4960	39.9	3.96	3.94	1.005
Low point	4980	20.0	1.98	1.99	0.999
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	7.92	7.94	0.997
Average Correction Factor					1.001

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002451	0.997119
THC Cal Offset:	0.015044	0.013262
CH ₄ Cal Slope:	1.000657	0.999590
CH ₄ Cal Offset:	-0.011935	-0.002739
NMHC Cal Slope:	1.003673	0.995046
NMHC Cal Offset:	0.027380	0.016600

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 10, 2024	Previous Calibration:	March 26, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:45	End Time (MST):	14:27
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999994	<i>≥0.995</i>
16.91	16.88	1.0021	Slope	0.997119	<i>0.90 - 1.10</i>
8.46	8.44	1.0026	Intercept	0.013262	<i>+/-0.5</i>
4.24	4.27	0.9939			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 10, 2024	Previous Calibration:	March 26, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:45	End Time (MST):	14:27
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999992	<i>>0.995</i>
7.92	7.92	1.0001	Slope	0.999590	<i>0.90 - 1.10</i>
3.96	3.94	1.0046	Intercept	-0.002739	<i>+/-0.5</i>
1.98	1.99	0.9990			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 10, 2024	Previous Calibration:	March 26, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:45	End Time (MST):	14:27
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999984	<i>>0.995</i>
9.00	8.96	1.0039	Slope	0.995046	<i>0.90 - 1.10</i>
4.50	4.50	1.0004	Intercept	0.016600	<i>+/-0.5</i>
2.26	2.28	0.9890			



NMHC Calibration Plot

Date: April 10, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Athabasca Valley
 Station number: AMS 07
 Calibration Date: April 9, 2024
 Last Cal Date: February 27, 2024
 Start time (MST): 9:00
 End time (MST): 14:33
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0033919
 NOX Cal Gas Conc: 60.10 ppm
 Removed Cylinder #: N/A
 Removed Gas NOX Conc: 60.10 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701H
 Cal Gas Expiry Date: January 9, 2032
 NO Cal Gas Conc: 59.90 ppm
 Removed Gas Exp Date: N/A
 Removed Gas NO Conc: 59.90 ppm
 NO gas Diff:
 Serial Number: 3805
 Serial Number: 198

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
AF High point	4933	66.8	803.0	800.3	2.7	803.7	795.6	7.9	0.9991	1.0058
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 804.5 ppb		NO = 800.1 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -0.1%	
Baseline Corr 1st pt	NO _x = 803.7 ppb		NO = 795.7 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -0.6%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1160120024

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999973	0.997568
NO _x Cal Offset:	1.541897	1.871922
NO Cal Slope:	0.998014	0.997614
NO Cal Offset:	1.431946	1.471946
NO ₂ Cal Slope:	1.008548	1.006328
NO ₂ Cal Offset:	1.193739	1.745224

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.075	1.077	NO bkgnd or offset:	7.6	7.6
NOX coeff or slope:	1.005	1.003	NOX bkgnd or offset:	7.9	7.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	214.5	212.0

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.2	----	----
High point	4933	66.8	803.0	800.3	2.7	801.7	798.7	3.0	1.0016	1.0020
Mid point	4966	33.4	401.5	400.2	1.3	404.4	402.8	1.6	0.9929	0.9935
Low point	4983	16.7	200.7	200.1	0.7	202.9	201.5	1.4	0.9894	0.9929
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	-0.1	----	----
As left span	4933	66.8	803.0	393.8	409.2	804.0	393.8	410.2	0.9987	1.0000
Average Correction Factor									0.9946	0.9961

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	794.0	397.2	399.5	402.9	0.9915	100.9%
Mid GPT point	794.0	598.8	197.9	201.9	0.9801	102.0%
Low GPT point	794.0	698.3	98.4	102.1	0.9635	103.8%
Average Correction Factor					0.9783	102.2%

Notes:

Span adjusted slightly.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 9, 2024	Previous Calibration:	February 27, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:00	End Time (MST):	14:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999976	≥0.995
803.0	801.7	1.0016	Slope	0.997568	0.90 - 1.10
401.5	404.4	0.9929	Intercept	1.871922	+/-20
200.7	202.9	0.9894			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 9, 2024	Previous Calibration:	February 27, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:00	End Time (MST):	14:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999932	≥0.995
399.5	402.9	0.9915	Slope	1.006328	0.90 - 1.10
197.9	201.9	0.9801	Intercept	1.745224	+/-20
98.4	102.1	0.9635			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 9, 2024	Previous Calibration:	February 27, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:00	End Time (MST):	14:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999978	<i>≥0.995</i>
800.3	798.7	1.0020	Slope	0.997614	<i>0.90 - 1.10</i>
400.2	402.8	0.9935	Intercept	1.471946	<i>+/-20</i>
200.1	201.5	0.9929			



NO_x Calibration Plot

Date: April 9, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	April 12, 2024	Last Cal Date:	March 17, 2024
Start time (MST):	9:55	End time (MST):	14:29
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3805
Calibrator Make/Model:	T700	Serial Number:	198
ZAG Make/Model:	T701H		

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993629	0.999857	Backgd or Offset:	-1.6	-1.6
Calibration intercept:	1.040000	0.500000	Coeff or Slope:	1.549	1.549

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-2.0	----
As found High point	5000	1522.8	400.0	397.8	1.001
As found Mid point					
As found Low point					
Baseline Corr As found:	399.8	Previous response	398.5	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.5	----
High point	5000	1523.6	400.0	400.0	1.000
Mid point	5000	1088.1	200.0	200.8	0.996
Low point	5000	880.5	100.0	101.6	0.984
As left zero	5000	0.0	0.0	-0.7	----
As left span	5000	1522.4	400.0	404.5	0.989
Average Correction Factor					0.993

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 12, 2024	Previous Calibration:	March 17, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:55	End Time (MST):	14:29
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.5	----	Correlation Coefficient	0.999971	≥0.995
400.0	400.0	1.0000	Slope	0.999857	0.90 - 1.10
200.0	200.8	0.9960	Intercept	0.500000	+/- 5
100.0	101.6	0.9843			



O₃ Calibration Plot

Date: April 12, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Athabasca Valley Station number: AMS 07
Calibration Date: April 15, 2024 Last Cal Date: March 27, 2024
Start time (MST): 13:19 End time (MST): 14:29
Analyzer Make: S/N: 645
Particulate Fraction: PM2.5
Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	5.4	5.8	5.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	736.1	735	736.1	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.05	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	----	37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.5	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 6, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 27, 2024
Date Disposable Filter Changed: February 27, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022
Date RH/T Sensor Cleaned: December 5, 2022

Notes: No adjustments needed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS 07
Calibration Date:	April 22, 2024	Last Cal Date:	March 27, 2024
Start time (MST):	12:05	End time (MST):	15:00
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date: December 12, 2026
Cal Gas Cylinder #:	LL66942		
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	API T700		Serial Number: 3805
ZAG Make/Model:	API 700H		Serial Number: 198

Analyzer Information

Analyzer make:	Thermo 48i-TLE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994233	0.990260	Backgd or Offset:	4.642	4.819
Calibration intercept:	0.090507	0.092551	Coeff or Slope:	1.087	1.087

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4933	66.7	40.0	40.1	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	39.98	Prev response:	39.88	*% change:	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4933	66.7	40.0	39.7	1.009
Mid point	4967	33.3	20.0	20.0	0.998
Low point	4983	16.7	10.0	10.1	0.997
As left zero	5000	0.0	0.0	0.0	----
As left span	4933	66.7	40.0	39.7	1.009
Average Correction Factor					1.002

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


CO Calibration Summary

Station Information

Calibration Date:	April 22, 2024	Previous Calibration:	March 27, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	12:05	End Time (MST):	15:00
Analyzer make:	Thermo 48i-TLE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999961	≥0.995
40.0	39.7	1.0094	Slope	0.990260	0.90 - 1.10
20.0	20.0	0.9984	Intercept	0.092551	+/-1.5
10.0	10.1	0.9971			



CO Calibration Plot

Date: April 22, 2024

Location: Athabasca Valley





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	April 17, 2024	Last Cal Date:	March 20, 2024
Start time (MST):	13:49	End time (MST):	16:28
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date: January 6, 2030
Cal Gas Cylinder #:	CC196697		
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	TELEDYNE API T700		Serial Number: 3252
Zero Air Gen Model:	TELEDYNE API T701		Serial Number: 135

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Serial Number:	1136451241
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003858	0.999602	Backgd or Offset:	1.8	1.8
Calibration intercept:	0.615265	0.656198	Coeff or Slope:	0.989	0.989

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000			-0.5	----
As found High point	4920			799.1	
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	799.6	Previous response	804.1	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	4920	80.3	800.4	801.0	0.999
Mid point	4960	40.2	400.7	399.4	1.003
Low point	4980	20.1	200.4	203.4	0.985
As left zero	5000	0.0	0.0	-0.2	----
As left span	4920	80.3	800.4	799.2	1.001
Average Correction Factor:					0.996

Notes: Changed out inlet filter after as founds. No adjustments made

Calibration Performed By: Matthew Courtoreille



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 17, 2024	Previous Calibration:	March 20, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	13:49	End Time (MST):	16:28
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999971	≥0.995
800.4	801.0	0.9992	Slope	0.999602	0.90 - 1.10
400.7	399.4	1.0032	Intercept	0.656198	+/-30
200.4	203.4	0.9850			



SO2 Calibration Plot

Date: April 17, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS 08
Calibration Date:	April 18, 2024	Last Cal Date:	March 20, 2024
Start time (MST):	14:12	End time (MST):	17:57
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.97	ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	EY0002276			
Removed Cal Gas Conc:	4.97	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3252
ZAG Make/Model:	Teledyne API T701		Serial Number:	135

Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744
Converter make:	CDN-101	Converter serial #:	14639
Analyzer Range	0 - 100 ppb	Converter Temp:	834 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003284	1.007424	Backgd or Offset:	1.0	1.0
Calibration intercept:	0.498794	0.578976	Coeff or Slope:	0.754	0.754

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.6	----
As found High point	4920	80.5	80.0	80.5	1.001
As found Mid point	4960	40.2	40.0	40.8	0.994
As found Low point	4980	20.1	20.0	20.8	0.989
New cylinder response					
Baseline Corr As found:	79.9	Prev response:	80.77	*% change:	-1.1%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	0.997996	AF Intercept:	0.758823
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999979	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.5	----
High point	4920	80.5	80.0	81.1	0.987
Mid point	4960	40.2	40.0	41.0	0.975
Low point	4980	20.1	20.0	20.7	0.965
As left zero	5000	0.0	0.0	0.6	----
As left span	4920	80.5	80.0	81.2	0.985
SO2 Scrubber Check	4919.7	80.3	803.0	0.1	----
Date of last scrubber change:		7-Mar-22		Ave Corr Factor	0.975
Date of last converter efficiency test:		March 15, 2022		100.7%	efficiency

Notes: Sampled inlet filter after as founds. Scrubber checked passed. No adjustments made.

Calibration Performed By: Matthew Courtoreille



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 20, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	14:12	End Time (MST):	17:57
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999988	≥ 0.995
80.0	81.1	0.9865	Slope	1.007424	$0.90 - 1.10$
40.0	41.0	0.9746	Intercept	0.578976	± 3
20.0	20.7	0.9652			



TRS Calibration Plot

Date: April 18, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan
 Station number: AMS 08
 Calibration Date: April 18, 2024
 Last Cal Date: March 20, 2024
 Start time (MST): 8:24
 End time (MST): 13:13
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DTO046831
 NOX Cal Gas Conc: 60.20 ppm
 Removed Cylinder #: DT0046831
 Removed Gas NOX Conc: 60.20 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701H
 Cal Gas Expiry Date: January 9, 2032
 NO Cal Gas Conc: 60.00 ppm
 Removed Gas Exp Date: January 9, 2032
 Removed Gas NO Conc: 60.00 ppm
 NO gas Diff:
 Serial Number: 3252
 Serial Number: 135

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	1.4	1.4	0.0	----	----
AF High point	4933	66.7	803.1	800.4	2.7	779.6	778.0	1.7	1.0320	1.0307
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 802.5 ppb		NO = 800.4 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -3.1%	
Baseline Corr 1st pt	NO _x = 778.2 ppb		NO = 776.6 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -3.1%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: API T200
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 4460

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995037	1.000160
NO _x Cal Offset:	3.395408	2.195593
NO Cal Slope:	0.998540	0.992944
NO Cal Offset:	1.155021	2.314346
NO ₂ Cal Slope:	0.993301	0.997333
NO ₂ Cal Offset:	0.722793	1.324008

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.200	1.231	NO bkgnd or offset:	-2.5	2.5
NOX coeff or slope:	1.194	1.229	NOX bkgnd or offset:	-2.2	2.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	-2.9	-1.8

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	1.6	1.5	0.1	----	----
High point	4933	66.7	803.1	800.4	2.7	804.7	796.7	8.0	0.9980	1.0047
Mid point	4967	33.3	400.9	399.6	1.3	404.8	399.5	5.4	0.9904	1.0002
Low point	4983	16.7	201.1	200.4	0.7	203.0	202.1	0.9	0.9905	0.9916
As left zero	5000	0.0	0.0	0.0	0.0	1.3	1.3	0.0	----	----
As left span	4933	66.7	803.1	385.4	417.7	795.3	385.4	409.7	1.0098	1.0000
Average Correction Factor									0.9930	0.9988

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	798.4	389.4	411.7	412.0	0.9992	100.1%
Mid GPT point	798.4	597.9	203.2	202.5	1.0033	99.7%
Low GPT point	798.4	703.3	97.8	101.4	0.9642	103.7%
Average Correction Factor					0.9889	101.2%

Notes: Changed inlet filter after as founds. Adjusted high point span. Low point needs further investigating from senior technician.

Calibration Performed By: Matthew Courtoreille



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 20, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	8:24	End Time (MST):	13:13
Analyzer make:	API T200	Analyzer serial #:	4460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	1.6	----	Correlation Coefficient	0.999990	≥0.995
803.1	804.7	0.9980	Slope	1.000160	0.90 - 1.10
400.9	404.8	0.9904	Intercept	2.195593	+/-20
201.1	203.0	0.9905			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 20, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	8:24	End Time (MST):	13:13
Analyzer make:	API T200	Analyzer serial #:	4460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999890	≥0.995
411.7	412.0	0.9992	Slope	0.997333	0.90 - 1.10
203.2	202.5	1.0033	Intercept	1.324008	+/-20
97.8	101.4	0.9642			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 20, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	8:24	End Time (MST):	13:13
Analyzer make:	API T200	Analyzer serial #:	4460

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	1.5	----	Correlation Coefficient	0.999995	≥0.995
800.4	796.7	1.0047	Slope	0.992944	0.90 - 1.10
399.6	399.5	1.0002	Intercept	2.314346	+/-20
200.4	202.1	0.9916			



NO_x Calibration Plot

Date: April 18, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS 08
Calibration Date:	April 16, 2024	Last Cal Date:	March 19, 2024
Start time (MST):	7:02	End time (MST):	9:44
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3252
Calibrator Make/Model:	Teledyne API T700	Serial Number:	135
ZAG Make/Model:	Teledyne API T701		

Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #:	3872
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994571	0.999086	Backgd or Offset:	-2.0	-1.9
Calibration intercept:	0.100000	0.160000	Coeff or Slope:	1.036	1.019

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	5000	913.0	400.0	409.3	0.978
As found Mid point					
As found Low point					
Baseline Corr As found:	409.1	Previous response	397.9	*% change	2.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.7	----
High point	5000	914.7	400.0	400.1	1.000
Mid point	5000	786.4	200.0	199.6	1.002
Low point	5000	701.3	100.0	99.6	1.004
As left zero	5000	0.0	0.0	0.7	----
As left span	5000	963.3	400.0	401.2	0.997
Average Correction Factor					1.002

Notes: Changed out inlet filter after as found. Adjustments made to span.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 16, 2024	Previous Calibration:	March 19, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	7:02	End Time (MST):	9:44
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.7	----	Correlation Coefficient	0.999991	≥0.995
400.0	400.1	0.9998	Slope	0.999086	0.90 - 1.10
200.0	199.6	1.0020	Intercept	0.160000	+/- 5
100.0	99.6	1.0040			



O₃ Calibration Plot

Date: April 16, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
 Calibration Date: April 29, 2024 Last Cal Date: March 20, 2024
 Start time (MST): 14:28 End time (MST): 15:05

Analyzer Make: Teledyne API T640 S/N: 319
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	8.60	7.88	8.60	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	734.20	734.9	734.20	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.04	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39%			<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	4.00	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: 10-Jun-24
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.90	0.00	10.80	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 20, 2024
 Date Disposable Filter Changed: February 20, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 25, 2023
 Date RH/T Sensor Cleaned: July 25, 2023

Notes: No adjustments needed.

Calibration by: Morgan Voyageur



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS 08
Calibration Date:	April 29, 2024	Last Cal Date:	March 19, 2024
Start time (MST):	12:20	End time (MST):	15:09
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #:	ALM014846		
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 3252
ZAG Make/Model:	Teledyne API T701H		Serial Number: 135

Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997054	1.003563	Backgd or Offset:	-0.015	-0.015
Calibration intercept:	0.100939	0.210880	Coeff or Slope:	1.007	1.007

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4934	66.7	40.4	40.7	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	40.63	Prev response:	40.40	*% change:	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4934	66.7	40.4	40.7	0.992
Mid point	4966.7	33.3	20.2	20.5	0.983
Low point	4983.3	16.7	10.1	10.4	0.970
As left zero	5000	0.0	0.0	0.1	----
As left span	2960	40.0	40.4	40.5	0.999
Average Correction Factor					0.982

Notes: Changed inlet filter after as found, no adjustments made.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association


CO Calibration Summary

Station Information

Calibration Date:	April 29, 2024	Previous Calibration:	March 19, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	12:20	End Time (MST):	15:09
Analyzer make:	Teledyne API T300	Analyzer serial #:	3505

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999981	≥0.995
40.4	40.7	0.9925	Slope	1.003563	0.90 - 1.10
20.2	20.5	0.9829	Intercept	0.210880	+/-1.5
10.1	10.4	0.9703			



CO Calibration Plot

Date: April 29, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

CO₂ Calibration Report

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS 08
Calibration Date:	April 16, 2024	Last Cal Date:	March 19, 2024
Start time (MST):	9:50	End time (MST):	13:43
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #:	ALM014846		
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 3252
N2 Gen Make/Model:	Peak Scientific		Serial Number: 135

Analyzer Information

Analyzer make:	Teledyne API T360	Analyzer serial #:	289
Analyzer Range	0 - 2,000 ppm		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011498	1.003378	Backgd or Offset:	-0.063
Calibration intercept:	-16.020000	-4.540000	Coeff or Slope:	1.094
				-0.063
				1.087

CO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	3000	0.0	0.0	-4.9	----
As found High Point	2920	80.0	1605.9	1646.4	0.972
As found Mid Point					
As found Low Point					
New cylinder response					
Baseline Corr As found:	1651.3	Prev response:	1608.3	*% change:	2.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	3000	0.0	0.0	-4.9	----
High point	2920	80.0	1605.9	1606.8	0.999
Mid point	2960	40.0	802.9	800.6	1.003
Low point	2980	20.0	401.5	399.1	1.006
As left zero	3000	0.0	0.0	-4.8	----
As left span	2960	40.0	802.9	768.0	1.045
Average Correction Factor					1.003

Notes: Changed inlet filter after as found, Adjustments span, mid point and low point.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association


CO₂ Calibration Summary

Station Information

Calibration Date	April 16, 2024	Previous Calibration	March 19, 2024
Station Name	Fort Chipewyan	Station Number	AMS 08
Start Time (MST)	9:50	End Time (MST)	13:43
Analyzer make	Teledyne API T360	Analyzer serial #	289

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	-4.9	----	Correlation Coefficient	≥0.995
1605.9	1606.8	0.9994	Slope	0.90 - 1.10
802.9	800.6	1.0029	Intercept	+/-20
401.5	399.1	1.0059		



CO₂ Calibration Plot

Date: April 16, 2024

Location: Fort Chipewyan





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Barge Landing	Station number:	AMS 09
Calibration Date:	April 2, 2024	Last Cal Date:	March 6, 2024
Start time (MST):	9:18	End time (MST):	12:44
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.96	ppm	Cal Gas Exp Date: January 5, 2025
Cal Gas Cylinder #:	CC151285		
Removed Cal Gas Conc:	49.96	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	API T700		Serial Number: 3812
Zero Air Gen Model:	APIT701		Serial Number: 4888

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1118148498
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001110	0.997204	Backgd or Offset:	10.2	10.1
Calibration intercept:	0.391536	0.710524	Coeff or Slope:	0.963	0.956

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4919	80.2	801.5	803.1	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	802.9	Previous response	802.8	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4919	80.2	801.5	799.7	1.002
Mid point	4959	40.1	400.8	400.8	1.000
Low point	4980	20.0	199.8	200.1	0.999
As left zero	5000	0.0	0.0	0.4	----
As left span	4919	80.2	801.5	802.3	0.999
Average Correction Factor:					1.000

Notes: Changed sample inlet filter after as founds. Span adjusted.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 6, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:18	End Time (MST):	12:44
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.4	----	Correlation Coefficient	0.999999	≥0.995
801.5	799.7	1.0022	Slope	0.997204	0.90 - 1.10
400.8	400.8	0.9999	Intercept	0.710524	+/-30
199.8	200.1	0.9987			



SO2 Calibration Plot

Date: April 2, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Barge Landing	Station number: AMS 09
Calibration Date: April 3, 2024	Last Cal Date: March 7, 2024
Start time (MST): 8:53	End time (MST): 13:05
Reason: Routine	

Calibration Standards

Cal Gas Concentration: 5.17 ppm	Cal Gas Exp Date: August 22, 2026
Cal Gas Cylinder #: CC511415	
Removed Cal Gas Conc: 5.17 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 3812
ZAG Make/Model: API T701	Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i-TLE	Analyzer serial #: 1331259320
Converter make: CDN-101	Converter serial #: 519
Analyzer Range: 0 - 100 ppb	Converter Temp: 830 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000835	0.990985	Backgd or Offset:	2.840	2.830
Calibration intercept:	0.139405	0.059198	Coeff or Slope:	1.170	1.170

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.0	0.0	----
As found High point	4923	77.4	80.0	79.1	1.012
As found Mid point	4961	38.7	40.0	39.7	1.008
As found Low point	4981	19.3	20.0	19.7	1.013
New cylinder response					
Baseline Corr As found:	79.1	Prev response:	80.25	*% change:	-1.4%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.988559	AF Intercept:	0.019148
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/(Ic-AFzero))
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4923	77.4	80.0	79.3	1.009
Mid point	4961	38.7	40.0	39.9	1.003
Low point	4981	19.3	20.0	19.8	1.008
As left zero	5000	0.0	0.0	0.2	----
As left span	4923	77.4	80.0	79.4	1.008
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:					Ave Corr Factor
Date of last converter efficiency test:					1.007

Notes: Changed inlet filter after as founds. Pump was changed due to low flow. No adjustments made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 7, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	8:53	End Time (MST):	13:05
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999988	≥ 0.995
80.0	79.3	1.0093	Slope	0.990985	0.90 - 1.10
40.0	39.9	1.0032	Intercept	0.059198	+/-3
20.0	19.8	1.0080			



TRS Calibration Plot

Date: April 3, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Barge Landing	Station number:	AMS 09
Calibration Date:	April 2, 2024	Last Cal Date:	March 6, 2024
Start time (MST):	9:18	End time (MST):	12:44
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
Zero Air Gen model:	APIT701	Serial Number:	4888

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1331259521
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.58E-04	2.60E-04	NMHC SP Ratio:	4.37E-05	4.50E-05
CH4 Retention time:	15.00	15.20	NMHC Peak Area:	209298	205331
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	17.12	16.90	1.013
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.90	Prev response	17.04	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	17.12	17.14	0.999
Mid point	4960	40.1	8.56	8.55	1.000
Low point	4980	20.0	4.27	4.28	0.997
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	17.12	17.12	1.000
Average Correction Factor					0.999

Notes: Changed inlet filter after as founds. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	9.14	9.00	1.015
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.00	Prev response	9.08	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	9.14	9.16	0.998
Mid point	4960	40.1	4.57	4.57	0.999
Low point	4980	20.0	2.28	2.29	0.996
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	9.14	9.15	0.999
Average Correction Factor					0.998

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	7.98	7.90	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.90	Prev response	7.96	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	7.98	7.98	1.000
Mid point	4960	40.1	3.99	3.98	1.002
Low point	4980	20.0	1.99	1.99	0.998
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	7.98	7.98	1.001
Average Correction Factor					1.000

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.995217	1.000957
THC Cal Offset:	0.001033	0.000059
CH ₄ Cal Slope:	0.997716	0.999506
CH ₄ Cal Offset:	-0.000741	0.000263
NMHC Cal Slope:	0.993146	1.001926
NMHC Cal Offset:	0.001575	0.000995

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 6, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:18	End Time (MST):	12:44
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥0.995
17.12	17.14	0.9988	Slope	1.000957	0.90 - 1.10
8.56	8.55	1.0005	Intercept	0.000059	+/-0.5
4.27	4.28	0.9971			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 6, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:18	End Time (MST):	12:44
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥0.995
7.98	7.98	1.0002	Slope	0.999506	0.90 - 1.10
3.99	3.98	1.0019	Intercept	0.000263	+/-0.5
1.99	1.99	0.9982			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 6, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:18	End Time (MST):	12:44
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999999	<i>≥0.995</i>
9.14	9.16	0.9979	Slope	1.001926	<i>0.90 - 1.10</i>
4.57	4.57	0.9986	Intercept	0.000995	<i>+/-0.5</i>
2.28	2.29	0.9961			



NMHC Calibration Plot

Date: April 2, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Barge Landing
 Station number: AMS 09
 Calibration Date: April 18, 2024
 Last Cal Date: March 22, 2024
 Start time (MST): 10:38
 End time (MST): 15:03
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1KDH
 NOX Cal Gas Conc: 47.38 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 47.38 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: Api T701
 Cal Gas Expiry Date: November 17, 2026
 NO Cal Gas Conc: 46.94 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 46.94 ppm
 NO gas Diff:
 Serial Number: 3812
 Serial Number: 4888

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
AF High point	4915	85.3	808.3	800.7	7.5	805.3	793.7	11.6	1.0032	1.0088
AF Mid point										
AF Low point										
New cyl resp										

Previous Response NO_x = 808.7 ppb NO = 799.8 ppb * = > +/-5% change initiates investigation *Percent Change NO_x = -0.4%
 Baseline Corr 1st pt NO_x = 805.7 ppb NO = 793.8 ppb As Found Statistics *Percent Change NO = -0.8%
 Baseline Corr 2nd pt NO_x = NA ppb NO = NA ppb As found NO_x r²: Nx SI: Nx Int:
 Baseline Corr 3rd pt NO_x = NA ppb NO = NA ppb As found NO r²: NO SI: NO Int:
 As found NO₂ r²: NO₂ SI: NO₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1426262593

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000294	0.999233
NO _x Cal Offset:	0.198498	0.498364
NO Cal Slope:	0.999838	0.998482
NO Cal Offset:	-0.783809	-0.583944
NO ₂ Cal Slope:	0.999826	1.006238
NO ₂ Cal Offset:	-1.024158	0.877238

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.094	1.101	NO bkgnd or offset:	10.0	10.1
NOX coeff or slope:	0.998	0.998	NOX bkgnd or offset:	10.3	10.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	181.0	182.2

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
High point	4915	85.3	808.3	800.7	7.5	807.8	799.2	8.6	1.0006	1.0019
Mid point	4957	42.6	403.7	400.0	3.7	404.2	398.6	5.6	0.9988	1.0034
Low point	4979	21.3	201.8	200.0	1.9	202.9	198.4	4.4	0.9947	1.0078
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
As left span	4915	85.3	808.3	417.7	390.6	806.6	417.7	388.8	1.0021	1.0000
Average Correction Factor									0.9980	1.0044

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	795.6	414.7	388.4	391.0	0.9934	100.7%
Mid GPT point	795.6	607.1	196.0	199.1	0.9845	101.6%
Low GPT point	795.6	702.5	100.6	102.9	0.9777	102.3%
Average Correction Factor					0.9852	101.5%

Notes:

Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 22, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:38	End Time (MST):	15:03
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999997	≥0.995
808.3	807.8	1.0006	Slope	0.999233	0.90 - 1.10
403.7	404.2	0.9988	Intercept	0.498364	+/-20
201.8	202.9	0.9947			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 22, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:38	End Time (MST):	15:03
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999961	≥0.995
388.4	391.0	0.9934	Slope	1.006238	0.90 - 1.10
196.0	199.1	0.9845	Intercept	0.877238	+/-20
100.6	102.9	0.9777			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 22, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:38	End Time (MST):	15:03
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999997	≥0.995
800.7	799.2	1.0019	Slope	0.998482	0.90 - 1.10
400.0	398.6	1.0034	Intercept	-0.583944	+/-20
200.0	198.4	1.0078			



NO_x Calibration Plot

Date: April 18, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Barge Landing Station number: AMS 09
 Calibration Date: April 13, 2024 Last Cal Date: March 22, 2024
 Start time (MST): 10:20 End time (MST): 11:15

Analyzer Make: API T640 S/N: 844
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.00	9.40	10.00	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	719.40	727.38	719.40	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.95	4.96	4.95	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	36.00	----	36.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	5.80	PM w/ HEPA: _____	6.70	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	5.00	11.00	11.00	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 13, 2024
 Date Disposable Filter Changed: April 13, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 23, 2023
 Date RH/T Sensor Cleaned: August 23, 2023

Notes: Inlet head looks good. No adjustments. Leak check passed.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS11 LOWER CAMP APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Lower Camp	Station number:	AMS 11
Calibration Date:	April 24, 2024	Last Cal Date:	March 26, 2024
Start time (MST):	10:04	End time (MST):	14:09
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.25	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC2216			
Removed Cal Gas Conc:	49.25	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	3807
Zero Air Gen Model:	Teledyne API T701		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	100841398
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994478	0.988759	Backgd or Offset:	14.7	14.7
Calibration intercept:	0.097406	-0.098992	Coeff or Slope:	1.034	1.034

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	1.3	----
As found High point	4932	81.4	799.6	788.8	1.015
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	787.5	Previous response	795.3	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	1.2	----
High point	4932	81.4	799.6	791.3	1.011
Mid point	4959	40.7	400.9	395.4	1.014
Low point	4981	20.4	200.9	197.4	1.018
As left zero	5000	0.0	0.0	1.4	----
As left span	4932	81.4	799.6	792.4	1.009
Average Correction Factor:					1.014

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 24, 2024	Previous Calibration:	March 26, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:04	End Time (MST):	14:09
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	1.2	----	Correlation Coefficient	0.999987	≥0.995
799.6	791.3	1.0105	Slope	0.988759	0.90 - 1.10
400.9	395.4	1.0140	Intercept	-0.098992	+/-30
200.9	197.4	1.0176			



SO2 Calibration Plot

Date: April 24, 2024

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Lower Camp	Station number: AMS 11
Calibration Date: April 23, 2024	Last Cal Date: March 25, 2024
Start time (MST): 10:14	End time (MST): 14:37
Reason: Routine	

Calibration Standards

Cal Gas Concentration: 5.43 ppm	Cal Gas Exp Date: January 4, 2025
Cal Gas Cylinder #: CC501097	
Removed Cal Gas Conc: 5.43 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 3807
ZAG Make/Model: API T701H	Serial Number: 196

Analyzer Information

Analyzer make: Thermo 43iQ	Analyzer serial #: 1203169745
Converter make: Global G150	Converter serial #: 2022-223
Analyzer Range: 0 - 100 ppb	Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.012278	0.990021	Backgd or Offset:	3.0	2.5
Calibration intercept:	-0.364532	0.156017	Coeff or Slope:	0.817	0.817

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.6	----
As found High point	4926	73.6	79.9	79.3	1.000
As found Mid point	4963	36.8	40.0	39.5	0.996
As found Low point	4982	18.6	20.2	19.5	1.005
New cylinder response					
Baseline Corr As found:	79.9	Prev response:	80.54	*% change:	-0.8%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	1.000323	AF Intercept:	-0.604758
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999992	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4926	73.6	79.9	79.2	1.009
Mid point	4963	36.8	40.0	39.9	1.001
Low point	4982	18.6	20.2	20.1	1.005
As left zero	5000	0.0	0.0	0.1	----
As left span	4926	73.6	79.9	79.9	1.000
SO2 Scrubber Check	4935	81.5	812.3	-0.1	----
Date of last scrubber change:				Ave Corr Factor	1.005
Date of last converter efficiency test:					

Notes: Changed sample inlet filter after as founds. Adjusted zero only. Ran scrubber check after calibrator zero and it passed.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 23, 2024	Previous Calibration:	March 25, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:14	End Time (MST):	14:37
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999987	≥0.995
79.9	79.2	1.0091	Slope	0.990021	0.90 - 1.10
40.0	39.9	1.0015	Intercept	0.156017	+/-3
20.2	20.1	1.0046			



H₂S Calibration Plot

Date: April 23, 2024

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH4 / NMHC Calibration Report

Station Information

Station Name:	Lower Camp	Station number:	AMS 11
Calibration Date:	April 24, 2024	Last Cal Date:	March 26, 2024
Start time (MST):	10:04	End time (MST):	14:09
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	502.0 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	502.0 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
Zero Air Gen model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1505164381
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.68E-04	2.75E-04	NMHC SP Ratio: 4.81E-05	4.90E-05
CH4 Retention time:	14.6	14.8	NMHC Peak Area: 190907	187273
Zero Chromatogram:	OFF	OFF	Flat Baseline: OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	81.4	17.33	16.82	1.030
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.82	Prev response	17.29	*% change	-2.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	81.4	17.33	17.30	1.002
Mid point	4959	40.7	8.69	8.63	1.006
Low point	4981	20.4	4.35	4.31	1.011
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	81.4	17.33	17.44	0.993
Average Correction Factor					1.006

Notes: Changed sample inlet filter and nitrogen cylinder after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	81.4	9.18	8.96	1.025
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.96	Prev response	9.16	*% change	-2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	81.4	9.18	9.15	1.002
Mid point	4959	40.7	4.60	4.57	1.006
Low point	4981	20.4	2.31	2.28	1.009
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	81.4	9.18	9.23	0.994
Average Correction Factor					1.006

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	81.4	8.15	7.86	1.036
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.86	Prev response	8.13	*% change	-3.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	81.4	8.15	8.14	1.001
Mid point	4959	40.7	4.09	4.06	1.007
Low point	4981	20.4	2.05	2.02	1.013
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	81.4	8.15	8.22	0.992
Average Correction Factor					1.007

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999083	0.998938
THC Cal Offset:	-0.021522	-0.023918
CH ₄ Cal Slope:	0.998771	0.999584
CH ₄ Cal Offset:	-0.011567	-0.013970
NMHC Cal Slope:	0.999149	0.997916
NMHC Cal Offset:	-0.009354	-0.009146

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 24, 2024	Previous Calibration:	March 26, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:04	End Time (MST):	14:09
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999991	<i>>0.995</i>
17.33	17.30	1.0016	Slope	0.998938	<i>0.90 - 1.10</i>
8.69	8.63	1.0062	Intercept	-0.023918	<i>+/-0.5</i>
4.35	4.31	1.0108			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 24, 2024	Previous Calibration:	March 26, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:04	End Time (MST):	14:09
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999985	<i>≥0.995</i>
8.15	8.14	1.0011	Slope	0.999584	<i>0.90 - 1.10</i>
4.09	4.06	1.0068	Intercept	-0.013970	<i>+/-0.5</i>
2.05	2.02	1.0127			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 24, 2024	Previous Calibration:	March 26, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:04	End Time (MST):	14:09
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999995	<i>>0.995</i>
9.18	9.15	1.0025	Slope	0.997916	<i>0.90 - 1.10</i>
4.60	4.57	1.0058	Intercept	-0.009146	<i>+/-0.5</i>
2.31	2.28	1.0092			



NMHC Calibration Plot

Date: April 24, 2024

Location: Lower Camp





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13 FORT MCKAY SOUTH APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	April 4, 2024	Last Cal Date:	March 12, 2024
Start time (MST):	8:47	End time (MST):	14:27
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.55	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC260812			
Removed Cal Gas Conc:	50.55	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	2448
Zero Air Gen Model:	API T701		Serial Number:	1117

Analyzer Information

Analyzer make:	API T100	Serial Number:	599
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002127	1.004442	Backgd or Offset:	90.0	90.0
Calibration intercept:	-2.338170	-2.497906	Coeff or Slope:	0.711	0.711

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4921	79.1	799.7	800.9	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	800.8	Previous response	799.0	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4921	79.1	799.7	801.8	0.997
Mid point	4961	39.5	399.3	398.0	1.003
Low point	4980	19.8	200.2	195.5	1.024
As left zero	5000	0.0	0.0	0.3	----
As left span	4921	79.1	799.7	801.0	0.998
Average Correction Factor:					1.008

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 12, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:47	End Time (MST):	14:27
Analyzer make:	API T100	Analyzer serial #:	599

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999950	≥0.995
799.7	801.8	0.9974	Slope	1.004442	0.90 - 1.10
399.3	398.0	1.0033	Intercept	-2.497906	+/-30
200.2	195.5	1.0240			



SO2 Calibration Plot

Date: April 4, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort McKay South	Station number: AMS 13
Calibration Date: April 10, 2024	Last Cal Date: March 21, 2024
Start time (MST): 9:04	End time (MST): 13:13
Reason: Routine	

Calibration Standards

Cal Gas Concentration: 5.34 ppm	Cal Gas Exp Date: January 1, 2025
Cal Gas Cylinder #: CC500241	
Removed Cal Gas Conc: 5.34 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 2448
ZAG Make/Model: API T701	Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i TLE	Analyzer serial #: 1180540017
Converter make: CDN-101	Converter serial #: 521
Analyzer Range: 0 - 100 ppb	Converter Temp: degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007879	1.010145	Backgd or Offset:	4.2	4.1
Calibration intercept:	-0.422268	-0.602184	Coeff or Slope:	1.157	1.143

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.5	----
As found High point	4925	75.5	80.6	82.1	0.976
As found Mid point	4962	37.7	40.3	40.4	0.984
As found Low point	4981	18.9	20.2	19.6	1.004
New cylinder response					
Baseline Corr As found:	82.6	Prev response:	80.84	*% change:	2.1%
Baseline Corr 2nd AF pt:	40.9	AF Slope:	1.026451	AF Intercept:	-0.802285
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999939	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.4	----
High point	4925	75.5	80.6	80.9	0.997
Mid point	4962	37.7	40.3	40.1	1.004
Low point	4981	18.9	20.2	19.5	1.035
As left zero	5000	0.0	0.0	-0.2	----
As left span	4925	75.5	80.6	80.5	1.002
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:	20-Jan-20			Ave Corr Factor	1.012
Date of last converter efficiency test:					

Notes: Changed inlet filter after as founds. Scrubber checked after calibrator zero. Span adjusted only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	April 10, 2024	Previous Calibration:	March 21, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:04	End Time (MST):	13:13
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	0.999965	≥ 0.995
80.6	80.9	0.9966	Slope	1.010145	$0.90 - 1.10$
40.3	40.1	1.0041	Intercept	-0.602184	± 3
20.2	19.5	1.0352			



TRS Calibration Plot

Date: April 10, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	April 4, 2024	Last Cal Date:	March 12, 2024
Start time (MST):	8:47	End time (MST):	14:27
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC2608112	Cal Gas Expiry Date:	
CH4 Cal Gas Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
C3H8 Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
Removed C3H8 Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
Zero Air Gen model:	API T701	Serial Number:	1117

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1172750023
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.95E-04	3.05E-04	NMHC SP Ratio:	5.07E-05	5.21E-05
CH4 Retention time:	15.20	15.40	NMHC Peak Area:	179256	174169
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF	ON

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.02	----
As found High point	4921	79.1	17.05	16.61	1.028
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.59	Prev response	16.99	*% change	-2.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	17.05	17.07	0.999
Mid point	4961	39.5	8.51	8.52	0.999
Low point	4980	19.8	4.27	4.19	1.019
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	17.05	17.14	0.994
Average Correction Factor					1.006

Notes: Changed inlet filter after as founds. Do zero chromatogram and adjusted span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	9.08	8.82	1.029
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.82	Prev response	9.10	*% change	-3.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	9.08	9.11	0.997
Mid point	4961	39.5	4.53	4.54	0.998
Low point	4980	19.8	2.27	2.26	1.005
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	9.08	9.19	0.988
Average Correction Factor					1.000

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.02	----
As found High point	4921	79.1	7.97	7.96	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.95	Prev response	7.89	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	7.97	7.96	1.001
Mid point	4961	39.5	3.98	3.98	1.001
Low point	4980	19.8	1.99	1.93	1.035
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	7.97	7.95	1.002
Average Correction Factor					1.012

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001286	1.002927
THC Cal Offset:	-0.078556	-0.034543
CH ₄ Cal Slope:	0.994805	1.002287
CH ₄ Cal Offset:	-0.034188	-0.026764
NMHC Cal Slope:	1.007213	1.003450
NMHC Cal Offset:	-0.045569	-0.007378

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 12, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:47	End Time (MST):	14:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999971	<i>>0.995</i>
17.05	17.07	0.9987	Slope	1.002927	<i>0.90 - 1.10</i>
8.51	8.52	0.9994	Intercept	-0.034543	<i>+/-0.5</i>
4.27	4.19	1.0187			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 12, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:47	End Time (MST):	14:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999911	<i>≥0.995</i>
7.97	7.96	1.0006	Slope	1.002287	<i>0.90 - 1.10</i>
3.98	3.98	1.0005	Intercept	-0.026764	<i>+/-0.5</i>
1.99	1.93	1.0355			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 12, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:47	End Time (MST):	14:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999996	<i>≥0.995</i>
9.08	9.11	0.9972	Slope	1.003450	<i>0.90 - 1.10</i>
4.53	4.54	0.9977	Intercept	-0.007378	<i>+/-0.5</i>
2.27	2.26	1.0048			



NMHC Calibration Plot

Date: April 4, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort McKay South
 Station number: AMS 13
 Calibration Date: April 16, 2024
 Last Cal Date: March 27, 2024
 Start time (MST): 8:38
 End time (MST): 13:10
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2UP1RP
 NOX Cal Gas Conc: 48.25 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.25 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: APIT701
 Cal Gas Expiry Date: November 17, 2026
 NO Cal Gas Conc: 47.88 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 47.88 ppm
 NO gas Diff:
 Serial Number: 2448
 Serial Number: 1117

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
AF High point	4917	83.5	805.7	799.5	6.2	797.9	792.3	5.5	1.0098	1.0091
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 799.0 ppb	NO = 794.7 ppb	<i>* = > +/-5% change initiates investigation</i>				*Percent Change	NO _x = -0.1%		
Baseline Corr 1st pt	NO _x = 797.9 ppb	NO = 792.3 ppb	<u>As Found Statistics</u>				*Percent Change	NO = -0.3%		
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :		Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :		NO SI:	NO Int:			
			As found	NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1410661329

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.994114	1.000782
NO _x Cal Offset:	-1.991485	-1.593202
NO Cal Slope:	0.997692	1.003482
NO Cal Offset:	-2.930061	-2.631535
NO ₂ Cal Slope:	0.999151	1.001146
NO ₂ Cal Offset:	-1.075362	-0.447686

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.128	1.140	NO bkgnd or offset:	10.1	10.2
NOX coeff or slope:	0.999	1.002	NOX bkgnd or offset:	10.2	10.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.3	163.5

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.0	----	----
High point	4917	83.5	805.7	799.5	6.2	805.6	801.1	4.5	1.0001	0.9980
Mid point	4958	41.8	403.4	400.3	3.1	401.4	397.7	3.7	1.0049	1.0065
Low point	4979	20.9	201.7	200.1	1.5	198.2	195.2	3.0	1.0176	1.0253
As left zero	5000	0.0	0.0	0.0	0.0	0.3	0.2	0.1	----	----
As left span	4917	83.5	805.7	382.8	422.9	809.6	382.8	426.8	0.9952	1.0000
Average Correction Factor									1.0076	1.0100

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	795.6	381.3	420.5	420.8	0.9992	100.1%
Mid GPT point	795.6	586.7	215.1	214.5	1.0027	99.7%
Low GPT point	795.6	690.8	111.0	110.3	1.0062	99.4%
Average Correction Factor					1.0027	99.7%

Notes: Changed inlet filter after as founds. Adjusted span only. Used 2nd NO reference point due to drift.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 16, 2024	Previous Calibration:	March 27, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:38	End Time (MST):	13:10
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999975	≥0.995
805.7	805.6	1.0001	Slope	1.000782	0.90 - 1.10
403.4	401.4	1.0049	Intercept	-1.593202	+/-20
201.7	198.2	1.0176			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 16, 2024	Previous Calibration:	March 27, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:38	End Time (MST):	13:10
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999994	≥0.995
420.5	420.8	0.9992	Slope	1.001146	0.90 - 1.10
215.1	214.5	1.0027	Intercept	-0.447686	+/-20
111.0	110.3	1.0062			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 16, 2024	Previous Calibration:	March 27, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:38	End Time (MST):	13:10
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999939	≥0.995
799.5	801.1	0.9980	Slope	1.003482	0.90 - 1.10
400.3	397.7	1.0065	Intercept	-2.631535	+/-20
200.1	195.2	1.0253			



NO_x Calibration Plot

Date: April 16, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	April 15, 2024	Last Cal Date:	April 5, 2024
Start time (MST):	9:37	End time (MST):	13:08
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	2448
Calibrator Make/Model:	Teledyne API T700	Serial Number:	1117
ZAG Make/Model:	Teledyne API T701		

Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #:	3871
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999371	1.001029	Backgd or Offset:	4.0	2.7
Calibration intercept:	0.560000	1.320000	Coeff or Slope:	0.984	0.973

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) Limit = 0.90-1.10
As found zero	5000	0.0	0.0	-1.2	----
As found High point	5000	994.5	400.0	405.2	0.984
As found Mid point					
As found Low point					
Baseline Corr As found:	406.4	Previous response	400.3	*% change	1.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.0	0.7	----
High point	5000	995.4	400.0	401.3	0.997
Mid point	5000	849.9	200.0	202.2	0.989
Low point	5000	745.1	100.0	101.8	0.982
As left zero	5000	0.0	0.0	0.4	----
As left span	5000	995.4	400.0	401.7	0.996
Average Correction Factor					0.989

Notes: Changed inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 15, 2024	Previous Calibration:	April 5, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:37	End Time (MST):	13:08
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.7	----	Correlation Coefficient	0.999987	≥0.995
400.0	401.3	0.9968	Slope	1.001029	0.90 - 1.10
200.0	202.2	0.9891	Intercept	1.320000	+/- 5
100.0	101.8	0.9823			



O₃ Calibration Plot

Date: April 15, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay South Station number: AMS 13
 Calibration Date: April 16, 2024 Last Cal Date: March 27, 2024
 Start time (MST): 10:04 End time (MST): 11:59

Analyzer Make: API T640 S/N: 1335
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388746
 Temp/RH standard: Alicat FP-25BT S/N: 388746

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-0.80	-0.60	-0.80	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	738.80	740.50	738.80	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	4.95	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	46.00	----	46.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	1.50	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	7.00	10.00	10.90	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ April 16, 2024
 Date Disposable Filter Changed: _____ April 16, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ June 29, 2023
 Date RH/T Sensor Cleaned: _____ June 29, 2023

Notes: Inlet head cleaned. PMT peak test adjusted. Leak check passed.

Calibration by: Sean Bala



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay South Station number: AMS 13
 Calibration Date: April 19, 2024 Last Cal Date: April 16, 2024
 Start time (MST): 9:02 End time (MST): 9:45

Analyzer Make: API T640 S/N: 1335
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388746
 Temp/RH standard: Alicat FP-25BT S/N: 388746

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2.50	2.47	2.50	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	749.60	752.03	749.60	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.94	4.96	4.94	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	45.00	----	45.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	9.20	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.00	11.00	11.00	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 19, 2024
 Date Disposable Filter Changed: April 19, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: June 29, 2023
 Date RH/T Sensor Cleaned: June 29, 2023

Notes: Elevated readings since yesterday. Diagnostics and rest of the readings are normal. Completed annual maintenance again. Readings were closer to baseline after cycling power to the instrument.

Calibration by: Max Farrell



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC
APRIL 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	April 2, 2024	Last Cal Date:	March 7, 2024
Start time (MST):	9:54	End time (MST):	13:38
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.95	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC279389			
Removed Cal Gas Conc:	49.95	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	3060
Zero Air Gen Model:	API T701H		Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	0710321322
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005467	0.999403	Backgd or Offset:	26.4	26.5
Calibration intercept:	-1.163959	-1.240981	Coeff or Slope:	0.836	0.836

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.5	----
As found High point	4938	80.3	799.3	802.7	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	802.2	Previous response	802.5	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	1.0	----
High point	4938	80.3	799.3	798.8	1.001
Mid point	4979	40.2	400.1	397.1	1.007
Low point	4998	20.2	201.1	197.7	1.017
As left zero	5000	0.0	0.0	0.8	----
As left span	4938	80.3	799.3	795.4	1.005
Average Correction Factor:					1.008

Notes: Sample inlet filter changed after as founds. No adjustments made. High point was noisier than usual; maintenance will be performed during the next visit.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 7, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:54	End Time (MST):	13:38
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	1.0	----	Correlation Coefficient	0.999963	≥0.995
799.3	798.8	1.0006	Slope	0.999403	0.90 - 1.10
400.1	397.1	1.0075	Intercept	-1.240981	+/-30
201.1	197.7	1.0170			



SO2 Calibration Plot

Date: April 2, 2024

Location: Anzac





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	April 16, 2024	Last Cal Date:	April 2, 2024
Start time (MST):	9:03	End time (MST):	10:25
Reason:	As Found		

Calibration Standards

Cal Gas Concentration:	49.95	ppm	Cal Gas Exp Date: January 5, 2025
Cal Gas Cylinder #:	CC279389		
Removed Cal Gas Conc:	49.95	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	API T700		Serial Number: 3060
Zero Air Gen Model:	API T701H		Serial Number: 357

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	0710321322
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999403	NA	Backgd or Offset:	26.5	NA
Calibration intercept:	-1.240981	NA	Coeff or Slope:	0.836	NA

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.5	----
As found High point	4938	80.3	799.3	772.1	1.036
As found Mid point	4979	40.2	400.1	385.9	1.038
As found Low point	4998	20.2	201.1	187.8	1.073
New cylinder response					
Baseline Corr As found:	771.6	Previous response	797.6	*% change	-3.4%
Baseline Corr 2nd AF pt:	385.4	AF Slope:	0.968168	AF Intercept:	-2.380666
Baseline Corr 3rd AF pt:	187.3	AF Correlation:	0.999909	* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor:

Notes: Before performing maintenance on the instrument, as-found were conducted. Investigation was made. No alarms were noticed on the instrument, diagnostics were similar to last calibration, and there were no issues to report with the setup. Suspecting faulty lamp/socket.

Calibration Performed By: Mohammed Kashif

SO2 Calibration Plot

Date: April 16, 2024

Location: Anzac





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Anzac	Station number: AMS 14
Calibration Date:	April 16, 2024	Last Cal Date: April 2, 2024
Start time (MST):	14:15	End time (MST): 17:10
Reason:	Maintenance	

Calibration Standards

Cal Gas Concentration:	49.95 ppm	Cal Gas Exp Date: January 5, 2025
Cal Gas Cylinder #:	CC279389	
Removed Cal Gas Conc:	49.95 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA	Diff between cyl:
Calibrator Model:	API T700	Serial Number: 3060
Zero Air Gen Model:	API T701H	Serial Number: 357

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 0710321322
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.009311	Backgd or Offset:	NA
Calibration intercept:	NA	-4.184570	Coeff or Slope:	NA
				24.6
				1.043

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.4	----
High point	4938	80.3	799.3	804.6	0.993
Mid point	4979	40.2	400.1	397.2	1.007
Low point	4998	20.2	201.1	195.3	1.030
As left zero	5000	0.0	0.0	-0.6	----
As left span	4938	80.3	799.3	806.5	0.991
Average Correction Factor:					1.010

Notes: In response to the noisy readings during the last calibration, the lamp and socket were replaced. Following this, a PMT and lamp voltages adjustment were carried out, resulting in changes to both flash and PMT voltages values, which are now within limits. A stable and less noisy response was observed post maintenance. Adjusted both zero and span.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 16, 2024	Previous Calibration:	April 2, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	14:15	End Time (MST):	17:10
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.4	----	Correlation Coefficient	0.999898	≥0.995
799.3	804.6	0.9934	Slope	1.009311	0.90 - 1.10
400.1	397.2	1.0072	Intercept	-4.184570	+/-30
201.1	195.3	1.0295			



SO2 Calibration Plot

Date: April 16, 2024

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	April 11, 2024	Last Cal Date:	March 27, 2024
Start time (MST):	9:47	End time (MST):	14:19
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.15	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC510379			
Removed Cal Gas Conc:	5.15	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API 701H		Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582
Converter make:	CD Nova CDN-101	Converter serial #:	503
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000445	1.015603	Backgd or Offset:	2.3	2.3
Calibration intercept:	-0.045344	-0.225550	Coeff or Slope:	0.984	0.984

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4938	77.9	80.0	81.0	0.986
As found Mid point	4973	38.9	40.0	40.5	0.985
As found Low point	4997	19.5	20.0	20.1	0.991
New cylinder response					
Baseline Corr As found:	81.1	Prev response:	79.97	*% change:	1.4%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.014491	AF Intercept:	-0.125462
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999996	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4938	77.9	80.0	81.1	0.986
Mid point	4973	38.9	40.0	40.1	0.996
Low point	4997	19.5	20.0	20.1	0.996
As left zero	5000	0.0	0.0	0.1	----
As left span	4938	77.9	80.0	79.3	1.008
SO2 Scrubber Check	4936	80.3	800.4	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.993
Date of last converter efficiency test:					

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 27, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:47	End Time (MST):	14:19
Analyzer make:	CD Nova CDN-101	Analyzer serial #:	503

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999973	≥ 0.995
80.0	81.1	0.9858	Slope	1.015603	$0.90 - 1.10$
40.0	40.1	0.9964	Intercept	-0.225550	± 3
20.0	20.1	0.9956			



TRS Calibration Plot

Date: April 11, 2024

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	April 2, 2024	Last Cal Date:	March 7, 2024
Start time (MST):	9:54	End time (MST):	13:38
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
Zero Air Gen model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1118148494
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.25E-04	2.25E-04	NMHC SP Ratio:	4.11E-05	4.11E-05
CH4 Retention time:	13.30	13.30	NMHC Peak Area:	221451	221451
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4938	80.3	17.10	16.75	1.021
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.75	Prev response	16.97	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4938	80.3	17.10	16.74	1.021
Mid point	4979	40.2	8.56	8.29	1.033
Low point	4998	20.2	4.30	4.11	1.047
As left zero	5000	0.0	0.00	0.00	----
As left span	4938	80.3	17.10	16.75	1.021
Average Correction Factor					1.034

Notes: Sample inlet filter changed after as founds. No adjustments needed.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4938	80.3	9.11	8.89	1.025
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.89	Prev response	9.05	*% change	-1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4938	80.3	9.11	8.89	1.025
Mid point	4979	40.2	4.56	4.41	1.035
Low point	4998	20.2	2.29	2.19	1.045
As left zero	5000	0.0	0.00	0.00	----
As left span	4938	80.3	9.11	8.89	1.025
Average Correction Factor					1.035

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4938	80.3	7.99	7.86	1.017
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.86	Prev response	7.91	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4938	80.3	7.99	7.86	1.017
Mid point	4979	40.2	4.00	3.88	1.031
Low point	4998	20.2	2.01	1.92	1.050
As left zero	5000	0.0	0.00	0.00	----
As left span	4938	80.3	7.99	7.86	1.017
Average Correction Factor					1.033

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.995891	0.980408
THC Cal Offset:	-0.064367	-0.060379
CH ₄ Cal Slope:	0.995346	0.984987
CH ₄ Cal Offset:	-0.040562	-0.035307
NMHC Cal Slope:	0.995954	0.976268
NMHC Cal Offset:	-0.023403	-0.025072

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 7, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:54	End Time (MST):	13:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999938 ≥0.995
17.10	16.74	1.0215	Slope	0.980408 0.90 - 1.10
8.56	8.29	1.0331	Intercept	-0.060379 +/-0.5
4.30	4.11	1.0473		





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 7, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:54	End Time (MST):	13:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999904	<i>≥0.995</i>
7.99	7.86	1.0171	Slope	0.984987	<i>0.90 - 1.10</i>
4.00	3.88	1.0315	Intercept	-0.035307	<i>+/-0.5</i>
2.01	1.92	1.0495			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 7, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:54	End Time (MST):	13:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999961	<i>>0.995</i>
9.11	8.89	1.0255	Slope	0.976268	<i>0.90 - 1.10</i>
4.56	4.41	1.0348	Intercept	-0.025072	<i>+/-0.5</i>
2.29	2.19	1.0454			



NMHC Calibration Plot

Date: April 2, 2024

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	April 30, 2024	Last Cal Date:	April 2, 2024
Start time (MST):	11:12	End time (MST):	13:45
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
Zero Air Gen model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1118148494
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.25E-04	2.25E-04	NMHC SP Ratio:	4.11E-05	4.11E-05
CH4 Retention time:	13.30	13.30	NMHC Peak Area:	221451	221451
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4938	80.3	17.10	16.31	1.048
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.31	Prev response	16.71	*% change	-2.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4938	80.3	17.10	16.33	1.047
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.047

Notes: Hydrogen cylinder changed out.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4938	80.3	9.11	8.56	1.065
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.56	Prev response	8.87	*% change	-3.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4938	80.3	9.11	8.55	1.065
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.065

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4938	80.3	7.99	7.75	1.030
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.75	Prev response	7.83	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4938	80.3	7.99	7.78	1.027
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.027

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.980408	0.954935
THC Cal Offset:	-0.060379	0.000000
CH ₄ Cal Slope:	0.984987	0.973275
CH ₄ Cal Offset:	-0.035307	0.000000
NMHC Cal Slope:	0.976268	0.938637
NMHC Cal Offset:	-0.025072	0.000000

Calibration Performed By: Mohammed Kashif

NMHC Calibration Plot

Date: April 30, 2024

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Anzac
 Station number: AMS 14
 Calibration Date: April 3, 2024
 Last Cal Date: March 14, 2024
 Start time (MST): 10:17
 End time (MST): 15:07
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0037092
 NOX Cal Gas Conc: 60.7 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 60.70 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T700H
 Cal Gas Expiry Date: May 16, 2031
 NO Cal Gas Conc: 60.40 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 60.40 ppm
 NO gas Diff:
 Serial Number: 3060
 Serial Number: 357

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.3	0.0	0.3	----	----
AF High point	4934	66.3	804.8	800.9	4.0	798.1	797.0	1.0	1.0088	1.0048
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 799.1 ppb		NO = 798.7 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -0.2%	
Baseline Corr 1st pt	NO _x = 797.8 ppb		NO = 797.0 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -0.2%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1426262592

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.993607	0.992088
NO _x Cal Offset:	-0.629791	-0.369777
NO Cal Slope:	0.999912	0.998384
NO Cal Offset:	-2.109531	-1.649243
NO ₂ Cal Slope:	0.994068	0.993217
NO ₂ Cal Offset:	-1.527274	-1.136094

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.411	1.411	NO bkgnd or offset:	3.8	3.8
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.8	3.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.0	158.8

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.0	0.3	----	----
High point	4934	66.3	804.8	800.9	4.0	798.3	798.5	-0.2	1.0082	1.0030
Mid point	4985	33.2	401.6	399.6	2.0	398.1	397.2	0.9	1.0088	1.0060
Low point	5004	16.7	201.9	200.9	1.0	199.0	196.8	2.2	1.0146	1.0209
As left zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.3	----	----
As left span	4934	66.3	804.8	403.8	401.0	797.8	403.8	394.0	1.0088	1.0000
Average Correction Factor									1.0105	1.0100

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.3	----	----
High GPT point	795.5	408.1	391.4	388.2	1.0082	99.2%
Mid GPT point	795.5	610.6	188.9	186.0	1.0155	98.5%
Low GPT point	795.5	703.2	96.3	92.9	1.0364	96.5%
Average Correction Factor					1.0200	98.1%

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 14, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:17	End Time (MST):	15:07
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999996	≥0.995
804.8	798.3	1.0082	Slope	0.992088	0.90 - 1.10
401.6	398.1	1.0088	Intercept	-0.369777	+/-20
201.9	199.0	1.0146			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 14, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:17	End Time (MST):	15:07
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999937	≥0.995
391.4	388.2	1.0082	Slope	0.993217	0.90 - 1.10
188.9	186.0	1.0155	Intercept	-1.136094	+/-20
96.3	92.9	1.0364			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 14, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:17	End Time (MST):	15:07
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999978	<i>≥0.995</i>
800.9	798.5	1.0030	Slope	0.998384	<i>0.90 - 1.10</i>
399.6	397.2	1.0060	Intercept	-1.649243	<i>+/-20</i>
200.9	196.8	1.0209			



NO_x Calibration Plot

Date: April 3, 2024

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	April 9, 2024	Last Cal Date:	March 6, 2024
Start time (MST):	9:31	End time (MST):	14:19
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3060
Calibrator Make/Model:	API T700	Serial Number:	357
ZAG Make/Model:	API 701H		

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004629	0.997086	Backgd or Offset:	1.4
Calibration intercept:	0.840000	2.260000	Coeff or Slope:	1.620
				1.594

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.9	----
As found High point	5000	918.8	400.0	406.2	0.987
As found Mid point	5000	803.8	200.0	206.8	0.971
As found Low point	5000	709.8	100.0	104.6	0.964
Baseline Corr As found:	405.3	Previous response	402.7	*% change	0.6%
Baseline Corr 2nd AF pt:	205.9	AF Slope:	1.012143	AF Intercept:	2.500000
Baseline Corr 3rd AF pt:	103.7	AF Correlation:	0.999909	<i>* = > +/-5% change initiates investigation</i>	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.5	----
High point	5000	918.8	400.0	400.1	1.000
Mid point	5000	803.8	200.0	202.9	0.986
Low point	5000	709.8	100.0	103.5	0.966
As left zero	5000	0.0	0.0	1.1	----
As left span	5000	918.8	400.0	401.0	0.998
Average Correction Factor					0.984

Notes: Sample inlet filter changed after as founds. The ozone scrubber was replaced as part of maintenance procedure to address linearity. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 9, 2024	Previous Calibration:	March 6, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:31	End Time (MST):	14:19
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999909	≥0.995
400.0	400.1	0.9998	Slope	0.997086	0.90 - 1.10
200.0	202.9	0.9857	Intercept	2.260000	+/- 5
100.0	103.5	0.9662			



O₃ Calibration Plot

Date: April 9, 2024

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Anzac Station number: AMS 14
 Calibration Date: April 30, 2024 Last Cal Date: March 27, 2024
 Start time (MST): 10:26 End time (MST): 12:42

Analyzer Make: AP T640 S/N: 825
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
 Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2.6	2.1	2.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	711.9	712.89	711.9	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.09	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	38	----	38	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	4.8	PM w/ HEPA: _____	0.1	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 11.3 Expiry Date: October 6, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: December 6, 2023
 Date Disposable Filter Changed: December 6, 2023

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 6, 2023
 Date RH/T Sensor Cleaned: July 6, 2023

Notes: No adjustments made. Leak check passed. Head cleaned.

Calibration by: Mohammed Kashif



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Anzac	Station Number:	AMS 14
Calibration Date:	April 8, 2024	Prev Cal Date:	August 23, 2023
Start Time (MST):	12:15	End Time (MST):	12:30
Tower Height (m):	20.0	Reason:	Removal

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	E5132
WS Calibrator:	MetOne 053	Serial Number:	R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.2	0.2%
400	39.4	39.4	0.1%
600	58.6	58.6	0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		1.000000	≥0.9995
Calculated slope		0.999465	0.90 - 1.10
Calculated intercept		-0.013446	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	Z1048
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:34	Calc Declination*:	13.23 Degrees
Deadband calc:	#VALUE!	degrees (<i>Limit 4 deg</i>)	* - calculated declination as per NOAA website

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	NA	---
90	NA	
180	NA	
270	NA	
356	NA	

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)			≥0.9995
Calculated slope			0.90 - 1.10
Calculated intercept			+/- 4

Notes: Wind direction sensor was non-functional on arrival due to broken potentiometer connection. Wind speed removal cal only.

Calibration Performed By: Kelly Baragar



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Anzac	Station Number:	AMS 14
Calibration Date:	April 8, 2024	Prev Cal Date:	March 11, 2024
Start Time (MST):	13:15	End Time (MST):	14:30
Tower Height (m):	20.0	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	U11126
WS Calibrator:	MetOne 053	Serial Number:	R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.2	0.2%
400	39.4	39.4	0.1%
600	58.6	58.6	0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		1.000000	≥0.9995
Calculated slope		0.999465	0.90 - 1.10
Calculated intercept		-0.013446	+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	C21020
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	12:26	Calc Declination*:	13.23 Degrees
Deadband calc:	1.9 degrees (<i>Limit 4 deg</i>)	* - calculated declination as per NOAA website	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.9	---
90	89.5	-0.1%
180	179.9	0.0%
270	270.2	0.1%
356	356.0	0.0%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)		0.999994	≥0.9995
Calculated slope		1.001229	0.90 - 1.10
Calculated intercept		-0.320328	+/- 4

Notes: Replaced all sensors and cabling.

Calibration Performed By: Kelly Baragar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS17
WAPASU
APRIL 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Wapasu	Station number: AMS17
Calibration Date:	April 2, 2024	Last Cal Date: March 6, 2024
Start time (MST):	10:35	End time (MST): 13:38
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	50.38 ppm	Cal Gas Exp Date: January 12, 2029
Cal Gas Cylinder #:	ALM066507	
Removed Cal Gas Conc:	50.38 ppm	Rem Gas Exp Date: N/A
Removed Gas Cyl #:	N/A	Diff between cyl:
Calibrator Model:	API T700	Serial Number: 2449
Zero Air Gen Model:	API 701H	Serial Number: 359

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 1218153459
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003695	0.998796	Backgd or Offset:	13.2	13.3
Calibration intercept:	-1.998962	-1.859943	Coeff or Slope:	1.098	1.098

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4921	79.4	800.0	790.0	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	790.2	Previous response	800.9	*% change	-1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4921	79.4	800.0	798.6	1.002
Mid point	4960	39.7	400.0	395.4	1.012
Low point	4980	19.8	199.5	196.2	1.017
As left zero	5000	0.0	0.0	0.3	----
As left span	4920	79.4	800.1	800.0	1.000
Average Correction Factor:					1.010

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 6, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:35	End Time (MST):	13:38
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999962	≥0.995
800.0	798.6	1.0017	Slope	0.998796	0.90 - 1.10
400.0	395.4	1.0117	Intercept	-1.859943	+/-30
199.5	196.2	1.0169			



SO2 Calibration Plot

Date: April 2, 2024

Location: Wapasu





Wood Buffalo Environmental Association

H₂S/TRS Calibration Report

Station Information

Station Name:	Wapasu	Station number:	AMS 17
Calibration Date:	April 8, 2024	Last Cal Date:	March 12, 2024
Start time (MST):	10:28	End time (MST):	16:13
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.08	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511852			
Removed Cal Gas Conc:	5.08	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2449
ZAG Make/Model:	API T701H		Serial Number:	359

Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583
Converter make:	CD Nova	Converter serial #:	N/A
Analyzer Range	0 - 100 ppb	Converter Temp:	degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995282	0.992711	Backgd or Offset:	11.7	11.8
Calibration intercept:	0.340790	-0.119219	Coeff or Slope:	1.096	1.096

H₂S/TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero)) Limit = 0.90-1.10
As found zero	5000	0.0	0.0	0.6	----
As found High point	4921	78.8	80.0	80.5	1.001
As found Mid point	4961	39.4	40.0	40.4	1.005
As found Low point	4980	19.7	20.0	20.4	1.010
New cylinder response					
Baseline Corr As found:	79.9	Prev response:	79.96	*% change:	-0.1%
Baseline Corr 2nd AF pt:	39.8	AF Slope:	0.999282	AF Intercept:	0.500796
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999993	* = > +/-5% change initiates investigation	

H₂S/TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4921	78.8	80.0	79.5	1.006
Mid point	4961	39.4	40.0	39.3	1.018
Low point	4980	19.7	20.0	19.5	1.026
As left zero	5000	0.0	0.0	0.4	----
As left span	4921	78.8	80.0	78.9	1.014
SO2 Scrubber Check	4921	79.4	793.9	-0.1	----
Date of last scrubber change:		N/A		Ave Corr Factor	1.017
Date of last converter efficiency test:		N/A			

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


H₂S/TRS Calibration Summary

Station Information

Calibration Date:	April 8, 2024	Previous Calibration:	March 12, 2024
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:28	End Time (MST):	16:13
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999919	≥0.995
80.0	79.5	1.0063	Slope	0.992711	0.90 - 1.10
40.0	39.3	1.0177	Intercept	-0.119219	+/-3
20.0	19.5	1.0257			



H₂S/TRS Calibration Plot

Date: April 8, 2024

Location: Wapasu





Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 6, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:35	End Time (MST):	13:38
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.10	----	Correlation Coefficient	0.999897	≥0.995
17.09	17.01	1.0047	Slope	0.990526	0.90 - 1.10
8.55	8.41	1.0162	Intercept	0.030412	+/-1.5
4.26	4.22	1.0108			



THC Calibration Plot

Date: April 2, 2024

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Wapasu
 Station number: AMS 17
 Calibration Date: April 16, 2024
 Last Cal Date: March 26, 2024
 Start time (MST): 10:35
 End time (MST): N/A
 Reason: As Found

Calibration Standards

NO Gas Cylinder #: T375YK8
 NOX Cal Gas Conc: 49.11 ppm
 Removed Cylinder #: T375YK8
 Removed Gas NOX Conc: 49.11 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701H
 Cal Gas Expiry Date: April 13, 2025
 NO Cal Gas Conc: 48.07 ppm
 Removed Gas Exp Date: N/A
 Removed Gas NO Conc: 48.07 ppm
 NO gas Diff:
 Serial Number: 2449
 Serial Number: 359

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
AF High point	4917	83.2	817.2	799.9	17.3	809.8	791.0	18.7	1.0088	1.0110
AF Mid point	4958	41.6	408.6	399.9	8.7	395.7	385.3	10.4	1.0318	1.0375
AF Low point	4979	20.8	204.3	200.0	4.3	193.5	187.4	6.2	1.0542	1.0659

New cyl resp

Previous Response	NO _x = 816.2 ppb	NO = 796.2 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -0.8%
Baseline Corr 1st pt	NO _x = 810.1 ppb	NO = 791.2 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -0.6%
Baseline Corr 2nd pt	NO _x = 396.0 ppb	NO = 385.5 ppb	As found	NO _x r ² : 0.999787	Nx SI: 0.993718	Nx Int: -5.600
Baseline Corr 3rd pt	NO _x = 193.8 ppb	NO = 187.6 ppb	As found	NO r ² : 0.999710	NO SI: 0.992071	NO Int: -6.300
			As found	NO ₂ r ² : 1.000000	NO ₂ SI: 0.991618	NO ₂ Int: -0.100

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	-0.1	----	----
As found high GPT point	791.0	390.1	418.2	414.6	1.0087	99.1%
As found mid GPT point	791.0					
As found low GPT point	791.0					



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo Scientific 42iQ
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12300522720

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.004319	
NO _x Cal Offset:	-4.540000	
NO Cal Slope:	1.001930	
NO Cal Offset:	-5.200000	
NO ₂ Cal Slope:	1.007870	
NO ₂ Cal Offset:	1.467173	

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.345	N/A	NO bkgnd or offset:	8.1	N/A
NOX coeff or slope:	0.994	N/A	NOX bkgnd or offset:	8.1	N/A
NO2 coeff or slope:	1.000	N/A	Reaction cell Press:	404.0	N/A

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero
 High point
 Mid point
 Low point
 As left zero
 As left span

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero
 High GPT point
 Mid GPT point
 Low GPT point

Average Correction Factor

Notes:

Pump changed out after as founds. Will calibrate instrument tomorrow.


Calibration Performed By:

Aswin Sasi Kumar

NO_x Calibration Plot

Date: April 16, 2024

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Wapasu
 Station number: AMS 17
 Calibration Date: April 18, 2024
 Last Cal Date: N/A
 Start time (MST): 10:37
 End time (MST): 15:02
 Reason: Install

Calibration Standards

NO Gas Cylinder #: T375YK8
 NOX Cal Gas Conc: 49.11 ppm
 Removed Cylinder #: T375YK8
 Removed Gas NOX Conc: 49.11 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701H
 Cal Gas Expiry Date: April 13, 2025
 NO Cal Gas Conc: 48.07 ppm
 Removed Gas Exp Date: N/A
 Removed Gas NO Conc: 48.07 ppm
 NO gas Diff:
 Serial Number: 2449
 Serial Number: 359

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = NA	ppb	NO = NA	ppb					*Percent Change	NO _x = NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb					*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb					Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb					NO SI:	NO Int:
									NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo Scientific 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153460

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	N/A	0.999550
NO _x Cal Offset:	N/A	-0.560000
NO Cal Slope:	N/A	1.000573
NO Cal Offset:	N/A	-1.400000
NO ₂ Cal Slope:	N/A	0.995649
NO ₂ Cal Offset:	N/A	-0.696163

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	N/A	1.100	NO bkgnd or offset:	N/A	3.8
NOX coeff or slope:	N/A	0.992	NOX bkgnd or offset:	N/A	4.2
NO2 coeff or slope:	N/A	1.000	Reaction cell Press:	N/A	261.0

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.0	----	----
High point	4917	83.2	817.2	799.9	17.3	816.8	799.8	17.0	1.0005	1.0001
Mid point	4958	41.6	408.6	399.9	8.7	406.9	397.6	9.3	1.0042	1.0059
Low point	4979	20.8	204.3	200.0	4.3	203.4	197.6	5.8	1.0044	1.0120
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
As left span	4917	83.2	817.2	401.2	416.0	817.0	401.2	415.8	1.0002	1.0000
Average Correction Factor									1.0030	1.0060

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	797.8	400.1	415.0	413.3	1.0041	99.6%
Mid GPT point	797.8	593.7	221.4	219.0	1.0110	98.9%
Low GPT point	797.8	593.7	221.4	219.0	1.0110	98.9%
Average Correction Factor					1.0087	99.1%

Notes:

Install calibration. Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	N/A
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:37	End Time (MST):	15:02
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999995	≥0.995
817.2	816.8	1.0005	Slope	0.999550	0.90 - 1.10
408.6	406.9	1.0042	Intercept	-0.560000	+/-20
204.3	203.4	1.0044			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	N/A
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:37	End Time (MST):	15:02
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999974	≥0.995
415.0	413.3	1.0041	Slope	0.995649	0.90 - 1.10
221.4	219.0	1.0110	Intercept	-0.696163	+/-20
221.4	219.0	1.0110			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	N/A
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:37	End Time (MST):	15:02
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999985	≥0.995
799.9	799.8	1.0001	Slope	1.000573	0.90 - 1.10
399.9	397.6	1.0059	Intercept	-1.400000	+/-20
200.0	197.6	1.0120			



NO_x Calibration Plot

Date: April 18, 2024

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	April 4, 2024	Last Cal Date:	March 11, 2024
Start time (MST):	10:03	End time (MST):	13:20
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	2449
Calibrator Make/Model:	API T700	Serial Number:	359
ZAG Make/Model:	API T701H		

Analyzer Information

Analyzer make:	API T400	Analyzer serial #:	3870
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000657	1.000714	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.340000	-0.600000	Coeff or Slope:	1.013	1.013

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	5000	1077.3	400.0	399.4	1.002
As found Mid point					
As found Low point					
Baseline Corr As found:	399.2	Previous response	399.9	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	5000	1077.3	400.0	400.1	1.000
Mid point	5000	900.3	200.0	199.3	1.004
Low point	5000	789.5	100.0	98.3	1.017
As left zero	5000	0.0	0.0	0.6	----
As left span	5000	1077.3	400.0	404.0	0.990
Average Correction Factor					1.007

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 11, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:03	End Time (MST):	13:20
Analyzer make:	API T400	Analyzer serial #:	3870

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999970	≥0.995
400.0	400.1	0.9998	Slope	1.000714	0.90 - 1.10
200.0	199.3	1.0035	Intercept	-0.600000	+/- 5
100.0	98.3	1.0173			



O₃ Calibration Plot

Date: April 4, 2024

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Wapasu Station number: AMS 17
 Calibration Date: April 17, 2024 Last Cal Date: March 26, 2024
 Start time (MST): 10:36 End time (MST): 12:28

Analyzer Make: Teledyne API T640 S/N: 1183
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-3.8	-3.68	-3.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	726.3	729.6	726.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.98	5.00	4.98	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	41	----	41	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.1	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: October 6, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	12.7	11.1	11.1	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 13, 2024
 Date Disposable Filter Changed: February 13, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____
 Date RH/T Sensor Cleaned: _____

Notes: Temp. pressure and flow checked. Leak check passed. Chamber cleaning completed. PMT adjusted.

Calibration by: Aswin Sasi Kumar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	April 3, 2024	Last Cal Date:	March 1, 2024
Start time (MST):	10:20	End time (MST):	14:11
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.40	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC463851			
Removed Cal Gas Conc:	49.40	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	2658
Zero Air Gen Model:	Teledyne API 701H		Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	JC1501301453
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000733	1.002389	Backgd or Offset:	22.9	22.9
Calibration intercept:	0.276595	0.496756	Coeff or Slope:	0.800	0.800

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5009	0.0	0.0	0.8	----
As found High point	4919	81.0	800.3	801.2	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	800.4	Previous response	801.1	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.8	----
High point	4919	81.0	800.3	802.5	0.997
Mid point	4959	40.5	400.2	402.5	0.994
Low point	4979	20.2	199.6	199.6	1.000
As left zero	5000	0.0	0.0	1.1	----
As left span	4919	81.0	800.3	802.6	0.997
Average Correction Factor:					0.997

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 1, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:20	End Time (MST):	14:11
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.8	----	Correlation Coefficient	0.999995	≥0.995
800.3	802.5	0.9972	Slope	1.002389	0.90 - 1.10
400.2	402.5	0.9942	Intercept	0.496756	+/-30
199.6	199.6	1.0000			



SO2 Calibration Plot

Date: April 3, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Stony Mountain	Station number:	AMS18
Calibration Date:	April 24, 2024	Last Cal Date:	March 27, 2024
Start time (MST):	10:10	End time (MST):	15:25
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.48	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC500395			
Removed Cal Gas Conc:	5.48	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701		Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359
Converter make:	CD Nova CDN-101	Converter serial #:	555
Analyzer Range	0 - 100 ppb	Converter Temp:	799 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002157	0.993725	Backgd or Offset:	2.7	2.6
Calibration intercept:	0.240997	0.261187	Coeff or Slope:	1.157	1.137

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4927	73.0	80.0	82.8	0.968
As found Mid point	4964	36.5	40.0	41.7	0.964
As found Low point	4983	18.3	20.0	20.7	0.978
New cylinder response					
Baseline Corr As found:	82.6	Prev response:	80.41	*% change:	2.7%
Baseline Corr 2nd AF pt:	41.5	AF Slope:	1.033737	AF Intercept:	0.160426
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999979	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4927	73.0	80.0	79.7	1.004
Mid point	4964	36.5	40.0	40.2	0.995
Low point	4983	18.3	20.0	20.0	1.002
As left zero	5000	0.0	0.0	0.3	----
As left span	4927	73.0	80.0	78.6	1.018
SO2 Scrubber Check	4923	77.1	771.0	0.0	----
Date of last scrubber change:	17-Dec-21			Ave Corr Factor	1.000
Date of last converter efficiency test:					

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


H₂S/TRS Calibration Summary

Station Information

Calibration Date:	April 24, 2024	Previous Calibration:	March 27, 2024
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:10	End Time (MST):	15:25
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999978	≥0.995
80.0	79.7	1.0037	Slope	0.993725	0.90 - 1.10
40.0	40.2	0.9948	Intercept	0.261187	+/-3
20.0	20.0	1.0024			



H₂S/TRS Calibration Plot

Date: April 24, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	April 3, 2024	Last Cal Date:	March 1, 2024
Start time (MST):	10:20	End time (MST):	14:11
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.8 ppm	CH4 Equiv Conc.	1066.8 ppm
C3H8 Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	500.8 ppm	CH4 Equiv Conc.	1066.8 ppm
Removed C3H8 Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
Zero Air Gen model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1193585647
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.10E-04	3.23E-04	NMHC SP Ratio:	5.89E-05	5.94E-05
CH4 Retention time:	16.2	16.4	NMHC Peak Area:	156523	154457
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.08	----
As found High point	4919	81.0	17.28	16.83	1.032
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.75	Prev response	17.37	*% change	-3.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.07	----
High point	4919	81.0	17.28	17.34	0.997
Mid point	4959	40.5	8.64	8.69	0.994
Low point	4979	20.2	4.31	4.38	0.984
As left zero	5000	0.0	0.00	0.03	----
As left span	4919	81.0	17.28	17.68	0.977
Average Correction Factor					0.992

Notes: As found CH₄ 5.4% out. Diagnostics all normal, possible shift after zero chromatogram procedure last month. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	81.0	9.17	9.06	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.06	Prev response	9.23	*% change	-1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	81.0	9.17	9.19	0.997
Mid point	4959	40.5	4.58	4.62	0.993
Low point	4979	20.2	2.29	2.30	0.995
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	81.0	9.17	9.41	0.974
Average Correction Factor					0.995

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.08	----
As found High point	4919	81.0	8.11	7.77	1.054
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.70	Prev response	8.15	*% change	-5.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.07	----
High point	4919	81.0	8.11	8.15	0.996
Mid point	4959	40.5	4.06	4.07	0.996
Low point	4979	20.2	2.02	2.08	0.971
As left zero	5000	0.0	0.00	0.03	----
As left span	4919	81.0	8.11	8.27	0.981
Average Correction Factor					0.988

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.003678	0.999037
THC Cal Offset:	0.027639	0.068218
CH ₄ Cal Slope:	1.002169	0.994973
CH ₄ Cal Offset:	0.018204	0.061988
NMHC Cal Slope:	1.005151	1.002583
NMHC Cal Offset:	0.009635	0.006430

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 1, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:20	End Time (MST):	14:11
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.07	----	Correlation Coefficient	0.999999	<i>>0.995</i>
17.28	17.34	0.9968	Slope	0.999037	<i>0.90 - 1.10</i>
8.64	8.69	0.9944	Intercept	0.068218	<i>+/-0.5</i>
4.31	4.38	0.9839			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 1, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:20	End Time (MST):	14:11
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.07	----	Correlation Coefficient	0.999975 ≥0.995
8.11	8.15	0.9961	Slope	0.994973 0.90 - 1.10
4.06	4.07	0.9960	Intercept	0.061988 +/-0.5
2.02	2.08	0.9715		





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 1, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:20	End Time (MST):	14:11
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999994	<i>>0.995</i>
9.17	9.19	0.9974	Slope	1.002583	<i>0.90 - 1.10</i>
4.58	4.62	0.9932	Intercept	0.006430	<i>+/-0.5</i>
2.29	2.30	0.9947			



NMHC Calibration Plot

Date: April 3, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	April 26, 2024	Last Cal Date:	April 3, 2024
Start time (MST):	13:45	End time (MST):	15:18
Reason:	Cylinder Change N2		

Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.8 ppm	CH4 Equiv Conc.	1066.8 ppm
C3H8 Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	500.8 ppm	CH4 Equiv Conc.	1066.8 ppm
Removed C3H8 Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
Zero Air Gen model:	Teledyne API T701H	Serial Number:	360

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1193585647
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.23E-04	3.23E-04	NMHC SP Ratio:	5.94E-05	5.94E-05
CH4 Retention time:	16.4	16.4	NMHC Peak Area:	154457	154457
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.03	----
As found High point	4919	81.0	17.28	17.38	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.35	Prev response	17.33	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.03	----
High point	4919	81.0	17.28	17.44	0.991
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 0.991

Notes: N2 cylinder change



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	81.0	9.17	9.24	0.992
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.24	Prev response	9.20	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	81.0	9.17	9.17	1.000
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.000

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.03	----
As found High point	4919	81.0	8.11	8.15	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.12	Prev response	8.13	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.03	----
High point	4919	81.0	8.11	8.27	0.981
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 0.981

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999037	1.007271
THC Cal Offset:	0.068218	0.033000
CH ₄ Cal Slope:	0.994973	1.015289
CH ₄ Cal Offset:	0.061988	0.033000
NMHC Cal Slope:	1.002583	1.000176
NMHC Cal Offset:	0.006430	0.000000

Calibration Performed By: Aswin Sasi Kumar

NMHC Calibration Plot

Date: April 26, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Stony Mountain
 Station number: AMS 18
 Calibration Date: April 23, 2024
 Last Cal Date: March 26, 2024
 Start time (MST): 9:52
 End time (MST): 14:53
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T26DHGA
 NOX Cal Gas Conc: 48.28 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.28 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: November 17, 2026
 NO Cal Gas Conc: 47.58 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 47.58 ppm
 NO gas Diff:
 Serial Number: 2658
 Serial Number: 13779

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
AF High point	4916	84.0	811.1	799.3	11.8	811.7	800.3	11.3	0.9991	0.9986
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 809.1 ppb	NO = 797.8 ppb	<i>* = > +/-5% change initiates investigation</i>				*Percent Change	NO _x = 0.3%		
Baseline Corr 1st pt	NO _x = 811.8 ppb	NO = 800.5 ppb	<u>As Found Statistics</u>				*Percent Change	NO = 0.3%		
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :		Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :		NO SI:	NO Int:			
			As found	NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1336160088

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998146	1.003895
NO _x Cal Offset:	-0.500000	-0.640000
NO Cal Slope:	0.998819	1.005239
NO Cal Offset:	-0.600000	-1.220000
NO ₂ Cal Slope:	0.999802	1.001293
NO ₂ Cal Offset:	0.793514	-0.223061

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.119	1.119	NO bkgnd or offset:	3.1	3.1
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	3.2	3.1
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	253.2	256.4

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.1	----	----
High point	4916	84.0	811.1	799.3	11.8	813.7	802.7	11.1	0.9968	0.9958
Mid point	4958	42.0	405.6	399.7	5.9	407.0	400.5	6.6	0.9964	0.9979
Low point	4979	21.0	202.8	199.8	2.9	201.6	198.2	3.4	1.0058	1.0083
As left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
As left span	4916	84.0	811.1	371.4	439.7	816.5	371.4	445.1	0.9934	1.0000
Average Correction Factor									0.9997	1.0007

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	801.4	372.2	441.0	441.6	0.9986	100.1%
Mid GPT point	801.4	588.1	225.1	224.6	1.0020	99.8%
Low GPT point	801.4	696.6	116.6	116.4	1.0014	99.9%
Average Correction Factor					1.0007	99.9%

Notes:

No adjustments needed.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 23, 2024	Previous Calibration:	March 26, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:52	End Time (MST):	14:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999993	≥0.995
811.1	813.7	0.9968	Slope	1.003895	0.90 - 1.10
405.6	407.0	0.9964	Intercept	-0.640000	+/-20
202.8	201.6	1.0058			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 23, 2024	Previous Calibration:	March 26, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:52	End Time (MST):	14:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999995	≥0.995
441.0	441.6	0.9986	Slope	1.001293	0.90 - 1.10
225.1	224.6	1.0020	Intercept	-0.223061	+/-20
116.6	116.4	1.0014			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 23, 2024	Previous Calibration:	March 26, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:52	End Time (MST):	14:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999990	≥0.995
799.3	802.7	0.9958	Slope	1.005239	0.90 - 1.10
399.7	400.5	0.9979	Intercept	-1.220000	+/-20
199.8	198.2	1.0083			



NO_x Calibration Plot

Date: April 23, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	April 11, 2024	Last Cal Date:	March 20, 2024
Start time (MST):	10:35	End time (MST):	14:02
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	2658
Calibrator Make/Model:	Teledyne API T700	Serial Number:	360
ZAG Make/Model:	Teledyne API 701H		

Analyzer Information

Analyzer make:	API T400	Analyzer serial #:	825
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995743	1.000029	Backgd or Offset:	0.3	0.3
Calibration intercept:	0.220000	-0.180000	Coeff or Slope:	0.982	0.982

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	NA	0.0	-0.4	----
As found High point	4804	1141.9	400.0	402.0	0.994
As found Mid point					
As found Low point					
Baseline Corr As found:	402.4	Previous response	398.5	*% change	1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	NA	0.0	-0.2	----
High point	4888	1138.1	400.0	399.7	1.001
Mid point	4888	884.5	200.0	200.2	0.999
Low point	4888	741.4	100.0	99.6	1.004
As left zero	5000	NA	0.0	-0.1	----
As left span	4812	1097.9	400.0	403.0	0.993
Average Correction Factor					1.001

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 20, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:35	End Time (MST):	14:02
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999998	≥0.995
400.0	399.7	1.0008	Slope	1.000029	0.90 - 1.10
200.0	200.2	0.9990	Intercept	-0.180000	+/- 5
100.0	99.6	1.0040			



O₃ Calibration Plot

Date: April 11, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: April 26, 2024 Last Cal Date: March 27, 2024
Start time (MST): 13:07 End time (MST): 13:55

Analyzer Make: API T640 S/N: 1162
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	7.4	7.6	7.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	701.2	702.4	701.2	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	5.03	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	----	37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.4	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	N/A		N/A	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 21, 2024
Date Disposable Filter Changed: February 21, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022
Date RH/T Sensor Cleaned: August 30, 2022

Notes: Flow, pressure and temp checked. Leak check passed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	April 5, 2024	Last Cal Date:	March 14, 2024
Start time (MST):	11:20	End time (MST):	14:03
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,080	ppm	Cal Gas Exp Date: November 4, 2028
Cal Gas Cylinder #:	EB0065608		
Removed Cal Gas Conc:	3,080	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 2658
ZAG Make/Model:	Teledyne API T701H		Serial Number: 355

Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996462	1.006642	Backgd or Offset:	-0.010	-0.010
Calibration intercept:	0.037795	0.249837	Coeff or Slope:	0.905	0.905

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4933	66.7	41.1	41.5	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	41.27	Prev response:	40.98	*% change:	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4933	66.7	41.1	41.5	0.990
Mid point	4966	33.3	20.5	21.1	0.970
Low point	4983	16.7	10.3	10.5	0.980
As left zero	5000	0.0	0.0	0.2	----
As left span	4933	66.7	41.1	41.4	0.992
Average Correction Factor					0.980

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


CO Calibration Summary

Station Information

Calibration Date:	April 5, 2024	Previous Calibration:	March 14, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:20	End Time (MST):	14:03
Analyzer make:	API T300	Analyzer serial #:	3504

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999916	≥0.995
41.1	41.5	0.9896	Slope	1.006642	0.90 - 1.10
20.5	21.1	0.9705	Intercept	0.249837	+/-1.5
10.3	10.5	0.9798			



CO Calibration Plot

Date: April 5, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Report

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	April 26, 2024	Last Cal Date:	March 21, 2024
Start time (MST):	10:05	End time (MST):	13:47
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	59,100	ppm	Cal Gas Exp Date: November 4, 2028
Cal Gas Cylinder #:	EB0065608		
Removed Cal Gas Conc:	59,100	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 2658
N2 Gen Make/Model:	Peak Scientific		Serial Number: 771048317

Analyzer Information

Analyzer make:	API T360	Analyzer serial #: 489
Analyzer Range	0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000660	1.006157	Backgd or Offset:	-0.037
Calibration intercept:	-4.780000	-3.120000	Coeff or Slope:	0.939

CO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	3000	0.0	0.0	0.1	----
As found High Point	2920	80.0	1576.0	1581.2	0.997
As found Mid Point					
As found Low Point					
New cylinder response					
Baseline Corr As found:	1581.1	Prev response:	1572.3	*% change:	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	3000	0.0	0.0	0.5	----
High point	2920	80.0	1576.0	1586.2	0.994
Mid point	2960	40.0	788.0	782.5	1.007
Low point	2980	20.0	394.0	393.3	1.002
As left zero	3000	0.0	0.0	0.0	----
As left span	2930	80.0	1570.8	1584.0	0.992
Average Correction Factor					1.001

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


CO₂ Calibration Summary

Station Information

Calibration Date	April 26, 2024	Previous Calibration	March 21, 2024
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:05	End Time (MST)	13:47
Analyzer make	API T360	Analyzer serial #	489

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.5	----	Correlation Coefficient	0.999943	≥0.995
1576.0	1586.2	0.9936	Slope	1.006157	0.90 - 1.10
788.0	782.5	1.0070	Intercept	-3.1	+/-20
394.0	393.3	1.0018			



CO₂ Calibration Plot

Date: April 26, 2024

Location: Stony Mountain





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS19 FIREBAG APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	April 22, 2024	Last Cal Date:	March 12, 2024
Start time (MST):	10:11	End time (MST):	13:21
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.29	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC716618			
Removed Cal Gas Conc:	49.29	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	1607
Zero Air Gen Model:	Teledyne API T701H		Serial Number:	201

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1410661308
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997705	0.996719	Backgd or Offset:	10.5	10.6
Calibration intercept:	0.418240	0.438192	Coeff or Slope:	0.991	0.991

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	4999	0.0	0.0	-0.1	----
As found High point	4919	81.1	799.5	797.0	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.1	Previous response	798.1	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	4999	0.0	0.0	-0.1	----
High point	4919	81.1	799.5	796.8	1.003
Mid point	4959	40.6	400.3	400.3	1.000
Low point	4980	20.3	200.1	200.0	1.001
As left zero	4999	0.0	0.0	0.2	----
As left span	4919	81.1	799.5	799.2	1.000
Average Correction Factor:					1.001

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 22, 2024	Previous Calibration:	March 12, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:11	End Time (MST):	13:21
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999996	≥0.995
799.5	796.8	1.0033	Slope	0.996719	0.90 - 1.10
400.3	400.3	0.9999	Intercept	0.438192	+/-30
200.1	200.0	1.0005			



SO2 Calibration Plot

Date: April 22, 2024

Location: Firebag





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	April 16, 2024	Last Cal Date:	March 5, 2024
Start time (MST):	10:19	End time (MST):	14:48
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.114	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517427			
Removed Cal Gas Conc:	5.114	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1607
ZAG Make/Model:	Teledyne API T701		Serial Number:	201

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032
Converter make:	Global	Converter serial #:	2022-222
Analyzer Range	0 - 100 ppb	Converter Temp:	350 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999337	1.001197	Backgd or Offset:	3.22	2.72
Calibration intercept:	-0.081535	0.078370	Coeff or Slope:	1.198	1.163

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.5	----
As found High point	4922	78.2	80.0	83.0	0.958
As found Mid point	4961	39.1	40.0	41.2	0.959
As found Low point	4980	19.6	20.0	20.4	0.959
New cylinder response					
Baseline Corr As found:	83.5	Prev response:	79.85	*% change:	4.4%
Baseline Corr 2nd AF pt:	41.7	AF Slope:	1.044074	AF Intercept:	-0.522523
Baseline Corr 3rd AF pt:	20.9	AF Correlation:	1.000000	<i>* = > +/-5% change initiates investigation</i>	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4922	78.2	80.0	80.1	0.998
Mid point	4961	39.1	40.0	40.2	0.995
Low point	4980	19.6	20.0	20.2	0.992
As left zero	5000	0.0	0.0	0.1	----
As left span	4922	78.2	80.0	79.3	1.009
SO2 Scrubber Check	4922	78.3	783.0	0.0	----
Date of last scrubber change:	<u>18-Jan-23</u>			Ave Corr Factor	0.995
Date of last converter efficiency test:	<u>n/a</u>				

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	April 16, 2024	Previous Calibration:	March 5, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:19	End Time (MST):	14:48
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999995	≥ 0.995
80.0	80.1	0.9985	Slope	1.001197	$0.90 - 1.10$
40.0	40.2	0.9948	Intercept	0.078370	± 3
20.0	20.2	0.9925			



H2S Calibration Plot

Date: April 16, 2024

Location: Firebag





Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 22, 2024	Previous Calibration:	March 12, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:11	End Time (MST):	13:21
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.05	----	Correlation Coefficient	0.999982	≥0.995
17.31	17.27	1.0020	Slope	0.996237	0.90 - 1.10
8.66	8.62	1.0051	Intercept	0.011064	+/-1.5
4.33	4.30	1.0083			



THC Calibration Plot

Date: April 22, 2024

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Firebag
 Station number: AMS 19
 Calibration Date: April 3, 2024
 Last Cal Date: March 21, 2024
 Start time (MST): 10:40
 End time (MST): 13:45
 Reason: Maintenance

Calibration Standards

NO Gas Cylinder #: DT0044018
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: November 3, 2031
 NO Cal Gas Conc: 48.70 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 48.70 ppm
 NO gas Diff:
 Serial Number: 1607
 Serial Number: 201

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
AF High point	4918	82.1	802.9	799.7	3.3	772.9	767.5	5.4	1.0386	1.0418
AF Mid point	4959	41.1	402.0	400.3	1.6	387.6	383.8	3.7	1.0365	1.0428
AF Low point	4980	20.5	200.5	199.7	0.8	194.0	191.6	2.4	1.0324	1.0416

New cyl resp

Previous Response	NO _x = 808.4 ppb	NO = 805.5 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -4.6%
Baseline Corr 1st pt	NO _x = 773.1 ppb	NO = 767.6 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -4.9%
Baseline Corr 2nd pt	NO _x = 387.8 ppb	NO = 383.9 ppb	As found	NO _x r ² : 0.999997	Nx SI: 0.962454	Nx Int: 0.420
Baseline Corr 3rd pt	NO _x = 194.2 ppb	NO = 191.7 ppb	As found	NO r ² : 1.000000	NO SI: 0.959848	NO Int: -0.160
			As found	NO ₂ r ² : 0.999965	NO ₂ SI: 0.999204	NO ₂ Int: -0.649

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	-0.1	----	----
As found high GPT point	765.1	371.3	397.1	396.8	1.0007	99.9%
As found mid GPT point	765.1	568.4	200.0	197.7	1.0116	98.9%
As found low GPT point	765.1	668.2	100.2	99.7	1.0049	99.5%



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1410661309

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.006289	NA
NO _x Cal Offset:	0.400360	NA
NO Cal Slope:	1.006923	NA
NO Cal Offset:	0.340247	NA
NO ₂ Cal Slope:	0.998051	NA
NO ₂ Cal Offset:	-0.640112	NA

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.040	1.040	NO bkgnd or offset:	7.3	7.3
NOX coeff or slope:	0.995	0.995	NOX bkgnd or offset:	7.4	7.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	216.5	210.3

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero
 High point
 Mid point
 Low point
 As left zero
 As left span

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero
 High GPT point
 Mid GPT point
 Low GPT point

Average Correction Factor


Notes: As founds done for maintenance. Changing reaction cell O-rings, pump and charcoal filter.

Calibration Performed By: Braiden Boutilier

NO_x Calibration Plot

Date: April 3, 2024

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Firebag
 Station number: AMS 19
 Calibration Date: April 4, 2024
 Last Cal Date: April 3, 2024
 Start time (MST): 10:15
 End time (MST): 15:04
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0044018
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: November 3, 2031
 NO Cal Gas Conc: 48.70 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 48.70 ppm
 NO gas Diff:
 Serial Number: 1607
 Serial Number: 201

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = NA	ppb	NO = NA	ppb					*Percent Change	NO _x = NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb					*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb						
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb						
* = > +/-5% change initiates investigation										
<u>As Found Statistics</u>										
						As found NO _x r ² :			Nx SI:	Nx Int:
						As found NO r ² :			NO SI:	NO Int:
						As found NO ₂ r ² :			NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1410661309

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.006289	0.998971
NO _x Cal Offset:	0.400360	1.739938
NO Cal Slope:	1.006923	1.001345
NO Cal Offset:	0.340247	0.620021
NO ₂ Cal Slope:	0.998051	1.000081
NO ₂ Cal Offset:	-0.640112	0.547188

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.040	0.990	NO bkgnd or offset:	7.3	5.0
NOX coeff or slope:	0.995	0.995	NOX bkgnd or offset:	7.4	5.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	216.5	177.5

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
High point	4918	82.1	802.9	799.7	3.3	803.0	801.0	1.8	0.9999	0.9983
Mid point	4959	41.1	402.0	400.3	1.6	404.2	401.8	2.4	0.9945	0.9963
Low point	4980	20.5	200.5	199.7	0.8	203.6	201.3	2.3	0.9847	0.9919
As left zero	5000	0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
As left span	4918	82.1	802.9	384.9	418.0	804.0	384.9	419.3	0.9987	1.0000
Average Correction Factor									0.9930	0.9955

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	798.0	383.8	417.5	417.7	0.9995	100.1%
Mid GPT point	798.0	593.1	208.2	209.5	0.9937	100.6%
Low GPT point	798.0	697.5	103.8	104.3	0.9951	100.5%
Average Correction Factor					0.9961	100.4%

Notes: Calibration after maintenance done on April 3. Changed sample inlet filter. Reset calibration factors. Adjusted PMT voltage, zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	April 3, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:15	End Time (MST):	15:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999981	≥0.995
802.9	803.0	0.9999	Slope	0.998971	0.90 - 1.10
402.0	404.2	0.9945	Intercept	1.739938	+/-20
200.5	203.6	0.9847			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	April 3, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:15	End Time (MST):	15:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999991	≥0.995
417.5	417.7	0.9995	Slope	1.000081	0.90 - 1.10
208.2	209.5	0.9937	Intercept	0.547188	+/-20
103.8	104.3	0.9951			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	April 3, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:15	End Time (MST):	15:04
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999996	≥0.995
799.7	801.0	0.9983	Slope	1.001345	0.90 - 1.10
400.3	401.8	0.9963	Intercept	0.620021	+/-20
199.7	201.3	0.9919			



NO_x Calibration Plot

Date: April 4, 2024

Location: Firebag





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS20 MACKAY RIVER APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	MacKay River	Station number:	AMS 20
Calibration Date:	April 10, 2024	Last Cal Date:	March 7, 2024
Start time (MST):	7:46	End time (MST):	10:29
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.22	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC30686			
Removed Cal Gas Conc:	49.22	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	1220
Zero Air Gen Model:	API 701		Serial Number:	4522

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1501301450
Analyzer Range:	0-1000ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004703	0.994567	Backgd or Offset:	19.7	19.3
Calibration intercept:	2.311780	3.111267	Coeff or Slope:	0.965	0.950

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4919	81.3	800.3	814.1	0.983
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	814.0	Previous response	806.3	*% change	0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.6	----
High point	4919	81.3	800.3	798.2	1.003
Mid point	4959	40.7	400.7	401.6	0.998
Low point	4980	20.3	199.8	205.2	0.974
As left zero	5000	0.0	0.0	0.6	----
As left span	4919	81.3	800.3	801.7	0.998
Average Correction Factor:					0.991

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 10, 2024	Previous Calibration:	March 7, 2024
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	7:46	End Time (MST):	10:29
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999947	≥0.995
800.3	798.2	1.0026	Slope	0.994567	0.90 - 1.10
400.7	401.6	0.9977	Intercept	3.111267	+/-30
199.8	205.2	0.9738			



SO2 Calibration Plot

Date: April 10, 2024

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	MacKay River	Station number:	AMS 20
Calibration Date:	April 9, 2024	Last Cal Date:	March 5, 2024
Start time (MST):	6:43	End time (MST):	11:00
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.12	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC515997			
Removed Cal Gas Conc:	5.12	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1220
ZAG Make/Model:	API 701		Serial Number:	4522

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117
Converter make:	Global	Converter serial #:	2022-226
Analyzer Range	0 - 100 ppb	Converter Temp:	325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.975111	0.984302	Backgd or Offset:	3.2	3.2
Calibration intercept:	0.699032	0.679249	Coeff or Slope:	1.113	1.113

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4922	78.1	80.0	80.2	0.998
As found Mid point	4961	39.0	39.9	40.6	0.986
As found Low point	4980	19.5	20.0	20.7	0.969
New cylinder response					
Baseline Corr As found:	80.1	Prev response:	78.68	*% change:	1.8%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	0.999736	AF Intercept:	0.439543
Baseline Corr 3rd AF pt:	20.6	AF Correlation:	0.999916	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4922	78.1	80.0	79.2	1.010
Mid point	4961	39.0	39.9	40.2	0.993
Low point	4980	19.5	20.0	20.7	0.965
As left zero	5000	0.0	0.0	0.3	----
As left span	4922	78.1	80.0	78.6	1.017
SO ₂ Scrubber Check	4982	81.3	802.8	0.0	----
Date of last scrubber change:	25-May-23			Ave Corr Factor	0.989
Date of last converter efficiency test:					

Notes: Flow sensor replaced and calibrated. Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 9, 2024	Previous Calibration:	March 5, 2024
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	6:43	End Time (MST):	11:00
Analyzer make:	Global	Analyzer serial #:	2022-226

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999894	≥ 0.995
80.0	79.2	1.0098	Slope	0.984302	$0.90 - 1.10$
39.9	40.2	0.9934	Intercept	0.679249	± 3
20.0	20.7	0.9647			



H₂S Calibration Plot

Date: April 9, 2024

Location: MacKay River





Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 10, 2024	Previous Calibration:	March 7, 2024
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	8:45	End Time (MST):	10:28
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.01	----	Correlation Coefficient	0.999986	≥0.995
17.34	17.27	1.0040	Slope	0.995490	0.90 - 1.10
8.68	8.60	1.0094	Intercept	-0.003380	+/-1.5
4.33	4.33	0.9990			



THC Calibration Plot

Date: April 10, 2024

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: MacKay River
 Station number: AMS 20
 Calibration Date: April 8, 2024
 Last Cal Date: March 6, 2024
 Start time (MST): 6:45
 End time (MST): 11:07
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T376265
 NOX Cal Gas Conc: 49.19 ppm
 Removed Cylinder #:
 Removed Gas NOX Conc: 49.19 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701
 Cal Gas Expiry Date: April 13, 2025
 NO Cal Gas Conc: 48.04 ppm
 Removed Gas Exp Date:
 Removed Gas NO Conc: 48.04 ppm
 NO gas Diff:
 Serial Number: 1220
 Serial Number: 4522

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.2	----	----
AF High point	4917	83.3	819.5	800.3	19.2	817.8	797.4	20.4	1.0019	1.0036
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 824.0 ppb		NO = 804.2 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -0.7%	
Baseline Corr 1st pt	NO _x = 817.9 ppb		NO = 797.4 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -0.8%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1505164379

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001037	0.994162
NO _x Cal Offset:	3.662215	3.902461
NO Cal Slope:	1.001142	0.993845
NO Cal Offset:	2.942582	2.922811
NO ₂ Cal Slope:	0.998215	1.003001
NO ₂ Cal Offset:	-0.831656	-0.362701

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.990	0.990	NO bkgnd or offset:	2.8	2.8
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	3.0	3.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	164.2	164.2

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
High point	4917	83.3	819.5	800.3	19.2	816.9	797.1	19.6	1.0031	1.0040
Mid point	4958	41.7	410.3	400.7	9.6	412.9	401.9	11.0	0.9936	0.9970
Low point	4979	20.8	204.6	199.9	4.8	211.8	204.8	7.1	0.9662	0.9759
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
As left span	4917	83.3	819.5	438.5	381.0	814.7	438.5	376.1	1.0058	1.0000
Average Correction Factor									0.9876	0.9923

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	793.3	438.0	374.5	375.2	0.9980	100.2%
Mid GPT point	793.3	609.3	203.2	203.8	0.9968	100.3%
Low GPT point	793.3	697.9	114.6	113.9	1.0058	99.4%
Average Correction Factor					1.0002	100.0%

Notes: No adjustments and maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 8, 2024	Previous Calibration:	March 6, 2024
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	6:45	End Time (MST):	11:07
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999892	≥0.995
819.5	816.9	1.0031	Slope	0.994162	0.90 - 1.10
410.3	412.9	0.9936	Intercept	3.902461	+/-20
204.6	211.8	0.9662			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 8, 2024	Previous Calibration:	March 6, 2024
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	6:45	End Time (MST):	11:07
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
374.5	375.2	0.9980	Slope	1.003001	<i>0.90 - 1.10</i>
203.2	203.8	0.9968	Intercept	-0.362701	<i>+/-20</i>
114.6	113.9	1.0058			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 8, 2024	Previous Calibration:	March 6, 2024
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	6:45	End Time (MST):	11:07
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999941	≥0.995
800.3	797.1	1.0040	Slope	0.993845	0.90 - 1.10
400.7	401.9	0.9970	Intercept	2.922811	+/-20
199.9	204.8	0.9759			



NO_x Calibration Plot

Date: April 8, 2024

Location: MacKay River





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS21
CONKLIN
APRIL 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	April 3, 2024	Last Cal Date:	March 18, 2024
Start time (MST):	9:14	End time (MST):	12:13
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.93	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	<u>CC259455</u>			
Removed Cal Gas Conc:	49.93	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	3810
Zero Air Gen Model:	Teledyne API 701		Serial Number:	953

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1428701363
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997419	0.992525	Backgd or Offset:	28.3	28.3
Calibration intercept:	1.855817	1.395547	Coeff or Slope:	0.901	0.901

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	80.2	800.8	794.0	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	794.0	Previous response	800.6	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4920	80.2	800.8	795.5	1.007
Mid point	4960	40.1	400.4	400.1	1.001
Low point	4980	20.0	200.1	200.5	0.998
As left zero	5005	0.0	0.0	0.5	----
As left span	4920	80.2	800.8	798.2	1.003
Average Correction Factor:					1.002

Notes: Sample inlet filter changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 18, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:14	End Time (MST):	12:13
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999990	≥0.995
800.8	795.5	1.0067	Slope	0.992525	0.90 - 1.10
400.4	400.1	1.0008	Intercept	1.395547	+/-30
200.1	200.5	0.9981			



SO2 Calibration Plot

Date: April 3, 2024

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	April 18, 2024	Last Cal Date:	March 21, 2024
Start time (MST):	10:03	End time (MST):	14:41
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.00	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC501204			
Removed Cal Gas Conc:	5.00	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3810
ZAG Make/Model:	Teledyne API 701H		Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116
Converter make:	CD-Nova 101	Converter serial #:	NA
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000286	0.999714	Backgd or Offset:	2.8	2.8
Calibration intercept:	0.140000	0.060000	Coeff or Slope:	0.998	0.998

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4920	80.0	80.0	80.0	0.998
As found Mid point	4960	40.0	40.0	40.0	0.995
As found Low point	4980	20.0	20.0	20.0	0.990
New cylinder response					
Baseline Corr As found:	80.2	Prev response:	80.16	*% change:	0.0%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	1.002000	AF Intercept:	-0.120000
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.0	80.0	80.0	1.000
Mid point	4960	40.0	40.0	40.1	0.998
Low point	4980	20.0	20.0	20.1	0.995
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	80.0	80.0	79.9	1.001
SO2 Scrubber Check	4920	80.2	802.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.998
Date of last converter efficiency test:					

Notes: Sample inlet filters was changed after multipoint as founds. SO2 scrubber check done after calibrator zero and passed. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 21, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:03	End Time (MST):	14:41
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999997	≥ 0.995
80.0	80.0	1.0000	Slope	0.999714	$0.90 - 1.10$
40.0	40.1	0.9975	Intercept	0.060000	± 3
20.0	20.1	0.9950			



TRS Calibration Plot

Date: April 18, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	April 3, 2024	Last Cal Date:	March 20, 2024
Start time (MST):	9:14	End time (MST):	12:13
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	
CH4 Cal Gas Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
C3H8 Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
Removed C3H8 Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
Zero Air Gen model:	Teledyne API 701H	Serial Number:	953

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1331259520
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.55E-04	2.26E-04	NMHC SP Ratio:	5.96E-05	5.81E-05
CH4 Retention time:	15.0	14.2	NMHC Peak Area:	153262	157380
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	17.13	17.39	0.985
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.39	Prev response	17.07	*% change	1.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	17.13	17.11	1.001
Mid point	4960	40.1	8.56	8.58	0.998
Low point	4980	20.0	4.28	4.33	0.988
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	17.13	17.15	0.999
Average Correction Factor					0.995

Notes: Sample inlet filter changed after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	9.14	9.22	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.22	Prev response	9.13	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	9.14	9.13	1.001
Mid point	4960	40.1	4.57	4.60	0.994
Low point	4980	20.0	2.28	2.34	0.978
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	9.14	9.15	0.999
Average Correction Factor					0.991

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	7.99	8.17	0.977
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.17	Prev response	7.94	*% change	2.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	7.99	7.99	1.000
Mid point	4960	40.1	3.99	3.99	1.001
Low point	4980	20.0	2.00	2.00	1.000
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	7.99	8.00	0.998
Average Correction Factor					1.000

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.993227	0.998192
THC Cal Offset:	0.057164	0.027967
CH ₄ Cal Slope:	0.991351	0.999780
CH ₄ Cal Offset:	0.026556	-0.000643
NMHC Cal Slope:	0.995229	0.997354
NMHC Cal Offset:	0.029409	0.027410

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 20, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:14	End Time (MST):	12:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999987	≥0.995
17.13	17.11	1.0009	Slope	0.998192	0.90 - 1.10
8.56	8.58	0.9975	Intercept	0.027967	+/-0.5
4.28	4.33	0.9881			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 20, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:14	End Time (MST):	12:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	1.000000	<i>≥0.995</i>
7.99	7.99	1.0001	Slope	0.999780	<i>0.90 - 1.10</i>
3.99	3.99	1.0013	Intercept	-0.000643	<i>+/-0.5</i>
2.00	2.00	0.9998			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 3, 2024	Previous Calibration:	March 20, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:14	End Time (MST):	12:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999958	<i>≥0.995</i>
9.14	9.13	1.0010	Slope	0.997354	<i>0.90 - 1.10</i>
4.57	4.60	0.9945	Intercept	0.027410	<i>+/-0.5</i>
2.28	2.34	0.9780			



NMHC Calibration Plot

Date: April 3, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	April 19, 2024	Last Cal Date:	April 3, 2024
Start time (MST):	10:12	End time (MST):	12:26
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	
CH4 Cal Gas Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
C3H8 Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
Removed C3H8 Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
Zero Air Gen model:	Teledyne API 701H	Serial Number:	953

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1331259520
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	NA	2.48E-04	NMHC SP Ratio:	NA	4.50E-05
CH4 Retention time:	NA	15.4	NMHC Peak Area:	NA	202994
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF	ON

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.03	----
High point	4920	80.2	17.13	17.15	0.999
Mid point	4960	40.1	8.56	8.62	0.993
Low point	4980	20.0	4.28	4.35	0.983
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	17.13	17.18	0.997
Average Correction Factor					0.992

Notes: Actuator was change and inlet filters before calibrator zero. Use flat baseline. Adjusted span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero					<i>Limit = 0.90-1.10</i>
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	9.14	9.12	1.002
Mid point	4960	40.1	4.57	4.59	0.997
Low point	4980	20.0	2.28	2.32	0.986
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	9.14	9.13	1.001
Average Correction Factor					0.995

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero					<i>Limit = 0.90-1.10</i>
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.03	----
High point	4920	80.2	7.99	8.02	0.996
Mid point	4960	40.1	3.99	4.03	0.990
Low point	4980	20.0	2.00	2.04	0.980
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	7.99	8.05	0.992
Average Correction Factor					0.989

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	NA	0.999013
THC Cal Offset:	NA	0.053566
CH ₄ Cal Slope:	NA	1.000152
CH ₄ Cal Offset:	NA	0.035557
NMHC Cal Slope:	NA	0.996817
NMHC Cal Offset:	NA	0.019809

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 19, 2024	Previous Calibration:	April 3, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:12	End Time (MST):	12:26
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.00	0.03	----	Correlation Coefficient	≥0.995
17.13	17.15	0.9986	Slope	0.90 - 1.10
8.56	8.62	0.9934	Intercept	+/-0.5
4.28	4.35	0.9833		





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 19, 2024	Previous Calibration:	April 3, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:12	End Time (MST):	12:26
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.03	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
7.99	8.02	0.9958	Slope	1.000152	<i>0.90 - 1.10</i>
3.99	4.03	0.9903	Intercept	0.035557	<i>+/-0.5</i>
2.00	2.04	0.9797			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 19, 2024	Previous Calibration:	April 3, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:12	End Time (MST):	12:26
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999978	<i>≥0.995</i>
9.14	9.12	1.0021	Slope	0.996817	<i>0.90 - 1.10</i>
4.57	4.59	0.9967	Intercept	0.019809	<i>+/-0.5</i>
2.28	2.32	0.9861			



NMHC Calibration Plot

Date: April 19, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	April 22, 2024	Last Cal Date:	NA
Start time (MST):	11:59	End time (MST):	14:12
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	
CH4 Cal Gas Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
C3H8 Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
Removed C3H8 Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
Zero Air Gen model:	Teledyne API 701H	Serial Number:	953

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1181490018
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	NA	2.64E-04	NMHC SP Ratio:	NA
CH4 Retention time:	NA	14.4	NMHC Peak Area:	NA
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* => +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	17.13	17.18	0.997
Mid point	4960	40.1	8.56	8.62	0.993
Low point	4980	20.0	4.28	4.38	0.977
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	17.13	17.20	0.995
Average Correction Factor					0.989

Notes: Install Calibrations. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	9.14	9.19	0.994
Mid point	4960	40.1	4.57	4.64	0.986
Low point	4980	20.0	2.28	2.36	0.969
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	9.14	9.22	0.992
Average Correction Factor					0.983

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	7.99	7.99	1.000
Mid point	4960	40.1	3.99	3.99	1.002
Low point	4980	20.0	2.00	2.02	0.988
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	7.99	7.99	1.000
Average Correction Factor					0.996

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	NA	1.001709
THC Cal Offset:	NA	0.040371
CH ₄ Cal Slope:	NA	0.998993
CH ₄ Cal Offset:	NA	0.008358
NMHC Cal Slope:	NA	1.004194
NMHC Cal Offset:	NA	0.031813

Calibration Performed By: Jan Castro and Sean Bala



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 22, 2024	Previous Calibration:	NA
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	11:59	End Time (MST):	14:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999972	<i>>0.995</i>
17.13	17.18	0.9968	Slope	1.001709	<i>0.90 - 1.10</i>
8.56	8.62	0.9931	Intercept	0.040371	<i>+/-0.5</i>
4.28	4.38	0.9775			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 22, 2024	Previous Calibration:	NA
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	11:59	End Time (MST):	14:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999985	<i>≥0.995</i>
7.99	7.99	0.9999	Slope	0.998993	<i>0.90 - 1.10</i>
3.99	3.99	1.0015	Intercept	0.008358	<i>+/-0.5</i>
2.00	2.02	0.9879			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 22, 2024	Previous Calibration:	NA
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	11:59	End Time (MST):	14:12
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999945	<i>≥0.995</i>
9.14	9.19	0.9940	Slope	1.004194	<i>0.90 - 1.10</i>
4.57	4.64	0.9859	Intercept	0.031813	<i>+/-0.5</i>
2.28	2.36	0.9685			



NMHC Calibration Plot

Date: April 22, 2024

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Conklin
 Station number: AMS 21
 Calibration Date: April 12, 2024
 Last Cal Date: March 26, 2024
 Start time (MST): 10:29
 End time (MST): 14:43
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: SA18828
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701H
 Cal Gas Expiry Date: November 3, 2031
 NO Cal Gas Conc: 48.80 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 48.80 ppm
 NO gas Diff:
 Serial Number: 3810
 Serial Number: 953

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.4	0.0	----	----
AF High point	4918	82.0	802.0	800.3	1.6	865.0	864.0	1.5	0.9267	0.9259
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 800.1 ppb		NO = 801.3 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 7.6%	
Baseline Corr 1st pt	NO _x = 865.4 ppb		NO = 864.4 ppb			<u>As Found Statistics</u>		*Percent Change	NO = 7.3%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1501663731

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.994854	0.996407
NO _x Cal Offset:	2.228020	1.908027
NO Cal Slope:	1.000020	0.998578
NO Cal Offset:	1.008029	1.088012
NO ₂ Cal Slope:	0.996734	0.999882
NO ₂ Cal Offset:	-0.362609	0.066334

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.004	0.929	NO bkgnd or offset:	10.1	9.3
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	10.2	9.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	147.3	146.2

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
High point	4918	82.0	802.0	800.3	1.6	799.8	799.6	0.2	1.0027	1.0009
Mid point	4959	41.0	401.0	400.2	0.8	402.9	401.5	1.4	0.9952	0.9967
Low point	4980	20.5	200.5	200.1	0.4	203.5	201.9	1.6	0.9851	0.9909
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
As left span	4918	82.0	802.0	387.9	414.1	797.1	387.9	409.3	1.0061	1.0000
Average Correction Factor									0.9943	0.9961

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	798.2	384.4	415.4	415.3	1.0003	100.0%
Mid GPT point	798.2	596.2	203.6	204.0	0.9982	100.2%
Low GPT point	798.2	697.8	102.0	102.1	0.9994	100.1%
Average Correction Factor					0.9993	100.1%

Notes: Sample inlet filters changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 12, 2024	Previous Calibration:	March 26, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:29	End Time (MST):	14:43
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999967	≥0.995
802.0	799.8	1.0027	Slope	0.996407	0.90 - 1.10
401.0	402.9	0.9952	Intercept	1.908027	+/-20
200.5	203.5	0.9851			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 12, 2024	Previous Calibration:	March 26, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:29	End Time (MST):	14:43
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
415.4	415.3	1.0003	Slope	0.999882	<i>0.90 - 1.10</i>
203.6	204.0	0.9982	Intercept	0.066334	<i>+/-20</i>
102.0	102.1	0.9994			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 12, 2024	Previous Calibration:	March 26, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:29	End Time (MST):	14:43
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999990	≥0.995
800.3	799.6	1.0009	Slope	0.998578	0.90 - 1.10
400.2	401.5	0.9967	Intercept	1.088012	+/-20
200.1	201.9	0.9909			



NO_x Calibration Plot

Date: April 12, 2024

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Conklin
 Station number: AMS 21
 Calibration Date: April 16, 2024
 Last Cal Date: April 12, 2024
 Start time (MST): 8:55
 End time (MST): 11:00
 Reason: As Found

Calibration Standards

NO Gas Cylinder #: SA18828
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701H
 Cal Gas Expiry Date: November 3, 2031
 NO Cal Gas Conc: 48.80 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 48.80 ppm
 NO gas Diff:
 Serial Number: 3810
 Serial Number: 953

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.5	0.0	----	----
AF High point	4918	82.0	802.0	800.3	1.6	712.5	709.0	3.4	1.1248	1.1280
AF Mid point	4959	41.0	401.0	400.2	0.8	354.5	352.1	2.5	1.1295	1.1349
AF Low point	4980	20.5	200.5	200.1	0.4	181.3	179.1	2.2	1.1027	1.1139
New cyl resp										
Previous Response	NO _x = 801.0 ppb	NO = 800.3 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -12.3%	
Baseline Corr 1st pt	NO _x = 713.0 ppb	NO = 709.5 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -12.8%	
Baseline Corr 2nd pt	NO _x = 355.0 ppb	NO = 352.6 ppb				As found	NO _x r ² : 0.999954	Nx SI: 0.887333	Nx Int: 0.627	
Baseline Corr 3rd pt	NO _x = 181.8 ppb	NO = 179.6 ppb				As found	NO r ² : 0.999967	NO SI: 0.885252	NO Int: -0.033	
						As found	NO ₂ r ² : 1.000000	NO ₂ SI: 1.002046	NO ₂ Int: 0.000	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	0.0	----	----
As found high GPT point	704.4	334.5	371.5	372.3	0.9980	100.2%
As found mid GPT point	704.4					
As found low GPT point	704.4					



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1501663731

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996407	NA
NO _x Cal Offset:	1.908027	NA
NO Cal Slope:	0.998578	NA
NO Cal Offset:	1.088012	NA
NO ₂ Cal Slope:	0.999882	NA
NO ₂ Cal Offset:	0.066334	NA

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.929	NA	NO bkgnd or offset:	9.3	NA
NOX coeff or slope:	0.993	NA	NOX bkgnd or offset:	9.4	NA
NO2 coeff or slope:	1.000	NA	Reaction cell Press:	146.2	NA

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero
 High point
 Mid point
 Low point
 As left zero
 As left span

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero
 High GPT point
 Mid GPT point
 Low GPT point

Average Correction Factor


Notes: Maintenance is done due to +/- 10% high reading. After multipoint as founds pump and charcoal was replaced. Reaction chamber was cleaned and orings replaced.

Calibration Performed By: Jan Castro

NO_x Calibration Plot

Date: April 16, 2024

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Conklin
 Station number: AMS 21
 Calibration Date: April 17, 2024
 Last Cal Date: April 12, 2024
 Start time (MST): 9:03
 End time (MST): 13:09
 Reason: Maintenance

Calibration Standards

NO Gas Cylinder #: SA18828
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701H
 Cal Gas Expiry Date: November 3, 2031
 NO Cal Gas Conc: 48.80 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 48.80 ppm
 NO gas Diff:
 Serial Number: 3810
 Serial Number: 953

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero))	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero))
As found zero									<i>Limit = 0.90 - 1.10</i>	<i>Limit = 0.90 - 1.10</i>
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = NA	ppb	NO = NA	ppb					*Percent Change	NO _x = NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb					*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb						
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb						
						* = > +/-5% change initiates investigation				
						<u>As Found Statistics</u>				
						As found	NO _x r ² :	Nx SI:		Nx Int:
						As found	NO r ² :	NO SI:		NO Int:
						As found	NO ₂ r ² :	NO ₂ SI:		NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero))	Converter Efficiency
As Found GPT zero					<i>Limit = 0.90 - 1.10</i>	<i>Limit = 96-104%</i>
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1501663731

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	NA	0.999799
NO _x Cal Offset:	NA	2.168069
NO Cal Slope:	NA	0.998620
NO Cal Offset:	NA	0.948023
NO ₂ Cal Slope:	NA	1.011507
NO ₂ Cal Offset:	NA	-1.127282

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	NA	1.098	NO bkgnd or offset:	NA	12.0
NOX coeff or slope:	NA	0.999	NOX bkgnd or offset:	NA	12.2
NO2 coeff or slope:	NA	1.000	Reaction cell Press:	NA	157.8

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
High point	4918	82.0	802.0	800.3	1.6	803.0	800.0	2.9	0.9987	1.0004
Mid point	4959	41.0	401.0	400.2	0.8	403.7	400.1	3.6	0.9933	1.0001
Low point	4980	20.5	200.5	200.1	0.4	205.2	202.3	2.9	0.9769	0.9889
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
As left span	4918	82.0	802.0	385.7	416.3	804.0	385.7	418.6	0.9975	1.0000
Average Correction Factor									0.9896	0.9965

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	792.8	380.9	413.5	417.6	0.9903	101.0%
Mid GPT point	792.8	586.4	208.0	209.1	0.9949	100.5%
Low GPT point	792.8	687.0	107.4	106.3	1.0107	98.9%
Average Correction Factor					0.9986	100.1%

Notes: Maintenance calibration. Sample inlet filters changed before calibrator zero. Adjusted zero and span. Used 2nd NO reference point because of drift.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 17, 2024	Previous Calibration:	April 12, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:03	End Time (MST):	13:09
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999962	≥0.995
802.0	803.0	0.9987	Slope	0.999799	0.90 - 1.10
401.0	403.7	0.9933	Intercept	2.168069	+/-20
200.5	205.2	0.9769			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 17, 2024	Previous Calibration:	April 12, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:03	End Time (MST):	13:09
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999970	≥0.995
413.5	417.6	0.9903	Slope	1.011507	0.90 - 1.10
208.0	209.1	0.9949	Intercept	-1.127282	+/-20
107.4	106.3	1.0107			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 17, 2024	Previous Calibration:	April 12, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:03	End Time (MST):	13:09
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999990	≥0.995
800.3	800.0	1.0004	Slope	0.998620	0.90 - 1.10
400.2	400.1	1.0001	Intercept	0.948023	+/-20
200.1	202.3	0.9889			



NO_x Calibration Plot

Date: April 17, 2024

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	April 8, 2024	Last Cal Date:	March 15, 2024
Start time (MST):	10:23	End time (MST):	13:39
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3810
Calibrator Make/Model:	Teledyne API T700	Serial Number:	691
ZAG Make/Model:	Teledyne API 701H		

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999171	0.999943	Backgd or Offset:	-1.1	-1.1
Calibration intercept:	0.320000	0.060000	Coeff or Slope:	0.998	0.998

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	0.0	----
As found High point	5000	952.3	400.0	399.9	1.000
As found Mid point					
As found Low point					
Baseline Corr As found:	399.9	Previous response	400.0	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.0	----
High point	5000	950.7	400.0	400.0	1.000
Mid point	5000	806.9	200.0	200.1	1.000
Low point	5000	704.5	100.0	100.1	0.999
As left zero	5000	800.0	0.0	0.0	----
As left span	5000	951.8	400.0	401.1	0.997
Average Correction Factor					1.000

Notes: Changed sample inlet filters after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 8, 2024	Previous Calibration:	March 15, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:23	End Time (MST):	13:39
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	1.000000	≥0.995
400.0	400.0	1.0000	Slope	0.999943	0.90 - 1.10
200.0	200.1	0.9995	Intercept	0.060000	+/- 5
100.0	100.1	0.9990			



O₃ Calibration Plot

Date: April 8, 2024

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Conklin Station number: AMS 21
 Calibration Date: April 18, 2024 Last Cal Date: March 18, 2024
 Start time (MST): 11:47 End time (MST): 12:24

Analyzer Make: API T640 S/N: 326
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-1.00	-1.36	-1.00	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	721.90	723.58	721.90	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.05	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	40.00	----	40.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.80	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.90** Expiry Date: September 29, 2024
 Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.10	11.00	11.00	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ April 18, 2024
 Date Disposable Filter Changed: _____ April 18, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ December 7, 2023
 Date RH/T Sensor Cleaned: _____ December 7, 2023

Notes: Verified flow, pressure, temperature and pump power. Leak check passed. No adjustment made.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER
APRIL 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	April 12, 2024	Last Cal Date:	March 11, 2024
Start time (MST):	11:19	End time (MST):	14:40
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.11	ppm	Cal Gas Exp Date: January 18, 2029
Cal Gas Cylinder #:	CC281519		
Removed Cal Gas Conc:	50.11	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	Teledyne API T700		Serial Number: 3806
Zero Air Gen Model:	Teledyne API T701		Serial Number: 4890

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 1152430006
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999806	1.002236	Backgd or Offset:	23.6	24.1
Calibration intercept:	0.664437	-0.036002	Coeff or Slope:	1.022	1.013

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4920	79.8	799.8	805.6	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	805.3	Previous response	800.3	*% change	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	79.8	799.8	801.5	0.998
Mid point	4960	39.9	399.9	400.9	0.997
Low point	4980	20.0	200.4	200.7	0.999
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	79.8	799.8	799.7	1.000
Average Correction Factor:					0.998

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 12, 2024	Previous Calibration:	March 11, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:19	End Time (MST):	14:40
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	1.000000
799.8	801.5	0.9979	Slope	1.002236
399.9	400.9	0.9975	Intercept	-0.036002
200.4	200.7	0.9987		
				≥0.995
				0.90 - 1.10
				+/-30



SO2 Calibration Plot

Date: April 12, 2024

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	April 18, 2024	Last Cal Date:	March 26, 2024
Start time (MST):	11:09	End time (MST):	15:32
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.02	ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	CC424047			
Removed Cal Gas Conc:	5.02	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	4890

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031
Converter make:	CDN-101	Converter serial #:	620
Analyzer Range:	0 - 100 ppb	Converter Temp:	850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.989096	0.995809	Backgd or Offset:	3.65	3.57
Calibration intercept:	0.440430	0.280545	Coeff or Slope:	1.188	1.188

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	79.7	80.0	79.7	1.004
As found Mid point	4960	39.8	40.0	40.4	0.989
As found Low point	4980	19.9	20.0	20.2	0.989
New cylinder response					
Baseline Corr As found:	79.7	Prev response:	79.59	*% change:	0.1%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	0.995519	AF Intercept:	0.240679
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999929	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	79.7	80.0	79.9	1.002
Mid point	4960	39.8	40.0	40.2	0.994
Low point	4980	19.9	20.0	20.2	0.989
As left zero	5000	0.0	0.0	0.3	----
As left span	4920	79.7	80.0	79.5	1.007
SO2 Scrubber Check	4920	79.8	798.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.995
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Scrubber test passed. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 26, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:09	End Time (MST):	15:32
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999992	≥ 0.995
80.0	79.9	1.0015	Slope	0.995809	$0.90 - 1.10$
40.0	40.2	0.9940	Intercept	0.280545	± 3
20.0	20.2	0.9891			



TRS Calibration Plot

Date: April 18, 2024

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	April 12, 2024	Last Cal Date:	March 25, 2024
Start time (MST):	11:19	End time (MST):	14:40
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH4 Cal Gas Conc.	502.8 ppm	CH4 Equiv Conc.	1075.9 ppm
C3H8 Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	502.8 ppm	CH4 Equiv Conc.	1075.9 ppm
Removed C3H8 Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
Zero Air Gen model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1317958219
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.360E-04	2.49E-04	NMHC SP Ratio:	5.52E-05	5.65E-05
CH4 Retention time:	11.6	11.8	NMHC Peak Area:	165650	162086
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	17.17	17.10	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.10	Prev response	17.15	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	17.17	17.17	1.000
Mid point	4960	39.9	8.59	8.56	1.003
Low point	4980	20.0	4.30	4.29	1.002
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	17.17	17.14	1.002
Average Correction Factor					1.002

Notes: Changed the inlet filter and H2 cylinder after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.15	9.25	0.989
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.25	Prev response	9.14	*% change	1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.15	9.16	0.999
Mid point	4960	39.9	4.57	4.58	0.999
Low point	4980	20.0	2.29	2.30	0.998
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.15	9.11	1.004
Average Correction Factor					0.998

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	8.03	7.84	1.023
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.84	Prev response	8.01	*% change	-2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.03	8.01	1.002
Mid point	4960	39.9	4.01	3.98	1.008
Low point	4980	20.0	2.01	2.00	1.007
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.03	8.03	0.999
Average Correction Factor					1.006

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000249	0.999704
THC Cal Offset:	-0.023996	-0.008399
CH ₄ Cal Slope:	1.001651	0.998349
CH ₄ Cal Offset:	-0.024958	-0.009359
NMHC Cal Slope:	0.999256	1.001105
NMHC Cal Offset:	0.000762	0.001359

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 12, 2024	Previous Calibration:	March 25, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:19	End Time (MST):	14:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999997	<i>>0.995</i>
17.17	17.17	1.0003	Slope	0.999704	<i>0.90 - 1.10</i>
8.59	8.56	1.0033	Intercept	-0.008399	<i>+/-0.5</i>
4.30	4.29	1.0022			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 12, 2024	Previous Calibration:	March 25, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:19	End Time (MST):	14:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999988 ≥0.995
8.03	8.01	1.0017	Slope	0.998349 0.90 - 1.10
4.01	3.98	1.0081	Intercept	-0.009359 +/-0.5
2.01	2.00	1.0071		





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 12, 2024	Previous Calibration:	March 25, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:19	End Time (MST):	14:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	1.000000	<i>≥0.995</i>
9.15	9.16	0.9988	Slope	1.001105	<i>0.90 - 1.10</i>
4.57	4.58	0.9986	Intercept	0.001359	<i>+/-0.5</i>
2.29	2.30	0.9976			



NMHC Calibration Plot

Date: April 12, 2024

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	April 19, 2024	Last Cal Date:	April 12, 2024
Start time (MST):	11:03	End time (MST):	14:47
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH4 Cal Gas Conc.	502.8 ppm	CH4 Equiv Conc.	1075.9 ppm
C3H8 Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	502.8 ppm	CH4 Equiv Conc.	1075.9 ppm
Removed C3H8 Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
Zero Air Gen model:	Teledyne API 701	Serial Number:	4890

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1317958219
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.49E-04	2.41E-04	NMHC SP Ratio:	5.65E-05	5.65E-05
CH4 Retention time:	11.8	11.6	NMHC Peak Area:	162086	161932
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	17.17	17.42	0.986
As found Mid point	4960	39.9	8.59	8.72	0.985
As found Low point	4980	20.0	4.30	4.37	0.985
New cylinder response					
Baseline Corr AF:	17.42	Prev response	17.16	*% change	1.5%
Baseline Corr 2nd AF:	8.72	AF Slope:	1.014275	AF Intercept:	0.002340
Baseline Corr 3rd AF:	4.37	AF Correlation:	1.000000	* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	17.17	17.19	0.999
Mid point	4960	39.9	8.59	8.52	1.007
Low point	4980	20.0	4.30	4.24	1.015
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	17.17	17.03	1.008
Average Correction Factor					1.007

Notes: Changed the actuator after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.15	9.15	1.000
As found Mid point	4960	39.9	4.57	4.60	0.994
As found Low point	4980	20.0	2.29	2.31	0.992
New cylinder response					
Baseline Corr AF:	9.15	Prev response	9.16	*% change	-0.1%
Baseline Corr 2nd AF:	4.60	AF Slope:	0.999618	AF Intercept:	0.012560
Baseline Corr 3rd AF:	2.31	AF Correlation:	0.999989	* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted
					Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.15	9.16	0.999
Mid point	4960	39.9	4.57	4.56	1.003
Low point	4980	20.0	2.29	2.27	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.15	9.08	1.007
Average Correction Factor					1.004

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	8.03	8.27	0.970
As found Mid point	4960	39.9	4.01	4.12	0.975
As found Low point	4980	20.0	2.01	2.06	0.979
New cylinder response					
Baseline Corr AF:	8.27	Prev response	8.00	*% change	3.2%
Baseline Corr 2nd AF:	4.12	AF Slope:	1.030868	AF Intercept:	-0.010820
Baseline Corr 3rd AF:	2.06	AF Correlation:	0.999991	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted
					Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.03	8.04	0.999
Mid point	4960	39.9	4.01	3.97	1.012
Low point	4980	20.0	2.01	1.97	1.021
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.03	7.95	1.010
Average Correction Factor					1.010

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999704	1.002111
THC Cal Offset:	-0.008399	-0.041995
CH ₄ Cal Slope:	0.998349	1.002620
CH ₄ Cal Offset:	-0.009359	-0.028359
NMHC Cal Slope:	1.001105	1.001665
NMHC Cal Offset:	0.001359	-0.013635

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 19, 2024	Previous Calibration:	April 12, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:03	End Time (MST):	14:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999970	<i>≥0.995</i>
17.17	17.19	0.9988	Slope	1.002111	<i>0.90 - 1.10</i>
8.59	8.52	1.0073	Intercept	-0.041995	<i>+/-0.5</i>
4.30	4.24	1.0150			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 19, 2024	Previous Calibration:	April 12, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:03	End Time (MST):	14:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999933 ≥0.995
8.03	8.04	0.9985	Slope	1.002620 0.90 - 1.10
4.01	3.97	1.0120	Intercept	-0.028359 +/-0.5
2.01	1.97	1.0209		





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 19, 2024	Previous Calibration:	April 12, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:03	End Time (MST):	14:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999990	<i>≥0.995</i>
9.15	9.16	0.9990	Slope	1.001665	<i>0.90 - 1.10</i>
4.57	4.56	1.0032	Intercept	-0.013635	<i>+/-0.5</i>
2.29	2.27	1.0099			



NMHC Calibration Plot

Date: April 19, 2024

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Janvier
 Station number: AMS 22
 Calibration Date: April 16, 2024
 Last Cal Date: March 27, 2024
 Start time (MST): 10:54
 End time (MST): 15:52
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0047765
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: March 11, 2031
 NO Cal Gas Conc: 48.80 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 48.80 ppm
 NO gas Diff:
 Serial Number: 3806
 Serial Number: 691

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	1.0	-0.1	1.0	----	----
AF High point	4918	82.0	802.0	800.3	1.6	807.8	802.6	5.2	0.9940	0.9970
AF Mid point										
AF Low point										
New cyl resp										

Previous Response NO_x = 805.1 ppb NO = 804.1 ppb * = > +/-5% change initiates investigation *Percent Change NO_x = 0.2%
 Baseline Corr 1st pt NO_x = 806.8 ppb NO = 802.7 ppb As Found Statistics *Percent Change NO = -0.2%
 Baseline Corr 2nd pt NO_x = NA ppb NO = NA ppb As found NO_x r²: Nx SI: Nx Int:
 Baseline Corr 3rd pt NO_x = NA ppb NO = NA ppb As found NO r²: NO SI: NO Int:
 As found NO₂ r²: NO₂ SI: NO₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 833

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003829	0.997901
NO _x Cal Offset:	0.024109	0.704052
NO Cal Slope:	1.005929	0.999703
NO Cal Offset:	-0.915862	0.464053
NO ₂ Cal Slope:	0.998553	0.997737
NO ₂ Cal Offset:	-0.654801	-0.534162

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.853	0.846	NO bkgnd or offset:	0.0	-0.7
NOX coeff or slope:	0.845	0.840	NOX bkgnd or offset:	-0.2	0.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	6.4	6.7

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
High point	4918	82.0	802.0	800.3	1.6	800.6	800.2	0.2	1.0017	1.0001
Mid point	4960	41.0	400.9	400.1	0.8	401.2	401.1	0.1	0.9993	0.9975
Low point	4980	20.5	200.5	200.1	0.4	201.4	200.5	0.9	0.9954	0.9978
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.4	-0.5	----	----
As left span	4918	82.0	802.0	400.1	401.9	795.7	400.1	395.6	1.0079	1.0000
Average Correction Factor									0.9988	0.9985

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	798.1	400.4	399.3	398.1	1.0031	99.7%
Mid GPT point	798.1	600.6	199.1	198.0	1.0058	99.4%
Low GPT point	798.1	698.7	101.0	99.8	1.0124	98.8%
Average Correction Factor					1.0071	99.3%

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 16, 2024	Previous Calibration:	March 27, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:54	End Time (MST):	15:52
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999996	<i>≥0.995</i>
802.0	800.6	1.0017	Slope	0.997901	<i>0.90 - 1.10</i>
400.9	401.2	0.9993	Intercept	0.704052	<i>+/-20</i>
200.5	201.4	0.9954			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 16, 2024	Previous Calibration:	March 27, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:54	End Time (MST):	15:52
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999994	≥0.995
399.3	398.1	1.0031	Slope	0.997737	0.90 - 1.10
199.1	198.0	1.0058	Intercept	-0.534162	+/-20
101.0	99.8	1.0124			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 16, 2024	Previous Calibration:	March 27, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:54	End Time (MST):	15:52
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
800.3	800.2	1.0001	Slope	0.999703	<i>0.90 - 1.10</i>
400.1	401.1	0.9975	Intercept	0.464053	<i>+/-20</i>
200.1	200.5	0.9978			



NO_x Calibration Plot

Date: April 16, 2024

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	April 15, 2024	Last Cal Date:	March 22, 2024
Start time (MST):	10:46	End time (MST):	13:40
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3806
Calibrator Make/Model:	Teledyne API T700	Serial Number:	691
ZAG Make/Model:	Teledyne API T701H		

Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #:	7046
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003829	1.003571	Backgd or Offset:	2.2	2.2
Calibration intercept:	0.480000	0.300000	Coeff or Slope:	1.027	1.027

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-1.3	----
As found High point	5000	916.2	400.0	401.0	0.994
As found Mid point					
As found Low point					
Baseline Corr As found:	402.3	Previous response	402.0	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	-0.1	----
High point	5000	916.2	400.0	401.4	0.997
Mid point	5000	763.7	200.0	201.6	0.992
Low point	5000	656.1	100.0	100.8	0.992
As left zero	5000	800.0	0.0	0.2	----
As left span	5000	916.2	400.0	403.7	0.991
Average Correction Factor					0.994

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 15, 2024	Previous Calibration:	March 22, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:46	End Time (MST):	13:40
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999993	≥0.995
400.0	401.4	0.9965	Slope	1.003571	0.90 - 1.10
200.0	201.6	0.9921	Intercept	0.300000	+/- 5
100.0	100.8	0.9921			



O₃ Calibration Plot

Date: April 15, 2024

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: April 18, 2024 Last Cal Date: March 26, 2024
 Start time (MST): 13:33 End time (MST): 15:22

Analyzer Make: Teledyne API T640 S/N: 325
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752
 Temp/RH standard: Alicat FP-25BT S/N: 388752

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	0.6	0.16	0.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	731.2	731.78	731.2	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.074	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	38	----	38	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	1.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.2	11.9	10.9	<input checked="" type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: April 18, 2024
 Date Disposable Filter Changed: April 18, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 26, 2023
 Date RH/T Sensor Cleaned: April 18, 2024

Notes: Verified flow, temperature, and pressure. Leak check passed. PMT peak voltage adjusted. Optical chamber and RH/T sensor cleaned. Disposable filter changed.

Calibration by: Rene Chamberland



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23 FORT HILLS APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Fort Hills	Station number:	AMS 23
Calibration Date:	April 11, 2024	Last Cal Date:	March 8, 2024
Start time (MST):	7:55	End time (MST):	10:48
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.76	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC281425			
Removed Cal Gas Conc:	49.76	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	451
Zero Air Gen Model:	API T701		Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1160290012
Analyzer Range:	0-1000ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003512	1.009318	Backgd or Offset:	18.5	18.6
Calibration intercept:	-0.524217	-1.305054	Coeff or Slope:	1.063	1.063

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4920	80.3	799.1	802.6	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	803.0	Previous response	801.4	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4920	80.3	799.1	805.9	0.992
Mid point	4960	40.2	400.1	401.7	0.996
Low point	4980	20.1	200.0	199.5	1.003
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	80.3	799.1	804.8	0.993
Average Correction Factor:					0.997

Notes: Zero adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 8, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:55	End Time (MST):	10:48
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999990	≥0.995
799.1	805.9	0.9916	Slope	1.009318	0.90 - 1.10
400.1	401.7	0.9959	Intercept	-1.305054	+/-30
200.0	199.5	1.0027			



SO2 Calibration Plot

Date: April 11, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Fort Hills	Station number:	AMS 23
Calibration Date:	April 12, 2024	Last Cal Date:	March 13, 2024
Start time (MST):	6:50	End time (MST):	10:53
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.20	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517372			
Removed Cal Gas Conc:	5.20	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	451
ZAG Make/Model:	API T701		Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232
Converter make:	CDN-101	Converter serial #:	594
Analyzer Range	0 - 100 ppb	Converter Temp:	750 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.991459	1.008213	Backgd or Offset:	2.0	2.1
Calibration intercept:	-0.038309	0.074071	Coeff or Slope:	1.160	1.190

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4923	77.0	80.1	77.4	1.035
As found Mid point	4962	38.5	40.0	38.5	1.040
As found Low point	4981	19.2	20.0	19.2	1.040
New cylinder response					
Baseline Corr As found:	77.4	Prev response:	79.36	*% change:	-2.5%
Baseline Corr 2nd AF pt:	38.5	AF Slope:	0.966678	AF Intercept:	-0.078827
Baseline Corr 3rd AF pt:	19.2	AF Correlation:	0.999992	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4923	77.0	80.1	80.7	0.992
Mid point	4962	38.5	40.0	40.7	0.984
Low point	4981	19.2	20.0	20.1	0.992
As left zero	5000	0.0	0.0	0.0	----
As left span	4923	77.0	80.1	83.1	0.964
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.989
Date of last converter efficiency test:	March 13, 2024			102.7%	efficiency

Notes: SOx scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	April 12, 2024	Previous Calibration:	March 13, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:50	End Time (MST):	10:53
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999974	≥ 0.995
80.1	80.7	0.9923	Slope	1.008213	$0.90 - 1.10$
40.0	40.7	0.9837	Intercept	0.074071	± 3
20.0	20.1	0.9919			



TRS Calibration Plot

Date: April 12, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Fort Hills	Station number:	AMS 23
Calibration Date:	April 11, 2024	Last Cal Date:	March 8, 2024
Start time (MST):	7:55	End time (MST):	10:47
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	500.2 ppm	CH4 Equiv Conc.	1070.6 ppm
C3H8 Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.2 ppm	CH4 Equiv Conc.	1070.6 ppm
Removed C3H8 Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
Zero Air Gen model:	API T701	Serial Number:	5611

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1193585648
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.35E-04	2.33E-04	NMHC SP Ratio:	4.93E-05	5.01E-05
CH4 Retention time:	13.2	13.2	NMHC Peak Area:	183614	182937
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	---
As found High point	4920	80.3	17.19	17.28	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.28	Prev response	17.27	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	---
High point	4920	80.3	17.19	17.20	0.999
Mid point	4960	40.2	8.61	8.64	0.996
Low point	4980	20.1	4.30	4.32	0.996
As left zero	5000	0.0	0.00	0.00	---
As left span	4920	80.3	17.19	17.18	1.001
Average Correction Factor					0.997

Notes: Nitrogen Cylinder changed. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	9.16	9.11	1.006
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.11	Prev response	9.17	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	9.16	9.15	1.001
Mid point	4960	40.2	4.59	4.66	0.985
Low point	4980	20.1	2.29	2.36	0.972
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	9.16	9.13	1.003
Average Correction Factor					0.986

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	8.03	8.17	0.983
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.17	Prev response	8.10	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	8.03	8.06	0.997
Mid point	4960	40.2	4.02	3.98	1.010
Low point	4980	20.1	2.01	1.96	1.025
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	8.03	8.04	0.999
Average Correction Factor					1.010

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.004093	1.000451
THC Cal Offset:	0.011596	0.012003
CH ₄ Cal Slope:	1.012136	1.004637
CH ₄ Cal Offset:	-0.025657	-0.032039
NMHC Cal Slope:	0.997614	0.997117
NMHC Cal Offset:	0.036452	0.043442

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 8, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:55	End Time (MST):	10:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999997	<i>≥0.995</i>
17.19	17.20	0.9994	Slope	1.000451	<i>0.90 - 1.10</i>
8.61	8.64	0.9962	Intercept	0.012003	<i>+/-0.5</i>
4.30	4.32	0.9957			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 8, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:55	End Time (MST):	10:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999924	<i>≥0.995</i>
8.03	8.06	0.9970	Slope	1.004637	<i>0.90 - 1.10</i>
4.02	3.98	1.0097	Intercept	-0.032039	<i>+/-0.5</i>
2.01	1.96	1.0249			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 8, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:55	End Time (MST):	10:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999883	<i>≥0.995</i>
9.16	9.15	1.0012	Slope	0.997117	<i>0.90 - 1.10</i>
4.59	4.66	0.9846	Intercept	0.043442	<i>+/-0.5</i>
2.29	2.36	0.9715			



NMHC Calibration Plot

Date: April 11, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Hills
 Station number: AMS 23
 Calibration Date: April 4, 2024
 Last Cal Date: March 12, 2024
 Start time (MST): 6:32
 End time (MST): 10:53
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC358149
 NOX Cal Gas Conc: 60.30 ppm
 Removed Cylinder #:
 Removed Gas NOX Conc: 60.30 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701
 Cal Gas Expiry Date: January 5, 2032
 NO Cal Gas Conc: 60.10 ppm
 Removed Gas Exp Date:
 Removed Gas NO Conc: 60.10 ppm
 NO gas Diff:
 Serial Number: 451
 Serial Number: 5611

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
AF High point	4934	66.3	799.5	796.9	2.7	777.6	770.3	7.3	1.0279	1.0344
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 798.3 ppb		NO = 794.1 ppb			<i>* = > +/-5% change initiates investigation</i>			*Percent Change	NO _x = -2.6%
Baseline Corr 1st pt	NO _x = 777.8 ppb		NO = 770.4 ppb			<u>As Found Statistics</u>			*Percent Change	NO = -3.1%
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :		Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :		NO SI:	NO Int:
						As found	NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430007

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998626	1.000499
NO _x Cal Offset:	-0.173464	0.045790
NO Cal Slope:	0.998303	1.001459
NO Cal Offset:	-1.451679	-1.552494
NO ₂ Cal Slope:	1.002264	1.003459
NO ₂ Cal Offset:	0.618019	-1.071483

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.132	1.169	NO bkgnd or offset:	3.1	3.2
NOX coeff or slope:	0.996	0.995	NOX bkgnd or offset:	3.3	3.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	177.0	177.0

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
High point	4934	66.3	799.5	796.9	2.7	799.9	797.3	2.7	0.9995	0.9995
Mid point	4967	33.2	400.4	399.0	1.3	400.9	397.3	3.6	0.9987	1.0044
Low point	4983	16.6	200.2	199.5	0.7	200.1	196.6	3.5	1.0006	1.0150
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
As left span	4934	66.3	799.5	388.0	411.5	804.5	388.0	416.4	0.9938	1.0000
Average Correction Factor									0.9996	1.0063

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	796.2	386.2	412.7	414.1	0.9965	100.4%
Mid GPT point	796.2	588.0	210.9	208.4	1.0118	98.8%
Low GPT point	796.2	690.9	108.0	107.2	1.0070	99.3%
Average Correction Factor					1.0051	99.5%

Notes:

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 12, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:32	End Time (MST):	10:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	1.000000	≥0.995
799.5	799.9	0.9995	Slope	1.000499	0.90 - 1.10
400.4	400.9	0.9987	Intercept	0.045790	+/-20
200.2	200.1	1.0006			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 12, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:32	End Time (MST):	10:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999927	≥0.995
412.7	414.1	0.9965	Slope	1.003459	0.90 - 1.10
210.9	208.4	1.0118	Intercept	-1.071483	+/-20
108.0	107.2	1.0070			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 12, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:32	End Time (MST):	10:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999980	≥0.995
796.9	797.3	0.9995	Slope	1.001459	0.90 - 1.10
399.0	397.3	1.0044	Intercept	-1.552494	+/-20
199.5	196.6	1.0150			



NO_x Calibration Plot

Date: April 4, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Hills Station number: AMS 23
 Calibration Date: April 11, 2024 Last Cal Date: March 8, 2024
 Start time (MST): 6:56 End time (MST): 7:56

Analyzer Make: API T640 S/N: 1546
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	0.4	-0.6	0.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	737.6	737.5	737.6	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.97	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	42	----	42	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	5.3	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: 10-Jun-24
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.1	11	11	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ April 11, 2024
 Date Disposable Filter Changed: _____ April 11, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ October 17, 2023
 Date RH/T Sensor Cleaned: _____ October 17, 2023

Notes: No adjustments done. Leak check passed before and after cleaning.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25 WASKŌW OHCI PIMÂTISIWIN APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number: AMS 25
Calibration Date:	April 22, 2024	Last Cal Date: March 11, 2024
Start time (MST):	7:54	End time (MST): 10:53
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	49.70	ppm	Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #:	CC342445		
Removed Cal Gas Conc:	49.70	ppm	Rem Gas Exp Date:
Removed Gas Cyl #:			Diff between cyl:
Calibrator Model:	API T700		Serial Number: 747
Zero Air Gen Model:	API T701		Serial Number: 4765

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 1118148497
Analyzer Range:	0-1000ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002469	0.998541	Backgd or Offset:	11.2	10.8
Calibration intercept:	0.048260	-0.052708	Coeff or Slope:	1.056	1.048

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4920	80.5	800.1	803.0	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	803.4	Previous response	802.1	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4920	80.5	800.1	798.9	1.001
Mid point	4960	40.2	399.6	399.0	1.001
Low point	4980	20.1	199.8	199.2	1.003
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	80.5	800.1	800.2	1.000
Average Correction Factor:					1.002

Notes: No maintenance done. Zero and span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 22, 2024	Previous Calibration:	March 11, 2024
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS 25
Start Time (MST):	7:54	End Time (MST):	10:53
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	1.000000	≥0.995
800.1	798.9	1.0015	Slope	0.998541	0.90 - 1.10
399.6	399.0	1.0014	Intercept	-0.052708	+/-30
199.8	199.2	1.0030			



SO2 Calibration Plot

Date: April 22, 2024

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number:	AMS 25
Calibration Date:	April 17, 2024	Last Cal Date:	March 18, 2024
Start time (MST):	6:15	End time (MST):	10:58
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.97 ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC517099		
Removed Cal Gas Conc:	4.97 ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:		Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	747
ZAG Make/Model:	API T701	Serial Number:	261

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146
Converter make:	Global G-150	Converter serial #:	2022-219
Analyzer Range	0 - 100 ppb	Converter Temp:	350 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008429	1.001868	Backgd or Offset:	3.3	3.2
Calibration intercept:	0.120000	0.120000	Coeff or Slope:	1.113	1.095

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	80.0	79.5	81.3	0.978
As found Mid point	4960	40.0	39.8	40.6	0.979
As found Low point	4980	20.0	19.9	20.4	0.975
New cylinder response					
Baseline Corr As found:	81.3	Prev response:	80.31	*% change:	1.2%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.021989	AF Intercept:	0.020000
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999998	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4920	80.0	79.5	79.7	0.998
Mid point	4960	40.0	39.8	40.2	0.989
Low point	4980	20.0	19.9	19.9	0.999
As left zero	5000	0.0	0.0	0.2	----
As left span	4912	88.3	800.0	819.8	0.976
SO ₂ Scrubber Check	4921	79.2	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.995
Date of last converter efficiency test:					

Notes: SO_x scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 17, 2024	Previous Calibration:	March 18, 2024
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS 25
Start Time (MST):	6:15	End Time (MST):	10:58
Analyzer make:	Global G-150	Analyzer serial #:	2022-219

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999975	≥ 0.995
79.5	79.7	0.9977	Slope	1.001868	$0.90 - 1.10$
39.8	40.2	0.9891	Intercept	0.120000	± 3
19.9	19.9	0.9990			



H₂S Calibration Plot

Date: April 17, 2024

Location: Waskow ohci Pimatisiwin





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS26 CHRISTINA LAKE APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Christina Lake	Station number:	AMS 26
Calibration Date:	April 9, 2024	Last Cal Date:	March 22, 2024
Start time (MST):	9:46	End time (MST):	12:36
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.56	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC362134			
Removed Cal Gas Conc:	49.56	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	281
Zero Air Gen Model:	API T701H		Serial Number:	832

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1152430005
Analyzer Range:	0- 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997322	0.999449	Backgd or Offset:	25.8	26.5
Calibration intercept:	0.975976	0.556005	Coeff or Slope:	0.956	0.944

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4919	80.8	800.9	805.6	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	805.3	Previous response	799.8	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4919	80.8	800.9	801.0	1.000
Mid point	4960	40.4	400.4	400.4	1.000
Low point	4980	20.2	200.2	201.5	0.994
As left zero	5000	0.0	0.0	0.1	----
As left span	4919	80.8	800.9	800.0	1.001
Average Correction Factor:					0.998

Notes: Changed sample inlet filters after as founds. Adjusted zero and span.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 9, 2024	Previous Calibration:	March 22, 2024
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	9:46	End Time (MST):	12:36
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430005

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999997	≥0.995
800.9	801.0	0.9999	Slope	0.999449	0.90 - 1.10
400.4	400.4	1.0000	Intercept	0.556005	+/-30
200.2	201.5	0.9936			



SO2 Calibration Plot

Date: April 9, 2024

Location: Christina Lake





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Christina Lake	Station number: AMS 26
Calibration Date: April 10, 2024	Last Cal Date: March 12, 2024
Start time (MST): 9:43	End time (MST): 11:42
Reason: Removal	

Calibration Standards

Cal Gas Concentration: 5.05 ppm	Cal Gas Exp Date: November 15, 2026
Cal Gas Cylinder #: DT0014831	
Removed Cal Gas Conc: 5.05 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 5258
ZAG Make/Model: API T701H	Serial Number: 832

Analyzer Information

Analyzer make: Thermo 450i	Analyzer serial #: 1180030032
Converter make: NA	Converter serial #: NA
Analyzer Range: 0 - 100 ppb	Converter Temp: 322 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996996	NA	Backgd or Offset:	34.5	NA
Calibration intercept:	0.358415	NA	Coeff or Slope:	1.064	NA

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.5	----
As found High point	4921	79.2	80.0	80.2	1.004
As found Mid point	4960	39.6	40.0	40.2	1.008
As found Low point	4980	19.8	20.0	20.0	1.026
New cylinder response					
Baseline Corr As found:	79.7	Prev response:	80.11	*% change:	-0.5%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.997997	AF Intercept:	0.298408
Baseline Corr 3rd AF pt:	19.5	AF Correlation:	0.999968	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					
SO2 Scrubber Check					
Date of last scrubber change:				Ave Corr Factor	
Date of last converter efficiency test:					


Notes: Removing instrument for upgrade. Will install new analyzer with external converter.

Calibration Performed By: Jan Castro

H2S Calibration Plot

Date: April 10, 2024

Location: Christina Lake





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Christina Lake	Station number: AMS 26
Calibration Date: April 11, 2024	Last Cal Date: NA
Start time (MST): 10:57	End time (MST): 13:53
Reason: Install	

Calibration Standards

Cal Gas Concentration: 5.05 ppm	Cal Gas Exp Date: November 15, 2026
Cal Gas Cylinder #: DT0014831	
Removed Cal Gas Conc: 5.05 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 5258
ZAG Make/Model: API T701H	Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL	Analyzer serial #: 12333331547
Converter make: Global 150	Converter serial #: 2022-196
Analyzer Range: 0 - 100 ppb	Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.004141	Backgd or Offset:	NA	1.5
Calibration intercept:	NA	-0.141605	Coeff or Slope:	NA	1.030

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4921	79.2	80.0	80.2	0.997
Mid point	4960	39.6	40.0	40.0	1.000
Low point	4980	19.8	20.0	19.9	1.005
As left zero	5000	0.0	0.0	-0.1	----
As left span	4921	79.2	80.0	80.0	1.000
SO2 Scrubber Check	4919	80.8	808.0	0.0	----
Date of last scrubber change:	11-Apr-24			Ave Corr Factor	1.001
Date of last converter efficiency test:					

Notes: Install calibrations. Changed sample inlet filters before calibrator zero. Scrubber check done after calibrator zero and passed. Adjusted zero and span.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	NA
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	10:57	End Time (MST):	13:53
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331547

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999999	≥ 0.995
80.0	80.2	0.9974	Slope	1.004141	$0.90 - 1.10$
40.0	40.0	1.0000	Intercept	-0.141605	± 3
20.0	19.9	1.0050			



H2S Calibration Plot

Date: April 11, 2024

Location: Christina Lake





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Christina Lake
 Station number: AMS 26
 Calibration Date: April 23, 2024
 Last Cal Date: March 13, 2024
 Start time (MST): 9:37
 End time (MST): 14:24
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC755290
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701H
 Cal Gas Expiry Date: January 3, 2031
 NO Cal Gas Conc: 48.70 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 48.70 ppm
 NO gas Diff:
 Serial Number: 3253
 Serial Number: 832

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.2	-0.7	----	----
AF High point	4918	82.1	802.9	799.6	3.3	808.9	803.3	5.5	0.9916	0.9952
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 803.7 ppb	NO = 799.5 ppb				<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 0.7%	
Baseline Corr 1st pt	NO _x = 809.7 ppb	NO = 803.5 ppb				<u>As Found Statistics</u>		*Percent Change	NO = 0.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found NO _x r ² :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found NO r ² :		NO SI:	NO Int:	
						As found NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1173480006

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000066	1.000009
NO _x Cal Offset:	0.766423	-1.813460
NO Cal Slope:	0.999129	0.998685
NO Cal Offset:	0.606305	-2.513223
NO ₂ Cal Slope:	1.004655	1.022718
NO ₂ Cal Offset:	0.153346	0.962157

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.318	1.329	NO bkgnd or offset:	2.5	2.6
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	3.3	2.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	162.8	162.5

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	----	----
High point	4918	82.1	802.9	799.6	3.3	802.1	797.5	4.6	1.0010	1.0027
Mid point	4959	41.1	401.9	400.3	1.6	399.0	395.4	3.6	1.0074	1.0124
Low point	4980	20.5	200.5	199.7	0.8	197.0	194.9	2.1	1.0176	1.0244
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
As left span	4918	82.1	802.9	406.0	396.9	802.7	406.0	396.7	1.0003	1.0000
Average Correction Factor									1.0087	1.0132

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	775.7	403.3	375.7	384.1	0.9781	102.2%
Mid GPT point	775.7	582.2	196.8	204.3	0.9632	103.8%
Low GPT point	775.7	682.4	96.6	99.7	0.9687	103.2%
Average Correction Factor					0.9700	103.1%

Notes: Sample inlet filters changed after as founds. Adjusted zero and span. Used 2nd NO reference point because of drift.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 23, 2024	Previous Calibration:	March 13, 2024
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	9:37	End Time (MST):	14:24
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999976	≥0.995
802.9	802.1	1.0010	Slope	1.000009	0.90 - 1.10
401.9	399.0	1.0074	Intercept	-1.813460	+/-20
200.5	197.0	1.0176			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 23, 2024	Previous Calibration:	March 13, 2024
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	9:37	End Time (MST):	14:24
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999920	<i>≥0.995</i>
375.7	384.1	0.9781	Slope	1.022718	<i>0.90 - 1.10</i>
196.8	204.3	0.9632	Intercept	0.962157	<i>+/-20</i>
96.6	99.7	0.9687			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 23, 2024	Previous Calibration:	March 13, 2024
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	9:37	End Time (MST):	14:24
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999956	≥0.995
799.6	797.5	1.0027	Slope	0.998685	0.90 - 1.10
400.3	395.4	1.0124	Intercept	-2.513223	+/-20
199.7	194.9	1.0244			



NO_x Calibration Plot

Date: April 23, 2024

Location: Christina Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27 JACKFISH 2/3 APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	April 15, 2024	Last Cal Date:	March 8, 2024
Start time (MST):	11:03	End time (MST):	14:10
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	SG9133974BAL			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	3811
Zero Air Gen Model:	API 701		Serial Number:	268

Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Serial Number:	12124313138
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003248	0.999989	Backgd or Offset:	8.7	8.4
Calibration intercept:	-1.638441	-1.697880	Coeff or Slope:	0.983	0.955

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4921	79.1	800.2	821.4	0.974
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	821.6	Previous response	801.1	*% change	2.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4921	79.1	800.2	799.2	1.001
Mid point	4961	39.5	399.5	397.6	1.005
Low point	4980	19.8	200.3	196.1	1.021
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	79.1	800.2	803.7	0.996
Average Correction Factor:					1.009

Notes: Changed the sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 15, 2024	Previous Calibration:	March 8, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:03	End Time (MST):	14:10
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999969	≥0.995
800.2	799.2	1.0012	Slope	0.999989	0.90 - 1.10
399.5	397.6	1.0049	Intercept	-1.697880	+/-30
200.3	196.1	1.0214			



SO2 Calibration Plot

Date: April 15, 2024

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Jackfish 2/3	Station number: AMS 27
Calibration Date: April 17, 2024	Last Cal Date: March 12, 2024
Start time (MST): 10:34	End time (MST): 15:17
Reason: Routine	

Calibration Standards

Cal Gas Concentration: 5.41 ppm	Cal Gas Exp Date: January 4, 2025
Cal Gas Cylinder #: CC345023	
Removed Cal Gas Conc: 5.41 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 3811
ZAG Make/Model: API 701	Serial Number: 268

Analyzer Information

Analyzer make: API T101	Analyzer serial #: 621
Converter make: NA	Converter serial #: NA
Analyzer Range: 0 - 100 ppb	Converter Temp: 316 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.014173	0.990223	Backgd or Offset:	29.9	29.9
Calibration intercept:	-0.277753	-0.338155	Coeff or Slope:	0.965	0.965

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4926	74.1	80.2	81.0	0.986
As found Mid point	4963	37.0	40.0	40.4	0.984
As found Low point	4982	18.5	20.0	19.4	1.016
New cylinder response					
Baseline Corr As found:	81.3	Prev response:	81.03	*% change:	0.3%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.016734	AF Intercept:	-0.517520
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999923	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4926	74.1	80.2	79.1	1.014
Mid point	4963	37.0	40.0	39.4	1.016
Low point	4982	18.5	20.0	19.1	1.048
As left zero	5000	0.0	0.0	0.0	----
As left span	4926	74.1	80.2	79.5	1.008
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.026
Date of last converter efficiency test:					

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 17, 2024	Previous Calibration:	March 12, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	10:34	End Time (MST):	15:17
Analyzer make:	API T101	Analyzer serial #:	621

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999938	≥ 0.995
80.2	79.1	1.0136	Slope	0.990223	$0.90 - 1.10$
40.0	39.4	1.0161	Intercept	-0.338155	± 3
20.0	19.1	1.0479			



H₂S Calibration Plot

Date: April 17, 2024

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Jackfish 2/3
 Station number: AMS 27
 Calibration Date: April 18, 2024
 Last Cal Date: March 19, 2024
 Start time (MST): 11:47
 End time (MST): 17:26
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC757838
 NOX Cal Gas Conc: 60.30 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 60.30 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701
 Cal Gas Expiry Date: January 9, 2023
 NO Cal Gas Conc: 60.20 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 60.20 ppm
 NO gas Diff:
 Serial Number: 3811
 Serial Number: 268

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
AF High point	4942	66.5	800.6	799.3	1.3	771.6	764.6	7.0	1.0372	1.0450
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 795.9 ppb		NO = 796.6 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -3.1%	
Baseline Corr 1st pt	NO _x = 771.9 ppb		NO = 764.9 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -4.1%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: API T200
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 722

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998635	1.004943
NO _x Cal Offset:	-3.655643	-4.114844
NO Cal Slope:	1.002024	0.996905
NO Cal Offset:	-4.335475	-3.695611
NO ₂ Cal Slope:	0.991424	0.980873
NO ₂ Cal Offset:	-0.986298	1.007817

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.236	1.271	NO bkgnd or offset:	0.3	0.3
NOX coeff or slope:	1.217	1.260	NOX bkgnd or offset:	1.2	1.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	3.2	3.2

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	0.5	-0.7	----	----
High point	4942	66.5	800.6	799.3	1.3	802.2	795.3	6.9	0.9980	1.0050
Mid point	4979	33.3	400.6	399.9	0.7	397.2	392.7	4.5	1.0086	1.0185
Low point	4996	16.6	199.7	199.4	0.3	192.2	191.0	1.2	1.0390	1.0438
As left zero	5000	0.0	0.0	0.0	0.0	1.1	2.2	-1.1	----	----
As left span	4942	66.5	800.6	415.6	385.0	799.6	415.6	384.0	1.0013	1.0000
Average Correction Factor									1.0152	1.0224

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.7	----	----
High GPT point	801.8	413.5	389.6	382.7	1.0181	98.2%
Mid GPT point	801.8	627.2	175.9	173.5	1.0140	98.6%
Low GPT point	801.8	714.4	88.7	90.3	0.9826	101.8%
Average Correction Factor					1.0049	99.5%

Notes: Changed the inlet filter after as founds. Adjusted span only

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 19, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:47	End Time (MST):	17:26
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999890	≥0.995
800.6	802.2	0.9980	Slope	1.004943	0.90 - 1.10
400.6	397.2	1.0086	Intercept	-4.114844	+/-20
199.7	192.2	1.0390			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 19, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:47	End Time (MST):	17:26
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.7	----	Correlation Coefficient	0.999897	≥0.995
389.6	382.7	1.0181	Slope	0.980873	0.90 - 1.10
175.9	173.5	1.0140	Intercept	1.007817	+/-20
88.7	90.3	0.9826			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 19, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:47	End Time (MST):	17:26
Analyzer make:	API T200	Analyzer serial #:	722

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999873	≥0.995
799.3	795.3	1.0050	Slope	0.996905	0.90 - 1.10
399.9	392.7	1.0185	Intercept	-3.695611	+/-20
199.4	191.0	1.0438			



NO_x Calibration Plot

Date: April 18, 2024

Location: Jackfish 2/3





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS29 SURMONT 2 APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Surmont 2	Station number:	AMS 29
Calibration Date:	April 12, 2024	Last Cal Date:	March 8, 2024
Start time (MST):	9:45	End time (MST):	14:18
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC356008			
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	5472
Zero Air Gen Model:	Teledyne API T701		Serial Number:	4698

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1170050150
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007141	1.006014	Backgd or Offset:	12.9	13.0
Calibration intercept:	-1.805499	-1.785690	Coeff or Slope:	0.939	0.939

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4919	81.3	800.1	803.0	0.996
As found Mid point	4959	40.7	400.6	400.1	1.001
As found Low point	4979	20.3	199.8	197.6	1.011
New cylinder response					
Baseline Corr As found:	803.1	Previous response	804.0	*% change	-0.1%
Baseline Corr 2nd AF pt:	400.2	AF Slope:	1.004700	AF Intercept:	-1.625629
Baseline Corr 3rd AF pt:	197.7	AF Correlation:	0.999984	<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4919	81.3	800.1	804.0	0.995
Mid point	4959	40.7	400.6	400.4	1.000
Low point	4979	20.3	199.8	197.4	1.012
As left zero	5000	0.0	0.0	0.1	----
As left span	4919	81.3	800.1	803.0	0.996
Average Correction Factor:					1.003

Notes: No maintenance done. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 12, 2024	Previous Calibration:	March 8, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:45	End Time (MST):	14:18
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999977	≥0.995
800.1	804.0	0.9952	Slope	1.006014	0.90 - 1.10
400.6	400.4	1.0005	Intercept	-1.785690	+/-30
199.8	197.4	1.0123			



SO2 Calibration Plot

Date: April 12, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Surmont 2	Station number: AMS 29
Calibration Date: April 15, 2024	Last Cal Date: March 7, 2024
Start time (MST): 10:24	End time (MST): 15:06
Reason: Routine	

Calibration Standards

Cal Gas Concentration: <u>5.391</u> ppm	Cal Gas Exp Date: January 4, 2025
Cal Gas Cylinder #: <u>CC508338</u>	
Removed Cal Gas Conc: <u>5.391</u> ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: <u>CC508338</u>	Diff between cyl:
Calibrator Make/Model: Teledyne API T700	Serial Number: 5472
ZAG Make/Model: Teledyne API T701	Serial Number: 4698

Analyzer Information

Analyzer make: Thermo 43iQ-TLE	Analyzer serial #: 1200326170
Converter make: Global	Converter serial #: 2022-220
Analyzer Range: 0 - 100 ppb	Converter Temp: 325.0 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997613	0.998614	Backgd or Offset:	0.92	0.90
Calibration intercept:	0.017447	-0.142639	Coeff or Slope:	1.074	1.050

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4925.8	74.2	80.0	82.0	0.974
As found Mid point	4962.9	37.2	40.1	41.0	0.976
As found Low point	4981.5	18.6	20.1	20.3	0.983
New cylinder response					
Baseline Corr As found:	82.1	Prev response:	79.83	*% change:	2.8%
Baseline Corr 2nd AF pt:	41.1	AF Slope:	1.026903	AF Intercept:	-0.183896
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999995	* = +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4925.8	74.2	80.0	79.8	1.003
Mid point	4962.9	37.2	40.1	39.8	1.008
Low point	4981.5	18.6	20.1	19.9	1.008
As left zero	5000	0.0	0.0	0.0	----
As left span	4925.8	74.2	80.0	79.3	1.009
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.006
Date of last converter efficiency test:					

Notes: Adjusted span. Changed sample inlet filter after as founds.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	April 15, 2024	Previous Calibration:	March 7, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:24	End Time (MST):	15:06
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999995	≥ 0.995
80.0	79.8	1.0025	Slope	0.998614	$0.90 - 1.10$
40.1	39.8	1.0077	Intercept	-0.142639	± 3
20.1	19.9	1.0077			



H2S Calibration Plot

Date: April 15, 2024

Location: Surmont 2





Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 12, 2024	Previous Calibration:	March 8, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:45	End Time (MST):	14:18
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.02	----	Correlation Coefficient	0.999979	≥0.995
17.31	17.23	1.0049	Slope	0.997593	0.90 - 1.10
8.65	8.61	1.0042	Intercept	-0.041046	+/-1.5
4.32	4.23	1.0232			



THC Calibration Plot

Date: April 12, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Surmont 2
 Station number: AMS 29
 Calibration Date: April 2, 2024
 Last Cal Date: March 6, 2024
 Start time (MST): 9:50
 End time (MST): 14:33
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T12YYFE
 NOX Cal Gas Conc: 47.46 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 47.46 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: October 30, 2024
 NO Cal Gas Conc: 47.46 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 47.46 ppm
 NO gas Diff:
 Serial Number: 5472
 Serial Number: 4698

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
AF High point	4916	84.3	800.1	800.1	0.0	790.4	789.8	0.6	1.0121	1.0128
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 799.7 ppb		NO = 799.0 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -1.2%	
Baseline Corr 1st pt	NO _x = 790.6 ppb		NO = 790.0 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -1.1%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1170050148

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000767	1.001053
NO _x Cal Offset:	-1.012410	-0.812441
NO Cal Slope:	1.000380	1.000036
NO Cal Offset:	-1.452150	-1.731680
NO ₂ Cal Slope:	0.998848	1.002282
NO ₂ Cal Offset:	0.144949	0.146547

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.414	1.429	NO bkgnd or offset:	1.4	1.4
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	1.4	1.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	174.7	173.8

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4916	84.2	799.2	799.2	0.0	799.2	797.8	1.3	1.0000	1.0017
Mid point	4958	42.1	399.6	399.6	0.0	400.1	398.7	1.4	0.9988	1.0023
Low point	4979	21.1	200.3	200.3	0.0	198.0	195.7	2.3	1.0115	1.0234
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
As left span	4916	84.2	799.2	408.8	390.4	795.1	408.8	386.3	1.0051	1.0000
Average Correction Factor									1.0034	1.0091

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	793.2	406.1	387.1	388.2	0.9972	100.3%
Mid GPT point	793.2	608.7	184.5	184.7	0.9989	100.1%
Low GPT point	793.2	701.5	91.7	92.5	0.9914	100.9%
Average Correction Factor					0.9958	100.4%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 6, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:50	End Time (MST):	14:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999988	<i>≥0.995</i>
799.2	799.2	1.0000	Slope	1.001053	<i>0.90 - 1.10</i>
399.6	400.1	0.9988	Intercept	-0.812441	<i>+/-20</i>
200.3	198.0	1.0115			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 6, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:50	End Time (MST):	14:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999996	≥0.995
387.1	388.2	0.9972	Slope	1.002282	0.90 - 1.10
184.5	184.7	0.9989	Intercept	0.146547	+/-20
91.7	92.5	0.9914			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 6, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:50	End Time (MST):	14:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999966	≥0.995
799.2	797.8	1.0017	Slope	1.000036	0.90 - 1.10
399.6	398.7	1.0023	Intercept	-1.731680	+/-20
200.3	195.7	1.0234			



NO_x Calibration Plot

Date: April 2, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Surmont 2 Station number: AMS 29
 Calibration Date: April 15, 2024 Last Cal Date: March 8, 2024
 Start time (MST): 12:56 End time (MST): 14:13

Analyzer Make: API T640 S/N: 253
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2.5	2.07	2.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	707.3	708.45	707.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.088	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	36	----	36	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.9	10.8	10.8	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 15, 2024
 Date Disposable Filter Changed: April 15, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 25, 2023
 Date RH/T Sensor Cleaned: October 25, 2023

Notes: No adjustments made.

Calibration by: Braiden Boutilier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS30 ELLS RIVER APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	April 4, 2024	Last Cal Date:	March 6, 2024
Start time (MST):	8:45	End time (MST):	11:43
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.53	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC494126			
Removed Cal Gas Conc:	50.53	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	3061
Zero Air Gen Model:	API T701H		Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1008841397
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000331	0.997233	Backgd or Offset:	9.5	9.5
Calibration intercept:	-2.815894	-2.355844	Coeff or Slope:	0.982	0.982

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4921	79.2	800.4	796.5	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	796.6	Previous response	797.8	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	79.2	800.4	797.2	1.004
Mid point	4960	39.6	400.2	394.9	1.013
Low point	4980	19.8	200.1	195.3	1.025
As left zero	5000	0.0	0.0	0.0	----
As left span	4921	79.2	800.4	797.9	1.003
Average Correction Factor:					1.014

Notes: Sample inlet filter changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 6, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:45	End Time (MST):	11:43
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999958	≥0.995
800.4	797.2	1.0040	Slope	0.997233	0.90 - 1.10
400.2	394.9	1.0135	Intercept	-2.355844	+/-30
200.1	195.3	1.0246			



SO2 Calibration Plot

Date: April 4, 2024

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	April 2, 2024	Last Cal Date:	March 11, 2024
Start time (MST):	9:37	End time (MST):	13:52
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.99	ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	CC505806			
Removed Cal Gas Conc:	4.99	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3061
ZAG Make/Model:	API T701H		Serial Number:	358

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331
Converter make:	CDN- 101	Converter serial #:	562
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002902	1.005758	Backgd or Offset:	1.6	1.6
Calibration intercept:	-0.080433	-0.080404	Coeff or Slope:	1.060	1.060

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	80.2	80.0	80.0	1.000
As found Mid point	4960	40.1	40.0	39.7	1.008
As found Low point	4980	20.0	20.0	19.8	1.008
New cylinder response					
Baseline Corr As found:	80.0	Prev response:	80.19	*% change:	-0.2%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.999763	AF Intercept:	-0.120540
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999982	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.2	80.0	80.5	0.994
Mid point	4960	40.1	40.0	40.0	1.000
Low point	4980	20.0	20.0	20.0	0.998
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	80.2	80.0	80.2	0.998
SO2 Scrubber Check	4921	79.2	792.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.998
Date of last converter efficiency test:					

Notes: Change sample inlet filters after multipoint as founds. Sox scrubber check done after calibrator zero and passed. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	April 2, 2024	Previous Calibration:	March 11, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:37	End Time (MST):	13:52
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999988	≥ 0.995
80.0	80.5	0.9942	Slope	1.005758	$0.90 - 1.10$
40.0	40.0	1.0005	Intercept	-0.080404	± 3
20.0	20.0	0.9980			



TRS Calibration Plot

Date: April 2, 2024

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	April 4, 2024	Last Cal Date:	March 28, 2024
Start time (MST):	8:45	End time (MST):	11:43
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH4 Cal Gas Conc.	499.7 ppm	CH4 Equiv Conc.	1075.0 ppm
C3H8 Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	499.7 ppm	CH4 Equiv Conc.	1075.0 ppm
Removed C3H8 Conc.	209.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
Zero Air Gen model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1152430011
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.07E-04	3.07E-04	NMHC SP Ratio:	6.43E-05	6.43E-05
CH4 Retention time:	17.4	17.4	NMHC Peak Area:	141658	141658
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	---
As found High point	4921	79.2	17.03	17.16	0.992
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.16	Prev response	17.04	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	---
High point	4921	79.2	17.03	17.15	0.993
Mid point	4960	39.6	8.51	8.51	1.001
Low point	4980	19.8	4.26	4.20	1.013
As left zero	5000	0.0	0.00	0.00	---
As left span	4921	79.2	17.03	17.18	0.991
Average Correction Factor					1.002

Notes: Sample inlet filters changed after as founds. No adjustment made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.2	9.11	9.30	0.980
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.30	Prev response	9.15	47116	
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.2	9.11	9.29	0.981
Mid point	4960	39.6	4.56	4.62	0.987
Low point	4980	19.8	2.28	2.28	0.998
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.2	9.11	9.30	0.980
Average Correction Factor					0.989

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.2	7.91	7.86	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.86	Prev response	7.89	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.2	7.91	7.86	1.007
Mid point	4960	39.6	3.96	3.89	1.017
Low point	4980	19.8	1.98	1.92	1.030
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.2	7.91	7.88	1.005
Average Correction Factor					1.018

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002121	1.008470
THC Cal Offset:	-0.024340	-0.047141
CH ₄ Cal Slope:	1.000048	0.994402
CH ₄ Cal Offset:	-0.029357	-0.025556
NMHC Cal Slope:	1.003946	1.020664
NMHC Cal Offset:	0.005416	-0.021985

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 28, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:45	End Time (MST):	11:43
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999965	<i>≥0.995</i>
17.03	17.15	0.9928	Slope	1.008470	<i>0.90 - 1.10</i>
8.51	8.51	1.0006	Intercept	-0.047141	<i>+/-0.5</i>
4.26	4.20	1.0129			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 28, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:45	End Time (MST):	11:43
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999950	<i>≥0.995</i>
7.91	7.86	1.0070	Slope	0.994402	<i>0.90 - 1.10</i>
3.96	3.89	1.0172	Intercept	-0.025556	<i>+/-0.5</i>
1.98	1.92	1.0301			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 4, 2024	Previous Calibration:	March 28, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:45	End Time (MST):	11:43
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999974	<i>≥0.995</i>
9.11	9.29	0.9809	Slope	1.020664	<i>0.90 - 1.10</i>
4.56	4.62	0.9869	Intercept	-0.021985	<i>+/-0.5</i>
2.28	2.28	0.9984			



NMHC Calibration Plot

Date: April 4, 2024

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Ells River
 Station number: AMS 30
 Calibration Date: April 5, 2024
 Last Cal Date: March 8, 2024
 Start time (MST): 9:44
 End time (MST): 14:16
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0027487
 NOX Cal Gas Conc: 59.30 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 59.30 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701H
 Cal Gas Expiry Date: January 9, 2032
 NO Cal Gas Conc: 59.10 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 59.10 ppm
 NO gas Diff:
 Serial Number: 3061
 Serial Number: 358

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
AF High point	4932	67.7	803.0	800.3	2.7	794.6	792.1	2.5	1.0104	1.0100
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 802.0 ppb		NO = 800.8 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -0.9%	
Baseline Corr 1st pt	NO _x = 794.7 ppb		NO = 792.3 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -1.1%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 710321429

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999616	0.996856
NO _x Cal Offset:	-0.638284	-0.719188
NO Cal Slope:	1.002443	0.999360
NO Cal Offset:	-1.418930	-1.439899
NO ₂ Cal Slope:	0.996723	0.994435
NO ₂ Cal Offset:	-0.460536	0.194273

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.182	1.195	NO bkgnd or offset:	13.9	14.0
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	13.8	13.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	192.4	193.6

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
High point	4932	67.7	803.0	800.3	2.7	800.1	799.1	1.0	1.0036	1.0015
Mid point	4966	33.8	400.9	399.5	1.4	398.6	396.9	1.7	1.0057	1.0066
Low point	4983	16.9	200.4	199.8	0.7	198.2	196.9	1.3	1.0113	1.0145
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.1	----	----
As left span	4932	67.7	803.0	419.1	383.9	801.0	419.1	382.1	1.0025	1.0000
Average Correction Factor									1.0069	1.0075

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	797.0	417.3	382.4	380.4	1.0053	99.5%
Mid GPT point	797.0	613.3	186.4	185.7	1.0038	99.6%
Low GPT point	797.0	703.4	96.3	96.0	1.0032	99.7%
Average Correction Factor					1.0041	99.6%

Notes: Sample inlet filters changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 5, 2024	Previous Calibration:	March 8, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:44	End Time (MST):	14:16
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999995	<i>≥0.995</i>
803.0	800.1	1.0036	Slope	0.996856	<i>0.90 - 1.10</i>
400.9	398.6	1.0057	Intercept	-0.719188	<i>+/-20</i>
200.4	198.2	1.0113			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 5, 2024	Previous Calibration:	March 8, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:44	End Time (MST):	14:16
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	1.000000	≥0.995
382.4	380.4	1.0053	Slope	0.994435	0.90 - 1.10
186.4	185.7	1.0038	Intercept	0.194273	+/-20
96.3	96.0	1.0032			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 5, 2024	Previous Calibration:	March 8, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:44	End Time (MST):	14:16
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999985	<i>≥0.995</i>
800.3	799.1	1.0015	Slope	0.999360	<i>0.90 - 1.10</i>
399.5	396.9	1.0066	Intercept	-1.439899	<i>+/-20</i>
199.8	196.9	1.0145			



NO_x Calibration Plot

Date: April 5, 2024

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Ells River Station number: AMS 30
 Calibration Date: April 2, 2024 Last Cal Date: March 8, 2024
 Start time (MST): 10:11 End time (MST): 12:02

Analyzer Make: API T640 S/N: 875
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	6.50	6.33	6.50	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	726.80	728.54	726.80	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.43	4.50	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	100.00	----	39.00	<input checked="" type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	5.70	PM w/ HEPA: _____	1.60	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: September 29, 2024
 Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	6.00	10.60	10.60	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 2, 2024
 Date Disposable Filter Changed: January 22, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 27, 2023
 Date RH/T Sensor Cleaned: February 23, 2024

Notes: Verified flow, temperature and pressure. Pump changed. Leak check passed. No adjustment made.

Calibration by: Jan Castro



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Ells River Station number: AMS 30
 Calibration Date: April 4, 2024 Last Cal Date: April 2, 2024
 Start time (MST): 11:57 End time (MST): 12:12

Analyzer Make: API T640 S/N: 875
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2.60	1.89	2.60	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	737.40	739.59	737.40	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.03	5.02	5.03	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39.00	----	39.00	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	4.80	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: September 29, 2024
 Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 2, 2024
 Date Disposable Filter Changed: April 4, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 27, 2023
 Date RH/T Sensor Cleaned: February 23, 2024

Notes: Verified flow, temperature and pressure. Disposable filter changed. Leak check passed. No adjustment made.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS505 SAWBONES BAY APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Sawbones Bay	Station number:	AMS 505
Calibration Date:	April 17, 2024	Last Cal Date:	March 18, 2024
Start time (MST):	9:48	End time (MST):	12:35
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	51.40	ppm	Cal Gas Exp Date:	February 15, 2029
Cal Gas Cylinder #:	EY0000672			
Removed Cal Gas Conc:	51.40	ppm	Rem Gas Exp Date:	February 15, 2029
Removed Gas Cyl #:	EY0000672		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	5112
Zero Air Gen Model:	Teledyne API T701H		Serial Number:	690

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	710321323
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999393	1.008070	Backgd or Offset:	20.5	20.3
Calibration intercept:	-1.051659	-1.013890	Coeff or Slope:	1.008	1.001

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4922	77.8	799.8	811.7	0.985
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	812.0	Previous response	798.3	*% change	1.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4922	77.8	799.8	806.1	0.992
Mid point	4961	38.9	399.9	400.9	0.998
Low point	4981	19.5	200.4	200.1	1.002
As left zero	5000	0.0	0.0	0.1	----
As left span	4922	77.8	799.8	805.3	0.993
Average Correction Factor:					0.997

Notes: Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 17, 2024	Previous Calibration:	March 18, 2024
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:48	End Time (MST):	12:35
Analyzer make:	Thermo 43i	Analyzer serial #:	710321323

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999986	≥0.995
799.8	806.1	0.9922	Slope	1.008070	0.90 - 1.10
399.9	400.9	0.9975	Intercept	-1.013890	+/-30
200.4	200.1	1.0017			



SO2 Calibration Plot

Date: April 17, 2024

Location: Sawbones Bay





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Sawbones Bay	Station number: AMS 505
Calibration Date: April 24, 2024	Last Cal Date: March 19, 2024
Start time (MST): 8:45	End time (MST): 12:22
Reason: Routine	

Calibration Standards

Cal Gas Concentration: 5.15 ppm	Cal Gas Exp Date: February 5, 2024
Cal Gas Cylinder #: CC517397	
Removed Cal Gas Conc: 5.15 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: Teledyne API T700	Serial Number: 5112
ZAG Make/Model: Teledyne API T701	Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43iQ	Analyzer serial #: 12113311965
Converter make: Global 150	Converter serial #: 2022-224
Analyzer Range: 0 - 100 ppb	Converter Temp: degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.015068	0.998937	Backgd or Offset:	1.010	0.990
Calibration intercept:	0.002254	-0.158258	Coeff or Slope:	1.119	1.092

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4922	77.7	80.0	82.6	0.968
As found Mid point	4961	38.8	40.0	40.8	0.977
As found Low point	4981	19.4	20.0	20.1	0.989
New cylinder response					
Baseline Corr As found:	82.7	Prev response:	81.24	*% change:	1.8%
Baseline Corr 2nd AF pt:	40.9	AF Slope:	1.034635	AF Intercept:	-0.357524
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999954	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4922	77.7	80.0	79.9	1.002
Mid point	4961	38.8	40.0	39.6	1.009
Low point	4981	19.4	20.0	19.7	1.014
As left zero	5000	0.0	0.0	0.0	----
As left span	4922	77.7	80.0	79.7	1.004
SO2 Scrubber Check	4922	77.8	778.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.008
Date of last converter efficiency test:					

Notes: Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	April 24, 2024	Previous Calibration:	March 19, 2024
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:45	End Time (MST):	12:22
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12113311965

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999979	≥ 0.995
80.0	79.9	1.0017	Slope	0.998937	$0.90 - 1.10$
40.0	39.6	1.0092	Intercept	-0.158258	± 3
20.0	19.7	1.0142			



H2S Calibration Plot

Date: April 24, 2024

Location: Sawbones Bay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Sawbones Bay
 Station number: AMS 505
 Calibration Date: April 11, 2024
 Last Cal Date: March 20, 2024
 Start time (MST): 8:44
 End time (MST): 13:23
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T1FY3PK
 NOX Cal Gas Conc: 47.94 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 47.94 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701H
 Cal Gas Expiry Date: March 14, 2025
 NO Cal Gas Conc: 47.94 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 47.94 ppm
 NO gas Diff:
 Serial Number: 5112
 Serial Number: 690

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	1.0	0.5	0.6	----	----
AF High point	4917	83.4	799.6	799.6	0.0	805.0	801.2	3.9	0.9945	0.9986
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 799.3 ppb	NO = 798.5 ppb	<i>* = > +/-5% change initiates investigation</i>				*Percent Change	NO _x = 0.6%		
Baseline Corr 1st pt	NO _x = 804.0 ppb	NO = 800.7 ppb	<u>As Found Statistics</u>				*Percent Change	NO = 0.3%		
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :		Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :		NO SI:	NO Int:			
			As found	NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: API T200
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 4260

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001550	1.001007
NO _x Cal Offset:	-1.470566	-0.830518
NO Cal Slope:	1.000349	1.002164
NO Cal Offset:	-1.350176	-1.110354
NO ₂ Cal Slope:	1.002397	0.998476
NO ₂ Cal Offset:	0.386621	0.748401

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.045	1.044	NO bkgnd or offset:	-1.1	-0.2
NOX coeff or slope:	1.044	1.039	NOX bkgnd or offset:	-0.1	0.7
NO2 coeff or slope:	NA	NA	Reaction cell Press:	8.0	8.1

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.2	----	----
High point	4917	83.4	799.6	799.6	0.0	799.6	800.3	-0.6	1.0000	0.9991
Mid point	4958	41.7	399.8	399.8	0.0	400.0	400.5	-0.5	0.9996	0.9984
Low point	4979	20.9	200.4	200.4	0.0	198.4	197.5	0.9	1.0100	1.0146
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.2	-0.3	----	----
As left span	4916	83.4	799.7	341.6	458.0	796.6	341.6	455.0	1.0039	1.0000
Average Correction Factor									1.0032	1.0040

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	796.8	340.4	456.4	456.0	1.0009	99.9%
Mid GPT point	796.8	549.6	247.2	247.6	0.9984	100.2%
Low GPT point	796.8	651.3	145.5	147.4	0.9871	101.3%
Average Correction Factor					0.9955	100.5%

Notes: Changed inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 20, 2024
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:44	End Time (MST):	13:23
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
799.6	799.6	1.0000	Slope	1.001007	<i>0.90 - 1.10</i>
399.8	400.0	0.9996	Intercept	-0.830518	<i>+/-20</i>
200.4	198.4	1.0100			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 20, 2024
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:44	End Time (MST):	13:23
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999973	<i>≥0.995</i>
456.4	456.0	1.0009	Slope	0.998476	<i>0.90 - 1.10</i>
247.2	247.6	0.9984	Intercept	0.748401	<i>+/-20</i>
145.5	147.4	0.9871			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 11, 2024	Previous Calibration:	March 20, 2024
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:44	End Time (MST):	13:23
Analyzer make:	API T200	Analyzer serial #:	4260

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999979	<i>≥0.995</i>
799.6	800.3	0.9991	Slope	1.002164	<i>0.90 - 1.10</i>
399.8	400.5	0.9984	Intercept	-1.110354	<i>+/-20</i>
200.4	197.5	1.0146			



NO_x Calibration Plot

Date: April 11, 2024

Location: Sawbones Bay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS507 KIRBY SOUTH APRIL 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 31, 2024



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Kirby South	Station number:	AMS 507
Calibration Date:	April 18, 2024	Last Cal Date:	March 14, 2024
Start time (MST):	7:45	End time (MST):	11:13
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.18	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC303554</u>			
Removed Cal Gas Conc:	49.18	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	3804
Zero Air Gen Model:	Teledyne API T701H		Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1173410001
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995804	1.000467	Backgd or Offset:	21.7	25.5
Calibration intercept:	0.311657	-1.470231	Coeff or Slope:	0.906	0.906

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	3.4	----
As found High point	4919	81.3	799.6	801.6	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	798.2	Previous response	796.6	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	4919	81.3	799.6	798.8	1.001
Mid point	4959	40.7	400.3	399.4	1.002
Low point	4980	20.3	199.7	196.5	1.016
As left zero	5000	0.0	0.0	-0.2	----
As left span	4919	81.3	799.6	801.0	0.998
Average Correction Factor:					1.006

Notes: Changed sample inlet filter after as founds. Adjusted zero.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 14, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	7:45	End Time (MST):	11:13
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999986	≥0.995
799.6	798.8	1.0010	Slope	1.000467	0.90 - 1.10
400.3	399.4	1.0024	Intercept	-1.470231	+/-30
199.7	196.5	1.0161			



SO2 Calibration Plot

Date: April 18, 2024

Location: Kirby South





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name:	Kirby South	Station number:	AMS 507
Calibration Date:	April 17, 2024	Last Cal Date:	March 13, 2024
Start time (MST):	11:15	End time (MST):	16:07
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.05	ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	<u>DT0019762</u>			
Removed Cal Gas Conc:	5.05	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T750		Serial Number:	281
ZAG Make/Model:	Teledyne API T751		Serial Number:	321

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012
Converter make:	Global	Converter serial #:	2022-197
Analyzer Range	0 - 100 ppb	Converter Temp:	350 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996813	0.999242	Backgd or Offset:	1.73	1.72
Calibration intercept:	0.139039	-0.020960	Coeff or Slope:	1.041	1.041

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4921	79.2	80.0	81.1	0.985
As found Mid point	4960	39.6	40.0	40.0	0.997
As found Low point	4980	19.8	20.0	20.0	0.995
New cylinder response	5000	0.0	0.0		
Baseline Corr As found:	81.2	Prev response:	79.88	*% change:	1.6%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	1.015243	AF Intercept:	-0.280964
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999953	<i>* = > +/-5% change initiates investigation</i>	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	79.2	80.0	79.9	1.001
Mid point	4960	39.6	40.0	40.0	1.000
Low point	4980	19.8	20.0	19.9	1.005
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	79.2	80.0	78.2	1.023
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:		July 25, 2023		Ave Corr Factor	1.002
Date of last converter efficiency test:		NA			

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	April 17, 2024	Previous Calibration:	March 13, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	11:15	End Time (MST):	16:07
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999998	≥ 0.995
80.0	79.9	1.0012	Slope	0.999242	$0.90 - 1.10$
40.0	40.0	1.0000	Intercept	-0.020960	± 3
20.0	19.9	1.0050			



H2S Calibration Plot

Date: April 17, 2024

Location: Kirby South





Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 18, 2024	Previous Calibration:	March 14, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	7:45	End Time (MST):	11:13
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.05	----	Correlation Coefficient	0.999989	≥0.995
17.26	17.23	1.0019	Slope	1.001763	0.90 - 1.10
8.64	8.57	1.0085	Intercept	-0.076111	+/-1.5
4.31	4.21	1.0229			



THC Calibration Plot

Date: April 18, 2024

Location: Kirby South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Kirby South
 Station number: AMS 507
 Calibration Date: April 17, 2024
 Last Cal Date: March 15, 2024
 Start time (MST): 11:05
 End time (MST): 16:11
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T34ULGL
 NOX Cal Gas Conc: 49.39 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 49.39 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: March 8, 2025
 NO Cal Gas Conc: 49.02 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 49.02 ppm
 NO gas Diff:
 Serial Number: 3804
 Serial Number: 880

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.3	-0.7	-0.6	----	----
AF High point	4919	81.0	800.1	794.1	6.0	832.0	826.0	6.2	0.9602	0.9606
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 796.1 ppb	NO = 790.9 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 4.5%
Baseline Corr 1st pt	NO _x = 833.3 ppb	NO = 826.7 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 4.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12400232071

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000646	1.000289
NO _x Cal Offset:	-4.504120	-4.604106
NO Cal Slope:	1.002372	1.001623
NO Cal Offset:	-5.105630	-5.445289
NO ₂ Cal Slope:	0.996340	0.999497
NO ₂ Cal Offset:	0.258912	1.161308

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.097	1.057	NO bkgnd or offset:	1.8	1.1
NOX coeff or slope:	0.996	0.998	NOX bkgnd or offset:	2.5	1.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	196.86	197.21

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4919	81.0	800.1	794.1	6.0	798.6	793.7	4.9	1.0019	1.0005
Mid point	4960	40.5	400.1	397.1	3.0	391.6	386.5	5.1	1.0216	1.0273
Low point	4980	20.2	199.5	198.0	1.5	191.5	189.5	2.1	1.0420	1.0451
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.3	-0.1	----	----
As left span	4919	81.0	800.1	408.3	391.8	802.0	408.3	393.8	0.9977	1.0000
Average Correction Factor									1.0218	1.0243

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	787.3	403.7	389.6	389.9	0.9992	100.1%
Mid GPT point	787.3	619.2	174.1	176.3	0.9875	101.3%
Low GPT point	787.3	706.5	86.8	88.6	0.9796	102.1%
Average Correction Factor					0.9888	101.1%

Notes: Changed sample inlet filters after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 17, 2024	Previous Calibration:	March 15, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	11:05	End Time (MST):	16:11
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999837	<i>≥0.995</i>
800.1	798.6	1.0019	Slope	1.000289	<i>0.90 - 1.10</i>
400.1	391.6	1.0216	Intercept	-4.604106	<i>+/-20</i>
199.5	191.5	1.0420			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 17, 2024	Previous Calibration:	March 15, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	11:05	End Time (MST):	16:11
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999958	≥0.995
389.6	389.9	0.9992	Slope	0.999497	0.90 - 1.10
174.1	176.3	0.9875	Intercept	1.161308	+/-20
86.8	88.6	0.9796			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 17, 2024	Previous Calibration:	March 15, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	11:05	End Time (MST):	16:11
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999745	≥0.995
794.1	793.7	1.0005	Slope	1.001623	0.90 - 1.10
397.1	386.5	1.0273	Intercept	-5.445289	+/-20
198.0	189.5	1.0451			



NO_x Calibration Plot

Date: April 17, 2024

Location: Kirby South





End of Report