



**WOOD BUFFALO  
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive  
Fort McMurray, AB T9K 0K4  
P: 780.799.4420 E: info@wbea.org  
[wbea.org](http://wbea.org)

Wood Buffalo Environmental Association

# MAY 2024

# MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

June 28, 2024

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS01 BERTHA GANTER - FORT MCKAY MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number: AMS 01
Calibration Date:	May 16, 2024	Last Cal Date: April 3, 2024
Start time (MST):	9:22	End time (MST): 12:35
Reason:	Routine	

### Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #:	CC418809		
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	Teledyne API T700		Serial Number: 3565
Zero Air Gen Model:	Teledyne API T701		Serial Number: 4890

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: JC1501301448
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001305	1.000564	Backgd or Offset:	20.8	20.7
Calibration intercept:	-0.794719	-1.035075	Coeff or Slope:	0.892	0.881

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	1.0	----
As found High point	4918	81.3	800.3	808.6	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	807.6	Previous response	800.5	*% change	0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4918	81.3	800.3	800.6	1.000
Mid point	4959	40.7	400.6	398.3	1.006
Low point	4979	20.3	199.8	198.2	1.008
As left zero	5000	0.0	0.0	0.4	----
As left span	4918	81.3	800.3	800.2	1.000
Average Correction Factor:					1.005

Notes:                   Calibrating remotely. Adjusted span only.

Calibration Performed By:                   Rene Chamberland



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 16, 2024	Previous Calibration:	April 3, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:22	End Time (MST):	12:35
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999985	<b>≥0.995</b>
800.3	800.6	0.9996	Slope	1.000564	<b>0.90 - 1.10</b>
400.6	398.3	1.0058	Intercept	-1.035075	<b>+/-30</b>
199.8	198.2	1.0082			



SO2 Calibration Plot

Date: May 16, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	May 13, 2024	Last Cal Date:	April 9, 2024
Start time (MST):	10:27	End time (MST):	15:00
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.10 ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511749		
Removed Cal Gas Conc:	5.10 ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	3565
ZAG Make/Model:	Teledyne API T701	Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461
Converter make:	CD Nova	Converter serial #:	470
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001846	0.996559	Backgd or Offset:	2.42	2.39
Calibration intercept:	0.119998	0.179997	Coeff or Slope:	0.918	0.903

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4921	78.4	80.0	79.8	1.003
As found Mid point	4960	39.2	40.0	40.2	0.997
As found Low point	4980	19.6	20.0	20.2	0.995
New cylinder response					
Baseline Corr As found:	79.7	Prev response:	80.25	*% change:	-0.7%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	0.996130	AF Intercept:	0.219998
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999986	<i>* = &gt; +/-5% change initiates investigation</i>	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4921	78.4	80.0	79.8	1.002
Mid point	4960	39.2	40.0	40.2	0.995
Low point	4980	19.6	20.0	20.1	0.995
As left zero	5000	0.0	0.0	0.3	----
As left span	4921	78.4	80.0	79.6	1.005
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	<b>0.997</b>
Date of last converter efficiency test:					

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association


## TRS Calibration Summary

### Station Information

Calibration Date:	May 13, 2024	Previous Calibration:	April 9, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:27	End Time (MST):	15:00
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999988	$\geq 0.995$
80.0	79.8	1.0022	Slope	0.996559	$0.90 - 1.10$
40.0	40.2	0.9948	Intercept	0.179997	$\pm 3$
20.0	20.1	0.9947			



TRS Calibration Plot

Date: May 13, 2024

Location: Bertha Ganter-Fort McKay







# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	May 13, 2024	Last Cal Date:	April 9, 2024
Start time (MST):	10:27	End time (MST):	15:00
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.10 ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511749		
Removed Cal Gas Conc:	5.10 ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	3565
ZAG Make/Model:	Teledyne API T701	Serial Number:	4890

### Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1200326167
Converter make:	CD Nova	Converter serial #:	2022-221
Analyzer Range	0 - 100 ppb	Converter Temp:	350 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002768	1.002910	Backgd or Offset:	2.03	2.01
Calibration intercept:	0.016774	-0.063214	Coeff or Slope:	0.985	0.974

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4922	78.4	80.0	80.8	0.987
As found Mid point	4960	39.2	40.0	40.6	0.980
As found Low point	4980	19.6	20.0	19.9	0.995
New cylinder response					
Baseline Corr As found:	81.0	Prev response:	80.20	*% change:	1.0%
Baseline Corr 2nd AF pt:	40.8	AF Slope:	1.014059	AF Intercept:	-0.203262
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999972	<i>* = +/-5% change initiates investigation</i>	

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4922	78.4	80.0	80.1	0.998
Mid point	4960	39.2	40.0	40.1	0.997
Low point	4980	19.6	20.0	20.0	1.000
As left zero	5000	0.0	0.0	0.0	----
As left span	4922	78.4	80.0	79.9	1.001
SO2 Scrubber Check	4919	81.3	813.0	0.1	----
Date of last scrubber change:	January 25, 2024			Ave Corr Factor	0.998
Date of last converter efficiency test:					

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association


## H<sub>2</sub>S Calibration Summary

### Station Information

Calibration Date:	May 13, 2024	Previous Calibration:	April 9, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:27	End Time (MST):	15:00
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1200326167

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999998	≥0.995
80.0	80.1	0.9983	Slope	1.002910	0.90 - 1.10
40.0	40.1	0.9973	Intercept	-0.063214	+/-3
20.0	20.0	0.9997			



H<sub>2</sub>S Calibration Plot

Date: May 13, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	May 16, 2024	Last Cal Date:	April 3, 2024
Start time (MST):	9:22	End time (MST):	12:35
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	497.2 ppm	CH4 Equiv Conc.	1061.8 ppm
C3H8 Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.2 ppm	CH4 Equiv Conc.	1061.8 ppm
Removed C3H8 Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
Zero Air Gen model:	Teledyne API T701	Serial Number:	4890

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320040
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.31E-04	4.36E-04	NMHC SP Ratio:	7.53E-05	7.78E-05
CH4 Retention time:	16.7	16.7	NMHC Peak Area:	122095	118187
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	17.27	16.83	1.026
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.83	Prev response	17.25	*% change	-2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	17.27	17.24	1.002
Mid point	4959	40.7	8.64	8.49	1.019
Low point	4979	20.3	4.31	4.21	1.023
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	17.27	17.09	1.010
Average Correction Factor					1.015

Notes: Calibrating remotely. N2 cylinder was changed after as founds. Adjusted span only.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	9.18	8.88	1.034
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.88	Prev response	9.21	*% change	-3.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	9.18	9.18	1.000
Mid point	4959	40.7	4.60	4.54	1.012
Low point	4979	20.3	2.29	2.28	1.005
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	9.18	9.05	1.014
Average Correction Factor					1.006

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	8.09	7.95	1.018
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.95	Prev response	8.05	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	8.09	8.06	1.003
Mid point	4959	40.7	4.05	3.94	1.026
Low point	4979	20.3	2.02	1.93	1.046
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	8.09	8.04	1.005
Average Correction Factor					1.025

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001141	0.999301
THC Cal Offset:	-0.034037	-0.065629
CH <sub>4</sub> Cal Slope:	1.000742	0.998751
CH <sub>4</sub> Cal Offset:	-0.042118	-0.050322
NMHC Cal Slope:	1.001953	0.999822
NMHC Cal Offset:	0.007481	-0.015708

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 16, 2024	Previous Calibration:	April 3, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:22	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999906	≥0.995
17.27	17.24	1.0016	Slope	0.999301	0.90 - 1.10
8.64	8.49	1.0187	Intercept	-0.065629	+/-0.5
4.31	4.21	1.0234			





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 16, 2024	Previous Calibration:	April 3, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:22	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999800	≥0.995
8.09	8.06	1.0034	Slope	0.998751	0.90 - 1.10
4.05	3.94	1.0262	Intercept	-0.050322	+/-0.5
2.02	1.93	1.0455			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 16, 2024	Previous Calibration:	April 3, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:22	End Time (MST):	12:35
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999955	<span style="color: red;">≥0.995</span>
9.18	9.18	0.9999	Slope	0.999822	<span style="color: red;">0.90 - 1.10</span>
4.60	4.54	1.0121	Intercept	-0.015708	<span style="color: red;">+/-0.5</span>
2.29	2.28	1.0050			






NMHC Calibration Plot

Date: May 16, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Bertha Ganter-Fort McKay  
 Station number: AMS 01  
 Calibration Date: May 8, 2024  
 Last Cal Date: April 8, 2024  
 Start time (MST): 10:11  
 End time (MST): 13:06  
 Reason: Removal

### Calibration Standards

NO Gas Cylinder #: CC335700  
 NOX Cal Gas Conc: 59.40 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 59.40 ppm  
 NOX gas Diff:  
 Calibrator Model: Teledyne API T700  
 ZAG make/model: Teledyne API T701  
 Cal Gas Expiry Date: September 1, 2032  
 NO Cal Gas Conc: 59.20 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 59.20 ppm  
 NO gas Diff:  
 Serial Number: 3565  
 Serial Number: 4890

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	26.4	26.4	0.0	----	----
AF High point	4932	67.6	803.1	800.4	2.7	820.8	811.2	9.6	1.0109	1.0199
AF Mid point	4966	33.8	401.5	400.2	1.4	422.7	417.4	5.3	1.0132	1.0235
AF Low point	4983	16.9	200.8	200.1	0.7	224.6	220.7	3.9	1.0130	1.0298

New cyl resp

Previous Response	NO <sub>x</sub> = 802.8 ppb	NO = 799.3 ppb	<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -1.1%
Baseline Corr 1st pt	NO <sub>x</sub> = 794.4 ppb	NO = 784.8 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -1.9%
Baseline Corr 2nd pt	NO <sub>x</sub> = 396.3 ppb	NO = 391.0 ppb	As found	NO <sub>x</sub> r <sup>2</sup> : 0.999998	Nx SI: 0.989224	Nx Int: 26.060
Baseline Corr 3rd pt	NO <sub>x</sub> = 198.2 ppb	NO = 194.3 ppb	As found	NO r <sup>2</sup> : 0.999992	NO SI: 0.981143	NO Int: 25.360
			As found	NO <sub>2</sub> r <sup>2</sup> : 0.999850	NO <sub>2</sub> SI: 0.998389	NO <sub>2</sub> Int: 2.327

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	0.0	----	----
As found high GPT point	807.6	406.5	403.8	404.5	0.9983	100.2%
As found mid GPT point	807.6	614.6	195.7	198.5	0.9859	101.4%
As found low GPT point	807.6	714.2	96.1	100.8	0.9534	<b>104.9%</b>



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153357

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.999129	
NO <sub>x</sub> Cal Offset:	0.380000	
NO Cal Slope:	0.999991	
NO Cal Offset:	-1.040000	
NO <sub>2</sub> Cal Slope:	0.995614	
NO <sub>2</sub> Cal Offset:	1.058457	

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.522		NO bkgnd or offset:	7.5	
NOX coeff or slope:	0.997		NOX bkgnd or offset:	7.6	
NO2 coeff or slope:	1.000		Reaction cell Press:	201.3	

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero  
 High point  
 Mid point  
 Low point  
 As left zero  
 As left span

Average Correction Factor

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero  
 High GPT point  
 Mid GPT point  
 Low GPT point

Average Correction Factor


Notes: PMT cooler failed. Removing the instrument.

Calibration Performed By: Rene Chamberland

**NO<sub>x</sub> Calibration Plot**

Date: May 8, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Bertha Ganter-Fort McKay  
 Station number: AMS 01  
 Calibration Date: May 9, 2024  
 Last Cal Date: NA  
 Start time (MST): 10:15  
 End time (MST): 14:54  
 Reason: Install

### Calibration Standards

NO Gas Cylinder #: CC335700  
 NOX Cal Gas Conc: 59.40 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 59.40 ppm  
 NOX gas Diff:  
 Calibrator Model: Teledyne API T700  
 ZAG make/model: Teledyne API T701  
 Cal Gas Expiry Date: September 1, 2032  
 NO Cal Gas Conc: 59.20 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 59.20 ppm  
 NO gas Diff:  
 Serial Number: 3565  
 Serial Number: 4890

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero))	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero))
As found zero									<i>Limit = 0.90 - 1.10</i>	<i>Limit = 0.90 - 1.10</i>
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = NA	ppb	NO = NA	ppb					*Percent Change	NO <sub>x</sub> = NA
Baseline Corr 1st pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb					*Percent Change	NO = NA
Baseline Corr 2nd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb						
Baseline Corr 3rd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb						
<b>* = &gt; +/-5% change initiates investigation</b>										
<u>As Found Statistics</u>										
					As found	NO <sub>x</sub> r <sup>2</sup> :		Nx SI:		Nx Int:
					As found	NO r <sup>2</sup> :		NO SI:		NO Int:
					As found	NO <sub>2</sub> r <sup>2</sup> :		NO <sub>2</sub> SI:		NO <sub>2</sub> Int:

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero))	Converter Efficiency
					<i>Limit = 0.90 - 1.10</i>	<i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Teledyne API T200  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 7117

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:		1.000652
NO <sub>x</sub> Cal Offset:		-2.280000
NO Cal Slope:		1.000563
NO Cal Offset:		-2.540000
NO <sub>2</sub> Cal Slope:		0.998079
NO <sub>2</sub> Cal Offset:		-0.755680

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.106		NO bkgnd or offset:	-0.5	
NOX coeff or slope:	1.103		NOX bkgnd or offset:	-0.4	
NO2 coeff or slope:	1.000		Reaction cell Press:	6.0	

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
High point	4932	67.6	803.1	800.4	2.7	802.8	799.9	2.9	1.0004	1.0006
Mid point	4966	33.8	401.5	400.2	1.4	397.5	395.7	1.8	1.0102	1.0114
Low point	4983	16.9	200.8	200.1	0.7	196.8	195.6	1.2	1.0202	1.0230
As left zero	5000	0.0	0.0	0.0	0.0	0.4	0.5	-0.1	----	----
As left span	4932	67.6	803.1	390.3	412.8	794.6	390.3	404.3	1.0107	1.0000
Average Correction Factor									1.0102	1.0116

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	795.4	386.7	411.4	410.2	1.0029	99.7%
Mid GPT point	795.4	591.3	206.8	205.4	1.0068	99.3%
Low GPT point	795.4	693.2	104.9	103.1	1.0175	98.3%
Average Correction Factor					1.0091	99.1%

Notes: Installing a new instrument. Changed the inlet filter. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	NA
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:15	End Time (MST):	14:54
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999957	≥0.995
803.1	802.8	1.0004	Slope	1.000652	0.90 - 1.10
401.5	397.5	1.0102	Intercept	-2.280000	+/-20
200.8	196.8	1.0202			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	NA
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:15	End Time (MST):	14:54
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999984	≥0.995
411.4	410.2	1.0029	Slope	0.998079	0.90 - 1.10
206.8	205.4	1.0068	Intercept	-0.755680	+/-20
104.9	103.1	1.0175			







# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	NA
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:15	End Time (MST):	14:54
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999947	<span style="color: red;">≥0.995</span>
800.4	799.9	1.0006	Slope	1.000563	<span style="color: red;">0.90 - 1.10</span>
400.2	395.7	1.0114	Intercept	-2.540000	<span style="color: red;">+/-20</span>
200.1	195.6	1.0230			



NO<sub>x</sub> Calibration Plot

Date: May 9, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number: AMS 01
Calibration Date:	May 1, 2024	Last Cal Date: April 2, 2024
Start time (MST):	9:48	End time (MST): 12:50
Reason:	Routine	

### Calibration Standards

O3 generation mode:	Photometer	
Calibrator Make/Model:	Teledyne API T700	Serial Number: 3565
ZAG Make/Model:	Teledyne API T701	Serial Number: 4890

### Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #: 1107
Analyzer Range	0 - 500 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998829	1.000114	Backgd or Offset:	4.6	4.6
Calibration intercept:	0.180000	0.080000	Coeff or Slope:	1.027	1.017

### O<sub>3</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	5000	863.1	400.0	405.6	0.985
As found Mid point					
As found Low point					
Baseline Corr As found:	405.9	Previous response	399.7	*% change	1.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	5000	863.1	400.0	399.8	1.001
Mid point	5000	742.5	200.0	200.7	0.997
Low point	5000	651.7	100.0	100.2	0.998
As left zero	5000	0.0	0.0	-0.1	----
As left span	5000	863.1	400.0	402.3	0.994
Average Correction Factor					0.998

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association


## O<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 1, 2024	Previous Calibration:	April 2, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:48	End Time (MST):	12:50
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999993	<span style="color: red;">≥0.995</span>
400.0	399.8	1.0005	Slope	1.000114	<span style="color: red;">0.90 - 1.10</span>
200.0	200.7	0.9965	Intercept	0.080000	<span style="color: red;">+/- 5</span>
100.0	100.2	0.9980			



O<sub>3</sub> Calibration Plot

Date: May 1, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01  
Calibration Date: May 19, 2024 Last Cal Date: N/A  
Start time (MST): 14:45 End time (MST): 14:58  
  
Analyzer Make: Teledyne API T640 S/N: 216  
Particulate Fraction: PM2.5  
  
Flow Meter Make/Model: Alicat FP-25BT S/N: 388748  
Temp/RH standard: Alicat FP-25BT S/N: 388748

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.1	9.0	10.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	733	735.46	733	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.96	5.03	4.96	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	70	----	70	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	1.4	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024  
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.0	N/A	11.0	<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: \_\_\_\_\_ N/A  
Date Disposable Filter Changed: \_\_\_\_\_ N/A

Post- maintenance Zero Verification: PM w/ HEPA: \_\_\_\_\_ 0.0 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_ September 14, 2023  
Date RH/T Sensor Cleaned: \_\_\_\_\_ April 11, 2024

Notes: Installing new instrument, previous instrument had water damage. Flow, pressure and temp checked. Leak check passed. PMT check done. No adjustments needed.

Calibration by: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Report

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	May 10, 2024	Last Cal Date:	April 10, 2024
Start time (MST):	10:07	End time (MST):	13:50
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,040	ppm	Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #:	ALM042207		
Removed Cal Gas Conc:	3,040	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 3565
ZAG Make/Model:	Teledyne API T701		Serial Number: 4890

### Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000946	1.002580	Backgd or Offset:	-0.013	-0.013
Calibration intercept:	0.155852	0.111862	Coeff or Slope:	0.991	0.990

### CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4933	66.7	40.6	40.8	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	40.76	Prev response:	40.75	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4933	66.7	40.6	40.6	0.998
Mid point	4966	33.3	20.2	20.7	0.978
Low point	4983	16.7	10.2	10.2	0.993
As left zero	5000	0.0	0.0	0.0	----
As left span	2960	40.0	40.5	40.1	1.010
Average Correction Factor					0.989

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association


## CO Calibration Summary

### Station Information

Calibration Date:	May 10, 2024	Previous Calibration:	April 10, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:07	End Time (MST):	13:50
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999867
40.6	40.6	0.9979	Slope	1.002580
20.2	20.7	0.9778	Intercept	0.111862
10.2	10.2	0.9926		
				≥0.995
				0.90 - 1.10
				+/-1.5






CO Calibration Plot

Date: May 10, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	May 15, 2024	Last Cal Date:	April 4, 2024
Start time (MST):	8:55	End time (MST):	11:57
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #:	ALM042207		
Removed Cal Gas Conc:	60,200	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 3565
N2 Gen Make/Model:	Peak Scientific		Serial Number: 7220900034

### Analyzer Information

Analyzer make:	Teledyne API 360	Analyzer serial #: 442
Analyzer Range	0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001476	1.001149	Backgd or Offset:	0.045	0.045
Calibration intercept:	-5.720000	-6.240000	Coeff or Slope:	0.876	0.876

### CO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	3000	0.0	0.0	-0.2	----
As found High Point	2920	80.0	1605.3	1601.0	1.003
As found Mid Point					
As found Low Point					
New cylinder response					
Baseline Corr As found:	1601.2	Prev response:	1602.0	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	3000	0.0	0.0	0.2	----
High point	2920	80.0	1605.3	1604.4	1.001
Mid point	2960	40.0	802.7	793.4	1.012
Low point	2980	20.0	401.3	389.6	1.030
As left zero	3000	0.0	0.0	0.1	----
As left span	2960	40.0	802.7	784.4	1.023
Average Correction Factor					1.014

Notes:                   Calibrating remotely. No adjustments made.

Calibration Performed By:                   Rene Chamberland



# Wood Buffalo Environmental Association


## CO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date	May 15, 2024	Previous Calibration	April 4, 2024
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS 01
Start Time (MST)	8:55	End Time (MST)	11:57
Analyzer make	Teledyne API 360	Analyzer serial #	442

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999926	<b>≥0.995</b>
1605.3	1604.4	1.0006	Slope	1.001149	<b>0.90 - 1.10</b>
802.7	793.4	1.0117	Intercept	-6.2	<b>+/-20</b>
401.3	389.6	1.0301			



CO<sub>2</sub> Calibration Plot

Date: May 15, 2024

Location: Bertha Ganter-Fort McKay





# Wood Buffalo Environmental Association

## Nt - NOX - NH3 Calibration Report

### Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
NOX Cal Date:	May 21, 2024	Last Cal Date:	April 11, 2024
Start time (MST):	9:17	End time (MST):	14:30
NH3 Cal Date:	May 22, 2024	Last Cal Date:	April 11, 2024
Start time (MST):	10:12	End time (MST):	14:08
Reason:	Routine		

### Calibration Standards

NOX Cal Gas Conc:	59.40	ppm	NO Gas Cylinder #:	CC335700
NO Cal Gas Conc:	59.20	ppm	NO Cal Gas Expiry:	September 1, 2032
Removed NOX Conc:	59.40	ppm	Removed Cylinder #:	NA
Removed NO Conc:	59.20	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.58	ppm	NH3 Gas Cylinder #:	CC743587
Removed NH3 Conc:	76.58	ppm	NH3 Cal Gas Expiry:	August 22, 2024
NH3 gas Diff:			Removed Cylinder #:	NA
Calibrator Model:	API T700		Removed cyl Expiry:	NA
ZAG make/model:	API T701		Serial Number:	3565
			Serial Number:	4890

### Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	475
Converter model:	API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	7.70
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	510

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.952	0.907	Nt coefficient:	0.962	0.922
NOX coefficient:	0.959	0.913	NO bkgrnd:	-0.9	-0.9
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	-0.3	-0.3
NH3 coefficient:	0.946	0.946	Nt bkgrnd:	1.2	1.2

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000623	0.999115
NO <sub>x</sub> Cal Offset:	-2.120000	-2.240000
NO Cal Slope:	0.997950	0.998449
NO Cal Offset:	-2.600000	-2.700000
NO <sub>2</sub> Cal Slope:	1.001154	1.000883
NO <sub>2</sub> Cal Offset:	0.886739	-0.637947
NH3 Cal Slope:	0.997027	0.997495
NH3 Cal Offset:	2.325094	1.189973
Nt Cal Slope:	1.000070	1.000452
Nt Cal Offset:	2.166761	1.101631



# Wood Buffalo Environmental Association

## NO<sub>x</sub> - NO - NO<sub>2</sub> Calibration Report

### NO<sub>x</sub> / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>	Baseline corr NO Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.8	0.8	0.8	----	----
As found span	4932	67.6	803.1	800.4	803.1	845.1	834.0	837.8	0.9503	0.9597
AF GPT span	4932	67.6	803.1	-----	803.1	801.0	-----	803.0	1.0026	-----

new NO cyl rp

Baseline Corr As Fd      Nt = 837 ppb      NO<sub>x</sub> = 844.3 ppb      NO = 833.2 ppb      \*Percent Change      Nt<sub>(NO)</sub> = 3.8%

Previous Response      Nt = 805.31 ppb      NO<sub>x</sub> = 801.5 ppb      NO = 796.1 ppb      \*Percent Change      NO<sub>x</sub> = 5.1%

\*\*NO<sub>x</sub> Δ (NO to GPT response) = -5.2%      \*Percent Change      NO = 4.4%

*\*\* = > +/-2% difference initiates investigation*      *\* = > +/-5% change initiates investigation*

### NO<sub>x</sub> / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	-0.3	0.0	-0.5	----	----
High point	4932	67.6	803.1	800.4	803.1	801.4	798.2	801.7	1.0021	1.0027
Mid point	4966	33.8	401.5	400.2	401.5	397.1	394.3	398.8	1.0112	1.0149
Low point	4983	16.9	200.8	200.1	200.8	197.0	195.2	198.2	1.0191	1.0251
Average Correction Factor									1.0108	1.0143

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found zero	----	----	0.0	0.0	----	----
Calibration zero	----	----	0.0	-0.3	----	----
High GPT point (400 ppb O3)	795.3	385.2	412.8	412.4	1.0010	99.9%
Mid GPT point (200 ppb O3)	795.3	591.9	206.1	206.4	0.9986	100.1%
Low GPT point (100 ppb O3)	795.3	692.7	105.3	103.8	1.0145	98.6%
Average Correction Factor					1.0047	99.5%



## Wood Buffalo Environmental Association NH<sub>3</sub> - N<sub>T</sub> Calibration Report

### NH<sub>3</sub> As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH3 Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.0	-1.0	0.0	----	----
AF High point										
AF Mid point										
AF Low point										
new NH3 cyl rp										
Baseline Corr As Fd	Nt =	NA ppb	NH3 =	NA ppb					*Percent Change	Nt <sub>(NH3)</sub> = NA
Previous Response	Nt =	NA ppb	NH3 =	NA ppb					*Percent Change	NH3 = NA

\* => +/-5% change initiates investigation

### NH<sub>3</sub> Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.3	-0.2	----	----
High point	3418	82.2	1798.5	0.0	1798.5	1798.2	5.2	1793.1	1.000	1.003
Mid point	3454	45.7	1000.0	0.0	1000.0	1005.8	2.7	1002.8	0.994	0.997
Low point	3477	22.8	498.9	0.0	498.9	499.8	1.5	498.2	0.998	1.001
								Average Correction Factor	0.9975	1.0005
NH3 Previous Converter Efficiency =		90.8 %								
NH3 Current Converter Efficiency =		90.8 %								

Notes: Changed the inlet filter after as founds. Adjusted the NOx/Nt span.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association


## Nt Calibration Summary

### Station Information

Calibration Date:	May 22, 2024	Previous Calibration:	April 11, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:17	End Time (MST):	14:30
Analyzer make:	API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.5	----	Correlation Coefficient	0.999985	<i>≥0.995</i>
1798.5	1798.2	1.0002	Slope	1.000452	<i>0.90 - 1.10</i>
1000.0	1005.8	0.9942	Intercept	1.101631	<i>+/-20</i>
498.9	499.8	0.9981			







# Wood Buffalo Environmental Association


## NH<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 22, 2024	Previous Calibration:	April 11, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:17	End Time (MST):	14:30
Analyzer make:	API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999987	<i>≥0.995</i>
1798.5	1793.1	1.0030	Slope	0.997495	<i>0.90 - 1.10</i>
1000.0	1002.8	0.9972	Intercept	1.189973	<i>+/-20</i>
498.9	498.2	1.0013			





# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 21, 2024	Previous Calibration:	April 11, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:17	End Time (MST):	14:30
Analyzer make:	API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.3	----	Correlation Coefficient	0.999970	<i>≥0.995</i>
803.1	801.4	1.0021	Slope	0.999115	<i>0.90 - 1.10</i>
401.5	397.1	1.0112	Intercept	-2.240000	<i>+/-20</i>
200.8	197.0	1.0191			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 21, 2024	Previous Calibration:	April 11, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:17	End Time (MST):	14:30
Analyzer make:	API T201	Analyzer serial #:	475

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999941	<i>≥0.995</i>
800.4	798.2	1.0027	Slope	0.998449	<i>0.90 - 1.10</i>
400.2	394.3	1.0149	Intercept	-2.700000	<i>+/-20</i>
200.1	195.2	1.0251			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 21, 2024	Previous Calibration:	April 11, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:17	End Time (MST):	14:30
Analyzer make:	API T201	Analyzer serial #:	475

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.3	----	Correlation Coefficient	0.999983	<i>≥0.995</i>
412.8	412.4	1.0010	Slope	1.000883	<i>0.90 - 1.10</i>
206.1	206.4	0.9986	Intercept	-0.637947	<i>+/-20</i>
105.3	103.8	1.0145			



NO<sub>x</sub> Calibration Plot

Date: May 21, 2024


Location: Bertha Ganter-Fort McKay



NH<sub>3</sub> Calibration Plot

Date: May 22, 2024

Location: Bertha Ganter-Fort McKay





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS02 MILDRED LAKE MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Mildred Lake	Station number:	AMS 02
Calibration Date:	May 3, 2024	Last Cal Date:	April 5, 2024
Start time (MST):	9:41	End time (MST):	12:43
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date: August 12, 2024
Cal Gas Cylinder #:	CC501209		
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	Teledyne API T700		Serial Number: 1185
Zero Air Gen Model:	Teledyne API T701		Serial Number: 4891

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	JC1404901075
Analyzer Range:	0-1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993915	1.009337	Backgd or Offset:	18.7	18.7
Calibration intercept:	-0.766439	-0.823191	Coeff or Slope:	0.787	0.787

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4920	80.2	801.6	797.7	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.8	Previous response	796.0	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4920	80.2	801.6	808.8	0.991
Mid point	4960	40.1	400.8	402.9	0.995
Low point	4980	20.0	199.9	200.7	0.996
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	80.2	801.6	803.2	0.998
Average Correction Factor:					0.994

Notes: Changed sample inlet filter after As Found. No adjustments made.

Calibration Performed By: Ryan Power





# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 5, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:41	End Time (MST):	12:43
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>	
0.0	-0.2	----	Correlation Coefficient	0.999996	<b>≥0.995</b>
801.6	808.8	0.9912	Slope	1.009337	<b>0.90 - 1.10</b>
400.8	402.9	0.9949	Intercept	-0.823191	<b>+/-30</b>
199.9	200.7	0.9961			



SO2 Calibration Plot

Date: May 3, 2024

Location: Mildred Lake





# Wood Buffalo Environmental Association

## H2S Calibration Report

### Station Information

Station Name:	Mildred Lake	Station number:	AMS 02
Calibration Date:	May 28, 2024	Last Cal Date:	April 23, 2024
Start time (MST):	9:20	End time (MST):	14:01
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.29	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC345191			
Removed Cal Gas Conc:	5.29	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1185
ZAG Make/Model:	Teledyne API T701		Serial Number:	4891

### Analyzer Information

Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966
Converter make:	Global G150	Converter serial #:	2022-198
Analyzer Range	0 - 100 ppb	Converter Temp:	325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008965	1.005108	Backgd or Offset:	1.95	1.93
Calibration intercept:	-0.259192	-0.199197	Coeff or Slope:	0.731	0.722

### H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4924	75.6	80.0	81.1	0.986
As found Mid point	4962	37.8	40.0	40.6	0.985
As found Low point	4981	18.9	20.0	20.1	0.995
New cylinder response					
Baseline Corr As found:	81.1	Prev response:	80.45	*% change:	0.8%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.014680	AF Intercept:	-0.059189
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999993	<i>* = +/-5% change initiates investigation</i>	

### H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4924	75.6	80.0	80.2	0.997
Mid point	4962	37.8	40.0	40.1	0.997
Low point	4981	18.9	20.0	19.7	1.015
As left zero	5000	0.0	0.0	-0.1	----
As left span	4924	75.6	80.0	79.8	1.002
SO2 Scrubber Check	4920	80.2	802.0	-0.1	----
Date of last scrubber change:	September 20, 2023			Ave Corr Factor	1.003
Date of last converter efficiency test:	April 23, 2024			107.1% efficiency	

Notes: Changed inlet filter after MPAFs. Adjusted span.

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association


## H2S Calibration Summary

### Station Information

Calibration Date:	May 28, 2024	Previous Calibration:	April 23, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:20	End Time (MST):	14:01
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311966

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999983	$\geq 0.995$
80.0	80.2	0.9974	Slope	1.005108	$0.90 - 1.10$
40.0	40.1	0.9974	Intercept	-0.199197	$\pm 3$
20.0	19.7	1.0151			



H2S Calibration Plot

Date: May 28, 2024

Location: Mildred Lake





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Mildred Lake	Station number:	AMS 02
Calibration Date:	May 3, 2024	Last Cal Date:	April 11, 2024
Start time (MST):	9:41	End time (MST):	12:41
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH4 Cal Gas Conc.	500.2 ppm	CH4 Equiv Conc.	1048.6 ppm
C3H8 Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	500.2 ppm	CH4 Equiv Conc.	1048.6 ppm
Removed C3H8 Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
Zero Air Gen model:	Teledyne API T701	Serial Number:	4891

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320039
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.08E-04	3.19E-04	NMHC SP Ratio:	6.25E-05	6.29E-05
CH4 Retention time:	17.6	17.8	NMHC Peak Area:	140693	139909
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	---
As found High point	4920	80.2	16.82	16.62	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.62	Prev response	16.83	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	
High point	4920	80.2	16.82	16.87	0.997
Mid point	4960	40.1	8.41	8.46	0.994
Low point	4980	20.0	4.19	4.22	0.995
As left zero	5000	0.0	0.00	0.00	---
As left span	4920	80.2	16.82	16.89	0.996
Average Correction Factor					0.995

Notes: Filter changed after As Finds. Minor adjustment to span. Station temp lowered 3 degrees for better analyzer operation.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	8.80	8.80	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.80	Prev response	8.82	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	8.80	8.81	0.998
Mid point	4960	40.1	4.40	4.44	0.990
Low point	4980	20.0	2.19	2.22	0.989
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	8.80	8.84	0.995
Average Correction Factor					0.993

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	8.02	7.82	1.026
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.82	Prev response	8.02	*% change	-2.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	8.02	8.06	0.995
Mid point	4960	40.1	4.01	4.02	0.997
Low point	4980	20.0	2.00	2.00	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	8.02	8.05	0.997
Average Correction Factor					0.998

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001093	1.003097
THC Cal Offset:	-0.005909	0.008106
CH <sub>4</sub> Cal Slope:	1.000042	1.005411
CH <sub>4</sub> Cal Offset:	-0.007450	-0.007040
NMHC Cal Slope:	1.002130	1.001362
NMHC Cal Offset:	0.001741	0.014946

Calibration Performed By: Ryan Power



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 11, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:41	End Time (MST):	12:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999997	<i>≥0.995</i>
16.82	16.87	0.9969	Slope	1.003097	<i>0.90 - 1.10</i>
8.41	8.46	0.9940	Intercept	0.008106	<i>+/-0.5</i>
4.19	4.22	0.9951			







# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 11, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:41	End Time (MST):	12:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999997	<i>≥0.995</i>
8.02	8.06	0.9950	Slope	1.005411	<i>0.90 - 1.10</i>
4.01	4.02	0.9974	Intercept	-0.007040	<i>+/-0.5</i>
2.00	2.00	1.0014			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 11, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:41	End Time (MST):	12:41
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999980 <span style="color: red;">≥0.995</span>
8.80	8.81	0.9983	Slope	1.001362 <span style="color: red;">0.90 - 1.10</span>
4.40	4.44	0.9905	Intercept	0.014946 <span style="color: red;">+/-0.5</span>
2.19	2.22	0.9894		



NMHC Calibration Plot

Date: May 3, 2024

Location: Mildred Lake





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Mildred Lake	Station number:	AMS 02
Calibration Date:	May 10, 2024	Last Cal Date:	May 3, 2024
Start time (MST):	9:35	End time (MST):	14:07
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH4 Cal Gas Conc.	500.2 ppm	CH4 Equiv Conc.	1048.6 ppm
C3H8 Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	500.2 ppm	CH4 Equiv Conc.	1048.6 ppm
Removed C3H8 Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
Zero Air Gen model:	Teledyne API T701	Serial Number:	4891

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320039
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.19E-04	3.19E-04	NMHC SP Ratio:	6.29E-05	6.25E-05
CH4 Retention time:	17.8	18.0	NMHC Peak Area:	139909	140773
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	---
As found High point	4920	80.2	16.82	16.74	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.74	Prev response	16.88	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	
High point	4920	80.2	16.82	16.70	1.007
Mid point	4960	40.1	8.41	8.42	0.999
Low point	4980	20.0	4.19	4.17	1.007
As left zero	5000	0.0	0.00	0.00	---
As left span	4920	80.2	16.82	16.85	0.998
Average Correction Factor					1.004

Notes: Ran new zero chromatogram, and turned on "use zero chromatogram" setting. Adjusted span.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	8.80	8.73	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.73	Prev response	8.82	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	8.80	8.70	1.011
Mid point	4960	40.1	4.40	4.42	0.994
Low point	4980	20.0	2.19	2.19	1.002
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	8.80	8.82	0.997
Average Correction Factor					1.002

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	8.02	8.01	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.01	Prev response	8.06	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	8.02	8.00	1.003
Mid point	4960	40.1	4.01	4.00	1.004
Low point	4980	20.0	2.00	1.98	1.012
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	8.02	8.02	1.000
Average Correction Factor					1.006

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.003097	0.993409
THC Cal Offset:	0.008106	0.014859
CH <sub>4</sub> Cal Slope:	1.005411	0.997693
CH <sub>4</sub> Cal Offset:	-0.007040	-0.007457
NMHC Cal Slope:	1.001362	0.989385
NMHC Cal Offset:	0.014946	0.022516

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 10, 2024	Previous Calibration:	May 3, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:35	End Time (MST):	14:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999977 <span style="color: red;">≥0.995</span>
16.82	16.70	1.0071	Slope	0.993409 <span style="color: red;">0.90 - 1.10</span>
8.41	8.42	0.9987	Intercept	0.014859 <span style="color: red;">+/-0.5</span>
4.19	4.17	1.0065		





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 10, 2024	Previous Calibration:	May 3, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:35	End Time (MST):	14:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999995	<i>≥0.995</i>
8.02	8.00	1.0030	Slope	0.997693	<i>0.90 - 1.10</i>
4.01	4.00	1.0039	Intercept	-0.007457	<i>+/-0.5</i>
2.00	1.98	1.0115			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 10, 2024	Previous Calibration:	May 3, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:35	End Time (MST):	14:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999912	<i>≥0.995</i>
8.80	8.70	1.0109	Slope	0.989385	<i>0.90 - 1.10</i>
4.40	4.42	0.9941	Intercept	0.022516	<i>+/-0.5</i>
2.19	2.19	1.0020			






NMHC Calibration Plot

Date: May 10, 2024

Location: Mildred Lake





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Mildred Lake	Station number:	AMS 02
Calibration Date:	May 30, 2024	Last Cal Date:	May 10, 2024
Start time (MST):	9:42	End time (MST):	13:19
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH4 Cal Gas Conc.	500.2 ppm	CH4 Equiv Conc.	1048.6 ppm
C3H8 Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	500.2 ppm	CH4 Equiv Conc.	1048.6 ppm
Removed C3H8 Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
Zero Air Gen model:	Teledyne API T701	Serial Number:	4891

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320039
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.19E-04	3.27E-04	NMHC SP Ratio:	6.25E-05	6.13E-05
CH4 Retention time:	18.0	18.2	NMHC Peak Area:	140773	143550
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	16.82	16.34	1.029
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.34	Prev response	16.72	*% change	-2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	
High point	4920	80.2	16.82	16.82	1.000
Mid point	4960	40.1	8.41	8.36	1.006
Low point	4980	20.0	4.19	4.19	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	16.82	16.71	1.006
Average Correction Factor					1.002

Notes: Swapped nitrogen cylinder. Ran new zero chromatogram and adjusted span.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	8.80	8.54	1.030
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.54	Prev response	8.72	*% change	-2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	8.80	8.87	0.992
Mid point	4960	40.1	4.40	4.38	1.005
Low point	4980	20.0	2.19	2.21	0.992
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	8.80	8.74	1.006
Average Correction Factor					0.996

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	8.02	7.80	1.029
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.80	Prev response	8.00	*% change	-2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	8.02	7.95	1.010
Mid point	4960	40.1	4.01	3.99	1.006
Low point	4980	20.0	2.00	1.98	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	8.02	7.98	1.006
Average Correction Factor					1.009

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.993409	0.999848
THC Cal Offset:	0.014859	-0.011503
CH <sub>4</sub> Cal Slope:	0.997693	0.990500
CH <sub>4</sub> Cal Offset:	-0.007457	0.002532
NMHC Cal Slope:	0.989385	1.007805
NMHC Cal Offset:	0.022516	-0.012836

Calibration Performed By: Braiden Boutillier



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	May 10, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:42	End Time (MST):	13:19
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999988	<i>≥0.995</i>
16.82	16.82	0.9999	Slope	0.999848	<i>0.90 - 1.10</i>
8.41	8.36	1.0059	Intercept	-0.011503	<i>+/-0.5</i>
4.19	4.19	1.0008			





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	May 10, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:42	End Time (MST):	13:19
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999996	<i>≥0.995</i>
8.02	7.95	1.0098	Slope	0.990500	<i>0.90 - 1.10</i>
4.01	3.99	1.0064	Intercept	0.002532	<i>+/-0.5</i>
2.00	1.98	1.0100			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	May 10, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:42	End Time (MST):	13:19
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999939	<i>≥0.995</i>
8.80	8.87	0.9916	Slope	1.007805	<i>0.90 - 1.10</i>
4.40	4.38	1.0052	Intercept	-0.012836	<i>+/-0.5</i>
2.19	2.21	0.9925			



NMHC Calibration Plot

Date: May 30, 2024

Location: Mildred Lake





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS04 BUFFALO VIEWPOINT MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024







# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Buffalo Viewpoint	Station number: AMS 04
Calibration Date:	May 30, 2024	Last Cal Date: April 15, 2024
Start time (MST):	7:20	End time (MST): 10:11
Reason:	Routine	

### Calibration Standards

Cal Gas Concentration:	50.87	ppm	Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #:	CC446753		
Removed Cal Gas Conc:	50.87	ppm	Rem Gas Exp Date:
Removed Gas Cyl #:			Diff between cyl:
Calibrator Model:	API T700		Serial Number: 3808
Zero Air Gen Model:	API T701		Serial Number: 362

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: JC1327300932
Analyzer Range:	0-1000ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006888	1.001458	Backgd or Offset:	25.1	25.0
Calibration intercept:	0.395801	0.594841	Coeff or Slope:	0.880	0.873

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4921	78.6	799.7	807.0	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	806.7	Previous response	805.6	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.5	----
High point	4921	78.6	799.7	801.2	0.998
Mid point	4961	39.3	399.8	401.8	0.995
Low point	4980	19.6	199.4	199.9	0.998
As left zero	5000	0.0	0.0	0.2	----
As left span	4921	78.6	799.7	801.1	0.998
Average Correction Factor:					0.997

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	April 15, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:20	End Time (MST):	10:11
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.5	----	Correlation Coefficient	0.999997	<span style="color: red;">≥0.995</span>
799.7	801.2	0.9982	Slope	1.001458	<span style="color: red;">0.90 - 1.10</span>
399.8	401.8	0.9951	Intercept	0.594841	<span style="color: red;">+/-30</span>
199.4	199.9	0.9976			



SO2 Calibration Plot

Date: May 30, 2024

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

### Station Information

Station Name: Buffalo Viewpoint	Station number: AMS 04
Calibration Date: May 13, 2024	Last Cal Date: April 5, 2024
Start time (MST): 7:50	End time (MST): 11:48
Reason: Routine	

### Calibration Standards

Cal Gas Concentration: 5.42 ppm	Cal Gas Exp Date: January 4, 2025
Cal Gas Cylinder #: CC345266	
Removed Cal Gas Conc: 5.42 ppm	Rem Gas Exp Date:
Removed Gas Cyl #:	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 3808
ZAG Make/Model: API T701H	Serial Number: 362

### Analyzer Information

Analyzer make: Thermo 43i-LTE	Analyzer serial #: 1008841400
Converter make: Global	Converter serial #: 2022-200
Analyzer Range: 0 - 100 ppb	Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997925	0.996076	Backgd or Offset:	1.09	1.07
Calibration intercept:	0.102175	0.042122	Coeff or Slope:	1.130	1.123

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4926	74.1	80.3	81.7	0.984
As found Mid point	4963	37.0	40.1	40.8	0.985
As found Low point	4982	18.5	20.1	20.0	1.008
New cylinder response					
Baseline Corr As found:	81.6	Prev response:	80.26	*% change:	1.6%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.017847	AF Intercept:	-0.097494
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999961	<i>* = &gt; +/-5% change initiates investigation</i>	

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4926	74.1	80.3	80.0	1.004
Mid point	4963	37.0	40.1	40.2	0.998
Low point	4982	18.5	20.1	19.8	1.013
As left zero	5000	0.0	0.0	0.1	----
As left span	4926	74.1	80.3	79.5	1.010
SO <sub>2</sub> Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	1.005
Date of last converter efficiency test:					

Notes: Sox scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association


## H<sub>2</sub>S Calibration Summary

### Station Information

Calibration Date:	May 13, 2024	Previous Calibration:	April 5, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:50	End Time (MST):	11:48
Analyzer make:	Global	Analyzer serial #:	2022-200

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999973	$\geq 0.995$
80.3	80.0	1.0040	Slope	0.996076	$0.90 - 1.10$
40.1	40.2	0.9977	Intercept	0.042122	$\pm 3$
20.1	19.8	1.0127			



H<sub>2</sub>S Calibration Plot

Date: May 13, 2024

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS 04
Calibration Date:	May 14, 2024	Last Cal Date:	April 15, 2024
Start time (MST):	7:20	End time (MST):	8:22
Reason:	Cylinder Change		

### Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	497.2 ppm	CH4 Equiv Conc.	1058.2 ppm
C3H8 Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	497.2 ppm	CH4 Equiv Conc.	1058.2 ppm
Removed C3H8 Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
Zero Air Gen model:	API T701	Serial Number:	362

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1426262594
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.22E-04	4.22E-04	NMHC SP Ratio:	1.07E-04	1.07E-04
CH4 Retention time:	13.7	13.7	NMHC Peak Area:	82687	82687
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	16.64	16.61	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.61	Prev response	16.60	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	16.64	16.50	1.009
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.009

Notes: Hydrogen Cylinder Changed.





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	8.82	8.80	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.80	Prev response	8.80	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	8.82	8.73	1.010
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.010

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	7.82	7.81	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.81	Prev response	7.80	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	7.82	7.77	1.006
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.006

### Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.997908	0.991571
THC Cal Offset:	0.002057	0.000000
CH <sub>4</sub> Cal Slope:	0.998894	0.993781
CH <sub>4</sub> Cal Offset:	-0.004105	0.000000
NMHC Cal Slope:	0.997215	0.989725
NMHC Cal Offset:	0.005963	0.000000

Calibration Performed By:      Melissa Lemay

NMHC Calibration Plot

Date: May 14, 2024

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS 04
Calibration Date:	May 30, 2024	Last Cal Date:	April 15, 2024
Start time (MST):	7:20	End time (MST):	10:10
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	497.2 ppm	CH4 Equiv Conc.	1058.2 ppm
C3H8 Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	497.2 ppm	CH4 Equiv Conc.	1058.2 ppm
Removed C3H8 Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
Zero Air Gen model:	API T701	Serial Number:	362

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1426262594
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.22E-04	4.24E-04	NMHC SP Ratio:	1.07E-04	1.07E-04
CH4 Retention time:	13.7	13.7	NMHC Peak Area:	82687	82267
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	---
As found High point	4921	78.6	16.64	16.67	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.67	Prev response	16.60	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	---
High point	4921	78.6	16.64	16.61	1.001
Mid point	4961	39.3	8.32	8.31	1.001
Low point	4980	19.6	4.15	4.16	0.998
As left zero	5000	0.0	0.00	0.00	---
As left span	4921	78.6	16.64	16.58	1.003
Average Correction Factor					1.000

Notes: No maintenance done. Span adjusted.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	8.82	8.86	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.86	Prev response	8.80	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	8.82	8.81	1.001
Mid point	4961	39.3	4.41	4.42	0.998
Low point	4980	19.6	2.20	2.22	0.991
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	8.82	8.79	1.003
Average Correction Factor					0.997

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	7.82	7.80	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.80	Prev response	7.80	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	7.82	7.80	1.002
Mid point	4961	39.3	3.91	3.89	1.004
Low point	4980	19.6	1.95	1.94	1.005
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	7.82	7.79	1.003
Average Correction Factor					1.004

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.997908	0.998333
THC Cal Offset:	0.002057	0.007463
CH <sub>4</sub> Cal Slope:	0.998894	0.998558
CH <sub>4</sub> Cal Offset:	-0.004105	-0.004706
NMHC Cal Slope:	0.997215	0.997874
NMHC Cal Offset:	0.005963	0.012169

Calibration Performed By:      Melissa Lemay



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	April 15, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:20	End Time (MST):	10:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999999 <span style="color: red;">≥0.995</span>
16.64	16.61	1.0014	Slope	0.998333 <span style="color: red;">0.90 - 1.10</span>
8.32	8.31	1.0006	Intercept	0.007463 <span style="color: red;">+/-0.5</span>
4.15	4.16	0.9977		





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	April 15, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:20	End Time (MST):	10:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
7.82	7.80	1.0016	Slope	0.998558	<i>0.90 - 1.10</i>
3.91	3.89	1.0040	Intercept	-0.004706	<i>+/-0.5</i>
1.95	1.94	1.0053			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	April 15, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:20	End Time (MST):	10:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999991	<i>≥0.995</i>
8.82	8.81	1.0014	Slope	0.997874	<i>0.90 - 1.10</i>
4.41	4.42	0.9980	Intercept	0.012169	<i>+/-0.5</i>
2.20	2.22	0.9911			



NMHC Calibration Plot

Date: May 30, 2024

Location: Buffalo Viewpoint







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Buffalo Viewpoint  
 Station number: AMS 04  
 Calibration Date: May 2, 2024  
 Last Cal Date: April 2, 2024  
 Start time (MST): 5:50  
 End time (MST): 10:55  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: CC324979  
 NOX Cal Gas Conc: 48.90 ppm  
 Removed Cylinder #:  
 Removed Gas NOX Conc: 48.90 ppm  
 NOX gas Diff:  
 Calibrator Model: API T700  
 ZAG make/model: APIT701  
 Cal Gas Expiry Date: November 3, 2032  
 NO Cal Gas Conc: 48.80 ppm  
 Removed Gas Exp Date:  
 Removed Gas NO Conc: 48.80 ppm  
 NO gas Diff:  
 Serial Number: 3808  
 Serial Number: 362

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
AF High point	4918	81.8	800.0	798.4	1.6	792.2	790.5	1.7	1.0100	1.0102
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 802.8 ppb		NO = 800.0 ppb			<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -1.4%	
Baseline Corr 1st pt	NO <sub>x</sub> = 792.1 ppb		NO = 790.3 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -1.2%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO r <sup>2</sup> :	NO SI:	NO Int:	
						As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: API T200  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 721

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.189	1.187	NO bkgnd or offset:	-0.6	-0.6
NOX coeff or slope:	1.177	1.199	NOX bkgnd or offset:	-0.3	-0.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.4	4.4

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.003822	0.996881
NO <sub>x</sub> Cal Offset:	-0.272631	0.805902
NO Cal Slope:	1.002903	0.998766
NO Cal Offset:	-0.733480	0.385887
NO <sub>2</sub> Cal Slope:	0.996313	0.984467
NO <sub>2</sub> Cal Offset:	-1.280139	1.356482

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.6	0.7	-0.1	----	----
High point	4918	81.8	800.0	798.4	1.6	797.9	797.9	0.0	1.0027	1.0006
Mid point	4959	40.9	400.0	399.2	0.8	400.7	399.1	1.6	0.9983	1.0002
Low point	4980	20.4	199.5	199.1	0.4	199.2	198.8	0.4	1.0015	1.0014
As left zero	5000	0.0	0.0	0.8	-0.8	0.7	0.8	-0.1	----	----
As left span	4918	81.8	800.0	400.1	800.0	792.5	400.1	392.3	1.0095	1.0000
Average Correction Factor									1.0008	1.0008

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	795.7	400.0	397.3	391.6	1.0146	98.6%
Mid GPT point	795.7	599.3	198.0	197.6	1.0022	99.8%
Low GPT point	795.7	695.2	102.1	103.0	0.9916	100.8%
Average Correction Factor					1.0028	99.7%

Notes: Span adjusted. Due to drifting during the GPT the 2nd NO ref point used.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 2, 2024	Previous Calibration:	April 2, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:50	End Time (MST):	10:55
Analyzer make:	API T200	Analyzer serial #:	721

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.6	----	Correlation Coefficient	0.999995	<i>≥0.995</i>
800.0	797.9	1.0027	Slope	0.996881	<i>0.90 - 1.10</i>
400.0	400.7	0.9983	Intercept	0.805902	<i>+/-20</i>
199.5	199.2	1.0015			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 2, 2024	Previous Calibration:	April 2, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:50	End Time (MST):	10:55
Analyzer make:	API T200	Analyzer serial #:	721

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999930	≥0.995
397.3	391.6	1.0146	Slope	0.984467	0.90 - 1.10
198.0	197.6	1.0022	Intercept	1.356482	+/-20
102.1	103.0	0.9916			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 2, 2024	Previous Calibration:	April 2, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:50	End Time (MST):	10:55
Analyzer make:	API T200	Analyzer serial #:	721

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.7	----	Correlation Coefficient	0.999999	<span style="color: red;">≥0.995</span>
798.4	797.9	1.0006	Slope	0.998766	<span style="color: red;">0.90 - 1.10</span>
399.2	399.1	1.0002	Intercept	0.385887	<span style="color: red;">+/-20</span>
199.1	198.8	1.0014			



NO<sub>x</sub> Calibration Plot

Date: May 2, 2024

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

### Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS 04
Calibration Date:	May 10, 2024	Last Cal Date:	April 15, 2024
Start time (MST):	8:35	End time (MST):	11:14
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3808
Calibrator Make/Model:	APIP T700	Serial Number:	362
ZAG Make/Model:	API T701		

### Analyzer Information

Analyzer make:	API T400	Analyzer serial #:	2961
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003629	1.002543	Backgd or Offset:	-2.2
Calibration intercept:	0.540000	0.780000	Coeff or Slope:	1.011

### O<sub>3</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	5000	995.6	400.0	400.1	0.999
As found Mid point					
As found Low point					
Baseline Corr As found:	400.5	Previous response	402.0	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	5000	998.1	400.0	401.6	0.996
Mid point	5000	822.5	200.0	201.4	0.993
Low point	5000	712.4	100.0	101.6	0.984
As left zero	5000	0.0	0.0	0.1	----
As left span	5000	997.1	400.0	402.9	0.993
Average Correction Factor					0.991

Notes: No adjustments and maintenance done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association


## O<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 10, 2024	Previous Calibration:	April 15, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	8:35	End Time (MST):	11:14
Analyzer make:	API T400	Analyzer serial #:	2961

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999993	<span style="color: red;">≥0.995</span>
400.0	401.6	0.9960	Slope	1.002543	<span style="color: red;">0.90 - 1.10</span>
200.0	201.4	0.9930	Intercept	0.780000	<span style="color: red;">+/- 5</span>
100.0	101.6	0.9843			






O<sub>3</sub> Calibration Plot

Date: May 10, 2024

Location: Buffalo Viewpoint





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04  
Calibration Date: April 24, 2024 Last Cal Date: March 20, 2024  
Start time (MST): 6:17 End time (MST): 7:20  
Analyzer Make: API T640 S/N: 321  
Particulate Fraction: PM2.5  
Flow Meter Make/Model: Alicat FP-25BT S/N: 1451  
Temp/RH standard: Alicat FP-25BT S/N: 1451

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.1	9.8	10.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	719.4	721.5	719.4	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.15	5.00	<input checked="" type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	----	42	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	4.6	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 6-10-2024  
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.9	12.2	10.9	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: May 30, 2024  
Date Disposable Filter Changed: May 30, 2024

Post- maintenance Zero Verification: PM w/ HEPA: \_\_\_\_\_ 0 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: February 27, 2024  
Date RH/T Sensor Cleaned: February 27, 2024

Notes: Flow and PMT adjusted. Leak check passed.

Calibration by: Melissa Lemay



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS05  
MANNIX  
MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	May 21, 2024	Last Cal Date:	April 4, 2024
Start time (MST):	9:06	End time (MST):	10:28
Reason:	Removal		

### Calibration Standards

Cal Gas Concentration:	50.02	ppm	Cal Gas Exp Date: January 12, 2029
Cal Gas Cylinder #:	XC026809B		
Removed Cal Gas Conc:	50.02	ppm	Rem Gas Exp Date: N/A
Removed Gas Cyl #:	N/A		Diff between cyl:
Calibrator Model:	API T700		Serial Number: 621
Zero Air Gen Model:	API T701H		Serial Number: 832

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1008841399
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999943		Backgd or Offset:	9.6	9.6
Calibration intercept:	0.380000		Coeff or Slope:	0.944	0.944

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4920	80.0	800.3	803.9	0.996
As found Mid point	4960	40.0	400.2	399.2	1.003
As found Low point	4980	20.0	200.1	199.6	1.003
New cylinder response					
Baseline Corr As found:	803.8	Previous response	800.7	*% change	0.4%
Baseline Corr 2nd AF pt:	399.1	AF Slope:	1.004570	AF Intercept:	-1.040000
Baseline Corr 3rd AF pt:	199.5	AF Correlation:	0.999985	<i>* = &gt; +/-5% change initiates investigation</i>	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor:


Notes: Removal calibration for station replacement.

Calibration Performed By: Max Farrell

SO2 Calibration Plot

Date: May 21, 2024

Location: Mannix





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Mannix	Station number: AMS 05
Calibration Date:	May 25, 2024	Last Cal Date: N/A
Start time (MST):	7:41	End time (MST): 10:00
Reason:	Install	

### Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC408659			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	5470
Zero Air Gen Model:	API T701		Serial Number:	361

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 1008841399
Analyzer Range:	1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		0.999272	Backgd or Offset:		9.7
Calibration intercept:		-0.306089	Coeff or Slope:		0.932

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
-----------	-------------------------------	-----------------------------	-------------------------------------	------------------------------------	--

As found zero  
 As found High point  
 As found Mid point  
 As found Low point  
 New cylinder response

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	-------------------------------	-----------------------------	-------------------------------------	------------------------------------	---

Calibrator zero	5000	0.0	0.0	1.0	----
High point	4920	80.3	800.4	800.0	1.000
Mid point	4960	40.1	399.7	398.9	1.002
Low point	4980	20.1	200.4	198.3	1.010
As left zero	5000	0.0	0.0	0.6	----
As left span	4920	80.3	800.4	802.3	0.998

Average Correction Factor: 1.004

Notes: Changed the inlet filter before installation. Adjusted the span only.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 25, 2024	Previous Calibration:	N/A
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	7:41	End Time (MST):	10:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	1.0	----	Correlation Coefficient	0.999987	<b>≥0.995</b>
800.4	800.0	1.0005	Slope	0.999272	<b>0.90 - 1.10</b>
399.7	398.9	1.0020	Intercept	-0.306089	<b>+/-30</b>
200.4	198.3	1.0104			



SO2 Calibration Plot

Date: May 25, 2024

Location: Mannix







# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

### Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	May 21, 2024	Last Cal Date:	April 19, 2024
Start time (MST):	10:33	End time (MST):	13:45
Reason:	Removal		

### Calibration Standards

Cal Gas Concentration:	4.96	ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	DT0037363			
Removed Cal Gas Conc:	4.96	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1845
ZAG Make/Model:	API T701H		Serial Number:	832

### Analyzer Information

Analyzer make:	Thermo 43iQTL	Analyzer serial #:	1200326169
Converter make:	Global	Converter serial #:	2022225
Analyzer Range	0 - 100 ppb	Converter Temp:	350 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995684		Backgd or Offset:	1.23	1.23
Calibration intercept:	-0.017567		Coeff or Slope:	0.978	0.978

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4919	80.6	80.0	80.7	0.991
As found Mid point	4960	40.3	40.0	40.5	0.987
As found Low point	4980	20.2	20.0	19.9	1.007
New cylinder response					
Baseline Corr As found:	80.7	Prev response:	79.60	*% change:	1.4%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	1.010837	AF Intercept:	-0.097821
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999969	* = > +/-5% change initiates investigation	

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					
SO2 Scrubber Check	4920	80.0	800.0		----
Date of last scrubber change:				Ave Corr Factor	
Date of last converter efficiency test:					


Notes: Removal calibration for station replacement. Completed the SO2 scrubber check after the third as found point.

Calibration Performed By: Max Farrell

H<sub>2</sub>S Calibration Plot

Date: May 21, 2024

Location: Mannix





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

### Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	May 24, 2024	Last Cal Date:	N/A
Start time (MST):	11:06	End time (MST):	15:11
Reason:	Install		

### Calibration Standards

Cal Gas Concentration:	4.96	ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	DT0037363			
Removed Cal Gas Conc:	4.96	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	5470
ZAG Make/Model:	API T701		Serial Number:	361

### Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	1200326169
Converter make:	Global	Converter serial #:	2022-225
Analyzer Range	0 - 100 ppb	Converter Temp:	325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		1.001686	Backgd or Offset:		1.2
Calibration intercept:		-0.377602	Coeff or Slope:		1.009

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4919	80.6	80.0	79.8	1.002
Mid point	4960	40.3	40.0	39.7	1.007
Low point	4980	20.2	20.0	19.3	1.038
As left zero	5000	0.0	0.0	0.1	----
As left span	4919	80.6	80.0	81.5	0.981
SO2 Scrubber Check	4920	80.3	803.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	1.016

Date of last converter efficiency test:

Notes: Changed the inlet filter. Completed SO2 scrubber check after calibrator zero. Turned off the pressure compensation to stop the random spikes. Adjusted the span. As left span spiked due to low pressure from zero air, increased the zero air pressure to fix the span.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association


## H<sub>2</sub>S Calibration Summary

### Station Information

Calibration Date:	May 24, 2024	Previous Calibration:	N/A
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	11:06	End Time (MST):	15:11
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1200326169

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999932	≥0.995
80.0	79.8	1.0020	Slope	1.001686	0.90 - 1.10
40.0	39.7	1.0069	Intercept	-0.377602	+/-3
20.0	19.3	1.0382			



H<sub>2</sub>S Calibration Plot

Date: May 24, 2024

Location: Mannix





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	May 7, 2024	Last Cal Date:	April 4, 2024
Start time (MST):	9:00	End time (MST):	13:30
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	
CH4 Cal Gas Conc.	504.9 ppm	CH4 Equiv Conc.	1076.6 ppm
C3H8 Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	504.9 ppm	CH4 Equiv Conc.	1076.6 ppm
Removed C3H8 Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
Zero Air Gen model:	API T701	Serial Number:	5613

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1170050130
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.73E-04	2.52E-04	5.89E-05	5.35E-05
CH4 Retention time:	14.4	14.4	155276	170930
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	17.23	17.15	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.15	Prev response	17.21	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.0	17.23	17.20	1.002
Mid point	4960	40.0	8.61	8.58	1.003
Low point	4980	20.0	4.31	4.27	1.008
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.0	17.23	16.97	1.015
Average Correction Factor					1.004

Notes: Changed the inlet filter after as founds. Readings became very unstable after as founds, possibly an issue with the actuator. Adjusted the span only. See docit note for more info.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	9.15	9.05	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.05	Prev response	9.13	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.0	9.15	9.10	1.005
Mid point	4960	40.0	4.57	4.54	1.007
Low point	4980	20.0	2.29	2.27	1.008
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.0	9.15	8.94	1.023
Average Correction Factor					1.007

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	8.08	8.11	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.11	Prev response	8.08	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.0	8.08	8.10	0.998
Mid point	4960	40.0	4.04	4.04	1.000
Low point	4980	20.0	2.02	2.01	1.007
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.0	8.08	8.03	1.006
Average Correction Factor					1.001

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999340	0.998663
THC Cal Offset:	-0.004400	-0.012800
CH <sub>4</sub> Cal Slope:	1.001641	1.003084
CH <sub>4</sub> Cal Offset:	-0.007600	-0.009200
NMHC Cal Slope:	0.997707	0.995196
NMHC Cal Offset:	0.002600	-0.004600

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 4, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:00	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999997	<span style="color: red;">≥0.995</span>
17.23	17.20	1.0017	Slope	0.998663	<span style="color: red;">0.90 - 1.10</span>
8.61	8.58	1.0034	Intercept	-0.012800	<span style="color: red;">+/-0.5</span>
4.31	4.27	1.0076			







# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 4, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:00	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999994	<i>≥0.995</i>
8.08	8.10	0.9976	Slope	1.003084	<i>0.90 - 1.10</i>
4.04	4.04	0.9998	Intercept	-0.009200	<i>+/-0.5</i>
2.02	2.01	1.0068			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 4, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:00	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999999	<i>≥0.995</i>
9.15	9.10	1.0050	Slope	0.995196	<i>0.90 - 1.10</i>
4.57	4.54	1.0068	Intercept	-0.004600	<i>+/-0.5</i>
2.29	2.27	1.0083			



NMHC Calibration Plot

Date: May 7, 2024

Location: Mannix





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	May 16, 2024	Last Cal Date:	May 7, 2024
Start time (MST):	9:33	End time (MST):	11:21
Reason:	Cylinder Change		

### Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	
CH <sub>4</sub> Cal Gas Conc.	504.9 ppm	CH <sub>4</sub> Equiv Conc.	1076.6 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH <sub>4</sub> Conc.	504.9 ppm	CH <sub>4</sub> Equiv Conc.	1076.6 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
Zero Air Gen model:	API T701	Serial Number:	5613

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1170050130
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.52E-04	2.52E-04	NMHC SP Ratio:	5.35E-05	5.35E-05
CH <sub>4</sub> Retention time:	14.4	14.4	NMHC Peak Area:	170930	170930
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	17.23	17.25	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.25	Prev response	17.19	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.0	17.23	17.18	1.003
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.003

Notes: Changed N2 cylinder only.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	9.15	9.25	0.989
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.25	Prev response	9.10	*% change	1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.0	9.15	9.19	0.996
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 0.996

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	8.08	8.00	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.00	Prev response	8.09	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.0	8.08	7.99	1.011
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.011

### Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998663	0.997214
THC Cal Offset:	-0.012800	0.000000
CH <sub>4</sub> Cal Slope:	1.003084	0.989305
CH <sub>4</sub> Cal Offset:	-0.009200	0.000000
NMHC Cal Slope:	0.995196	1.004307
NMHC Cal Offset:	-0.004600	0.000000

Calibration Performed By: Mohammed Kashif

NMHC Calibration Plot

Date: May 16, 2024

Location: Mannix





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	May 21, 2024	Last Cal Date:	May 7, 2024
Start time (MST):	9:06	End time (MST):	10:27
Reason:	Removal		

### Calibration Standards

Gas Cert Reference:	XCO268098	Cal Gas Expiry Date:	
CH4 Cal Gas Conc.	504.9 ppm	CH4 Equiv Conc.	1076.6 ppm
C3H8 Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	504.9 ppm	CH4 Equiv Conc.	1076.6 ppm
Removed C3H8 Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	621
Zero Air Gen model:	API T701	Serial Number:	5613

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1170050130
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.52E-04	2.52E-04	NMHC SP Ratio:	5.35E-05	5.35E-05
CH4 Retention time:	14.4	14.4	NMHC Peak Area:	170930	170930
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	17.23	17.13	1.005
As found Mid point	4960	40.0	8.61	8.54	1.009
As found Low point	4980	20.0	4.31	4.28	1.006
New cylinder response					
Baseline Corr AF:	17.13	Prev response	17.19	*% change	-0.3%
Baseline Corr 2nd AF:	8.54	AF Slope:	0.994444	AF Intercept:	-0.006000
Baseline Corr 3rd AF:	4.28	AF Correlation:	0.999997	<i>* = &gt; +/-5% change initiates investigation</i>	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Notes: Removal calibration for station replacement.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	9.15	9.20	0.994
As found Mid point	4960	40.0	4.57	4.59	0.996
As found Low point	4980	20.0	2.29	2.30	0.993
New cylinder response					
Baseline Corr AF:	9.20	Prev response	9.10	*% change	1.1%
Baseline Corr 2nd AF:	4.59	AF Slope:	1.005953	AF Intercept:	-0.000400
Baseline Corr 3rd AF:	2.30	AF Correlation:	0.999999	* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	8.08	7.93	1.019
As found Mid point	4960	40.0	4.04	3.95	1.024
As found Low point	4980	20.0	2.02	1.98	1.021
New cylinder response					
Baseline Corr AF:	7.93	Prev response	8.09	*% change	-2.1%
Baseline Corr 2nd AF:	3.95	AF Slope:	0.981425	AF Intercept:	-0.005400
Baseline Corr 3rd AF:	1.98	AF Correlation:	0.999993	* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998663	
THC Cal Offset:	-0.012800	
CH <sub>4</sub> Cal Slope:	1.003084	
CH <sub>4</sub> Cal Offset:	-0.009200	
NMHC Cal Slope:	0.995196	
NMHC Cal Offset:	-0.004600	


Calibration Performed By: Max Farrell



NMHC Calibration Plot

Date: May 21, 2024

Location: Mannix





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	May 25, 2024	Last Cal Date:	N/A
Start time (MST):	7:41	End time (MST):	10:00
Reason:	Install		

### Calibration Standards

Gas Cert Reference:	CC408659	Cal Gas Expiry Date:	January 6, 2030
CH4 Cal Gas Conc.	507.2 ppm	CH4 Equiv Conc.	1057.8 ppm
C3H8 Cal Gas Conc.	200.2 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH4 Conc.	507.2 ppm	CH4 Equiv Conc.	1057.8 ppm
Removed C3H8 Conc.	200.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	5470
Zero Air Gen model:	API T701	Serial Number:	361

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1170050130
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:		2.8E-04	NMHC SP Ratio:	5.39E-05
CH4 Retention time:		14.6	NMHC Peak Area:	164047
Zero Chromatogram:		ON	Flat Baseline:	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	16.99	16.93	1.003
Mid point	4960	40.1	8.48	8.44	1.005
Low point	4980	20.1	4.25	4.21	1.011
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	16.99	16.95	1.002
Average Correction Factor					1.006

Notes: Changed the inlet filter before the install. Adjusted the span only.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	8.84	8.77	1.008
Mid point	4960	40.1	4.42	4.37	1.009
Low point	4980	20.1	2.21	2.18	1.015
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	8.84	8.78	1.007
Average Correction Factor					1.011

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	8.15	8.16	0.998
Mid point	4960	40.1	4.07	4.07	1.000
Low point	4980	20.1	2.04	2.03	1.006
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	8.15	8.17	0.997
Average Correction Factor					1.001

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		0.997154
THC Cal Offset:		-0.014732
CH <sub>4</sub> Cal Slope:		1.002564
CH <sub>4</sub> Cal Offset:		-0.008063
NMHC Cal Slope:		0.992300
NMHC Cal Offset:		-0.006668

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 25, 2024	Previous Calibration:	N/A
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	7:41	End Time (MST):	10:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999996	<span style="color: red;">≥0.995</span>
16.99	16.93	1.0034	Slope	0.997154	<span style="color: red;">0.90 - 1.10</span>
8.48	8.44	1.0049	Intercept	-0.014732	<span style="color: red;">+/-0.5</span>
4.25	4.21	1.0107			





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 25, 2024	Previous Calibration:	N/A
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	7:41	End Time (MST):	10:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999995	<i>≥0.995</i>
8.15	8.16	0.9981	Slope	1.002564	<i>0.90 - 1.10</i>
4.07	4.07	0.9997	Intercept	-0.008063	<i>+/-0.5</i>
2.04	2.03	1.0064			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 25, 2024	Previous Calibration:	N/A
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	7:41	End Time (MST):	10:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999997	<i>≥0.995</i>
8.84	8.77	1.0082	Slope	0.992300	<i>0.90 - 1.10</i>
4.42	4.37	1.0094	Intercept	-0.006668	<i>+/-0.5</i>
2.21	2.18	1.0147			



NMHC Calibration Plot

Date: May 25, 2024

Location: Mannix





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS06**  
**PATRICIA MCINNES**  
**MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	May 3, 2024	Last Cal Date:	April 23, 2024
Start time (MST):	10:17	End time (MST):	13:50
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.78	ppm	Cal Gas Exp Date: September 9, 2024
Cal Gas Cylinder #:	AAL070632		
Removed Cal Gas Conc:	49.78	ppm	Rem Gas Exp Date:
Removed Gas Cyl #:			Diff between cyl:
Calibrator Model:	API T700		Serial Number: 3566
Zero Air Gen Model:	API T701		Serial Number: 5608

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1160290013
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004978	1.005521	Backgd or Offset:	17.7	17.7
Calibration intercept:	1.779365	1.439462	Coeff or Slope:	0.922	0.922

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4919.7	80.3	799.5	802.4	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	802.2	Previous response	805.2	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4919.7	80.3	799.5	805.0	0.993
Mid point	4959.8	40.2	400.2	403.7	0.991
Low point	4979.9	20.1	200.1	204.3	0.980
As left zero	5000	0.0	0.0	0.4	----
As left span	4919.7	80.3	799.5	805.3	0.993
				Average Correction Factor:	0.988

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 23, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:17	End Time (MST):	13:50
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999988
799.5	805.0	0.9931	Slope	1.005521
400.2	403.7	0.9914	Intercept	1.439462
200.1	204.3	0.9795		
				≥0.995
				0.90 - 1.10
				+/-30



SO2 Calibration Plot

Date: May 3, 2024

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	May 10, 2024	Last Cal Date:	April 18, 2024
Start time (MST):	8:21 AM	End time (MST):	12:30
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.328	ppm	Cal Gas Exp Date:	February 14, 2025
Cal Gas Cylinder #:	CC506659			
Removed Cal Gas Conc:	5.328	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

### Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb	Converter Temp:	825 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996746	1.000456	Backgd or Offset:	1.98	1.98
Calibration intercept:	0.460147	0.180275	Coeff or Slope:	1.170	1.170

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4925	75.1	80.0	79.9	1.003
As found Mid point	4963	37.5	40.0	40.0	1.001
As found Low point	4981	18.8	20.0	20.2	0.997
New cylinder response					
Baseline Corr As found:	79.8	Prev response:	80.22	*% change:	-0.5%
Baseline Corr 2nd AF pt:	39.9	AF Slope:	0.996746	AF Intercept:	0.160146
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999997	<i>* = +/-5% change initiates investigation</i>	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4925	75.1	80.0	80.2	0.998
Mid point	4963	37.5	40.0	40.3	0.991
Low point	4981	18.8	20.0	20.1	0.997
As left zero	5000	0.0	0.0	0.2	----
As left span	4925	75.1	80.0	79.2	1.010
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	<b>0.995</b>
Date of last converter efficiency test:					

Notes: Changed the inlet filter and the external pump after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association


## TRS Calibration Summary

### Station Information

Calibration Date:	May 10, 2024	Previous Calibration:	April 18, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:21	End Time (MST):	12:30
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999989	$\geq 0.995$
80.0	80.2	0.9978	Slope	1.000456	$0.90 - 1.10$
40.0	40.3	0.9915	Intercept	0.180275	$\pm 3$
20.0	20.1	0.9967			



TRS Calibration Plot

Date: May 10, 2024

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	May 11, 2024	Last Cal Date:	May 10, 2024
Start time (MST):	8:50	End time (MST):	11:32
Reason:	Maintenance		

### Calibration Standards

Cal Gas Concentration:	5.328	ppm	Cal Gas Exp Date:	February 14, 2025
Cal Gas Cylinder #:	CC506659			
Removed Cal Gas Conc:	5.328	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

### Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	621
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000456	0.995746	Backgd or Offset:	1.98	1.92
Calibration intercept:	0.180275	0.220146	Coeff or Slope:	1.170	1.115

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	<i>* = &gt; +/-5% change initiates investigation</i>	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4925	75.1	80.0	79.9	1.002
Mid point	4963	37.5	40.0	40.0	0.999
Low point	4981	18.8	20.0	20.2	0.992
As left zero	5000	0.0	0.0	0.2	----
As left span	4925	75.1	80.0	78.3	1.022
SO2 Scrubber Check	4920	80.3	803.0		----
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	0.997
Date of last converter efficiency test:					

Notes: Swapping out the converter. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association


## TRS Calibration Summary

### Station Information

Calibration Date:	May 11, 2024	Previous Calibration:	May 10, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:50	End Time (MST):	11:32
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	1.000000	$\geq 0.995$
80.0	79.9	1.0016	Slope	0.995746	$0.90 - 1.10$
40.0	40.0	0.9989	Intercept	0.220146	$\pm 3$
20.0	20.2	0.9918			






TRS Calibration Plot

Date: May 11, 2024

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	May 3, 2024	Last Cal Date:	April 23, 2024
Start time (MST):	10:17 AM	End time (MST):	13:50
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH4 Cal Gas Conc.	501.6 ppm	CH4 Equiv Conc.	1066.2 ppm
C3H8 Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	501.6 ppm	CH4 Equiv Conc.	1066.2 ppm
Removed C3H8 Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
Zero Air Gen model:	API T701	Serial Number:	4602

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1118148495
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.26E-04	2.26E-04	NMHC SP Ratio:	4.92E-05
CH4 Retention time:	14.4	14.4	NMHC Peak Area:	184279
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	---
As found High point	4920	80.3	17.12	17.16	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.15	Prev response	17.14	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	---
High point	4920	80.3	17.12	17.10	1.001
Mid point	4960	40.2	8.57	8.56	1.001
Low point	4980	20.1	4.29	4.34	0.988
As left zero	5000	0.0	0.00	0.01	---
As left span	4920	80.3	17.12	17.04	1.005
Average Correction Factor					0.997

Notes: Changed the inlet filter and the N2 cylinder after as founds. Adjusted the span only.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	9.07	9.15	0.992
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.14	Prev response	9.10	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	9.07	9.05	1.002
Mid point	4960	40.2	4.54	4.55	0.997
Low point	4980	20.1	2.27	2.31	0.981
As left zero	5000	0.0	0.00	0.01	----
As left span	4920	80.3	9.07	9.00	1.008
Average Correction Factor					0.993

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919.7	80.3	8.06	8.01	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.01	Prev response	8.04	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	8.06	8.05	1.001
Mid point	4960	40.2	4.03	4.01	1.006
Low point	4980	20.1	2.02	2.02	0.996
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	8.06	8.03	1.003
Average Correction Factor					1.001

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999948	0.997419
THC Cal Offset:	0.023034	0.024244
CH <sub>4</sub> Cal Slope:	0.998696	0.998370
CH <sub>4</sub> Cal Offset:	-0.002350	0.000050
NMHC Cal Slope:	1.001388	0.996687
NMHC Cal Offset:	0.024584	0.023994

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 23, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:17	End Time (MST):	13:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999986	<span style="color: red;">≥0.995</span>
17.12	17.10	1.0014	Slope	0.997419	<span style="color: red;">0.90 - 1.10</span>
8.57	8.56	1.0009	Intercept	0.024244	<span style="color: red;">+/-0.5</span>
4.29	4.34	0.9880			





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 23, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:17	End Time (MST):	13:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999988	<i>≥0.995</i>
8.06	8.05	1.0010	Slope	0.998370	<i>0.90 - 1.10</i>
4.03	4.01	1.0057	Intercept	0.000050	<i>+/-0.5</i>
2.02	2.02	0.9962			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 23, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:17	End Time (MST):	13:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999967	<i>≥0.995</i>
9.07	9.05	1.0018	Slope	0.996687	<i>0.90 - 1.10</i>
4.54	4.55	0.9967	Intercept	0.023994	<i>+/-0.5</i>
2.27	2.31	0.9808			



NMHC Calibration Plot

Date: May 3, 2024

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Patricia McInnes  
 Station number: AMS 06  
 Calibration Date: May 8, 2024  
 Last Cal Date: April 10, 2024  
 Start time (MST): 9:45 AM  
 End time (MST): 14:27  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T30YCWN  
 NOX Cal Gas Conc: 47.94 ppm  
 Removed Cylinder #: N/A  
 Removed Gas NOX Conc: 47.94 ppm  
 NOX gas Diff:  
 Calibrator Model: Teledyne API T700  
 ZAG make/model: Teledyne API T701  
 Cal Gas Expiry Date: April 11, 2025  
 NO Cal Gas Conc: 46.39 ppm  
 Removed Gas Exp Date: N/A  
 Removed Gas NO Conc: 46.39 ppm  
 NO gas Diff:  
 Serial Number: 3566  
 Serial Number: 4602

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.3	0.5	-0.2	----	----
AF High point	4914	86.2	826.5	799.7	26.7	826.1	797.4	28.7	1.0008	1.0036
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO <sub>x</sub> = 827.2 ppb	NO = 801.9 ppb	<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -0.2%
Baseline Corr 1st pt	NO <sub>x</sub> = 825.8 ppb	NO = 796.9 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -0.6%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :	NO SI:	NO Int:
			As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1172750022

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.997990	0.997948
NO <sub>x</sub> Cal Offset:	2.376066	2.316043
NO Cal Slope:	1.000731	0.999015
NO Cal Offset:	1.562605	1.163165
NO <sub>2</sub> Cal Slope:	1.000336	0.999537
NO <sub>2</sub> Cal Offset:	-0.255207	0.049129

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.836	0.836	NO bkgnd or offset:	3.3	3.3
NOX coeff or slope:	0.987	0.987	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	155.4	155.4

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.4	-0.1	----	----
High point	4914	86.2	826.5	799.7	26.7	826.1	799.6	26.6	1.0004	1.0002
Mid point	4957	43.1	413.2	399.9	13.4	415.6	401.4	14.3	0.9943	0.9962
Low point	4978	21.6	207.1	200.4	6.7	211.1	201.9	9.1	0.9811	0.9927
As left zero	5000	0.0	0.0	0.0	0.0	0.5	0.6	-0.1	----	----
As left span	4914	86.2	826.5	398.9	427.6	826.8	398.9	427.8	0.9996	1.0000
Average Correction Factor									0.9920	0.9963

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	798.2	397.7	427.2	427.0	1.0005	99.9%
Mid GPT point	798.2	599.5	225.4	225.4	1.0001	100.0%
Low GPT point	798.2	701.0	123.9	124.1	0.9986	100.1%
Average Correction Factor					0.9997	100.0%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 10, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:45	End Time (MST):	14:27
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999973	≥0.995
826.5	826.1	1.0004	Slope	0.997948	0.90 - 1.10
413.2	415.6	0.9943	Intercept	2.316043	+/-20
207.1	211.1	0.9811			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 10, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:45	End Time (MST):	14:27
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999999	≥0.995
427.2	427.0	1.0005	Slope	0.999537	0.90 - 1.10
225.4	225.4	1.0001	Intercept	0.049129	+/-20
123.9	124.1	0.9986			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 10, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:45	End Time (MST):	14:27
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.4	----	Correlation Coefficient	0.999995	<span style="color: red;">≥0.995</span>
799.7	799.6	1.0002	Slope	0.999015	<span style="color: red;">0.90 - 1.10</span>
399.9	401.4	0.9962	Intercept	1.163165	<span style="color: red;">+/-20</span>
200.4	201.9	0.9927			



NO<sub>x</sub> Calibration Plot

Date: May 8, 2024

Location: Patricia McInnes





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

### Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	May 6, 2024	Last Cal Date:	April 12, 2024
Start time (MST):	10:06	End time (MST):	13:20
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3566
Calibrator Make/Model:	API T700	Serial Number:	4602
ZAG Make/Model:	API T701		

### Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006514	1.003343	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	-0.840000	-0.060000	Coeff or Slope:	1.026	1.026

### O<sub>3</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-0.2	----
As found High point	5000	1303.0	400.0	401.6	0.996
As found Mid point					
As found Low point					
Baseline Corr As found:	401.8	Previous response	401.8	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.1	----
High point	5000	1303.0	400.0	401.3	0.997
Mid point	5000	966.5	200.0	200.7	0.997
Low point	5000	794.3	100.0	100.0	1.000
As left zero	5000	800.0	0.0	0.1	----
As left span	5000	1303.0	400.0	403.8	0.991
Average Correction Factor					<b>0.998</b>

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association


## O<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 6, 2024	Previous Calibration:	April 12, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:06	End Time (MST):	13:20
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999999	<span style="color: red;">≥0.995</span>
400.0	401.3	0.9968	Slope	1.003343	<span style="color: red;">0.90 - 1.10</span>
200.0	200.7	0.9965	Intercept	-0.060000	<span style="color: red;">+/- 5</span>
100.0	100.0	1.0000			



O<sub>3</sub> Calibration Plot

Date: May 6, 2024

Location: Patricia McInnes







# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Patricia McInnes Station number: AMS 06  
 Calibration Date: May 23, 2024 Last Cal Date: April 22, 2024  
 Start time (MST): 8:42 End time (MST): 9:13

Analyzer Make: API T640 S/N: 766  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388755  
 Temp/RH standard: Alicat FP-25BT S/N: 388755

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	8.8	8.4	8.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	726.8	728.3	726.8	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.04	5.14	5.04	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	40	----	40	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	6.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: 11-23-2023  
 Lot No.: 100128-050-035

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: \_\_\_\_\_ March 22, 2024  
 Date Disposable Filter Changed: \_\_\_\_\_ May 23, 2024

Post- maintenance Zero Verification: PM w/ HEPA: \_\_\_\_\_ 0 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_ April 13, 2023  
 Date RH/T Sensor Cleaned: \_\_\_\_\_ April 13, 2023

Notes: Quarterly calibrations completed in March. Leak check passed, no adjustments made.

Calibration by: Max Farrell



# Wood Buffalo Environmental Association

## Nt - NOX - NH3 Calibration Report

### Station Information

Station Name:	Patricia McInnes	Station number:	
NOX Cal Date:	May 14, 2024	Last Cal Date:	April 11, 2024
Start time (MST):	8:12	End time (MST):	12:30
NH3 Cal Date:	May 15, 2024	Last Cal Date:	April 11, 2024
Start time (MST):	13:49	End time (MST):	16:00
Reason:	Routine		

### Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	N/A
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	N/A
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.3	ppm	NH3 Gas Cylinder #:	EB0108520
			NH3 Cal Gas Expiry:	August 22, 2024
Removed NH3 Conc:	76.3	ppm	Removed Cylinder #:	N/A
NH3 gas Diff:			Removed cyl Expiry:	N/A
Calibrator Model:	API T700		Serial Number:	3566
ZAG make/model:	API T701		Serial Number:	4602

### Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	808
Converter model:	API T501	Converter serial #:	484
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	4.80
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	26.6

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.854	0.854	Nt coefficient:	0.849	0.849
NOX coefficient:	0.848	0.848	NO bkgrnd:	-1.0	-1.0
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	-0.6	-0.6
NH3 coefficient:	0.896	0.896	Nt bkgrnd:	5.0	5.0

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.993922	0.993729
NO <sub>x</sub> Cal Offset:	1.997132	1.117125
NO Cal Slope:	0.997329	0.994355
NO Cal Offset:	0.703361	0.744064
NO <sub>2</sub> Cal Slope:	0.997014	1.000404
NO <sub>2</sub> Cal Offset:	-0.514270	-0.068951
NH3 Cal Slope:	0.990318	1.000753
NH3 Cal Offset:	4.179726	0.220004
Nt Cal Slope:	0.997067	1.006983
Nt Cal Offset:	5.735735	1.679101



# Wood Buffalo Environmental Association

## NO<sub>x</sub> - NO - NO<sub>2</sub> Calibration Report

### NO<sub>x</sub> / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>	Baseline corr NO Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.2	-2.0	----	----
As found span	4914	86.2	826.5	799.7	826.5	822.0	792.7	822.0	1.0054	1.0089
AF GPT span										
new NO cyl rp										

Baseline Corr As Fd      Nt = 824 ppb      NO<sub>x</sub> = 822.0 ppb      NO = 792.5 ppb  
 Previous Response      Nt = 829.76 ppb      NO<sub>x</sub> = 823.4 ppb      NO = 798.3 ppb

\*Percent Change      Nt<sub>(NO)</sub> = -0.7%

\*Percent Change      NO<sub>x</sub> = -0.2%

\*Percent Change      NO = -0.7%

\*\*NO<sub>x</sub> Δ (NO to GPT response) =

\* \* = > +/-2% difference initiates investigation

\* = > +/-5% change initiates investigation

### NO<sub>x</sub> / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.4	0.4	-1.0	----	----
High point	4914	86.2	826.5	799.7	826.5	822.5	795.9	821.0	1.0048	1.0048
Mid point	4957	43.1	413.2	399.9	413.2	410.7	398.2	409.6	1.0062	1.0042
Low point	4978	21.6	207.1	200.4	207.1	208.6	200.6	208.2	0.9929	0.9991
Average Correction Factor									1.0013	1.0027

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found zero	----	----	0.0	-0.2	----	----
Calibration zero	----	----	0.0	0.0	----	----
High GPT point (400 ppb O3)	792.8	392.0	427.5	427.7	0.9996	100.0%
Mid GPT point (200 ppb O3)	792.8	595.2	224.3	224.2	1.0005	99.9%
Low GPT point (100 ppb O3)	792.8	696.6	122.9	122.9	1.0002	100.0%
Average Correction Factor					1.0001	100.0%



## Wood Buffalo Environmental Association NH<sub>3</sub> - N<sub>T</sub> Calibration Report

### NH<sub>3</sub> As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH3 Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	-2.0	0.0	-2.1	----	----
AF High point	3417	82.6	1800.6	0.0	1800.6	1813.6	11.9	1801.6	0.992	0.998
AF Mid point										
AF Low point										
new NH3 cyl rp										
Baseline Corr As Fd	Nt = 1815.6 ppb	NH3 = 1803.7 ppb							*Percent Change	Nt <sub>(NH3)</sub> = 0.8%
Previous Response	Nt = 1801.1 ppb	NH3 = 1787.4 ppb							*Percent Change	NH3 = 0.9%

\* = > +/-5% change initiates investigation

### NH<sub>3</sub> Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	-1.0	0.4	-1.4	----	----
High point	3417	82.6	1800.6	0.0	1800.6	1813.6	11.9	1801.6	0.993	0.999
Mid point	3454	45.9	1000.5	0.0	1000.5	1009.9	8.4	1001.5	0.991	0.999
Low point	3477	22.9	499.2	0.0	499.2	507.6	5.5	502.0	0.983	0.994
Average Correction Factor									0.9890	0.9976
NH3 Previous Converter Efficiency =	90.8 %									
NH3 Current Converter Efficiency =	90.8 %									

Notes: Changed the inlet filter after as founds. No adjustments made. NH<sub>3</sub> portion of the calibration completed remotely on the 15th due to forest fire evacuation.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association


## Nt Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 11, 2024
Station Name:	Patricia McInnes	Station Number:	
Start Time (MST):	8:12	End Time (MST):	12:30
Analyzer make:	API T201	Analyzer serial #:	808

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-1.0	----	Correlation Coefficient	0.999989	<i>≥0.995</i>
1800.6	1813.6	0.9929	Slope	1.006983	<i>0.90 - 1.10</i>
1000.5	1009.9	0.9907	Intercept	1.679101	<i>+/-20</i>
499.2	507.6	0.9834			





# Wood Buffalo Environmental Association


## NH<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 11, 2024
Station Name:	Patricia McInnes	Station Number:	
Start Time (MST):	8:12	End Time (MST):	12:30
Analyzer make:	API T201	Analyzer serial #:	808

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-1.4	----	Correlation Coefficient	0.999995	≥0.995
1800.6	1801.6	0.9995	Slope	1.000753	0.90 - 1.10
1000.5	1001.5	0.9990	Intercept	0.220004	+/-20
499.2	502.0	0.9944			





# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 14, 2024	Previous Calibration:	April 11, 2024
Station Name:	Patricia McInnes	Station Number:	
Start Time (MST):	8:12	End Time (MST):	12:30
Analyzer make:	API T201	Analyzer serial #:	808

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999988	≥0.995
826.5	822.5	1.0048	Slope	0.993729	0.90 - 1.10
413.2	410.7	1.0062	Intercept	1.117125	+/-20
207.1	208.6	0.9929			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 14, 2024	Previous Calibration:	April 11, 2024
Station Name:	Patricia McInnes	Station Number:	
Start Time (MST):	8:12	End Time (MST):	12:30
Analyzer make:	API T201	Analyzer serial #:	808

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999999	<i>≥0.995</i>
799.7	795.9	1.0048	Slope	0.994355	<i>0.90 - 1.10</i>
399.9	398.2	1.0042	Intercept	0.744064	<i>+/-20</i>
200.4	200.6	0.9991			







# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 14, 2024	Previous Calibration:	April 11, 2024
Station Name:	Patricia McInnes	Station Number:	
Start Time (MST):	8:12	End Time (MST):	12:30
Analyzer make:	API T201	Analyzer serial #:	808

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	1.000000	<i>≥0.995</i>
427.5	427.7	0.9996	Slope	1.000404	<i>0.90 - 1.10</i>
224.3	224.2	1.0005	Intercept	-0.068951	<i>+/-20</i>
122.9	122.9	1.0002			



NO<sub>x</sub> Calibration Plot

Date: May 14, 2024


Location: Patricia McInnes



NH<sub>3</sub> Calibration Plot

Date: May 15, 2024

Location: Patricia McInnes





## **WOOD BUFFALO ENVIRONMENTAL ASSOCIATION**

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS07 ATHABASCA VALLEY MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Athabasca Valley	Station number: AMS07
Calibration Date:	May 3, 2024	Last Cal Date: April 10, 2024
Start time (MST):	10:05	End time (MST): 14:56
Reason:	Routine	

### Calibration Standards

Cal Gas Concentration:	50.06	ppm	Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #:	CC320556		
Removed Cal Gas Conc:	50.06	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	API T700		Serial Number: 3805
Zero Air Gen Model:	API 701H		Serial Number: 198

### Analyzer Information

Analyzer make:	Thermo 43i-LTE	Serial Number: 1507864683
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000104	1.002151	Backgd or Offset:	2.6	2.6
Calibration intercept:	1.984432	2.043866	Coeff or Slope:	0.845	0.845

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	79.8	799.0	800.0	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	800.0	Previous response	801.1	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4920	79.8	799.0	801.3	0.997
Mid point	4960	39.9	399.5	405.0	0.986
Low point	4980	20.0	200.2	203.3	0.985
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	79.8	799.0	801.5	0.997
Average Correction Factor:					0.989

Notes: No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 10, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:05	End Time (MST):	14:56
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.3	----	Correlation Coefficient	0.999965	<b>≥0.995</b>
799.0	801.3	0.9971	Slope	1.002151	<b>0.90 - 1.10</b>
399.5	405.0	0.9864	Intercept	2.043866	<b>+/-30</b>
200.2	203.3	0.9849			



SO2 Calibration Plot

Date: May 3, 2024

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name: Athabasca Valley	Station number: AMS07
Calibration Date: May 29, 2024	Last Cal Date: April 15, 2024
Start time (MST): 8:29	End time (MST): 13:13
Reason: Routine	

### Calibration Standards

Cal Gas Concentration: 5.25 ppm	Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC504080	
Removed Cal Gas Conc: 5.25 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 3805
ZAG Make/Model: API T701H	Serial Number: 198

### Analyzer Information

Analyzer make: Thermo 43i LTE	Analyzer serial #: 1180540018
Converter make: CDN-101	Converter serial #: 551
Analyzer Range: 0 - 100 ppb	Converter Temp: 840 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.013471	1.012031	Backgd or Offset:	2.4	2.4
Calibration intercept:	0.037852	-0.062173	Coeff or Slope:	0.901	0.901

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4925	75.5	79.3	80.4	0.985
As found Mid point	4962	37.7	39.6	40.2	0.982
As found Low point	4981	18.9	19.8	19.9	0.992
New cylinder response					
Baseline Corr As found:	80.5	Prev response:	80.37	*% change:	0.2%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.016358	AF Intercept:	-0.142169
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999992	* = > +/-5% change initiates investigation	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4925	75.5	79.3	80.2	0.989
Mid point	4962	37.7	39.6	40.1	0.988
Low point	4981	18.9	19.9	19.9	0.998
As left zero	5000	0.0	0.0	0.0	----
As left span	4925	75.5	79.3	80.1	0.990
SO2 Scrubber Check	4920	79.2	792.1	0.1	----
Date of last scrubber change:	25-Feb-22			Ave Corr Factor	0.992
Date of last converter efficiency test:	April 22, 2022				

Notes: No adjustments needed.

Calibration Performed By: Mohammed Kashif





# Wood Buffalo Environmental Association


## TRS Calibration Summary

### Station Information

Calibration Date:	May 29, 2024	Previous Calibration:	April 15, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:29	End Time (MST):	13:13
Analyzer make:	Thermo 43i LTE	Analyzer serial #:	1180540018

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999992	$\geq 0.995$
79.3	80.2	0.9889	Slope	1.012031	$0.90 - 1.10$
39.6	40.1	0.9878	Intercept	-0.062173	$\pm 3$
19.9	19.9	0.9978			



TRS Calibration Plot

Date: May 29, 2024

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS 07
Calibration Date:	May 3, 2024	Last Cal Date:	April 10, 2024
Start time (MST):	10:05	End time (MST):	14:56
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	496.0 ppm	CH4 Equiv Conc.	1059.8 ppm
C3H8 Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	496.0 ppm	CH4 Equiv Conc.	1059.8 ppm
Removed C3H8 Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3805
Zero Air Gen model:	Teledyne API T701H	Serial Number:	198

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 12227620777
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.60E-05	2.64E-04	NMHC SP Ratio:	5.34E-05	5.50E-05
CH4 Retention time:	13.4	13.4	NMHC Peak Area:	168530	163554
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.91	16.53	1.023
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.53	Prev response	16.88	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.91	16.92	1.000
Mid point	4960	39.9	8.46	8.43	1.003
Low point	4980	20.0	4.24	4.20	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.91	16.88	1.002
Average Correction Factor					1.004

Notes: Nitrogen cylinder changed out. Span adjusted.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.00	8.79	1.023
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.79	Prev response	8.97	*% change	-2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.00	9.03	0.997
Mid point	4960	39.9	4.50	4.52	0.996
Low point	4980	20.0	2.26	2.25	1.003
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.00	9.03	0.996
Average Correction Factor					0.998

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	7.92	7.74	1.023
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.74	Prev response	7.91	*% change	-2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	7.92	7.90	1.002
Mid point	4960	39.9	3.96	3.91	1.011
Low point	4980	20.0	1.98	1.95	1.019
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	7.92	7.85	1.008
Average Correction Factor					1.011

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.997119	1.001164
THC Cal Offset:	0.013262	-0.023931
CH <sub>4</sub> Cal Slope:	0.999590	0.998822
CH <sub>4</sub> Cal Offset:	-0.002739	-0.020327
NMHC Cal Slope:	0.995046	1.003694
NMHC Cal Offset:	0.016600	-0.004204

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 10, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:05	End Time (MST):	14:56
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999991	<i>&gt;0.995</i>
16.91	16.92	0.9995	Slope	1.001164	<i>0.90 - 1.10</i>
8.46	8.43	1.0031	Intercept	-0.023931	<i>+/-0.5</i>
4.24	4.20	1.0102			





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 10, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:05	End Time (MST):	14:56
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999966	<i>&gt;0.995</i>
7.92	7.90	1.0021	Slope	0.998822	<i>0.90 - 1.10</i>
3.96	3.91	1.0113	Intercept	-0.020327	<i>+/-0.5</i>
1.98	1.95	1.0190			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 10, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:05	End Time (MST):	14:56
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999996	<i>&gt;0.995</i>
9.00	9.03	0.9969	Slope	1.003694	<i>0.90 - 1.10</i>
4.50	4.52	0.9958	Intercept	-0.004204	<i>+/-0.5</i>
2.26	2.25	1.0027			



NMHC Calibration Plot

Date: May 3, 2024

Location: Athabasca Valley







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Athabasca Valley  
 Station number: AMS 07  
 Calibration Date: May 24, 2024  
 Last Cal Date: April 9, 2024  
 Start time (MST): 9:20  
 End time (MST): 14:25  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: DT0033919  
 NOX Cal Gas Conc: 60.10 ppm  
 Removed Cylinder #: N/A  
 Removed Gas NOX Conc: 60.10 ppm  
 NOX gas Diff:  
 Calibrator Model: API T700  
 ZAG make/model: API T701H  
 Cal Gas Expiry Date: January 9, 2032  
 NO Cal Gas Conc: 59.90 ppm  
 Removed Gas Exp Date: N/A  
 Removed Gas NO Conc: 59.90 ppm  
 NO gas Diff:  
 Serial Number: 3805  
 Serial Number: 198

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
AF High point	4933	66.8	803.0	800.3	2.7	803.7	795.6	7.9	0.9991	1.0058
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO <sub>x</sub> = 802.9 ppb	NO = 799.9 ppb	<i>* = &gt; +/-5% change initiates investigation</i>				*Percent Change	NO <sub>x</sub> = 0.1%
Baseline Corr 1st pt	NO <sub>x</sub> = 803.7 ppb	NO = 795.7 ppb	<u>As Found Statistics</u>				*Percent Change	NO = -0.5%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :		NO SI:	NO Int:	
			As found	NO <sub>2</sub> r <sup>2</sup> :		NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1160120024

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.997568	0.998991
NO <sub>x</sub> Cal Offset:	1.871922	2.071875
NO Cal Slope:	0.997614	1.000699
NO Cal Offset:	1.471946	1.891889
NO <sub>2</sub> Cal Slope:	1.006328	1.002535
NO <sub>2</sub> Cal Offset:	1.745224	1.346655

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.077	1.107	NO bkgnd or offset:	7.6	7.8
NOX coeff or slope:	1.003	1.001	NOX bkgnd or offset:	7.8	8.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	212.0	216.6

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
High point	4933	66.8	803.0	800.3	2.7	802.5	801.1	1.3	1.0006	0.9990
Mid point	4966	33.4	401.5	400.2	1.3	406.3	405.5	0.7	0.9882	0.9869
Low point	4983	16.7	200.7	200.1	0.7	203.3	202.4	0.9	0.9874	0.9885
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
As left span	4933	66.8	803.0	393.8	409.2	804.0	393.8	410.2	0.9987	1.0000
Average Correction Factor									0.9921	0.9915

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	798.8	397.0	404.5	406.2	0.9957	100.4%
Mid GPT point	798.8	599.2	202.3	204.6	0.9886	101.2%
Low GPT point	798.8	699.3	102.2	105.4	0.9694	103.2%
Average Correction Factor					0.9846	101.6%

Notes:

Span adjusted slightly.

Calibration Performed By:

Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 24, 2024	Previous Calibration:	April 9, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:20	End Time (MST):	14:25
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999950	≥0.995
803.0	802.5	1.0006	Slope	0.998991	0.90 - 1.10
401.5	406.3	0.9882	Intercept	2.071875	+/-20
200.7	203.3	0.9874			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 24, 2024	Previous Calibration:	April 9, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:20	End Time (MST):	14:25
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999940	≥0.995
404.5	406.2	0.9957	Slope	1.002535	0.90 - 1.10
202.3	204.6	0.9886	Intercept	1.346655	+/-20
102.2	105.4	0.9694			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 24, 2024	Previous Calibration:	April 9, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:20	End Time (MST):	14:25
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999955	<span style="color: red;">≥0.995</span>
800.3	801.1	0.9990	Slope	1.000699	<span style="color: red;">0.90 - 1.10</span>
400.2	405.5	0.9869	Intercept	1.891889	<span style="color: red;">+/-20</span>
200.1	202.4	0.9885			



NO<sub>x</sub> Calibration Plot

Date: May 24, 2024

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	May 6, 2024	Last Cal Date:	April 12, 2024
Start time (MST):	11:00	End time (MST):	N/A
Reason:	As Found		

### Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3805
Calibrator Make/Model:	T700	Serial Number:	198
ZAG Make/Model:	T701H		

### Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999857	N/A	Backgd or Offset:	-1.6
Calibration intercept:	0.500000	N/A	Coeff or Slope:	1.549
				N/A

### O<sub>3</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-2.2	----
As found High point	5000	1523.6	400.0	398.4	0.999
As found Mid point	5000	1088.1	200.0	199.5	0.992
As found Low point	5000	880.5	100.0	100.1	0.978
Baseline Corr As found:	400.6	Previous response	400.4	*% change	0.0%
Baseline Corr 2nd AF pt:	201.7	AF Slope:	1.000057	AF Intercept:	-1.060000
Baseline Corr 3rd AF pt:	102.3	AF Correlation:	0.999963	* = > +/-5% change initiates investigation	

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor


Notes: As founds done before maintance.

Calibration Performed By: Aswin Sasi Kumar

O<sub>3</sub> Calibration Plot

Date: May 6, 2024

Location: Athabasca Valley







# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	May 7, 2024	Last Cal Date:	April 12, 2024
Start time (MST):	9:30	End time (MST):	12:22
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3805
Calibrator Make/Model:	T700	Serial Number:	198
ZAG Make/Model:	T701H		

### Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	N/A	1.001400	Backgd or Offset:	N/A
Calibration intercept:	N/A	0.980000	Coeff or Slope:	N/A
				1.549

### O<sub>3</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <span style="color: red;">Limit = 0.90-1.10</span>
As found zero					
As found High point					
As found Mid point					
As found Low point					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <span style="color: red;">Limit = 0.95-1.05</span>
Calibrator zero	5000	0.0	0.0	0.6	----
High point	5000	1523.6	400.0	401.4	0.997
Mid point	5000	1088.1	200.0	201.3	0.994
Low point	5000	880.5	100.0	101.6	0.984
As left zero	5000	0.0	0.0	0.2	----
As left span	5000	1522.4	400.0	402.0	0.995
Average Correction Factor					0.991

Notes: Reaction cells cleaned, "blue can" replaced, windows and gaskets swapped after multipoint as founds May 6, 2024. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## O<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 12, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:30	End Time (MST):	12:22
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999995	<span style="color: red;">≥0.995</span>
400.0	401.4	0.9965	Slope	1.001400	<span style="color: red;">0.90 - 1.10</span>
200.0	201.3	0.9935	Intercept	0.980000	<span style="color: red;">+/- 5</span>
100.0	101.6	0.9843			



O<sub>3</sub> Calibration Plot

Date: May 7, 2024

Location: Athabasca Valley





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Athabasca Valley Station number: AMS 07  
Calibration Date: May 29, 2024 Last Cal Date: April 15, 2024  
Start time (MST): 12:14 End time (MST): 13:01  
Analyzer Make: API T640 S/N: 645  
Particulate Fraction: PM2.5  
Flow Meter Make/Model: Alicat FP-25BT S/N: 388749  
Temp/RH standard: Alicat FP-25BT S/N: 388749

### Monthly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Measured</u>	<u>As left</u>	<u>Adjusted</u>	<b><i>(Limits)</i></b>
T (°C)	16	15.59	16	<input type="checkbox"/>	<b><i>+/- 2 °C</i></b>
P (mmHg)	728.6	727.06	728.6	<input type="checkbox"/>	<b><i>+/- 10 mmHg</i></b>
Flow (LPM)	4.99	4.984	4.99	<input type="checkbox"/>	<b><i>+/- 0.25 LPM</i></b>
PW% (pump)	36	----	36	<input type="checkbox"/>	<b><i>&gt;80%</i></b>
Zero Verification	PM w/o HEPA: _____	12.4	PM w/ HEPA: _____	0.0	<b><i>&lt;0.2 ug/m3</i></b>

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: October 6, 2024  
Lot No.: 100128-050-042

<u>Parameter</u>	<u>As found</u>	<u>Post maintenance</u>	<u>As left</u>	<u>Adjusted</u>	<b><i>(Limits)</i></b>
PMT Peak Test				<input type="checkbox"/>	<b><i>+/- 0.5</i></b>

Date Optical Chamber Cleaned: February 27, 2024  
Date Disposable Filter Changed: February 27, 2024

Post- maintenance Zero Verification: PM w/ HEPA: \_\_\_\_\_ ***<0.2 ug/m3***

### Annual Maintenance

Date Sample Tube Cleaned: December 5, 2022  
Date RH/T Sensor Cleaned: December 5, 2022

Notes: No adjustments needed.

Calibration by: Mohammed Kashif



# Wood Buffalo Environmental Association

## CO Calibration Report

### Station Information

Station Name:	Athabasca Valley	Station number:	AMS 07
Calibration Date:	May 2, 2024	Last Cal Date:	April 22, 2024
Start time (MST):	11:20	End time (MST):	15:00
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,000	ppm	Cal Gas Exp Date: December 12, 2026
Cal Gas Cylinder #:	LL66942		
Removed Cal Gas Conc:	3,000	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	API T700		Serial Number: 3805
ZAG Make/Model:	API 700H		Serial Number: 198

### Analyzer Information

Analyzer make:	Thermo 48i-TLE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.990260	0.998628	Backgd or Offset:	4.819	4.831
Calibration intercept:	0.092551	0.118562	Coeff or Slope:	1.087	1.095

### CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4933	66.7	40.0	39.7	1.013
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	39.51	Prev response:	39.73	*% change:	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4933	66.7	40.0	40.0	1.000
Mid point	4967	33.3	20.0	20.2	0.989
Low point	4983	16.7	10.0	10.2	0.987
As left zero	5000	0.0	0.0	0.0	----
As left span	4933	66.7	40.0	40.0	1.001
Average Correction Factor					0.992

Notes: Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## CO Calibration Summary

### Station Information

Calibration Date:	May 2, 2024	Previous Calibration:	April 22, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	11:20	End Time (MST):	15:00
Analyzer make:	Thermo 48i-TLE	Analyzer serial #:	1408761381

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999961	<span style="color: red;">≥0.995</span>
40.0	40.0	1.0003	Slope	0.998628	<span style="color: red;">0.90 - 1.10</span>
20.0	20.2	0.9886	Intercept	0.118562	<span style="color: red;">+/-1.5</span>
10.0	10.2	0.9873			



CO Calibration Plot

Date: May 2, 2024

Location: Athabasca Valley





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS08 FORT CHIPEWYAN MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS08
Calibration Date:	May 14, 2024	Last Cal Date:	April 17, 2024
Start time (MST):	12:29	End time (MST):	15:19
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date: January 6, 2030
Cal Gas Cylinder #:	CC196697		
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	Teledyne API T700		Serial Number: 3252
Zero Air Gen Model:	Teledyne API T701		Serial Number: 135

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Serial Number:	1136451241
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999602	0.997919	Backgd or Offset:	1.8	1.8
Calibration intercept:	0.656198	1.195944	Coeff or Slope:	0.989	0.976

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4920	80.3	800.4	809.4	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	799.6	Previous response	804.1	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4920	80.3	800.4	799.6	1.001
Mid point	4960	40.2	400.7	400.6	1.000
Low point	4980	20.1	200.4	203.3	0.986
As left zero	5000	0.0	0.0	-0.3	----
As left span	4920	80.3	800.4	800.1	1.000
Average Correction Factor:					0.996

Notes: Changed out inlet filter after as founds. adjustments made to span high point.

Calibration Performed By: Matthew Courtoreille, brett g, jeremy c.



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 14, 2024	Previous Calibration:	April 17, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	12:29	End Time (MST):	15:19
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999980	<span style="color: red;">≥0.995</span>
800.4	799.6	1.0010	Slope	0.997919	<span style="color: red;">0.90 - 1.10</span>
400.7	400.6	1.0002	Intercept	1.195944	<span style="color: red;">+/-30</span>
200.4	203.3	0.9855			



SO2 Calibration Plot

Date: May 14, 2024

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name: Fort Chipewyan	Station number: AMS 08
Calibration Date: May 27, 2024	Last Cal Date: April 18, 2024
Start time (MST): 9:20	End time (MST): 12:51
Reason: Routine	

### Calibration Standards

Cal Gas Concentration: 4.97 ppm	Cal Gas Exp Date: February 9, 2024
Cal Gas Cylinder #: EY0002276	
Removed Cal Gas Conc: 4.97 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: Teledyne API T700	Serial Number: 3252
ZAG Make/Model: Teledyne API T701	Serial Number: 135

### Analyzer Information

Analyzer make: Thermo 43iQ-TL	Analyzer serial #: 1203169744
Converter make: CDN-101	Converter serial #: 14639
Analyzer Range: 0 - 100 ppb	Converter Temp: 834 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007424	0.998848	Backgd or Offset:	1.0	1.0
Calibration intercept:	0.578976	0.879017	Coeff or Slope:	0.754	0.754

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.5	----
As found High point	4920	80.5	80.0	81.0	0.994
As found Mid point	4960	40.2	40.0	41.1	0.984
As found Low point	4980	20.1	20.0	20.8	0.984
New cylinder response					
Baseline Corr As found:	80.5	Prev response:	81.18	*% change:	-0.8%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.005851	AF Intercept:	0.658998
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999969	<i>* = &gt; +/-5% change initiates investigation</i>	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.5	----
High point	4920	80.5	80.0	80.5	0.994
Mid point	4960	40.2	40.0	41.3	0.967
Low point	4980	20.1	20.0	21.0	0.951
As left zero	5000	0.0	0.0	0.6	----
As left span	4920	80.5	80.0		
SO2 Scrubber Check	4919.7	80.3	803.0	0.0	----
Date of last scrubber change:		7-Mar-22		Ave Corr Factor	<b>0.971</b>
Date of last converter efficiency test:		March 15, 2022		100.7%	efficiency

Notes: Changed inlet filter after as founds. Scrubber checked passed. No adjustments made.

Calibration Performed By: Morgan Voyageur, Sabian Voyageur



# Wood Buffalo Environmental Association


## TRS Calibration Summary

### Station Information

Calibration Date:	May 27, 2024	Previous Calibration:	April 18, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	9:20	End Time (MST):	12:51
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.5	----	Correlation Coefficient	0.999852	$\geq 0.995$
80.0	80.5	0.9939	Slope	0.998848	$0.90 - 1.10$
40.0	41.3	0.9675	Intercept	0.879017	$\pm 3$
20.0	21.0	0.9514			



TRS Calibration Plot

Date: May 27, 2024

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Fort Chipewyan  
 Station number: AMS 08  
 Calibration Date: May 30, 2024  
 Last Cal Date: April 18, 2024  
 Start time (MST): 7:55  
 End time (MST): 12:06  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: DTO046831  
 NOX Cal Gas Conc: 60.20 ppm  
 Removed Cylinder #: DT0046831  
 Removed Gas NOX Conc: 60.20 ppm  
 NOX gas Diff:  
 Calibrator Model: Teledyne API T700  
 ZAG make/model: Teledyne API T701H  
 Cal Gas Expiry Date: January 9, 2032  
 NO Cal Gas Conc: 60.00 ppm  
 Removed Gas Exp Date: January 9, 2032  
 Removed Gas NO Conc: 60.00 ppm  
 NO gas Diff:  
 Serial Number: 3252  
 Serial Number: 135

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	1.4	1.4	0.0	----	----
AF High point	4933	66.7	803.1	800.4	2.7	804.1	798.8	5.2	1.0005	1.0038
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 805.4 ppb	NO = 797.1 ppb	<i>* = &gt; +/-5% change initiates investigation</i>				*Percent Change	NO <sub>x</sub> = -0.3%		
Baseline Corr 1st pt	NO <sub>x</sub> = 802.7 ppb	NO = 797.4 ppb	<u>As Found Statistics</u>				*Percent Change	NO = 0.0%		
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :		Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :		NO SI:	NO Int:			
			As found	NO <sub>2</sub> r <sup>2</sup> :		NO <sub>2</sub> SI:	NO <sub>2</sub> Int:			

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: API T200  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 4460

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000160	1.000873
NO <sub>x</sub> Cal Offset:	2.195593	3.095431
NO Cal Slope:	0.992944	0.997458
NO Cal Offset:	2.314346	2.334079
NO <sub>2</sub> Cal Slope:	0.997333	1.006095
NO <sub>2</sub> Cal Offset:	1.324008	0.117234

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.231	1.231	NO bkgnd or offset:	2.5	2.5
NOX coeff or slope:	1.229	1.229	NOX bkgnd or offset:	2.2	2.2
NO2 coeff or slope:	1.000	2.100	Reaction cell Press:	-1.8	-1.7

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	1.8	1.8	0.0	----	----
High point	4933	66.7	803.1	800.4	2.7	805.9	800.6	5.0	0.9965	0.9998
Mid point	4967	33.3	400.9	399.6	1.3	406.1	400.8	5.3	0.9872	0.9969
Low point	4983	16.7	201.1	200.4	0.7	204.9	203.0	1.9	0.9813	0.9872
As left zero	5000	0.0	0.0	0.0	0.0	1.6	1.7	-0.2	----	----
As left span	4933	66.7	803.1	395.9	407.2	802.0	395.9	406.1	1.0014	1.0000
Average Correction Factor									0.9884	0.9946

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	802.1	397.6	407.2	409.9	0.9933	100.7%
Mid GPT point	802.1	602.7	202.1	202.9	0.9959	100.4%
Low GPT point	802.1	704.2	100.6	101.8	0.9879	101.2%
Average Correction Factor					0.9924	100.8%

Notes: Changed inlet filter after as founds. No adjustment made

Calibration Performed By: Morgan Voyageur





# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	April 18, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	7:55	End Time (MST):	12:06
Analyzer make:	API T200	Analyzer serial #:	4460

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	1.8	----	Correlation Coefficient	0.999983	≥0.995
803.1	805.9	0.9965	Slope	1.000873	0.90 - 1.10
400.9	406.1	0.9872	Intercept	3.095431	+/-20
201.1	204.9	0.9813			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	April 18, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	7:55	End Time (MST):	12:06
Analyzer make:	API T200	Analyzer serial #:	4460

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999994	≥0.995
407.2	409.9	0.9933	Slope	1.006095	0.90 - 1.10
202.1	202.9	0.9959	Intercept	0.117234	+/-20
100.6	101.8	0.9879			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	April 18, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	7:55	End Time (MST):	12:06
Analyzer make:	API T200	Analyzer serial #:	4460

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	1.8	----	Correlation Coefficient	0.999997	<span style="color: red;">≥0.995</span>
800.4	800.6	0.9998	Slope	0.997458	<span style="color: red;">0.90 - 1.10</span>
399.6	400.8	0.9969	Intercept	2.334079	<span style="color: red;">+/-20</span>
200.4	203.0	0.9872			



NO<sub>x</sub> Calibration Plot

Date: May 30, 2024

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS 08
Calibration Date:	May 10, 2024	Last Cal Date:	April 16, 2024
Start time (MST):	9:07	End time (MST):	11:53
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3252
Calibrator Make/Model:	Teledyne API T700	Serial Number:	135
ZAG Make/Model:	Teledyne API T701		

### Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #:	3872
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999086	0.999829	Backgd or Offset:	-1.9	-1.9
Calibration intercept:	0.160000	-0.520000	Coeff or Slope:	1.019	1.011

### O<sub>3</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.4	----
As found High point	5000	913.0	400.0	403.0	0.994
As found Mid point					
As found Low point					
Baseline Corr As found:	402.6	Previous response	399.8	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.8	----
High point	5000	914.7	400.0	399.9	1.000
Mid point	5000	786.4	200.0	199.3	1.004
Low point	5000	701.3	100.0	97.8	1.022
As left zero	5000	0.0	0.0	-0.1	----
As left span	5000	963.3	400.0	401.9	0.995
Average Correction Factor					1.009

Notes: Changed out inlet filter after as found. Made calibration zero run longer. Adjusted span high point.

Calibration Performed By: Matthew C



# Wood Buffalo Environmental Association


## O<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 10, 2024	Previous Calibration:	April 16, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	9:07	End Time (MST):	11:53
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.8	----	Correlation Coefficient	0.999945	<span style="color: red;">≥0.995</span>
400.0	399.9	1.0003	Slope	0.999829	<span style="color: red;">0.90 - 1.10</span>
200.0	199.3	1.0035	Intercept	-0.520000	<span style="color: red;">+/- 5</span>
100.0	97.8	1.0225			



O<sub>3</sub> Calibration Plot

Date: May 10, 2024

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Fort Chipewyan Station number: AMS 08  
 Calibration Date: May 30, 2024 Last Cal Date: April 29, 2024  
 Start time (MST): 10:05 End time (MST): 11:19

Analyzer Make: Teledyne API T640 S/N: 319  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754  
 Temp/RH standard: Alicat FP-25BT S/N: 388754

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.70	11.12	10.70	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	727.10	729.51	727.10	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	4.98	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	38%			<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	6.60	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: 10-Jun-24  
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.91	0.00	10.80	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: May 30, 2024  
 Date Disposable Filter Changed: May 30, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: July 25, 2023  
 Date RH/T Sensor Cleaned: July 25, 2023

Notes: No adjustments needed.

Calibration by: Morgan Voyageur





# Wood Buffalo Environmental Association

## CO Calibration Report

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS 08
Calibration Date:	May 13, 2024	Last Cal Date:	April 29, 2024
Start time (MST):	13:11	End time (MST):	15:59
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #:	ALM014846		
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 3252
ZAG Make/Model:	Teledyne API T701H		Serial Number: 135

### Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003563	0.994765	Backgd or Offset:	-0.015	-0.015
Calibration intercept:	0.210880	0.218907	Coeff or Slope:	1.007	1.003

### CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4934	66.7	40.4	40.8	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	40.67	Prev response:	40.77	*% change:	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4934	66.7	40.4	40.3	1.002
Mid point	4966.7	33.3	20.2	20.4	0.989
Low point	4983.3	16.7	10.1	10.4	0.975
As left zero	5000	0.0	0.0	0.1	----
As left span	2960	40.0	40.4	40.2	1.005
Average Correction Factor					0.989

Notes: Changed inlet filter after as found, Adjustments made to Zero and span high point.

Calibration Performed By: Matt c, jermey c, brett g.



# Wood Buffalo Environmental Association


## CO Calibration Summary

### Station Information

Calibration Date:	May 13, 2024	Previous Calibration:	April 29, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	13:11	End Time (MST):	15:59
Analyzer make:	Teledyne API T300	Analyzer serial #:	3505

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999948	<b>≥0.995</b>
40.4	40.3	1.0018	Slope	0.994765	<b>0.90 - 1.10</b>
20.2	20.4	0.9887	Intercept	0.218907	<b>+/-1.5</b>
10.1	10.4	0.9750			



CO Calibration Plot

Date: May 13, 2024

Location: Fort Chipewyan





# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Fort Chipewyan	Station number:	AMS 08
Calibration Date:	May 16, 2024	Last Cal Date:	April 16, 2024
Start time (MST):	7:20	End time (MST):	11:51
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #:	ALM014846		
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 3252
N2 Gen Make/Model:	Peak Scientific		Serial Number: 135

### Analyzer Information

Analyzer make:	Teledyne API T360	Analyzer serial #:	289
Analyzer Range	0 - 2,000 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003378	1.000610	Backgd or Offset:	-0.063	-0.011
Calibration intercept:	-4.540000	-5.520000	Coeff or Slope:	1.087	1.033

### CO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	3000	0.0	0.0	-5.0	----
As found High Point	2920	80.0	1605.9	1621.0	0.988
As found Mid Point					
As found Low Point					
New cylinder response					
Baseline Corr As found:	1626.0	Prev response:	1606.8	*% change:	1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	3000	0.0	0.0	-0.3	----
High point	2920	80.0	1605.9	1608.4	0.998
Mid point	2960	40.0	802.9	781.9	1.027
Low point	2980	20.0	401.5	399.9	1.004
As left zero	3000	0.0	0.0	-0.5	----
As left span	2960	40.0	802.9	781.7	1.027
Average Correction Factor					1.010

Notes: Changed inlet filter after as found, Adjustments made to zero and span high point.

Calibration Performed By: Matthew c, brett g, jeremy c.



# Wood Buffalo Environmental Association


## CO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date	May 16, 2024	Previous Calibration	April 16, 2024
Station Name	Fort Chipewyan	Station Number	AMS 08
Start Time (MST)	7:20	End Time (MST)	11:51
Analyzer make	Teledyne API T360	Analyzer serial #	289

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999754	<b>≥0.995</b>
1605.9	1608.4	0.9984	Slope	1.000610	<b>0.90 - 1.10</b>
802.9	781.9	1.0269	Intercept	-5.5	<b>+/-20</b>
401.5	399.9	1.0039			



CO<sub>2</sub> Calibration Plot

Date: May 16, 2024

Location: Fort Chipewyan





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS09 BARGE LANDING MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Barge Landing	Station number:	AMS 09
Calibration Date:	May 13, 2024	Last Cal Date:	April 2, 2024
Start time (MST):	10:01	End time (MST):	13:51
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.96	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC151285			
Removed Cal Gas Conc:	49.96	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	3812
Zero Air Gen Model:	APIT701		Serial Number:	4888

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1118148498
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997204	0.996431	Backgd or Offset:	10.1	10.3
Calibration intercept:	0.710524	0.426529	Coeff or Slope:	0.956	0.968

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4919	80.2	801.5	798.1	1.005
As found Mid point	4960	40.1	400.7	398.5	1.006
As found Low point	4980	20.0	199.8	198.5	1.008
New cylinder response					
Baseline Corr As found:	797.8	Previous response	800.0	*% change	-0.3%
Baseline Corr 2nd AF pt:	398.2	AF Slope:	0.995620	AF Intercept:	-0.114112
Baseline Corr 3rd AF pt:	198.2	AF Correlation:	0.999999	<i>* = &gt; +/-5% change initiates investigation</i>	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4919	80.2	801.5	799.2	1.003
Mid point	4960	40.1	400.7	399.1	1.004
Low point	4980	20.0	199.8	200.1	0.999
As left zero	5000	0.0	0.0	0.5	----
As left span	4919	80.2	801.5	800.2	1.002
Average Correction Factor:					1.002

Notes: Changed sample inlet filter after as founds. Span adjusted.

Calibration Performed By: Sean Bala





# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 13, 2024	Previous Calibration:	April 2, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:01	End Time (MST):	13:51
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999998
801.5	799.2	1.0029	Slope	0.996431
400.7	399.1	1.0039	Intercept	0.426529
199.8	200.1	0.9987		
				≥0.995
				0.90 - 1.10
				+/-30



SO2 Calibration Plot

Date: May 13, 2024

Location: Barge Landing





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name: Barge Landing	Station number: AMS 09
Calibration Date: May 14, 2024	Last Cal Date: April 3, 2024
Start time (MST): 8:48	End time (MST): 11:00
Reason: As Found	

### Calibration Standards

Cal Gas Concentration: 5.17 ppm	Cal Gas Exp Date: August 22, 2026
Cal Gas Cylinder #: CC511415	
Removed Cal Gas Conc: 5.17 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 3812
ZAG Make/Model: API T701	Serial Number: 4888

### Analyzer Information

Analyzer make: Thermo 43i-TLE	Analyzer serial #: 1331259320
Converter make: CDN-101	Converter serial #: 519
Analyzer Range: 0 - 100 ppb	Converter Temp: 830 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.990985		Backgd or Offset:	2.830
Calibration intercept:	0.059198		Coeff or Slope:	1.170

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4923	77.4	80.0	78.3	1.022
As found Mid point	4961	38.7	40.0	39.2	1.021
As found Low point	4981	19.3	20.0	19.5	1.024
New cylinder response					
Baseline Corr As found:	78.3	Prev response:	79.38	*% change:	-1.4%
Baseline Corr 2nd AF pt:	39.2	AF Slope:	0.978423	AF Intercept:	-0.001033
Baseline Corr 3rd AF pt:	19.5	AF Correlation:	0.999999	<i>* = &gt; +/-5% change initiates investigation</i>	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point					
Mid point					
Low point					
As left zero					
As left span					
SO2 Scrubber Check					
Date of last scrubber change:				Ave Corr Factor	
Date of last converter efficiency test:					


Notes: Changed inlet filter after as founds. Pump was changed due to low flow.

Calibration Performed By: Sean Bala

TRS Calibration Plot

Date: May 14, 2024

Location: Barge Landing





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name: Barge Landing	Station number: AMS 09
Calibration Date: May 15, 2024	Last Cal Date: April 3, 2024
Start time (MST): 6:45	End time (MST): 9:29
Reason: Routine	

### Calibration Standards

Cal Gas Concentration: 5.17 ppm	Cal Gas Exp Date: August 22, 2026
Cal Gas Cylinder #: CC511415	
Removed Cal Gas Conc: 5.17 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 3812
ZAG Make/Model: API T701	Serial Number: 4888

### Analyzer Information

Analyzer make: Thermo 43i-TLE	Analyzer serial #: 1331259320
Converter make: CDN-101	Converter serial #: 519
Analyzer Range: 0 - 100 ppb	Converter Temp: 830 degC

	<u>Start</u>	<u>Finish</u>	
Calibration slope:	0.990985	0.968001	Backgd or Offset: 2.830
Calibration intercept:	0.059198	0.038801	Coeff or Slope: 1.170

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4923	77.4	80.0	77.5	1.033
Mid point	4961	38.7	40.0	38.8	1.032
Low point	4981	19.3	20.0	19.4	1.029
As left zero	5000	0.0	0.0	0.0	----
As left span	4923	77.4	80.0	77.4	1.034
SO2 Scrubber Check	4920	80.2	802.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	1.031
Date of last converter efficiency test:					

Notes: No as founds as it was done yesterday. Filter was changed yesterday as well. No adjustment made.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association


## TRS Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 3, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	6:45	End Time (MST):	9:29
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999999	$\geq 0.995$
80.0	77.5	1.0328	Slope	0.968001	$0.90 - 1.10$
40.0	38.8	1.0316	Intercept	0.038801	$\pm 3$
20.0	19.4	1.0288			



TRS Calibration Plot

Date: May 15, 2024

Location: Barge Landing





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Barge Landing	Station number:	AMS 09
Calibration Date:	May 8, 2024	Last Cal Date:	April 2, 2024
Start time (MST):	8:48	End time (MST):	12:47
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
Zero Air Gen model:	APIT701	Serial Number:	4888

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1331259521
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.60E-04	2.96E-04	NMHC SP Ratio:	4.50E-05	5.00E-05
CH4 Retention time:	15.20	16.00	NMHC Peak Area:	205331	182732
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	17.12	12.37	<b>1.384</b>
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	12.37	Prev response	17.14	*% change	<b>-38.5%</b>
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	17.12	17.20	0.995
Mid point	4960	40.1	8.56	8.63	0.992
Low point	4980	20.0	4.27	4.29	0.996
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	17.12	17.38	0.985
Average Correction Factor					<b>0.994</b>

Notes: Changed inlet filter after as founds. Chromatogram shifted. Possible cause was the N2 cylinder pressure was around 70 psi. Adjusted span.





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	9.14	8.17	1.119
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.17	Prev response	9.16	*% change	-12.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	9.14	9.17	0.996
Mid point	4960	40.1	4.57	4.60	0.994
Low point	4980	20.0	2.28	2.29	0.997
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	9.14	9.26	0.987
Average Correction Factor					0.996

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	7.98	4.21	1.898
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	4.21	Prev response	7.98	*% change	-89.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	7.98	8.03	0.994
Mid point	4960	40.1	3.99	4.03	0.990
Low point	4980	20.0	1.99	2.00	0.995
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	7.98	8.12	0.983
Average Correction Factor					0.993

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000957	1.005116
THC Cal Offset:	0.000059	0.003673
CH <sub>4</sub> Cal Slope:	0.999506	1.006735
CH <sub>4</sub> Cal Offset:	0.000263	0.002275
NMHC Cal Slope:	1.001926	1.004065
NMHC Cal Offset:	0.000995	0.001198

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 2, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	8:48	End Time (MST):	12:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999997	<i>≥0.995</i>
17.12	17.20	0.9951	Slope	1.005116	<i>0.90 - 1.10</i>
8.56	8.63	0.9923	Intercept	0.003673	<i>+/-0.5</i>
4.27	4.29	0.9959			





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 2, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	8:48	End Time (MST):	12:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999994	≥0.995
7.98	8.03	0.9936	Slope	1.006735	0.90 - 1.10
3.99	4.03	0.9898	Intercept	0.002275	+/-0.5
1.99	2.00	0.9947			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 2, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	8:48	End Time (MST):	12:47
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
9.14	9.17	0.9961	Slope	1.004065	<i>0.90 - 1.10</i>
4.57	4.60	0.9940	Intercept	0.001198	<i>+/-0.5</i>
2.28	2.29	0.9970			



NMHC Calibration Plot

Date: May 8, 2024

Location: Barge Landing





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Barge Landing	Station number:	AMS 09
Calibration Date:	May 9, 2024	Last Cal Date:	May 8, 2024
Start time (MST):	9:27	End time (MST):	12:28
Reason:	Maintenance low response during daily zero and span		

### Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
Zero Air Gen model:	APIT701	Serial Number:	4888

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1331259521
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.96E-04	2.62E-04	NMHC SP Ratio:	5.00E-05	4.46E-05
CH4 Retention time:	16.00	15.00	NMHC Peak Area:	182732	204877
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	17.12	14.72	<b>1.163</b>
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	14.72	Prev response	17.21	*% change	<b>-17.0%</b>
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	17.12	17.13	0.999
Mid point	4960	40.1	8.56	8.53	1.004
Low point	4980	20.0	4.27	4.26	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	17.12	17.03	1.005
Average Correction Factor					<b>1.001</b>

Notes: Noticed there's a leak from zero air line to the analyzer. Replaced the line. Adjusted the span.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	9.14	10.24	<b>0.893</b>
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	10.24	Prev response	9.17	<b>*% change</b>	<b>10.4%</b>
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<b>* = &gt; +/-5% change initiates investigation</b>	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	9.14	9.15	0.999
Mid point	4960	40.1	4.57	4.56	1.003
Low point	4980	20.0	2.28	2.28	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	9.14	9.10	1.005
Average Correction Factor					1.001

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	7.98	4.48	<b>1.781</b>
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	4.48	Prev response	8.04	<b>*% change</b>	<b>-79.4%</b>
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<b>* = &gt; +/-5% change initiates investigation</b>	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	7.98	7.98	1.000
Mid point	4960	40.1	3.99	3.97	1.005
Low point	4980	20.0	1.99	1.99	1.000
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	7.98	7.94	1.006
Average Correction Factor					1.002

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.005116	1.000391
THC Cal Offset:	0.003673	-0.009947
CH <sub>4</sub> Cal Slope:	1.006735	0.999763
CH <sub>4</sub> Cal Offset:	0.002275	-0.004138
NMHC Cal Slope:	1.004065	1.001251
NMHC Cal Offset:	0.001198	-0.005809

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	May 8, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:27	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999994	<i>&gt;0.995</i>
17.12	17.13	0.9995	Slope	1.000391	<i>0.90 - 1.10</i>
8.56	8.53	1.0038	Intercept	-0.009947	<i>+/-0.5</i>
4.27	4.26	1.0011			







# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	May 8, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:27	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
7.98	7.98	1.0000	Slope	0.999763	<i>0.90 - 1.10</i>
3.99	3.97	1.0050	Intercept	-0.004138	<i>+/-0.5</i>
1.99	1.99	1.0002			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	May 8, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:27	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999996	<i>≥0.995</i>
9.14	9.15	0.9988	Slope	1.001251	<i>0.90 - 1.10</i>
4.57	4.56	1.0025	Intercept	-0.005809	<i>+/-0.5</i>
2.28	2.28	1.0014			



NMHC Calibration Plot

Date: May 9, 2024

Location: Barge Landing





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Barge Landing	Station number:	AMS 09
Calibration Date:	May 10, 2024	Last Cal Date:	NA
Start time (MST):	9:51	End time (MST):	12:05
Reason:	Install		

### Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
Zero Air Gen model:	APIT701	Serial Number:	4888

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320038
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	NA	2.40E-04	NMHC SP Ratio:	NA
CH4 Retention time:	NA	14.20	NMHC Peak Area:	NA
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	17.12	17.17	0.997
Mid point	4960	40.1	8.56	8.61	0.994
Low point	4980	20.0	4.27	4.36	0.980
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	17.12	17.57	0.975
Average Correction Factor					0.990

Notes: Install calibrations. Adjusted span only.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFZero))
<i>Baseline Adjusted</i>					
<i>Limit = 0.90-1.10</i>					
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
<i>Baseline Adjusted</i>					
<i>Limit = 0.95-1.05</i>					
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	9.14	9.19	0.994
Mid point	4960	40.1	4.57	4.64	0.984
Low point	4980	20.0	2.28	2.37	0.960
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	9.14	9.58	0.954
Average Correction Factor					0.980

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFZero))
<i>Baseline Adjusted</i>					
<i>Limit = 0.90-1.10</i>					
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
<i>Baseline Adjusted</i>					
<i>Limit = 0.95-1.05</i>					
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	7.98	7.98	1.000
Mid point	4960	40.1	3.99	3.97	1.005
Low point	4980	20.0	1.99	1.98	1.004
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	7.98	7.98	1.000
Average Correction Factor					1.003

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	NA	1.001789
THC Cal Offset:	NA	0.035083
CH <sub>4</sub> Cal Slope:	NA	0.999950
CH <sub>4</sub> Cal Offset:	NA	-0.006540
NMHC Cal Slope:	NA	1.003133
NMHC Cal Offset:	NA	0.042422

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 10, 2024	Previous Calibration:	NA
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:51	End Time (MST):	12:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999979	<i>≥0.995</i>
17.12	17.17	0.9969	Slope	1.001789	<i>0.90 - 1.10</i>
8.56	8.61	0.9936	Intercept	0.035083	<i>+/-0.5</i>
4.27	4.36	0.9801			





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 10, 2024	Previous Calibration:	NA
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:51	End Time (MST):	12:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999994	<i>≥0.995</i>
7.98	7.98	1.0001	Slope	0.999950	<i>0.90 - 1.10</i>
3.99	3.97	1.0047	Intercept	-0.006540	<i>+/-0.5</i>
1.99	1.98	1.0037			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 10, 2024	Previous Calibration:	NA
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:51	End Time (MST):	12:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999902	<i>≥0.995</i>
9.14	9.19	0.9944	Slope	1.003133	<i>0.90 - 1.10</i>
4.57	4.64	0.9842	Intercept	0.042422	<i>+/-0.5</i>
2.28	2.37	0.9600			






NMHC Calibration Plot

Date: May 10, 2024

Location: Barge Landing





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Barge Landing	Station number:	AMS 09
Calibration Date:	May 31, 2024	Last Cal Date:	May 10, 2024
Start time (MST):	8:05	End time (MST):	11:05
Reason:	Cylinder Change		

### Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
Zero Air Gen model:	APIT701	Serial Number:	4888

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320038
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.40E-04	2.36E-04	NMHC SP Ratio:	6.02E-05	5.67E-05
CH4 Retention time:	14.20	13.80	NMHC Peak Area:	151938	161257
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	17.12	17.68	0.968
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.68	Prev response	17.19	*% change	2.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	17.12	17.08	1.003
Mid point	4959	40.1	8.56	8.53	1.003
Low point	4980	20.0	4.27	4.29	0.996
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	17.12	17.08	1.002
Average Correction Factor					1.001

Notes: Nitrogen Cylinder Changed. Spans dipping after nitrogen change. Chromatograms show Span had a slight move. Full calibration done.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	9.14	9.62	0.950
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.62	Prev response	9.21	<b>*% change</b>	4.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<b>* = &gt; +/-5% change initiates investigation</b>	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	9.14	9.08	1.007
Mid point	4959	40.1	4.57	4.55	1.004
Low point	4980	20.0	2.28	2.29	0.995
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	9.14	9.13	1.001
Average Correction Factor					1.002

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	7.98	8.06	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.06	Prev response	7.98	<b>*% change</b>	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<b>* = &gt; +/-5% change initiates investigation</b>	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	7.98	8.00	0.998
Mid point	4959	40.1	3.99	3.98	1.003
Low point	4980	20.0	1.99	2.00	0.997
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	7.98	7.95	1.004
Average Correction Factor					0.999

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001789	0.996814
THC Cal Offset:	0.035083	0.009905
CH <sub>4</sub> Cal Slope:	0.999950	1.001770
CH <sub>4</sub> Cal Offset:	-0.006540	-0.003091
NMHC Cal Slope:	1.003133	0.992658
NMHC Cal Offset:	0.042422	0.012796

Calibration Performed By: Melissa Lemay and Jan Castro



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 31, 2024	Previous Calibration:	May 10, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	8:05	End Time (MST):	11:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999996	<i>≥0.995</i>
17.12	17.08	1.0026	Slope	0.996814	<i>0.90 - 1.10</i>
8.56	8.53	1.0033	Intercept	0.009905	<i>+/-0.5</i>
4.27	4.29	0.9961			





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 31, 2024	Previous Calibration:	May 10, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	8:05	End Time (MST):	11:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999991	≥0.995
7.98	8.00	0.9978	Slope	1.001770	0.90 - 1.10
3.99	3.98	1.0029	Intercept	-0.003091	+/-0.5
1.99	2.00	0.9967			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 31, 2024	Previous Calibration:	May 10, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	8:05	End Time (MST):	11:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999991	<i>≥0.995</i>
9.14	9.08	1.0066	Slope	0.992658	<i>0.90 - 1.10</i>
4.57	4.55	1.0038	Intercept	0.012796	<i>+/-0.5</i>
2.28	2.29	0.9952			



NMHC Calibration Plot

Date: May 31, 2024

Location: Barge Landing





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Barge Landing  
 Station number: AMS 09  
 Calibration Date: May 21, 2024  
 Last Cal Date: April 18, 2024  
 Start time (MST): 9:21  
 End time (MST): 13:44  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T2Y1KDH  
 NOX Cal Gas Conc: 47.38 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 47.38 ppm  
 NOX gas Diff:  
 Calibrator Model: API T700  
 ZAG make/model: Api T701

Cal Gas Expiry Date: November 17, 2026  
 NO Cal Gas Conc: 46.94 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 46.94 ppm  
 NO gas Diff:  
 Serial Number: 3812  
 Serial Number: 4888

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
AF High point	4915	85.3	808.3	800.7	7.5	796.8	785.6	11.3	1.0139	1.0192
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 808.1 ppb		NO = 798.9 ppb			<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -1.4%	
Baseline Corr 1st pt	NO <sub>x</sub> = 797.2 ppb		NO = 785.7 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -1.7%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO r <sup>2</sup> :	NO SI:	NO Int:	
						As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1426262593

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.999233	0.999445
NO <sub>x</sub> Cal Offset:	0.498364	0.698545
NO Cal Slope:	0.998482	0.998311
NO Cal Offset:	-0.583944	-0.423947
NO <sub>2</sub> Cal Slope:	1.006238	1.005694
NO <sub>2</sub> Cal Offset:	0.877238	0.635782

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.101	1.119	NO bkgnd or offset:	10.1	10.2
NOX coeff or slope:	0.998	0.998	NOX bkgnd or offset:	10.4	10.6
NO2 coeff or slope:	1.000	0.998	Reaction cell Press:	182.2	177.1

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
High point	4915	85.3	808.3	800.7	7.5	808.0	799.2	8.8	1.0003	1.0019
Mid point	4957	42.6	403.7	400.0	3.7	404.8	398.6	6.1	0.9973	1.0034
Low point	4979	21.3	201.8	200.0	1.9	203.2	198.8	4.3	0.9932	1.0058
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.1	----	----
As left span	4915	85.3	808.3	418.5	389.8	807.4	418.5	389.0	1.0011	1.0000
Average Correction Factor									0.9970	1.0037

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	796.5	416.3	387.7	390.0	0.9941	100.6%
Mid GPT point	796.5	603.5	200.5	203.0	0.9877	101.2%
Low GPT point	796.5	700.4	103.6	105.5	0.9820	101.8%
Average Correction Factor					0.9880	101.2%

Notes:

Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By:

Sean Bala



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 21, 2024	Previous Calibration:	April 18, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:21	End Time (MST):	13:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999994	<span style="color: red;">≥0.995</span>
808.3	808.0	1.0003	Slope	0.999445	<span style="color: red;">0.90 - 1.10</span>
403.7	404.8	0.9973	Intercept	0.698545	<span style="color: red;">+/-20</span>
201.8	203.2	0.9932			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 21, 2024	Previous Calibration:	April 18, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:21	End Time (MST):	13:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999976	≥0.995
387.7	390.0	0.9941	Slope	1.005694	0.90 - 1.10
200.5	203.0	0.9877	Intercept	0.635782	+/-20
103.6	105.5	0.9820			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 21, 2024	Previous Calibration:	April 18, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:21	End Time (MST):	13:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999999	≥0.995
800.7	799.2	1.0019	Slope	0.998311	0.90 - 1.10
400.0	398.6	1.0034	Intercept	-0.423947	+/-20
200.0	198.8	1.0058			



NO<sub>x</sub> Calibration Plot

Date: May 21, 2024

Location: Barge Landing





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Barge Landing Station number: AMS 09  
 Calibration Date: May 21, 2024 Last Cal Date: April 13, 2024  
 Start time (MST): 10:55 End time (MST): 11:52

Analyzer Make: API T640 S/N: 844  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754  
 Temp/RH standard: Alicat FP-25BT S/N: 388754

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	11.3	10.9	10.0	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	726.0	734.0	719.4	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.97	4.88	4.97	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	35	----	36	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.6	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024  
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	8.0	11.2	11.0	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: \_\_\_\_\_ May 21, 2024  
 Date Disposable Filter Changed: \_\_\_\_\_ May 21, 2024

Post- maintenance Zero Verification: PM w/ HEPA: \_\_\_\_\_ 0.0 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_ August 23, 2023  
 Date RH/T Sensor Cleaned: \_\_\_\_\_ August 23, 2023

Notes: Inlet head looks good. No adjustments. Leak check passed.

Calibration by: Sean Bala



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS11 LOWER CAMP MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Lower Camp	Station number:	AMS 11
Calibration Date:	May 24, 2024	Last Cal Date:	April 24, 2024
Start time (MST):	10:16	End time (MST):	13:32
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.25	ppm	Cal Gas Exp Date: February 23, 2025
Cal Gas Cylinder #:	CC2216		
Removed Cal Gas Conc:	49.25	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	Teledyne API T700		Serial Number: 3807
Zero Air Gen Model:	Teledyne API T701		Serial Number: 196

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 100841398
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.988759	0.998178	Backgd or Offset:	14.7	16.7
Calibration intercept:	-0.098992	-1.624170	Coeff or Slope:	1.034	1.034

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	1.4	----
As found High point	4932	81.4	799.6	794.7	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	793.3	Previous response	790.6	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	4932	81.4	799.6	797.7	1.002
Mid point	4959	40.7	400.9	396.5	1.011
Low point	4981	20.4	200.9	198.5	1.012
As left zero	5000	0.0	0.0	-0.2	----
As left span	4932	81.4	799.6	800.4	0.999
Average Correction Factor:					1.009

Notes: Changed sample inlet filter after as founds. Adjusted zero only.

Calibration Performed By: Mohammed Kashif





# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 24, 2024	Previous Calibration:	April 24, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:16	End Time (MST):	13:32
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999979	<b>≥0.995</b>
799.6	797.7	1.0024	Slope	0.998178	<b>0.90 - 1.10</b>
400.9	396.5	1.0111	Intercept	-1.624170	<b>+/-30</b>
200.9	198.5	1.0120			



SO2 Calibration Plot

Date: May 24, 2024

Location: Lower Camp





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

### Station Information

Station Name: Lower Camp	Station number: AMS 11
Calibration Date: May 23, 2024	Last Cal Date: April 23, 2024
Start time (MST): 9:48	End time (MST): 13:54
Reason: Routine	

### Calibration Standards

Cal Gas Concentration: 5.43 ppm	Cal Gas Exp Date: January 4, 2025
Cal Gas Cylinder #: CC501097	
Removed Cal Gas Conc: 5.43 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 3807
ZAG Make/Model: API T701H	Serial Number: 196

### Analyzer Information

Analyzer make: Thermo 43iQ	Analyzer serial #: 1203169745
Converter make: Global G150	Converter serial #: 2022-223
Analyzer Range: 0 - 100 ppb	Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.990021	1.005616	Backgd or Offset:	2.5	2.5
Calibration intercept:	0.156017	-0.165109	Coeff or Slope:	0.817	0.817

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4926	73.6	79.9	80.3	0.994
As found Mid point	4963	36.8	40.0	39.5	1.009
As found Low point	4982	18.6	20.2	20.0	1.005
New cylinder response					
Baseline Corr As found:	80.4	Prev response:	79.28	*% change:	1.4%
Baseline Corr 2nd AF pt:	39.6	AF Slope:	1.006054	AF Intercept:	-0.305434
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999932	<i>* = &gt; +/-5% change initiates investigation</i>	

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4926	73.6	79.9	80.2	0.997
Mid point	4963	36.8	40.0	40.1	0.996
Low point	4982	18.6	20.2	20.0	1.010
As left zero	5000	0.0	0.0	0.0	----
As left span	4926	73.6	79.9	79.8	1.002
SO2 Scrubber Check	4935	81.5	812.3	0.2	----
Date of last scrubber change:				Ave Corr Factor	1.001
Date of last converter efficiency test:					

Notes: Changed sample inlet filter after as founds. No adjustments made. Ran scrubber check after calibrator zero and it passed.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association


## H<sub>2</sub>S Calibration Summary

### Station Information

Calibration Date:	May 23, 2024	Previous Calibration:	April 23, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	9:48	End Time (MST):	13:54
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1203169745

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999991	$\geq 0.995$
79.9	80.2	0.9965	Slope	1.005616	$0.90 - 1.10$
40.0	40.1	0.9965	Intercept	-0.165109	$\pm 3$
20.2	20.0	1.0097			



H<sub>2</sub>S Calibration Plot

Date: May 23, 2024

Location: Lower Camp





# Wood Buffalo Environmental Association

## THC / CH4 / NMHC Calibration Report

### Station Information

Station Name:	Lower Camp	Station number:	AMS 11
Calibration Date:	May 24, 2024	Last Cal Date:	April 24, 2024
Start time (MST):	10:16	End time (MST):	13:32
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	502.0 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	502.0 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
Zero Air Gen model:	API T701	Serial Number:	196

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1505164381
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.75E-04	2.75E-04	4.90E-05	4.90E-05
CH4 Retention time:	14.8	14.8	187273	187273
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	---
As found High point	4932	81.4	17.33	17.50	0.990
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.50	Prev response	17.28	*% change	1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	---
High point	4932	81.4	17.33	17.50	0.990
Mid point	4959	40.7	8.69	8.68	1.000
Low point	4981	20.4	4.35	4.32	1.007
As left zero	5000	0.0	0.00	0.00	---
As left span	4932	81.4	17.33	17.59	0.985
Average Correction Factor					0.999

Notes: Changed sample inlet filter after as founds. No adjustments made.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	81.4	9.18	9.30	0.987
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.30	Prev response	9.15	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	81.4	9.18	9.29	0.988
Mid point	4959	40.7	4.60	4.62	0.996
Low point	4981	20.4	2.31	2.30	1.003
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	81.4	9.18	9.34	0.982
Average Correction Factor					0.996

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	81.4	8.15	8.20	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.20	Prev response	8.13	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	81.4	8.15	8.21	0.993
Mid point	4959	40.7	4.09	4.07	1.005
Low point	4981	20.4	2.05	2.02	1.012
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	81.4	8.15	8.25	0.988
Average Correction Factor					1.003

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998938	1.010841
THC Cal Offset:	-0.023918	-0.047031
CH <sub>4</sub> Cal Slope:	0.999584	1.008203
CH <sub>4</sub> Cal Offset:	-0.013970	-0.025003
NMHC Cal Slope:	0.997916	1.013246
NMHC Cal Offset:	-0.009146	-0.022029

Calibration Performed By:

Mohammed Kashif



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 24, 2024	Previous Calibration:	April 24, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:16	End Time (MST):	13:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999960	<i>≥0.995</i>
17.33	17.50	0.9901	Slope	1.010841	<i>0.90 - 1.10</i>
8.69	8.68	1.0005	Intercept	-0.047031	<i>+/-0.5</i>
4.35	4.32	1.0066			







# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 24, 2024	Previous Calibration:	April 24, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:16	End Time (MST):	13:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999949	<i>≥0.995</i>
8.15	8.21	0.9928	Slope	1.008203	<i>0.90 - 1.10</i>
4.09	4.07	1.0046	Intercept	-0.025003	<i>+/-0.5</i>
2.05	2.02	1.0117			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 24, 2024	Previous Calibration:	April 24, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:16	End Time (MST):	13:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999970	<i>≥0.995</i>
9.18	9.29	0.9877	Slope	1.013246	<i>0.90 - 1.10</i>
4.60	4.62	0.9964	Intercept	-0.022029	<i>+/-0.5</i>
2.31	2.30	1.0026			



NMHC Calibration Plot

Date: May 24, 2024

Location: Lower Camp





# Wood Buffalo Environmental Association

## THC / CH4 / NMHC Calibration Report

### Station Information

Station Name:	Lower Camp	Station number:	AMS 11
Calibration Date:	May 31, 2024	Last Cal Date:	May 24, 2024
Start time (MST):	10:53	End time (MST):	12:43
Reason:	Cylinder Change		

### Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	502.0 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	502.0 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
Zero Air Gen model:	API T701	Serial Number:	196

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1505164381
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.75E-04	2.75E-04	4.90E-05	4.90E-05
CH4 Retention time:	14.8	14.8	187273	187273
Zero Chromatogram:	OFF	OFF	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	81.4	17.33	17.45	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.45	Prev response	17.47	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	81.4	17.33	17.46	0.992
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 0.992

Notes: Changed Hydrogen cylinder.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	81.4	9.18	9.25	0.992
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.25	Prev response	9.28	<b>*% change</b>	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<b>* = &gt; +/-5% change initiates investigation</b>	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	81.4	9.18	9.27	0.990
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 0.990

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	81.4	8.15	8.20	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.20	Prev response	8.19	<b>*% change</b>	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		<b>* = &gt; +/-5% change initiates investigation</b>	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	81.4	8.15	8.19	0.996
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 0.996

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.010841	1.007713
THC Cal Offset:	-0.047031	0.000000
CH <sub>4</sub> Cal Slope:	1.008203	1.004452
CH <sub>4</sub> Cal Offset:	-0.025003	0.000000
NMHC Cal Slope:	1.013246	1.010610
NMHC Cal Offset:	-0.022029	0.000000


Calibration Performed By:

Mohammed Kashif

NMHC Calibration Plot

Date: May 31, 2024

Location: Lower Camp





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS13 FORT MCKAY SOUTH MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	May 17, 2024	Last Cal Date:	April 4, 2024
Start time (MST):	10:57	End time (MST):	13:05
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.55	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC260812			
Removed Cal Gas Conc:	50.55	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	2448
Zero Air Gen Model:	API T701		Serial Number:	1117

### Analyzer Information

Analyzer make:	API T100	Serial Number:	599
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004442	1.008015	Backgd or Offset:	90.0	90.0
Calibration intercept:	-2.497906	-1.197668	Coeff or Slope:	0.711	0.711

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	1.1	----
As found High point	4921	79.1	799.7	805.7	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	804.6	Previous response	800.7	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	1.1	----
High point	4921	79.1	799.7	805.7	0.993
Mid point	4961	39.5	399.3	401.2	0.995
Low point	4980	19.8	200.2	197.6	1.013
As left zero	5000	0.0	0.0	1.3	----
As left span	4921	79.1	799.7	809.4	0.988
Average Correction Factor:					1.000

Notes: Remote calibrations. No adjustment made.

Calibration Performed By: Jan Castro





# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 17, 2024	Previous Calibration:	April 4, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:57	End Time (MST):	13:05
Analyzer make:	API T100	Analyzer serial #:	599

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	1.1	----	Correlation Coefficient	0.999958	<b>≥0.995</b>
799.7	805.7	0.9925	Slope	1.008015	<b>0.90 - 1.10</b>
399.3	401.2	0.9953	Intercept	-1.197668	<b>+/-30</b>
200.2	197.6	1.0131			



SO2 Calibration Plot

Date: May 17, 2024

Location: Fort McKay South





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	May 2, 2024	Last Cal Date:	April 10, 2024
Start time (MST):	8:59	End time (MST):	12:53
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.34	ppm	Cal Gas Exp Date:	January 1, 2025
Cal Gas Cylinder #:	CC500241			
Removed Cal Gas Conc:	5.34	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2448
ZAG Make/Model:	API T701		Serial Number:	1117

### Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017
Converter make:	CDN-101	Converter serial #:	521
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010145	1.000504	Backgd or Offset:	4.1	3.7
Calibration intercept:	-0.602184	-0.262173	Coeff or Slope:	1.143	1.131

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/Ic - AFzero))
As found zero	5000	0.0	0.0	-0.5	----
As found High point	4925	75.5	80.6	81.2	0.987
As found Mid point	4962	37.7	40.3	40.0	0.994
As found Low point	4981	18.9	20.2	19.3	1.019
New cylinder response					
Baseline Corr As found:	81.7	Prev response:	80.84	*% change:	1.1%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	1.015674	AF Intercept:	-0.822218
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999926	* = > +/-5% change initiates investigation	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic - Limit = 0.95-1.05)
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4925	75.5	80.6	80.5	1.002
Mid point	4962	37.7	40.3	40.0	1.007
Low point	4981	18.9	20.2	19.6	1.030
As left zero	5000	0.0	0.0	0.0	----
As left span	4925	75.5	80.6	79.4	1.015
SO2 Scrubber Check	4921	79.1	791.0	0.1	----
Date of last scrubber change:	20-Jan-20		Ave Corr Factor		1.013
Date of last converter efficiency test:					

Notes: Changed inlet filter after as founds. Scrubber checked after as left zero. Adjusted zero and span.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association


## TRS Calibration Summary

### Station Information

Calibration Date:	May 2, 2024	Previous Calibration:	April 10, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:59	End Time (MST):	12:53
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999947	$\geq 0.995$
80.6	80.5	1.0016	Slope	1.000504	0.90 - 1.10
40.3	40.0	1.0067	Intercept	-0.262173	+/-3
20.2	19.6	1.0299			



TRS Calibration Plot

Date: May 2, 2024

Location: Fort McKay South





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	May 17, 2024	Last Cal Date:	April 4, 2024
Start time (MST):	10:57	End time (MST):	13:05
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC2608112	Cal Gas Expiry Date:	
CH4 Cal Gas Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
C3H8 Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
Removed C3H8 Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	2448
Zero Air Gen model:	API T701	Serial Number:	1117

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1172750023
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.05E-04	3.05E-04	NMHC SP Ratio:	5.21E-05	5.21E-05
CH4 Retention time:	15.40	15.40	NMHC Peak Area:	174169	174169
Zero Chromatogram:	ON	ON	Flat Baseline:	ON	ON

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.01	----
As found High point	4921	79.1	17.05	17.05	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.04	Prev response	17.06	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.01	----
High point	4921	79.1	17.05	17.05	1.000
Mid point	4961	39.5	8.51	8.53	0.998
Low point	4980	19.8	4.27	4.27	1.000
As left zero	5000	0.0	0.00	0.04	----
As left span	4921	79.1	17.05	17.09	0.997
Average Correction Factor					0.999

Notes: Remote calibrations. No adjustment made.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	9.08	9.00	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.00	Prev response	9.10	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	9.08	9.00	1.008
Mid point	4961	39.5	4.53	4.52	1.003
Low point	4980	19.8	2.27	2.27	0.999
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	9.08	9.01	1.008
Average Correction Factor					1.004

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.01	----
As found High point	4921	79.1	7.97	8.06	0.990
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.05	Prev response	7.96	*% change	1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.01	----
High point	4921	79.1	7.97	8.06	0.989
Mid point	4961	39.5	3.98	4.01	0.992
Low point	4980	19.8	1.99	2.00	1.000
As left zero	5000	0.0	0.00	0.04	----
As left span	4921	79.1	7.97	8.09	0.985
Average Correction Factor					0.993

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002927	1.000019
THC Cal Offset:	-0.034543	0.008642
CH <sub>4</sub> Cal Slope:	1.002287	1.011045
CH <sub>4</sub> Cal Offset:	-0.026764	-0.005782
NMHC Cal Slope:	1.003450	0.991238
NMHC Cal Offset:	-0.007378	0.012622

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 17, 2024	Previous Calibration:	April 4, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:57	End Time (MST):	13:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999999	<span style="color: red;">≥0.995</span>
17.05	17.05	0.9997	Slope	1.000019	<span style="color: red;">0.90 - 1.10</span>
8.51	8.53	0.9976	Intercept	0.008642	<span style="color: red;">+/-0.5</span>
4.27	4.27	0.9996			







# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 17, 2024	Previous Calibration:	April 4, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:57	End Time (MST):	13:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999986	≥0.995
7.97	8.06	0.9890	Slope	1.011045	0.90 - 1.10
3.98	4.01	0.9915	Intercept	-0.005782	+/-0.5
1.99	2.00	0.9997			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 17, 2024	Previous Calibration:	April 4, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:57	End Time (MST):	13:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1172750023

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999990	<span style="color: red;">≥0.995</span>
9.08	9.00	1.0084	Slope	0.991238	<span style="color: red;">0.90 - 1.10</span>
4.53	4.52	1.0032	Intercept	0.012622	<span style="color: red;">+/-0.5</span>
2.27	2.27	0.9995			



NMHC Calibration Plot

Date: May 17, 2024

Location: Fort McKay South





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Fort McKay South  
 Station number: AMS 13  
 Calibration Date: May 7, 2024  
 Last Cal Date: April 16, 2024  
 Start time (MST): 9:40  
 End time (MST): 13:55  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T2UP1RP  
 NOX Cal Gas Conc: 48.25 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 48.25 ppm  
 NOX gas Diff:  
 Calibrator Model: API T700  
 ZAG make/model: APIT701  
 Cal Gas Expiry Date: November 17, 2026  
 NO Cal Gas Conc: 47.88 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 47.88 ppm  
 NO gas Diff:  
 Serial Number: 2448  
 Serial Number: 1117

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
AF High point	4917	83.5	805.7	799.5	6.2	805.8	798.3	7.5	0.9999	1.0015
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 804.7 ppb	NO = 799.7 ppb	<i>* = &gt; +/-5% change initiates investigation</i>				*Percent Change	NO <sub>x</sub> = 0.1%		
Baseline Corr 1st pt	NO <sub>x</sub> = 805.8 ppb	NO = 798.3 ppb	<u>As Found Statistics</u>				*Percent Change	NO = -0.2%		
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :		Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :		NO SI:	NO Int:			
			As found	NO <sub>2</sub> r <sup>2</sup> :		NO <sub>2</sub> SI:	NO <sub>2</sub> Int:			

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1410661329

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000782	0.998924
NO <sub>x</sub> Cal Offset:	-1.593202	-1.312891
NO Cal Slope:	1.003482	1.000524
NO Cal Offset:	-2.631535	-2.371115
NO <sub>2</sub> Cal Slope:	1.001146	0.996093
NO <sub>2</sub> Cal Offset:	-0.447686	-0.942456

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.140	1.140	NO bkgnd or offset:	10.2	10.2
NOX coeff or slope:	1.002	1.002	NOX bkgnd or offset:	10.3	10.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	163.5	157.3

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.0	----	----
High point	4917	83.5	805.7	799.5	6.2	804.3	798.8	5.4	1.0017	1.0009
Mid point	4958	41.8	403.4	400.3	3.1	400.9	397.1	3.8	1.0062	1.0080
Low point	4979	20.9	201.7	200.1	1.5	198.5	195.0	3.5	1.0161	1.0264
As left zero	5000	0.0	0.0	0.0	0.0	0.5	0.4	0.1	----	----
As left span	4917	83.5	805.7	379.9	425.8	808.9	379.9	429.0	0.9960	1.0000
Average Correction Factor									1.0080	1.0118

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	797.9	376.9	427.2	425.0	1.0051	99.5%
Mid GPT point	797.9	585.2	218.9	216.8	1.0096	99.1%
Low GPT point	797.9	691.2	112.9	110.4	1.0224	97.8%
Average Correction Factor					1.0124	98.8%

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 16, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:40	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999982	≥0.995
805.7	804.3	1.0017	Slope	0.998924	0.90 - 1.10
403.4	400.9	1.0062	Intercept	-1.312891	+/-20
201.7	198.5	1.0161			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 16, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:40	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999976	≥0.995
427.2	425.0	1.0051	Slope	0.996093	0.90 - 1.10
218.9	216.8	1.0096	Intercept	-0.942456	+/-20
112.9	110.4	1.0224			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 16, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:40	End Time (MST):	13:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999949	≥0.995
799.5	798.8	1.0009	Slope	1.000524	0.90 - 1.10
400.3	397.1	1.0080	Intercept	-2.371115	+/-20
200.1	195.0	1.0264			






NO<sub>x</sub> Calibration Plot

Date: May 7, 2024

Location: Fort McKay South





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

### Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	May 6, 2024	Last Cal Date:	April 15, 2024
Start time (MST):	10:22	End time (MST):	13:29
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer	Serial Number:	2448
Calibrator Make/Model:	Teledyne API T700	Serial Number:	1117
ZAG Make/Model:	Teledyne API T701		

### Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #:	3871
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001029	0.999914	Backgd or Offset:	2.7	2.7
Calibration intercept:	1.320000	1.540000	Coeff or Slope:	0.973	0.973

### O<sub>3</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	5000	994.5	400.0	401.9	0.996
As found Mid point					
As found Low point					
Baseline Corr As found:	401.8	Previous response	401.7	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.6	----
High point	5000	995.4	400.0	401.0	0.998
Mid point	5000	849.9	200.0	202.1	0.990
Low point	5000	746.6	100.0	102.4	0.977
As left zero	5000	0.0	0.0	0.6	----
As left span	5000	995.4	400.0	402.1	0.995
Average Correction Factor					<b>0.988</b>

Notes: Changed inlet filter after as founds. No adjustment.

Calibration Performed By: Sean Bala



# Wood Buffalo Environmental Association


## O<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 6, 2024	Previous Calibration:	April 15, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:22	End Time (MST):	13:29
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999975	<span style="color: red;">≥0.995</span>
400.0	401.0	0.9975	Slope	0.999914	<span style="color: red;">0.90 - 1.10</span>
200.0	202.1	0.9896	Intercept	1.540000	<span style="color: red;">+/- 5</span>
100.0	102.4	0.9766			



O<sub>3</sub> Calibration Plot

Date: May 6, 2024

Location: Fort McKay South





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Fort McKay South Station number: AMS 13  
 Calibration Date: May 22, 2024 Last Cal Date: April 19, 2024  
 Start time (MST): 9:05 End time (MST): 9:22

Analyzer Make: API T640 S/N: 1335  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388746  
 Temp/RH standard: Alicat FP-25BT S/N: 388746

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	9.30	8.91	9.30	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	735.50	737.70	735.50	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.93	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	45	----	45	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024  
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 19, 2024  
 Date Disposable Filter Changed: April 19, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: June 29, 2023  
 Date RH/T Sensor Cleaned: June 29, 2023

Notes: No adjustments made. Leak check passed. Filter is still clean.

Calibration by: Sean Bala



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS14  
ANZAC  
MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	May 3, 2024	Last Cal Date:	April 16, 2024
Start time (MST):	9:04	End time (MST):	14:04
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.95	ppm	Cal Gas Exp Date: January 5, 2025
Cal Gas Cylinder #:	CC279389		
Removed Cal Gas Conc:	49.95	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	API T700		Serial Number: 3060
Zero Air Gen Model:	API T701H		Serial Number: 357

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	0710321322
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009311	1.019487	Backgd or Offset:	24.6	24.8
Calibration intercept:	-4.184570	-2.972409	Coeff or Slope:	1.043	1.043

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4938	80.3	799.3	809.3	0.987
As found Mid point	4979	40.2	400.1	400.2	0.999
As found Low point	4998	20.2	201.1	198.2	1.014
New cylinder response					
Baseline Corr As found:	809.4	Previous response	802.5	*% change	0.8%
Baseline Corr 2nd AF pt:	400.3	AF Slope:	1.014249	AF Intercept:	-3.188607
Baseline Corr 3rd AF pt:	198.3	AF Correlation:	0.999930	<i>* = &gt; +/-5% change initiates investigation</i>	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	4938	80.3	799.3	813.6	0.982
Mid point	4979	40.2	400.1	402.4	0.994
Low point	4998	20.2	201.1	200.1	1.005
As left zero	5000	0.0	0.0	0.2	----
As left span	4938	80.3	799.3	813.8	0.982
Average Correction Factor:					0.994

Notes: Sample inlet filter changed after as founds. No adjustments needed.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	April 16, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:04	End Time (MST):	14:04
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.3	----	Correlation Coefficient	0.999945	<b>≥0.995</b>
799.3	813.6	0.9824	Slope	1.019487	<b>0.90 - 1.10</b>
400.1	402.4	0.9942	Intercept	-2.972409	<b>+/-30</b>
201.1	200.1	1.0048			






SO2 Calibration Plot

Date: May 3, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name: Anzac	Station number: AMS 14
Calibration Date: May 9, 2024	Last Cal Date: April 11, 2024
Start time (MST): 9:22	End time (MST): 14:02
Reason: Routine	

### Calibration Standards

Cal Gas Concentration: 5.15 ppm	Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC510379	
Removed Cal Gas Conc: 5.15 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 3060
ZAG Make/Model: API 701H	Serial Number: 357

### Analyzer Information

Analyzer make: Thermo 43i-TLE	Analyzer serial #: 1218153582
Converter make: CD Nova CDN-101	Converter serial #: 503
Analyzer Range: 0 - 100 ppb	Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>	
Calibration slope:	1.015603	1.003735	Backgd or Offset: 2.3
Calibration intercept:	-0.225550	-0.185416	Coeff or Slope: 0.984

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4938	77.9	80.0	80.5	0.992
As found Mid point	4973	38.9	40.0	40.4	0.987
As found Low point	4997	19.5	20.0	20.0	0.996
New cylinder response					
Baseline Corr As found:	80.6	Prev response:	81.01	*% change:	-0.5%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	1.008346	AF Intercept:	-0.085404
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999987	<i>* = &gt; +/-5% change initiates investigation</i>	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4938	77.9	80.0	80.1	0.998
Mid point	4973	38.9	40.0	39.9	1.001
Low point	4997	19.5	20.0	19.8	1.011
As left zero	5000	0.0	0.0	0.1	----
As left span	4938	77.9	80.0	78.4	1.020
SO2 Scrubber Check	4936	80.3	800.4	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.003
Date of last converter efficiency test:					

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association


## TRS Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	April 11, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:22	End Time (MST):	14:02
Analyzer make:	CD Nova CDN-101	Analyzer serial #:	503

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999995	$\geq 0.995$
80.0	80.1	0.9981	Slope	1.003735	$0.90 - 1.10$
40.0	39.9	1.0014	Intercept	-0.185416	$\pm 3$
20.0	19.8	1.0107			



TRS Calibration Plot

Date: May 9, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	May 3, 2024	Last Cal Date:	April 2, 2024
Start time (MST):	9:04	End time (MST):	10:47
Reason:	Removal		

### Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
Zero Air Gen model:	API 701H	Serial Number:	357

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1118148494
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.25E-04	NA	NMHC SP Ratio:	4.11E-05	NA
CH4 Retention time:	13.30	NA	NMHC Peak Area:	221451	NA
Zero Chromatogram:	OFF		Flat Baseline:	OFF	

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4938	80.3	17.10	16.24	1.053
As found Mid point	4979	40.2	8.56	8.05	1.063
As found Low point	4998	20.2	4.30	3.99	1.078
New cylinder response					
Baseline Corr AF:	16.24	Prev response	16.71	*% change	-2.9%
Baseline Corr 2nd AF:	8.05	AF Slope:	0.951160	AF Intercept:	-0.053021
Baseline Corr 3rd AF:	3.99	AF Correlation:	0.999950	* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Notes:

Removal calibration. Conducted investigation and found no alarms; diagnostics closely resembled previous calibration; Chromatograms appeared normal. Observations revealed a drop in air pressure (air and actuator lines (T- Connection) on the back of analyzer from ZAG) post backflush mode, decreasing to approximately 18 psi from the normal 24 psi. This decline is resulting in reduced flame count and flame temperature, though neither reaches zero. Suspecting an internal issue, to be resolved at the repair shop.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	---- <i>Limit = 0.90-1.10</i>
As found High point	4938	80.3	9.11	8.50	1.072
As found Mid point	4979	40.2	4.56	4.23	1.079
As found Low point	4998	20.2	2.29	2.11	1.085
New cylinder response					
Baseline Corr AF:	8.50	Prev response	8.87	*% change	-4.4%
Baseline Corr 2nd AF:	4.23	AF Slope:	0.933538	AF Intercept:	-0.016001
Baseline Corr 3rd AF:	2.11	AF Correlation:	0.999982	* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					<i>Limit = 0.95-1.05</i>
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	---- <i>Limit = 0.90-1.10</i>
As found High point	4938	80.3	7.99	7.74	1.032
As found Mid point	4979	40.2	4.00	3.83	1.045
As found Low point	4998	20.2	2.01	1.88	1.070
New cylinder response					
Baseline Corr AF:	7.74	Prev response	7.83	*% change	-1.2%
Baseline Corr 2nd AF:	3.83	AF Slope:	0.971146	AF Intercept:	-0.037620
Baseline Corr 3rd AF:	1.88	AF Correlation:	0.999892	* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					<i>Limit = 0.95-1.05</i>
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

### Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.980408	NA
THC Cal Offset:	-0.060379	NA
CH <sub>4</sub> Cal Slope:	0.984987	NA
CH <sub>4</sub> Cal Offset:	-0.035307	NA
NMHC Cal Slope:	0.976268	NA
NMHC Cal Offset:	-0.025072	NA

Calibration Performed By: Mohammed Kashif

NMHC Calibration Plot

Date: May 3, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	May 3, 2024	Last Cal Date:	NA
Start time (MST):	11:30	End time (MST):	14:04
Reason:	Install		

### Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
Zero Air Gen model:	API 701H	Serial Number:	357

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1193585649
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	NA	2.63E-04	NMHC SP Ratio:	NA
CH4 Retention time:	NA	14.90	NMHC Peak Area:	NA
Zero Chromatogram:		OFF	Flat Baseline:	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4938	80.3	17.10	17.04	1.004
Mid point	4979	40.2	8.56	8.36	1.024
Low point	4998	20.2	4.30	4.13	1.041
As left zero	5000	0.0	0.00	0.00	----
As left span	4938	80.3	17.10	16.94	1.010
Average Correction Factor					1.023

Notes: Replaced the sample inlet filter and Nitrogen cylinder, then conducted the installation. Adjusted span only.





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4938	80.3	9.11	9.08	1.003
Mid point	4979	40.2	4.56	4.47	1.020
Low point	4998	20.2	2.29	2.21	1.035
As left zero	5000	0.0	0.00	0.00	----
As left span	4938	80.3	9.11	9.02	1.010
Average Correction Factor					1.020

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4938	80.3	7.99	7.96	1.004
Mid point	4979	40.2	4.00	3.89	1.028
Low point	4998	20.2	2.01	1.92	1.048
As left zero	5000	0.0	0.00	0.00	----
As left span	4938	80.3	7.99	7.91	1.010
Average Correction Factor					1.026

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	NA	0.998493
THC Cal Offset:	NA	-0.096363
CH <sub>4</sub> Cal Slope:	NA	0.998448
CH <sub>4</sub> Cal Offset:	NA	-0.052166
NMHC Cal Slope:	NA	0.998382
NMHC Cal Offset:	NA	-0.043597

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	NA
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:30	End Time (MST):	14:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999837	≥0.995
17.10	17.04	1.0035	Slope	0.998493	0.90 - 1.10
8.56	8.36	1.0239	Intercept	-0.096363	+/-0.5
4.30	4.13	1.0415			





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	NA
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:30	End Time (MST):	14:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999779	≥0.995
7.99	7.96	1.0038	Slope	0.998448	0.90 - 1.10
4.00	3.89	1.0278	Intercept	-0.052166	±0.5
2.01	1.92	1.0479			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 3, 2024	Previous Calibration:	NA
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:30	End Time (MST):	14:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999883	<i>≥0.995</i>
9.11	9.08	1.0033	Slope	0.998382	<i>0.90 - 1.10</i>
4.56	4.47	1.0205	Intercept	-0.043597	<i>+/-0.5</i>
2.29	2.21	1.0355			



NMHC Calibration Plot

Date: May 3, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Anzac  
 Station number: AMS 14  
 Calibration Date: May 7, 2024  
 Last Cal Date: April 3, 2024  
 Start time (MST): 9:20  
 End time (MST): 14:12  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: DT0037092  
 NOX Cal Gas Conc: 60.7 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 60.70 ppm  
 NOX gas Diff:  
 Calibrator Model: Teledyne API T700  
 ZAG make/model: Teledyne API T700H  
 Cal Gas Expiry Date: May 16, 2031  
 NO Cal Gas Conc: 60.40 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 60.40 ppm  
 NO gas Diff:  
 Serial Number: 3060  
 Serial Number: 357

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
AF High point	4934	66.3	804.8	800.9	4.0	799.5	796.3	3.1	1.0067	1.0056
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO <sub>x</sub> = 798.1 ppb	NO = 797.9 ppb	<i>* = &gt; +/-5% change initiates investigation</i>				*Percent Change	NO <sub>x</sub> = 0.2%
Baseline Corr 1st pt	NO <sub>x</sub> = 799.5 ppb	NO = 796.4 ppb	<u>As Found Statistics</u>				*Percent Change	NO = -0.2%
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO <sub>x</sub> r <sup>2</sup> :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb	As found	NO r <sup>2</sup> :		NO SI:	NO Int:	
			As found	NO <sub>2</sub> r <sup>2</sup> :		NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430008

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.992088	0.992016
NO <sub>x</sub> Cal Offset:	-0.369777	-0.569604
NO Cal Slope:	0.998384	0.997285
NO Cal Offset:	-1.649243	-2.189386
NO <sub>2</sub> Cal Slope:	0.993217	0.995331
NO <sub>2</sub> Cal Offset:	-1.136094	-1.206129

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.411	1.411	NO bkgnd or offset:	3.8	3.8
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.5	3.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	158.8	158.8

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.1	----	----
High point	4934	66.3	804.8	800.9	4.0	798.0	797.4	0.5	1.0086	1.0043
Mid point	4985	33.2	401.6	399.6	2.0	398.0	395.7	2.3	1.0090	1.0099
Low point	5004	16.7	201.9	200.9	1.0	198.7	195.8	2.8	1.0161	1.0261
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
As left span	4934	66.3	804.8	406.8	398.0	799.9	406.8	393.3	1.0062	1.0000
Average Correction Factor									1.0112	1.0134

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	796.1	402.3	397.8	395.4	1.0060	99.4%
Mid GPT point	796.1	595.7	204.4	201.6	1.0138	98.6%
Low GPT point	796.1	693.5	106.6	103.5	1.0297	97.1%
Average Correction Factor					1.0165	98.4%

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 3, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:20	End Time (MST):	14:12
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999996	≥0.995
804.8	798.0	1.0086	Slope	0.992016	0.90 - 1.10
401.6	398.0	1.0090	Intercept	-0.569604	+/-20
201.9	198.7	1.0161			







# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 3, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:20	End Time (MST):	14:12
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999948	≥0.995
397.8	395.4	1.0060	Slope	0.995331	0.90 - 1.10
204.4	201.6	1.0138	Intercept	-1.206129	+/-20
106.6	103.5	1.0297			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 3, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:20	End Time (MST):	14:12
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999968	<span style="color: red;">≥0.995</span>
800.9	797.4	1.0043	Slope	0.997285	<span style="color: red;">0.90 - 1.10</span>
399.6	395.7	1.0099	Intercept	-2.189386	<span style="color: red;">+/-20</span>
200.9	195.8	1.0261			



NO<sub>x</sub> Calibration Plot

Date: May 7, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

### Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	May 6, 2024	Last Cal Date:	April 9, 2024
Start time (MST):	10:24	End time (MST):	13:32
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3060
Calibrator Make/Model:	API T700	Serial Number:	357
ZAG Make/Model:	API 701H		

### Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997086	0.996257	Backgd or Offset:	1.4
Calibration intercept:	2.260000	3.380000	Coeff or Slope:	1.594

### O<sub>3</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <span style="color: red;">Limit = 0.90-1.10</span>
As found zero	5000	0.0	0.0	0.9	----
As found High point	5000	918.8	400.0	401.6	0.998
As found Mid point					
As found Low point					
Baseline Corr As found:	400.7	Previous response	401.1	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<span style="color: red;">* = &gt; +/-5% change initiates investigation</span>	

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <span style="color: red;">Limit = 0.95-1.05</span>
Calibrator zero	5000	0.0	0.0	0.7	----
High point	5000	918.8	400.0	400.3	0.999
Mid point	5000	803.8	200.0	204.7	0.977
Low point	5000	709.8	100.0	105.2	0.951
As left zero	5000	0.0	0.0	1.0	----
As left span	5000	918.8	400.0	402.3	0.994
Average Correction Factor					0.976

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association


## O<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 6, 2024	Previous Calibration:	April 9, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:24	End Time (MST):	13:32
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.7	----	Correlation Coefficient	0.999784	<span style="color: red;">≥0.995</span>
400.0	400.3	0.9993	Slope	0.996257	<span style="color: red;">0.90 - 1.10</span>
200.0	204.7	0.9770	Intercept	3.380000	<span style="color: red;">+/- 5</span>
100.0	105.2	0.9506			



O<sub>3</sub> Calibration Plot

Date: May 6, 2024

Location: Anzac





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Anzac Station number: AMS 14  
 Calibration Date: May 10, 2024 Last Cal Date: April 30, 2024  
 Start time (MST): 10:31 End time (MST): 11:45

Analyzer Make: AP T640 S/N: 825  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749  
 Temp/RH standard: Alicat FP-25BT S/N: 388749

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	21.9	21.57	21.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	711.6	712.48	711.6	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	4.98	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	36	----	36	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.5	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: December 15, 2024  
 Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.8	11.0	10.9	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: May 10, 2024  
 Date Disposable Filter Changed: May 10, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: July 6, 2023  
 Date RH/T Sensor Cleaned: July 6, 2023

Notes: Performed quarterly calibration test. Leak check passed. Head cleaned.

Calibration by: Mohammed Kashif



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS17  
WAPASU  
MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	May 16, 2024	Last Cal Date:	April 2, 2024
Start time (MST):	10:50	End time (MST):	14:02
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.38	ppm	Cal Gas Exp Date: January 12, 2029
Cal Gas Cylinder #:	ALM066507		
Removed Cal Gas Conc:	50.38	ppm	Rem Gas Exp Date: N/A
Removed Gas Cyl #:	N/A		Diff between cyl:
Calibrator Model:	API T700		Serial Number: 2449
Zero Air Gen Model:	API 701H		Serial Number: 359

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1218153459
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998796	0.997012	Backgd or Offset:	13.3	13.3
Calibration intercept:	-1.859943	-2.260549	Coeff or Slope:	1.098	1.098

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4921	79.4	800.0	793.0	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	793.1	Previous response	797.1	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	79.4	800.0	796.7	1.004
Mid point	4960	39.7	400.0	394.7	1.014
Low point	4980	19.8	199.5	194.9	1.024
As left zero	5000	0.0	0.0	-0.1	----
As left span	4920	79.4	800.1	796.0	1.005
Average Correction Factor:					1.014

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 16, 2024	Previous Calibration:	April 2, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:50	End Time (MST):	14:02
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999961	<b>≥0.995</b>
800.0	796.7	1.0041	Slope	0.997012	<b>0.90 - 1.10</b>
400.0	394.7	1.0135	Intercept	-2.260549	<b>+/-30</b>
199.5	194.9	1.0237			



SO2 Calibration Plot

Date: May 16, 2024

Location: Wapasu





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

### Station Information

Station Name:	Wapasu	Station number:	AMS 17
Calibration Date:	May 22, 2024	Last Cal Date:	April 8, 2024
Start time (MST):	9:49	End time (MST):	15:00
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.08	ppm	Cal Gas Exp Date:	September 16, 2024
Cal Gas Cylinder #:	CC511852			
Removed Cal Gas Conc:	5.08	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2449
ZAG Make/Model:	API T701H		Serial Number:	359

### Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583
Converter make:	CD Nova	Converter serial #:	N/A
Analyzer Range	0 - 100 ppb	Converter Temp:	degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992711	1.010854	Backgd or Offset:	11.8	11.8
Calibration intercept:	-0.119219	-0.579198	Coeff or Slope:	1.096	1.096

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4921	78.8	80.0	79.2	1.013
As found Mid point	4961	39.4	40.0	39.6	1.015
As found Low point	4980	19.7	20.0	19.8	1.020
New cylinder response					
Baseline Corr As found:	79.0	Prev response:	79.30	*% change:	-0.4%
Baseline Corr 2nd AF pt:	39.4	AF Slope:	0.987997	AF Intercept:	0.120789
Baseline Corr 3rd AF pt:	19.6	AF Correlation:	0.999995	<i>* = +/-5% change initiates investigation</i>	

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.4	----
High point	4921	78.8	80.0	80.4	0.995
Mid point	4961	39.4	40.0	39.7	1.007
Low point	4980	19.7	20.0	19.5	1.026
As left zero	5000	0.0	0.0	-0.4	----
As left span	4921	78.8	80.0	79.7	1.004
SO <sub>2</sub> Scrubber Check	4921	79.4	793.9	-0.1	----
Date of last scrubber change:		N/A		Ave Corr Factor	1.009
Date of last converter efficiency test:		N/A			

Notes: Pump changed out after as founds. Scrubber check passed. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## H<sub>2</sub>S/TRS Calibration Summary

### Station Information

Calibration Date:	May 22, 2024	Previous Calibration:	April 8, 2024
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	9:49	End Time (MST):	15:00
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	0.999976	≥0.995
80.0	80.4	0.9950	Slope	1.010854	0.90 - 1.10
40.0	39.7	1.0074	Intercept	-0.579198	+/-3
20.0	19.5	1.0257			



H<sub>2</sub>S Calibration Plot

Date: May 22, 2024

Location: Wapasu







# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 16, 2024	Previous Calibration:	April 2, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:50	End Time (MST):	14:02
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

### Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999989	<span style="color: red;">≥0.995</span>
17.09	17.10	0.9993	Slope	0.999180	<span style="color: red;">0.90 - 1.10</span>
8.55	8.53	1.0022	Intercept	0.018473	<span style="color: red;">+/-1.5</span>
4.26	4.31	0.9899			






THC Calibration Plot

Date: May 16, 2024

Location: Wapasu





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Wapasu  
 Station number: AMS 17  
 Calibration Date: May 17, 2024  
 Last Cal Date: April 18, 2024  
 Start time (MST): 7:04  
 End time (MST): 12:28  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T375YK8  
 NOX Cal Gas Conc: 49.11 ppm  
 Removed Cylinder #: T375YK8  
 Removed Gas NOX Conc: 49.11 ppm  
 NOX gas Diff:  
 Calibrator Model: API T700  
 ZAG make/model: API T701H  
 Cal Gas Expiry Date: April 13, 2025  
 NO Cal Gas Conc: 48.07 ppm  
 Removed Gas Exp Date: N/A  
 Removed Gas NO Conc: 48.07 ppm  
 NO gas Diff:  
 Serial Number: 2449  
 Serial Number: 359

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.3	-0.3	----	----
AF High point	4917	83.2	817.2	799.9	17.3	820.5	803.4	17.0	0.9952	0.9953
AF Mid point										
AF Low point										
New cyl resp										

Previous Response NO<sub>x</sub> = 816.3 ppb NO = 798.9 ppb \* = > +/-5% change initiates investigation \*Percent Change NO<sub>x</sub> = 0.6%  
 Baseline Corr 1st pt NO<sub>x</sub> = 821.1 ppb NO = 803.7 ppb As Found Statistics \*Percent Change NO = 0.6%  
 Baseline Corr 2nd pt NO<sub>x</sub> = NA ppb NO = NA ppb As found NO<sub>x</sub> r<sup>2</sup>: Nx SI: Nx Int:  
 Baseline Corr 3rd pt NO<sub>x</sub> = NA ppb NO = NA ppb As found NO r<sup>2</sup>: NO SI: NO Int:  
 As found NO<sub>2</sub> r<sup>2</sup>: NO2 SI: NO2 Int:

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo Scientific 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153460

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.999550	1.004026
NO <sub>x</sub> Cal Offset:	-0.560000	-1.060000
NO Cal Slope:	1.000573	1.005288
NO Cal Offset:	-1.400000	-1.300000
NO <sub>2</sub> Cal Slope:	0.995649	0.997935
NO <sub>2</sub> Cal Offset:	-0.696163	-0.116856

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.100	1.100	NO bkgnd or offset:	3.8	3.8
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	4.2	4.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	261.0	261.0

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
High point	4917	83.2	817.2	799.9	17.3	819.8	803.4	16.5	0.9968	0.9956
Mid point	4958	41.6	408.6	399.9	8.7	408.7	400.2	8.5	0.9997	0.9994
Low point	4979	20.8	204.3	200.0	4.3	203.5	198.5	4.9	1.0039	1.0074
As left zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
As left span	4917	83.2	817.2	397.7	419.5	819.1	397.7	421.6	0.9977	1.0000
Average Correction Factor									1.0002	1.0008

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.3	----	----
High GPT point	802.3	399.2	420.4	419.3	1.0026	99.7%
Mid GPT point	802.3	603.5	216.1	215.7	1.0019	99.8%
Low GPT point	802.3	705.9	113.7	113.5	1.0018	99.8%
Average Correction Factor					1.0021	99.8%

Notes:

Remote calibration. No adjustments needed.

Calibration Performed By:

Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 17, 2024	Previous Calibration:	April 18, 2024
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	7:04	End Time (MST):	12:28
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	0.999997	≥0.995
817.2	819.8	0.9968	Slope	1.004026	0.90 - 1.10
408.6	408.7	0.9997	Intercept	-1.060000	+/-20
204.3	203.5	1.0039			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 17, 2024	Previous Calibration:	April 18, 2024
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	7:04	End Time (MST):	12:28
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999999	≥0.995
420.4	419.3	1.0026	Slope	0.997935	0.90 - 1.10
216.1	215.7	1.0019	Intercept	-0.116856	+/-20
113.7	113.5	1.0018			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 17, 2024	Previous Calibration:	April 18, 2024
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	7:04	End Time (MST):	12:28
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999990	≥0.995
799.9	803.4	0.9956	Slope	1.005288	0.90 - 1.10
399.9	400.2	0.9994	Intercept	-1.300000	+/-20
200.0	198.5	1.0074			



NO<sub>x</sub> Calibration Plot

Date: May 17, 2024

Location: Wapasu





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

### Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	May 10, 2024	Last Cal Date:	April 4, 2024
Start time (MST):	9:55	End time (MST):	13:14
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer	Serial Number:	2449
Calibrator Make/Model:	API T700	Serial Number:	359
ZAG Make/Model:	API T701H		

### Analyzer Information

Analyzer make:	API T400	Analyzer serial #:	3870
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000714	1.012486	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.600000	-0.860000	Coeff or Slope:	1.013	1.013

### O<sub>3</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <span style="color: red;">Limit = 0.90-1.10</span>
As found zero	5000	0.0	0.0	0.2	----
As found High point	5000	1077.3	400.0	403.0	0.993
As found Mid point					
As found Low point					
Baseline Corr As found:	402.8	Previous response	399.7	*% change	0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<span style="color: red;">* = &gt; +/-5% change initiates investigation</span>	

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <span style="color: red;">Limit = 0.95-1.05</span>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	5000	1077.3	400.0	404.8	0.988
Mid point	5000	900.3	200.0	200.6	0.997
Low point	5000	789.5	100.0	99.8	1.002
As left zero	5000	0.0	0.0	0.8	----
As left span	5000	1077.3	400.0	404.0	0.990
Average Correction Factor					0.996

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar





# Wood Buffalo Environmental Association


## O<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 10, 2024	Previous Calibration:	April 4, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	9:55	End Time (MST):	13:14
Analyzer make:	API T400	Analyzer serial #:	3870

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999969	<b>≥0.995</b>
400.0	404.8	0.9881	Slope	1.012486	<b>0.90 - 1.10</b>
200.0	200.6	0.9970	Intercept	-0.860000	<b>+/- 5</b>
100.0	99.8	1.0020			



O<sub>3</sub> Calibration Plot

Date: May 10, 2024

Location: Wapasu





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Wapasu Station number: AMS 17  
Calibration Date: May 22, 2024 Last Cal Date: April 17, 2024  
Start time (MST): 14:16 End time (MST): 14:55

Analyzer Make: Teledyne API T640 S/N: 1183  
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753  
Temp/RH standard: Alicat FP-25BT S/N: 388753

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.5	10.0	10.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	715	713.2	715	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	5.05	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	41	----	41	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.2	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 6, 2024  
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 13, 2024  
Date Disposable Filter Changed: February 13, 2024

Post- maintenance Zero Verification: PM w/ HEPA: \_\_\_\_\_ <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_  
Date RH/T Sensor Cleaned: \_\_\_\_\_

Notes: Temp. pressure and flow checked. Leak check passed.

Calibration by: Aswin Sasi Kumar



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS18 STONY MOUNTAIN MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	May 8, 2024	Last Cal Date:	April 3, 2024
Start time (MST):	10:15	End time (MST):	15:42
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.40	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC463851			
Removed Cal Gas Conc:	49.40	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	2658
Zero Air Gen Model:	Teledyne API 701H		Serial Number:	360

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	JC1501301453
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002389	1.005230	Backgd or Offset:	22.9	23.4
Calibration intercept:	0.496756	0.277512	Coeff or Slope:	0.800	0.800

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.6	----
As found High point	4919	81.0	800.3	807.1	0.992
As found Mid point	4959	40.5	400.2	403.3	0.994
As found Low point	4979	20.2	199.6	199.9	1.002
New cylinder response					
Baseline Corr As found:	806.5	Previous response	802.7	*% change	0.5%
Baseline Corr 2nd AF pt:	402.7	AF Slope:	1.008486	AF Intercept:	-0.262204
Baseline Corr 3rd AF pt:	199.3	AF Correlation:	0.999994	* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.8	----
High point	4919	81.0	800.3	805.0	0.994
Mid point	4959	40.5	400.2	402.3	0.995
Low point	4979	20.2	199.6	200.4	0.996
As left zero	5000	0.0	0.0	1.1	----
As left span	4919	81.0	800.3	802.6	0.997
Average Correction Factor:					0.995

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 3, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:15	End Time (MST):	15:42
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.8	----	Correlation Coefficient	0.999998	<b>≥0.995</b>
800.3	805.0	0.9941	Slope	1.005230	<b>0.90 - 1.10</b>
400.2	402.3	0.9947	Intercept	0.277512	<b>+/-30</b>
199.6	200.4	0.9960			



SO2 Calibration Plot

Date: May 8, 2024

Location: Stony Mountain





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name:	Stony Mountain	Station number:	AMS18
Calibration Date:	May 28, 2024	Last Cal Date:	April 24, 2024
Start time (MST):	10:15	End time (MST):	15:05
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.48	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC500395			
Removed Cal Gas Conc:	5.48	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701		Serial Number:	360

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359
Converter make:	CD Nova CDN-101	Converter serial #:	555
Analyzer Range	0 - 100 ppb	Converter Temp:	799 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993725	0.996870	Backgd or Offset:	2.6	2.6
Calibration intercept:	0.261187	0.301076	Coeff or Slope:	1.137	1.167

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4927	73.0	80.0	79.6	1.007
As found Mid point	4964	36.5	40.0	39.9	1.007
As found Low point	4983	18.3	20.0	20.1	1.007
New cylinder response					
Baseline Corr As found:	79.4	Prev response:	79.75	*% change:	-0.4%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.992584	AF Intercept:	0.201110
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	1.000000	<i>* = +/-5% change initiates investigation</i>	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4927	73.0	80.0	80.0	1.000
Mid point	4964	36.5	40.0	40.3	0.992
Low point	4983	18.3	20.0	20.2	0.992
As left zero	5000	0.0	0.0	0.5	----
As left span	4927	73.0	80.0	78.2	1.023
SO2 Scrubber Check	4923	77.1	771.0	0.0	----
Date of last scrubber change:	17-Dec-21			Ave Corr Factor	0.995
Date of last converter efficiency test:					

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar





# Wood Buffalo Environmental Association


## H<sub>2</sub>S/TRS Calibration Summary

### Station Information

Calibration Date:	May 28, 2024	Previous Calibration:	April 24, 2024
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:15	End Time (MST):	15:05
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999992	≥0.995
80.0	80.0	0.9999	Slope	0.996870	0.90 - 1.10
40.0	40.3	0.9924	Intercept	0.301076	+/-3
20.0	20.2	0.9925			



H<sub>2</sub>S/TRS Calibration Plot

Date: May 28, 2024

Location: Stony Mountain





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	May 9, 2024	Last Cal Date:	May 8, 2024
Start time (MST):	10:30	End time (MST):	14:30
Reason:	Maintenance		

### Calibration Standards

Gas Cert Reference:	CC463851	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.8 ppm	CH4 Equiv Conc.	1066.8 ppm
C3H8 Cal Gas Conc.	205.8 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	500.8 ppm	CH4 Equiv Conc.	1066.8 ppm
Removed C3H8 Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
Zero Air Gen model:	Teledyne API T701H	Serial Number:	360

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1193585647
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.23E-04	2.96E-04	5.94E-05	4.35E-05
CH4 Retention time:	16.4	16.4	154457	210670
Zero Chromatogram:	ON	OFF	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.05</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	81.0	17.28	17.26	1.002
Mid point	4959	40.5	8.64	8.67	0.997
Low point	4979	20.2	4.31	4.33	0.997
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	81.0	17.28	17.27	1.001
Average Correction Factor					0.998

Notes: Resolved linearity issue, reset all parameters of instrument before calibration.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero					<i>Limit = 0.90-1.10</i>
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	81.0	9.17	9.17	1.000
Mid point	4959	40.5	4.58	4.61	0.994
Low point	4979	20.2	2.29	2.32	0.985
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	81.0	9.17	9.19	0.997
Average Correction Factor					0.993

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero					<i>Limit = 0.90-1.10</i>
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	81.0	8.11	8.09	1.003
Mid point	4959	40.5	4.06	4.06	1.000
Low point	4979	20.2	2.02	2.00	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	81.0	8.11	8.07	1.005
Average Correction Factor					1.004

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999037	0.998336
THC Cal Offset:	0.068218	0.017013
CH <sub>4</sub> Cal Slope:	0.994973	0.997807
CH <sub>4</sub> Cal Offset:	0.061988	-0.002818
NMHC Cal Slope:	1.002583	0.999129
NMHC Cal Offset:	0.006430	0.019031

Calibration Performed By: Kelly Baragar



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	May 8, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:30	End Time (MST):	14:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
17.28	17.26	1.0015	Slope	0.998336	<i>0.90 - 1.10</i>
8.64	8.67	0.9966	Intercept	0.017013	<i>+/-0.5</i>
4.31	4.33	0.9966			





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	May 8, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:30	End Time (MST):	14:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999991	<i>≥0.995</i>
8.11	8.09	1.0030	Slope	0.997807	<i>0.90 - 1.10</i>
4.06	4.06	0.9997	Intercept	-0.002818	<i>+/-0.5</i>
2.02	2.00	1.0098			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	May 8, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:30	End Time (MST):	14:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999980	<i>≥0.995</i>
9.17	9.17	0.9999	Slope	0.999129	<i>0.90 - 1.10</i>
4.58	4.61	0.9941	Intercept	0.019031	<i>+/-0.5</i>
2.29	2.32	0.9853			



NMHC Calibration Plot

Date: May 9, 2024

Location: Stony Mountain







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Stony Mountain  
 Station number: AMS 18  
 Calibration Date: May 15, 2024  
 Last Cal Date: April 23, 2024  
 Start time (MST): 8:23  
 End time (MST): 13:08  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T26DHGA  
 NOX Cal Gas Conc: 48.28 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 48.28 ppm  
 NOX gas Diff:  
 Calibrator Model: Teledyne API T700  
 ZAG make/model: Teledyne API T701  
 Cal Gas Expiry Date: November 17, 2026  
 NO Cal Gas Conc: 47.58 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 47.58 ppm  
 NO gas Diff:  
 Serial Number: 2658  
 Serial Number: 13779

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
AF High point	4916	84.0	811.1	799.3	11.8	801.3	789.7	11.6	1.0121	1.0120
AF Mid point										
AF Low point										
New cyl resp										

Previous Response NO<sub>x</sub> = 813.6 ppb NO = 802.3 ppb \* = > +/-5% change initiates investigation \*Percent Change NO<sub>x</sub> = -1.5%  
 Baseline Corr 1st pt NO<sub>x</sub> = 801.4 ppb NO = 789.9 ppb As Found Statistics \*Percent Change NO = -1.6%  
 Baseline Corr 2nd pt NO<sub>x</sub> = NA ppb NO = NA ppb As found NO<sub>x</sub> r<sup>2</sup>: Nx SI: Nx Int:  
 Baseline Corr 3rd pt NO<sub>x</sub> = NA ppb NO = NA ppb As found NO r<sup>2</sup>: NO SI: NO Int:  
 As found NO<sub>2</sub> r<sup>2</sup>: NO2 SI: NO2 Int:

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1336160088

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.003895	0.985690
NO <sub>x</sub> Cal Offset:	-0.640000	-1.180000
NO Cal Slope:	1.005239	0.986466
NO Cal Offset:	-1.220000	-1.380000
NO <sub>2</sub> Cal Slope:	1.001293	1.001655
NO <sub>2</sub> Cal Offset:	-0.223061	-0.094534

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.119	1.119	NO bkgnd or offset:	3.1	3.1
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	3.1	3.1
NO2 coeff or slope:	0.999	0.999	Reaction cell Press:	256.4	256.4

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4916	84.0	811.1	799.3	11.8	798.6	787.6	11.1	1.0157	1.0149
Mid point	4958	42.0	405.6	399.7	5.9	398.9	392.9	6.0	1.0167	1.0172
Low point	4979	21.0	202.8	199.8	2.9	196.9	193.9	3.0	1.0298	1.0306
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
As left span	4916	84.0	811.1	344.5	466.6	798.8	344.5	454.4	1.0154	1.0000
Average Correction Factor									1.0207	1.0209

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	787.0	346.7	452.1	452.6	0.9988	100.1%
Mid GPT point	787.0	580.7	218.1	218.8	0.9966	100.3%
Low GPT point	787.0	684.7	114.1	113.7	1.0032	99.7%
Average Correction Factor					0.9995	100.0%

Notes:

No adjustments needed.

Calibration Performed By:

Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 23, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	8:23	End Time (MST):	13:08
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999986	≥0.995
811.1	798.6	1.0157	Slope	0.985690	0.90 - 1.10
405.6	398.9	1.0167	Intercept	-1.180000	+/-20
202.8	196.9	1.0298			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 23, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	8:23	End Time (MST):	13:08
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999996	≥0.995
452.1	452.6	0.9988	Slope	1.001655	0.90 - 1.10
218.1	218.8	0.9966	Intercept	-0.094534	+/-20
114.1	113.7	1.0032			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 23, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	8:23	End Time (MST):	13:08
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999984	$\geq 0.995$
799.3	787.6	1.0149	Slope	0.986466	0.90 - 1.10
399.7	392.9	1.0172	Intercept	-1.380000	$\pm 20$
199.8	193.9	1.0306			



NO<sub>x</sub> Calibration Plot

Date: May 15, 2024

Location: Stony Mountain





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	May 21, 2024	Last Cal Date:	April 11, 2024
Start time (MST):	10:50	End time (MST):	14:42
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer	Serial Number:	2658
Calibrator Make/Model:	Teledyne API T700	Serial Number:	360
ZAG Make/Model:	Teledyne API 701H		

### Analyzer Information

Analyzer make:	API T400	Analyzer serial #:	825
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000029	1.000800	Backgd or Offset:	0.3	0.3
Calibration intercept:	-0.180000	-0.240000	Coeff or Slope:	0.982	0.966

### O<sub>3</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	NA	0.0	-0.3	----
As found High point	4804	1141.9	400.0	407.6	0.981
As found Mid point					
As found Low point					
Baseline Corr As found:	407.9	Previous response	399.8	*% change	2.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	NA	0.0	-0.1	----
High point	4888	1138.1	400.0	400.1	1.000
Mid point	4888	884.5	200.0	200.0	1.000
Low point	4888	741.4	100.0	99.6	1.004
As left zero	5000	NA	0.0	0.1	----
As left span	4812	1097.9	400.0	402.3	0.994
Average Correction Factor					1.001

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## O<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 21, 2024	Previous Calibration:	April 11, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:50	End Time (MST):	14:42
Analyzer make:	API T400	Analyzer serial #:	825

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999999	<span style="color: red;">≥0.995</span>
400.0	400.1	0.9998	Slope	1.000800	<span style="color: red;">0.90 - 1.10</span>
200.0	200.0	1.0000	Intercept	-0.240000	<span style="color: red;">+/- 5</span>
100.0	99.6	1.0040			






O<sub>3</sub> Calibration Plot

Date: May 21, 2024

Location: Stony Mountain





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Stony Mountain Station number: AMS 18  
 Calibration Date: May 16, 2024 Last Cal Date: April 26, 2024  
 Start time (MST): 11:10 End time (MST): 12:20

Analyzer Make: API T640 S/N: 1162  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388755  
 Temp/RH standard: Alicat FP-25BT S/N: 388755

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	12.6	12.15	12.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	692.2	691.75	692.2	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	5.04	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	36	----	36	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	20.0	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: October 10, 2024  
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.9		10.9	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: May 16, 2024  
 Date Disposable Filter Changed: February 21, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022  
 Date RH/T Sensor Cleaned: August 30, 2022

Notes: Instrument reading zero. Flow, pressure and temp checked, no issues. No response from PMT. Reset instrument and readings came back. All checks completed again. No issues. PMT check good. Leak check passed.

Calibration by: Devin Russell



# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Stony Mountain Station number: AMS 18  
 Calibration Date: May 21, 2024 Last Cal Date: May 16, 2024  
 Start time (MST): 13:55 End time (MST): 14:27

Analyzer Make: API T640 S/N: 1162  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748  
 Temp/RH standard: Alicat FP-25BT S/N: 388748

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.4	9.44	10.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	698.9	698.91	698.9	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.05	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	36	----	36	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.0	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: October 10, 2024  
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.9		10.9	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: May 16, 2024  
 Date Disposable Filter Changed: February 21, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: August 30, 2022  
 Date RH/T Sensor Cleaned: August 30, 2022

Notes: Flow, temp and pressure checked. Heated stack replaced. Leak check passed and PMT checked.

Calibration by: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Stony Mountain Station number: AMS 18  
 Calibration Date: May 29, 2024 Last Cal Date: N/A  
 Start time (MST): 12:40 End time (MST): 13:45

Analyzer Make: API T640 S/N: 766  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748  
 Temp/RH standard: Alicat FP-25BT S/N: 388748

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	16.3	16.24	16.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	691.2	692.99	691.2	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.97	5.09	4.97	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	----	37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.2	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: October 10, 2024  
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	N/A		N/A	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: \_\_\_\_\_ N/A  
 Date Disposable Filter Changed: \_\_\_\_\_ N/A

Post- maintenance Zero Verification: PM w/ HEPA: \_\_\_\_\_ 0.0 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_ August 30, 2022  
 Date RH/T Sensor Cleaned: \_\_\_\_\_ August 30, 2022

Notes: Flow, temp and pressure checked. Leak check passed.

Calibration by: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Stony Mountain Station number: AMS 18  
 Calibration Date: May 31, 2024 Last Cal Date: May 29, 2024  
 Start time (MST): 11:05 End time (MST): 12:16

Analyzer Make: API T640 S/N: 324  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748  
 Temp/RH standard: Alicat FP-25BT S/N: 388748

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.2	9.7	10.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	693.8	692.2	693.8	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.98	5.02	4.98	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	----	37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	8.2	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: October 10, 2024  
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.0		11.0	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: \_\_\_\_\_ N/A  
 Date Disposable Filter Changed: \_\_\_\_\_ N/A

Post- maintenance Zero Verification: PM w/ HEPA: \_\_\_\_\_ 0.0 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_ August 30, 2022  
 Date RH/T Sensor Cleaned: \_\_\_\_\_ August 30, 2022

Notes: Flow, temp and pressure checked. PMT checked and leak check passed.

Calibration by: Aswin Sasi Kumar



# Wood Buffalo Environmental Association

## CO Calibration Report

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	May 1, 2024	Last Cal Date:	April 5, 2024
Start time (MST):	9:15	End time (MST):	12:30
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	3,080	ppm	Cal Gas Exp Date: November 4, 2028
Cal Gas Cylinder #:	EB0065608		
Removed Cal Gas Conc:	3,080	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 2658
ZAG Make/Model:	Teledyne API T701H		Serial Number: 355

### Analyzer Information

Analyzer make:	API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006642	0.994403	Backgd or Offset:	-0.010	-0.011
Calibration intercept:	0.249837	0.079799	Coeff or Slope:	0.905	0.902

### CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4933	66.7	41.1	41.5	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	41.26	Prev response:	41.61	*% change:	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4933	66.7	41.1	40.9	1.005
Mid point	4966	33.3	20.5	20.7	0.993
Low point	4983	16.7	10.3	10.2	1.006
As left zero	5000	0.0	0.0	0.0	----
As left span	4933	66.7	41.1	40.7	1.010
Average Correction Factor					1.001

Notes: Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## CO Calibration Summary

### Station Information

Calibration Date:	May 1, 2024	Previous Calibration:	April 5, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:15	End Time (MST):	12:30
Analyzer make:	API T300	Analyzer serial #:	3504

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999951	<b>≥0.995</b>
41.1	40.9	1.0054	Slope	0.994403	<b>0.90 - 1.10</b>
20.5	20.7	0.9930	Intercept	0.079799	<b>+/-1.5</b>
10.3	10.2	1.0057			



CO Calibration Plot

Date: May 1, 2024

Location: Stony Mountain







# Wood Buffalo Environmental Association

## CO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	May 23, 2024	Last Cal Date:	April 26, 2024
Start time (MST):	10:44	End time (MST):	14:06
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	59,100	ppm	Cal Gas Exp Date: November 4, 2028
Cal Gas Cylinder #:	EB0065608		
Removed Cal Gas Conc:	59,100	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 2658
N2 Gen Make/Model:	Peak Scientific		Serial Number: 771048317

### Analyzer Information

Analyzer make:	API T360	Analyzer serial #: 489
Analyzer Range	0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006157	1.007099	Backgd or Offset:	-0.037
Calibration intercept:	-3.120000	-1.320000	Coeff or Slope:	0.939

### CO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	3000	0.0	0.0	1.0	----
As found High Point	2920	80.0	1576.0	1592.0	0.991
As found Mid Point					
As found Low Point					
New cylinder response					
Baseline Corr As found:	1591.0	Prev response:	1582.6	*% change:	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	3000	0.0	0.0	1.0	----
High point	2920	80.0	1576.0	1588.8	0.992
Mid point	2960	40.0	788.0	785.8	1.003
Low point	2980	20.0	394.0	396.7	0.993
As left zero	3000	0.0	0.0	0.7	----
As left span	2930	80.0	1570.8	1586.4	0.990
Average Correction Factor					0.996

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



# Wood Buffalo Environmental Association


## CO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date	May 23, 2024	Previous Calibration	April 26, 2024
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:44	End Time (MST)	14:06
Analyzer make	API T360	Analyzer serial #	489

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.0	----	Correlation Coefficient	0.999958	<b>≥0.995</b>
1576.0	1588.8	0.9919	Slope	1.007099	<b>0.90 - 1.10</b>
788.0	785.8	1.0028	Intercept	-1.3	<b>+/-20</b>
394.0	396.7	0.9932			



CO<sub>2</sub> Calibration Plot

Date: May 23, 2024

Location: Stony Mountain





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS19 FIREBAG MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	May 15, 2024	Last Cal Date:	April 22, 2024
Start time (MST):	10:11	End time (MST):	14:18
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.29	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC716618			
Removed Cal Gas Conc:	49.29	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	1607
Zero Air Gen Model:	Teledyne API T701H		Serial Number:	201

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1410661308
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996719	0.995775	Backgd or Offset:	10.6	10.6
Calibration intercept:	0.438192	0.818481	Coeff or Slope:	0.991	0.991

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	4999	0.0	0.0	0.0	----
As found High point	4919	81.1	799.5	797.3	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.3	Previous response	797.3	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	4999	0.0	0.0	0.0	----
High point	4919	81.1	799.5	796.5	1.004
Mid point	4959	40.6	400.3	399.8	1.001
Low point	4980	20.3	200.1	200.9	0.996
As left zero	4999	0.0	0.0	0.2	----
As left span	4919	81.1	799.5	797.4	1.003
				Average Correction Factor:	1.000

Notes: Calibrated remotely due to evacuation notice. No adjustments made.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 22, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:11	End Time (MST):	14:18
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999995	<b>≥0.995</b>
799.5	796.5	1.0037	Slope	0.995775	<b>0.90 - 1.10</b>
400.3	399.8	1.0012	Intercept	0.818481	<b>+/-30</b>
200.1	200.9	0.9960			



SO2 Calibration Plot

Date: May 15, 2024

Location: Firebag





# Wood Buffalo Environmental Association

## H2S Calibration Report

### Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	May 16, 2024	Last Cal Date:	April 16, 2024
Start time (MST):	10:51	End time (MST):	16:55
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.114	ppm	Cal Gas Exp Date:	February 5, 2024
Cal Gas Cylinder #:	CC517427			
Removed Cal Gas Conc:	5.114	ppm	Rem Gas Exp Date:	n/a
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1607
ZAG Make/Model:	Teledyne API T701		Serial Number:	201

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032
Converter make:	Global	Converter serial #:	2022-222
Analyzer Range	0 - 100 ppb	Converter Temp:	350 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001197	1.002480	Backgd or Offset:	2.72	2.72
Calibration intercept:	0.078370	-0.041548	Coeff or Slope:	1.163	1.163

### H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4922	78.2	80.0	81.6	0.979
As found Mid point	4961	39.1	40.0	40.7	0.980
As found Low point	4980	19.6	20.0	20.0	0.997
New cylinder response					
Baseline Corr As found:	81.7	Prev response:	80.15	*% change:	1.9%
Baseline Corr 2nd AF pt:	40.8	AF Slope:	1.023060	AF Intercept:	-0.261942
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999975	<i>* = &gt; +/-5% change initiates investigation</i>	

### H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4922	78.2	80.0	80.1	0.998
Mid point	4961	39.1	40.0	40.2	0.995
Low point	4980	19.6	20.0	19.9	1.007
As left zero	5000	0.0	0.0	0.0	----
As left span	4922	78.2	80.0	78.8	1.015
SO2 Scrubber Check	4922	78.3	783.0	0.0	----
Date of last scrubber change:		18-Jan-23		Ave Corr Factor	1.000
Date of last converter efficiency test:		n/a			

Notes: Calibrated remotely due to evacuation notice. Scrubber check passed. No adjustments made.

Calibration Performed By: Braiden Boutillier





# Wood Buffalo Environmental Association


## H2S Calibration Summary

### Station Information

Calibration Date:	May 16, 2024	Previous Calibration:	April 16, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:51	End Time (MST):	16:55
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999986	$\geq 0.995$
80.0	80.1	0.9985	Slope	1.002480	$0.90 - 1.10$
40.0	40.2	0.9948	Intercept	-0.041548	$\pm 3$
20.0	19.9	1.0075			



H2S Calibration Plot

Date: May 16, 2024

Location: Firebag







# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 22, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:11	End Time (MST):	14:18
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.08	----	Correlation Coefficient	0.999992	<span style="color: red;">≥0.995</span>
17.31	17.29	1.0009	Slope	1.003351	<span style="color: red;">0.90 - 1.10</span>
8.66	8.58	1.0098	Intercept	-0.083575	<span style="color: red;">+/-1.5</span>
4.33	4.28	1.0120			



THC Calibration Plot

Date: May 15, 2024

Location: Firebag





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Firebag  
 Station number: AMS 19  
 Calibration Date: May 23, 2024  
 Last Cal Date: April 4, 2024  
 Start time (MST): 10:17  
 End time (MST): 14:54  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: DT0044018  
 NOX Cal Gas Conc: 48.90 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 48.90 ppm  
 NOX gas Diff:  
 Calibrator Model: Teledyne API T700  
 ZAG make/model: Teledyne API T701

Cal Gas Expiry Date: November 3, 2031  
 NO Cal Gas Conc: 48.70 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 48.70 ppm  
 NO gas Diff:  
 Serial Number: 1607  
 Serial Number: 201

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
AF High point	4918	82.1	802.9	799.7	3.3	861.0	855.0	5.2	0.9326	0.9352
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 803.9 ppb		NO = 801.3 ppb			<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = 6.6%	
Baseline Corr 1st pt	NO <sub>x</sub> = 861.0 ppb		NO = 855.1 ppb			<u>As Found Statistics</u>		*Percent Change	NO = 6.3%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO r <sup>2</sup> :	NO SI:	NO Int:	
						As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1410661309

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.998971	0.997818
NO <sub>x</sub> Cal Offset:	1.739938	1.520105
NO Cal Slope:	1.001345	1.001331
NO Cal Offset:	0.620021	0.699868
NO <sub>2</sub> Cal Slope:	1.000081	0.999250
NO <sub>2</sub> Cal Offset:	0.547188	-0.951936

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.990	0.927	NO bkgnd or offset:	5.0	4.6
NOX coeff or slope:	0.995	0.993	NOX bkgnd or offset:	5.0	4.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	177.5	166.7

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4918	82.1	802.9	799.7	3.3	802.0	801.0	1.1	1.0012	0.9983
Mid point	4959	41.1	402.0	400.3	1.6	403.2	402.1	1.1	0.9969	0.9956
Low point	4980	20.5	200.5	199.7	0.8	203.2	201.2	2.0	0.9867	0.9924
As left zero	5000	0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
As left span	4918	82.1	802.9	388.6	414.3	804.0	388.6	415.1	0.9987	1.0000
Average Correction Factor									0.9949	0.9954

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	799.1	388.3	414.1	413.4	1.0017	99.8%
Mid GPT point	799.1	594.4	208.0	206.1	1.0091	99.1%
Low GPT point	799.1	698.8	103.6	101.8	1.0175	98.3%
Average Correction Factor					1.0094	99.1%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 23, 2024	Previous Calibration:	April 4, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:17	End Time (MST):	14:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999983	≥0.995
802.9	802.0	1.0012	Slope	0.997818	0.90 - 1.10
402.0	403.2	0.9969	Intercept	1.520105	+/-20
200.5	203.2	0.9867			







# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 23, 2024	Previous Calibration:	April 4, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:17	End Time (MST):	14:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999974	≥0.995
414.1	413.4	1.0017	Slope	0.999250	0.90 - 1.10
208.0	206.1	1.0091	Intercept	-0.951936	+/-20
103.6	101.8	1.0175			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 23, 2024	Previous Calibration:	April 4, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:17	End Time (MST):	14:54
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999996	<i>≥0.995</i>
799.7	801.0	0.9983	Slope	1.001331	<i>0.90 - 1.10</i>
400.3	402.1	0.9956	Intercept	0.699868	<i>+/-20</i>
199.7	201.2	0.9924			



NO<sub>x</sub> Calibration Plot

Date: May 23, 2024

Location: Firebag





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS20 MACKAY RIVER MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	MacKay River	Station number:	AMS 20
Calibration Date:	May 9, 2024	Last Cal Date:	April 10, 2024
Start time (MST):	7:15	End time (MST):	10:02
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.22	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC30686			
Removed Cal Gas Conc:	49.22	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	1220
Zero Air Gen Model:	API 701		Serial Number:	4522

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1501301450
Analyzer Range:	0-1000ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994567	1.003462	Backgd or Offset:	19.3	19.3
Calibration intercept:	3.111267	3.271175	Coeff or Slope:	0.950	0.950

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4919	81.3	800.3	804.4	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	804.1	Previous response	799.0	*% change	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4919	81.3	800.3	805.3	0.994
Mid point	4959	40.7	400.7	405.5	0.988
Low point	4980	20.3	199.8	207.5	0.963
As left zero	5000	0.0	0.0	0.4	----
As left span	4919	81.3	800.3	806.8	0.992
Average Correction Factor:					0.982

Notes: No adjustments or maintenance done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	April 10, 2024
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	7:15	End Time (MST):	10:02
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.4	----	Correlation Coefficient	0.999935	<b>≥0.995</b>
800.3	805.3	0.9938	Slope	1.003462	<b>0.90 - 1.10</b>
400.7	405.5	0.9881	Intercept	3.271175	<b>+/-30</b>
199.8	207.5	0.9630			



SO2 Calibration Plot

Date: May 9, 2024

Location: MacKay River





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

### Station Information

Station Name: MacKay River	Station number: AMS 20
Calibration Date: May 8, 2024	Last Cal Date: April 9, 2024
Start time (MST): 7:00	End time (MST): 11:57
Reason: Routine	

### Calibration Standards

Cal Gas Concentration: 5.12 ppm	Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC515997	
Removed Cal Gas Conc: 5.12 ppm	Rem Gas Exp Date:
Removed Gas Cyl #:	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 1220
ZAG Make/Model: API 701	Serial Number: 4522

### Analyzer Information

Analyzer make: Thermo 43i TLE	Analyzer serial #: 1236656117
Converter make: Global	Converter serial #: 2022-226
Analyzer Range: 0 - 100 ppb	Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.984302	0.992309	Backgd or Offset:	3.19	2.99
Calibration intercept:	0.679249	0.699268	Coeff or Slope:	1.113	1.051

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4922	78.1	80.0	79.9	1.002
As found Mid point	4961	39.0	39.9	40.3	0.993
As found Low point	4980	19.5	20.0	20.7	0.969
New cylinder response					
Baseline Corr As found:	79.8	Prev response:	79.40	*% change:	0.5%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	0.995450	AF Intercept:	0.439415
Baseline Corr 3rd AF pt:	20.6	AF Correlation:	0.999915	<i>* = &gt; +/-5% change initiates investigation</i>	

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.6	----
High point	4922	78.1	80.0	80.0	1.000
Mid point	4961	39.0	39.9	40.4	0.989
Low point	4980	19.5	20.0	20.6	0.969
As left zero	5000	0.0	0.0	0.6	----
As left span	4922	78.1	80.0	79.1	1.011
SO <sub>2</sub> Scrubber Check	4982	81.3	802.8	-0.1	----
Date of last scrubber change:	25-May-23			Ave Corr Factor	<b>0.986</b>
Date of last converter efficiency test:					

Notes: Lamp Voltage and intensity have been fluctuating. Lamp and Socket replaced. Sox scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay





# Wood Buffalo Environmental Association


## H<sub>2</sub>S Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 9, 2024
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	7:00	End Time (MST):	11:57
Analyzer make:	Global	Analyzer serial #:	2022-226

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999993	≥0.995
80.0	80.0	0.9997	Slope	0.992309	0.90 - 1.10
39.9	40.4	0.9885	Intercept	0.699268	+/-3
20.0	20.6	0.9694			



H<sub>2</sub>S Calibration Plot

Date: May 8, 2024

Location: MacKay River





# Wood Buffalo Environmental Association

## THC Calibration Report

### Station Information

Station Name: MacKay River  
Calibration Date: May 9, 2024  
Start time (MST): 7:15  
Reason: Routine  
Station number: AMS 20  
Last Cal Date: April 10, 2024  
End time (MST): 10:01

### Calibration Standards

Gas Cert Reference: CC306868  
CH4 Cal Gas Conc. 499.4 ppm  
C3H8 Cal Gas Conc. 206.2 ppm  
Removed Gas Cert: 499.4 ppm  
Removed CH4 Conc. 206.2 ppm  
Calibrator Make/Model: API T700  
ZAG Make/Model: API 701  
Cal Gas Expiry Date: February 23, 2024  
CH4 Equiv Conc. 1066.5 ppm  
Removed Gas Expiry: 1066.5 ppm  
CH4 Equiv Conc. 1066.5 ppm  
Diff between cyl:  
Serial Number: 1220  
Serial Number: 4522

### Analyzer Information

Analyzer make: Thermo 51i-LT  
Analyzer Range: 0 - 20 ppm  
Analyzer serial #: 1501663727

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995490	1.000613	Background: 3.680	3.680
Calibration intercept:	-0.003380	0.009001	Coefficient: 6.124	6.124

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted
					Correction factor (Cc/Ic-AFzero)
As found zero	5000	0.0	0.00	-0.03	---- <i>Limit = 0.90-1.10</i>
As found High point	4919	81.3	17.34	17.48	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.50	Previous response	17.26	*% change	1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)
					<i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	-0.05	----
High point	4919	81.3	17.34	17.34	1.000
Mid point	4959	40.7	8.68	8.71	0.997
Low point	4980	20.3	4.33	4.41	0.983
As left zero	5000	0.0	0.00	-0.10	----
As left span	4919	81.3	17.34	17.40	0.996
Average Correction Factor					<b>0.993</b>

Notes: Hydrogen Cylinder changed. No adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	April 10, 2024
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	7:15	End Time (MST):	10:01
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.05	----	Correlation Coefficient	0.999954	<span style="color: red;">≥0.995</span>
17.34	17.34	1.0001	Slope	1.000613	<span style="color: red;">0.90 - 1.10</span>
8.68	8.71	0.9969	Intercept	0.009001	<span style="color: red;">+/-1.5</span>
4.33	4.41	0.9829			



THC Calibration Plot

Date: May 9, 2024

Location: MacKay River





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: MacKay River  
 Station number: AMS 20  
 Calibration Date: May 7, 2024  
 Last Cal Date: April 8, 2024  
 Start time (MST): 7:10  
 End time (MST): 11:33  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T376265  
 NOX Cal Gas Conc: 49.19 ppm  
 Removed Cylinder #:  
 Removed Gas NOX Conc: 49.19 ppm  
 NOX gas Diff:  
 Calibrator Model: API T700  
 ZAG make/model: API T701  
 Cal Gas Expiry Date: April 13, 2025  
 NO Cal Gas Conc: 48.04 ppm  
 Removed Gas Exp Date:  
 Removed Gas NO Conc: 48.04 ppm  
 NO gas Diff:  
 Serial Number: 1220  
 Serial Number: 4522

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
AF High point	4917	83.3	819.5	800.3	19.2	815.2	794.3	21.0	1.0051	1.0076
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 818.6 ppb		NO = 798.3 ppb			<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -0.4%	
Baseline Corr 1st pt	NO <sub>x</sub> = 815.3 ppb		NO = 794.3 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -0.5%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO r <sup>2</sup> :	NO SI:	NO Int:	
						As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1505164379

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.994162	0.996087
NO <sub>x</sub> Cal Offset:	3.902461	4.362190
NO Cal Slope:	0.993845	0.996287
NO Cal Offset:	2.922811	3.342620
NO <sub>2</sub> Cal Slope:	1.003001	1.002860
NO <sub>2</sub> Cal Offset:	-0.362701	0.059293

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.990	0.997	NO bkgnd or offset:	2.8	2.8
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	3.0	3.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	164.2	164.2

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	0.0	----	----
High point	4917	83.3	819.5	800.3	19.2	818.6	799.2	19.4	1.0010	1.0014
Mid point	4958	41.7	410.3	400.7	9.6	414.7	403.7	11.0	0.9893	0.9925
Low point	4979	20.8	204.6	199.9	4.8	212.9	206.0	6.9	0.9612	0.9702
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	-0.1	----	----
As left span	4917	83.3	819.5	440.1	379.4	819.8	440.1	379.6	0.9996	1.0000
Average Correction Factor									0.9839	0.9880

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	796.8	439.5	376.5	377.5	0.9972	100.3%
Mid GPT point	796.8	611.9	204.1	204.9	0.9959	100.4%
Low GPT point	796.8	701.0	115.0	115.3	0.9970	100.3%
Average Correction Factor					0.9967	100.3%

Notes: Span adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 8, 2024
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	7:10	End Time (MST):	11:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999868	≥0.995
819.5	818.6	1.0010	Slope	0.996087	0.90 - 1.10
410.3	414.7	0.9893	Intercept	4.362190	+/-20
204.6	212.9	0.9612			







# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 8, 2024
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	7:10	End Time (MST):	11:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999999	≥0.995
376.5	377.5	0.9972	Slope	1.002860	0.90 - 1.10
204.1	204.9	0.9959	Intercept	0.059293	+/-20
115.0	115.3	0.9970			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 7, 2024	Previous Calibration:	April 8, 2024
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	7:10	End Time (MST):	11:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999924	≥0.995
800.3	799.2	1.0014	Slope	0.996287	0.90 - 1.10
400.7	403.7	0.9925	Intercept	3.342620	+/-20
199.9	206.0	0.9702			



NO<sub>x</sub> Calibration Plot

Date: May 7, 2024

Location: MacKay River





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS21  
CONKLIN  
MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	May 2, 2024	Last Cal Date:	April 3, 2024
Start time (MST):	10:11	End time (MST):	13:15
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.93	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC259455			
Removed Cal Gas Conc:	49.93	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	3810
Zero Air Gen Model:	Teledyne API 701		Serial Number:	953

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1428701363
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992525	1.002413	Backgd or Offset:	28.3	28.3
Calibration intercept:	1.395547	2.056037	Coeff or Slope:	0.901	0.901

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.5	----
As found High point	4920	80.2	800.8	802.0	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	801.5	Previous response	796.3	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.6	----
High point	4920	80.2	800.8	804.3	0.996
Mid point	4960	40.1	400.4	403.6	0.992
Low point	4980	20.0	200.1	204.5	0.979
As left zero	5005	0.0	0.0	0.5	----
As left span	4920	80.2	800.8	804.3	0.996
Average Correction Factor:					0.989

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 2, 2024	Previous Calibration:	April 3, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:11	End Time (MST):	13:15
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999983
800.8	804.3	0.9957	Slope	1.002413
400.4	403.6	0.9921	Intercept	2.056037
200.1	204.5	0.9786		
				<span style="color: red;">≥0.995</span> <span style="color: red;">0.90 - 1.10</span> <span style="color: red;">+/-30</span>



SO2 Calibration Plot

Date: May 2, 2024

Location: Conklin





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	May 28, 2024	Last Cal Date:	April 18, 2024
Start time (MST):	9:30	End time (MST):	14:00
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.00	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC501204			
Removed Cal Gas Conc:	5.00	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3810
ZAG Make/Model:	Teledyne API 701H		Serial Number:	691

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116
Converter make:	CD-Nova 101	Converter serial #:	NA
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999714	0.997857	Backgd or Offset:	2.8	2.4
Calibration intercept:	0.060000	0.200000	Coeff or Slope:	0.998	0.923

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4920	80.0	80.0	84.0	0.949
As found Mid point	4960	40.0	40.0	42.1	0.943
As found Low point	4980	20.0	20.0	21.1	0.935
New cylinder response					
Baseline Corr As found:	84.3	Prev response:	80.04	*% change:	5.1%
Baseline Corr 2nd AF pt:	42.4	AF Slope:	1.052714	AF Intercept:	-0.120000
Baseline Corr 3rd AF pt:	21.4	AF Correlation:	0.999979	* = > +/-5% change initiates investigation	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.0	80.0	79.9	1.001
Mid point	4960	40.0	40.0	40.3	0.993
Low point	4980	20.0	20.0	20.3	0.985
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	80.0	80.0	80.0	1.000
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:	28-May-24			Ave Corr Factor	0.993
Date of last converter efficiency test:					

Notes: Sample inlet filters and scrubber was changed after multipoint as founds. SO2 scrubber check done after calibrator zero and passed. Adjusted zero and span.

Calibration Performed By: Jan Castro





# Wood Buffalo Environmental Association


## TRS Calibration Summary

### Station Information

Calibration Date:	May 28, 2024	Previous Calibration:	April 18, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:30	End Time (MST):	14:00
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999968	$\geq 0.995$
80.0	79.9	1.0013	Slope	0.997857	$0.90 - 1.10$
40.0	40.3	0.9926	Intercept	0.200000	$\pm 3$
20.0	20.3	0.9852			



TRS Calibration Plot

Date: May 28, 2024

Location: Conklin





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	May 2, 2024	Last Cal Date:	April 19, 2024
Start time (MST):	10:11	End time (MST):	13:14
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	
CH4 Cal Gas Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
C3H8 Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
Removed C3H8 Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
Zero Air Gen model:	Teledyne API 701H	Serial Number:	953

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1331259520
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.48E-04	2.63E-04	NMHC SP Ratio:	4.50E-05	6.48E-05
CH4 Retention time:	15.4	14.4	NMHC Peak Area:	202994	140997
Zero Chromatogram:	ON	ON	Flat Baseline:	ON	ON

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	17.13	17.93	0.955
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.93	Prev response	17.16	*% change	4.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	17.13	17.15	0.998
Mid point	4960	40.1	8.56	8.59	0.997
Low point	4980	20.0	4.28	4.37	0.979
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	17.13	17.10	1.001
Average Correction Factor					0.991

Notes: Changed the inlet filter after as founds. Adjusted the span.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	9.14	9.89	0.924
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.89	Prev response	9.13	*% change	7.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	9.14	9.13	1.001
Mid point	4960	40.1	4.57	4.58	0.998
Low point	4980	20.0	2.28	2.34	0.976
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	9.14	9.09	1.006
Average Correction Factor					0.992

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	7.99	8.04	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.04	Prev response	8.02	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	7.99	8.02	0.995
Mid point	4960	40.1	3.99	4.01	0.996
Low point	4980	20.0	2.00	2.03	0.981
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	7.99	8.01	0.997
Average Correction Factor					0.991

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999013	0.999907
THC Cal Offset:	0.053566	0.037370
CH <sub>4</sub> Cal Slope:	1.000152	1.003372
CH <sub>4</sub> Cal Offset:	0.035557	0.011560
NMHC Cal Slope:	0.996817	0.997092
NMHC Cal Offset:	0.019809	0.025211

Calibration Performed By: Max Farrell



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 2, 2024	Previous Calibration:	April 19, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:11	End Time (MST):	13:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999970	≥0.995
17.13	17.15	0.9984	Slope	0.999907	0.90 - 1.10
8.56	8.59	0.9969	Intercept	0.037370	+/-0.5
4.28	4.37	0.9786			





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 2, 2024	Previous Calibration:	April 19, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:11	End Time (MST):	13:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999983	<i>≥0.995</i>
7.99	8.02	0.9954	Slope	1.003372	<i>0.90 - 1.10</i>
3.99	4.01	0.9955	Intercept	0.011560	<i>+/-0.5</i>
2.00	2.03	0.9811			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 2, 2024	Previous Calibration:	April 19, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:11	End Time (MST):	13:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999955	<i>≥0.995</i>
9.14	9.13	1.0009	Slope	0.997092	<i>0.90 - 1.10</i>
4.57	4.58	0.9984	Intercept	0.025211	<i>+/-0.5</i>
2.28	2.34	0.9764			



NMHC Calibration Plot

Date: May 2, 2024

Location: Conklin







# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	May 24, 2024	Last Cal Date:	May 2, 2024
Start time (MST):	9:37	End time (MST):	11:24
Reason:	Cylinder Change Support Gas N2 and H2		

### Calibration Standards

Gas Cert Reference:	CC259455	Cal Gas Expiry Date:	
CH4 Cal Gas Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
C3H8 Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
Removed C3H8 Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3810
Zero Air Gen model:	Teledyne API 701H	Serial Number:	953

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1331259520
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.63E-04	2.63E-04	NMHC SP Ratio:	6.48E-05	6.48E-05
CH4 Retention time:	14.4	14.4	NMHC Peak Area:	140997	140997
Zero Chromatogram:	ON	ON	Flat Baseline:	ON	ON

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	17.13	17.50	0.979
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.50	Prev response	17.16	*% change	1.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	17.13	17.43	0.983
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 0.983

Notes: N2 and H2 cylinder changed after as founds. No adjustment made.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	9.14	9.48	0.964
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.48	Prev response	9.14	*% change	3.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	9.14	9.47	0.965
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 0.965

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	7.99	8.02	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.02	Prev response	8.02	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	7.99	7.96	1.003
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.003

### Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999907	1.017505
THC Cal Offset:	0.037370	0.000000
CH <sub>4</sub> Cal Slope:	1.003372	0.996620
CH <sub>4</sub> Cal Offset:	0.011560	0.000000
NMHC Cal Slope:	0.997092	1.035864
NMHC Cal Offset:	0.025211	0.000000

Calibration Performed By: Jan Castro

NMHC Calibration Plot

Date: May 24, 2024

Location: Conklin





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Conklin  
 Station number: AMS 21  
 Calibration Date: May 8, 2024  
 Last Cal Date: April 17, 2024  
 Start time (MST): 8:59  
 End time (MST): 13:24  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: SA18828  
 NOX Cal Gas Conc: 48.90 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 48.90 ppm  
 NOX gas Diff:  
 Calibrator Model: Teledyne API T700  
 ZAG make/model: Teledyne API T701H  
 Cal Gas Expiry Date: November 3, 2031  
 NO Cal Gas Conc: 48.80 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 48.80 ppm  
 NO gas Diff:  
 Serial Number: 3810  
 Serial Number: 953

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.8	-1.7	-0.1	----	----
AF High point	4918	82.0	802.0	800.3	1.6	758.8	749.9	8.8	1.0544	1.0648
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 804.0 ppb		NO = 800.2 ppb			<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -5.7%	
Baseline Corr 1st pt	NO <sub>x</sub> = 760.6 ppb		NO = 751.6 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -6.5%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO r <sup>2</sup> :	NO SI:	NO Int:	
						As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1501663731

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.999799	0.998117
NO <sub>x</sub> Cal Offset:	2.168069	2.108054
NO Cal Slope:	0.998620	0.999906
NO Cal Offset:	0.948023	1.148032
NO <sub>2</sub> Cal Slope:	1.011507	1.002080
NO <sub>2</sub> Cal Offset:	-1.127282	-1.209969

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.098	1.173	NO bkgnd or offset:	12.0	11.5
NOX coeff or slope:	0.999	0.993	NOX bkgnd or offset:	12.2	11.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.8	151.2

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.1	----	----
High point	4918	82.0	802.0	800.3	1.6	801.5	800.8	0.6	1.0006	0.9994
Mid point	4959	41.0	401.0	400.2	0.8	403.2	401.7	1.5	0.9945	0.9962
Low point	4980	20.5	200.5	200.1	0.4	204.7	202.7	2.0	0.9793	0.9870
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
As left span	4918	82.0	802.0	386.6	415.4	803.0	386.6	416.4	0.9987	1.0000
Average Correction Factor									0.9915	0.9942

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	803.6	383.3	421.9	421.9	1.0001	100.0%
Mid GPT point	803.6	601.1	204.1	203.8	1.0017	99.8%
Low GPT point	803.6	700.6	104.6	101.6	1.0299	97.1%
Average Correction Factor					1.0106	99.0%

Notes: Sample inlet filters changed after as founds. Adjusted zero and span.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 17, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	8:59	End Time (MST):	13:24
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999962	≥0.995
802.0	801.5	1.0006	Slope	0.998117	0.90 - 1.10
401.0	403.2	0.9945	Intercept	2.108054	+/-20
200.5	204.7	0.9793			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 17, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	8:59	End Time (MST):	13:24
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999937	≥0.995
421.9	421.9	1.0001	Slope	1.002080	0.90 - 1.10
204.1	203.8	1.0017	Intercept	-1.209969	+/-20
104.6	101.6	1.0299			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 17, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	8:59	End Time (MST):	13:24
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999987	<span style="color: red;">≥0.995</span>
800.3	800.8	0.9994	Slope	0.999906	<span style="color: red;">0.90 - 1.10</span>
400.2	401.7	0.9962	Intercept	1.148032	<span style="color: red;">+/-20</span>
200.1	202.7	0.9870			






NO<sub>x</sub> Calibration Plot

Date: May 8, 2024

Location: Conklin





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

### Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	May 17, 2024	Last Cal Date:	April 8, 2024
Start time (MST):	7:35	End time (MST):	10:08
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3810
Calibrator Make/Model:	Teledyne API T700	Serial Number:	691
ZAG Make/Model:	Teledyne API 701H		

### Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999943	0.999857	Backgd or Offset:	-1.1	-1.1
Calibration intercept:	0.060000	0.100000	Coeff or Slope:	0.998	0.998

### O<sub>3</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-0.3	----
As found High point	5000	955.4	400.0	399.8	1.000
As found Mid point					
As found Low point					
Baseline Corr As found:	400.1	Previous response	400.0	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	-0.3	----
High point	5000	955.4	400.0	399.8	1.001
Mid point	5000	809.3	200.0	200.4	0.998
Low point	5000	706.5	100.0	100.4	0.996
As left zero	5000	800.0	0.0	0.3	----
As left span	5000	951.8	400.0	401.4	0.997
Average Correction Factor					<b>0.998</b>

Notes: Remote calibrations. No adjustment made.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association


## O<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 17, 2024	Previous Calibration:	April 8, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	7:35	End Time (MST):	10:08
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999995	<span style="color: red;">≥0.995</span>
400.0	399.8	1.0005	Slope	0.999857	<span style="color: red;">0.90 - 1.10</span>
200.0	200.4	0.9980	Intercept	0.100000	<span style="color: red;">+/- 5</span>
100.0	100.4	0.9960			



O<sub>3</sub> Calibration Plot

Date: May 17, 2024

Location: Conklin





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Conklin Station number: AMS 21  
 Calibration Date: May 28, 2024 Last Cal Date: April 18, 2024  
 Start time (MST): 10:34 End time (MST): 11:32

Analyzer Make: API T640 S/N: 326  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754  
 Temp/RH standard: Alicat FP-25BT S/N: 388754

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	20.90	21.9	20.90	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	707.80	709.88	707.80	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.11	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	----	37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	5.70	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: September 29, 2024  
Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.10	10.90	10.90	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: May 28, 2024  
Date Disposable Filter Changed: May 28, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: December 7, 2023  
Date RH/T Sensor Cleaned: December 7, 2023

Notes: Verified flow, pressure, temperature and pump power. Disposable filter changed. Leak check passed. No adjustment made.

Calibration by: Jan Castro



# Wood Buffalo Environmental Association

## Wind Speed/Direction Calibration Report

Version-10-2022

### Station Information

Station Name:	Conklin	Station Number:	AMS 21
Calibration Date:	May 17, 2024	Prev Cal Date:	August 30, 2023
Start Time (MST):	11:30	End Time (MST):	12:40
Tower Height (m):	10.0	Reason:	Routine

### Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	J4337
WS Calibrator:	MetOne 053-120	Serial Number:	R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.5%
400	39.4	39.4	0.1%
600	58.6	58.7	0.2%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999999	$\geq 0.9995$
Calculated slope		0.998853	$0.90 - 1.10$
Calculated intercept		-0.027978	$\pm 2$

### Wind Direction Information

Sensor make/model:	Met One 020D	Serial Number:	D14062
As Found Declination (deg east of True North):	<u>13</u>	As Left Declination (deg east of True North):	<u>13</u>
Solar noon time (MST):	NA	Calc Declination*:	13 Degrees
Deadband calc:	-2.7 degrees ( <i>Limit 4 deg</i> )	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	-0.4	---
90	88.3	-0.5%
180	178.8	-0.3%
270	269.6	-0.1%
358	359.4	0.4%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r <sup>2</sup> )		0.999980	$\geq 0.9995$
Calculated slope		0.994697	$0.90 - 1.10$
Calculated intercept		1.427878	$\pm 4$

Notes: Bearings still good. Confirmed declination with a compass. Replaced Wind Direction sensor as previous unit with water damage

Calibration Performed By: Ryan Power



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

**AMS22  
JANVIER  
MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	May 28, 2024	Last Cal Date:	April 12, 2024
Start time (MST):	10:36	End time (MST):	14:08
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.11	ppm	Cal Gas Exp Date: January 18, 2029
Cal Gas Cylinder #:	CC281519		
Removed Cal Gas Conc:	50.11	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	Teledyne API T700		Serial Number: 3806
Zero Air Gen Model:	Teledyne API T701		Serial Number: 4890

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1152430006
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002236	1.000307	Backgd or Offset:	24.1	23.9
Calibration intercept:	-0.036002	0.464235	Coeff or Slope:	1.013	1.009

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4920	79.8	799.8	802.1	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	802.0	Previous response	801.5	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	79.8	799.8	800.3	0.999
Mid point	4960	39.9	399.9	400.8	0.998
Low point	4980	20.0	200.4	201.1	0.997
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	79.8	799.8	799.8	1.000
Average Correction Factor:					0.998

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland





# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 28, 2024	Previous Calibration:	April 12, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:36	End Time (MST):	14:08
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999999	<span style="color: red;">≥0.995</span>
799.8	800.3	0.9994	Slope	1.000307	<span style="color: red;">0.90 - 1.10</span>
399.9	400.8	0.9977	Intercept	0.464235	<span style="color: red;">+/-30</span>
200.4	201.1	0.9967			



SO2 Calibration Plot

Date: May 28, 2024

Location: Janvier





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name: Janvier	Station number: AMS 22
Calibration Date: May 23, 2024	Last Cal Date: April 18, 2024
Start time (MST): 10:50	End time (MST): 14:55
Reason: Routine	

### Calibration Standards

Cal Gas Concentration: 5.02 ppm	Cal Gas Exp Date: November 15, 2026
Cal Gas Cylinder #: CC424047	
Removed Cal Gas Conc: 5.02 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: Teledyne API T700	Serial Number: 3806
ZAG Make/Model: Teledyne API T701	Serial Number: 4890

### Analyzer Information

Analyzer make: Thermo 43i-TLE	Analyzer serial #: 1151680031
Converter make: CDN-101	Converter serial #: 620
Analyzer Range: 0 - 100 ppb	Converter Temp: 850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995809	0.998806	Backgd or Offset:	3.57	3.52
Calibration intercept:	0.280545	0.100668	Coeff or Slope:	1.188	1.158

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4920	79.7	80.0	82.2	0.971
As found Mid point	4960	39.8	40.0	41.5	0.958
As found Low point	4980	19.9	20.0	20.6	0.961
New cylinder response					
Baseline Corr As found:	82.4	Prev response:	79.97	*% change:	3.0%
Baseline Corr 2nd AF pt:	41.7	AF Slope:	1.029509	AF Intercept:	0.001358
Baseline Corr 3rd AF pt:	20.8	AF Correlation:	0.999945	<i>* = +/-5% change initiates investigation</i>	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4920	79.7	80.0	79.9	1.002
Mid point	4960	39.8	40.0	40.2	0.994
Low point	4980	19.9	20.0	20.2	0.989
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	79.7	80.0	79.7	1.004
SO2 Scrubber Check	4920	79.8	798.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.995
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Scrubber test passed. Adjusted span only.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association


## TRS Calibration Summary

### Station Information

Calibration Date:	May 23, 2024	Previous Calibration:	April 18, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:50	End Time (MST):	14:55
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999968	$\geq 0.995$
80.0	79.9	1.0015	Slope	0.998806	$0.90 - 1.10$
40.0	40.2	0.9940	Intercept	0.100668	$\pm 3$
20.0	20.2	0.9891			



TRS Calibration Plot

Date: May 23, 2024

Location: Janvier





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	May 28, 2024	Last Cal Date:	April 19, 2024
Start time (MST):	10:36	End time (MST):	14:08
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH <sub>4</sub> Cal Gas Conc.	502.8 ppm	CH <sub>4</sub> Equiv Conc.	1075.9 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	502.8 ppm	CH <sub>4</sub> Equiv Conc.	1075.9 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
Zero Air Gen model:	Teledyne API 701	Serial Number:	4890

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1317958219
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	2.41E-04	2.46E-04	5.65E-05	5.74E-05
CH <sub>4</sub> Retention time:	11.6	11.6	161932	159535
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	17.17	16.81	1.022
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.81	Prev response	17.17	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	17.17	17.14	1.002
Mid point	4960	39.9	8.59	8.52	1.008
Low point	4980	20.0	4.30	4.26	1.011
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	17.17	17.05	1.007
Average Correction Factor					1.007

Notes: Changed the inlet filter and N<sub>2</sub>/H<sub>2</sub> cylinders after as founds. Adjusted span only.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.15	9.01	1.015
As found Mid point					
As found Low point				2.31	
New cylinder response					
Baseline Corr AF:	9.01	Prev response	9.15	*% change	-1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	2.31	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.15	9.13	1.002
Mid point	4960	39.9	4.57	4.56	1.004
Low point	4980	20.0	2.29	2.28	1.005
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.15	9.07	1.008
Average Correction Factor					1.004

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	8.03	7.80	1.030
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.80	Prev response	8.02	*% change	-2.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.03	8.01	1.002
Mid point	4960	39.9	4.01	3.96	1.012
Low point	4980	20.0	2.01	1.98	1.017
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.03	7.98	1.006
Average Correction Factor					1.010

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002111	0.998258
THC Cal Offset:	-0.041995	-0.024786
CH <sub>4</sub> Cal Slope:	1.002620	0.998675
CH <sub>4</sub> Cal Offset:	-0.028359	-0.019755
NMHC Cal Slope:	1.001665	0.998093
NMHC Cal Offset:	-0.013635	-0.004832

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 28, 2024	Previous Calibration:	April 19, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:36	End Time (MST):	14:08
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999988	<i>≥0.995</i>
17.17	17.14	1.0022	Slope	0.998258	<i>0.90 - 1.10</i>
8.59	8.52	1.0080	Intercept	-0.024786	<i>+/-0.5</i>
4.30	4.26	1.0109			







# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 28, 2024	Previous Calibration:	April 19, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:36	End Time (MST):	14:08
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999964	<i>≥0.995</i>
8.03	8.01	1.0020	Slope	0.998675	<i>0.90 - 1.10</i>
4.01	3.96	1.0122	Intercept	-0.019755	<i>+/-0.5</i>
2.01	1.98	1.0168			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 28, 2024	Previous Calibration:	April 19, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	10:36	End Time (MST):	14:08
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999999	<i>≥0.995</i>
9.15	9.13	1.0021	Slope	0.998093	<i>0.90 - 1.10</i>
4.57	4.56	1.0040	Intercept	-0.004832	<i>+/-0.5</i>
2.29	2.28	1.0054			



NMHC Calibration Plot

Date: May 28, 2024

Location: Janvier





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Janvier  
 Station number: AMS 22  
 Calibration Date: May 27, 2024  
 Last Cal Date: April 16, 2024  
 Start time (MST): 9:55  
 End time (MST): 14:34  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: DT0047765  
 NOX Cal Gas Conc: 48.90 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 48.90 ppm  
 NOX gas Diff:  
 Calibrator Model: Teledyne API T700  
 ZAG make/model: Teledyne API T701  
 Cal Gas Expiry Date: March 11, 2031  
 NO Cal Gas Conc: 48.80 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 48.80 ppm  
 NO gas Diff:  
 Serial Number: 3806  
 Serial Number: 691

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.5	-0.1	0.6	----	----
AF High point	4918	82.0	802.0	800.3	1.6	797.1	783.2	13.9	1.0067	1.0217
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 801.0 ppb		NO = 800.5 ppb			<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -0.5%	
Baseline Corr 1st pt	NO <sub>x</sub> = 796.6 ppb		NO = 783.3 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -2.2%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO r <sup>2</sup> :	NO SI:	NO Int:	
						As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Teledyne API T200  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 833

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.997901	0.995663
NO <sub>x</sub> Cal Offset:	0.704052	0.964017
NO Cal Slope:	0.999703	0.999503
NO Cal Offset:	0.464053	-0.016008
NO <sub>2</sub> Cal Slope:	0.997737	1.002632
NO <sub>2</sub> Cal Offset:	-0.534162	0.961073

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.846	0.854	NO bkgnd or offset:	-0.7	-0.7
NOX coeff or slope:	0.840	0.846	NOX bkgnd or offset:	0.5	0.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	6.7	7.1

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.4	0.4	0.1	----	----
High point	4918	82.0	802.0	800.3	1.6	799.0	800.0	-1.1	1.0037	1.0004
Mid point	4960	41.0	400.9	400.1	0.8	400.9	400.0	0.9	1.0000	1.0002
Low point	4980	20.5	200.5	200.1	0.4	200.8	199.3	1.5	0.9984	1.0038
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.3	-0.3	----	----
As left span	4918	82.0	802.0	393.7	408.3	794.9	393.7	401.2	1.0089	1.0000
Average Correction Factor									1.0007	1.0015

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	791.6	395.6	397.6	399.1	0.9963	100.4%
Mid GPT point	791.6	591.8	201.4	203.7	0.9889	101.1%
Low GPT point	791.6	694.9	98.3	100.2	0.9814	101.9%
Average Correction Factor					0.9889	101.1%

Notes: Changed the inlet filter after as founds. Adjusted span only. Used the 2nd GPT reference point.

Calibration Performed By: Rene Chamberland



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 27, 2024	Previous Calibration:	April 16, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:55	End Time (MST):	14:34
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.4	----	Correlation Coefficient	0.999997	≥0.995
802.0	799.0	1.0037	Slope	0.995663	0.90 - 1.10
400.9	400.9	1.0000	Intercept	0.964017	+/-20
200.5	200.8	0.9984			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 27, 2024	Previous Calibration:	April 16, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:55	End Time (MST):	14:34
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999977	≥0.995
397.6	399.1	0.9963	Slope	1.002632	0.90 - 1.10
201.4	203.7	0.9889	Intercept	0.961073	+/-20
98.3	100.2	0.9814			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 27, 2024	Previous Calibration:	April 16, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:55	End Time (MST):	14:34
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.4	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
800.3	800.0	1.0004	Slope	0.999503	<i>0.90 - 1.10</i>
400.1	400.0	1.0002	Intercept	-0.016008	<i>+/-20</i>
200.1	199.3	1.0038			






NO<sub>x</sub> Calibration Plot

Date: May 27, 2024

Location: Janvier





# Wood Buffalo Environmental Association

## O<sub>3</sub> Calibration Report

### Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	May 17, 2024	Last Cal Date:	April 15, 2024
Start time (MST):	9:49	End time (MST):	12:35
Reason:	Routine		

### Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3806
Calibrator Make/Model:	Teledyne API T700	Serial Number:	691
ZAG Make/Model:	Teledyne API T701H		

### Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #:	7046
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003571	1.011571	Backgd or Offset:	2.2	2.2
Calibration intercept:	0.300000	0.000000	Coeff or Slope:	1.027	1.027

### O<sub>3</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-1.3	----
As found High point	5000	916.2	400.0	404.5	0.986
As found Mid point					
As found Low point					
Baseline Corr As found:	405.8	Previous response	401.7	*% change	1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### O<sub>3</sub> Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.1	----
High point	5000	916.2	400.0	404.7	0.988
Mid point	5000	763.7	200.0	202.2	0.989
Low point	5000	656.1	100.0	101.1	0.989
As left zero	5000	800.0	0.0	0.5	----
As left span	5000	916.2	400.0	404.2	0.990
Average Correction Factor					<b>0.989</b>

Notes:                   Calibrating remotely. No adjustments made.

Calibration Performed By:                   Rene Chamberland



# Wood Buffalo Environmental Association


## O<sub>3</sub> Calibration Summary

### Station Information

Calibration Date:	May 17, 2024	Previous Calibration:	April 15, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:49	End Time (MST):	12:35
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	1.000000	<b>≥0.995</b>
400.0	404.7	0.9884	Slope	1.011571	<b>0.90 - 1.10</b>
200.0	202.2	0.9891	Intercept	0.000000	<b>+/- 5</b>
100.0	101.1	0.9891			



O<sub>3</sub> Calibration Plot

Date: May 17, 2024

Location: Janvier





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Janvier Station number: AMS 22  
 Calibration Date: May 27, 2024 Last Cal Date: April 18, 2024  
 Start time (MST): 11:48 End time (MST): 12:17

Analyzer Make: Teledyne API T640 S/N: 325  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752  
 Temp/RH standard: Alicat FP-25BT S/N: 388752

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	20.1	19.43	20.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	717.5	718.4	717.5	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.04	5.042	5.04	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	35	----	35	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.8	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024  
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 18, 2024  
 Date Disposable Filter Changed: April 18, 2024

Post- maintenance Zero Verification: PM w/ HEPA: \_\_\_\_\_ <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: July 26, 2023  
 Date RH/T Sensor Cleaned: April 18, 2024

Notes: Verified flow, temperature, and pressure. Leak check passed.

Calibration by: Rene Chamberland



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS23 FORT HILLS MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Fort Hills	Station number:	AMS 23
Calibration Date:	May 21, 2024	Last Cal Date:	April 11, 2024
Start time (MST):	8:50	End time (MST):	12:16
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.76	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC281425			
Removed Cal Gas Conc:	49.76	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	451
Zero Air Gen Model:	API T701		Serial Number:	5611

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1160290012
Analyzer Range:	0-1000ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009318	1.000622	Backgd or Offset:	18.6	18.4
Calibration intercept:	-1.305054	-1.263478	Coeff or Slope:	1.063	1.046

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.5	----
As found High point	4920	80.3	799.1	813.0	0.982
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	813.5	Previous response	805.2	*% change	1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.4	----
High point	4920	80.3	799.1	798.8	1.000
Mid point	4960	40.2	400.1	398.5	1.004
Low point	4980	20.1	200.0	198.1	1.010
As left zero	5000	0.0	0.0	-0.3	----
As left span	4920	80.3	799.1	800.9	0.998
Average Correction Factor:					1.005

Notes: Span adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 21, 2024	Previous Calibration:	April 11, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	8:50	End Time (MST):	12:16
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.4	----	Correlation Coefficient	0.999995	<b>≥0.995</b>
799.1	798.8	1.0004	Slope	1.000622	<b>0.90 - 1.10</b>
400.1	398.5	1.0039	Intercept	-1.263478	<b>+/-30</b>
200.0	198.1	1.0097			






SO2 Calibration Plot

Date: May 21, 2024

Location: Fort Hills





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name: Fort Hills	Station number: AMS 23
Calibration Date: May 1, 2024	Last Cal Date: April 12, 2024
Start time (MST): 7:05	End time (MST): 11:03
Reason: Routine	

### Calibration Standards

Cal Gas Concentration: 5.20 ppm	Cal Gas Exp Date: February 5, 2024
Cal Gas Cylinder #: CC517372	
Removed Cal Gas Conc: 5.20 ppm	Rem Gas Exp Date:
Removed Gas Cyl #:	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 451
ZAG Make/Model: API T701	Serial Number: 5611

### Analyzer Information

Analyzer make: Thermo 43i TLE	Analyzer serial #: 1300156232
Converter make: CDN-101	Converter serial #: 594
Analyzer Range: 0 - 100 ppb	Converter Temp: 750 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008213	1.009197	Backgd or Offset:	2.1	2.1
Calibration intercept:	0.074071	0.082105	Coeff or Slope:	1.190	1.190

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4923	77.0	80.1	79.2	1.010
As found Mid point	4962	38.5	40.0	39.5	1.011
As found Low point	4981	19.2	20.0	19.6	1.014
New cylinder response					
Baseline Corr As found:	79.3	Prev response:	80.81	*% change:	-1.9%
Baseline Corr 2nd AF pt:	39.6	AF Slope:	0.990506	AF Intercept:	-0.138309
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999999	<i>* = &gt; +/-5% change initiates investigation</i>	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4923	77.0	80.1	80.9	0.990
Mid point	4962	38.5	40.0	40.5	0.989
Low point	4981	19.2	20.0	20.2	0.988
As left zero	5000	0.0	0.0	0.0	----
As left span	4923	77.0	80.1	83.8	0.956
SO2 Scrubber Check	4920	80.3	803.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	<b>0.989</b>
Date of last converter efficiency test:	March 13, 2024			102.7%	efficiency

Notes: SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association


## TRS Calibration Summary

### Station Information

Calibration Date:	May 1, 2024	Previous Calibration:	April 12, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:05	End Time (MST):	11:03
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	1.000000	$\geq 0.995$
80.1	80.9	0.9899	Slope	1.009197	$0.90 - 1.10$
40.0	40.5	0.9885	Intercept	0.082105	$\pm 3$
20.0	20.2	0.9885			



TRS Calibration Plot

Date: May 1, 2024

Location: Fort Hills





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Fort Hills	Station number:	AMS 23
Calibration Date:	May 21, 2024	Last Cal Date:	April 11, 2024
Start time (MST):	8:50	End time (MST):	12:16
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	500.2 ppm	CH4 Equiv Conc.	1070.6 ppm
C3H8 Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.2 ppm	CH4 Equiv Conc.	1070.6 ppm
Removed C3H8 Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
Zero Air Gen model:	API T701	Serial Number:	5611

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1193585648
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.33E-04	2.35E-04	NMHC SP Ratio:	5.01E-05	5.11E-05
CH4 Retention time:	13.2	13.2	NMHC Peak Area:	182937	179091
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	17.19	16.93	1.015
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.93	Prev response	17.21	*% change	-1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	17.19	17.23	0.998
Mid point	4960	40.2	8.61	8.64	0.996
Low point	4980	20.1	4.30	4.33	0.994
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	17.19	17.21	0.999
Average Correction Factor					0.996

Notes: Span adjusted. Had trouble getting the NM zero back down to 0.00ppm on calibrator zero. Changed filter holder and filter.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	9.16	8.97	1.021
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.97	Prev response	9.18	*% change	-2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	9.16	9.19	0.997
Mid point	4960	40.2	4.59	4.65	0.987
Low point	4980	20.1	2.29	2.37	0.968
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	9.16	9.20	0.996
Average Correction Factor					0.984

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	8.03	7.96	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.96	Prev response	8.04	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	8.03	8.05	0.998
Mid point	4960	40.2	4.02	4.00	1.007
Low point	4980	20.1	2.01	1.96	1.026
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	8.03	8.02	1.002
Average Correction Factor					1.010

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000451	1.002212
THC Cal Offset:	0.012003	0.008001
CH <sub>4</sub> Cal Slope:	1.004637	1.003614
CH <sub>4</sub> Cal Offset:	-0.032039	-0.028442
NMHC Cal Slope:	0.997117	1.001457
NMHC Cal Offset:	0.043442	0.036042

Calibration Performed By:      Melissa Lemay



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 21, 2024	Previous Calibration:	April 11, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	8:50	End Time (MST):	12:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999999	<i>≥0.995</i>
17.19	17.23	0.9976	Slope	1.002212	<i>0.90 - 1.10</i>
8.61	8.64	0.9964	Intercept	0.008001	<i>+/-0.5</i>
4.30	4.33	0.9941			





# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 21, 2024	Previous Calibration:	April 11, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	8:50	End Time (MST):	12:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999943 <span style="color: red;">≥0.995</span>
8.03	8.05	0.9982	Slope	1.003614 <span style="color: red;">0.90 - 1.10</span>
4.02	4.00	1.0066	Intercept	-0.028442 <span style="color: red;">+/-0.5</span>
2.01	1.96	1.0259		







# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 21, 2024	Previous Calibration:	April 11, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	8:50	End Time (MST):	12:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999929	<i>≥0.995</i>
9.16	9.19	0.9965	Slope	1.001457	<i>0.90 - 1.10</i>
4.59	4.65	0.9872	Intercept	0.036042	<i>+/-0.5</i>
2.29	2.37	0.9678			



NMHC Calibration Plot

Date: May 21, 2024

Location: Fort Hills





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Fort Hills  
 Station number: AMS 23  
 Calibration Date: May 22, 2024  
 Last Cal Date: April 4, 2024  
 Start time (MST): 6:50  
 End time (MST): 11:08  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: CC358149  
 NOX Cal Gas Conc: 60.30 ppm  
 Removed Cylinder #:  
 Removed Gas NOX Conc: 60.30 ppm  
 NOX gas Diff:  
 Calibrator Model: API T700  
 ZAG make/model: API T701  
 Cal Gas Expiry Date: January 5, 1932  
 NO Cal Gas Conc: 60.10 ppm  
 Removed Gas Exp Date:  
 Removed Gas NO Conc: 60.10 ppm  
 NO gas Diff:  
 Serial Number: 451  
 Serial Number: 5611

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
AF High point	4934	66.3	799.5	796.9	2.7	796.3	789.3	7.1	1.0039	1.0096
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 800.0 ppb		NO = 796.5 ppb			<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -0.4%	
Baseline Corr 1st pt	NO <sub>x</sub> = 796.4 ppb		NO = 789.3 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -0.9%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO r <sup>2</sup> :	NO SI:	NO Int:	
						As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430007

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000499	0.999098
NO <sub>x</sub> Cal Offset:	0.045790	-0.213753
NO Cal Slope:	1.001459	1.003237
NO Cal Offset:	-1.552494	-1.772812
NO <sub>2</sub> Cal Slope:	1.003459	0.998852
NO <sub>2</sub> Cal Offset:	-1.071483	-0.832061

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.169	1.181	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.995	0.991	NOX bkgnd or offset:	3.4	3.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	177.7	177.7

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
High point	4934	66.3	799.5	796.9	2.7	798.6	798.6	0.0	1.0012	0.9978
Mid point	4967	33.2	400.4	399.0	1.3	400.0	397.7	2.3	1.0009	1.0034
Low point	4983	16.6	200.2	199.5	0.7	199.4	196.5	2.9	1.0041	1.0155
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.1	----	----
As left span	4934	66.3	799.5	390.9	408.6	801.4	390.9	410.5	0.9977	1.0000
Average Correction Factor									1.0021	1.0056

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	795.6	389.5	408.8	407.8	1.0023	99.8%
Mid GPT point	795.6	592.0	206.3	204.9	1.0066	99.3%
Low GPT point	795.6	692.5	105.8	104.0	1.0168	98.3%
Average Correction Factor					1.0086	99.2%

Notes: Span adjusted. No Maintenance done.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 22, 2024	Previous Calibration:	April 4, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:50	End Time (MST):	11:08
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999999	≥0.995
799.5	798.6	1.0012	Slope	0.999098	0.90 - 1.10
400.4	400.0	1.0009	Intercept	-0.213753	+/-20
200.2	199.4	1.0041			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 22, 2024	Previous Calibration:	April 4, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:50	End Time (MST):	11:08
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999985	≥0.995
408.8	407.8	1.0023	Slope	0.998852	0.90 - 1.10
206.3	204.9	1.0066	Intercept	-0.832061	+/-20
105.8	104.0	1.0168			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 22, 2024	Previous Calibration:	April 4, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:50	End Time (MST):	11:08
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999975	<span style="color: red;">≥0.995</span>
796.9	798.6	0.9978	Slope	1.003237	<span style="color: red;">0.90 - 1.10</span>
399.0	397.7	1.0034	Intercept	-1.772812	<span style="color: red;">+/-20</span>
199.5	196.5	1.0155			



NO<sub>x</sub> Calibration Plot

Date: May 22, 2024

Location: Fort Hills







# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Fort Hills Station number: AMS 23  
 Calibration Date: May 21, 2024 Last Cal Date: April 11, 2024  
 Start time (MST): 7:52 End time (MST): 8:47

Analyzer Make: API T640 S/N: 1546  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753  
 Temp/RH standard: Alicat FP-25BT S/N: 388753

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	7.9	8.32	7.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	734.1	733.1	734.1	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	4.67	5.02	<input checked="" type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	40	----	42	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	4.6	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: 10-Jun-24  
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10	11	11	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: May 21, 2024  
 Date Disposable Filter Changed: May 21, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: October 17, 2023  
 Date RH/T Sensor Cleaned: October 17, 2023

Notes: Flow adjusted after cleaning. Flow after cleaning was 4.78LPM. Leak check passed before and after cleaning. Sample Temp and Sample RH Warning.

Calibration by: Melissa Lemay



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS25 WASKŌW OHCI PIMÂTISIWIN MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number: AMS 25
Calibration Date:	May 29, 2024	Last Cal Date: April 22, 2024
Start time (MST):	7:44	End time (MST): 10:45
Reason:	Routine	

### Calibration Standards

Cal Gas Concentration:	49.70	ppm	Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #:	CC342445		
Removed Cal Gas Conc:	49.70	ppm	Rem Gas Exp Date:
Removed Gas Cyl #:			Diff between cyl:
Calibrator Model:	API T700		Serial Number: 747
Zero Air Gen Model:	API T701		Serial Number: 4765

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 1118148497
Analyzer Range:	0-1000ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998541	1.002056	Backgd or Offset:	10.8	10.8
Calibration intercept:	-0.052708	-0.332206	Coeff or Slope:	1.048	1.048

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4920	80.5	800.1	820.7	0.974
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	821.1	Previous response	798.9	*% change	2.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	80.5	800.1	801.6	0.998
Mid point	4960	40.2	399.6	400.0	0.999
Low point	4980	20.1	199.8	199.2	1.003
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	80.5	800.1	803.5	0.996
Average Correction Factor:					1.000

Notes: No maintenance done. Zero and span adjusted.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 29, 2024	Previous Calibration:	April 22, 2024
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS 25
Start Time (MST):	7:44	End Time (MST):	10:45
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999998	<b>≥0.995</b>
800.1	801.6	0.9981	Slope	1.002056	<b>0.90 - 1.10</b>
399.6	400.0	0.9989	Intercept	-0.332206	<b>+/-30</b>
199.8	199.2	1.0030			



SO2 Calibration Plot

Date: May 29, 2024

Location: Waskow ohci Pimatisiwin





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

### Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number:	AMS 25
Calibration Date:	May 28, 2024	Last Cal Date:	April 17, 2024
Start time (MST):	6:50	End time (MST):	12:02
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	4.97	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC517099			
Removed Cal Gas Conc:	4.97	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	747
ZAG Make/Model:	API T701		Serial Number:	261

### Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146
Converter make:	Global G-150	Converter serial #:	2022-219
Analyzer Range	0 - 100 ppb	Converter Temp:	350 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001868	1.004168	Backgd or Offset:	3.2	3.1
Calibration intercept:	0.120000	-0.060000	Coeff or Slope:	1.095	1.105

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4920	80.0	79.5	78.6	1.010
As found Mid point	4960	40.0	39.8	39.4	1.007
As found Low point	4980	20.0	19.9	19.7	1.004
New cylinder response					
Baseline Corr As found:	78.7	Prev response:	79.79	*% change:	-1.4%
Baseline Corr 2nd AF pt:	39.5	AF Slope:	0.989365	AF Intercept:	-0.020000
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999994	<i>* = &gt; +/-5% change initiates investigation</i>	

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.0	79.5	79.9	0.995
Mid point	4960	40.0	39.8	39.6	1.004
Low point	4980	20.0	19.9	20.0	0.994
As left zero	5000	0.0	0.0	0.3	----
As left span	4912	88.3	800.0	803.4	0.996
SO <sub>2</sub> Scrubber Check	4921	79.2	800.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	<b>0.998</b>
Date of last converter efficiency test:					

Notes: SO<sub>x</sub> scrubber checked after the calibrator zero. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



# Wood Buffalo Environmental Association


## H<sub>2</sub>S Calibration Summary

### Station Information

Calibration Date:	May 28, 2024	Previous Calibration:	April 17, 2024
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS 25
Start Time (MST):	6:50	End Time (MST):	12:02
Analyzer make:	Global G-150	Analyzer serial #:	2022-219

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999973	≥0.995
79.5	79.9	0.9952	Slope	1.004168	0.90 - 1.10
39.8	39.6	1.0040	Intercept	-0.060000	+/-3
19.9	20.0	0.9940			



H<sub>2</sub>S Calibration Plot

Date: May 28, 2024

Location: Waskow ohci Pimatisiwin







## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS26 CHRISTINA LAKE MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Christina Lake	Station number:	AMS 26
Calibration Date:	May 15, 2024	Last Cal Date:	April 9, 2024
Start time (MST):	11:06	End time (MST):	1:18
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.56	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC362134			
Removed Cal Gas Conc:	49.56	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	281
Zero Air Gen Model:	API T701H		Serial Number:	832

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1152430005
Analyzer Range:	0- 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999449	1.012862	Backgd or Offset:	26.5	26.5
Calibration intercept:	0.556005	-1.543821	Coeff or Slope:	0.944	0.944

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4919	80.8	800.9	810.3	0.988
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	810.3	Previous response	801.0	*% change	1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4919	80.8	800.9	810.3	0.988
Mid point	4960	40.4	400.4	403.7	0.992
Low point	4980	20.2	200.2	199.4	1.004
As left zero	5000	0.0	0.0	0.3	----
As left span	4919	80.8	800.9	810.4	0.988
Average Correction Factor:					0.995

Notes: Remote calibrations. No adjustment made.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 9, 2024
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	11:06	End Time (MST):	1:18
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430005

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999983
800.9	810.3	0.9884	Slope	1.012862
400.4	403.7	0.9919	Intercept	-1.543821
200.2	199.4	1.0041		
				≥0.995
				0.90 - 1.10
				+/-30



SO2 Calibration Plot

Date: May 15, 2024

Location: Christina Lake





# Wood Buffalo Environmental Association

## H2S Calibration Report

### Station Information

Station Name: Christina Lake	Station number: AMS 26
Calibration Date: May 23, 2024	Last Cal Date: April 11, 2024
Start time (MST): 9:40	End time (MST): 13:29
Reason: Routine	

### Calibration Standards

Cal Gas Concentration: 5.05 ppm	Cal Gas Exp Date: November 15, 2026
Cal Gas Cylinder #: DT0014831	
Removed Cal Gas Conc: 5.05 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 5258
ZAG Make/Model: API T701H	Serial Number: 832

### Analyzer Information

Analyzer make: Thermo 43iQTL	Analyzer serial #: 12333331547
Converter make: Global 150	Converter serial #: 2022-196
Analyzer Range: 0 - 100 ppb	Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004141	1.009856	Backgd or Offset:	1.5	1.3
Calibration intercept:	-0.141605	-0.241611	Coeff or Slope:	1.030	1.037

### H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4921	79.2	80.0	79.0	1.007
As found Mid point	4960	39.6	40.0	38.9	1.018
As found Low point	4980	19.8	20.0	19.0	1.031
New cylinder response					
Baseline Corr As found:	79.4	Prev response:	80.18	*% change:	-1.0%
Baseline Corr 2nd AF pt:	39.3	AF Slope:	0.993996	AF Intercept:	-0.661577
Baseline Corr 3rd AF pt:	19.4	AF Correlation:	0.999949	<i>* = &gt; +/-5% change initiates investigation</i>	

### H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4921	79.2	80.0	80.6	0.992
Mid point	4960	39.6	40.0	40.1	0.997
Low point	4980	19.8	20.0	19.8	1.010
As left zero	5000	0.0	0.0	-0.1	----
As left span	4921	79.2	80.0	81.0	0.988
SO2 Scrubber Check	4919	80.8	808.0	-0.1	----
Date of last scrubber change:	11-Apr-24			Ave Corr Factor	1.000
Date of last converter efficiency test:					

Notes: Changed sample inlet filters after multipoint as founds. Scrubber check done after calibrator zero and passed. Adjusted zero and span.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association


## H2S Calibration Summary

### Station Information

Calibration Date:	May 23, 2024	Previous Calibration:	April 11, 2024
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	9:40	End Time (MST):	13:29
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331547

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999986	$\geq 0.995$
80.0	80.6	0.9924	Slope	1.009856	$0.90 - 1.10$
40.0	40.1	0.9975	Intercept	-0.241611	$\pm 3$
20.0	19.8	1.0100			



H2S Calibration Plot

Date: May 23, 2024

Location: Christina Lake





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Christina Lake  
 Station number: AMS 26  
 Calibration Date: May 14, 2024  
 Last Cal Date: April 23, 2024  
 Start time (MST): 9:28  
 End time (MST): 10:25  
 Reason: As Found

### Calibration Standards

NO Gas Cylinder #: CC755290  
 NOX Cal Gas Conc: 48.90 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 48.90 ppm  
 NOX gas Diff:  
 Calibrator Model: API T700  
 ZAG make/model: API T701H  
 Cal Gas Expiry Date: January 3, 2031  
 NO Cal Gas Conc: 48.70 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 48.70 ppm  
 NO gas Diff:  
 Serial Number: 3253  
 Serial Number: 832

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	-0.1	----	----
AF High point	4918	82.1	802.9	799.6	3.3	819.9	812.6	7.1	0.9791	0.9838
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 801.1 ppb		NO = 796.1 ppb			<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = 2.3%	
Baseline Corr 1st pt	NO <sub>x</sub> = 820.1 ppb		NO = 812.8 ppb			<u>As Found Statistics</u>		*Percent Change	NO = 2.1%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO r <sup>2</sup> :	NO SI:	NO Int:	
						As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1173480006

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000009	
NO <sub>x</sub> Cal Offset:	-1.813460	
NO Cal Slope:	0.998685	
NO Cal Offset:	-2.513223	
NO <sub>2</sub> Cal Slope:	1.022718	
NO <sub>2</sub> Cal Offset:	0.962157	

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.329	NA	NO bkgnd or offset:	2.6	NA
NOX coeff or slope:	0.994	NA	NOX bkgnd or offset:	2.7	NA
NO2 coeff or slope:	1.000	NA	Reaction cell Press:	162.5	NA

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero  
 High point  
 Mid point  
 Low point  
 As left zero  
 As left span

Average Correction Factor

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero  
 High GPT point  
 Mid GPT point  
 Low GPT point

Average Correction Factor


Notes: Sample inlet filters changed after as founds. Calibrations was aborted due to wildfire concerns.

Calibration Performed By: Jan Castro

NO<sub>x</sub> Calibration Plot

Date: May 14, 2024

Location: Christina Lake





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Christina Lake  
 Station number: AMS 26  
 Calibration Date: May 15, 2024  
 Last Cal Date: April 23, 2024  
 Start time (MST): 6:37  
 End time (MST): 10:44  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: CC755290  
 NOX Cal Gas Conc: 48.90 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 48.90 ppm  
 NOX gas Diff:  
 Calibrator Model: API T700  
 ZAG make/model: API T701H  
 Cal Gas Expiry Date: January 3, 2031  
 NO Cal Gas Conc: 48.70 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 48.70 ppm  
 NO gas Diff:  
 Serial Number: 3253  
 Serial Number: 832

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = NA	ppb	NO = NA	ppb					*Percent Change	NO <sub>x</sub> = NA
Baseline Corr 1st pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb					*Percent Change	NO = NA
Baseline Corr 2nd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb						
Baseline Corr 3rd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb						
						<b>* = &gt; +/-5% change initiates investigation</b>				
						<u>As Found Statistics</u>				
					As found	NO <sub>x</sub> r <sup>2</sup> :		Nx SI:		Nx Int:
					As found	NO r <sup>2</sup> :		NO SI:		NO Int:
					As found	NO <sub>2</sub> r <sup>2</sup> :		NO <sub>2</sub> SI:		NO <sub>2</sub> Int:

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1173480006

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	NA	1.005033
NO <sub>x</sub> Cal Offset:	NA	-0.853732
NO Cal Slope:	NA	1.003844
NO Cal Offset:	NA	-1.793642
NO <sub>2</sub> Cal Slope:	NA	1.000810
NO <sub>2</sub> Cal Offset:	NA	-1.663798

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	NA	1.329	NO bkgnd or offset:	NA	2.6
NOX coeff or slope:	NA	0.994	NOX bkgnd or offset:	NA	2.7
NO2 coeff or slope:	NA	1.000	Reaction cell Press:	NA	162.5

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
High point	4918	82.1	802.9	799.6	3.3	806.3	801.7	4.6	0.9958	0.9974
Mid point	4959	41.1	401.9	400.3	1.6	403.1	399.3	3.8	0.9971	1.0025
Low point	4980	20.5	200.5	199.7	0.8	199.9	197.0	2.9	1.0029	1.0135
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
As left span	4918	82.1	802.9	385.1	417.8	802.7	385.1	417.4	1.0003	1.0000
Average Correction Factor									0.9986	1.0045

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	797.0	377.3	423.0	422.4	1.0014	99.9%
Mid GPT point	797.0	587.2	213.1	211.0	1.0099	99.0%
Low GPT point	797.0	691.7	108.6	105.3	1.0312	97.0%
Average Correction Factor					1.0141	98.6%

Notes: Remote calibrations. No adjustment made.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 23, 2024
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	6:37	End Time (MST):	10:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.3	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
802.9	806.3	0.9958	Slope	1.005033	<i>0.90 - 1.10</i>
401.9	403.1	0.9971	Intercept	-0.853732	<i>+/-20</i>
200.5	199.9	1.0029			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 23, 2024
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	6:37	End Time (MST):	10:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999936	≥0.995
423.0	422.4	1.0014	Slope	1.000810	0.90 - 1.10
213.1	211.0	1.0099	Intercept	-1.663798	+/-20
108.6	105.3	1.0312			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 23, 2024
Station Name:	Christina Lake	Station Number:	AMS 26
Start Time (MST):	6:37	End Time (MST):	10:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999982	≥0.995
799.6	801.7	0.9974	Slope	1.003844	0.90 - 1.10
400.3	399.3	1.0025	Intercept	-1.793642	+/-20
199.7	197.0	1.0135			



NO<sub>x</sub> Calibration Plot

Date: May 15, 2024

Location: Christina Lake







## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS27 JACKFISH 2/3 MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	May 13, 2024	Last Cal Date:	April 15, 2024
Start time (MST):	11:47	End time (MST):	14:46
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	SG9133974BAL			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	3811
Zero Air Gen Model:	API 701		Serial Number:	268

### Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Serial Number:	12124313138
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999989	1.016707	Backgd or Offset:	8.4	8.5
Calibration intercept:	-1.697880	-2.198983	Coeff or Slope:	0.955	0.955

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4921	79.1	800.2	812.7	0.985
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	812.7	Previous response	798.5	*% change	1.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4921	79.1	800.2	812.8	0.984
Mid point	4961	39.5	399.5	402.0	0.994
Low point	4980	19.8	200.3	199.6	1.004
As left zero	5000	0.0	0.0	0.0	----
As left span	4921	79.1	800.2	816.4	0.980
Average Correction Factor:					0.994

Notes: Changed the sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 13, 2024	Previous Calibration:	April 15, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:47	End Time (MST):	14:46
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	12124313138

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999957	<b>≥0.995</b>
800.2	812.8	0.9844	Slope	1.016707	<b>0.90 - 1.10</b>
399.5	402.0	0.9939	Intercept	-2.198983	<b>+/-30</b>
200.3	199.6	1.0035			



SO2 Calibration Plot

Date: May 13, 2024

Location: Jackfish 2/3





# Wood Buffalo Environmental Association

## H<sub>2</sub>S Calibration Report

### Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	May 14, 2024	Last Cal Date:	April 17, 2024
Start time (MST):	8:33	End time (MST):	18:55
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.41	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC345023			
Removed Cal Gas Conc:	5.41	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3811
ZAG Make/Model:	API 701		Serial Number:	268

### Analyzer Information

Analyzer make:	API T101	Analyzer serial #:	621
Converter make:	NA	Converter serial #:	NA
Analyzer Range	0 - 100 ppb	Converter Temp:	316 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.990223	0.981525	Backgd or Offset:	29.9	29.9
Calibration intercept:	-0.338155	0.241749	Coeff or Slope:	0.965	0.965

### H<sub>2</sub>S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4926	74.1	80.2	83.0	0.968
As found Mid point	4963	37.0	40.0	41.4	0.972
As found Low point	4982	18.5	20.0	20.2	1.001
New cylinder response					
Baseline Corr As found:	82.8	Prev response:	79.05	*% change:	4.5%
Baseline Corr 2nd AF pt:	41.2	AF Slope:	1.035409	AF Intercept:	-0.097195
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999926	<i>* =&gt; +/-5% change initiates investigation</i>	

### H<sub>2</sub>S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4926	74.1	80.2	78.9	1.016
Mid point	4963	37.0	40.0	39.6	1.011
Low point	4982	18.5	20.0	19.9	1.006
As left zero	5000	0.0	0.0	0.0	----
As left span	4926	74.1	80.2	79.6	1.007
SO2 Scrubber Check	4921	79.1	791.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	1.011
Date of last converter efficiency test:					

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Mohammed Kashif



# Wood Buffalo Environmental Association


## H<sub>2</sub>S Calibration Summary

### Station Information

Calibration Date:	May 14, 2024	Previous Calibration:	April 17, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	8:33	End Time (MST):	18:55
Analyzer make:	API T101	Analyzer serial #:	621

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999998	≥0.995
80.2	78.9	1.0162	Slope	0.981525	0.90 - 1.10
40.0	39.6	1.0110	Intercept	0.241749	+/-3
20.0	19.9	1.0058			



H<sub>2</sub>S Calibration Plot

Date: May 14, 2024

Location: Jackfish 2/3





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Jackfish 2/3  
 Station number: AMS 27  
 Calibration Date: May 15, 2024  
 Last Cal Date: April 18, 2024  
 Start time (MST): 9:07  
 End time (MST): 14:42  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: CC757838  
 NOX Cal Gas Conc: 60.30 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 60.30 ppm  
 NOX gas Diff:  
 Calibrator Model: API T700  
 ZAG make/model: API T701  
 Cal Gas Expiry Date: January 9, 2023  
 NO Cal Gas Conc: 60.20 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 60.20 ppm  
 NO gas Diff:  
 Serial Number: 3811  
 Serial Number: 268

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	1.0	-0.5	1.4	----	----
AF High point	4942	66.5	800.6	799.3	1.3	797.0	783.8	13.1	1.0058	1.0191
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 800.5 ppb		NO = 793.1 ppb			<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -0.6%	
Baseline Corr 1st pt	NO <sub>x</sub> = 796.0 ppb		NO = 784.3 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -1.1%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO r <sup>2</sup> :	NO SI:	NO Int:	
						As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: API T200  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 722

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.004943	1.003401
NO <sub>x</sub> Cal Offset:	-4.114844	-3.974590
NO Cal Slope:	0.996905	0.996415
NO Cal Offset:	-3.695611	-4.574196
NO <sub>2</sub> Cal Slope:	0.980873	1.000868
NO <sub>2</sub> Cal Offset:	1.007817	1.963831

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.271	1.271	NO bkgnd or offset:	0.3	0.3
NOX coeff or slope:	1.260	1.260	NOX bkgnd or offset:	1.2	1.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	3.2	3.2

### Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	1.2	0.3	0.9	----	----
High point	4942	66.5	800.6	799.3	1.3	802.2	795.3	6.9	0.9980	1.0050
Mid point	4979	33.3	400.6	399.9	0.7	394.7	388.5	6.2	1.0150	1.0295
Low point	4996	16.6	199.7	199.4	0.3	191.7	191.2	0.5	1.0417	1.0427
As left zero	5000	0.0	0.0	0.0	0.0	1.5	1.5	0.0	----	----
As left span	4942	66.5	800.6	405.5	395.1	791.9	405.5	386.4	1.0110	1.0000
Average Correction Factor									1.0182	1.0257

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.9	----	----
High GPT point	789.4	405.4	385.3	387.3	0.9949	100.5%
Mid GPT point	789.4	616.3	174.4	176.6	0.9877	101.2%
Low GPT point	789.4	703.9	86.8	90.2	0.9626	103.9%
Average Correction Factor					0.9817	101.9%

Notes:

No adjustments made.

Calibration Performed By:

Mohammed Kashif



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 18, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	9:07	End Time (MST):	14:42
Analyzer make:	API T200	Analyzer serial #:	722

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	1.2	----	Correlation Coefficient	0.999808	≥0.995
800.6	802.2	0.9980	Slope	1.003401	0.90 - 1.10
400.6	394.7	1.0150	Intercept	-3.974590	+/-20
199.7	191.7	1.0417			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 18, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	9:07	End Time (MST):	14:42
Analyzer make:	API T200	Analyzer serial #:	722

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.9	----	Correlation Coefficient	0.999963	≥0.995
385.3	387.3	0.9949	Slope	1.000868	0.90 - 1.10
174.4	176.6	0.9877	Intercept	1.963831	+/-20
86.8	90.2	0.9626			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 15, 2024	Previous Calibration:	April 18, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	9:07	End Time (MST):	14:42
Analyzer make:	API T200	Analyzer serial #:	722

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999788	<i>≥0.995</i>
799.3	795.3	1.0050	Slope	0.996415	<i>0.90 - 1.10</i>
399.9	388.5	1.0295	Intercept	-4.574196	<i>+/-20</i>
199.4	191.2	1.0427			



NO<sub>x</sub> Calibration Plot

Date: May 15, 2024

Location: Jackfish 2/3





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS29 SURMONT 2 MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Surmont 2	Station number:	AMS 29
Calibration Date:	May 13, 2024	Last Cal Date:	April 12, 2024
Start time (MST):	10:51	End time (MST):	14:32
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC356008			
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	5472
Zero Air Gen Model:	Teledyne API T701		Serial Number:	4698

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1170050150
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006014	0.997045	Backgd or Offset:	13.0	12.8
Calibration intercept:	-1.785690	-1.145686	Coeff or Slope:	0.939	0.928

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4919	81.3	800.1	804.0	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	804.1	Previous response	803.1	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4919	81.3	800.1	797.1	1.004
Mid point	4959	40.7	400.6	398.0	1.007
Low point	4979	20.3	199.8	196.6	1.016
As left zero	5000	0.0	0.0	0.2	----
As left span	4919	81.3	800.1	797.5	1.003
Average Correction Factor:					1.009

Notes: Changed filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 13, 2024	Previous Calibration:	April 12, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:51	End Time (MST):	14:32
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999988	<span style="color: red;">≥0.995</span>
800.1	797.1	1.0038	Slope	0.997045	<span style="color: red;">0.90 - 1.10</span>
400.6	398.0	1.0065	Intercept	-1.145686	<span style="color: red;">+/-30</span>
199.8	196.6	1.0164			






SO2 Calibration Plot

Date: May 13, 2024

Location: Surmont 2





# Wood Buffalo Environmental Association

## H2S Calibration Report

### Station Information

Station Name:	Surmont 2	Station number:	AMS 29
Calibration Date:	May 9, 2024	Last Cal Date:	April 15, 2024
Start time (MST):	10:07	End time (MST):	14:51
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	<u>5.391</u>	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	<u>CC508338</u>			
Removed Cal Gas Conc:	<u>5.391</u>	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>CC508338</u>		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	5472
ZAG Make/Model:	Teledyne API T701		Serial Number:	4698

### Analyzer Information

Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170
Converter make:	Global	Converter serial #:	2022-220
Analyzer Range:	0 - 100 ppb	Converter Temp:	325.0 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998614	0.998331	Backgd or Offset:	0.90	0.90
Calibration intercept:	-0.142639	-0.082706	Coeff or Slope:	1.050	1.056

### H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4925.8	74.2	80.0	80.7	0.990
As found Mid point	4962.9	37.2	40.1	40.2	0.995
As found Low point	4981.5	18.6	20.1	19.8	1.008
New cylinder response					
Baseline Corr As found:	80.8	Prev response:	79.75	*% change:	1.3%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.011184	AF Intercept:	-0.283103
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999976	<i>* = +/-5% change initiates investigation</i>	

### H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4925.8	74.2	80.0	79.8	1.003
Mid point	4962.9	37.2	40.1	40.0	1.003
Low point	4981.5	18.6	20.1	19.8	1.013
As left zero	5000	0.0	0.0	0.0	----
As left span	4925.8	74.2	80.0	79.2	1.010
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.006
Date of last converter efficiency test:					

Notes: Adjusted span. Changed sample inlet filter after as founds.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association


## H2S Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	April 15, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:07	End Time (MST):	14:51
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999992	$\geq 0.995$
80.0	79.8	1.0025	Slope	0.998331	$0.90 - 1.10$
40.1	40.0	1.0027	Intercept	-0.082706	$\pm 3$
20.1	19.8	1.0128			



H2S Calibration Plot

Date: May 9, 2024

Location: Surmont 2







# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 13, 2024	Previous Calibration:	April 12, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:51	End Time (MST):	14:32
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

### Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.03	----	Correlation Coefficient	0.999941	<span style="color: red;">≥0.995</span>
17.31	17.43	0.9933	Slope	1.005876	<span style="color: red;">0.90 - 1.10</span>
8.65	8.61	1.0042	Intercept	-0.025252	<span style="color: red;">+/-1.5</span>
4.32	4.29	1.0082			



THC Calibration Plot

Date: May 13, 2024

Location: Surmont 2





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Surmont 2  
 Station number: AMS 29  
 Calibration Date: May 8, 2024  
 Last Cal Date: April 2, 2024  
 Start time (MST): 9:41  
 End time (MST): 14:13  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T12YYFE  
 NOX Cal Gas Conc: 47.46 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 47.46 ppm  
 NOX gas Diff:  
 Calibrator Model: Teledyne API T700  
 ZAG make/model: Teledyne API T701  
 Cal Gas Expiry Date: October 30, 2024  
 NO Cal Gas Conc: 47.46 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 47.46 ppm  
 NO gas Diff:  
 Serial Number: 5472  
 Serial Number: 4698

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0	0.0	0.0	0.0	-0.3	-0.3	0.0	----	----
AF High point	4916	84.3	800.1	800.1	0.0	776.5	773.2	3.3	1.0300	1.0344
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 800.2 ppb		NO = 798.4 ppb			<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -3.0%	
Baseline Corr 1st pt	NO <sub>x</sub> = 776.8 ppb		NO = 773.5 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -3.2%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO r <sup>2</sup> :	NO SI:	NO Int:	
						As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1170050148

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.001053	1.001124
NO <sub>x</sub> Cal Offset:	-0.812441	-1.212303
NO Cal Slope:	1.000036	1.002811
NO Cal Offset:	-1.731680	-1.852292
NO <sub>2</sub> Cal Slope:	1.002282	1.000762
NO <sub>2</sub> Cal Offset:	0.146547	-0.703548

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.429	1.474	NO bkgnd or offset:	1.4	1.4
NOX coeff or slope:	0.996	0.993	NOX bkgnd or offset:	1.4	1.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	173.8	177.4

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
High point	4916	84.2	799.2	799.2	0.0	799.0	800.0	-1.2	1.0002	0.9990
Mid point	4958	42.1	399.6	399.6	0.0	399.6	399.4	0.2	1.0000	1.0005
Low point	4979	21.1	200.3	200.3	0.0	197.3	196.3	1.0	1.0151	1.0203
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
As left span	4916	84.2	799.2	799.2	0.0	796.5	400.7	395.8	1.0034	1.9945
Average Correction Factor									1.0051	1.0066

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	794.5	403.0	391.5	391.3	1.0005	99.9%
Mid GPT point	794.5	605.8	188.7	188.2	1.0027	99.7%
Low GPT point	794.5	699.9	94.6	93.0	1.0172	98.3%
Average Correction Factor					1.0068	99.3%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 2, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:41	End Time (MST):	14:13
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999983	≥0.995
799.2	799.0	1.0002	Slope	1.001124	0.90 - 1.10
399.6	399.6	1.0000	Intercept	-1.212303	+/-20
200.3	197.3	1.0151			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 2, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:41	End Time (MST):	14:13
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999982	≥0.995
391.5	391.3	1.0005	Slope	1.000762	0.90 - 1.10
188.7	188.2	1.0027	Intercept	-0.703548	+/-20
94.6	93.0	1.0172			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 8, 2024	Previous Calibration:	April 2, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:41	End Time (MST):	14:13
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999969	<span style="color: red;">≥0.995</span>
799.2	800.0	0.9990	Slope	1.002811	<span style="color: red;">0.90 - 1.10</span>
399.6	399.4	1.0005	Intercept	-1.852292	<span style="color: red;">+/-20</span>
200.3	196.3	1.0203			



NO<sub>x</sub> Calibration Plot

Date: May 8, 2024

Location: Surmont 2





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Surmont 2 Station number: AMS 29  
 Calibration Date: May 9, 2024 Last Cal Date: April 15, 2024  
 Start time (MST): 12:49 End time (MST): 13:16

Analyzer Make: API T640 S/N: 253  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754  
 Temp/RH standard: Alicat FP-25BT S/N: 388754

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	17.7	18.51	17.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	714.4	715.67	714.4	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.111	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	34	----	34	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	9.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024  
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: \_\_\_\_\_ April 15, 2024  
 Date Disposable Filter Changed: \_\_\_\_\_ April 15, 2024

Post- maintenance Zero Verification: PM w/ HEPA: \_\_\_\_\_ <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: \_\_\_\_\_ October 25, 2023  
 Date RH/T Sensor Cleaned: \_\_\_\_\_ October 25, 2023

Notes: Verified temperature, pressure and flow. No adjustments made.

Calibration by: Braiden Boutilier



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS30 ELLS RIVER MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	May 16, 2024	Last Cal Date:	April 4, 2024
Start time (MST):	9:43	End time (MST):	13:19
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	50.53	ppm	Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #:	CC494126		
Removed Cal Gas Conc:	50.53	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	API T700		Serial Number: 3061
Zero Air Gen Model:	API T701H		Serial Number: 358

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1008841397
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997233	1.002830	Backgd or Offset:	9.5	9.5
Calibration intercept:	-2.355844	-2.515950	Coeff or Slope:	0.982	0.992

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4921	79.2	800.4	792.8	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	793.0	Previous response	795.8	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	79.2	800.4	801.4	0.999
Mid point	4960	39.6	400.2	397.5	1.007
Low point	4980	19.8	200.1	195.7	1.023
As left zero	5000	0.0	0.0	0.0	----
As left span	4921	79.2	800.4	803.1	0.997
Average Correction Factor:					1.009

Notes: Sample inlet filter changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro





# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 16, 2024	Previous Calibration:	April 4, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:43	End Time (MST):	13:19
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999955
800.4	801.4	0.9987	Slope	1.002830
400.2	397.5	1.0069	Intercept	-2.515950
200.1	195.7	1.0225		
				<span style="color: red;">≥0.995</span> <span style="color: red;">0.90 - 1.10</span> <span style="color: red;">+/-30</span>



SO2 Calibration Plot

Date: May 16, 2024

Location: Ells River





# Wood Buffalo Environmental Association

## TRS Calibration Report

### Station Information

Station Name: Ells River	Station number: AMS 30
Calibration Date: May 9, 2024	Last Cal Date: April 2, 2024
Start time (MST): 9:03	End time (MST): 13:10
Reason: Routine	

### Calibration Standards

Cal Gas Concentration: 4.99 ppm	Cal Gas Exp Date: November 15, 2026
Cal Gas Cylinder #: CC505806	
Removed Cal Gas Conc: 4.99 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 3061
ZAG Make/Model: API T701H	Serial Number: 358

### Analyzer Information

Analyzer make: Thermo 43i TLE	Analyzer serial #: 1410661331
Converter make: CDN- 101	Converter serial #: 562
Analyzer Range: 0 - 100 ppb	Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005758	1.009186	Backgd or Offset:	1.6	1.7
Calibration intercept:	-0.080404	-0.200410	Coeff or Slope:	1.060	1.082

### TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4920	80.2	80.0	78.9	1.013
As found Mid point	4960	40.1	40.0	39.2	1.018
As found Low point	4980	20.0	20.0	19.4	1.024
New cylinder response					
Baseline Corr As found:	79.0	Prev response:	80.42	*% change:	-1.8%
Baseline Corr 2nd AF pt:	39.3	AF Slope:	0.987629	AF Intercept:	-0.220806
Baseline Corr 3rd AF pt:	19.5	AF Correlation:	0.999989	<i>* = &gt; +/-5% change initiates investigation</i>	

### TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.2	80.0	80.7	0.992
Mid point	4960	40.1	40.0	40.0	1.000
Low point	4980	20.0	20.0	19.8	1.008
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	80.2	80.0	80.0	1.000
SO2 Scrubber Check	4921	79.2	792.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.000
Date of last converter efficiency test:					

Notes: Change sample inlet filters and pump after multipoint as founds. Sox scrubber check done after calibrator zero and passed. Adjusted span only.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association


## TRS Calibration Summary

### Station Information

Calibration Date:	May 9, 2024	Previous Calibration:	April 2, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:03	End Time (MST):	13:10
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999969	$\geq 0.995$
80.0	80.7	0.9918	Slope	1.009186	$0.90 - 1.10$
40.0	40.0	1.0005	Intercept	-0.200410	$\pm 3$
20.0	19.8	1.0081			



TRS Calibration Plot

Date: May 9, 2024

Location: Ells River





# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	May 16, 2024	Last Cal Date:	April 4, 2024
Start time (MST):	9:43	End time (MST):	13:19
Reason:	Routine		

### Calibration Standards

Gas Cert Reference:	CC494126	Cal Gas Expiry Date:	December 29, 2028
CH <sub>4</sub> Cal Gas Conc.	499.7 ppm	CH <sub>4</sub> Equiv Conc.	1075.0 ppm
C <sub>3</sub> H <sub>8</sub> Cal Gas Conc.	209.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH <sub>4</sub> Conc.	499.7 ppm	CH <sub>4</sub> Equiv Conc.	1075.0 ppm
Removed C <sub>3</sub> H <sub>8</sub> Conc.	209.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH <sub>4</sub> ):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
Zero Air Gen model:	API T701H	Serial Number:	358

### Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1152430011
THC Range: 0 - 20 ppm	NMHC/CH <sub>4</sub> Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH <sub>4</sub> SP Ratio:	3.07E-04	3.12E-04	NMHC SP Ratio:	6.43E-05	6.31E-05
CH <sub>4</sub> Retention time:	17.4	17.4	NMHC Peak Area:	141658	144334
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	---
As found High point	4921	79.2	17.03	17.05	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.05	Prev response	17.12	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	---
High point	4921	79.2	17.03	17.03	1.000
Mid point	4960	39.6	8.51	8.44	1.008
Low point	4980	19.8	4.26	4.19	1.016
As left zero	5000	0.0	0.00	0.00	---
As left span	4921	79.2	17.03	17.03	1.000
Average Correction Factor					1.008

Notes: Sample inlet filters, H<sub>2</sub> cylinder, and N<sub>2</sub> cylinder was changed after as founds. Adjusted span only.



# Wood Buffalo Environmental Association

## THC / CH<sub>4</sub> / NMHC Calibration Report

### NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.2	9.11	9.25	0.985
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.25	Prev response	9.28	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.2	9.11	9.11	1.000
Mid point	4960	39.6	4.56	4.53	1.006
Low point	4980	19.8	2.28	2.26	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.2	9.11	9.11	1.000
Average Correction Factor					1.005

### CH<sub>4</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.2	7.91	7.79	1.016
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.79	Prev response	7.85	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### CH<sub>4</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.2	7.91	7.92	0.999
Mid point	4960	39.6	3.96	3.92	1.011
Low point	4980	19.8	1.98	1.93	1.024
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.2	7.91	7.91	1.000
Average Correction Factor					1.011

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.008470	1.001295
THC Cal Offset:	-0.047141	-0.042939
CH <sub>4</sub> Cal Slope:	0.994402	1.002344
CH <sub>4</sub> Cal Offset:	-0.025556	-0.028557
NMHC Cal Slope:	1.020664	1.000672
NMHC Cal Offset:	-0.021985	-0.014781

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 16, 2024	Previous Calibration:	April 4, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:43	End Time (MST):	13:19
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999968	<i>≥0.995</i>
17.03	17.03	0.9996	Slope	1.001295	<i>0.90 - 1.10</i>
8.51	8.44	1.0085	Intercept	-0.042939	<i>+/-0.5</i>
4.26	4.19	1.0163			







# Wood Buffalo Environmental Association


## CH<sub>4</sub> Calibration Summary

### Station Information

Calibration Date:	May 16, 2024	Previous Calibration:	April 4, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:43	End Time (MST):	13:19
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

### Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999938	<i>≥0.995</i>
7.91	7.92	0.9991	Slope	1.002344	<i>0.90 - 1.10</i>
3.96	3.92	1.0107	Intercept	-0.028557	<i>+/-0.5</i>
1.98	1.93	1.0243			





# Wood Buffalo Environmental Association


## NMHC Calibration Summary

### Station Information

Calibration Date:	May 16, 2024	Previous Calibration:	April 4, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:43	End Time (MST):	13:19
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

### Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999985 <span style="color: red;">≥0.995</span>
9.11	9.11	0.9998	Slope	1.000672 <span style="color: red;">0.90 - 1.10</span>
4.56	4.53	1.0061	Intercept	-0.014781 <span style="color: red;">+/-0.5</span>
2.28	2.26	1.0099		



NMHC Calibration Plot

Date: May 16, 2024

Location: Ells River





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Ells River  
 Station number: AMS 30  
 Calibration Date: May 22, 2024  
 Last Cal Date: April 5, 2024  
 Start time (MST): 8:00  
 End time (MST): 12:45  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: DT0027487  
 NOX Cal Gas Conc: 59.30 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 59.30 ppm  
 NOX gas Diff:  
 Calibrator Model: API T700  
 ZAG make/model: API T701H  
 Cal Gas Expiry Date: January 9, 2032  
 NO Cal Gas Conc: 59.10 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 59.10 ppm  
 NO gas Diff:  
 Serial Number: 3061  
 Serial Number: 358

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
AF High point	4932	67.7	803.0	800.3	2.7	806.3	801.1	5.3	0.9957	0.9988
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 799.7 ppb		NO = 798.3 ppb			<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = 0.8%	
Baseline Corr 1st pt	NO <sub>x</sub> = 806.4 ppb		NO = 801.2 ppb			<u>As Found Statistics</u>		*Percent Change	NO = 0.4%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb		NO = NA ppb			As found	NO r <sup>2</sup> :	NO SI:	NO Int:	
						As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 710321429

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	0.996856	0.999915
NO <sub>x</sub> Cal Offset:	-0.719188	-0.918381
NO Cal Slope:	0.999360	1.001445
NO Cal Offset:	-1.439899	-2.419656
NO <sub>2</sub> Cal Slope:	0.994435	0.998908
NO <sub>2</sub> Cal Offset:	0.194273	-0.395536

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.195	1.195	NO bkgnd or offset:	14.0	14.0
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	13.9	13.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	193.6	193.6

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
High point	4932	67.7	803.0	800.3	2.7	802.3	800.0	2.2	1.0008	1.0003
Mid point	4966	33.8	400.9	399.5	1.4	400.1	397.2	2.9	1.0020	1.0059
Low point	4983	16.9	200.4	199.8	0.7	197.9	194.6	3.4	1.0128	1.0265
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
As left span	4932	67.7	803.0	414.4	388.6	803.7	414.4	389.3	0.9991	1.0000
Average Correction Factor									1.0052	1.0109

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	798.3	410.2	390.8	390.1	1.0018	99.8%
Mid GPT point	798.3	612.8	188.2	187.8	1.0022	99.8%
Low GPT point	798.3	700.2	100.8	99.5	1.0131	98.7%
Average Correction Factor					1.0057	99.4%

Notes: Sample inlet filters changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 22, 2024	Previous Calibration:	April 5, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:00	End Time (MST):	12:45
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999989	≥0.995
803.0	802.3	1.0008	Slope	0.999915	0.90 - 1.10
400.9	400.1	1.0020	Intercept	-0.918381	+/-20
200.4	197.9	1.0128			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 22, 2024	Previous Calibration:	April 5, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:00	End Time (MST):	12:45
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999989	≥0.995
390.8	390.1	1.0018	Slope	0.998908	0.90 - 1.10
188.2	187.8	1.0022	Intercept	-0.395536	+/-20
100.8	99.5	1.0131			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 22, 2024	Previous Calibration:	April 5, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:00	End Time (MST):	12:45
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999952	<span style="color: red;">≥0.995</span>
800.3	800.0	1.0003	Slope	1.001445	<span style="color: red;">0.90 - 1.10</span>
399.5	397.2	1.0059	Intercept	-2.419656	<span style="color: red;">+/-20</span>
199.8	194.6	1.0265			






NO<sub>x</sub> Calibration Plot

Date: May 22, 2024

Location: Ells River





# Wood Buffalo Environmental Association

## T640 PM<sub>2.5</sub> CALIBRATION

Version-01-2024

### Station Information

Station Name: Ells River Station number: AMS 30  
 Calibration Date: May 21, 2024 Last Cal Date: April 4, 2024  
 Start time (MST): 11:34 End time (MST): 12:31

Analyzer Make: API T640 S/N: 875  
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754  
 Temp/RH standard: Alicat FP-25BT S/N: 388754

### Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.30	9.84	10.30	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	730.40	732.16	730.40	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.98	5.05	4.98	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39	----	39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	1.90	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean  Alignment Factor On :

### Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: September 29, 2024  
 Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.00	12.20	11.20	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: May 21, 2024  
 Date Disposable Filter Changed: May 21, 2024

Post- maintenance Zero Verification: PM w/ HEPA: \_\_\_\_\_ 0.00 <0.2 ug/m3

### Annual Maintenance

Date Sample Tube Cleaned: October 27, 2023  
 Date RH/T Sensor Cleaned: February 23, 2024

Notes: Verified flow, temperature and pressure. Disposable filter changed. Leak check passed. No adjustment made.

Calibration by: Jan Castro



## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS505 SAWBONES BAY MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024



# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Sawbones Bay	Station number:	AMS 505
Calibration Date:	May 31, 2024	Last Cal Date:	April 17, 2024
Start time (MST):	10:07	End time (MST):	13:05
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	51.40	ppm	Cal Gas Exp Date:	February 15, 2029
Cal Gas Cylinder #:	EY0000672			
Removed Cal Gas Conc:	51.40	ppm	Rem Gas Exp Date:	February 15, 2029
Removed Gas Cyl #:	EY0000672		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	5112
Zero Air Gen Model:	Teledyne API T701H		Serial Number:	690

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	710321323
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008070	1.003853	Backgd or Offset:	20.3	20.8
Calibration intercept:	-1.013890	-1.712657	Coeff or Slope:	1.001	0.995

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4922	77.8	799.8	803.0	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	803.1	Previous response	805.3	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4922	77.8	799.8	802.0	0.997
Mid point	4961	38.9	399.9	399.0	1.002
Low point	4981	19.5	200.4	197.7	1.014
As left zero	5000	0.0	0.0	-0.3	----
As left span	4922	77.8	799.8	803.0	0.996
Average Correction Factor:					1.004

Notes: Changed inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 31, 2024	Previous Calibration:	April 17, 2024
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	10:07	End Time (MST):	13:05
Analyzer make:	Thermo 43i	Analyzer serial #:	710321323

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999979	<b>≥0.995</b>
799.8	802.0	0.9973	Slope	1.003853	<b>0.90 - 1.10</b>
399.9	399.0	1.0023	Intercept	-1.712657	<b>+/-30</b>
200.4	197.7	1.0139			



SO2 Calibration Plot

Date: May 31, 2024

Location: Sawbones Bay





# Wood Buffalo Environmental Association

## H2S Calibration Report

### Station Information

Station Name:	Sawbones Bay	Station number:	AMS 505
Calibration Date:	May 29, 2024	Last Cal Date:	April 24, 2024
Start time (MST):	9:42	End time (MST):	14:45
Reason:	Routine H2S cylinder changed		

### Calibration Standards

Cal Gas Concentration:	5.26 ppm	Cal Gas Exp Date:	March 19, 2027
Cal Gas Cylinder #:	DT0034141	Rem Gas Exp Date:	February 5, 2024
Removed Cal Gas Conc:	5.15 ppm	Diff between cyl:	-1.4%
Removed Gas Cyl #:	CC517397	Serial Number:	5112
Calibrator Make/Model:	Teledyne API T700	Serial Number:	690
ZAG Make/Model:	Teledyne API T701		

### Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	12113311965
Converter make:	Global 150	Converter serial #:	2022-224
Analyzer Range	0 - 100 ppb	Converter Temp:	325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998937	0.999743	Backgd or Offset:	0.990	1.450
Calibration intercept:	-0.158258	-0.320000	Coeff or Slope:	1.092	1.105

### H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4922	77.7	80.0	80.1	0.998
As found Mid point	4961	38.8	40.0	40.0	0.997
As found Low point	4981	19.4	20.0	19.7	1.009
New cylinder response	4924	76.0	80.0	78.9	1.013
Baseline Corr As found:	80.2	Prev response:	79.79	*% change:	0.5%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	1.003075	AF Intercept:	-0.178053
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999988	* = > +/-5% change initiates investigation	

### H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4924	76.0	80.0	79.7	1.003
Mid point	4962	38.0	40.0	39.5	1.012
Low point	4981	19.0	20.0	19.6	1.020
As left zero	5000	0.0	0.0	-0.2	----
As left span	4924	76.0	80.0	79.1	1.011
SO2 Scrubber Check	4922	77.8	778.0	-0.2	----
Date of last scrubber change:				Ave Corr Factor	1.012
Date of last converter efficiency test:					

Notes: Changed inlet filter and H2S cylinder after multipoint as founds. SO2 scrubber check done after calibrator zero and passed. Adjusted zero and span.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association


## H2S Calibration Summary

### Station Information

Calibration Date:	May 29, 2024	Previous Calibration:	April 24, 2024
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:42	End Time (MST):	14:45
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12113311965

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999986	$\geq 0.995$
80.0	79.7	1.0032	Slope	0.999743	$0.90 - 1.10$
40.0	39.5	1.0121	Intercept	-0.320000	$\pm 3$
20.0	19.6	1.0198			






H2S Calibration Plot

Date: May 29, 2024

Location: Sawbones Bay





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Sawbones Bay  
 Station number: AMS 505  
 Calibration Date: May 30, 2024  
 Last Cal Date: April 11, 2024  
 Start time (MST): 9:47  
 End time (MST): 14:09  
 Reason: Routine

### Calibration Standards

NO Gas Cylinder #: T1FY3PK  
 NOX Cal Gas Conc: 47.94 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 47.94 ppm  
 NOX gas Diff:  
 Calibrator Model: API T700  
 ZAG make/model: API T701H  
 Cal Gas Expiry Date: March 14, 2025  
 NO Cal Gas Conc: 47.94 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 47.94 ppm  
 NO gas Diff:  
 Serial Number: 5112  
 Serial Number: 690

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	1.1	0.0	1.1	----	----
AF High point	4917	83.4	799.6	799.6	0.0	799.9	799.9	0.0	1.0010	0.9996
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = 799.6 ppb	NO = 800.2 ppb				<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = -0.1%	
Baseline Corr 1st pt	NO <sub>x</sub> = 798.8 ppb	NO = 799.9 ppb				<u>As Found Statistics</u>		*Percent Change	NO = 0.0%	
Baseline Corr 2nd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb				As found	NO <sub>x</sub> r <sup>2</sup> :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO <sub>x</sub> = NA ppb	NO = NA ppb				As found	NO r <sup>2</sup> :	NO SI:	NO Int:	
						As found	NO <sub>2</sub> r <sup>2</sup> :	NO <sub>2</sub> SI:	NO <sub>2</sub> Int:	

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: API T200  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 4260

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.001007	1.005153
NO <sub>x</sub> Cal Offset:	-0.830518	-1.131241
NO Cal Slope:	1.002164	1.006082
NO Cal Offset:	-1.110354	-1.231378
NO <sub>2</sub> Cal Slope:	0.998476	0.996813
NO <sub>2</sub> Cal Offset:	0.748401	-0.401090

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.044	1.044	NO bkgnd or offset:	-0.2	-0.2
NOX coeff or slope:	1.039	1.039	NOX bkgnd or offset:	0.7	0.7
NO2 coeff or slope:	NA	NA	Reaction cell Press:	8.1	8.1

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
High point	4917	83.4	799.6	799.6	0.0	803.0	803.8	-0.7	0.9957	0.9947
Mid point	4958	41.7	399.8	399.8	0.0	400.7	400.7	0.0	0.9979	0.9979
Low point	4979	20.9	200.4	200.4	0.0	198.7	198.7	0.0	1.0085	1.0085
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.1	-0.3	----	----
As left span	4916	83.4	799.7	345.2	454.4	796.7	345.2	451.5	1.0038	1.0000
Average Correction Factor									1.0007	1.0004

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	799.7	345.4	454.3	452.6	1.0038	99.6%
Mid GPT point	799.7	554.8	244.9	243.7	1.0049	99.5%
Low GPT point	799.7	653.2	146.5	145.2	1.0090	99.1%
Average Correction Factor					1.0059	99.4%

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Jan Castro



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	April 11, 2024
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:47	End Time (MST):	14:09
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999988	≥0.995
799.6	803.0	0.9957	Slope	1.005153	0.90 - 1.10
399.8	400.7	0.9979	Intercept	-1.131241	+/-20
200.4	198.7	1.0085			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	April 11, 2024
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:47	End Time (MST):	14:09
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999997	<i>≥0.995</i>
454.3	452.6	1.0038	Slope	0.996813	<i>0.90 - 1.10</i>
244.9	243.7	1.0049	Intercept	-0.401090	<i>+/-20</i>
146.5	145.2	1.0090			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 30, 2024	Previous Calibration:	April 11, 2024
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:47	End Time (MST):	14:09
Analyzer make:	API T200	Analyzer serial #:	4260

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999985	<i>≥0.995</i>
799.6	803.8	0.9947	Slope	1.006082	<i>0.90 - 1.10</i>
399.8	400.7	0.9979	Intercept	-1.231378	<i>+/-20</i>
200.4	198.7	1.0085			



NO<sub>x</sub> Calibration Plot

Date: May 30, 2024

Location: Sawbones Bay





## WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY  
MONITORING PROGRAM  
CALIBRATION REPORT

### **AMS507 KIRBY SOUTH MAY 2024**

Operations, Data Collection,  
QA/QC, Data Validation and Reporting by:  
Wood Buffalo Environmental Association  
Fort McMurray, Alberta

June 28, 2024





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Kirby South	Station number:	AMS 507
Calibration Date:	May 1, 2024	Last Cal Date:	April 18, 2024
Start time (MST):	15:30	End time (MST):	17:10
Reason:	Removal		

### Calibration Standards

Cal Gas Concentration:	49.18	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC303554</u>		Rem Gas Exp Date:	NA
Removed Cal Gas Conc:	49.18	ppm	Diff between cyl:	
Removed Gas Cyl #:	<u>NA</u>		Serial Number:	3804
Calibrator Model:	Teledyne API T700		Serial Number:	880
Zero Air Gen Model:	Teledyne API T701H			

### Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1173410001
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000467		Backgd or Offset:	25.5	25.7
Calibration intercept:	-1.470231		Coeff or Slope:	0.906	0.906

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4919	81.3	799.6	801.0	0.999
As found Mid point	4959	40.7	400.3	398.5	1.005
As found Low point	4980	20.3	199.7	198.8	1.005
New cylinder response					
Baseline Corr As found:	800.8	Previous response	798.5	*% change	0.3%
Baseline Corr 2nd AF pt:	398.3	AF Slope:	1.001678	AF Intercept:	-0.868812
Baseline Corr 3rd AF pt:	198.6	AF Correlation:	0.999986	<i>* = &gt; +/-5% change initiates investigation</i>	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor:


Notes: Removed due to zero drift. All as found values passed.

Calibration Performed By: Braiden Boutilier

SO2 Calibration Plot

Date: May 1, 2024

Location: Kirby South





# Wood Buffalo Environmental Association

## SO<sub>2</sub> Calibration Report

### Station Information

Station Name:	Kirby South	Station number:	AMS 507
Calibration Date:	May 2, 2024	Last Cal Date:	May 1, 2024
Start time (MST):	13:15	End time (MST):	16:15
Reason:	Install		

### Calibration Standards

Cal Gas Concentration:	49.18	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC303554</u>			
Removed Cal Gas Conc:	49.18	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	3804
Zero Air Gen Model:	Teledyne API T701H		Serial Number:	880

### Analyzer Information

Analyzer make:	Thermo 43iQ	Serial Number:	11823400017
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.001378	Backgd or Offset:	NA	24.7
Calibration intercept:	NA	-0.288825	Coeff or Slope:	NA	1.041

### SO<sub>2</sub> As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
-----------	-------------------------------	-----------------------------	-------------------------------------	------------------------------------	--

As found zero  
 As found High point  
 As found Mid point  
 As found Low point  
 New cylinder response

Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = &gt; +/-5% change initiates investigation</i>	

### SO<sub>2</sub> Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4919	81.3	799.6	801.0	0.998
Mid point	4959	40.7	400.3	399.3	1.003
Low point	4980	20.3	199.7	200.0	0.998
As left zero	5000	0.0	0.0	-0.1	----
As left span	4919	81.3	799.6	804.0	0.995
Average Correction Factor:					1.000

Notes: Swapped external pump, adjusted span.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association


## SO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 2, 2024	Previous Calibration:	May 1, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	13:15	End Time (MST):	16:15
Analyzer make:	Thermo 43iQ	Analyzer serial #:	11823400017

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999993	<b>≥0.995</b>
799.6	801.0	0.9983	Slope	1.001378	<b>0.90 - 1.10</b>
400.3	399.3	1.0026	Intercept	-0.288825	<b>+/-30</b>
199.7	200.0	0.9983			



SO2 Calibration Plot

Date: May 2, 2024

Location: Kirby South





# Wood Buffalo Environmental Association

## H2S Calibration Report

### Station Information

Station Name:	Kirby South	Station number:	AMS 507
Calibration Date:	May 1, 2024	Last Cal Date:	April 17, 2024
Start time (MST):	11:55	End time (MST):	17:50
Reason:	Routine		

### Calibration Standards

Cal Gas Concentration:	5.05	ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	<u>DT0019762</u>			
Removed Cal Gas Conc:	5.05	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	n/a		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T750		Serial Number:	281
ZAG Make/Model:	Teledyne API T751H		Serial Number:	322

### Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012
Converter make:	Global	Converter serial #:	2022-197
Analyzer Range	0 - 100 ppb	Converter Temp:	350 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999242	1.002242	Backgd or Offset:	1.72	1.75
Calibration intercept:	-0.020960	0.099034	Coeff or Slope:	1.041	1.041

### H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4921	79.2	80.0	81.3	0.983
As found Mid point	4960	39.6	40.0	40.8	0.978
As found Low point	4980	19.8	20.0	20.3	0.980
New cylinder response	5000	0.0	0.0		
Baseline Corr As found:	81.4	Prev response:	79.91	*% change:	1.8%
Baseline Corr 2nd AF pt:	40.9	AF Slope:	1.017672	AF Intercept:	-0.040981
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999993	<i>* = &gt; +/-5% change initiates investigation</i>	

### H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	79.2	80.0	80.2	0.997
Mid point	4960	39.6	40.0	40.3	0.993
Low point	4980	19.8	20.0	20.2	0.990
As left zero	5000	0.0	0.0	0.0	----
As left span	4921	79.2	80.0	80.0	1.000
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:		July 25, 2023		Ave Corr Factor	<b>0.993</b>
Date of last converter efficiency test:		n/a			

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association


## H2S Calibration Summary

### Station Information

Calibration Date:	May 1, 2024	Previous Calibration:	April 17, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	11:55	End Time (MST):	17:50
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999991	$\geq 0.995$
80.0	80.2	0.9974	Slope	1.002242	$0.90 - 1.10$
40.0	40.3	0.9925	Intercept	0.099034	$\pm 3$
20.0	20.2	0.9900			



H2S Calibration Plot

Date: May 1, 2024

Location: Kirby South







# Wood Buffalo Environmental Association

## THC Calibration Report

### Station Information

Station Name: Kirby South	Station number: AMS 507
Calibration Date: May 1, 2024	Last Cal Date: April 18, 2024
Start time (MST): 15:30	End time (MST): 17:10
Reason: As Found	

### Calibration Standards

Gas Cert Reference: CC303554	Cal Gas Expiry Date: March 23, 2025
CH4 Cal Gas Conc. 496.6 ppm	CH4 Equiv Conc. 1061.7 ppm
C3H8 Cal Gas Conc. 205.5 ppm	
Removed Gas Cert: NA	Removed Gas Expiry: NA
Removed CH4 Conc. 496.6 ppm	CH4 Equiv Conc. 1061.7 ppm
Removed C3H8 Conc. 205.5 ppm	Diff between cyl:
Calibrator Make/Model: Teledyne API T700	Serial Number: 3804
ZAG Make/Model: Teledyne API T701H	Serial Number: 880

### Analyzer Information

Analyzer make: Thermo 51i-LT	Analyzer serial #: 1182340005
Analyzer Range: 0 - 20 ppm	

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001763		Background:	2.07
Calibration intercept:	-0.076111		Coefficient:	3.685

### THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
<i>Limit = 0.90-1.10</i>					
As found zero	5000	0.0	0.00	-0.04	----
As found High point	4919	81.3	17.26	17.05	1.010
As found Mid point	4959	40.7	8.64	8.43	1.020
As found Low point	4980	20.3	4.31	4.22	1.010
New cylinder response					
Baseline Corr As found:	17.09	Previous response	17.22	*% change	-0.7%
Baseline Corr 2nd AF pt:	8.47	AF Slope:	0.989698	AF Intercept:	-0.061470
Baseline Corr 3rd AF pt:	4.27	AF Correlation:	0.999967	<i>* = &gt; +/-5% change initiates investigation</i>	

### THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)
<i>Limit = 0.95-1.05</i>					
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor


Notes: As founds completed with SO2 removal.

Calibration Performed By: Braiden Boutillier

THC Calibration Plot

Date: May 1, 2024

Location: Kirby South







# Wood Buffalo Environmental Association


## THC Calibration Summary

### Station Information

Calibration Date:	May 2, 2024	Previous Calibration:	May 1, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	13:15	End Time (MST):	16:15
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1182340005

### Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.05	----	Correlation Coefficient	0.999968	<span style="color: red;">≥0.995</span>
17.26	17.24	1.0013	Slope	1.002250	<span style="color: red;">0.90 - 1.10</span>
8.64	8.52	1.0144	Intercept	-0.091288	<span style="color: red;">+/-1.5</span>
4.31	4.21	1.0234			



THC Calibration Plot

Date: May 2, 2024

Location: Kirby South





# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Kirby South  
 Station number: AMS 507  
 Calibration Date: May 1, 2024  
 Last Cal Date: April 17, 2024  
 Start time (MST): 11:46  
 End time (MST): 14:50  
 Reason: Removal

### Calibration Standards

NO Gas Cylinder #: T34ULGL  
 NOX Cal Gas Conc: 49.39 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 49.39 ppm  
 NOX gas Diff:  
 Calibrator Model: Teledyne API T700  
 ZAG make/model: Teledyne API T701  
 Cal Gas Expiry Date: March 8, 2025  
 NO Cal Gas Conc: 49.02 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 49.02 ppm  
 NO gas Diff:  
 Serial Number: 3804  
 Serial Number: 880

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
AF High point	4919	81.0	800.1	794.1	6.0	806.0	796.1	9.7	0.9926	0.9974
AF Mid point	4960	40.5	400.1	397.1	3.0	394.9	389.0	5.9	1.0128	1.0205
AF Low point	4980	20.2	199.5	198.0	1.5	192.7	189.0	3.7	1.0349	1.0473

New cyl resp

Previous Response	NO <sub>x</sub> = 795.7 ppb	NO = 790.0 ppb	<i>* = &gt; +/-5% change initiates investigation</i>		*Percent Change	NO <sub>x</sub> = 1.3%
Baseline Corr 1st pt	NO <sub>x</sub> = 806.1 ppb	NO = 796.2 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.8%
Baseline Corr 2nd pt	NO <sub>x</sub> = 395.0 ppb	NO = 389.1 ppb	As found	NO <sub>x</sub> r <sup>2</sup> : 0.999820	Nx SI: 1.009857	Nx Int: -5.002
Baseline Corr 3rd pt	NO <sub>x</sub> = 192.8 ppb	NO = 189.1 ppb	As found	NO r <sup>2</sup> : 0.999763	NO SI: 1.005408	NO Int: -5.685
			As found	NO <sub>2</sub> r <sup>2</sup> : 0.999852	NO <sub>2</sub> SI: 0.998731	NO <sub>2</sub> Int: -2.008

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	0.0	----	----
As found high GPT point	799.0	405.7	399.3	398.1	1.0030	99.7%
As found mid GPT point	799.0	626.3	178.7	173.9	1.0276	97.3%
As found low GPT point	799.0	713.9	91.1	88.2	1.0328	96.8%



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42iQ  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12400232071

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	1.000289	
NO <sub>x</sub> Cal Offset:	-4.604106	
NO Cal Slope:	1.001623	
NO Cal Offset:	-5.445289	
NO <sub>2</sub> Cal Slope:	0.999497	
NO <sub>2</sub> Cal Offset:	1.161308	

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.057		NO bkgnd or offset:	1.1	
NOX coeff or slope:	0.998		NOX bkgnd or offset:	1.2	
NO2 coeff or slope:	1.000		Reaction cell Press:	197.21	

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>x</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero  
 High point  
 Mid point  
 Low point  
 As left zero  
 As left span

Average Correction Factor

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	NO <sub>2</sub> Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero  
 High GPT point  
 Mid GPT point  
 Low GPT point

Average Correction Factor


Notes: Removed due to linearity issues. As found values passed.

Calibration Performed By: Braiden Boutilier

NO<sub>x</sub> Calibration Plot

Date: May 1, 2024

Location: Kirby South







# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Station Information

Station Name: Kirby South  
 Station number: AMS 507  
 Calibration Date: May 4, 2024  
 Last Cal Date: May 1, 2024  
 Start time (MST): 8:40  
 End time (MST): 12:50  
 Reason: Install

### Calibration Standards

NO Gas Cylinder #: T34ULGL  
 NOX Cal Gas Conc: 49.39 ppm  
 Removed Cylinder #: NA  
 Removed Gas NOX Conc: 49.39 ppm  
 NOX gas Diff:  
 Calibrator Model: Teledyne API T700  
 ZAG make/model: Teledyne API T701  
 Cal Gas Expiry Date: March 8, 2025  
 NO Cal Gas Conc: 49.02 ppm  
 Removed Gas Exp Date: NA  
 Removed Gas NO Conc: 49.02 ppm  
 NO gas Diff:  
 Serial Number: 5240  
 Serial Number: 880

### As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO <sub>x</sub> concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>x</sub> concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>x</sub> Correction factor (Cc/(Ic-AFzero))	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero))
As found zero									<i>Limit = 0.90 - 1.10</i>	<i>Limit = 0.90 - 1.10</i>
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO <sub>x</sub> = NA	ppb	NO = NA	ppb						NO <sub>x</sub> = NA
Baseline Corr 1st pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb						NO = NA
Baseline Corr 2nd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb						
Baseline Corr 3rd pt	NO <sub>x</sub> = NA	ppb	NO = NA	ppb						

### As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO <sub>2</sub> concentration (ppb) (Cc)	Indicated NO <sub>2</sub> concentration (ppb) (Ic)	Baseline Adjusted NO <sub>2</sub> Correction factor (Cc/(Ic-AFzero))	Converter Efficiency
As Found GPT zero					<i>Limit = 0.90 - 1.10</i>	<i>Limit = 96-104%</i>
As found high GPT point						
As found mid GPT point						
As found low GPT point						



# Wood Buffalo Environmental Association

## NO<sub>x</sub> \ NO \ NO<sub>2</sub> Calibration Report

### Analyzer Information

Analyzer Make: Thermo 42i  
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12400232071

### Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO <sub>x</sub> Cal Slope:	NA	1.002136
NO <sub>x</sub> Cal Offset:	NA	-4.264015
NO Cal Slope:	NA	1.002087
NO Cal Offset:	NA	-4.324137
NO <sub>2</sub> Cal Slope:	NA	1.002404
NO <sub>2</sub> Cal Offset:	NA	1.422929

### Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	NA	0.889	NO bkgnd or offset:	NA	8.7
NOX coeff or slope:	NA	0.997	NOX bkgnd or offset:	NA	8.8
NO2 coeff or slope:	NA	1.000	Reaction cell Press:	NA	149.20

### Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
High point	4918	81.6	806.1	800.1	6.0	806.0	800.0	6.0	1.0001	1.0001
Mid point	4959	40.8	403.0	400.0	3.0	396.7	393.2	3.6	1.0160	1.0173
Low point	4980	20.4	201.5	200.0	1.5	193.8	192.4	1.4	1.0397	1.0394
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
As left span	4918	81.6	806.1	800.1	6.0	809.0	405.3	403.5	0.9964	1.9740
Average Correction Factor									1.0186	1.0190

### GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	798.2	402.4	401.8	403.3	0.9964	100.4%
Mid GPT point	798.2	604.3	199.9	203.2	0.9840	101.6%
Low GPT point	798.2	702.2	102.0	104.7	0.9746	102.6%
Average Correction Factor					0.9850	101.5%

Notes: Swapped calibrator due to a leak (removed SN: 3804, installed SN: 5240). Replaced external pump.

Calibration Performed By: Braiden Boutilier



# Wood Buffalo Environmental Association


## NO<sub>x</sub> Calibration Summary

### Station Information

Calibration Date:	May 4, 2024	Previous Calibration:	May 1, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	8:40	End Time (MST):	12:50
Analyzer make:	Thermo 42i	Analyzer serial #:	12400232071

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999864	≥0.995
806.1	806.0	1.0001	Slope	1.002136	0.90 - 1.10
403.0	396.7	1.0160	Intercept	-4.264015	+/-20
201.5	193.8	1.0397			





# Wood Buffalo Environmental Association


## NO<sub>2</sub> Calibration Summary

### Station Information

Calibration Date:	May 4, 2024	Previous Calibration:	May 1, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	8:40	End Time (MST):	12:50
Analyzer make:	Thermo 42i	Analyzer serial #:	12400232071

### Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999935	≥0.995
401.8	403.3	0.9964	Slope	1.002404	0.90 - 1.10
199.9	203.2	0.9840	Intercept	1.422929	+/-20
102.0	104.7	0.9746			





# Wood Buffalo Environmental Association


## NO Calibration Summary

### Station Information

Calibration Date:	May 4, 2024	Previous Calibration:	May 1, 2024
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	8:40	End Time (MST):	12:50
Analyzer make:	Thermo 42i	Analyzer serial #:	12400232071

### Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999855	≥0.995
800.1	800.0	1.0001	Slope	1.002087	0.90 - 1.10
400.0	393.2	1.0173	Intercept	-4.324137	+/-20
200.0	192.4	1.0394			



NO<sub>x</sub> Calibration Plot

Date: May 4, 2024

Location: Kirby South





End of Report