



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

DECEMBER 2024 MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

January 31, 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01 BERTHA GANTER - FORT MCKAY DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number: AMS 01
Calibration Date:	December 6, 2024	Last Cal Date: November 15, 2024
Start time (MST):	11:52	End time (MST): 15:10
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #:	CC418809		
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	Teledyne API T700		Serial Number: 3565
Zero Air Gen Model:	Teledyne API T701		Serial Number: 4766

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: JC1501301448
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994626	0.994626	Backgd or Offset:	20.5	20.8
Calibration intercept:	-0.213467	-0.213467	Coeff or Slope:	0.875	0.888

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.4	----
As found High point	4918	81.3	800.3	790.1	1.013
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	789.7	Previous response	795.8	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.6	----
High point	4918	81.3	800.3	801.9	0.998
Mid point	4959	40.7	400.6	399.0	1.004
Low point	4979	20.3	199.8	200.5	0.997
As left zero	5000	0.0	0.0	0.6	----
As left span	4918	81.3	800.3	800.8	0.999
Average Correction Factor:					1.000

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 15, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:52	End Time (MST):	15:10
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999985
800.3	801.9	0.9980	Slope	1.000978
400.6	399.0	1.0040	Intercept	-0.012554
199.8	200.5	0.9966		
				≥0.995 0.90 - 1.10 +/-30



SO2 Calibration Plot

Date: December 6, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	December 9, 2024	Last Cal Date:	November 6, 2024
Start time (MST):	10:45	End time (MST):	15:25
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.84	ppm	Cal Gas Exp Date:	September 5, 2027
Cal Gas Cylinder #:	CC738239			
Removed Cal Gas Conc:	4.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4766

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461
Converter make:	CD Nova	Converter serial #:	470
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994520	0.995667	Backgd or Offset:	2.60	2.50
Calibration intercept:	0.021668	0.081535	Coeff or Slope:	0.910	0.893

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4917	82.6	80.0	80.9	0.987
As found Mid point	4959	41.3	40.0	40.6	0.982
As found Low point	4979	20.7	20.0	20.3	0.982
New cylinder response					
Baseline Corr As found:	81.0	Prev response:	79.55	*% change:	1.8%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.012825	AF Intercept:	-0.018317
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999992	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4917	82.6	80.0	79.7	1.003
Mid point	4959	41.3	40.0	39.9	1.002
Low point	4979	20.7	20.0	20.0	1.002
As left zero	5000	0.0	0.0	0.2	----
As left span	4921	78.4	75.9	78.4	0.968
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:		December 17, 2021		Ave Corr Factor	1.002
Date of last converter efficiency test:					

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 6, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:45	End Time (MST):	15:25
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	1.000000	≥ 0.995
80.0	79.7	1.0033	Slope	0.995667	$0.90 - 1.10$
40.0	39.9	1.0020	Intercept	0.081535	± 3
20.0	20.0	1.0019			



TRS Calibration Plot

Date: December 9, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	December 9, 2024	Last Cal Date:	November 7, 2024
Start time (MST):	10:45	End time (MST):	15:25
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.84	ppm	Cal Gas Exp Date:	September 5, 2027
Cal Gas Cylinder #:	CC738239			
Removed Cal Gas Conc:	4.84	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3565
ZAG Make/Model:	Teledyne API T701		Serial Number:	4766

Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1200326167
Converter make:	CD Nova	Converter serial #:	2022-221
Analyzer Range	0 - 100 ppb	Converter Temp:	315 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996526	0.992381	Backgd or Offset:	1.58	1.57
Calibration intercept:	0.202067	0.322108	Coeff or Slope:	0.969	0.969

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4917	82.6	80.0	79.8	1.006
As found Mid point	4959	41.3	40.0	40.0	1.007
As found Low point	4979	20.7	20.0	20.1	1.012
New cylinder response					
Baseline Corr As found:	79.5	Prev response:	79.89	*% change:	-0.5%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.994669	AF Intercept:	0.242050
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999997	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4917	82.6	80.0	79.7	1.003
Mid point	4959	41.3	40.0	40.0	0.999
Low point	4979	20.7	20.0	20.1	0.997
As left zero	5000	0.0	0.0	0.5	----
As left span	4917	82.6	80.0	79.8	1.002
SO ₂ Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	January 25, 2024			Ave Corr Factor	1.000
Date of last converter efficiency test:	November 7, 2024			107.9% efficiency	

Notes: Inlet filter change and scrubber check completed after as founds.No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 7, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:45	End Time (MST):	15:25
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999995	≥ 0.995
80.0	79.7	1.0033	Slope	0.992381	$0.90 - 1.10$
40.0	40.0	0.9994	Intercept	0.322108	± 3
20.0	20.1	0.9970			



H₂S Calibration Plot

Date: December 9, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	December 6, 2024	Last Cal Date:	November 15, 2024
Start time (MST):	11:52	End time (MST):	15:10
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	497.2 ppm	CH4 Equiv Conc.	1061.8 ppm
C3H8 Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.2 ppm	CH4 Equiv Conc.	1061.8 ppm
Removed C3H8 Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
Zero Air Gen model:	Teledyne API T701	Serial Number:	4766

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320040
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.48E-04	4.45E-04	NMHC SP Ratio:	8.38E-05	8.40E-05
CH4 Retention time:	16.7	16.7	NMHC Peak Area:	109721	109431
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	17.27	17.39	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.39	Prev response	17.20	*% change	1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	17.27	17.27	1.000
Mid point	4959	40.7	8.64	8.49	1.018
Low point	4979	20.3	4.31	4.25	1.014
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	17.27	17.40	0.992
Average Correction Factor					1.011

Notes: Changed the inlet filter after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	9.18	9.25	0.992
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.25	Prev response	9.15	*% change	1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	9.18	9.19	1.000
Mid point	4959	40.7	4.60	4.56	1.007
Low point	4979	20.3	2.29	2.30	0.998
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	9.18	9.30	0.987
Average Correction Factor					1.002

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	8.09	8.14	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.14	Prev response	8.05	*% change	1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	8.09	8.09	1.000
Mid point	4959	40.7	4.05	3.92	1.031
Low point	4979	20.3	2.02	1.96	1.032
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	8.09	8.10	0.999
Average Correction Factor					1.021

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999339	1.000516
THC Cal Offset:	-0.055421	-0.055816
CH ₄ Cal Slope:	1.001489	1.000794
CH ₄ Cal Offset:	-0.043511	-0.049303
NMHC Cal Slope:	0.997471	0.999886
NMHC Cal Offset:	-0.011511	-0.005712

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 15, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:52	End Time (MST):	15:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999899	<i>≥0.995</i>
17.27	17.27	0.9996	Slope	1.000516	<i>0.90 - 1.10</i>
8.64	8.49	1.0182	Intercept	-0.055816	<i>+/-0.5</i>
4.31	4.25	1.0140			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 15, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:52	End Time (MST):	15:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999710	≥ 0.995
8.09	8.09	1.0001	Slope	1.000794	$0.90 - 1.10$
4.05	3.92	1.0315	Intercept	-0.049303	± 0.5
2.02	1.96	1.0316			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 15, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:52	End Time (MST):	15:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999981	<i>≥0.995</i>
9.18	9.19	0.9996	Slope	0.999886	<i>0.90 - 1.10</i>
4.60	4.56	1.0070	Intercept	-0.005712	<i>+/-0.5</i>
2.29	2.30	0.9985			



NMHC Calibration Plot

Date: December 6, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	December 17, 2024	Last Cal Date:	December 6, 2024
Start time (MST):	12:45	End time (MST):	15:10
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	497.2 ppm	CH4 Equiv Conc.	1061.8 ppm
C3H8 Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	497.2 ppm	CH4 Equiv Conc.	1061.8 ppm
Removed C3H8 Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
Zero Air Gen model:	Teledyne API T701	Serial Number:	4766

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320040
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.45E-04	4.45E-04	NMHC SP Ratio:	8.40E-05	8.40E-05
CH4 Retention time:	16.7	16.7	NMHC Peak Area:	109431	109431
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	17.27	17.37	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.37	Prev response	17.22	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	17.27	17.38	0.993
Average Correction Factor					

Notes: Changing the N2 cylinder.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	9.18	9.34	0.984
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.34	Prev response	9.17	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	9.18	9.35	0.982
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	8.09	8.03	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.03	Prev response	8.04	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	8.09	8.04	1.006
Average Correction Factor					

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000516	
THC Cal Offset:	-0.055816	
CH ₄ Cal Slope:	1.000794	
CH ₄ Cal Offset:	-0.049303	
NMHC Cal Slope:	0.999886	
NMHC Cal Offset:	-0.005712	

Calibration Performed By: Rene Chamberland

NMHC Calibration Plot

Date: December 17, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay
 Station number: AMS 01
 Calibration Date: December 5, 2024
 Last Cal Date: November 5, 2024
 Start time (MST): 11:53
 End time (MST): 16:48
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC335700
 NOX Cal Gas Conc: 59.40 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 59.40 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: September 1, 2032
 NO Cal Gas Conc: 59.20 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 59.20 ppm
 NO gas Diff:
 Serial Number: 3565
 Serial Number: 4766

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.5	0.1	----	----
AF High point	4932	67.6	803.1	800.4	2.7	794.8	789.3	5.5	1.0099	1.0134
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 801.8 ppb	NO = 799.5 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -0.8%
Baseline Corr 1st pt	NO _x = 795.2 ppb	NO = 789.8 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -1.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 7117

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999286	1.001961
NO _x Cal Offset:	-0.700000	-0.240000
NO Cal Slope:	1.000977	1.001534
NO Cal Offset:	-1.660000	-1.580000
NO ₂ Cal Slope:	1.000376	1.000167
NO ₂ Cal Offset:	-0.520742	0.917072

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.148	1.157	NO bkgnd or offset:	-3.1	-3.1
NOX coeff or slope:	1.148	1.161	NOX bkgnd or offset:	-2.9	-2.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	6.8	6.9

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.1	----	----
High point	4932	67.6	803.1	800.4	2.7	804.7	801.0	3.6	0.9980	0.9992
Mid point	4966	33.8	401.5	400.2	1.4	401.8	398.2	3.6	0.9994	1.0050
Low point	4983	16.9	200.8	200.1	0.7	200.4	197.0	3.4	1.0019	1.0157
As left zero	5000	0.0	0.0	0.0	0.0	1.3	1.2	0.1	----	----
As left span	4932	67.6	803.1	389.1	414.0	797.0	389.1	407.9	1.0076	1.0000
Average Correction Factor									0.9997	1.0066

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	796.3	389.8	409.2	409.9	0.9983	100.2%
Mid GPT point	796.3	594.1	204.9	205.9	0.9952	100.5%
Low GPT point	796.3	696.1	102.9	104.9	0.9810	101.9%
Average Correction Factor					0.9915	100.9%

Notes: Changed the inlet filter after as founds. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 5, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:53	End Time (MST):	16:48
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999998	≥0.995
803.1	804.7	0.9980	Slope	1.001961	0.90 - 1.10
401.5	401.8	0.9994	Intercept	-0.240000	+/-20
200.8	200.4	1.0019			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 5, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:53	End Time (MST):	16:48
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999979	<i>≥0.995</i>
409.2	409.9	0.9983	Slope	1.000167	<i>0.90 - 1.10</i>
204.9	205.9	0.9952	Intercept	0.917072	<i>+/-20</i>
102.9	104.9	0.9810			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 5, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:53	End Time (MST):	16:48
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999975	<i>≥0.995</i>
800.4	801.0	0.9992	Slope	1.001534	<i>0.90 - 1.10</i>
400.2	398.2	1.0050	Intercept	-1.580000	<i>+/-20</i>
200.1	197.0	1.0157			



NO_x Calibration Plot

Date: December 5, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	December 2, 2024	Last Cal Date:	November 1, 2024
Start time (MST):	11:30	End time (MST):	13:57
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3565
Calibrator Make/Model:	Teledyne API T700	Serial Number:	4766
ZAG Make/Model:	Teledyne API T701		

Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #:	1107
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003114	1.003514	Backgd or Offset:	5.4	5.4
Calibration intercept:	0.380000	0.160000	Coeff or Slope:	1.010	1.010

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	5000	863.1	400.0	400.9	0.998
As found Mid point					
As found Low point					
Baseline Corr As found:	400.8	Previous response	401.6	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	5000	863.1	400.0	401.3	0.997
Mid point	5000	744.0	200.0	201.6	0.992
Low point	5000	651.7	100.0	100.1	0.999
As left zero	5000	0.0	0.0	-0.1	----
As left span	5000	863.1	400.0	404.3	0.989
Average Correction Factor					0.996

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 1, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:30	End Time (MST):	13:57
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999991	≥0.995
400.0	401.3	0.9968	Slope	1.003514	0.90 - 1.10
200.0	201.6	0.9921	Intercept	0.160000	+/- 5
100.0	100.1	0.9990			



O₃ Calibration Plot

Date: December 2, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
Calibration Date: December 12, 2024 Last Cal Date: November 15, 2024
Start time (MST): 13:59 End time (MST): 15:09

Analyzer Make: Teledyne API T640 S/N: 322
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752
Temp/RH standard: Alicat FP-25BT S/N: 388752

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-12	-12.7	-12	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	729.7	731.96	729.7	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.98	4.967	4.98	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	40		40	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	5.5	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: October 24, 2024
Date Disposable Filter Changed: October 24, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 24, 2024
Date RH/T Sensor Cleaned: October 24, 2024

Notes: Flow, temperature, and pressure were verified. Leak check passed.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	December 3, 2024	Last Cal Date:	November 4, 2024
Start time (MST):	11:52	End time (MST):	14:52
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,040	ppm	Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #:	ALM042207		
Removed Cal Gas Conc:	3,040	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 3565
ZAG Make/Model:	Teledyne API T701		Serial Number: 4766

Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3520
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002581	1.002243	Backgd or Offset:	-0.014	-0.014
Calibration intercept:	0.161850	0.117838	Coeff or Slope:	0.990	0.996

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4933	66.7	40.6	40.7	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	40.56	Prev response:	40.82	*% change:	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4933	66.7	40.6	40.7	0.998
Mid point	4966	33.3	20.2	20.7	0.980
Low point	4983	16.7	10.2	10.2	0.992
As left zero	5000	0.0	0.0	0.0	----
As left span	2960	40.0	40.5	40.2	1.009
Average Correction Factor					0.990

Notes: Changed the inlet filter after as founds. Adjusted the zero.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


CO Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 4, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:52	End Time (MST):	14:52
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.0	----	Correlation Coefficient	0.999900	≥0.995
40.6	40.7	0.9977	Slope	1.002243	0.90 - 1.10
20.2	20.7	0.9796	Intercept	0.117838	+/-1.5
10.2	10.2	0.9916			



CO Calibration Plot

Date: December 3, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

CO₂ Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number: AMS 01
Calibration Date:	December 4, 2024	Last Cal Date: November 12, 2024
Start time (MST):	11:37	End time (MST): 15:00
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	60,200	ppm	Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #:	ALM042207		
Removed Cal Gas Conc:	60,200	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 3565
N2 Gen Make/Model:	Peak Scientific		Serial Number: 7220900034

Analyzer Information

Analyzer make:	Teledyne API 360	Analyzer serial #: 442
Analyzer Range	0 - 2,000 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998806	0.998885	Backgd or Offset:	0.002	0.002
Calibration intercept:	-1.320000	-1.400000	Coeff or Slope:	0.911	0.913

CO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	3000	0.0	0.0	0.4	----
As found High Point	2920	80.0	1605.3	1587.6	1.011
As found Mid Point					
As found Low Point					
New cylinder response					
Baseline Corr As found:	1587.2	Prev response:	1602.1	*% change:	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	3000	0.0	0.0	0.6	----
High point	2920	80.0	1605.3	1605.2	1.000
Mid point	2960	40.0	802.7	793.2	1.012
Low point	2980	20.0	401.3	401.6	0.999
As left zero	3000	0.0	0.0	-2.5	----
As left span	2960	40.0	802.7	783.8	1.024
Average Correction Factor					1.004

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


CO₂ Calibration Summary

Station Information

Calibration Date	December 4, 2024	Previous Calibration	November 12, 2024
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS 01
Start Time (MST)	11:37	End Time (MST)	15:00
Analyzer make	Teledyne API 360	Analyzer serial #	442

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.6	----	Correlation Coefficient	0.999951	≥0.995
1605.3	1605.2	1.0001	Slope	0.998885	0.90 - 1.10
802.7	793.2	1.0119	Intercept	-1.4	+/-20
401.3	401.6	0.9993			



CO₂ Calibration Plot

Date: December 4, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

Nt - NOX - NH3 Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
NOX Cal Date:	December 11, 2024	Last Cal Date:	November 19, 2024
Start time (MST):	11:19	End time (MST):	16:05
NH3 Cal Date:	December 12, 2024	Last Cal Date:	November 20, 2024
Start time (MST):	12:15	End time (MST):	16:04
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	59.40	ppm	NO Gas Cylinder #:	CC335700
NO Cal Gas Conc:	59.20	ppm	NO Cal Gas Expiry:	September 1, 2032
Removed NOX Conc:	59.40	ppm	Removed Cylinder #:	NA
Removed NO Conc:	59.20	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.58	ppm	NH3 Gas Cylinder #:	CC743587
Removed NH3 Conc:	76.58	ppm	NH3 Cal Gas Expiry:	August 22, 2024
NH3 gas Diff:			Removed Cylinder #:	NA
Calibrator Model:	API T700		Removed cyl Expiry:	NA
ZAG make/model:	API T701		Serial Number:	3565
			Serial Number:	4766

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	475
Converter model:	API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	11.00
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	510

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.838	0.838	Nt coefficient:	0.851	0.847
NOX coefficient:	0.844	0.841	NO bkgrnd:	-0.9	-0.9
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	-0.3	-0.3
NH3 coefficient:	0.946	0.946	Nt bkgrnd:	1.2	1.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999769	1.009375
NO _x Cal Offset:	-2.020000	-3.420000
NO Cal Slope:	0.997664	1.003233
NO Cal Offset:	-2.300000	-2.800000
NO ₂ Cal Slope:	0.996402	1.000150
NO ₂ Cal Offset:	-0.671037	0.920283
NH3 Cal Slope:	1.003838	1.002081
NH3 Cal Offset:	-5.989321	-2.741292
Nt Cal Slope:	1.006936	1.005363
Nt Cal Offset:	-6.268143	-3.121782



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

NO_x / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO _x Correction factor (Cc/Ic)	Baseline corr NO Correction factor (Cc/Ic)
									Limit = 0.9 - 1.0	Limit = 0.9 - 1.0
As found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.4	-0.8	----	----
As found span	4932	67.6	803.1	800.4	803.1	807.2	797.4	808.7	0.9949	1.0037
AF GPT span	4932	67.6	803.1	----	803.1	800.8	----	802.0	1.0029	----

new NO cyl rp

Baseline Corr As Fd	Nt = 809.5 ppb	NO _x = 807.7 ppb	NO = 797.8 ppb	*Percent Change	Nt _(NO) = 0.9%
Previous Response	Nt = 802.39 ppb	NO _x = 800.9 ppb	NO = 796.2 ppb	*Percent Change	NO _x = 0.8%
**NO _x Δ (NO to GPT response) =	-0.8%			*Percent Change	NO = 0.2%

** = > +/-2% difference initiates investigation

* = > +/-5% change initiates investigation

NO_x / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic)	NO Correction factor (Cc/Ic)
									Limit = 0.95-1.05	Limit = 0.95-1.05
Calibration zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.4	-0.9	----	----
High point	4932	67.6	803.1	800.4	803.1	809.2	801.2	810.0	0.9924	0.9990
Mid point	4966	33.8	401.5	400.2	401.5	398.7	398.0	398.9	1.0071	1.0055
Low point	4983	16.9	200.8	200.1	200.8	197.6	195.2	198.0	1.0161	1.0251
Average Correction Factor									1.0052	1.0099

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic)	Converter Efficiency
					Limit = 0.95-1.05	Limit = 96-104%
Calibration zero	----	----	0.0	-0.2	----	----
High GPT point (400 ppb O3)	793.4	389.1	407.0	407.4	0.9990	100.1%
Mid GPT point (200 ppb O3)	793.4	591.0	205.1	206.7	0.9923	100.8%
Low GPT point (100 ppb O3)	793.4	696.2	99.9	101.9	0.9804	102.0%
Average Correction Factor					0.9906	101.0%



Wood Buffalo Environmental Association NH₃ - N_T Calibration Report

NH₃ As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH3 Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.8	0.1	----	----
AF High point	3418	82.2	1798.5	----	1798.5	1806.2	----	1800.8	0.995	0.999
AF Mid point							----			
AF Low point							----			
new NH3 cyl rp										
Baseline Corr As Fd		Nt = 1806.8 ppb	NH3 = 1800.7 ppb						*Percent Change	Nt _(NH3) = 0.1%
Previous Response		Nt = 1804.7 ppb	NH3 = 1799.4 ppb			* = > +/-5% change initiates investigation			*Percent Change	NH3 = 0.1%

NH₃ Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	-0.9	-0.6	-0.3	----	----
High point	3418	82.2	1798.5	----	1798.5	1806.2	----	1800.8	0.996	0.999
Mid point	3454	45.7	1000.0	----	1000.0	1001.5	----	998.5	0.999	1.002
Low point	3477	22.8	498.9	----	498.9	495.8	----	494.3	1.006	1.009
Average Correction Factor									1.0001	1.0032
NH3 Previous Converter Efficiency =		90.8 %								
NH3 Current Converter Efficiency =		90.8 %								

Notes:

Changed the inlet filter after as founds. Adjusted the NOx/Nt span.

Calibration Performed By:

Rene Chamberland



Wood Buffalo Environmental Association


Nt Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 19, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:19	End Time (MST):	16:05
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.9	----	Correlation Coefficient	0.999992	≥ 0.995
1798.5	1806.2	0.9958	Slope	1.005363	0.90 - 1.10
1000.0	1001.5	0.9985	Intercept	-3.121782	+/-20
498.9	495.8	1.0062			





Wood Buffalo Environmental Association


NH₃ Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 19, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:19	End Time (MST):	16:05
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999991	≥0.995
1798.5	1800.8	0.9987	Slope	1.002081	0.90 - 1.10
1000.0	998.5	1.0015	Intercept	-2.741292	+/-20
498.9	494.3	1.0092			





Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 19, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:19	End Time (MST):	16:05
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.6	----	Correlation Coefficient	0.999931	≥0.995
803.1	809.2	0.9924	Slope	1.009375	0.90 - 1.10
401.5	398.7	1.0071	Intercept	-3.420000	+/-20
200.8	197.6	1.0161			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 19, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:19	End Time (MST):	16:05
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	0.999958	≥ 0.995
800.4	801.2	0.9990	Slope	1.003233	0.90 - 1.10
400.2	398.0	1.0055	Intercept	-2.800000	+/-20
200.1	195.2	1.0251			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 19, 2024
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:19	End Time (MST):	16:05
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999966	≥ 0.995
407.0	407.4	0.9990	Slope	1.000150	0.90 - 1.10
205.1	206.7	0.9923	Intercept	0.920283	+/-20
99.9	101.9	0.9804			



NO_x Calibration Plot

Date: December 11, 2024


Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: December 12, 2024

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Fort McKay - Bertha Ganter	Station Number:	AMS 01
Calibration Date:	December 4, 2024	Prev Cal Date:	November 29, 2024
Start Time (MST):	12:20	End Time (MST):	14:18
Tower Height (m):	10.0	Reason:	Install

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	P10041
WS Calibrator:	MetOne 053	Serial Number:	

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0		---
200	20.2		
400	39.4		
600	58.6		
800	77.8		

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999999		≥0.9995
Calculated slope	0.999594		0.90 - 1.10
Calculated intercept	-0.028293		+/- 2

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	R14655
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):		Calc Declination*:	13.74 Degrees
Deadband calc:	2.6 degrees (<i>Limit 4 deg</i>)		<i>* - calculated declination as per NOAA website</i>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.4	---
90	91.8	0.5%
180	180.2	0.0%
270	270.0	0.0%
354	354.8	0.2%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r ²)	0.999987		≥0.9995
Calculated slope	1.001036		0.90 - 1.10
Calculated intercept	-0.841792		+/- 4

Notes:

Sensor install.

Calibration Performed By:

Rene Chamberland




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Mildred Lake	Station number:	AMS 02
Calibration Date:	December 20, 2024	Last Cal Date:	November 4, 2024
Start time (MST):	10:59	End time (MST):	14:24
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.98	ppm	Cal Gas Exp Date:	August 12, 2024
Cal Gas Cylinder #:	CC501209			
Removed Cal Gas Conc:	49.98	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	1185
Zero Air Gen Model:	Teledyne API T701		Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	JC1404901075
Analyzer Range:	0-1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999445	0.994713	Backgd or Offset:	18.9	18.9
Calibration intercept:	-0.705190	-0.546122	Coeff or Slope:	0.776	0.781

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4920	80.2	801.6	792.4	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	792.8	Previous response	800.5	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.2	801.6	797.2	1.006
Mid point	4960	40.1	400.8	397.7	1.008
Low point	4980	20.0	199.9	197.9	1.010
As left zero	5000	0.0	0.0	-0.4	----
As left span	4920	80.2	801.6	800.0	1.002
Average Correction Factor:					1.008

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 20, 2024	Previous Calibration:	November 4, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:59	End Time (MST):	14:24
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999998	≥0.995
801.6	797.2	1.0056	Slope	0.994713	0.90 - 1.10
400.8	397.7	1.0079	Intercept	-0.546122	+/-30
199.9	197.9	1.0102			



SO2 Calibration Plot

Date: December 20, 2024

Location: Mildred Lake





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name:	Mildred Lake	Station number:	AMS 02
Calibration Date:	December 18, 2024	Last Cal Date:	November 21, 2024
Start time (MST):	11:00	End time (MST):	14:53
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.29	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC345191			
Removed Cal Gas Conc:	5.29	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	1185
ZAG Make/Model:	Teledyne API T701		Serial Number:	4891

Analyzer Information

Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331546
Converter make:	Global G150	Converter serial #:	2023-267
Analyzer Range	0 - 100 ppb	Converter Temp:	325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998678	0.991106	Backgd or Offset:	1.46	1.46
Calibration intercept:	-0.099199	-0.059207	Coeff or Slope:	1.007	1.007

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4924	75.6	80.0	79.3	1.009
As found Mid point	4962	37.8	40.0	39.5	1.013
As found Low point	4981	18.9	20.0	19.6	1.020
New cylinder response					
Baseline Corr As found:	79.3	Prev response:	79.79	*% change:	-0.6%
Baseline Corr 2nd AF pt:	39.5	AF Slope:	0.992106	AF Intercept:	-0.119209
Baseline Corr 3rd AF pt:	19.6	AF Correlation:	0.999990	<i>* = > +/-5% change initiates investigation</i>	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4924	75.6	80.0	79.2	1.010
Mid point	4962	37.8	40.0	39.7	1.007
Low point	4981	18.9	20.0	19.6	1.020
As left zero	5000	0.0	0.0	0.1	----
As left span	4924	75.6	80.0	79.2	1.010
SO2 Scrubber Check	4920	80.2	802.0	0.1	----
Date of last scrubber change:	July 16, 2024			Ave Corr Factor	1.013
Date of last converter efficiency test:	NA				

Notes: Changed sample inlet filter after as founds. SO2 scrubber test done after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	July 5, 2024	Previous Calibration:	June 24, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	11:00	End Time (MST):	14:53
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331546

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999987	≥ 0.995
80.0	79.2	1.0100	Slope	0.991106	$0.90 - 1.10$
40.0	39.7	1.0074	Intercept	-0.059207	± 3
20.0	19.6	1.0202			



H2S Calibration Plot

Date: December 18, 2024

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Mildred Lake	Station number:	AMS 02
Calibration Date:	December 20, 2024	Last Cal Date:	November 4, 2024
Start time (MST):	10:59	End time (MST):	14:24
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC501209	Cal Gas Expiry Date:	August 12, 2024
CH4 Cal Gas Conc.	500.2 ppm	CH4 Equiv Conc.	1048.6 ppm
C3H8 Cal Gas Conc.	199.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	500.2 ppm	CH4 Equiv Conc.	1048.6 ppm
Removed C3H8 Conc.	199.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
Zero Air Gen model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1170050131
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.60E-04	2.60E-04	NMHC SP Ratio:	6.85E-05	6.97E-05
CH4 Retention time:	14.6	14.6	NMHC Peak Area:	128504	126172
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	16.82	16.66	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.66	Prev response	16.87	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	
High point	4920	80.2	16.82	16.82	1.000
Mid point	4960	40.0	8.39	8.41	0.997
Low point	4980	20.0	4.19	4.19	1.002
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	16.82	16.75	1.004
Average Correction Factor					1.000

Notes: Changed sample inlet filter after as founds. Adjusted span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	8.80	8.67	1.014
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.67	Prev response	8.77	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	8.80	8.80	0.999
Mid point	4960	40.0	4.39	4.39	0.999
Low point	4980	20.0	2.19	2.19	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	8.80	8.77	1.003
Average Correction Factor					1.000

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	8.02	8.00	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.00	Prev response	8.11	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	8.02	8.02	1.000
Mid point	4960	40.0	4.00	4.02	0.996
Low point	4980	20.0	2.00	2.00	1.003
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	8.02	7.98	1.005
Average Correction Factor					1.000

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.005571	1.000410
THC Cal Offset:	-0.037095	0.001069
CH ₄ Cal Slope:	1.013671	1.000108
CH ₄ Cal Offset:	-0.022023	0.001548
NMHC Cal Slope:	0.999351	1.000698
NMHC Cal Offset:	-0.017070	-0.000278

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 20, 2024	Previous Calibration:	November 4, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:59	End Time (MST):	14:24
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050131

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995
16.82	16.82	0.9999	Slope	1.000410	$0.90 - 1.10$
8.39	8.41	0.9974	Intercept	0.001069	± 0.5
4.19	4.19	1.0017			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 20, 2024	Previous Calibration:	November 4, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:59	End Time (MST):	14:24
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050131

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
8.02	8.02	1.0004	Slope	1.000108	<i>0.90 - 1.10</i>
4.00	4.02	0.9959	Intercept	0.001548	<i>+/-0.5</i>
2.00	2.00	1.0029			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 20, 2024	Previous Calibration:	November 4, 2024
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:59	End Time (MST):	14:24
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050131

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	1.000000	<i>≥0.995</i>
8.80	8.80	0.9995	Slope	1.000698	<i>0.90 - 1.10</i>
4.39	4.39	0.9986	Intercept	-0.000278	<i>+/-0.5</i>
2.19	2.19	1.0006			



NMHC Calibration Plot

Date: December 20, 2024

Location: Mildred Lake






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Buffalo Viewpoint	Station number: AMS 04
Calibration Date:	December 11, 2024	Last Cal Date: November 6, 2024
Start time (MST):	7:25	End time (MST): 10:10
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	50.87	ppm	Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #:	CC446753		
Removed Cal Gas Conc:	50.87	ppm	Rem Gas Exp Date:
Removed Gas Cyl #:			Diff between cyl:
Calibrator Model:	API T700		Serial Number: 3808
Zero Air Gen Model:	API T701		Serial Number: 362

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: JC1327300932
Analyzer Range:	0-1000ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002759	0.996587	Backgd or Offset:	26.5	26.5
Calibration intercept:	-0.385234	-0.826674	Coeff or Slope:	0.877	0.877

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4921	78.6	799.7	799.2	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	799.1	Previous response	801.6	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4921	78.6	799.7	796.6	1.004
Mid point	4961	39.3	399.8	397.4	1.006
Low point	4980	19.6	199.4	196.7	1.014
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	78.6	799.7	794.2	1.007
Average Correction Factor:					1.008

Notes: No Maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 6, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:25	End Time (MST):	10:10
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999992	≥0.995
799.7	796.6	1.0039	Slope	0.996587	0.90 - 1.10
399.8	397.4	1.0061	Intercept	-0.826674	+/-30
199.4	196.7	1.0139			



SO2 Calibration Plot

Date: December 11, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS 04
Calibration Date:	December 19, 2024	Last Cal Date:	November 26, 2024
Start time (MST):	7:08	End time (MST):	11:50
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.42	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC345266			
Removed Cal Gas Conc:	5.42	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3808
ZAG Make/Model:	Teledyne API T701H		Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400
Converter make:	Global	Converter serial #:	2022-200
Analyzer Range	0 - 100 ppb	Converter Temp:	325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999922	0.990525	Backgd or Offset:	1.93	1.95
Calibration intercept:	-0.117954	0.362080	Coeff or Slope:	1.118	1.140

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4926	74.1	80.3	78.5	1.025
As found Mid point	4963	37.0	40.1	39.5	1.018
As found Low point	4982	18.5	20.1	19.5	1.034
New cylinder response					
Baseline Corr As found:	78.4	Prev response:	80.20	*% change:	-2.3%
Baseline Corr 2nd AF pt:	39.4	AF Slope:	0.977152	AF Intercept:	0.081725
Baseline Corr 3rd AF pt:	19.4	AF Correlation:	0.999974	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4926	74.1	80.3	79.8	1.007
Mid point	4963	37.0	40.1	40.4	0.993
Low point	4982	18.5	20.1	20.1	0.998
As left zero	5000	0.0	0.0	0.4	----
As left span	4926	74.1	80.3	80.2	1.002
SO2 Scrubber Check	4920	80.0	800.0	0.1	----
Date of last scrubber change:	16-May-23			Ave Corr Factor	0.999

Date of last converter efficiency test:

Notes: Accidentally hit the high point instead of the SO2 Scrubber check on the calibrator, went back down to zero then did the Sox scrubber check. Sox scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	December 19, 2024	Previous Calibration:	November 26, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:08	End Time (MST):	11:50
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999962	≥0.995
80.3	79.8	1.0066	Slope	0.990525	0.90 - 1.10
40.1	40.4	0.9928	Intercept	0.362080	+/-3
20.1	20.1	0.9976			



H₂S Calibration Plot

Date: December 19, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS 04
Calibration Date:	December 11, 2024	Last Cal Date:	November 6, 2024
Start time (MST):	7:25	End time (MST):	10:10
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	497.2 ppm	CH4 Equiv Conc.	1058.2 ppm
C3H8 Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	497.2 ppm	CH4 Equiv Conc.	1058.2 ppm
Removed C3H8 Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
Zero Air Gen model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1426262594
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.51E-04	4.56E-04	NMHC SP Ratio:	1.15E-04	1.16E-04
CH4 Retention time:	13.7	13.7	NMHC Peak Area:	76662	75886
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	16.64	16.35	1.018
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.35	Prev response	16.65	*% change	-1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	16.64	16.57	1.004
Mid point	4961	39.3	8.32	8.22	1.011
Low point	4980	19.6	4.15	4.11	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	16.64	16.55	1.006
Average Correction Factor					1.009

Notes: Hydrogen Cylinder changed out. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	8.82	8.66	1.019
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.66	Prev response	8.84	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	8.82	8.77	1.005
Mid point	4961	39.3	4.41	4.36	1.012
Low point	4980	19.6	2.20	2.19	1.006
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	8.82	8.77	1.006
Average Correction Factor					1.008

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	7.82	7.69	1.016
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.69	Prev response	7.81	*% change	-1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	7.82	7.80	1.003
Mid point	4961	39.3	3.91	3.86	1.012
Low point	4980	19.6	1.95	1.92	1.015
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	7.82	7.78	1.005
Average Correction Factor					1.010

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001514	0.996136
THC Cal Offset:	-0.013931	-0.022555
CH ₄ Cal Slope:	1.001161	0.997784
CH ₄ Cal Offset:	-0.015104	-0.016312
NMHC Cal Slope:	1.001840	0.994430
NMHC Cal Offset:	0.001373	-0.006044

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 6, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:25	End Time (MST):	10:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999983	<i>≥0.995</i>
16.64	16.57	1.0040	Slope	0.996136	<i>0.90 - 1.10</i>
8.32	8.22	1.0114	Intercept	-0.022555	<i>+/-0.5</i>
4.15	4.11	1.0103			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 6, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:25	End Time (MST):	10:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999974	<i>≥0.995</i>
7.82	7.80	1.0028	Slope	0.997784	<i>0.90 - 1.10</i>
3.91	3.86	1.0116	Intercept	-0.016312	<i>+/-0.5</i>
1.95	1.92	1.0152			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 6, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:25	End Time (MST):	10:10
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999988	<i>≥0.995</i>
8.82	8.77	1.0053	Slope	0.994430	<i>0.90 - 1.10</i>
4.41	4.36	1.0115	Intercept	-0.006044	<i>+/-0.5</i>
2.20	2.19	1.0061			



NMHC Calibration Plot

Date: December 11, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Buffalo Viewpoint
 Station number: AMS 04
 Calibration Date: December 18, 2024
 Last Cal Date: November 15, 2024
 Start time (MST): 7:20
 End time (MST): 12:07
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC324979
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #:
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: APIT701
 Cal Gas Expiry Date: November 3, 2032
 NO Cal Gas Conc: 48.80 ppm
 Removed Gas Exp Date:
 Removed Gas NO Conc: 48.80 ppm
 NO gas Diff:
 Serial Number: 3808
 Serial Number: 362

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.2	----	----
AF High point	4918	81.8	800.0	798.4	1.6	795.7	792.9	2.8	1.0060	1.0071
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 803.9 ppb	NO = 801.3 ppb				<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -1.1%	
Baseline Corr 1st pt	NO _x = 795.3 ppb	NO = 792.8 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -1.1%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found NO _x r ² :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found NO r ² :		NO SI:	NO Int:	
						As found NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 721

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.004121	1.004479
NO _x Cal Offset:	0.547607	-0.452723
NO Cal Slope:	1.003031	0.996506
NO Cal Offset:	0.446673	-1.575097
NO ₂ Cal Slope:	0.992934	0.989876
NO ₂ Cal Offset:	0.795535	0.217112

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.240	1.250	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.228	1.239	NOX bkgnd or offset:	-0.2	-0.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.8	4.8

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.2	----	----
High point	4918	81.8	800.0	798.4	1.6	803.6	795.0	8.5	0.9956	1.0043
Mid point	4959	40.9	400.0	399.2	0.8	400.9	395.0	5.9	0.9978	1.0106
Low point	4980	20.4	199.5	199.1	0.4	199.1	195.4	3.7	1.0020	1.0189
As left zero	5000	0.0	0.0	0.6	-0.6	0.5	0.6	-0.1	----	----
As left span	4918	81.8	800.0	400.6	800.0	793.0	400.6	392.3	1.0089	1.0000
Average Correction Factor									0.9984	1.0113

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	796.1	399.0	398.7	394.7	1.0102	99.0%
Mid GPT point	796.1	592.9	204.8	203.6	1.0061	99.4%
Low GPT point	796.1	692.4	105.3	104.1	1.0119	98.8%
Average Correction Factor					1.0094	99.1%

Notes: No maintenance done. Span adjusted. Due to drifting during the GPT the 2nd NO ref point used.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 18, 2024	Previous Calibration:	November 15, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:20	End Time (MST):	12:07
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.4	----	Correlation Coefficient	0.999995	<i>≥0.995</i>
800.0	803.6	0.9956	Slope	1.004479	<i>0.90 - 1.10</i>
400.0	400.9	0.9978	Intercept	-0.452723	<i>+/-20</i>
199.5	199.1	1.0020			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 18, 2024	Previous Calibration:	November 15, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:20	End Time (MST):	12:07
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999993	<i>≥0.995</i>
398.7	394.7	1.0102	Slope	0.989876	<i>0.90 - 1.10</i>
204.8	203.6	1.0061	Intercept	0.217112	<i>+/-20</i>
105.3	104.1	1.0119			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 18, 2024	Previous Calibration:	November 15, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:20	End Time (MST):	12:07
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999979	<i>≥0.995</i>
798.4	795.0	1.0043	Slope	0.996506	<i>0.90 - 1.10</i>
399.2	395.0	1.0106	Intercept	-1.575097	<i>+/-20</i>
199.1	195.4	1.0189			



NO_x Calibration Plot

Date: December 18, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS 04
Calibration Date:	December 11, 2024	Last Cal Date:	November 25, 2024
Start time (MST):	10:08	End time (MST):	12:17
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3808
Calibrator Make/Model:	APIP T700	Serial Number:	362
ZAG Make/Model:	API T701		

Analyzer Information

Analyzer make:	API T400	Analyzer serial #:	2961
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998371	0.999200	Backgd or Offset:	-2.2	-2.2
Calibration intercept:	0.260000	0.240000	Coeff or Slope:	1.020	1.020

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.5	----
As found High point	5000	997.9	400.0	398.8	1.002
As found Mid point					
As found Low point					
Baseline Corr As found:	399.3	Previous response	399.6	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	5000	997.1	400.0	399.7	1.001
Mid point	5000	821.2	200.0	200.4	0.998
Low point	5000	709.9	100.0	100.4	0.996
As left zero	5000	0.0	0.0	-0.8	----
As left span	5000	996.1	400.0	401.3	0.997
Average Correction Factor					0.998

Notes: No adjustments or maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 25, 2024
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	10:08	End Time (MST):	12:17
Analyzer make:	API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999996	≥0.995
400.0	399.7	1.0008	Slope	0.999200	0.90 - 1.10
200.0	200.4	0.9980	Intercept	0.240000	+/- 5
100.0	100.4	0.9960			



O₃ Calibration Plot

Date: December 11, 2024

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
 Calibration Date: December 19, 2024 Last Cal Date: November 26, 2024
 Start time (MST): 8:32 End time (MST): 9:45

Analyzer Make: Teledyne API T640 S/N: 321
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
 Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-24.6	-25.6	-24.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	743.0	743.3	743.0	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.20	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	42	----	42	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	0.1	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: 6-10-2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.7	10.7	10.7	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: December 19, 2024
 Date Disposable Filter Changed: December 19, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: February 27, 2024
 Date RH/T Sensor Cleaned: February 27, 2024

No adjustments done. Leak Check, Flow and PMT checked before and after optical cleaning.

Notes:

Calibration by: Melissa Lemay




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS05
MANNIX
DECEMBER 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	December 12, 2024	Last Cal Date:	November 21, 2024
Start time (MST):	10:02	End time (MST):	13:00
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.84	ppm	Cal Gas Exp Date:	January 6, 2030
Cal Gas Cylinder #:	CC408659			
Removed Cal Gas Conc:	49.84	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	5470
Zero Air Gen Model:	API T701		Serial Number:	361

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1008841399
Analyzer Range:	1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997157	0.992430	Backgd or Offset:	10.0	10.0
Calibration intercept:	-0.665731	-0.685495	Coeff or Slope:	0.935	0.935

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	80.3	800.4	794.2	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	794.2	Previous response	797.4	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4920	80.3	800.4	793.6	1.009
Mid point	4960	40.1	399.7	396.9	1.007
Low point	4980	20.1	200.4	196.5	1.020
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	80.3	800.4	794.8	1.007
Average Correction Factor:					1.012

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 21, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:02	End Time (MST):	13:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999988	≥0.995
800.4	793.6	1.0085	Slope	0.992430	0.90 - 1.10
399.7	396.9	1.0071	Intercept	-0.685495	+/-30
200.4	196.5	1.0196			



SO2 Calibration Plot

Date: December 12, 2024

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	December 3, 2024	Last Cal Date:	November 20, 2024
Start time (MST):	9:54	End time (MST):	14:30
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.96	ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	DT0037363			
Removed Cal Gas Conc:	4.96	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	5470
ZAG Make/Model:	API T701		Serial Number:	361

Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	1200326169
Converter make:	Global	Converter serial #:	2022-225
Analyzer Range	0 - 100 ppb	Converter Temp:	325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.987394	0.996970	Backgd or Offset:	1.22	1.25
Calibration intercept:	0.022526	0.062422	Coeff or Slope:	1.009	1.029

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4919	80.6	80.0	78.4	1.020
As found Mid point	4960	40.3	40.0	39.3	1.017
As found Low point	4980	20.2	20.0	19.6	1.022
New cylinder response					
Baseline Corr As found:	78.4	Prev response:	78.98	*% change:	-0.7%
Baseline Corr 2nd AF pt:	39.3	AF Slope:	0.980819	AF Intercept:	0.002612
Baseline Corr 3rd AF pt:	19.6	AF Correlation:	0.999996	* = +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4919	80.6	80.0	79.7	1.003
Mid point	4960	40.3	40.0	40.2	0.994
Low point	4980	20.2	20.0	19.8	1.012
As left zero	5000	0.0	0.0	0.3	----
As left span	4919	80.6	80.0	79.0	1.012
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.003
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 20, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:54	End Time (MST):	14:30
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1200326169

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999958	≥0.995
80.0	79.7	1.0033	Slope	0.996970	0.90 - 1.10
40.0	40.2	0.9944	Intercept	0.062422	+/-3
20.0	19.8	1.0120			



H₂S Calibration Plot

Date: December 3, 2024

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Mannix	Station number:	AMS 05
Calibration Date:	December 12, 2024	Last Cal Date:	November 21, 2024
Start time (MST):	10:02	End time (MST):	13:00
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC408659	Cal Gas Expiry Date:	January 6, 2030
CH4 Cal Gas Conc.	507.2 ppm	CH4 Equiv Conc.	1057.8 ppm
C3H8 Cal Gas Conc.	200.2 ppm		
Removed Gas Cert:	N/A	Removed Gas Expiry:	N/A
Removed CH4 Conc.	507.2 ppm	CH4 Equiv Conc.	1057.8 ppm
Removed C3H8 Conc.	200.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	5470
Zero Air Gen model:	API T701	Serial Number:	361

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1193585649
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.94E-04	2.94E-04	NMHC SP Ratio:	5.37E-05	5.37E-05
CH4 Retention time:	15.4	15.4	NMHC Peak Area:	164658	164658
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	16.99	17.03	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.03	Prev response	17.07	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	16.99	17.01	0.998
Mid point	4960	40.1	8.48	8.53	0.995
Low point	4980	20.1	4.25	4.23	1.005
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	16.99	17.05	0.996
Average Correction Factor					0.999

Notes: Changed the inlet filter after as founds. No adjustments made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	8.84	8.87	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.87	Prev response	8.97	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	8.84	8.85	0.999
Mid point	4960	40.1	4.42	4.44	0.996
Low point	4980	20.1	2.21	2.20	1.006
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	8.84	8.87	0.997
Average Correction Factor					1.000

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	8.15	8.17	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.17	Prev response	8.10	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	8.15	8.16	0.998
Mid point	4960	40.1	4.07	4.09	0.994
Low point	4980	20.1	2.04	2.03	1.004
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	8.15	8.18	0.996
Average Correction Factor					0.999

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.005080	1.002401
THC Cal Offset:	-0.003124	-0.004715
CH ₄ Cal Slope:	0.994044	1.002491
CH ₄ Cal Offset:	0.001544	-0.001054
NMHC Cal Slope:	1.014886	1.002085
NMHC Cal Offset:	-0.004268	-0.003261

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 21, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:02	End Time (MST):	13:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999989	≥0.995
16.99	17.01	0.9984	Slope	1.002401	0.90 - 1.10
8.48	8.53	0.9946	Intercept	-0.004715	+/-0.5
4.25	4.23	1.0050			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 21, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:02	End Time (MST):	13:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999986	≥0.995
8.15	8.16	0.9983	Slope	1.002491	0.90 - 1.10
4.07	4.09	0.9936	Intercept	-0.001054	+/-0.5
2.04	2.03	1.0044			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 21, 2024
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:02	End Time (MST):	13:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999991	<i>≥0.995</i>
8.84	8.85	0.9987	Slope	1.002085	<i>0.90 - 1.10</i>
4.42	4.44	0.9956	Intercept	-0.003261	<i>+/-0.5</i>
2.21	2.20	1.0055			



NMHC Calibration Plot

Date: December 12, 2024

Location: Mannix





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number: AMS 06
Calibration Date:	December 2, 2024	Last Cal Date: November 4, 2024
Start time (MST):	10:30	End time (MST): 13:45
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	49.78	ppm	Cal Gas Exp Date: September 9, 2024
Cal Gas Cylinder #:	AAL070632		
Removed Cal Gas Conc:	49.78	ppm	Rem Gas Exp Date:
Removed Gas Cyl #:			Diff between cyl:
Calibrator Model:	API T700		Serial Number: 3566
Zero Air Gen Model:	API T701		Serial Number: 5608

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 1160290013
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998832	0.996115	Backgd or Offset:	18.1	18.1
Calibration intercept:	1.880203	1.781114	Coeff or Slope:	0.910	0.910

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4919.7	80.3	799.5	796.5	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	796.7	Previous response	800.4	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4919.7	80.3	799.5	797.6	1.002
Mid point	4959.8	40.2	400.2	400.5	0.999
Low point	4979.9	20.1	200.1	203.2	0.985
As left zero	5000	0.0	0.0	0.2	----
As left span	4919.7	80.3	799.5	797.7	1.002
Average Correction Factor:					0.995

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 4, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:30	End Time (MST):	13:45
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999979
799.5	797.6	1.0023	Slope	0.996115
400.2	400.5	0.9993	Intercept	1.781114
200.1	203.2	0.9848		
				≥0.995
				0.90 - 1.10
				+/-30



SO2 Calibration Plot

Date: December 2, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	December 13, 2024	Last Cal Date:	November 27, 2024
Start time (MST):	9:36	End time (MST):	13:45
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.328	ppm	Cal Gas Exp Date:	February 14, 2025
Cal Gas Cylinder #:	CC506659			
Removed Cal Gas Conc:	5.328	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3566
ZAG Make/Model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358
Converter make:	CDN-101	Converter serial #:	517
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002736	1.001163	Backgd or Offset:	1.99	1.99
Calibration intercept:	0.300485	0.280517	Coeff or Slope:	1.153	1.153

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4925	75.1	80.0	82.2	0.975
As found Mid point	4963	37.5	40.0	41.5	0.965
As found Low point	4981	18.8	20.0	20.7	0.973
New cylinder response					
Baseline Corr As found:	82.1	Prev response:	80.54	*% change:	1.9%
Baseline Corr 2nd AF pt:	41.4	AF Slope:	1.026308	AF Intercept:	0.200345
Baseline Corr 3rd AF pt:	20.6	AF Correlation:	0.999968	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4925	75.1	80.0	80.2	0.998
Mid point	4963	37.5	40.0	40.8	0.979
Low point	4981	18.8	20.0	20.1	0.997
As left zero	5000	0.0	0.0	0.3	----
As left span	4925	75.1	80.0	80.9	0.989
SO2 Scrubber Check	4920	80.3	803.0	0.1	----
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	0.991
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber test after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	December 13, 2024	Previous Calibration:	November 27, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:36	End Time (MST):	13:45
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999895	≥ 0.995
80.0	80.2	0.9978	Slope	1.001163	$0.90 - 1.10$
40.0	40.8	0.9793	Intercept	0.280517	± 3
20.0	20.1	0.9967			



TRS Calibration Plot

Date: December 13, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	December 1, 2024	Last Cal Date:	
Start time (MST):	6:45	End time (MST):	8:45
Reason:	Install		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH4 Cal Gas Conc.	501.6 ppm	CH4 Equiv Conc.	1066.2 ppm
C3H8 Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	501.6 ppm	CH4 Equiv Conc.	1066.2 ppm
Removed C3H8 Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
Zero Air Gen model:	API T701	Serial Number:	4602

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320037
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.14E-04		NMHC SP Ratio:	5.56E-05
CH4 Retention time:	15.6		NMHC Peak Area:	163047
Zero Chromatogram:	OFF		Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	17.12	17.10	1.001
Mid point	4960	40.2	8.57	8.61	0.995
Low point	4980	20.1	4.29	4.38	0.979
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	17.12	16.99	1.008
Average Correction Factor					0.992

Notes: Install calibration removed 55i not working properly.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFZero))
Baseline Adjusted <i>Limit = 0.90-1.10</i>					
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Baseline Adjusted <i>Limit = 0.95-1.05</i>					
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	9.07	9.07	1.000
Mid point	4960	40.2	4.54	4.59	0.990
Low point	4980	20.1	2.27	2.33	0.972
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	9.07	8.98	1.009
Average Correction Factor					0.987

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFZero))
Baseline Adjusted <i>Limit = 0.90-1.10</i>					
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Baseline Adjusted <i>Limit = 0.95-1.05</i>					
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	8.06	8.04	1.002
Mid point	4960	40.2	4.03	4.03	1.001
Low point	4980	20.1	2.02	2.05	0.986
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	8.06	8.01	1.006
Average Correction Factor					0.997

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		0.997000
THC Cal Offset:		0.049629
CH ₄ Cal Slope:		0.996570
CH ₄ Cal Offset:		0.013646
NMHC Cal Slope:		0.998050
NMHC Cal Offset:		0.035582

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 1, 2024	Previous Calibration:	
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	6:45	End Time (MST):	8:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999961 ≥0.995
17.12	17.10	1.0014	Slope	0.997000 0.90 - 1.10
8.57	8.61	0.9953	Intercept	0.049629 +/-0.5
4.29	4.38	0.9790		





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 1, 2024	Previous Calibration:	
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	6:45	End Time (MST):	8:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999980	<i>≥0.995</i>
8.06	8.04	1.0021	Slope	0.996570	<i>0.90 - 1.10</i>
4.03	4.03	1.0014	Intercept	0.013646	<i>+/-0.5</i>
2.02	2.05	0.9860			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 1, 2024	Previous Calibration:	
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	6:45	End Time (MST):	8:45
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999929 ≥0.995
9.07	9.07	1.0001	Slope	0.998050 0.90 - 1.10
4.54	4.59	0.9895	Intercept	0.035582 +/-0.5
2.27	2.33	0.9724		



NMHC Calibration Plot

Date: December 1, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	December 20, 2024	Last Cal Date:	December 1, 2024
Start time (MST):	12:55	End time (MST):	14:44
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	AAL070632	Cal Gas Expiry Date:	September 9, 2024
CH4 Cal Gas Conc.	501.6 ppm	CH4 Equiv Conc.	1066.2 ppm
C3H8 Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	501.6 ppm	CH4 Equiv Conc.	1066.2 ppm
Removed C3H8 Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
Zero Air Gen model:	API T701	Serial Number:	4602

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320037
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.14E-04	3.14E-04	NMHC SP Ratio: 5.56E-05	5.56E-05
CH4 Retention time:	15.6	15.6	NMHC Peak Area: 163047	163047
Zero Chromatogram:	OFF	OFF	Flat Baseline: OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	17.12	17.11	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.11	Prev response	17.12	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	17.12	17.08	1.003
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.003

Notes: Changed the N2 cylinder after as founds. Small dips occurring. Will monitor.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	9.07	9.18	0.987
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.18	Prev response	9.08	*% change	1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	9.07	9.34	0.971
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 0.971

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	8.06	7.92	1.017
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.92	Prev response	8.04	*% change	-1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	8.06	7.74	1.041
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.041

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.997000	0.997269
THC Cal Offset:	0.049629	0.000000
CH ₄ Cal Slope:	0.996570	0.960438
CH ₄ Cal Offset:	0.013646	0.000000
NMHC Cal Slope:	0.998050	1.030101
NMHC Cal Offset:	0.035582	0.000000

Calibration Performed By: Max Farrell

NMHC Calibration Plot

Date: December 20, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Patricia McInnes
 Station number: AMS 06
 Calibration Date: December 4, 2024
 Last Cal Date: November 13, 2024
 Start time (MST): 10:00
 End time (MST): 15:00
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T30YCWN
 NOX Cal Gas Conc: 47.94 ppm
 Removed Cylinder #: N/A
 Removed Gas NOX Conc: 47.94 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: April 11, 2025
 NO Cal Gas Conc: 46.39 ppm
 Removed Gas Exp Date: N/A
 Removed Gas NO Conc: 46.39 ppm
 NO gas Diff:
 Serial Number: 3566
 Serial Number: 4602

As Found Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.0	-0.2	-0.8	----	----
AF High point	4914	86.2	826.5	799.7	26.7	833.2	802.5	30.7	0.9907	0.9963
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 825.4 ppb	NO = 797.5 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 1.1%
Baseline Corr 1st pt	NO _x = 834.2 ppb	NO = 802.7 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.6%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1172750022

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996717	1.003204
NO _x Cal Offset:	1.636275	1.714916
NO Cal Slope:	0.995700	1.003246
NO Cal Offset:	1.223377	0.882218
NO ₂ Cal Slope:	1.005126	1.002822
NO ₂ Cal Offset:	0.370755	-0.465504

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.832	0.832	NO bkgnd or offset:	3.7	3.7
NOX coeff or slope:	0.989	0.989	NOX bkgnd or offset:	4.7	4.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	155.1	155.1

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.1	-0.5	----	----
High point	4914	86.2	826.5	799.7	26.7	829.6	802.6	27.1	0.9962	0.9964
Mid point	4957	43.1	413.2	399.9	13.4	417.6	402.9	14.7	0.9895	0.9925
Low point	4978	21.6	207.1	200.4	6.7	211.7	202.7	9.0	0.9784	0.9888
As left zero	5000	0.0	0.0	0.0	0.0	-0.9	-0.2	-0.7	----	----
As left span	4914	86.2	826.5	405.9	420.6	829.0	405.9	423.1	0.9969	1.0000
Average Correction Factor									0.9880	0.9926

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.5	----	----
High GPT point	800.6	403.2	424.1	424.6	0.9989	100.1%
Mid GPT point	800.6	599.5	227.8	228.7	0.9962	100.4%
Low GPT point	800.6	700.0	127.3	126.8	1.0041	99.6%
Average Correction Factor					0.9997	100.0%

Notes: Changed the inlet filter after as founds. No adjustments needed.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 4, 2024	Previous Calibration:	November 13, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:00	End Time (MST):	15:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.6	----	Correlation Coefficient	0.999964	<i>≥0.995</i>
826.5	829.6	0.9962	Slope	1.003204	<i>0.90 - 1.10</i>
413.2	417.6	0.9895	Intercept	1.714916	<i>+/-20</i>
207.1	211.7	0.9784			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 4, 2024	Previous Calibration:	November 13, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:00	End Time (MST):	15:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.5	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
424.1	424.6	0.9989	Slope	1.002822	<i>0.90 - 1.10</i>
227.8	228.7	0.9962	Intercept	-0.465504	<i>+/-20</i>
127.3	126.8	1.0041			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 4, 2024	Previous Calibration:	November 13, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:00	End Time (MST):	15:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999993	<i>≥0.995</i>
799.7	802.6	0.9964	Slope	1.003246	<i>0.90 - 1.10</i>
399.9	402.9	0.9925	Intercept	0.882218	<i>+/-20</i>
200.4	202.7	0.9888			



NO_x Calibration Plot

Date: December 4, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	December 11, 2024	Last Cal Date:	November 5, 2024
Start time (MST):	11:37	End time (MST):	14:37
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3566
Calibrator Make/Model:	API T700	Serial Number:	4602
ZAG Make/Model:	API T701		

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000771	1.002771	Backgd or Offset:	-0.9	-0.9
Calibration intercept:	0.140000	0.140000	Coeff or Slope:	1.020	1.020

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	0.2	----
As found High point	5000	1031.0	400.0	400.5	0.999
As found Mid point					
As found Low point					
Baseline Corr As found:	400.3	Previous response	400.4	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.4	----
High point	5000	1031.0	400.0	401.4	0.997
Mid point	5000	821.4	200.0	200.5	0.998
Low point	5000	699.5	100.0	100.2	0.998
As left zero	5000	800.0	0.0	0.7	----
As left span	5000	1031.0	400.0	403.1	0.992
Average Correction Factor					0.997

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 5, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	11:37	End Time (MST):	14:37
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999998	≥0.995
400.0	401.4	0.9965	Slope	1.002771	0.90 - 1.10
200.0	200.5	0.9975	Intercept	0.140000	+/- 5
100.0	100.2	0.9980			



O₃ Calibration Plot

Date: December 11, 2024

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Patricia McInnes Station number: AMS 06
 Calibration Date: December 13, 2024 Last Cal Date: November 26, 2024
 Start time (MST): 14:07 End time (MST): 15:12

Analyzer Make: API T640 S/N: 766
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388755
 Temp/RH standard: Alicat FP-25BT S/N: 388755

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-2.1	-2.2	-2.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	717.00	719.00	717.00	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	5.16	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	41	----	41	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	15.4	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: 6-10-2024
 Lot No.: 100128-050-035

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.8	11.8	10.8	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: December 13, 2024
 Date Disposable Filter Changed: December 13, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: April 13, 2023
 Date RH/T Sensor Cleaned: April 13, 2023

Notes: Parameters within limits. Leak check passed. Adjusted the PMT peak.

Calibration by: Max Farrell



Wood Buffalo Environmental Association

Nt - NOX - NH3 Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	December 9, 2024	Last Cal Date:	November 19, 2024
Start time (MST):	9:48	End time (MST):	14:00
NH3 Cal Date:	December 9, 2024	Last Cal Date:	November 19, 2024
Start time (MST):	14:00	End time (MST):	16:15
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	N/A
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	N/A
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	76.3	ppm	NH3 Gas Cylinder #:	EB0108520
			NH3 Cal Gas Expiry:	August 22, 2024
Removed NH3 Conc:	76.3	ppm	Removed Cylinder #:	N/A
NH3 gas Diff:			Removed cyl Expiry:	N/A
Calibrator Model:	API T700		Serial Number:	3566
ZAG make/model:	API T701		Serial Number:	4602

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	215
Converter model:	API T501	Converter serial #:	147
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.20
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	25.4

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	1.009	1.009	Nt coefficient:	1.700	1.700
NOX coefficient:	1.002	1.002	NO bkgrnd:	0.2	0.2
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	0.1	0.1
NH3 coefficient:	1.200	1.200	Nt bkgrnd:	0.0	0.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.005599	0.996150
NO _x Cal Offset:	1.473664	3.016603
NO Cal Slope:	1.003405	0.996442
NO Cal Offset:	-0.298339	0.463717
NO ₂ Cal Slope:	1.002194	0.999502
NO ₂ Cal Offset:	2.833619	0.782057
NH3 Cal Slope:	0.995698	0.988331
NH3 Cal Offset:	7.440608	8.443668
Nt Cal Slope:	0.999133	0.991747
Nt Cal Offset:	7.756362	8.825682



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

NO_x / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO _x Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>	Baseline corr NO Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.2	-0.8	----	----
As found span	4914	86.2	826.5	799.7	826.5	829.5	796.5	834.9	0.9963	1.0041
AF GPT span										
new NO cyl rp										

Baseline Corr As Fd Nt = 835.7 ppb NO_x = 829.5 ppb NO = 796.7 ppb
 Previous Response Nt = 833.49 ppb NO_x = 832.6 ppb NO = 802.2 ppb

**NO_x Δ (NO to GPT response) =

* * = > +/-2% difference initiates investigation

*Percent Change Nt(NO) = 0.3%

*Percent Change NO_x = -0.4%

*Percent Change NO = -0.7%

* = > +/-5% change initiates investigation

NO_x / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.4	0.0	1.5	----	----
High point	4914	86.2	826.5	799.7	826.5	824.5	797.1	831.1	1.0024	1.0033
Mid point	4957	43.1	413.2	399.9	413.2	417.4	399.2	419.8	0.9900	1.0017
Low point	4978	21.6	207.1	200.4	207.1	211.0	200.6	213.6	0.9816	0.9991
Average Correction Factor									0.9913	1.0014

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found zero	----	----	0.0	0.2	----	----
Calibration zero	----	----	0.0	0.5	----	----
High GPT point (400 ppb O3)	800.3	397.1	429.9	430.3	0.9991	100.1%
Mid GPT point (200 ppb O3)	800.3	598.8	228.2	229.1	0.9962	100.4%
Low GPT point (100 ppb O3)	800.3	699.6	127.4	128.4	0.9924	100.8%
Average Correction Factor					0.9959	100.4%



Wood Buffalo Environmental Association NH₃ - N_T Calibration Report

NH₃ As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH ₃ concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH ₃ concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH ₃ Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.8	0.0	-0.8	----	----
AF High point	3417	82.6	1800.6	0.0	1800.6	1787.6	6.5	1781.1	1.007	1.011
AF Mid point										
AF Low point										
new NH ₃ cyl rp										
Baseline Corr As Fd	Nt = 1788.4 ppb	NH ₃ = 1781.9 ppb							*Percent Change	Nt _(NH₃) = -1.0%
Previous Response	Nt = 1806.8 ppb	NH ₃ = 1800.3 ppb							*Percent Change	NH ₃ = -1.0%

* = > +/-5% change initiates investigation

NH₃ Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH ₃ concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH ₃ concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH ₃ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	1.5	0.4	1.1	----	----
High point	3417	82.6	1800.6	0.0	1800.6	1787.6	6.5	1781.1	1.007	1.011
Mid point	3454	45.9	1000.5	0.0	1000.5	1012.0	3.8	1008.1	0.989	0.992
Low point	3477	22.9	499.2	0.0	499.2	507.3	2.0	505.3	0.984	0.988
Average Correction Factor									0.9933	0.9971
NH ₃ Previous Converter Efficiency =	90.2 %									
NH ₃ Current Converter Efficiency =	90.2 %									

Notes:

Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By:

Max Farrell



Wood Buffalo Environmental Association


Nt Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 19, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:48	End Time (MST):	14:00
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.5	----	Correlation Coefficient	0.999866	≥ 0.995
1800.6	1787.6	1.0073	Slope	0.991747	0.90 - 1.10
1000.5	1012.0	0.9887	Intercept	8.825682	+/-20
499.2	507.3	0.9840			





Wood Buffalo Environmental Association


NH₃ Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 19, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:48	End Time (MST):	14:00
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.1	----	Correlation Coefficient	0.999866	≥0.995
1800.6	1781.1	1.0110	Slope	0.988331	0.90 - 1.10
1000.5	1008.1	0.9925	Intercept	8.443668	+/-20
499.2	505.3	0.9879			





Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 19, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:48	End Time (MST):	14:00
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999945	<i>≥0.995</i>
826.5	824.5	1.0024	Slope	0.996150	<i>0.90 - 1.10</i>
413.2	417.4	0.9900	Intercept	3.016603	<i>+/-20</i>
207.1	211.0	0.9816			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 19, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:48	End Time (MST):	14:00
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999998	≥ 0.995
799.7	797.1	1.0033	Slope	0.996442	0.90 - 1.10
399.9	399.2	1.0017	Intercept	0.463717	+/-20
200.4	200.6	0.9991			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 19, 2024
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:48	End Time (MST):	14:00
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
429.9	430.3	0.9991	Slope	0.999502	<i>0.90 - 1.10</i>
228.2	229.1	0.9962	Intercept	0.782057	<i>+/-20</i>
127.4	128.4	0.9924			



NO_x Calibration Plot

Date: December 9, 2024

Location: Patricia McInnes



NH₃ Calibration Plot

Date: December 9, 2024

Location: Patricia McInnes






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	December 13, 2024	Last Cal Date:	November 15, 2024
Start time (MST):	10:28	End time (MST):	14:45
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.06	ppm	Cal Gas Exp Date:	March 10, 2031
Cal Gas Cylinder #:	CC320556			
Removed Cal Gas Conc:	50.06	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	3805
Zero Air Gen Model:	API 701H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Serial Number:	1507864683
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000034	0.999948	Backgd or Offset:	2.66	2.70
Calibration intercept:	1.583979	2.064237	Coeff or Slope:	0.847	0.854

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	79.8	799.0	791.4	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	791.4	Previous response	800.6	*% change	-1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	79.8	799.0	799.5	0.999
Mid point	4960	39.9	399.5	404.2	0.988
Low point	4980	20.0	200.2	203.0	0.986
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	79.8	799.0	800.3	0.998
Average Correction Factor:					0.991

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 13, 2024	Previous Calibration:	November 15, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:28	End Time (MST):	14:45
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999962	≥0.995
799.0	799.5	0.9994	Slope	0.999948	0.90 - 1.10
399.5	404.2	0.9883	Intercept	2.064237	+/-30
200.2	203.0	0.9864			



SO2 Calibration Plot

Date: December 13, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	December 10, 2024	Last Cal Date:	November 5, 2024
Start time (MST):	10:11	End time (MST):	N/A
Reason:	As Found		

Calibration Standards

Cal Gas Concentration:	5.25	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC504080			
Removed Cal Gas Conc:	5.25	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3805
ZAG Make/Model:	API T701H		Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 43i LTE	Analyzer serial #:	1180540018
Converter make:	CDN-101	Converter serial #:	551
Analyzer Range	0 - 100 ppb	Converter Temp:	840 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004828		Backgd or Offset:	2.7	N/A
Calibration intercept:	-0.162299		Coeff or Slope:	0.900	N/A

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4925	75.5	79.3	78.0	1.014
As found Mid point	4962	37.7	39.6	39.1	1.007
As found Low point	4981	18.9	19.8	19.3	1.018
New cylinder response					
Baseline Corr As found:	78.2	Prev response:	79.49	*% change:	-1.6%
Baseline Corr 2nd AF pt:	39.3	AF Slope:	0.987227	AF Intercept:	-0.182066
Baseline Corr 3rd AF pt:	19.5	AF Correlation:	0.999983	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					
SO2 Scrubber Check	4920	79.2	792.1		----
Date of last scrubber change:		25-Feb-22		Ave Corr Factor	
Date of last converter efficiency test:		April 22, 2022			


Notes: MPAF done, Sox scrubber failed.

Calibration Performed By: Aswin Sasi Kumar

TRS Calibration Plot

Date: December 10, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Athabasca Valley	Station number: AMS07
Calibration Date: December 11, 2024	Last Cal Date: November 5, 2024
Start time (MST): 10:25	End time (MST): 13:15
Reason: Routine	

Calibration Standards

Cal Gas Concentration: 5.25 ppm	Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC504080	
Removed Cal Gas Conc: 5.25 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: API T700	Serial Number: 3805
ZAG Make/Model: API T701H	Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE	Analyzer serial #: 1180540018
Converter make: CDN-101	Converter serial #: 551
Analyzer Range: 0 - 100 ppb	Converter Temp: 840 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004828	1.004677	Backgd or Offset:	N/A	2.7
Calibration intercept:	-0.162299	-0.182049	Coeff or Slope:	N/A	0.915

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4925	75.5	79.3	79.4	0.999
Mid point	4962	37.7	39.6	39.9	0.993
Low point	4981	18.9	19.9	19.6	1.013
As left zero	5000	0.0	0.0	0.0	----
As left span	4925	75.5	79.3	79.5	0.998
SO2 Scrubber Check	4920	79.2	792.1	0.1	----
Date of last scrubber change:	25-Feb-22		Ave Corr Factor		1.002
Date of last converter efficiency test:	April 22, 2022				

Notes: MPAF done yesterday, SOx scrubber swapped out. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 5, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:25	End Time (MST):	13:15
Analyzer make:	Thermo 43i LTE	Analyzer serial #:	1180540018

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999965	≥ 0.995
79.3	79.4	0.9989	Slope	1.004677	$0.90 - 1.10$
39.6	39.9	0.9927	Intercept	-0.182049	± 3
19.9	19.6	1.0131			



TRS Calibration Plot

Date: December 11, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS 07
Calibration Date:	December 13, 2024	Last Cal Date:	November 15, 2024
Start time (MST):	10:28	End time (MST):	14:45
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	496.0 ppm	CH4 Equiv Conc.	1059.8 ppm
C3H8 Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	496.0 ppm	CH4 Equiv Conc.	1059.8 ppm
Removed C3H8 Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3805
Zero Air Gen model:	Teledyne API T701H	Serial Number:	198

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 12227620777
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.90E-04	2.95E-04	NMHC SP Ratio:	6.44E-05	6.06E-05
CH4 Retention time:	13.6	13.6	NMHC Peak Area:	139822	148545
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.91	17.54	0.964
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.54	Prev response	16.79	*% change	4.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.91	16.79	1.007
Mid point	4960	39.9	8.46	8.31	1.018
Low point	4980	20.0	4.24	4.19	1.012
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.91	16.76	1.009
Average Correction Factor					1.012

Notes: N2 cylinder changed out. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.00	9.68	0.930
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.68	Prev response	8.94	*% change	7.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.00	8.93	1.008
Mid point	4960	39.9	4.50	4.45	1.011
Low point	4980	20.0	2.26	2.25	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.00	8.92	1.009
Average Correction Factor					1.007

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	7.92	7.86	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.86	Prev response	7.86	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	7.92	7.87	1.007
Mid point	4960	39.9	3.96	3.86	1.025
Low point	4980	20.0	1.98	1.94	1.025
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	7.92	7.84	1.010
Average Correction Factor					1.019

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.994272	0.992664
THC Cal Offset:	-0.022911	-0.025510
CH ₄ Cal Slope:	0.995631	0.994014
CH ₄ Cal Offset:	-0.025522	-0.027921
NMHC Cal Slope:	0.993088	0.991590
NMHC Cal Offset:	0.002812	0.003211

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 13, 2024	Previous Calibration:	November 15, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:28	End Time (MST):	14:45
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999967	<i>≥0.995</i>
16.91	16.79	1.0072	Slope	0.992664	<i>0.90 - 1.10</i>
8.46	8.31	1.0177	Intercept	-0.025510	<i>+/-0.5</i>
4.24	4.19	1.0122			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 13, 2024	Previous Calibration:	November 15, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:28	End Time (MST):	14:45
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999902	<i>≥0.995</i>
7.92	7.87	1.0065	Slope	0.994014	<i>0.90 - 1.10</i>
3.96	3.86	1.0246	Intercept	-0.027921	<i>+/-0.5</i>
1.98	1.94	1.0248			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 13, 2024	Previous Calibration:	November 15, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:28	End Time (MST):	14:45
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999990	≥ 0.995
9.00	8.93	1.0077	Slope	0.991590	$0.90 - 1.10$
4.50	4.45	1.0110	Intercept	0.003211	± 0.5
2.26	2.25	1.0009			



NMHC Calibration Plot

Date: December 13, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS 07
Calibration Date:	December 24, 2024	Last Cal Date:	December 13, 2024
Start time (MST):	6:27	End time (MST):	8:53
Reason:	Maintenance	Dipping on CH4 channel	

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	March 10, 2031
CH4 Cal Gas Conc.	496.0 ppm	CH4 Equiv Conc.	1059.8 ppm
C3H8 Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	496.0 ppm	CH4 Equiv Conc.	1059.8 ppm
Removed C3H8 Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3805
Zero Air Gen model:	Teledyne API T701H	Serial Number:	198

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 12227620777
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.95E-04	2.99E-04	NMHC SP Ratio:	6.06E-05	5.75E-05
CH4 Retention time:	13.6	14.0	NMHC Peak Area:	148545	156648
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.91	17.46	0.969
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.46	Prev response	16.76	*% change	4.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.91	16.91	1.000
Mid point	4960	39.9	8.46	8.39	1.008
Low point	4980	20.0	4.24	4.21	1.007
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.91	16.82	1.006
Average Correction Factor					1.005

Notes: Dipping on the CH4 channel. RT and NM bump has slightly moved over.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.00	9.58	0.939
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.58	Prev response	8.93	*% change	6.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.00	8.99	1.001
Mid point	4960	39.9	4.50	4.49	1.003
Low point	4980	20.0	2.26	2.27	0.996
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.00	8.95	1.005
Average Correction Factor					1.000

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	7.92	7.88	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.88	Prev response	7.84	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	7.92	7.92	0.999
Mid point	4960	39.9	3.96	3.90	1.014
Low point	4980	20.0	1.98	1.95	1.020
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	7.92	7.86	1.007
Average Correction Factor					1.011

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.992664	0.999834
THC Cal Offset:	-0.025510	-0.024336
CH ₄ Cal Slope:	0.994014	1.001695
CH ₄ Cal Offset:	-0.027921	-0.027533
NMHC Cal Slope:	0.991590	0.998323
NMHC Cal Offset:	0.003211	0.003197

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 24, 2024	Previous Calibration:	December 13, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	6:27	End Time (MST):	8:53
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999980	<i>≥0.995</i>
16.91	16.91	1.0003	Slope	0.999834	<i>0.90 - 1.10</i>
8.46	8.39	1.0083	Intercept	-0.024336	<i>+/-0.5</i>
4.24	4.21	1.0066			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 24, 2024	Previous Calibration:	December 13, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	6:27	End Time (MST):	8:53
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999926	<i>≥0.995</i>
7.92	7.92	0.9992	Slope	1.001695	<i>0.90 - 1.10</i>
3.96	3.90	1.0141	Intercept	-0.027533	<i>+/-0.5</i>
1.98	1.95	1.0195			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 24, 2024	Previous Calibration:	December 13, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	6:27	End Time (MST):	8:53
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999995	<i>≥0.995</i>
9.00	8.99	1.0011	Slope	0.998323	<i>0.90 - 1.10</i>
4.50	4.49	1.0031	Intercept	0.003197	<i>+/-0.5</i>
2.26	2.27	0.9956			



NMHC Calibration Plot

Date: December 24, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Athabasca Valley
 Station number: AMS 07
 Calibration Date: December 6, 2024
 Last Cal Date: November 22, 2024
 Start time (MST): 10:37
 End time (MST): 15:16
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0033919
 NOX Cal Gas Conc: 60.10 ppm
 Removed Cylinder #: N/A
 Removed Gas NOX Conc: 60.10 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701H
 Cal Gas Expiry Date: January 9, 2032
 NO Cal Gas Conc: 59.90 ppm
 Removed Gas Exp Date: N/A
 Removed Gas NO Conc: 59.90 ppm
 NO gas Diff:
 Serial Number: 3805
 Serial Number: 198

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
AF High point	4933	66.8	803.0	800.3	2.7	800.8	798.0	2.9	1.0027	1.0028
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 805.1 ppb		NO = 800.7 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -0.5%	
Baseline Corr 1st pt	NO _x = 800.8 ppb		NO = 798.1 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -0.3%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1160120024

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000827	1.000855
NO _x Cal Offset:	1.451920	1.591906
NO Cal Slope:	0.998899	0.999328
NO Cal Offset:	1.271961	1.371944
NO ₂ Cal Slope:	1.004263	1.001200
NO ₂ Cal Offset:	0.892899	0.988044

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.187	1.162	NO bkgnd or offset:	8.4	8.3
NOX coeff or slope:	1.005	1.005	NOX bkgnd or offset:	8.6	8.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	229.8	227.4

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
High point	4933	66.8	803.0	800.3	2.7	804.2	800.1	4.2	0.9985	1.0002
Mid point	4966	33.4	401.5	400.2	1.3	405.1	403.0	2.1	0.9912	0.9930
Low point	4983	16.7	200.7	200.1	0.7	203.4	202.0	1.4	0.9870	0.9905
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
As left span	4933	66.8	803.0	399.6	403.4	805.2	399.6	405.6	0.9972	1.0000
Average Correction Factor									0.9922	0.9946

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	798.1	398.2	402.6	403.6	0.9975	100.3%
Mid GPT point	798.1	608.3	192.5	194.2	0.9911	100.9%
Low GPT point	798.1	702.9	97.9	99.8	0.9807	102.0%
Average Correction Factor					0.9897	101.0%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 22, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:37	End Time (MST):	15:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999981	<i>≥0.995</i>
803.0	804.2	0.9985	Slope	1.000855	<i>0.90 - 1.10</i>
401.5	405.1	0.9912	Intercept	1.591906	<i>+/-20</i>
200.7	203.4	0.9870			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 22, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:37	End Time (MST):	15:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999978	<i>≥0.995</i>
402.6	403.6	0.9975	Slope	1.001200	<i>0.90 - 1.10</i>
192.5	194.2	0.9911	Intercept	0.988044	<i>+/-20</i>
97.9	99.8	0.9807			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 22, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:37	End Time (MST):	15:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999982	<i>≥0.995</i>
800.3	800.1	1.0002	Slope	0.999328	<i>0.90 - 1.10</i>
400.2	403.0	0.9930	Intercept	1.371944	<i>+/-20</i>
200.1	202.0	0.9905			



NO_x Calibration Plot

Date: December 6, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS07
Calibration Date:	December 23, 2024	Last Cal Date:	November 1, 2024
Start time (MST):	9:04	End time (MST):	11:47
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3805
Calibrator Make/Model:	T700	Serial Number:	198
ZAG Make/Model:	T701H		

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997857	1.001223	Backgd or Offset:	-2.0	-2.0
Calibration intercept:	1.000000	1.996000	Coeff or Slope:	1.560	1.560

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	0.8	----
As found High point	5000	1652.0	400.0	401.4	0.999
As found Mid point					
As found Low point					
Baseline Corr As found:	400.6	Previous response	400.1	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	1.5	----
High point	5000	1654.3	400.0	402.3	0.994
Mid point	5000	1152.0	200.0	202.3	0.989
Low point	5000	913.0	100.0	102.7	0.973
As left zero	5000	1652.9	0.0	1.7	----
As left span	5000	1582.6	400.0	402.4	0.994
Average Correction Factor					0.985

Notes: No Maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	December 23, 2024	Previous Calibration:	November 1, 2024
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:04	End Time (MST):	11:47
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.5	----	Correlation Coefficient	0.999992	≥0.995
400.0	402.3	0.9943	Slope	1.001223	0.90 - 1.10
200.0	202.3	0.9886	Intercept	1.996000	+/- 5
100.0	102.7	0.9733			



O₃ Calibration Plot

Date: December 23, 2024

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Athabasca Valley Station number: AMS 07
Calibration Date: December 23, 2024 Last Cal Date: November 22, 2024
Start time (MST): 8:41 End time (MST): 9:05
Analyzer Make: API T640 S/N: 645
Particulate Fraction: PM2.5
Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-13.4	-14	-13.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	730.7	729.7	730.7	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.94	4.93	4.94	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39	----	39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	17.9	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 6, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: November 22, 2024
Date Disposable Filter Changed: August 16, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 8, 2024
Date RH/T Sensor Cleaned: July 8, 2024

Notes: No adjustments done.

Calibration by: Melissa Lemay



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS 07
Calibration Date:	December 11, 2024	Last Cal Date:	November 12, 2024
Start time (MST):	13:15	End time (MST):	16:03
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	2,953	ppm	Cal Gas Exp Date: September 30, 2029
Cal Gas Cylinder #:	T1TWKRN		
Removed Cal Gas Conc:	2,953	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T750		Serial Number: 282
ZAG Make/Model:	Teledyne API 751H		Serial Number: 321

Analyzer Information

Analyzer make:	Thermo 48i-TLE	Analyzer serial #:	1408761381
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998781	0.996926	Backgd or Offset:	5.219	5.219
Calibration intercept:	0.102000	0.191983	Coeff or Slope:	1.068	1.068

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4932	67.8	40.0	40.0	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	39.92	Prev response:	40.10	*% change:	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4932	67.8	40.0	40.0	1.001
Mid point	4966	33.9	20.0	20.3	0.986
Low point	4983	16.9	10.0	10.2	0.977
As left zero	5000	0.0	0.0	0.1	----
As left span	4932	67.8	40.0	40.0	1.001
Average Correction Factor					0.988

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


CO Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 12, 2024
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	13:15	End Time (MST):	16:03
Analyzer make:	Thermo 48i-TLE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.1	----	Correlation Coefficient	0.999938	≥0.995
40.0	40.0	1.0006	Slope	0.996926	0.90 - 1.10
20.0	20.3	0.9863	Intercept	0.191983	+/-1.5
10.0	10.2	0.9766			



CO Calibration Plot

Date: December 11, 2024

Location: Athabasca Valley






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Fort Chipewyan	Station number: AMS08
Calibration Date:	December 12, 2024	Last Cal Date: November 15, 2024
Start time (MST):	13:15	End time (MST): 16:04
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	49.84 ppm	Cal Gas Exp Date: January 6, 2030
Cal Gas Cylinder #:	CC196697	
Removed Cal Gas Conc:	49.84 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA	Diff between cyl:
Calibrator Model:	Teledyne API T700	Serial Number: 3252
Zero Air Gen Model:	Teledyne API T701	Serial Number: 135

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Serial Number: 1136451241
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994277	0.995662	Backgd or Offset:	1.8	1.8
Calibration intercept:	0.297009	0.436778	Coeff or Slope:	0.959	0.901

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4920	80.3	800.4	780.6	1.025
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	781.0	Previous response	796.1	*% change	-1.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4920	80.3	800.4	797.4	1.004
Mid point	4960	40.2	400.7	398.6	1.005
Low point	4980	20.1	200.4	201.3	0.995
As left zero	5000	0.0	0.0	-0.3	----
As left span	4920	80.3	800.4	801.2	0.999
Average Correction Factor:					1.001

Notes: Changed out inlet filter after as founds. Adjustments made to span.

Calibration Performed By: Morgan Voyageur, Sabian Voyageur



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 15, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	13:15	End Time (MST):	16:04
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1136451241

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999992	≥0.995
800.4	797.4	1.0037	Slope	0.995662	0.90 - 1.10
400.7	398.6	1.0053	Intercept	0.436778	+/-30
200.4	201.3	0.9953			



SO2 Calibration Plot

Date: December 12, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS 08
Calibration Date:	December 11, 2024	Last Cal Date:	November 29, 2024
Start time (MST):	9:02	End time (MST):	13:15
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.97 ppm	Cal Gas Exp Date:	February 9, 2024
Cal Gas Cylinder #:	EY0002276		
Removed Cal Gas Conc:	4.97 ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	3810
ZAG Make/Model:	Teledyne API T701	Serial Number:	135

Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744
Converter make:	CDN-101	Converter serial #:	14639
Analyzer Range	0 - 100 ppb	Converter Temp:	833 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.985428	0.997567	Backgd or Offset:	1.4	1.4
Calibration intercept:	0.198526	0.298821	Coeff or Slope:	0.742	0.742

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4920	80.5	80.0	79.2	1.014
As found Mid point	4960	40.2	40.0	39.9	1.009
As found Low point	4980	20.1	20.0	19.8	1.025
New cylinder response					
Baseline Corr As found:	78.9	Prev response:	79.04	*% change:	-0.2%
Baseline Corr 2nd AF pt:	39.6	AF Slope:	0.987284	AF Intercept:	0.258569
Baseline Corr 3rd AF pt:	19.5	AF Correlation:	0.999978	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	80.5	80.0	80.0	1.000
Mid point	4960	40.2	40.0	40.4	0.989
Low point	4980	20.1	20.0	20.2	0.989
As left zero	5000	0.0	0.0	0.3	----
As left span	4920	80.5	80.0	80.4	0.995
SO2 Scrubber Check	4919.7	80.3	803.0	0.2	----
Date of last scrubber change:		March 7, 2022		Ave Corr Factor	0.993
Date of last converter efficiency test:		March 15, 2022		100.7% efficiency	

Notes: Changed inlet filter after multipoint as founds. SO2 scrubber check passed no adjustment made.

Calibration Performed By: Morgan Voyageur, Sabian Voyageur,



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 29, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	9:02	End Time (MST):	13:15
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999977	≥ 0.995
80.0	80.0	1.0001	Slope	0.997567	$0.90 - 1.10$
40.0	40.4	0.9890	Intercept	0.298821	± 3
20.0	20.2	0.9891			



TRS Calibration Plot

Date: December 11, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan
 Station number: AMS 08
 Calibration Date: December 12, 2024
 Last Cal Date: November 16, 2024
 Start time (MST): 8:55
 End time (MST): 13:17
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0046831
 NOX Cal Gas Conc: 60.20 ppm
 Removed Cylinder #: DT0046831
 Removed Gas NOX Conc: 60.20 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701H
 Cal Gas Expiry Date: January 9, 2032
 NO Cal Gas Conc: 60.00 ppm
 Removed Gas Exp Date: January 9, 2032
 Removed Gas NO Conc: 60.00 ppm
 NO gas Diff:
 Serial Number: 3252
 Serial Number: 135

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
AF High point	4933	66.7	803.1	800.4	2.7	820.1	815.0	5.1	0.9792	0.9819
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 798.5 ppb		NO = 797.1 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 2.6%	
Baseline Corr 1st pt	NO _x = 820.2 ppb		NO = 815.2 ppb			<u>As Found Statistics</u>		*Percent Change	NO = 2.2%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 4460

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.992491	0.997601
NO _x Cal Offset:	1.414756	-0.805395
NO Cal Slope:	0.996043	0.997455
NO Cal Offset:	-0.145628	-1.165086
NO ₂ Cal Slope:	1.010685	1.007149
NO ₂ Cal Offset:	0.893977	1.334457

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.195	1.168	NO bkgnd or offset:	1.2	0.9
NOX coeff or slope:	1.194	1.167	NOX bkgnd or offset:	1.2	0.6
NO2 coeff or slope:	2.100	1.300	Reaction cell Press:	2.1	1.2

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	-0.4	0.4	----	----
High point	4933	66.7	803.1	800.4	2.7	800.7	797.4	2.9	1.0030	1.0038
Mid point	4967	33.3	400.9	399.6	1.3	399.0	397.7	1.3	1.0048	1.0047
Low point	4983	16.7	201.1	200.4	0.7	198.8	197.5	2.3	1.0114	1.0147
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.2	----	----
As left span	4933	66.7	803.1	383.5	419.6	797.8	383.5	414.2	1.0067	1.0000
Average Correction Factor									1.0064	1.0077

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.4	----	----
High GPT point	796.2	377.8	421.1	425.1	0.9905	101.0%
Mid GPT point	796.2	589.9	209.0	211.8	0.9866	101.4%
Low GPT point	796.2	693.5	105.4	108.7	0.9693	103.2%
Average Correction Factor					0.9822	101.8%

Notes: Changed inlet filter after as founds. Made adjustments to span high point.

Calibration Performed By: Morgan Voyageur, Sabian Voyageur



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 16, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	8:55	End Time (MST):	13:17
Analyzer make:	Teledyne API T200	Analyzer serial #:	4460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999995	<i>≥0.995</i>
803.1	800.7	1.0030	Slope	0.997601	<i>0.90 - 1.10</i>
400.9	399.0	1.0048	Intercept	-0.805395	<i>+/-20</i>
201.1	198.8	1.0114			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 16, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	8:55	End Time (MST):	13:17
Analyzer make:	Teledyne API T200	Analyzer serial #:	4460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999974	≥ 0.995
421.1	425.1	0.9905	Slope	1.007149	0.90 - 1.10
209.0	211.8	0.9866	Intercept	1.334457	+/-20
105.4	108.7	0.9693			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 16, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	8:55	End Time (MST):	13:17
Analyzer make:	Teledyne API T200	Analyzer serial #:	4460

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.4	----	Correlation Coefficient	0.999994	≥0.995
800.4	797.4	1.0038	Slope	0.997455	0.90 - 1.10
399.6	397.7	1.0047	Intercept	-1.165086	+/-20
200.4	197.5	1.0147			



NO_x Calibration Plot

Date: December 12, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS 08
Calibration Date:	December 10, 2024	Last Cal Date:	November 5, 2024
Start time (MST):	8:42	End time (MST):	11:22
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3810
Calibrator Make/Model:	Teledyne API T700	Serial Number:	135
ZAG Make/Model:	Teledyne API T701		

Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #:	3872
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.987743	0.979971	Backgd or Offset:	-2.2	-2.2
Calibration intercept:	1.220000	1.480000	Coeff or Slope:	0.982	0.982

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.9	----
As found High point	5000	913.0	400.0	394.6	1.016
As found Mid point					
As found Low point					
Baseline Corr As found:	393.7	Previous response	396.3	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	1.3	----
High point	5000	914.7	400.0	393.3	1.017
Mid point	5000	786.4	200.0	197.8	1.011
Low point	5000	701.3	100.0	99.5	1.005
As left zero	5000	0.0	0.0	1.3	----
As left span	5000	914.7	400.0	393.3	1.017
Average Correction Factor					1.011

Notes: Changed out inlet filter after as found. No adjustments made.

Calibration Performed By: Morgan Voyageur, Sabian Voyageur



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	December 10, 2024	Previous Calibration:	November 5, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	8:42	End Time (MST):	11:22
Analyzer make:	Teledyne API T400	Analyzer serial #:	3872

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.3	----	Correlation Coefficient	0.999998	≥0.995
400.0	393.3	1.0170	Slope	0.979971	0.90 - 1.10
200.0	197.8	1.0111	Intercept	1.480000	+/- 5
100.0	99.5	1.0050			



O₃ Calibration Plot

Date: December 10, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
 Calibration Date: December 10, 2024 Last Cal Date: October 22, 2024
 Start time (MST): 11:15 End time (MST): 11:45

Analyzer Make: Teledyne API T640 S/N: 319
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 14719
 Temp/RH standard: Alicat FP-25BT S/N: 14719

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-22.10	-23.9	-22.10	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	751.90	753.6	751.90	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.97	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	42%		39%	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	4.40	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: 10-Jun-24
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.00	11.00	11.00	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: September 18, 2024
 Date Disposable Filter Changed: September 18, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 29, 2024
 Date RH/T Sensor Cleaned: August 29, 2024

Notes:

No adjustment made

Calibration by: Morgan Voyageur, Sabian Voyageur



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS 08
Calibration Date:	December 10, 2024	Last Cal Date:	November 5, 2024
Start time (MST):	11:25	End time (MST):	14:01
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,030	ppm	Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #:	ALM014846		
Removed Cal Gas Conc:	3,030	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 3810
ZAG Make/Model:	Teledyne API T701H		Serial Number: 135

Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3505
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008286	1.005457	Backgd or Offset:	-0.015	-0.015
Calibration intercept:	-0.085114	0.234888	Coeff or Slope:	1.005	1.005

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4934	66.7	40.4	40.9	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	40.67	Prev response:	40.66	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* => +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4934	66.7	40.4	40.8	0.990
Mid point	4966.7	33.3	20.2	20.6	0.980
Low point	4983.3	16.7	10.1	10.4	0.969
As left zero	5000	0.0	0.0	0.2	----
As left span	2960	40.0	40.4	40.8	0.989
Average Correction Factor					0.980

Notes: Changed inlet filter after as found. No adjustments made

Calibration Performed By: Morgan Voyageur, Sabian Voyageur.



Wood Buffalo Environmental Association


CO Calibration Summary

Station Information

Calibration Date:	December 10, 2024	Previous Calibration:	November 5, 2024
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	11:25	End Time (MST):	14:01
Analyzer make:	Teledyne API T300	Analyzer serial #:	3505

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.2	----	Correlation Coefficient	0.999989	≥0.995
40.4	40.8	0.9898	Slope	1.005457	0.90 - 1.10
20.2	20.6	0.9801	Intercept	0.234888	+/-1.5
10.1	10.4	0.9694			



CO Calibration Plot

Date: December 10, 2024

Location: Fort Chipewyan





Wood Buffalo Environmental Association

CO₂ Calibration Report

Station Information

Station Name:	Fort Chipewyan	Station number:	AMS 08
Calibration Date:	December 10, 2024	Last Cal Date:	November 18, 2024
Start time (MST):	14:08	End time (MST):	16:51
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	60,220	ppm	Cal Gas Exp Date:	December 1, 2028
Cal Gas Cylinder #:	ALM014846			
Removed Cal Gas Conc:	60,220	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3810
N2 Gen Make/Model:	Peak Scientific		Serial Number:	135

Analyzer Information

Analyzer make:	Teledyne API T360	Analyzer serial #:	289
Analyzer Range	0 - 2,000 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008701	1.003841	Backgd or Offset:	-0.016	-0.016
Calibration intercept:	-6.680000	-7.440000	Coeff or Slope:	1.034	1.034

CO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	3000	0.0	0.0	0.2	----
As found High Point	2920	80.0	1605.9	1614.0	0.995
As found Mid Point					
As found Low Point					
New cylinder response					
Baseline Corr As found:	1613.8	Prev response:	1613.2	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	3000	0.0	0.0	0.5	----
High point	2920	80.0	1605.9	1612.7	0.996
Mid point	2960	40.0	802.9	782.2	1.027
Low point	2980	20.0	401.5	395.9	1.014
As left zero	3000	0.0	0.0	-0.1	----
As left span	2960	40.0	802.9	778.5	1.031
Average Correction Factor					1.012

Notes: Changed inlet filter after as found, No adjustments made.

Calibration Performed By: Morgan Voyageur, Sabian Voyageur



Wood Buffalo Environmental Association


CO₂ Calibration Summary

Station Information

Calibration Date	December 10, 2024	Previous Calibration	November 18, 2024
Station Name	Fort Chipewyan	Station Number	AMS 08
Start Time (MST)	14:08	End Time (MST)	16:51
Analyzer make	Teledyne API T360	Analyzer serial #	289

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	≥0.995
1605.9	1612.7	0.9958	Slope	0.90 - 1.10
802.9	782.2	1.0265	Intercept	+/-20
401.5	395.9	1.0141		



CO₂ Calibration Plot

Date: December 10, 2024

Location: Fort Chipewyan






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Barge Landing	Station number: AMS 09
Calibration Date:	December 3, 2024	Last Cal Date: November 7, 2024
Start time (MST):	10:05	End time (MST): 13:05
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	49.96 ppm	Cal Gas Exp Date: January 5, 2025
Cal Gas Cylinder #:	CC151285	
Removed Cal Gas Conc:	49.96 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA	Diff between cyl:
Calibrator Model:	API T700	Serial Number: 3812
Zero Air Gen Model:	APIT701	Serial Number: 4888

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 1118148498
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994978	1.002005	Backgd or Offset:	10.9	10.9
Calibration intercept:	0.085896	0.047698	Coeff or Slope:	0.963	0.963

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4919	80.2	801.5	802.6	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	802.8	Previous response	797.5	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4919	80.2	801.5	803.5	0.997
Mid point	4960	40.1	400.7	400.6	1.000
Low point	4980	20.0	199.8	200.7	0.996
As left zero	5000	0.0	0.0	-0.1	----
As left span	4919	80.2	801.5	803.0	0.998
Average Correction Factor:					0.998

Notes: Inlet filter changed after as founds.

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 7, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:05	End Time (MST):	13:05
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999997	≥0.995
801.5	803.5	0.9975	Slope	1.002005	0.90 - 1.10
400.7	400.6	1.0002	Intercept	0.047698	+/-30
199.8	200.7	0.9957			



SO2 Calibration Plot

Date: December 3, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Barge Landing	Station number:	AMS 09
Calibration Date:	December 5, 2024	Last Cal Date:	November 12, 2024
Start time (MST):	9:55	End time (MST):	14:25
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.17	ppm	Cal Gas Exp Date:	August 22, 2026
Cal Gas Cylinder #:	CC511415			
Removed Cal Gas Conc:	5.17	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3812
ZAG Make/Model:	API T701		Serial Number:	4888

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320
Converter make:	CDN-101	Converter serial #:	519
Analyzer Range	0 - 100 ppb	Converter Temp:	830 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001976	1.001834	Backgd or Offset:	2.860	2.880
Calibration intercept:	0.099442	0.179444	Coeff or Slope:	1.182	1.182

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4923	77.4	80.0	80.0	1.002
As found Mid point	4961	38.7	40.0	40.2	0.998
As found Low point	4981	19.3	20.0	20.1	0.998
New cylinder response					
Baseline Corr As found:	79.9	Prev response:	80.30	*% change:	-0.5%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	0.998122	AF Intercept:	0.159370
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999996	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4923	77.4	80.0	80.3	0.997
Mid point	4961	38.7	40.0	40.4	0.991
Low point	4981	19.3	20.0	20.2	0.988
As left zero	5000	0.0	0.0	0.1	----
As left span	4923	77.4	80.0	80.6	0.993
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.992
Date of last converter efficiency test:					

Notes: Sample inlet filter changed after multipoint as founds. SO2 scrubber check completed after calibrator zero. No adjustments made.

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 12, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:55	End Time (MST):	14:25
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999993	≥ 0.995
80.0	80.3	0.9968	Slope	1.001834	$0.90 - 1.10$
40.0	40.4	0.9907	Intercept	0.179444	± 3
20.0	20.2	0.9881			



TRS Calibration Plot

Date: December 5, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Barge Landing	Station number:	AMS 09
Calibration Date:	December 3, 2024	Last Cal Date:	November 7, 2024
Start time (MST):	10:05	End time (MST):	13:05
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
Zero Air Gen model:	APIT701	Serial Number:	4888

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320038
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.29E-04	2.41E-04	NMHC SP Ratio:	5.49E-05	5.82E-05
CH4 Retention time:	14.0	14.0	NMHC Peak Area:	166530	157157
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.02	----
As found High point	4919	80.2	17.12	16.15	1.062
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.13	Prev response	17.25	*% change	-7.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	17.12	17.09	1.002
Mid point	4960	40.1	8.56	8.50	1.007
Low point	4980	20.0	4.27	4.27	0.999
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	17.12	17.09	1.002
Average Correction Factor					1.002

Notes: Changed the inlet filter after as founds. Span adjusted



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	9.14	8.56	1.067
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.56	Prev response	9.21	*% change	-7.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	9.14	9.13	1.001
Mid point	4960	40.1	4.57	4.52	1.010
Low point	4980	20.0	2.28	2.28	1.000
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	9.14	9.15	0.999
Average Correction Factor					1.004

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.02	----
As found High point	4919	80.2	7.98	7.59	1.056
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.56	Prev response	8.04	*% change	-6.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	80.2	7.98	7.96	1.002
Mid point	4960	40.1	3.99	3.98	1.003
Low point	4980	20.0	1.99	2.00	0.997
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	7.98	7.94	1.006
Average Correction Factor					1.001

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.004139	0.997813
THC Cal Offset:	0.055488	-0.004153
CH ₄ Cal Slope:	1.002924	0.997172
CH ₄ Cal Offset:	0.029078	0.003659
NMHC Cal Slope:	1.004963	0.998236
NMHC Cal Offset:	0.026610	-0.007012

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 7, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:05	End Time (MST):	13:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999989	≥ 0.995
17.12	17.09	1.0017	Slope	0.997813	$0.90 - 1.10$
8.56	8.50	1.0068	Intercept	-0.004153	± 0.5
4.27	4.27	0.9989			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 7, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:05	End Time (MST):	13:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	≥0.995
7.98	7.96	1.0024	Slope	0.90 - 1.10
3.99	3.98	1.0029	Intercept	+/-0.5
1.99	2.00	0.9972		





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 7, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:05	End Time (MST):	13:05
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999973	<i>≥0.995</i>
9.14	9.13	1.0012	Slope	0.998236	<i>0.90 - 1.10</i>
4.57	4.52	1.0101	Intercept	-0.007012	<i>+/-0.5</i>
2.28	2.28	1.0000			



NMHC Calibration Plot

Date: December 3, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Barge Landing	Station number:	AMS 09
Calibration Date:	December 23, 2024	Last Cal Date:	December 3, 2024
Start time (MST):	10:15	End time (MST):	12:00
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC151285	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	497.6 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
Zero Air Gen model:	APIT701	Serial Number:	4888

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320038
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.41E-04	2.41E-04	5.82E-05	5.82E-05
CH4 Retention time:	14.0	14.0	157157	157157
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.02	----
As found High point	4919	80.2	17.12	16.06	1.068
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.03	Prev response	17.08	*% change	-6.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	17.12	15.96	1.073

Average Correction Factor

Notes: N2 Cylinder Change



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4919	80.2	9.14	8.58	1.064
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.58	Prev response	9.11	*% change	-6.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	9.14	8.52	1.073
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.02	----
As found High point	4919	80.2	7.98	7.47	1.072
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.45	Prev response	7.96	*% change	-6.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	80.2	7.98	7.44	1.073
Average Correction Factor					

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.997813	
THC Cal Offset:	-0.004153	
CH ₄ Cal Slope:	0.997172	
CH ₄ Cal Offset:	0.003659	
NMHC Cal Slope:	0.998236	
NMHC Cal Offset:	-0.007012	

Calibration Performed By: Devin Russell

NMHC Calibration Plot

Date: December 23, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Barge Landing
 Station number: AMS 09
 Calibration Date: December 9, 2024
 Last Cal Date: November 15, 2024
 Start time (MST): 10:05
 End time (MST): 15:00
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1KDH
 NOX Cal Gas Conc: 47.38 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 47.38 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: Api T701
 Cal Gas Expiry Date: November 17, 2026
 NO Cal Gas Conc: 46.94 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 46.94 ppm
 NO gas Diff:
 Serial Number: 3812
 Serial Number: 4888

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.2	----	----
AF High point	4915	85.3	808.3	800.7	7.5	798.5	790.0	8.5	1.0118	1.0135
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 806.6 ppb	NO = 800.1 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -1.0%
Baseline Corr 1st pt	NO _x = 798.8 ppb	NO = 790.1 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -1.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1426262593

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997622	0.997861
NO _x Cal Offset:	0.317920	0.658322
NO Cal Slope:	0.999938	0.998554
NO Cal Offset:	-0.643839	-0.483887
NO ₂ Cal Slope:	0.999704	1.003546
NO ₂ Cal Offset:	-0.753159	0.547567

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.095	1.107	NO bkgnd or offset:	10.0	10.1
NOX coeff or slope:	1.000	0.999	NOX bkgnd or offset:	10.3	10.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	176.2	179.5

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	0.1	-0.2	----	----
High point	4915	85.3	808.3	800.7	7.5	806.6	799.3	7.3	1.0021	1.0018
Mid point	4957	42.6	403.7	400.0	3.7	404.4	398.9	5.5	0.9983	1.0027
Low point	4979	21.3	201.8	200.0	1.9	202.6	198.4	4.3	0.9962	1.0078
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
As left span	4915	85.3	808.3	421.1	387.2	806.8	421.1	385.7	1.0018	1.0000
Average Correction Factor									0.9988	1.0041

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	795.4	419.6	383.3	384.8	0.9961	100.4%
Mid GPT point	795.4	610.7	192.2	193.9	0.9913	100.9%
Low GPT point	795.4	703.9	99.0	100.6	0.9842	101.6%
Average Correction Factor					0.9905	101.0%

Notes:

Inlet filter changed after as founds. Span adjusted.

Calibration Performed By:

Devin Russell



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 15, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:05	End Time (MST):	15:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999994	<i>≥0.995</i>
808.3	806.6	1.0021	Slope	0.997861	<i>0.90 - 1.10</i>
403.7	404.4	0.9983	Intercept	0.658322	<i>+/-20</i>
201.8	202.6	0.9962			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 15, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:05	End Time (MST):	15:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.2	----	Correlation Coefficient	0.999982	<i>≥0.995</i>
383.3	384.8	0.9961	Slope	1.003546	<i>0.90 - 1.10</i>
192.2	193.9	0.9913	Intercept	0.547567	<i>+/-20</i>
99.0	100.6	0.9842			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 15, 2024
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:05	End Time (MST):	15:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999997	<i>≥0.995</i>
800.7	799.3	1.0018	Slope	0.998554	<i>0.90 - 1.10</i>
400.0	398.9	1.0027	Intercept	-0.483887	<i>+/-20</i>
200.0	198.4	1.0078			



NO_x Calibration Plot

Date: December 9, 2024

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Barge Landing Station number: AMS 09
 Calibration Date: December 9, 2024 Last Cal Date: November 26, 2024
 Start time (MST): 11:45 End time (MST): 12:50

Analyzer Make: API T640 S/N: 844
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388746
 Temp/RH standard: Alicat FP-25BT S/N: 388746

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-6.40	-6.40	-6.40	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	731.10	737.10	731.10	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.98	5.12	4.98	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	40	----	40	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.00	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: October 18, 2024
 Date Disposable Filter Changed: October 18, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 18, 2024
 Date RH/T Sensor Cleaned: October 18, 2024

Notes: Verified flow, pressure, temperature and pump power. No adjustment needed. Leak check passed.

Calibration by: Devin Russell



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Barge Landing Station number: AMS 09
 Calibration Date: December 29, 2024 Last Cal Date: December 9, 2024
 Start time (MST): 12:41 End time (MST): 13:15

Analyzer Make: API T640 S/N: 844
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
 Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-6.80	-6.40	-6.80	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	721.80	725.80	721.80	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.05	5.02	5.05	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39	----	39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	6.30	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.20	10.20	11.00	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: December 29, 2024
 Date Disposable Filter Changed: December 29, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 18, 2024
 Date RH/T Sensor Cleaned: October 18, 2024

Notes: Here to address cyclic readings on PM2.5 channel. Verified flow, temp and pressure. Leak check passed. Chamber cleaned. DFU filter changed out. PMT adjusted.

Calibration by: Aswin Sasi Kumar




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS11 LOWER CAMP DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Lower Camp	Station number:	AMS 11
Calibration Date:	December 11, 2024	Last Cal Date:	November 18, 2024
Start time (MST):	11:50	End time (MST):	15:55
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.25	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC2216			
Removed Cal Gas Conc:	49.25	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	3807
Zero Air Gen Model:	Teledyne API T701		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	100841398
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005641	0.994133	Backgd or Offset:	15.8	15.6
Calibration intercept:	-0.988705	-1.481721	Coeff or Slope:	0.977	0.977

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4932	81.4	799.6	790.7	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	790.6	Previous response	803.2	*% change	-1.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4932	81.4	799.6	794.5	1.006
Mid point	4959	40.7	400.9	395.5	1.014
Low point	4981	20.4	200.9	197.3	1.018
As left zero	5000	0.0	0.0	0.3	----
As left span	4932	81.4	799.6	792.7	1.009
Average Correction Factor:					1.013

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 18, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	11:50	End Time (MST):	15:55
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999981	≥0.995
799.6	794.5	1.0065	Slope	0.994133	0.90 - 1.10
400.9	395.5	1.0137	Intercept	-1.481721	+/-30
200.9	197.3	1.0182			



SO2 Calibration Plot

Date: December 11, 2024

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	Lower Camp	Station number:	AMS 11
Calibration Date:	December 10, 2024	Last Cal Date:	November 26, 2024
Start time (MST):	10:59	End time (MST):	15:10
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.43	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC501097			
Removed Cal Gas Conc:	5.43	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3807
ZAG Make/Model:	API T701H		Serial Number:	196

Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	1203169745
Converter make:	Global G150	Converter serial #:	2022-223
Analyzer Range	0 - 100 ppb	Converter Temp:	325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.991875	0.998327	Backgd or Offset:	2.6
Calibration intercept:	-0.183931	0.015156	Coeff or Slope:	0.769

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4926	73.6	79.9	79.7	0.999
As found Mid point	4963	36.8	40.0	39.8	0.996
As found Low point	4982	18.6	20.2	20.1	0.990
New cylinder response					
Baseline Corr As found:	80.0	Prev response:	79.09	*% change:	1.1%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	1.000335	AF Intercept:	-0.205188
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999993	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4926	73.6	79.9	79.7	1.003
Mid point	4963	36.8	40.0	40.1	0.996
Low point	4982	18.6	20.2	20.2	1.000
As left zero	5000	0.0	0.0	0.0	----
As left span	4926	73.6	79.9	79.6	1.004
SO2 Scrubber Check	4935	81.5	812.3	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.000
Date of last converter efficiency test:					

Notes: Changed sample inlet filter after as founds. Ran scrubber check after calibrator zero and it passed.
No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	December 10, 2024	Previous Calibration:	November 26, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:59	End Time (MST):	15:10
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999982	≥0.995
79.9	79.7	1.0028	Slope	0.998327	0.90 - 1.10
40.0	40.1	0.9965	Intercept	0.015156	+/-3
20.2	20.2	0.9997			



H₂S Calibration Plot

Date: December 10, 2024

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH4 / NMHC Calibration Report

Station Information

Station Name:	Lower Camp	Station number:	AMS 11
Calibration Date:	December 11, 2024	Last Cal Date:	November 18, 2024
Start time (MST):	11:50	End time (MST):	15:55
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC2216	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	502.0 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	205.5 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	502.0 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	205.5 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
Zero Air Gen model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1505164381
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.63E-04	2.90E-04	NMHC SP Ratio:	4.48E-05	4.65E-05
CH4 Retention time:	15.4	15.8	NMHC Peak Area:	204852	197066
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	81.4	17.33	16.90	1.025
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.90	Prev response	17.29	*% change	-2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	81.4	17.33	17.27	1.003
Mid point	4959	40.7	8.69	8.62	1.008
Low point	4981	20.4	4.35	4.32	1.008
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	81.4	17.33	17.33	1.000

Average Correction Factor 1.007

Notes: Changed sample inlet filter and both Hydrogen and Nitrogen cylinders after as founds. Adjusted both zero and span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	81.4	9.18	9.11	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.11	Prev response	9.16	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	81.4	9.18	9.14	1.003
Mid point	4959	40.7	4.60	4.55	1.010
Low point	4981	20.4	2.31	2.28	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	81.4	9.18	9.15	1.003
Average Correction Factor					1.008

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	81.4	8.15	7.78	1.047
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.78	Prev response	8.13	*% change	-4.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	81.4	8.15	8.13	1.003
Mid point	4959	40.7	4.09	4.06	1.006
Low point	4981	20.4	2.05	2.04	1.006
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	81.4	8.15	8.19	0.996
Average Correction Factor					1.005

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999787	0.996781
THC Cal Offset:	-0.030112	-0.016298
CH ₄ Cal Slope:	1.000126	0.996965
CH ₄ Cal Offset:	-0.021159	-0.004369
NMHC Cal Slope:	0.999609	0.996730
NMHC Cal Offset:	-0.008955	-0.012128

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 18, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	11:50	End Time (MST):	15:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999993	<i>≥0.995</i>
17.33	17.27	1.0034	Slope	0.996781	<i>0.90 - 1.10</i>
8.69	8.62	1.0080	Intercept	-0.016298	<i>+/-0.5</i>
4.35	4.32	1.0083			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 18, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	11:50	End Time (MST):	15:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
8.15	8.13	1.0032	Slope	0.996965	<i>0.90 - 1.10</i>
4.09	4.06	1.0055	Intercept	-0.004369	<i>+/-0.5</i>
2.05	2.04	1.0062			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 18, 2024
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	11:50	End Time (MST):	15:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1505164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999986	<i>≥0.995</i>
9.18	9.14	1.0035	Slope	0.996730	<i>0.90 - 1.10</i>
4.60	4.55	1.0102	Intercept	-0.012128	<i>+/-0.5</i>
2.31	2.28	1.0101			



NMHC Calibration Plot

Date: July 18, 2024

Location: Lower Camp






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13 FORT MCKAY SOUTH DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	December 5, 2024	Last Cal Date:	November 1, 2024
Start time (MST):	10:20	End time (MST):	12:55
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.55	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC260812			
Removed Cal Gas Conc:	50.55	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	2448
Zero Air Gen Model:	Teledyne API T701		Serial Number:	1118

Analyzer Information

Analyzer make:	Teledyne API T100	Serial Number:	599
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003744	1.002885	Backgd or Offset:	99.5	99.5
Calibration intercept:	-3.278551	-2.978200	Coeff or Slope:	0.697	0.697

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4921	79.1	799.7	799.5	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	799.9	Previous response	799.4	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.4	----
High point	4921	79.1	799.7	800.2	0.999
Mid point	4961	39.5	399.3	396.5	1.007
Low point	4980	19.8	200.2	195.0	1.027
As left zero	5000	0.0	0.0	-0.1	----
As left span	4921	79.1	799.7	800.0	1.000
Average Correction Factor:					1.011

Notes: Sample inlet filters changed after as founds. No adjustment made.

Calibration Performed By: Kelly Baragar



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 1, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:20	End Time (MST):	12:55
Analyzer make:	Teledyne API T100	Analyzer serial #:	599

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.4	----	Correlation Coefficient	0.999952	≥0.995
799.7	800.2	0.9994	Slope	1.002885	0.90 - 1.10
399.3	396.5	1.0071	Intercept	-2.978200	+/-30
200.2	195.0	1.0266			



SO2 Calibration Plot

Date: December 5, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	December 13, 2024	Last Cal Date:	November 13, 2024
Start time (MST):	10:09	End time (MST):	14:34
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.34	ppm	Cal Gas Exp Date:	January 1, 2025
Cal Gas Cylinder #:	CC500241			
Removed Cal Gas Conc:	5.34	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2448
ZAG Make/Model:	Teledyne API T701		Serial Number:	5609

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017
Converter make:	CDN-101	Converter serial #:	521
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.986610	1.000080	Backgd or Offset:	3.69	3.77
Calibration intercept:	-0.122144	0.077786	Coeff or Slope:	1.13	1.14

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4925	75.5	80.6	80.2	1.005
As found Mid point	4962	37.7	40.3	40.0	1.007
As found Low point	4981	18.9	20.2	19.9	1.014
New cylinder response					
Baseline Corr As found:	80.2	Prev response:	79.42	*% change:	1.0%
Baseline Corr 2nd AF pt:	40.0	AF Slope:	0.995402	AF Intercept:	-0.082214
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999994	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4925	75.5	80.6	80.7	0.999
Mid point	4962	37.7	40.3	40.4	0.997
Low point	4981	18.9	20.2	20.2	0.999
As left zero	5000	0.0	0.0	0.3	----
As left span	4925	75.5	80.6	80.0	1.008
SO2 Scrubber Check	4921	79.1	791.0	0.1	----
Date of last scrubber change:	20-Jan-20			Ave Corr Factor	0.998
Date of last converter efficiency test:					

Notes: Changed inlet filter and pump after multipoint as founds. SO2 scrubber check done after calibrator zero and passed. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	December 13, 2024	Previous Calibration:	November 13, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:09	End Time (MST):	14:34
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999998	≥ 0.995
80.6	80.7	0.9991	Slope	1.000080	$0.90 - 1.10$
40.3	40.4	0.9967	Intercept	0.077786	± 3
20.2	20.2	0.9993			



TRS Calibration Plot

Date: December 13, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	December 5, 2024	Last Cal Date:	November 25, 2024
Start time (MST):	10:20	End time (MST):	12:55
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC2608112	Cal Gas Expiry Date:	
CH4 Cal Gas Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
C3H8 Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH4 Conc.	503.6 ppm	CH4 Equiv Conc.	1077.5 ppm
Removed C3H8 Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2448
Zero Air Gen model:	Teledyne API T701	Serial Number:	1118

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1181490018
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.58E-04	2.58E-04	NMHC SP Ratio:	5.87E-05	4.75E-05
CH4 Retention time:	14.60	14.60	NMHC Peak Area:	154800	191241
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	17.05	17.82	0.957
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.82	Prev response	17.08	*% change	4.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	17.05	17.08	0.998
Mid point	4961	39.5	8.51	8.48	1.004
Low point	4980	19.8	4.27	4.18	1.021
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	17.05	17.14	0.995

Average Correction Factor 1.008

Notes: Replaced N2 carrier gas and adjusted NM window from 33 to 32 seconds. Performed span adjustment.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	9.08	9.89	0.918
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.89	Prev response	9.11	*% change	7.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	9.08	9.12	0.996
Mid point	4961	39.5	4.53	4.52	1.002
Low point	4980	19.8	2.27	2.24	1.016
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	9.08	9.13	0.995
Average Correction Factor					1.005

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	7.97	7.93	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.93	Prev response	7.97	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	7.97	7.97	1.000
Mid point	4961	39.5	3.98	3.96	1.006
Low point	4980	19.8	1.99	1.94	1.026
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	7.97	8.01	0.995
Average Correction Factor					1.011

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.003277	1.003815
THC Cal Offset:	-0.023150	-0.049161
CH ₄ Cal Slope:	1.002838	1.002163
CH ₄ Cal Offset:	-0.023183	-0.025579
NMHC Cal Slope:	1.003297	1.005604
NMHC Cal Offset:	0.000232	-0.024182

Calibration Performed By: Kelly Baragar



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 25, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:20	End Time (MST):	12:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999962	<i>≥0.995</i>
17.05	17.08	0.9978	Slope	1.003815	<i>0.90 - 1.10</i>
8.51	8.48	1.0040	Intercept	-0.049161	<i>+/-0.5</i>
4.27	4.18	1.0209			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 25, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:20	End Time (MST):	12:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999951	≥ 0.995
7.97	7.97	0.9997	Slope	1.002163	0.90 - 1.10
3.98	3.96	1.0058	Intercept	-0.025579	± 0.5
1.99	1.94	1.0264			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 25, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:20	End Time (MST):	12:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999968	<i>≥0.995</i>
9.08	9.12	0.9958	Slope	1.005604	<i>0.90 - 1.10</i>
4.53	4.52	1.0023	Intercept	-0.024182	<i>+/-0.5</i>
2.27	2.24	1.0160			



NMHC Calibration Plot

Date: December 5, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort McKay South
 Station number: AMS 13
 Calibration Date: December 10, 2024
 Last Cal Date: November 6, 2024
 Start time (MST): 10:45
 End time (MST): 14:50
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2UP1RP
 NOX Cal Gas Conc: 48.25 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.25 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne APIT701
 Cal Gas Expiry Date: November 17, 2026
 NO Cal Gas Conc: 47.88 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 47.88 ppm
 NO gas Diff:
 Serial Number: 2448
 Serial Number: 1118

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	1.9	1.9	0.0	----	----
AF High point	4917	83.5	805.7	799.5	6.2	806.1	797.4	8.7	1.0019	1.0050
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 801.7 ppb	NO = 792.6 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 0.3%
Baseline Corr 1st pt	NO _x = 804.2 ppb	NO = 795.5 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1410661329

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997562	0.996427
NO _x Cal Offset:	-2.032369	-1.932112
NO Cal Slope:	0.994805	0.996592
NO Cal Offset:	-2.769719	-2.870130
NO ₂ Cal Slope:	1.000308	1.011196
NO ₂ Cal Offset:	0.284103	1.206318

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.148	1.148	NO bkgnd or offset:	10.2	12.4
NOX coeff or slope:	1.004	1.004	NOX bkgnd or offset:	10.4	12.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	161.1	161.1

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
High point	4917	83.5	805.7	799.5	6.2	801.9	795.3	6.5	1.0047	1.0053
Mid point	4958	41.8	403.4	400.3	3.1	399.1	394.9	4.2	1.0107	1.0137
Low point	4979	20.9	201.7	200.1	1.5	196.8	193.4	3.4	1.0248	1.0349
As left zero	5000	0.0	0.0	0.0	0.0	1.2	1.0	0.2	----	----
As left span	4917	83.5	805.7	374.3	431.4	794.8	374.3	420.5	1.0137	1.0000
Average Correction Factor									1.0134	1.0179

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	783.1	373.5	415.8	420.9	0.9878	101.2%
Mid GPT point	783.1	577.0	212.3	216.9	0.9787	102.2%
Low GPT point	783.1	681.4	107.9	111.1	0.9710	103.0%
Average Correction Factor					0.9792	102.1%

Notes: Changed inlet filter. Performed zero adjustment, suspect the slightly high PMT temperature has affected the baseline slightly. Used second GPT reference point due to slight drift.

Calibration Performed By: Kelly Baragar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 10, 2024	Previous Calibration:	November 6, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:45	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999968	<i>≥0.995</i>
805.7	801.9	1.0047	Slope	0.996427	<i>0.90 - 1.10</i>
403.4	399.1	1.0107	Intercept	-1.932112	<i>+/-20</i>
201.7	196.8	1.0248			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 10, 2024	Previous Calibration:	November 6, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:45	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999964	≥0.995
415.8	420.9	0.9878	Slope	1.011196	0.90 - 1.10
212.3	216.9	0.9787	Intercept	1.206318	+/-20
107.9	111.1	0.9710			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 10, 2024	Previous Calibration:	November 6, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:45	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661329

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999936	<i>≥0.995</i>
799.5	795.3	1.0053	Slope	0.996592	<i>0.90 - 1.10</i>
400.3	394.9	1.0137	Intercept	-2.870130	<i>+/-20</i>
200.1	193.4	1.0349			



NO_x Calibration Plot

Date: December 10, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	December 2, 2024	Last Cal Date:	November 4, 2024
Start time (MST):	10:28	End time (MST):	13:26
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	2448
Calibrator Make/Model:	Teledyne API T700	Serial Number:	1118
ZAG Make/Model:	Teledyne API T701		

Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #:	3871
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002457	1.002486	Backgd or Offset:	2.7	2.7
Calibration intercept:	1.220000	0.940000	Coeff or Slope:	0.973	0.973

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.5	----
As found High point	5000	997.5	400.0	401.8	0.994
As found Mid point					
As found Low point					
Baseline Corr As found:	402.3	Previous response	402.2	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	5000	996.1	400.0	401.4	0.997
Mid point	5000	850.2	200.0	202.2	0.989
Low point	5000	751.7	100.0	101.8	0.982
As left zero	5000	0.0	0.0	0.3	----
As left span	5000	996.1	400.0	403.0	0.993
Average Correction Factor					0.989

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 4, 2024
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:28	End Time (MST):	13:26
Analyzer make:	Teledyne API T400	Analyzer serial #:	3871

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999978	≥0.995
400.0	401.4	0.9965	Slope	1.002486	0.90 - 1.10
200.0	202.2	0.9891	Intercept	0.940000	+/- 5
100.0	101.8	0.9823			



O₃ Calibration Plot

Date: December 2, 2024

Location: Fort McKay South





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay South Station number: AMS 13
 Calibration Date: December 10, 2024 Last Cal Date: November 13, 2024
 Start time (MST): 11:55 End time (MST): 13:07

Analyzer Make: Teledyne API T640 S/N: 1335
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388746
 Temp/RH standard: Alicat FP-25BT S/N: 388746

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-19.6	-21.23	-19.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	747.7	749.56	747.7	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.87	5.00	<input checked="" type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	44	----	44	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.6	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	7.0		11.0	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ October 1, 2024
 Date Disposable Filter Changed: _____ October 1, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ October 1, 2024
 Date RH/T Sensor Cleaned: _____ October 1, 2024

Notes: Made slight adjustment to flow, performed peak test and adjusted PMT from 1599 to 1630. Leak check performed before and after PMT adjustment, no issues.

Calibration by: Kelly Baragar




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC
DECEMBER 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	December 3, 2024	Last Cal Date:	November 7, 2024
Start time (MST):	11:02	End time (MST):	14:33
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.95	ppm	Cal Gas Exp Date:	January 5, 2025
Cal Gas Cylinder #:	CC279389			
Removed Cal Gas Conc:	49.95	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	3060
Zero Air Gen Model:	API T701H		Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	0710321322
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006657	0.996286	Backgd or Offset:	24.8	24.1
Calibration intercept:	-2.605499	-1.499686	Coeff or Slope:	1.043	1.043

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.6	----
As found High point	4938	80.3	799.3	795.8	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	796.4	Previous response	802.0	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4938	80.3	799.3	795.9	1.004
Mid point	4979	40.2	400.1	395.7	1.011
Low point	4998	20.2	201.1	197.2	1.020
As left zero	5000	0.0	0.0	0.3	----
As left span	4938	80.3	799.3	800.9	0.998
Average Correction Factor:					1.012

Notes: Sample inlet filter changed after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 7, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:02	End Time (MST):	14:33
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.4	----	Correlation Coefficient	0.999973	≥0.995
799.3	795.9	1.0042	Slope	0.996286	0.90 - 1.10
400.1	395.7	1.0110	Intercept	-1.499686	+/-30
201.1	197.2	1.0196			



SO2 Calibration Plot

Date: December 3, 2024

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	December 12, 2024	Last Cal Date:	November 27, 2024
Start time (MST):	10:07	End time (MST):	14:58
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.15	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC510379			
Removed Cal Gas Conc:	5.15	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3060
ZAG Make/Model:	API 701H		Serial Number:	357

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582
Converter make:	CD Nova CDN-101	Converter serial #:	503
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.975285	1.000876	Backgd or Offset:	2.3	2.5
Calibration intercept:	-0.065266	0.014569	Coeff or Slope:	0.995	1.065

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4938	77.9	80.0	76.1	1.048
As found Mid point	4973	38.9	40.0	37.7	1.055
As found Low point	4997	19.5	20.0	18.6	1.065
New cylinder response					
Baseline Corr As found:	76.3	Prev response:	77.94	*% change:	-2.2%
Baseline Corr 2nd AF pt:	37.9	AF Slope:	0.954900	AF Intercept:	-0.365158
Baseline Corr 3rd AF pt:	18.8	AF Correlation:	0.999978	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4938	77.9	80.0	80.1	0.998
Mid point	4973	38.9	40.0	39.9	1.001
Low point	4997	19.5	20.0	20.0	1.001
As left zero	5000	0.0	0.0	0.2	----
As left span	4938	77.9	80.0	79.3	1.008
SO2 Scrubber Check	4936	80.3	800.4	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.000
Date of last converter efficiency test:					

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 27, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:07	End Time (MST):	14:58
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999993	≥ 0.995
80.0	80.1	0.9981	Slope	1.000876	$0.90 - 1.10$
40.0	39.9	1.0014	Intercept	0.014569	± 3
20.0	20.0	1.0006			



TRS Calibration Plot

Date: December 12, 2024

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	December 3, 2024	Last Cal Date:	November 21, 2024
Start time (MST):	11:02	End time (MST):	14:33
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
Zero Air Gen model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1118148494
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.60E-04	4.40E-04	NMHC SP Ratio:	4.49E-05	4.48E-05
CH4 Retention time:	14.80	14.60	NMHC Peak Area:	203545	203902
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	17.14	17.57	0.976
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.57	Prev response	17.14	*% change	2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	17.14	17.15	1.000
Mid point	4960	40.1	8.57	8.53	1.005
Low point	4980	20.0	4.28	4.27	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	17.14	17.08	1.004
Average Correction Factor					1.002

Notes: Sample inlet filter changed after as founds. Adjusted zero and span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	9.13	9.17	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.17	Prev response	9.11	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	9.13	9.13	1.001
Mid point	4960	40.1	4.57	4.55	1.005
Low point	4980	20.0	2.28	2.27	1.004
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	9.13	9.14	1.000
Average Correction Factor					1.003

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	8.01	8.40	0.953
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.40	Prev response	8.03	*% change	4.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	8.01	8.02	0.998
Mid point	4960	40.1	4.00	3.98	1.006
Low point	4980	20.0	2.00	2.00	0.997
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	8.01	7.94	1.008
Average Correction Factor					1.000

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999637	1.000257
THC Cal Offset:	-0.001623	-0.012023
CH ₄ Cal Slope:	1.001585	1.001044
CH ₄ Cal Offset:	0.004958	-0.004645
NMHC Cal Slope:	0.997917	0.999556
NMHC Cal Offset:	-0.006781	-0.007578

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 21, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:02	End Time (MST):	14:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999990	<i>≥0.995</i>
17.14	17.15	0.9995	Slope	1.000257	<i>0.90 - 1.10</i>
8.57	8.53	1.0051	Intercept	-0.012023	<i>+/-0.5</i>
4.28	4.27	1.0010			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 21, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:02	End Time (MST):	14:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999982	<i>≥0.995</i>
8.01	8.02	0.9984	Slope	1.001044	<i>0.90 - 1.10</i>
4.00	3.98	1.0056	Intercept	-0.004645	<i>+/-0.5</i>
2.00	2.00	0.9971			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 21, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:02	End Time (MST):	14:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999994	≥ 0.995
9.13	9.13	1.0005	Slope	0.999556	$0.90 - 1.10$
4.57	4.55	1.0050	Intercept	-0.007578	± 0.5
2.28	2.27	1.0045			



NMHC Calibration Plot

Date: December 3, 2024

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	December 16, 2024	Last Cal Date:	December 3, 2024
Start time (MST):	10:39	End time (MST):	14:58
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
Zero Air Gen model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1118148494
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.40E-04	4.96E-04	NMHC SP Ratio:	4.48E-05	4.66E-05
CH4 Retention time:	14.6	14.8	NMHC Peak Area:	203902	195844
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	17.14	15.92	1.077
As found Mid point	4960	40.1	8.57	7.93	1.081
As found Low point	4980	20.0	4.28	3.97	1.076
New cylinder response					
Baseline Corr AF:	15.92	Prev response	17.14	*% change	-7.6%
Baseline Corr 2nd AF:	7.93	AF Slope:	0.928328	AF Intercept:	-0.004722
Baseline Corr 3rd AF:	3.97	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	17.14	17.18	0.998
Mid point	4960	40.1	8.57	8.55	1.003
Low point	4980	20.0	4.28	4.28	0.999
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	17.14	17.22	0.995
Average Correction Factor					1.000

Notes: Replaced the actuator and the pump due to dips that started occurring and the drift on the span.
Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	9.13	8.89	1.027
As found Mid point	4960	40.1	4.57	4.41	1.035
As found Low point	4980	20.0	2.28	2.20	1.035
New cylinder response					
Baseline Corr AF:	8.89	Prev response	9.12	*% change	-2.6%
Baseline Corr 2nd AF:	4.41	AF Slope:	0.973864	AF Intercept:	-0.013437
Baseline Corr 3rd AF:	2.20	AF Correlation:	0.999983	* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	9.13	9.14	0.999
Mid point	4960	40.1	4.57	4.55	1.004
Low point	4980	20.0	2.28	2.28	1.000
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	9.13	9.15	0.998
Average Correction Factor					1.001

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	8.01	7.03	1.140
As found Mid point	4960	40.1	4.00	3.52	1.139
As found Low point	4980	20.0	2.00	1.77	1.128
New cylinder response					
Baseline Corr AF:	7.03	Prev response	8.01	*% change	-14.0%
Baseline Corr 2nd AF:	3.52	AF Slope:	0.876687	AF Intercept:	0.007915
Baseline Corr 3rd AF:	1.77	AF Correlation:	0.999992	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	8.01	8.03	0.997
Mid point	4960	40.1	4.00	4.00	1.001
Low point	4980	20.0	2.00	2.00	0.998
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	8.01	8.07	0.992
Average Correction Factor					0.999

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000257	1.001557
THC Cal Offset:	-0.012023	-0.008018
CH ₄ Cal Slope:	1.001044	1.002871
CH ₄ Cal Offset:	-0.004645	-0.003045
NMHC Cal Slope:	0.999556	1.000581
NMHC Cal Offset:	-0.007578	-0.005173

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 16, 2024	Previous Calibration:	December 3, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:39	End Time (MST):	14:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999993	≥ 0.995
17.14	17.18	0.9982	Slope	1.001557	$0.90 - 1.10$
8.57	8.55	1.0028	Intercept	-0.008018	± 0.5
4.28	4.28	0.9989			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 16, 2024	Previous Calibration:	December 3, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:39	End Time (MST):	14:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999994	<i>≥0.995</i>
8.01	8.03	0.9968	Slope	1.002871	<i>0.90 - 1.10</i>
4.00	4.00	1.0008	Intercept	-0.003045	<i>+/-0.5</i>
2.00	2.00	0.9981			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 16, 2024	Previous Calibration:	December 3, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:39	End Time (MST):	14:58
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999992	≥ 0.995
9.13	9.14	0.9992	Slope	1.000581	$0.90 - 1.10$
4.57	4.55	1.0041	Intercept	-0.005173	± 0.5
2.28	2.28	1.0000			



NMHC Calibration Plot

Date: December 16, 2024

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	December 31, 2024	Last Cal Date:	December 16, 2024
Start time (MST):	9:30	End time (MST):	13:30
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC279389	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
C3H8 Cal Gas Conc.	207.1 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	499.3 ppm	CH4 Equiv Conc.	1068.8 ppm
Removed C3H8 Conc.	207.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
Zero Air Gen model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1118148494
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	4.96E-04	5.89E-04	NMHC SP Ratio:	4.66E-05	4.87E-05
CH4 Retention time:	14.8	15.0	NMHC Peak Area:	195844	187508
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.01	----
As found High point	4920	80.2	17.14	15.63	1.097
As found Mid point	4960	40.1	8.57	7.77	1.105
As found Low point	4980	20.0	4.28	3.89	1.102
New cylinder response					
Baseline Corr AF:	15.63	Prev response	17.16	*% change	-9.8%
Baseline Corr 2nd AF:	7.76	AF Slope:	0.911559	AF Intercept:	-0.011996
Baseline Corr 3rd AF:	3.88	AF Correlation:	0.999986	* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	17.14	17.26	0.993
Mid point	4960	40.1	8.57	8.55	1.003
Low point	4980	20.0	4.28	4.27	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	17.14	17.22	0.995
Average Correction Factor					0.999

Notes: Calibrating the instrument due to a baseline shift yesterday.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	9.13	8.82	1.035
As found Mid point	4960	40.1	4.57	4.38	1.042
As found Low point	4980	20.0	2.28	2.19	1.039
New cylinder response					
Baseline Corr AF:	8.82	Prev response	9.13	*% change	-3.5%
Baseline Corr 2nd AF:	4.38	AF Slope:	0.965883	AF Intercept:	-0.009052
Baseline Corr 3rd AF:	2.19	AF Correlation:	0.999986	<i>* = > +/-5% change initiates investigation</i>	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	9.13	9.13	1.000
Mid point	4960	40.1	4.57	4.54	1.006
Low point	4980	20.0	2.28	2.27	1.003
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	9.13	9.15	0.998
Average Correction Factor					1.003

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.01	----
As found High point	4920	80.2	8.01	6.81	1.178
As found Mid point	4960	40.1	4.00	3.38	1.185
As found Low point	4980	20.0	2.00	1.69	1.184
New cylinder response					
Baseline Corr AF:	6.80	Prev response	8.03	*% change	-18.1%
Baseline Corr 2nd AF:	3.38	AF Slope:	0.849166	AF Intercept:	-0.001945
Baseline Corr 3rd AF:	1.69	AF Correlation:	0.999987	<i>* = > +/-5% change initiates investigation</i>	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	8.01	8.13	0.985
Mid point	4960	40.1	4.00	4.01	0.998
Low point	4980	20.0	2.00	2.00	0.998
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	8.01	8.07	0.992
Average Correction Factor					0.994

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001557	1.007143
THC Cal Offset:	-0.008018	-0.029399
CH ₄ Cal Slope:	1.002871	1.015499
CH ₄ Cal Offset:	-0.003045	-0.021023
NMHC Cal Slope:	1.000581	0.999793
NMHC Cal Offset:	-0.005173	-0.007776

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 31, 2024	Previous Calibration:	December 16, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:30	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999972	≥0.995
17.14	17.26	0.9930	Slope	1.007143	0.90 - 1.10
8.57	8.55	1.0026	Intercept	-0.029399	+/-0.5
4.28	4.27	1.0005			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 31, 2024	Previous Calibration:	December 16, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:30	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999945	≥0.995
8.01	8.13	0.9850	Slope	1.015499	0.90 - 1.10
4.00	4.01	0.9983	Intercept	-0.021023	+/-0.5
2.00	2.00	0.9976			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 31, 2024	Previous Calibration:	December 16, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:30	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999989	≥0.995
9.13	9.13	1.0001	Slope	0.999793	0.90 - 1.10
4.57	4.54	1.0061	Intercept	-0.007776	+/-0.5
2.28	2.27	1.0027			



NMHC Calibration Plot

Date: December 31, 2024

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Anzac
 Station number: AMS 14
 Calibration Date: December 2, 2024
 Last Cal Date: November 6, 2024
 Start time (MST): 11:09
 End time (MST): 15:46
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0037092
 NOX Cal Gas Conc: 60.7 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 60.70 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T700H
 Cal Gas Expiry Date: May 16, 2031
 NO Cal Gas Conc: 60.40 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 60.40 ppm
 NO gas Diff:
 Serial Number: 3060
 Serial Number: 357

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
AF High point	4934	66.3	804.8	800.9	4.0	813.4	808.3	5.1	0.9895	0.9907
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 802.4 ppb	NO = 799.9 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 1.4%
Baseline Corr 1st pt	NO _x = 813.4 ppb	NO = 808.4 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 1.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430008

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.424	1.424	NO bkgnd or offset:	3.9	3.9
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.8	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	155.7	157.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997812	1.007671
NO _x Cal Offset:	-0.710232	-0.731231
NO Cal Slope:	1.001538	1.011561
NO Cal Offset:	-2.229291	-2.190787
NO ₂ Cal Slope:	1.001925	0.995747
NO ₂ Cal Offset:	-1.159515	-1.424813

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.1	----	----
High point	4934	66.3	804.8	800.9	4.0	810.5	808.9	1.6	0.9930	0.9901
Mid point	4985	33.2	401.6	399.6	2.0	404.1	401.4	2.7	0.9938	0.9955
Low point	5004	16.7	201.9	200.9	1.0	201.5	198.4	3.1	1.0020	1.0126
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
As left span	4934	66.3	804.8	426.2	378.6	805.7	426.2	379.6	0.9989	1.0000
Average Correction Factor									0.9963	0.9994

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	805.2	415.8	393.4	391.0	1.0061	99.4%
Mid GPT point	805.2	612.2	197.0	194.1	1.0148	98.5%
Low GPT point	805.2	709.7	99.5	96.0	1.0362	96.5%
Average Correction Factor					1.0190	98.1%

Notes: Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 6, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:09	End Time (MST):	15:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999994	≥0.995
804.8	810.5	0.9930	Slope	1.007671	0.90 - 1.10
401.6	404.1	0.9938	Intercept	-0.731231	+/-20
201.9	201.5	1.0020			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 6, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:09	End Time (MST):	15:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999930	<i>≥0.995</i>
393.4	391.0	1.0061	Slope	0.995747	<i>0.90 - 1.10</i>
197.0	194.1	1.0148	Intercept	-1.424813	<i>+/-20</i>
99.5	96.0	1.0362			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 6, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:09	End Time (MST):	15:46
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999962	<i>≥0.995</i>
800.9	808.9	0.9901	Slope	1.011561	<i>0.90 - 1.10</i>
399.6	401.4	0.9955	Intercept	-2.190787	<i>+/-20</i>
200.9	198.4	1.0126			



NO_x Calibration Plot

Date: December 2, 2024

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Anzac	Station number:	AMS 14
Calibration Date:	December 9, 2024	Last Cal Date:	November 25, 2024
Start time (MST):	10:26	End time (MST):	13:32
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3060
Calibrator Make/Model:	API T700	Serial Number:	357
ZAG Make/Model:	API 701H		

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002457	1.004286	Backgd or Offset:	2.1	2.1
Calibration intercept:	-1.180000	-1.200000	Coeff or Slope:	1.667	1.667

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	5000	918.8	400.0	399.1	1.003
As found Mid point					
As found Low point					
Baseline Corr As found:	398.8	Previous response	399.8	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	5000	918.8	400.0	400.9	0.998
Mid point	5000	803.8	200.0	199.4	1.003
Low point	5000	709.8	100.0	98.2	1.018
As left zero	5000	0.0	0.0	-0.5	----
As left span	5000	918.8	400.0	402.8	0.993
Average Correction Factor					1.006

Notes: Sample inlet filter changed after asfinds. No adjustments needed.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 25, 2024
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:26	End Time (MST):	13:32
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999976	≥0.995
400.0	400.9	0.9978	Slope	1.004286	0.90 - 1.10
200.0	199.4	1.0030	Intercept	-1.200000	+/- 5
100.0	98.2	1.0183			



O₃ Calibration Plot

Date: December 9, 2024

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Anzac Station number: AMS 14
 Calibration Date: December 12, 2024 Last Cal Date: November 27, 2024
 Start time (MST): 15:15 End time (MST): 15:29

Analyzer Make: AP T640 S/N: 825
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
 Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-11.1	-11.16	-11.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	709.8	710.72	709.8	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.000	4.963	5.000	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	41	-----	41	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.0	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: _____ Expiry Date: _____
 Lot No.: _____

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ August 29, 2024
 Date Disposable Filter Changed: _____ August 29, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ August 29, 2024
 Date RH/T Sensor Cleaned: _____ August 29, 2024

Notes: No adjustments made. Leak check passed. Head cleaned.

Calibration by: Mohammed Kashif




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS17
WAPASU
DECEMBER 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	December 2, 2024	Last Cal Date:	November 4, 2024
Start time (MST):	11:43	End time (MST):	15:00
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.38	ppm	Cal Gas Exp Date:	January 12, 2029
Cal Gas Cylinder #:	ALM066507			
Removed Cal Gas Conc:	50.38	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	2449
Zero Air Gen Model:	Teledyne API 701H		Serial Number:	359

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1218153459
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999668	1.000869	Backgd or Offset:	13.4	15.0
Calibration intercept:	-1.939863	-2.760072	Coeff or Slope:	1.109	1.109

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.7	----
As found High point	4921	79.4	800.0	799.5	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	798.8	Previous response	797.8	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.6	----
High point	4921	79.4	800.0	799.0	1.001
Mid point	4960	39.7	400.0	396.5	1.009
Low point	4980	19.8	199.5	194.8	1.024
As left zero	5000	0.0	0.0	-0.9	----
As left span	4920	79.4	800.1	803.5	0.996
Average Correction Factor:					1.011

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 4, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:43	End Time (MST):	15:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.6	----	Correlation Coefficient	0.999967	≥0.995
800.0	799.0	1.0012	Slope	1.000869	0.90 - 1.10
400.0	396.5	1.0089	Intercept	-2.760072	+/-30
199.5	194.8	1.0242			



SO2 Calibration Plot

Date: December 2, 2024

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	Wapasu	Station number:	AMS 17
Calibration Date:	December 16, 2024	Last Cal Date:	November 26, 2024
Start time (MST):	12:19	End time (MST):	16:55
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.77	ppm	Cal Gas Exp Date:	August 28, 2027
Cal Gas Cylinder #:	DT20029267			
Removed Cal Gas Conc:	4.77	ppm	Rem Gas Exp Date:	N/A
Removed Gas Cyl #:	N/A		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2449
ZAG Make/Model:	API T701H		Serial Number:	359

Analyzer Information

Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583
Converter make:	CD Nova	Converter serial #:	N/A
Analyzer Range	0 - 100 ppb	Converter Temp:	340 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997214	1.011211	Backgd or Offset:	13.5	13.5
Calibration intercept:	-0.239842	-0.079935	Coeff or Slope:	1.141	1.141

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero)) Limit = 0.90-1.10
As found zero	5000	0.0	0.0	0.1	----
As found High point	4916	83.9	80.0	83.1	0.964
As found Mid point	4958	41.9	40.0	41.3	0.970
As found Low point	4979	21.0	20.0	20.3	0.992
New cylinder response					
Baseline Corr As found:	83.0	Prev response:	79.58	*% change:	4.1%
Baseline Corr 2nd AF pt:	41.2	AF Slope:	1.039054	AF Intercept:	-0.179781
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999945	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4916	83.9	80.0	81.0	0.988
Mid point	4958	41.9	40.0	40.1	0.997
Low point	4979	21.0	20.0	20.1	0.997
As left zero	5000	0.0	0.0	0.3	----
As left span	4916	83.9	80.0	80.3	0.997
SO2 Scrubber Check	4921	79.4	793.9	0.1	----
Date of last scrubber change:		N/A		Ave Corr Factor	0.994
Date of last converter efficiency test:		N/A			

Notes: Changed the sample inlet filter after as founds. Completed a SO2 scrubber check after calibrator zero. No adjustments performed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	December 16, 2024	Previous Calibration:	November 26, 2024
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	12:19	End Time (MST):	16:55
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999967	≥0.995
80.0	81.0	0.9882	Slope	1.011211	0.90 - 1.10
40.0	40.1	0.9968	Intercept	-0.079935	+/-3
20.0	20.1	0.9967			



H₂S Calibration Plot

Date: December 16, 2024

Location: Wapasu





Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 4, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:43	End Time (MST):	15:00
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.07	----	Correlation Coefficient	0.999950	≥0.995
17.09	17.18	0.9948	Slope	1.002146	0.90 - 1.10
8.55	8.54	1.0010	Intercept	0.019052	+/-1.5
4.26	4.25	1.0034			



THC Calibration Plot

Date: December 2, 2024

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Wapasu
 Station number: AMS 17
 Calibration Date: December 12, 2024
 Last Cal Date: November 13, 2024
 Start time (MST): 10:50
 End time (MST): 14:55
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T375YK8
 NOX Cal Gas Conc: 49.11 ppm
 Removed Cylinder #: N/A
 Removed Gas NOX Conc: 49.11 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701H
 Cal Gas Expiry Date: April 13, 2025
 NO Cal Gas Conc: 48.07 ppm
 Removed Gas Exp Date: N/A
 Removed Gas NO Conc: 48.07 ppm
 NO gas Diff:
 Serial Number: 2449
 Serial Number: 359

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.3	-0.4	----	----
AF High point	4917	83.2	817.2	799.9	17.3	816.3	796.5	19.9	1.0004	1.0039
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 815.9 ppb		NO = 798.0 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 0.1%	
Baseline Corr 1st pt	NO _x = 816.9 ppb		NO = 796.8 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -0.2%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found NO _x r ² :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found NO r ² :		NO SI:	NO Int:	
						As found NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo Scientific 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153460

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000138	1.001214
NO _x Cal Offset:	-1.420000	-1.480000
NO Cal Slope:	1.000015	1.001016
NO Cal Offset:	-1.880000	-1.780000
NO ₂ Cal Slope:	1.002948	1.001334
NO ₂ Cal Offset:	-0.302443	-0.701642

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.087	1.091	NO bkgnd or offset:	3.8	3.8
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	4.2	4.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	241.5	239.7

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.1	-0.4	----	----
High point	4917	83.2	817.2	799.9	17.3	817.3	799.8	17.4	0.9999	1.0001
Mid point	4958	41.6	408.6	399.9	8.7	406.8	397.6	9.2	1.0044	1.0059
Low point	4979	20.8	204.3	200.0	4.3	202.3	196.8	5.6	1.0099	1.0161
As left zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.1	-0.4	----	----
As left span	4917	83.2	817.2	404.8	412.4	817.0	404.8	412.3	1.0002	1.0000
Average Correction Factor									1.0047	1.0074

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.4	----	----
High GPT point	797.7	399.5	415.5	415.6	0.9998	100.0%
Mid GPT point	797.7	601.5	213.5	212.7	1.0038	99.6%
Low GPT point	797.7	701.9	113.1	112.4	1.0063	99.4%
Average Correction Factor					1.0033	99.7%

Notes: Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 13, 2024
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:50	End Time (MST):	14:55
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.5	----	Correlation Coefficient	0.999993	≥0.995
817.2	817.3	0.9999	Slope	1.001214	0.90 - 1.10
408.6	406.8	1.0044	Intercept	-1.480000	+/-20
204.3	202.3	1.0099			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 13, 2024
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:50	End Time (MST):	14:55
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	0.999997	≥0.995
415.5	415.6	0.9998	Slope	1.001334	0.90 - 1.10
213.5	212.7	1.0038	Intercept	-0.701642	+/-20
113.1	112.4	1.0063			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 13, 2024
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:50	End Time (MST):	14:55
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999980	<i>≥0.995</i>
799.9	799.8	1.0001	Slope	1.001016	<i>0.90 - 1.10</i>
399.9	397.6	1.0059	Intercept	-1.780000	<i>+/-20</i>
200.0	196.8	1.0161			



NO_x Calibration Plot

Date: December 12, 2024

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Wapasu	Station number:	AMS17
Calibration Date:	December 9, 2024	Last Cal Date:	November 6, 2024
Start time (MST):	11:26	End time (MST):	15:25
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer		
Calibrator Make/Model:	API T700	Serial Number:	2449
ZAG Make/Model:	API T701H	Serial Number:	359

Analyzer Information

Analyzer make:	API T400	Analyzer serial #:	7045
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001086	1.002114	Backgd or Offset:	0.6	0.6
Calibration intercept:	-0.740000	-0.420000	Coeff or Slope:	1.018	1.028

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	5000	1104.7	400.0	398.0	1.004
As found Mid point					
As found Low point					
Baseline Corr As found:	398.4	Previous response	399.7	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	5000	1104.7	400.0	401.0	0.998
Mid point	5000	917.3	200.0	199.0	1.005
Low point	5000	797.9	100.0	99.5	1.005
As left zero	5000	0.0	0.0	0.6	----
As left span	5000	1104.0	400.0	405.4	0.987
Average Correction Factor					1.003

Notes: Inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 6, 2024
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:26	End Time (MST):	15:25
Analyzer make:	API T400	Analyzer serial #:	7045

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999978	≥0.995
400.0	401.0	0.9975	Slope	1.002114	0.90 - 1.10
200.0	199.0	1.0050	Intercept	-0.420000	+/- 5
100.0	99.5	1.0050			



O₃ Calibration Plot

Date: December 9, 2024

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Wapasu Station number: AMS 17
 Calibration Date: December 19, 2024 Last Cal Date: November 26, 2024
 Start time (MST): 11:52 End time (MST): 14:00

Analyzer Make: Teledyne API T640 S/N: 1183
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
 Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-15.30	-16.20	-15.60	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	725.20	728.05	728.05	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.16	4.20	5.00	<input checked="" type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	100	----	39	<input checked="" type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	1.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: October 6, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: October 18, 2024
 Date Disposable Filter Changed: December 19, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 23, 2024
 Date RH/T Sensor Cleaned: July 23, 2024

Notes:

Pump swapped out after as founds. Flow adjusted. Leak check passed. DFU filter changed out.

Calibration by: Aswin Sasi Kumar




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Stony Mountain	Station number: AMS 18
Calibration Date:	December 5, 2024	Last Cal Date: November 21, 2024
Start time (MST):	11:55	End time (MST): 15:07
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	50.02 ppm	Cal Gas Exp Date: January 12, 2029
Cal Gas Cylinder #:	XC026809B	
Removed Cal Gas Conc:	50.02 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	XC026809B	Diff between cyl:
Calibrator Model:	Teledyne API T700	Serial Number: 2658
Zero Air Gen Model:	Teledyne API 701H	Serial Number: 360

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: JC1501301453
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997772	0.996102	Backgd or Offset:	23.6	23.6
Calibration intercept:	-0.160000	-0.800000	Coeff or Slope:	0.797	0.797

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.6	----
As found High point	4920	80.0	800.3	794.5	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	793.9	Previous response	798.4	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.8	----
High point	4920	80.0	800.3	796.9	1.004
Mid point	4960	40.0	400.2	397.9	1.006
Low point	4980	20.0	200.1	196.3	1.019
As left zero	5000	0.0	0.0	0.7	----
As left span	4920	80.0	800.3	797.1	1.004
Average Correction Factor:					1.010

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 21, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:55	End Time (MST):	15:07
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.8	----	Correlation Coefficient	0.999978	≥0.995
800.3	796.9	1.0043	Slope	0.996102	0.90 - 1.10
400.2	397.9	1.0057	Intercept	-0.800000	+/-30
200.1	196.3	1.0193			



SO2 Calibration Plot

Date: December 5, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Stony Mountain	Station number:	AMS18
Calibration Date:	December 3, 2024	Last Cal Date:	November 28, 2024
Start time (MST):	11:36	End time (MST):	16:17
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.48	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC500395			
Removed Cal Gas Conc:	5.48	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	2658
ZAG Make/Model:	Teledyne API T701		Serial Number:	360

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359
Converter make:	CD Nova CDN-101	Converter serial #:	555
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001158	0.996582	Backgd or Offset:	3.3	3.09
Calibration intercept:	0.400939	0.161164	Coeff or Slope:	1.335	1.248

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4927	73.0	80.0	87.8	0.911
As found Mid point	4964	36.5	40.0	44.4	0.901
As found Low point	4983	18.3	20.0	21.7	0.924
New cylinder response					
Baseline Corr As found:	87.8	Prev response:	80.49	*% change:	8.3%
Baseline Corr 2nd AF pt:	44.4	AF Slope:	1.099611	AF Intercept:	-0.020748
Baseline Corr 3rd AF pt:	21.7	AF Correlation:	0.999924	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4927	73.0	80.0	79.8	1.002
Mid point	4964	36.5	40.0	40.3	0.992
Low point	4983	18.3	20.0	19.9	1.007
As left zero	5000	0.0	0.0	0.3	----
As left span	4927	73.0	80.0	79.6	1.005
SO2 Scrubber Check	4923	77.1	771.0	0.1	----
Date of last scrubber change:	17-Dec-21			Ave Corr Factor	1.001
Date of last converter efficiency test:					

Notes: Asfinds high due to weather. Changed sample inlet filter after as finds. Adjusted span.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 28, 2024
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	11:36	End Time (MST):	16:17
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999958	≥ 0.995
80.0	79.8	1.0024	Slope	0.996582	$0.90 - 1.10$
40.0	40.3	0.9924	Intercept	0.161164	± 3
20.0	19.9	1.0074			



TRS Calibration Plot

Date: December 3, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	December 5, 2024	Last Cal Date:	November 21, 2024
Start time (MST):	11:55	End time (MST):	15:07
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	XC026809B	Cal Gas Expiry Date:	January 12, 2029
CH4 Cal Gas Conc.	504.9 ppm	CH4 Equiv Conc.	1076.6 ppm
C3H8 Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:	CC463851	Removed Gas Expiry:	February 23, 2025
Removed CH4 Conc.	500.8 ppm	CH4 Equiv Conc.	1066.8 ppm
Removed C3H8 Conc.	205.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T750	Serial Number:	282
Zero Air Gen model:	Teledyne API T751H	Serial Number:	321

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1193585647
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.71E-04	2.71E-04	NMHC SP Ratio:	4.16E-05	4.16E-05
CH4 Retention time:	16.4	16.4	NMHC Peak Area:	223110	223110
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.02	----
As found High point	4920	80.0	17.07	17.02	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.00	Prev response	17.05	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.02	----
High point	4920	80.0	17.23	16.97	1.015
Mid point	4960	40.0	8.61	8.52	1.011
Low point	4980	20.0	4.31	4.27	1.009
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.0	17.23	17.17	1.003
Average Correction Factor					1.012

Notes:

No adjustments needed.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.0	9.06	9.05	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.05	Prev response	9.07	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.0	9.15	9.03	1.013
Mid point	4960	40.0	4.57	4.54	1.008
Low point	4980	20.0	2.29	2.27	1.007
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.0	9.15	9.14	1.001
Average Correction Factor					1.009

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.02	----
As found High point	4920	80.0	8.01	7.97	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.95	Prev response	7.98	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.02	----
High point	4920	80.0	8.08	7.94	1.017
Mid point	4960	40.0	4.04	3.98	1.015
Low point	4980	20.0	2.02	2.00	1.012
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.0	8.08	8.03	1.007
Average Correction Factor					1.015

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.997150	0.984207
THC Cal Offset:	0.029600	0.027400
CH ₄ Cal Slope:	0.993818	0.980986
CH ₄ Cal Offset:	0.016800	0.017400
NMHC Cal Slope:	1.000106	0.987150
NMHC Cal Offset:	0.013000	0.009600

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 21, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:55	End Time (MST):	15:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.02	----	Correlation Coefficient	0.999998	≥ 0.995
17.23	16.97	1.0148	Slope	0.984207	$0.90 - 1.10$
8.61	8.52	1.0110	Intercept	0.027400	± 0.5
4.31	4.27	1.0095			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 21, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:55	End Time (MST):	15:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	Limits	
0.00	0.02	----	Correlation Coefficient	0.999999	≥0.995
8.08	7.94	1.0170	Slope	0.980986	0.90 - 1.10
4.04	3.98	1.0149	Intercept	0.017400	+/-0.5
2.02	2.00	1.0123			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 21, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:55	End Time (MST):	15:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585647

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
9.15	9.03	1.0128	Slope	0.987150	<i>0.90 - 1.10</i>
4.57	4.54	1.0079	Intercept	0.009600	<i>+/-0.5</i>
2.29	2.27	1.0070			



NMHC Calibration Plot

Date: December 5, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Stony Mountain
 Station number: AMS 18
 Calibration Date: December 20, 2024
 Last Cal Date: November 28, 2024
 Start time (MST): 9:42
 End time (MST): 14:42
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0045516
 NOX Cal Gas Conc: 60.30 ppm
 Removed Cylinder #: N/A
 Removed Gas NOX Conc: 60.30 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API 701H
 Cal Gas Expiry Date: November 17, 2026
 NO Cal Gas Conc: 60.10 ppm
 Removed Gas Exp Date: N/A
 Removed Gas NO Conc: 60.10 ppm
 NO gas Diff:
 Serial Number: 2658
 Serial Number: 355

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-2.1	-1.4	-0.7	----	----
AF High point	4933	66.6	803.3	800.6	2.7	782.9	778.5	4.5	1.0233	1.0265
AF Mid point										
AF Low point										
New cyl resp										
Previous Respo 4933	NO _x = 804.3 ppb		NO = 802.6 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -2.5%	
Baseline Corr 1st pt	NO _x = 785.0 ppb		NO = 779.9 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -2.9%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found NO _x r ² :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found NO r ² :		NO SI:	NO Int:	
						As found NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1035

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001325	0.999821
NO _x Cal Offset:	-0.027026	-0.148573
NO Cal Slope:	1.002817	1.001637
NO Cal Offset:	-0.207696	-0.969502
NO ₂ Cal Slope:	0.990356	1.001892
NO ₂ Cal Offset:	1.039482	1.135453

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.932	0.932	NO bkgnd or offset:	-38.0	-38.0
NOX coeff or slope:	0.929	0.929	NOX bkgnd or offset:	-37.1	-37.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	7.2	7.2

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	1.2	1.7	-0.5	----	----
High point	4933	66.6	803.3	800.6	2.7	803.5	802.4	1.0	0.9997	0.9978
Mid point	4967	33.3	401.6	400.2	1.3	401.1	398.2	2.9	1.0012	1.0051
Low point	4983	16.6	200.2	199.5	0.7	198.4	196.5	1.8	1.0091	1.0155

As left zero
 As left span

Average Correction Factor	1.0033	1.0061
---------------------------	--------	--------

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.5	----	----
High GPT point	798.8	382.2	419.3	420.6	0.9968	100.3%
Mid GPT point	798.8	595.3	206.2	208.0	0.9912	100.9%
Low GPT point	798.8	705.2	96.3	99.5	0.9675	103.4%

Average Correction Factor	0.9852	101.5%
---------------------------	--------	--------

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 20, 2024	Previous Calibration:	November 28, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:42	End Time (MST):	14:42
Analyzer make:	Teledyne API T200	Analyzer serial #:	1035

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	1.2	----	Correlation Coefficient	0.999986	≥0.995
803.3	803.5	0.9997	Slope	0.999821	0.90 - 1.10
401.6	401.1	1.0012	Intercept	-0.148573	+/-20
200.2	198.4	1.0091			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 20, 2024	Previous Calibration:	November 28, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:42	End Time (MST):	14:42
Analyzer make:	Teledyne API T200	Analyzer serial #:	1035

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.5	----	Correlation Coefficient	0.999930	≥0.995
419.3	420.6	0.9968	Slope	1.001892	0.90 - 1.10
206.2	208.0	0.9912	Intercept	1.135453	+/-20
96.3	99.5	0.9675			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 20, 2024	Previous Calibration:	November 28, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	9:42	End Time (MST):	14:42
Analyzer make:	Teledyne API T200	Analyzer serial #:	1035

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.7	----	Correlation Coefficient	0.999949	≥0.995
800.6	802.4	0.9978	Slope	1.001637	0.90 - 1.10
400.2	398.2	1.0051	Intercept	-0.969502	+/-20
199.5	196.5	1.0155			
		5000		0	
		4933		66.6	



NO_x Calibration Plot

Date: December 20, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	December 4, 2024	Last Cal Date:	November 20, 2024
Start time (MST):	11:49	End time (MST):	14:56
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	2658
Calibrator Make/Model:	Teledyne API T700	Serial Number:	355
ZAG Make/Model:	Teledyne API 701H		

Analyzer Information

Analyzer make:	API T400	Analyzer serial #:	825
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999629	1.003457	Backgd or Offset:	2.0	2.0
Calibration intercept:	0.540000	0.220000	Coeff or Slope:	0.994	1.009

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	NA	0.0	-0.1	----
As found High point	4888	1138.1	400.0	396.3	1.009
As found Mid point					
As found Low point					
Baseline Corr As found:	396.4	Previous response	400.4	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	NA	0.0	0.2	----
High point	4888	1138.1	400.0	401.6	0.996
Mid point	4888	884.5	200.0	200.9	0.996
Low point	4888	741.4	100.0	100.6	0.994
As left zero	5000	NA	0.0	0.5	----
As left span	4812	1097.9	400.0	401.6	0.996
Average Correction Factor					0.995

Notes: Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	December 4, 2024	Previous Calibration:	November 20, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:49	End Time (MST):	14:56
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	1.000000	≥0.995
400.0	401.6	0.9960	Slope	1.003457	0.90 - 1.10
200.0	200.9	0.9955	Intercept	0.220000	+/- 5
100.0	100.6	0.9940			



O₃ Calibration Plot

Date: December 4, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Stony Mountain Station number: AMS 18
 Calibration Date: December 20, 2024 Last Cal Date: November 28, 2024
 Start time (MST): 13:15 End time (MST): 14:06

Analyzer Make: API T640 S/N: 324
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388750
 Temp/RH standard: Alicat FP-25BT S/N: 388750

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-13.3	-14	-13.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	700.1	697.50	700.1	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.05	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	57	----	57	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	N/A	N/A	N/A	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: September 24, 2024
 Date Disposable Filter Changed: December 20, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ N/A <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 4, 2024
 Date RH/T Sensor Cleaned: July 4, 2024

Notes: Flow, temp and pressure checked. Leak check passed. DFU filter changed out.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	December 17, 2024	Last Cal Date:	November 7, 2024
Start time (MST):	12:06	End time (MST):	15:08
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	3,080	ppm	Cal Gas Exp Date: November 4, 2028
Cal Gas Cylinder #:	EB0065608		
Removed Cal Gas Conc:	3,080	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 2658
ZAG Make/Model:	Teledyne API T701H		Serial Number: 355

Analyzer Information

Analyzer make:	Teledyne API T300	Analyzer serial #:	3504
Analyzer Range:	0 - 50 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005671	1.010873	Backgd or Offset:	-0.011	-0.012
Calibration intercept:	0.149779	0.123780	Coeff or Slope:	0.909	0.909

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4933	66.7	41.1	41.4	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	41.23	Prev response:	41.47	*% change:	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* => +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4933	66.7	41.1	41.6	0.988
Mid point	4966	33.3	20.5	21.0	0.976
Low point	4983	16.7	10.3	10.5	0.984
As left zero	5000	0.0	0.0	0.0	----
As left span	4933	66.7	41.1	41.6	0.987
Average Correction Factor					0.983

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


CO Calibration Summary

Station Information

Calibration Date:	December 17, 2024	Previous Calibration:	November 7, 2024
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	12:06	End Time (MST):	15:08
Analyzer make:	Teledyne API T300	Analyzer serial #:	3504

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999964	≥0.995
41.1	41.6	0.9877	Slope	1.010873	0.90 - 1.10
20.5	21.0	0.9760	Intercept	0.123780	+/-1.5
10.3	10.5	0.9845			



CO Calibration Plot

Date: December 17, 2024

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Report

Station Information

Station Name:	Stony Mountain	Station number:	AMS 18
Calibration Date:	December 18, 2024	Last Cal Date:	November 19, 2024
Start time (MST):	12:04	End time (MST):	15:00
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	59,100	ppm	Cal Gas Exp Date: November 4, 2028
Cal Gas Cylinder #:	EB0065608		
Removed Cal Gas Conc:	59,100	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Make/Model:	Teledyne API T700		Serial Number: 2658
N2 Gen Make/Model:	Peak Scientific		Serial Number: 771048317

Analyzer Information

Analyzer make:	API T360	Analyzer serial #:	489
Analyzer Range	0 - 2,000 ppm		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001022	1.000819	Backgd or Offset:	-0.068	-0.068
Calibration intercept:	-5.080000	-4.740000	Coeff or Slope:	0.961	0.961

CO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	3000	0.0	0.0	-0.6	----
As found High Point	2920	80.0	1576.0	1573.0	1.002
As found Mid Point					
As found Low Point					
New cylinder response					
Baseline Corr As found:	1573.6	Prev response:	1572.5	*% change:	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	3000	0.0	0.0	-0.1	----
High point	2920	80.0	1576.0	1577.0	0.999
Mid point	2960	40.0	788.0	775.2	1.017
Low point	2980	20.0	394.0	389.2	1.012
As left zero	3000	0.0	0.0	-0.6	----
As left span	2930	80.0	1570.8	1573.0	0.999
Average Correction Factor					1.009

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


CO₂ Calibration Summary

Station Information

Calibration Date	December 18, 2024	Previous Calibration	November 19, 2024
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	12:04	End Time (MST)	15:00
Analyzer make	API T360	Analyzer serial #	489

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999914	≥0.995
1576.0	1577.0	0.9994	Slope	1.000819	0.90 - 1.10
788.0	775.2	1.0165	Intercept	-4.7	+/-20
394.0	389.2	1.0123			



CO₂ Calibration Plot

Date: December 18, 2024

Location: Stony Mountain






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS19 FIREBAG DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025







Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	December 16, 2024	Last Cal Date:	November 18, 2024
Start time (MST):	12:09	End time (MST):	15:47
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.29	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC716618			
Removed Cal Gas Conc:	49.29	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	1607
Zero Air Gen Model:	Teledyne API T701H		Serial Number:	201

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1410661308
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002779	1.003024	Backgd or Offset:	10.5	10.5
Calibration intercept:	0.417366	0.856640	Coeff or Slope:	0.981	0.981

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	4999	0.0	0.0	0.1	----
As found High point	4919	81.1	799.5	803.0	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	802.9	Previous response	802.1	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	4999	0.0	0.0	0.1	----
High point	4919	81.1	799.5	802.0	0.997
Mid point	4959	40.6	400.3	403.8	0.991
Low point	4980	20.3	200.1	201.6	0.993
As left zero	4999	0.0	0.0	0.1	----
As left span	4919	81.1	799.5	806.0	0.992
Average Correction Factor:					0.994

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 16, 2024	Previous Calibration:	November 18, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	12:09	End Time (MST):	15:47
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999991	≥0.995
799.5	802.0	0.9968	Slope	1.003024	0.90 - 1.10
400.3	403.8	0.9913	Intercept	0.856640	+/-30
200.1	201.6	0.9926			



SO2 Calibration Plot

Date: December 16, 2024

Location: Firebag





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Firebag	Station number: AMS 19
Calibration Date: December 6, 2024	Last Cal Date: November 26, 2024
Start time (MST): 11:11	End time (MST): 15:17
Reason: Routine	

Calibration Standards

Cal Gas Concentration: 5.29 ppm	Cal Gas Exp Date: March 19, 2027
Cal Gas Cylinder #: DT0010492	
Removed Cal Gas Conc: 5.29 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: Teledyne API T700	Serial Number: 1607
ZAG Make/Model: Teledyne API T701	Serial Number: 201

Analyzer Information

Analyzer make: Thermo 43i-TLE	Analyzer serial #: 1151680032
Converter make: Global	Converter serial #: 2022-222
Analyzer Range: 0 - 100 ppb	Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994189	0.998904	Backgd or Offset:	2.94	2.98
Calibration intercept:	0.160000	0.220000	Coeff or Slope:	1.205	1.215

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4924	75.6	80.0	80.1	0.999
As found Mid point	4962	37.8	40.0	40.3	0.992
As found Low point	4981	18.9	20.0	20.1	0.995
New cylinder response					
Baseline Corr As found:	80.1	Prev response:	79.68	*% change:	0.5%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.001476	AF Intercept:	0.080000
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999988	* = +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4924	75.6	80.0	80.0	1.000
Mid point	4962	37.8	40.0	40.4	0.990
Low point	4981	18.9	20.0	20.2	0.990
As left zero	5000	0.0	0.0	0.2	----
As left span	4924	75.6	80.0	79.5	1.006
SO2 Scrubber Check	4922	78.3	783.0	0.0	----
Date of last scrubber change:	18-Jan-23		Ave Corr Factor		0.993
Date of last converter efficiency test:	November 26, 2024		106.2% efficiency		

Notes: Changed sample inlet filter after as founds. Completed scrubber test after calibrator zero. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 26, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:11	End Time (MST):	15:17
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999977	≥ 0.995
80.0	80.0	0.9998	Slope	0.998904	$0.90 - 1.10$
40.0	40.4	0.9899	Intercept	0.220000	± 3
20.0	20.2	0.9899			



H2S Calibration Plot

Date: December 6, 2024

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name:	Firebag	Station number:	AMS 19
Calibration Date:	December 16, 2024	Last Cal Date:	November 18, 2024
Start time (MST):	12:09	End time (MST):	15:47
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC716618	Cal Gas Expiry Date:	February 23, 2025
CH4 Cal Gas Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
C3H8 Cal Gas Conc.	205.9 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.7 ppm	CH4 Equiv Conc.	1066.9 ppm
Removed C3H8 Conc.	205.9 ppm	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	1607
ZAG Make/Model:	Teledyne API T701H	Serial Number:	201

Analyzer Information

Analyzer make: Thermo 51i-LT	Analyzer serial #: 1336160089
Analyzer Range: 0 - 20 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998247	0.993942	Background:	1.93	1.91
Calibration intercept:	-0.033912	0.054452	Coefficient:	3.877	3.851

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted
					Correction factor (Cc/Ic-AFzero)
					<i>Limit = 0.90-1.10</i>
As found zero	4999	0.0	0.00	0.10	----
As found High point	4919	81.1	17.31	17.43	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.33	Previous response	17.24	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)
					<i>Limit = 0.95-1.05</i>
Calibrator zero	4999	0.0	0.00	0.01	----
High point	4919	81.1	17.31	17.25	1.003
Mid point	4959	40.6	8.66	8.64	1.003
Low point	4980	20.3	4.33	4.43	0.978
As left zero	5000	0.0	0.00	0.13	----
As left span	4919	81.1	17.31	17.45	0.992
Average Correction Factor					0.995

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 16, 2024	Previous Calibration:	November 18, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	12:09	End Time (MST):	15:47
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999954	≥0.995
17.31	17.25	1.0032	Slope	0.993942	0.90 - 1.10
8.66	8.64	1.0028	Intercept	0.054452	+/-1.5
4.33	4.43	0.9775			



THC Calibration Plot

Date: December 16, 2024

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Firebag
 Station number: AMS 19
 Calibration Date: December 17, 2024
 Last Cal Date: November 12, 2024
 Start time (MST): 11:18
 End time (MST): 15:36
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0044018
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701H
 Cal Gas Expiry Date: November 3, 2031
 NO Cal Gas Conc: 48.70 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 48.70 ppm
 NO gas Diff:
 Serial Number: 1607
 Serial Number: 201

As Found Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
AF High point	4918	82.1	802.9	799.7	3.3	820.0	815.0	4.5	0.9790	0.9809
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 804.5 ppb		NO = 801.2 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 1.9%	
Baseline Corr 1st pt	NO _x = 820.2 ppb		NO = 815.2 ppb			<u>As Found Statistics</u>		*Percent Change	NO = 1.7%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1410661309

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000536	0.999383
NO _x Cal Offset:	1.140119	0.920189
NO Cal Slope:	1.001187	1.001559
NO Cal Offset:	0.600188	0.219979
NO ₂ Cal Slope:	1.003587	1.007368
NO ₂ Cal Offset:	0.606716	-0.236154

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.918	0.903	NO bkgnd or offset:	4.6	4.5
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	4.6	4.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.5	164.2

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4918	82.1	802.9	799.7	3.3	803.0	801.0	2.0	0.9999	0.9983
Mid point	4959	41.1	402.0	400.3	1.6	402.8	401.3	1.4	0.9979	0.9975
Low point	4980	20.5	200.5	199.7	0.8	202.4	200.4	2.0	0.9906	0.9964
As left zero	5000	0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
As left span	4918	82.1	802.9	413.6	389.3	804.0	413.6	390.5	0.9987	1.0000
Average Correction Factor									0.9961	0.9974

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	796.8	413.6	386.5	389.1	0.9933	100.7%
Mid GPT point	796.8	606.1	194.0	195.4	0.9928	100.7%
Low GPT point	796.8	702.9	97.2	97.2	0.9998	100.0%
Average Correction Factor					0.9953	100.5%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 17, 2024	Previous Calibration:	November 12, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:18	End Time (MST):	15:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999994	<i>≥0.995</i>
802.9	803.0	0.9999	Slope	0.999383	<i>0.90 - 1.10</i>
402.0	402.8	0.9979	Intercept	0.920189	<i>+/-20</i>
200.5	202.4	0.9906			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 17, 2024	Previous Calibration:	November 12, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:18	End Time (MST):	15:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999996	<i>≥0.995</i>
386.5	389.1	0.9933	Slope	1.007368	<i>0.90 - 1.10</i>
194.0	195.4	0.9928	Intercept	-0.236154	<i>+/-20</i>
97.2	97.2	0.9998			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 17, 2024	Previous Calibration:	November 12, 2024
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:18	End Time (MST):	15:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	1.000000	<i>≥0.995</i>
799.7	801.0	0.9983	Slope	1.001559	<i>0.90 - 1.10</i>
400.3	401.3	0.9975	Intercept	0.219979	<i>+/-20</i>
199.7	200.4	0.9964			



NO_x Calibration Plot

Date: December 17, 2024

Location: Firebag






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS20 MACKAY RIVER DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	MacKay River	Station number:	AMS 20
Calibration Date:	December 5, 2024	Last Cal Date:	November 19, 2024
Start time (MST):	8:53	End time (MST):	11:46
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.22	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC30686			
Removed Cal Gas Conc:	49.22	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	1220
Zero Air Gen Model:	API 701		Serial Number:	4522

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1501301450
Analyzer Range:	0-1000ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996067	0.996980	Backgd or Offset:	18.9	19.3
Calibration intercept:	3.811121	3.691185	Coeff or Slope:	0.930	0.939

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4919	81.3	800.3	792.7	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	792.5	Previous response	800.9	*% change	-1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.6	----
High point	4919	81.3	800.3	800.4	1.000
Mid point	4959	40.7	400.7	403.5	0.993
Low point	4980	20.3	199.8	206.8	0.966
As left zero	5000	0.0	0.0	0.6	----
As left span	4919	81.3	800.3	799.3	1.001
Average Correction Factor:					0.986

Notes: No Maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 19, 2024
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	8:53	End Time (MST):	11:46
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	0.6	----	Correlation Coefficient	0.999925	≥0.995
800.3	800.4	0.9998	Slope	0.996980	0.90 - 1.10
400.7	403.5	0.9930	Intercept	3.691185	+/-30
199.8	206.8	0.9663			



SO2 Calibration Plot

Date: December 5, 2024

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	MacKay River	Station number:	AMS 20
Calibration Date:	December 4, 2024	Last Cal Date:	November 13, 2024
Start time (MST):	8:28	End time (MST):	12:33
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.12	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC515997			
Removed Cal Gas Conc:	5.12	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	1220
ZAG Make/Model:	API 701		Serial Number:	4522

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117
Converter make:	Global	Converter serial #:	2022-226
Analyzer Range	0 - 100 ppb	Converter Temp:	325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.981587	0.968153	Backgd or Offset:	3.43	3.43
Calibration intercept:	0.599198	0.718985	Coeff or Slope:	1.051	1.051

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4922	78.1	80.0	78.6	1.019
As found Mid point	4961	39.0	39.9	39.8	1.006
As found Low point	4980	19.5	20.0	20.5	0.979
New cylinder response					
Baseline Corr As found:	78.5	Prev response:	79.10	*% change:	-0.8%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.978872	AF Intercept:	0.519148
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999869	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4922	78.1	80.0	77.9	1.027
Mid point	4961	39.0	39.9	39.7	1.006
Low point	4980	19.5	20.0	20.4	0.979
As left zero	5000	0.0	0.0	0.4	----
As left span	4922	78.1	80.0	78.4	1.020
SO ₂ Scrubber Check	4982	81.3	802.8	0.0	----
Date of last scrubber change:	25-May-23			Ave Corr Factor	1.004
Date of last converter efficiency test:					

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	December 4, 2024	Previous Calibration:	November 13, 2024
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	8:28	End Time (MST):	12:33
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999861	≥ 0.995
80.0	77.9	1.0266	Slope	0.968153	$0.90 - 1.10$
39.9	39.7	1.0059	Intercept	0.718985	± 3
20.0	20.4	0.9789			



H₂S Calibration Plot

Date: December 4, 2024

Location: MacKay River





Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 19, 2024
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	8:53	End Time (MST):	11:45
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.03	----	Correlation Coefficient	0.999975	≥0.995
17.34	17.29	1.0031	Slope	0.996948	0.90 - 1.10
8.68	8.64	1.0043	Intercept	0.004814	+/-1.5
4.33	4.37	0.9898			



THC Calibration Plot

Date: December 5, 2024

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name:	MacKay River	Station number:	AMS 20
Calibration Date:	December 12, 2024	Last Cal Date:	December 5, 2024
Start time (MST):	10:40	End time (MST):	13:50
Reason:	Maintenance		

Calibration Standards

Gas Cert Reference:	CC306868	Cal Gas Expiry Date:	February 23, 2024
CH4 Cal Gas Conc.	499.4 ppm	CH4 Equiv Conc.	1066.5 ppm
C3H8 Cal Gas Conc.	206.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	499.4 ppm	CH4 Equiv Conc.	1066.5 ppm
Removed C3H8 Conc.	206.2 ppm	Diff between cyl:	
Calibrator Make/Model:	API T700	Serial Number:	1220
ZAG Make/Model:	API 701	Serial Number:	4522

Analyzer Information

Analyzer make: Thermo 51i-LT	Analyzer serial #: 1501663727
Analyzer Range: 0 - 20 ppm	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996948	0.992322	Background:	4.690	4.140
Calibration intercept:	0.004814	0.086409	Coefficient:	5.924	5.026

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.03	----
As found High point	4919	81.3	17.34	19.34	0.898
As found Mid point	4959	40.7	8.68	9.76	0.892
As found Low point	4979	20.3	4.33	5.05	0.864
New cylinder response					
Baseline Corr As found:	19.31	Previous response	17.29	*% change	10.4%
Baseline Corr 2nd AF pt:	9.73	AF Slope:	1.110141	AF Intercept:	0.122207
Baseline Corr 3rd AF pt:	5.01	AF Correlation:	0.999888	<i>* = > +/-5% change initiates investigation</i>	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.03	----
High point	4919	81.3	17.34	17.27	1.004
Mid point	4959	40.7	8.68	8.70	0.997
Low point	4980	20.3	4.33	4.45	0.972
As left zero	5000	0.0	0.00	-0.07	----
As left span	4919	81.3	17.34	17.26	1.005
Average Correction Factor					0.991

Notes: Pump replaced after as founds. Zero and span adjusted

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	December 5, 2024
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	10:40	End Time (MST):	13:50
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.03	----	Correlation Coefficient	0.999951	≥0.995
17.34	17.27	1.0038	Slope	0.992322	0.90 - 1.10
8.68	8.70	0.9974	Intercept	0.086409	+/-1.5
4.33	4.45	0.9725			



THC Calibration Plot

Date: December 12, 2024

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: MacKay River
 Station number: AMS 20
 Calibration Date: December 3, 2024
 Last Cal Date: November 14, 2024
 Start time (MST): 7:55
 End time (MST): 11:58
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T376265
 NOX Cal Gas Conc: 49.19 ppm
 Removed Cylinder #:
 Removed Gas NOX Conc: 49.19 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701
 Cal Gas Expiry Date: April 13, 2025
 NO Cal Gas Conc: 48.04 ppm
 Removed Gas Exp Date:
 Removed Gas NO Conc: 48.04 ppm
 NO gas Diff:
 Serial Number: 1220
 Serial Number: 4522

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
AF High point	4917	83.3	819.5	800.3	19.2	819.9	797.5	22.5	0.9993	1.0035
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 820.6 ppb		NO = 800.5 ppb			<i>* = > +/-5% change initiates investigation</i>			*Percent Change	NO _x = -0.1%
Baseline Corr 1st pt	NO _x = 820.0 ppb		NO = 797.5 ppb			<u>As Found Statistics</u>			*Percent Change	NO = -0.4%
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :		Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :		NO SI:	NO Int:
						As found	NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1505164379

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996324	0.997202
NO _x Cal Offset:	4.202261	4.362344
NO Cal Slope:	0.996144	0.995687
NO Cal Offset:	3.242757	3.402729
NO ₂ Cal Slope:	1.002283	1.002574
NO ₂ Cal Offset:	-0.864534	-0.544534

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.001	1.001	NO bkgnd or offset:	2.8	2.8
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	3.0	3.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	161.5	161.5

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
High point	4917	83.3	819.5	800.3	19.2	819.6	798.8	20.8	0.9998	1.0019
Mid point	4958	41.7	410.3	400.7	9.6	414.9	403.4	11.5	0.9888	0.9933
Low point	4979	20.8	204.6	199.9	4.8	213.3	206.1	7.1	0.9594	0.9697
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
As left span	4917	83.3	819.5	436.6	382.9	817.0	436.6	380.3	1.0030	1.0000
Average Correction Factor									0.9827	0.9883

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	795.6	436.2	378.6	379.2	0.9983	100.2%
Mid GPT point	795.6	608.6	206.2	206.1	1.0003	100.0%
Low GPT point	795.6	698.5	116.3	115.4	1.0074	99.3%
Average Correction Factor					1.0020	99.8%

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 14, 2024
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	7:55	End Time (MST):	11:58
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999867	<i>≥0.995</i>
819.5	819.6	0.9998	Slope	0.997202	<i>0.90 - 1.10</i>
410.3	414.9	0.9888	Intercept	4.362344	<i>+/-20</i>
204.6	213.3	0.9594			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 14, 2024
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	7:55	End Time (MST):	11:58
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
378.6	379.2	0.9983	Slope	1.002574	<i>0.90 - 1.10</i>
206.2	206.1	1.0003	Intercept	-0.544534	<i>+/-20</i>
116.3	115.4	1.0074			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 14, 2024
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	7:55	End Time (MST):	11:58
Analyzer make:	Thermo 42i	Analyzer serial #:	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999920	<i>≥0.995</i>
800.3	798.8	1.0019	Slope	0.995687	<i>0.90 - 1.10</i>
400.7	403.4	0.9933	Intercept	3.402729	<i>+/-20</i>
199.9	206.1	0.9697			



NO_x Calibration Plot

Date: December 3, 2024

Location: MacKay River






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS21
CONKLIN
DECEMBER 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	December 12, 2024	Last Cal Date:	November 28, 2024
Start time (MST):	9:51	End time (MST):	13:33
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.34	ppm	Cal Gas Exp Date:	October 9, 2023
Cal Gas Cylinder #:	CC340840			
Removed Cal Gas Conc:	49.93	ppm	Rem Gas Exp Date:	January 5, 2025
Removed Gas Cyl #:	CC259455		Diff between cyl:	-1.9%
Calibrator Model:	Teledyne API T700		Serial Number:	5252
Zero Air Gen Model:	Teledyne API T701		Serial Number:	953

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1428701363
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999503	0.998751	Backgd or Offset:	28.6	29.3
Calibration intercept:	-0.024227	-0.083970	Coeff or Slope:	0.886	0.899

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	80.2	800.8	802.3	0.998
As found Mid point					
As found Low point					
New cylinder response	4920	80.2	807.4	793.5	1.018
Baseline Corr As found:	802.3	Previous response	800.4	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	80.2	807.4	806.6	1.001
Mid point	4960	40.1	403.7	402.6	1.003
Low point	4980	20.0	201.8	201.4	1.002
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	80.2	807.4	807.2	1.000
Average Correction Factor:					1.002

Notes: Sample inlet filter and SO₂ cylinder was changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 28, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:51	End Time (MST):	13:33
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999999	≥0.995
807.4	806.6	1.0010	Slope	0.998751	0.90 - 1.10
403.7	402.6	1.0028	Intercept	-0.083970	+/-30
201.8	201.4	1.0018			



SO2 Calibration Plot

Date: December 12, 2024

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	December 2, 2024	Last Cal Date:	November 13, 2024
Start time (MST):	10:15	End time (MST):	14:56
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.00	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC501204			
Removed Cal Gas Conc:	5.00	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne T700		Serial Number:	5252
ZAG Make/Model:	Teledyne T701		Serial Number:	953

Analyzer Information

Analyzer make:	Thermo 43i-QTL	Analyzer serial #:	12228021058
Converter make:	CD-Nova 101	Converter serial #:	565
Analyzer Range	0 - 100 ppb	Converter Temp:	850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001714	1.011143	Backgd or Offset:	3.1	3.1
Calibration intercept:	-0.060000	-0.040000	Coeff or Slope:	1.517	1.556

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	80.0	80.0	78.8	1.015
As found Mid point	4960	40.0	40.0	39.5	1.013
As found Low point	4980	20.0	20.0	19.7	1.015
New cylinder response					
Baseline Corr As found:	78.8	Prev response:	80.08	*% change:	-1.6%
Baseline Corr 2nd AF pt:	39.5	AF Slope:	0.985143	AF Intercept:	0.020000
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999998	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4920	80.0	80.0	81.0	0.988
Mid point	4960	40.0	40.0	40.1	0.998
Low point	4980	20.0	20.0	20.2	0.990
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	80.0	80.0	81.3	0.984
SO2 Scrubber Check	4920	80.2	802.0	0.1	----
Date of last scrubber change:	November 13, 2024			Ave Corr Factor	0.992
Date of last converter efficiency test:					

Notes: Sample inlet filter was changed after multipoint as founds. Sox scrubber check done after calibrator zero and passed. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 13, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:15	End Time (MST):	14:56
Analyzer make:	Thermo 43i-QTL	Analyzer serial #:	12228021058

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999962	≥ 0.995
80.0	81.0	0.9877	Slope	1.011143	$0.90 - 1.10$
40.0	40.1	0.9975	Intercept	-0.040000	± 3
20.0	20.2	0.9901			



TRS Calibration Plot

Date: December 2, 2024

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	December 12, 2024	Last Cal Date:	November 18, 2024
Start time (MST):	9:51	End time (MST):	13:33
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC340840	Cal Gas Expiry Date:	October 9, 2032
CH4 Cal Gas Conc.	503.8 ppm	CH4 Equiv Conc.	1067.6 ppm
C3H8 Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	CC259455	Removed Gas Expiry:	January 5, 2025
Removed CH4 Conc.	497.9 ppm	CH4 Equiv Conc.	1067.7 ppm
Removed C3H8 Conc.	207.2 ppm	Diff between cyl (THC):	0.8%
Diff between cyl (CH ₄):	-0.7%	Diff between cyl (NM):	2.2%
Calibrator Model:	Teledyne API T700	Serial Number:	5252
Zero Air Gen model:	Teledyne API T701	Serial Number:	953

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1180320039
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.27E-04	2.30E-04	NMHC SP Ratio:	4.60E-05	4.73E-05
CH4 Retention time:	15.0	15.2	NMHC Peak Area:	198650	190954
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	17.13	16.62	1.031
As found Mid point					
As found Low point					
New cylinder response	4920	80.2	17.12	16.74	1.023
Baseline Corr AF:	16.62	Prev response	17.06	*% change	-2.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	17.12	17.11	1.001
Mid point	4960	40.1	8.56	8.54	1.003
Low point	4980	20.0	4.28	4.29	0.998
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	17.12	17.11	1.001
Average Correction Factor					1.001

Notes: Sample inlet filter and SO₂ cylinder was changed after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	9.14	8.66	1.055
As found Mid point					
As found Low point					
New cylinder response	4920	80.2	9.04	8.75	1.033
Baseline Corr AF:	8.66	Prev response	9.09	*% change	-4.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	9.04	9.03	1.001
Mid point	4960	40.1	4.52	4.53	0.998
Low point	4980	20.0	2.26	2.28	0.989
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	9.04	9.04	1.001
Average Correction Factor					0.996

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.2	7.99	7.95	1.004
As found Mid point					
As found Low point					
New cylinder response	4920	80.2	8.08	7.99	1.012
Baseline Corr AF:	7.95	Prev response	7.98	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.2	8.08	8.08	1.000
Mid point	4960	40.1	4.04	4.01	1.008
Low point	4980	20.0	2.02	2.00	1.009
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.2	8.08	8.08	1.000
Average Correction Factor					1.006

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.992600	0.998793
THC Cal Offset:	0.064362	0.000765
CH ₄ Cal Slope:	1.000525	0.999952
CH ₄ Cal Offset:	-0.013244	-0.012639
NMHC Cal Slope:	0.985800	0.997884
NMHC Cal Offset:	0.076606	0.013405

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 18, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:51	End Time (MST):	13:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999997 ≥0.995
17.12	17.11	1.0009	Slope	0.998793 0.90 - 1.10
8.56	8.54	1.0030	Intercept	0.000765 +/-0.5
4.28	4.29	0.9983		





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 18, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:51	End Time (MST):	13:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999982	<i>≥0.995</i>
8.08	8.08	1.0003	Slope	0.999952	<i>0.90 - 1.10</i>
4.04	4.01	1.0078	Intercept	-0.012639	<i>+/-0.5</i>
2.02	2.00	1.0086			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 18, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:51	End Time (MST):	13:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999989	≥ 0.995
9.04	9.03	1.0012	Slope	0.997884	$0.90 - 1.10$
4.52	4.53	0.9985	Intercept	0.013405	± 0.5
2.26	2.28	0.9893			



NMHC Calibration Plot

Date: December 12, 2024

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Conklin
 Station number: AMS 21
 Calibration Date: December 11, 2024
 Last Cal Date: November 27, 2024
 Start time (MST): 9:20
 End time (MST): 13:42
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: SA18828
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701H
 Cal Gas Expiry Date: November 3, 2031
 NO Cal Gas Conc: 48.80 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 48.80 ppm
 NO gas Diff:
 Serial Number: 5252
 Serial Number: 953

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
AF High point	4918	82.0	802.0	800.3	1.6	761.0	757.0	4.0	1.0538	1.0572
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 802.3 ppb	NO = 801.0 ppb	<i>* = > +/-5% change initiates investigation</i>				*Percent Change	NO _x = -5.4%		
Baseline Corr 1st pt	NO _x = 761.0 ppb	NO = 757.0 ppb	<u>As Found Statistics</u>				*Percent Change	NO = -5.8%		
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :		Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :		NO SI:	NO Int:			
			As found	NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1501663731

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999101	0.993172
NO _x Cal Offset:	1.088026	1.067977
NO Cal Slope:	1.001291	0.994694
NO Cal Offset:	-0.312001	0.047961
NO ₂ Cal Slope:	1.003757	0.999973
NO ₂ Cal Offset:	-0.410756	0.024329

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.982	1.034	NO bkgnd or offset:	9.6	10.1
NOX coeff or slope:	0.995	0.996	NOX bkgnd or offset:	9.5	10.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	148.8	148.8

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
High point	4918	82.0	802.0	800.3	1.6	796.9	796.1	0.8	1.0063	1.0053
Mid point	4959	41.0	401.0	400.2	0.8	400.3	398.1	2.2	1.0017	1.0052
Low point	4980	20.5	200.5	200.1	0.4	200.8	199.1	1.7	0.9984	1.0048
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
As left span	4918	82.0	802.0	436.7	365.3	793.7	436.7	357.0	1.0104	1.0000
Average Correction Factor									1.0021	1.0051

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	791.0	423.9	368.7	368.8	0.9998	100.0%
Mid GPT point	791.0	603.6	189.0	189.0	1.0002	100.0%
Low GPT point	791.0	702.5	90.1	90.1	1.0004	100.0%
Average Correction Factor					1.0002	100.0%

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 27, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:20	End Time (MST):	13:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
802.0	796.9	1.0063	Slope	0.993172	<i>0.90 - 1.10</i>
401.0	400.3	1.0017	Intercept	1.067977	<i>+/-20</i>
200.5	200.8	0.9984			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 27, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:20	End Time (MST):	13:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	1.000000	<i>≥0.995</i>
368.7	368.8	0.9998	Slope	0.999973	<i>0.90 - 1.10</i>
189.0	189.0	1.0002	Intercept	0.024329	<i>+/-20</i>
90.1	90.1	1.0004			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 27, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:20	End Time (MST):	13:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	1.000000	<i>≥0.995</i>
800.3	796.1	1.0053	Slope	0.994694	<i>0.90 - 1.10</i>
400.2	398.1	1.0052	Intercept	0.047961	<i>+/-20</i>
200.1	199.1	1.0048			



NO_x Calibration Plot

Date: December 11, 2024

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Conklin	Station number:	AMS 21
Calibration Date:	December 2, 2024	Last Cal Date:	November 7, 2024
Start time (MST):	10:15	End time (MST):	13:28
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	281
Calibrator Make/Model:	Teledyne API T750	Serial Number:	77
ZAG Make/Model:	Teledyne API 751H		

Analyzer Information

Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006886	1.000571	Backgd or Offset:	2.3	1.3
Calibration intercept:	-0.980000	-0.500000	Coeff or Slope:	1.119	1.097

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-1.1	----
As found High point	5000	1095.7	400.0	401.7	0.993
As found Mid point					
As found Low point					
Baseline Corr As found:	402.8	Previous response	401.8	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	-0.1	----
High point	5000	1096.6	400.0	399.9	1.000
Mid point	5000	813.9	200.0	199.5	1.003
Low point	5000	685.3	100.0	99.1	1.009
As left zero	5000	800.0	0.0	-0.6	----
As left span	5000	908.4	400.0	409.9	0.976
Average Correction Factor					1.004

Notes: Sample inlet filter changed after as founds. Adjusted zero and span.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 7, 2024
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:15	End Time (MST):	13:28
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999995	≥0.995
400.0	399.9	1.0003	Slope	1.000571	0.90 - 1.10
200.0	199.5	1.0025	Intercept	-0.500000	+/- 5
100.0	99.1	1.0091			



O₃ Calibration Plot

Date: December 2, 2024

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Conklin Station number: AMS 21
 Calibration Date: December 11, 2024 Last Cal Date: November 27, 2024
 Start time (MST): 12:26 End time (MST): 12:50

Analyzer Make: API T640 S/N: 326
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-14.70	-15.41	-14.70	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	713.20	714.91	713.20	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	5.06	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	43	----	43	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.20	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: September 29, 2024
Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: November 27, 2024
Date Disposable Filter Changed: November 27, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 9, 2024
Date RH/T Sensor Cleaned: August 9, 2024

Notes: Verified flow, pressure, temperature and pump power. No adjustment needed. Leak check passed.

Calibration by: Jan Castro




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER
DECEMBER 2024**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	December 19, 2024	Last Cal Date:	November 18, 2024
Start time (MST):	12:53	End time (MST):	16:00
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.11	ppm	Cal Gas Exp Date: January 18, 2029
Cal Gas Cylinder #:	CC281519		
Removed Cal Gas Conc:	50.11	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	Teledyne API T700		Serial Number: 3806
Zero Air Gen Model:	Teledyne API T701		Serial Number: 691

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1152430006
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002936	1.000706	Backgd or Offset:	25.0	25.2
Calibration intercept:	0.043822	-0.075626	Coeff or Slope:	1.024	1.046

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4920	79.8	799.8	780.3	1.025
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	780.2	Previous response	802.2	*% change	-2.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	79.8	799.8	800.2	0.999
Mid point	4960	39.9	399.9	400.4	0.999
Low point	4980	20.0	200.4	200.2	1.001
As left zero	5000	0.0	0.0	0.3	----
As left span	4920	79.8	799.8	804.8	0.994
Average Correction Factor:					1.000

Notes: Changed the inlet filter after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 19, 2024	Previous Calibration:	November 18, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:53	End Time (MST):	16:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999999	≥0.995
799.8	800.2	0.9995	Slope	1.000706	0.90 - 1.10
399.9	400.4	0.9987	Intercept	-0.075626	+/-30
200.4	200.2	1.0012			



SO2 Calibration Plot

Date: December 19, 2024

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	December 20, 2024	Last Cal Date:	November 21, 2024
Start time (MST):	12:53	End time (MST):	16:42
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.02	ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	CC424047			
Removed Cal Gas Conc:	5.02	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700		Serial Number:	3806
ZAG Make/Model:	Teledyne API T701		Serial Number:	691

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031
Converter make:	CDN-101	Converter serial #:	620
Analyzer Range	0 - 100 ppb	Converter Temp:	850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.987240	0.979100	Backgd or Offset:	3.74	3.77
Calibration intercept:	0.280383	0.440224	Coeff or Slope:	1.197	1.197

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4920	79.7	80.0	79.0	1.016
As found Mid point	4960	39.8	40.0	40.2	0.999
As found Low point	4980	19.9	20.0	20.0	1.009
New cylinder response					
Baseline Corr As found:	78.8	Prev response:	79.28	*% change:	-0.6%
Baseline Corr 2nd AF pt:	40.0	AF Slope:	0.985094	AF Intercept:	0.380470
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999915	<i>* = +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4920	79.7	80.0	78.7	1.017
Mid point	4960	39.8	40.0	39.8	1.004
Low point	4980	19.9	20.0	19.9	1.004
As left zero	5000	0.0	0.0	0.6	----
As left span	4920	79.7	80.0	79.1	1.012
SO2 Scrubber Check	4920	79.8	798.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.008
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Scrubber test performed after zero point, no issues. No adjustments were made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	December 20, 2024	Previous Calibration:	November 21, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:53	End Time (MST):	16:42
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999978	≥ 0.995
80.0	78.7	1.0168	Slope	0.979100	$0.90 - 1.10$
40.0	39.8	1.0040	Intercept	0.440224	± 3
20.0	19.9	1.0040			



TRS Calibration Plot

Date: December 20, 2024

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	December 19, 2024	Last Cal Date:	November 18, 2024
Start time (MST):	12:53	End time (MST):	16:00
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH4 Cal Gas Conc.	502.8 ppm	CH4 Equiv Conc.	1075.9 ppm
C3H8 Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	502.8 ppm	CH4 Equiv Conc.	1075.9 ppm
Removed C3H8 Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
Zero Air Gen model:	Teledyne API 701	Serial Number:	691

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1317958219
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.49E-04	2.44E-04	NMHC SP Ratio:	5.95E-05	5.75E-05
CH4 Retention time:	11.6	11.6	NMHC Peak Area:	153693	159068
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	17.17	17.51	0.981
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.51	Prev response	17.14	*% change	2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	17.17	17.13	1.003
Mid point	4960	39.9	8.59	8.48	1.012
Low point	4980	20.0	4.30	4.24	1.015
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	17.17	17.13	1.003
Average Correction Factor					1.010

Notes: Changed the inlet filter and N2 cylinder after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.15	9.36	0.978
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.36	Prev response	9.13	*% change	2.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.15	9.11	1.005
Mid point	4960	39.9	4.57	4.53	1.009
Low point	4980	20.0	2.29	2.27	1.009
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.15	9.12	1.003
Average Correction Factor					1.008

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	8.03	8.15	0.984
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.15	Prev response	8.01	*% change	1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.03	8.02	1.001
Mid point	4960	39.9	4.01	3.95	1.015
Low point	4980	20.0	2.01	1.97	1.022
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.03	8.01	1.002
Average Correction Factor					1.013

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999803	0.997965
THC Cal Offset:	-0.027392	-0.037583
CH ₄ Cal Slope:	1.001053	0.999957
CH ₄ Cal Offset:	-0.024358	-0.027756
NMHC Cal Slope:	0.999092	0.995631
NMHC Cal Offset:	-0.003835	-0.008227

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 19, 2024	Previous Calibration:	November 18, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:53	End Time (MST):	16:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999970	≥ 0.995
17.17	17.13	1.0026	Slope	0.997965	$0.90 - 1.10$
8.59	8.48	1.0121	Intercept	-0.037583	± 0.5
4.30	4.24	1.0152			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 19, 2024	Previous Calibration:	November 18, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:53	End Time (MST):	16:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999930	<i>≥0.995</i>
8.03	8.02	1.0010	Slope	0.999957	<i>0.90 - 1.10</i>
4.01	3.95	1.0153	Intercept	-0.027756	<i>+/-0.5</i>
2.01	1.97	1.0220			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 19, 2024	Previous Calibration:	November 18, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:53	End Time (MST):	16:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999993	<i>≥0.995</i>
9.15	9.11	1.0045	Slope	0.995631	<i>0.90 - 1.10</i>
4.57	4.53	1.0091	Intercept	-0.008227	<i>+/-0.5</i>
2.29	2.27	1.0090			



NMHC Calibration Plot

Date: December 19, 2024

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	December 23, 2024	Last Cal Date:	December 19, 2024
Start time (MST):	11:52	End time (MST):	14:21
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281519	Cal Gas Expiry Date:	January 18, 2029
CH4 Cal Gas Conc.	502.8 ppm	CH4 Equiv Conc.	1075.9 ppm
C3H8 Cal Gas Conc.	208.4 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH4 Conc.	502.8 ppm	CH4 Equiv Conc.	1075.9 ppm
Removed C3H8 Conc.	208.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API 700	Serial Number:	3806
Zero Air Gen model:	Teledyne API 701	Serial Number:	691

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1317958219
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.44E-04	2.44E-04	NMHC SP Ratio:	5.75E-05
CH4 Retention time:	11.6	11.6	NMHC Peak Area:	159068
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	17.17	16.87	1.018
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.87	Prev response	17.10	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	17.17	16.62	1.033
Average Correction Factor					1.033

Notes: Swapped out the H2 cylinder after as founds.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.15	8.94	1.024
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.94	Prev response	9.10	*% change	-1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.15	8.78	1.042
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	8.03	7.93	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.93	Prev response	8.00	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.03	7.83	1.024
Average Correction Factor					

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.997965	
THC Cal Offset:	-0.037583	
CH ₄ Cal Slope:	0.999957	
CH ₄ Cal Offset:	-0.027756	
NMHC Cal Slope:	0.995631	
NMHC Cal Offset:	-0.008227	

Calibration Performed By: Rene Chamberland

NMHC Calibration Plot

Date: December 23, 2024

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Janvier
 Station number: AMS 22
 Calibration Date: December 16, 2024
 Last Cal Date: November 26, 2024
 Start time (MST): 11:45
 End time (MST): 16:00
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0047765
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: March 11, 2031
 NO Cal Gas Conc: 48.80 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 48.80 ppm
 NO gas Diff:
 Serial Number: 3806
 Serial Number: 691

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
AF High point	4918	82.0	802.0	800.3	1.6	807.8	795.3	12.6	0.9929	1.0062
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 795.1 ppb		NO = 790.3 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 1.6%	
Baseline Corr 1st pt	NO _x = 807.7 ppb		NO = 795.4 ppb			<u>As Found Statistics</u>		*Percent Change	NO = 0.6%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 833

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.991246	1.002119
NO _x Cal Offset:	0.163866	1.424245
NO Cal Slope:	0.988179	0.999960
NO Cal Offset:	-0.576326	0.424054
NO ₂ Cal Slope:	1.001375	0.997993
NO ₂ Cal Offset:	-0.096837	1.353493

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.880	0.880	NO bkgnd or offset:	-0.7	-0.7
NOX coeff or slope:	0.879	0.879	NOX bkgnd or offset:	0.5	0.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	8.4	8.4

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
High point	4918	82.0	802.0	800.3	1.6	804.1	800.4	3.8	0.9973	0.9999
Mid point	4960	41.0	400.9	400.1	0.8	404.7	401.1	3.6	0.9906	0.9975
Low point	4980	20.5	200.5	200.1	0.4	203.2	200.5	2.7	0.9866	0.9978
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.2	----	----
As left span	4918	82.0	802.0	399.0	403.0	802.0	399.0	402.9	1.0000	1.0000
Average Correction Factor									0.9915	0.9984

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	800.1	396.8	404.9	404.5	1.0011	99.9%
Mid GPT point	800.1	596.3	205.4	207.8	0.9886	101.1%
Low GPT point	800.1	697.7	104.0	106.2	0.9797	102.1%
Average Correction Factor					0.9898	101.0%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 16, 2024	Previous Calibration:	November 26, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:45	End Time (MST):	16:00
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999983	<i>≥0.995</i>
802.0	804.1	0.9973	Slope	1.002119	<i>0.90 - 1.10</i>
400.9	404.7	0.9906	Intercept	1.424245	<i>+/-20</i>
200.5	203.2	0.9866			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 16, 2024	Previous Calibration:	November 26, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:45	End Time (MST):	16:00
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999931	<i>≥0.995</i>
404.9	404.5	1.0011	Slope	0.997993	<i>0.90 - 1.10</i>
205.4	207.8	0.9886	Intercept	1.353493	<i>+/-20</i>
104.0	106.2	0.9797			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 16, 2024	Previous Calibration:	November 26, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:45	End Time (MST):	16:00
Analyzer make:	Teledyne API T200	Analyzer serial #:	833

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
800.3	800.4	0.9999	Slope	0.999960	<i>0.90 - 1.10</i>
400.1	401.1	0.9975	Intercept	0.424054	<i>+/-20</i>
200.1	200.5	0.9978			



NO_x Calibration Plot

Date: December 16, 2024

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name:	Janvier	Station number:	AMS 22
Calibration Date:	December 13, 2024	Last Cal Date:	November 13, 2024
Start time (MST):	13:16	End time (MST):	15:55
Reason:	Routine		

Calibration Standards

O3 generation mode:	Photometer	Serial Number:	3806
Calibrator Make/Model:	Teledyne API T700	Serial Number:	691
ZAG Make/Model:	Teledyne API T701H		

Analyzer Information

Analyzer make:	Teledyne API T400	Analyzer serial #:	7046
Analyzer Range	0 - 500 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999771	0.996314	Backgd or Offset:	1.5	1.5
Calibration intercept:	1.440000	1.220000	Coeff or Slope:	1.011	1.011

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-0.6	----
As found High point	5000	922.9	400.0	399.1	1.001
As found Mid point					
As found Low point					
Baseline Corr As found:	399.7	Previous response	401.3	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.0	----
High point	5000	922.9	400.0	399.1	1.002
Mid point	5000	768.8	200.0	201.2	0.994
Low point	5000	656.1	100.0	102.0	0.980
As left zero	5000	800.0	0.0	0.7	----
As left span	5000	916.2	400.0	402.3	0.994
Average Correction Factor					0.992

Notes: Changed the inlet filter after the as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	December 13, 2024	Previous Calibration:	November 13, 2024
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	13:16	End Time (MST):	15:55
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999957	≥0.995
400.0	399.1	1.0023	Slope	0.996314	0.90 - 1.10
200.0	201.2	0.9940	Intercept	1.220000	+/- 5
100.0	102.0	0.9804			



O₃ Calibration Plot

Date: December 13, 2024

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Janvier Station number: AMS 22
 Calibration Date: December 20, 2024 Last Cal Date: November 21, 2024
 Start time (MST): 13:27 End time (MST): 14:35

Analyzer Make: Teledyne API T640 S/N: 325
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752
 Temp/RH standard: Alicat FP-25BT S/N: 388752

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-11.3	-11.63	-11.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	717.7	718.35	717.7	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	4.961	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	40	----	40	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.5	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input checked="" type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: _____ October 29, 2024
 Date Disposable Filter Changed: _____ October 29, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ October 29, 2024
 Date RH/T Sensor Cleaned: _____ October 29, 2024

Notes: Verified flow, temperature, and pressure. Leak checks passed.

Calibration by: Rene Chamberland




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23 FORT HILLS DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Fort Hills	Station number:	AMS 23
Calibration Date:	December 17, 2024	Last Cal Date:	November 21, 2024
Start time (MST):	8:25	End time (MST):	12:17
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.76	ppm	Cal Gas Exp Date: January 5, 2025
Cal Gas Cylinder #:	CC281425		
Removed Cal Gas Conc:	49.76	ppm	Rem Gas Exp Date:
Removed Gas Cyl #:			Diff between cyl:
Calibrator Model:	API T700		Serial Number: 451
Zero Air Gen Model:	API T701		Serial Number: 5611

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1160290012
Analyzer Range:	0-1000ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000937	0.992885	Backgd or Offset:	18.3	18.3
Calibration intercept:	-0.623634	-0.282101	Coeff or Slope:	1.056	1.056

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4920	80.3	799.1	794.4	1.006
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	794.1	Previous response	799.2	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.5	----
High point	4920	80.3	799.1	793.4	1.007
Mid point	4960	40.2	400.1	396.9	1.008
Low point	4980	20.1	200.0	197.3	1.014
As left zero	5000	0.0	0.0	0.6	----
As left span	4920	80.3	799.1	793.4	1.007
Average Correction Factor:					1.010

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 17, 2024	Previous Calibration:	November 21, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	8:25	End Time (MST):	12:17
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999995	≥0.995
799.1	793.4	1.0072	Slope	0.992885	0.90 - 1.10
400.1	396.9	1.0079	Intercept	-0.282101	+/-30
200.0	197.3	1.0138			



SO2 Calibration Plot

Date: December 17, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Fort Hills	Station number:	AMS 23
Calibration Date:	December 10, 2024	Last Cal Date:	November 1, 2024
Start time (MST):	8:00	End time (MST):	12:05
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.84	ppm	Cal Gas Exp Date:	August 28, 2027
Cal Gas Cylinder #:	DT0021910			
Removed Cal Gas Conc:	5.20	ppm	Rem Gas Exp Date:	February 5, 2024
Removed Gas Cyl #:	CC517372		Diff between cyl:	1.3%
Calibrator Make/Model:	API T700		Serial Number:	451
ZAG Make/Model:	API T701		Serial Number:	5611

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232
Converter make:	CDN-101	Converter serial #:	594
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000783	1.017684	Backgd or Offset:	2.00
Calibration intercept:	-0.198212	-0.138341	Coeff or Slope:	1.172

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4923	77.0	80.1	81.2	0.986
As found Mid point	4962	38.5	40.0	40.5	0.989
As found Low point	4981	19.2	20.0	19.9	1.003
New cylinder response	4917	82.6	80.0	82.2	0.973
Baseline Corr As found:	81.2	Prev response:	79.94	*% change:	1.5%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	1.015335	AF Intercept:	-0.157857
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999980	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4917	82.6	80.0	81.4	0.982
Mid point	4959	41.3	40.0	40.3	0.992
Low point	4979	20.7	20.0	20.1	0.997
As left zero	5000	0.0	0.0	0.1	----
As left span	4923	77.0	74.5	82.6	0.902
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.990
Date of last converter efficiency test:	March 13, 2024			110.3% efficiency	

Notes: SOx scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	December 10, 2024	Previous Calibration:	November 1, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	8:00	End Time (MST):	12:05
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999954	≥ 0.995
80.0	81.4	0.9823	Slope	1.017684	$0.90 - 1.10$
40.0	40.3	0.9920	Intercept	-0.138341	± 3
20.0	20.1	0.9970			



TRS Calibration Plot

Date: December 10, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Fort Hills	Station number:	AMS 23
Calibration Date:	December 12, 2024	Last Cal Date:	November 21, 2024
Start time (MST):	8:35	End time (MST):	9:37
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	500.2 ppm	CH4 Equiv Conc.	1070.6 ppm
C3H8 Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.2 ppm	CH4 Equiv Conc.	1070.6 ppm
Removed C3H8 Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
Zero Air Gen model:	API T701	Serial Number:	5611

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1152430012
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.00E-05	2.00E-05	NMHC SP Ratio:	7.19E-05	7.19E-05
CH4 Retention time:	12.0	12.0	NMHC Peak Area:	127432	127432
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	17.19	16.96	1.014
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.96	Prev response	17.22	*% change	-1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	17.19	16.97	1.013
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.013

Notes: Nitrogen Cylinder Changed.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	9.16	8.96	1.023
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.96	Prev response	9.19	*% change	-2.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	9.16	8.97	1.021
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.021

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	8.03	8.00	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.00	Prev response	8.03	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	8.03	8.00	1.005
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.005

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998719	0.986971
THC Cal Offset:	0.052792	0.000000
CH ₄ Cal Slope:	1.001737	0.995303
CH ₄ Cal Offset:	-0.013844	0.000000
NMHC Cal Slope:	0.996196	0.979445
NMHC Cal Offset:	0.065637	0.000000

Calibration Performed By: Melissa Lemay

NMHC Calibration Plot

Date: December 12, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Fort Hills	Station number:	AMS 23
Calibration Date:	December 17, 2024	Last Cal Date:	November 21, 2024
Start time (MST):	8:25	End time (MST):	12:16
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC281425	Cal Gas Expiry Date:	January 5, 2025
CH4 Cal Gas Conc.	500.2 ppm	CH4 Equiv Conc.	1070.6 ppm
C3H8 Cal Gas Conc.	207.4 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	500.2 ppm	CH4 Equiv Conc.	1070.6 ppm
Removed C3H8 Conc.	207.4 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
Zero Air Gen model:	API T701	Serial Number:	5611

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1152430012
THC Range: 0 - 20 ppm	NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.00E-05	2.02E-05	NMHC SP Ratio:	7.19E-05	7.43E-05
CH4 Retention time:	12.0	12.0	NMHC Peak Area:	127432	122718
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	17.19	16.80	1.023
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.80	Prev response	17.22	*% change	-2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	17.19	17.14	1.003
Mid point	4960	40.2	8.61	8.56	1.006
Low point	4980	20.1	4.30	4.26	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	17.19	17.15	1.003
Average Correction Factor					1.006

Notes: Linear issues on NM channel. Zero chromatogram turned to ON. Zero and span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	9.16	8.79	1.042
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.79	Prev response	9.19	*% change	-4.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	9.16	9.08	1.009
Mid point	4960	40.2	4.59	4.57	1.004
Low point	4980	20.1	2.29	2.29	1.002
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	9.16	9.10	1.006
Average Correction Factor					1.005

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	80.3	8.03	8.01	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.01	Prev response	8.03	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	80.3	8.03	8.06	0.997
Mid point	4960	40.2	4.02	3.99	1.008
Low point	4980	20.1	2.01	1.97	1.019
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	80.3	8.03	8.05	0.998
Average Correction Factor					1.008

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.998719	0.997331
THC Cal Offset:	0.052792	-0.015765
CH ₄ Cal Slope:	1.001737	1.004381
CH ₄ Cal Offset:	-0.013844	-0.026641
NMHC Cal Slope:	0.996196	0.991173
NMHC Cal Offset:	0.065637	0.010277

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 17, 2024	Previous Calibration:	November 21, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	8:25	End Time (MST):	12:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995
17.19	17.14	1.0031	Slope	0.997331	$0.90 - 1.10$
8.61	8.56	1.0057	Intercept	-0.015765	± 0.5
4.30	4.26	1.0097			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 17, 2024	Previous Calibration:	November 21, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	8:25	End Time (MST):	12:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999946	<i>≥0.995</i>
8.03	8.06	0.9969	Slope	1.004381	<i>0.90 - 1.10</i>
4.02	3.99	1.0081	Intercept	-0.026641	<i>+/-0.5</i>
2.01	1.97	1.0191			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 17, 2024	Previous Calibration:	November 21, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	8:25	End Time (MST):	12:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
9.16	9.08	1.0086	Slope	0.991173	<i>0.90 - 1.10</i>
4.59	4.57	1.0038	Intercept	0.010277	<i>+/-0.5</i>
2.29	2.29	1.0021			



NMHC Calibration Plot

Date: December 17, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Hills
 Station number: AMS 23
 Calibration Date: December 12, 2024
 Last Cal Date: November 4, 2024
 Start time (MST): 9:35
 End time (MST): 12:03
 Reason: As Found

Calibration Standards

NO Gas Cylinder #: CC358149
 NOX Cal Gas Conc: 60.30 ppm
 Removed Cylinder #:
 Removed Gas NOX Conc: 60.30 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701
 Cal Gas Expiry Date: January 5, 1932
 NO Cal Gas Conc: 60.10 ppm
 Removed Gas Exp Date:
 Removed Gas NO Conc: 60.10 ppm
 NO gas Diff:
 Serial Number: 451
 Serial Number: 5611

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.2	----	----
AF High point	4934	66.3	799.5	796.9	2.7	730.4	727.0	3.4	1.0942	1.0960
AF Mid point	4967	33.2	400.4	399.0	1.3	367.0	363.9	3.1	1.0901	1.0963
AF Low point	4983	16.6	200.2	199.5	0.7	183.4	180.1	3.3	1.0899	1.1074

New cyl resp

Previous Response	NO _x = 798.5 ppb	NO = 798.2 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -9.3%
Baseline Corr 1st pt	NO _x = 730.7 ppb	NO = 727.1 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -9.8%
Baseline Corr 2nd pt	NO _x = 367.3 ppb	NO = 364.0 ppb	As found	NO _x r ² : 0.999995	Nx SI: 0.913802	Nx Int: 0.267
Baseline Corr 3rd pt	NO _x = 183.7 ppb	NO = 180.2 ppb	As found	NO r ² : 0.999992	NO SI: 0.913226	NO Int: -0.871
			As found	NO ₂ r ² : 0.999997	NO ₂ SI: 0.997840	NO ₂ Int: 0.079

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	-0.2	----	----
As found high GPT point	730.5	370.5	362.7	361.8	1.0024	99.8%
As found mid GPT point	730.5	551.3	181.9	181.7	1.0008	99.9%
As found low GPT point	730.5	642.0	91.2	91.3	0.9984	100.2%



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430007

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.327	1.327	NO bkgnd or offset:	4.0	4.0
NOX coeff or slope:	0.990	0.990	NOX bkgnd or offset:	4.2	4.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	212.1	212.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999069	
NO _x Cal Offset:	-0.253684	
NO Cal Slope:	1.003782	
NO Cal Offset:	-1.713073	
NO ₂ Cal Slope:	0.994533	
NO ₂ Cal Offset:	-1.185414	

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero
 High point
 Mid point
 Low point
 As left zero
 As left span

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero
 High GPT point
 Mid GPT point
 Low GPT point

Average Correction Factor


Notes: As founds to change the pump and charcoal.

Calibration Performed By: Melissa Lemay

NO_x Calibration Plot

Date: December 12, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Hills
 Station number: AMS 23
 Calibration Date: December 13, 2024
 Last Cal Date: December 12, 2024
 Start time (MST): 7:32
 End time (MST): 11:11
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC358149
 NOX Cal Gas Conc: 60.30 ppm
 Removed Cylinder #:
 Removed Gas NOX Conc: 60.30 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701
 Cal Gas Expiry Date: January 5, 1932
 NO Cal Gas Conc: 60.10 ppm
 Removed Gas Exp Date:
 Removed Gas NO Conc: 60.10 ppm
 NO gas Diff:
 Serial Number: 451
 Serial Number: 5611

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = NA	ppb	NO = NA	ppb		* = > +/-5% change initiates investigation		*Percent Change	NO _x =	NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb		<u>As Found Statistics</u>		*Percent Change	NO =	NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb		As found NO _x r ² :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb		As found NO r ² :		NO SI:	NO Int:	
						As found NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430007

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999069	0.999069
NO _x Cal Offset:	-0.253684	-0.253684
NO Cal Slope:	1.003782	1.003782
NO Cal Offset:	-1.713073	-1.713073
NO ₂ Cal Slope:	0.994533	0.994533
NO ₂ Cal Offset:	-1.185414	-1.185414

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.327	0.955	NO bkgnd or offset:	4.0	2.9
NOX coeff or slope:	0.990	0.999	NOX bkgnd or offset:	4.2	3.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	145.3	145.3

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.2	----	----
High point	4934	66.3	799.5	796.9	2.7	798.0	797.8	0.2	1.0019	0.9988
Mid point	4967	33.2	400.4	399.0	1.3	399.4	397.2	2.2	1.0024	1.0047
Low point	4983	16.6	200.2	199.5	0.7	199.4	196.4	3.0	1.0041	1.0160
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	-0.1	----	----
As left span	4934	66.3	799.5	404.0	395.5	790.8	404.0	386.9	1.0110	1.0000
Average Correction Factor									1.0028	1.0065

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	789.8	405.4	387.1	386.8	1.0007	99.9%
Mid GPT point	789.8	597.9	194.6	193.6	1.0049	99.5%
Low GPT point	789.8	693.8	98.7	96.0	1.0276	97.3%
Average Correction Factor					1.0111	98.9%

Notes: Calibration after pump and charcoal replaced. Span adjusted. Due to drifting during the GPT the 2nd NO ref point was used.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 13, 2024	Previous Calibration:	December 12, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:32	End Time (MST):	11:11
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	1.000000	≥0.995
799.5	798.0	1.0019	Slope	0.998240	0.90 - 1.10
400.4	399.4	1.0024	Intercept	-0.213435	+/-20
200.2	199.4	1.0041			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 13, 2024	Previous Calibration:	December 12, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:32	End Time (MST):	11:11
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999943	≥0.995
387.1	386.8	1.0007	Slope	1.000947	0.90 - 1.10
194.6	193.6	1.0049	Intercept	-1.074996	+/-20
98.7	96.0	1.0276			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 13, 2024	Previous Calibration:	December 12, 2024
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:32	End Time (MST):	11:11
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999977	≥0.995
796.9	797.8	0.9988	Slope	1.002376	0.90 - 1.10
399.0	397.2	1.0047	Intercept	-1.872616	+/-20
199.5	196.4	1.0160			



NO_x Calibration Plot

Date: December 13, 2024

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Hills Station number: AMS 23
 Calibration Date: December 17, 2024 Last Cal Date: November 17, 2024
 Start time (MST): 7:57 End time (MST): 8:30

Analyzer Make: API T640 S/N: 320
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388744
 Temp/RH standard: Alicat FP-25BT S/N: 388744

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-26.4	-27.4	-26.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	749.9	749.4	749.9	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.93	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39	----	39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.2	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: **10.9** Expiry Date: 10-Jun-24
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ November 21, 2024
 Date Disposable Filter Changed: _____ November 21, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ October 17, 2023
 Date RH/T Sensor Cleaned: _____ October 17, 2023

Notes: No adjustments done.

Calibration by: Melissa Lemay




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25 WASKŌW OHCI PIMÂTISIWIN DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number: AMS 25
Calibration Date:	December 20, 2024	Last Cal Date: November 28, 2024
Start time (MST):	7:16	End time (MST): 10:07
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	49.70	ppm	Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #:	CC342445		
Removed Cal Gas Conc:	49.70	ppm	Rem Gas Exp Date:
Removed Gas Cyl #:			Diff between cyl:
Calibrator Model:	API T700		Serial Number: 747
Zero Air Gen Model:	API T701		Serial Number: 4765

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 1118148497
Analyzer Range:	0-1000ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999070	0.998456	Backgd or Offset:	11.6	11.8
Calibration intercept:	-0.112739	-0.372807	Coeff or Slope:	1.062	1.071

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4920	80.5	800.1	791.3	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	791.5	Previous response	799.2	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4920	80.5	800.1	798.6	1.002
Mid point	4960	40.2	399.6	398.5	1.003
Low point	4980	20.1	199.8	198.8	1.005
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	80.5	800.1	799.6	1.001
Average Correction Factor:					1.003

Notes: No Maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 20, 2024	Previous Calibration:	November 28, 2024
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS 25
Start Time (MST):	7:16	End Time (MST):	10:07
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999999	≥0.995
800.1	798.6	1.0019	Slope	0.998456	0.90 - 1.10
399.6	398.5	1.0027	Intercept	-0.372807	+/-30
199.8	198.8	1.0050			



SO2 Calibration Plot

Date: December 20, 2024

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	Waskow ohci Pimatisiwin	Station number:	AMS 25
Calibration Date:	December 16, 2024	Last Cal Date:	November 27, 2024
Start time (MST):	8:27	End time (MST):	12:30
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.97	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC517099			
Removed Cal Gas Conc:	4.97	ppm	Rem Gas Exp Date:	
Removed Gas Cyl #:			Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	747
ZAG Make/Model:	API T701		Serial Number:	261

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146
Converter make:	Global G-150	Converter serial #:	2022-219
Analyzer Range	0 - 100 ppb	Converter Temp:	350 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.988502	1.003306	Backgd or Offset:	3.42	3.50
Calibration intercept:	0.360000	0.320000	Coeff or Slope:	1.074	1.096

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	80.0	79.5	78.2	1.017
As found Mid point	4960	40.0	39.8	39.1	1.017
As found Low point	4980	20.0	19.9	19.7	1.009
New cylinder response					
Baseline Corr As found:	78.2	Prev response:	78.97	*% change:	-1.0%
Baseline Corr 2nd AF pt:	39.1	AF Slope:	0.982754	AF Intercept:	0.060000
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999995	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	80.0	79.5	80.0	0.994
Mid point	4960	40.0	39.8	40.4	0.984
Low point	4980	20.0	19.9	20.3	0.979
As left zero	5000	0.0	0.0	0.3	----
As left span	4912	88.3	800.0	810.4	0.987
SO ₂ Scrubber Check	4921	79.2	800.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.986
Date of last converter efficiency test:					

Notes: SO_x scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	December 16, 2024	Previous Calibration:	November 27, 2024
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS 25
Start Time (MST):	8:27	End Time (MST):	12:30
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999982	≥0.995
79.5	80.0	0.9940	Slope	1.003306	0.90 - 1.10
39.8	40.4	0.9842	Intercept	0.320000	+/-3
19.9	20.3	0.9793			



H₂S Calibration Plot

Date: December 16, 2024

Location: Waskow ohci Pimatisiwin






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27
JACKFISH 2/3
DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	December 4, 2024	Last Cal Date:	November 12, 2024
Start time (MST):	11:29	End time (MST):	14:32
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.58	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	SG9133974BAL			
Removed Cal Gas Conc:	50.58	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	API T700		Serial Number:	3811
Zero Air Gen Model:	API 701		Serial Number:	268

Analyzer Information

Analyzer make:	Thermo 43iQ-TL	Serial Number:	12124313138
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993016	1.001040	Backgd or Offset:	8.2	8.3
Calibration intercept:	-1.166634	-1.325883	Coeff or Slope:	0.927	0.947

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	1.0	----
As found High point	4919	79.1	800.5	780.8	1.027
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	779.8	Previous response	793.7	*% change	-1.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	1.3	----
High point	4919	79.1	800.5	801.0	0.999
Mid point	4960	39.5	399.6	398.3	1.003
Low point	4979	19.8	200.3	196.0	1.022
As left zero	5000	0.0	0.0	1.1	----
As left span	4921	79.1	800.2	802.9	0.997
Average Correction Factor:					1.008

Notes: Changed the sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 4, 2024	Previous Calibration:	November 12, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:29	End Time (MST):	14:32
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	1.3	----	Correlation Coefficient	0.999947	≥0.995
800.5	801.0	0.9994	Slope	1.001040	0.90 - 1.10
399.6	398.3	1.0033	Intercept	-1.325883	+/-30
200.3	196.0	1.0222			



SO2 Calibration Plot

Date: December 4, 2024

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	Jackfish 2/3	Station number:	AMS 27
Calibration Date:	December 5, 2024	Last Cal Date:	November 19, 2024
Start time (MST):	10:56	End time (MST):	15:19
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.41	ppm	Cal Gas Exp Date:	January 4, 2025
Cal Gas Cylinder #:	CC345023			
Removed Cal Gas Conc:	5.41	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	3811
ZAG Make/Model:	API T701H		Serial Number:	268

Analyzer Information

Analyzer make:	Thermo 43iQ	Analyzer serial #:	12228021055
Converter make:	Global G150	Converter serial #:	2022-195
Analyzer Range	0 - 100 ppb	Converter Temp:	325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.022160	1.014598	Backgd or Offset:	3.6	3.2
Calibration intercept:	-0.657722	-0.317640	Coeff or Slope:	1.110	1.162

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.8	----
As found High point	4926	74.1	80.2	78.7	1.008
As found Mid point	4963	37.0	40.0	38.5	1.019
As found Low point	4982	18.5	20.0	18.5	1.037
New cylinder response					
Baseline Corr As found:	79.5	Prev response:	81.29	*% change:	-2.3%
Baseline Corr 2nd AF pt:	39.3	AF Slope:	0.993362	AF Intercept:	-1.098199
Baseline Corr 3rd AF pt:	19.3	AF Correlation:	0.999936	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4926	74.1	80.2	81.1	0.989
Mid point	4963	37.0	40.0	40.3	0.993
Low point	4982	18.5	20.0	19.7	1.016
As left zero	5000	0.0	0.0	0.1	----
As left span	4926	74.1	80.2	81.6	0.983
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.999
Date of last converter efficiency test:					

Notes: Changed sample inlet filter. Ran scrubber check after calibrator zero and it passed. Adjusted zero and span.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 19, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	10:56	End Time (MST):	15:19
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12228021055

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999962	≥0.995
80.2	81.1	0.9886	Slope	1.014598	0.90 - 1.10
40.0	40.3	0.9934	Intercept	-0.317640	+/-3
20.0	19.7	1.0160			



H₂S Calibration Plot

Date: December 5, 2024

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Jackfish 2/3
 Station number: AMS 27
 Calibration Date: December 6, 2024
 Last Cal Date: November 15, 2024
 Start time (MST): 8:48
 End time (MST): 14:37
 Reason: Install

Calibration Standards

NO Gas Cylinder #: CC757838
 NOX Cal Gas Conc: 60.30 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 60.30 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701

Cal Gas Expiry Date: January 9, 2032
 NO Cal Gas Conc: 60.20 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 60.20 ppm
 NO gas Diff:
 Serial Number: 3811
 Serial Number: 268

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
AF High point	4942	66.5	800.6	799.3	1.3	816.0	815.0	0.8	0.9812	0.9807
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 805.5 ppb		NO = 806.6 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 1.3%	
Baseline Corr 1st pt	NO _x = 816.0 ppb		NO = 815.0 ppb			<u>As Found Statistics</u>		*Percent Change	NO = 1.0%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153357

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.008779	0.995064
NO _x Cal Offset:	-2.133379	-1.654761
NO Cal Slope:	1.014488	1.001235
NO Cal Offset:	-4.293618	-3.734605
NO ₂ Cal Slope:	1.012471	0.996639
NO ₂ Cal Offset:	-0.121108	-0.410548

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.360	1.332	NO bkgnd or offset:	4.6	4.6
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	4.7	4.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	156.9	156.3

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.8	0.2	0.6	----	----
High point	4942	66.5	800.6	799.3	1.3	796.5	798.8	-2.3	1.0052	1.0006
Mid point	4979	33.3	400.6	399.9	0.7	395.0	393.9	1.0	1.0142	1.0154
Low point	4996	16.6	199.7	199.4	0.3	195.1	192.5	2.6	1.0235	1.0356
As left zero	5000	0.0	0.0	0.0	0.0	1.5	0.2	1.3	----	----
As left span	4942	66.5	800.6	404.9	395.7	804.2	404.9	399.2	0.9956	1.0000
Average Correction Factor									1.0143	1.0172

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.6	----	----
High GPT point	800.8	403.6	398.5	397.6	1.0023	99.8%
Mid GPT point	800.8	622.3	179.8	177.0	1.0160	98.4%
Low GPT point	800.8	713.8	88.3	87.6	1.0083	99.2%
Average Correction Factor					1.0089	99.1%

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 15, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	8:48	End Time (MST):	14:37
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.8	----	Correlation Coefficient	0.999954	<i>≥0.995</i>
800.6	796.5	1.0052	Slope	0.995064	<i>0.90 - 1.10</i>
400.6	395.0	1.0142	Intercept	-1.654761	<i>+/-20</i>
199.7	195.1	1.0235			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 15, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	8:48	End Time (MST):	14:37
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.6	----	Correlation Coefficient	0.999943	<i>≥0.995</i>
398.5	397.6	1.0023	Slope	0.996639	<i>0.90 - 1.10</i>
179.8	177.0	1.0160	Intercept	-0.410548	<i>+/-20</i>
88.3	87.6	1.0083			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 15, 2024
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	8:48	End Time (MST):	14:37
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999886	<i>≥0.995</i>
799.3	798.8	1.0006	Slope	1.001235	<i>0.90 - 1.10</i>
399.9	393.9	1.0154	Intercept	-3.734605	<i>+/-20</i>
199.4	192.5	1.0356			



NO_x Calibration Plot

Date: December 6, 2024

Location: Jackfish 2/3






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS29 SURMONT 2 DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Surmont 2	Station number:	AMS 29
Calibration Date:	December 2, 2024	Last Cal Date:	November 7, 2024
Start time (MST):	11:20	End time (MST):	14:40
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.21	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	CC356008			
Removed Cal Gas Conc:	49.21	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	5472
Zero Air Gen Model:	Teledyne API T701		Serial Number:	4698

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1170050150
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000130	1.006441	Backgd or Offset:	13.0	14.1
Calibration intercept:	-1.425614	-2.285476	Coeff or Slope:	0.925	0.970

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4919	81.3	800.1	763.7	1.048
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	763.4	Previous response	798.8	*% change	-4.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.4	----
High point	4919	81.3	800.1	804.0	0.995
Mid point	4959	40.7	400.6	399.7	1.002
Low point	4979	20.3	199.8	197.1	1.014
As left zero	5000	0.0	0.0	-0.2	----
As left span	4919	81.3	800.1	808.0	0.990
Average Correction Factor:					1.004

Notes: Adjusted zero and span. Large span shift likely caused by a dropping lamp voltage, maintenance planned for January.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 7, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	11:20	End Time (MST):	14:40
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.4	----	Correlation Coefficient	0.999975	≥0.995
800.1	804.0	0.9952	Slope	1.006441	0.90 - 1.10
400.6	399.7	1.0022	Intercept	-2.285476	+/-30
199.8	197.1	1.0138			



SO2 Calibration Plot

Date: December 2, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Surmont 2	Station number: AMS 29
Calibration Date: December 4, 2024	Last Cal Date: November 5, 2024
Start time (MST): 12:28	End time (MST): 14:33
Reason: As Found	

Calibration Standards

Cal Gas Concentration: <u>5.391</u> ppm	Cal Gas Exp Date: January 4, 2025
Cal Gas Cylinder #: <u>CC508338</u>	
Removed Cal Gas Conc: <u>5.391</u> ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: <u>CC508338</u>	Diff between cyl:
Calibrator Make/Model: Teledyne API T700	Serial Number: 5472
ZAG Make/Model: Teledyne API T701	Serial Number: 4698

Analyzer Information

Analyzer make: Thermo 43iQ-TLE	Analyzer serial #: 1200326170
Converter make: Global	Converter serial #: 2022-220
Analyzer Range: 0 - 100 ppb	Converter Temp: 350.0 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000611		Backgd or Offset:	0.88
Calibration intercept:	-0.062596		Coeff or Slope:	1.030

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4926	74.2	80.0	81.2	0.984
As found Mid point	4963	37.2	40.1	40.5	0.988
As found Low point	4982	18.6	20.1	20.1	0.993
New cylinder response					
Baseline Corr As found:	81.3	Prev response:	79.99	*% change:	1.6%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.016757	AF Intercept:	-0.203396
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999992	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					
SO2 Scrubber Check					
Date of last scrubber change:				Ave Corr Factor	
Date of last converter efficiency test:					


Notes: As founds completed before changing the converter temperature from 350C to 325C. Calibration and converter efficiency test planned for Dec 5.

Calibration Performed By: Braiden Boutilier

H2S Calibration Plot

Date: December 4, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Surmont 2	Station number: AMS 29
Calibration Date: December 5, 2024	Last Cal Date: December 4, 2024
Start time (MST): 10:39	End time (MST): 15:21
Reason: Routine	

Calibration Standards

Cal Gas Concentration: <u>5.391</u> ppm	Cal Gas Exp Date: January 4, 2025
Cal Gas Cylinder #: <u>CC508338</u>	
Removed Cal Gas Conc: <u>5.391</u> ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: <u>CC508338</u>	Diff between cyl:
Calibrator Make/Model: Teledyne API T700	Serial Number: 5472
ZAG Make/Model: Teledyne API T701	Serial Number: 4698

Analyzer Information

Analyzer make: Thermo 43iQ-TLE	Analyzer serial #: 1200326170
Converter make: Global	Converter serial #: 2022-220
Analyzer Range: 0 - 100 ppb	Converter Temp: 325.0 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000611	1.001184	Backgd or Offset:	0.88	0.88
Calibration intercept:	-0.062596	-0.082675	Coeff or Slope:	1.030	1.030

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4926	74.2	80.0	80.1	0.999
Mid point	4963	37.2	40.1	39.9	1.005
Low point	4982	18.6	20.1	20.0	1.003
As left zero	5000	0.0	0.0	0.1	----
As left span	4926	74.2	80.0	79.3	1.009
SO2 Scrubber Check				0.0	
Date of last scrubber change:				Ave Corr Factor	1.002
Date of last converter efficiency test:		December 5, 2024		108.1% efficiency	

Notes: Adjusted span during and after converter efficiency test. SOx scrubber check done after converter efficiency test, before the span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	December 4, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:39	End Time (MST):	15:21
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999987	≥ 0.995
80.0	80.1	0.9988	Slope	1.001184	$0.90 - 1.10$
40.1	39.9	1.0052	Intercept	-0.082675	± 3
20.1	20.0	1.0027			



H2S Calibration Plot

Date: December 5, 2024

Location: Surmont 2





Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 7, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	11:20	End Time (MST):	14:40
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.00	-0.02	----	Correlation Coefficient	0.999955	≥0.995
17.31	17.39	0.9956	Slope	1.007016	0.90 - 1.10
8.65	8.59	1.0065	Intercept	-0.076638	+/-1.5
4.32	4.23	1.0215			



THC Calibration Plot

Date: December 2, 2024


Location: Surmont 2



THC Calibration Plot

Date: December 4, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Surmont 2
 Station number: AMS 29
 Calibration Date: December 18, 2024
 Last Cal Date: November 19, 2024
 Start time (MST): 11:11
 End time (MST): 16:47
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T12YYFE
 NOX Cal Gas Conc: 47.46 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 47.46 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: October 30, 2024
 NO Cal Gas Conc: 47.46 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 47.46 ppm
 NO gas Diff:
 Serial Number: 5472
 Serial Number: 4698

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
AF High point	4916	84.3	800.1	800.1	0.0	794.1	792.3	1.8	1.0073	1.0096
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 800.6 ppb		NO = 801.0 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -0.8%	
Baseline Corr 1st pt	NO _x = 794.3 ppb		NO = 792.5 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -1.1%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found NO _x r ² :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found NO r ² :		NO SI:	NO Int:	
						As found NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1170050148

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001911	1.001023
NO _x Cal Offset:	-1.012618	-0.652175
NO Cal Slope:	1.004112	1.000922
NO Cal Offset:	-2.432496	-1.891558
NO ₂ Cal Slope:	1.002053	1.006053
NO ₂ Cal Offset:	-0.296979	1.038885

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.639	1.652	NO bkgnd or offset:	1.4	1.4
NOX coeff or slope:	0.994	0.998	NOX bkgnd or offset:	1.5	1.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	194.1	198.0

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
High point	4916	84.2	799.2	799.2	0.0	799.1	798.3	0.8	1.0001	1.0011
Mid point	4958	42.1	399.6	399.6	0.0	401.0	399.3	1.8	0.9965	1.0008
Low point	4979	21.1	200.3	200.3	0.0	197.6	195.1	2.5	1.0135	1.0265
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
As left span	4916	84.2	799.2	599.1	200.1	793.2	599.1	194.0	1.0076	1.0000
Average Correction Factor									1.0034	1.0095

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	790.0	411.5	378.5	381.2	0.9929	100.7%
Mid GPT point	790.0	599.1	190.9	194.0	0.9840	101.6%
Low GPT point	790.0	691.2	98.8	101.1	0.9773	102.3%
Average Correction Factor					0.9847	101.6%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 18, 2024	Previous Calibration:	November 19, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	11:11	End Time (MST):	16:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999976	<i>≥0.995</i>
799.2	799.1	1.0001	Slope	1.001023	<i>0.90 - 1.10</i>
399.6	401.0	0.9965	Intercept	-0.652175	<i>+/-20</i>
200.3	197.6	1.0135			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 18, 2024	Previous Calibration:	November 19, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	11:11	End Time (MST):	16:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999968	<i>≥0.995</i>
378.5	381.2	0.9929	Slope	1.006053	<i>0.90 - 1.10</i>
190.9	194.0	0.9840	Intercept	1.038885	<i>+/-20</i>
98.8	101.1	0.9773			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 18, 2024	Previous Calibration:	November 19, 2024
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	11:11	End Time (MST):	16:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999950	<i>≥0.995</i>
799.2	798.3	1.0011	Slope	1.000922	<i>0.90 - 1.10</i>
399.6	399.3	1.0008	Intercept	-1.891558	<i>+/-20</i>
200.3	195.1	1.0265			



NO_x Calibration Plot

Date: December 18, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Surmont 2
 Station number: AMS 29
 Calibration Date: December 19, 2024
 Last Cal Date: December 18, 2024
 Start time (MST): 11:40
 End time (MST): 13:35
 Reason: Cylinder Change

Calibration Standards

NO Gas Cylinder #: CC218007
 NOX Cal Gas Conc: 60.20 ppm
 Removed Cylinder #: T12YYFE
 Removed Gas NOX Conc: 47.46 ppm
 NOX gas Diff: -1.2%
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701

Cal Gas Expiry Date: January 9, 2032
 NO Cal Gas Conc: 60.00 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 47.46 ppm
 NO gas Diff: -2.7%
 Serial Number: 5472
 Serial Number: 4698

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
AF High point	4916	84.3	800.1	800.1	0.0	794.1	792.3	1.8	1.0073	1.0096
AF Mid point										
AF Low point										
New cyl resp	4933	66.7	803.1	800.4	2.7	787.4	771.9	15.5	1.0111	1.0100
Previous Response	NO _x = 800.6 ppb		NO = 801.0 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -0.8%	
Baseline Corr 1st pt	NO _x = 794.3 ppb		NO = 792.5 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -1.1%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1170050148

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001911	
NO _x Cal Offset:	-1.012618	
NO Cal Slope:	1.004112	
NO Cal Offset:	-2.432496	
NO ₂ Cal Slope:	1.002053	
NO ₂ Cal Offset:	-0.296979	

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.652	1.652	NO bkgnd or offset:	1.4	1.4
NOX coeff or slope:	0.998	0.998	NOX bkgnd or offset:	1.5	1.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	198.0	198.0

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero										
High point										
Mid point										
Low point										
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
As left span	4933	66.7	803.1	384.9		785.0	384.9	400.1	1.0231	1.0000
Average Correction Factor										

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero						
High GPT point						
Mid GPT point						
Low GPT point						
Average Correction Factor						

Notes:

Swapped NO cylinder.


Calibration Performed By:

Braiden Boutilier

NO_x Calibration Plot

Date: December 19, 2024

Location: Surmont 2





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Surmont 2 Station number: AMS 29
 Calibration Date: December 13, 2024 Last Cal Date: November 19, 2024
 Start time (MST): 10:45 End time (MST): 14:00

Analyzer Make: API T640 S/N: 323
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-7.6	-6.41	-7.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	704.3	702.84	704.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.98	4.958	4.98	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39	----	39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	8.6	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
 Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: October 31, 2024
 Date Disposable Filter Changed: October 31, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 30, 2024
 Date RH/T Sensor Cleaned: October 30, 2024

Notes: Verified temperature, pressure and flow. Leak check passed.

Calibration by: Braiden Boutilier




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS30 ELLS RIVER DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Ells River	Station number: AMS 30
Calibration Date:	December 6, 2024	Last Cal Date: November 1, 2024
Start time (MST):	11:12	End time (MST): 14:07
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	48.75	ppm	Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #:	CC350110		
Removed Cal Gas Conc:	48.75	ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA		Diff between cyl:
Calibrator Model:	API T700		Serial Number: 3061
Zero Air Gen Model:	API T701H		Serial Number: 358

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 1008841397
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003404	1.001932	Backgd or Offset:	9.7	10.1
Calibration intercept:	-2.392023	-3.552064	Coeff or Slope:	0.992	0.992

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4918	82.0	799.5	799.3	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	799.3	Previous response	799.8	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4918	82.0	799.5	799.1	1.001
Mid point	4959	41.0	399.8	395.6	1.010
Low point	4980	20.5	199.9	193.0	1.036
As left zero	5000	0.0	0.0	-0.2	----
As left span	4918	82.0	799.5	799.8	1.000
Average Correction Factor:					1.016

Notes: Sample inlet filter was changed after as founds. Adjusted zero only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 1, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	11:12	End Time (MST):	14:07
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999915	≥0.995
799.5	799.1	1.0005	Slope	1.001932	0.90 - 1.10
399.8	395.6	1.0105	Intercept	-3.552064	+/-30
199.9	193.0	1.0355			



SO2 Calibration Plot

Date: December 6, 2024

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	December 3, 2024	Last Cal Date:	November 4, 2024
Start time (MST):	9:23	End time (MST):	14:44
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.99	ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	CC505806			
Removed Cal Gas Conc:	4.99	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T750		Serial Number:	281
ZAG Make/Model:	API 751H		Serial Number:	77

Analyzer Information

Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331
Converter make:	CDN- 101	Converter serial #:	562
Analyzer Range	0 - 100 ppb	Converter Temp:	800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008757	0.991618	Backgd or Offset:	1.7	1.6
Calibration intercept:	-0.060386	0.439545	Coeff or Slope:	1.094	1.042

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4920	80.2	80.0	82.8	0.965
As found Mid point	4960	40.1	40.0	41.4	0.964
As found Low point	4980	20.0	20.0	20.8	0.955
New cylinder response					
Baseline Corr As found:	82.9	Prev response:	80.68	*% change:	2.7%
Baseline Corr 2nd AF pt:	41.5	AF Slope:	1.034880	AF Intercept:	0.000230
Baseline Corr 3rd AF pt:	20.9	AF Correlation:	0.999992	<i>* = > +/-5% change initiates investigation</i>	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.2	80.0	79.6	1.005
Mid point	4960	40.1	40.0	40.3	0.993
Low point	4980	20.0	20.0	20.7	0.964
As left zero	5000	0.0	0.0	0.5	----
As left span	4920	80.2	80.0	78.5	1.020
SO2 Scrubber Check	4918	82.0	820.0	0.2	----
Date of last scrubber change:				Ave Corr Factor	0.988
Date of last converter efficiency test:					

Notes: Change sample inlet filter after multipoint as founds. Sox scrubber check done after calibrator zero and passed. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 4, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:23	End Time (MST):	14:44
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999859	≥ 0.995
80.0	79.6	1.0055	Slope	0.991618	$0.90 - 1.10$
40.0	40.3	0.9930	Intercept	0.439545	± 3
20.0	20.7	0.9643			



TRS Calibration Plot

Date: December 3, 2024

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	December 6, 2024	Last Cal Date:	November 1, 2024
Start time (MST):	11:12	End time (MST):	14:07
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC350110	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	496.6 ppm	CH ₄ Equiv Conc.	1066.4 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	496.6 ppm	CH ₄ Equiv Conc.	1066.4 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
Zero Air Gen model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1152430011
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.12E-04	3.12E-04	5.94E-05	5.94E-05
CH ₄ Retention time:	17.6	17.6	157107	157107
Zero Chromatogram:	ON	ON	Flat Baseline: OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	17.49	17.48	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.48	Prev response	17.49	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	82.0	17.49	17.43	1.003
Mid point	4959	41.0	8.74	8.63	1.013
Low point	4980	20.5	4.37	4.27	1.024
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	17.49	17.44	1.003
Average Correction Factor					1.014

Notes: Sample inlet filter was changed after as founds. No adjustment made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	9.34	9.35	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.35	Prev response	9.36	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	82.0	9.34	9.31	1.004
Mid point	4959	41.0	4.67	4.62	1.011
Low point	4980	20.5	2.34	2.30	1.015
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	9.34	9.31	1.004
Average Correction Factor					1.010

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	8.14	8.14	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.14	Prev response	8.13	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	82.0	8.14	8.12	1.003
Mid point	4959	41.0	4.07	4.01	1.016
Low point	4980	20.5	2.04	1.97	1.036
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	8.14	8.13	1.001
Average Correction Factor					1.018

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002015	0.997832
THC Cal Offset:	-0.036226	-0.052226
CH ₄ Cal Slope:	1.002263	0.998966
CH ₄ Cal Offset:	-0.033119	-0.036119
NMHC Cal Slope:	1.002360	0.997065
NMHC Cal Offset:	-0.004906	-0.017507

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 1, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	11:12	End Time (MST):	14:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999957	<i>≥0.995</i>
17.49	17.43	1.0034	Slope	0.997832	<i>0.90 - 1.10</i>
8.74	8.63	1.0130	Intercept	-0.052226	<i>+/-0.5</i>
4.37	4.27	1.0243			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 1, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	11:12	End Time (MST):	14:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999908	<i>≥0.995</i>
8.14	8.12	1.0030	Slope	0.998966	<i>0.90 - 1.10</i>
4.07	4.01	1.0163	Intercept	-0.036119	<i>+/-0.5</i>
2.04	1.97	1.0355			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 1, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	11:12	End Time (MST):	14:07
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999980	≥ 0.995
9.34	9.31	1.0035	Slope	0.997065	$0.90 - 1.10$
4.67	4.62	1.0109	Intercept	-0.017507	± 0.5
2.34	2.30	1.0152			



NMHC Calibration Plot

Date: December 6, 2024

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Ells River	Station number:	AMS 30
Calibration Date:	December 23, 2024	Last Cal Date:	December 6, 2024
Start time (MST):	11:35	End time (MST):	13:35
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC350110	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	496.6 ppm	CH ₄ Equiv Conc.	1066.4 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	496.6 ppm	CH ₄ Equiv Conc.	1066.4 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
Zero Air Gen model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make: Thermo 55i	Analyzer serial #: 1152430011
THC Range: 0 - 20 ppm	NMHC/CH ₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.12E-04	3.12E-04	5.94E-05	5.94E-05
CH ₄ Retention time:	17.6	17.6	157107	157107
Zero Chromatogram:	ON	ON	Flat Baseline: OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	17.49	17.33	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.33	Prev response	17.40	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	17.49	17.28	1.012

Average Correction Factor

Notes: N2 Cylinder Replaced.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	9.34	9.25	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.25	Prev response	9.30	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	9.34	9.19	1.017
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	8.14	8.08	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.08	Prev response	8.10	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	8.14	8.09	1.007
Average Correction Factor					

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.997832	
THC Cal Offset:	-0.052226	
CH ₄ Cal Slope:	0.998966	
CH ₄ Cal Offset:	-0.036119	
NMHC Cal Slope:	0.997065	
NMHC Cal Offset:	-0.017507	

Calibration Performed By: Devin Russell

NMHC Calibration Plot

Date: December 23, 2024

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Ells River
 Station number: AMS 30
 Calibration Date: December 3, 2024
 Last Cal Date: November 12, 2024
 Start time (MST): 9:23
 End time (MST): 13:59
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0027487
 NOX Cal Gas Conc: 59.30 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 59.30 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701H
 Cal Gas Expiry Date: January 9, 2032
 NO Cal Gas Conc: 59.10 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 59.10 ppm
 NO gas Diff:
 Serial Number: 3061
 Serial Number: 358

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.5	-0.1	----	----
AF High point	4932	67.7	803.0	800.3	2.7	793.7	791.5	2.2	1.0109	1.0104
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 802.0 ppb		NO = 800.6 ppb			<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = -1.0%	
Baseline Corr 1st pt	NO _x = 794.3 ppb		NO = 792.0 ppb			<u>As Found Statistics</u>		*Percent Change	NO = -1.1%	
Baseline Corr 2nd pt	NO _x = NA ppb		NO = NA ppb			As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb		NO = NA ppb			As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 710321429

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000086	1.001750
NO _x Cal Offset:	-1.078126	-1.237513
NO Cal Slope:	1.002943	0.998473
NO Cal Offset:	-2.018813	-2.379819
NO ₂ Cal Slope:	1.000104	0.999496
NO ₂ Cal Offset:	-0.123124	0.002264

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.130	1.143	NO bkgnd or offset:	13.4	13.4
NOX coeff or slope:	0.992	0.999	NOX bkgnd or offset:	13.7	13.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	190.6	198.1

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
High point	4932	67.7	803.0	800.3	2.7	803.1	797.2	5.8	0.9998	1.0038
Mid point	4966	33.8	400.9	399.5	1.4	401.3	397.0	4.3	0.9990	1.0064
Low point	4983	16.9	200.4	199.8	0.7	197.8	194.0	3.8	1.0133	1.0297
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.2	----	----
As left span	4932	67.7	803.0	429.0	374.0	803.0	429.0	374.0	1.0000	1.0000
Average Correction Factor									1.0040	1.0133

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	797.2	427.5	372.4	371.6	1.0022	99.8%
Mid GPT point	797.2	610.6	189.3	191.0	0.9911	100.9%
Low GPT point	797.2	699.9	100.0	98.9	1.0112	98.9%
Average Correction Factor					1.0015	99.9%

Notes: Sample inlet filter changed after as founds. Adjusted span only. Used 2nd NO reference point because of drift.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 12, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:23	End Time (MST):	13:59
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.4	----	Correlation Coefficient	0.999987	<i>≥0.995</i>
803.0	803.1	0.9998	Slope	1.001750	<i>0.90 - 1.10</i>
400.9	401.3	0.9990	Intercept	-1.237513	<i>+/-20</i>
200.4	197.8	1.0133			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 12, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:23	End Time (MST):	13:59
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999937	<i>≥0.995</i>
372.4	371.6	1.0022	Slope	0.999496	<i>0.90 - 1.10</i>
189.3	191.0	0.9911	Intercept	0.002264	<i>+/-20</i>
100.0	98.9	1.0112			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 12, 2024
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:23	End Time (MST):	13:59
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999959	≥0.995
800.3	797.2	1.0038	Slope	0.998473	0.90 - 1.10
399.5	397.0	1.0064	Intercept	-2.379819	+/-20
199.8	194.0	1.0297			



NO_x Calibration Plot

Date: December 3, 2024

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Ells River Station number: AMS 30
 Calibration Date: December 6, 2024 Last Cal Date: November 12, 2024
 Start time (MST): 13:11 End time (MST): 13:34

Analyzer Make: API T640 S/N: 875
 Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
 Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	4.00	3.27	4.00	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	721.60	723.30	721.60	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.03	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	33	----	33	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.10	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean Alignment Factor On :

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: September 29, 2024
 Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: October 11, 2024
 Date Disposable Filter Changed: October 11, 2024

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: December 6, 2024
 Date RH/T Sensor Cleaned: February 23, 2024

Notes: Verified flow, temperature, pump power and pressure No adjustment made. Leak check passed.

Calibration by: Jan Castro




WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS33 MONDAY CREEK DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Monday Creek	Station number: AMS 33
Calibration Date:	December 5, 2024	Last Cal Date: November 7, 2024
Start time (MST):	9:07	End time (MST): 12:01
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	50.62 ppm	Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #:	EB0008522	
Removed Cal Gas Conc:	50.62 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA	Diff between cyl:
Calibrator Model:	Teledyne T750	Serial Number: 281
Zero Air Gen Model:	Teledyne 751H	Serial Number: 321

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number: 1152430005
Analyzer Range:	0- 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002587	1.004012	Backgd or Offset:	29.0	28.8
Calibration intercept:	1.445233	-1.253595	Coeff or Slope:	0.977	0.960

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.4	----
As found High point	4921	79.0	799.8	811.0	0.987
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	810.6	Previous response	803.3	*% change	0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4921	79.0	799.8	802.4	0.997
Mid point	4960	39.5	399.9	399.9	1.000
Low point	4980	19.8	200.5	198.2	1.011
As left zero	5000	0.0	0.0	0.4	----
As left span	4921	79.0	799.8	801.2	0.998
Average Correction Factor:					1.003

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 5, 2024	Previous Calibration:	November 7, 2024
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	9:07	End Time (MST):	12:01
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430005

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999982
799.8	802.4	0.9968	Slope	1.004012
399.9	399.9	1.0001	Intercept	-1.253595
200.5	198.2	1.0114		
				≥0.995
				0.90 - 1.10
				+/-30



SO2 Calibration Plot

Date: December 5, 2024

Location: Monday Creek





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Monday Creek	Station number: AMS 33
Calibration Date: December 4, 2024	Last Cal Date: November 7, 2024
Start time (MST): 11:03	End time (MST): 16:30
Reason: Routine	

Calibration Standards

Cal Gas Concentration: 5.05 ppm	Cal Gas Exp Date: November 15, 2026
Cal Gas Cylinder #: DT0014831	
Removed Cal Gas Conc: 5.05 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #: NA	Diff between cyl:
Calibrator Make/Model: Teledyne API T750	Serial Number: 281
ZAG Make/Model: Teledyne T701H	Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL	Analyzer serial #: 12333331547
Converter make: Global 150	Converter serial #: 2022-196
Analyzer Range: 0 - 100 ppb	Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003998	0.996854	Backgd or Offset:	1.3	1.5
Calibration intercept:	0.138391	0.038399	Coeff or Slope:	1.063	1.063

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4921	79.2	80.0	80.0	1.000
As found Mid point	4960	39.6	40.0	40.1	0.997
As found Low point	4980	19.8	20.0	20.2	0.990
New cylinder response					
Baseline Corr As found:	80.0	Prev response:	80.45	*% change:	-0.6%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	0.999426	AF Intercept:	0.098398
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999993	<i>* = > +/-5% change initiates investigation</i>	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4921	79.2	80.0	79.7	1.004
Mid point	4960	39.6	40.0	40.0	1.000
Low point	4980	19.8	20.0	20.1	0.995
As left zero	5000	0.0	0.0	-0.2	----
As left span	4921	79.2	80.0	80.2	0.997
SO2 Scrubber Check	4921	79.0	790.0	0.1	----
Date of last scrubber change:	11-Apr-24			Ave Corr Factor	1.000
Date of last converter efficiency test:					

Notes: Sample inlet filter changed after multipoint as founds. Sox scrubber check done after calibrator zero and passed. Adjusted zero only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	December 4, 2024	Previous Calibration:	November 7, 2024
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	11:03	End Time (MST):	16:30
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331547

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999986	≥ 0.995
80.0	79.7	1.0036	Slope	0.996854	$0.90 - 1.10$
40.0	40.0	1.0000	Intercept	0.038399	± 3
20.0	20.1	0.9950			



H2S Calibration Plot

Date: December 4, 2024

Location: Monday Creek





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Monday Creek
 Station number: AMS 33
 Calibration Date: December 4, 2024
 Last Cal Date: November 7, 2024
 Start time (MST): 11:03
 End time (MST): 15:48
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC755290
 NOX Cal Gas Conc: 48.90 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 48.90 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701H
 Cal Gas Expiry Date: January 3, 2031
 NO Cal Gas Conc: 48.70 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 48.70 ppm
 NO gas Diff:
 Serial Number: 5258
 Serial Number: 832

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.5	0.1	-1.6	----	----
AF High point	4918	82.1	802.9	799.6	3.3	804.2	806.3	-2.0	0.9966	0.9919
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 799.6 ppb	NO = 801.7 ppb	<i>* = > +/-5% change initiates investigation</i>				*Percent Change	NO _x = 0.8%		
Baseline Corr 1st pt	NO _x = 805.7 ppb	NO = 806.2 ppb	<u>As Found Statistics</u>				*Percent Change	NO = 0.6%		
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :		Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :		NO SI:	NO Int:			
			As found	NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1173480006

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997620	0.996610
NO _x Cal Offset:	-1.409395	-1.429437
NO Cal Slope:	1.004775	0.999817
NO Cal Offset:	-1.729194	-2.169452
NO ₂ Cal Slope:	0.995663	0.998596
NO ₂ Cal Offset:	-1.104038	-0.279105

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.315	1.307	NO bkgnd or offset:	2.5	2.7
NOX coeff or slope:	0.984	0.990	NOX bkgnd or offset:	4.1	2.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	161.7	162.2

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
High point	4918	82.1	802.9	799.6	3.3	799.5	798.4	1.1	1.0043	1.0016
Mid point	4959	41.1	401.9	400.3	1.6	398.3	396.9	1.3	1.0092	1.0086
Low point	4979	20.5	200.5	199.7	0.8	197.2	195.5	1.7	1.0168	1.0214
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
As left span	4918	82.1	802.9	360.5	442.4	795.2	360.5	434.7	1.0097	1.0000
Average Correction Factor									1.0101	1.0105

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	788.3	379.1	412.5	412.3	1.0004	100.0%
Mid GPT point	788.3	574.2	217.4	215.2	1.0101	99.0%
Low GPT point	788.3	683.8	107.8	108.0	0.9980	100.2%
Average Correction Factor					1.0029	99.7%

Notes: Sample inlet filter changed after as founds. Adjusted zero and span. Used 2nd NO reference point because of drift.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 4, 2024	Previous Calibration:	November 7, 2024
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	11:03	End Time (MST):	15:48
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999987	≥0.995
802.9	799.5	1.0043	Slope	0.996610	0.90 - 1.10
401.9	398.3	1.0092	Intercept	-1.429437	+/-20
200.5	197.2	1.0168			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 4, 2024	Previous Calibration:	November 7, 2024
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	11:03	End Time (MST):	15:48
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999962	≥0.995
412.5	412.3	1.0004	Slope	0.998596	0.90 - 1.10
217.4	215.2	1.0101	Intercept	-0.279105	+/-20
107.8	108.0	0.9980			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 4, 2024	Previous Calibration:	November 7, 2024
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	11:03	End Time (MST):	15:48
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.1	----	Correlation Coefficient	0.999969	<i>≥0.995</i>
799.6	798.4	1.0016	Slope	0.999817	<i>0.90 - 1.10</i>
400.3	396.9	1.0086	Intercept	-2.169452	<i>+/-20</i>
199.7	195.5	1.0214			



NO_x Calibration Plot

Date: December 4, 2024

Location: Monday Creek






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS506
JACKFISH 1
DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Jackfish 1	Station number:	AMS 506
Calibration Date:	December 3, 2024	Last Cal Date:	November 21, 2024
Start time (MST):	13:34	End time (MST):	16:07
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	50.52	ppm	Cal Gas Exp Date:	December 29, 2028
Cal Gas Cylinder #:	CC274266			
Removed Cal Gas Conc:	50.52	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	2659
Zero Air Gen Model:	Teledyne API T701		Serial Number:	4427

Analyzer Information

Analyzer make:	Thermo 43i	Serial Number:	1160290011
Analyzer Range:	0-1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001358	0.996659	Backgd or Offset:	20.1	20.1
Calibration intercept:	0.143923	0.464029	Coeff or Slope:	0.979	0.979

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4921	79.2	800.2	796.5	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	796.2	Previous response	801.4	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4921	79.2	800.2	797.8	1.003
Mid point	4960	39.6	400.2	399.7	1.001
Low point	4980	19.8	200.1	199.8	1.001
As left zero	5000	0.0	0.0	0.2	----
As left span	4921	79.2	800.2	799.7	1.001
Average Correction Factor:					1.002

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 21, 2024
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	13:34	End Time (MST):	16:07
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995
800.2	797.8	1.0030	Slope	0.996659	0.90 - 1.10
400.2	399.7	1.0011	Intercept	0.464029	+/-30
200.1	199.8	1.0013			



SO2 Calibration Plot

Date: December 3, 2024

Location: Jackfish 1





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name:	Jackfish 1	Station number:	AMS 506
Calibration Date:	December 4, 2024	Last Cal Date:	November 20, 2024
Start time (MST):	8:41	End time (MST):	12:42
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	4.89	ppm	Cal Gas Exp Date:	September 5, 2027
Cal Gas Cylinder #:	CC737971			
Removed Cal Gas Conc:	4.89	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	Teledyne 750		Serial Number:	282
ZAG Make/Model:	Teledyne 751H		Serial Number:	321

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020
Converter make:	Global G150	Converter serial #:	2022-218
Analyzer Range	0 - 100 ppb	Converter Temp:	325.0 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998310	0.989169	Backgd or Offset:	3.56	3.56
Calibration intercept:	0.020574	0.040416	Coeff or Slope:	1.154	1.154

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4918	81.8	80.0	79.0	1.010
As found Mid point	4959	40.9	40.0	39.4	1.010
As found Low point	4980	20.4	19.9	19.6	1.008
New cylinder response					
Baseline Corr As found:	79.2	Prev response:	79.89	*% change:	-0.9%
Baseline Corr 2nd AF pt:	39.6	AF Slope:	0.989739	AF Intercept:	-0.179516
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999999	<i>* = > +/-5% change initiates investigation</i>	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4918	81.8	80.0	79.2	1.010
Mid point	4959	40.9	40.0	39.6	1.010
Low point	4980	20.4	19.9	19.7	1.013
As left zero	5000	0.0	0.0	0.1	----
As left span	4918	81.8	80.0	79.4	1.008
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.011
Date of last converter efficiency test:					

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	December 4, 2024	Previous Calibration:	November 20, 2024
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	8:41	End Time (MST):	12:42
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999997	≥ 0.995
80.0	79.2	1.0101	Slope	0.989169	$0.90 - 1.10$
40.0	39.6	1.0101	Intercept	0.040416	± 3
19.9	19.7	1.0127			



H2S Calibration Plot

Date: December 4, 2024

Location: Jackfish 1





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Jackfish 1
 Station number: AMS 506
 Calibration Date: December 3, 2024
 Last Cal Date: November 19, 2024
 Start time (MST): 9:30
 End time (MST): 13:38
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0022706
 NOX Cal Gas Conc: 60.20 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 60.20 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T750
 ZAG make/model: Teledyne API T751H
 Cal Gas Expiry Date: January 5, 2032
 NO Cal Gas Conc: 60.10 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 60.10 ppm
 NO gas Diff:
 Serial Number: 282
 Serial Number: 321

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.2	-0.3	-0.9	----	----
AF High point	4933	66.6	801.9	800.6	1.3	836.0	827.0	9.8	0.9579	0.9677
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x =	801.7 ppb	NO =	797.9 ppb	<i>* = > +/-5% change initiates investigation</i>			*Percent Change	NO _x =	4.2%
Baseline Corr 1st pt	NO _x =	837.2 ppb	NO =	827.3 ppb	<u>As Found Statistics</u>			*Percent Change	NO =	3.5%
Baseline Corr 2nd pt	NO _x =	NA ppb	NO =	NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:		
Baseline Corr 3rd pt	NO _x =	NA ppb	NO =	NA ppb	As found	NO r ² :	NO SI:	NO Int:		
					As found	NO ₂ r ² :	NO2 SI:	NO2 Int:		

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153356

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002537	1.001681
NO _x Cal Offset:	-2.268674	-1.068623
NO Cal Slope:	1.000867	1.004422
NO Cal Offset:	-3.349973	-2.869782
NO ₂ Cal Slope:	1.005533	1.003838
NO ₂ Cal Offset:	1.350171	0.762410

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.726	1.493	NO bkgnd or offset:	5.3	4.9
NOX coeff or slope:	0.989	0.993	NOX bkgnd or offset:	6.2	6.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	167.4	171.0

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.3	0.0	-0.3	----	----
High point	4933	66.6	801.9	800.6	1.3	802.0	802.0	-0.3	0.9999	0.9982
Mid point	4967	33.3	400.9	400.2	0.7	401.9	399.8	2.0	0.9975	1.0011
Low point	4983	16.6	199.9	199.5	0.3	197.2	193.3	3.9	1.0136	1.0323
As left zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.2	-0.5	----	----
As left span	4933	66.6	801.9	409.7	392.2	814.0	409.7	404.3	0.9852	1.0000
Average Correction Factor									1.0037	1.0106

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.3	----	----
High GPT point	799.8	399.1	402.0	403.6	0.9961	100.4%
Mid GPT point	799.8	606.2	194.9	197.7	0.9860	101.4%
Low GPT point	799.8	706.5	94.6	96.3	0.9827	101.8%
Average Correction Factor					0.9883	101.2%

Notes: Changed inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 19, 2024
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	9:30	End Time (MST):	13:38
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.3	----	Correlation Coefficient	0.999982	<i>≥0.995</i>
801.9	802.0	0.9999	Slope	1.001681	<i>0.90 - 1.10</i>
400.9	401.9	0.9975	Intercept	-1.068623	<i>+/-20</i>
199.9	197.2	1.0136			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 19, 2024
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	9:30	End Time (MST):	13:38
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	-0.3	----	Correlation Coefficient	0.999960	<i>≥0.995</i>
402.0	403.6	0.9961	Slope	1.003838	<i>0.90 - 1.10</i>
194.9	197.7	0.9860	Intercept	0.762410	<i>+/-20</i>
94.6	96.3	0.9827			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 3, 2024	Previous Calibration:	November 19, 2024
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	9:30	End Time (MST):	13:38
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999923	<i>≥0.995</i>
800.6	802.0	0.9982	Slope	1.004422	<i>0.90 - 1.10</i>
400.2	399.8	1.0011	Intercept	-2.869782	<i>+/-20</i>
199.5	193.3	1.0323			



NO_x Calibration Plot

Date: December 3, 2024

Location: Jackfish 1






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS508 KIRBY NORTH DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Kirby North	Station number:	AMS 508
Calibration Date:	December 12, 2024	Last Cal Date:	November 13, 2024
Start time (MST):	9:10	End time (MST):	12:35
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	49.18	ppm	Cal Gas Exp Date:	February 23, 2025
Cal Gas Cylinder #:	<u>CC303554</u>			
Removed Cal Gas Conc:	49.18	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	<u>NA</u>		Diff between cyl:	
Calibrator Model:	Teledyne API T700		Serial Number:	5240
Zero Air Gen Model:	Teledyne API T701H		Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 43iQ	Serial Number:	1182340007
Analyzer Range:	0 - 1000 ppb		

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002407	0.994819	Backgd or Offset:	26.6	27.3
Calibration intercept:	0.151064	0.431080	Coeff or Slope:	1.065	1.072

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.4	----
As found High point	4919	81.3	799.6	792.7	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	792.3	Previous response	801.7	*% change	-1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4919	81.3	799.6	796.0	1.005
Mid point	4959	40.7	400.3	398.0	1.006
Low point	4980	20.3	199.7	200.1	0.998
As left zero	5000	0.0	0.0	-0.1	----
As left span	4919	81.3	799.6	798.5	1.001
Average Correction Factor:					1.003

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 13, 2024
Station Name:	Kirby North	Station Number:	AMS 508
Start Time (MST):	9:10	End Time (MST):	12:35
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999995
799.6	796.0	1.0045	Slope	0.994819
400.3	398.0	1.0059	Intercept	0.431080
199.7	200.1	0.9978		
				≥0.995
				0.90 - 1.10
				+/-30



SO2 Calibration Plot

Date: December 12, 2024

Location: Kirby North





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name:	Kirby North	Station number:	AMS 508
Calibration Date:	December 11, 2024	Last Cal Date:	November 13, 2024
Start time (MST):	9:10	End time (MST):	11:30
Reason:	As Found		

Calibration Standards

Cal Gas Concentration:	5.05 ppm	Cal Gas Exp Date:	November 15, 2026
Cal Gas Cylinder #:	<u>DT0019762</u>		
Removed Cal Gas Conc:	5.05 ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	n/a	Diff between cyl:	
Calibrator Make/Model:	Teledyne API T700	Serial Number:	5240
ZAG Make/Model:	Teledyne API T701H	Serial Number:	880

Analyzer Information

Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012
Converter make:	Global	Converter serial #:	2022-197
Analyzer Range	0 - 100 ppb	Converter Temp:	325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996242	1.001242	Backgd or Offset:	1.75	1.74
Calibration intercept:	0.259036	-0.240953	Coeff or Slope:	1.037	1.030

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4921	79.2	80.0	79.8	1.001
As found Mid point	4960	39.6	40.0	39.8	1.001
As found Low point	4980	19.8	20.0	19.6	1.013
New cylinder response					
Baseline Corr As found:	79.9	Prev response:	79.95	*% change:	-0.1%
Baseline Corr 2nd AF pt:	39.9	AF Slope:	0.999849	AF Intercept:	-0.195960
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999989	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	79.2	80.0	80.0	1.000
Mid point	4960	39.6	40.0	39.6	1.010
Low point	4980	19.8	20.0	19.6	1.020
As left zero	5000	0.0	0.0	0.0	----
As left span	4921	79.2	80.0	79.4	1.007
SO2 Scrubber Check	4919	80.0	800.2	-0.1	----
Date of last scrubber change:		July 25, 2023		Ave Corr Factor	1.010
Date of last converter efficiency test:		n/a			

Notes: Calibration completed. Station lost power from 12:30 to 16:30 MST. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	December 11, 2024	Previous Calibration:	November 13, 2024
Station Name:	Kirby North	Station Number:	AMS 508
Start Time (MST):	9:10	End Time (MST):	11:30
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999955	≥ 0.995
80.0	80.0	0.9999	Slope	1.001242	$0.90 - 1.10$
40.0	39.6	1.0101	Intercept	-0.240953	± 3
20.0	19.6	1.0203			





Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	December 12, 2024	Previous Calibration:	November 13, 2024
Station Name:	Kirby North	Station Number:	AMS 508
Start Time (MST):	9:10	End Time (MST):	12:35
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1182340005

Calibration Data


Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.06	----	Correlation Coefficient	0.999988	≥0.995
17.26	17.24	1.0013	Slope	0.996210	0.90 - 1.10
8.64	8.63	1.0015	Intercept	0.030089	+/-1.5
4.31	4.29	1.0039			



THC Calibration Plot

Date: December 12, 2024

Location: Kirby North





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Kirby North
 Station number: AMS 508
 Calibration Date: December 10, 2024
 Last Cal Date: November 13, 2024
 Start time (MST): 15:30
 End time (MST): 20:50
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: T34ULGL
 NOX Cal Gas Conc: 49.39 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 49.39 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701H
 Cal Gas Expiry Date: March 8, 2025
 NO Cal Gas Conc: 49.02 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 49.02 ppm
 NO gas Diff:
 Serial Number: 5240
 Serial Number: 880

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
AF High point	4918	81.6	806.1	800.1	6.0	822.0	814.0	8.0	0.9807	0.9829
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 807.1 ppb	NO = 800.8 ppb	<i>* = > +/-5% change initiates investigation</i>				*Percent Change	NO _x = 1.8%		
Baseline Corr 1st pt	NO _x = 822.0 ppb	NO = 814.0 ppb	<u>As Found Statistics</u>				*Percent Change	NO = 1.6%		
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :		Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :		NO SI:	NO Int:			
			As found	NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1118148496

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003355	1.001923
NO _x Cal Offset:	-1.743898	-1.463910
NO Cal Slope:	1.003858	1.003129
NO Cal Offset:	-2.344036	-1.964029
NO ₂ Cal Slope:	0.987139	0.978262
NO ₂ Cal Offset:	1.475739	2.102038

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.711	0.698	NO bkgnd or offset:	7.7	7.6
NOX coeff or slope:	0.994	0.993	NOX bkgnd or offset:	7.7	7.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	146.5	146.8

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
High point	4918	81.6	806.1	800.1	6.0	807.0	801.8	5.2	0.9989	0.9978
Mid point	4959	40.8	403.0	400.0	3.0	401.5	397.8	3.7	1.0038	1.0056
Low point	4980	20.4	201.5	200.0	1.5	198.9	196.9	2.0	1.0130	1.0157
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.0	----	----
As left span	4918	81.6	806.1	390.5	415.6	795.2	390.5	404.7	1.0137	1.0000
Average Correction Factor									1.0053	1.0064

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	797.0	390.0	413.0	404.6	1.0207	98.0%
Mid GPT point	797.0	609.6	193.4	194.1	0.9963	100.4%
Low GPT point	797.0	702.1	100.9	101.6	0.9930	100.7%
Average Correction Factor					1.0034	99.7%

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 10, 2024	Previous Calibration:	November 13, 2024
Station Name:	Kirby North	Station Number:	AMS 508
Start Time (MST):	15:30	End Time (MST):	20:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1118148496

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999983	<i>≥0.995</i>
806.1	807.0	0.9989	Slope	1.001923	<i>0.90 - 1.10</i>
403.0	401.5	1.0038	Intercept	-1.463910	<i>+/-20</i>
201.5	198.9	1.0130			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 10, 2024	Previous Calibration:	November 13, 2024
Station Name:	Kirby North	Station Number:	AMS 508
Start Time (MST):	15:30	End Time (MST):	20:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1118148496

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999829	<i>≥0.995</i>
413.0	404.6	1.0207	Slope	0.978262	<i>0.90 - 1.10</i>
193.4	194.1	0.9963	Intercept	2.102038	<i>+/-20</i>
100.9	101.6	0.9930			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 10, 2024	Previous Calibration:	November 13, 2024
Station Name:	Kirby North	Station Number:	AMS 508
Start Time (MST):	15:30	End Time (MST):	20:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1118148496

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999969	<i>≥0.995</i>
800.1	801.8	0.9978	Slope	1.003129	<i>0.90 - 1.10</i>
400.0	397.8	1.0056	Intercept	-1.964029	<i>+/-20</i>
200.0	196.9	1.0157			



NO_x Calibration Plot

Date: December 10, 2024

Location: Kirby North






WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS512 HANGINGSTONE EXPANSION DECEMBER 2024

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

January 31, 2025





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name:	Hangingsstone Expansion	Station number: AMS 512
Calibration Date:	December 9, 2024	Last Cal Date: November 22, 2024
Start time (MST):	7:48	End time (MST): 10:46
Reason:	Routine	

Calibration Standards

Cal Gas Concentration:	50.06 ppm	Cal Gas Exp Date: January 5, 2029
Cal Gas Cylinder #:	CC147416	
Removed Cal Gas Conc:	50.06 ppm	Rem Gas Exp Date: NA
Removed Gas Cyl #:	NA	Diff between cyl:
Calibrator Model:	Teledyne API T700	Serial Number: 2445
Zero Air Gen Model:	Teledyne API 701	Serial Number: 138

Analyzer Information

Analyzer make:	Thermo scientific	Serial Number: 1173410001
Analyzer Range:	0 - 1000 ppb	

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002887	1.003116	Backgd or Offset:	13.6	14.1
Calibration intercept:	-1.243382	-1.623537	Coeff or Slope:	1.164	1.157

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.4	----
As found High point	4920	79.8	799.0	802.1	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	801.7	Previous response	800.1	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		<i>* = > +/-5% change initiates investigation</i>	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4920	79.8	799.0	800.9	0.998
Mid point	4960	39.9	399.5	397.5	1.005
Low point	4987	20.0	200.0	198.0	1.010
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	79.8	799.0	801.9	0.996
Average Correction Factor:					1.004

Notes: No maintenance done. Zero and Span adjusted

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	December 9, 2024	Previous Calibration:	November 22, 2024
Station Name:	Hangystone Expansion	Station Number:	AMS 512
Start Time (MST):	7:48	End Time (MST):	10:46
Analyzer make:	Thermo scientific	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	<u>Limits</u>	
0.0	-0.1	----	Correlation Coefficient	0.999980	≥0.995
799.0	800.9	0.9976	Slope	1.003116	0.90 - 1.10
399.5	397.5	1.0050	Intercept	-1.623537	+/-30
200.0	198.0	1.0099			



SO2 Calibration Plot

Date: December 9, 2024

Location: Hangingstone Expansion





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name:	Hangingstone Expansion	Station number:	AMS 512
Calibration Date:	December 6, 2024	Last Cal Date:	November 20, 2024
Start time (MST):	7:30	End time (MST):	11:41
Reason:	Routine		

Calibration Standards

Cal Gas Concentration:	5.139	ppm	Cal Gas Exp Date:	January 3, 2026
Cal Gas Cylinder #:	CC511397			
Removed Cal Gas Conc:	5.139	ppm	Rem Gas Exp Date:	NA
Removed Gas Cyl #:	NA		Diff between cyl:	
Calibrator Make/Model:	API T700		Serial Number:	2445
ZAG Make/Model:	API T701		Serial Number:	138

Analyzer Information

Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1336160090
Converter make:	Global G150	Converter serial #:	2022-227
Analyzer Range	0 - 100 ppb	Converter Temp:	325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.983476	0.997057	Backgd or Offset:	3.51	3.57
Calibration intercept:	0.181186	-0.019091	Coeff or Slope:	1.194	1.210

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4922	77.8	80.0	79.3	1.011
As found Mid point	4961	38.9	40.0	39.6	1.015
As found Low point	4981	19.5	20.0	19.6	1.033
New cylinder response					
Baseline Corr As found:	79.1	Prev response:	78.83	*% change:	0.3%
Baseline Corr 2nd AF pt:	39.4	AF Slope:	0.990767	AF Intercept:	0.001028
Baseline Corr 3rd AF pt:	19.4	AF Correlation:	0.999968	<i>* = > +/-5% change initiates investigation</i>	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4922	77.8	80.0	79.8	1.002
Mid point	4961	38.9	40.0	39.8	1.005
Low point	4981	19.5	20.0	19.7	1.017
As left zero	5000	0.0	0.0	0.4	----
As left span	4922	77.8	80.0	78.9	1.014
SO ₂ Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.008
Date of last converter efficiency test:					

Notes: Sox scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	December 6, 2024	Previous Calibration:	November 20, 2024
Station Name:	Hangingstone Expansion	Station Number:	AMS 512
Start Time (MST):	7:30	End Time (MST):	11:41
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1336160090

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999964	≥0.995
80.0	79.8	1.0021	Slope	0.997057	0.90 - 1.10
40.0	39.8	1.0046	Intercept	-0.019091	+/-3
20.0	19.7	1.0173			



H₂S Calibration Plot

Date: December 6, 2024

Location: Hangingstone Expansion





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Hangingstone Expansion
 Station number: AMS 512
 Calibration Date: December 2, 2024
 Last Cal Date: November 18, 2024
 Start time (MST): 7:55
 End time (MST): 12:01
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: TOF8P52
 NOX Cal Gas Conc: 47.43 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 47.43 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: August 16, 2026
 NO Cal Gas Conc: 47.43 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 47.43 ppm
 NO gas Diff:
 Serial Number: 2445
 Serial Number: 138

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
AF High point	4916	84.4	800.6	800.6	0.0	799.5	799.1	0.5	1.0012	1.0016
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 799.9 ppb	NO = 799.5 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 0.0%
Baseline Corr 1st pt	NO _x = 799.6 ppb	NO = 799.3 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
 NOX Range (ppb): 0 - 1000 ppb

Serial Number: 7029

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000766	1.001494
NO _x Cal Offset:	-1.272780	-0.852815
NO Cal Slope:	1.000880	1.001080
NO Cal Offset:	-1.712772	-1.232796
NO ₂ Cal Slope:	0.997612	0.996045
NO ₂ Cal Offset:	0.073951	0.110093

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.056	1.056	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.052	1.052	NOX bkgnd or offset:	0.4	0.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.6	4.6

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4916	84.4	800.6	800.6	0.0	801.1	800.6	0.6	0.9993	0.9999
Mid point	4958	42.2	400.3	400.3	0.0	400.3	399.5	0.8	1.0000	1.0020
Low point	4979	21.1	200.2	200.2	0.0	198.3	197.5	0.8	1.0094	1.0134
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
As left span	4916	84.4	800.6	406.4	394.2	795.8	406.4	389.4	1.0060	1.0000
Average Correction Factor									1.0029	1.0051

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	799.7	404.1	395.6	394.2	1.0036	99.6%
Mid GPT point	799.7	618.2	181.5	180.6	1.0050	99.5%
Low GPT point	799.7	709.2	90.5	90.6	0.9989	100.1%
Average Correction Factor					1.0025	99.8%

Notes:

No adjustments and maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 18, 2024
Station Name:	Hangingsstone Expansion	Station Number:	AMS 512
Start Time (MST):	7:55	End Time (MST):	12:01
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999993	≥0.995
800.6	801.1	0.9993	Slope	1.001494	0.90 - 1.10
400.3	400.3	1.0000	Intercept	-0.852815	+/-20
200.2	198.3	1.0094			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 18, 2024
Station Name:	Hangingsstone Expansion	Station Number:	AMS 512
Start Time (MST):	7:55	End Time (MST):	12:01
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999997	<i>≥0.995</i>
395.6	394.2	1.0036	Slope	0.996045	<i>0.90 - 1.10</i>
181.5	180.6	1.0050	Intercept	0.110093	<i>+/-20</i>
90.5	90.6	0.9989			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	December 2, 2024	Previous Calibration:	November 18, 2024
Station Name:	Hangingsstone Expansion	Station Number:	AMS 512
Start Time (MST):	7:55	End Time (MST):	12:01
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999988	<i>≥0.995</i>
800.6	800.6	0.9999	Slope	1.001080	<i>0.90 - 1.10</i>
400.3	399.5	1.0020	Intercept	-1.232796	<i>+/-20</i>
200.2	197.5	1.0134			



NO_x Calibration Plot

Date: December 2, 2024

Location: Hangingstone Expansion





End of Report