



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

APRIL 2025 MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

May 30, 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01 BERTHA GANTER - FORT MCKAY APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: April 9, 2025 Last Cal Date: March 7, 2025
Start time (MST): 11:39 End time (MST): 15:18
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.21 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC418809
Removed Cal Gas Conc: 49.21 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3565
Zero Air Gen Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1501301448
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001209	1.001108	Backgd or Offset:	20.8	21.7
Calibration intercept:	-0.193570	-0.433285	Coeff or Slope:	0.887	0.887

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.8	----
As found High point	4918	81.3	800.3	802.6	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	801.8	Previous response	801.0	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4918	81.3	800.3	801.1	0.999
Mid point	4959	40.7	400.6	399.9	1.002
Low point	4979	20.3	199.8	199.5	1.002
As left zero	5000	0.0	0.0	0.0	----
As left span	4918	81.3	800.3	801.7	0.998
Average Correction Factor:					1.001

Notes: Changed the inlet filter after as founds. Adjusted the zero.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	March 7, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:39	End Time (MST):	15:18
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999998	≥0.995
800.3	801.1	0.9990	Slope	1.001108	0.90 - 1.10
400.6	399.9	1.0017	Intercept	-0.433285	+/-30
199.8	199.5	1.0016			



SO2 Calibration Plot

Date: April 9, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: April 28, 2025 Last Cal Date: March 27, 2025
Start time (MST): 10:09 End time (MST): 14:53
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: September 5, 2027
Cal Gas Cylinder #: CC738239
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 12113311966
Converter make: CD Nova Converter serial #: 470
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995103	1.001390	Backgd or Offset:	2.22	2.18
Calibration intercept:	0.201884	-0.118138	Coeff or Slope:	1.024	1.148

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4917	82.6	80.0	78.3	1.019
As found Mid point	4959	41.3	40.0	39.0	1.020
As found Low point	4979	20.7	20.0	19.6	1.012
New cylinder response					
Baseline Corr As found:	78.5	Prev response:	79.77	*% change:	-1.6%
Baseline Corr 2nd AF pt:	39.2	AF Slope:	0.981091	AF Intercept:	-0.157790
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4917	82.6	80.0	80.0	1.000
Mid point	4959	41.3	40.0	39.8	1.004
Low point	4979	20.7	20.0	20.0	1.002
As left zero	5000	0.0	0.0	0.0	----
As left span	4917	82.6	80.0	79.0	1.012
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	1.002
Date of last converter efficiency test:					

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	April 28, 2025	Previous Calibration:	March 27, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:09	End Time (MST):	14:53
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	12113311966

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999995		≥ 0.995
80.0	80.0	0.9995	Slope	1.001390		$0.90 - 1.10$
40.0	39.8	1.0044	Intercept	-0.118138		± 3
20.0	20.0	1.0019				


TRS Calibration Curve



TRS Calibration Plot

Date: April 28, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: April 28, 2025 Last Cal Date: March 19, 2025
Start time (MST): 10:09 End time (MST): 14:53
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: September 5, 2027
Cal Gas Cylinder #: CC738239
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 1200326167
Converter make: CD Nova Converter serial #: 2022-221
Analyzer Range: 0 - 100 ppb Converter Temp: 315 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000243	1.002388	Backgd or Offset:	2.03
Calibration intercept:	0.021999	-0.178072	Coeff or Slope:	0.983

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4917	82.6	80.0	80.3	0.995
As found Mid point	4959	41.3	40.0	40.1	0.994
As found Low point	4979	20.7	20.0	20.0	0.997
New cylinder response					
Baseline Corr As found:	80.4	Prev response:	80.00	*% change:	0.5%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	1.005677	AF Intercept:	-0.118167
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	1.000000	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4917	82.6	80.0	80.0	1.000
Mid point	4959	41.3	40.0	39.9	1.002
Low point	4979	20.7	20.0	19.8	1.012
As left zero	5000	0.0	0.0	0.0	----
As left span	4917	82.6	80.0	79.4	1.007
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	January 25, 2024		Ave Corr Factor		1.005
Date of last converter efficiency test:	November 7, 2024		107.9% efficiency		

Notes: Inlet filter change and scrubber check completed after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 28, 2025	Previous Calibration:	March 19, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:09	End Time (MST):	14:53
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999995		≥ 0.995
80.0	80.0	0.9995	Slope	1.002388		$0.90 - 1.10$
40.0	39.9	1.0019	Intercept	-0.178072		± 3
20.0	19.8	1.0121				



H₂S Calibration Plot

Date: April 28, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
 Calibration Date: April 9, 2025 Last Cal Date: March 9, 2025
 Start time (MST): 11:39 End time (MST): 15:18
 Reason: Routine

Calibration Standards

Gas Cert Reference: CC418809 Cal Gas Expiry Date: March 10, 2031
 CH₄ Cal Gas Conc. 497.2 ppm CH₄ Equiv Conc. 1061.8 ppm
 C₃H₈ Cal Gas Conc. 205.3 ppm
 Removed Gas Cert: NA Removed Gas Expiry: NA
 Removed CH₄ Conc. 497.2 ppm CH₄ Equiv Conc. 1061.8 ppm
 Removed C₃H₈ Conc. 205.3 ppm Diff between cyl (THC):
 Diff between cyl (CH₄): Diff between cyl (NMHC):
 Calibrator Model: Teledyne API T700 Serial Number: 3565
 Zero Air Gen model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 1193585648
 THC Range: 0 - 20 ppm NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.47E-04	2.53E-04	NMHC SP Ratio:	4.82E-05
CH ₄ Retention time:	14.8	15.0	NMHC Peak Area:	190523
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	17.27	16.99	1.016
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.99	Prev response	17.29	*% change	-1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	17.27	17.32	0.997
Mid point	4959	40.7	8.64	8.70	0.994
Low point	4979	20.3	4.31	4.36	0.990
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	17.27	17.31	0.998
Average Correction Factor					0.994

Notes: Changed the inlet filter after as founds. Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4918	81.3	9.18	9.09	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.09	Prev response	9.17	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	9.18	9.22	0.996
Mid point	4959	40.7	4.60	4.65	0.989
Low point	4979	20.3	2.29	2.34	0.981
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	9.18	9.22	0.996
Average Correction Factor					0.989

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4918	81.3	8.09	7.91	1.023
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.91	Prev response	8.11	*% change	-2.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	8.09	8.10	0.998
Mid point	4959	40.7	4.05	4.05	0.999
Low point	4979	20.3	2.02	2.02	0.999
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	8.09	8.09	1.000
Average Correction Factor					0.999

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.999822	1.002190
THC Cal Offset:	0.026931	0.021536
CH ₄ Cal Slope:	1.004040	1.001397
CH ₄ Cal Offset:	-0.003534	0.001066
NMHC Cal Slope:	0.995772	1.002902
NMHC Cal Offset:	0.031064	0.020670

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	March 9, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:39	End Time (MST):	15:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995
17.27	17.32	0.9972	Slope	1.002190	$0.90 - 1.10$
8.64	8.70	0.9938	Intercept	0.021536	± 0.5
4.31	4.36	0.9895			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	March 9, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:39	End Time (MST):	15:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999999	<i>≥0.995</i>
8.09	8.10	0.9982	Slope	1.001397	<i>0.90 - 1.10</i>
4.05	4.05	0.9991	Intercept	0.001066	<i>+/-0.5</i>
2.02	2.02	0.9990			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	March 9, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:39	End Time (MST):	15:18
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999975	<i>≥0.995</i>
9.18	9.22	0.9962	Slope	1.002902	<i>0.90 - 1.10</i>
4.60	4.65	0.9890	Intercept	0.020670	<i>+/-0.5</i>
2.29	2.34	0.9814			



NMHC Calibration Plot

Date: April 9, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay
Station number: AMS 01
Calibration Date: April 17, 2025
Last Cal Date: March 11, 2025
Start time (MST): 10:36
End time (MST): 16:12
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC335700
NOX Cal Gas Conc: 59.40 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.40 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: September 1, 2032
NO Cal Gas Conc: 59.20 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.20 ppm
NO gas Diff:
Serial Number: 3565
Serial Number: 146

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.5	-0.1	----	----
AF High point	4932	67.6	803.1	800.4	2.7	803.0	796.2	6.8	0.9994	1.0046
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 801.9 ppb	NO = 799.9 ppb				* = > +/-5% change initiates investigation			*Percent Change	NO _x = 0.2%
Baseline Corr 1st pt	NO _x = 803.6 ppb	NO = 796.7 ppb				<u>As Found Statistics</u>			*Percent Change	NO = -0.4%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found NO _x r ² :			Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found NO r ² :			NO SI:	NO Int:
						As found NO ₂ r ² :			NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 7117

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.176	1.176	NO bkgnd or offset:	-3.1	-3.1
NOX coeff or slope:	1.178	1.178	NOX bkgnd or offset:	-2.9	-2.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	7.7	8.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999328	1.001733
NO _x Cal Offset:	-0.640000	-0.160000
NO Cal Slope:	1.001476	1.003133
NO Cal Offset:	-1.660000	-1.240000
NO ₂ Cal Slope:	1.000095	1.000579
NO ₂ Cal Offset:	0.729865	-0.003919

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.0	----	----
High point	4932	67.6	803.1	800.4	2.7	804.6	802.4	2.3	0.9981	0.9975
Mid point	4966	33.8	401.5	400.2	1.4	401.6	399.3	2.3	0.9999	1.0022
Low point	4983	16.9	200.8	200.1	0.7	200.8	198.3	2.5	0.9999	1.0091
As left zero	5000	0.0	0.0	0.0	0.0	1.8	1.8	0.0	----	----
As left span	4932	67.6	803.1	371.4	431.7	782.2	371.4	410.8	1.0267	1.0000
Average Correction Factor									0.9993	1.0029

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	795.1	404.9	392.9	393.1	0.9995	100.0%
Mid GPT point	795.1	594.3	203.5	203.7	0.9990	100.1%
Low GPT point	795.1	696.7	101.1	101.1	1.0000	100.0%
Average Correction Factor					0.9995	100.0%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 17, 2025	Previous Calibration:	March 11, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:36	End Time (MST):	16:12
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999999	≥0.995
803.1	804.6	0.9981	Slope	1.001733	0.90 - 1.10
401.5	401.6	0.9999	Intercept	-0.160000	+/-20
200.8	200.8	0.9999			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 17, 2025	Previous Calibration:	March 11, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:36	End Time (MST):	16:12
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	1.000000	≥0.995
392.9	393.1	0.9995	Slope	1.000579	0.90 - 1.10
203.5	203.7	0.9990	Intercept	-0.003919	+/-20
101.1	101.1	1.0000			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 17, 2025	Previous Calibration:	March 11, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:36	End Time (MST):	16:12
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.1	----	Correlation Coefficient	0.999987	≥ 0.995
800.4	802.4	0.9975	Slope	1.003133	0.90 - 1.10
400.2	399.3	1.0022	Intercept	-1.240000	+/-20
200.1	198.3	1.0091			



NO_x Calibration Plot

Date: April 17, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: April 1, 2025 Last Cal Date: March 5, 2025
Start time (MST): 11:02 End time (MST): 14:24
Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 1107
Analyzer Range: 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001943	0.996600	Backgd or Offset:	7.6	6.8
Calibration intercept:	0.260000	0.920000	Coeff or Slope:	1.021	1.031

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.6	----
As found High point	5000	863.1	400.0	395.7	1.009
As found Mid point					
As found Low point					
Baseline Corr As found:	396.3	Previous response	401.0	*% change	-1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	5000	863.1	400.0	399.2	1.002
Mid point	5000	744.0	200.0	200.5	0.998
Low point	5000	651.7	100.0	101.5	0.985
As left zero	5000	0.0	0.0	0.0	----
As left span	5000	863.1	400.0	400.2	1.000
Average Correction Factor					0.995

Notes: Changed the inlet filter after as founds. Adjusted both zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 1, 2025	Previous Calibration:	March 5, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:02	End Time (MST):	14:24
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999980	≥0.995
400.0	399.2	1.0020	Slope	0.996600	0.90 - 1.10
200.0	200.5	0.9975	Intercept	0.920000	+/- 5
100.0	101.5	0.9852			



O₃ Calibration Plot

Date: April 1, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
Calibration Date: April 2, 2025 Last Cal Date: March 20, 2025
Start time (MST): 11:59 End time (MST): 14:17

Analyzer Make: Teledyne API T640 S/N: 322
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752
Temp/RH standard: Alicat FP-25BT S/N: 388752

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-0.4	-0.9	-0.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	728.2	730.53	728.2	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	4.978	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39		39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA:	17.2	PM w/ HEPA:	17.2	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	7.5	10.2	10.8	<input checked="" type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: April 2, 2025
Date Disposable Filter Changed: April 2, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 24, 2024
Date RH/T Sensor Cleaned: April 2, 2025

Notes: Flow, temperature, and pressure were verified. Initial leak check did not pass. Optical chamber and RH/T sensor cleaned. Disposable filter changed. PMT peak adjusted. Post maintenance leak check passed.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: April 2, 2025 Last Cal Date: March 3, 2025
Start time (MST): 11:08 End time (MST): 14:45
Reason: Routine

Calibration Standards

Cal Gas Concentration: 3,040 ppm Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #: ALM042207
Removed Cal Gas Conc: 3,040 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Teledyne API T300 Analyzer serial #: 3520
Analyzer Range: 0 - 50 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003849	1.002893	Backgd or Offset:	-0.015
Calibration intercept:	0.071854	0.133812	Coeff or Slope:	0.994

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4933	66.7	40.6	41.0	0.992
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	40.89	Prev response:	40.78	*% change:	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4933	66.7	40.6	40.7	0.997
Mid point	4966	33.3	20.2	20.7	0.979
Low point	4983	16.7	10.2	10.3	0.983
As left zero	5000	0.0	0.0	0.0	----
As left span	2960	40.0	40.5	40.5	1.000
Average Correction Factor					0.986

Notes: Changed the inlet filter after as founds. Adjusted the zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

CO Calibration Summary


Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	March 3, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:08	End Time (MST):	14:45
Analyzer make:	Teledyne API T300	Analyzer serial #:	3520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999894	≥ 0.995
40.6	40.7	0.9970	Slope	1.002893	$0.90 - 1.10$
20.2	20.7	0.9787	Intercept	0.133812	± 1.5
10.2	10.3	0.9830			


CO Calibration Curve



CO Calibration Plot

Date: April 2, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

CO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: April 4, 2025 Last Cal Date: March 4, 2025
Start time (MST): 10:58 End time (MST): 14:09
Reason: Routine

Calibration Standards

Cal Gas Concentration: 60,200 ppm Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #: ALM042207
Removed Cal Gas Conc: 60,200 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
N2 Gen Make/Model: Peak Scientific Serial Number: 7220900034

Analyzer Information

Analyzer make: Teledyne API 360 Analyzer serial #: 442
Analyzer Range: 0 - 2,000 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005093	1.000358	Backgd or Offset:	-0.011	-0.011
Calibration intercept:	-5.360000	-3.460000	Coeff or Slope:	0.922	0.922

CO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero) Limit = 0.90-1.10
As found zero	3000	0.0	0.0	0.5	----
As found High Point	2920	80.0	1605.3	1606.1	1.000
As found Mid Point					
As found Low Point					
New cylinder response					
Baseline Corr As found:	1605.6	Prev response:	1608.1	*% change:	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	3000	0.0	0.0	0.6	----
High point	2920	80.0	1605.3	1607.1	0.999
Mid point	2960	40.0	802.7	789.6	1.017
Low point	2980	20.0	401.3	399.2	1.005
As left zero	3000	0.0	0.0	-3.1	----
As left span	2960	40.0	802.7	791.3	1.014

Average Correction Factor **1.007**

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


CO₂ Calibration Summary

Station Information

Calibration Date	April 4, 2025	Previous Calibration	March 4, 2025
Station Name	Bertha Ganter-Fort McKay	Station Number	AMS 01
Start Time (MST)	10:58	End Time (MST)	14:09
Analyzer make	Teledyne API 360	Analyzer serial #	442

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999903	≥0.995
1605.3	1607.1	0.9989	Slope	1.000358	0.90 - 1.10
802.7	789.6	1.0165	Intercept	-3.5	+/-20
401.3	399.2	1.0053			



CO₂ Calibration Plot

Date: April 4, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

Nt - NOX - NH3 Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
NOX Cal Date:	April 14, 2025	Last Cal Date:	March 12, 2025
Start time (MST):	11:00	End time (MST):	15:55
NH3 Cal Date:	April 15, 2025	Last Cal Date:	March 13, 2025
Start time (MST):	11:10	End time (MST):	15:00
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	59.40	ppm	NO Gas Cylinder #:	CC335700
NO Cal Gas Conc:	59.20	ppm	NO Cal Gas Expiry:	September 1, 2032
Removed NOX Conc:	59.40	ppm	Removed Cylinder #:	NA
Removed NO Conc:	59.20	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	77.80	ppm	NH3 Gas Cylinder #:	CC711249
			NH3 Cal Gas Expiry:	December 31, 2025
Removed NH3 Conc:	77.80	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	API T700		Serial Number:	3565
ZAG make/model:	API T701		Serial Number:	146

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	808
Converter model:	API T501	Converter serial #:	484
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	4.40
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	393

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.911	0.903	Nt coefficient:	0.918	0.907
NOX coefficient:	0.912	0.905	NO bkgnd:	-1.8	0.5
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-1.7	0.8
NH3 coefficient:	0.983	0.983	Nt bkgnd:	-0.6	2.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998844	1.001577
NO _x Cal Offset:	-1.020000	-1.280000
NO Cal Slope:	1.001505	1.002933
NO Cal Offset:	-1.720000	-1.820000
NO ₂ Cal Slope:	1.004855	1.002698
NO ₂ Cal Offset:	-0.045831	-0.301647
NH3 Cal Slope:	0.998529	0.998379
NH3 Cal Offset:	1.260645	-6.890779
Nt Cal Slope:	1.002958	1.003583
Nt Cal Offset:	3.555325	-6.335266



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

NO_x / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO _x Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>	Baseline corr NO Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>
As found zero	5000	0.0	0.0	0.0	0.0	1.5	1.1	1.3	----	----
As found span	4932	67.6	803.1	800.4	803.1	814.7	803.7	817.0	0.9857	0.9959
AF GPT span	4932	67.6	803.1	-----	803.1	796.0	-----	795.6	1.0089	-----

new NO cyl rp

Baseline Corr As Fd Nt = 815.7 ppb NO_x = 813.2 ppb NO = 802.6 ppb

Previous Response Nt = 809.02 ppb NO_x = 801.1 ppb NO = 799.9 ppb

****NO_x Δ (NO to GPT response) = -2.3%**

*** = > +/-2% difference initiates investigation**

***Percent Change Nt_(NO) = 0.8%**

***Percent Change NO_x = 1.5%**

***Percent Change NO = 0.3%**

*** = > +/-5% change initiates investigation**

NO_x / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	-0.1	----	----
High point	4932	67.6	803.1	800.4	803.1	803.9	801.8	804.0	0.9990	0.9982
Mid point	4966	33.8	401.5	400.2	401.5	399.5	398.5	401.1	1.0051	1.0042
Low point	4983	16.9	200.8	200.1	200.8	199.3	197.4	199.9	1.0074	1.0137
Average Correction Factor									1.0038	1.0054

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Calibration zero	----	----	0.0	0.0	----	----
High GPT point (400 ppb O3)	797.2	403.6	396.3	397.4	0.9972	100.3%
Mid GPT point (200 ppb O3)	797.2	596.7	203.2	202.8	1.0020	99.8%
Low GPT point (100 ppb O3)	797.2	697.3	102.6	102.6	1.0000	100.0%
Average Correction Factor					0.9998	100.0%



Wood Buffalo Environmental Association NH₃ - N_T Calibration Report

NH₃ As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH3 Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.3	-0.6	----	----
AF High point	2929	70.5	1828.6	----	1828.6	1805.4	----	1795.9	1.012	1.018
AF Mid point							----			
AF Low point							----			
new NH3 cyl rp										
Baseline Corr As Fd	Nt =	1806.2 ppb	NH3 =	1796.5 ppb					*Percent Change	Nt _(NH3) = -1.7%
Previous Response	Nt =	1837.6 ppb	NH3 =	1827.2 ppb					*Percent Change	NH3 = -1.7%

* = > +/-5% change initiates investigation

NH₃ Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
High point	2931	69.4	1799.8	----	1799.8	1805.4	----	1795.9	0.997	1.002
Mid point	2961	38.6	1001.0	----	1001.0	990.2	----	984.1	1.011	1.017
Low point	2981	19.3	500.5	----	500.5	492.3	----	488.3	1.017	1.025
Average Correction Factor									1.0082	1.0148
NH3 Previous Converter Efficiency =		98.3 %								
NH3 Current Converter Efficiency =		98.3 %								

Notes:

Changed the inlet filter after as founds. Adjusted the NOx/NT zero and span.

Calibration Performed By:

Rene Chamberland



Wood Buffalo Environmental Association


Nt Calibration Summary

Station Information

Calibration Date:	April 15, 2025	Previous Calibration:	March 12, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:00	End Time (MST):	15:55
Analyzer make:	API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999917	≥ 0.995
1799.8	1805.4	0.9969	Slope	1.003583	0.90 - 1.10
1001.0	990.2	1.0109	Intercept	-6.335266	+/-20
500.5	492.3	1.0167			





Wood Buffalo Environmental Association


NH₃ Calibration Summary

Station Information

Calibration Date:	April 15, 2025	Previous Calibration:	March 12, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:00	End Time (MST):	15:55
Analyzer make:	API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999901	≥ 0.995
1799.8	1795.9	1.0022	Slope	0.998379	0.90 - 1.10
1001.0	984.1	1.0172	Intercept	-6.890779	+/-20
500.5	488.3	1.0250			





Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 14, 2025	Previous Calibration:	March 12, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:00	End Time (MST):	15:55
Analyzer make:	API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999989	≥0.995
803.1	803.9	0.9990	Slope	1.001577	0.90 - 1.10
401.5	399.5	1.0051	Intercept	-1.280000	+/-20
200.8	199.3	1.0074			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 14, 2025	Previous Calibration:	March 12, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:00	End Time (MST):	15:55
Analyzer make:	API T201	Analyzer serial #:	808

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999981	≥ 0.995
800.4	801.8	0.9982	Slope	1.002933	0.90 - 1.10
400.2	398.5	1.0042	Intercept	-1.820000	+/-20
200.1	197.4	1.0137			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 14, 2025	Previous Calibration:	March 12, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:00	End Time (MST):	15:55
Analyzer make:	API T201	Analyzer serial #:	808

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999993	≥0.995
396.3	397.4	0.9972	Slope	1.002698	0.90 - 1.10
203.2	202.8	1.0020	Intercept	-0.301647	+/-20
102.6	102.6	1.0000			




NO_x Calibration Plot

Date: April 14, 2025

Location: Bertha Ganter-Fort McKay







WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Mildred Lake
Calibration Date: April 17, 2025
Start time (MST): 9:49
Reason: Routine

Station number: AMS 02
Last Cal Date: March 3, 2025
End time (MST): 13:24

Calibration Standards

Cal Gas Concentration: 50.99 ppm
Cal Gas Cylinder #: EB0112903
Removed Cal Gas Conc: 50.99 ppm
Removed Gas Cyl #:
Calibrator Model: Teledyne API T700
Zero Air Gen Model: Teledyne API T701

Cal Gas Exp Date: October 9, 2032
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 1185
Serial Number: 4891

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0-1000 ppb

Serial Number: JC1404901075

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004323	1.008050	Backgd or Offset:	24.4	24.0
Calibration intercept:	-1.490497	-1.170425	Coeff or Slope:	0.791	0.763

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.7	----
As found High point	4913	78.6	803.0	830.0	0.968
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	829.3	Previous response	805.0	*% change	2.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4913	78.4	801.0	807.0	0.993
Mid point	4961	39.2	399.8	400.7	0.998
Low point	4980	19.6	199.9	199.5	1.002
As left zero	5000	0.0	0.0	0.3	----
As left span	4913	78.4	801.0	802.0	0.999
Average Correction Factor:					0.997

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 17, 2025	Previous Calibration:	March 3, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:49	End Time (MST):	13:24
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999989	≥0.995
801.0	807.0	0.9925	Slope	1.008050	0.90 - 1.10
399.8	400.7	0.9977	Intercept	-1.170425	+/-30
199.9	199.5	1.0019			



SO2 Calibration Plot

Date: April 17, 2025

Location: Mildred Lake





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Mildred Lake
Calibration Date: April 16, 2025
Start time (MST): 10:23
Reason: Routine

Station number: AMS 02
Last Cal Date: March 28, 2025
End time (MST): 15:46

Calibration Standards

Cal Gas Concentration: 4.75 ppm
Cal Gas Cylinder #: CC700774
Removed Cal Gas Conc: 4.75 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 1185
Serial Number: 4891

Analyzer Information

Analyzer make: Thermo 43iQTL
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12333331546
Converter serial #: 2023-267
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999268	1.001840	Backgd or Offset:	1.41	1.42
Calibration intercept:	0.180000	0.240000	Coeff or Slope:	0.972	0.977

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4916	84.2	80.0	79.8	1.004
As found Mid point	4958	42.1	40.0	39.8	1.007
As found Low point	4979	21.1	20.0	19.9	1.010
New cylinder response					
Baseline Corr As found:	79.7	Prev response:	80.11	*% change:	-0.5%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.996696	AF Intercept:	0.020000
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4916	84.2	80.0	80.3	0.996
Mid point	4958	42.1	40.0	40.4	0.990
Low point	4979	21.1	20.0	20.4	0.980
As left zero	5000	0.0	0.0	0.1	----
As left span	4916	84.2	80.0	81.2	0.985
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:		July 16, 2024		Ave Corr Factor	0.989
Date of last converter efficiency test:		NA			

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 16, 2025	Previous Calibration:	March 28, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:23	End Time (MST):	15:46
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331546

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999986		≥ 0.995
80.0	80.3	0.9961	Slope	1.001840		0.90 - 1.10
40.0	40.4	0.9900	Intercept	0.240000		+/-3
20.0	20.4	0.9803				



H2S Calibration Plot

Date: April 16, 2025

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Mildred Lake
 Calibration Date: April 17, 2025
 Start time (MST): 9:49
 Reason: Routine

Station number: AMS 02
 Last Cal Date: March 6, 2025
 End time (MST): 13:24

Calibration Standards

Gas Cert Reference:	EB0112903	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	503.1 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	503.1 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
Zero Air Gen model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 12227620776
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.14E-04	3.17E-04	NMHC SP Ratio:	5.59E-05
CH ₄ Retention time:	14.8	14.8	NMHC Peak Area:	158755
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4913	78.4	16.76	16.59	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.59	Prev response	16.82	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	
High point	4913	78.4	16.76	16.79	0.998
Mid point	4961	39.2	8.37	8.33	1.004
Low point	4980	19.6	4.18	4.10	1.021
As left zero	5000	0.0	0.00	0.00	----
As left span	4913	78.4	16.76	16.78	0.999
Average Correction Factor					1.008

Notes: Changed nitrogen cylinder and sample inlet filter after as founds. Adjusted span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4913	78.4	8.86	8.76	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.76	Prev response	8.89	*% change	-1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4913	78.4	8.86	8.88	0.998
Mid point	4961	39.2	4.42	4.43	0.998
Low point	4980	19.6	2.21	2.21	1.002
As left zero	5000	0.0	0.00	0.00	----
As left span	4913	78.4	8.86	8.91	0.994
Average Correction Factor					0.999

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4913	78.4	7.90	7.83	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.83	Prev response	7.93	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4913	78.4	7.90	7.91	0.999
Mid point	4961	39.2	3.94	3.90	1.011
Low point	4980	19.6	1.97	1.89	1.042
As left zero	5000	0.0	0.00	0.00	----
As left span	4913	78.4	7.90	7.87	1.005
Average Correction Factor					1.018

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.005204	1.003153
THC Cal Offset:	-0.033153	-0.046376
CH ₄ Cal Slope:	1.007405	1.003647
CH ₄ Cal Offset:	-0.033157	-0.042173
NMHC Cal Slope:	1.003755	1.002609
NMHC Cal Offset:	0.000006	-0.003803

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 17, 2025	Previous Calibration:	March 6, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:49	End Time (MST):	13:24
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620776

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999965	≥ 0.995
16.76	16.79	0.9984	Slope	1.003153	$0.90 - 1.10$
8.37	8.33	1.0044	Intercept	-0.046376	± 0.5
4.18	4.10	1.0205			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 17, 2025	Previous Calibration:	March 6, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:49	End Time (MST):	13:24
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620776

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999870	<i>≥0.995</i>
7.90	7.91	0.9992	Slope	1.003647	<i>0.90 - 1.10</i>
3.94	3.90	1.0114	Intercept	-0.042173	<i>+/-0.5</i>
1.97	1.89	1.0424			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 17, 2025	Previous Calibration:	March 6, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:49	End Time (MST):	13:24
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620776

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999999	<i>≥0.995</i>
8.86	8.88	0.9977	Slope	1.002609	<i>0.90 - 1.10</i>
4.42	4.43	0.9980	Intercept	-0.003803	<i>+/-0.5</i>
2.21	2.21	1.0018			



NMHC Calibration Plot

Date: April 17, 2025

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Mildred Lake
Calibration Date: April 28, 2025
Start time (MST): 10:15
Reason: Cylinder Change

Station number: AMS 02
Last Cal Date: April 17, 2025
End time (MST): 12:15

Calibration Standards

Gas Cert Reference:	EB0112903	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	503.1 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	503.1 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
Zero Air Gen model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 12227620776
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.17E-04	3.17E-04	NMHC SP Ratio:	5.66E-05
CH ₄ Retention time:	14.8	14.8	NMHC Peak Area:	156929
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4913	78.4	16.76	16.63	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.63	Prev response	16.77	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4913	78.4	16.76	16.70	1.004
Average Correction Factor					

Notes:

Changed the H2 cylinder after as founds.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4913	78.4	8.86	8.77	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.77	Prev response	8.88	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4913	78.4	8.86	8.79	1.009
Average Correction Factor					<input type="text"/>

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4913	78.4	7.90	7.86	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.86	Prev response	7.89	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4913	78.4	7.90	7.92	0.998
Average Correction Factor					<input type="text"/>

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.003153	
THC Cal Offset:	-0.046376	
CH ₄ Cal Slope:	1.003647	
CH ₄ Cal Offset:	-0.042173	
NMHC Cal Slope:	1.002609	
NMHC Cal Offset:	-0.003803	

Calibration Performed By: Max Farrell

NMHC Calibration Plot

Date: April 28, 2025

Location: Mildred Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
Calibration Date: April 24, 2025 Last Cal Date: March 19, 2025
Start time (MST): 5:42 End time (MST): 8:05
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.87 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC446753
Removed Cal Gas Conc: 50.87 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 3808
Zero Air Gen Model: API T701 Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1327300932
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008431	1.000030	Backgd or Offset:	28.0	27.2
Calibration intercept:	-0.643895	-0.005582	Coeff or Slope:	0.892	0.879

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4921	78.6	799.7	812.0	0.985
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	811.8	Previous response	805.8	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.5	----
High point	4921	78.6	799.7	800.0	1.000
Mid point	4961	39.3	399.8	399.6	1.001
Low point	4980	19.6	199.4	198.9	1.003
As left zero	5000	0.0	0.0	0.2	----
As left span	4921	78.6	799.7	801.5	0.998
Average Correction Factor:					1.001

Notes: No Maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 24, 2025	Previous Calibration:	March 19, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:42	End Time (MST):	8:05
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999998	≥0.995
799.7	800.0	0.9997	Slope	1.000030	0.90 - 1.10
399.8	399.6	1.0005	Intercept	-0.005582	+/-30
199.4	198.9	1.0026			



SO2 Calibration Plot

Date: April 24, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: April 28, 2025
Start time (MST): 7:28
Reason: Routine

Station number: AMS 04
Last Cal Date: March 18, 2025
End time (MST): 11:15

Calibration Standards

Cal Gas Concentration: 4.80 ppm
Cal Gas Cylinder #: DT0037528
Removed Cal Gas Conc: 4.80 ppm
Removed Gas Cyl #:
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i-LTE
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1008841400
Converter serial #: 2022-200
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007150	0.999433	Backgd or Offset:	1.97	1.92
Calibration intercept:	0.158201	0.078228	Coeff or Slope:	1.130	1.110

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4917	83.3	80.0	82.7	0.968
As found Mid point	4958	41.7	40.0	41.6	0.965
As found Low point	4979	20.8	20.0	20.8	0.965
New cylinder response					
Baseline Corr As found:	82.6	Prev response:	80.69	*% change:	2.3%
Baseline Corr 2nd AF pt:	41.5	AF Slope:	1.032872	AF Intercept:	0.158138
Baseline Corr 3rd AF pt:	20.7	AF Correlation:	0.999996	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4917	83.3	80.0	80.0	1.000
Mid point	4958	41.7	40.0	40.1	0.998
Low point	4979	20.8	20.0	20.0	0.998
As left zero	5000	0.0	0.0	0.2	----
As left span	4917	83.3	80.0	79.5	1.006
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23		Ave Corr Factor		0.999
Date of last converter efficiency test:					

Notes: Sox scrubber checked after calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	April 28, 2025	Previous Calibration:	March 18, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:28	End Time (MST):	11:15
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999999		≥0.995
80.0	80.0	0.9995	Slope	0.999433		0.90 - 1.10
40.0	40.1	0.9984	Intercept	0.078228		+/-3
20.0	20.0	0.9984				


H₂S Calibration Curve



H₂S Calibration Plot

Date: April 28, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Buffalo Viewpoint
 Calibration Date: April 24, 2025
 Start time (MST): 5:42
 Reason: Routine

Station number: AMS 04
 Last Cal Date: March 19, 2025
 End time (MST): 8:04

Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3808
Zero Air Gen model:	API T701	Serial Number:	362

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1426262594
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.58E-04	4.79E-04	NMHC SP Ratio:	9.45E-04
CH ₄ Retention time:	13.7	13.9	NMHC Peak Area:	93285
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	16.64	16.52	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.52	Prev response	16.54	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	16.64	16.57	1.004
Mid point	4961	39.3	8.32	8.24	1.010
Low point	4980	19.6	4.15	4.07	1.019
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	16.64	16.57	1.004
Average Correction Factor					1.011

Notes:

No Maintenance done. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	78.6	8.82	8.92	0.989
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.92	Prev response	8.76	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	8.82	8.77	1.005
Mid point	4961	39.3	4.41	4.38	1.007
Low point	4980	19.6	2.20	2.15	1.021
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	8.82	8.77	1.005
Average Correction Factor					1.011

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	78.6	7.82	7.61	1.027
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.61	Prev response	7.77	*% change	-2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	7.82	7.79	1.003
Mid point	4961	39.3	3.91	3.86	1.012
Low point	4980	19.6	1.95	1.92	1.016
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	7.82	7.79	1.003
Average Correction Factor					1.011

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.997100	0.996742
THC Cal Offset:	-0.049569	-0.032965
CH ₄ Cal Slope:	0.997360	0.997448
CH ₄ Cal Offset:	-0.022114	-0.016913
NMHC Cal Slope:	0.996753	0.995910
NMHC Cal Offset:	-0.027256	-0.015253

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 24, 2025	Previous Calibration:	March 19, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:42	End Time (MST):	8:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999982	≥ 0.995
16.64	16.57	1.0042	Slope	0.996742	$0.90 - 1.10$
8.32	8.24	1.0098	Intercept	-0.032965	± 0.5
4.15	4.07	1.0188			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 24, 2025	Previous Calibration:	March 19, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:42	End Time (MST):	8:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999972	<i>≥0.995</i>
7.82	7.79	1.0032	Slope	0.997448	<i>0.90 - 1.10</i>
3.91	3.86	1.0121	Intercept	-0.016913	<i>+/-0.5</i>
1.95	1.92	1.0163			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 24, 2025	Previous Calibration:	March 19, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:42	End Time (MST):	8:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999983	≥ 0.995
8.82	8.77	1.0053	Slope	0.995910	$0.90 - 1.10$
4.41	4.38	1.0074	Intercept	-0.015253	± 0.5
2.20	2.15	1.0210			



NMHC Calibration Plot

Date: April 24, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Station number: AMS 04
Calibration Date: April 11, 2025
Last Cal Date: March 14, 2025
Start time (MST): 5:55
End time (MST): 10:49
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC324979
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: APIT701
Cal Gas Expiry Date: November 3, 2032
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3808
Serial Number: 362

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.2	0.3	----	----
AF High point	4918	81.8	800.0	798.4	1.6	791.0	788.4	2.6	1.0116	1.0124
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 801.5 ppb	NO = 798.2 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.3%
Baseline Corr 1st pt	NO _x = 790.9 ppb	NO = 788.6 ppb	<u>As Found Statistics</u>	*Percent Change	NO = -1.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :	NO SI:	NO Int:
			As found NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 721

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.276	1.291	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.266	1.287	NOX bkgnd or offset:	-0.2	-0.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.6	4.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001880	1.003949
NO _x Cal Offset:	0.006724	1.107618
NO Cal Slope:	1.000957	1.000915
NO Cal Offset:	-0.954223	-0.414394
NO ₂ Cal Slope:	0.992096	1.002895
NO ₂ Cal Offset:	0.653706	1.450870

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.5	-0.1	----	----
High point	4918	81.8	800.0	798.4	1.6	803.5	798.7	4.6	0.9957	0.9996
Mid point	4959	40.9	400.0	399.2	0.8	404.3	400.1	4.2	0.9894	0.9977
Low point	4980	20.4	199.5	199.1	0.4	201.4	197.0	4.4	0.9905	1.0106
As left zero	5000	0.0	0.0	1.2	-1.2	0.8	1.2	-0.4	----	----
As left span	4918	81.8	800.0	404.7	800.0	799.7	404.7	395.0	1.0004	1.0000
Average Correction Factor									0.9919	1.0027

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	797.4	404.6	394.4	396.3	0.9953	100.5%
Mid GPT point	797.4	605.6	193.4	196.1	0.9864	101.4%
Low GPT point	797.4	700.0	99.0	102.4	0.9671	103.4%
Average Correction Factor					0.9830	101.7%

Notes: No maintenance done. Span adjusted. Due to drifting during the GPT the 2nd NO ref point used.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 14, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:55	End Time (MST):	10:49
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999989	≥0.995
800.0	803.5	0.9957	Slope	1.003949	0.90 - 1.10
400.0	404.3	0.9894	Intercept	1.107618	+/-20
199.5	201.4	0.9905			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 14, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:55	End Time (MST):	10:49
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999930	≥0.995
394.4	396.3	0.9953	Slope	1.002895	0.90 - 1.10
193.4	196.1	0.9864	Intercept	1.450870	+/-20
99.0	102.4	0.9671			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 14, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:55	End Time (MST):	10:49
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999985	≥ 0.995
798.4	798.7	0.9996	Slope	1.000915	$0.90 - 1.10$
399.2	400.1	0.9977	Intercept	-0.414394	± 20
199.1	197.0	1.0106			



NO_x Calibration Plot

Date: April 11, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: April 24, 2025
Start time (MST): 8:02
Reason: Routine

Station number: AMS 04
Last Cal Date: March 19, 2025
End time (MST): 10:15

Calibration Standards

O3 generation mode: Photometer
Calibrator Make/Model: APIP T700
ZAG Make/Model: API T701

Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: API T400
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 2961

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000914	0.995971	Backgd or Offset:	-1.2	-1.2
Calibration intercept:	0.640000	-0.020000	Coeff or Slope:	1.054	1.022

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-1.3	----
As found High point	5000	1003.6	400.0	412.0	0.968
As found Mid point					
As found Low point					
Baseline Corr As found:	413.3	Previous response	401.0	*% change	3.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.7	----
High point	5000	1005.9	400.0	397.9	1.005
Mid point	5000	827.4	200.0	199.9	1.001
Low point	5000	714.8	100.0	100.0	1.000
As left zero	5000	0.0	0.0	0.0	----
As left span	5000	1005.0	400.0	396.7	1.008
Average Correction Factor					1.002

Notes: No Maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 24, 2025	Previous Calibration:	March 19, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	8:02	End Time (MST):	10:15
Analyzer make:	API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.7	----	Correlation Coefficient	0.999984	≥0.995
400.0	397.9	1.0053	Slope	0.995971	0.90 - 1.10
200.0	199.9	1.0005	Intercept	-0.020000	+/- 5
100.0	100.0	1.0000			



O₃ Calibration Plot

Date: April 24, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
Calibration Date: April 28, 2025 Last Cal Date: March 26, 2025
Start time (MST): 6:19 End time (MST): 6:49

Analyzer Make: Teledyne API T640 S/N: 321
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-2.6	-2.7	-2.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	730.9	733.0	730.9	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.94	5.10	4.94	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39	----	39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 8.2		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 16-Jul-26
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: March 26, 2025
Date Disposable Filter Changed: March 26, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: March 26, 2025
Date RH/T Sensor Cleaned: March 26, 2025

No adjustments done.

Notes:

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS05
MANNIX
APRIL 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Mannix
Calibration Date: April 15, 2025
Start time (MST): 9:39
Reason: Routine

Station number: AMS 05
Last Cal Date: March 12, 2025
End time (MST): 13:45

Calibration Standards

Cal Gas Concentration: 50.06 ppm
Cal Gas Cylinder #: CC308040
Removed Cal Gas Conc: 50.06 ppm
Removed Gas Cyl #:
Calibrator Model: API T700
Zero Air Gen Model: API T701

Cal Gas Exp Date: October 22, 2032
Rem Gas Exp Date: October 22, 2032
Diff between cyl:
Serial Number: 5470
Serial Number: 361

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 1000 ppb

Serial Number: 1008841399

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006894	1.005853	Backgd or Offset:	10.3	10.3
Calibration intercept:	-0.937958	0.001847	Coeff or Slope:	0.950	0.950

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4920	79.9	800.0	805.8	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	805.5	Previous response	804.6	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4920	79.9	800.0	804.7	0.994
Mid point	4960	40.0	400.5	403.1	0.994
Low point	4980	20.0	200.2	200.7	0.998
As left zero	5000	0.0	0.0	0.5	----
As left span	4920	79.9	800.0	808.6	0.989
Average Correction Factor:					0.995

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 15, 2025	Previous Calibration:	March 12, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:39	End Time (MST):	13:45
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999998	≥0.995
800.0	804.7	0.9941	Slope	1.005853	0.90 - 1.10
400.5	403.1	0.9935	Intercept	0.001847	+/-30
200.2	200.7	0.9977			



SO2 Calibration Plot

Date: April 15, 2025

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Mannix
Calibration Date: April 2, 2025
Start time (MST): 9:23
Reason: Routine

Station number: AMS 05
Last Cal Date: March 5, 2025
End time (MST): 14:00

Calibration Standards

Cal Gas Concentration: 4.96 ppm
Cal Gas Cylinder #: DT0037363
Removed Cal Gas Conc: 4.96 ppm
Removed Gas Cyl #: N/A
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: N/A
Diff between cyl:
Serial Number: 5470
Serial Number: 361

Analyzer Information

Analyzer make: Thermo 43iQ
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1200326169
Converter serial #: 2022-225
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005120	0.999974	Backgd or Offset:	1.25	1.25
Calibration intercept:	0.102228	-0.017707	Coeff or Slope:	1.029	1.029

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4919	80.6	80.0	79.9	1.001
As found Mid point	4960	40.3	40.0	40.2	0.994
As found Low point	4980	20.2	20.0	19.9	1.007
New cylinder response					
Baseline Corr As found:	79.9	Prev response:	80.47	*% change:	-0.7%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	1.000117	AF Intercept:	0.002317
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999979	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4919	80.6	80.0	80.0	1.000
Mid point	4960	40.3	40.0	39.9	1.002
Low point	4980	20.2	20.0	19.9	1.007
As left zero	5000	0.0	0.0	0.2	----
As left span	4919	80.6	80.0	79.2	1.010
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.003
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	March 5, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:23	End Time (MST):	14:00
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1200326169

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999990		≥ 0.995
80.0	80.0	0.9995	Slope	0.999974		$0.90 - 1.10$
40.0	39.9	1.0019	Intercept	-0.017707		± 3
20.0	19.9	1.0069				



H₂S Calibration Plot

Date: April 2, 2025

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Mannix
 Calibration Date: April 15, 2025
 Start time (MST): 9:38
 Reason: Routine

Station number: AMS 05
 Last Cal Date: March 12, 2025
 End time (MST): 13:50

Calibration Standards

Gas Cert Reference: CC308040
 CH₄ Cal Gas Conc. 500.3 ppm
 C₃H₈ Cal Gas Conc. 199.0 ppm
 Removed Gas Cert:
 Removed CH₄ Conc. 500.3 ppm
 Removed C₃H₈ Conc. 199.0 ppm
 Diff between cyl (CH₄):
 Calibrator Model: API T700
 Zero Air Gen model: API T701

Cal Gas Expiry Date: October 22, 2032
 CH₄ Equiv Conc. 1047.6 ppm
 Removed Gas Expiry:
 CH₄ Equiv Conc. 1047.6 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 5470
 Serial Number: 361

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1193585649
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.79E-04	3.74E-04	NMHC SP Ratio:	7.45E-05	7.39E-05
CH ₄ Retention time:	15.6	15.6	NMHC Peak Area:	117387	118483
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.9	16.74	16.63	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.63	Prev response	16.64	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.9	16.74	16.65	1.005
Mid point	4960	40.0	8.38	8.20	1.022
Low point	4980	20.0	4.19	4.05	1.035
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	16.74	16.52	1.013
Average Correction Factor					1.021

Notes: Changed the inlet filter and the H₂/N₂ cylinders after as founds. Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.9	8.75	8.68	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.68	Prev response	8.73	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.9	8.75	8.70	1.006
Mid point	4960	40.0	4.38	4.32	1.014
Low point	4980	20.0	2.19	2.14	1.025
As left zero	5000	0.0	0.00	0.00	1.016
As left span	4920	79.9	8.75	8.61	1.015
Average Correction Factor					1.015

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.9	7.99	7.95	1.006
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.95	Prev response	7.91	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.9	7.99	7.96	1.005
Mid point	4960	40.0	4.00	3.88	1.032
Low point	4980	20.0	2.00	1.91	1.047
As left zero	5000	0.0	0.00	0.00	1.011
As left span	4920	79.9	7.99	7.91	1.028
Average Correction Factor					1.028

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.996286	0.996242
THC Cal Offset:	-0.038478	-0.075658
CH ₄ Cal Slope:	0.994597	0.997482
CH ₄ Cal Offset:	-0.040730	-0.053327
NMHC Cal Slope:	0.997646	0.995382
NMHC Cal Offset:	0.002452	-0.023131

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 15, 2025	Previous Calibration:	March 12, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:38	End Time (MST):	13:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999891	≥ 0.995
16.74	16.65	1.0053	Slope	0.996242	$0.90 - 1.10$
8.38	8.20	1.0222	Intercept	-0.075658	± 0.5
4.19	4.05	1.0351			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 15, 2025	Previous Calibration:	March 12, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:38	End Time (MST):	13:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999749	≥ 0.995
7.99	7.96	1.0045	Slope	0.997482	$0.90 - 1.10$
4.00	3.88	1.0315	Intercept	-0.053327	± 0.5
2.00	1.91	1.0472			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 15, 2025	Previous Calibration:	March 12, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:38	End Time (MST):	13:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999967	≥ 0.995
8.75	8.70	1.0058	Slope	0.995382	$0.90 - 1.10$
4.38	4.32	1.0139	Intercept	-0.023131	± 0.5
2.19	2.14	1.0248			



NMHC Calibration Plot

Date: April 15, 2025

Location: Mannix





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: April 11, 2025
Start time (MST): 8:40
Reason: Routine

Station number: AMS 06
Last Cal Date: March 13, 2025
End time (MST): 12:00

Calibration Standards

Cal Gas Concentration: 50.08 ppm
Cal Gas Cylinder #: CC255448
Removed Cal Gas Conc: 50.08 ppm
Removed Gas Cyl #:
Calibrator Model: API T700
Zero Air Gen Model: API T701

Cal Gas Exp Date: October 22, 2032
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 1160290013

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001008	1.001251	Backgd or Offset:	18.5	18.4
Calibration intercept:	1.638757	1.678702	Coeff or Slope:	0.928	0.920

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4920.2	79.8	799.3	807.7	0.989
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	807.9	Previous response	801.7	*% change	0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920.2	79.8	799.3	801.0	0.998
Mid point	4960.1	39.9	399.6	403.0	0.992
Low point	4980	20.0	200.3	203.7	0.983
As left zero	5000	0.0	0.0	-0.1	----
As left span	4919.7	80.3	804.3	805.1	0.999
Average Correction Factor:					0.991

Notes: Changed the inlet filter after as founds. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 13, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:40	End Time (MST):	12:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999979	≥0.995
799.3	801.0	0.9978	Slope	1.001251	0.90 - 1.10
399.6	403.0	0.9917	Intercept	1.678702	+/-30
200.3	203.7	0.9834			



SO2 Calibration Plot

Date: April 11, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: April 16, 2025
Start time (MST): 9:03
Reason: Routine

Station number: AMS 06
Last Cal Date: March 20, 2025
End time (MST): 15:00

Calibration Standards

Cal Gas Concentration: 4.760 ppm
Cal Gas Cylinder #: DT0014585
Removed Cal Gas Conc: 5.328 ppm
Removed Gas Cyl #: CC506659
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date: February 14, 2025
Diff between cyl:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1218153358
Converter serial #: 517
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001258	1.002115	Backgd or Offset:	1.94	1.99
Calibration intercept:	0.120000	0.340000	Coeff or Slope:	1.117	1.146

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4925	75.1	80.0	78.7	1.018
As found Mid point	4963	37.5	40.0	40.0	1.001
As found Low point	4981	18.8	20.0	20.1	1.002
New cylinder response					
Baseline Corr As found:	78.6	Prev response:	80.25	*% change:	-2.1%
Baseline Corr 2nd AF pt:	39.9	AF Slope:	0.981741	AF Intercept:	0.360369
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999913	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4916	84.0	80.0	80.3	0.996
Mid point	4958	42.0	40.0	40.8	0.980
Low point	4979	21.0	20.0	20.3	0.985
As left zero	5000	0.0	0.0	0.4	----
As left span	4916	84.0	80.0	79.1	1.011
SO2 Scrubber Check				0.0	
Date of last scrubber change:		December 20, 2021		Ave Corr Factor	0.987
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. Adjusted the span only.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	April 16, 2025	Previous Calibration:	March 20, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:03	End Time (MST):	15:00
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999940	≥ 0.995
80.0	80.3	0.9959	Slope	1.002115	$0.90 - 1.10$
40.0	40.8	0.9800	Intercept	0.340000	± 3
20.0	20.3	0.9848			

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: April 11, 2025
Start time (MST): 8:40
Reason: Routine

Station number: AMS 06
Last Cal Date: March 14, 2025
End time (MST): 12:00

Calibration Standards

Gas Cert Reference: AAL070632
CH₄ Cal Gas Conc. 501.4 ppm
C₃H₈ Cal Gas Conc. 199.3 ppm
Removed Gas Cert:
Removed CH₄ Conc. 501.4 ppm
Removed C₃H₈ Conc. 199.3 ppm
Diff between cyl (CH₄):
Calibrator Model: API T700
Zero Air Gen model: API T701

Cal Gas Expiry Date: September 9, 2024
CH₄ Equiv Conc. 1049.5 ppm
Removed Gas Expiry:
CH₄ Equiv Conc. 1049.5 ppm
Diff between cyl (THC):
Diff between cyl (NM):
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1118148494
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.25E-04	4.46E-04	NMHC SP Ratio:	4.12E-05
CH ₄ Retention time:	14.0	14.2	NMHC Peak Area:	212421
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.75	16.20	1.034
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.20	Prev response	16.84	*% change	-4.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.75	16.81	0.996
Mid point	4960	39.9	8.37	8.49	0.987
Low point	4980	20.0	4.20	4.32	0.972
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.75	16.82	0.996
Average Correction Factor					0.985

Notes: Instrument was installed last month. Chnaged the inlet filter after as founds. Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	8.75	8.56	1.022
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.56	Prev response	8.78	*% change	-2.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.8	8.75	8.77	0.997
Mid point	4960.1	39.9	4.37	4.43	0.987
Low point	4980	20.0	2.19	2.26	0.971
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.75	8.77	0.998
Average Correction Factor					0.985

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	8.00	7.64	1.047
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.64	Prev response	8.06	*% change	-5.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.8	8.00	8.04	0.995
Mid point	4960.1	39.9	4.00	4.06	0.986
Low point	4980	20.0	2.01	2.06	0.973
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.00	8.05	0.994
Average Correction Factor					0.984

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.002148	1.001964
THC Cal Offset:	0.059170	0.059767
CH ₄ Cal Slope:	1.003496	1.003725
CH ₄ Cal Offset:	0.028976	0.026176
NMHC Cal Slope:	1.000667	1.000863
NMHC Cal Offset:	0.031394	0.033391

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 14, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:40	End Time (MST):	12:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999943		≥ 0.995
16.75	16.81	0.9964	Slope	1.001964		0.90 - 1.10
8.37	8.49	0.9865	Intercept	0.059767		± 0.5
4.20	4.32	0.9720				





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 14, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:40	End Time (MST):	12:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999954	≥ 0.995
8.00	8.04	0.9948	Slope	1.003725	$0.90 - 1.10$
4.00	4.06	0.9855	Intercept	0.026176	± 0.5
2.01	2.06	0.9731			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 14, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:40	End Time (MST):	12:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999933	≥ 0.995
8.75	8.77	0.9974	Slope	1.000863	$0.90 - 1.10$
4.37	4.43	0.9868	Intercept	0.033391	± 0.5
2.19	2.26	0.9709			



NMHC Calibration Plot

Date: April 11, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: April 17, 2025
Start time (MST): 9:44
Reason: Cylinder Change

Station number: AMS 06
Last Cal Date: April 11, 2025
End time (MST): 11:42

Calibration Standards

Gas Cert Reference: AAL070632
CH₄ Cal Gas Conc. 501.4 ppm
C₃H₈ Cal Gas Conc. 199.3 ppm
Removed Gas Cert:
Removed CH₄ Conc. 501.4 ppm
Removed C₃H₈ Conc. 199.3 ppm
Diff between cyl (CH₄):
Calibrator Model: API T700
Zero Air Gen model: API T701

Cal Gas Expiry Date: September 9, 2024
CH₄ Equiv Conc. 1049.5 ppm
Removed Gas Expiry:
CH₄ Equiv Conc. 1049.5 ppm
Diff between cyl (THC):
Diff between cyl (NM):
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1118148494
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.46E-04	4.46E-04	NMHC SP Ratio:	4.23E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	207068
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.75	16.74	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.74	Prev response	16.84	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.02	----
As left span	4920	79.8	16.75	16.77	0.999
Average Correction Factor					

Notes:

Changed the H2 cylinder after as founds.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	8.75	8.80	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.80	Prev response	8.79	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					Limit = 0.95-1.05
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.75	8.78	0.996
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	8.00	7.94	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.94	Prev response	8.06	*% change	-1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					Limit = 0.95-1.05
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.02	----
As left span	4920	79.8	8.00	7.99	1.002
Average Correction Factor					

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001964	
THC Cal Offset:	0.059767	
CH ₄ Cal Slope:	1.003725	
CH ₄ Cal Offset:	0.026176	
NMHC Cal Slope:	1.000863	
NMHC Cal Offset:	0.033391	

Calibration Performed By: Max Farrell

NMHC Calibration Plot

Date: April 17, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Patricia McInnes
Station number: AMS 06
Calibration Date: April 3, 2025
Last Cal Date: March 4, 2025
Start time (MST): 8:39
End time (MST): 14:30
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T30YCWN
NOX Cal Gas Conc: 47.94 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 47.94 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: April 11, 2025
NO Cal Gas Conc: 46.39 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 46.39 ppm
NO gas Diff:
Serial Number: 3566
Serial Number: 4602

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.4	-0.5	-0.9	----	----
AF High point	4914	86.2	826.5	799.7	26.7	831.4	803.6	27.8	0.9924	0.9946
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 828.6 ppb	NO = 801.9 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = 0.5%			
Baseline Corr 1st pt	NO _x = 832.8 ppb	NO = 804.1 ppb	<u>As Found Statistics</u>			*Percent Change	NO = 0.3%			
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:	NO Int:			
			As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1172750022

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000894	1.002857
NO _x Cal Offset:	1.375479	1.615390
NO Cal Slope:	1.001960	1.004732
NO Cal Offset:	0.582447	0.662208
NO ₂ Cal Slope:	1.000905	1.000499
NO ₂ Cal Offset:	-0.618864	-0.488997

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.841	0.841	NO bkgnd or offset:	3.8	3.5
NOX coeff or slope:	0.990	0.990	NOX bkgnd or offset:	4.7	3.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	155.1	155.1

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.1	----	----
High point	4914	86.2	826.5	799.7	26.7	829.4	803.7	25.7	0.9964	0.9951
Mid point	4957	43.1	413.2	399.9	13.4	417.4	403.3	14.1	0.9900	0.9915
Low point	4978	21.6	207.1	200.4	6.7	210.7	202.2	8.5	0.9830	0.9912
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
As left span	4914	86.2	826.5	402.8	423.7	827.8	402.8	425.0	0.9984	1.0000
Average Correction Factor									0.9898	0.9926

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	798.7	399.2	426.2	426.2	1.0000	100.0%
Mid GPT point	798.7	593.8	231.6	231.0	1.0027	99.7%
Low GPT point	798.7	695.7	129.7	128.9	1.0064	99.4%
Average Correction Factor					1.0030	99.7%

Notes: Changed the inlet filter after as founds. Adjusted the zero only.

Calibration Performed By:

Max Farrell



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 3, 2025	Previous Calibration:	March 4, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:39	End Time (MST):	14:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999979	≥0.995
826.5	829.4	0.9964	Slope	1.002857	0.90 - 1.10
413.2	417.4	0.9900	Intercept	1.615390	+/-20
207.1	210.7	0.9830			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 3, 2025	Previous Calibration:	March 4, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:39	End Time (MST):	14:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999996	≥0.995
426.2	426.2	1.0000	Slope	1.000499	0.90 - 1.10
231.6	231.0	1.0027	Intercept	-0.488997	+/-20
129.7	128.9	1.0064			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 3, 2025	Previous Calibration:	March 4, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:39	End Time (MST):	14:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999996	≥ 0.995
799.7	803.7	0.9951	Slope	1.004732	0.90 - 1.10
399.9	403.3	0.9915	Intercept	0.662208	+/-20
200.4	202.2	0.9912			



NO_x Calibration Plot

Date: April 3, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: April 4, 2025
Start time (MST): 8:14
Reason: Routine

Station number: AMS 06
Last Cal Date: March 17, 2025
End time (MST): 11:30

Calibration Standards

O3 generation mode: Photometer
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1300156234

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001200	1.000714	Backgd or Offset:	-0.5	-0.5
Calibration intercept:	0.340000	0.200000	Coeff or Slope:	1.020	1.020

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	0.0	----
As found High point	5000	1031.0	400.0	400.9	0.998
As found Mid point					
As found Low point					
Baseline Corr As found:	400.9	Previous response	400.8	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.3	----
High point	5000	1031.0	400.0	400.5	0.999
Mid point	5000	821.4	200.0	200.4	0.998
Low point	5000	699.5	100.0	100.1	0.999
As left zero	5000	800.0	0.0	0.4	----
As left span	5000	1031.0	400.0	402.2	0.995
Average Correction Factor					0.999

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 4, 2025	Previous Calibration:	March 17, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:14	End Time (MST):	11:30
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	1.000000	≥0.995
400.0	400.5	0.9988	Slope	1.000714	0.90 - 1.10
200.0	200.4	0.9980	Intercept	0.200000	+/- 5
100.0	100.1	0.9990			



O₃ Calibration Plot

Date: April 4, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Patricia McInnes Station number: AMS 06
Calibration Date: April 4, 2025 Last Cal Date: March 17, 2025
Start time (MST): 11:56 End time (MST): 13:37

Analyzer Make: API T640 S/N: 766
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388755
Temp/RH standard: Alicat FP-25BT S/N: 388755

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2.9	3.1	2.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	732.50	734.80	732.50	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	5.11	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	86	----	38	<input checked="" type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	5.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 07-16-2026
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.8	10.8	10.8	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 4, 2025
Date Disposable Filter Changed: April 4, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: April 4, 2025
Date RH/T Sensor Cleaned: April 4, 2025

Notes: Replaced the pump. Completed both quarterly and annual maintenance. No issues.

Calibration by: Max Farrell



Wood Buffalo Environmental Association

Nt - NOX - NH3 Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	April 7, 2025	Last Cal Date:	March 10, 2025
Start time (MST):	8:30	End time (MST):	12:54
NH3 Cal Date:	April 7, 2025	Last Cal Date:	March 11, 2025
Start time (MST):	12:56	End time (MST):	15:22
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	47.94	ppm	NO Gas Cylinder #:	T30YCWN
NO Cal Gas Conc:	46.39	ppm	NO Cal Gas Expiry:	April 11, 2025
Removed NOX Conc:	47.94	ppm	Removed Cylinder #:	N/A
Removed NO Conc:	46.39	ppm	Removed cyl Expiry:	N/A
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	75.0	ppm	NH3 Gas Cylinder #:	CC709372
Removed NH3 Conc:	75.0	ppm	NH3 Cal Gas Expiry:	December 31, 2025
NH3 gas Diff:			Removed Cylinder #:	
Calibrator Model:	API T700		Removed cyl Expiry:	
ZAG make/model:	API T701		Serial Number:	3566
			Serial Number:	4602

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	215
Converter model:	API T501	Converter serial #:	147
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.20
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	25.4

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	1.002	1.027	Nt coefficient:	0.995	1.019
NOX coefficient:	0.993	1.017	NO bkgrnd:	0.2	0.2
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	-0.1	-0.1
NH3 coefficient:	0.922	0.922	Nt bkgrnd:	1.7	1.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996053	0.998888
NO _x Cal Offset:	1.276664	1.376032
NO Cal Slope:	0.991297	0.998531
NO Cal Offset:	0.864516	0.982601
NO ₂ Cal Slope:	1.003410	0.999928
NO ₂ Cal Offset:	-0.382664	0.923159
NH3 Cal Slope:	1.005770	1.004786
NH3 Cal Offset:	4.646188	9.258156
Nt Cal Slope:	1.009569	1.008514
Nt Cal Offset:	4.888735	9.809090



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

NO_x / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO _x Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>	Baseline corr NO Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.3	-0.7	----	----
As found span	4914	86.2	826.5	799.7	826.5	803.5	776.6	808.9	1.0286	1.0298
AF GPT span										
new NO cyl rp										

Baseline Corr As Fd Nt = 809.6 ppb NO_x = 803.6 ppb NO = 776.9 ppb

Previous Response Nt = 839.25 ppb NO_x = 824.5 ppb NO = 793.6 ppb

****NO_x Δ (NO to GPT response) =**

*** = > +/-2% difference initiates investigation**

***Percent Change Nt_(NO) = -3.7%**

***Percent Change NO_x = -2.6%**

***Percent Change NO = -2.2%**

*** = > +/-5% change initiates investigation**

NO_x / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.5	----	----
High point	4914	86.2	826.5	799.7	826.5	826.5	799.5	826.4	0.9999	1.0003
Mid point	4957	43.1	413.2	399.9	413.2	414.1	399.4	414.8	0.9979	1.0012
Low point	4978	21.6	207.1	200.4	207.1	210.0	203.0	210.2	0.9863	0.9873
Average Correction Factor									0.9947	0.9963

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found zero	----	----	0.0	0.2	----	----
Calibration zero	----	----	0.0	0.1	----	----
High GPT point (400 ppb O3)	797.8	397.7	426.8	427.5	0.9984	100.2%
Mid GPT point (200 ppb O3)	797.8	603.2	221.3	221.8	0.9978	100.2%
Low GPT point (100 ppb O3)	797.8	697.6	126.9	129.3	0.9816	101.9%
Average Correction Factor					0.9926	100.8%



Wood Buffalo Environmental Association NH₃ - N_T Calibration Report

NH₃ As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH ₃ concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH ₃ concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH ₃ Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.1	-0.6	----	----
AF High point	3416	84.0	1799.0	0.0	1799.0	1827.4	7.8	1819.6	0.984	0.988
AF Mid point										
AF Low point										
new NH ₃ cyl rp										
Baseline Corr As Fd	Nt = 1828.1 ppb	NH ₃ = 1820.2 ppb							*Percent Change	Nt _(NH₃) = 0.4%
Previous Response	Nt = 1821.1 ppb	NH ₃ = 1814.1 ppb							*Percent Change	NH ₃ = 0.3%

* = > +/-5% change initiates investigation

NH₃ Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH ₃ concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH ₃ concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH ₃ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.5	0.1	0.4	----	----
High point	3416	84.0	1799.0	0.0	1799.0	1818.7	7.0	1811.7	0.989	0.993
Mid point	3453	46.7	1000.3	0.0	1000.3	1023.5	4.5	1019.0	0.977	0.982
Low point	3477	23.3	499.0	0.0	499.0	522.9	2.8	520.0	0.954	0.960
Average Correction Factor									0.9736	0.9781
NH ₃ Previous Converter Efficiency =	92.2 %									
NH ₃ Current Converter Efficiency =	92.2 %									

Notes:

Changed the inlet filter after as founds. Adjusted the NOX span.

Calibration Performed By:

Max Farrell



Wood Buffalo Environmental Association


Nt Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 10, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:30	End Time (MST):	12:54
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999868	≥ 0.995
1799.0	1818.7	0.9892	Slope	1.008514	0.90 - 1.10
1000.3	1023.5	0.9773	Intercept	9.809090	+/-20
499.0	522.9	0.9542			





Wood Buffalo Environmental Association


NH₃ Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 10, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:30	End Time (MST):	12:54
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.4	----	Correlation Coefficient	0.999879	<i>≥0.995</i>
1799.0	1811.7	0.9930	Slope	1.004786	<i>0.90 - 1.10</i>
1000.3	1019.0	0.9816	Intercept	9.258156	<i>+/-20</i>
499.0	520.0	0.9596			





Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 10, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:30	End Time (MST):	12:54
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999987	<i>≥0.995</i>
826.5	826.5	0.9999	Slope	0.998888	<i>0.90 - 1.10</i>
413.2	414.1	0.9979	Intercept	1.376032	<i>+/-20</i>
207.1	210.0	0.9863			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 10, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:30	End Time (MST):	12:54
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999985	≥ 0.995
799.7	799.5	1.0003	Slope	0.998531	0.90 - 1.10
399.9	399.4	1.0012	Intercept	0.982601	+/-20
200.4	203.0	0.9873			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 10, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:30	End Time (MST):	12:54
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999969	≥0.995
426.8	427.5	0.9984	Slope	0.999928	0.90 - 1.10
221.3	221.8	0.9978	Intercept	0.923159	+/-20
126.9	129.3	0.9816			



NO_x Calibration Plot

Date: April 7, 2025

Location: Patricia McInnes



NH₃ Calibration Plot

Date: April 7, 2025

Location: Patricia McInnes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Athabasca Valley Station number: AMS07
Calibration Date: April 8, 2025 Last Cal Date: March 11, 2025
Start time (MST): 8:20 End time (MST): 11:30
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.06 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC320556
Removed Cal Gas Conc: 50.06 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3805
Zero Air Gen Model: API 701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i-LTE Serial Number: 1507864683
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997988	0.999033	Backgd or Offset:	2.70	2.73
Calibration intercept:	2.224556	2.384086	Coeff or Slope:	0.859	0.866

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	79.8	799.0	793.8	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	793.8	Previous response	799.6	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	79.8	799.0	799.1	1.000
Mid point	4960	39.9	399.5	403.8	0.989
Low point	4980	20.0	200.2	203.8	0.983
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	79.8	799.0	799.2	1.000
Average Correction Factor:					0.991

Notes: Inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 8, 2025	Previous Calibration:	March 11, 2025
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	8:20	End Time (MST):	11:30
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.2	----	Correlation Coefficient	0.999959	≥0.995
799.0	799.1	0.9999	Slope	0.999033	0.90 - 1.10
399.5	403.8	0.9893	Intercept	2.384086	+/-30
200.2	203.8	0.9825			



SO2 Calibration Plot

Date: April 8, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Athabasca Valley Station number: AMS07
Calibration Date: April 4, 2025 Last Cal Date: March 5, 2025
Start time (MST): 9:52 End time (MST): 14:16
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.25 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC504080
Removed Cal Gas Conc: 5.25 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3805
ZAG Make/Model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE Analyzer serial #: 1180540018
Converter make: CDN-101 Converter serial #: 551
Analyzer Range: 0 - 100 ppb Converter Temp: 840 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011167	1.013763	Backgd or Offset:	2.7
Calibration intercept:	-0.282221	-0.122264	Coeff or Slope:	0.908

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4925	75.5	79.3	80.1	0.987
As found Mid point	4962	37.7	39.6	40.0	0.985
As found Low point	4981	18.9	19.8	19.7	0.997
New cylinder response					
Baseline Corr As found:	80.3	Prev response:	79.87	*% change:	0.5%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	1.014050	AF Intercept:	-0.262148
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999987	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4925	75.5	79.3	80.3	0.988
Mid point	4962	37.7	39.6	40.0	0.990
Low point	4981	18.9	19.9	20.0	0.993
As left zero	5000	0.0	0.0	0.2	----
As left span	4925	75.5	79.3	79.9	0.993
SO2 Scrubber Check	4920	79.2	792.1	0.2	----
Date of last scrubber change:		21-Feb-25		Ave Corr Factor	0.990
Date of last converter efficiency test:		April 22, 2022			

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	April 4, 2025	Previous Calibration:	March 5, 2025
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:52	End Time (MST):	14:16
Analyzer make:	Thermo 43i LTE	Analyzer serial #:	1180540018

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999999	≥ 0.995
79.3	80.3	0.9877	Slope	1.013763	$0.90 - 1.10$
39.6	40.0	0.9902	Intercept	-0.122264	± 3
19.9	20.0	0.9928			

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Athabasca Valley
 Calibration Date: April 8, 2025
 Start time (MST): 8:20
 Reason: Routine

Station number: AMS 07
 Last Cal Date: March 11, 2025
 End time (MST): 11:30

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3805
Zero Air Gen model:	Teledyne API T701H	Serial Number:	198

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1331259520
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.99E-03	3.02E-03	NMHC SP Ratio:	5.57E-05
CH ₄ Retention time:	14.4	14.4	NMHC Peak Area:	161721
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.91	16.72	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.72	Prev response	16.92	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.91	16.93	0.999
Mid point	4960	39.9	8.46	8.51	0.994
Low point	4980	20.0	4.24	4.31	0.984
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.91	17.00	0.995
Average Correction Factor					0.992

Notes:

Inlet filter changed after as founds. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.00	8.90	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.90	Prev response	9.00	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.00	9.03	0.997
Mid point	4960	39.9	4.50	4.55	0.988
Low point	4980	20.0	2.26	2.31	0.977
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.00	9.07	0.993
Average Correction Factor					0.987

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	7.92	7.82	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.82	Prev response	7.91	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	7.92	7.90	1.002
Mid point	4960	39.9	3.96	3.95	1.001
Low point	4980	20.0	1.98	2.00	0.993
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	7.92	7.93	0.999
Average Correction Factor					0.999

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.998416	1.000066
THC Cal Offset:	0.030657	0.033446
CH ₄ Cal Slope:	0.998146	0.997613
CH ₄ Cal Offset:	0.012263	0.006863
NMHC Cal Slope:	0.998641	1.002237
NMHC Cal Offset:	0.018194	0.025783

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 8, 2025	Previous Calibration:	March 11, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	8:20	End Time (MST):	11:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999982	≥ 0.995
16.91	16.93	0.9990	Slope	1.000066	$0.90 - 1.10$
8.46	8.51	0.9940	Intercept	0.033446	± 0.5
4.24	4.31	0.9844			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 8, 2025	Previous Calibration:	March 11, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	8:20	End Time (MST):	11:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995
7.92	7.90	1.0017	Slope	0.997613	$0.90 - 1.10$
3.96	3.95	1.0013	Intercept	0.006863	± 0.5
1.98	2.00	0.9935			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 8, 2025	Previous Calibration:	March 11, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	8:20	End Time (MST):	11:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999962	<i>≥0.995</i>
9.00	9.03	0.9965	Slope	1.002237	<i>0.90 - 1.10</i>
4.50	4.55	0.9881	Intercept	0.025783	<i>+/-0.5</i>
2.26	2.31	0.9770			



NMHC Calibration Plot

Date: April 8, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Athabasca Valley
Calibration Date: April 17, 2025
Start time (MST): 9:30
Reason: Cylinder Change

Station number: AMS 07
Last Cal Date: April 8, 2025
End time (MST): 11:00

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3805
Zero Air Gen model:	Teledyne API T701H	Serial Number:	198

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1331259520
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.02E-03	3.02E-03	NMHC SP Ratio:	5.65E-05
CH ₄ Retention time:	14.4	14.4	NMHC Peak Area:	159257
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.91	17.12	0.988
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.12	Prev response	16.95	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.91	16.71	1.012
Average Correction Factor					

Notes:

H2 cylinder change, minimal change in response.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	9.00	9.15	0.984
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.15	Prev response	9.04	*% change	1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					Limit = 0.95-1.05
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.00	8.91	1.010
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	7.92	7.97	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.97	Prev response	7.90	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					Limit = 0.95-1.05
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	7.92	7.79	1.016
Average Correction Factor					

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000066	
THC Cal Offset:	0.033446	
CH ₄ Cal Slope:	0.997613	
CH ₄ Cal Offset:	0.006863	
NMHC Cal Slope:	1.002237	
NMHC Cal Offset:	0.025783	

Calibration Performed By: Ryan Power

NMHC Calibration Plot

Date: April 17, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Athabasca Valley
Station number: AMS 07
Calibration Date: April 1, 2025
Last Cal Date: March 10, 2025
Start time (MST): 8:44
End time (MST): 14:44
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0033919
NOX Cal Gas Conc: 60.10 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 60.10 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.90 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 59.90 ppm
NO gas Diff:
Serial Number: 3805
Serial Number: 198

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
AF High point	4933	66.8	803.0	800.3	2.7	804.4	801.3	3.2	0.9982	0.9987
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 802.9 ppb	NO = 801.2 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 0.2%
Baseline Corr 1st pt	NO _x = 804.4 ppb	NO = 801.3 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1160120024

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997767	1.001766
NO _x Cal Offset:	1.751943	2.271883
NO Cal Slope:	0.999427	1.004283
NO Cal Offset:	1.411957	1.711898
NO ₂ Cal Slope:	1.002402	0.997518
NO ₂ Cal Offset:	1.579936	0.612052

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.155	1.161	NO bkgnd or offset:	8.3	8.3
NOX coeff or slope:	1.004	1.004	NOX bkgnd or offset:	8.6	8.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	229.8	226.7

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.2	0.1	----	----
High point	4933	66.8	803.0	800.3	2.7	805.3	804.3	1.0	0.9971	0.9950
Mid point	4966	33.4	401.5	400.2	1.3	406.6	405.5	1.1	0.9875	0.9869
Low point	4983	16.7	200.7	200.1	0.7	204.6	203.4	1.2	0.9812	0.9837
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	0.0	----	----
As left span	4933	66.8	803.0	395.4	407.6	806.7	395.4	411.4	0.9954	1.0000
Average Correction Factor									0.9886	0.9885

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	800.0	396.5	406.2	405.9	1.0007	99.9%
Mid GPT point	800.0	599.3	203.4	202.6	1.0038	99.6%
Low GPT point	800.0	701.3	101.4	103.0	0.9842	101.6%
Average Correction Factor					0.9962	100.4%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 1, 2025	Previous Calibration:	March 10, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	8:44	End Time (MST):	14:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999967	≥0.995
803.0	805.3	0.9971	Slope	1.001766	0.90 - 1.10
401.5	406.6	0.9875	Intercept	2.271883	+/-20
200.7	204.6	0.9812			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 1, 2025	Previous Calibration:	March 10, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	8:44	End Time (MST):	14:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999970	≥0.995
406.2	405.9	1.0007	Slope	0.997518	0.90 - 1.10
203.4	202.6	1.0038	Intercept	0.612052	+/-20
101.4	103.0	0.9842			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 1, 2025	Previous Calibration:	March 10, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	8:44	End Time (MST):	14:44
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.2	----	Correlation Coefficient	0.999978	≥ 0.995
800.3	804.3	0.9950	Slope	1.004283	$0.90 - 1.10$
400.2	405.5	0.9869	Intercept	1.711898	± 20
200.1	203.4	0.9837			



NO_x Calibration Plot

Date: April 1, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Athabasca Valley
Calibration Date: April 7, 2025
Start time (MST): 9:05
Reason: Routine

Station number: AMS07
Last Cal Date: March 17, 2025
End time (MST): 12:40

Calibration Standards

O3 generation mode: Photometer
Calibrator Make/Model: T700
ZAG Make/Model: T701H

Serial Number: 3805
Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1152220023

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001457	0.999600	Backgd or Offset:	-1.2	-1.1
Calibration intercept:	0.820000	1.020000	Coeff or Slope:	1.605	1.556

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	0.1	----
As found High point	5000	1704.0	400.0	412.7	0.969
As found Mid point					
As found Low point					
Baseline Corr As found:	412.6	Previous response	401.4	*% change	2.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.1	----
High point	5000	1706.0	400.0	400.4	0.999
Mid point	5000	1175.0	200.0	201.4	0.993
Low point	5000	926.0	100.0	101.9	0.981
As left zero	5000	1652.9	0.0	0.1	----
As left span	5000	1582.6	400.0	402.8	0.993
Average Correction Factor					0.991

Notes: Inlet filter changed after as founds. Adjusted span.

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association

O₃ Calibration Summary


Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 17, 2025
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:05	End Time (MST):	12:40
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999976	≥ 0.995
400.0	400.4	0.9990	Slope	0.999600	$0.90 - 1.10$
200.0	201.4	0.9930	Intercept	1.020000	± 5
100.0	101.9	0.9814			


O₃ Calibration Curve



O₃ Calibration Plot

Date: April 7, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Athabasca Valley Station number: AMS 07
Calibration Date: April 1, 2025 Last Cal Date: March 24, 2025
Start time (MST): 8:43 End time (MST): 9:45

Analyzer Make: API T640 S/N: 645
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-2.4	-2.9		<input type="checkbox"/>	+/- 2 °C
P (mmHg)	729.6	729.1		<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.75		<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	38	----		<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 12.2	PM w/ HEPA: 10.7			<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 6, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 27, 2025
Date Disposable Filter Changed: February 27, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 8, 2024
Date RH/T Sensor Cleaned: July 8, 2024

Notes: Removal calibration. Flow, temp and pressure checked. Leak check failed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Athabasca Valley Station number: AMS 07
Calibration Date: April 1, 2025 Last Cal Date: N/A
Start time (MST): 12:05 End time (MST): 13:54

Analyzer Make: API T640 S/N: 2235
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	3.5	3.5	3.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	729.0	727.9	729.0	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.95	4.78	5.05	<input checked="" type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	38	----	38	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 5.4		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 6, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.9		10.9	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 1, 2025
Date Disposable Filter Changed: April 1, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 8, 2024
Date RH/T Sensor Cleaned: July 8, 2024

Notes: Install calibration completed. Flow adjusted. Nothing else to note.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name: Athabasca Valley Station number: AMS 07
Calibration Date: April 25, 2025 Last Cal Date: March 24, 2025
Start time (MST): 10:42 End time (MST): 14:07
Reason: Routine

Calibration Standards

Cal Gas Concentration: 2,953 ppm Cal Gas Exp Date: September 30, 2029
Cal Gas Cylinder #: T1TWKRN
Removed Cal Gas Conc: 2,953 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T750 Serial Number: 282
ZAG Make/Model: Teledyne API 751H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 48i-TLE Analyzer serial #: 1408761381
Analyzer Range: 0 - 50 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997952	1.002518	Backgd or Offset:	5.450
Calibration intercept:	0.154017	0.164069	Coeff or Slope:	1.073

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4932	67.8	40.0	40.2	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	40.05	Prev response:	40.12	*% change:	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4932	67.8	40.0	40.2	0.996
Mid point	4966	33.9	20.0	20.4	0.981
Low point	4983	16.9	10.0	10.3	0.971
As left zero	5000	0.0	0.0	0.1	----
As left span	4932	67.8	40.0	40.0	1.001
Average Correction Factor					0.983

Notes: No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


CO Calibration Summary

Station Information

Calibration Date:	April 25, 2025	Previous Calibration:	March 24, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:42	End Time (MST):	14:07
Analyzer make:	Thermo 48i-TLE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999912	≥ 0.995
40.0	40.2	0.9961	Slope	1.002518	$0.90 - 1.10$
20.0	20.4	0.9815	Intercept	0.164069	± 1.5
10.0	10.3	0.9709			



CO Calibration Plot

Date: April 25, 2025

Location: Athabasca Valley





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan Station number: AMS08
Calibration Date: April 9, 2025 Last Cal Date: March 17, 2025
Start time (MST): 12:23 End time (MST): 14:38
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.84 ppm Cal Gas Exp Date: January 6, 2030
Cal Gas Cylinder #: CC196697
Removed Cal Gas Conc: 49.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3810
Zero Air Gen Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43i-TLE Serial Number: 1236656116
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001018	1.005001	Backgd or Offset:	2.0
Calibration intercept:	0.635300	0.414722	Coeff or Slope:	1.048

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4920	80.3	800.4	804.4	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	804.3	Previous response	801.8	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4920	80.3	800.4	804.7	0.995
Mid point	4960	40.2	400.7	403.3	0.994
Low point	4980	20.1	200.4	201.8	0.993
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	80.3	800.4	807.0	0.992
Average Correction Factor:					0.994

Notes: Changed out inlet filter after as founds.

Calibration Performed By: Sabian V, Jeremy C, Morgan V,



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	March 17, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS08
Start Time (MST):	12:23	End Time (MST):	14:38
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	1.000000	≥0.995
800.4	804.7	0.9946	Slope	1.005001	0.90 - 1.10
400.7	403.3	0.9935	Intercept	0.414722	+/-30
200.4	201.8	0.9928			



SO2 Calibration Plot

Date: April 9, 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort Chipewyan
Calibration Date: April 11, 2025
Start time (MST): 7:38
Reason: Routine

Station number: AMS 08
Last Cal Date: March 10, 2025
End time (MST): 11:26

Calibration Standards

Cal Gas Concentration: 4.84 ppm
Cal Gas Cylinder #: SA7549
Removed Cal Gas Conc: 4.84 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3810
Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43iQ-TL
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1203169744
Converter serial #: 580
Converter Temp: 850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005392	0.982087	Backgd or Offset:	2.0	2.0
Calibration intercept:	-0.058192	-0.317646	Coeff or Slope:	0.779	0.779

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4917	82.6	80.0	76.4	1.044
As found Mid point	4959	41.3	40.0	38.0	1.046
As found Low point	4979	20.7	20.0	18.6	1.066
New cylinder response					
Baseline Corr As found:	76.6	Prev response:	80.34	*% change:	-4.9%
Baseline Corr 2nd AF pt:	38.2	AF Slope:	0.959500	AF Intercept:	-0.377228
Baseline Corr 3rd AF pt:	18.8	AF Correlation:	0.999970	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4917	82.6	80.0	78.4	1.020
Mid point	4959	41.3	40.0	38.7	1.033
Low point	4979	20.7	20.0	19.1	1.049
As left zero	5000	0.0	0.0	-0.1	----
As left span	4917	82.6	80.0	79.0	1.012
SO2 Scrubber Check	4919.7	80.3	803.0	-0.1	----
Date of last scrubber change:		March 7, 2022		Ave Corr Factor	1.034
Date of last converter efficiency test:		March 15, 2022		103.4% efficiency	

Notes: Changed inlet filter after as founds. Scrubber check passed no issues. No Adjustments made.

Calibration Performed By: Sabian Voyageur Jeremy Cardinal



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 10, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	7:38	End Time (MST):	11:26
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999922		≥ 0.995
80.0	78.4	1.0199	Slope	0.982087		$0.90 - 1.10$
40.0	38.7	1.0330	Intercept	-0.317646		± 3
20.0	19.1	1.0492				


TRS Calibration Curve



TRS Calibration Plot

Date: April 11, 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort Chipewyan
Calibration Date: April 15, 2025
Start time (MST): 9:16
Reason: Maintenance

Station number: AMS 08
Last Cal Date: April 11, 2025
End time (MST): 11:58

Calibration Standards

Cal Gas Concentration: 4.84 ppm
Cal Gas Cylinder #: SA7549
Removed Cal Gas Conc: 4.84 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3810
Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43iQ-TL
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1203169744
Converter serial #: 580
Converter Temp: 850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.982087	1.015539	Backgd or Offset:	2.0	2.2
Calibration intercept:	-0.317646	-0.538270	Coeff or Slope:	0.779	0.853

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4917	82.6	80.0	72.8	1.094
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	73.1	Prev response:	78.21	*% change:	-7.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	4917	82.6	80.0	80.8	0.990
Mid point	4959	41.3	40.0	39.9	1.002
Low point	4979	20.7	20.0	19.6	1.022
As left zero	5000	0.0	0.0	-0.2	----
As left span	4917	82.6	80.0	81.8	0.978
SO2 Scrubber Check	4919.7	80.3	803.0		----
Date of last scrubber change:		March 7, 2022		Ave Corr Factor	1.005
Date of last converter efficiency test:		March 15, 2022		103.4% efficiency	

Notes: Redid calibration span was out of range. Adjusted span.

Calibration Performed By: Morgan Voyageur



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	April 15, 2025	Previous Calibration:	April 11, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	9:16	End Time (MST):	11:58
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1203169744

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999960		≥ 0.995
80.0	80.8	0.9896	Slope	1.015539		$0.90 - 1.10$
40.0	39.9	1.0019	Intercept	-0.538270		± 3
20.0	19.6	1.0224				

TRS Calibration Curve







Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan
Station number: AMS 08
Calibration Date: April 8, 2025
Last Cal Date: March 20, 2025
Start time (MST): 9:11
End time (MST): 12:09
Reason: Removal

Calibration Standards

NO Gas Cylinder #: DT0046831
NOX Cal Gas Conc: 60.20 ppm
Removed Cylinder #: DT0046831
Removed Gas NOX Conc: 60.20 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: January 9, 2032
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 135

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.2	-0.2	----	----
AF High point	4933	66.7	803.1	800.4	2.7	781.2	778.1	3.1	1.0273	1.0285
AF Mid point	4967	33.3	400.9	399.6	1.3	392.5	389.4	3.1	1.0200	1.0256
AF Low point	4983	16.7	201.1	200.4	0.7	198.2	195.1	3.1	1.0117	1.0261

New cyl resp

Previous Response	NO _x = 801.8 ppb	NO = 799.3 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = -2.6%
Baseline Corr 1st pt	NO _x = 781.7 ppb	NO = 778.3 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -2.7%
Baseline Corr 2nd pt	NO _x = 393.0 ppb	NO = 389.6 ppb	As found	NO _x r ² : 0.999975	Nx SI: 0.972490	Nx Int: 1.231
Baseline Corr 3rd pt	NO _x = 198.7 ppb	NO = 195.3 ppb	As found	NO r ² : 0.999998	NO SI: 0.972293	NO Int: 0.194
			As found	NO ₂ r ² : 0.999986	NO2 SI: 0.961115	NO ₂ Int: 0.164

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	-0.2	----	----
As found high GPT point	773.4	377.0	399.1	383.3	1.0411	96.0%
As found mid GPT point	773.4	575.3	200.8	194.0	1.0349	96.6%
As found low GPT point	773.4	669.7	106.4	102.3	1.0398	96.2%



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12400232072

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.727	
NOX coeff or slope:	0.996	
NO2 coeff or slope:	1.000	

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	0.8	
NOX bkgnd or offset:	1.1	
Reaction cell Press:	113.7	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996890	
NO _x Cal Offset:	1.194493	
NO Cal Slope:	0.997843	
NO Cal Offset:	0.574281	
NO ₂ Cal Slope:	0.971567	
NO ₂ Cal Offset:	1.371312	

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero										
High point										
Mid point										
Low point										
As left zero										
As left span										

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero						
High GPT point						
Mid GPT point						
Low GPT point						

Average Correction Factor


Notes: No adjustments or maintenance performed.

Calibration Performed By: Sabian Voyageur, Jerney Cardinal, Morgan Voyageur.

NO_x Calibration Plot

Date: April 8, 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan
Station number: AMS 08
Calibration Date: April 9, 2025
Last Cal Date: N/A
Start time (MST): 7:53
End time (MST): 12:07
Reason: Install

Calibration Standards

NO Gas Cylinder #: DT0046831
NOX Cal Gas Conc: 60.20 ppm
Removed Cylinder #: DT0046831
Removed Gas NOX Conc: 60.20 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: January 9, 2032
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 135

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
-----------	------------------------------	--------------------------------	---	--	---	--	---	--	--	---

As found zero
AF High point
AF Mid point
AF Low point
New cyl resp

Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>		*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	---	--	--	---	--	--

As Found GPT zero
As found high GPT point
As found mid GPT point
As found low GPT point



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12124313137

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:		1.023	NO bkgnd or offset:		1.5
NOX coeff or slope:		0.993	NOX bkgnd or offset:		1.5
NO2 coeff or slope:		1.000	Reaction cell Press:		113.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:		1.000364
NO _x Cal Offset:		-0.525927
NO Cal Slope:		1.002985
NO Cal Offset:		-2.025916
NO ₂ Cal Slope:		1.001492
NO ₂ Cal Offset:		1.255111

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.9	0.0	----	----
High point	4933	66.7	803.1	800.4	2.7	802.9	801.5	1.4	1.0003	0.9987
Mid point	4967	33.3	400.9	399.6	1.3	400.2	397.8	2.4	1.0018	1.0045
Low point	4983	16.7	201.1	200.4	0.7	201.2	198.1	3.0	0.9993	1.0116
As left zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.7	0.1	----	----
As left span	4933	66.7	803.1	393.4	409.7	804.8	393.4	411.3	0.9979	1.0000
Average Correction Factor									1.0005	1.0049

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	798.8	389.6	411.9	412.9	0.9975	100.3%
Mid GPT point	798.8	598.6	202.9	205.7	0.9862	101.4%
Low GPT point	798.8	695.7	105.8	108.0	0.9793	102.1%
Average Correction Factor					0.9877	101.3%

Notes: Adjustments made to span. Changed filter before the start of the calibration.

Calibration Performed By: Sabian V, Jermy C, Morgan V.



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	N/A
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	7:53	End Time (MST):	12:07
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12124313137

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.8	----	Correlation Coefficient	0.999999	≥ 0.995
803.1	802.9	1.0003	Slope	1.000364	0.90 - 1.10
400.9	400.2	1.0018	Intercept	-0.525927	+/-20
201.1	201.2	0.9993			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	N/A
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	7:53	End Time (MST):	12:07
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12124313137

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999951	≥ 0.995
411.9	412.9	0.9975	Slope	1.001492	$0.90 - 1.10$
202.9	205.7	0.9862	Intercept	1.255111	± 20
105.8	108.0	0.9793			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	N/A
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	7:53	End Time (MST):	12:07
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12124313137

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.9	----	Correlation Coefficient	0.999990	≥ 0.995
800.4	801.5	0.9987	Slope	1.002985	$0.90 - 1.10$
399.6	397.8	1.0045	Intercept	-2.025916	± 20
200.4	198.1	1.0116			



NO_x Calibration Plot

Date: April 9, 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Fort Chipewyan
Calibration Date: April 7, 2025
Start time (MST): 12:11
Reason: Routine

Station number: AMS 08
Last Cal Date: March 11, 2025
End time (MST): 15:22

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Serial Number: 3810
Serial Number: 135

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1152220026

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993514	0.986800	Backgd or Offset:	-0.3	-0.3
Calibration intercept:	0.060000	0.260000	Coeff or Slope:	1.005	1.005

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	5000	968.7	400.0	395.4	1.012
As found Mid point					
As found Low point					
Baseline Corr As found:	395.1	Previous response	397.5	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
					<i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	5000	968.7	400.0	395.1	1.012
Mid point	5000	820.5	200.0	197.4	1.013
Low point	5000	720.0	100.0	98.9	1.011
As left zero	5000	0.0	0.0	0.2	----
As left span	5000	968.7	400.0	396.1	1.010
Average Correction Factor					1.012

Notes: Changed Filter after asfound. No adjustments needed.

Calibration Performed By: Sabian Voyageur, JermeY Cardinal, Morgan Voyageur.



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 11, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	12:11	End Time (MST):	15:22
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220026

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999999	≥0.995
400.0	395.1	1.0124	Slope	0.986800	0.90 - 1.10
200.0	197.4	1.0132	Intercept	0.260000	+/- 5
100.0	98.9	1.0111			



O₃ Calibration Plot

Date: April 7, 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
Calibration Date: April 25, 2025 Last Cal Date: March 11, 2025
Start time (MST): 8:10 End time (MST): 8:43

Analyzer Make: Teledyne API T640 S/N: 319
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 14719
Temp/RH standard: Alicat FP-25BT S/N: 14719

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	9.30	8.75	9.30	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	736.30	738.5	736.30	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	4.74	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	49%		49%	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 5.10		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: 10-Jun-24
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	NA	10.80	10.80	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: March 11, 2025
Date Disposable Filter Changed: September 18, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 29, 2024
Date RH/T Sensor Cleaned: August 29, 2024

Notes:

No adjustment made

Calibration by: Jeremy Cardinal, Morgan Voyageur



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name: Fort Chipewyan
Calibration Date: April 10, 2025
Start time (MST): 11:57
Reason: Routine

Station number: AMS 08
Last Cal Date: March 17, 2025
End time (MST): 14:37

Calibration Standards

Cal Gas Concentration: 3,030 ppm
Cal Gas Cylinder #: ALM014846
Removed Cal Gas Conc: 3,030 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Exp Date: December 1, 2028
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3810
Serial Number: 135

Analyzer Information

Analyzer make: Teledyne API T300
Analyzer Range: 0 - 50 ppm

Analyzer serial #: 3505

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998216	0.998612	Backgd or Offset:	-0.016	-0.016
Calibration intercept:	0.172906	0.100910	Coeff or Slope:	1.003	1.003

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4934	66.7	40.4	40.5	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	40.42	Prev response:	40.52	*% change:	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4934	66.7	40.4	40.4	1.000
Mid point	4966.7	33.3	20.2	20.3	0.992
Low point	4983.3	16.7	10.1	10.2	0.989
As left zero	5000	0.0	0.0	0.2	----
As left span	2960	40.0	40.4	40.2	1.004
Average Correction Factor					0.994

Notes: Changed inlet filter after as found. No Adjustments made

Calibration Performed By: Sabian V, Jeremy C,



Wood Buffalo Environmental Association

CO Calibration Summary


Station Information

Calibration Date:	April 10, 2025	Previous Calibration:	March 17, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	11:57	End Time (MST):	14:37
Analyzer make:	Teledyne API T300	Analyzer serial #:	3505

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999984	≥ 0.995
40.4	40.4	1.0001	Slope	0.998612	$0.90 - 1.10$
20.2	20.3	0.9921	Intercept	0.100910	± 1.5
10.1	10.2	0.9893			


CO Calibration Curve



CO Calibration Plot

Date: April 10, 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

CO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
Calibration Date: April 10, 2025 Last Cal Date: March 19, 2025
Start time (MST): 8:19 End time (MST): 11:37
Reason: Routine

Calibration Standards

Cal Gas Concentration: 60,220 ppm Cal Gas Exp Date: December 1, 2028
Cal Gas Cylinder #: ALM014846
Removed Cal Gas Conc: 60,220 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3810
N2 Gen Make/Model: Peak Scientific Serial Number: 135

Analyzer Information

Analyzer make: Teledyne API T360 Analyzer serial #: 289
Analyzer Range: 0 - 2,000 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998354	1.004061	Backgd or Offset:	-0.014	-0.014
Calibration intercept:	-4.560000	-5.320000	Coeff or Slope:	1.033	1.033

CO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	3000	0.0	0.0	0.1	----
As found High Point	2920	80.0	1605.9	1608.0	0.999
As found Mid Point					
As found Low Point					
New cylinder response					
Baseline Corr As found:	1607.9	Prev response:	1598.7	*% change:	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	3000	0.0	0.0	0.4	----
High point	2920	80.0	1605.9	1613.7	0.995
Mid point	2960	40.0	802.9	786.7	1.021
Low point	2980	20.0	401.5	399.6	1.005
As left zero	3000	0.0	0.0	0.2	----
As left span	2960	40.0	802.9	787.8	1.019

Average Correction Factor **1.007**

Notes: Changed inlet filter after as found, no adjustments made

Calibration Performed By: Sabian V, Jeremy C



Wood Buffalo Environmental Association


CO₂ Calibration Summary

Station Information

Calibration Date	April 10, 2025	Previous Calibration	March 19, 2025
Station Name	Fort Chipewyan	Station Number	AMS 08
Start Time (MST)	8:19	End Time (MST)	11:37
Analyzer make	Teledyne API T360	Analyzer serial #	289

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999802		≥0.995
1605.9	1613.7	0.9951	Slope	1.004061		0.90 - 1.10
802.9	786.7	1.0206	Intercept	-5.3		+/-20
401.5	399.6	1.0047				






CO₂ Calibration Plot

Date: April 10, 2025

Location: Fort Chipewyan





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Barge Landing
Calibration Date: April 7, 2025
Start time (MST): 9:09
Reason: Routine

Station number: AMS 09
Last Cal Date: March 6, 2025
End time (MST): 12:05

Calibration Standards

Cal Gas Concentration: 50.56 ppm
Cal Gas Cylinder #: CC705748
Removed Cal Gas Conc: 50.56 ppm
Removed Gas Cyl #: NA
Calibrator Model: API T700
Zero Air Gen Model: APIT701

Cal Gas Exp Date: October 9, 2032
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3812
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 1118148498

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999602	1.006817	Backgd or Offset:	11.4	11.5
Calibration intercept:	-0.799117	-0.598200	Coeff or Slope:	0.998	1.010

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4921	79.1	799.8	792.9	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	792.8	Previous response	798.7	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4921	79.1	799.8	805.0	0.994
Mid point	4961	39.5	399.4	401.4	0.995
Low point	4980	19.8	200.2	200.0	1.001
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	79.1	799.8	806.4	0.992
Average Correction Factor:					0.997

Notes:

Inlet filter changed after as founds. Adjusted span only.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 6, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:09	End Time (MST):	12:05
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999995	≥0.995
799.8	805.0	0.9936	Slope	1.006817	0.90 - 1.10
399.4	401.4	0.9950	Intercept	-0.598200	+/-30
200.2	200.0	1.0011			



SO2 Calibration Plot

Date: April 7, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Barge Landing
Calibration Date: April 1, 2025
Start time (MST): 8:54
Reason: Routine

Station number: AMS 09
Last Cal Date: March 24, 2025
End time (MST): 12:56

Calibration Standards

Cal Gas Concentration: 5.17 ppm
Cal Gas Cylinder #: CC511415
Removed Cal Gas Conc: 5.17 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: August 22, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3812
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12426335708
Converter serial #: 519
Converter Temp: 830 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999979	0.999833	Backgd or Offset:	2.130
Calibration intercept:	-0.280625	0.099474	Coeff or Slope:	1.069

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.8	----
As found High point	4923	77.4	80.0	72.3	1.095
As found Mid point	4961	38.7	40.0	35.8	1.094
As found Low point	4981	19.3	20.0	17.6	1.085
New cylinder response					
Baseline Corr As found:	73.1	Prev response:	79.76	*% change:	-9.1%
Baseline Corr 2nd AF pt:	36.6	AF Slope:	0.912612	AF Intercept:	-0.722238
Baseline Corr 3rd AF pt:	18.4	AF Correlation:	0.999994	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4923	77.4	80.0	80.1	0.999
Mid point	4961	38.7	40.0	40.1	0.998
Low point	4981	19.3	20.0	20.2	0.988
As left zero	5000	0.0	0.0	0.1	----
As left span	4923	77.4	80.0	80.5	0.994
SO2 Scrubber Check	4920	80.2	802.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	0.995
Date of last converter efficiency test:					

Notes: Low span response probably due to drift. Diagnostics seems fine. Adjusted zero and span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	April 1, 2025	Previous Calibration:	March 24, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	8:54	End Time (MST):	12:56
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	12426335708

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999991		≥ 0.995
80.0	80.1	0.9993	Slope	0.999833		$0.90 - 1.10$
40.0	40.1	0.9982	Intercept	0.099474		± 3
20.0	20.2	0.9881				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Barge Landing
 Calibration Date: April 5, 2025
 Start time (MST): 12:05
 Reason: Install

Station number: AMS 09
 Last Cal Date: March 6, 2025
 End time (MST): 14:50

Calibration Standards

Gas Cert Reference:	CC705748	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	505.6 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	204.8 ppm		
Removed Gas Cert:	CC151285	Removed Gas Expiry:	January 5, 2025
Removed CH ₄ Conc.	505.6 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	204.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
Zero Air Gen model:	APIT701	Serial Number:	5613

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1193585650
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	N/A	2.31E-04	NMHC SP Ratio:	N/A	4.90E+05
CH ₄ Retention time:	N/A	15.2	NMHC Peak Area:	N/A	181998
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	16.91	16.96	0.997
Mid point	4961	39.5	8.44	8.46	0.998
Low point	4980	19.8	4.23	4.22	1.004
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	16.91	17.05	0.991
Average Correction Factor					1.000

Notes:

Install calibration. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	8.91	8.96	0.995
Mid point	4961	39.5	4.45	4.48	0.994
Low point	4980	19.8	2.23	2.24	0.994
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	8.91	9.03	0.987
Average Correction Factor					0.994

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	8.00	8.01	0.998
Mid point	4961	39.5	3.99	3.98	1.004
Low point	4980	19.8	2.00	1.97	1.015
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	8.00	8.02	0.997
Average Correction Factor					1.006

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	N/A	1.003755
THC Cal Offset:	N/A	-0.014363
CH ₄ Cal Slope:	N/A	1.002645
CH ₄ Cal Offset:	N/A	-0.017384
NMHC Cal Slope:	N/A	1.005407
NMHC Cal Offset:	N/A	0.001219

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 5, 2025	Previous Calibration:	March 6, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	12:05	End Time (MST):	14:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995
16.91	16.96	0.9968	Slope	1.003755	$0.90 - 1.10$
8.44	8.46	0.9983	Intercept	-0.014363	± 0.5
4.23	4.22	1.0037			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 5, 2025	Previous Calibration:	March 6, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	12:05	End Time (MST):	14:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999979	≥ 0.995
8.00	8.01	0.9984	Slope	1.002645	$0.90 - 1.10$
3.99	3.98	1.0040	Intercept	-0.017384	± 0.5
2.00	1.97	1.0148			






Wood Buffalo Environmental Association NMHC Calibration Summary

Station Information

Calibration Date:	April 5, 2025	Previous Calibration:	March 6, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	12:05	End Time (MST):	14:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	1.000000	≥ 0.995
8.91	8.96	0.9946	Slope	1.005407	$0.90 - 1.10$
4.45	4.48	0.9939	Intercept	0.001219	± 0.5
2.23	2.24	0.9939			



NMHC Calibration Plot

Date: April 5, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Barge Landing
 Calibration Date: April 21, 2025
 Start time (MST): 6:25
 Reason: Routine

Station number: AMS 09
 Last Cal Date: April 5, 2025
 End time (MST): 8:40

Calibration Standards

Gas Cert Reference:	CC705748	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	505.6 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	204.8 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	505.6 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	204.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
Zero Air Gen model:	APIT701	Serial Number:	5613

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1193585650
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	N/A	2.31E-04	NMHC SP Ratio:	N/A
CH ₄ Retention time:	N/A	15.2	NMHC Peak Area:	N/A
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	16.91	16.05	1.053
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.05	Prev response	16.96	*% change	-5.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	16.91	17.01	0.994
Mid point	4961	39.5	8.44	8.41	1.004
Low point	4980	19.8	4.23	4.15	1.020
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	16.91	16.85	1.003
Average Correction Factor					1.006

Notes:

CH₄ channel baseline dipping. RT has moved. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	8.91	8.12	1.098
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.12	Prev response	8.96	*% change	-10.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	8.91	9.00	0.991
Mid point	4961	39.5	4.45	4.45	0.999
Low point	4980	19.8	2.23	2.19	1.018
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	8.91	8.92	0.999
Average Correction Factor					1.002

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	8.00	7.94	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.94	Prev response	8.00	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	8.00	8.02	0.998
Mid point	4961	39.5	3.99	3.96	1.010
Low point	4980	19.8	2.00	1.96	1.023
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	8.00	7.94	1.008
Average Correction Factor					1.010

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.003755	1.007610
THC Cal Offset:	-0.014363	-0.060372
CH ₄ Cal Slope:	1.002645	1.003661
CH ₄ Cal Offset:	-0.017384	-0.029190
NMHC Cal Slope:	1.005407	1.011373
NMHC Cal Offset:	0.001219	-0.031782

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 21, 2025	Previous Calibration:	April 5, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	6:25	End Time (MST):	8:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999942	<i>≥0.995</i>
16.91	17.01	0.9941	Slope	1.007610	<i>0.90 - 1.10</i>
8.44	8.41	1.0040	Intercept	-0.060372	<i>+/-0.5</i>
4.23	4.15	1.0202			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 21, 2025	Previous Calibration:	April 5, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	6:25	End Time (MST):	8:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999936	<i>≥0.995</i>
8.00	8.02	0.9978	Slope	1.003661	<i>0.90 - 1.10</i>
3.99	3.96	1.0096	Intercept	-0.029190	<i>+/-0.5</i>
2.00	1.96	1.0231			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 21, 2025	Previous Calibration:	April 5, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	6:25	End Time (MST):	8:40
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			Limits
0.00	0.00	----	Correlation Coefficient	0.999943		≥ 0.995
8.91	9.00	0.9905	Slope	1.011373		0.90 - 1.10
4.45	4.45	0.9993	Intercept	-0.031782		± 0.5
2.23	2.19	1.0175				



NMHC Calibration Plot

Date: April 21, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Barge Landing
 Calibration Date: April 29, 2025
 Start time (MST): 9:08
 Reason: Maintenance

Station number: AMS 09
 Last Cal Date: April 21, 2025
 End time (MST): 12:22

Calibration Standards

Gas Cert Reference:	CC705748	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	505.6 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	204.8 ppm		
Removed Gas Cert:	CC151285	Removed Gas Expiry:	January 5, 2025
Removed CH ₄ Conc.	505.6 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	204.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
Zero Air Gen model:	APIT701	Serial Number:	5613

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1193585650
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.31E-04	2.28E-04	NMHC SP Ratio:	4.90E+05
CH ₄ Retention time:	15.2	14.2	NMHC Peak Area:	181998
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	16.91	17.38	0.973
As found Mid point	4961	39.5	8.44	8.65	0.976
As found Low point	4980	19.8	4.23	4.31	0.981
New cylinder response					
Baseline Corr AF:	17.38	Prev response	16.98	*% change	2.3%
Baseline Corr 2nd AF:	8.65	AF Slope:	1.028459	AF Intercept:	-0.020568
Baseline Corr 3rd AF:	4.31	AF Correlation:	0.999994	* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	16.91	16.96	0.997
Mid point	4961	39.5	8.44	8.40	1.005
Low point	4980	19.8	4.23	4.17	1.014
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	16.91	17.00	0.995
Average Correction Factor					1.005

Notes:

Few dips. Carrier pressure increased. Span adjusted. RT moved.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.1	8.91	9.39	0.949
As found Mid point	4961	39.5	4.45	4.68	0.951
As found Low point	4980	19.8	2.23	2.34	0.955
New cylinder response					
Baseline Corr AF:	9.39	Prev response	8.98	*% change	4.4%
Baseline Corr 2nd AF:	4.68	AF Slope:	1.054634	AF Intercept:	-0.008381
Baseline Corr 3rd AF:	2.34	AF Correlation:	0.999996	* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.1	8.91	8.93	0.998
Mid point	4961	39.5	4.45	4.43	1.004
Low point	4980	19.8	2.23	2.20	1.012
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	8.91	8.96	0.994
Average Correction Factor					1.005

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.1	8.00	7.99	1.001
As found Mid point	4961	39.5	3.99	3.97	1.007
As found Low point	4980	19.8	2.00	1.98	1.012
New cylinder response					
Baseline Corr AF:	7.99	Prev response	8.00	*% change	-0.1%
Baseline Corr 2nd AF:	3.97	AF Slope:	0.999458	AF Intercept:	-0.012987
Baseline Corr 3rd AF:	1.98	AF Correlation:	0.999987	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.1	8.00	8.03	0.997
Mid point	4961	39.5	3.99	3.97	1.006
Low point	4980	19.8	2.00	1.97	1.016
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	8.00	8.04	0.995
Average Correction Factor					1.006

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.007610	1.003757
THC Cal Offset:	-0.060372	-0.041375
CH ₄ Cal Slope:	1.003661	1.004447
CH ₄ Cal Offset:	-0.029190	-0.023190
NMHC Cal Slope:	1.011373	1.003150
NMHC Cal Offset:	-0.031782	-0.018986

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 29, 2025	Previous Calibration:	April 21, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:08	End Time (MST):	12:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999971	<i>≥0.995</i>
16.91	16.96	0.9972	Slope	1.003757	<i>0.90 - 1.10</i>
8.44	8.40	1.0051	Intercept	-0.041375	<i>+/-0.5</i>
4.23	4.17	1.0140			





Wood Buffalo Environmental Association

CH₄ Calibration Summary


Station Information

Calibration Date: April 29, 2025
Station Name: Barge Landing
Start Time (MST): 9:08
Analyzer make: Thermo 55i

Previous Calibration: April 21, 2025
Station Number: AMS 09
End Time (MST): 12:22
Analyzer serial #: 1193585650

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999959	≥ 0.995
8.00	8.03	0.9967	Slope	1.004447	$0.90 - 1.10$
3.99	3.97	1.0063	Intercept	-0.023190	± 0.5
2.00	1.97	1.0164			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 29, 2025	Previous Calibration:	April 21, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:08	End Time (MST):	12:22
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999978	<i>≥0.995</i>
8.91	8.93	0.9977	Slope	1.003150	<i>0.90 - 1.10</i>
4.45	4.43	1.0045	Intercept	-0.018986	<i>+/-0.5</i>
2.23	2.20	1.0124			



NMHC Calibration Plot

Date: April 29, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Barge Landing
Station number: AMS 09
Calibration Date: April 11, 2025
Last Cal Date: March 17, 2025
Start time (MST): 9:09
End time (MST): 13:16
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1KDH
NOX Cal Gas Conc: 47.38 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.38 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 46.94 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 46.94 ppm
NO gas Diff:
Serial Number: 3812
Serial Number: 4888

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.2	----	----
AF High point	4915	85.3	808.3	800.7	7.5	810.0	798.5	11.5	0.9975	1.0026
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 807.3 ppb	NO = 799.0 ppb				* = > +/-5% change initiates investigation			*Percent Change	NO _x = 0.4%
Baseline Corr 1st pt	NO _x = 810.3 ppb	NO = 798.7 ppb				<u>As Found Statistics</u>			*Percent Change	NO = 0.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found NO _x r ² :			Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found NO r ² :			NO SI:	NO Int:
						As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1426262593

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.124	1.124	NO bkgnd or offset:	10.3	10.3
NOX coeff or slope:	0.999	0.999	NOX bkgnd or offset:	10.6	10.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	176.8	176.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998427	0.997861
NO _x Cal Offset:	0.358142	1.058268
NO Cal Slope:	0.998439	0.998054
NO Cal Offset:	-0.543944	-0.083947
NO ₂ Cal Slope:	1.002354	1.025270
NO ₂ Cal Offset:	-0.248929	-0.569655

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
High point	4915	85.3	808.3	800.7	7.5	807.1	799.3	7.8	1.0014	1.0018
Mid point	4957	42.6	403.7	400.0	3.7	404.1	398.6	5.5	0.9990	1.0034
Low point	4979	21.3	201.8	200.0	1.9	204.0	199.7	4.2	0.9893	1.0013
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
As left span	4915	85.3	808.3	438.2	370.1	801.2	438.2	362.9	1.0088	1.0000
Average Correction Factor									0.9966	1.0022

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	791.9	442.0	357.4	366.2	0.9760	102.5%
Mid GPT point	791.9	618.5	180.9	184.3	0.9816	101.9%
Low GPT point	791.9	706.1	93.3	95.0	0.9822	101.8%
Average Correction Factor					0.9799	102.1%

Notes: Inlet filter changed after as founds. No adjustment. 2nd NO reference point was used due to drift.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 17, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:09	End Time (MST):	13:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999988	≥0.995
808.3	807.1	1.0014	Slope	0.997861	0.90 - 1.10
403.7	404.1	0.9990	Intercept	1.058268	+/-20
201.8	204.0	0.9893			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 17, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:09	End Time (MST):	13:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999991	≥0.995
357.4	366.2	0.9760	Slope	1.025270	0.90 - 1.10
180.9	184.3	0.9816	Intercept	-0.569655	+/-20
93.3	95.0	0.9822			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 17, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:09	End Time (MST):	13:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999999	≥ 0.995
800.7	799.3	1.0018	Slope	0.998054	$0.90 - 1.10$
400.0	398.6	1.0034	Intercept	-0.083947	± 20
200.0	199.7	1.0013			



NO_x Calibration Plot

Date: April 11, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Barge Landing Station number: AMS 09
Calibration Date: April 9, 2025 Last Cal Date: March 24, 2025
Start time (MST): 14:15 End time (MST): 14:30

Analyzer Make: API T640 S/N: 844
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	16.10	16.10	16.10	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	724.00	732.50	724.00	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.03	4.90	5.03	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	36	----	36	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 36.40		PM w/ HEPA: 21.60		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	NA	NA	NA	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: March 24, 2025
Date Disposable Filter Changed: March 7, 2025

Post- maintenance Zero Verification: PM w/ HEPA: NA <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 18, 2024
Date RH/T Sensor Cleaned: October 18, 2024

Notes: Removal calibration due to erroneous readings. PMT check showed no response, and leak check failed.

Calibration by: Devin Russell



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Barge Landing Station number: AMS 09
Calibration Date: April 10, 2025 Last Cal Date: NA
Start time (MST): 8:50 End time (MST): 9:10

Analyzer Make: API T640 S/N: 2237
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	9.10	8.88	9.10	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	739.20	737.70	739.20	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	4.97	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	34	----	34	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 3.80		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: January 30, 2027
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.00	11.00	11.00	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 10, 2025
Date Disposable Filter Changed: April 10, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 18, 2024
Date RH/T Sensor Cleaned: October 18, 2024

Notes: Install calibration. Flow, temperature, pressure, and PMT verified.

Calibration by: Devin Russell



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS11 LOWER CAMP APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Lower Camp Station number: AMS 11
Calibration Date: April 9, 2025 Last Cal Date: March 3, 2025
Start time (MST): 12:07 End time (MST): 15:46
Reason: Routine

Calibration Standards

Cal Gas Concentration: 48.75 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC741503
Removed Cal Gas Conc: 48.75 ppm Rem Gas Exp Date:
Removed Gas Cyl #: CC741503 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3807
Zero Air Gen Model: Teledyne API T701 Serial Number: 196

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 100841398
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003845	1.018981	Backgd or Offset:	16.2	16.2
Calibration intercept:	-0.495900	0.056971	Coeff or Slope:	1.005	1.005

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.6	----
As found High point	4939	81.5	791.4	804.4	0.985
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	803.8	Previous response	793.9	*% change	1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.8	----
High point	4939	81.5	791.4	807.2	0.980
Mid point	4972	40.8	396.8	402.9	0.985
Low point	4990	20.4	198.5	202.3	0.981
As left zero	5000	0.0	0.0	0.7	----
As left span	4939	81.5	791.4	807.4	0.980
Average Correction Factor:					0.982

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	March 3, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	12:07	End Time (MST):	15:46
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.8	----	Correlation Coefficient	0.999991	≥0.995
791.4	807.2	0.9804	Slope	1.018981	0.90 - 1.10
396.8	402.9	0.9848	Intercept	0.056971	+/-30
198.5	202.3	0.9812			



SO2 Calibration Plot

Date: April 9, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Lower Camp Station number: AMS 11
Calibration Date: April 22, 2025 Last Cal Date: March 24, 2025
Start time (MST): 10:31 End time (MST): 14:50
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.83 ppm Cal Gas Exp Date: August 28, 2028
Cal Gas Cylinder #: CC737863
Removed Cal Gas Conc: 4.83 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3807
ZAG Make/Model: API T701H Serial Number: 196

Analyzer Information

Analyzer make: Thermo 43iQ Analyzer serial #: 1203169745
Converter make: Global G150 Converter serial #: 2022-223
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000358	1.003220	Backgd or Offset:	2.6
Calibration intercept:	-0.202641	-0.202864	Coeff or Slope:	0.752

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4935	83.0	79.9	80.1	0.995
As found Mid point	4977	41.6	40.0	40.1	0.993
As found Low point	4993	20.9	20.1	20.0	0.997
New cylinder response					
Baseline Corr As found:	80.3	Prev response:	79.72	*% change:	0.7%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.005367	AF Intercept:	-0.203054
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999999	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4935	83.0	79.9	79.9	1.000
Mid point	4977	41.6	40.0	40.1	0.998
Low point	4993	20.9	20.1	19.9	1.012
As left zero	5000	0.0	0.0	-0.2	----
As left span	4935	83.0	79.9	79.3	1.007
SO2 Scrubber Check	4935	81.5	812.3	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.003
Date of last converter efficiency test:					

Notes: Changed sample inlet filter after as founds. Ran scrubber check after calibrator zero and it passed.
No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 22, 2025	Previous Calibration:	March 24, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:31	End Time (MST):	14:50
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999991		≥ 0.995
79.9	79.9	0.9999	Slope	1.003220		$0.90 - 1.10$
40.0	40.1	0.9984	Intercept	-0.202864		± 3
20.1	19.9	1.0117				



H₂S Calibration Plot

Date: April 22, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Lower Camp
Calibration Date: April 9, 2025
Start time (MST): 12:07
Reason: Routine

Station number: AMS 11
Last Cal Date: March 16, 2025
End time (MST): 15:46

Calibration Standards

Gas Cert Reference:	CC741503	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	504.8 ppm	CH ₄ Equiv Conc.	1071.9 ppm
C ₃ H ₈ Cal Gas Conc.	206.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.8 ppm	CH ₄ Equiv Conc.	1071.9 ppm
Removed C ₃ H ₈ Conc.	206.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3807
Zero Air Gen model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1118148495
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.04E-04	2.04E-04	NMHC SP Ratio:	4.27E-05
CH ₄ Retention time:	14.3	14.3	NMHC Peak Area:	215713
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4939	81.5	17.40	16.94	1.027
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.94	Prev response	17.39	*% change	-2.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4939	81.5	17.40	16.85	1.033
Mid point	4972	40.8	8.72	8.43	1.035
Low point	4990	20.4	4.36	4.25	1.028
As left zero	5000	0.0	0.00	0.00	----
As left span	4939	81.5	17.40	16.88	1.031
Average Correction Factor					1.032

Notes: Changed sample inlet filter after as founds. No adjustments made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4939	81.5	9.21	8.86	1.039
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.86	Prev response	9.23	*% change	-4.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4939	81.5	9.21	8.80	1.046
Mid point	4972	40.8	4.62	4.43	1.042
Low point	4990	20.4	2.31	2.25	1.027
As left zero	5000	0.0	0.00	0.00	----
As left span	4939	81.5	9.21	8.83	1.042
Average Correction Factor					1.038

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4939	81.5	8.19	8.08	1.014
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.08	Prev response	8.16	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4939	81.5	8.19	8.04	1.019
Mid point	4972	40.8	4.11	4.00	1.027
Low point	4990	20.4	2.06	2.00	1.029
As left zero	5000	0.0	0.00	0.00	----
As left span	4939	81.5	8.19	8.05	1.018
Average Correction Factor					1.025

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.998443	0.967607
THC Cal Offset:	0.017156	0.004689
CH ₄ Cal Slope:	0.998463	0.981979
CH ₄ Cal Offset:	-0.017882	-0.014211
NMHC Cal Slope:	0.998401	0.954688
NMHC Cal Offset:	0.035637	0.019898

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	March 16, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	12:07	End Time (MST):	15:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995
17.40	16.85	1.0329	Slope	0.967607	$0.90 - 1.10$
8.72	8.43	1.0350	Intercept	0.004689	± 0.5
4.36	4.25	1.0280			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	March 16, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	12:07	End Time (MST):	15:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			Limits
0.00	0.00	----	Correlation Coefficient	0.999979		≥0.995
8.19	8.04	1.0187	Slope	0.981979		0.90 - 1.10
4.11	4.00	1.0269	Intercept	-0.014211		+/-0.5
2.06	2.00	1.0287				





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	March 16, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	12:07	End Time (MST):	15:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999975	<i>≥0.995</i>
9.21	8.80	1.0460	Slope	0.954688	<i>0.90 - 1.10</i>
4.62	4.43	1.0418	Intercept	0.019898	<i>+/-0.5</i>
2.31	2.25	1.0270			



NMHC Calibration Plot

Date: April 9, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Lower Camp
Calibration Date: April 25, 2025
Start time (MST): 13:47
Reason: Maintenance

Station number: AMS 11
Last Cal Date: April 9, 2025
End time (MST): 17:13

Calibration Standards

Gas Cert Reference:	CC741503	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	504.8 ppm	CH ₄ Equiv Conc.	1071.9 ppm
C ₃ H ₈ Cal Gas Conc.	206.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.8 ppm	CH ₄ Equiv Conc.	1071.9 ppm
Removed C ₃ H ₈ Conc.	206.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3811
Zero Air Gen model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1118148495
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.04E-04	2.27E-04	NMHC SP Ratio:	4.27E-05
CH ₄ Retention time:	14.3	14.3	NMHC Peak Area:	215713
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	82.2	17.57	15.32	1.147
As found Mid point	4971	41.2	8.81	7.64	1.153
As found Low point	4996	20.6	4.40	3.85	1.143
New cylinder response					
Baseline Corr AF:	15.32	Prev response	17.01	*% change	-11.0%
Baseline Corr 2nd AF:	7.64	AF Slope:	0.871265	AF Intercept:	-0.002098
Baseline Corr 3rd AF:	3.85	AF Correlation:	0.999987	* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	82.2	17.57	17.45	1.007
Mid point	4971	41.2	8.81	8.72	1.010
Low point	4996	20.6	4.40	4.40	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	17.57	17.26	1.018
Average Correction Factor					1.006

Notes: No alarms were detected during the investigation. Diagnostics were consistent, and the chromatograms showed no discrepancies, aside from a slight alignment drift. The issue is likely due to the new calibrator, prompting only a span adjustment.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	82.2	9.30	7.96	1.167
As found Mid point	4971	41.2	4.66	3.99	1.168
As found Low point	4996	20.6	2.33	2.03	1.148
New cylinder response					
Baseline Corr AF:	7.96	Prev response	8.89	*% change	-11.7%
Baseline Corr 2nd AF:	3.99	AF Slope:	0.855410	AF Intercept:	0.012797
Baseline Corr 3rd AF:	2.03	AF Correlation:	0.999975	* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	82.2	9.30	9.24	1.006
Mid point	4971	41.2	4.66	4.65	1.002
Low point	4996	20.6	2.33	2.36	0.985
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	9.30	9.14	1.017
Average Correction Factor					0.998

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	82.2	8.28	7.35	1.125
As found Mid point	4971	41.2	4.15	3.65	1.136
As found Low point	4996	20.6	2.07	1.82	1.137
New cylinder response					
Baseline Corr AF:	7.35	Prev response	8.11	*% change	-10.3%
Baseline Corr 2nd AF:	3.65	AF Slope:	0.888951	AF Intercept:	-0.014694
Baseline Corr 3rd AF:	1.82	AF Correlation:	0.999971	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	82.2	8.28	8.21	1.009
Mid point	4971	41.2	4.15	4.07	1.020
Low point	4996	20.6	2.07	2.04	1.018
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	8.28	8.12	1.019
Average Correction Factor					1.015

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.967607	0.992250
THC Cal Offset:	0.004689	0.006074
CH ₄ Cal Slope:	0.981979	0.991671
CH ₄ Cal Offset:	-0.014211	-0.016495
NMHC Cal Slope:	0.954688	0.992864
NMHC Cal Offset:	0.019898	0.022170

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 25, 2025	Previous Calibration:	April 9, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	13:47	End Time (MST):	17:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999991	≥ 0.995
17.57	17.45	1.0071	Slope	0.992250	$0.90 - 1.10$
8.81	8.72	1.0102	Intercept	0.006074	± 0.5
4.40	4.40	1.0005			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 25, 2025	Previous Calibration:	April 9, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	13:47	End Time (MST):	17:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999964	≥ 0.995
8.28	8.21	1.0086	Slope	0.991671	0.90 - 1.10
4.15	4.07	1.0195	Intercept	-0.016495	± 0.5
2.07	2.04	1.0181			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 25, 2025	Previous Calibration:	April 9, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	13:47	End Time (MST):	17:13
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999970	<i>≥0.995</i>
9.30	9.24	1.0056	Slope	0.992864	<i>0.90 - 1.10</i>
4.66	4.65	1.0022	Intercept	0.022170	<i>+/-0.5</i>
2.33	2.36	0.9854			



NMHC Calibration Plot

Date: April 25, 2025

Location: Lower Camp





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13 FORT MCKAY SOUTH APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC
APRIL 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Anzac Station number: AMS 14
Calibration Date: April 7, 2025 Last Cal Date: March 5, 2025
Start time (MST): 10:08 End time (MST): 14:46
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.32 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC462030
Removed Cal Gas Conc: 50.32 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3060
Zero Air Gen Model: API T701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 0710321322
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004124	0.995812	Backgd or Offset:	24.4	24.8
Calibration intercept:	-0.860151	-1.079277	Coeff or Slope:	1.074	1.074

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.8	----
As found High point	4941	79.7	798.8	792.4	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	791.6	Previous response	801.2	*% change	-1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.9	----
High point	4941	79.7	798.8	795.7	1.004
Mid point	4980	39.9	400.0	395.2	1.012
Low point	4994	19.9	199.7	196.5	1.016
As left zero	5000	0.0	0.0	0.7	----
As left span	4941	79.7	798.8	796.7	1.003
Average Correction Factor:					1.011

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 5, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:08	End Time (MST):	14:46
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.9	----	Correlation Coefficient	0.999967	≥0.995
798.8	795.7	1.0039	Slope	0.995812	0.90 - 1.10
400.0	395.2	1.0120	Intercept	-1.079277	+/-30
199.7	196.5	1.0164			



SO2 Calibration Plot

Date: April 7, 2025

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Anzac
Calibration Date: April 16, 2025
Start time (MST): 11:08
Reason: Routine

Station number: AMS 14
Last Cal Date: March 11, 2025
End time (MST): 18:45

Calibration Standards

Cal Gas Concentration: 5.15 ppm
Cal Gas Cylinder #: CC510379
Removed Cal Gas Conc: 5.15 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API 701H

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3060
Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: CD Nova CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1218153582
Converter serial #: 503
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.981575	1.010026	Backgd or Offset: 2.4	2.4
Calibration intercept:	-0.085266	-0.005494	Coeff or Slope: 1.027	1.026

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4938	77.9	80.0	82.1	0.973
As found Mid point	4973	38.9	40.0	41.3	0.966
As found Low point	4997	19.5	20.0	20.5	0.972
New cylinder response					
Baseline Corr As found:	82.2	Prev response:	78.42	*% change:	4.6%
Baseline Corr 2nd AF pt:	41.4	AF Slope:	1.028067	AF Intercept:	-0.025511
Baseline Corr 3rd AF pt:	20.6	AF Correlation:	0.999980	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4938	77.9	80.0	80.9	0.988
Mid point	4973	38.9	40.0	40.1	0.996
Low point	4997	19.5	20.0	20.1	0.996
As left zero	5000	0.0	0.0	0.3	----
As left span	4938	77.9	80.0	78.9	1.013
SO2 Scrubber Check	4936	80.3	800.4	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.993
Date of last converter efficiency test:		April 16, 2025	86.9% efficiency		

Notes: Changed the sample inlet filter and performed converter efficiency test(13:00 to 16:00) after as founds. Completed a SO2 scrubber check after calibrator zero. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	April 16, 2025	Previous Calibration:	March 11, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:08	End Time (MST):	18:45
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999961		≥ 0.995
80.0	80.9	0.9883	Slope	1.010026		$0.90 - 1.10$
40.0	40.1	0.9964	Intercept	-0.005494		± 3
20.0	20.1	0.9956				


TRS Calibration Curve



TRS Calibration Plot

Date: April 16, 2025

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Anzac
 Calibration Date: April 7, 2025
 Start time (MST): 10:08
 Reason: Routine

Station number: AMS 14
 Last Cal Date: March 5, 2025
 End time (MST): 14:46

Calibration Standards

Gas Cert Reference:	CC462030	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	505.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	204.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	505.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	204.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
Zero Air Gen model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1331259521
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.92E-04	2.88E-04	NMHC SP Ratio:	5.51E-05
CH ₄ Retention time:	14.9	14.9	NMHC Peak Area:	162132
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4941	79.7	16.97	16.98	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.98	Prev response	16.85	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4941	79.7	16.97	16.95	1.001
Mid point	4980	39.9	8.50	8.30	1.024
Low point	4994	19.9	4.24	4.09	1.037
As left zero	5000	0.0	0.00	0.00	----
As left span	4941	79.7	16.97	16.96	1.000
Average Correction Factor					1.021

Notes:

Changed sample inlet filter after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4941	79.7	8.94	8.81	1.015
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.81	Prev response	8.81	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4941	79.7	8.94	8.93	1.002
Mid point	4980	39.9	4.48	4.39	1.020
Low point	4994	19.9	2.24	2.17	1.030
As left zero	5000	0.0	0.00	0.00	----
As left span	4941	79.7	8.94	8.93	1.001
Average Correction Factor					1.017

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4941	79.7	8.02	8.17	0.982
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.17	Prev response	8.04	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4941	79.7	8.02	8.02	1.001
Mid point	4980	39.9	4.02	3.91	1.029
Low point	4994	19.9	2.01	1.92	1.045
As left zero	5000	0.0	0.00	0.00	----
As left span	4941	79.7	8.02	8.03	0.999
Average Correction Factor					1.025

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.998574	1.000487
THC Cal Offset:	-0.092906	-0.096108
CH ₄ Cal Slope:	1.008920	1.001383
CH ₄ Cal Offset:	-0.051592	-0.055381
NMHC Cal Slope:	0.989475	0.999223
NMHC Cal Offset:	-0.041514	-0.039928

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 5, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:08	End Time (MST):	14:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999819	<i>≥0.995</i>
16.97	16.95	1.0012	Slope	1.000487	<i>0.90 - 1.10</i>
8.50	8.30	1.0240	Intercept	-0.096108	<i>+/-0.5</i>
4.24	4.09	1.0369			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 5, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:08	End Time (MST):	14:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999732	<i>≥0.995</i>
8.02	8.02	1.0007	Slope	1.001383	<i>0.90 - 1.10</i>
4.02	3.91	1.0285	Intercept	-0.055381	<i>+/-0.5</i>
2.01	1.92	1.0445			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 5, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:08	End Time (MST):	14:46
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999888	<i>≥0.995</i>
8.94	8.93	1.0021	Slope	0.999223	<i>0.90 - 1.10</i>
4.48	4.39	1.0200	Intercept	-0.039928	<i>+/-0.5</i>
2.24	2.17	1.0301			



NMHC Calibration Plot

Date: April 7, 2025

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Anzac
 Station number: AMS 14
 Calibration Date: April 4, 2025
 Last Cal Date: March 6, 2025
 Start time (MST): 10:16
 End time (MST): 14:50
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0037092
 NOX Cal Gas Conc: 60.7 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 60.70 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T700H
 Cal Gas Expiry Date: May 16, 2031
 NO Cal Gas Conc: 60.40 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 60.40 ppm
 NO gas Diff:
 Serial Number: 3060
 Serial Number: 357

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
AF High point	4934	66.3	804.8	800.9	4.0	806.2	801.0	5.0	0.9982	0.9996
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 800.5 ppb	NO = 798.4 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.7%	
Baseline Corr 1st pt	NO _x = 806.3 ppb	NO = 801.2 ppb				<u>As Found Statistics</u>		*Percent Change	NO = 0.3%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430008

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.424	1.424	NO bkgnd or offset:	3.9	3.9
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	156.9	159.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995143	0.998863
NO _x Cal Offset:	-0.470500	-0.430260
NO Cal Slope:	0.999427	1.002566
NO Cal Offset:	-1.989787	-2.089249
NO ₂ Cal Slope:	0.994537	0.994974
NO ₂ Cal Offset:	-1.309208	-1.530514

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
High point	4934	66.3	804.8	800.9	4.0	803.6	801.6	1.9	1.0015	0.9991
Mid point	4985	33.2	401.6	399.6	2.0	401.0	398.4	2.5	1.0015	1.0030
Low point	5004	16.7	201.9	200.9	1.0	200.2	196.5	3.6	1.0085	1.0224
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
As left span	4934	66.3	804.8	416.8	388.0	801.9	416.8	385.0	1.0037	1.0000
Average Correction Factor									1.0038	1.0082

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	800.6	411.8	392.8	390.0	1.0071	99.3%
Mid GPT point	800.6	601.7	202.9	199.9	1.0149	98.5%
Low GPT point	800.6	701.6	103.0	99.0	1.0402	96.1%
Average Correction Factor					1.0207	98.0%

Notes:

Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 4, 2025	Previous Calibration:	March 6, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:16	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999996	≥0.995
804.8	803.6	1.0015	Slope	0.998863	0.90 - 1.10
401.6	401.0	1.0015	Intercept	-0.430260	+/-20
201.9	200.2	1.0085			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 4, 2025	Previous Calibration:	March 6, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:16	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999915	≥0.995
392.8	390.0	1.0071	Slope	0.994974	0.90 - 1.10
202.9	199.9	1.0149	Intercept	-1.530514	+/-20
103.0	99.0	1.0402			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 4, 2025	Previous Calibration:	March 6, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:16	End Time (MST):	14:50
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999962	≥ 0.995
800.9	801.6	0.9991	Slope	1.002566	0.90 - 1.10
399.6	398.4	1.0030	Intercept	-2.089249	+/-20
200.9	196.5	1.0224			



NO_x Calibration Plot

Date: April 4, 2025

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Anzac
Calibration Date: April 1, 2025
Start time (MST): 10:56
Reason: Routine

Station number: AMS 14
Last Cal Date: March 4, 2025
End time (MST): 14:06

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: API T700
ZAG Make/Model: API 701H

Serial Number: 3060
Serial Number: 357

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1426262595

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.991800	0.990971	Backgd or Offset:	1.5	1.6
Calibration intercept:	0.060000	0.380000	Coeff or Slope:	1.668	1.668

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.9	----
As found High point	5000	935.9	400.0	397.8	1.008
As found Mid point					
As found Low point					
Baseline Corr As found:	396.9	Previous response	396.8	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	5000	935.9	400.0	396.7	1.008
Mid point	5000	817.5	200.0	198.7	1.007
Low point	5000	722.8	100.0	99.5	1.005
As left zero	5000	0.0	0.0	0.3	----
As left span	5000	935.9	400.0	401.4	0.997
Average Correction Factor					1.007

Notes: Sample inlet filter changed after asfound. No adjustment made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 1, 2025	Previous Calibration:	March 4, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:56	End Time (MST):	14:06
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	1.000000	≥0.995
400.0	396.7	1.0083	Slope	0.990971	0.90 - 1.10
200.0	198.7	1.0065	Intercept	0.380000	+/- 5
100.0	99.5	1.0050			



O₃ Calibration Plot

Date: April 1, 2025

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Anzac
Calibration Date: April 29, 2025
Start time (MST): 11:50
Station number: AMS 14
Last Cal Date: March 14, 2025
End time (MST): 12:12
Analyzer Make: AP T640
S/N: 825
Particulate Fraction: PM2.5
Flow Meter Make/Model: Alicat FP-25BT
S/N: 388749
Temp/RH standard: Alicat FP-25BT
S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	8.6	8.18	8.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	705.4	706.1	705.4	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.000	4.95	5.000	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	36	-----	36	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 0.3		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: Expiry Date:
Lot No.:

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: January 30, 2025
Date Disposable Filter Changed: January 30, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 29, 2024
Date RH/T Sensor Cleaned: August 29, 2024

Notes: No adjustments made. Leak check passed. Head cleaned

Calibration by: Mohammed Kashif



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS17
WAPASU
APRIL 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Wapasu Station number: AMS17
Calibration Date: April 7, 2025 Last Cal Date: March 4, 2025
Start time (MST): 10:15 End time (MST): 13:59
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.38 ppm Cal Gas Exp Date: January 12, 2029
Cal Gas Cylinder #: ALM066507
Removed Cal Gas Conc: 50.38 ppm Rem Gas Exp Date: N/A
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 2449
Zero Air Gen Model: Teledyne API 701H Serial Number: 1238

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1218153459
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994240	0.995597	Backgd or Offset:	14.0	14.0
Calibration intercept:	-1.540718	-1.540479	Coeff or Slope:	1.109	1.109

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4921	79.4	800.0	794.8	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	794.8	Previous response	793.8	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4921	79.4	800.0	796.2	1.005
Mid point	4960	39.7	400.0	394.7	1.014
Low point	4980	19.8	199.5	196.0	1.018
As left zero	5000	0.0	0.0	0.3	----
As left span	4920	79.4	800.1	797.9	1.003
Average Correction Factor:					1.012

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 4, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:15	End Time (MST):	13:59
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999970	≥0.995
800.0	796.2	1.0047	Slope	0.995597	0.90 - 1.10
400.0	394.7	1.0135	Intercept	-1.540479	+/-30
199.5	196.0	1.0179			



SO2 Calibration Plot

Date: April 7, 2025

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Wapasu
Calibration Date: April 15, 2025
Start time (MST): 10:33
Reason: Routine

Station number: AMS 17
Last Cal Date: March 20, 2025
End time (MST): 15:20

Calibration Standards

Cal Gas Concentration: 4.77 ppm
Cal Gas Cylinder #: DT20029267
Removed Cal Gas Conc: 4.77 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API T701H

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: Thermo 450i
Converter make: CD Nova
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1218153583
Converter serial #: N/A
Converter Temp: 340 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997355	1.001218	Backgd or Offset:	13.1	13.1
Calibration intercept:	0.680200	-0.180032	Coeff or Slope:	1.099	1.099

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4921	83.9	80.0	80.7	0.988
As found Mid point	4961	41.9	39.9	40.5	0.982
As found Low point	4980	21.0	20.0	19.7	1.007
New cylinder response					
Baseline Corr As found:	80.9	Prev response:	80.43	*% change:	0.6%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.013690	AF Intercept:	-0.289348
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999945	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4916	83.9	80.0	80.2	0.998
Mid point	4958	41.9	40.0	39.4	1.015
Low point	4979	21.0	20.0	19.8	1.012
As left zero	5000	0.0	0.0	0.2	----
As left span	4916	83.9	80.0	78.3	1.022
SO2 Scrubber Check	4921	79.4	793.9	0.1	----
Date of last scrubber change:		N/A		Ave Corr Factor	1.008
Date of last converter efficiency test:		N/A			

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	April 15, 2025	Previous Calibration:	March 20, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:33	End Time (MST):	15:20
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999904		≥ 0.995
80.0	80.2	0.9980	Slope	1.001218		$0.90 - 1.10$
40.0	39.4	1.0146	Intercept	-0.180032		± 3
20.0	19.8	1.0118				


H₂S Calibration Curve



H₂S Calibration Plot

Date: April 15, 2025

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Wapasu Station number: AMS17
Calibration Date: April 7, 2025 Last Cal Date: March 4, 2025
Start time (MST): 10:15 End time (MST): 13:59
Reason: Routine

Calibration Standards

Gas Cert Reference: ALM066507 Cal Gas Expiry Date: January 12, 2029
CH4 Cal Gas Conc. 503.5 ppm CH4 Equiv Conc. 1076.3 ppm
C3H8 Cal Gas Conc. 208.3 ppm
Removed Gas Cert: n/a Removed Gas Expiry:
Removed CH4 Conc. 503.5 ppm CH4 Equiv Conc. 1076.3 ppm
Removed C3H8 Conc. 208.3 ppm Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 2449
ZAG Make/Model: Teledyne API 701H Serial Number: 1238

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1218153352
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993740	1.003335	Background:	3.230	3.340
Calibration intercept:	-0.053360	-0.176337	Coefficient:	4.337	4.476

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	-0.15	----
As found High point	4921	79.4	17.09	16.50	1.027
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	16.65	Previous response	16.93	*% change	-1.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	-0.13	----
High point	4921	79.4	17.09	17.01	1.005
Mid point	4960	39.7	8.55	8.34	1.025
Low point	4980	19.8	4.26	4.07	1.047
As left zero	5000	0.0	0.00	-0.16	----
As left span	4921	79.4	17.09	17.03	1.004
Average Correction Factor					1.026

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 4, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:15	End Time (MST):	13:59
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.13	----	Correlation Coefficient	0.999949	≥ 0.995
17.09	17.01	1.0049	Slope	1.003335	$0.90 - 1.10$
8.55	8.34	1.0248	Intercept	-0.176337	± 1.5
4.26	4.07	1.0470			


THC Calibration Curve



THC Calibration Plot

Date: April 7, 2025

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Wapasu
Station number: AMS 17
Calibration Date: April 8, 2025
Last Cal Date: March 19, 2025
Start time (MST): 10:25
End time (MST): 15:44
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T375YK8
NOX Cal Gas Conc: 49.11 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 49.11 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: April 13, 2025
NO Cal Gas Conc: 48.07 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 48.07 ppm
NO gas Diff:
Serial Number: 2449
Serial Number: 359

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.2	-0.4	----	----
AF High point	4917	83.2	817.2	799.9	17.3	818.4	798.2	20.2	0.9978	1.0019
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 817.0 ppb	NO = 797.9 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = 0.2%			
Baseline Corr 1st pt	NO _x = 819.0 ppb	NO = 798.4 ppb	<u>As Found Statistics</u>			*Percent Change	NO = 0.1%			
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:	NO Int:			
			As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo Scientific 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153460

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.084	1.084
NOX coeff or slope:	0.996	0.996
NO2 coeff or slope:	1.000	1.000

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	3.8	3.8
NOX bkgnd or offset:	4.2	4.2
Reaction cell Press:	240.6	238.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001354	0.999201
NO _x Cal Offset:	-1.280000	-1.360000
NO Cal Slope:	0.999687	0.999815
NO Cal Offset:	-1.740000	-1.860000
NO ₂ Cal Slope:	1.002988	0.998689
NO ₂ Cal Offset:	0.366820	-0.611431

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.4	0.0	-0.4	----	----
High point	4917	83.2	817.2	799.9	17.3	815.8	798.9	17.0	1.0017	1.0012
Mid point	4958	41.6	408.6	399.9	8.7	406.0	396.8	9.2	1.0064	1.0079
Low point	4979	20.8	204.3	200.0	4.3	202.1	196.4	5.7	1.0109	1.0182
As left zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.4	----	----
As left span	4917	83.2	817.2	396.7	420.5	814.5	396.7	417.8	1.0033	1.0000
Average Correction Factor									1.0063	1.0091

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.4	----	----
High GPT point	796.3	398.9	414.7	413.8	1.0022	99.8%
Mid GPT point	796.3	600.8	212.8	211.4	1.0066	99.3%
Low GPT point	796.3	700.5	113.1	112.4	1.0063	99.4%
Average Correction Factor					1.0050	99.5%

Notes:

Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 8, 2025	Previous Calibration:	March 19, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:25	End Time (MST):	15:44
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	0.999993	≥0.995
817.2	815.8	1.0017	Slope	0.999201	0.90 - 1.10
408.6	406.0	1.0064	Intercept	-1.360000	+/-20
204.3	202.1	1.0109			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 8, 2025	Previous Calibration:	March 19, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:25	End Time (MST):	15:44
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	0.999996	≥0.995
414.7	413.8	1.0022	Slope	0.998689	0.90 - 1.10
212.8	211.4	1.0066	Intercept	-0.611431	+/-20
113.1	112.4	1.0063			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 8, 2025	Previous Calibration:	March 19, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:25	End Time (MST):	15:44
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999975	≥ 0.995
799.9	798.9	1.0012	Slope	0.999815	$0.90 - 1.10$
399.9	396.8	1.0079	Intercept	-1.860000	± 20
200.0	196.4	1.0182			



NO_x Calibration Plot

Date: April 8, 2025

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Wapasu
Calibration Date: April 3, 2025
Start time (MST): 10:35
Reason: Routine

Station number: AMS17
Last Cal Date: March 13, 2025
End time (MST): 14:25

Calibration Standards

O3 generation mode: Photometer
Calibrator Make/Model: API T700
ZAG Make/Model: API T701H

Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: API T400
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 7045

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010314	0.995943	Backgd or Offset:	0.6	0.6
Calibration intercept:	-0.280000	0.160000	Coeff or Slope:	1.046	1.027

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	5000	1104.7	400.0	406.7	0.983
As found Mid point					
As found Low point					
Baseline Corr As found:	406.9	Previous response	403.8	*% change	0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	5000	1104.7	400.0	398.6	1.004
Mid point	5000	917.3	200.0	199.0	1.005
Low point	5000	797.9	100.0	100.2	0.998
As left zero	5000	0.0	0.0	0.7	----
As left span	5000	1104.0	400.0	399.4	1.002
Average Correction Factor					1.002

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 3, 2025	Previous Calibration:	March 13, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:35	End Time (MST):	14:25
Analyzer make:	API T400	Analyzer serial #:	7045

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999996	≥0.995
400.0	398.6	1.0035	Slope	0.995943	0.90 - 1.10
200.0	199.0	1.0050	Intercept	0.160000	+/- 5
100.0	100.2	0.9980			



O₃ Calibration Plot

Date: April 3, 2025

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Wapasu
Calibration Date: April 13, 2025
Start time (MST): 11:20
Reason: Maintenance

Station number: AMS17
Last Cal Date: April 3, 2025
End time (MST): 13:50

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: API T700
ZAG Make/Model: API T701H

Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: API T400
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 7045

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995943	1.001457	Backgd or Offset:	0.6	0.6
Calibration intercept:	0.160000	-0.680000	Coeff or Slope:	1.027	1.025

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	5000	1104.7	400.0	400.1	1.000
Mid point	5000	917.3	200.0	199.5	1.003
Low point	5000	797.9	100.0	98.9	1.011
As left zero	5000	0.0	0.0	0.1	----
As left span	5000	1104.0	400.0	400.3	0.999
Average Correction Factor					1.004

Notes: O₃ Reference was below limits causing no readings. No as founds as the instrument was not operating upon arrival. Adjusted the lamp to increase the O₃ ref. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 13, 2025	Previous Calibration:	April 3, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:20	End Time (MST):	13:50
Analyzer make:	API T400	Analyzer serial #:	7045

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999993	≥0.995
400.0	400.1	0.9998	Slope	1.001457	0.90 - 1.10
200.0	199.5	1.0025	Intercept	-0.680000	+/- 5
100.0	98.9	1.0111			



O₃ Calibration Plot

Date: April 13, 2025

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Wapasu Station number: AMS 17
Calibration Date: April 15, 2025 Last Cal Date: March 20, 2025
Start time (MST): 14:31 End time (MST): 15:12

Analyzer Make: Teledyne API T640 S/N: 1183
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.20	9.80	10.20	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	705.40	706.80	705.40	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.98	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	33	----	36	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 0.2		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 6, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	N/A	N/A	N/A	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 26, 2025
Date Disposable Filter Changed: March 20, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 23, 2024
Date RH/T Sensor Cleaned: July 23, 2024

Notes:

Flow, temp and pressure checked.

Calibration by: Aswin Sasi Kumar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: April 28, 2025 Last Cal Date: March 12, 2025
Start time (MST): 11:05 End time (MST): 14:59
Reason: Routine

Calibration Standards

Cal Gas Concentration: 51.22 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC417455
Removed Cal Gas Conc: 51.22 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: CC417455 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 282
Zero Air Gen Model: Teledyne API 701H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1501301453
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003718	1.008274	Backgd or Offset:	25.6	25.3
Calibration intercept:	-3.279867	-2.906028	Coeff or Slope:	0.830	0.818

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4921	78.1	800.2	814.1	0.983
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	813.8	Previous response	799.9	*% change	1.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4921	78.1	800.2	804.6	0.995
Mid point	4960	39.1	400.6	402.4	0.996
Low point	4980	20.0	204.9	198.3	1.033
As left zero	5000	0.0	0.0	0.6	----
As left span	4921	78.1	800.2	805.4	0.994
Average Correction Factor:					1.008

Notes: Changed sample inlet filter. Span adjusted.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 28, 2025	Previous Calibration:	March 12, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:05	End Time (MST):	14:59
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999881	≥0.995
800.2	804.6	0.9945	Slope	1.008274	0.90 - 1.10
400.6	402.4	0.9956	Intercept	-2.906028	+/-30
204.9	198.3	1.0332			



SO2 Calibration Plot

Date: April 28, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS18
Calibration Date: April 24, 2025 Last Cal Date: March 25, 2025
Start time (MST): 11:00 End time (MST): 16:06
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.86 ppm Cal Gas Exp Date: May 9, 2027
Cal Gas Cylinder #: CC523103
Removed Cal Gas Conc: 4.86 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
ZAG Make/Model: Teledyne API T701 Serial Number: 360

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359
Converter make: CD Nova CDN-101 Converter serial #: 555
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007792	1.002370	Backgd or Offset:	2.9	2.90
Calibration intercept:	-0.218991	0.020908	Coeff or Slope:	1.172	1.181

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4917	82.3	80.0	82.7	0.969
As found Mid point	4958	41.2	40.1	41.3	0.972
As found Low point	4979	20.6	20.0	20.4	0.986
New cylinder response					
Baseline Corr As found:	82.6	Prev response:	80.41	*% change:	2.6%
Baseline Corr 2nd AF pt:	41.2	AF Slope:	1.033797	AF Intercept:	-0.079704
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999977	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4917	82.3	80.0	80.3	0.996
Mid point	4958	41.2	40.1	40.2	0.996
Low point	4979	20.6	20.0	19.7	1.016
As left zero	5000	0.0	0.0	0.4	----
As left span	4917	82.3	80.0	80.4	0.995
SO2 Scrubber Check	4923	77.1	771.0	0.1	----
Date of last scrubber change:	17-Dec-21		Ave Corr Factor		1.003
Date of last converter efficiency test:					

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	April 24, 2025	Previous Calibration:	March 25, 2025
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	11:00	End Time (MST):	16:06
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999932		≥ 0.995
80.0	80.3	0.9963	Slope	1.002370		$0.90 - 1.10$
40.1	40.2	0.9963	Intercept	0.020908		± 3
20.0	19.7	1.0165				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS 18
 Calibration Date: April 17, 2025 Last Cal Date: March 18, 2025
 Start time (MST): 11:30 End time (MST): 16:05
 Reason: Cylinder Change N2 cylinder change out

Calibration Standards

Gas Cert Reference: XC026809B Cal Gas Expiry Date: January 12, 2029
 CH4 Cal Gas Conc. 504.9 ppm CH4 Equiv Conc. 1076.6 ppm
 C3H8 Cal Gas Conc. 207.9 ppm
 Removed Gas Cert: NA Removed Gas Expiry: NA
 Removed CH4 Conc. 504.9 ppm CH4 Equiv Conc. 1076.6 ppm
 Removed C3H8 Conc. 207.9 ppm Diff between cyl (THC):
 Diff between cyl (CH₄): Diff between cyl (NM):
 Calibrator Model: Teledyne API T750 Serial Number: 282
 Zero Air Gen model: Teledyne API T751H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 1218153355
 THC Range: 0 - 20 ppm NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.17E-04	N/A	NMHC SP Ratio:	4.10E-05
CH4 Retention time:	14.2	N/A	NMHC Peak Area:	217649
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.1	16.82	16.49	1.020
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.49	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	16.82	16.52	1.018
Average Correction Factor					

Notes: N2 cylinder changed out after as founds.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.1	8.93	8.73	1.023
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.73	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	8.93	8.74	1.022
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.1	7.89	7.76	1.016
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.76	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	7.89	7.77	1.015
Average Correction Factor					


Calibration Statistics

Start

Finish

THC Cal Slope:
THC Cal Offset:
CH₄ Cal Slope:
CH₄ Cal Offset:
NMHC Cal Slope:
NMHC Cal Offset:

Calibration Performed By: Aswin Sasi Kumar





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Stony Mountain
 Calibration Date: April 28, 2025
 Start time (MST): 11:05
 Reason: Routine

Station number: AMS 18
 Last Cal Date: March 18, 2025
 End time (MST): 14:59

Calibration Standards

Gas Cert Reference:	XC026809B	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T750	Serial Number:	282
Zero Air Gen model:	Teledyne API T751H	Serial Number:	321

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1218153355
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.17E-04	2.21E-04	NMHC SP Ratio:	4.10E-05
CH ₄ Retention time:	14.2	14.4	NMHC Peak Area:	217649
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.02	----
As found High point	4921	78.1	16.82	16.38	1.028
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.36	Prev response	16.83	*% change	-2.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.1	16.82	16.82	1.000
Mid point	4960	39.1	8.42	8.39	1.004
Low point	4980	20.0	4.31	4.17	1.032
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	16.82	16.80	1.001
Average Correction Factor					1.012

Notes: Changed H2 cylinder and sample inlet filter. Adjusted zero and span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	78.1	8.93	8.66	1.032
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.66	Prev response	8.96	*% change	-3.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	78.1	8.93	8.92	1.001
Mid point	4960	39.1	4.47	4.46	1.002
Low point	4980	20.0	2.29	2.23	1.027
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	8.93	8.96	0.997
Average Correction Factor					1.010

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.02	Limit = 0.90-1.10
As found High point	4921	78.1	7.89	7.72	1.024
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.70	Prev response	7.87	*% change	-2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	78.1	7.89	7.89	0.999
Mid point	4960	39.1	3.95	3.92	1.006
Low point	4980	20.0	2.02	1.95	1.038
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	7.89	7.84	1.007
Average Correction Factor					1.015

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.003220	1.001998
THC Cal Offset:	-0.041072	-0.057041
CH ₄ Cal Slope:	1.001011	1.003000
CH ₄ Cal Offset:	-0.021650	-0.033289
NMHC Cal Slope:	1.005632	1.000998
NMHC Cal Offset:	-0.020230	-0.023550

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 28, 2025	Previous Calibration:	March 18, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:05	End Time (MST):	14:59
Analyzer make:	Thermo 55i	Analyzer serial #:	1218153355

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999927	<i>≥0.995</i>
16.82	16.82	1.0003	Slope	1.001998	<i>0.90 - 1.10</i>
8.42	8.39	1.0040	Intercept	-0.057041	<i>+/-0.5</i>
4.31	4.17	1.0320			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 28, 2025	Previous Calibration:	March 18, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:05	End Time (MST):	14:59
Analyzer make:	Thermo 55i	Analyzer serial #:	1218153355

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999892	<i>≥0.995</i>
7.89	7.89	0.9995	Slope	1.003000	<i>0.90 - 1.10</i>
3.95	3.92	1.0064	Intercept	-0.033289	<i>+/-0.5</i>
2.02	1.95	1.0378			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 28, 2025	Previous Calibration:	March 18, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:05	End Time (MST):	14:59
Analyzer make:	Thermo 55i	Analyzer serial #:	1218153355

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999950	<i>≥0.995</i>
8.93	8.92	1.0011	Slope	1.000998	<i>0.90 - 1.10</i>
4.47	4.46	1.0019	Intercept	-0.023550	<i>+/-0.5</i>
2.29	2.23	1.0269			



NMHC Calibration Plot

Date: April 28, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Stony Mountain
Station number: AMS 18
Calibration Date: April 24, 2025
Last Cal Date: March 25, 2025
Start time (MST): 11:00
End time (MST): 16:55
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0045516
NO_x Cal Gas Conc: 60.30 ppm
Removed Cylinder #: N/A
Removed Gas NO_x Conc: 60.30 ppm
NO_x gas Diff:
Calibrator Model: Teledyne API T750
ZAG make/model: Teledyne API 751H
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 60.10 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 60.10 ppm
NO gas Diff:
Serial Number: 282
Serial Number: 321

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.0	-0.9	-0.1	----	----
AF High point	4933	66.6	803.3	800.6	2.7	792.0	795.3	-3.3	1.0129	1.0055
AF Mid point										
AF Low point										
New cyl resp										
Previous Respo 4933	NO _x = 801.8 ppb	NO = 801.9 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -1.1%	
Baseline Corr 1st pt	NO _x = 793.0 ppb	NO = 796.2 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -0.7%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found NO _x r ² :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found NO r ² :		NO SI:	NO Int:	
						As found NO ₂ r ² :		NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1035

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.938	0.949	NO bkgnd or offset:	-28.3	-25.9
NOX coeff or slope:	0.935	0.941	NOX bkgnd or offset:	-28.0	-25.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	7.3	7.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999208	0.995141
NO _x Cal Offset:	-0.808436	-1.829828
NO Cal Slope:	1.003246	1.002849
NO Cal Offset:	-1.307881	-2.669053
NO ₂ Cal Slope:	0.987109	0.990662
NO ₂ Cal Offset:	-0.087128	-1.030634

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
High point	4933	66.6	803.3	800.6	2.7	798.3	801.7	-3.4	1.0062	0.9986
Mid point	4967	33.3	401.6	400.2	1.3	397.1	396.8	0.3	1.0113	1.0087
Low point	4983	16.6	200.2	199.5	0.7	195.7	195.3	0.4	1.0231	1.0218
As left zero	5000	0.0	0.0	0.0	0.0	0.5	0.7	-0.2	----	----
As left span	4933	66.6	803.3	345.5	457.8	787.5	345.5	442.0	1.0200	1.0000
Average Correction Factor									1.0135	1.0097

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	802.1	404.2	400.6	396.3	1.0108	98.9%
Mid GPT point	802.1	604.8	200.0	196.5	1.0176	98.3%
Low GPT point	802.1	701.7	103.1	100.2	1.0286	97.2%
Average Correction Factor					1.0190	98.1%

Notes:

Portable calibration system used. Zero and Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 24, 2025	Previous Calibration:	March 25, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:00	End Time (MST):	16:55
Analyzer make:	Teledyne API T200	Analyzer serial #:	1035

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999981	≥0.995
803.3	798.3	1.0062	Slope	0.995141	0.90 - 1.10
401.6	397.1	1.0113	Intercept	-1.829828	+/-20
200.2	195.7	1.0231			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 24, 2025	Previous Calibration:	March 25, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:00	End Time (MST):	16:55
Analyzer make:	Teledyne API T200	Analyzer serial #:	1035

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999974	≥0.995
400.6	396.3	1.0108	Slope	0.990662	0.90 - 1.10
200.0	196.5	1.0176	Intercept	-1.030634	+/-20
103.1	100.2	1.0286			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 24, 2025	Previous Calibration:	March 25, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:00	End Time (MST):	16:55
Analyzer make:	Teledyne API T200	Analyzer serial #:	1035

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999952	≥ 0.995
800.6	801.7	0.9986	Slope	1.002849	0.90 - 1.10
400.2	396.8	1.0087	Intercept	-2.669053	+/-20
199.5	195.3	1.0218			



NO_x Calibration Plot

Date: April 24, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Stony Mountain
Calibration Date: April 16, 2025
Start time (MST): 10:57
Reason: Routine

Station number: AMS 18
Last Cal Date: March 6, 2025
End time (MST): 14:38

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API 701H

Serial Number: 2658
Serial Number: 355

Analyzer Information

Analyzer make: API T400
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 825

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997629	0.997657	Backgd or Offset:	2.0	2.0
Calibration intercept:	0.840000	0.960000	Coeff or Slope:	1.027	1.012

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	NA	0.0	0.0	----
As found High point	4888	1138.1	400.0	407.0	0.983
As found Mid point					
As found Low point					
Baseline Corr As found:	407.0	Previous response	399.9	*% change	1.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	NA	0.0	0.4	----
High point	4888	1138.1	400.0	399.5	1.001
Mid point	4888	884.5	200.0	201.5	0.993
Low point	4888	741.4	100.0	100.8	0.992
As left zero	5000	NA	0.0	0.4	----
As left span	4812	1097.9	400.0	399.8	1.001
Average Correction Factor					0.995

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 16, 2025	Previous Calibration:	March 6, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:57	End Time (MST):	14:38
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999982	≥0.995
400.0	399.5	1.0013	Slope	0.997657	0.90 - 1.10
200.0	201.5	0.9926	Intercept	0.960000	+/- 5
100.0	100.8	0.9921			



O₃ Calibration Plot

Date: April 16, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: April 28, 2025 Last Cal Date: March 27, 2025
Start time (MST): 15:00 End time (MST): 15:13

Analyzer Make: API T640 S/N: 324
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	13.9	13.7	13.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	697.8	697.51	697.8	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	5.123	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	71	----	71	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 2.2		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.0	11	11.0	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 27, 2025
Date Disposable Filter Changed: December 20, 2024

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 4, 2024
Date RH/T Sensor Cleaned: July 4, 2024

Notes: Flow, temp and pressure checked. Leak check passed.

Calibration by: Mohammed Kashif



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: April 23, 2025 Last Cal Date: March 3, 2025
Start time (MST): 10:50 End time (MST): 13:37
Reason: Routine

Calibration Standards

Cal Gas Concentration: 3,080 ppm Cal Gas Exp Date: November 4, 2028
Cal Gas Cylinder #: EB0065608
Removed Cal Gas Conc: 3,080 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
ZAG Make/Model: Teledyne API T701H Serial Number: 355

Analyzer Information

Analyzer make: Teledyne API T300 Analyzer serial #: 3504
Analyzer Range: 0 - 50 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997355	1.002362	Backgd or Offset:	-0.012
Calibration intercept:	0.081757	0.201757	Coeff or Slope:	0.907

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4933	66.7	41.1	41.3	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	41.16	Prev response:	41.06	*% change:	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4933	66.7	41.1	41.3	0.994
Mid point	4966	33.3	20.5	20.9	0.982
Low point	4983	16.7	10.3	10.5	0.982
As left zero	5000	0.0	0.0	0.2	----
As left span	4933	66.7	41.1	41.5	0.991
Average Correction Factor					0.986

Notes: No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


CO Calibration Summary

Station Information

Calibration Date:	April 23, 2025	Previous Calibration:	March 3, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:50	End Time (MST):	13:37
Analyzer make:	Teledyne API T300	Analyzer serial #:	3504

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999974	≥ 0.995
41.1	41.3	0.9942	Slope	1.002362	$0.90 - 1.10$
20.5	20.9	0.9816	Intercept	0.201757	± 1.5
10.3	10.5	0.9817			



CO Calibration Plot

Date: April 23, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

CO₂ Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: April 22, 2025 Last Cal Date: March 27, 2025
Start time (MST): 10:46 End time (MST): 14:03
Reason: Routine

Calibration Standards

Cal Gas Concentration: 59,100 ppm Cal Gas Exp Date: November 4, 2028
Cal Gas Cylinder #: EB0065608
Removed Cal Gas Conc: 59,100 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
N2 Gen Make/Model: Peak Scientific Serial Number: 771048318

Analyzer Information

Analyzer make: API T360 Analyzer serial #: 489
Analyzer Range: 0 - 2,000 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998644	1.001298	Backgd or Offset:	-0.068	-0.068
Calibration intercept:	-1.640000	-2.520000	Coeff or Slope:	0.960	0.960

CO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	3000	0.0	0.0	0.0	----
As found High Point	2920	80.0	1576.0	1571.6	1.003
As found Mid Point					
As found Low Point					
New cylinder response					
Baseline Corr As found:	1571.6	Prev response:	1572.2	*% change:	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	3000	0.0	0.0	0.4	----
High point	2920	80.0	1576.0	1578.5	0.998
Mid point	2960	40.0	788.0	780.5	1.010
Low point	2980	20.0	394.0	392.1	1.005
As left zero	3000	0.0	0.0	0.0	----
As left span	2930	80.0	1570.8	1578.9	0.995
Average Correction Factor					1.004

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


CO₂ Calibration Summary

Station Information

Calibration Date	April 22, 2025	Previous Calibration	March 27, 2025
Station Name	Stony Mountain	Station Number	AMS 18
Start Time (MST)	10:46	End Time (MST)	14:03
Analyzer make	API T360	Analyzer serial #	489

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999961		≥0.995
1576.0	1578.5	0.9984	Slope	1.001298		0.90 - 1.10
788.0	780.5	1.0096	Intercept	-2.5		+/-20
394.0	392.1	1.0048				



CO₂ Calibration Plot

Date: April 22, 2025

Location: Stony Mountain





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS19
FIREBAG
APRIL 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Firebag Station number: AMS 19
Calibration Date: April 28, 2025 Last Cal Date: March 15, 2025
Start time (MST): 10:55 End time (MST): 14:09
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.97 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC705799
Removed Cal Gas Conc: 50.97 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 1607
Zero Air Gen Model: Teledyne API T701H Serial Number: 201

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1410661308
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002848	0.994511	Backgd or Offset:	10.9	10.8
Calibration intercept:	0.800000	0.940000	Coeff or Slope:	1.005	1.005

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	4999	0.0	0.0	0.3	----
As found High point	4922	78.4	799.2	796.3	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	796.0	Previous response	802.3	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	4999	0.0	0.0	0.4	----
High point	4922	78.4	799.2	795.5	1.005
Mid point	4961	39.2	399.6	398.6	1.003
Low point	4980	19.6	199.8	200.2	0.998
As left zero	4999	0.0	0.0	0.2	----
As left span	4922	78.4	799.2	797.6	1.002
Average Correction Factor:					1.002

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 28, 2025	Previous Calibration:	March 15, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:55	End Time (MST):	14:09
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999998	≥0.995
799.2	795.5	1.0047	Slope	0.994511	0.90 - 1.10
399.6	398.6	1.0025	Intercept	0.940000	+/-30
199.8	200.2	0.9980			



SO2 Calibration Plot

Date: April 28, 2025

Location: Firebag





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Firebag
Calibration Date: April 29, 2025
Start time (MST): 10:30
Reason: Routine

Station number: AMS 19
Last Cal Date: March 4, 2025
End time (MST): 14:54

Calibration Standards

Cal Gas Concentration: 5.29 ppm
Cal Gas Cylinder #: DT0010492
Removed Cal Gas Conc: 5.29 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Exp Date: March 19, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 1607
Serial Number: 201

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1151680032
Converter serial #: 2022-222
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001762	1.007763	Backgd or Offset:	3.07	3.09
Calibration intercept:	0.020000	-0.040000	Coeff or Slope:	1.214	1.214

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4924	75.6	80.0	81.7	0.977
As found Mid point	4962	37.8	40.0	40.9	0.973
As found Low point	4981	18.9	20.0	20.1	0.985
New cylinder response					
Baseline Corr As found:	81.9	Prev response:	80.15	*% change:	2.1%
Baseline Corr 2nd AF pt:	41.1	AF Slope:	1.024909	AF Intercept:	-0.240000
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999986	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4924	75.6	80.0	80.4	0.995
Mid point	4962	37.8	40.0	40.6	0.985
Low point	4981	18.9	20.0	20.1	0.995
As left zero	5000	0.0	0.0	-0.1	----
As left span	4924	75.6	80.0	79.8	1.002
SO2 Scrubber Check	4922	78.4	784.0	0.0	----
Date of last scrubber change:		18-Jan-23		Ave Corr Factor	0.992
Date of last converter efficiency test:		November 26, 2024		106.2% efficiency	

Notes: Changed sample inlet filter after as founds. No adjustments made. SOx scrubber check done after cal zero.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	April 29, 2025	Previous Calibration:	March 4, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:30	End Time (MST):	14:54
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999953		≥ 0.995
80.0	80.4	0.9948	Slope	1.007763		$0.90 - 1.10$
40.0	40.6	0.9850	Intercept	-0.040000		± 3
20.0	20.1	0.9948				



H2S Calibration Plot

Date: April 29, 2025

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Firebag
Calibration Date: April 28, 2025
Start time (MST): 10:55
Reason: Routine

Station number: AMS 19
Last Cal Date: March 15, 2025
End time (MST): 14:08

Calibration Standards

Gas Cert Reference: CC705799
CH4 Cal Gas Conc. 505.1 ppm
C3H8 Cal Gas Conc. 204.3 ppm
Removed Gas Cert:
Removed CH4 Conc. 505.1 ppm
Removed C3H8 Conc. 204.3 ppm
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Expiry Date: October 9, 2032
CH4 Equiv Conc. 1066.9 ppm
Removed Gas Expiry:
CH4 Equiv Conc. 1066.9 ppm
Diff between cyl:
Serial Number: 1607
Serial Number: 201

Analyzer Information

Analyzer make: Thermo 51i-LT
Analyzer Range: 0 - 20 ppm

Analyzer serial #: 1336160089

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992690	0.993468	Background:	1.99	2.04
Calibration intercept:	-0.023933	0.013867	Coefficient:	3.718	3.818

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	4999	0.0	0.00	0.03	----
As found High point	4922	78.4	16.73	16.25	1.031
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	16.22	Previous response	16.58	*% change	-2.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	4999	0.0	0.00	0.02	----
High point	4922	78.4	16.73	16.64	1.005
Mid point	4961	39.2	8.36	8.31	1.007
Low point	4980	19.6	4.18	4.17	1.002
As left zero	4999	0.0	0.00	0.04	----
As left span	4922	78.4	16.73	16.80	0.996
Average Correction Factor					1.005

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	April 28, 2025	Previous Calibration:	March 15, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:55	End Time (MST):	14:08
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.02	----	Correlation Coefficient	0.999998	≥ 0.995
16.73	16.64	1.0054	Slope	0.993468	$0.90 - 1.10$
8.36	8.31	1.0066	Intercept	0.013867	± 1.5
4.18	4.17	1.0023			


THC Calibration Curve



THC Calibration Plot

Date: April 28, 2025

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Firebag
Station number: AMS 19
Calibration Date: April 1, 2025
Last Cal Date: March 11, 2025
Start time (MST): 9:48
End time (MST): 14:43
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0044018
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.70 ppm
NO gas Diff:
Serial Number: 1607
Serial Number: 201

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
AF High point	4918	82.1	802.9	799.7	3.3	796.2	791.3	4.8	1.0085	1.0106
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 798.3 ppb	NO = 793.9 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = -0.3%			
Baseline Corr 1st pt	NO _x = 796.2 ppb	NO = 791.3 ppb	<u>As Found Statistics</u>			*Percent Change	NO = -0.3%			
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:	NO Int:			
			As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1410661309

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.905	0.916	NO bkgnd or offset:	4.5	4.6
NOX coeff or slope:	0.994	0.993	NOX bkgnd or offset:	4.5	4.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.5	160.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.993178	0.998046
NO _x Cal Offset:	0.800270	1.240021
NO Cal Slope:	0.992470	1.001402
NO Cal Offset:	0.300227	0.500049
NO ₂ Cal Slope:	1.001499	0.997186
NO ₂ Cal Offset:	-0.313391	-0.688529

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
High point	4918	82.1	802.9	799.7	3.3	802.0	801.0	1.4	1.0012	0.9983
Mid point	4959	41.1	402.0	400.3	1.6	403.0	401.6	1.4	0.9974	0.9968
Low point	4980	20.5	200.5	199.7	0.8	202.6	201.1	1.5	0.9896	0.9929
As left zero	5000	0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
As left span	4918	82.1	802.9	418.9	384.0	802.0	418.9	382.8	1.0012	1.0000
Average Correction Factor									0.9961	0.9960

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	798.9	418.2	384.0	382.6	1.0036	99.6%
Mid GPT point	798.9	609.1	193.1	191.5	1.0083	99.2%
Low GPT point	798.9	704.9	97.3	95.5	1.0187	98.2%
Average Correction Factor					1.0102	99.0%

Notes:

Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 1, 2025	Previous Calibration:	March 11, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:48	End Time (MST):	14:43
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999989	≥0.995
802.9	802.0	1.0012	Slope	0.998046	0.90 - 1.10
402.0	403.0	0.9974	Intercept	1.240021	+/-20
200.5	202.6	0.9896			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 1, 2025	Previous Calibration:	March 11, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:48	End Time (MST):	14:43
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999980	≥ 0.995
384.0	382.6	1.0036	Slope	0.997186	$0.90 - 1.10$
193.1	191.5	1.0083	Intercept	-0.688529	± 20
97.3	95.5	1.0187			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 1, 2025	Previous Calibration:	March 11, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:48	End Time (MST):	14:43
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999997	≥ 0.995
799.7	801.0	0.9983	Slope	1.001402	$0.90 - 1.10$
400.3	401.6	0.9968	Intercept	0.500049	± 20
199.7	201.1	0.9929			



NO_x Calibration Plot

Date: April 1, 2025

Location: Firebag





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS20 MACKAY RIVER APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: MacKay River Station number: AMS 20
Calibration Date: April 14, 2025 Last Cal Date: March 10, 2025
Start time (MST): 7:40 End time (MST): 10:26
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.15 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC409669
Removed Cal Gas Conc: 49.15 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 5706
Zero Air Gen Model: API 701 Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1501301450
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.009842	0.997962	Backgd or Offset:	20.5	19.9
Calibration intercept:	-0.254232	-0.196724	Coeff or Slope:	0.961	0.945

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4919	81.4	800.1	811.5	0.986
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	811.2	Previous response	807.7	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4919	81.4	800.1	798.6	1.002
Mid point	4959	40.7	400.1	398.6	1.004
Low point	4980	20.3	199.5	198.6	1.005
As left zero	5000	0.0	0.0	0.2	----
As left span	4919	81.4	800.1	799.9	1.000
Average Correction Factor:					1.003

Notes: No Maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 14, 2025	Previous Calibration:	March 10, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	7:40	End Time (MST):	10:26
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999998	≥0.995
800.1	798.6	1.0019	Slope	0.997962	0.90 - 1.10
400.1	398.6	1.0038	Intercept	-0.196724	+/-30
199.5	198.6	1.0047			



SO2 Calibration Plot

Date: April 14, 2025

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: MacKay River
Calibration Date: April 3, 2025
Start time (MST): 6:48
Reason: Routine

Station number: AMS 20
Last Cal Date: March 6, 2025
End time (MST): 10:56

Calibration Standards

Cal Gas Concentration: 5.12 ppm
Cal Gas Cylinder #: CC515997
Removed Cal Gas Conc: 5.12 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API 701

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 5706
Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1236656117
Converter serial #: 2022-226
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004456	0.999883	Backgd or Offset:	3.84	3.84
Calibration intercept:	-0.100522	-0.040619	Coeff or Slope:	1.086	1.086

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4922	78.1	80.0	80.7	0.990
As found Mid point	4961	39.0	39.9	40.3	0.989
As found Low point	4980	19.5	20.0	20.0	0.994
New cylinder response					
Baseline Corr As found:	80.8	Prev response:	80.23	*% change:	0.7%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.010744	AF Intercept:	-0.120418
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999998	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4922	78.1	80.0	79.9	1.001
Mid point	4961	39.0	39.9	39.9	1.001
Low point	4980	19.5	20.0	20.0	0.998
As left zero	5000	0.0	0.0	0.0	----
As left span	4922	78.1	80.0	79.7	1.003
SO2 Scrubber Check	4982	81.3	802.8	0.1	----
Date of last scrubber change:	25-May-23		Ave Corr Factor		1.000
Date of last converter efficiency test:					

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	April 3, 2025	Previous Calibration:	March 6, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	6:48	End Time (MST):	10:56
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999997	≥ 0.995
80.0	79.9	1.0009	Slope	0.999883	$0.90 - 1.10$
39.9	39.9	1.0009	Intercept	-0.040619	± 3
20.0	20.0	0.9985			


H₂S Calibration Curve



H₂S Calibration Plot

Date: April 3, 2025

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: MacKay River
Calibration Date: April 14, 2025
Start time (MST): 7:40
Reason: Routine

Station number: AMS 20
Last Cal Date: March 21, 2025
End time (MST): 10:25

Calibration Standards

Gas Cert Reference: CC409669
CH4 Cal Gas Conc. 505.1 ppm
C3H8 Cal Gas Conc. 206.4 ppm
Removed Gas Cert:
Removed CH4 Conc. 505.1 ppm
Removed C3H8 Conc. 206.4 ppm
Calibrator Make/Model: API T700
ZAG Make/Model: API 701

Cal Gas Expiry Date: October 9, 2032
CH4 Equiv Conc. 1072.7 ppm
Removed Gas Expiry:
CH4 Equiv Conc. 1072.7 ppm
Diff between cyl:
Serial Number: 5706
Serial Number: 4522

Analyzer Information

Analyzer make: Thermo 51i-LT
Analyzer Range: 0 - 20 ppm

Analyzer serial #: 1501663727

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999215	0.997991	Background:	3.140	2.990
Calibration intercept:	0.020404	0.016001	Coefficient:	4.900	4.871

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic-
					AFzero)
					<i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	-0.10	----
As found High point	4919	81.4	17.46	17.38	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.48	Previous response	17.47	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)
					<i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.03	----
High point	4919	81.4	17.46	17.46	1.000
Mid point	4959	40.7	8.73	8.71	1.003
Low point	4980	20.3	4.35	4.36	1.000
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	81.4	17.46	17.56	0.995
Average Correction Factor					1.001

Notes:

Zero and Span adjusted. No maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	April 14, 2025	Previous Calibration:	March 21, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	7:40	End Time (MST):	10:25
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.03	----	Correlation Coefficient	0.999993	≥ 0.995
17.46	17.46	1.0003	Slope	0.997991	$0.90 - 1.10$
8.73	8.71	1.0030	Intercept	0.016001	± 1.5
4.35	4.36	0.9997			


THC Calibration Curve



THC Calibration Plot

Date: April 14, 2025

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: MacKay River
Station number: AMS 20
Calibration Date: April 2, 2025
Last Cal Date: March 5, 2025
Start time (MST): 6:50
End time (MST): 10:55
Reason:

Calibration Standards

NO Gas Cylinder #: T376265
NOX Cal Gas Conc: 49.19 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 49.19 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: April 13, 2025
NO Cal Gas Conc: 48.04 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 48.04 ppm
NO gas Diff:
Serial Number: 5706
Serial Number: 4522

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
AF High point	4917	83.3	819.5	800.3	19.2	826.3	806.2	20.0	0.9915	0.9927
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 817.4 ppb	NO = 798.9 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = 1.1%			
Baseline Corr 1st pt	NO _x = 826.5 ppb	NO = 806.2 ppb	<u>As Found Statistics</u>			*Percent Change	NO = 0.9%			
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:	NO Int:			
			As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1505164379

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.285	1.002	NO bkgnd or offset:	2.8	2.8
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	3.0	3.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	159.4	159.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996966	1.005402
NO _x Cal Offset:	0.422044	0.721954
NO Cal Slope:	0.998643	1.006853
NO Cal Offset:	-0.357427	-0.357714
NO ₂ Cal Slope:	0.997985	0.997243
NO ₂ Cal Offset:	-1.492859	-1.552636

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.2	-0.2	----	----
High point	4917	83.3	819.5	800.3	19.2	824.3	805.7	18.5	0.9941	0.9933
Mid point	4958	41.7	410.3	400.7	9.6	413.4	402.8	10.5	0.9924	0.9947
Low point	4979	20.8	204.6	199.9	4.8	207.3	200.3	7.0	0.9872	0.9978
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.2	-0.2	----	----
As left span	4917	83.3	819.5	409.7	409.8	818.2	409.7	408.5	1.0015	1.0000
Average Correction Factor									0.9912	0.9953

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	799.6	407.9	410.9	408.9	1.0048	99.5%
Mid GPT point	799.6	601.2	217.6	214.9	1.0124	98.8%
Low GPT point	799.6	698.8	120.0	116.5	1.0297	97.1%
Average Correction Factor					1.0156	98.5%

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	March 5, 2025
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	6:50	End Time (MST):	10:55
Analyzer make:	Thermo 42i	6:50:00 AM	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999996	≥ 0.995
819.5	824.3	0.9941	Slope	1.005402	$0.90 - 1.10$
410.3	413.4	0.9924	Intercept	0.721954	± 20
204.6	207.3	0.9872			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	March 5, 2025
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	6:50	End Time (MST):	10:55
Analyzer make:	Thermo 42i	6:50:00 AM	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999943	≥0.995
410.9	408.9	1.0048	Slope	0.997243	0.90 - 1.10
217.6	214.9	1.0124	Intercept	-1.552636	+/-20
120.0	116.5	1.0297			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	March 5, 2025
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	6:50	End Time (MST):	10:55
Analyzer make:	Thermo 42i	6:50:00 AM	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999998	≥ 0.995
800.3	805.7	0.9933	Slope	1.006853	$0.90 - 1.10$
400.7	402.8	0.9947	Intercept	-0.357714	± 20
199.9	200.3	0.9978			



NO_x Calibration Plot

Date: April 2, 2025

Location: MacKay River





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS21
APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Conklin
Calibration Date: April 4, 2025
Start time (MST): 9:50
Reason: Routine

Station number: AMS 21
Last Cal Date: March 17, 2025
End time (MST): 13:00

Calibration Standards

Cal Gas Concentration: 50.34 ppm
Cal Gas Cylinder #: CC340840
Removed Cal Gas Conc: 50.34 ppm
Removed Gas Cyl #: NA
Calibrator Model: Teledyne API T700P
Zero Air Gen Model: Teledyne API T701

Cal Gas Exp Date: October 9, 2032
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 2659
Serial Number: 953

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 1428701363

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995774	1.000944	Backgd or Offset:	29.3	29.3
Calibration intercept:	-0.820824	-1.182060	Coeff or Slope:	0.899	0.899

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.5	----
As found High point	4921	79.5	800.3	801.0	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	800.5	Previous response	796.1	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.5	----
High point	4921	79.5	800.3	800.2	1.000
Mid point	4960	39.8	400.7	400.7	1.000
Low point	4980	19.9	200.4	196.6	1.019
As left zero	5000	0.0	0.0	0.6	----
As left span	4921	79.5	800.3	800.0	1.000
Average Correction Factor:					1.006

Notes: Sample inlet filter and H₂/N₂ was changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 4, 2025	Previous Calibration:	March 17, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:50	End Time (MST):	13:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999968	≥0.995
800.3	800.2	1.0002	Slope	1.000944	0.90 - 1.10
400.7	400.7	1.0001	Intercept	-1.182060	+/-30
200.4	196.6	1.0191			



SO2 Calibration Plot

Date: April 4, 2025

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Conklin
Calibration Date: April 9, 2025
Start time (MST): 9:33
Reason: Routine

Station number: AMS 21
Last Cal Date: March 6, 2025
End time (MST): 13:20

Calibration Standards

Cal Gas Concentration: 5.14 ppm
Cal Gas Cylinder #: CC501204
Removed Cal Gas Conc: 5.14 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne T700P
ZAG Make/Model: Teledyne T701

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 2659
Serial Number: 953

Analyzer Information

Analyzer make: Thermo 43i-QTL
Converter make: CD-Nova 101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12228021058
Converter serial #: 565
Converter Temp: 850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006000	1.000789	Backgd or Offset:	3.3	3.3
Calibration intercept:	-0.160000	-0.061610	Coeff or Slope:	1.545	1.571

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4922	78.4	80.6	80.2	1.004
As found Mid point	4961	39.2	40.3	40.1	1.002
As found Low point	4980	19.6	20.2	20.0	1.003
New cylinder response					
Baseline Corr As found:	80.3	Prev response:	80.91	*% change:	-0.8%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	0.996393	AF Intercept:	-0.081607
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	1.000000	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4922	78.4	80.6	80.6	1.000
Mid point	4961	39.2	40.3	40.3	1.000
Low point	4980	19.6	20.2	20.0	1.008
As left zero	5000	0.0	0.0	0.0	----
As left span	4922	78.4	80.6	81.1	0.994
SO2 Scrubber Check	4921	79.5	794.9	0.0	----
Date of last scrubber change:	November 13, 2024			Ave Corr Factor	1.002
Date of last converter efficiency test:					

Notes: Sample inlet filter was changed after multipoint as founds. SO2 scrubber check done after calibrator zero and passed. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	March 6, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:33	End Time (MST):	13:20
Analyzer make:	Thermo 43i-QTL	Analyzer serial #:	12228021058

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999996	≥ 0.995
80.6	80.6	0.9999	Slope	1.000789	$0.90 - 1.10$
40.3	40.3	0.9999	Intercept	-0.061610	± 3
20.2	20.0	1.0075			



TRS Calibration Plot

Date: April 9, 2025

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Conklin
 Calibration Date: April 4, 2025
 Start time (MST): 9:50
 Reason: Routine

Station number: AMS 21
 Last Cal Date: March 17, 2025
 End time (MST): 13:00

Calibration Standards

Gas Cert Reference:	CC340840	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	503.8 ppm	CH ₄ Equiv Conc.	1067.6 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	503.8 ppm	CH ₄ Equiv Conc.	1067.6 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700P	Serial Number:	2659
Zero Air Gen model:	Teledyne API T701	Serial Number:	953

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1180320039
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.30E-04	2.30E-04	NMHC SP Ratio:	4.73E-05
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	190954
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.5	16.97	17.05	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.05	Prev response	17.00	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.5	16.97	17.08	0.994
Mid point	4960	39.8	8.50	8.52	0.998
Low point	4980	19.9	4.25	4.19	1.015
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	16.97	17.10	0.993
Average Correction Factor					1.002

Notes: Sample inlet filter and H₂/N₂ was changed after as founds. No adjustment made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.5	8.96	9.00	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.00	Prev response	8.98	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.5	8.96	9.00	0.996
Mid point	4960	39.8	4.49	4.51	0.996
Low point	4980	19.9	2.24	2.23	1.008
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	8.96	9.04	0.992
Average Correction Factor					1.000

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.5	8.01	8.05	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.05	Prev response	8.03	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.5	8.01	8.07	0.993
Mid point	4960	39.8	4.01	4.01	1.000
Low point	4980	19.9	2.01	1.96	1.023
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	8.01	8.06	0.993
Average Correction Factor					1.005

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.004359	1.007955
THC Cal Offset:	-0.043713	-0.042932
CH ₄ Cal Slope:	1.005753	1.009706
CH ₄ Cal Offset:	-0.029221	-0.029831
NMHC Cal Slope:	1.003137	1.005229
NMHC Cal Offset:	-0.015091	-0.011299

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 4, 2025	Previous Calibration:	March 17, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:50	End Time (MST):	13:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999968	<i>≥0.995</i>
16.97	17.08	0.9937	Slope	1.007955	<i>0.90 - 1.10</i>
8.50	8.52	0.9975	Intercept	-0.042932	<i>+/-0.5</i>
4.25	4.19	1.0153			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 4, 2025	Previous Calibration:	March 17, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:50	End Time (MST):	13:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999935	<i>≥0.995</i>
8.01	8.07	0.9925	Slope	1.009706	<i>0.90 - 1.10</i>
4.01	4.01	0.9996	Intercept	-0.029831	<i>+/-0.5</i>
2.01	1.96	1.0230			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 4, 2025	Previous Calibration:	March 17, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:50	End Time (MST):	13:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999988	<i>≥0.995</i>
8.96	9.00	0.9959	Slope	1.005229	<i>0.90 - 1.10</i>
4.49	4.51	0.9959	Intercept	-0.011299	<i>+/-0.5</i>
2.24	2.23	1.0084			



NMHC Calibration Plot

Date: April 4, 2025

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Conklin
Station number: AMS 21
Calibration Date: April 24, 2025
Last Cal Date: March 18, 2025
Start time (MST): 9:45
End time (MST): 14:00
Reason: Routine

Calibration Standards

NO Gas Cylinder #: SA18828
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700P
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 2659
Serial Number: 953

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
AF High point	4918	82.0	802.0	800.3	1.6	790.5	787.2	3.3	1.0142	1.0163
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 801.3 ppb	NO = 800.1 ppb				* = > +/-5% change initiates investigation			*Percent Change	NO _x = -1.3%
Baseline Corr 1st pt	NO _x = 790.7 ppb	NO = 787.5 ppb				<u>As Found Statistics</u>			*Percent Change	NO = -1.6%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found NO _x r ² :			Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found NO r ² :			NO SI:	NO Int:
						As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1501663731

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.043	1.060	NO bkgnd or offset:	10.1	10.3
NOX coeff or slope:	0.995	0.995	NOX bkgnd or offset:	10.1	10.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	148.5	150.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000027	0.998189
NO _x Cal Offset:	-0.712016	-0.392024
NO Cal Slope:	1.002448	0.999592
NO Cal Offset:	-2.192041	-1.692057
NO ₂ Cal Slope:	1.001058	0.996573
NO ₂ Cal Offset:	-0.135051	-0.812981

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.1	----	----
High point	4918	82.0	802.0	800.3	1.6	799.7	798.5	1.2	1.0028	1.0023
Mid point	4959	41.0	401.0	400.2	0.8	401.4	399.3	2.2	0.9990	1.0022
Low point	4980	20.5	200.5	200.1	0.4	198.3	195.5	2.8	1.0109	1.0233
As left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
As left span	4918	82.0	802.0	403.8	398.2	793.4	403.8	389.6	1.0108	1.0000
Average Correction Factor									1.0042	1.0093

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	795.8	402.3	395.1	393.1	1.0052	99.5%
Mid GPT point	795.8	601.0	196.4	195.5	1.0048	99.5%
Low GPT point	795.8	698.1	99.3	96.6	1.0284	97.2%
Average Correction Factor					1.0128	98.7%

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 24, 2025	Previous Calibration:	March 18, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:45	End Time (MST):	14:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999987	≥0.995
802.0	799.7	1.0028	Slope	0.998189	0.90 - 1.10
401.0	401.4	0.9990	Intercept	-0.392024	+/-20
200.5	198.3	1.0109			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 24, 2025	Previous Calibration:	March 18, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:45	End Time (MST):	14:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999957	≥0.995
395.1	393.1	1.0052	Slope	0.996573	0.90 - 1.10
196.4	195.5	1.0048	Intercept	-0.812981	+/-20
99.3	96.6	1.0284			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 24, 2025	Previous Calibration:	March 18, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:45	End Time (MST):	14:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1501663731

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999968	≥ 0.995
800.3	798.5	1.0023	Slope	0.999592	$0.90 - 1.10$
400.2	399.3	1.0022	Intercept	-1.692057	± 20
200.1	195.5	1.0233			



NO_x Calibration Plot

Date: April 24, 2025

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Conklin
Calibration Date: April 14, 2025
Start time (MST): 11:14
Reason: Routine

Station number: AMS 21
Last Cal Date: March 3, 2025
End time (MST): 14:26

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700P
ZAG Make/Model: Teledyne API T701

Serial Number: 2659
Serial Number: 953

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1501663734

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003886	1.005686	Backgd or Offset:	1.9	2.0
Calibration intercept:	-0.480000	-0.120000	Coeff or Slope:	1.100	1.113

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	0.4	----
As found High point	5000	918.9	400.0	399.5	1.002
As found Mid point					
As found Low point					
Baseline Corr As found:	399.1	Previous response	401.1	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.4	----
High point	5000	921.1	400.0	402.4	0.994
Mid point	5000	757.7	200.0	200.8	0.996
Low point	5000	653.1	100.0	99.9	1.001
As left zero	5000	800.0	0.0	-0.6	----
As left span	5000	921.9	400.0	418.6	0.956
Average Correction Factor					0.997

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 14, 2025	Previous Calibration:	March 3, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	11:14	End Time (MST):	14:26
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999992	≥0.995
400.0	402.4	0.9940	Slope	1.005686	0.90 - 1.10
200.0	200.8	0.9960	Intercept	-0.120000	+/- 5
100.0	99.9	1.0010			



O₃ Calibration Plot

Date: April 14, 2025

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Conklin
Calibration Date: April 14, 2025
Start time (MST): 12:01
Station number: AMS 21
Last Cal Date: March 6, 2025
End time (MST): 12:17
Analyzer Make: API T640
Particulate Fraction: PM2.5
S/N: 326
Flow Meter Make/Model: Alicat FP-25BT
Temp/RH standard: Alicat FP-25BT
S/N: 388754
S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	13.00	12.97	13.00	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	711.10	713.50	711.10	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.08	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	38	----	38	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 3.50		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: July 16, 2026
Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: February 25, 2025
Date Disposable Filter Changed: February 25, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 9, 2024
Date RH/T Sensor Cleaned: August 9, 2024

Notes:

Verified flow, pressure, temperature and pump power. No adjustment needed. Leak check passed.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER
APRIL 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: April 11, 2025 Last Cal Date: March 26, 2025
Start time (MST): 12:00 End time (MST): 15:24
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.11 ppm Cal Gas Exp Date: January 18, 2029
Cal Gas Cylinder #: CC281519
Removed Cal Gas Conc: 50.11 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3806
Zero Air Gen Model: Teledyne API T701 Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1152430006
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999521	0.999692	Backgd or Offset:	26.1	26.4
Calibration intercept:	1.164268	0.304361	Coeff or Slope:	1.017	1.017

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	1.0	----
As found High point	4920	79.8	799.8	798.5	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.5	Previous response	800.6	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	79.8	799.8	799.8	1.000
Mid point	4960	39.9	399.9	400.1	0.999
Low point	4980	20.0	200.4	200.8	0.998
As left zero	5000	0.0	0.0	0.3	----
As left span	4920	79.8	799.8	797.3	1.003

Average Correction Factor: 0.999

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 26, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:00	End Time (MST):	15:24
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	1.000000	≥0.995
799.8	799.8	1.0000	Slope	0.999692	0.90 - 1.10
399.9	400.1	0.9995	Intercept	0.304361	+/-30
200.4	200.8	0.9982			



SO2 Calibration Plot

Date: April 11, 2025

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: April 29, 2025 Last Cal Date: March 27, 2025
Start time (MST): 11:50 End time (MST): 16:30
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.02 ppm Cal Gas Exp Date: November 15, 2026
Cal Gas Cylinder #: CC424047
Removed Cal Gas Conc: 5.02 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
ZAG Make/Model: Teledyne API T701 Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680031
Converter make: CDN-101 Converter serial #: 620
Analyzer Range: 0 - 100 ppb Converter Temp: 850 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>	
Calibration slope:	0.970959	0.999093	Backgd or Offset:	3.77	3.98
Calibration intercept:	0.300065	0.140635	Coeff or Slope:	1.197	1.251

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	79.7	80.0	77.5	1.033
As found Mid point	4960	39.8	40.0	38.9	1.027
As found Low point	4980	19.9	20.0	19.4	1.030
New cylinder response					
Baseline Corr As found:	77.5	Prev response:	78.00	*% change:	-0.6%
Baseline Corr 2nd AF pt:	38.9	AF Slope:	0.968533	AF Intercept:	0.059981
Baseline Corr 3rd AF pt:	19.4	AF Correlation:	0.999992	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4920	79.7	80.0	80.0	1.000
Mid point	4960	39.8	40.0	40.3	0.992
Low point	4980	19.9	20.0	20.0	0.999
As left zero	5000	0.0	0.0	0.4	----
As left span	4920	79.7	80.0	78.9	1.014
SO2 Scrubber Check	4920	79.8	798.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.997
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Scrubber test performed after zero point, no issues.
Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	April 29, 2025	Previous Calibration:	March 27, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:50	End Time (MST):	16:30
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999978		≥ 0.995
80.0	80.0	1.0003	Slope	0.999093		$0.90 - 1.10$
40.0	40.3	0.9916	Intercept	0.140635		± 3
20.0	20.0	0.9990				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Janvier
 Calibration Date: April 11, 2025
 Start time (MST): 12:00
 Reason: Routine

Station number: AMS 22
 Last Cal Date: March 26, 2025
 End time (MST): 15:24

Calibration Standards

Gas Cert Reference: CC281519
 CH₄ Cal Gas Conc. 502.8 ppm
 C₃H₈ Cal Gas Conc. 208.4 ppm
 Removed Gas Cert: NA
 Removed CH₄ Conc. 502.8 ppm
 Removed C₃H₈ Conc. 208.4 ppm
 Diff between cyl (CH₄):
 Calibrator Model: Teledyne API 700
 Zero Air Gen model: Teledyne API 701

Cal Gas Expiry Date: January 18, 2029
 CH₄ Equiv Conc. 1075.9 ppm
 Removed Gas Expiry: NA
 CH₄ Equiv Conc. 1075.9 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 3806
 Serial Number: 691

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1317958219
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.51E-04	2.54E-04	NMHC SP Ratio:	6.02E-05
CH ₄ Retention time:	11.6	11.6	NMHC Peak Area:	152054
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				149432
				OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	17.17	17.01	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.01	Prev response	17.15	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	17.17	17.20	0.999
Mid point	4960	39.9	8.59	8.51	1.009
Low point	4980	20.0	4.30	4.27	1.008
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	17.17	17.21	0.998
Average Correction Factor					1.005

Notes: Changed the inlet filter and H₂ cylinder after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	9.15	9.04	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.04	Prev response	9.14	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.8	9.15	9.15	1.000
Mid point	4960	39.9	4.57	4.55	1.006
Low point	4980	20.0	2.29	2.29	1.000
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.15	9.16	0.999
Average Correction Factor					1.002

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	8.03	7.97	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.97	Prev response	8.01	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.8	8.03	8.05	0.997
Mid point	4960	39.9	4.01	3.96	1.013
Low point	4980	20.0	2.01	1.98	1.018
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.03	8.06	0.996
Average Correction Factor					1.009

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.000628	1.001574
THC Cal Offset:	-0.035596	-0.034204
CH ₄ Cal Slope:	1.001452	1.003860
CH ₄ Cal Offset:	-0.029759	-0.028964
NMHC Cal Slope:	0.999880	0.999793
NMHC Cal Offset:	-0.006237	-0.005640

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 26, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:00	End Time (MST):	15:24
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999967	≥ 0.995
17.17	17.20	0.9987	Slope	1.001574	$0.90 - 1.10$
8.59	8.51	1.0090	Intercept	-0.034204	± 0.5
4.30	4.27	1.0083			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 26, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:00	End Time (MST):	15:24
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999919	≥ 0.995
8.03	8.05	0.9970	Slope	1.003860	$0.90 - 1.10$
4.01	3.96	1.0127	Intercept	-0.028964	± 0.5
2.01	1.98	1.0178			






Wood Buffalo Environmental Association NMHC Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 26, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:00	End Time (MST):	15:24
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999988	≥ 0.995
9.15	9.15	0.9999	Slope	0.999793	$0.90 - 1.10$
4.57	4.55	1.0058	Intercept	-0.005640	± 0.5
2.29	2.29	1.0002			



NMHC Calibration Plot

Date: April 11, 2025

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Janvier
Station number: AMS 22
Calibration Date: April 30, 2025
Last Cal Date: March 25, 2025
Start time (MST): 11:41
End time (MST): 16:17
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0047765
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: March 11, 2031
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3806
Serial Number: 691

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
AF High point	4918	82.0	802.0	800.3	1.6	771.4	764.6	6.7	1.0393	1.0464
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 802.5 ppb	NO = 799.9 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -4.0%	
Baseline Corr 1st pt	NO _x = 771.6 ppb	NO = 764.8 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -4.6%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1229254994

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.950	0.988	NO bkgnd or offset:	2.6	2.7
NOX coeff or slope:	0.996	0.998	NOX bkgnd or offset:	2.7	2.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	173.8	172.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001849	1.001321
NO _x Cal Offset:	-0.955969	0.204103
NO Cal Slope:	1.001603	1.000003
NO Cal Offset:	-1.676088	-0.716024
NO ₂ Cal Slope:	1.004645	1.004611
NO ₂ Cal Offset:	-1.110485	0.607262

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4918	82.0	802.0	800.3	1.6	803.1	799.9	3.1	0.9986	1.0005
Mid point	4960	41.0	400.9	400.1	0.8	401.8	399.2	2.5	0.9978	1.0022
Low point	4980	20.5	200.5	200.1	0.4	201.1	198.5	2.7	0.9969	1.0079
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	0.0	----	----
As left span	4918	82.0	802.0	398.1	403.9	800.8	398.1	402.7	1.0014	1.0000
Average Correction Factor									0.9977	1.0035

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	797.9	396.1	403.4	405.4	0.9952	100.5%
Mid GPT point	797.9	599.7	199.8	202.3	0.9878	101.2%
Low GPT point	797.9	697.8	101.7	103.0	0.9878	101.2%
Average Correction Factor					0.9903	101.0%

Notes: Inlet filter was changed after as founds. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 30, 2025	Previous Calibration:	March 25, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:41	End Time (MST):	16:17
Analyzer make:	Thermo 42i	Analyzer serial #:	1229254994

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	1.000000	≥0.995
802.0	803.1	0.9986	Slope	1.001321	0.90 - 1.10
400.9	401.8	0.9978	Intercept	0.204103	+/-20
200.5	201.1	0.9969			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 30, 2025	Previous Calibration:	March 25, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:41	End Time (MST):	16:17
Analyzer make:	Thermo 42i	Analyzer serial #:	1229254994

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999983	≥0.995
403.4	405.4	0.9952	Slope	1.004611	0.90 - 1.10
199.8	202.3	0.9878	Intercept	0.607262	+/-20
101.7	103.0	0.9878			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 30, 2025	Previous Calibration:	March 25, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:41	End Time (MST):	16:17
Analyzer make:	Thermo 42i	Analyzer serial #:	1229254994

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999996	≥ 0.995
800.3	799.9	1.0005	Slope	1.000003	$0.90 - 1.10$
400.1	399.2	1.0022	Intercept	-0.716024	± 20
200.1	198.5	1.0079			



NO_x Calibration Plot

Date: April 30, 2025

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Janvier
Calibration Date: April 22, 2025
Start time (MST): 9:51
Reason: Routine

Station number: AMS 22
Last Cal Date: March 18, 2025
End time (MST): 12:42

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Serial Number: 3806
Serial Number: 691

Analyzer Information

Analyzer make: Teledyne API T400
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 7046

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000486	0.996029	Backgd or Offset:	1.5	1.5
Calibration intercept:	1.040000	1.320000	Coeff or Slope:	1.011	1.011

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	0.3	----
As found High point	5000	922.9	400.0	399.0	1.003
As found Mid point					
As found Low point					
Baseline Corr As found:	398.7	Previous response	401.2	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.6	----
High point	5000	922.9	400.0	399.2	1.002
Mid point	5000	768.8	200.0	201.4	0.993
Low point	5000	656.1	100.0	101.3	0.987
As left zero	5000	800.0	0.0	0.6	----
As left span	5000	916.2	400.0	401.2	0.997
Average Correction Factor					0.994

Notes: Changed the inlet filter after the as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	April 22, 2025	Previous Calibration:	March 18, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	9:51	End Time (MST):	12:42
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999980	≥ 0.995
400.0	399.2	1.0020	Slope	0.996029	$0.90 - 1.10$
200.0	201.4	0.9930	Intercept	1.320000	± 5
100.0	101.3	0.9872			



O₃ Calibration Plot

Date: April 22, 2025

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: April 30, 2025 Last Cal Date: March 24, 2025
Start time (MST): 12:27 End time (MST): 14:19

Analyzer Make: Teledyne API T640 S/N: 325
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	12.3	11.51	12.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	713.7	714.94	713.7	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.04	4.896	5.04	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	----	37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 1.1		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: January 31, 2025
Date Disposable Filter Changed: January 31, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 29, 2024
Date RH/T Sensor Cleaned: October 29, 2024

Notes: Verified flow, temperature, and pressure. Leak check passed.

Calibration by: Rene Chamberland



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23
FORT HILLS
APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Fort Hills Station number: AMS 23
Calibration Date: April 17, 2025 Last Cal Date: March 17, 2025
Start time (MST): 6:30 End time (MST): 9:05
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.35 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC484463
Removed Cal Gas Conc: 50.35 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 451
Zero Air Gen Model: API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1160290012
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.988558	0.998705	Backgd or Offset:	18.9	18.9
Calibration intercept:	-0.121868	0.080120	Coeff or Slope:	1.071	1.071

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4921	79.4	799.5	797.8	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.6	Previous response	790.2	*% change	0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4921	79.4	799.5	798.5	1.001
Mid point	4960	39.7	399.8	399.7	1.000
Low point	4980	19.8	199.4	198.7	1.003
As left zero	5000	0.0	0.0	0.4	----
As left span	4921	79.4	799.5	798.5	1.001
Average Correction Factor:					1.002

Notes: Calibration gas changed out. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 17, 2025	Previous Calibration:	March 17, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:30	End Time (MST):	9:05
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995
799.5	798.5	1.0012	Slope	0.998705	0.90 - 1.10
399.8	399.7	1.0003	Intercept	0.080120	+/-30
199.4	198.7	1.0035			



SO2 Calibration Plot

Date: April 17, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort Hills
Calibration Date: April 16, 2025
Start time (MST): 6:32
Reason: Routine

Station number: AMS 23
Last Cal Date: March 11, 2025
End time (MST): 10:19

Calibration Standards

Cal Gas Concentration: 4.84 ppm
Cal Gas Cylinder #: DT0021910
Removed Cal Gas Conc: 4.84 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 451
Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1300156232
Converter serial #: 594
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001247	1.002388	Backgd or Offset:	2.06	2.06
Calibration intercept:	-0.038151	-0.078072	Coeff or Slope:	1.160	1.160

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4917	82.6	80.0	79.7	1.002
As found Mid point	4959	41.3	40.0	39.6	1.007
As found Low point	4979	20.7	20.0	19.4	1.028
New cylinder response					
Baseline Corr As found:	79.8	Prev response:	80.02	*% change:	-0.3%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.999812	AF Intercept:	-0.337937
Baseline Corr 3rd AF pt:	19.5	AF Correlation:	0.999956	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4917	82.6	80.0	80.1	0.998
Mid point	4959	41.3	40.0	40.0	0.999
Low point	4979	20.7	20.0	19.9	1.007
As left zero	5000	0.0	0.0	-0.1	----
As left span	4917	82.6	80.0	82.1	0.974
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.002
Date of last converter efficiency test:		March 13, 2024		110.3% efficiency	

Notes: SOx scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	April 16, 2025	Previous Calibration:	March 11, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:32	End Time (MST):	10:19
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999995		≥ 0.995
80.0	80.1	0.9983	Slope	1.002388		$0.90 - 1.10$
40.0	40.0	0.9994	Intercept	-0.078072		± 3
20.0	19.9	1.0070				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Fort Hills
 Calibration Date: April 17, 2025
 Start time (MST): 6:30
 Reason: Routine

Station number: AMS 23
 Last Cal Date: March 17, 2025
 End time (MST): 9:05

Calibration Standards

Gas Cert Reference:	CC484463	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	504.3 ppm	CH ₄ Equiv Conc.	1065.6 ppm
C ₃ H ₈ Cal Gas Conc.	204.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.3 ppm	CH ₄ Equiv Conc.	1065.6 ppm
Removed C ₃ H ₈ Conc.	204.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	451
Zero Air Gen model:	API T701	Serial Number:	1117

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 12227620777
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.55E-04	3.55E-04	NMHC SP Ratio:	5.42E-05
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	164497
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.4	16.92	17.00	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.00	Prev response	16.86	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.4	16.92	16.90	1.001
Mid point	4960	39.7	8.46	8.40	1.007
Low point	4980	19.8	4.22	4.21	1.003
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	16.92	16.85	1.004
Average Correction Factor					1.004

Notes: Calibration Gas changed out. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.4	8.91	9.00	0.990
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.00	Prev response	8.90	*% change	1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.4	8.91	8.95	0.996
Mid point	4960	39.7	4.46	4.47	0.997
Low point	4980	19.8	2.22	2.25	0.987
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	8.91	8.90	1.002
Average Correction Factor					0.994

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.4	8.01	7.99	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.99	Prev response	7.97	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.4	8.01	7.95	1.007
Mid point	4960	39.7	4.00	3.93	1.019
Low point	4980	19.8	2.00	1.96	1.021
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	8.01	7.95	1.007
Average Correction Factor					1.016

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.997066	0.998551
THC Cal Offset:	-0.009783	-0.012778
CH ₄ Cal Slope:	0.997435	0.993810
CH ₄ Cal Offset:	-0.021799	-0.020604
NMHC Cal Slope:	0.996952	1.003055
NMHC Cal Offset:	0.011416	0.006627

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 17, 2025	Previous Calibration:	March 17, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:30	End Time (MST):	9:05
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999989	≥ 0.995
16.92	16.90	1.0012	Slope	0.998551	$0.90 - 1.10$
8.46	8.40	1.0073	Intercept	-0.012778	± 0.5
4.22	4.21	1.0028			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 17, 2025	Previous Calibration:	March 17, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:30	End Time (MST):	9:05
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999954	≥ 0.995
8.01	7.95	1.0067	Slope	0.993810	$0.90 - 1.10$
4.00	3.93	1.0189	Intercept	-0.020604	± 0.5
2.00	1.96	1.0210			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	April 17, 2025	Previous Calibration:	March 17, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:30	End Time (MST):	9:05
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
8.91	8.95	0.9961	Slope	1.003055	<i>0.90 - 1.10</i>
4.46	4.47	0.9975	Intercept	0.006627	<i>+/-0.5</i>
2.22	2.25	0.9874			



NMHC Calibration Plot

Date: April 17, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Hills
 Station number: AMS 23
 Calibration Date: April 15, 2025
 Last Cal Date: March 12, 2025
 Start time (MST): 6:50
 End time (MST): 11:33
 Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC358149
 NOX Cal Gas Conc: 60.30 ppm
 Removed Cylinder #:
 Removed Gas NOX Conc: 60.30 ppm
 NOX gas Diff:
 Calibrator Model: API T700
 ZAG make/model: API T701
 Cal Gas Expiry Date: January 5, 2032
 NO Cal Gas Conc: 60.10 ppm
 Removed Gas Exp Date:
 Removed Gas NO Conc: 60.10 ppm
 NO gas Diff:
 Serial Number: 451
 Serial Number: 1117

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.4	0.0	----	----
AF High point	4934	66.3	799.5	796.9	2.7	775.0	770.9	4.0	1.0311	1.0332
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 794.8 ppb	NO = 793.8 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = -2.5%			
Baseline Corr 1st pt	NO _x = 775.4 ppb	NO = 771.3 ppb	<u>As Found Statistics</u>			*Percent Change	NO = -2.9%			
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:	NO Int:			
			As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430007

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.940	0.970	NO bkgnd or offset:	2.8	2.5
NOX coeff or slope:	0.990	0.990	NOX bkgnd or offset:	3.0	2.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	146.3	146.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995194	0.999841
NO _x Cal Offset:	-0.872416	0.026011
NO Cal Slope:	0.999005	1.002405
NO Cal Offset:	-2.271408	-1.432704
NO ₂ Cal Slope:	0.998438	0.997503
NO ₂ Cal Offset:	-1.622293	-1.163127

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	0.0	----	----
High point	4934	66.3	799.5	796.9	2.7	799.4	798.2	1.2	1.0002	0.9983
Mid point	4967	33.2	400.4	399.0	1.3	400.5	397.7	2.7	0.9997	1.0034
Low point	4983	16.6	200.2	199.5	0.7	200.0	197.0	3.0	1.0011	1.0129
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
As left span	4934	66.3	799.5	410.8	388.7	796.8	410.8	386.0	1.0034	1.0000
Average Correction Factor									1.0003	1.0049

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	795.0	411.1	386.6	384.9	1.0043	99.6%
Mid GPT point	795.0	603.0	194.7	192.8	1.0096	99.0%
Low GPT point	795.0	696.9	100.8	97.9	1.0291	97.2%
Average Correction Factor					1.0143	98.6%

Notes: Zero and Span Adjusted. During second point the NO_x and NO were spiking down, Blow out inside of analyzer. Continued Calibration.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 15, 2025	Previous Calibration:	March 12, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:50	End Time (MST):	11:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	1.000000	≥0.995
799.5	799.4	1.0002	Slope	0.999841	0.90 - 1.10
400.4	400.5	0.9997	Intercept	0.026011	+/-20
200.2	200.0	1.0011			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 15, 2025	Previous Calibration:	March 12, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:50	End Time (MST):	11:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999954	≥ 0.995
386.6	384.9	1.0043	Slope	0.997503	$0.90 - 1.10$
194.7	192.8	1.0096	Intercept	-1.163127	± 20
100.8	97.9	1.0291			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 15, 2025	Previous Calibration:	March 12, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:50	End Time (MST):	11:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999981	≥ 0.995
796.9	798.2	0.9983	Slope	1.002405	$0.90 - 1.10$
399.0	397.7	1.0034	Intercept	-1.432704	± 20
199.5	197.0	1.0129			



NO_x Calibration Plot

Date: April 15, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Hills Station number: AMS 23
Calibration Date: April 17, 2025 Last Cal Date: March 17, 2025
Start time (MST): 7:00 End time (MST): 7:52

Analyzer Make: API T640 S/N: 320
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388744
Temp/RH standard: Alicat FP-25BT S/N: 388744

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-3.8	-4.1	-3.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	740.4	739.4	740.4	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.98	4.96	4.98	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	55	----	55	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	4.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 16-Jul-26
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	8	11.2	11.2	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 17, 2025
Date Disposable Filter Changed: April 17, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: April 17, 2025
Date RH/T Sensor Cleaned: April 17, 2025

Notes: No adjustments done. Leak Check, Flow and PMT checked before and after cleaning.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25 WASKŌW OHCI PIMÂTISIWIN APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS 25
Calibration Date: April 9, 2025 Last Cal Date: March 25, 2025
Start time (MST): 7:04 End time (MST): 9:50
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.70 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC342445
Removed Cal Gas Conc: 49.70 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 621
Zero Air Gen Model: API T701 Serial Number: 4765

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1118148497
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003113	1.007511	Backgd or Offset:	11.3	11.3
Calibration intercept:	-0.352267	0.109107	Coeff or Slope:	1.065	1.065

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4920	80.5	800.1	802.9	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	802.8	Previous response	802.2	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	80.5	800.1	806.1	0.993
Mid point	4960	40.2	399.6	403.1	0.991
Low point	4980	20.1	199.8	201.0	0.994
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	80.5	800.1	807.2	0.991
Average Correction Factor:					0.993

Notes: No Maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 9, 2025	Previous Calibration:	March 25, 2025
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS 25
Start Time (MST):	7:04	End Time (MST):	9:50
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999999	≥0.995
800.1	806.1	0.9925	Slope	1.007511	0.90 - 1.10
399.6	403.1	0.9912	Intercept	0.109107	+/-30
199.8	201.0	0.9940			



SO2 Calibration Plot

Date: April 9, 2025

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS 25
Calibration Date: April 23, 2025 Last Cal Date: March 24, 2025
Start time (MST): 7:10 End time (MST): 11:07
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC517099
Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 747
ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
Converter make: Global G-150 Converter serial #: 2022-219
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011498	1.007473	Backgd or Offset:	3.50
Calibration intercept:	-0.140000	0.000000	Coeff or Slope:	1.108

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4920	80.0	79.5	82.8	0.959
As found Mid point	4960	40.0	39.8	41.4	0.958
As found Low point	4980	20.0	19.9	20.6	0.960
New cylinder response					
Baseline Corr As found:	82.9	Prev response:	80.29	*% change:	3.1%
Baseline Corr 2nd AF pt:	41.5	AF Slope:	1.042685	AF Intercept:	-0.100000
Baseline Corr 3rd AF pt:	20.7	AF Correlation:	0.999999	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.0	79.5	80.1	0.993
Mid point	4960	40.0	39.8	40.1	0.992
Low point	4980	20.0	19.9	20.0	0.994
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	80.0	800.0	811.7	0.986
SO ₂ Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.993
Date of last converter efficiency test:	February 12, 2025			111.0%	efficiency

Notes: SO_x Scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 23, 2025	Previous Calibration:	March 24, 2025
Station Name:	Waskow ohci Pimatisiwin	Station Number:	10:47:00 AM
Start Time (MST):	7:10	End Time (MST):	11:07
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999999	≥ 0.995
79.5	80.1	0.9928	Slope	1.007473	$0.90 - 1.10$
39.8	40.1	0.9915	Intercept	0.000000	± 3
19.9	20.0	0.9940			



H₂S Calibration Plot

Date: April 23, 2025

Location: Waskow ohci Pimatisiwin





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27
JACKFISH 2/3
APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Jackfish 2/3 Station number: AMS 27
Calibration Date: April 11, 2025 Last Cal Date: March 13, 2025
Start time (MST): 10:47 End time (MST): 14:04
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.58 ppm Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #: SG9133974BAL
Removed Cal Gas Conc: 50.58 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 5252
Zero Air Gen Model: API 701 Serial Number: 268

Analyzer Information

Analyzer make: Thermo 43iQ-TL Serial Number: 12124313138
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997385	1.010079	Backgd or Offset:	8.5	8.4
Calibration intercept:	-0.546041	0.842963	Coeff or Slope:	0.947	0.947

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	1.2	----
As found High point	4913	78.9	799.4	808.6	0.990
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	807.4	Previous response	796.8	*% change	1.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	1.3	----
High point	4913	78.9	799.4	808.8	0.988
Mid point	4955	39.5	400.0	404.0	0.990
Low point	4971	19.7	199.7	202.5	0.986
As left zero	5000	0.0	0.0	1.3	----
As left span	4913	78.9	799.4	807.3	0.990
Average Correction Factor:					0.988

Notes: Changed the sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 13, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	10:47	End Time (MST):	14:04
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.3	----	Correlation Coefficient	0.999997	≥0.995
799.4	808.8	0.9884	Slope	1.010079	0.90 - 1.10
400.0	404.0	0.9902	Intercept	0.842963	+/-30
199.7	202.5	0.9860			



SO2 Calibration Plot

Date: April 11, 2025

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Jackfish 2/3
Calibration Date: April 23, 2025
Start time (MST): 11:48
Reason: Routine

Station number: AMS 27
Last Cal Date: March 19, 2025
End time (MST): 18:33

Calibration Standards

Cal Gas Concentration: 4.87 ppm
Cal Gas Cylinder #: CC523090
Removed Cal Gas Conc: 4.87 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API T701H

Cal Gas Exp Date: September 5, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 5252
Serial Number: 268

Analyzer Information

Analyzer make: Thermo 43iQ
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12228021055
Converter serial #: 2022-195
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011680	1.004231	Backgd or Offset:	3.6
Calibration intercept:	-0.227800	0.159197	Coeff or Slope:	1.136

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4911	82.0	80.0	81.1	0.983
As found Mid point	4951	41.0	40.0	40.1	0.990
As found Low point	4973	20.5	20.0	19.7	1.000
New cylinder response					
Baseline Corr As found:	81.4	Prev response:	80.69	*% change:	0.9%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.018806	AF Intercept:	-0.500808
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999971	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4911	82.0	80.0	80.5	0.994
Mid point	4951	41.0	40.0	40.3	0.993
Low point	4973	20.5	20.0	20.2	0.990
As left zero	5000	0.0	0.0	0.2	----
As left span	4911	82.0	80.0	79.2	1.010
SO2 Scrubber Check	4915	78.9	790.0	-0.2	----
Date of last scrubber change:	21-Feb-25		Ave Corr Factor		0.992
Date of last converter efficiency test:	April 23, 2025		91.4% efficiency		

Notes: Performed a converter efficiency test. Changed sample inlet filter after as founds. Ran scrubber check after calibrator zero and it passed. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	April 23, 2025	Previous Calibration:	March 19, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:48	End Time (MST):	18:33
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12228021055

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999999	≥ 0.995
80.0	80.5	0.9935	Slope	1.004231	$0.90 - 1.10$
40.0	40.3	0.9925	Intercept	0.159197	± 3
20.0	20.2	0.9898			



H₂S Calibration Plot

Date: April 23, 2025

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Jackfish 2/3
Station number: AMS 27
Calibration Date: April 14, 2025
Last Cal Date: March 18, 2025
Start time (MST): 11:12
End time (MST): 16:28
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC757838
NOX Cal Gas Conc: 60.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.30 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 60.20 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.20 ppm
NO gas Diff:
Serial Number: 5252
Serial Number: 268

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.4	0.5	-0.1	----	----
AF High point	4924	66.3	801.1	799.8	1.3	810.2	804.9	5.4	0.9893	0.9943
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 799.3 ppb	NO = 796.9 ppb							*Percent Change	NO _x = 1.3%
Baseline Corr 1st pt	NO _x = 809.8 ppb	NO = 804.4 ppb							*Percent Change	NO = 0.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb								
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb								
As Found Statistics										
	As found	NO _x r ² :							Nx SI:	Nx Int:
	As found	NO r ² :							NO SI:	NO Int:
	As found	NO ₂ r ² :							NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153357

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.279	1.272	NO bkgnd or offset:	4.3	4.3
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	4.4	4.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	158.1	159.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000532	0.996775
NO _x Cal Offset:	-2.295044	1.250382
NO Cal Slope:	1.001195	0.999158
NO Cal Offset:	-3.895489	-0.729210
NO ₂ Cal Slope:	1.000485	0.999829
NO ₂ Cal Offset:	-0.127047	0.510255

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	1.2	0.6	0.5	----	----
High point	4924	66.3	801.1	799.8	1.3	799.8	799.0	0.8	1.0017	1.0010
Mid point	4958	33.2	401.1	400.4	0.7	401.0	398.9	2.1	1.0002	1.0038
Low point	4976	16.6	200.5	200.2	0.3	201.2	197.8	3.4	0.9965	1.0119
As left zero	5000	0.0	0.0	0.0	0.0	2.9	0.8	2.1	----	----
As left span	4924	66.3	801.1	374.4	426.7	796.8	374.4	422.4	1.0054	1.0000
Average Correction Factor									0.9995	1.0056

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.5	----	----
High GPT point	793.4	373.0	421.7	422.2	0.9989	100.1%
Mid GPT point	793.4	577.3	217.4	217.8	0.9983	100.2%
Low GPT point	793.4	691.7	103.0	103.6	0.9945	100.6%
Average Correction Factor					0.9972	100.3%

Notes:

Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 14, 2025	Previous Calibration:	March 18, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:12	End Time (MST):	16:28
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.2	----	Correlation Coefficient	1.000000	≥0.995
801.1	799.8	1.0017	Slope	0.996775	0.90 - 1.10
401.1	401.0	1.0002	Intercept	1.250382	+/-20
200.5	201.2	0.9965			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 14, 2025	Previous Calibration:	March 18, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:12	End Time (MST):	16:28
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	1.000000	≥0.995
421.7	422.2	0.9989	Slope	0.999829	0.90 - 1.10
217.4	217.8	0.9983	Intercept	0.510255	+/-20
103.0	103.6	0.9945			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 14, 2025	Previous Calibration:	March 18, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:12	End Time (MST):	16:28
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999987	≥ 0.995
799.8	799.0	1.0010	Slope	0.999158	$0.90 - 1.10$
400.4	398.9	1.0038	Intercept	-0.729210	± 20
200.2	197.8	1.0119			



NO_x Calibration Plot

Date: April 14, 2025

Location: Jackfish 2/3





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS29
SURMONT 2
APRIL 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: April 14, 2025 Last Cal Date: March 17, 2025
Start time (MST): 10:28 End time (MST): 14:14
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.95 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC356229
Removed Cal Gas Conc: 49.95 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 5472
Zero Air Gen Model: Teledyne API T701 Serial Number: 4428

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1170050150
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000431	1.006144	Backgd or Offset:	14.5	14.3
Calibration intercept:	-1.000571	-1.799656	Coeff or Slope:	0.962	0.956

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4919.9	80.1	800.2	806.0	0.992
As found Mid point					
As found Low point					
New cylinder response	4920	80.1	800.2		
Baseline Corr As found:	806.3	Previous response	799.5	*% change	0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	80.1	800.2	804.0	0.995
Mid point	4960	40.0	399.6	400.2	0.999
Low point	4980	20.0	199.8	196.6	1.016
As left zero	5000	0.0	0.0	-0.2	----
As left span	4920	80.1	800.2	798.1	1.003
Average Correction Factor:					1.003

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 14, 2025	Previous Calibration:	March 17, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:28	End Time (MST):	14:14
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.2	----	Correlation Coefficient	0.999968	≥0.995
800.2	804.0	0.9953	Slope	1.006144	0.90 - 1.10
399.6	400.2	0.9985	Intercept	-1.799656	+/-30
199.8	196.6	1.0163			



SO2 Calibration Plot

Date: April 14, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: April 8, 2025 Last Cal Date: March 26, 2025
Start time (MST): 9:59 End time (MST): 15:34
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.750 ppm Cal Gas Exp Date: August 28, 2027
Cal Gas Cylinder #: CC737848
Removed Cal Gas Conc: 4.750 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 5472
ZAG Make/Model: Teledyne API T701 Serial Number: 4428

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 1200326170
Converter make: Global Converter serial #: 2022-220
Analyzer Range: 0 - 100 ppb Converter Temp: 325.0 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006025	1.003882	Backgd or Offset:	0.94
Calibration intercept:	-0.080482	-0.180480	Coeff or Slope:	1.039

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4916	84.2	80.0	81.7	0.978
As found Mid point	4958	42.1	40.0	40.5	0.985
As found Low point	4979	21.1	20.0	20.4	0.975
New cylinder response					
Baseline Corr As found:	81.8	Prev response:	80.39	*% change:	1.7%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.022028	AF Intercept:	-0.140489
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999979	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4916	84.2	80.0	80.2	0.997
Mid point	4958	42.1	40.0	39.8	1.005
Low point	4979	21.1	20.0	19.9	1.005
As left zero	5000	0.0	0.0	0.1	----
As left span	4916	84.2	80.0	79.1	1.011
SO2 Scrubber Check	4919	81.3	813.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	1.002
Date of last converter efficiency test:		December 5, 2024		108.1%	efficiency

Notes: Changed sample inlet filter after as founds. Ran SOx scrubber check after cal zero, passed. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	April 8, 2025	Previous Calibration:	March 26, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:59	End Time (MST):	15:34
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999988		≥ 0.995
80.0	80.2	0.9973	Slope	1.003882		$0.90 - 1.10$
40.0	39.8	1.0049	Intercept	-0.180480		± 3
20.0	19.9	1.0049				



H2S Calibration Plot

Date: April 8, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: April 14, 2025 Last Cal Date: March 17, 2025
Start time (MST): 10:28 End time (MST): 14:14
Reason: Routine

Calibration Standards

Gas Cert Reference: CC356229 Cal Gas Expiry Date: October 9, 2032
CH4 Cal Gas Conc. 503.7 ppm CH4 Equiv Conc. 1066.9 ppm
C3H8 Cal Gas Conc. 204.8 ppm
Removed Gas Cert: NA Removed Gas Expiry: NA
Removed CH4 Conc. 503.7 ppm CH4 Equiv Conc. 1066.9 ppm
Removed C3H8 Conc. 204.8 ppm Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 5472
ZAG Make/Model: Teledyne API T701 Serial Number: 4428

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1170050149
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999635	1.001408	Background:	3.45	3.54
Calibration intercept:	-0.017156	-0.029160	Coefficient:	3.890	3.872

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.07	----
As found High point	4920	80.1	17.09	17.21	0.997
As found Mid point					
As found Low point					
New cylinder response	4920	80.1	17.09		
Baseline Corr As found:	17.14	Previous response	17.07	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.03	----
High point	4920	80.1	17.09	17.12	0.998
Mid point	4960	40.0	8.54	8.48	1.007
Low point	4980	20.0	4.27	4.19	1.019
As left zero	5000	0.0	0.00	-0.02	----
As left span	4920	80.1	17.09	17.14	0.997

Average Correction Factor 1.008

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	April 14, 2025	Previous Calibration:	March 17, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:28	End Time (MST):	14:14
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.03	----	Correlation Coefficient	0.999939	≥ 0.995
17.09	17.12	0.9983	Slope	1.001408	$0.90 - 1.10$
8.54	8.48	1.0065	Intercept	-0.029160	± 1.5
4.27	4.19	1.0193			


THC Calibration Curve



THC Calibration Plot

Date: April 14, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1170050148

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.001	1.001	NO bkgnd or offset:	1.2	1.2
NOX coeff or slope:	0.989	0.989	NOX bkgnd or offset:	1.2	1.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	150.2	148.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999329	0.997004
NO _x Cal Offset:	-0.449184	-0.809963
NO Cal Slope:	0.999498	0.994903
NO Cal Offset:	-1.708431	-2.049505
NO ₂ Cal Slope:	0.995539	1.002832
NO ₂ Cal Offset:	-0.464424	0.553598

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
High point	4933	66.7	800.4	800.4	0.0	797.5	795.0	2.6	1.0037	1.0069
Mid point	4967	33.3	399.6	399.6	0.0	397.7	395.6	2.1	1.0047	1.0101
Low point	4983	16.7	200.4	200.4	0.0	197.7	194.4	3.4	1.0137	1.0309
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
As left span	4933	66.7	800.4	403.2	397.2	791.4	403.2	388.2	1.0114	1.0000
Average Correction Factor									1.0074	1.0159

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	788.8	401.8	387.0	388.5	0.9961	100.4%
Mid GPT point	788.8	602.8	186.0	187.0	0.9947	100.5%
Low GPT point	788.8	695.9	92.9	94.5	0.9831	101.7%
Average Correction Factor					0.9913	100.9%

Notes:

Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 25, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:20	End Time (MST):	14:32
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999993	≥0.995
800.4	797.5	1.0037	Slope	0.997004	0.90 - 1.10
399.6	397.7	1.0047	Intercept	-0.809963	+/-20
200.4	197.7	1.0137			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 25, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:20	End Time (MST):	14:32
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999989	≥0.995
387.0	388.5	0.9961	Slope	1.002832	0.90 - 1.10
186.0	187.0	0.9947	Intercept	0.553598	+/-20
92.9	94.5	0.9831			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 25, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:20	End Time (MST):	14:32
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999960	≥ 0.995
800.4	795.0	1.0069	Slope	0.994903	$0.90 - 1.10$
399.6	395.6	1.0101	Intercept	-2.049505	± 20
200.4	194.4	1.0309			



NO_x Calibration Plot

Date: April 7, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: April 14, 2025 Last Cal Date: March 24, 2025
Start time (MST): 10:36 End time (MST): 13:57

Analyzer Make: API T640 S/N: 323
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.4	10.17	10.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	715.5	714.43	715.5	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.931	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	----	37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 2.7		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.8			<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 14, 2025
Date Disposable Filter Changed: April 14, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: April 14, 2025
Date RH/T Sensor Cleaned: October 30, 2024

Notes: Verified temperature, pressure and flow. Leak check passed. After cleaning optical chamber, analyzer can no longer run the PMT peak test. Will replace tomorrow.

Calibration by: Braiden Boutilier



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: April 15, 2025 Last Cal Date: April 14, 2025
Start time (MST): 10:26 End time (MST): 12:07

Analyzer Make: API T640 S/N: 2236
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.7	11.21	10.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	707.3	705.61	707.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.97	5.041	4.97	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	33	----	33	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 2.7		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test			11.2	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 15, 2025
Date Disposable Filter Changed: April 15, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: April 14, 2025
Date RH/T Sensor Cleaned: October 30, 2024

Notes: Installed T640. Verified pressure, temperature, and flow. Conducted PMT peak test, adjusted and passed.

Calibration by: Braiden Boutilier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS30
ELLS RIVER
APRIL 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Ells River Station number: AMS 30
Calibration Date: April 11, 2025 Last Cal Date: March 11, 2025
Start time (MST): 9:25 End time (MST): 12:33
Reason: Routine

Calibration Standards

Cal Gas Concentration: 48.75 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC350110
Removed Cal Gas Conc: 48.75 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3061
Zero Air Gen Model: API T701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1008841397
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002175	1.000403	Backgd or Offset:	10.1	10.1
Calibration intercept:	-3.212042	-2.892063	Coeff or Slope:	0.991	0.991

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4918	82.0	799.5	795.6	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	795.9	Previous response	798.0	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4918	82.0	799.5	798.2	1.002
Mid point	4959	41.0	399.8	396.0	1.009
Low point	4980	20.5	199.9	194.0	1.030
As left zero	5000	0.0	0.0	0.1	----
As left span	4918	82.0	799.5	798.9	1.001
Average Correction Factor:					1.014

Notes: Sample inlet filter and N2 cylinder was changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 11, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:25	End Time (MST):	12:33
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999944	≥0.995
799.5	798.2	1.0016	Slope	1.000403	0.90 - 1.10
399.8	396.0	1.0095	Intercept	-2.892063	+/-30
199.9	194.0	1.0302			



SO2 Calibration Plot

Date: April 11, 2025

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Ells River Station number: AMS 30
Calibration Date: April 1, 2025 Last Cal Date: March 14, 2025
Start time (MST): 8:38 End time (MST): 12:27
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.99 ppm Cal Gas Exp Date: November 15, 2026
Cal Gas Cylinder #: CC505806
Removed Cal Gas Conc: 4.99 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3061
ZAG Make/Model: API 701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1410661331
Converter make: CDN- 101 Converter serial #: 562
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007901	0.999336	Backgd or Offset:	1.7	1.7
Calibration intercept:	-0.180432	-0.180614	Coeff or Slope:	1.080	1.080

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4920	80.2	80.0	80.0	0.999
As found Mid point	4960	40.1	40.0	39.6	1.008
As found Low point	4980	20.0	20.0	19.5	1.018
New cylinder response					
Baseline Corr As found:	80.1	Prev response:	80.49	*% change:	-0.5%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	1.001906	AF Intercept:	-0.320569
Baseline Corr 3rd AF pt:	19.6	AF Correlation:	0.999964	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.2	80.0	79.9	1.002
Mid point	4960	40.1	40.0	39.7	1.008
Low point	4980	20.0	20.0	19.6	1.018
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	80.2	80.0	79.9	1.002
SO2 Scrubber Check	4918	82.0	820.0	0.0	----
Date of last scrubber change:	14-Mar-25		Ave Corr Factor		1.009
Date of last converter efficiency test:					

Notes: Changed sample inlet filter after multipoint as founds. SO2 scrubber check done after calibrator zero and passed. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	April 1, 2025	Previous Calibration:	March 14, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	8:38	End Time (MST):	12:27
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999977		≥ 0.995
80.0	79.9	1.0017	Slope	0.999336		$0.90 - 1.10$
40.0	39.7	1.0080	Intercept	-0.180614		± 3
20.0	19.6	1.0184				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Ells River
 Calibration Date: April 11, 2025
 Start time (MST): 9:25
 Reason: Routine

Station number: AMS 30
 Last Cal Date: March 10, 2025
 End time (MST): 12:33

Calibration Standards

Gas Cert Reference: CC350110
 CH₄ Cal Gas Conc. 496.6 ppm
 C₃H₈ Cal Gas Conc. 207.2 ppm
 Removed Gas Cert: NA
 Removed CH₄ Conc. 496.6 ppm
 Removed C₃H₈ Conc. 207.2 ppm
 Diff between cyl (CH₄):
 Calibrator Model: API T700
 Zero Air Gen model: API T701H

Cal Gas Expiry Date: March 10, 2031
 CH₄ Equiv Conc. 1066.4 ppm
 Removed Gas Expiry: NA
 CH₄ Equiv Conc. 1066.4 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 3061
 Serial Number: 358

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1152430011
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.11E-04	3.11E-04	NMHC SP Ratio:	5.96E-05
CH ₄ Retention time:	17.4	17.4	NMHC Peak Area:	156612
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	17.49	17.40	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.40	Prev response	17.45	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	82.0	17.49	17.44	1.003
Mid point	4959	41.0	8.74	8.63	1.013
Low point	4980	20.5	4.37	4.25	1.028
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	17.49	17.45	1.002
Average Correction Factor					1.015

Notes: Sample inlet filter and N₂ cylinder was changed after as founds. No adjustment made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4918	82.0	9.34	9.26	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.26	Prev response	9.33	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4918	82.0	9.34	9.31	1.003
Mid point	4959	41.0	4.67	4.63	1.010
Low point	4980	20.5	2.34	2.29	1.019
As left zero	5000	0.0	0.00	0.00	1.004
As left span	4918	82.0	9.34	9.31	1.004
Average Correction Factor					1.011

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4918	82.0	8.14	8.13	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.13	Prev response	8.12	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4918	82.0	8.14	8.13	1.002
Mid point	4959	41.0	4.07	4.01	1.016
Low point	4980	20.5	2.04	1.96	1.039
As left zero	5000	0.0	0.00	0.00	1.001
As left span	4918	82.0	8.14	8.13	1.001
Average Correction Factor					1.019

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.001453	0.998643
THC Cal Offset:	-0.062426	-0.060426
CH ₄ Cal Slope:	1.001772	1.000159
CH ₄ Cal Offset:	-0.039119	-0.040119
NMHC Cal Slope:	1.001382	0.997542
NMHC Cal Offset:	-0.023907	-0.020707

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 10, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:25	End Time (MST):	12:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999945	<i>≥0.995</i>
17.49	17.44	1.0029	Slope	0.998643	<i>0.90 - 1.10</i>
8.74	8.63	1.0129	Intercept	-0.060426	<i>+/-0.5</i>
4.37	4.25	1.0284			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 10, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:25	End Time (MST):	12:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999888	<i>≥0.995</i>
8.14	8.13	1.0021	Slope	1.000159	<i>0.90 - 1.10</i>
4.07	4.01	1.0160	Intercept	-0.040119	<i>+/-0.5</i>
2.04	1.96	1.0392			






Wood Buffalo Environmental Association NMHC Calibration Summary

Station Information

Calibration Date:	April 11, 2025	Previous Calibration:	March 10, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:25	End Time (MST):	12:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999977	≥ 0.995
9.34	9.31	1.0034	Slope	0.997542	$0.90 - 1.10$
4.67	4.63	1.0102	Intercept	-0.020707	± 0.5
2.34	2.29	1.0192			



NMHC Calibration Plot

Date: April 11, 2025

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Ells River
Station number: AMS 30
Calibration Date: April 16, 2025
Last Cal Date: March 13, 2025
Start time (MST): 9:45
End time (MST): 14:00
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0027487
NOX Cal Gas Conc: 59.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.30 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.10 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.10 ppm
NO gas Diff:
Serial Number: 3061
Serial Number: 358

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
AF High point	4932	67.7	803.0	800.3	2.7	782.1	777.0	5.1	1.0267	1.0301
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 800.6 ppb	NO = 797.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -2.4%
Baseline Corr 1st pt	NO _x = 782.1 ppb	NO = 776.9 ppb	<u>As Found Statistics</u>	*Percent Change	NO = -2.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :	NO SI:	NO Int:
			As found NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 710321429

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.053	1.092	NO bkgnd or offset:	12.3	12.7
NOX coeff or slope:	0.994	0.993	NOX bkgnd or offset:	12.4	12.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	181.8	186.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000203	1.001140
NO _x Cal Offset:	-2.519480	-1.398323
NO Cal Slope:	1.001320	1.003617
NO Cal Offset:	-3.700986	-2.779401
NO ₂ Cal Slope:	0.985498	1.001702
NO ₂ Cal Offset:	0.120532	-0.164440

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.3	0.0	----	----
High point	4932	67.7	803.0	800.3	2.7	803.0	801.6	1.4	1.0000	0.9983
Mid point	4966	33.8	400.9	399.5	1.4	400.0	397.6	2.4	1.0022	1.0049
Low point	4983	16.9	200.4	199.8	0.7	197.1	194.0	3.1	1.0169	1.0297
As left zero	5000	0.0	0.0	0.0	0.0	0.3	0.4	0.0	----	----
As left span	4932	67.7	803.0	428.9	374.1	802.5	428.9	373.6	1.0006	1.0000
Average Correction Factor									1.0064	1.0110

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	799.0	424.6	377.1	377.6	0.9987	100.1%
Mid GPT point	799.0	611.4	190.3	190.6	0.9985	100.2%
Low GPT point	799.0	703.3	98.4	98.1	1.0031	99.7%
Average Correction Factor					1.0001	100.0%

Notes: Sample inlet filter changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 16, 2025	Previous Calibration:	March 13, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:45	End Time (MST):	14:00
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999979	≥0.995
803.0	803.0	1.0000	Slope	1.001140	0.90 - 1.10
400.9	400.0	1.0022	Intercept	-1.398323	+/-20
200.4	197.1	1.0169			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 16, 2025	Previous Calibration:	March 13, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:45	End Time (MST):	14:00
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999998	≥0.995
377.1	377.6	0.9987	Slope	1.001702	0.90 - 1.10
190.3	190.6	0.9985	Intercept	-0.164440	+/-20
98.4	98.1	1.0031			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 16, 2025	Previous Calibration:	March 13, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:45	End Time (MST):	14:00
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999929	≥ 0.995
800.3	801.6	0.9983	Slope	1.003617	$0.90 - 1.10$
399.5	397.6	1.0049	Intercept	-2.779401	± 20
199.8	194.0	1.0297			



NO_x Calibration Plot

Date: April 16, 2025

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Ells River Station number: AMS 30
Calibration Date: April 16, 2025 Last Cal Date: March 14, 2025
Start time (MST): 10:46 End time (MST): 11:51

Analyzer Make: API T640 S/N: 875
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	3.80	3.30	3.80	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	728.70	730.47	728.70	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.05	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	33	----	33	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 4.10		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: September 29, 2024
Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.00	11.00	11.00	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 16, 2025
Date Disposable Filter Changed: April 16, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: December 6, 2024
Date RH/T Sensor Cleaned: February 23, 2024

Notes: Verified flow, temperature, pump power and pressure No adjustment made. Leak check passed.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS31
BLACKROD
APRIL 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Blackrod Station number: AMS 31
Calibration Date: April 7, 2025 Last Cal Date: March 26, 2025
Start time (MST): 13:29 End time (MST): 16:05
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.25 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC327023
Removed Cal Gas Conc: 50.25 ppm Rem Gas Exp Date: N/A
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Model: Teledyne T700 Serial Number: 5762
Zero Air Gen Model: Teledyne N701H Serial Number: 72

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1160290014
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999971	0.998942	Backgd or Offset: 39.1	39.5
Calibration intercept:	-0.052005	0.007990	Coeff or Slope: 1.019	1.019

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4920	79.6	800.0	799.0	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	799.1	Previous response	800.0	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4920	79.6	800.0	799.2	1.001
Mid point	4960	39.8	400.0	399.9	1.000
Low point	4980	19.9	200.0	199.2	1.004
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	79.6	800.0	800.0	1.000
Average Correction Factor:					1.002

Notes: Sample inlet filter was changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 26, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	13:29	End Time (MST):	16:05
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290014

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999998	≥0.995
800.0	799.2	1.0011	Slope	0.998942	0.90 - 1.10
400.0	399.9	1.0003	Intercept	0.007990	+/-30
200.0	199.2	1.0040			



SO2 Calibration Plot

Date: April 7, 2025

Location: Blackrod





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Blackrod Station number: AMS 31
Calibration Date: April 7, 2025 Last Cal Date: March 24, 2025
Start time (MST): 9:38 End time (MST): 13:29
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.42 ppm Cal Gas Exp Date: March 19, 2027
Cal Gas Cylinder #: DT0016926
Removed Cal Gas Conc: 5.42 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 5762
ZAG Make/Model: Teledyne API N701H Serial Number: 72

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 12228021056
Converter make: Global Converter serial #: 2023-266
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010620	1.007906	Backgd or Offset:	2.77
Calibration intercept:	-0.060561	-0.140523	Coeff or Slope:	1.030

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4926	73.8	80.0	80.3	0.995
As found Mid point	4963	36.9	40.0	40.6	0.983
As found Low point	4982	18.5	20.1	20.0	0.998
New cylinder response					
Baseline Corr As found:	80.4	Prev response:	80.79	*% change:	-0.5%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.005905	AF Intercept:	-0.020438
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999944	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4926	73.8	80.0	80.5	0.994
Mid point	4963	36.9	40.0	40.2	0.995
Low point	4982	18.5	20.1	20.0	1.003
As left zero	5000	0.0	0.0	0.1	----
As left span	4926	73.8	80.0	80.2	0.998
SO2 Scrubber Check	4920	79.6	796.1	0.1	----
Date of last scrubber change:				Ave Corr Factor	0.997
Date of last converter efficiency test:					

Notes: Sample inlet filter was changed after multipoint as founds. SO2 scrubber check done after calibrator zero and passed. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	April 7, 2025	Previous Calibration:	March 24, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	9:38	End Time (MST):	13:29
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12228021056

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999998		≥ 0.995
80.0	80.5	0.9938	Slope	1.007906		$0.90 - 1.10$
40.0	40.2	0.9950	Intercept	-0.140523		± 3
20.1	20.0	1.0026				


H₂S Calibration Curve



H₂S Calibration Plot

Date: April 7, 2025

Location: Blackrod





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Blackrod
Station number: AMS 31
Calibration Date: April 8, 2025
Last Cal Date: March 25, 2025
Start time (MST): 7:17
End time (MST): 11:36
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0035071
NOX Cal Gas Conc: 59.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.30 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API N701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.10 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.10 ppm
NO gas Diff:
Serial Number: 5762
Serial Number: 72

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
AF High point	4932	67.7	803.0	800.3	2.7	793.6	790.8	2.8	1.0118	1.0121
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 799.3 ppb	NO = 796.8 ppb	* = > +/-5% change initiates investigation			*Percent Change		NO _x = -0.7%		
Baseline Corr 1st pt	NO _x = 793.6 ppb	NO = 790.7 ppb	<u>As Found Statistics</u>			*Percent Change		NO = -0.8%		
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:		Nx Int:		
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:		NO Int:		
			As found NO ₂ r ² :			NO2 SI:		NO ₂ Int:		

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1426262592

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.959	0.974	NO bkgnd or offset:	12.4	12.5
NOX coeff or slope:	0.996	0.998	NOX bkgnd or offset:	12.8	12.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	200.4	192.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995387	1.002729
NO _x Cal Offset:	0.021490	0.743770
NO Cal Slope:	0.996585	1.003697
NO Cal Offset:	-0.719133	-0.157471
NO ₂ Cal Slope:	1.000119	1.005362
NO ₂ Cal Offset:	-0.101363	0.109061

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.3	-0.2	----	----
High point	4932	67.7	803.0	800.3	2.7	805.0	802.7	2.3	0.9975	0.9970
Mid point	4966	33.8	400.9	399.5	1.4	404.8	402.4	2.4	0.9903	0.9929
Low point	4983	16.9	200.4	199.8	0.7	201.2	198.7	2.5	0.9962	1.0053
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.2	-0.1	----	----
As left span	4932	67.7	803.0	394.0	409.0	795.1	394.0	401.1	1.0099	1.0000
Average Correction Factor									0.9947	0.9984

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	794.7	385.9	411.5	413.4	0.9954	100.5%
Mid GPT point	794.7	595.7	201.7	203.9	0.9893	101.1%
Low GPT point	794.7	694.6	102.8	103.2	0.9962	100.4%
Average Correction Factor					0.9936	100.6%

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 8, 2025	Previous Calibration:	March 25, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	7:17	End Time (MST):	11:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999984	≥0.995
803.0	805.0	0.9975	Slope	1.002729	0.90 - 1.10
400.9	404.8	0.9903	Intercept	0.743770	+/-20
200.4	201.2	0.9962			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 8, 2025	Previous Calibration:	March 25, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	7:17	End Time (MST):	11:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999986	≥0.995
411.5	413.4	0.9954	Slope	1.005362	0.90 - 1.10
201.7	203.9	0.9893	Intercept	0.109061	+/-20
102.8	103.2	0.9962			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 8, 2025	Previous Calibration:	March 25, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	7:17	End Time (MST):	11:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.3	----	Correlation Coefficient	0.999985	≥ 0.995
800.3	802.7	0.9970	Slope	1.003697	0.90 - 1.10
399.5	402.4	0.9929	Intercept	-0.157471	+/-20
199.8	198.7	1.0053			



NO_x Calibration Plot

Date: April 8, 2025

Location: Blackrod





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS33 MONDAY CREEK APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Monday Creek
Calibration Date: April 2, 2025
Start time (MST): 12:44
Reason: Routine

Station number: AMS 33
Last Cal Date: March 5, 2025
End time (MST): 15:35

Calibration Standards

Cal Gas Concentration: 50.62 ppm
Cal Gas Cylinder #: EB0008522
Removed Cal Gas Conc: 50.62 ppm
Removed Gas Cyl #: NA
Calibrator Model: Teledyne T700
Zero Air Gen Model: Teledyne T701H

Cal Gas Exp Date: March 10, 2031
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3253
Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0- 1000 ppb

Serial Number: 1152430005

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997815	1.005166	Backgd or Offset:	30.7	31.2
Calibration intercept:	-0.238150	-0.537953	Coeff or Slope:	0.988	1.001

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.6	----
As found High point	4921	79.1	800.8	791.2	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	791.8	Previous response	798.8	*% change	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.5	----
High point	4921	79.1	800.8	804.2	0.996
Mid point	4961	39.5	399.9	402.0	0.995
Low point	4980	19.8	200.5	200.5	1.000
As left zero	5000	0.0	0.0	-0.3	----
As left span	4921	79.1	800.8	803.5	0.997
Average Correction Factor:					0.997

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	March 5, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	12:44	End Time (MST):	15:35
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430005

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.5	----	Correlation Coefficient	0.999998	≥0.995
800.8	804.2	0.9958	Slope	1.005166	0.90 - 1.10
399.9	402.0	0.9947	Intercept	-0.537953	+/-30
200.5	200.5	0.9998			



SO2 Calibration Plot

Date: April 2, 2025

Location: Monday Creek





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Monday Creek
Calibration Date: April 15, 2025
Start time (MST): 9:44
Reason: Routine

Station number: AMS 33
Last Cal Date: March 4, 2025
End time (MST): 13:18

Calibration Standards

Cal Gas Concentration: 5.05 ppm
Cal Gas Cylinder #: DT0014831
Removed Cal Gas Conc: 5.05 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne T701H

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3253
Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL
Converter make: Global 150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12333331547
Converter serial #: 2022-196
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006570	0.999283	Backgd or Offset:	1.6	1.6
Calibration intercept:	-0.101612	-0.021599	Coeff or Slope:	1.076	1.076

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4921	79.2	80.0	80.1	0.996
As found Mid point	4960	39.6	40.0	40.1	0.993
As found Low point	4980	19.8	20.0	20.0	0.990
New cylinder response					
Baseline Corr As found:	80.3	Prev response:	80.41	*% change:	-0.1%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.003569	AF Intercept:	-0.121610
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4921	79.2	80.0	79.9	1.001
Mid point	4960	39.6	40.0	39.9	1.002
Low point	4980	19.8	20.0	20.1	0.995
As left zero	5000	0.0	0.0	0.0	----
As left span	4921	79.2	80.0	79.7	1.004
SO2 Scrubber Check	4921	79.1	791.0	0.1	----
Date of last scrubber change:	11-Apr-24		Ave Corr Factor		1.000
Date of last converter efficiency test:					

Notes: Sample inlet filter changed after multipoint as founds. SO2 scrubber check done after calibrator zero and passed. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	April 15, 2025	Previous Calibration:	March 4, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	9:44	End Time (MST):	13:18
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331547

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999992		≥ 0.995
80.0	79.9	1.0011	Slope	0.999283		$0.90 - 1.10$
40.0	39.9	1.0025	Intercept	-0.021599		± 3
20.0	20.1	0.9950				



H2S Calibration Plot

Date: April 15, 2025

Location: Monday Creek





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Monday Creek
Station number: AMS 33
Calibration Date: April 2, 2025
Last Cal Date: March 13, 2025
Start time (MST): 10:15
End time (MST): 12:44
Reason: Removal

Calibration Standards

NO Gas Cylinder #: CC755290
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: January 3, 2031
NO Cal Gas Conc: 48.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.70 ppm
NO gas Diff:
Serial Number: 3253
Serial Number: 832

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.6	-0.1	----	----
AF High point	4918	82.1	802.9	799.6	3.3	799.2	794.3	4.9	1.0038	1.0060
AF Mid point	4959	41.1	401.9	400.3	1.6	396.9	393.4	3.4	1.0109	1.0160
AF Low point	4980	20.5	200.5	199.7	0.8	196.8	194.0	2.8	1.0150	1.0260

New cyl resp

Previous Response	NO _x = 797.5 ppb	NO = 792.9 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.3%
Baseline Corr 1st pt	NO _x = 799.9 ppb	NO = 794.9 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.3%
Baseline Corr 2nd pt	NO _x = 397.6 ppb	NO = 394.0 ppb	As found	NO _x r ² : 0.999982	Nx SI: 0.996777	Nx Int: -2.153
Baseline Corr 3rd pt	NO _x = 197.5 ppb	NO = 194.6 ppb	As found	NO r ² : 0.999957	NO SI: 0.995169	NO Int: -2.933
			As found	NO ₂ r ² : 0.999985	NO2 SI: 0.995165	NO ₂ Int: 0.586

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	-0.1	----	----
As found high GPT point	792.7	381.2	414.8	412.9	1.0046	99.5%
As found mid GPT point	792.7	590.4	205.6	205.9	0.9985	100.2%
As found low GPT point	792.7	693.0	103.0	103.5	0.9950	100.5%



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1182340006

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.202	NA	NO bkgnd or offset:	1.0	NA
NOX coeff or slope:	0.992	NA	NOX bkgnd or offset:	1.2	NA
NO2 coeff or slope:	1.000	NA	Reaction cell Press:	104.9	NA

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.994870	
NO _x Cal Offset:	-1.333022	
NO Cal Slope:	0.994154	
NO Cal Offset:	-2.053068	
NO ₂ Cal Slope:	0.980054	
NO ₂ Cal Offset:	2.225929	

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero
High point
Mid point
Low point
As left zero
As left span

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero
High GPT point
Mid GPT point
Low GPT point

Average Correction Factor

Notes:

Removing the instrument to do further troubleshooting at the shop.


Calibration Performed By:

Jan Castro

NO_x Calibration Plot

Date: April 2, 2025

Location: Monday Creek





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Monday Creek
Station number: AMS 33
Calibration Date: April 3, 2025
Last Cal Date: NA
Start time (MST): 7:19
End time (MST): 12:27
Reason: Install

Calibration Standards

NO Gas Cylinder #: CC755290
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: March 11, 2031
NO Cal Gas Conc: 48.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.70 ppm
NO gas Diff:
Serial Number: 3253
Serial Number: 832

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation		*Percent Change		NO _x =	NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>		*Percent Change		NO =	NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:		Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:		NO Int:	
					As found	NO ₂ r ² :	NO2 SI:		NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12426335704

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:		0.999496
NO _x Cal Offset:		-0.133286
NO Cal Slope:		1.004314
NO Cal Offset:		-0.933140
NO ₂ Cal Slope:		0.975877
NO ₂ Cal Offset:		0.255076

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	NA	1.096	NO bkgnd or offset:	NA	1.6
NOX coeff or slope:	NA	0.998	NOX bkgnd or offset:	NA	1.7
NO2 coeff or slope:	NA	0.990	Reaction cell Press:	NA	149.0

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.5	-0.1	----	----
High point	4918	82.1	802.9	799.6	3.3	802.3	802.6	-0.3	1.0008	0.9963
Mid point	4959	41.1	401.9	400.3	1.6	401.4	400.2	1.3	1.0014	1.0003
Low point	4980	20.5	200.5	199.7	0.8	201.0	199.6	1.5	0.9974	1.0003
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
As left span	4918	82.1	802.9	390.8	412.1	785.2	390.8	394.4	1.0226	1.0000
Average Correction Factor									0.9998	0.9989

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	799.8	394.8	408.3	398.0	1.0258	97.5%
Mid GPT point	799.8	606.1	197.0	194.3	1.0138	98.6%
Low GPT point	799.8	704.8	98.3	95.4	1.0302	97.1%
Average Correction Factor					1.0233	97.7%

Notes:

Sample inlet filter and charcoal was changed before calibrator zero. Adjusted zero and span.

Calibration Performed By:

Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 3, 2025	Previous Calibration:	NA
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	7:19	End Time (MST):	12:27
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12426335704

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.6	----	Correlation Coefficient	0.999998	≥0.995
802.9	802.3	1.0008	Slope	0.999496	0.90 - 1.10
401.9	401.4	1.0014	Intercept	-0.133286	+/-20
200.5	201.0	0.9974			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 3, 2025	Previous Calibration:	NA
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	7:19	End Time (MST):	12:27
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12426335704

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999949	≥0.995
408.3	398.0	1.0258	Slope	0.975877	0.90 - 1.10
197.0	194.3	1.0138	Intercept	0.255076	+/-20
98.3	95.4	1.0302			





Wood Buffalo Environmental Association

NO Calibration Summary


Station Information

Calibration Date:	April 3, 2025	Previous Calibration:	NA
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	7:19	End Time (MST):	12:27
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12426335704

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.5	----	Correlation Coefficient	0.999997	≥ 0.995
799.6	802.6	0.9963	Slope	1.004314	$0.90 - 1.10$
400.3	400.2	1.0003	Intercept	-0.933140	± 20
199.7	199.6	1.0003			


NO Calibration Curve



NO_x Calibration Plot

Date: April 3, 2025

Location: Monday Creek





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS505 SAWBONES BAY APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Sawbones Bay
Calibration Date: March 28, 2025
Start time (MST): 9:30
Reason: Install

Station number: AMS 505
Last Cal Date: NA
End time (MST): 11:42

Calibration Standards

Cal Gas Concentration: 51.40 ppm
Cal Gas Cylinder #: EY0000672
Removed Cal Gas Conc: 51.40 ppm
Removed Gas Cyl #: EY0000672
Calibrator Model: Teledyne API T700
Zero Air Gen Model: Teledyne API T701

Cal Gas Exp Date: February 15, 2029
Rem Gas Exp Date: February 15, 2029
Diff between cyl:
Serial Number: 5112
Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 710321323

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		1.000238	Backgd or Offset:		19.8
Calibration intercept:		-0.072137	Coeff or Slope:		1.020

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.7	----
High point	4922	77.8	799.8	800.0	1.000
Mid point	4961	38.9	399.9	400.5	0.999
Low point	4981	19.5	200.4	199.0	1.007
As left zero	5000	0.0	0.0	1.2	----
As left span	4922	77.8	799.8	809.1	0.989
Average Correction Factor:					1.002

Notes:

Install calibration. Adjusted span and zero.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	March 28, 2025	Previous Calibration:	NA
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:30	End Time (MST):	11:42
Analyzer make:	Thermo 43i	Analyzer serial #:	710321323

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.7	----	Correlation Coefficient	0.999992	≥0.995
799.8	800.0	0.9998	Slope	1.000238	0.90 - 1.10
399.9	400.5	0.9985	Intercept	-0.072137	+/-30
200.4	199.0	1.0072			



SO2 Calibration Plot

Date: March 28, 2025

Location: Sawbones Bay





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Sawbones Bay Station number: AMS 505
Calibration Date: April 25, 2025 Last Cal Date: March 28, 2025
Start time (MST): 8:22 End time (MST): 11:04
Reason: Routine

Calibration Standards

Cal Gas Concentration: 51.40 ppm Cal Gas Exp Date: February 15, 2029
Cal Gas Cylinder #: EY0000672
Removed Cal Gas Conc: 51.40 ppm Rem Gas Exp Date: February 15, 2029
Removed Gas Cyl #: EY0000672 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 5112
Zero Air Gen Model: Teledyne API T701 Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 710321323
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000238	0.999694	Backgd or Offset:	19.8	19.8
Calibration intercept:	-0.072137	0.168148	Coeff or Slope:	1.020	1.020

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.8	----
As found High point	4922	77.8	799.8	803.9	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	803.1	Previous response	799.9	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4922	77.8	799.8	798.7	1.001
Mid point	4961	38.9	399.9	402.9	0.993
Low point	4981	19.5	200.4	198.8	1.008
As left zero	5000	0.0	0.0	-0.1	----
As left span	4922	77.8	799.8	809.7	0.988
Average Correction Factor:					1.001

Notes: Changed inelt filter after as founds. Adjusted zero and span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 25, 2025	Previous Calibration:	March 28, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:22	End Time (MST):	11:04
Analyzer make:	Thermo 43i	Analyzer serial #:	710321323

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999963	≥0.995
799.8	798.7	1.0014	Slope	0.999694	0.90 - 1.10
399.9	402.9	0.9926	Intercept	0.168148	+/-30
200.4	198.8	1.0082			



SO2 Calibration Plot

Date: April 25, 2025

Location: Sawbones Bay





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Sawbones Bay
Calibration Date: March 26, 2025
Start time (MST): 10:30
Reason: Install

Station number: AMS 505
Last Cal Date: NA
End time (MST): 13:13

Calibration Standards

Cal Gas Concentration: 5.26 ppm
Cal Gas Cylinder #: DT0034141
Removed Cal Gas Conc: 5.26 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T750
ZAG Make/Model: Teledyne API T751H

Cal Gas Exp Date: March 19, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 282
Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43iQTL
Converter make: Global 150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12113311965
Converter serial #: 2022-224
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		0.992596	Backgd or Offset:		0.920
Calibration intercept:		-0.020000	Coeff or Slope:		1.105

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4924	76.0	80.0	79.4	1.007
Mid point	4962	38.0	40.0	39.5	1.012
Low point	4981	19.0	20.0	19.9	1.004
As left zero	5000	0.0	0.0	0.1	----
As left span	4924	76.0	80.0	79.0	1.012
SO2 Scrubber Check	4922	77.8	778.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.008
Date of last converter efficiency test:					

Notes:

Install calibration.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	March 26, 2025	Previous Calibration:	NA
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	10:30	End Time (MST):	13:13
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311965

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999990		≥0.995
80.0	79.4	1.0070	Slope	0.992596		0.90 - 1.10
40.0	39.5	1.0121	Intercept	-0.020000		+/-3
20.0	19.9	1.0044				



H2S Calibration Plot

Date: March 26, 2025

Location: Sawbones Bay





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Sawbones Bay Station number: AMS 505
Calibration Date: April 24, 2025 Last Cal Date: March 26, 2025
Start time (MST): 8:09 End time (MST): 11:56
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.26 ppm Cal Gas Exp Date: March 19, 2027
Cal Gas Cylinder #: DT0034141
Removed Cal Gas Conc: 5.26 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T750 Serial Number: 282
ZAG Make/Model: Teledyne API T751H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43iQ Analyzer serial #: 12113311965
Converter make: Global 150 Converter serial #: 2022-224
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992596	0.994025	Backgd or Offset:	0.920	0.920
Calibration intercept:	-0.020000	-0.020000	Coeff or Slope:	1.105	1.105

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4924	76.0	80.0	79.3	1.008
As found Mid point	4962	38.0	40.0	39.6	1.009
As found Low point	4981	19.0	20.0	19.6	1.020
New cylinder response					
Baseline Corr As found:	79.3	Prev response:	79.34	*% change:	-0.1%
Baseline Corr 2nd AF pt:	39.6	AF Slope:	0.992739	AF Intercept:	-0.100000
Baseline Corr 3rd AF pt:	19.6	AF Correlation:	0.999991	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4924	76.0	80.0	79.5	1.006
Mid point	4962	38.0	40.0	39.7	1.007
Low point	4981	19.0	20.0	19.7	1.015
As left zero	5000	0.0	0.0	0.1	----
As left span	4924	76.0	80.0	79.1	1.011
SO2 Scrubber Check	4922	77.8	778.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	1.009
Date of last converter efficiency test:					

Notes: Changed inlet filter after as founds. Scrubber test was done after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H2S Calibration Summary


Station Information

Calibration Date:	April 24, 2025	Previous Calibration:	March 26, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:09	End Time (MST):	11:56
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12113311965

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999989		≥ 0.995
80.0	79.5	1.0057	Slope	0.994025		$0.90 - 1.10$
40.0	39.7	1.0070	Intercept	-0.020000		± 3
20.0	19.7	1.0146				


H₂S Calibration Curve



H2S Calibration Plot

Date: April 24, 2025

Location: Sawbones Bay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Sawbones Bay
Station number: AMS 505
Calibration Date: April 2, 2025
Last Cal Date: NA
Start time (MST): 8:49
End time (MST): 12:42
Reason: Install

Calibration Standards

NO Gas Cylinder #: DT0009786
NOX Cal Gas Conc: 60.10 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.10 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 5112
Serial Number: 690

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
-----------	------------------------------	--------------------------------	---	--	---	--	---	--	--	---

As found zero
AF High point
AF Mid point
AF Low point
New cyl resp

Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>		*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	---	--	--	---	--	--

As Found GPT zero
As found high GPT point
As found mid GPT point
As found low GPT point



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 4259

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:		0.941	NO bkgnd or offset:		3.4
NOX coeff or slope:		0.941	NOX bkgnd or offset:		3.8
NO2 coeff or slope:		1.000	Reaction cell Press:		3.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:		1.003271
NO _x Cal Offset:		-3.589772
NO Cal Slope:		1.003045
NO Cal Offset:		-4.150059
NO ₂ Cal Slope:		0.998577
NO ₂ Cal Offset:		0.523117

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.5	0.0	----	----
High point	4933	66.7	801.8	800.4	1.3	802.1	800.4	1.7	0.9996	1.0001
Mid point	4967	33.3	400.2	399.6	0.7	397.2	395.3	1.9	1.0077	1.0108
Low point	4983	16.7	200.7	200.4	0.3	194.2	192.9	1.3	1.0337	1.0389
As left zero	5000	0.0	0.0	0.0	0.0	-1.6	-1.7	0.1	----	----
As left span	4933	66.7	801.8	355.4	446.4	794.6	355.4	439.3	1.0090	1.0000
Average Correction Factor									1.0137	1.0166

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	794.0	348.4	446.9	446.5	1.0010	99.9%
Mid GPT point	794.0	551.0	244.3	244.7	0.9985	100.1%
Low GPT point	794.0	646.9	148.4	149.4	0.9935	100.7%
Average Correction Factor					0.9977	100.2%

Notes:

Install Calibration.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	NA
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:49	End Time (MST):	12:42
Analyzer make:	API T200	Analyzer serial #:	4259

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.5	----	Correlation Coefficient	0.999930	≥ 0.995
801.8	802.1	0.9996	Slope	1.003271	0.90 - 1.10
400.2	397.2	1.0077	Intercept	-3.589772	+/-20
200.7	194.2	1.0337			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	NA
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:49	End Time (MST):	12:42
Analyzer make:	API T200	Analyzer serial #:	4259

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999992	<i>≥0.995</i>
446.9	446.5	1.0010	Slope	0.998577	<i>0.90 - 1.10</i>
244.3	244.7	0.9985	Intercept	0.523117	<i>+/-20</i>
148.4	149.4	0.9935			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	NA
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:49	End Time (MST):	12:42
Analyzer make:	API T200	Analyzer serial #:	4259

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.5	----	Correlation Coefficient	0.999904	≥ 0.995
800.4	800.4	1.0001	Slope	1.003045	0.90 - 1.10
399.6	395.3	1.0108	Intercept	-4.150059	+/-20
200.4	192.9	1.0389			



NO_x Calibration Plot

Date: April 2, 2025

Location: Sawbones Bay





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Station Information

Station Name:	Sawbones Bay	Station Number:	AMS 505
Calibration Date:	March 27, 2025	Prev Cal Date:	NA
Start Time (MST):	9:36	End Time (MST):	12:50
Tower Height (m):	10.0	Reason:	Install

Wind Speed Calibration

Sensor make/model:	Met One 010C-1	Serial Number:	P10040
WS Calibrator:	MetOne 053	Serial Number:	CA 03845

Shaft RPM (Hz)	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.7%
400	39.4	39.4	0.1%
600	58.6	58.7	0.2%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)		0.999998	≥ 0.9995
Calculated slope		0.999451	$0.98 - 1.02$
Calculated intercept		-0.052868	± 2

Wind Direction Calibration

Sensor make/model:	Met One 020C-1	Serial Number:	B4693
As Found Declination (deg east of True North):	NA	As Left Declination (deg east of True North):	13
Solar noon (MST):	12:28	Calc Declination*:	12.72 Degrees
WD Calibrator:	Met One 040		<small>* - calculated declination as per NOAA website</small>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 360° FS) <i>Limit = +/- 1%</i>
0	0.8	0.2%
90	86.3	-1.0%
180	177.6	-0.7%
270	270.5	0.1%
355	356.3	0.4%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)		0.999894	≥ 0.9995
Calculated slope		0.994042	$0.97 - 1.03$
Calculated intercept		1.762340	± 5

Notes: Install Calibration.

Calibration Performed By: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS506
JACKFISH 1
APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Jackfish 1 Station number: AMS 506
Calibration Date: April 23, 2025 Last Cal Date: March 19, 2025
Start time (MST): 8:58 End time (MST): 11:39
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.52 ppm Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #: CC274266
Removed Cal Gas Conc: 50.52 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 2659
Zero Air Gen Model: Teledyne API T701 Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1160290011
Analyzer Range: 0-1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001515	0.999944	Backgd or Offset:	20.2	20.2
Calibration intercept:	0.763956	0.664017	Coeff or Slope:	0.985	0.985

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.4	----
As found High point	4921	79.2	800.2	797.1	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	796.7	Previous response	802.2	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.6	----
High point	4921	79.2	800.2	801.0	0.999
Mid point	4960	39.6	400.2	400.2	1.000
Low point	4980	19.8	200.1	201.2	0.994
As left zero	5000	0.0	0.0	0.3	----
As left span	4921	79.2	800.2	801.0	0.999
Average Correction Factor:					0.998

Notes: Changed inlet filter after as founds. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 23, 2025	Previous Calibration:	March 19, 2025
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	8:58	End Time (MST):	11:39
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999998	≥0.995
800.2	801.0	0.9990	Slope	0.999944	0.90 - 1.10
400.2	400.2	0.9999	Intercept	0.664017	+/-30
200.1	201.2	0.9944			



SO2 Calibration Plot

Date: April 23, 2025

Location: Jackfish 1





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Jackfish 1
Calibration Date: April 15, 2025
Start time (MST): 8:59
Reason: Routine

Station number: AMS 506
Last Cal Date: March 20, 2025
End time (MST): 12:53

Calibration Standards

Cal Gas Concentration: 4.89 ppm
Cal Gas Cylinder #: CC737971
Removed Cal Gas Conc: 4.89 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne 750
ZAG Make/Model: Teledyne 751H

Cal Gas Exp Date: September 5, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 282
Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1180540020
Converter serial #: 2022-218
Converter Temp: 325.0 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005447	0.993022	Backgd or Offset:	3.72
Calibration intercept:	0.020854	0.080592	Coeff or Slope:	1.179

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4918	81.8	80.0	81.5	0.978
As found Mid point	4959	40.9	40.0	40.2	0.988
As found Low point	4980	20.4	19.9	20.2	0.973
New cylinder response					
Baseline Corr As found:	81.8	Prev response:	80.46	*% change:	1.6%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	1.021443	AF Intercept:	-0.338798
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999961	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4918	81.8	80.0	79.4	1.008
Mid point	4959	40.9	40.0	40.0	1.000
Low point	4980	20.4	19.9	20.0	0.997
As left zero	5000	0.0	0.0	-0.1	----
As left span	4918	81.8	80.0	79.8	1.003
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.002
Date of last converter efficiency test:					

Notes: Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H2S Calibration Summary


Station Information

Calibration Date:	April 15, 2025	Previous Calibration:	March 20, 2025
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	8:59	End Time (MST):	12:53
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999971	≥ 0.995
80.0	79.4	1.0076	Slope	0.993022	$0.90 - 1.10$
40.0	40.0	1.0000	Intercept	0.080592	± 3
19.9	20.0	0.9975			


H₂S Calibration Curve



H2S Calibration Plot

Date: April 15, 2025

Location: Jackfish 1



0



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Jackfish 1
Station number: AMS 506
Calibration Date: April 16, 2025
Last Cal Date: March 25, 2025
Start time (MST): 8:30
End time (MST): 12:44
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0022706
NOX Cal Gas Conc: 60.20 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.20 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API 701
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.10 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.10 ppm
NO gas Diff:
Serial Number: 3252
Serial Number: 4427

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
AF High point	4933	66.6	801.9	800.6	1.3	802.0	797.1	5.0	0.9998	1.0044
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x =	802.1 ppb	NO =	798.4 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = 0.0%	
Baseline Corr 1st pt	NO _x =	802.1 ppb	NO =	797.1 ppb	<u>As Found Statistics</u>			*Percent Change	NO = -0.2%	
Baseline Corr 2nd pt	NO _x =	NA ppb	NO =	NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x =	NA ppb	NO =	NA ppb	As found NO r ² :			NO SI:	NO Int:	
					As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12400232071

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.912	0.912	NO bkgnd or offset:	0.7	0.7
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	0.9	0.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	194.7	192.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996743	0.995577
NO _x Cal Offset:	2.813160	2.172098
NO Cal Slope:	0.996494	0.996708
NO Cal Offset:	0.631010	0.231152
NO ₂ Cal Slope:	0.993078	0.985570
NO ₂ Cal Offset:	2.112455	1.533148

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.0	----	----
High point	4933	66.6	801.9	800.6	1.3	799.5	798.5	1.0	1.0030	1.0026
Mid point	4967	33.3	400.9	400.2	0.7	402.5	398.2	4.3	0.9960	1.0051
Low point	4983	16.6	199.9	199.5	0.3	203.0	199.8	3.2	0.9846	0.9987
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
As left span	4933	66.6	801.9	380.4	421.5	789.5	380.4	409.1	1.0157	1.0000
Average Correction Factor									0.9946	1.0022

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	794.4	380.7	415.0	409.7	1.0130	98.7%
Mid GPT point	794.4	585.8	209.9	209.4	1.0025	99.7%
Low GPT point	794.4	684.1	111.6	113.0	0.9879	101.2%
Average Correction Factor					1.0012	99.9%

Notes:

Changed inlet filter. No adjustment made.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 16, 2025	Previous Calibration:	March 25, 2025
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	8:30	End Time (MST):	12:44
Analyzer make:	Thermo 42i	Analyzer serial #:	12400232071

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999972	≥0.995
801.9	799.5	1.0030	Slope	0.995577	0.90 - 1.10
400.9	402.5	0.9960	Intercept	2.172098	+/-20
199.9	203.0	0.9846			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 16, 2025	Previous Calibration:	March 25, 2025
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	8:30	End Time (MST):	12:44
Analyzer make:	Thermo 42i	Analyzer serial #:	12400232071

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999932	≥0.995
415.0	409.7	1.0130	Slope	0.985570	0.90 - 1.10
209.9	209.4	1.0025	Intercept	1.533148	+/-20
111.6	113.0	0.9879			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 16, 2025	Previous Calibration:	March 25, 2025
Station Name:	Jackfish 1	Station Number:	AMS 506
Start Time (MST):	8:30	End Time (MST):	12:44
Analyzer make:	Thermo 42i	Analyzer serial #:	12400232071

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999996	≥ 0.995
800.6	798.5	1.0026	Slope	0.996708	0.90 - 1.10
400.2	398.2	1.0051	Intercept	0.231152	+/-20
199.5	199.8	0.9987			



NO_x Calibration Plot

Date: April 16, 2025

Location: Jackfish 1





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS508
KIRBY NORTH
APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Kirby North
Calibration Date: April 3, 2025
Start time (MST): 10:54
Reason: Routine

Station number: AMS 508
Last Cal Date: March 13, 2025
End time (MST): 13:58

Calibration Standards

Cal Gas Concentration: 50.74 ppm
Cal Gas Cylinder #: CC255918
Removed Cal Gas Conc: 50.74 ppm
Removed Gas Cyl #:
Calibrator Model: Teledyne API T700
Zero Air Gen Model: Teledyne API T701H

Cal Gas Exp Date: October 9, 2032
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 5240
Serial Number: 880

Analyzer Information

Analyzer make: Thermo 43iQ
Analyzer Range: 0 - 1000 ppb

Serial Number: 1182340007

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998619	1.002663	Backgd or Offset:	29.1	29.0
Calibration intercept:	0.327990	-0.512059	Coeff or Slope:	1.117	1.117

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4921	78.8	799.7	803.0	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	802.9	Previous response	798.9	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	78.8	799.7	802.0	0.997
Mid point	4961	39.4	399.8	398.8	1.003
Low point	4980	19.7	199.9	200.3	0.998
As left zero	5000	0.0	0.0	-0.2	----
As left span	4921	78.8	799.7	804.0	0.995
Average Correction Factor:					0.999

Notes: Changed sample inlet filter after as founds. Adjusted zero.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 3, 2025	Previous Calibration:	March 13, 2025
Station Name:	Kirby North	Station Number:	AMS 508
Start Time (MST):	10:54	End Time (MST):	13:58
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999991	≥0.995
799.7	802.0	0.9971	Slope	1.002663	0.90 - 1.10
399.8	398.8	1.0025	Intercept	-0.512059	+/-30
199.9	200.3	0.9981			



SO2 Calibration Plot

Date: April 3, 2025

Location: Kirby North





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Kirby North
Calibration Date: April 2, 2025
Start time (MST): 9:25
Reason: Routine

Station number: AMS 508
Last Cal Date: March 12, 2025
End time (MST): 14:40

Calibration Standards

Cal Gas Concentration: 5.05 ppm
Cal Gas Cylinder #: DT0019762
Removed Cal Gas Conc: 5.05 ppm
Removed Gas Cyl #: n/a
Calibrator Make/Model: Teledyne API T750
ZAG Make/Model: Teledyne API T751H

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 282
Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1150840012
Converter serial #: 2022-197
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006814	0.999956	Backgd or Offset:	1.74	1.76
Calibration intercept:	-0.160959	-0.120959	Coeff or Slope:	1.035	1.043

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4921	79.2	80.0	80.3	0.995
As found Mid point	4960	39.6	40.0	40.2	0.993
As found Low point	4980	19.8	20.0	19.6	1.015
New cylinder response					
Baseline Corr As found:	80.4	Prev response:	80.38	*% change:	0.0%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.006957	AF Intercept:	-0.240970
Baseline Corr 3rd AF pt:	19.7	AF Correlation:	0.999962	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	79.2	80.0	79.9	1.001
Mid point	4960	39.6	40.0	39.9	1.002
Low point	4980	19.8	20.0	19.7	1.015
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	79.2	80.0	78.9	1.014
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:		July 25, 2023		Ave Corr Factor	1.006
Date of last converter efficiency test:		n/a			

Notes: Changed sample inlet filter and conducted scrubber test after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	March 12, 2025
Station Name:	Kirby North	Station Number:	AMS 508
Start Time (MST):	9:25	End Time (MST):	14:40
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999986		≥ 0.995
80.0	79.9	1.0012	Slope	0.999956		$0.90 - 1.10$
40.0	39.9	1.0025	Intercept	-0.120959		± 3
20.0	19.7	1.0152				



H2S Calibration Plot

Date: April 2, 2025

Location: Kirby North





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Kirby North
Calibration Date: April 3, 2025
Start time (MST): 10:54
Reason: Routine

Station number: AMS 508
Last Cal Date: March 13, 2025
End time (MST): 13:58

Calibration Standards

Gas Cert Reference: CC255918
CH4 Cal Gas Conc. 506.4 ppm
C3H8 Cal Gas Conc. 205.0 ppm
Removed Gas Cert:
Removed CH4 Conc. 506.4 ppm
Removed C3H8 Conc. 205.0 ppm
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Expiry Date: October 9, 2032
CH4 Equiv Conc. 1070.2 ppm
Removed Gas Expiry:
CH4 Equiv Conc. 1070.2 ppm
Diff between cyl:
Serial Number: 5240
Serial Number: 880

Analyzer Information

Analyzer make: Thermo 51i-LT
Analyzer Range: 0 - 20 ppm

Analyzer serial #: 1182340005

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000987	0.994170	Background:	2.23	1.96
Calibration intercept:	-0.082032	0.019768	Coefficient:	3.697	3.644

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	-0.20	----
As found High point	4921	78.8	16.87	16.77	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	16.97	Previous response	16.80	*% change	1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	
High point	4921	78.8	16.87	16.78	1.005
Mid point	4961	39.4	8.43	8.41	1.003
Low point	4980	19.7	4.22	4.23	0.997
As left zero	5000	0.0	0.00	0.06	----
As left span	4921	78.8	16.87	16.83	1.002
Average Correction Factor					1.002

Notes: Changed sample inlet filter after as founds. Adjusted zero.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	April 3, 2025	Previous Calibration:	March 13, 2025
Station Name:	Kirby North	Station Number:	AMS 508
Start Time (MST):	10:54	End Time (MST):	13:58
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1182340005

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995
16.87	16.78	1.0051	Slope	0.994170	$0.90 - 1.10$
8.43	8.41	1.0026	Intercept	0.019768	± 1.5
4.22	4.23	0.9973			


THC Calibration Curve



THC Calibration Plot

Date: April 3, 2025

Location: Kirby North





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Kirby North
Station number: AMS 508
Calibration Date: April 2, 2025
Last Cal Date: March 12, 2025
Start time (MST): 9:05
End time (MST): 13:49
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0019572
NOX Cal Gas Conc: 60.00 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.00 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 59.90 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.90 ppm
NO gas Diff:
Serial Number: 5240
Serial Number: 880

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
AF High point	4933	66.8	801.6	800.3	1.3	805.0	804.0	1.1	0.9958	0.9954
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 799.8 ppb	NO = 799.5 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 0.6%
Baseline Corr 1st pt	NO _x = 805.0 ppb	NO = 804.0 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.6%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1118148496

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998604	0.993215
NO _x Cal Offset:	-0.693582	-0.053616
NO Cal Slope:	1.000714	0.997815
NO Cal Offset:	-1.373584	-0.933611
NO ₂ Cal Slope:	0.981281	0.977286
NO ₂ Cal Offset:	1.476756	0.360528

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.707	0.703	NO bkgnd or offset:	8.0	8.0
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	8.1	8.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	147.1	146.2

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4933	66.8	801.6	800.3	1.3	795.7	797.6	-1.8	1.0075	1.0034
Mid point	4967	33.4	400.8	400.1	0.7	399.4	399.3	0.1	1.0035	1.0021
Low point	4983	16.7	200.4	200.1	0.3	198.0	196.8	1.2	1.0121	1.0166
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
As left span	4933	66.8	801.6	381.9	419.7	788.0	381.9	406.1	1.0173	1.0000
Average Correction Factor									1.0077	1.0074

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	794.8	385.6	410.5	401.4	1.0228	97.8%
Mid GPT point	794.8	610.9	185.2	181.6	1.0200	98.0%
Low GPT point	794.8	698.5	97.6	96.1	1.0160	98.4%
Average Correction Factor					1.0196	98.1%

Notes:

Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	March 12, 2025
Station Name:	Kirby North	Station Number:	AMS 508
Start Time (MST):	9:05	End Time (MST):	13:49
Analyzer make:	Thermo 42i	Analyzer serial #:	1118148496

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999991	≥0.995
801.6	795.7	1.0075	Slope	0.993215	0.90 - 1.10
400.8	399.4	1.0035	Intercept	-0.053616	+/-20
200.4	198.0	1.0121			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	March 12, 2025
Station Name:	Kirby North	Station Number:	AMS 508
Start Time (MST):	9:05	End Time (MST):	13:49
Analyzer make:	Thermo 42i	Analyzer serial #:	1118148496

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999997	≥0.995
410.5	401.4	1.0228	Slope	0.977286	0.90 - 1.10
185.2	181.6	1.0200	Intercept	0.360528	+/-20
97.6	96.1	1.0160			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 2, 2025	Previous Calibration:	March 12, 2025
Station Name:	Kirby North	Station Number:	AMS 508
Start Time (MST):	9:05	End Time (MST):	13:49
Analyzer make:	Thermo 42i	Analyzer serial #:	1118148496

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999984	≥ 0.995
800.3	797.6	1.0034	Slope	0.997815	$0.90 - 1.10$
400.1	399.3	1.0021	Intercept	-0.933611	± 20
200.1	196.8	1.0166			



NO_x Calibration Plot

Date: April 2, 2025

Location: Kirby North





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS512 HANGINGSTONE EXPANSION APRIL 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

May 30, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Hangingstone Expansion Station number: AMS 512
Calibration Date: April 10, 2025 Last Cal Date: March 7, 2025
Start time (MST): 6:10 End time (MST): 8:38
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.06 ppm Cal Gas Exp Date: January 5, 2029
Cal Gas Cylinder #: CC147416
Removed Cal Gas Conc: 50.06 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 2445
Zero Air Gen Model: Teledyne API 701 Serial Number: 138

Analyzer Information

Analyzer make: Thermo scientific Serial Number: 1173410001
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006091	1.009553	Backgd or Offset:	14.2	14.2
Calibration intercept:	-1.263661	-1.623881	Coeff or Slope:	1.175	1.175

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4920	79.8	799.0	806.4	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	806.3	Previous response	802.6	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4920	79.8	799.0	806.0	0.991
Mid point	4960	39.9	399.5	400.6	0.997
Low point	4987	20.0	200.0	198.4	1.008
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	79.8	799.0	809.6	0.987
Average Correction Factor:					0.999

Notes: No adjustments or maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	April 10, 2025	Previous Calibration:	March 7, 2025
Station Name:	Hangingstone Expansion	Station Number:	AMS 512
Start Time (MST):	6:10	End Time (MST):	8:38
Analyzer make:	Thermo scientific	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999974	≥0.995
799.0	806.0	0.9913	Slope	1.009553	0.90 - 1.10
399.5	400.6	0.9972	Intercept	-1.623881	+/-30
200.0	198.4	1.0079			



SO2 Calibration Plot

Date: April 10, 2025

Location: Hangingstone Expansion





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Hangingstone Expansion Station number: AMS 512
Calibration Date: April 4, 2025 Last Cal Date: March 13, 2025
Start time (MST): 6:27 End time (MST): 10:38
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.139 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC511397
Removed Cal Gas Conc: 5.139 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 2445
ZAG Make/Model: API T701 Serial Number: 138

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1336160090
Converter make: Global G150 Converter serial #: 2022-227
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995485	1.000488	Backgd or Offset:	3.56
Calibration intercept:	0.060913	0.160819	Coeff or Slope:	1.235

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4922	77.8	80.0	79.8	1.005
As found Mid point	4961	38.9	40.0	40.3	0.997
As found Low point	4981	19.5	20.0	20.1	1.007
New cylinder response					
Baseline Corr As found:	79.6	Prev response:	79.67	*% change:	-0.1%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	0.996058	AF Intercept:	0.240853
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999978	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4922	77.8	80.0	80.2	0.997
Mid point	4961	38.9	40.0	40.2	0.995
Low point	4981	19.5	20.0	20.0	1.002
As left zero	5000	0.0	0.0	0.4	----
As left span	4922	77.8	80.0	80.0	1.000
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.998
Date of last converter efficiency test:					

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	April 4, 2025	Previous Calibration:	March 13, 2025
Station Name:	Hangingstone Expansion	Station Number:	AMS 512
Start Time (MST):	6:27	End Time (MST):	10:38
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1336160090

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999981		≥ 0.995
80.0	80.2	0.9971	Slope	1.000488		$0.90 - 1.10$
40.0	40.2	0.9946	Intercept	0.160819		± 3
20.0	20.0	1.0020				


H₂S Calibration Curve



H₂S Calibration Plot

Date: April 4, 2025

Location: Hangingstone Expansion





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Hangingstone Expansion
Station number: AMS 512
Calibration Date: April 1, 2025
Last Cal Date: March 3, 2025
Start time (MST): 6:28
End time (MST): 10:43
Reason: Routine

Calibration Standards

NO Gas Cylinder #: TOF8P52
NOX Cal Gas Conc: 47.43 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.43 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: August 16, 2026
NO Cal Gas Conc: 47.43 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.43 ppm
NO gas Diff:
Serial Number: 2445
Serial Number: 138

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
AF High point	4916	84.4	800.6	800.6	0.0	801.9	801.2	0.5	0.9981	0.9989
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 807.3 ppb	NO = 807.2 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = -0.6%			
Baseline Corr 1st pt	NO _x = 802.1 ppb	NO = 801.4 ppb	<u>As Found Statistics</u>			*Percent Change	NO = -0.7%			
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:	NO Int:			
			As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 7029

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.056	1.056	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.052	1.052	NOX bkgnd or offset:	0.4	0.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.6	4.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.008889	1.005306
NO _x Cal Offset:	-0.392928	-0.812870
NO Cal Slope:	1.010359	1.005634
NO Cal Offset:	-1.632904	-1.552843
NO ₂ Cal Slope:	0.999078	0.999674
NO ₂ Cal Offset:	0.427715	0.103787

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
High point	4916	84.4	800.6	800.6	0.0	804.1	804.1	0.2	0.9956	0.9956
Mid point	4958	42.2	400.3	400.3	0.0	402.0	400.7	1.3	0.9958	0.9990
Low point	4979	21.1	200.2	200.2	0.0	199.2	198.0	1.2	1.0048	1.0109
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	-0.1	----	----
As left span	4916	84.4	800.6	406.9	393.7	799.9	406.9	393.0	1.0008	1.0000
Average Correction Factor									0.9987	1.0018

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	802.1	410.9	391.2	391.1	1.0003	100.0%
Mid GPT point	802.1	624.1	178.0	178.2	0.9989	100.1%
Low GPT point	802.1	712.2	89.9	90.0	0.9989	100.1%
Average Correction Factor					0.9993	100.1%

Notes:

No adjustments and maintenance done.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	April 1, 2025	Previous Calibration:	March 3, 2025
Station Name:	Hangingstone Expansion	Station Number:	AMS 512
Start Time (MST):	6:28	End Time (MST):	10:43
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999994	≥0.995
800.6	804.1	0.9956	Slope	1.005306	0.90 - 1.10
400.3	402.0	0.9958	Intercept	-0.812870	+/-20
200.2	199.2	1.0048			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	April 1, 2025	Previous Calibration:	March 3, 2025
Station Name:	Hangingstone Expansion	Station Number:	AMS 512
Start Time (MST):	6:28	End Time (MST):	10:43
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	1.000000	≥ 0.995
391.2	391.1	1.0003	Slope	0.999674	$0.90 - 1.10$
178.0	178.2	0.9989	Intercept	0.103787	± 20
89.9	90.0	0.9989			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	April 1, 2025	Previous Calibration:	March 3, 2025
Station Name:	Hangingstone Expansion	Station Number:	AMS 512
Start Time (MST):	6:28	End Time (MST):	10:43
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999984	≥ 0.995
800.6	804.1	0.9956	Slope	1.005634	$0.90 - 1.10$
400.3	400.7	0.9990	Intercept	-1.552843	± 20
200.2	198.0	1.0109			



NO_x Calibration Plot

Date: April 1, 2025

Location: Hangingstone Expansion





End of Report