



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

JULY 2025

MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

August 29, 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01 BERTHA GANTER - FORT MCKAY JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: July 7, 2025 Last Cal Date: June 9, 2025
Start time (MST): 11:00 End time (MST): 14:11
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.21 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC418809
Removed Cal Gas Conc: 49.21 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3565
Zero Air Gen Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1501301448
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997895	0.997010	Backgd or Offset:	21.7	21.5
Calibration intercept:	-0.932900	-0.373313	Coeff or Slope:	0.876	0.879

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4918	81.3	800.3	792.7	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	792.9	Previous response	797.6	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4918	81.3	800.3	797.9	1.003
Mid point	4959	40.7	400.6	398.3	1.006
Low point	4979	20.3	199.8	198.7	1.006
As left zero	5000	0.0	0.0	0.0	----
As left span	4918	81.3	800.3	799.2	1.001

Average Correction Factor: 1.005

Notes: Changed the inlet filter after as founds. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 9, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:00	End Time (MST):	14:11
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999997		≥0.995
800.3	797.9	1.0030	Slope	0.997010		0.90 - 1.10
400.6	398.3	1.0058	Intercept	-0.373313		+/-30
199.8	198.7	1.0056				



SO2 Calibration Plot

Date: July 7, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: July 14, 2025 Last Cal Date: June 5, 2025
Start time (MST): 9:51 End time (MST): 13:35
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: September 5, 2027
Cal Gas Cylinder #: CC738239
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 12113311966
Converter make: CD Nova Converter serial #: 470
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002391	1.008099	Backgd or Offset:	2.20	2.14
Calibration intercept:	-0.078174	-0.577930	Coeff or Slope:	1.159	1.123

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4917	82.6	80.0	83.5	0.955
As found Mid point	4959	41.3	40.0	41.2	0.966
As found Low point	4979	20.7	20.0	20.0	0.992
New cylinder response					
Baseline Corr As found:	83.7	Prev response:	80.08	*% change:	4.3%
Baseline Corr 2nd AF pt:	41.4	AF Slope:	1.049416	AF Intercept:	-0.598777
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999896	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4917	82.6	80.0	80.2	0.997
Mid point	4959	41.3	40.0	39.7	1.007
Low point	4979	20.7	20.0	19.0	1.055
As left zero					
As left span					
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:		December 17, 2021		Ave Corr Factor	1.020
Date of last converter efficiency test:					

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted the span. Converter failed during the 3rd point. Will replace tomorrow.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	July 14, 2025	Previous Calibration:	June 5, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:51	End Time (MST):	13:35
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	12113311966

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999818		≥ 0.995
80.0	80.2	0.9970	Slope	1.008099		$0.90 - 1.10$
40.0	39.7	1.0070	Intercept	-0.577930		± 3
20.0	19.0	1.0547				


TRS Calibration Curve



TRS Calibration Plot

Date: July 14, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: July 15, 2025 Last Cal Date: July 14, 2025
Start time (MST): 9:51 End time (MST): 15:12
Reason: Maintenance

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: September 5, 2027
Cal Gas Cylinder #: CC738239
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 12113311966
Converter make: CD Nova Converter serial #: 580
Analyzer Range: 0 - 100 ppb Converter Temp: 850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008099	1.003817	Backgd or Offset:	2.14	2.13
Calibration intercept:	-0.577930	-0.278082	Coeff or Slope:	1.123	1.121

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4917	82.6	80.0	80.0	1.000
Mid point	4959	41.3	40.0	39.9	1.002
Low point	4979	20.7	20.0	19.7	1.017
As left zero	5000	0.0	0.0	-0.1	----
As left span	4917	82.6	80.0	78.6	1.017
SO2 Scrubber Check					
Date of last scrubber change:		December 17, 2021		Ave Corr Factor	1.006
Date of last converter efficiency test:					

Notes: Installing a new TRS converter. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	July 14, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:51	End Time (MST):	15:12
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	12113311966

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999992	≥ 0.995
80.0	80.0	0.9995	Slope	1.003817	$0.90 - 1.10$
40.0	39.9	1.0019	Intercept	-0.278082	± 3
20.0	19.7	1.0172			


TRS Calibration Curve



TRS Calibration Plot

Date: July 15, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: July 14, 2025 Last Cal Date: June 5, 2025
Start time (MST): 9:51 End time (MST): 14:22
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: September 5, 2027
Cal Gas Cylinder #: CC738239
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 1200326167
Converter make: CD Nova Converter serial #: 2022-221
Analyzer Range: 0 - 100 ppb Converter Temp: 315 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006678	1.006105	Backgd or Offset:	2.03	2.00
Calibration intercept:	-0.178203	-0.258141	Coeff or Slope:	0.983	9.770

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4917	82.6	80.0	81.3	0.980
As found Mid point	4959	41.3	40.0	40.3	0.985
As found Low point	4979	20.7	20.0	19.8	0.997
New cylinder response					
Baseline Corr As found:	81.6	Prev response:	80.32	*% change:	1.6%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.021686	AF Intercept:	-0.478384
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999978	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4917	82.6	80.0	80.2	0.997
Mid point	4959	41.3	40.0	40.0	0.999
Low point	4979	20.7	20.0	19.8	1.012
As left zero	5000	0.0	0.0	-0.1	----
As left span	4917	82.6	80.0	79.5	1.006
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	January 25, 2024		Ave Corr Factor		1.003
Date of last converter efficiency test:	November 7, 2024		107.9% efficiency		

Notes: Inlet filter change and scrubber check completed after as founds. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	July 14, 2025	Previous Calibration:	June 5, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	9:51	End Time (MST):	14:22
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1200326167

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999996		≥0.995
80.0	80.2	0.9970	Slope	1.006105		0.90 - 1.10
40.0	40.0	0.9994	Intercept	-0.258141		+/-3
20.0	19.8	1.0121				


H₂S Calibration Curve



H₂S Calibration Plot

Date: July 14, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
 Calibration Date: July 7, 2025 Last Cal Date: June 9, 2025
 Start time (MST): 11:00 End time (MST): 14:11
 Reason: Routine

Calibration Standards

Gas Cert Reference: CC418809 Cal Gas Expiry Date: March 10, 2031
 CH₄ Cal Gas Conc. 497.2 ppm CH₄ Equiv Conc. 1061.8 ppm
 C₃H₈ Cal Gas Conc. 205.3 ppm
 Removed Gas Cert: NA Removed Gas Expiry: NA
 Removed CH₄ Conc. 497.2 ppm CH₄ Equiv Conc. 1061.8 ppm
 Removed C₃H₈ Conc. 205.3 ppm
 Diff between cyl (CH₄): Diff between cyl (NM):
 Calibrator Model: Teledyne API T700 Serial Number: 3565
 Zero Air Gen model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 1193585648
 THC Range: 0 - 20 ppm NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.52E-04	2.48E-04	NMHC SP Ratio:	4.95E-05	4.96E-05
CH ₄ Retention time:	15.0	15.0	NMHC Peak Area:	185536	185016
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	17.27	17.37	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.37	Prev response	17.28	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	17.27	17.27	1.000
Mid point	4959	40.7	8.64	8.65	0.999
Low point	4979	20.3	4.31	4.35	0.991
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	17.27	17.29	0.999
Average Correction Factor					0.997

Notes: Changed the inlet filter after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	9.18	9.17	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.17	Prev response	9.14	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	9.18	9.17	1.001
Mid point	4959	40.7	4.60	4.62	0.995
Low point	4979	20.3	2.29	2.33	0.983
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	9.18	9.15	1.003
Average Correction Factor					0.993

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	8.09	8.20	0.986
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.20	Prev response	8.13	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	8.09	8.10	0.999
Mid point	4959	40.7	4.05	4.03	1.004
Low point	4979	20.3	2.02	2.02	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	8.09	8.14	0.993
Average Correction Factor					1.001

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	0.999481	0.999310
THC Cal Offset:	0.017754	0.016545
CH ₄ Cal Slope:	1.005593	1.001042
CH ₄ Cal Offset:	0.002469	-0.005928
NMHC Cal Slope:	0.994224	0.997823
NMHC Cal Offset:	0.015285	0.022073

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 9, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:00	End Time (MST):	14:11
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995
17.27	17.27	1.0000	Slope	0.999310	$0.90 - 1.10$
8.64	8.65	0.9991	Intercept	0.016545	± 0.5
4.31	4.35	0.9914			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 9, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:00	End Time (MST):	14:11
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999992	≥ 0.995
8.09	8.10	0.9988	Slope	1.001042	$0.90 - 1.10$
4.05	4.03	1.0041	Intercept	-0.005928	± 0.5
2.02	2.02	1.0009			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 9, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:00	End Time (MST):	14:11
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999973	≥ 0.995
9.18	9.17	1.0010	Slope	0.997823	$0.90 - 1.10$
4.60	4.62	0.9948	Intercept	0.022073	± 0.5
2.29	2.33	0.9835			



NMHC Calibration Plot

Date: July 7, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay
Calibration Date: July 29, 2025
Start time (MST): 10:52
Reason: Cylinder Change
Station number: AMS 01
Last Cal Date: July 7, 2025
End time (MST): 13:40

Calibration Standards

Gas Cert Reference: CC418809
CH₄ Cal Gas Conc. 497.2 ppm
C₃H₈ Cal Gas Conc. 205.3 ppm
Removed Gas Cert: NA
Removed CH₄ Conc. 497.2 ppm
Removed C₃H₈ Conc. 205.3 ppm
Diff between cyl (CH₄):
Calibrator Model: Teledyne API T700
Zero Air Gen model: Teledyne API T701
Cal Gas Expiry Date: March 10, 2031
CH₄ Equiv Conc. 1061.8 ppm
Removed Gas Expiry: NA
CH₄ Equiv Conc. 1061.8 ppm
Diff between cyl (THC):
Diff between cyl (NM):
Serial Number: 3565
Serial Number: 146

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm
Analyzer serial #: 1193585648
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.48E-04	2.48E-04	NMHC SP Ratio:	4.96E-05	4.96E-05
CH ₄ Retention time:	15.0	15.0	NMHC Peak Area:	185016	185016
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	17.27	17.44	0.990
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.44	Prev response	17.27	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	17.27	17.43	0.991
Average Correction Factor					

Notes: Changed the N₂/H₂ cylinders after as founds.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	9.18	9.40	0.977
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.40	Prev response	9.18	*% change	2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	9.18	9.33	0.984
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	8.09	8.05	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.05	Prev response	8.09	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	8.09	8.10	0.998
Average Correction Factor					

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999310	
THC Cal Offset:	0.016545	
CH ₄ Cal Slope:	1.001042	
CH ₄ Cal Offset:	-0.005928	
NMHC Cal Slope:	0.997823	
NMHC Cal Offset:	0.022073	

Calibration Performed By: Rene Chamberland

NMHC Calibration Plot

Date: July 29, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay
Station number: AMS 01
Calibration Date: July 8, 2025
Last Cal Date: June 6, 2025
Start time (MST): 11:09
End time (MST): 15:58
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC335700
NOX Cal Gas Conc: 59.40 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.40 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: September 1, 2032
NO Cal Gas Conc: 59.20 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.20 ppm
NO gas Diff:
Serial Number: 3565
Serial Number: 146

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.4	-1.8	0.4	----	----
AF High point	4932	67.6	803.1	800.4	2.7	791.8	779.1	12.8	1.0125	1.0250
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 806.8 ppb	NO = 802.3 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -1.7%
Baseline Corr 1st pt	NO _x = 793.2 ppb	NO = 780.9 ppb	<u>As Found Statistics</u>	*Percent Change	NO = -2.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :	NO SI:	NO Int:
			As found NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 7117

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.190	1.191	NO bkgnd or offset:	-3.1	-3.1
NOX coeff or slope:	1.197	1.193	NOX bkgnd or offset:	-2.9	-2.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	8.2	8.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.004708	1.003085
NO _x Cal Offset:	-0.080000	-0.060000
NO Cal Slope:	1.004232	1.003161
NO Cal Offset:	-1.500000	-1.300000
NO ₂ Cal Slope:	1.002436	1.002345
NO ₂ Cal Offset:	1.555533	1.862876

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.4	0.3	----	----
High point	4932	67.6	803.1	800.4	2.7	805.3	802.4	2.9	0.9973	0.9975
Mid point	4966	33.8	401.5	400.2	1.4	403.3	398.7	4.6	0.9956	1.0037
Low point	4983	16.9	200.8	200.1	0.7	201.0	199.2	1.8	0.9989	1.0045
As left zero	5000	0.0	0.0	0.0	0.0	0.7	0.6	0.2	----	----
As left span	4932	67.6	803.1	390.4	412.7	807.5	390.4	417.2	0.9945	1.0000
Average Correction Factor									0.9973	1.0019

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.3	----	----
High GPT point	799.8	402.9	399.6	401.3	0.9958	100.4%
Mid GPT point	799.8	603.0	199.5	203.6	0.9799	102.1%
Low GPT point	799.8	703.8	98.7	101.7	0.9705	103.0%
Average Correction Factor					0.9821	101.8%

Notes: Changed the inlet filter after as founds. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 8, 2025	Previous Calibration:	June 6, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:09	End Time (MST):	15:58
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	-0.1	----	Correlation Coefficient	0.999999	≥0.995
803.1	805.3	0.9973	Slope	1.003085	0.90 - 1.10
401.5	403.3	0.9956	Intercept	-0.060000	+/-20
200.8	201.0	0.9989			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 8, 2025	Previous Calibration:	June 6, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:09	End Time (MST):	15:58
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999914	≥ 0.995
399.6	401.3	0.9958	Slope	1.002345	$0.90 - 1.10$
199.5	203.6	0.9799	Intercept	1.862876	± 20
98.7	101.7	0.9705			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 8, 2025	Previous Calibration:	June 6, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:09	End Time (MST):	15:58
Analyzer make:	Teledyne API T200	Analyzer serial #:	7117

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	0.999990	≥ 0.995
800.4	802.4	0.9975	Slope	1.003161	$0.90 - 1.10$
400.2	398.7	1.0037	Intercept	-1.300000	± 20
200.1	199.2	1.0045			



NO_x Calibration Plot

Date: July 8, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: July 4, 2025 Last Cal Date: June 3, 2025
Start time (MST): 11:16 End time (MST): 14:02
Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 1107
Analyzer Range: 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007286	1.004457	Backgd or Offset:	6.4	6.4
Calibration intercept:	0.300000	0.020000	Coeff or Slope:	1.013	1.013

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/Ic-AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	5000	863.1	400.0	402.6	0.993
As found Mid point					
As found Low point					
Baseline Corr As found:	402.8	Previous response	403.2	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	5000	863.1	400.0	401.5	0.996
Mid point	5000	744.0	200.0	201.6	0.992
Low point	5000	651.7	100.0	100.3	0.997
As left zero	5000	0.0	0.0	0.0	----
As left span	5000	863.1	400.0	405.9	0.985
Average Correction Factor					0.995

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	July 4, 2025	Previous Calibration:	June 3, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:16	End Time (MST):	14:02
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999993	≥0.995
400.0	401.5	0.9963	Slope	1.004457	0.90 - 1.10
200.0	201.6	0.9921	Intercept	0.020000	+/- 5
100.0	100.3	0.9970			



O₃ Calibration Plot

Date: July 4, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
Calibration Date: July 14, 2025 Last Cal Date: June 25, 2025
Start time (MST): 11:30 End time (MST): 12:12

Analyzer Make: Teledyne API T640 S/N: 322
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752
Temp/RH standard: Alicat FP-25BT S/N: 388752

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	17.6	16.6	17.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	740.6	743.14	740.6	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.07	5.127	5.07	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	46		46	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA:	4.1	PM w/ HEPA:	4.1	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: May 14, 2025
Date Disposable Filter Changed: May 14, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 24, 2024
Date RH/T Sensor Cleaned: May 14, 2025

Notes: Flow, temperature, and pressure were verified. Leak check passed.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

Nt - NOX - NH3 Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
NOX Cal Date:	July 10, 2025	Last Cal Date:	June 26, 2025
Start time (MST):	11:18	End time (MST):	15:43
NH3 Cal Date:	July 11, 2025	Last Cal Date:	June 26, 2025
Start time (MST):	11:10	End time (MST):	14:15
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	59.40	ppm	NO Gas Cylinder #:	CC335700
NO Cal Gas Conc:	59.20	ppm	NO Cal Gas Expiry:	September 1, 2032
Removed NOX Conc:	59.40	ppm	Removed Cylinder #:	NA
Removed NO Conc:	59.20	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	77.80	ppm	NH3 Gas Cylinder #:	CC711249
			NH3 Cal Gas Expiry:	December 31, 2025
Removed NH3 Conc:	77.80	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	API T700		Serial Number:	3565
ZAG make/model:	API T701		Serial Number:	146

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	475
Converter model:	API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	4.90
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	530

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	1.073	1.048	Nt coefficient:	1.080	1.063
NOX coefficient:	1.073	1.053	NO bkgnd:	-3.0	-3.0
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-2.8	-2.8
NH3 coefficient:	0.978	0.978	Nt bkgnd:	-1.9	-1.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999229	1.001221
NO _x Cal Offset:	-3.680000	-3.580000
NO Cal Slope:	1.001491	1.001262
NO Cal Offset:	-3.840000	-4.360000
NO ₂ Cal Slope:	0.990395	0.994509
NO ₂ Cal Offset:	-1.403149	-1.233420
NH3 Cal Slope:	0.998778	1.000525
NH3 Cal Offset:	0.855341	5.113005
Nt Cal Slope:	1.002337	1.003858
Nt Cal Offset:	1.317737	6.087688



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

NO_x / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO _x Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>	Baseline corr NO Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>
As found zero	5000	0.0	0.0	0.0	0.0	1.1	0.9	1.6	----	----
As found span	4932	67.6	803.2	800.4	803.2	821.7	818.0	820.3	0.9774	0.9785
AF GPT span	4932	67.6	803.2	-----	803.2	808.0	-----	809.9	0.9940	-----
new NO cyl rp										

Baseline Corr As Fd Nt = 818.7 ppb NO_x = 820.6 ppb NO = 817.1 ppb

Previous Response Nt = 806.35 ppb NO_x = 798.9 ppb NO = 797.8 ppb

****NO_x Δ (NO to GPT response) = -1.7%**

*** = > +/-2% difference initiates investigation**

***Percent Change Nt_(NO) = 1.5%**

***Percent Change NO_x = 2.7%**

***Percent Change NO = 2.4%**

*** = > +/-5% change initiates investigation**

NO_x / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	0.1	----	----
High point	4932	67.6	803.1	800.4	803.1	803.0	799.9	802.8	1.0001	1.0006
Mid point	4966	33.8	401.5	400.2	401.5	394.6	392.3	395.7	1.0176	1.0201
Low point	4983	16.9	200.8	200.1	200.8	195.1	192.6	195.0	1.0291	1.0389
Average Correction Factor									1.0156	1.0199

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Calibration zero	----	----	0.0	-0.1	----	----
High GPT point (400 ppb O3)	797.0	390.1	409.6	406.5	1.0076	99.2%
Mid GPT point (200 ppb O3)	797.0	592.0	207.7	205.4	1.0112	98.9%
Low GPT point (100 ppb O3)	797.0	692.5	107.2	103.8	1.0328	96.8%

Average Correction Factor	1.0172	98.3%
---------------------------	--------	-------



Wood Buffalo Environmental Association NH₃ - N_T Calibration Report

NH₃ As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH3 Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.1	0.3	-0.2	----	----
AF High point	2931	69.4	1799.5	----	1799.5	1808.1	----	1802.0	0.995	0.999
AF Mid point							----			
AF Low point							----			
new NH3 cyl rp										
Baseline Corr As Fd	Nt = 1808.0 ppb		NH3 = 1802.2 ppb					*Percent Change	Nt _(NH3) = 0.2%	
Previous Response	Nt = 1805.1 ppb		NH3 = 1798.2 ppb				* = > +/-5% change initiates investigation	*Percent Change	NH3 = 0.2%	

NH₃ Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
High point	2931	69.4	1799.8	----	1799.8	1808.1	----	1802.0	0.995	0.999
Mid point	2961	38.6	1001.0	----	1001.0	1017.2	----	1011.5	0.984	0.990
Low point	2981	19.3	500.5	----	500.5	513.0	----	510.0	0.976	0.981
Average Correction Factor									0.9851	0.9899
NH3 Previous Converter Efficiency =		97.8 %								
NH3 Current Converter Efficiency =		97.8 %								

Notes:

Changed the inlet filter after as founds. Adjusted the NOx, NO, and NT span.

Calibration Performed By:

Rene Chamberland



Wood Buffalo Environmental Association


Nt Calibration Summary

Station Information

Calibration Date:	July 11, 2025	Previous Calibration:	June 26, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:18	End Time (MST):	15:43
Analyzer make:	API T201	Analyzer serial #:	824

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999934	≥ 0.995
1799.8	1808.1	0.9954	Slope	1.003858	$0.90 - 1.10$
1001.0	1017.2	0.9841	Intercept	6.087688	± 20
500.5	513.0	0.9757			





Wood Buffalo Environmental Association


NH₃ Calibration Summary

Station Information

Calibration Date:	July 11, 2025	Previous Calibration:	June 26, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:18	End Time (MST):	15:43
Analyzer make:	API T201	Analyzer serial #:	824

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999954	≥ 0.995
1799.8	1802.0	0.9988	Slope	1.000525	$0.90 - 1.10$
1001.0	1011.5	0.9896	Intercept	5.113005	± 20
500.5	510.0	0.9814			





Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 26, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:18	End Time (MST):	15:43
Analyzer make:	API T201	Analyzer serial #:	824

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999886	≥0.995
803.1	803.0	1.0001	Slope	1.001221	0.90 - 1.10
401.5	394.6	1.0176	Intercept	-3.580000	+/-20
200.8	195.1	1.0291			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 26, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:18	End Time (MST):	15:43
Analyzer make:	API T201	Analyzer serial #:	824

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999838	≥ 0.995
800.4	799.9	1.0006	Slope	1.001262	$0.90 - 1.10$
400.2	392.3	1.0201	Intercept	-4.360000	± 20
200.1	192.6	1.0389			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 26, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:18	End Time (MST):	15:43
Analyzer make:	API T201	Analyzer serial #:	824

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999956	≥ 0.995
409.6	406.5	1.0076	Slope	0.994509	$0.90 - 1.10$
207.7	205.4	1.0112	Intercept	-1.233420	± 20
107.2	103.8	1.0328			



NO_x Calibration Plot

Date: July 10, 2025


Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: July 11, 2025

Location: Bertha Ganter-Fort McKay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Mildred Lake
Calibration Date: July 21, 2025
Start time (MST): 9:38
Reason: Routine

Station number: AMS 02
Last Cal Date: June 4, 2025
End time (MST): 13:28

Calibration Standards

Cal Gas Concentration: 50.99 ppm
Cal Gas Cylinder #: EB0112903
Removed Cal Gas Conc: 50.99 ppm
Removed Gas Cyl #:
Calibrator Model: Teledyne API T700
Zero Air Gen Model: Teledyne API T701

Cal Gas Exp Date: October 9, 2032
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 1185
Serial Number: 4891

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0-1000 ppb

Serial Number: JC1404901075

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003722	1.001940	Backgd or Offset:	24.0	23.7
Calibration intercept:	-0.730267	-0.631255	Coeff or Slope:	0.769	0.739

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.9	----
As found High point	4913	78.6	803.0	825.0	0.974
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	824.1	Previous response	805.2	*% change	2.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4913	78.4	801.0	802.0	0.999
Mid point	4961	39.2	399.8	400.3	0.999
Low point	4980	19.6	199.9	198.4	1.007
As left zero	5000	0.0	0.0	-0.3	----
As left span	4913	78.4	801.0	802.0	0.999
Average Correction Factor:					1.002

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 21, 2025	Previous Calibration:	June 4, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:38	End Time (MST):	13:28
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999994	≥0.995
801.0	802.0	0.9987	Slope	1.001940	0.90 - 1.10
399.8	400.3	0.9987	Intercept	-0.631255	+/-30
199.9	198.4	1.0075			



SO2 Calibration Plot

Date: July 21, 2025

Location: Mildred Lake





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Mildred Lake
Calibration Date: July 29, 2025
Start time (MST): 9:00
Reason: Routine

Station number: AMS 02
Last Cal Date: June 25, 2025
End time (MST): 13:40

Calibration Standards

Cal Gas Concentration: 4.75 ppm
Cal Gas Cylinder #: CC700774
Removed Cal Gas Conc: 4.75 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 1185
Serial Number: 4891

Analyzer Information

Analyzer make: Thermo 43iQTL
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12333331546
Converter serial #: 2023-267
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002125	1.005269	Backgd or Offset:	1.40	1.38
Calibration intercept:	-0.020000	-0.080000	Coeff or Slope:	0.964	0.953

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4916	84.2	80.0	80.6	0.996
As found Mid point	4958	42.1	40.0	40.6	0.992
As found Low point	4979	21.1	20.0	20.3	1.000
New cylinder response					
Baseline Corr As found:	80.3	Prev response:	80.14	*% change:	0.2%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.004411	AF Intercept:	0.300000
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999993	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4916	84.2	80.0	80.3	0.996
Mid point	4958	42.1	40.0	40.2	0.995
Low point	4979	21.1	20.0	20.0	1.000
As left zero	5000	0.0	0.0	-0.1	----
As left span	4916	84.2	80.0	80.4	0.995
SO2 Scrubber Check	4920	80.2	802.0	0.1	----
Date of last scrubber change:		July 16, 2024		Ave Corr Factor	0.997
Date of last converter efficiency test:		NA			

Notes: Changed sample inlet filter after as founds. Adjusted span. Scrubber check completed after calibrator zero point.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	July 29, 2025	Previous Calibration:	June 25, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:00	End Time (MST):	13:40
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331546

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999998		≥ 0.995
80.0	80.3	0.9961	Slope	1.005269		$0.90 - 1.10$
40.0	40.2	0.9949	Intercept	-0.080000		± 3
20.0	20.0	0.9999				



H2S Calibration Plot

Date: July 29, 2025

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Mildred Lake
 Calibration Date: July 21, 2025
 Start time (MST): 9:38
 Reason: Routine

Station number: AMS 02
 Last Cal Date: June 9, 2025
 End time (MST): 13:28

Calibration Standards

Gas Cert Reference:	EB0112903	Cal Gas Expiry Date:	October 9, 2032
CH4 Cal Gas Conc.	503.1 ppm	CH4 Equiv Conc.	1067.1 ppm
C3H8 Cal Gas Conc.	205.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	503.1 ppm	CH4 Equiv Conc.	1067.1 ppm
Removed C3H8 Conc.	205.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
Zero Air Gen model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1180320037
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.38E-04	3.82E-04	NMHC SP Ratio:	5.26E-05
CH4 Retention time:	16.4	16.8	NMHC Peak Area:	168464
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4913	78.4	16.76	15.59	1.075
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	15.59	Prev response	16.71	*% change	-7.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	
High point	4913	78.4	16.76	16.76	1.000
Mid point	4961	39.2	8.37	8.28	1.010
Low point	4980	19.6	4.18	4.09	1.023
As left zero	5000	0.0	0.00	0.00	----
As left span	4913	78.4	16.76	16.73	1.002
Average Correction Factor					1.011

Notes:

Changed sample inlet filter after as founds. Adjusted span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4913	78.4	8.86	8.11	1.092
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.11	Prev response	8.81	*% change	-8.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4913	78.4	8.86	8.86	1.000
Mid point	4961	39.2	4.42	4.39	1.008
Low point	4980	19.6	2.21	2.16	1.022
As left zero	5000	0.0	0.00	0.00	----
As left span	4913	78.4	8.86	8.81	1.006
Average Correction Factor					1.010

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4913	78.4	7.90	7.48	1.057
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.48	Prev response	7.90	*% change	-5.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4913	78.4	7.90	7.90	1.000
Mid point	4961	39.2	3.94	3.90	1.012
Low point	4980	19.6	1.97	1.93	1.023
As left zero	5000	0.0	0.00	0.00	----
As left span	4913	78.4	7.90	7.92	0.998
Average Correction Factor					1.012

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.998843	1.001137
THC Cal Offset:	-0.028795	-0.053605
CH ₄ Cal Slope:	1.001748	1.000855
CH ₄ Cal Offset:	-0.014363	-0.026777
NMHC Cal Slope:	0.996394	1.001440
NMHC Cal Offset:	-0.014231	-0.026025

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 21, 2025	Previous Calibration:	June 9, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:38	End Time (MST):	13:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999951	≥ 0.995
16.76	16.76	1.0002	Slope	1.001137	$0.90 - 1.10$
8.37	8.28	1.0104	Intercept	-0.053605	± 0.5
4.18	4.09	1.0225			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 21, 2025	Previous Calibration:	June 9, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:38	End Time (MST):	13:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999942	<i>≥0.995</i>
7.90	7.90	1.0004	Slope	1.000855	<i>0.90 - 1.10</i>
3.94	3.90	1.0124	Intercept	-0.026777	<i>+/-0.5</i>
1.97	1.93	1.0229			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 21, 2025	Previous Calibration:	June 9, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:38	End Time (MST):	13:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320037

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999961	<i>≥0.995</i>
8.86	8.86	1.0000	Slope	1.001440	<i>0.90 - 1.10</i>
4.42	4.39	1.0077	Intercept	-0.026025	<i>+/-0.5</i>
2.21	2.16	1.0222			



NMHC Calibration Plot

Date: July 21, 2025

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Mildred Lake
Calibration Date: July 22, 2025
Start time (MST): 9:53
Reason: Removal

Station number: AMS 02
Last Cal Date: July 21, 2025
End time (MST): 11:42

Calibration Standards

Gas Cert Reference:	EB0112903	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	503.1 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	503.1 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
Zero Air Gen model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1180320037
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.82E-04	NA	NMHC SP Ratio:	6.24E-05
CH ₄ Retention time:	16.8	NA	NMHC Peak Area:	141917
Zero Chromatogram:	ON		Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4913	78.4	16.76	16.82	0.997
As found Mid point	4961	39.2	8.37	8.31	1.007
As found Low point	4980	19.6	4.18	4.11	1.019
New cylinder response					
Baseline Corr AF:	16.82	Prev response	16.73	*% change	0.5%
Baseline Corr 2nd AF:	8.31	AF Slope:	1.004737	AF Intercept:	-0.053983
Baseline Corr 3rd AF:	4.11	AF Correlation:	0.999951	* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Notes:

Swapped 55i due to CH₄ dips seen during sampling.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4913	78.4	8.86	8.91	0.994
As found Mid point	4961	39.2	4.42	4.41	1.003
As found Low point	4980	19.6	2.21	2.18	1.016
New cylinder response					
Baseline Corr AF:	8.91	Prev response	8.85	*% change	0.7%
Baseline Corr 2nd AF:	4.41	AF Slope:	1.006987	AF Intercept:	-0.026009
Baseline Corr 3rd AF:	2.18	AF Correlation:	0.999961	* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4913	78.4	7.90	7.91	0.999
As found Mid point	4961	39.2	3.94	3.90	1.011
As found Low point	4980	19.6	1.97	1.93	1.023
New cylinder response					
Baseline Corr AF:	7.91	Prev response	7.88	*% change	0.4%
Baseline Corr 2nd AF:	3.90	AF Slope:	1.002764	AF Intercept:	-0.028373
Baseline Corr 3rd AF:	1.93	AF Correlation:	0.999936	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001137	
THC Cal Offset:	-0.053605	
CH ₄ Cal Slope:	1.000855	
CH ₄ Cal Offset:	-0.026777	
NMHC Cal Slope:	1.001440	
NMHC Cal Offset:	-0.026025	

Calibration Performed By: Braiden Boutilier

NMHC Calibration Plot

Date: July 22, 2025

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Mildred Lake
 Calibration Date: July 23, 2025
 Start time (MST): 10:05
 Reason: Install

Station number: AMS 02
 Last Cal Date:
 End time (MST): 12:30

Calibration Standards

Gas Cert Reference:	EB0112903	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	503.1 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	503.1 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
Zero Air Gen model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1180320040
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	NA	4.23E-04	NMHC SP Ratio:	NA	7.62E-05
CH ₄ Retention time:	NA	15.6	NMHC Peak Area:	NA	116215
Zero Chromatogram:		OFF	Flat Baseline:		OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	
High point	4913	78.4	16.76	16.73	1.002
Mid point	4961	39.2	8.37	8.37	1.000
Low point	4980	19.6	4.18	4.17	1.004
As left zero	5000	0.0	0.00	0.00	----
As left span	4913	78.4	16.76	16.81	0.997
Average Correction Factor					1.002

Notes:

Install calibration was unsuccessful yesterday due to some contamination from N2 cylinder change.
 Column bake out seems to have cleared out the contamination. Adjusted the span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4913	78.4	8.86	8.85	1.001
Mid point	4961	39.2	4.42	4.46	0.991
Low point	4980	19.6	2.21	2.24	0.986
As left zero	5000	0.0	0.00	0.00	----
As left span	4913	78.4	8.86	8.89	0.997
Average Correction Factor					0.993

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4913	78.4	7.90	7.88	1.003
Mid point	4961	39.2	3.94	3.90	1.011
Low point	4980	19.6	1.97	1.93	1.023
As left zero	5000	0.0	0.00	0.00	----
As left span	4913	78.4	7.90	7.93	0.997
Average Correction Factor					1.013

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		0.998417
THC Cal Offset:		-0.000169
CH ₄ Cal Slope:		0.997845
CH ₄ Cal Offset:		-0.021378
NMHC Cal Slope:		0.998630
NMHC Cal Offset:		0.021607

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:05	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995
16.76	16.73	1.0018	Slope	0.998417	$0.90 - 1.10$
8.37	8.37	1.0002	Intercept	-0.000169	± 0.5
4.18	4.17	1.0036			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:05	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999966	<i>≥0.995</i>
7.90	7.88	1.0034	Slope	0.997845	<i>0.90 - 1.10</i>
3.94	3.90	1.0111	Intercept	-0.021378	<i>+/-0.5</i>
1.97	1.93	1.0234			





Wood Buffalo Environmental Association


NMHC Calibration Summary


Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:05	End Time (MST):	12:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999966	<i>≥0.995</i>
8.86	8.85	1.0007	Slope	0.998630	<i>0.90 - 1.10</i>
4.42	4.46	0.9910	Intercept	0.021607	<i>+/-0.5</i>
2.21	2.24	0.9862			







WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
Calibration Date: July 4, 2025 Last Cal Date: June 12, 2025
Start time (MST): 5:36 End time (MST): 8:17
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.87 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC446753
Removed Cal Gas Conc: 50.87 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3808
Zero Air Gen Model: Teledyne API T701 Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1327300932
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997429	0.997828	Backgd or Offset:	27.6	27.2
Calibration intercept:	0.154012	0.814546	Coeff or Slope:	0.867	0.867

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.5	----
As found High point	4921	78.6	799.7	794.8	1.006
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	795.3	Previous response	797.8	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4921	78.6	799.7	798.5	1.002
Mid point	4961	39.3	399.8	400.1	0.999
Low point	4980	19.6	199.4	200.4	0.995
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	78.6	799.7	797.8	1.002
Average Correction Factor:					0.999

Notes: Zero adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 4, 2025	Previous Calibration:	June 12, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:36	End Time (MST):	8:17
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1327300932

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999997	≥0.995
799.7	798.5	1.0016	Slope	0.997828	0.90 - 1.10
399.8	400.1	0.9993	Intercept	0.814546	+/-30
199.4	200.4	0.9951			



SO2 Calibration Plot

Date: July 4, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: July 24, 2025
Start time (MST): 7:00
Reason: Routine

Station number: AMS 04
Last Cal Date: June 10, 2025
End time (MST): 10:40

Calibration Standards

Cal Gas Concentration: 4.80 ppm
Cal Gas Cylinder #: DT0037528
Removed Cal Gas Conc: 4.80 ppm
Removed Gas Cyl #:
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i-LTE
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1008841400
Converter serial #: 2022-200
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004149	1.010581	Backgd or Offset:	1.89	1.89
Calibration intercept:	0.138203	0.038163	Coeff or Slope:	1.100	1.100

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4917	83.3	80.0	82.0	0.974
As found Mid point	4958	41.7	40.0	41.0	0.974
As found Low point	4979	20.8	20.0	20.5	0.969
New cylinder response					
Baseline Corr As found:	82.1	Prev response:	80.43	*% change:	2.0%
Baseline Corr 2nd AF pt:	41.1	AF Slope:	1.026297	AF Intercept:	-0.061777
Baseline Corr 3rd AF pt:	20.6	AF Correlation:	0.999998	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4917	83.3	80.0	80.8	0.990
Mid point	4958	41.7	40.0	40.6	0.986
Low point	4979	20.8	20.0	20.2	0.989
As left zero	5000	0.0	0.0	0.1	----
As left span	4917	83.3	80.0	79.8	1.002
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23		Ave Corr Factor		0.988
Date of last converter efficiency test:					

Notes: Sox scrubber checked after calibrator zero. No adjustment made.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	July 24, 2025	Previous Calibration:	June 10, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:00	End Time (MST):	10:40
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999996		≥0.995
80.0	80.8	0.9896	Slope	1.010581		0.90 - 1.10
40.0	40.6	0.9861	Intercept	0.038163		+/-3
20.0	20.2	0.9886				



H₂S Calibration Plot

Date: July 24, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Buffalo Viewpoint
 Calibration Date: July 4, 2025
 Start time (MST): 5:36
 Reason: Routine

Station number: AMS 04
 Last Cal Date: June 12, 2025
 End time (MST): 8:14

Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3808
Zero Air Gen model:	Teledyne API T701	Serial Number:	362

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1426262594
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.78E-04	4.73E-04	NMHC SP Ratio:	9.87E-04
CH ₄ Retention time:	14.1	14.1	NMHC Peak Area:	89385
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	16.64	16.42	1.013
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.42	Prev response	16.64	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	16.64	16.63	1.000
Mid point	4961	39.3	8.32	8.27	1.006
Low point	4980	19.6	4.15	4.08	1.018
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	16.64	16.58	1.004
Average Correction Factor					1.008

Notes:

Span adjusted. Hydrogen Cylinder changed.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	78.6	8.82	8.48	1.041
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.48	Prev response	8.84	*% change	-4.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	78.6	8.82	8.83	0.999
Mid point	4961	39.3	4.41	4.41	0.999
Low point	4980	19.6	2.20	2.16	1.018
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	8.82	8.77	1.006
Average Correction Factor					1.005

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	78.6	7.82	7.95	0.983
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.95	Prev response	7.80	*% change	1.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	78.6	7.82	7.80	1.002
Mid point	4961	39.3	3.91	3.85	1.014
Low point	4980	19.6	1.95	1.92	1.017
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	7.82	7.81	1.001
Average Correction Factor					1.011

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.003818	1.000630
THC Cal Offset:	-0.059943	-0.037751
CH ₄ Cal Slope:	1.001879	0.998457
CH ₄ Cal Offset:	-0.031309	-0.021111
NMHC Cal Slope:	1.005550	1.002220
NMHC Cal Offset:	-0.028434	-0.015841

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 4, 2025	Previous Calibration:	June 12, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:36	End Time (MST):	8:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999977	<i>≥0.995</i>
16.64	16.63	1.0005	Slope	1.000630	<i>0.90 - 1.10</i>
8.32	8.27	1.0063	Intercept	-0.037751	<i>+/-0.5</i>
4.15	4.08	1.0178			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 4, 2025	Previous Calibration:	June 12, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:36	End Time (MST):	8:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999951	<i>≥0.995</i>
7.82	7.80	1.0021	Slope	0.998457	<i>0.90 - 1.10</i>
3.91	3.85	1.0145	Intercept	-0.021111	<i>+/-0.5</i>
1.95	1.92	1.0173			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 4, 2025	Previous Calibration:	June 12, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:36	End Time (MST):	8:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1426262594

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999973	<i>≥0.995</i>
8.82	8.83	0.9994	Slope	1.002220	<i>0.90 - 1.10</i>
4.41	4.41	0.9989	Intercept	-0.015841	<i>+/-0.5</i>
2.20	2.16	1.0182			



NMHC Calibration Plot

Date: July 4, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: July 28, 2025
Start time (MST): 5:50
Reason: Removal

Station number: AMS 04
Last Cal Date: July 4, 2025
End time (MST): 7:12

Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3808
Zero Air Gen model:	Teledyne API T701	Serial Number:	362

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1426262594
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.73E-04	4.73E-04	NMHC SP Ratio:	1.04E-04
CH ₄ Retention time:	14.1	14.1	NMHC Peak Area:	85156
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	16.64	15.91	1.045
As found Mid point	4961	39.3	8.32	7.92	1.050
As found Low point	4980	19.8	4.19	3.85	1.089
New cylinder response					
Baseline Corr AF:	15.91	Prev response	16.61	*% change	-4.4%
Baseline Corr 2nd AF:	7.92	AF Slope:	0.959640	AF Intercept:	-0.072143
Baseline Corr 3rd AF:	3.85	AF Correlation:	0.999884	* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Notes:

Removal Due to Noisy NM Channel



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	78.6	8.82	8.25	1.069
As found Mid point	4961	39.3	4.41	4.13	1.067
As found Low point	4980	19.8	2.22	1.97	1.126
New cylinder response					
Baseline Corr AF:	8.25	Prev response	8.82	*% change	-6.9%
Baseline Corr 2nd AF:	4.13	AF Slope:	0.939703	AF Intercept:	-0.040463
Baseline Corr 3rd AF:	1.97	AF Correlation:	0.999785	* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	78.6	7.82	7.66	1.020
As found Mid point	4961	39.3	3.91	3.79	1.032
As found Low point	4980	19.8	1.97	1.87	1.051
New cylinder response					
Baseline Corr AF:	7.66	Prev response	7.78	*% change	-1.6%
Baseline Corr 2nd AF:	3.79	AF Slope:	0.982266	AF Intercept:	-0.031881
Baseline Corr 3rd AF:	1.87	AF Correlation:	0.999919	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000630	
THC Cal Offset:	-0.037751	
CH ₄ Cal Slope:	0.998457	
CH ₄ Cal Offset:	-0.021111	
NMHC Cal Slope:	1.002220	
NMHC Cal Offset:	-0.015841	

Calibration Performed By: Melissa Lemay

NMHC Calibration Plot

Date: July 28, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS 04
Calibration Date:	July 28, 2025	Last Cal Date:	
Start time (MST):	9:05	End time (MST):	11:11
Reason:	Install		

Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3808
Zero Air Gen model:	Teledyne API T701	Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038
THC Range:	0 - 20 ppm	NMHC/CH ₄ Range:	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:		2.97E-04	NMHC SP Ratio:	5.04E-04
CH ₄ Retention time:		15.6	NMHC Peak Area:	174857
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	16.64	16.62	1.001
Mid point	4961	39.3	8.32	8.34	0.998
Low point	4980	19.6	4.15	4.16	0.997
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	16.64	16.63	1.001
Average Correction Factor					0.998

Notes: Installed due to removed NMHC having unstable NM channel.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	8.82	8.81	1.001
Mid point	4961	39.3	4.41	4.44	0.994
Low point	4980	19.6	2.20	2.22	0.992
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	8.82	8.84	0.998
Average Correction Factor					0.996

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	7.82	7.81	1.001
Mid point	4961	39.3	3.91	3.90	1.002
Low point	4980	19.6	1.95	1.95	1.002
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	7.82	7.79	1.004
Average Correction Factor					1.002

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		0.998848
THC Cal Offset:		0.012465
CH ₄ Cal Slope:		0.999464
CH ₄ Cal Offset:		-0.002303
NMHC Cal Slope:		0.998562
NMHC Cal Offset:		0.013768

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 28, 2025	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	9:05	End Time (MST):	11:11
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999996	≥ 0.995
16.64	16.62	1.0010	Slope	0.998848	0.90 - 1.10
8.32	8.34	0.9977	Intercept	0.012465	± 0.5
4.15	4.16	0.9968			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 28, 2025	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	9:05	End Time (MST):	11:11
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995
7.82	7.81	1.0006	Slope	0.999464	$0.90 - 1.10$
3.91	3.90	1.0020	Intercept	-0.002303	± 0.5
1.95	1.95	1.0021			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 28, 2025	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	9:05	End Time (MST):	11:11
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999984	<i>≥0.995</i>
8.82	8.81	1.0011	Slope	0.998562	<i>0.90 - 1.10</i>
4.41	4.44	0.9942	Intercept	0.013768	<i>+/-0.5</i>
2.20	2.22	0.9925			



NMHC Calibration Plot

Date: July 28, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Station number: AMS 04
Calibration Date: July 23, 2025
Last Cal Date: June 18, 2025
Start time (MST): 6:06
End time (MST): 10:54
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC324979
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 3, 2032
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3808
Serial Number: 362

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.3	0.5	----	----
AF High point	4918	81.8	800.0	798.4	1.6	769.4	760.8	8.6	1.0400	1.0490
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 804.1 ppb	NO = 798.7 ppb							*Percent Change	NO _x = -4.5%
Baseline Corr 1st pt	NO _x = 769.3 ppb	NO = 761.1 ppb							*Percent Change	NO = -4.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb								
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb								
As Found Statistics										
	As found	NO _x r ² :							Nx SI:	Nx Int:
	As found	NO r ² :							NO SI:	NO Int:
	As found	NO ₂ r ² :							NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 721

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.380	1.440	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.375	1.434	NOX bkgnd or offset:	-0.2	-0.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.6	4.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002948	0.997835
NO _x Cal Offset:	1.708117	1.947157
NO Cal Slope:	0.999666	0.998465
NO Cal Offset:	0.546668	-0.034131
NO ₂ Cal Slope:	1.006266	0.996541
NO ₂ Cal Offset:	1.380378	1.708938

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.4	0.0	0.4	----	----
High point	4918	81.8	800.0	798.4	1.6	799.5	797.0	2.5	1.0007	1.0018
Mid point	4959	40.9	400.0	399.2	0.8	401.8	399.0	2.8	0.9955	1.0005
Low point	4980	20.4	199.5	199.1	0.4	202.6	198.4	4.2	0.9847	1.0035
As left zero	5000	0.0	0.0	1.1	-1.1	0.9	1.1	-0.2	----	----
As left span	4918	81.8	800.0	408.1	800.0	796.6	408.1	388.5	1.0043	1.0000
Average Correction Factor									0.9936	1.0019

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.4	----	----
High GPT point	794.5	406.9	389.2	388.8	1.0011	99.9%
Mid GPT point	794.5	602.7	193.4	195.6	0.9889	101.1%
Low GPT point	794.5	701.1	95.0	97.4	0.9757	102.5%
Average Correction Factor					0.9886	101.2%

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	June 18, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	6:06	End Time (MST):	10:54
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999983	≥0.995
800.0	799.5	1.0007	Slope	0.997835	0.90 - 1.10
400.0	401.8	0.9955	Intercept	1.947157	+/-20
199.5	202.6	0.9847			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	June 18, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	6:06	End Time (MST):	10:54
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999944	≥0.995
389.2	388.8	1.0011	Slope	0.996541	0.90 - 1.10
193.4	195.6	0.9889	Intercept	1.708938	+/-20
95.0	97.4	0.9757			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	June 18, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	6:06	End Time (MST):	10:54
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999999	≥ 0.995
798.4	797.0	1.0018	Slope	0.998465	$0.90 - 1.10$
399.2	399.0	1.0005	Intercept	-0.034131	± 20
199.1	198.4	1.0035			



NO_x Calibration Plot

Date: July 23, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: July 4, 2025
Start time (MST): 8:12
Reason: Routine

Station number: AMS 04
Last Cal Date: June 24, 2025
End time (MST): 10:15

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: Teledyne API T400
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 2961

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996400	1.003886	Backgd or Offset:	-2.2	-2.2
Calibration intercept:	1.680000	1.620000	Coeff or Slope:	1.019	1.019

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	1.3	----
As found High point	5000	1011.6	400.0	401.5	1.000
As found Mid point					
As found Low point					
Baseline Corr As found:	400.2	Previous response	400.2	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	1.5	----
High point	5000	1014.0	400.0	403.1	0.992
Mid point	5000	831.4	200.0	202.5	0.988
Low point	5000	720.0	100.0	102.1	0.979
As left zero	5000	0.0	0.0	1.8	----
As left span	5000	1014.0	400.0	403.8	0.991
Average Correction Factor					0.986

Notes: No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	July 4, 2025	Previous Calibration:	June 24, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	8:12	End Time (MST):	10:15
Analyzer make:	Teledyne API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.5	----	Correlation Coefficient	1.000000	≥ 0.995
400.0	403.1	0.9923	Slope	1.003886	$0.90 - 1.10$
200.0	202.5	0.9877	Intercept	1.620000	± 5
100.0	102.1	0.9794			



O₃ Calibration Plot

Date: July 4, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: July 24, 2025
Start time (MST): 6:32
Station number: AMS 04
Last Cal Date: June 24, 2025
End time (MST): 6:59
Analyzer Make: Teledyne API T640
Particulate Fraction: PM2.5
S/N: 321
Flow Meter Make/Model: Alicat FP-25BT
Temp/RH standard: Alicat FP-25BT
S/N: 388753
S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	14.0	13.6	14.0	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	724.8	727.1	724.8	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.03	5.09	5.03	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	36	----	36	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA:	0.9	PM w/ HEPA:	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 30-Jan-27
Lot No.: 100128-050-051

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: June 24, 2025
Date Disposable Filter Changed: June 24, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: June 24, 2025
Date RH/T Sensor Cleaned: June 24, 2025

No adjustments done.

Notes:

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS05
MANNIX
JULY 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Mannix
Calibration Date: July 10, 2025
Start time (MST): 9:29
Reason: Routine

Station number: AMS 05
Last Cal Date: June 11, 2025
End time (MST): 14:15

Calibration Standards

Cal Gas Concentration: 50.06 ppm
Cal Gas Cylinder #: CC308040
Removed Cal Gas Conc: 50.06 ppm
Removed Gas Cyl #:
Calibrator Model: API T700
Zero Air Gen Model: API T701

Cal Gas Exp Date: October 22, 2032
Rem Gas Exp Date: October 22, 2032
Diff between cyl:
Serial Number: 5470
Serial Number: 361

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 1000 ppb

Serial Number: 1008841399

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002165	1.002152	Backgd or Offset:	10.1	10.1
Calibration intercept:	-0.156926	-0.377188	Coeff or Slope:	0.939	0.939

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.7	----
As found High point	4920	79.9	800.0	799.4	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	798.7	Previous response	801.6	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4920	79.9	800.0	801.6	0.998
Mid point	4960	40.0	400.5	400.4	1.000
Low point	4980	20.0	200.2	200.3	1.000
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	79.9	800.0	802.7	0.997
Average Correction Factor:					0.999

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 11, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:29	End Time (MST):	14:15
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999999	≥0.995
800.0	801.6	0.9980	Slope	1.002152	0.90 - 1.10
400.5	400.4	1.0002	Intercept	-0.377188	+/-30
200.2	200.3	0.9997			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: July 10, 2025

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Mannix
Calibration Date: July 8, 2025
Start time (MST): 9:51
Reason: Routine

Station number: AMS 05
Last Cal Date: June 20, 2025
End time (MST): 14:10

Calibration Standards

Cal Gas Concentration: 4.96 ppm
Cal Gas Cylinder #: DT0037363
Removed Cal Gas Conc: 4.96 ppm
Removed Gas Cyl #: N/A
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: N/A
Diff between cyl:
Serial Number: 5470
Serial Number: 361

Analyzer Information

Analyzer make: Thermo 43iQ
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1200326169
Converter serial #: 2022-225
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998973	0.997832	Backgd or Offset:	1.25	1.25
Calibration intercept:	0.142351	0.182274	Coeff or Slope:	1.029	1.029

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4919	80.6	80.0	81.1	0.988
As found Mid point	4960	40.3	40.0	40.8	0.985
As found Low point	4980	20.2	20.0	20.1	1.007
New cylinder response					
Baseline Corr As found:	80.9	Prev response:	80.02	*% change:	1.1%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.013553	AF Intercept:	0.082148
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999961	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4919	80.6	80.0	79.9	1.001
Mid point	4960	40.3	40.0	40.2	0.994
Low point	4980	20.2	20.0	20.2	0.992
As left zero	5000	0.0	0.0	0.2	----
As left span	4919	80.6	80.0	79.4	1.007
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.996
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	July 8, 2025	Previous Calibration:	June 20, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:51	End Time (MST):	14:10
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1200326169

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999992		≥ 0.995
80.0	79.9	1.0008	Slope	0.997832		$0.90 - 1.10$
40.0	40.2	0.9944	Intercept	0.182274		± 3
20.0	20.2	0.9920				



H₂S Calibration Plot

Date: July 8, 2025

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Mannix
Calibration Date: July 10, 2025
Start time (MST): 9:29
Reason: Routine

Station number: AMS 05
Last Cal Date: June 11, 2025
End time (MST): 14:15

Calibration Standards

Gas Cert Reference: CC308040
CH₄ Cal Gas Conc. 500.3 ppm
C₃H₈ Cal Gas Conc. 199.0 ppm
Removed Gas Cert:
Removed CH₄ Conc. 500.3 ppm
Removed C₃H₈ Conc. 199.0 ppm
Diff between cyl (CH₄):
Calibrator Model: API T700
Zero Air Gen model: API T701

Cal Gas Expiry Date: October 22, 2032
CH₄ Equiv Conc. 1047.6 ppm
Removed Gas Expiry:
CH₄ Equiv Conc. 1047.6 ppm
Diff between cyl (THC):
Diff between cyl (NM):
Serial Number: 5470
Serial Number: 361

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1193585649
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.96E-04	3.96E-04	NMHC SP Ratio:	8.01E-05	8.05E-05
CH ₄ Retention time:	15.6	15.6	NMHC Peak Area:	109245	109395
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.9	16.74	16.99	0.985
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.99	Prev response	16.62	*% change	2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.9	16.74	16.80	0.997
Mid point	4960	40.0	8.38	8.32	1.008
Low point	4980	20.0	4.19	4.10	1.022
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	16.74	16.88	0.992
Average Correction Factor					1.009

Notes: Changed the H₂, N₂ cylinders and the inlet filter after as founds. Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.9	8.75	8.97	0.975
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.97	Prev response	8.71	*% change	3.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.9	8.75	8.81	0.993
Mid point	4960	40.0	4.38	4.41	0.993
Low point	4980	20.0	2.19	2.19	1.000
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	8.75	8.86	0.987
Average Correction Factor					0.995

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.9	7.99	8.02	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.02	Prev response	7.92	*% change	1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.9	7.99	7.99	1.001
Mid point	4960	40.0	4.00	3.90	1.026
Low point	4980	20.0	2.00	1.91	1.046
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	7.99	8.01	0.998
Average Correction Factor					1.024

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.996741	1.004805
THC Cal Offset:	-0.064064	-0.059908
CH ₄ Cal Slope:	0.996882	1.001185
CH ₄ Cal Offset:	-0.054726	-0.053537
NMHC Cal Slope:	0.996599	1.008206
NMHC Cal Offset:	-0.009538	-0.006971

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 11, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:29	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999939	≥ 0.995
16.74	16.80	0.9967	Slope	1.004805	$0.90 - 1.10$
8.38	8.32	1.0079	Intercept	-0.059908	± 0.5
4.19	4.10	1.0218			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 11, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:29	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999768	≥ 0.995
7.99	7.99	1.0011	Slope	1.001185	$0.90 - 1.10$
4.00	3.90	1.0257	Intercept	-0.053537	± 0.5
2.00	1.91	1.0461			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 11, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:29	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585649

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999995	<i>≥0.995</i>
8.75	8.81	0.9925	Slope	1.008206	<i>0.90 - 1.10</i>
4.38	4.41	0.9925	Intercept	-0.006971	<i>+/-0.5</i>
2.19	2.19	1.0005			



NMHC Calibration Plot

Date: July 10, 2025

Location: Mannix





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: July 4, 2025
Start time (MST): 8:40
Reason: Routine

Station number: AMS 06
Last Cal Date: June 27, 2025
End time (MST): 17:00

Calibration Standards

Cal Gas Concentration: 50.08 ppm
Cal Gas Cylinder #: CC255448
Removed Cal Gas Conc: 50.08 ppm
Removed Gas Cyl #:
Calibrator Model: API T700
Zero Air Gen Model: API T701

Cal Gas Exp Date: October 22, 2032
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 1160290013

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004025	1.000606	Backgd or Offset:	18.2	18.2
Calibration intercept:	1.158164	1.079118	Coeff or Slope:	0.907	0.907

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4920.2	79.8	799.3	797.0	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.3	Previous response	803.7	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4920.2	79.8	799.3	800.1	0.999
Mid point	4960.1	39.9	399.6	402.0	0.994
Low point	4980	20.0	200.3	202.4	0.990
As left zero	5000	0.0	0.0	-0.1	----
As left span	4919.7	80.3	804.3	802.7	1.002
Average Correction Factor:					0.994

Notes: Changed the inlet filter after as founds. No adjustments made. Calibration took longer due to issues with NHMC which was being calibrated at the same time.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 4, 2025	Previous Calibration:	June 27, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:40	End Time (MST):	17:00
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999989	≥0.995
799.3	800.1	0.9990	Slope	1.000606	0.90 - 1.10
399.6	402.0	0.9941	Intercept	1.079118	+/-30
200.3	202.4	0.9897			



SO2 Calibration Plot

Date: July 4, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: July 17, 2025
Start time (MST): 9:12
Reason: Routine

Station number: AMS 06
Last Cal Date: June 5, 2025
End time (MST): 14:15

Calibration Standards

Cal Gas Concentration: 4.760 ppm
Cal Gas Cylinder #: DT0014585
Removed Cal Gas Conc: 4.760 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1218153358
Converter serial #: 517
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.987252	1.025839	Backgd or Offset:	1.99	1.99
Calibration intercept:	0.160000	-0.140000	Coeff or Slope:	1.146	1.146

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4916	84.0	80.0	79.6	1.006
As found Mid point	4958	42.0	40.0	40.1	1.000
As found Low point	4979	21.0	20.0	20.2	0.995
New cylinder response					
Baseline Corr As found:	79.5	Prev response:	79.11	*% change:	0.5%
Baseline Corr 2nd AF pt:	40.0	AF Slope:	0.993540	AF Intercept:	0.240000
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999984	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4916	84.0	80.0	82.0	0.975
Mid point	4958	42.0	40.0	41.0	0.975
Low point	4979	21.0	20.0	19.7	1.015
As left zero	5000	0.0	0.0	0.3	----
As left span	4916	84.0	80.0	83.0	0.963
SO2 Scrubber Check				-0.1	
Date of last scrubber change:	December 20, 2021			Ave Corr Factor	0.988
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made. Points are drifting a bit, suspecting its due to scrubber moisture instability. Instrument diagnostics are normal.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	July 17, 2025	Previous Calibration:	June 5, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:12	End Time (MST):	14:15
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999819		≥ 0.995
80.0	82.0	0.9752	Slope	1.025839		$0.90 - 1.10$
40.0	41.0	0.9752	Intercept	-0.140000		± 3
20.0	19.7	1.0148				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Patricia McInnes
 Calibration Date: July 4, 2025
 Start time (MST): 8:41
 Reason: As Found

Station number: AMS 06
 Last Cal Date: June 3, 2025
 End time (MST): 12:00

Calibration Standards

Gas Cert Reference: CC255448
 CH₄ Cal Gas Conc. 501.4 ppm
 C₃H₈ Cal Gas Conc. 199.3 ppm
 Removed Gas Cert:
 Removed CH₄ Conc. 501.4 ppm
 Removed C₃H₈ Conc. 199.3 ppm
 Diff between cyl (CH₄):
 Calibrator Model: API T700
 Zero Air Gen model: API T701

Cal Gas Expiry Date: October 22, 2032
 CH₄ Equiv Conc. 1049.5 ppm
 Removed Gas Expiry:
 CH₄ Equiv Conc. 1049.5 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 3566
 Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1118148494
 NMHC/CH₄ Range: 0 - 10 ppm

Start
 CH₄ SP Ratio: 4.86E-04
 CH₄ Retention time: 14.4
 Zero Chromatogram: OFF

Finish

Start
 NMHC SP Ratio: 4.38E-05
 NMHC Peak Area: 199812
 Flat Baseline: OFF

Finish

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.04	----
As found High point	4920	79.8	16.75	16.68	1.006
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.64	Prev response	16.82	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.75	16.77	0.999
Mid point	4960	39.9	8.37	8.43	0.993
Low point	4980	20.0	4.20	4.50	0.933
As left zero					
As left span					

Average Correction Factor 0.975

Notes: Changed the inlet filter, N₂ and H₂ cylinders after as founds. Adjusted the zero and span. Third point is failing due to a contaminated cylinder, see docit note for further info.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	8.75	8.63	1.013
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.63	Prev response	8.76	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.75	8.83	0.990
Mid point	4960.1	39.9	4.37	4.44	0.985
Low point	4980	20.0	2.19	2.24	0.978
As left zero					
As left span					

Average Correction Factor 0.985

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.04	Limit = 0.90-1.10
As found High point	4920	79.8	8.00	8.04	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.01	Prev response	8.07	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.00	7.94	1.008
Mid point	4960.1	39.9	4.00	3.99	1.002
Low point	4980	20.0	2.01	2.26	0.889
As left zero					
As left span					

Average Correction Factor 0.966

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.002107	0.995518
THC Cal Offset:	0.039972	0.127524
CH ₄ Cal Slope:	1.006781	0.981106
CH ₄ Cal Offset:	0.012973	0.111143
NMHC Cal Slope:	0.998066	1.008977
NMHC Cal Offset:	0.026598	0.015581

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 4, 2025	Previous Calibration:	June 3, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:41	End Time (MST):	12:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999633	≥ 0.995
16.75	16.77	0.9987	Slope	0.995518	$0.90 - 1.10$
8.37	8.43	0.9935	Intercept	0.127524	± 0.5
4.20	4.50	0.9329			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 4, 2025	Previous Calibration:	June 3, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:41	End Time (MST):	12:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.998619	<i>≥0.995</i>
8.00	7.94	1.0079	Slope	0.981106	<i>0.90 - 1.10</i>
4.00	3.99	1.0023	Intercept	0.111143	<i>+/-0.5</i>
2.01	2.26	0.8886			






Wood Buffalo Environmental Association NMHC Calibration Summary

Station Information

Calibration Date:	July 4, 2025	Previous Calibration:	June 3, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:41	End Time (MST):	12:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995
8.75	8.83	0.9903	Slope	1.008977	$0.90 - 1.10$
4.37	4.44	0.9855	Intercept	0.015581	± 0.5
2.19	2.24	0.9778			



NMHC Calibration Plot

Date: July 4, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: July 5, 2025
Start time (MST): 8:45
Reason: Maintenance

Station number: AMS 06
Last Cal Date: July 4, 2025
End time (MST): 11:36

Calibration Standards

Gas Cert Reference: CC255448
CH₄ Cal Gas Conc. 501.4 ppm
C₃H₈ Cal Gas Conc. 199.3 ppm
Removed Gas Cert:
Removed CH₄ Conc. 501.4 ppm
Removed C₃H₈ Conc. 199.3 ppm
Diff between cyl (CH₄):
Calibrator Model: API T700
Zero Air Gen model: API T701

Cal Gas Expiry Date: October 22, 2032
CH₄ Equiv Conc. 1049.5 ppm
Removed Gas Expiry:
CH₄ Equiv Conc. 1049.5 ppm
Diff between cyl (THC):
Diff between cyl (NM):
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1118148494
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.86E-04	2.60E-04	NMHC SP Ratio:	4.38E-05
CH ₄ Retention time:	14.4	14.2	NMHC Peak Area:	199812
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.75	16.76	0.999
Mid point	4960	39.9	8.37	8.35	1.003
Low point	4980	20.0	4.20	4.20	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.75	16.86	0.993
Average Correction Factor					1.001

Notes:

Calibration failed yesterday due to a contaminated N2 cylinder. Installed new cylinder, adjusted the zero and span. Ran quick points to make sure it passed and full points after.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.75	8.78	0.996
Mid point	4960.1	39.9	4.37	4.40	0.993
Low point	4980	20.0	2.19	2.24	0.979
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.75	8.76	0.999
Average Correction Factor					0.990

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.00	7.98	1.003
Mid point	4960.1	39.9	4.00	3.95	1.013
Low point	4980	20.0	2.01	1.96	1.025
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.00	8.11	0.987
Average Correction Factor					1.014

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.002107	1.000535
THC Cal Offset:	0.039972	-0.008002
CH ₄ Cal Slope:	1.006781	0.998248
CH ₄ Cal Offset:	0.012973	-0.025392
NMHC Cal Slope:	0.998066	1.002405
NMHC Cal Offset:	0.026598	0.017991

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 5, 2025	Previous Calibration:	July 4, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:45	End Time (MST):	11:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999996		≥ 0.995
16.75	16.76	0.9994	Slope	1.000535		0.90 - 1.10
8.37	8.35	1.0029	Intercept	-0.008002		± 0.5
4.20	4.20	1.0007				





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 5, 2025	Previous Calibration:	July 4, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:45	End Time (MST):	11:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999952	<i>≥0.995</i>
8.00	7.98	1.0031	Slope	0.998248	<i>0.90 - 1.10</i>
4.00	3.95	1.0132	Intercept	-0.025392	<i>+/-0.5</i>
2.01	1.96	1.0254			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 5, 2025	Previous Calibration:	July 4, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:45	End Time (MST):	11:36
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999978	<i>≥0.995</i>
8.75	8.78	0.9963	Slope	1.002405	<i>0.90 - 1.10</i>
4.37	4.40	0.9933	Intercept	0.017991	<i>+/-0.5</i>
2.19	2.24	0.9791			



NMHC Calibration Plot

Date: July 5, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Patricia McInnes
Station number: AMS 06
Calibration Date: July 7, 2025
Last Cal Date: June 4, 2025
Start time (MST): 9:45
End time (MST): 14:35
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0036234
NOX Cal Gas Conc: 62.2 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 62.20 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: July 22, 2032
NO Cal Gas Conc: 61.90 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 61.90 ppm
NO gas Diff:
Serial Number: 3566
Serial Number: 4602

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.6	0.1	0.5	----	----
AF High point	4935	64.6	803.7	799.8	3.9	807.3	799.7	7.5	0.9963	1.0003
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 803.0 ppb	NO = 800.8 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.5%
Baseline Corr 1st pt	NO _x = 806.7 ppb	NO = 799.6 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -0.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1172750022

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996272	0.997467
NO _x Cal Offset:	2.312443	2.791865
NO Cal Slope:	1.000640	0.999884
NO Cal Offset:	0.473372	0.713155
NO ₂ Cal Slope:	0.998421	0.996589
NO ₂ Cal Offset:	-0.376238	-0.344283

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.887	0.887	NO bkgnd or offset:	3.5	3.5
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	162.4	162.4

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	1.0	0.4	0.6	----	----
High point	4935	64.6	803.7	799.8	3.9	803.4	800.2	3.2	1.0004	0.9995
Mid point	4968	32.3	401.8	399.9	1.9	404.9	400.9	4.1	0.9923	0.9974
Low point	4984	16.2	201.5	200.5	1.0	205.3	201.4	4.0	0.9816	0.9958
As left zero	5000	0.0	0.0	0.0	0.0	0.8	0.3	0.4	----	----
As left span	4935	64.6	803.7	397.7	406.0	797.8	397.7	400.1	1.0074	1.0000
Average Correction Factor									0.9914	0.9976

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.6	----	----
High GPT point	797.9	387.7	414.1	412.9	1.0028	99.7%
Mid GPT point	797.9	599.6	202.2	200.3	1.0094	99.1%
Low GPT point	797.9	699.2	102.6	101.2	1.0136	98.7%
Average Correction Factor					1.0086	99.1%

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 4, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:45	End Time (MST):	14:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.0	----	Correlation Coefficient	0.999976	≥0.995
803.7	803.4	1.0004	Slope	0.997467	0.90 - 1.10
401.8	404.9	0.9923	Intercept	2.791865	+/-20
201.5	205.3	0.9816			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 4, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:45	End Time (MST):	14:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.6	----	Correlation Coefficient	0.999974	<i>≥0.995</i>
414.1	412.9	1.0028	Slope	0.996589	<i>0.90 - 1.10</i>
202.2	200.3	1.0094	Intercept	-0.344283	<i>+/-20</i>
102.6	101.2	1.0136			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 4, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:45	End Time (MST):	14:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999999	≥ 0.995
799.8	800.2	0.9995	Slope	0.999884	$0.90 - 1.10$
399.9	400.9	0.9974	Intercept	0.713155	± 20
200.5	201.4	0.9958			



NO_x Calibration Plot

Date: July 7, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: July 9, 2025
Start time (MST): 9:26
Reason: Routine

Station number: AMS 06
Last Cal Date: June 13, 2025
End time (MST): 12:20

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1300156234

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002686	1.003086	Backgd or Offset:	-0.5	-0.5
Calibration intercept:	-0.120000	0.560000	Coeff or Slope:	1.020	1.020

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	0.4	----
As found High point	5000	1031.0	400.0	401.2	0.998
As found Mid point					
As found Low point					
Baseline Corr As found:	400.8	Previous response	401.0	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.5	----
High point	5000	1031.0	400.0	401.8	0.996
Mid point	5000	821.4	200.0	201.1	0.995
Low point	5000	699.5	100.0	101.0	0.990
As left zero	5000	800.0	0.0	0.2	----
As left span	5000	1031.0	400.0	403.0	0.993
Average Correction Factor					0.993

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	June 13, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:26	End Time (MST):	12:20
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	1.000000	≥0.995
400.0	401.8	0.9955	Slope	1.003086	0.90 - 1.10
200.0	201.1	0.9945	Intercept	0.560000	+/- 5
100.0	101.0	0.9901			



O₃ Calibration Plot

Date: July 9, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Patricia McInnes
Calibration Date: July 17, 2025
Start time (MST): 14:36
Station number: AMS 06
Last Cal Date: June 3, 2025
End time (MST): 15:09
Analyzer Make: API T640
Particulate Fraction: PM2.5
S/N: 1547
Flow Meter Make/Model: Alicat FP-25BT
Temp/RH standard: Alicat FP-25BT
S/N: 388755

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	26.0	25.2	26.0	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	728.50	727.44	728.50	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.91	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	29	----	29	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 53.2		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 07-16-2026
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: June 3, 2025
Date Disposable Filter Changed: June 3, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: April 4, 2025
Date RH/T Sensor Cleaned: April 4, 2025

No adjustments made. PMT peak test was done last month. Readings are within limits.

Notes:

Calibration by: Max Farrell



Wood Buffalo Environmental Association

Nt - NOX - NH3 Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	July 2, 2025	Last Cal Date:	June 30, 2025
Start time (MST):	8:49	End time (MST):	13:00
NH3 Cal Date:	July 2, 2025	Last Cal Date:	June 30, 2025
Start time (MST):	13:00	End time (MST):	15:45
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	62.20	ppm	NO Gas Cylinder #:	DT0036234
NO Cal Gas Conc:	61.90	ppm	NO Cal Gas Expiry:	July 22, 2032
Removed NOX Conc:	62.20	ppm	Removed Cylinder #:	N/A
Removed NO Conc:	61.90	ppm	Removed cyl Expiry:	N/A
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	75.0	ppm	NH3 Gas Cylinder #:	CC709372
Removed NH3 Conc:	75.0	ppm	NH3 Cal Gas Expiry:	December 31, 2025
NH3 gas Diff:			Removed Cylinder #:	
Calibrator Model:	API T700		Removed cyl Expiry:	
ZAG make/model:	API T701		Serial Number:	3566
			Serial Number:	4602

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	215
Converter model:	API T501	Converter serial #:	147
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.20
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	25.4

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	1.040	1.040	Nt coefficient:	1.039	1.039
NOX coefficient:	1.032	1.032	NO bkgrnd:	0.2	0.2
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	-0.1	-0.1
NH3 coefficient:	0.956	0.956	Nt bkgrnd:	1.7	1.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001134	0.993485
NO _x Cal Offset:	0.400000	2.392545
NO Cal Slope:	0.996109	1.002645
NO Cal Offset:	0.100000	0.771491
NO ₂ Cal Slope:	0.989970	0.999754
NO ₂ Cal Offset:	0.300000	0.861728
NH3 Cal Slope:	0.990623	1.001531
NH3 Cal Offset:	9.861309	7.366969
Nt Cal Slope:	0.994848	1.005961
Nt Cal Offset:	10.327730	7.739116



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

NO_x / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO _x Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>	Baseline corr NO Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.4	-0.1	----	----
As found span	4935	64.6	803.7	799.8	803.7	804.3	799.2	809.3	0.9992	1.0008
AF GPT span										
new NO cyl rp										

Baseline Corr As Fd Nt = 809.4 ppb NO_x = 804.4 ppb NO = 799.6 ppb

Previous Response Nt = 809.88 ppb NO_x = 805.0 ppb NO = 796.8 ppb

****NO_x Δ (NO to GPT response) =**

*** = > +/-2% difference initiates investigation**

***Percent Change Nt(NO) = -0.1%**

***Percent Change NO_x = -0.1%**

***Percent Change NO = 0.4%**

*** = > +/-5% change initiates investigation**

NO_x / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.4	----	----
High point	4935	64.6	803.7	799.8	803.7	799.6	802.9	807.2	1.0051	0.9962
Mid point	4968	32.3	401.8	399.9	401.8	403.1	400.4	407.6	0.9967	0.9986
Low point	4984	16.2	201.5	200.5	201.5	204.5	203.6	204.8	0.9854	0.9850
Average Correction Factor									0.9958	0.9933

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found zero	----	----	0.0	0.3	----	----
Calibration zero	----	----	0.0	0.1	----	----
High GPT point (400 ppb O3)	794.5	385.6	412.8	412.5	1.0007	99.9%
Mid GPT point (200 ppb O3)	794.5	595.7	202.7	205.9	0.9843	101.6%
Low GPT point (100 ppb O3)	794.5	699.0	99.4	99.6	0.9978	100.2%
Average Correction Factor					0.9943	100.6%



Wood Buffalo Environmental Association NH₃ - N_T Calibration Report

NH₃ As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH ₃ concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH ₃ concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH ₃ Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
AF High point	3416	84.0	1799.0	0.0	1799.0	1812.9	8.3	1804.7	0.992	0.997
AF Mid point										
AF Low point										
new NH ₃ cyl rp										
Baseline Corr As Fd	Nt = 1813.0 ppb	NH ₃ = 1804.7 ppb							*Percent Change	Nt _(NH₃) = 0.7%
Previous Response	Nt = 1800.1 ppb	NH ₃ = 1792.0 ppb				* = > +/-5% change initiates investigation			*Percent Change	NH ₃ = 0.7%

NH₃ Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH ₃ concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH ₃ concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH ₃ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.4	0.2	0.2	----	----
High point	3416	84.0	1799.0	0.0	1799.0	1812.9	8.3	1804.7	0.992	0.997
Mid point	3453	46.7	1000.3	0.0	1000.3	1018.6	4.9	1013.6	0.982	0.987
Low point	3477	23.3	499.0	0.0	499.0	517.0	2.7	514.3	0.965	0.970
Average Correction Factor									0.9798	0.9846
NH ₃ Previous Converter Efficiency =	95.6 %									
NH ₃ Current Converter Efficiency =	95.6 %									

Notes: Changed the inlet filter after as founds. No adjustments made. Had the wrong flow during NH₃ span which caused the jump on readings.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


Nt Calibration Summary

Station Information

Calibration Date:	July 2, 2025	Previous Calibration:	June 30, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:49	End Time (MST):	13:00
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999916	≥ 0.995
1799.0	1812.9	0.9924	Slope	1.005961	0.90 - 1.10
1000.3	1018.6	0.9820	Intercept	7.739116	+/-20
499.0	517.0	0.9651			





Wood Buffalo Environmental Association


NH₃ Calibration Summary

Station Information

Calibration Date:	July 2, 2025	Previous Calibration:	June 30, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:49	End Time (MST):	13:00
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999919	<i>≥0.995</i>
1799.0	1804.7	0.9969	Slope	1.001531	<i>0.90 - 1.10</i>
1000.3	1013.6	0.9868	Intercept	7.366969	<i>+/-20</i>
499.0	514.3	0.9702			





Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 2, 2025	Previous Calibration:	June 30, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:49	End Time (MST):	13:00
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.2	----	Correlation Coefficient	0.999965	<i>≥0.995</i>
803.7	799.6	1.0051	Slope	0.993485	<i>0.90 - 1.10</i>
401.8	403.1	0.9967	Intercept	2.392545	<i>+/-20</i>
201.5	204.5	0.9854			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 2, 2025	Previous Calibration:	June 30, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:49	End Time (MST):	13:00
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999985	≥ 0.995
799.8	802.9	0.9962	Slope	1.002645	0.90 - 1.10
399.9	400.4	0.9986	Intercept	0.771491	+/-20
200.5	203.6	0.9850			





Wood Buffalo Environmental Association


NO₂ Calibration Summary


Station Information

Calibration Date:	July 2, 2025	Previous Calibration:	June 30, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:49	End Time (MST):	13:00
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999916	<i>≥0.995</i>
412.8	412.5	1.0007	Slope	0.999754	<i>0.90 - 1.10</i>
202.7	205.9	0.9843	Intercept	0.861728	<i>+/-20</i>
99.4	99.6	0.9978			






NH₃ Calibration Plot

Date: July 2, 2025

Location: Patricia McInnes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Athabasca Valley Station number: AMS07
Calibration Date: July 18, 2025 Last Cal Date: June 6, 2025
Start time (MST): 11:00 End time (MST): 14:34
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.06 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC320556
Removed Cal Gas Conc: 50.06 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3805
Zero Air Gen Model: API 701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i-LTE Serial Number: 1507864683
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999447	0.999719	Backgd or Offset:	2.73	2.73
Calibration intercept:	1.964232	2.444307	Coeff or Slope:	0.845	0.845

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	79.8	799.0	798.8	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	798.8	Previous response	800.5	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4920	79.8	799.0	799.2	1.000
Mid point	4960	39.9	399.5	405.5	0.985
Low point	4980	20.0	200.2	203.3	0.985
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	79.8	799.0	800.6	0.998
Average Correction Factor:					0.990

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	July 18, 2025	Previous Calibration:	June 6, 2025
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	11:00	End Time (MST):	14:34
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999933	≥0.995
799.0	799.2	0.9997	Slope	0.999719	0.90 - 1.10
399.5	405.5	0.9852	Intercept	2.444307	+/-30
200.2	203.3	0.9849			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: July 18, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Athabasca Valley
Calibration Date: July 8, 2025
Start time (MST): 9:40
Reason: Routine

Station number: AMS07
Last Cal Date: June 3, 2025
End time (MST): 14:42

Calibration Standards

Cal Gas Concentration: 5.25 ppm
Cal Gas Cylinder #: CC504080
Removed Cal Gas Conc: 5.25 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API T701H

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3805
Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1180540018
Converter serial #: 551
Converter Temp: 840 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.012897	1.003528	Backgd or Offset:	2.7	2.7
Calibration intercept:	-0.242219	-0.242178	Coeff or Slope:	0.908	0.940

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4925	75.5	79.3	76.4	1.033
As found Mid point	4962	37.7	39.6	38.1	1.031
As found Low point	4981	18.9	19.8	18.7	1.044
New cylinder response					
Baseline Corr As found:	76.7	Prev response:	80.05	*% change:	-4.4%
Baseline Corr 2nd AF pt:	38.4	AF Slope:	0.968625	AF Intercept:	-0.362049
Baseline Corr 3rd AF pt:	19.0	AF Correlation:	0.999986	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4925	75.5	79.3	79.4	0.999
Mid point	4962	37.7	39.6	39.5	1.003
Low point	4981	18.9	19.9	19.5	1.018
As left zero	5000	0.0	0.0	0.0	----
As left span	4925	75.5	79.3	79.2	1.001
SO2 Scrubber Check	4920	79.2	792.1	-0.1	----
Date of last scrubber change:		21-Feb-25		Ave Corr Factor	1.007
Date of last converter efficiency test:		Friday, April 22, 2022			

Notes:

Span adjusted

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	July 8, 2025	Previous Calibration:	June 3, 2025
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:40	End Time (MST):	14:42
Analyzer make:	Thermo 43i LTE	Analyzer serial #:	1180540018

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999984		≥ 0.995
79.3	79.4	0.9989	Slope	1.003528		$0.90 - 1.10$
39.6	39.5	1.0028	Intercept	-0.242178		± 3
19.9	19.5	1.0183				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Athabasca Valley
 Calibration Date: July 18, 2025
 Start time (MST): 11:00
 Reason: Routine

Station number: AMS 07
 Last Cal Date: June 6, 2025
 End time (MST): 14:34

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	Monday, March 10, 2031
CH ₄ Cal Gas Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3805
Zero Air Gen model:	Teledyne API T701H	Serial Number:	198

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1331259520
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.99E-04	2.99E-04	NMHC SP Ratio:	5.66E-05
CH ₄ Retention time:	14.4	14.4	NMHC Peak Area:	158944
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.91	16.94	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.94	Prev response	16.93	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.91	16.88	1.002
Mid point	4960	39.9	8.46	8.50	0.995
Low point	4980	20.0	4.24	4.32	0.982
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.91	16.97	0.997
Average Correction Factor					0.993

Notes:

No adjustments needed.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	9.00	8.96	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.96	Prev response	9.01	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.8	9.00	8.92	1.008
Mid point	4960	39.9	4.50	4.51	0.999
Low point	4980	20.0	2.26	2.29	0.983
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.00	8.98	1.002
Average Correction Factor					0.997

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	7.92	7.97	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.97	Prev response	7.92	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.8	7.92	7.95	0.995
Mid point	4960	39.9	3.96	4.00	0.991
Low point	4980	20.0	1.98	2.02	0.982
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	7.92	8.00	0.990
Average Correction Factor					0.989

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.999531	0.996492
THC Cal Offset:	0.021654	0.046655
CH ₄ Cal Slope:	0.999850	1.003851
CH ₄ Cal Offset:	-0.000138	0.014249
NMHC Cal Slope:	0.999150	0.990157
NMHC Cal Offset:	0.021192	0.031606

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 18, 2025	Previous Calibration:	June 6, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	11:00	End Time (MST):	14:34
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999965	≥ 0.995
16.91	16.88	1.0022	Slope	0.996492	$0.90 - 1.10$
8.46	8.50	0.9948	Intercept	0.046655	± 0.5
4.24	4.32	0.9824			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 18, 2025	Previous Calibration:	June 6, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	11:00	End Time (MST):	14:34
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999985	≥ 0.995
7.92	7.95	0.9953	Slope	1.003851	$0.90 - 1.10$
3.96	4.00	0.9908	Intercept	0.014249	± 0.5
1.98	2.02	0.9822			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 18, 2025	Previous Calibration:	June 6, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	11:00	End Time (MST):	14:34
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999942	<i>≥0.995</i>
9.00	8.92	1.0083	Slope	0.990157	<i>0.90 - 1.10</i>
4.50	4.51	0.9986	Intercept	0.031606	<i>+/-0.5</i>
2.26	2.29	0.9830			



NMHC Calibration Plot

Date: July 18, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Athabasca Valley
Station number: AMS 07
Calibration Date: July 10, 2025
Last Cal Date: June 5, 2025
Start time (MST): 9:42
End time (MST): 15:20
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0033919
NOX Cal Gas Conc: 60.10 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 60.10 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.90 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 59.90 ppm
NO gas Diff:
Serial Number: 3805
Serial Number: 198

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
AF High point	4933	66.8	803.0	800.3	2.7	832.6	828.5	4.2	0.9645	0.9661
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 804.0 ppb	NO = 802.1 ppb							*Percent Change	NO _x = 3.4%
Baseline Corr 1st pt	NO _x = 832.5 ppb	NO = 828.4 ppb							*Percent Change	NO = 3.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb								
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb								
As Found Statistics										
	As found	NO _x r ² :							Nx SI:	Nx Int:
	As found	NO r ² :							NO SI:	NO Int:
	As found	NO ₂ r ² :							NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1160120024

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.945	0.912	NO bkgnd or offset:	6.6	6.4
NOX coeff or slope:	1.004	1.003	NOX bkgnd or offset:	6.8	6.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	182.6	183.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998550	0.996372
NO _x Cal Offset:	2.151910	2.391913
NO Cal Slope:	1.000199	0.998057
NO Cal Offset:	1.691929	1.991935
NO ₂ Cal Slope:	1.003031	1.003627
NO ₂ Cal Offset:	1.715926	0.404617

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.0	----	----
High point	4933	66.8	803.0	800.3	2.7	800.9	799.4	1.6	1.0026	1.0011
Mid point	4966	33.4	401.5	400.2	1.3	405.0	403.7	1.3	0.9914	0.9913
Low point	4983	16.7	200.7	200.1	0.7	203.5	202.4	1.1	0.9865	0.9885
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.0	----	----
As left span	4933	66.8	803.0	399.1	403.9	803.9	399.1	404.7	0.9988	1.0000
Average Correction Factor									0.9935	0.9936

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	797.8	398.4	402.1	403.7	0.9960	100.4%
Mid GPT point	797.8	597.8	202.7	204.1	0.9930	100.7%
Low GPT point	797.8	697.9	102.6	103.7	0.9891	101.1%
Average Correction Factor					0.9927	100.7%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 5, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:42	End Time (MST):	15:20
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999959	≥0.995
803.0	800.9	1.0026	Slope	0.996372	0.90 - 1.10
401.5	405.0	0.9914	Intercept	2.391913	+/-20
200.7	203.5	0.9865			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 5, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:42	End Time (MST):	15:20
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999995	≥0.995
402.1	403.7	0.9960	Slope	1.003627	0.90 - 1.10
202.7	204.1	0.9930	Intercept	0.404617	+/-20
102.6	103.7	0.9891			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 5, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:42	End Time (MST):	15:20
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999970	≥ 0.995
800.3	799.4	1.0011	Slope	0.998057	$0.90 - 1.10$
400.2	403.7	0.9913	Intercept	1.991935	± 20
200.1	202.4	0.9885			



NO_x Calibration Plot

Date: July 10, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Athabasca Valley
Calibration Date: July 3, 2025
Start time (MST): 10:23
Reason: Routine

Station number: AMS07
Last Cal Date: June 12, 2025
End time (MST): 14:01

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: T700
ZAG Make/Model: T701H

Serial Number: 3805
Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1152220023

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998086	1.002886	Backgd or Offset:	-1.1	-1.1
Calibration intercept:	0.560000	0.920000	Coeff or Slope:	1.556	1.556

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	NA	0.0	-0.2	----
As found High point	5000	1725.3	400.0	401.8	0.995
As found Mid point					
As found Low point					
Baseline Corr As found:	402.0	Previous response	399.8	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	NA	0.0	0.5	----
High point	5000	1705.1	400.0	401.9	0.995
Mid point	5000	1172.8	200.0	201.6	0.992
Low point	5000	921.2	100.0	101.7	0.983
As left zero	5000	NA	0.0	-0.7	----
As left span	5000	1582.6	400.0	401.6	0.996
Average Correction Factor					0.990

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	July 3, 2025	Previous Calibration:	June 12, 2025
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:23	End Time (MST):	14:01
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220023

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999995	≥0.995
400.0	401.9	0.9953	Slope	1.002886	0.90 - 1.10
200.0	201.6	0.9921	Intercept	0.920000	+/- 5
100.0	101.7	0.9833			



O₃ Calibration Plot

Date: July 3, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Athabasca Valley Station number: AMS 07
Calibration Date: July 31, 2025 Last Cal Date: June 20, 2025
Start time (MST): 12:09 End time (MST): 12:47

Analyzer Make: API T640 S/N: 2235
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	33.9	32.7	33.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	740.3	738.6	740.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	4.95	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	31	----	31	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 9.2		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 6, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: April 1, 2025
Date Disposable Filter Changed: July 31, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 8, 2024
Date RH/T Sensor Cleaned: July 8, 2024

Notes: Temp, pressure and flow checked. Leak check passed. DFU filter changed out.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name: Athabasca Valley Station number: AMS 07
Calibration Date: July 15, 2025 Last Cal Date: June 20, 2025
Start time (MST): 10:55 End time (MST): 14:22
Reason: Routine

Calibration Standards

Cal Gas Concentration: 2,953 ppm Cal Gas Exp Date: September 30, 2029
Cal Gas Cylinder #: T1TWKRN
Removed Cal Gas Conc: 2,953 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T750 Serial Number: 282
ZAG Make/Model: Teledyne API 751H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 48i-TLE Analyzer serial #: 1408761381
Analyzer Range: 0 - 50 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002547	0.999123	Backgd or Offset:	5.620	5.731
Calibration intercept:	0.176056	0.126011	Coeff or Slope:	1.079	1.081

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4932	67.8	40.0	40.2	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	40.11	Prev response:	40.32	*% change:	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4932	67.8	40.0	40.0	1.000
Mid point	4966	33.9	20.0	20.3	0.987
Low point	4983	16.9	10.0	10.2	0.980
As left zero	5000	0.0	0.0	0.0	----
As left span	4932	67.8	40.0	40.0	1.001
Average Correction Factor					0.989

Notes: Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


CO Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 20, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:55	End Time (MST):	14:22
Analyzer make:	Thermo 48i-TLE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999935	≥0.995
40.0	40.0	1.0001	Slope	0.999123	0.90 - 1.10
20.0	20.3	0.9873	Intercept	0.126011	+/-1.5
10.0	10.2	0.9805			



CO Calibration Plot

Date: July 15, 2025

Location: Athabasca Valley





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
Calibration Date: July 9, 2025 Last Cal Date: June 26 2025
Start time (MST): 9:11 End time (MST): 11:51
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.84 ppm Cal Gas Exp Date: January 6, 2030
Cal Gas Cylinder #: CC196697
Removed Cal Gas Conc: 49.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3810
Zero Air Gen Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43i-TLE Serial Number: 1236656116
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000545	1.003814	Backgd or Offset:	1.9	1.9
Calibration intercept:	-0.023960	-0.244419	Coeff or Slope:	1.029	1.029

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	80.3	800.4	802.7	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	802.7	Previous response	800.8	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4920	80.3	800.4	803.7	0.996
Mid point	4960	40.2	400.7	401.0	0.999
Low point	4980	20.1	200.4	200.8	0.998
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	80.3	800.4	801.2	0.999
Average Correction Factor:					0.998

Notes: Changed out inlet filter after as founds. No adjustments made.

Calibration Performed By: Sabian V Jeremy C



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	June 26 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	9:11	End Time (MST):	11:51
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999996	≥0.995
800.4	803.7	0.9959	Slope	1.003814	0.90 - 1.10
400.7	401.0	0.9992	Intercept	-0.244419	+/-30
200.4	200.8	0.9978			



SO2 Calibration Plot

Date: July 9, 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
Calibration Date: July 9, 2025 Last Cal Date: June 18, 2025
Start time (MST): 11:58 End time (MST): 15:34
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: August 28, 2027
Cal Gas Cylinder #: SA7549
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3810
ZAG Make/Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153461
Converter make: CDN-101 Converter serial #: 630
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.014823	1.034837	Backgd or Offset:	2.8	2.8
Calibration intercept:	-0.338230	-0.438593	Coeff or Slope:	1.144	1.144

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4917	82.6	80.0	81.7	0.978
As found Mid point	4959	41.3	40.0	40.5	0.985
As found Low point	4979	20.7	20.0	19.9	1.002
New cylinder response					
Baseline Corr As found:	81.8	Prev response:	80.81	*% change:	1.2%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.024687	AF Intercept:	-0.358402
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999954	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4917	82.6	80.0	82.6	0.968
Mid point	4959	41.3	40.0	40.5	0.987
Low point	4979	20.7	20.0	20.0	1.002
As left zero	5000	0.0	0.0	0.0	----
As left span	4917	82.6	80.0	82.2	0.973
SO2 Scrubber Check	4920	80.3	803.0	-0.1	----
Date of last scrubber change:	May 15, 2025			Ave Corr Factor	0.986
Date of last converter efficiency test:					

Notes: No adjustments performed. SO2 scrubber check was done passed.

Calibration Performed By: Sabian Voyageur Jeremy Cardinal



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	July 9, 2025	Previous Calibration:	June 18, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	11:58	End Time (MST):	15:34
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999853		≥ 0.995
80.0	82.6	0.9681	Slope	1.034837		$0.90 - 1.10$
40.0	40.5	0.9871	Intercept	-0.438593		± 3
20.0	20.0	1.0019				

TRS Calibration Curve







Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan
Station number: AMS 08
Calibration Date: July 10, 2025
Last Cal Date: June 13, 2025
Start time (MST): 8:46
End time (MST): 13:25
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0046831
NOX Cal Gas Conc: 60.20 ppm
Removed Cylinder #: DT0046831
Removed Gas NOX Conc: 60.20 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: January 9, 2032
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 135

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.0	-0.3	-0.7	----	----
AF High point	4933	66.7	803.1	800.4	2.7	832.6	829.5	3.1	0.9634	0.9646
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 802.9 ppb	NO = 800.2 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = 3.7%			
Baseline Corr 1st pt	NO _x = 833.6 ppb	NO = 829.8 ppb	<u>As Found Statistics</u>			*Percent Change	NO = 3.6%			
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:	NO Int:			
			As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12124313137

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.875	0.823	NO bkgnd or offset:	0.9	0.8
NOX coeff or slope:	0.996	0.995	NOX bkgnd or offset:	1.0	1.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	1.0	149.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999950	1.000349
NO _x Cal Offset:	-0.205598	-0.145464
NO Cal Slope:	1.000699	0.998300
NO Cal Offset:	-0.825756	-0.785790
NO ₂ Cal Slope:	1.006044	1.007238
NO ₂ Cal Offset:	0.044271	1.315166

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.1	-0.8	----	----
High point	4933	66.7	803.1	800.4	2.7	802.9	798.7	4.3	1.0003	1.0022
Mid point	4967	33.3	400.9	399.6	1.3	401.3	397.6	3.6	0.9990	1.0050
Low point	4983	16.7	201.1	200.4	0.7	201.6	198.7	2.9	0.9974	1.0086
As left zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.1	-0.7	----	----
As left span	4933	66.7	803.1	384.1	419.0	799.8	384.1	415.9	1.0041	1.0000
Average Correction Factor									0.9989	1.0052

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.8	----	----
High GPT point	796.1	385.1	413.7	416.8	0.9925	100.8%
Mid GPT point	796.1	592.9	205.9	210.1	0.9799	102.1%
Low GPT point	796.1	695.1	103.7	107.6	0.9635	103.8%
Average Correction Factor					0.9786	102.2%

Notes: Adjustments made to span. Changed filter after as found

Calibration Performed By: Sabian Voyageur Jeremy Cardinal



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 13, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	8:46	End Time (MST):	13:25
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12124313137

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.8	----	Correlation Coefficient	0.999997	≥0.995
803.1	802.9	1.0003	Slope	1.000349	0.90 - 1.10
400.9	401.3	0.9990	Intercept	-0.145464	+/-20
201.1	201.6	0.9974			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 13, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	8:46	End Time (MST):	13:25
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12124313137

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.8	----	Correlation Coefficient	0.999880	≥0.995
413.7	416.8	0.9925	Slope	1.007238	0.90 - 1.10
205.9	210.1	0.9799	Intercept	1.315166	+/-20
103.7	107.6	0.9635			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 13, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	8:46	End Time (MST):	13:25
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12124313137

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999997	≥ 0.995
800.4	798.7	1.0022	Slope	0.998300	$0.90 - 1.10$
399.6	397.6	1.0050	Intercept	-0.785790	± 20
200.4	198.7	1.0086			



NO_x Calibration Plot

Date: July 10, 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Fort Chipewyan
Calibration Date: July 7, 2025
Start time (MST): 13:14
Reason: Routine

Station number: AMS 08
Last Cal Date: June 11, 2025
End time (MST): 16:27

Calibration Standards

O3 generation mode: Photometer
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Serial Number: 3810
Serial Number: 135

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1152220026

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996400	1.015086	Backgd or Offset:	-0.3	-0.3
Calibration intercept:	0.420000	0.060000	Coeff or Slope:	1.022	0.983

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	5000	968.7	400.0	388.1	1.031
As found Mid point					
As found Low point					
Baseline Corr As found:	388.0	Previous response	399.0	*% change	-2.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	5000	968.7	400.0	406.3	0.984
Mid point	5000	820.5	200.0	202.8	0.986
Low point	5000	720.0	100.0	101.3	0.987
As left zero	5000	0.0	0.0	-1.3	----
As left span	5000	968.7	400.0	404.4	0.989
Average Correction Factor					0.986

Notes: Changed Filter after as founds. Adjustments made to span and O3 REF on calibrator.

Calibration Performed By: Sabian Voyageur



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 11, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	13:14	End Time (MST):	16:27
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220026

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999997	≥0.995
400.0	406.3	0.9845	Slope	1.015086	0.90 - 1.10
200.0	202.8	0.9862	Intercept	0.060000	+/- 5
100.0	101.3	0.9872			



O₃ Calibration Plot

Date: July 7, 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
Calibration Date: July 29, 2025 Last Cal Date: June 11, 2025
Start time (MST): 10:24 End time (MST): 11:26

Analyzer Make: Teledyne API T640 S/N: 319
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388744
Temp/RH standard: Alicat FP-25BT S/N: 388744

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	18.4	19.10	18.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	744.6	746.00	744.6	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	5.00	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	42%		42%	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 5.70		PM w/ HEPA: 0.20		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☐ Alignment Factor On : ☐

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: 10-Jun-24
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: June 11, 2025
Date Disposable Filter Changed: June 11, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 29, 2024
Date RH/T Sensor Cleaned: August 29, 2024

Notes:

No Adjustment made

Calibration by: Morgan Voyageur



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Barge Landing
Calibration Date: July 9, 2025
Start time (MST): 9:22
Reason: Routine

Station number: AMS 09
Last Cal Date: June 9, 2025
End time (MST): 12:27

Calibration Standards

Cal Gas Concentration: 50.56 ppm
Cal Gas Cylinder #: CC705748
Removed Cal Gas Conc: 50.56 ppm
Removed Gas Cyl #: NA
Calibrator Model: API T700
Zero Air Gen Model: APIT701

Cal Gas Exp Date: October 9, 2032
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3812
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 1118148498

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008116	1.006247	Backgd or Offset:	11.3	11.3
Calibration intercept:	-0.277874	-0.398966	Coeff or Slope:	0.988	0.984

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4921	79.1	799.8	806.1	0.992
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	806.2	Previous response	806.1	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4921	79.1	799.8	805.0	0.994
Mid point	4961	39.5	399.4	400.4	0.997
Low point	4980	19.8	200.2	201.0	0.996
As left zero	5000	0.0	0.0	-0.1	----
As left span	4921	79.1	799.8	805.1	0.993
Average Correction Factor:					0.996

Notes: Inlet filter changed after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	June 9, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:22	End Time (MST):	12:27
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999995	≥0.995
799.8	805.0	0.9936	Slope	1.006247	0.90 - 1.10
399.4	400.4	0.9975	Intercept	-0.398966	+/-30
200.2	201.0	0.9961			



SO2 Calibration Plot

Date: July 9, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Barge Landing
Calibration Date: July 8, 2025
Start time (MST): 8:33
Reason: Routine

Station number: AMS 09
Last Cal Date: June 17, 2025
End time (MST): 12:55

Calibration Standards

Cal Gas Concentration: 5.17 ppm
Cal Gas Cylinder #: CC511415
Removed Cal Gas Conc: 5.17 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: August 22, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3812
Serial Number: 5613

Analyzer Information

Analyzer make: Thermo 43iQ-TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1203169744
Converter serial #: 519
Converter Temp: 830 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003547	1.018393	Backgd or Offset:	1.820	1.880
Calibration intercept:	-0.180549	-0.300256	Coeff or Slope:	0.701	0.724

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4923	77.4	80.0	73.4	1.093
As found Mid point	4961	38.7	40.0	38.5	1.045
As found Low point	4981	19.3	20.0	19.6	1.029
New cylinder response					
Baseline Corr As found:	73.2	Prev response:	80.14	*% change:	-9.5%
Baseline Corr 2nd AF pt:	38.3	AF Slope:	0.912038	AF Intercept:	0.997844
Baseline Corr 3rd AF pt:	19.4	AF Correlation:	0.999264	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	4923	77.4	80.0	81.2	0.986
Mid point	4961	38.7	40.0	40.5	0.988
Low point	4981	19.3	20.0	20.0	0.998
As left zero	5000	0.0	0.0	-0.1	----
As left span	4923	77.4	80.0	81.6	0.981
SO2 Scrubber Check	4920	80.2	802.0	0.0	----
Date of last scrubber change:	30-May-25		Ave Corr Factor		0.991
Date of last converter efficiency test:					

Notes: Scrubber beads were changed to remedy low span response. Scrubber check passed and done after calibrator zero. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	July 8, 2025	Previous Calibration:	June 17, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	8:33	End Time (MST):	12:55
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1203169744

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999999		≥ 0.995
80.0	81.2	0.9857	Slope	1.018393		$0.90 - 1.10$
40.0	40.5	0.9883	Intercept	-0.300256		± 3
20.0	20.0	0.9979				


TRS Calibration Curve



TRS Calibration Plot

Date: July 8, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Barge Landing
 Calibration Date: July 9, 2025
 Start time (MST): 9:22
 Reason: Routine

Station number: AMS 09
 Last Cal Date: June 9, 2025
 End time (MST): 12:27

Calibration Standards

Gas Cert Reference: CC705748
 CH₄ Cal Gas Conc. 505.6 ppm
 C₃H₈ Cal Gas Conc. 204.8 ppm
 Removed Gas Cert: CC151285
 Removed CH₄ Conc. 505.6 ppm
 Removed C₃H₈ Conc. 204.8 ppm
 Diff between cyl (CH₄):
 Calibrator Model: API T700
 Zero Air Gen model: APIT701

Cal Gas Expiry Date: October 9, 2032
 CH₄ Equiv Conc. 1068.8 ppm
 Removed Gas Expiry: January 5, 2025
 CH₄ Equiv Conc. 1068.8 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 3812
 Serial Number: 5613

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1193585650
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.25E-04	2.22E-04	NMHC SP Ratio:	4.89E+05
CH ₄ Retention time:	14.4	14.4	NMHC Peak Area:	182254
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				4.87E+05
				182802
				OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	16.91	16.97	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.97	Prev response	16.94	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	16.91	16.90	1.000
Mid point	4961	39.5	8.44	8.42	1.003
Low point	4980	19.8	4.23	4.18	1.013
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	16.91	16.99	0.995
Average Correction Factor					1.006

Notes:

Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.1	8.91	8.89	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.89	Prev response	8.93	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.1	8.91	8.91	1.000
Mid point	4961	39.5	4.45	4.45	1.000
Low point	4980	19.8	2.23	2.21	1.008
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	8.91	8.98	0.993
Average Correction Factor					1.003

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.1	8.00	8.08	0.990
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.08	Prev response	8.01	*% change	0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.1	8.00	7.99	1.001
Mid point	4961	39.5	3.99	3.97	1.007
Low point	4980	19.8	2.00	1.97	1.018
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	8.00	8.02	0.998
Average Correction Factor					1.009

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.003382	1.000408
THC Cal Offset:	-0.026353	-0.025362
CH ₄ Cal Slope:	1.003172	1.000287
CH ₄ Cal Offset:	-0.015980	-0.018385
NMHC Cal Slope:	1.003544	1.000645
NMHC Cal Offset:	-0.009773	-0.006976

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	June 9, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:22	End Time (MST):	12:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999989	<i>≥0.995</i>
16.91	16.90	1.0005	Slope	1.000408	<i>0.90 - 1.10</i>
8.44	8.42	1.0033	Intercept	-0.025362	<i>+/-0.5</i>
4.23	4.18	1.0128			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	June 9, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:22	End Time (MST):	12:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999976	≥ 0.995
8.00	7.99	1.0008	Slope	1.000287	$0.90 - 1.10$
3.99	3.97	1.0070	Intercept	-0.018385	± 0.5
2.00	1.97	1.0179			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	June 9, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:22	End Time (MST):	12:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999994	<i>≥0.995</i>
8.91	8.91	1.0001	Slope	1.000645	<i>0.90 - 1.10</i>
4.45	4.45	0.9997	Intercept	-0.006976	<i>+/-0.5</i>
2.23	2.21	1.0083			



NMHC Calibration Plot

Date: July 9, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Barge Landing
Station number: AMS 09
Calibration Date: July 15, 2025
Last Cal Date: June 13, 2025
Start time (MST): 9:15
End time (MST): 13:22
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1KDH
NO_x Cal Gas Conc: 47.38 ppm
Removed Cylinder #: NA
Removed Gas NO_x Conc: 47.38 ppm
NO_x gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 46.94 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 46.94 ppm
NO gas Diff:
Serial Number: 3812
Serial Number: 4888

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
AF High point	4915	85.3	808.3	800.7	7.5	810.9	799.8	11.1	0.9965	1.0012
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 808.1 ppb	NO = 799.0 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.4%
Baseline Corr 1st pt	NO _x = 811.1 ppb	NO = 799.8 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1426262593

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999134	0.996914
NO _x Cal Offset:	0.558241	0.517899
NO Cal Slope:	0.998797	0.996456
NO Cal Offset:	-0.743986	-0.624358
NO ₂ Cal Slope:	1.003458	1.003372
NO ₂ Cal Offset:	-0.007940	-0.246299

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.103	1.103	NO bkgnd or offset:	10.1	10.1
NOX coeff or slope:	0.999	0.999	NOX bkgnd or offset:	10.4	10.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	178.3	178.3

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
High point	4915	85.3	808.3	800.7	7.5	806.0	797.6	8.3	1.0028	1.0039
Mid point	4957	42.6	403.7	400.0	3.7	403.3	397.7	5.5	1.0010	1.0057
Low point	4979	21.3	201.8	200.0	1.9	202.2	197.8	4.4	0.9982	1.0109
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.2	----	----
As left span	4915	85.3	808.3	404.7	403.6	803.9	404.7	399.3	1.0054	1.0000
Average Correction Factor									1.0007	1.0068

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	793.8	403.2	398.1	399.2	0.9973	100.3%
Mid GPT point	793.8	602.0	199.3	199.9	0.9970	100.3%
Low GPT point	793.8	696.7	104.6	104.4	1.0020	99.8%
Average Correction Factor					0.9988	100.1%

Notes:

Inlet filter changed after as founds. No adjustment.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 13, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:15	End Time (MST):	13:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999998	≥0.995
808.3	806.0	1.0028	Slope	0.996914	0.90 - 1.10
403.7	403.3	1.0010	Intercept	0.517899	+/-20
201.8	202.2	0.9982			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 13, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:15	End Time (MST):	13:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999998	≥0.995
398.1	399.2	0.9973	Slope	1.003372	0.90 - 1.10
199.3	199.9	0.9970	Intercept	-0.246299	+/-20
104.6	104.4	1.0020			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 13, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:15	End Time (MST):	13:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999996	≥ 0.995
800.7	797.6	1.0039	Slope	0.996456	$0.90 - 1.10$
400.0	397.7	1.0057	Intercept	-0.624358	± 20
200.0	197.8	1.0109			



NO_x Calibration Plot

Date: July 15, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Barge Landing Station number: AMS 09
Calibration Date: July 22, 2025 Last Cal Date: June 20, 2025
Start time (MST): 10:06 End time (MST): 11:21

Analyzer Make: API T640 S/N: 2237
Particulate Fraction: PM2.5

Flow Meter Make/Model: DeltaCal S/N: 1451
Temp/RH standard: DeltaCal S/N: 1451

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	18.60	17.70	18.60	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	736.70	734.50	736.70	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.03	4.96	5.03	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	29	----	29	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 3.00		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: January 30, 2027
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.8	11.9	10.9	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: July 22, 2025
Date Disposable Filter Changed: July 22, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 22, 2025
Date RH/T Sensor Cleaned: July 22, 2025

Notes: Verified flow, pressure, temperature and pump power. No issues. Leak check passed.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS11 LOWER CAMP JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Lower Camp Station number: AMS 11
Calibration Date: July 22, 2025 Last Cal Date: June 11, 2025
Start time (MST): 10:47 End time (MST): 15:05
Reason: Routine

Calibration Standards

Cal Gas Concentration: 48.75 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC741503
Removed Cal Gas Conc: 48.75 ppm Rem Gas Exp Date:
Removed Gas Cyl #: CC741503 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3811
Zero Air Gen Model: Teledyne API T701 Serial Number: 196

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 100841398
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003791	1.009463	Backgd or Offset:	17.1	17.3
Calibration intercept:	1.575920	0.715764	Coeff or Slope:	1.068	1.068

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	1.4	----
As found High point	4932	82.2	799.2	807.9	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	806.5	Previous response	803.8	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	1.5	----
High point	4932	82.2	799.2	808.3	0.989
Mid point	4971	41.2	400.7	403.5	0.993
Low point	4996	20.6	200.2	202.9	0.987
As left zero	5000	0.0	0.0	1.8	----
As left span	4932	82.2	799.2	808.8	0.988
Average Correction Factor:					0.989

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 22, 2025	Previous Calibration:	June 11, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:47	End Time (MST):	15:05
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.5	----	Correlation Coefficient	0.999988	≥0.995
799.2	808.3	0.9887	Slope	1.009463	0.90 - 1.10
400.7	403.5	0.9931	Intercept	0.715764	+/-30
200.2	202.9	0.9866			



SO2 Calibration Plot

Date: July 22, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Lower Camp
Calibration Date: July 25, 2025
Start time (MST): 10:53
Reason: Routine

Station number: AMS 11
Last Cal Date: June 10, 2025
End time (MST): 15:33

Calibration Standards

Cal Gas Concentration: 4.83 ppm
Cal Gas Cylinder #: CC737863
Removed Cal Gas Conc: 4.83 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API T701H

Cal Gas Exp Date: August 28, 2028
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3811
Serial Number: 196

Analyzer Information

Analyzer make: Thermo 43iQ
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1203169745
Converter serial #: 2022-223
Converter Temp: 325 degC

Start
Calibration slope: 1.002707
Calibration intercept: -0.151356

Finish
1.009574
0.008414

Start
Backgd or Offset: 2.8
Coeff or Slope: 0.808

Finish
2.5
0.808

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4923	82.8	79.9	80.6	0.986
As found Mid point	4967	41.5	40.0	40.1	0.988
As found Low point	4999	20.8	20.0	20.1	0.976
New cylinder response					
Baseline Corr As found:	81.0	Prev response:	79.96	*% change:	1.3%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	1.012861	AF Intercept:	-0.331575
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999988	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4923	82.8	79.9	80.6	0.991
Mid point	4967	41.5	40.0	40.5	0.988
Low point	4999	20.8	20.0	20.3	0.986
As left zero	5000	0.0	0.0	0.1	----
As left span	4923	82.8	79.9	80.7	0.990
SO2 Scrubber Check	4932	82.2	819.7		----
Date of last scrubber change:				Ave Corr Factor	0.988
Date of last converter efficiency test:					

Notes: Changed sample inlet filter after as founds. Ran scrubber check after calibrator zero and it passed.
Adjusted zero only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	July 25, 2025	Previous Calibration:	June 10, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:53	End Time (MST):	15:33
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999991		≥0.995
79.9	80.6	0.9912	Slope	1.009574		0.90 - 1.10
40.0	40.5	0.9882	Intercept	0.008414		+/-3
20.0	20.3	0.9859				



H₂S Calibration Plot

Date: July 25, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH4 / NMHC Calibration Report

Station Information

Station Name: Lower Camp
 Calibration Date: July 22, 2025
 Start time (MST): 10:47
 Reason: Routine

Station number: AMS 11
 Last Cal Date: June 28, 2025
 End time (MST): 15:05

Calibration Standards

Gas Cert Reference: CC741503
 CH4 Cal Gas Conc. 504.8 ppm
 C3H8 Cal Gas Conc. 206.2 ppm
 Removed Gas Cert:
 Removed CH4 Conc. 504.8 ppm
 Removed C3H8 Conc. 206.2 ppm
 Diff between cyl (CH₄):
 Calibrator Model: API T700
 Zero Air Gen model: API T701

Cal Gas Expiry Date: October 9, 2032
 CH4 Equiv Conc. 1071.9 ppm
 Removed Gas Expiry:
 CH4 Equiv Conc. 1071.9 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 3811
 Serial Number: 196

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 12227620776
 NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.64E-04	3.85E-04	NMHC SP Ratio:	7.70E-05	8.31E-05
CH4 Retention time:	15.6	15.8	NMHC Peak Area:	120823	111905
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	82.2	17.57	16.36	1.074
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.36	Prev response	17.54	*% change	-7.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	82.2	17.57	17.48	1.005
Mid point	4971	41.2	8.81	8.65	1.018
Low point	4996	20.6	4.40	4.30	1.023
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	17.57	17.40	1.010
Average Correction Factor					1.015

Notes: No alarms were detected during the investigation. Diagnostics were consistent, and the chromatograms showed no significant discrepancies, aside from a slight alignment drift. The system will be monitored for now, with maintenance planned if necessary in the near future. Changed sample inlet filter after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	82.2	9.30	8.52	1.091
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.52	Prev response	9.28	*% change	-9.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	82.2	9.30	9.21	1.009
Mid point	4971	41.2	4.66	4.56	1.023
Low point	4996	20.6	2.33	2.25	1.034
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	9.30	9.16	1.015
Average Correction Factor					1.022

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	82.2	8.28	7.84	1.056
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.84	Prev response	8.26	*% change	-5.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	82.2	8.28	8.27	1.001
Mid point	4971	41.2	4.15	4.10	1.013
Low point	4996	20.6	2.07	2.05	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	8.28	8.23	1.005
Average Correction Factor					1.008

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.000691	0.995715
THC Cal Offset:	-0.039137	-0.053089
CH ₄ Cal Slope:	1.000464	0.999555
CH ₄ Cal Offset:	-0.020116	-0.018320
NMHC Cal Slope:	1.000881	0.992296
NMHC Cal Offset:	-0.019221	-0.034769

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 22, 2025	Previous Calibration:	June 28, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:47	End Time (MST):	15:05
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620776

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999943	<i>≥0.995</i>
17.57	17.48	1.0051	Slope	0.995715	<i>0.90 - 1.10</i>
8.81	8.65	1.0182	Intercept	-0.053089	<i>+/-0.5</i>
4.40	4.30	1.0229			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 22, 2025	Previous Calibration:	June 28, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:47	End Time (MST):	15:05
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620776

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999951	<i>≥0.995</i>
8.28	8.27	1.0005	Slope	0.999555	<i>0.90 - 1.10</i>
4.15	4.10	1.0133	Intercept	-0.018320	<i>+/-0.5</i>
2.07	2.05	1.0102			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 22, 2025	Previous Calibration:	June 28, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:47	End Time (MST):	15:05
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620776

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999927	<i>≥0.995</i>
9.30	9.21	1.0091	Slope	0.992296	<i>0.90 - 1.10</i>
4.66	4.56	1.0226	Intercept	-0.034769	<i>+/-0.5</i>
2.33	2.25	1.0344			



NMHC Calibration Plot

Date: July 22, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH4 / NMHC Calibration Report

Station Information

Station Name: Lower Camp
Calibration Date: July 29, 2025
Start time (MST): 11:46
Reason: Cylinder Change

Station number: AMS 11
Last Cal Date: July 22, 2025
End time (MST): 13:41

Calibration Standards

Gas Cert Reference:	CC741503	Cal Gas Expiry Date:	October 9, 2032
CH4 Cal Gas Conc.	504.8 ppm	CH4 Equiv Conc.	1071.9 ppm
C3H8 Cal Gas Conc.	206.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	504.8 ppm	CH4 Equiv Conc.	1071.9 ppm
Removed C3H8 Conc.	206.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3811
Zero Air Gen model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 12227620776
NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	3.85E-04	3.85E-04	NMHC SP Ratio:	8.31E-05
CH4 Retention time:	15.8	15.8	NMHC Peak Area:	111905
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	82.2	17.57	17.20	1.022
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.20	Prev response	17.44	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	17.57	17.38	1.011
Average Correction Factor					

Notes:

Changed N2 cylinder



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	82.2	9.30	9.09	1.023
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.09	Prev response	9.19	*% change	-1.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	9.30	9.22	1.008
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	82.2	8.28	8.11	1.020
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.11	Prev response	8.25	*% change	-1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	8.28	8.15	1.015
Average Correction Factor					

Calibration Statistics

THC Cal Slope:	0.995715
THC Cal Offset:	-0.053089
CH ₄ Cal Slope:	0.999555
CH ₄ Cal Offset:	-0.018320
NMHC Cal Slope:	0.992296
NMHC Cal Offset:	-0.034769

Finish


Calibration Performed By:

Mohammed Kashif

NMHC Calibration Plot

Date: July 29, 2025

Location: Lower Camp





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13 FORT MCKAY SOUTH JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Fort McKay South
Calibration Date: July 7, 2025
Start time (MST): 9:33
Reason: Routine

Station number: AMS 13
Last Cal Date: June 3, 2025
End time (MST): 12:20

Calibration Standards

Cal Gas Concentration: 50.55 ppm
Cal Gas Cylinder #: CC260812
Removed Cal Gas Conc: 50.55 ppm
Removed Gas Cyl #: NA
Calibrator Model: Teledyne API T750
Zero Air Gen Model: Teledyne API 751H

Cal Gas Exp Date: December 29, 2028
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 281
Serial Number: 321

Analyzer Information

Analyzer make: Teledyne API T100
Analyzer Range: 0 - 1000 ppb

Serial Number: 599

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000559	1.006762	Backgd or Offset:	103.5	108.0
Calibration intercept:	1.265306	-2.854267	Coeff or Slope:	0.669	0.666

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4921	79.1	799.7	806.3	0.992
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	806.0	Previous response	801.4	*% change	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4921	79.1	799.7	803.5	0.995
Mid point	4960	39.5	399.4	398.1	1.003
Low point	4980	19.8	200.2	195.9	1.022
As left zero	5000	0.0	0.0	-0.4	----
As left span	4921	79.1	799.7	804.0	0.995
Average Correction Factor:					1.007

Notes:

Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 3, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:33	End Time (MST):	12:20
Analyzer make:	Teledyne API T100	Analyzer serial #:	599

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999950	≥0.995
799.7	803.5	0.9953	Slope	1.006762	0.90 - 1.10
399.4	398.1	1.0032	Intercept	-2.854267	+/-30
200.2	195.9	1.0219			



SO2 Calibration Plot

Date: July 7, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort McKay South
Calibration Date: July 10, 2025
Start time (MST): 9:09
Reason: Routine

Station number: AMS 13
Last Cal Date: June 3, 2025
End time (MST): 12:56

Calibration Standards

Cal Gas Concentration: 4.88 ppm
Cal Gas Cylinder #: CC500241
Removed Cal Gas Conc: 4.88 ppm
Removed Gas Cyl #: CC500241
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Exp Date: September 5, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 2657
Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1180540017
Converter serial #: 521
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007709	1.009574	Backgd or Offset:	3.97	3.97
Calibration intercept:	-0.138405	-0.678399	Coeff or Slope:	1.09	1.09

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.5	----
As found High point	4918	81.6	79.6	76.5	1.034
As found Mid point	4959	40.8	39.8	38.2	1.029
As found Low point	4980	20.4	19.9	18.9	1.026
New cylinder response					
Baseline Corr As found:	77.0	Prev response:	80.12	*% change:	-4.1%
Baseline Corr 2nd AF pt:	38.7	AF Slope:	0.966384	AF Intercept:	-0.398457
Baseline Corr 3rd AF pt:	19.4	AF Correlation:	0.999990	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.5	----
High point	4918	81.6	79.6	79.8	0.998
Mid point	4959	40.8	39.8	39.5	1.008
Low point	4980	20.4	19.9	19.2	1.037
As left zero	5000	0.0	0.0	-0.5	----
As left span	4918	81.6	79.6	80.3	0.992
SO2 Scrubber Check	4921	79.1	791.0	0.1	----
Date of last scrubber change:	3-Jun-25		Ave Corr Factor		1.014
Date of last converter efficiency test:					

Notes: Changed inlet filter after as found. SO2 scrubber check after calibrator zero and passed. No adjustment.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	July 10, 2025	Previous Calibration:	June 3, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:09	End Time (MST):	12:56
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.5	----	Correlation Coefficient	0.999976	≥ 0.995
79.6	79.8	0.9981	Slope	1.009574	$0.90 - 1.10$
39.8	39.5	1.0082	Intercept	-0.678399	± 3
19.9	19.2	1.0369			

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Fort McKay South	Station number:	AMS 13
Calibration Date:	July 7, 2025	Last Cal Date:	June 5, 2025
Start time (MST):	9:33	End time (MST):	12:20
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	December 29, 2028
CH ₄ Cal Gas Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T750	Serial Number:	281
Zero Air Gen model:	Teledyne API 751H	Serial Number:	321

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018
THC Range:	0 - 20 ppm	NMHC/CH ₄ Range:	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.94E-04	2.97E-04	NMHC SP Ratio:	4.70E-05
CH ₄ Retention time:	15.40	16.20	NMHC Peak Area:	193215
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	17.05	17.24	0.989
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.24	Prev response	17.06	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	17.05	17.05	1.000
Mid point	4960	39.5	8.51	8.47	1.005
Low point	4980	19.8	4.27	4.21	1.015
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	17.05	17.04	1.001
Average Correction Factor					1.007

Notes: Changed inlet filter after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.1	9.08	9.15	0.992
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.15	Prev response	9.09	*% change	0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	9.08	9.08	1.000
Mid point	4960	39.5	4.53	4.52	1.004
Low point	4980	19.8	2.27	2.25	1.012
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	9.08	9.07	1.001
Average Correction Factor					1.005

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.1	7.97	8.09	0.985
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.09	Prev response	7.97	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	7.97	7.97	0.999
Mid point	4960	39.5	3.98	3.95	1.007
Low point	4980	19.8	1.99	1.96	1.017
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	7.97	7.96	1.000
Average Correction Factor					1.008

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.003416	1.001276
THC Cal Offset:	-0.043114	-0.034909
CH ₄ Cal Slope:	1.004623	1.001953
CH ₄ Cal Offset:	-0.036351	-0.020547
NMHC Cal Slope:	1.002394	1.000682
NMHC Cal Offset:	-0.006162	-0.014362

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 5, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:33	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999981	≥ 0.995
17.05	17.05	0.9997	Slope	1.001276	$0.90 - 1.10$
8.51	8.47	1.0055	Intercept	-0.034909	± 0.5
4.27	4.21	1.0145			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 5, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:33	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999968	≥ 0.995
7.97	7.97	0.9991	Slope	1.001953	$0.90 - 1.10$
3.98	3.95	1.0073	Intercept	-0.020547	± 0.5
1.99	1.96	1.0170			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 5, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:33	End Time (MST):	12:20
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999989		≥ 0.995
9.08	9.08	1.0001	Slope	1.000682		$0.90 - 1.10$
4.53	4.52	1.0039	Intercept	-0.014362		± 0.5
2.27	2.25	1.0124				



NMHC Calibration Plot

Date: July 7, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort McKay South
Station number: AMS 13
Calibration Date: July 16, 2025
Last Cal Date: June 3, 2025
Start time (MST): 8:53
End time (MST): 11:29
Reason: As Found

Calibration Standards

NO Gas Cylinder #: T2UP1RP
NOX Cal Gas Conc: 48.25 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.25 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne APIT701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 47.88 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.88 ppm
NO gas Diff:
Serial Number: 2657
Serial Number: 1118

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.5	0.3	0.2	----	----
AF High point	4917	83.5	805.7	799.5	6.2	856.5	847.7	8.8	0.9412	0.9435
AF Mid point	4958	41.8	403.4	400.3	3.1	424.4	418.2	6.2	0.9516	0.9579
AF Low point	4979	20.9	201.7	200.1	1.5	209.9	204.1	5.7	0.9632	0.9821
New cyl resp										
Previous Response	NO _x = 804.7 ppb	NO = 800.3 ppb	* = > +/-5% change initiates investigation				*Percent Change		NO _x = 6.0%	
Baseline Corr 1st pt	NO _x = 856.0 ppb	NO = 847.4 ppb	<u>As Found Statistics</u>				*Percent Change		NO = 5.6%	
Baseline Corr 2nd pt	NO _x = 423.9 ppb	NO = 417.9 ppb	As found NO _x r ² : 0.999945				Nx SI: 1.063847		Nx Int: -2.386	
Baseline Corr 3rd pt	NO _x = 209.4 ppb	NO = 203.8 ppb	As found NO r ² : 0.999865				NO SI: 1.062545		NO Int: -4.303	
			As found NO ₂ r ² : 0.999948				NO ₂ SI: 0.997739		NO ₂ Int: 1.464	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	0.2	----	----
As found high GPT point	845.2	427.9	423.5	423.6	0.9997	100.0%
As found mid GPT point	845.2	635.4	216.0	216.8	0.9962	100.4%
As found low GPT point	845.2	742.5	108.9	111.9	0.9730	102.8%



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12300522720

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.244	
NOX coeff or slope:	1.001	
NO2 coeff or slope:	1.000	

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	3.3	
NOX bkgnd or offset:	3.3	
Reaction cell Press:	147.2	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.004251	
NO _x Cal Offset:	-4.391596	
NO Cal Slope:	1.007936	
NO Cal Offset:	-5.590352	
NO ₂ Cal Slope:	0.988380	
NO ₂ Cal Offset:	1.407333	

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero
High point
Mid point
Low point
As left zero
As left span

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero
High GPT point
Mid GPT point
Low GPT point

Average Correction Factor


Notes: Changed inlet filter, o-ring and capillaries after as founds. O-rings and capillaries were changed to see if it improve linearity.

Calibration Performed By: Sean Bala

NO_x Calibration Plot

Date: July 16, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort McKay South
Station number: AMS 13
Calibration Date: July 17, 2025
Last Cal Date: July 16, 2025
Start time (MST): 9:07
End time (MST): 12:53
Reason: Maintenance

Calibration Standards

NO Gas Cylinder #: T2UP1RP
NOX Cal Gas Conc: 48.25 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.25 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne APIT701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 47.88 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.88 ppm
NO gas Diff:
Serial Number: 2657
Serial Number: 1118

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation		*Percent Change		NO _x =	NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>		*Percent Change		NO =	NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:		Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:		NO Int:	
					As found	NO ₂ r ² :	NO2 SI:		NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12300522720

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.244	1.115
NOX coeff or slope:	1.001	1.002
NO2 coeff or slope:	1.000	1.000

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	3.3	3.1
NOX bkgnd or offset:	3.3	3.5
Reaction cell Press:	147.2	152.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.004251	1.002824
NO _x Cal Offset:	-4.391596	-1.613555
NO Cal Slope:	1.007936	1.003967
NO Cal Offset:	-5.590352	-3.151390
NO ₂ Cal Slope:	0.988380	0.991296
NO ₂ Cal Offset:	1.407333	0.082751

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	-0.1	----	----
High point	4917	83.5	805.7	799.5	6.2	807.3	801.2	6.2	0.9980	0.9979
Mid point	4958	41.8	403.4	400.3	3.1	401.6	396.8	4.8	1.0044	1.0088
Low point	4979	20.9	201.7	200.1	1.5	199.5	195.0	4.5	1.0110	1.0264
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.2	----	----
As left span	4917	83.5	805.7	407.0	398.7	801.0	407.0	394.0	1.0059	1.0000
Average Correction Factor									1.0045	1.0110

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	800.5	405.1	401.6	398.1	1.0087	99.1%
Mid GPT point	800.5	603.6	203.1	201.4	1.0083	99.2%
Low GPT point	800.5	702.9	103.8	103.2	1.0056	99.4%
Average Correction Factor					1.0076	99.3%

Notes:

Adjusted zero and span.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 17, 2025	Previous Calibration:	July 16, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:07	End Time (MST):	12:53
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999983	≥0.995
805.7	807.3	0.9980	Slope	1.002824	0.90 - 1.10
403.4	401.6	1.0044	Intercept	-1.613555	+/-20
201.7	199.5	1.0110			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 17, 2025	Previous Calibration:	July 16, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:07	End Time (MST):	12:53
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999999	≥0.995
401.6	398.1	1.0087	Slope	0.991296	0.90 - 1.10
203.1	201.4	1.0083	Intercept	0.082751	+/-20
103.8	103.2	1.0056			





Wood Buffalo Environmental Association

NO Calibration Summary


Station Information

Calibration Date:	July 17, 2025	Previous Calibration:	July 16, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:07	End Time (MST):	12:53
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999933	≥ 0.995
799.5	801.2	0.9979	Slope	1.003967	$0.90 - 1.10$
400.3	396.8	1.0088	Intercept	-3.151390	± 20
200.1	195.0	1.0264			


NO Calibration Curve



NO_x Calibration Plot

Date: July 17, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort McKay South
Station number: AMS 13
Calibration Date: July 29, 2025
Last Cal Date: July 17, 2025
Start time (MST): 10:34
End time (MST): 16:05
Reason: Maintenance

Calibration Standards

NO Gas Cylinder #: T2UP1RP
NOX Cal Gas Conc: 48.25 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.25 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne APIT701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 47.88 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.88 ppm
NO gas Diff:
Serial Number: 2657
Serial Number: 1118

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.5	-0.6	----	----
AF High point	4917	83.5	805.7	799.5	6.2	728.9	716.5	12.4	1.1052	1.1166
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 806.4 ppb	NO = 799.5 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -10.6%	
Baseline Corr 1st pt	NO _x = 729.0 ppb	NO = 716.0 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -11.7%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12300522720

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.115	1.115	NO bkgnd or offset:	3.1	3.1
NOX coeff or slope:	1.002	1.002	NOX bkgnd or offset:	3.5	3.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	152.8	152.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002824	1.004055
NO _x Cal Offset:	-1.613555	-3.172683
NO Cal Slope:	1.003967	1.005508
NO Cal Offset:	-3.151390	-4.490645
NO ₂ Cal Slope:	0.991296	0.996864
NO ₂ Cal Offset:	0.082751	1.668208

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.2	----	----
High point	4917	83.5	805.7	799.5	6.2	807.8	802.1	5.6	0.9974	0.9968
Mid point	4958	41.8	403.4	400.3	3.1	399.1	394.7	4.4	1.0107	1.0142
Low point	4979	20.9	201.7	200.1	1.5	196.8	192.7	4.1	1.0248	1.0386
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.4	-0.2	----	----
As left span	4917	83.5	805.7	410.1	395.6	810.5	410.1	400.4	0.9941	1.0000
Average Correction Factor									1.0110	1.0165

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	802.8	406.8	402.2	401.8	1.0009	99.9%
Mid GPT point	802.8	607.8	201.2	202.7	0.9925	100.8%
Low GPT point	802.8	708.3	100.7	104.2	0.9662	103.5%
Average Correction Factor					0.9865	101.4%

Notes: As found span is 11% low. No multipoint as founds since instrument is out of limits. Swapped out pump and charcoal pack. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 29, 2025	Previous Calibration:	July 17, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:34	End Time (MST):	16:05
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999920	≥0.995
805.7	807.8	0.9974	Slope	1.004055	0.90 - 1.10
403.4	399.1	1.0107	Intercept	-3.172683	+/-20
201.7	196.8	1.0248			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 29, 2025	Previous Calibration:	July 17, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:34	End Time (MST):	16:05
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999897	≥ 0.995
402.2	401.8	1.0009	Slope	0.996864	$0.90 - 1.10$
201.2	202.7	0.9925	Intercept	1.668208	± 20
100.7	104.2	0.9662			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 29, 2025	Previous Calibration:	July 17, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:34	End Time (MST):	16:05
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.2	----	Correlation Coefficient	0.999840	≥ 0.995
799.5	802.1	0.9968	Slope	1.005508	0.90 - 1.10
400.3	394.7	1.0142	Intercept	-4.490645	+/-20
200.1	192.7	1.0386			



NO_x Calibration Plot

Date: July 29, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Fort McKay South
Calibration Date: July 4, 2025
Start time (MST): 9:45
Reason: Routine

Station number: AMS 13
Last Cal Date: June 24, 2025
End time (MST): 12:52

Calibration Standards

O3 generation mode: Photometer
Calibrator Make/Model: Teledyne API T750
ZAG Make/Model: Teledyne API 751H

Serial Number: 281
Serial Number: 321

Analyzer Information

Analyzer make: Teledyne API T400
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 7413

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003314	1.007771	Backgd or Offset:	-2.6	-2.5
Calibration intercept:	0.720000	1.040000	Coeff or Slope:	1.063	1.041

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-0.7	----
As found High point	5000	950.2	400.0	411.8	0.970
As found Mid point					
As found Low point					
Baseline Corr As found:	412.5	Previous response	402.0	*% change	2.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.3	----
High point	5000	955.4	400.0	403.7	0.991
Mid point	5000	781.1	200.0	203.2	0.984
Low point	5000	668.6	100.0	102.4	0.977
As left zero	5000	800.0	0.0	-0.4	----
As left span	5000	963.0	400.0	406.9	0.983
Average Correction Factor					0.984

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	July 4, 2025	Previous Calibration:	June 24, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:45	End Time (MST):	12:52
Analyzer make:	Teledyne API T400	Analyzer serial #:	7413

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999984	≥ 0.995
400.0	403.7	0.9908	Slope	1.007771	$0.90 - 1.10$
200.0	203.2	0.9843	Intercept	1.040000	± 5
100.0	102.4	0.9766			



O₃ Calibration Plot

Date: July 4, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay South Station number: AMS 13
Calibration Date: July 21, 2025 Last Cal Date: June 2, 2025
Start time (MST): 10:34 End time (MST): 10:48

Analyzer Make: Teledyne API T640 S/N: 1335
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388755
Temp/RH standard: Alicat FP-25BT S/N: 388755

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	15.5	15.01	15.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	736.3	737.87	736.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.03	5.06	5.03	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	41	----	41	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.2	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: June 2, 2025
Date Disposable Filter Changed: June 2, 2025

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: June 2, 2025
Date RH/T Sensor Cleaned: June 2, 2025

Notes: No adjustments made.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC
JULY 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Anzac Station number: AMS 14
Calibration Date: July 10, 2025 Last Cal Date: June 9, 2025
Start time (MST): 10:04 End time (MST): 13:32
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.32 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC462030
Removed Cal Gas Conc: 50.32 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3060
Zero Air Gen Model: API T701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 0710321322
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004081	1.009216	Backgd or Offset:	25.6	25.5
Calibration intercept:	-1.920292	-1.025803	Coeff or Slope:	1.111	1.111

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4937	79.6	798.4	804.0	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	804.1	Previous response	799.8	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4937	79.6	798.4	805.5	0.991
Mid point	4977	39.7	398.2	399.5	0.997
Low point	4992	19.7	197.8	198.4	0.997
As left zero	5000	0.0	0.0	-0.1	----
As left span	4937	79.6	798.4	806.2	0.990
Average Correction Factor:					0.995

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 9, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:04	End Time (MST):	13:32
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999991	≥0.995
798.4	805.5	0.9912	Slope	1.009216	0.90 - 1.10
398.2	399.5	0.9968	Intercept	-1.025803	+/-30
197.8	198.4	0.9970			



SO2 Calibration Plot

Date: July 10, 2025

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Anzac Station number: AMS 14
Calibration Date: July 7, 2025 Last Cal Date: June 5, 2025
Start time (MST): 10:06 End time (MST): 16:48
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC510379
Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3060
ZAG Make/Model: API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153582
Converter make: CD Nova CDN-101 Converter serial #: 629
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011025	1.020456	Backgd or Offset:	2.2 2.4
Calibration intercept:	-0.464585	-0.444257	Coeff or Slope:	0.944 1.015

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4938	77.9	80.0	80.5	0.992
As found Mid point	4973	38.9	40.0	39.2	1.017
As found Low point	4997	19.4	19.9	19.0	1.043
New cylinder response					
Baseline Corr As found:	80.6	Prev response:	80.40	*% change:	0.2%
Baseline Corr 2nd AF pt:	39.3	AF Slope:	1.010489	AF Intercept:	-0.684586
Baseline Corr 3rd AF pt:	19.1	AF Correlation:	0.999741	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4938	77.9	80.0	81.3	0.983
Mid point	4973	38.9	40.0	40.3	0.991
Low point	4997	19.4	19.9	19.3	1.032
As left zero	5000	0.0	0.0	0.1	----
As left span	4938	77.9	80.0	78.2	1.022
SO2 Scrubber Check	4941	79.7	793.7	0.0	----
Date of last scrubber change:	28-May-25		Ave Corr Factor		1.002
Date of last converter efficiency test:					

Notes: Sample inlet filter changed after asfound. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


TRS Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 5, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:06	End Time (MST):	16:48
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999849		≥ 0.995
80.0	81.3	0.9834	Slope	1.020456		$0.90 - 1.10$
40.0	40.3	0.9915	Intercept	-0.444257		± 3
19.9	19.3	1.0316				



TRS Calibration Plot

Date: July 7, 2025

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Anzac
Calibration Date: July 10, 2025
Start time (MST): 10:04
Reason: Routine

Station number: AMS 14
Last Cal Date: June 9, 2025
End time (MST): 13:32

Calibration Standards

Gas Cert Reference:	CC462030	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	505.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	204.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	505.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	204.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
Zero Air Gen model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1331259521
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.95E-04	2.95E-04	NMHC SP Ratio:	5.82E-05
CH ₄ Retention time:	14.9	14.9	NMHC Peak Area:	153524
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4937	79.6	16.96	16.85	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.85	Prev response	16.87	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4937	79.6	16.96	16.80	1.010
Mid point	4977	39.7	8.46	8.20	1.031
Low point	4992	19.7	4.20	4.03	1.042
As left zero	5000	0.0	0.00	0.00	----
As left span	4937	79.6	16.96	16.69	1.016
				Average Correction Factor	1.028

Notes:

Changed sample inlet filter after as founds. No adjustments made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4937	79.6	8.94	8.87	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.87	Prev response	8.89	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4937	79.6	8.94	8.82	1.014
Mid point	4977	39.7	4.46	4.34	1.027
Low point	4992	19.7	2.21	2.14	1.035
As left zero	5000	0.0	0.00	0.00	----
As left span	4937	79.6	8.94	8.76	1.021
Average Correction Factor					1.025

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4937	79.6	8.02	7.98	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.98	Prev response	7.97	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4937	79.6	8.02	7.98	1.005
Mid point	4977	39.7	4.00	3.86	1.037
Low point	4992	19.7	1.99	1.89	1.050
As left zero	5000	0.0	0.00	0.00	----
As left span	4937	79.6	8.02	7.93	1.012
Average Correction Factor					1.031

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.000547	0.992013
THC Cal Offset:	-0.101304	-0.088253
CH ₄ Cal Slope:	1.001638	0.996805
CH ₄ Cal Offset:	-0.060777	-0.058244
NMHC Cal Slope:	0.999287	0.987422
NMHC Cal Offset:	-0.039927	-0.029611

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 9, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:04	End Time (MST):	13:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999841	≥ 0.995
16.96	16.80	1.0096	Slope	0.992013	$0.90 - 1.10$
8.46	8.20	1.0314	Intercept	-0.088253	± 0.5
4.20	4.03	1.0425			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 9, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:04	End Time (MST):	13:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999678	<i>≥0.995</i>
8.02	7.98	1.0051	Slope	0.996805	<i>0.90 - 1.10</i>
4.00	3.86	1.0370	Intercept	-0.058244	<i>+/-0.5</i>
1.99	1.89	1.0498			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 9, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:04	End Time (MST):	13:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999939	≥ 0.995
8.94	8.82	1.0138	Slope	0.987422	$0.90 - 1.10$
4.46	4.34	1.0270	Intercept	-0.029611	± 0.5
2.21	2.14	1.0355			



NMHC Calibration Plot

Date: July 10, 2025

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Anzac
Station number: AMS 14
Calibration Date: July 15, 2025
Last Cal Date: June 4, 2025
Start time (MST): 9:32
End time (MST): 14:16
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0037092
NOX Cal Gas Conc: 60.7 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.70 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T700H
Cal Gas Expiry Date: May 16, 2031
NO Cal Gas Conc: 60.40 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.40 ppm
NO gas Diff:
Serial Number: 3060
Serial Number: 357

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
AF High point	4934	66.3	804.8	800.9	4.0	813.7	808.1	5.5	0.9892	0.9909
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 813.3 ppb	NO = 810.7 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.0%
Baseline Corr 1st pt	NO _x = 813.6 ppb	NO = 808.2 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -0.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430008

Calibration Statistics

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.435	1.435	NO bkgnd or offset:	3.9	3.9	NO _x Cal Slope:	1.010781	1.008381
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.9	3.9	NO _x Cal Offset:	-0.251265	-0.531254
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.5	158.5	NO Cal Slope:	1.015030	1.011974
						NO Cal Offset:	-2.231101	-2.410523
						NO ₂ Cal Slope:	0.995053	0.996563
						NO ₂ Cal Offset:	-1.224921	-1.047395

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
High point	4934	66.3	804.8	800.9	4.0	811.2	809.0	2.2	0.9922	0.9899
Mid point	4985	33.2	401.6	399.6	2.0	404.7	401.5	3.3	0.9923	0.9953
Low point	5004	16.7	201.9	200.9	1.0	201.9	198.0	4.0	1.0000	1.0147
As left zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.2	----	----
As left span	4934	66.3	804.8	425.4	379.4	809.9	425.4	384.4	0.9937	1.0000
Average Correction Factor									0.9948	1.0000

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	806.3	420.8	389.5	387.8	1.0043	99.6%
Mid GPT point	806.3	598.4	211.9	209.5	1.0113	98.9%
Low GPT point	806.3	700.7	109.6	106.8	1.0260	97.5%
Average Correction Factor					1.0139	98.6%

Notes:

Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 4, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:32	End Time (MST):	14:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999994	≥0.995
804.8	811.2	0.9922	Slope	1.008381	0.90 - 1.10
401.6	404.7	0.9923	Intercept	-0.531254	+/-20
201.9	201.9	1.0000			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 4, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:32	End Time (MST):	14:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999948	≥0.995
389.5	387.8	1.0043	Slope	0.996563	0.90 - 1.10
211.9	209.5	1.0113	Intercept	-1.047395	+/-20
109.6	106.8	1.0260			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 4, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:32	End Time (MST):	14:16
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999957	≥ 0.995
800.9	809.0	0.9899	Slope	1.011974	$0.90 - 1.10$
399.6	401.5	0.9953	Intercept	-2.410523	± 20
200.9	198.0	1.0147			



NO_x Calibration Plot

Date: July 15, 2025

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Anzac
Calibration Date: July 4, 2025
Start time (MST): 10:54
Reason: Routine

Station number: AMS 14
Last Cal Date: June 2, 2025
End time (MST): 14:03

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: API T700
ZAG Make/Model: API 701H

Serial Number: 3060
Serial Number: 357

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1426262595

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.982886	0.994486	Backgd or Offset:	1.6	1.6
Calibration intercept:	1.220000	0.540000	Coeff or Slope:	1.691	1.691

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.6	----
As found High point	5000	935.9	400.0	399.2	1.004
As found Mid point					
As found Low point					
Baseline Corr As found:	398.6	Previous response	394.4	*% change	1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	5000	935.9	400.0	398.1	1.005
Mid point	5000	817.5	200.0	200.0	1.000
Low point	5000	722.8	100.0	99.8	1.002
As left zero	5000	0.0	0.0	0.4	----
As left span	5000	935.9	400.0	401.9	0.995
Average Correction Factor					1.002

Notes: Sample inlet filter changed after asfound. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	July 4, 2025	Previous Calibration:	June 2, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:54	End Time (MST):	14:03
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999995	≥ 0.995
400.0	398.1	1.0048	Slope	0.994486	$0.90 - 1.10$
200.0	200.0	1.0000	Intercept	0.540000	± 5
100.0	99.8	1.0020			



O₃ Calibration Plot

Date: July 4, 2025

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Anzac
Calibration Date: July 21, 2025
Start time (MST): 10:10
Station number: AMS 14
Last Cal Date: June 23, 2025
End time (MST): 10:28
Analyzer Make: AP T640
S/N: 825
Particulate Fraction: PM2.5
Flow Meter Make/Model: Alicat FP-25BT
S/N: 388749
Temp/RH standard: Alicat FP-25BT
S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	13.9	12.88	13.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	716.7	718.03	716.7	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.010	4.946	5.010	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	-----	37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 11.3		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: Expiry Date:
Lot No.:

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: May 22, 2025
Date Disposable Filter Changed: June 23, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 29, 2024
Date RH/T Sensor Cleaned: August 29, 2024

Notes: No adjustments made. Leak check passed. Head cleaned.

Calibration by: Mohammed Kashif



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS17
WAPASU
JULY 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Wapasu Station number: AMS17
Calibration Date: July 11, 2025 Last Cal Date: June 10, 2025
Start time (MST): 10:31 End time (MST): 13:42
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.38 ppm Cal Gas Exp Date: January 12, 2029
Cal Gas Cylinder #: ALM066507
Removed Cal Gas Conc: 50.38 ppm Rem Gas Exp Date: N/A
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 2449
Zero Air Gen Model: Teledyne API 701H Serial Number: 1238

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1218153459
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998497	0.997853	Backgd or Offset:	14.0	14.0
Calibration intercept:	-1.680288	-1.279869	Coeff or Slope:	1.109	1.109

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4921	79.4	800.0	797.8	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.5	Previous response	797.1	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4921	79.4	800.0	798.3	1.002
Mid point	4960	39.7	400.0	395.6	1.011
Low point	4980	19.8	199.5	197.1	1.012
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	79.4	800.1	801.7	0.998
Average Correction Factor:					1.009

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	July 11, 2025	Previous Calibration:	June 10, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:31	End Time (MST):	13:42
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999970	≥0.995
800.0	798.3	1.0021	Slope	0.997853	0.90 - 1.10
400.0	395.6	1.0112	Intercept	-1.279869	+/-30
199.5	197.1	1.0122			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: July 11, 2025

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Wapasu
Calibration Date: July 17, 2025
Start time (MST): 10:11
Reason: Routine

Station number: AMS 17
Last Cal Date: June 24, 2025
End time (MST): 15:21

Calibration Standards

Cal Gas Concentration: 4.77 ppm
Cal Gas Cylinder #: DT20029267
Removed Cal Gas Conc: 4.77 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API T701H

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: Thermo 450i
Converter make: CD Nova
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1218153583
Converter serial #: N/A
Converter Temp: 340 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.988078	0.990930	Backgd or Offset:	13.1
Calibration intercept:	0.080021	-0.219831	Coeff or Slope:	1.099

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4921	83.9	80.0	82.0	0.978
As found Mid point	4961	41.9	39.9	40.6	0.989
As found Low point	4980	21.0	20.0	20.3	0.997
New cylinder response					
Baseline Corr As found:	81.8	Prev response:	79.09	*% change:	3.3%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.023989	AF Intercept:	-0.049681
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999950	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4916	83.9	80.0	79.3	1.009
Mid point	4958	41.9	40.0	39.2	1.020
Low point	4979	21.0	20.0	19.2	1.043
As left zero	5000	0.0	0.0	0.4	----
As left span	4916	83.9	80.0	76.3	1.049
SO2 Scrubber Check	4921	79.4	793.9	-0.2	----
Date of last scrubber change:	N/A		Ave Corr Factor		
Date of last converter efficiency test:	N/A		1.024		

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	July 17, 2025	Previous Calibration:	June 24, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:11	End Time (MST):	15:21
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999872		≥ 0.995
80.0	79.3	1.0094	Slope	0.990930		0.90 - 1.10
40.0	39.2	1.0197	Intercept	-0.219831		+/-3
20.0	19.2	1.0434				



H₂S Calibration Plot

Date: July 17, 2025

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Wapasu Station number: AMS17
Calibration Date: July 11, 2025 Last Cal Date: June 10, 2025
Start time (MST): 10:31 End time (MST): 13:41
Reason: Routine

Calibration Standards

Gas Cert Reference: ALM066507 Cal Gas Expiry Date: January 12, 2029
CH4 Cal Gas Conc. 503.5 ppm CH4 Equiv Conc. 1076.3 ppm
C3H8 Cal Gas Conc. 208.3 ppm
Removed Gas Cert: n/a Removed Gas Expiry:
Removed CH4 Conc. 503.5 ppm CH4 Equiv Conc. 1076.3 ppm
Removed C3H8 Conc. 208.3 ppm Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 2449
ZAG Make/Model: Teledyne API 701H Serial Number: 1238

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1218153352
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001644	0.992956	Background:	3.340	3.140
Calibration intercept:	-0.125948	-0.019000	Coefficient:	4.476	4.433

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	-0.13	----
As found High point	4921	79.4	17.09	16.98	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.11	Previous response	16.99	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.05	----
High point	4921	79.4	17.09	16.98	1.006
Mid point	4960	39.7	8.55	8.44	1.013
Low point	4980	19.8	4.26	4.14	1.030
As left zero	5000	0.0	0.00	0.04	----
As left span	4921	79.4	17.09	16.96	1.008
Average Correction Factor					1.016

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	July 11, 2025	Previous Calibration:	June 10, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:31	End Time (MST):	13:41
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.05	----	Correlation Coefficient	0.999922	≥ 0.995
17.09	16.98	1.0063	Slope	0.992956	$0.90 - 1.10$
8.55	8.44	1.0127	Intercept	-0.019000	± 1.5
4.26	4.14	1.0298			


THC Calibration Curve



THC Calibration Plot

Date: July 11, 2025

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Wapasu
Station number: AMS 17
Calibration Date: July 21, 2025
Last Cal Date: June 18, 2025
Start time (MST): 10:59
End time (MST): 15:51
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T375YK8
NOX Cal Gas Conc: 49.11 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 49.11 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: April 13, 2025
NO Cal Gas Conc: 48.07 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 48.07 ppm
NO gas Diff:
Serial Number: 2449
Serial Number: 359

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
AF High point	4917	83.2	817.2	799.9	17.3	810.9	793.5	17.4	1.0080	1.0082
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 815.8 ppb	NO = 799.0 ppb				* = > +/-5% change initiates investigation			*Percent Change	NO _x = -0.6%
Baseline Corr 1st pt	NO _x = 810.7 ppb	NO = 793.4 ppb				<u>As Found Statistics</u>			*Percent Change	NO = -0.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found NO _x r ² :			Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found NO r ² :			NO SI:	NO Int:
						As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo Scientific 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153460

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.116	1.123	NO bkgnd or offset:	3.9	3.9
NOX coeff or slope:	0.993	0.996	NOX bkgnd or offset:	4.3	4.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	244.5	248.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999760	1.000599
NO _x Cal Offset:	-1.160000	-0.660000
NO Cal Slope:	1.000787	1.000658
NO Cal Offset:	-1.500000	-1.480000
NO ₂ Cal Slope:	0.998829	1.001614
NO ₂ Cal Offset:	-0.511740	-0.055111

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.2	----	----
High point	4917	83.2	817.2	799.9	17.3	817.2	799.6	17.5	1.0000	1.0004
Mid point	4958	41.6	408.6	399.9	8.7	408.0	398.1	9.9	1.0015	1.0046
Low point	4979	20.8	204.3	200.0	4.3	203.4	197.2	6.2	1.0044	1.0141
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
As left span	4917	83.2	817.2	402.1	415.1	819.1	402.1	417.0	0.9977	1.0000
Average Correction Factor									1.0020	1.0063

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	798.3	399.3	416.3	416.7	0.9991	100.1%
Mid GPT point	798.3	600.5	215.1	215.9	0.9963	100.4%
Low GPT point	798.3	701.4	114.2	114.2	1.0000	100.0%
Average Correction Factor					0.9985	100.2%

Notes:

Sample inlet filter changed after as founds. Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 21, 2025	Previous Calibration:	June 18, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:59	End Time (MST):	15:51
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999999	≥0.995
817.2	817.2	1.0000	Slope	1.000599	0.90 - 1.10
408.6	408.0	1.0015	Intercept	-0.660000	+/-20
204.3	203.4	1.0044			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 21, 2025	Previous Calibration:	June 18, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:59	End Time (MST):	15:51
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999996	≥ 0.995
416.3	416.7	0.9991	Slope	1.001614	$0.90 - 1.10$
215.1	215.9	0.9963	Intercept	-0.055111	± 20
114.2	114.2	1.0000			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 21, 2025	Previous Calibration:	June 18, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:59	End Time (MST):	15:51
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999986	≥ 0.995
799.9	799.6	1.0004	Slope	1.000658	$0.90 - 1.10$
399.9	398.1	1.0046	Intercept	-1.480000	± 20
200.0	197.2	1.0141			



NO_x Calibration Plot

Date: July 21, 2025

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Wapasu
Calibration Date: July 2, 2025
Start time (MST): 10:15
Reason: Routine

Station number: AMS17
Last Cal Date: June 4, 2025
End time (MST): 14:49

Calibration Standards

O3 generation mode: Photometer
Calibrator Make/Model: API T700
ZAG Make/Model: API T701H

Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: API T400
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 7045

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001286	1.000914	Backgd or Offset:	1.3	1.3
Calibration intercept:	-0.600000	-0.860000	Coeff or Slope:	1.025	1.025

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/Ic- AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	5000	1104.7	400.0	402.6	0.993
As found Mid point					
As found Low point					
Baseline Corr As found:	402.7	Previous response	399.9	*% change	0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
					<i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	5000	1104.7	400.0	400.2	1.000
Mid point	5000	917.3	200.0	198.4	1.008
Low point	5000	797.9	100.0	98.3	1.017
As left zero	5000	0.0	0.0	-0.1	----
As left span	5000	1104.0	400.0	407.1	0.983
Average Correction Factor					1.008

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	July 2, 2025	Previous Calibration:	June 4, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:15	End Time (MST):	14:49
Analyzer make:	API T400	Analyzer serial #:	7045

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999960	≥0.995
400.0	400.2	0.9995	Slope	1.000914	0.90 - 1.10
200.0	198.4	1.0081	Intercept	-0.860000	+/- 5
100.0	98.3	1.0173			



O₃ Calibration Plot

Date: July 2, 2025

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Wapasu Station number: AMS 17
Calibration Date: July 21, 2025 Last Cal Date: June 24, 2025
Start time (MST): 14:38 End time (MST): 15:29

Analyzer Make: Teledyne API T640 S/N: 1183
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	14.50	14.10	14.50	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	715.40	716.60	715.40	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	4.95	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	34	----	34	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 3.1		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 6, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	N/A	N/A	N/A	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: June 24, 2025
Date Disposable Filter Changed: July 21, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 21, 2025
Date RH/T Sensor Cleaned: July 21, 2025

Notes: Flow, temp and pressure checked. Leak check passed. DFU filter changed out. Annual maintenance done.

Calibration by: Aswin Sasi Kumar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: July 23, 2025 Last Cal Date: June 16, 2025
Start time (MST): 11:30 End time (MST): 15:29
Reason: Routine

Calibration Standards

Cal Gas Concentration: 51.22 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC417455
Removed Cal Gas Conc: 51.22 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: CC417455 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 282
Zero Air Gen Model: Teledyne API 701H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1501301453
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000278	1.005581	Backgd or Offset:	25.3	25.3
Calibration intercept:	-2.670930	-3.459390	Coeff or Slope:	0.818	0.818

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.4	----
As found High point	4921	78.1	800.2	800.5	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	800.1	Previous response	797.8	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.6	----
High point	4921	78.1	800.2	802.6	0.997
Mid point	4960	39.1	400.6	399.4	1.003
Low point	4980	20.0	204.9	197.1	1.039
As left zero	5000	0.0	0.0	0.5	----
As left span	4921	78.1	800.2	803.0	0.997
Average Correction Factor:					1.013

Notes: Changed sample inlet filter. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	June 16, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:30	End Time (MST):	15:29
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999863	≥0.995
800.2	802.6	0.9970	Slope	1.005581	0.90 - 1.10
400.6	399.4	1.0030	Intercept	-3.459390	+/-30
204.9	197.1	1.0395			



SO2 Calibration Plot

Date: July 23, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS18
Calibration Date: July 22, 2025 Last Cal Date: June 17, 2025
Start time (MST): 10:55 End time (MST): 15:35
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.86 ppm Cal Gas Exp Date: May 9, 2027
Cal Gas Cylinder #: CC523103
Removed Cal Gas Conc: 4.86 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
ZAG Make/Model: Teledyne API T701 Serial Number: 360

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359
Converter make: CD Nova CDN-101 Converter serial #: 555
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992940	1.003226	Backgd or Offset:	2.94
Calibration intercept:	0.001163	-0.059096	Coeff or Slope:	1.181

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4917	82.3	80.0	83.3	0.962
As found Mid point	4958	41.2	40.1	41.8	0.960
As found Low point	4979	20.6	20.0	20.5	0.982
New cylinder response					
Baseline Corr As found:	83.2	Prev response:	79.44	*% change:	4.5%
Baseline Corr 2nd AF pt:	41.7	AF Slope:	1.041798	AF Intercept:	-0.059918
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999965	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4917	82.3	80.0	80.3	0.996
Mid point	4958	41.2	40.1	40.1	0.999
Low point	4979	20.6	20.0	19.7	1.016
As left zero	5000	0.0	0.0	0.4	----
As left span	4917	82.3	80.0	79.5	1.006
SO2 Scrubber Check	4923	77.1	771.0	0.0	----
Date of last scrubber change:	17-Dec-21		Ave Corr Factor		1.004
Date of last converter efficiency test:					

Notes: No adjustment made. SOX scrubber tested post filter change, no issues to note.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	July 22, 2025	Previous Calibration:	June 17, 2025
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:55	End Time (MST):	15:35
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999947		≥ 0.995
80.0	80.3	0.9963	Slope	1.003226		0.90 - 1.10
40.1	40.1	0.9988	Intercept	-0.059096		+/-3
20.0	19.7	1.0165				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Stony Mountain
Calibration Date: July 6, 2025
Start time (MST): 10:55
Reason: Cylinder Change

Station number: AMS 18
Last Cal Date: June 16, 2025
End time (MST): 12:20

Calibration Standards

Gas Cert Reference:	XC026809B	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T750	Serial Number:	281
Zero Air Gen model:	Teledyne API T751	Serial Number:	529

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1170050130
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.68E-04	2.68E-04	NMHC SP Ratio:	4.79E-05
CH ₄ Retention time:	14.8	14.8	NMHC Peak Area:	186547
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.02	----
As found High point	4921	78.1	16.82	16.71	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.69	Prev response	16.78	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.02	----
As left span	4921	78.1	16.82	17.12	0.983
Average Correction Factor					

Notes:

Changed the H2 cylinder after as founds. No adjustments made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.1	8.93	8.85	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.85	Prev response	8.91	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	8.93	9.02	0.991
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.02	----
As found High point	4921	78.1	7.89	7.86	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.84	Prev response	7.87	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.02	----
As left span	4921	78.1	7.89	8.10	0.974
Average Correction Factor					

Calibration Statistics

THC Cal Slope:	0.999653
THC Cal Offset:	-0.029223
CH ₄ Cal Slope:	0.999071
CH ₄ Cal Offset:	-0.006679
NMHC Cal Slope:	1.000578
NMHC Cal Offset:	-0.023155


Calibration Performed By: Max Farrell

Finish

NMHC Calibration Plot

Date: July 6, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Stony Mountain
 Calibration Date: July 23, 2025
 Start time (MST): 11:30
 Reason: Routine

Station number: AMS 18
 Last Cal Date: June 16, 2025
 End time (MST): 15:29

Calibration Standards

Gas Cert Reference:	XC026809B	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T750	Serial Number:	281
Zero Air Gen model:	Teledyne API T751	Serial Number:	529

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1170050130
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.68E-04	2.62E-04	NMHC SP Ratio:	4.79E-05
CH ₄ Retention time:	14.8	14.8	NMHC Peak Area:	186547
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				191140

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.02	----
As found High point	4921	78.1	16.82	17.19	0.980
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.17	Prev response	16.78	*% change	2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.02	----
High point	4921	78.1	16.82	16.77	1.003
Mid point	4960	39.1	8.42	8.42	1.000
Low point	4980	20.0	4.31	4.18	1.031
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	16.82	16.85	0.999
Average Correction Factor					1.011

Notes:

Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	78.1	8.93	9.13	0.978
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.13	Prev response	8.91	*% change	2.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.1	8.93	8.90	1.003
Mid point	4960	39.1	4.47	4.47	1.001
Low point	4980	20.0	2.29	2.22	1.032
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	8.93	8.94	0.999
Average Correction Factor					1.012

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.02	Limit = 0.90-1.10
As found High point	4921	78.1	7.89	8.06	0.982
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.03	Prev response	7.87	*% change	2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.02	----
High point	4921	78.1	7.89	7.87	1.002
Mid point	4960	39.1	3.95	3.95	0.999
Low point	4980	20.0	2.02	1.96	1.029
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	7.89	7.90	0.998
Average Correction Factor					1.010

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.999653	0.998768
THC Cal Offset:	-0.029223	-0.029684
CH ₄ Cal Slope:	0.999071	0.998110
CH ₄ Cal Offset:	-0.006679	-0.006598
NMHC Cal Slope:	1.000578	0.999350
NMHC Cal Offset:	-0.023155	-0.023086

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	June 16, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:30	End Time (MST):	15:29
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.02	----	Correlation Coefficient	0.999918	≥ 0.995
16.82	16.77	1.0028	Slope	0.998768	$0.90 - 1.10$
8.42	8.42	1.0000	Intercept	-0.029684	± 0.5
4.31	4.18	1.0305			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	June 16, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:30	End Time (MST):	15:29
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.02	----	Correlation Coefficient	0.999904	<i>≥0.995</i>
7.89	7.87	1.0024	Slope	0.998110	<i>0.90 - 1.10</i>
3.95	3.95	0.9987	Intercept	-0.006598	<i>+/-0.5</i>
2.02	1.96	1.0294			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	June 16, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:30	End Time (MST):	15:29
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999930	<i>≥0.995</i>
8.93	8.90	1.0031	Slope	0.999350	<i>0.90 - 1.10</i>
4.47	4.47	1.0011	Intercept	-0.023086	<i>+/-0.5</i>
2.29	2.22	1.0315			



NMHC Calibration Plot

Date: July 23, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Stony Mountain
Station number: AMS 18
Calibration Date: July 29, 2025
Last Cal Date: June 26, 2025
Start time (MST): 11:29
End time (MST): 16:27
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0045516
NOX Cal Gas Conc: 60.30 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 60.30 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API 701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 60.10 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 60.10 ppm
NO gas Diff:
Serial Number: 2658
Serial Number: 4890

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	-0.1	----	----
AF High point	4933	66.6	803.3	800.6	2.7	801.5	798.9	2.5	1.0021	1.0020
AF Mid point										
AF Low point										
New cyl resp										
Previous Respo 4933	NO _x = 800.6 ppb	NO = 798.8 ppb	* = > +/-5% change initiates investigation			*Percent Change		NO _x = 0.1%		
Baseline Corr 1st pt	NO _x = 801.6 ppb	NO = 799.0 ppb	<u>As Found Statistics</u>			*Percent Change		NO = 0.0%		
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:		Nx Int:		
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:		NO Int:		
			As found NO ₂ r ² :			NO2 SI:		NO ₂ Int:		

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo Scientific 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1501663731

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.054	1.055
NOX coeff or slope:	0.996	0.999
NO2 coeff or slope:	1.000	1.000

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	10.4	10.4
NOX bkgnd or offset:	10.5	10.5
Reaction cell Press:	221.6	223.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997044	1.002081
NO _x Cal Offset:	-0.248284	-0.867552
NO Cal Slope:	0.999723	1.002178
NO Cal Offset:	-1.549402	-2.109018
NO ₂ Cal Slope:	0.995768	1.003367
NO ₂ Cal Offset:	-1.334028	0.362949

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
High point	4933	66.6	803.3	800.6	2.7	804.5	801.3	3.2	0.9985	0.9991
Mid point	4967	33.3	401.6	400.2	1.3	401.1	397.8	3.3	1.0012	1.0061
Low point	4983	16.6	200.2	199.5	0.7	198.9	196.0	2.9	1.0066	1.0181
As left zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.1	----	----
As left span	4933	66.6	803.3	383.2	420.1	806.8	383.2	423.5	0.9956	1.0000
Average Correction Factor									1.0021	1.0078

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	801.9	383.8	420.8	422.4	0.9961	100.4%
Mid GPT point	801.9	599.3	205.3	206.5	0.9940	100.6%
Low GPT point	801.9	701.3	103.3	104.2	0.9910	100.9%
Average Correction Factor					0.9937	100.6%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 29, 2025	Previous Calibration:	June 26, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:29	End Time (MST):	16:27
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999995	<i>≥0.995</i>
803.3	804.5	0.9985	Slope	1.002081	<i>0.90 - 1.10</i>
401.6	401.1	1.0012	Intercept	-0.867552	<i>+/-20</i>
200.2	198.9	1.0066			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 29, 2025	Previous Calibration:	June 26, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:29	End Time (MST):	16:27
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
420.8	422.4	0.9961	Slope	1.003367	<i>0.90 - 1.10</i>
205.3	206.5	0.9940	Intercept	0.362949	<i>+/-20</i>
103.3	104.2	0.9910			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 29, 2025	Previous Calibration:	June 26, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:29	End Time (MST):	16:27
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1501663731

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	-0.1	----	Correlation Coefficient	0.999971	≥ 0.995
800.6	801.3	0.9991	Slope	1.002178	0.90 - 1.10
400.2	397.8	1.0061	Intercept	-2.109018	+/-20
199.5	196.0	1.0181			



NO_x Calibration Plot

Date: July 29, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: July 16, 2025 Last Cal Date: June 11, 2025
Start time (MST): 11:06 End time (MST): 14:50
Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
ZAG Make/Model: Teledyne API 701H Serial Number: 355

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 825
Analyzer Range: 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999946	1.006771	Backgd or Offset:	3.8	2.3
Calibration intercept:	-0.898000	0.040000	Coeff or Slope:	1.004	1.004

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	NA	0.0	-1.8	----
As found High point	4888	1138.1	400.0	402.7	0.989
As found Mid point					
As found Low point					
Baseline Corr As found:	404.5	Previous response	399.1	*% change	1.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	NA	0.0	-0.1	----
High point	4888	1138.1	400.0	402.7	0.993
Mid point	4888	884.5	200.0	201.4	0.993
Low point	4888	741.4	100.0	100.9	0.991
As left zero	5000	NA	0.0	-0.1	----
As left span	4812	1097.9	400.0	404.2	0.990
Average Correction Factor					0.992

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	July 16, 2025	Previous Calibration:	June 11, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:06	End Time (MST):	14:50
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999999	≥0.995
400.0	402.7	0.9933	Slope	1.006771	0.90 - 1.10
200.0	201.4	0.9930	Intercept	0.040000	+/- 5
100.0	100.9	0.9911			



O₃ Calibration Plot

Date: July 16, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: July 29, 2025 Last Cal Date: June 27, 2025
Start time (MST): 14:47 End time (MST): 15:57

Analyzer Make: API T640 S/N: 324
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	23.2	24.4	23.2	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	708.3	705.40	708.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	5.1	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	33	----	33	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 6.5		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: June 27, 2025
Date Disposable Filter Changed: June 27, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 4, 2024
Date RH/T Sensor Cleaned: July 4, 2024

Notes: Flow, temp and pressure checked. Leak check passed.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Station Information

Station Name:	Stony Mountain	Station Number:	AMS 18
Calibration Date:	July 30, 2025	Prev Cal Date:	February 21, 2024
Start Time (MST):	13:00	End Time (MST):	N/A
Tower Height (m):	10.0	Reason:	Removal

Wind Speed Calibration

Sensor make/model:	Met One 010C-1	Serial Number:	D16124
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM (Hz)	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.4	-0.3%
800	77.8	77.6	-0.2%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)		0.999998	≥ 0.9995
Calculated slope		1.002044	$0.98 - 1.02$
Calculated intercept		-0.014312	± 2

Wind Direction Calibration

Sensor make/model:	Met One 020C-1	Serial Number:	D14060
As Found Declination (deg east of True North):		As Left Declination (deg east of True North):	
Solar noon (MST):	13:30	Calc Declination*:	13 Degrees
WD Calibrator:	Met One 040		<small>* - calculated declination as per NOAA website</small>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 360° FS) <i>Limit = +/- 1%</i>
10	9.6	-0.1%
90	89.3	-0.2%
180	179.8	-0.1%
270	269.7	-0.1%
350	348.6	-0.4%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)		0.999995	≥ 0.9995
Calculated slope		1.001805	$0.97 - 1.03$
Calculated intercept		0.276251	± 5

Notes: Removal calibration.

Calibration Performed By: Aswin Sasi Kumar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS19
FIREBAG
JULY 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Firebag Station number: AMS 19
Calibration Date: July 18, 2025 Last Cal Date: June 24, 2025
Start time (MST): 10:14 End time (MST): 13:37
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.97 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC705799
Removed Cal Gas Conc: 50.97 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 1607
Zero Air Gen Model: Teledyne API T701H Serial Number: 201

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1410661308
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000017	0.997271	Backgd or Offset:	11.5	11.5
Calibration intercept:	1.040000	0.300000	Coeff or Slope:	1.021	1.021

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	4999	0.0	0.0	-0.2	----
As found High point	4922	78.4	799.2	794.0	1.006
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	794.2	Previous response	800.3	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	4999	0.0	0.0	0.0	----
High point	4922	78.4	799.2	797.3	1.002
Mid point	4961	39.2	399.6	398.6	1.003
Low point	4980	19.6	199.8	200.1	0.999
As left zero	4999	0.0	0.0	-0.2	----
As left span	4922	78.4	799.2	796.8	1.003
Average Correction Factor:					1.001

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 18, 2025	Previous Calibration:	June 24, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:14	End Time (MST):	13:37
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.0	----	Correlation Coefficient	0.999999	≥0.995
799.2	797.3	1.0024	Slope	0.997271	0.90 - 1.10
399.6	398.6	1.0025	Intercept	0.300000	+/-30
199.8	200.1	0.9985			



SO2 Calibration Plot

Date: July 18, 2025

Location: Firebag





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Firebag Station number: AMS 19
Calibration Date: July 14, 2025 Last Cal Date: June 18, 2025
Start time (MST): 9:48 End time (MST): 15:32
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.29 ppm Cal Gas Exp Date: March 19, 2027
Cal Gas Cylinder #: DT0010492
Removed Cal Gas Conc: 5.29 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 1607
ZAG Make/Model: Teledyne API T701 Serial Number: 201

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680032
Converter make: Global Converter serial #: 2022-222
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002619	0.996904	Backgd or Offset:	2.86
Calibration intercept:	-0.060000	-0.060000	Coeff or Slope:	1.198

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4924	75.6	80.0	81.7	0.979
As found Mid point	4962	37.8	40.0	40.4	0.990
As found Low point	4981	18.9	20.0	20.2	0.990
New cylinder response					
Baseline Corr As found:	81.7	Prev response:	80.13	*% change:	1.9%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.021766	AF Intercept:	-0.180000
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999962	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4924	75.6	80.0	79.7	1.004
Mid point	4962	37.8	40.0	39.7	1.007
Low point	4981	18.9	20.0	20.0	1.000
As left zero	5000	0.0	0.0	0.1	----
As left span	4924	75.6	80.0	78.5	1.019
SO2 Scrubber Check	4922	78.4	784.0	-0.1	----
Date of last scrubber change:		18-Jan-23		Ave Corr Factor	1.004
Date of last converter efficiency test:		November 26, 2024		106.2%	efficiency

Notes: Changed sample inlet filter after as founds. No adjustments made. SOx scrubber check done after cal zero.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

H2S Calibration Summary


Station Information

Calibration Date:	July 14, 2025	Previous Calibration:	June 18, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:48	End Time (MST):	15:32
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999991		≥ 0.995
80.0	79.7	1.0036	Slope	0.996904		$0.90 - 1.10$
40.0	39.7	1.0074	Intercept	-0.060000		± 3
20.0	20.0	0.9998				


H₂S Calibration Curve



H2S Calibration Plot

Date: July 14, 2025

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Firebag Station number: AMS 19
Calibration Date: July 18, 2025 Last Cal Date: June 24, 2025
Start time (MST): 10:14 End time (MST): 13:37
Reason: Routine

Calibration Standards

Gas Cert Reference: CC705799 Cal Gas Expiry Date: October 9, 2032
CH4 Cal Gas Conc. 505.1 ppm CH4 Equiv Conc. 1066.9 ppm
C3H8 Cal Gas Conc. 204.3 ppm
Removed Gas Cert: Removed Gas Expiry:
Removed CH4 Conc. 505.1 ppm CH4 Equiv Conc. 1066.9 ppm
Removed C3H8 Conc. 204.3 ppm Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 1607
ZAG Make/Model: Teledyne API T701H Serial Number: 201

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1336160089
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994363	1.006674	Background:	2.16	2.30
Calibration intercept:	0.076067	-0.048535	Coefficient:	3.813	3.841

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	4999	0.0	0.00	0.11	----
As found High point	4922	78.4	16.73	16.79	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	16.68	Previous response	16.71	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	4999	0.0	0.00	-0.03	----
High point	4922	78.4	16.73	16.82	0.995
Mid point	4961	39.2	8.36	8.31	1.007
Low point	4980	19.6	4.18	4.18	1.002
As left zero	4999	0.0	0.00	0.02	----
As left span	4922	78.4	16.73	16.77	0.998
Average Correction Factor					1.001

Notes: Changed sample inlet filter after as founds. Adjusted zero.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	July 18, 2025	Previous Calibration:	June 24, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:14	End Time (MST):	13:37
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.03	----	Correlation Coefficient	0.999966	≥ 0.995
16.73	16.82	0.9946	Slope	1.006674	$0.90 - 1.10$
8.36	8.31	1.0066	Intercept	-0.048535	± 1.5
4.18	4.18	1.0016			


THC Calibration Curve



THC Calibration Plot

Date: July 18, 2025

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Firebag
Station number: AMS 19
Calibration Date: July 31, 2025
Last Cal Date: June 17, 2025
Start time (MST): 9:19
End time (MST): 13:52
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0044018
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.70 ppm
NO gas Diff:
Serial Number: 1607
Serial Number: 201

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
AF High point	4918	82.1	802.9	799.7	3.3	802.0	799.4	2.4	1.0013	1.0003
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 803.2 ppb	NO = 801.5 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = -0.2%			
Baseline Corr 1st pt	NO _x = 801.9 ppb	NO = 799.4 ppb	<u>As Found Statistics</u>			*Percent Change	NO = -0.3%			
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:	NO Int:			
			As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1410661309

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998316	1.000279
NO _x Cal Offset:	1.620007	1.080384
NO Cal Slope:	1.001516	1.001287
NO Cal Offset:	0.660105	0.639993
NO ₂ Cal Slope:	0.994398	1.000414
NO ₂ Cal Offset:	-0.361272	-0.537216

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.938	0.938	NO bkgnd or offset:	4.7	4.7
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	4.7	4.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	164.2	166.7

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
High point	4918	82.1	802.9	799.7	3.3	804.0	801.0	2.1	0.9987	0.9983
Mid point	4959	41.1	402.0	400.3	1.6	403.0	401.8	1.2	0.9974	0.9963
Low point	4980	20.5	200.5	199.7	0.8	202.9	201.2	1.8	0.9881	0.9924
As left zero	5000	0	0.0	0.0	0.0	0.2	-0.1	0.3	----	----
As left span	4918	82.1	802.9	392.7	410.2	803.0	392.7	410.2	0.9999	1.0000
Average Correction Factor									0.9947	0.9957

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	799.7	391.6	411.4	411.4	1.0000	100.0%
Mid GPT point	799.7	598.1	204.9	204.0	1.0043	99.6%
Low GPT point	799.7	702.7	100.3	99.1	1.0119	98.8%
Average Correction Factor					1.0054	99.5%

Notes:

Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 31, 2025	Previous Calibration:	June 17, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:19	End Time (MST):	13:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999993	≥0.995
802.9	804.0	0.9987	Slope	1.000279	0.90 - 1.10
402.0	403.0	0.9974	Intercept	1.080384	+/-20
200.5	202.9	0.9881			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 31, 2025	Previous Calibration:	June 17, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:19	End Time (MST):	13:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999985	≥0.995
411.4	411.4	1.0000	Slope	1.000414	0.90 - 1.10
204.9	204.0	1.0043	Intercept	-0.537216	+/-20
100.3	99.1	1.0119			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 31, 2025	Previous Calibration:	June 17, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:19	End Time (MST):	13:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999997	≥ 0.995
799.7	801.0	0.9983	Slope	1.001287	$0.90 - 1.10$
400.3	401.8	0.9963	Intercept	0.639993	± 20
199.7	201.2	0.9924			



NO_x Calibration Plot

Date: July 31, 2025

Location: Firebag





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS20 MACKAY RIVER JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: MacKay River Station number: AMS 20
Calibration Date: July 22, 2025 Last Cal Date: June 12, 2025
Start time (MST): 8:01 End time (MST): 10:31
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.15 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC409669
Removed Cal Gas Conc: 49.15 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 5706
Zero Air Gen Model: API 701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1501301450
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998504	1.002217	Backgd or Offset:	20.5	19.4
Calibration intercept:	-0.936338	0.964310	Coeff or Slope:	0.931	0.931

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-1.0	----
As found High point	4919	81.4	800.1	803.7	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	804.7	Previous response	798.0	*% change	0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.7	----
High point	4919	81.4	800.1	802.7	0.997
Mid point	4959	40.7	400.1	402.1	0.995
Low point	4980	20.3	199.5	201.2	0.992
As left zero	5000	0.0	0.0	0.7	----
As left span	4919	81.4	800.1	806.2	0.992
Average Correction Factor:					0.995

Notes: No Maintenance done. Zero adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 22, 2025	Previous Calibration:	June 12, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	8:01	End Time (MST):	10:31
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.7	----	Correlation Coefficient	1.000000	≥0.995
800.1	802.7	0.9968	Slope	1.002217	0.90 - 1.10
400.1	402.1	0.9950	Intercept	0.964310	+/-30
199.5	201.2	0.9917			



SO2 Calibration Plot

Date: July 22, 2025

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: MacKay River
Calibration Date: July 11, 2025
Start time (MST): 5:26
Reason: Routine

Station number: AMS 20
Last Cal Date: June 5, 2025
End time (MST): 9:07

Calibration Standards

Cal Gas Concentration: 5.12 ppm
Cal Gas Cylinder #: CC515997
Removed Cal Gas Conc: 5.12 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API 701

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 5706
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1236656117
Converter serial #: 2022-226
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001741	0.997310	Backgd or Offset:	3.77	3.77
Calibration intercept:	0.019427	-0.100631	Coeff or Slope:	1.071	1.071

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4922	78.1	80.0	80.4	0.992
As found Mid point	4961	39.0	39.9	40.1	0.991
As found Low point	4980	19.5	20.0	19.9	0.994
New cylinder response					
Baseline Corr As found:	80.6	Prev response:	80.13	*% change:	0.6%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.008029	AF Intercept:	-0.200468
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999999	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4922	78.1	80.0	79.6	1.005
Mid point	4961	39.0	39.9	39.9	1.001
Low point	4980	19.5	20.0	19.7	1.014
As left zero	5000	0.0	0.0	0.1	----
As left span	4922	78.1	80.0	78.5	1.019
SO2 Scrubber Check	4982	81.3	802.8	0.0	----
Date of last scrubber change:	25-May-23		Ave Corr Factor		1.006
Date of last converter efficiency test:					

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	July 11, 2025	Previous Calibration:	June 5, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	5:26	End Time (MST):	9:07
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999987	≥ 0.995
80.0	79.6	1.0047	Slope	0.997310	$0.90 - 1.10$
39.9	39.9	1.0009	Intercept	-0.100631	± 3
20.0	19.7	1.0137			



H₂S Calibration Plot

Date: July 11, 2025

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: MacKay River
Calibration Date: July 22, 2025
Start time (MST): 8:01
Reason: Routine

Station number: AMS 20
Last Cal Date: June 6, 2025
End time (MST): 10:30

Calibration Standards

Gas Cert Reference: CC409669
CH4 Cal Gas Conc. 505.1 ppm
C3H8 Cal Gas Conc. 206.4 ppm
Removed Gas Cert:
Removed CH4 Conc. 505.1 ppm
Removed C3H8 Conc. 206.4 ppm
Calibrator Make/Model: API T700
ZAG Make/Model: API 701

Cal Gas Expiry Date: October 9, 2032
CH4 Equiv Conc. 1072.7 ppm
Removed Gas Expiry:
CH4 Equiv Conc. 1072.7 ppm
Diff between cyl:
Serial Number: 5706
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 51i-LT
Analyzer Range: 0 - 20 ppm

Analyzer serial #: 1501663727

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000215	0.994954	Background:	2.980	2.980
Calibration intercept:	-0.073986	0.027194	Coefficient:	5.009	5.009

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic-AFzero))
					<i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.11	----
As found High point	4919	81.4	17.46	17.50	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.40	Previous response	17.39	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)
					<i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.03	----
High point	4919	81.4	17.46	17.41	1.003
Mid point	4959	40.7	8.73	8.70	1.003
Low point	4980	20.3	4.35	4.37	0.997
As left zero	5000	0.0	0.00	0.02	----
As left span	4919	81.4	17.46	17.51	0.997
Average Correction Factor					1.001

Notes:

Hydrogen cylinder changed. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	July 22, 2025	Previous Calibration:	June 6, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	8:01	End Time (MST):	10:30
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.03	----	Correlation Coefficient	0.999999	≥ 0.995
17.46	17.41	1.0032	Slope	0.994954	$0.90 - 1.10$
8.73	8.70	1.0034	Intercept	0.027194	± 1.5
4.35	4.37	0.9975			


THC Calibration Curve



THC Calibration Plot

Date: July 22, 2025

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: MacKay River
Station number: AMS 20
Calibration Date: July 21, 2025
Last Cal Date: June 2, 2025
Start time (MST): 7:22
End time (MST): 11:26
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0037393
NOX Cal Gas Conc: 62.00 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 62.00 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: July 22, 2032
NO Cal Gas Conc: 61.90 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 61.90 ppm
NO gas Diff:
Serial Number: 5706
Serial Number: 4888

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
AF High point	4935	64.6	801.1	799.8	1.3	804.7	803.6	1.1	0.9955	0.9954
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 805.3 ppb	NO = 804.6 ppb								NO _x = -0.1%
Baseline Corr 1st pt	NO _x = 804.7 ppb	NO = 803.5 ppb								NO = -0.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb								
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb								
As Found Statistics										
						As found NO _x r ² :				
						As found NO r ² :				
						As found NO ₂ r ² :				

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1505164379

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.016	1.016	NO bkgnd or offset:	2.8	2.8
NOX coeff or slope:	1.000	1.000	NOX bkgnd or offset:	3.0	3.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	161.8	161.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003936	1.001053
NO _x Cal Offset:	1.051925	0.812699
NO Cal Slope:	1.007256	1.005426
NO Cal Offset:	-0.967598	-0.827015
NO ₂ Cal Slope:	0.994795	0.997263
NO ₂ Cal Offset:	-1.179914	-0.843969

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.5	0.5	0.0	----	----
High point	4935	64.6	801.1	799.8	1.3	802.4	803.8	-1.4	0.9984	0.9950
Mid point	4968	32.3	400.5	399.9	0.6	402.5	401.1	1.5	0.9950	0.9969
Low point	4984	16.2	200.9	200.5	0.3	201.8	199.1	2.7	0.9954	1.0073
As left zero	5000	0.0	0.0	0.0	0.0	0.6	0.6	0.0	----	----
As left span	4935	64.6	801.1	406.4	394.7	803.1	406.4	396.7	0.9975	1.0000
Average Correction Factor									0.9963	0.9997

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	800.7	408.3	393.7	392.2	1.0038	99.6%
Mid GPT point	800.7	603.2	198.8	197.0	1.0091	99.1%
Low GPT point	800.7	700.0	102.0	100.0	1.0199	98.0%
Average Correction Factor					1.0109	98.9%

Notes: No adjustments or maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 21, 2025	Previous Calibration:	June 2, 2025
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	7:22	End Time (MST):	11:26
Analyzer make:	Thermo 42i		1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999998	≥0.995
801.1	802.4	0.9984	Slope	1.001053	0.90 - 1.10
400.5	402.5	0.9950	Intercept	0.812699	+/-20
200.9	201.8	0.9954			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 21, 2025	Previous Calibration:	June 2, 2025
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	7:22	End Time (MST):	11:26
Analyzer make:	Thermo 42i		1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999978	≥0.995
393.7	392.2	1.0038	Slope	0.997263	0.90 - 1.10
198.8	197.0	1.0091	Intercept	-0.843969	+/-20
102.0	100.0	1.0199			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 21, 2025	Previous Calibration:	June 2, 2025
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	7:22	End Time (MST):	11:26
Analyzer make:	Thermo 42i	6:50:00 AM	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.5	----	Correlation Coefficient	0.999986	≥ 0.995
799.8	803.8	0.9950	Slope	1.005426	0.90 - 1.10
399.9	401.1	0.9969	Intercept	-0.827015	+/-20
200.5	199.1	1.0073			



NO_x Calibration Plot

Date: July 21, 2025

Location: MacKay River





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS21
JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Conklin
Calibration Date: July 14, 2025
Start time (MST): 10:30
Reason: Routine

Station number: AMS 21
Last Cal Date: June 9, 2025
End time (MST): 13:27

Calibration Standards

Cal Gas Concentration: 50.34 ppm
Cal Gas Cylinder #: CC340840
Removed Cal Gas Conc: 50.34 ppm
Removed Gas Cyl #: NA
Calibrator Model: Teledyne API T700P
Zero Air Gen Model: Teledyne API T701

Cal Gas Exp Date: October 9, 2032
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 2659
Serial Number: 953

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 1428701363

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003885	1.008525	Backgd or Offset:	30.4	29.3
Calibration intercept:	-1.862570	-1.763281	Coeff or Slope:	0.891	0.891

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.7	----
As found High point	4921	79.5	800.3	805.0	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	805.7	Previous response	801.6	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4921	79.5	800.3	806.1	0.993
Mid point	4960	39.8	400.7	402.1	0.997
Low point	4980	19.9	200.4	198.0	1.012
As left zero	5000	0.0	0.0	0.0	----
As left span	4921	79.5	800.3	804.0	0.995
Average Correction Factor:					1.000

Notes: Sample inlet filter was changed after as founds. Adjusted zero only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 14, 2025	Previous Calibration:	June 9, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:30	End Time (MST):	13:27
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999974	≥0.995
800.3	806.1	0.9928	Slope	1.008525	0.90 - 1.10
400.7	402.1	0.9966	Intercept	-1.763281	+/-30
200.4	198.0	1.0119			



SO2 Calibration Plot

Date: July 14, 2025

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Conklin
Calibration Date: July 9, 2025
Start time (MST): 9:46
Reason: Routine

Station number: AMS 21
Last Cal Date: June 25, 2025
End time (MST): 13:46

Calibration Standards

Cal Gas Concentration: 5.14 ppm
Cal Gas Cylinder #: CC501204
Removed Cal Gas Conc: 5.14 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne T700P
ZAG Make/Model: Teledyne T701

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 2659
Serial Number: 953

Analyzer Information

Analyzer make: Thermo 43i-QTL
Converter make: CD-Nova 101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12228021058
Converter serial #: 565
Converter Temp: 850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010574	1.002775	Backgd or Offset:	3.3	3.2
Calibration intercept:	-0.081632	-0.181617	Coeff or Slope:	1.539	1.510

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4922	78.4	80.6	80.9	0.992
As found Mid point	4961	39.2	40.3	40.8	0.980
As found Low point	4980	19.6	20.2	20.2	0.983
New cylinder response					
Baseline Corr As found:	81.2	Prev response:	81.36	*% change:	-0.2%
Baseline Corr 2nd AF pt:	41.1	AF Slope:	1.007455	AF Intercept:	-0.121630
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999954	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4922	78.4	80.6	80.6	1.000
Mid point	4961	39.2	40.3	40.3	1.000
Low point	4980	19.6	20.2	20.0	1.008
As left zero	5000	0.0	0.0	-0.2	----
As left span	4922	78.4	80.6	80.6	1.000
SO2 Scrubber Check	4921	79.5	794.9	0.0	----
Date of last scrubber change:	November 13, 2024			Ave Corr Factor	1.002
Date of last converter efficiency test:					

Notes: Sample inlet filter was changed after multipoint as founds. SO2 scrubber check done after calibrator zero and passed. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


TRS Calibration Summary


Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	June 25, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:46	End Time (MST):	13:46
Analyzer make:	Thermo 43i-QTL	Analyzer serial #:	12228021058

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999998		≥ 0.995
80.6	80.6	0.9999	Slope	1.002775		$0.90 - 1.10$
40.3	40.3	0.9999	Intercept	-0.181617		± 3
20.2	20.0	1.0075				







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Conklin
 Calibration Date: July 14, 2025
 Start time (MST): 10:30
 Reason: Routine

Station number: AMS 21
 Last Cal Date: June 9, 2025
 End time (MST): 13:27

Calibration Standards

Gas Cert Reference:	CC340840	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	503.8 ppm	CH ₄ Equiv Conc.	1067.6 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	503.8 ppm	CH ₄ Equiv Conc.	1067.6 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700P	Serial Number:	2659
Zero Air Gen model:	Teledyne API T701	Serial Number:	953

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1180320039
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.32E-04	2.28E-04	NMHC SP Ratio:	4.94E-05
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	181376
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.5	16.97	17.27	0.983
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.27	Prev response	16.97	*% change	1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.5	16.97	16.98	0.999
Mid point	4960	39.8	8.50	8.46	1.004
Low point	4980	19.9	4.25	4.17	1.019
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	16.97	16.99	0.999
Average Correction Factor					1.008

Notes:

Sample inlet filter was changed after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.5	8.96	9.14	0.981
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.14	Prev response	8.97	*% change	1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.5	8.96	8.96	1.000
Mid point	4960	39.8	4.49	4.48	1.002
Low point	4980	19.9	2.24	2.22	1.013
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	8.96	8.97	0.999
Average Correction Factor					1.005

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.5	8.01	8.14	0.984
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.14	Prev response	7.99	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.5	8.01	8.02	0.999
Mid point	4960	39.8	4.01	3.98	1.007
Low point	4980	19.9	2.01	1.95	1.027
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	8.01	8.02	0.999
Average Correction Factor					1.011

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.001948	1.001901
THC Cal Offset:	-0.040304	-0.041701
CH ₄ Cal Slope:	1.001787	1.002943
CH ₄ Cal Offset:	-0.029812	-0.028615
NMHC Cal Slope:	1.001914	1.000995
NMHC Cal Offset:	-0.009294	-0.012686

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 14, 2025	Previous Calibration:	June 9, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:30	End Time (MST):	13:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999972	≥ 0.995
16.97	16.98	0.9995	Slope	1.001901	$0.90 - 1.10$
8.50	8.46	1.0045	Intercept	-0.041701	± 0.5
4.25	4.17	1.0194			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 14, 2025	Previous Calibration:	June 9, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:30	End Time (MST):	13:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999942	<i>≥0.995</i>
8.01	8.02	0.9990	Slope	1.002943	<i>0.90 - 1.10</i>
4.01	3.98	1.0071	Intercept	-0.028615	<i>+/-0.5</i>
2.01	1.95	1.0272			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 14, 2025	Previous Calibration:	June 9, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:30	End Time (MST):	13:27
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999989	<i>≥0.995</i>
8.96	8.96	1.0000	Slope	1.000995	<i>0.90 - 1.10</i>
4.49	4.48	1.0017	Intercept	-0.012686	<i>+/-0.5</i>
2.24	2.22	1.0125			



NMHC Calibration Plot

Date: July 14, 2025

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Conklin
Calibration Date: July 30, 2025
Start time (MST): 11:07
Reason: Maintenance

Station number: AMS 21
Last Cal Date: July 14, 2025
End time (MST): 13:30

Calibration Standards

Gas Cert Reference:	CC340840	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	503.8 ppm	CH ₄ Equiv Conc.	1067.6 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	503.8 ppm	CH ₄ Equiv Conc.	1067.6 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700P	Serial Number:	2659
Zero Air Gen model:	Teledyne API T701	Serial Number:	953

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1180320039
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.28E-04	2.28E-04	NMHC SP Ratio:	4.84E-05
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	185126
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.20	----
As found High point	4921	79.5	16.97	5.82	3.017
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	5.63	Prev response	16.96	*% change	-201.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.5	16.97	16.95	1.001
Mid point	4960	39.8	8.50	8.45	1.006
Low point	4980	19.9	4.25	4.16	1.022
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	16.97	16.95	1.001
Average Correction Factor					1.010

Notes:

As founds out of limits due to low N₂ cylinder pressure. N₂ cylinder was changed after as founds. No adjustment made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.5	8.96	5.75	1.560
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	5.75	Prev response	8.96	*% change	-55.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.5	8.96	8.94	1.002
Mid point	4960	39.8	4.49	4.47	1.003
Low point	4980	19.9	2.24	2.21	1.016
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	8.96	8.93	1.004
Average Correction Factor					1.007

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.20	Limit = 0.90-1.10
As found High point	4921	79.5	8.01	0.08	-66.195
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	-0.12	Prev response	8.00	*% change	6715.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.5	8.01	8.01	1.001
Mid point	4960	39.8	4.01	3.98	1.009
Low point	4980	19.9	2.01	1.95	1.029
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	8.01	8.02	0.999
Average Correction Factor					1.013

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.001901	1.000231
THC Cal Offset:	-0.041701	-0.043293
CH ₄ Cal Slope:	1.002943	1.001344
CH ₄ Cal Offset:	-0.028615	-0.029011
NMHC Cal Slope:	1.000995	0.998789
NMHC Cal Offset:	-0.012686	-0.013281

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 30, 2025	Previous Calibration:	July 14, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	11:07	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999969		≥ 0.995
16.97	16.95	1.0013	Slope	1.000231		$0.90 - 1.10$
8.50	8.45	1.0062	Intercept	-0.043293		± 0.5
4.25	4.16	1.0224				





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 30, 2025	Previous Calibration:	July 14, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	11:07	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999940	<i>≥0.995</i>
8.01	8.01	1.0006	Slope	1.001344	<i>0.90 - 1.10</i>
4.01	3.98	1.0089	Intercept	-0.029011	<i>+/-0.5</i>
2.01	1.95	1.0293			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 30, 2025	Previous Calibration:	July 14, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	11:07	End Time (MST):	13:30
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999987	<i>≥0.995</i>
8.96	8.94	1.0023	Slope	0.998789	<i>0.90 - 1.10</i>
4.49	4.47	1.0035	Intercept	-0.013281	<i>+/-0.5</i>
2.24	2.21	1.0162			



NMHC Calibration Plot

Date: July 30, 2025

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Conklin
Station number: AMS 21
Calibration Date: July 17, 2025
Last Cal Date: June 4, 2025
Start time (MST): 9:41
End time (MST): 13:42
Reason: Routine

Calibration Standards

NO Gas Cylinder #: SA18828
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700P
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 2659
Serial Number: 953

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.4	0.1	----	----
AF High point	4918	82.0	802.0	800.3	1.6	848.4	842.7	5.8	0.9449	0.9493
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 803.2 ppb	NO = 799.9 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = 5.4%	
Baseline Corr 1st pt	NO _x = 848.7 ppb	NO = 843.1 ppb				<u>As Found Statistics</u>		*Percent Change	NO = 5.1%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153356

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.706	0.669	NO bkgnd or offset:	3.9	3.7
NOX coeff or slope:	0.996	0.994	NOX bkgnd or offset:	3.9	3.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	163.2	163.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003747	1.004473
NO _x Cal Offset:	-1.791998	-1.371987
NO Cal Slope:	1.003262	1.005604
NO Cal Offset:	-3.052033	-2.472009
NO ₂ Cal Slope:	1.004631	1.000514
NO ₂ Cal Offset:	0.007506	0.104705

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
High point	4918	82.0	802.0	800.3	1.6	804.4	803.2	1.3	0.9970	0.9964
Mid point	4959	41.0	401.0	400.2	0.8	402.0	399.6	2.4	0.9975	1.0014
Low point	4980	20.5	200.5	200.1	0.4	197.9	195.9	2.0	1.0130	1.0212
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
As left span	4918	82.0	802.0	400.8	401.2	802.7	400.8	401.9	0.9991	1.0000
Average Correction Factor									1.0025	1.0064

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	800.5	396.6	405.5	405.9	0.9991	100.1%
Mid GPT point	800.5	603.2	198.9	199.0	0.9997	100.0%
Low GPT point	800.5	704.2	97.9	98.2	0.9974	100.3%
Average Correction Factor					0.9987	100.1%

Notes:

Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By:

Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 17, 2025	Previous Calibration:	June 4, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:41	End Time (MST):	13:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999982	≥0.995
802.0	804.4	0.9970	Slope	1.004473	0.90 - 1.10
401.0	402.0	0.9975	Intercept	-1.371987	+/-20
200.5	197.9	1.0130			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 17, 2025	Previous Calibration:	June 4, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:41	End Time (MST):	13:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	1.000000	≥0.995
405.5	405.9	0.9991	Slope	1.000514	0.90 - 1.10
198.9	199.0	0.9997	Intercept	0.104705	+/-20
97.9	98.2	0.9974			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 17, 2025	Previous Calibration:	June 4, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:41	End Time (MST):	13:42
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999961	≥ 0.995
800.3	803.2	0.9964	Slope	1.005604	$0.90 - 1.10$
400.2	399.6	1.0014	Intercept	-2.472009	± 20
200.1	195.9	1.0212			



NO_x Calibration Plot

Date: July 17, 2025

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Conklin
Calibration Date: July 2, 2025
Start time (MST): 9:45
Reason: Routine

Station number: AMS 21
Last Cal Date: June 16, 2025
End time (MST): 12:26

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700P
ZAG Make/Model: Teledyne API T701H

Serial Number: 2659
Serial Number: 355

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1501663734

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004600	0.997571	Backgd or Offset:	0.7	0.2
Calibration intercept:	-0.980000	-0.700000	Coeff or Slope:	1.054	1.054

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	0.5	----
As found High point	5000	920.6	400.0	399.7	1.002
As found Mid point					
As found Low point					
Baseline Corr As found:	399.2	Previous response	400.9	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.5	----
High point	5000	919.0	400.0	399.0	1.003
Mid point	5000	755.3	200.0	198.0	1.010
Low point	5000	646.4	100.0	98.0	1.020
As left zero	5000	800.0	0.0	-0.1	----
As left span	5000	917.0	400.0	407.0	0.983
Average Correction Factor					1.011

Notes: Sample inlet filter was changed after as founds. Adjusted zero only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	July 2, 2025	Previous Calibration:	June 16, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:45	End Time (MST):	12:26
Analyzer make:	Thermo 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999958	≥0.995
400.0	399.0	1.0025	Slope	0.997571	0.90 - 1.10
200.0	198.0	1.0101	Intercept	-0.700000	+/- 5
100.0	98.0	1.0204			



O₃ Calibration Plot

Date: July 2, 2025

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Conklin Station number: AMS 21
Calibration Date: July 9, 2025 Last Cal Date: June 30, 2025
Start time (MST): 11:51 End time (MST): 12:13

Analyzer Make: API T640 S/N: 1266
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	22.10	21.50	22.10	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	701.30	703.64	701.30	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.95	5.14	4.95	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	38	----	38	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 2.20		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: July 16, 2026
Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: June 30, 2025
Date Disposable Filter Changed: June 30, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: June 4, 2025
Date RH/T Sensor Cleaned: June 30, 2025

Notes: Verified flow, pressure, temperature and pump power. Leak check passed. No adjustment needed.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER
JULY 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: July 25, 2025 Last Cal Date: June 26, 2025
Start time (MST): 11:39 End time (MST): 14:53
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.11 ppm Cal Gas Exp Date: January 18, 2029
Cal Gas Cylinder #: CC281519
Removed Cal Gas Conc: 50.11 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3806
Zero Air Gen Model: Teledyne API T701 Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1152430006
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002479	0.998763	Backgd or Offset:	26.0	26.1
Calibration intercept:	1.803963	0.404805	Coeff or Slope:	1.006	0.986

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.7	----
As found High point	4920	79.8	799.8	811.0	0.987
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	810.3	Previous response	803.6	*% change	0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	79.8	799.8	798.8	1.001
Mid point	4960	39.9	399.9	400.8	0.998
Low point	4980	20.0	200.4	200.2	1.001
As left zero	5000	0.0	0.0	0.5	----
As left span	4920	79.8	799.8	797.5	1.003
Average Correction Factor:					1.000

Notes: Changed the inlet filter after as founds. Adjusted the zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 25, 2025	Previous Calibration:	June 26, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:39	End Time (MST):	14:53
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999996	≥0.995
799.8	798.8	1.0012	Slope	0.998763	0.90 - 1.10
399.9	400.8	0.9977	Intercept	0.404805	+/-30
200.4	200.2	1.0012			



SO2 Calibration Plot

Date: July 25, 2025

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Janvier
Calibration Date: July 22, 2025
Start time (MST): 11:23
Reason: Routine

Station number: AMS 22
Last Cal Date: June 25, 2025
End time (MST): 15:27

Calibration Standards

Cal Gas Concentration: 5.02 ppm
Cal Gas Cylinder #: CC424047
Removed Cal Gas Conc: 5.02 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3806
Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1151680031
Converter serial #: 620
Converter Temp: 850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005235	1.003521	Backgd or Offset:	3.81	3.79
Calibration intercept:	-0.199288	-0.039299	Coeff or Slope:	1.196	1.196

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4920	79.7	80.0	82.2	0.970
As found Mid point	4960	39.8	40.0	41.2	0.963
As found Low point	4980	19.9	20.0	20.2	0.975
New cylinder response					
Baseline Corr As found:	82.5	Prev response:	80.24	*% change:	2.7%
Baseline Corr 2nd AF pt:	41.5	AF Slope:	1.031797	AF Intercept:	-0.278695
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999976	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4920	79.7	80.0	80.2	0.998
Mid point	4960	39.8	40.0	40.1	0.997
Low point	4980	19.9	20.0	20.2	0.989
As left zero	5000	0.0	0.0	-0.1	----
As left span	4920	79.7	80.0	77.7	1.030
SO2 Scrubber Check	4920	79.8	798.0	0.2	----
Date of last scrubber change:				Ave Corr Factor	0.994
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Scrubber test performed, no issues. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	July 22, 2025	Previous Calibration:	June 25, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:23	End Time (MST):	15:27
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999981	≥ 0.995
80.0	80.2	0.9978	Slope	1.003521	$0.90 - 1.10$
40.0	40.1	0.9965	Intercept	-0.039299	± 3
20.0	20.2	0.9891			

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Janvier
 Calibration Date: July 25, 2025
 Start time (MST): 11:39
 Reason: Routine

Station number: AMS 22
 Last Cal Date: June 26, 2025
 End time (MST): 14:53

Calibration Standards

Gas Cert Reference: CC281519
 CH₄ Cal Gas Conc. 502.8 ppm
 C₃H₈ Cal Gas Conc. 208.4 ppm
 Removed Gas Cert: NA
 Removed CH₄ Conc. 502.8 ppm
 Removed C₃H₈ Conc. 208.4 ppm
 Diff between cyl (CH₄):
 Calibrator Model: Teledyne API 700
 Zero Air Gen model: Teledyne API 701

Cal Gas Expiry Date: January 18, 2029
 CH₄ Equiv Conc. 1075.9 ppm
 Removed Gas Expiry: NA
 CH₄ Equiv Conc. 1075.9 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 3806
 Serial Number: 691

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1317958219
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.61E-04	2.54E-04	NMHC SP Ratio:	6.23E-05
CH ₄ Retention time:	11.8	11.6	NMHC Peak Area:	146823
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				150044

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	17.17	17.32	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.32	Prev response	17.18	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	17.17	17.18	0.999
Mid point	4960	39.9	8.59	8.47	1.013
Low point	4980	20.0	4.30	4.24	1.015
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	17.17	16.93	1.014
Average Correction Factor					1.009

Notes:

Changed the inlet filter after as founds. Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.15	9.24	0.990
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.24	Prev response	9.16	*% change	0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.15	9.15	1.000
Mid point	4960	39.9	4.57	4.53	1.011
Low point	4980	20.0	2.29	2.27	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.15	8.97	1.020
Average Correction Factor					1.007

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	8.03	8.08	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.08	Prev response	8.02	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.03	8.03	0.999
Mid point	4960	39.9	4.01	3.95	1.016
Low point	4980	20.0	2.01	1.97	1.021
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.03	7.96	1.008
Average Correction Factor					1.012

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.000802	1.001194
THC Cal Offset:	-0.006404	-0.050597
CH ₄ Cal Slope:	1.002079	1.001994
CH ₄ Cal Offset:	-0.016962	-0.031160
NMHC Cal Slope:	0.999969	1.000342
NMHC Cal Offset:	0.010157	-0.018836

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 25, 2025	Previous Calibration:	June 26, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:39	End Time (MST):	14:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999939	≥ 0.995
17.17	17.18	0.9994	Slope	1.001194	$0.90 - 1.10$
8.59	8.47	1.0133	Intercept	-0.050597	± 0.5
4.30	4.24	1.0152			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 25, 2025	Previous Calibration:	June 26, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:39	End Time (MST):	14:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999902	<i>≥0.995</i>
8.03	8.03	0.9989	Slope	1.001994	<i>0.90 - 1.10</i>
4.01	3.95	1.0163	Intercept	-0.031160	<i>+/-0.5</i>
2.01	1.97	1.0209			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 25, 2025	Previous Calibration:	June 26, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:39	End Time (MST):	14:53
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999965	<i>≥0.995</i>
9.15	9.15	0.9999	Slope	1.000342	<i>0.90 - 1.10</i>
4.57	4.53	1.0107	Intercept	-0.018836	<i>+/-0.5</i>
2.29	2.27	1.0099			



NMHC Calibration Plot

Date: July 25, 2025

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Janvier
Station number: AMS 22
Calibration Date: July 23, 2025
Last Cal Date: June 30, 2025
Start time (MST): 11:34
End time (MST): 16:05
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0047765
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: March 11, 2031
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3806
Serial Number: 691

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
AF High point	4918	82.0	802.0	800.3	1.6	802.2	792.6	9.7	0.9995	1.0095
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 800.7 ppb	NO = 797.2 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.2%	
Baseline Corr 1st pt	NO _x = 802.4 ppb	NO = 792.8 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -0.6%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1229254994

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.998	1.003	NO bkgnd or offset:	2.7	2.7
NOX coeff or slope:	0.998	0.997	NOX bkgnd or offset:	2.8	2.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	172.9	173.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997687	1.000893
NO _x Cal Offset:	0.604037	0.604128
NO Cal Slope:	0.996904	1.001131
NO Cal Offset:	-0.656109	-0.936016
NO ₂ Cal Slope:	1.003200	1.001655
NO ₂ Cal Offset:	1.084488	0.801280

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
High point	4918	82.0	802.0	800.3	1.6	802.9	800.7	2.2	0.9988	0.9995
Mid point	4960	41.0	400.9	400.1	0.8	402.4	399.2	3.2	0.9963	1.0022
Low point	4980	20.5	200.5	200.1	0.4	201.7	198.5	3.2	0.9939	1.0079
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
As left span	4918	82.0	802.0	394.4	407.6	803.8	394.4	409.4	0.9977	1.0000
Average Correction Factor									0.9963	1.0032

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	798.5	395.2	404.9	406.1	0.9971	100.3%
Mid GPT point	798.5	597.9	202.2	203.6	0.9933	100.7%
Low GPT point	798.5	697.6	102.5	104.3	0.9831	101.7%
Average Correction Factor					0.9912	100.9%

Notes:

Inlet filter was changed after as founds. Adjusted the span.

Calibration Performed By:

Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	June 30, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:34	End Time (MST):	16:05
Analyzer make:	Thermo 42i	Analyzer serial #:	1229254994

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999997	≥0.995
802.0	802.9	0.9988	Slope	1.000893	0.90 - 1.10
400.9	402.4	0.9963	Intercept	0.604128	+/-20
200.5	201.7	0.9939			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	June 30, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:34	End Time (MST):	16:05
Analyzer make:	Thermo 42i	Analyzer serial #:	1229254994

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999986	≥0.995
404.9	406.1	0.9971	Slope	1.001655	0.90 - 1.10
202.2	203.6	0.9933	Intercept	0.801280	+/-20
102.5	104.3	0.9831			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	June 30, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:34	End Time (MST):	16:05
Analyzer make:	Thermo 42i	Analyzer serial #:	1229254994

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999995	≥ 0.995
800.3	800.7	0.9995	Slope	1.001131	$0.90 - 1.10$
400.1	399.2	1.0022	Intercept	-0.936016	± 20
200.1	198.5	1.0079			



NO_x Calibration Plot

Date: July 23, 2025

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Janvier
Calibration Date: July 9, 2025
Start time (MST): 11:52
Reason: Routine

Station number: AMS 22
Last Cal Date: June 27, 2025
End time (MST): 15:01

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Serial Number: 3806
Serial Number: 691

Analyzer Information

Analyzer make: Teledyne API T400
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 7046

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994829	1.003457	Backgd or Offset:	1.5	1.6
Calibration intercept:	1.480000	1.520000	Coeff or Slope:	1.011	1.018

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-0.3	----
As found High point	5000	926.2	400.0	398.6	1.003
As found Mid point					
As found Low point					
Baseline Corr As found:	398.9	Previous response	399.4	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.2	----
High point	5000	926.2	400.0	402.1	0.995
Mid point	5000	768.9	200.0	203.3	0.984
Low point	5000	666.4	100.0	102.9	0.972
As left zero	5000	800.0	0.0	0.1	----
As left span	5000	926.2	400.0	404.5	0.989
Average Correction Factor					0.983

Notes: Changed the inlet filter after the as founds. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	June 27, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:52	End Time (MST):	15:01
Analyzer make:	Teledyne API T400	Analyzer serial #:	7046

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999947	≥ 0.995
400.0	402.1	0.9948	Slope	1.003457	$0.90 - 1.10$
200.0	203.3	0.9838	Intercept	1.520000	± 5
100.0	102.9	0.9718			



O₃ Calibration Plot

Date: July 9, 2025

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: July 23, 2025 Last Cal Date: June 30, 2025
Start time (MST): 12:09 End time (MST): 13:15

Analyzer Make: Teledyne API T640 S/N: 325
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	23.3	23.44	23.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	716.6	717.5	716.6	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.023	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	35	----	35	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 14.9		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: Expiry Date:
Lot No.:

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: May 28, 2025
Date Disposable Filter Changed: May 28, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: May 28, 2025
Date RH/T Sensor Cleaned: May 28, 2025

Notes:

Verified flow, temperature, and pressure. Leak check passed.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Version-10-2022

Station Information

Station Name:	Janvier	Station Number:	AMS 22
Calibration Date:	July 9, 2025	Prev Cal Date:	October 8, 2024
Start Time (MST):	12:48	End Time (MST):	13:37
Tower Height (m):	10.0	Reason:	Removal

Wind Speed Information

Sensor make/model:	Met One 010C-1	Serial Number:	D16121
WS Calibrator:	MetOne 053	Serial Number:	R10866

Shaft RPM	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.2%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	0.999999	0.999999	≥ 0.9995
Calculated slope	0.998868	0.998868	<i>0.90 - 1.10</i>
Calculated intercept	0.028930	0.029112	<i>+/- 2</i>

Wind Direction Information

Sensor make/model:	Met One 020C-1	Serial Number:	D14528
As Found Declination (deg east of True North):	<u>14</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon time (MST):	13:28	Calc Declination*:	12.75 Degrees
Deadband calc:	1.1 degrees (<i>Limit 4 deg</i>)	* - calculated declination as per NOAA website	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 357° FS) <i>Limit = +/- 1.0%</i>
0	0.3	---
90	88.4	-0.5%
180	179.1	-0.3%
270	271.0	0.3%
357	356.2	-0.2%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	0.999972	0.999972	≥ 0.9995
Calculated slope	1.000103	0.999452	<i>0.90 - 1.10</i>
Calculated intercept	0.839965	0.508607	<i>+/- 4</i>

Notes: Removing WS/WD sensors. Solar noon was verified via compass before bringing the tower down.

Calibration Performed By: Rene Chamberland



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23 FORT HILLS JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Fort Hills Station number: AMS 23
Calibration Date: July 10, 2025 Last Cal Date: June 11, 2025
Start time (MST): 7:42 End time (MST): 10:21
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.35 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC484463
Removed Cal Gas Conc: 50.35 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 1222
Zero Air Gen Model: API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1160290012
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002120	0.999462	Backgd or Offset:	19.0	19.0
Calibration intercept:	-0.139048	0.540537	Coeff or Slope:	1.092	1.092

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4921	79.4	799.5	798.4	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	798.2	Previous response	801.0	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4921	79.4	799.5	799.4	1.000
Mid point	4960	39.7	399.8	400.5	0.998
Low point	4980	19.8	199.4	199.9	0.997
As left zero	5000	0.0	0.0	0.3	----
As left span	4921	79.4	799.5	803.6	0.995
Average Correction Factor:					0.999

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 11, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:42	End Time (MST):	10:21
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995
799.5	799.4	1.0001	Slope	0.999462	0.90 - 1.10
399.8	400.5	0.9983	Intercept	0.540537	+/-30
199.4	199.9	0.9975			



SO2 Calibration Plot

Date: July 10, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort Hills Station number: AMS 23
Calibration Date: July 9, 2025 Last Cal Date: June 10, 2025
Start time (MST): 6:32 End time (MST): 10:50
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: August 28, 2027
Cal Gas Cylinder #: DT0021910
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 1222
ZAG Make/Model: API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1300156232
Converter make: CDN-101 Converter serial #: 594
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010678	1.006535	Backgd or Offset:	1.99	1.92
Calibration intercept:	-0.318167	0.001795	Coeff or Slope:	1.137	1.000

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4917	82.6	80.0	77.7	1.028
As found Mid point	4959	41.3	40.0	38.7	1.030
As found Low point	4979	20.7	20.0	19.1	1.044
New cylinder response					
Baseline Corr As found:	77.8	Prev response:	80.50	*% change:	-3.5%
Baseline Corr 2nd AF pt:	38.8	AF Slope:	0.974082	AF Intercept:	-0.237526
Baseline Corr 3rd AF pt:	19.2	AF Correlation:	0.999984	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4917	82.6	80.0	80.4	0.995
Mid point	4959	41.3	40.0	40.4	0.990
Low point	4979	20.7	20.0	20.2	0.992
As left zero	5000	0.0	0.0	0.1	----
As left span	4917	82.6	80.0	81.8	0.978
SO2 Scrubber Check	4920	80.3	803.0	0.2	----
Date of last scrubber change:				Ave Corr Factor	0.992
Date of last converter efficiency test:		March 13, 2024		110.3% efficiency	

Notes: SOx scrubber checked after the calibrator zero. Lamp voltage and initial lamp adjusted. Lamp voltage was 952 now 1007. Lamp intensity was 91% now 96%. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


TRS Calibration Summary


Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	June 10, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:32	End Time (MST):	10:50
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999987		≥ 0.995
80.0	80.4	0.9946	Slope	1.006535		$0.90 - 1.10$
40.0	40.4	0.9895	Intercept	0.001795		± 3
20.0	20.2	0.9920				







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Fort Hills
 Calibration Date: July 10, 2025
 Start time (MST): 7:42
 Reason: Routine

Station number: AMS 23
 Last Cal Date: June 11, 2025
 End time (MST): 10:21

Calibration Standards

Gas Cert Reference:	CC484463	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	504.3 ppm	CH ₄ Equiv Conc.	1065.6 ppm
C ₃ H ₈ Cal Gas Conc.	204.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.3 ppm	CH ₄ Equiv Conc.	1065.6 ppm
Removed C ₃ H ₈ Conc.	204.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	1222
Zero Air Gen model:	API T701	Serial Number:	1117

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 12227620777
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.65E-04	3.70E-04	NMHC SP Ratio:	5.51E-05
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	161613
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				157868
				OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.4	16.92	16.70	1.013
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.70	Prev response	16.83	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.4	16.92	16.79	1.008
Mid point	4960	39.7	8.46	8.34	1.014
Low point	4980	19.8	4.22	4.20	1.004
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	16.92	16.77	1.009
Average Correction Factor					1.009

Notes: No Maintenance done. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.4	8.91	8.76	1.018
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.76	Prev response	8.87	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.4	8.91	8.83	1.009
Mid point	4960	39.7	4.46	4.41	1.010
Low point	4980	19.8	2.22	2.24	0.994
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	8.91	8.82	1.010
Average Correction Factor					1.005

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.4	8.01	7.94	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.94	Prev response	7.96	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.4	8.01	7.97	1.005
Mid point	4960	39.7	4.00	3.93	1.019
Low point	4980	19.8	2.00	1.97	1.015
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	8.01	7.95	1.007
Average Correction Factor					1.013

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.995390	0.991892
THC Cal Offset:	-0.011384	-0.004995
CH ₄ Cal Slope:	0.996307	0.994765
CH ₄ Cal Offset:	-0.019598	-0.018199
NMHC Cal Slope:	0.994579	0.989412
NMHC Cal Offset:	0.008414	0.012804

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 11, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:42	End Time (MST):	10:21
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999983	≥ 0.995
16.92	16.79	1.0075	Slope	0.991892	$0.90 - 1.10$
8.46	8.34	1.0140	Intercept	-0.004995	± 0.5
4.22	4.20	1.0040			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 11, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:42	End Time (MST):	10:21
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999948	<i>≥0.995</i>
8.01	7.97	1.0054	Slope	0.994765	<i>0.90 - 1.10</i>
4.00	3.93	1.0187	Intercept	-0.018199	<i>+/-0.5</i>
2.00	1.97	1.0153			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 11, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:42	End Time (MST):	10:21
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999980	<i>≥0.995</i>
8.91	8.83	1.0093	Slope	0.989412	<i>0.90 - 1.10</i>
4.46	4.41	1.0102	Intercept	0.012804	<i>+/-0.5</i>
2.22	2.24	0.9941			



NMHC Calibration Plot

Date: July 10, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Hills
Station number: AMS 23
Calibration Date: July 7, 2025
Last Cal Date: June 9, 2025
Start time (MST): 6:54
End time (MST): 10:51
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC358149
NOX Cal Gas Conc: 60.30 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 60.30 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.10 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 60.10 ppm
NO gas Diff:
Serial Number: 1222
Serial Number: 1117

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.4	-0.1	----	----
AF High point	4934	66.3	799.5	796.9	2.7	795.9	794.7	1.2	1.0041	1.0022
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 799.4 ppb	NO = 799.4 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -0.4%	
Baseline Corr 1st pt	NO _x = 796.3 ppb	NO = 795.1 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -0.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430007

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.992	0.997	NO bkgnd or offset:	3.2	3.2
NOX coeff or slope:	0.988	0.989	NOX bkgnd or offset:	3.6	3.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	147.8	147.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000355	0.998841
NO _x Cal Offset:	-0.453668	-0.173749
NO Cal Slope:	1.006004	1.003897
NO Cal Offset:	-2.213107	-1.753170
NO ₂ Cal Slope:	0.994493	0.998289
NO ₂ Cal Offset:	-1.516093	-1.291753

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
High point	4934	66.3	799.5	796.9	2.7	798.4	799.0	-0.6	1.0014	0.9973
Mid point	4967	33.2	400.4	399.0	1.3	399.6	398.0	1.7	1.0019	1.0026
Low point	4983	16.6	200.2	199.5	0.7	200.2	197.2	3.0	1.0001	1.0119
As left zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
As left span	4934	66.3	799.5	301.5	498.0	796.6	301.5	495.1	1.0037	1.0000
Average Correction Factor									1.0011	1.0040

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	797.3	304.8	495.2	493.5	1.0033	99.7%
Mid GPT point	797.3	551.9	248.1	246.1	1.0079	99.2%
Low GPT point	797.3	673.3	126.7	123.7	1.0239	97.7%
Average Correction Factor					1.0117	98.8%

Notes: No maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 9, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:54	End Time (MST):	10:51
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	0.999999	≥0.995
799.5	798.4	1.0014	Slope	0.998841	0.90 - 1.10
400.4	399.6	1.0019	Intercept	-0.173749	+/-20
200.2	200.2	1.0001			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 9, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:54	End Time (MST):	10:51
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999971	≥0.995
495.2	493.5	1.0033	Slope	0.998289	0.90 - 1.10
248.1	246.1	1.0079	Intercept	-1.291753	+/-20
126.7	123.7	1.0239			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 9, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:54	End Time (MST):	10:51
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999985	≥ 0.995
796.9	799.0	0.9973	Slope	1.003897	$0.90 - 1.10$
399.0	398.0	1.0026	Intercept	-1.753170	± 20
199.5	197.2	1.0119			



NO_x Calibration Plot

Date: July 7, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Hills Station number: AMS 23
Calibration Date: July 10, 2025 Last Cal Date: June 10, 2025
Start time (MST): 6:46 End time (MST): 7:43

Analyzer Make: API T640 S/N: 320
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388744
Temp/RH standard: Alicat FP-25BT S/N: 388744

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	16.9	16.8	16.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	730.1	729.3	730.1	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	4.97	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	49	----	51	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	14.8	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 16-Jul-26
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	8.1	11.4	11.4	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: July 10, 2025
Date Disposable Filter Changed: July 10, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 10, 2025
Date RH/T Sensor Cleaned: July 10, 2025

Notes: No adjustments done. Leak Check, Flow and PMT checked before and after cleaning.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25 WASKŌW OHCI PIMÂTISIWIN JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS 25
Calibration Date: July 29, 2025 Last Cal Date: June 25, 2025
Start time (MST): 6:58 End time (MST): 9:43
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.70 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC342445
Removed Cal Gas Conc: 49.70 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 621
Zero Air Gen Model: API T701 Serial Number: 4765

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1118148497
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999284	1.000255	Backgd or Offset:	10.2	10.5
Calibration intercept:	0.587369	0.447870	Coeff or Slope:	1.038	1.038

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.5	----
As found High point	4920	80.5	800.1	801.0	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	800.5	Previous response	800.1	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	80.5	800.1	800.5	0.999
Mid point	4960	40.2	399.6	400.6	0.997
Low point	4980	20.1	199.8	200.3	0.997
As left zero	5000	0.0	0.0	0.8	----
As left span	4920	80.5	800.1	803.6	0.996
Average Correction Factor:					0.998

Notes: No maintenance done. Zero adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 29, 2025	Previous Calibration:	June 25, 2025
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS 25
Start Time (MST):	6:58	End Time (MST):	9:43
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999999	≥0.995
800.1	800.5	0.9995	Slope	1.000255	0.90 - 1.10
399.6	400.6	0.9974	Intercept	0.447870	+/-30
199.8	200.3	0.9975			



SO2 Calibration Plot

Date: July 29, 2025

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS 25
Calibration Date: July 25, 2025 Last Cal Date: June 23, 2025
Start time (MST): 5:29 End time (MST): 9:42
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC517099
Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 747
ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
Converter make: Global G-150 Converter serial #: 2022-219
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>	
Calibration slope:	1.002651	1.005076	Backgd or Offset:	3.50	3.55
Calibration intercept:	0.042240	-0.017692	Coeff or Slope:	1.102	1.094

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4919	80.5	80.0	81.2	0.983
As found Mid point	4960	40.3	40.1	40.5	0.984
As found Low point	4980	20.1	20.0	20.4	0.970
New cylinder response					
Baseline Corr As found:	81.4	Prev response:	80.28	*% change:	1.4%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.015927	AF Intercept:	-0.097627
Baseline Corr 3rd AF pt:	20.6	AF Correlation:	0.999983	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4919	80.5	80.0	80.4	0.995
Mid point	4960	40.3	40.1	40.2	0.996
Low point	4980	20.1	20.0	20.2	0.989
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	80.0	800.0	809.5	0.988
SO ₂ Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.994
Date of last converter efficiency test:		February 12, 2025		111.0% efficiency	

Notes: SO_x Scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	July 25, 2025	Previous Calibration:	June 23, 2025
Station Name:	Waskow ohci Pimatisiwin	Station Number:	10:47:00 AM
Start Time (MST):	5:29	End Time (MST):	9:42
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999992	≥ 0.995
80.0	80.4	0.9953	Slope	1.005076	$0.90 - 1.10$
40.1	40.2	0.9964	Intercept	-0.017692	± 3
20.0	20.2	0.9891			


H₂S Calibration Curve



H₂S Calibration Plot

Date: July 25, 2025

Location: Waskow ohci Pimatisiwin





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27
JACKFISH 2/3
JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Jackfish 2/3 Station number: AMS 27
Calibration Date: July 14, 2025 Last Cal Date: June 18, 2025
Start time (MST): 10:54 End time (MST): 14:31
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.58 ppm Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #: SG9133974BAL
Removed Cal Gas Conc: 50.58 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 5252
Zero Air Gen Model: API 701 Serial Number: 268

Analyzer Information

Analyzer make: Thermo 43iQ-TL Serial Number: 12124313138
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.013181	0.998916	Backgd or Offset:	8.2	9.2
Calibration intercept:	0.883047	0.522481	Coeff or Slope:	0.923	0.907

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	1.0	----
As found High point	4913	78.9	799.4	813.6	0.984
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	812.6	Previous response	810.9	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4913	78.9	799.4	798.9	1.001
Mid point	4955	39.5	400.0	400.3	0.999
Low point	4971	19.7	199.7	200.4	0.996
As left zero	5000	0.0	0.0	-0.1	----
As left span	4913	78.9	799.4	800.4	0.999
Average Correction Factor:					0.999

Notes: Changed the sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 14, 2025	Previous Calibration:	June 18, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	10:54	End Time (MST):	14:31
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999999	≥0.995
799.4	798.9	1.0007	Slope	0.998916	0.90 - 1.10
400.0	400.3	0.9993	Intercept	0.522481	+/-30
199.7	200.4	0.9963			



SO2 Calibration Plot

Date: July 14, 2025

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Jackfish 2/3
Calibration Date: July 16, 2025
Start time (MST): 11:37
Reason: Routine

Station number: AMS 27
Last Cal Date: June 24, 2025
End time (MST): 16:44

Calibration Standards

Cal Gas Concentration: 4.87 ppm
Cal Gas Cylinder #: CC523090
Removed Cal Gas Conc: 4.87 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Exp Date: September 5, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 5252
Serial Number: 268

Analyzer Information

Analyzer make: Thermo 43iQ
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12228021055
Converter serial #: 2022-195
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005298	1.017160	Backgd or Offset:	3.26	3.26
Calibration intercept:	0.075924	-0.164112	Coeff or Slope:	1.228	1.228

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4918	82.1	80.0	79.1	1.010
As found Mid point	4951	41.0	40.0	39.4	1.013
As found Low point	4973	20.5	20.0	19.7	1.010
New cylinder response					
Baseline Corr As found:	79.2	Prev response:	80.46	*% change:	-1.6%
Baseline Corr 2nd AF pt:	39.5	AF Slope:	0.990291	AF Intercept:	-0.123983
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999997	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4918	82.1	80.0	81.2	0.985
Mid point	4951	41.0	40.0	40.6	0.985
Low point	4973	20.5	20.0	19.9	1.005
As left zero	5000	0.0	0.0	0.1	----
As left span	4911	82.0	80.0	80.8	0.990
SO2 Scrubber Check	4915	78.9	790.0	0.1	----
Date of last scrubber change:	21-Feb-25		Ave Corr Factor		0.992
Date of last converter efficiency test:	April 23, 2025		91.4% efficiency		

Notes: Changed sample inlet filter after as founds. Scrubber check passed. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	July 16, 2025	Previous Calibration:	June 24, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:37	End Time (MST):	16:44
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12228021055

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999970	≥ 0.995
80.0	81.2	0.9848	Slope	1.017160	$0.90 - 1.10$
40.0	40.6	0.9852	Intercept	-0.164112	± 3
20.0	19.9	1.0047			



H₂S Calibration Plot

Date: July 16, 2025

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Jackfish 2/3
Station number: AMS 27
Calibration Date: July 18, 2025
Last Cal Date: June 20, 2025
Start time (MST): 11:21
End time (MST): 16:20
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC757838
NOX Cal Gas Conc: 60.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.30 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 60.20 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.20 ppm
NO gas Diff:
Serial Number: 5252
Serial Number: 268

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	1.1	0.9	0.2	----	----
AF High point	4924	66.3	801.1	799.8	1.3	786.8	780.0	6.8	1.0196	1.0266
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 800.3 ppb	NO = 797.0 ppb				* = > +/-5% change initiates investigation			*Percent Change	NO _x = -1.9%
Baseline Corr 1st pt	NO _x = 785.7 ppb	NO = 779.1 ppb				<u>As Found Statistics</u>			*Percent Change	NO = -2.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:		Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:		NO Int:
						As found	NO ₂ r ² :	NO ₂ SI:		NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153357

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.306	1.333	NO bkgnd or offset:	4.4	4.4
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	4.4	4.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	167.9	169.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996304	0.998045
NO _x Cal Offset:	2.090330	2.629762
NO Cal Slope:	0.996514	0.998116
NO Cal Offset:	-0.028881	0.410595
NO ₂ Cal Slope:	1.002731	1.001080
NO ₂ Cal Offset:	0.242583	1.183955

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	1.8	1.0	0.8	----	----
High point	4924	66.3	801.1	799.8	1.3	801.8	798.9	3.0	0.9992	1.0011
Mid point	4958	33.2	401.1	400.4	0.7	403.3	400.1	3.2	0.9945	1.0008
Low point	4976	16.6	200.5	200.2	0.3	203.6	199.4	4.2	0.9847	1.0038
As left zero	5000	0.0	0.0	0.0	0.0	3.8	1.1	2.8	----	----
As left span	4924	66.3	801.1	401.4	399.7	799.2	401.4	397.8	1.0024	1.0000
Average Correction Factor									0.9928	1.0019

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.8	----	----
High GPT point	795.7	400.1	396.9	398.4	0.9963	100.4%
Mid GPT point	795.7	587.3	209.7	211.1	0.9935	100.7%
Low GPT point	795.7	683.1	113.9	115.8	0.9838	101.6%
Average Correction Factor					0.9912	100.9%

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 18, 2025	Previous Calibration:	June 20, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:21	End Time (MST):	16:20
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.8	----	Correlation Coefficient	0.999995	≥0.995
801.1	801.8	0.9992	Slope	0.998045	0.90 - 1.10
401.1	403.3	0.9945	Intercept	2.629762	+/-20
200.5	203.6	0.9847			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 18, 2025	Previous Calibration:	June 20, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:21	End Time (MST):	16:20
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.8	----	Correlation Coefficient	0.999994	≥0.995
396.9	398.4	0.9963	Slope	1.001080	0.90 - 1.10
209.7	211.1	0.9935	Intercept	1.183955	+/-20
113.9	115.8	0.9838			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 18, 2025	Previous Calibration:	June 20, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:21	End Time (MST):	16:20
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	1.0	----	Correlation Coefficient	0.999997	≥ 0.995
799.8	798.9	1.0011	Slope	0.998116	$0.90 - 1.10$
400.4	400.1	1.0008	Intercept	0.410595	± 20
200.2	199.4	1.0038			



NO_x Calibration Plot

Date: July 18, 2025

Location: Jackfish 2/3





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS29
SURMONT 2
JULY 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: July 7, 2025 Last Cal Date: June 16, 2025
Start time (MST): 10:43 End time (MST): 14:59
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.95 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC356229
Removed Cal Gas Conc: 49.95 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 5472
Zero Air Gen Model: Teledyne API T701 Serial Number: 4428

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1170050150
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002315	1.000545	Backgd or Offset:	13.7	14.0
Calibration intercept:	-1.059746	-0.940397	Coeff or Slope:	0.949	0.949

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4919.9	80.1	800.2	802.0	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	801.7	Previous response	801.0	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.1	800.2	800.0	1.000
Mid point	4960	40.0	399.6	398.9	1.002
Low point	4980	20.0	199.8	197.7	1.011
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	80.1	800.2	801.0	0.999
Average Correction Factor:					1.004

Notes: Changed sample inlet filter after as founds. Adjusted zero.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 16, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:43	End Time (MST):	14:59
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999993	≥0.995
800.2	800.0	1.0002	Slope	1.000545	0.90 - 1.10
399.6	398.9	1.0018	Intercept	-0.940397	+/-30
199.8	197.7	1.0106			



SO2 Calibration Plot

Date: July 7, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: July 3, 2025 Last Cal Date: June 12, 2025
Start time (MST): 10:06 End time (MST): 16:03
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.750 ppm Cal Gas Exp Date: August 28, 2027
Cal Gas Cylinder #: CC737848
Removed Cal Gas Conc: 4.750 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 5472
ZAG Make/Model: Teledyne API T701 Serial Number: 4428

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 1200326170
Converter make: Global Converter serial #: 2022-220
Analyzer Range: 0 - 100 ppb Converter Temp: 325.0 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002739	0.997310	Backgd or Offset:	0.94
Calibration intercept:	-0.040481	-0.100478	Coeff or Slope:	1.040

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4916	84.2	80.0	81.3	0.981
As found Mid point	4958	42.1	40.0	40.3	0.988
As found Low point	4979	21.1	20.0	20.0	0.990
New cylinder response					
Baseline Corr As found:	81.5	Prev response:	80.17	*% change:	1.6%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	1.019313	AF Intercept:	-0.320488
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999987	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4916	84.2	80.0	79.7	1.004
Mid point	4958	42.1	40.0	39.8	1.005
Low point	4979	21.1	20.0	19.7	1.015
As left zero	5000	0.0	0.0	0.0	----
As left span	4916	84.2	80.0	78.2	1.023
SO2 Scrubber Check	4919	81.3	813.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	1.008
Date of last converter efficiency test:		December 5, 2024		108.1% efficiency	

Notes: Changed sample inlet filter after as founds. Ran SOx scrubber check after cal zero, passed. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

H2S Calibration Summary


Station Information

Calibration Date:	July 3, 2025	Previous Calibration:	June 12, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:06	End Time (MST):	16:03
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999991		≥ 0.995
80.0	79.7	1.0036	Slope	0.997310		$0.90 - 1.10$
40.0	39.8	1.0049	Intercept	-0.100478		± 3
20.0	19.7	1.0151				


H₂S Calibration Curve



H2S Calibration Plot

Date: July 3, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Surmont 2
Calibration Date: July 7, 2025
Start time (MST): 10:43
Reason: Routine

Station number: AMS 29
Last Cal Date: June 16, 2025
End time (MST): 14:59

Calibration Standards

Gas Cert Reference: CC356229
CH4 Cal Gas Conc. 503.7 ppm
C3H8 Cal Gas Conc. 204.8 ppm
Removed Gas Cert: NA
Removed CH4 Conc. 503.7 ppm
Removed C3H8 Conc. 204.8 ppm
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Expiry Date: October 9, 2032
CH4 Equiv Conc. 1066.9 ppm
Removed Gas Expiry: NA
CH4 Equiv Conc. 1066.9 ppm
Diff between cyl:
Serial Number: 5472
Serial Number: 4428

Analyzer Information

Analyzer make: Thermo 51i-LT
Analyzer Range: 0 - 20 ppm

Analyzer serial #: 1170050149

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996526	1.005647	Background:	3.55	3.61
Calibration intercept:	-0.064169	-0.040090	Coefficient:	3.973	4.038

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	-0.02	----
As found High point	4920	80.1	17.09	16.83	1.014
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	16.85	Previous response	16.97	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	-0.04	----
High point	4920	80.1	17.09	17.13	0.998
Mid point	4960	40.0	8.54	8.60	0.992
Low point	4980	20.0	4.27	4.21	1.013
As left zero	5000	0.0	0.00	-0.02	----
As left span	4920	80.1	17.09	17.17	0.995

Average Correction Factor 1.001

Notes: Changed sample inlet filter and hydrogen cylinder after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 16, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:43	End Time (MST):	14:59
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.04	----	Correlation Coefficient	0.999968	≥ 0.995
17.09	17.13	0.9978	Slope	1.005647	$0.90 - 1.10$
8.54	8.60	0.9925	Intercept	-0.040090	± 1.5
4.27	4.21	1.0134			


THC Calibration Curve



THC Calibration Plot

Date: July 7, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Surmont 2
Station number: AMS 29
Calibration Date: July 15, 2025
Last Cal Date: June 3, 2025
Start time (MST): 9:47
End time (MST): 15:43
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC218007
NOX Cal Gas Conc: 60.20 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.20 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 5472
Serial Number: 4428

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0	0.0	0.0	0.0	-0.6	-0.1	-0.5	----	----
AF High point	4933	66.7	803.1	800.4	2.7	781.6	775.8	5.8	1.0267	1.0316
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 796.5 ppb	NO = 795.7 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -1.8%	
Baseline Corr 1st pt	NO _x = 782.2 ppb	NO = 775.9 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -2.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1170050148

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.024	1.050	NO bkgnd or offset:	1.3	1.4
NOX coeff or slope:	0.922	0.990	NOX bkgnd or offset:	1.9	1.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	149.3	151.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.990952	1.001518
NO _x Cal Offset:	0.652304	-1.709177
NO Cal Slope:	0.995198	1.003169
NO Cal Offset:	-0.927820	-2.468522
NO ₂ Cal Slope:	0.993556	0.990533
NO ₂ Cal Offset:	0.106812	0.265734

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4933	66.7	803.1	800.4	2.7	803.0	801.0	2.3	1.0001	0.9993
Mid point	4967	33.3	400.9	399.6	1.3	400.4	399.4	1.0	1.0013	1.0004
Low point	4983	16.7	201.1	200.4	0.7	197.0	194.6	2.5	1.0207	1.0299
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.0	0.1	----	----
As left span	4933	66.7	803.1	413.8	389.3	794.9	413.8	381.1	1.0103	1.0000
Average Correction Factor									1.0074	1.0099

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	795.2	408.3	389.6	385.8	1.0098	99.0%
Mid GPT point	795.2	610.9	187.0	186.3	1.0036	99.6%
Low GPT point	795.2	701.8	96.1	95.2	1.0091	99.1%
Average Correction Factor					1.0075	99.3%

Notes:

Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 3, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:47	End Time (MST):	15:43
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999970	≥0.995
803.1	803.0	1.0001	Slope	1.001518	0.90 - 1.10
400.9	400.4	1.0013	Intercept	-1.709177	+/-20
201.1	197.0	1.0207			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 3, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:47	End Time (MST):	15:43
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999988	≥ 0.995
389.6	385.8	1.0098	Slope	0.990533	0.90 - 1.10
187.0	186.3	1.0036	Intercept	0.265734	+/-20
96.1	95.2	1.0091			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 3, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:47	End Time (MST):	15:43
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999934	≥ 0.995
800.4	801.0	0.9993	Slope	1.003169	$0.90 - 1.10$
399.6	399.4	1.0004	Intercept	-2.468522	± 20
200.4	194.6	1.0299			



NO_x Calibration Plot

Date: July 15, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: July 15, 2025 Last Cal Date: June 16, 2025
Start time (MST): 10:09 End time (MST): 11:52

Analyzer Make: API T640 S/N: 2236
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	17.5	17.83	17.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	721.3	720.00	721.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	4.951	4.951	<input checked="" type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	31	----	31	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 3.9		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.1	11.0	11.0	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: July 15, 2025
Date Disposable Filter Changed: July 15, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: April 14, 2025
Date RH/T Sensor Cleaned: July 15, 2025

Notes: Verified temperature, pressure and flow. Cleaned RH/T sensor. Adjusted PMT voltage and flow.

Calibration by: Braiden Boutilier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS30 ELLS RIVER JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Ells River Station number: AMS 30
Calibration Date: July 7, 2025 Last Cal Date: June 6, 2025
Start time (MST): 9:48 End time (MST): 12:28
Reason: Routine

Calibration Standards

Cal Gas Concentration: 48.75 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC350110
Removed Cal Gas Conc: 48.75 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3061
Zero Air Gen Model: API T701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1008841397
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002418	1.003719	Backgd or Offset:	11.3	10.6
Calibration intercept:	-3.072052	-3.352039	Coeff or Slope:	0.991	0.991

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-1.1	----
As found High point	4918	82.0	799.5	797.1	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	798.2	Previous response	798.4	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4918	82.0	799.5	800.8	0.998
Mid point	4959	41.0	399.8	396.1	1.009
Low point	4980	20.5	199.9	194.1	1.030
As left zero	5000	0.0	0.0	-0.4	----
As left span	4918	82.0	799.5	802.0	0.997
Average Correction Factor:					1.012

Notes: Sample inlet filter was changed after as founds. Adjusted zero only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 6, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:48	End Time (MST):	12:28
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999925	≥0.995
799.5	800.8	0.9984	Slope	1.003719	0.90 - 1.10
399.8	396.1	1.0092	Intercept	-3.352039	+/-30
199.9	194.1	1.0296			



SO2 Calibration Plot

Date: July 7, 2025

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Ells River
Calibration Date: July 14, 2025
Start time (MST): 9:20
Reason: Routine

Station number: AMS 30
Last Cal Date: June 18, 2025
End time (MST): 13:07

Calibration Standards

Cal Gas Concentration: 4.99 ppm
Cal Gas Cylinder #: CC505806
Removed Cal Gas Conc: 4.99 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API 701H

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: CDN- 101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1410661331
Converter serial #: 562
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006617	0.998624	Backgd or Offset:	1.7	1.7
Calibration intercept:	-0.060455	-0.280669	Coeff or Slope:	1.089	1.089

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	80.2	80.0	78.9	1.014
As found Mid point	4960	40.1	40.0	39.2	1.021
As found Low point	4980	20.0	20.0	19.2	1.040
New cylinder response					
Baseline Corr As found:	78.9	Prev response:	80.51	*% change:	-2.0%
Baseline Corr 2nd AF pt:	39.2	AF Slope:	0.987488	AF Intercept:	-0.240896
Baseline Corr 3rd AF pt:	19.2	AF Correlation:	0.999957	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.2	80.0	79.8	1.003
Mid point	4960	40.1	40.0	39.5	1.013
Low point	4980	20.0	20.0	19.4	1.029
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	80.2	80.0	80.0	1.000
SO2 Scrubber Check	4918	82.0	820.0	0.0	----
Date of last scrubber change:	18-Jun-25		Ave Corr Factor		1.015
Date of last converter efficiency test:					

Notes: Changed sample inlet filter, scrubber and lines after multipoint as founds. SO2 scrubber check done after calibrator zero and passed. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	July 14, 2025	Previous Calibration:	June 18, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:20	End Time (MST):	13:07
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1410661331

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999943	≥ 0.995
80.0	79.8	1.0030	Slope	0.998624	$0.90 - 1.10$
40.0	39.5	1.0131	Intercept	-0.280669	± 3
20.0	19.4	1.0289			

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Ells River
Calibration Date: July 7, 2025
Start time (MST): 9:48
Reason: Routine

Station number: AMS 30
Last Cal Date: June 6, 2025
End time (MST): 12:28

Calibration Standards

Gas Cert Reference:	CC350110	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	496.6 ppm	CH ₄ Equiv Conc.	1066.4 ppm
C ₃ H ₈ Cal Gas Conc.	207.2 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	496.6 ppm	CH ₄ Equiv Conc.	1066.4 ppm
Removed C ₃ H ₈ Conc.	207.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3061
Zero Air Gen model:	API T701H	Serial Number:	358

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1152430011
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.11E-04	3.11E-04	NMHC SP Ratio:	5.96E-05
CH ₄ Retention time:	17.4	17.4	NMHC Peak Area:	156612
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	17.49	17.58	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.58	Prev response	17.33	*% change	1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	82.0	17.49	17.55	0.997
Mid point	4959	41.0	8.74	8.70	1.005
Low point	4980	20.5	4.37	4.29	1.018
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	17.49	17.53	0.998
Average Correction Factor					1.007

Notes: Sample inlet filter was changed after as founds. No adjustment made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4918	82.0	9.34	9.45	0.989
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.45	Prev response	9.26	*% change	2.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4918	82.0	9.34	9.43	0.991
Mid point	4959	41.0	4.67	4.70	0.995
Low point	4980	20.5	2.34	2.34	1.000
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	9.34	9.42	0.992
Average Correction Factor					0.995

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4918	82.0	8.14	8.13	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.13	Prev response	8.06	*% change	0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4918	82.0	8.14	8.12	1.003
Mid point	4959	41.0	4.07	4.01	1.017
Low point	4980	20.5	2.04	1.96	1.040
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	8.14	8.11	1.004
Average Correction Factor					1.020

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.993984	1.004720
THC Cal Offset:	-0.051027	-0.051425
CH ₄ Cal Slope:	0.994349	0.999022
CH ₄ Cal Offset:	-0.037920	-0.039319
NMHC Cal Slope:	0.992540	1.009588
NMHC Cal Offset:	-0.011507	-0.011706

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 6, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:48	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999960	≥ 0.995
17.49	17.55	0.9966	Slope	1.004720	$0.90 - 1.10$
8.74	8.70	1.0049	Intercept	-0.051425	± 0.5
4.37	4.29	1.0181			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 6, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:48	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999893	<i>≥0.995</i>
8.14	8.12	1.0032	Slope	0.999022	<i>0.90 - 1.10</i>
4.07	4.01	1.0168	Intercept	-0.039319	<i>+/-0.5</i>
2.04	1.96	1.0398			






Wood Buffalo Environmental Association NMHC Calibration Summary


Station Information

Calibration Date:	July 7, 2025	Previous Calibration:	June 6, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:48	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999993	≥ 0.995
9.34	9.43	0.9911	Slope	1.009588	$0.90 - 1.10$
4.67	4.70	0.9945	Intercept	-0.011706	± 0.5
2.34	2.34	1.0000			







Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Ells River
Station number: AMS 30
Calibration Date: July 11, 2025
Last Cal Date: June 16, 2025
Start time (MST): 9:04
End time (MST): 13:09
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0027487
NOX Cal Gas Conc: 59.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.30 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.10 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.10 ppm
NO gas Diff:
Serial Number: 3061
Serial Number: 358

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.0	----	----
AF High point	4932	67.7	803.0	800.3	2.7	808.7	804.4	4.3	0.9933	0.9952
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 800.9 ppb	NO = 797.6 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.9%
Baseline Corr 1st pt	NO _x = 808.4 ppb	NO = 804.1 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 710321429

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.176	1.268	NO bkgnd or offset:	13.5	14.4
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	13.6	14.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	197.8	196.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999206	1.005398
NO _x Cal Offset:	-1.419239	-2.018135
NO Cal Slope:	1.000634	1.003019
NO Cal Offset:	-3.160802	-3.220169
NO ₂ Cal Slope:	0.999422	0.998213
NO ₂ Cal Offset:	0.412998	0.317699

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.5	0.5	0.1	----	----
High point	4932	67.7	803.0	800.3	2.7	806.5	801.2	5.2	0.9956	0.9988
Mid point	4966	33.8	400.9	399.5	1.4	399.9	396.0	3.9	1.0025	1.0089
Low point	4983	16.9	200.4	199.8	0.7	196.9	193.2	3.6	1.0180	1.0340
As left zero	5000	0.0	0.0	0.0	0.0	0.5	0.5	0.0	----	----
As left span	4932	67.7	803.0	428.5	374.5	806.2	428.5	377.8	0.9960	1.0000
Average Correction Factor									1.0054	1.0139

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	798.9	422.5	379.1	378.7	1.0011	99.9%
Mid GPT point	798.9	610.1	191.5	191.4	1.0006	99.9%
Low GPT point	798.9	703.4	98.2	98.7	0.9950	100.5%
Average Correction Factor					0.9989	100.1%

Notes: Sample inlet filter changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 11, 2025	Previous Calibration:	June 16, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:04	End Time (MST):	13:09
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.5	----	Correlation Coefficient	0.999956	<i>≥0.995</i>
803.0	806.5	0.9956	Slope	1.005398	<i>0.90 - 1.10</i>
400.9	399.9	1.0025	Intercept	-2.018135	<i>+/-20</i>
200.4	196.9	1.0180			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 11, 2025	Previous Calibration:	June 16, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:04	End Time (MST):	13:09
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999998	≥0.995
379.1	378.7	1.0011	Slope	0.998213	0.90 - 1.10
191.5	191.4	1.0006	Intercept	0.317699	+/-20
98.2	98.7	0.9950			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 11, 2025	Previous Calibration:	June 16, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:04	End Time (MST):	13:09
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.5	----	Correlation Coefficient	0.999901	≥ 0.995
800.3	801.2	0.9988	Slope	1.003019	0.90 - 1.10
399.5	396.0	1.0089	Intercept	-3.220169	+/-20
199.8	193.2	1.0340			



NO_x Calibration Plot

Date: July 11, 2025

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Ells River Station number: AMS 30
Calibration Date: July 11, 2025 Last Cal Date: June 18, 2025
Start time (MST): 11:09 End time (MST): 12:09

Analyzer Make: API T640 S/N: 875
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	25.20	25.34	25.20	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	724.90	727.13	724.90	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	4.95	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	30	----	30	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 4.40		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: January 30, 2027
Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	0.20	10.50	10.50	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: July 11, 2025
Date Disposable Filter Changed: July 11, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: June 18, 2025
Date RH/T Sensor Cleaned: June 18, 2025

Notes: Verified flow, temperature, pump power and pressure No adjustment made. Leak check passed.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS31
BLACKROD
JULY 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Blackrod
Calibration Date: July 29, 2025
Start time (MST): 10:27
Reason: Routine

Station number: AMS 31
Last Cal Date: June 20, 2025
End time (MST): 13:12

Calibration Standards

Cal Gas Concentration: 50.25 ppm
Cal Gas Cylinder #: CC327023
Removed Cal Gas Conc: 50.25 ppm
Removed Gas Cyl #: N/A
Calibrator Model: Teledyne T700
Zero Air Gen Model: Teledyne N701H

Cal Gas Exp Date: March 10, 2031
Rem Gas Exp Date: N/A
Diff between cyl:
Serial Number: 5762
Serial Number: 72

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 1160290014

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008527	1.008913	Backgd or Offset:	37.8	37.8
Calibration intercept:	-0.871940	-0.031926	Coeff or Slope:	0.980	0.980

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4920	79.6	800.0	807.4	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	807.6	Previous response	806.0	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	79.6	800.0	807.0	0.991
Mid point	4960	39.8	400.0	404.2	0.990
Low point	4980	19.9	200.0	201.0	0.995
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	79.6	800.0	810.0	0.988
Average Correction Factor:					0.992

Notes: Sample inlet filter was changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	July 29, 2025	Previous Calibration:	June 20, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	10:27	End Time (MST):	13:12
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290014

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999997	≥0.995
800.0	807.0	0.9914	Slope	1.008913	0.90 - 1.10
400.0	404.2	0.9896	Intercept	-0.031926	+/-30
200.0	201.0	0.9950			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: July 29, 2025

Location: Blackrod





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Blackrod Station number: AMS 31
Calibration Date: July 16, 2025 Last Cal Date: June 11, 2025
Start time (MST): 10:16 End time (MST): 14:03
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.42 ppm Cal Gas Exp Date: March 19, 2027
Cal Gas Cylinder #: DT0016926
Removed Cal Gas Conc: 5.42 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 5762
ZAG Make/Model: Teledyne API N701H Serial Number: 72

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 12228021056
Converter make: Global Converter serial #: 2023-266
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001619	1.004049	Backgd or Offset:	2.62
Calibration intercept:	-0.020398	-0.080452	Coeff or Slope:	0.974

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4926	73.8	80.0	80.1	0.996
As found Mid point	4963	36.9	40.0	40.2	0.990
As found Low point	4982	18.5	20.1	20.0	0.993
New cylinder response					
Baseline Corr As found:	80.3	Prev response:	80.11	*% change:	0.2%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.003763	AF Intercept:	-0.120470
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999988	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4926	73.8	80.0	80.2	0.998
Mid point	4963	36.9	40.0	40.2	0.995
Low point	4982	18.5	20.1	20.0	1.003
As left zero	5000	0.0	0.0	0.2	----
As left span	4926	73.8	80.0	79.1	1.011
SO ₂ Scrubber Check	4920	79.6	796.1	0.1	----
Date of last scrubber change:				Ave Corr Factor	0.998
Date of last converter efficiency test:					

Notes: Sample inlet filter was changed after multipoint as founds. SO₂ scrubber check done after calibrator zero and passed. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	July 16, 2025	Previous Calibration:	June 11, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	10:16	End Time (MST):	14:03
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12228021056

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999995		≥ 0.995
80.0	80.2	0.9975	Slope	1.004049		$0.90 - 1.10$
40.0	40.2	0.9950	Intercept	-0.080452		± 3
20.1	20.0	1.0026				


H₂S Calibration Curve



H₂S Calibration Plot

Date: July 16, 2025

Location: Blackrod





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Blackrod
Station number: AMS 31
Calibration Date: July 15, 2025
Last Cal Date: June 10, 2025
Start time (MST): 9:35
End time (MST): 13:37
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0035071
NOX Cal Gas Conc: 59.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.30 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API N701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.10 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.10 ppm
NO gas Diff:
Serial Number: 5762
Serial Number: 72

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.3	0.2	-1.5	----	----
AF High point	4932	67.7	803.0	800.3	2.7	815.5	812.1	3.4	0.9831	0.9857
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 807.3 ppb	NO = 807.9 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = 1.2%
Baseline Corr 1st pt	NO _x = 816.8 ppb	NO = 811.9 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1426262592

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.015	1.015	NO bkgnd or offset:	13.9	14.3
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	15.5	14.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	178.2	174.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.006675	1.006453
NO _x Cal Offset:	-1.016585	2.286512
NO Cal Slope:	1.011769	1.008561
NO Cal Offset:	-1.757010	1.265522
NO ₂ Cal Slope:	0.993747	0.990265
NO ₂ Cal Offset:	0.261616	0.356789

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
High point	4932	67.7	803.0	800.3	2.7	808.0	806.5	1.5	0.9938	0.9923
Mid point	4966	33.8	400.9	399.5	1.4	410.6	408.6	2.0	0.9763	0.9778
Low point	4983	16.9	200.4	199.8	0.7	204.1	201.5	2.6	0.9821	0.9914
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
As left span	4932	67.7	803.0	419.6	383.4	813.0	419.6	393.7	0.9877	1.0000
Average Correction Factor									0.9841	0.9871

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	814.7	417.2	400.2	396.0	1.0106	98.9%
Mid GPT point	814.7	622.8	194.6	194.6	1.0000	100.0%
Low GPT point	814.7	718.7	98.7	97.8	1.0093	99.1%
Average Correction Factor					1.0067	99.3%

Notes: Sample inlet filter was changed after as founds. Adjusted zero only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 10, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	9:35	End Time (MST):	13:37
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999900	≥0.995
803.0	808.0	0.9938	Slope	1.006453	0.90 - 1.10
400.9	410.6	0.9763	Intercept	2.286512	+/-20
200.4	204.1	0.9821			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 10, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	9:35	End Time (MST):	13:37
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999963	≥0.995
400.2	396.0	1.0106	Slope	0.990265	0.90 - 1.10
194.6	194.6	1.0000	Intercept	0.356789	+/-20
98.7	97.8	1.0093			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 15, 2025	Previous Calibration:	June 10, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	9:35	End Time (MST):	13:37
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999927	≥ 0.995
800.3	806.5	0.9923	Slope	1.008561	0.90 - 1.10
399.5	408.6	0.9778	Intercept	1.265522	+/-20
199.8	201.5	0.9914			



NO_x Calibration Plot

Date: July 15, 2025

Location: Blackrod





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS33 MONDAY CREEK JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Monday Creek Station number: AMS 33
Calibration Date: July 10, 2025 Last Cal Date: June 26, 2025
Start time (MST): 9:54 End time (MST): 12:24
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.62 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: EB0008522
Removed Cal Gas Conc: 50.62 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne T700 Serial Number: 3253
Zero Air Gen Model: Teledyne T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1152430005
Analyzer Range: 0- 1000 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003239	0.999584	Backgd or Offset: 30.3	30.3
Calibration intercept:	-0.137941	0.342191	Coeff or Slope: 0.970	0.970

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4921	79.1	800.8	800.0	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	800.0	Previous response	803.2	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4921	79.1	800.8	800.4	1.000
Mid point	4961	39.5	399.9	401.0	0.997
Low point	4980	19.8	200.5	200.4	1.000
As left zero	5000	0.0	0.0	0.2	----
As left span	4921	79.1	800.8	801.0	1.000
Average Correction Factor:					0.999

Notes: Sample inlet filter was changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 10, 2025	Previous Calibration:	June 26, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	9:54	End Time (MST):	12:24
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430005

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999996	≥0.995
800.8	800.4	1.0005	Slope	0.999584	0.90 - 1.10
399.9	401.0	0.9972	Intercept	0.342191	+/-30
200.5	200.4	1.0003			



SO2 Calibration Plot

Date: July 10, 2025

Location: Monday Creek





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Monday Creek
Calibration Date: July 3, 2025
Start time (MST): 10:44
Reason: Routine

Station number: AMS 33
Last Cal Date: June 24, 2025
End time (MST): 14:15

Calibration Standards

Cal Gas Concentration: 5.05 ppm
Cal Gas Cylinder #: DT0014831
Removed Cal Gas Conc: 5.05 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne T701H

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3253
Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL
Converter make: Global 150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12333331547
Converter serial #: 2022-196
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003426	1.006141	Backgd or Offset:	1.6	1.3
Calibration intercept:	-0.141607	0.038389	Coeff or Slope:	1.062	1.062

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4921	79.2	80.0	79.0	1.009
As found Mid point	4960	39.6	40.0	40.0	0.993
As found Low point	4980	19.8	20.0	20.1	0.980
New cylinder response					
Baseline Corr As found:	79.3	Prev response:	80.12	*% change:	-1.0%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	0.989853	AF Intercept:	0.058396
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999893	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	79.2	80.0	80.5	0.994
Mid point	4960	39.6	40.0	40.3	0.993
Low point	4980	19.8	20.0	20.2	0.990
As left zero	5000	0.0	0.0	0.0	----
As left span	4921	79.2	80.0	80.0	1.000
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:	11-Apr-24		Ave Corr Factor		0.992
Date of last converter efficiency test:					

Notes: Sample inlet filter changed after multipoint as founds. SO2 scrubber check done after calibrator zero and passed. Adjusted zero only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

H2S Calibration Summary


Station Information


Calibration Date:	July 3, 2025	Previous Calibration:	June 24, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	10:44	End Time (MST):	14:15
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331547

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999999		≥ 0.995
80.0	80.5	0.9936	Slope	1.006141		$0.90 - 1.10$
40.0	40.3	0.9925	Intercept	0.038389		± 3
20.0	20.2	0.9900				

H₂S Calibration Curve







Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Monday Creek
Station number: AMS 33
Calibration Date: July 8, 2025
Last Cal Date: June 17, 2025
Start time (MST): 10:25
End time (MST): 14:25
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC755290
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: March 11, 2031
NO Cal Gas Conc: 48.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.70 ppm
NO gas Diff:
Serial Number: 3253
Serial Number: 832

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
AF High point	4918	82.1	802.9	799.6	3.3	820.2	812.2	7.9	0.9788	0.9844
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 799.6 ppb	NO = 794.8 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = 2.5%	
Baseline Corr 1st pt	NO _x = 820.3 ppb	NO = 812.3 ppb				<u>As Found Statistics</u>		*Percent Change	NO = 2.2%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12426335704

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.994647	0.999485
NO _x Cal Offset:	1.010251	0.710623
NO Cal Slope:	0.994201	1.000817
NO Cal Offset:	-0.229498	-0.069509
NO ₂ Cal Slope:	0.986412	0.989518
NO ₂ Cal Offset:	1.352716	1.303349

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.059	1.045	NO bkgnd or offset:	1.5	1.4
NOX coeff or slope:	1.002	0.999	NOX bkgnd or offset:	1.6	1.5
NO2 coeff or slope:	0.990	0.990	Reaction cell Press:	144.1	140.1

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4918	82.1	802.9	799.6	3.3	803.0	800.3	2.6	0.9999	0.9992
Mid point	4959	41.1	401.9	400.3	1.6	402.4	400.4	2.0	0.9989	0.9998
Low point	4979	20.5	200.5	199.7	0.8	202.1	199.8	2.3	0.9921	0.9994
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
As left span	4918	82.1	802.9	402.1	400.8	798.4	402.1	396.4	1.0057	1.0000
Average Correction Factor									0.9970	0.9995

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	799.4	398.5	404.2	400.5	1.0092	99.1%
Mid GPT point	799.4	604.0	198.7	198.9	0.9989	100.1%
Low GPT point	799.4	702.9	99.8	101.1	0.9870	101.3%
Average Correction Factor					0.9984	100.2%

Notes:

Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By:

Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 8, 2025	Previous Calibration:	June 17, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	10:25	End Time (MST):	14:25
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12426335704

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999996	≥0.995
802.9	803.0	0.9999	Slope	0.999485	0.90 - 1.10
401.9	402.4	0.9989	Intercept	0.710623	+/-20
200.5	202.1	0.9921			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 8, 2025	Previous Calibration:	June 17, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	10:25	End Time (MST):	14:25
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12426335704

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999950	≥0.995
404.2	400.5	1.0092	Slope	0.989518	0.90 - 1.10
198.7	198.9	0.9989	Intercept	1.303349	+/-20
99.8	101.1	0.9870			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 8, 2025	Previous Calibration:	June 17, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	10:25	End Time (MST):	14:25
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12426335704

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	1.000000	≥ 0.995
799.6	800.3	0.9992	Slope	1.000817	0.90 - 1.10
400.3	400.4	0.9998	Intercept	-0.069509	+/-20
199.7	199.8	0.9994			



NO_x Calibration Plot

Date: July 8, 2025

Location: Monday Creek





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS505 SAWBONES BAY JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Sawbones Bay Station number: AMS 505
Calibration Date: July 24, 2025 Last Cal Date: June 4, 2025
Start time (MST): 8:32 End time (MST): 11:20
Reason: Routine

Calibration Standards

Cal Gas Concentration: 51.40 ppm Cal Gas Exp Date: February 15, 2029
Cal Gas Cylinder #: EY0000672
Removed Cal Gas Conc: 51.40 ppm Rem Gas Exp Date: February 15, 2029
Removed Gas Cyl #: EY0000672 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 5112
Zero Air Gen Model: Teledyne API T701 Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 710321323
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011414	0.995221	Backgd or Offset:	22.0	20.7
Calibration intercept:	-1.034356	-0.691279	Coeff or Slope:	1.111	1.111

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-1.7	----
As found High point	4922	77.8	799.8	792.3	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	794.0	Previous response	807.9	*% change	-1.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4922	77.8	799.8	795.3	1.006
Mid point	4961	38.9	399.9	397.9	1.005
Low point	4981	19.5	200.4	197.6	1.014
As left zero	5000	0.0	0.0	-0.2	----
As left span	4922	77.8	799.8	795.4	1.006
Average Correction Factor:					1.008

Notes: Changed inlet filter after as founds. Adjusted zero only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	July 24, 2025	Previous Calibration:	June 4, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:32	End Time (MST):	11:20
Analyzer make:	Thermo 43i	Analyzer serial #:	710321323

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999994	≥0.995
799.8	795.3	1.0057	Slope	0.995221	0.90 - 1.10
399.9	397.9	1.0050	Intercept	-0.691279	+/-30
200.4	197.6	1.0144			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: July 24, 2025

Location: Sawbones Bay





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Sawbones Bay Station number: AMS 505
Calibration Date: July 23, 2025 Last Cal Date: June 11, 2025
Start time (MST): 8:22 End time (MST): 12:24
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.26 ppm Cal Gas Exp Date: March 19, 2027
Cal Gas Cylinder #: DT0034141
Removed Cal Gas Conc: 5.26 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T750 Serial Number: 282
ZAG Make/Model: Teledyne API T751H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 12113311965
Converter make: Global 150 Converter serial #: 2022-224
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998742	1.003888	Backgd or Offset:	0.920	0.920
Calibration intercept:	0.040000	0.060000	Coeff or Slope:	1.105	1.105

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4924	76.0	80.0	80.8	0.990
As found Mid point	4962	38.0	40.0	40.0	0.999
As found Low point	4981	19.0	20.0	19.9	1.004
New cylinder response					
Baseline Corr As found:	80.8	Prev response:	79.89	*% change:	1.1%
Baseline Corr 2nd AF pt:	40.0	AF Slope:	1.011321	AF Intercept:	-0.200000
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999965	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4924	76.0	80.0	80.4	0.994
Mid point	4962	38.0	40.0	40.0	0.999
Low point	4981	19.0	20.0	20.2	0.990
As left zero	5000	0.0	0.0	0.3	----
As left span	4924	76.0	80.0	79.4	1.007
SO2 Scrubber Check	4922	77.8	778.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.994
Date of last converter efficiency test:					

Notes: Changed inlet filter after as founds. Scrubber test was done after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	July 23, 2025	Previous Calibration:	June 11, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:22	End Time (MST):	12:24
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311965

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999986		≥0.995
80.0	80.4	0.9944	Slope	1.003888		0.90 - 1.10
40.0	40.0	0.9994	Intercept	0.060000		+/-3
20.0	20.2	0.9895				



H2S Calibration Plot

Date: July 23, 2025

Location: Sawbones Bay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Sawbones Bay
Station number: AMS 505
Calibration Date: July 3, 2025
Last Cal Date: June 12, 2025
Start time (MST): 8:15
End time (MST): 12:16
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0009786
NOX Cal Gas Conc: 60.10 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.10 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 5112
Serial Number: 690

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.5	0.0	----	----
AF High point	4933	66.7	801.8	800.4	1.3	801.0	797.7	3.3	1.0004	1.0028
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 805.1 ppb	NO = 802.7 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -0.5%	
Baseline Corr 1st pt	NO _x = 801.5 ppb	NO = 798.2 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -0.6%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 4259

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.940	0.940
NOX coeff or slope:	0.940	0.940
NO2 coeff or slope:	1.000	1.000

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	0.8	0.8
NOX bkgnd or offset:	1.3	1.3
Reaction cell Press:	3.8	3.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.005537	1.005807
NO _x Cal Offset:	-1.109170	-1.228895
NO Cal Slope:	1.005256	1.006240
NO Cal Offset:	-1.949049	-1.668830
NO ₂ Cal Slope:	1.000710	0.998322
NO ₂ Cal Offset:	1.201848	0.120580

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
High point	4933	66.7	801.8	800.4	1.3	805.1	803.9	1.3	0.9959	0.9957
Mid point	4967	33.3	400.2	399.6	0.7	402.6	401.5	1.1	0.9941	0.9952
Low point	4983	16.7	200.7	200.4	0.3	198.6	197.3	1.3	1.0108	1.0158
As left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	-0.1	----	----
As left span	4933	66.7	801.8	363.3	438.5	797.9	363.3	434.6	1.0049	1.0000
Average Correction Factor									1.0003	1.0022

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	800.1	362.8	438.6	437.7	1.0021	99.8%
Mid GPT point	800.1	561.3	240.1	240.5	0.9985	100.2%
Low GPT point	800.1	655.2	146.2	146.0	1.0016	99.8%
Average Correction Factor					1.0007	99.9%

Notes:

Changed inlet filter after as founds. No adjustment made.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 3, 2025	Previous Calibration:	June 12, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:15	End Time (MST):	12:16
Analyzer make:	API T200	Analyzer serial #:	4259

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999981	≥0.995
801.8	805.1	0.9959	Slope	1.005807	0.90 - 1.10
400.2	402.6	0.9941	Intercept	-1.228895	+/-20
200.7	198.6	1.0108			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 3, 2025	Previous Calibration:	June 12, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:15	End Time (MST):	12:16
Analyzer make:	API T200	Analyzer serial #:	4259

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999994	≥ 0.995
438.6	437.7	1.0021	Slope	0.998322	0.90 - 1.10
240.1	240.5	0.9985	Intercept	0.120580	+/-20
146.2	146.0	1.0016			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	July 3, 2025	Previous Calibration:	June 12, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:15	End Time (MST):	12:16
Analyzer make:	API T200	Analyzer serial #:	4259

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999970	≥ 0.995
800.4	803.9	0.9957	Slope	1.006240	$0.90 - 1.10$
399.6	401.5	0.9952	Intercept	-1.668830	± 20
200.4	197.3	1.0158			



NO_x Calibration Plot

Date: July 3, 2025

Location: Sawbones Bay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS507
KIRBY SOUTH**

JULY 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

August 29, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Kirby South
Calibration Date: July 9, 2025
Start time (MST): 15:40
Reason: Install

Station number: AMS 507
Last Cal Date: NA
End time (MST): 18:47

Calibration Standards

Cal Gas Concentration: 50.74 ppm
Cal Gas Cylinder #: CC255918
Removed Cal Gas Conc: 50.74 ppm
Removed Gas Cyl #:
Calibrator Model: Teledyne API T700
Zero Air Gen Model: Teledyne API T701H

Cal Gas Exp Date: October 9, 2032
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 2445
Serial Number: 880

Analyzer Information

Analyzer make: Thermo 43iQ
Analyzer Range: 0 - 1000 ppb

Serial Number: 1182340007

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.004045	Backgd or Offset:	NA	28.4
Calibration intercept:	NA	-0.483826	Coeff or Slope:	NA	1.106

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	78.8	799.7	803.0	0.996
Mid point	4961	39.4	399.8	399.8	1.000
Low point	4980	19.7	199.9	200.3	0.998
As left zero	5000	0.0	0.0	-0.1	----
As left span	4921	78.8	799.7	803.0	0.996

Average Correction Factor: 0.998

Notes: Install completed following the move to Kirby South. Swapped pump before calibrating. Adjusted zero.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	NA
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	15:40	End Time (MST):	18:47
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999995	≥0.995
799.7	803.0	0.9959	Slope	1.004045	0.90 - 1.10
399.8	399.8	1.0000	Intercept	-0.483826	+/-30
199.9	200.3	0.9981			



SO2 Calibration Plot

Date: July 9, 2025

Location: Kirby South





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Kirby South
Calibration Date: July 9, 2025
Start time (MST): 11:35
Reason: Install

Station number: AMS 507
Last Cal Date: NA
End time (MST): 16:00

Calibration Standards

Cal Gas Concentration: 5.05 ppm
Cal Gas Cylinder #: DT0019762
Removed Cal Gas Conc: 5.05 ppm
Removed Gas Cyl #: n/a
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 2445
Serial Number: 880

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1150840012
Converter serial #: 2022-197
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.014058	Backgd or Offset:	NA	1.70
Calibration intercept:	NA	-0.280000	Coeff or Slope:	NA	1.053

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4922	78.4	79.2	80.2	0.987
Mid point	4961	39.2	39.6	39.6	1.000
Low point	4980	19.6	19.8	19.6	1.010
As left zero	5000	0.0	0.0	0.1	----
As left span	4922	78.4	79.2	81.3	0.974
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:	July 25, 2023		Ave Corr Factor		0.999
Date of last converter efficiency test:	n/a				

Notes: Install completed for deployment to Kirby South. Swapped pump prior to calibrating. Scrubber check done after cal zero, passed.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	NA
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	11:35	End Time (MST):	16:00
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999937		≥ 0.995
79.2	80.2	0.9873	Slope	1.014058		0.90 - 1.10
39.6	39.6	0.9998	Intercept	-0.280000		+/-3
19.8	19.6	1.0100				



H2S Calibration Plot

Date: July 9, 2025

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Kirby South
Calibration Date: July 9, 2025
Start time (MST): 15:40
Reason: Install

Station number: AMS 507
Last Cal Date: NA
End time (MST): 19:05

Calibration Standards

Gas Cert Reference: CC255918
CH4 Cal Gas Conc. 506.4 ppm
C3H8 Cal Gas Conc. 205.0 ppm
Removed Gas Cert:
Removed CH4 Conc. 506.4 ppm
Removed C3H8 Conc. 205.0 ppm
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Expiry Date: October 9, 2032
CH4 Equiv Conc. 1070.2 ppm
Removed Gas Expiry:
CH4 Equiv Conc. 1070.2 ppm
Diff between cyl:
Serial Number: 2445
Serial Number: 880

Analyzer Information

Analyzer make: Thermo 51i-LT
Analyzer Range: 0 - 20 ppm

Analyzer serial #: 1182340005

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	0.996718	Background:	NA	1.91
Calibration intercept:	NA	-0.008033	Coefficient:	NA	3.774

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/Ic-AFzero) <i>Limit = 0.90-1.10</i>
As found zero					----
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.05	
High point	4921	78.8	16.87	16.83	1.002
Mid point	4961	39.4	8.43	8.37	1.007
Low point	4980	19.7	4.22	4.14	1.019
As left zero	5000	0.0	0.00	-0.01	----
As left span	4921	78.8	16.87	16.83	1.002

Average Correction Factor 1.009

Notes: Install calibration completed following the station move from Kirby North. Swapped internal pump.
Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	NA
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	15:40	End Time (MST):	19:05
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1182340005

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.05	----	Correlation Coefficient	0.999951	≥ 0.995
16.87	16.83	1.0022	Slope	0.996718	$0.90 - 1.10$
8.43	8.37	1.0074	Intercept	-0.008033	± 1.5
4.22	4.14	1.0188			


THC Calibration Curve



THC Calibration Plot

Date: July 9, 2025

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Kirby South
Calibration Date: July 16, 2025
Start time (MST): 11:53
Reason: Cylinder Change

Station number: AMS 507
Last Cal Date: July 9, 2025
End time (MST): 13:36

Calibration Standards

Gas Cert Reference: CC255918
CH4 Cal Gas Conc. 506.4 ppm
C3H8 Cal Gas Conc. 205.0 ppm
Removed Gas Cert:
Removed CH4 Conc. 506.4 ppm
Removed C3H8 Conc. 205.0 ppm
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Expiry Date: October 9, 2032
CH4 Equiv Conc. 1070.2 ppm
Removed Gas Expiry:
CH4 Equiv Conc. 1070.2 ppm
Diff between cyl:
Serial Number: 2445
Serial Number: 880

Analyzer Information

Analyzer make: Thermo 51i-LT
Analyzer Range: 0 - 20 ppm

Analyzer serial #: 1182340005

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA		Background:	NA	1.91
Calibration intercept:	NA		Coefficient:	NA	3.774

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic-AFzero))
					<i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.01	
As found High point	4921	78.8	16.87	17.20	0.981
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.19	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)
					<i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	-0.02	----
As left span	4921	78.8	16.87	17.16	0.983
Average Correction Factor					

Notes:

Swapped hydrogen cylinder. Leak check passed.


Calibration Performed By:

Braiden Boutilier

THC Calibration Plot

Date: July 16, 2025

Location: Kirby South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Kirby South
Station number: AMS 507
Calibration Date: July 9, 2025
Last Cal Date: NA
Start time (MST): 11:30
End time (MST): 18:30
Reason: Install

Calibration Standards

NO Gas Cylinder #: DT0019572
NOX Cal Gas Conc: 60.00 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.00 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T750
ZAG make/model: Teledyne API T751H
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 59.90 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.90 ppm
NO gas Diff:
Serial Number: 282
Serial Number: 530

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
-----------	------------------------------	--------------------------------	---	--	---	--	---	--	--	---

As found zero
AF High point
AF Mid point
AF Low point
New cyl resp

Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>		*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	---	--	--	---	--	--

As Found GPT zero
As found high GPT point
As found mid GPT point
As found low GPT point



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1173480006

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	NA	0.748
NOX coeff or slope:	NA	0.996
NO2 coeff or slope:	NA	1.000

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	NA	7.7
NOX bkgnd or offset:	NA	8.2
Reaction cell Press:	NA	143.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	NA	1.001128
NO _x Cal Offset:	NA	-0.453577
NO Cal Slope:	NA	1.000314
NO Cal Offset:	NA	-0.433592
NO ₂ Cal Slope:	NA	1.008592
NO ₂ Cal Offset:	NA	0.795640

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	-0.2	0.2	----	----
High point	4933	66.8	801.6	800.3	1.3	802.0	800.0	1.2	0.9995	1.0004
Mid point	4967	33.4	400.8	400.1	0.7	401.5	400.4	1.0	0.9983	0.9993
Low point	4983	16.7	200.4	200.1	0.3	199.1	199.0	0.1	1.0065	1.0054
As left zero	5000	0.0	0.0	0.0	0.0	-0.9	-0.5	-0.4	----	----
As left span	4933	66.8	801.6	417.0	384.6	794.2	417.0	377.2	1.0094	1.0000
Average Correction Factor									1.0014	1.0017

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	795.6	414.7	382.2	386.2	0.9897	101.0%
Mid GPT point	795.6	609.1	187.8	190.0	0.9886	101.2%
Low GPT point	795.6	702.5	94.4	97.0	0.9736	102.7%
Average Correction Factor					0.9840	101.6%

Notes: Install completed with station move to Kirby South. Swapped pump and charcoal scrubber. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	NA
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	11:30	End Time (MST):	18:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999995	≥ 0.995
801.6	802.0	0.9995	Slope	1.001128	$0.90 - 1.10$
400.8	401.5	0.9983	Intercept	-0.453577	± 20
200.4	199.1	1.0065			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	NA
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	11:30	End Time (MST):	18:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999983	≥0.995
382.2	386.2	0.9897	Slope	1.008592	0.90 - 1.10
187.8	190.0	0.9886	Intercept	0.795640	+/-20
94.4	97.0	0.9736			





Wood Buffalo Environmental Association

NO Calibration Summary


Station Information

Calibration Date:	July 9, 2025	Previous Calibration:	NA
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	11:30	End Time (MST):	18:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999997	≥ 0.995
800.3	800.0	1.0004	Slope	1.000314	$0.90 - 1.10$
400.1	400.4	0.9993	Intercept	-0.433592	± 20
200.1	199.0	1.0054			


NO Calibration Curve



NO_x Calibration Plot

Date: July 9, 2025

Location: Kirby South





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Station Information

Station Name:	Kirby South	Station Number:	AMS 507
Calibration Date:	July 9, 2025	Prev Cal Date:	NA
Start Time (MST):	12:45	End Time (MST):	15:30
Tower Height (m):	10.0	Reason:	Install

Wind Speed Calibration

Sensor make/model:	Met One 010C-1	Serial Number:	X16479
WS Calibrator:	MetOne 053	Serial Number:	CA 03988

Shaft RPM (Hz)	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.4%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	0.999999	0.999999	≥ 0.9995
Calculated slope	0.998909	0.999615	$0.98 - 1.02$
Calculated intercept	0.030357	-0.025296	± 2

Wind Direction Calibration

Sensor make/model:	Met One 020C-1	Serial Number:	D13601
As Found Declination (deg east of True North):	<u>13</u>	As Left Declination (deg east of True North):	<u>13</u>
Solar noon (MST):	13:28	Calc Declination*:	13 Degrees
WD Calibrator:	Met One 040	* - calculated declination as per NOAA website	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 360° FS) <i>Limit = +/- 1%</i>
10	9.3	-0.2%
90	87.9	-0.6%
180	178.0	-0.6%
270	271.5	0.4%
350	351.7	0.5%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	0.999977	0.999957	≥ 0.9995
Calculated slope	0.995429	0.990175	$0.97 - 1.03$
Calculated intercept	0.501444	2.083423	± 5

Notes: Installation completed. Tower aligned with solar noon. No issues to note.

Calibration Performed By: Braiden Boutilier



End of Report