



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

OCTOBER 2025 MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

November 28, 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01 BERTHA GANTER - FORT MCKAY OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: October 20, 2025 Last Cal Date: September 3, 2025
Start time (MST): 11:45 End time (MST): 14:59
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.21 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC418809
Removed Cal Gas Conc: 49.21 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3565
Zero Air Gen Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1501301448
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001893	1.000408	Backgd or Offset:	21.6	21.6
Calibration intercept:	-0.333051	-0.613104	Coeff or Slope:	0.879	0.879

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4918	81.3	800.3	799.6	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	799.8	Previous response	801.4	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4918	81.3	800.3	800.5	1.000
Mid point	4959	40.7	400.6	399.1	1.004
Low point	4979	20.3	199.8	199.3	1.003
As left zero	5000	0.0	0.0	0.0	----
As left span	4918	81.3	800.3	804.2	0.995

Average Correction Factor: 1.002

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 3, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:45	End Time (MST):	14:59
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999995		≥0.995
800.3	800.5	0.9997	Slope	1.000408		0.90 - 1.10
400.6	399.1	1.0037	Intercept	-0.613104		+/-30
199.8	199.3	1.0026				



SO2 Calibration Plot

Date: October 20, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: October 8, 2025 Last Cal Date: September 8, 2025
Start time (MST): 10:33 End time (MST): 15:17
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: September 5, 2027
Cal Gas Cylinder #: CC738239
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 12113311966
Converter make: CD Nova Converter serial #: 580
Analyzer Range: 0 - 100 ppb Converter Temp: 850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000671	0.997956	Backgd or Offset:	2.15	2.13
Calibration intercept:	-0.317986	-0.297954	Coeff or Slope:	1.131	1.119

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4917	82.6	80.0	80.8	0.988
As found Mid point	4959	41.3	40.0	40.3	0.990
As found Low point	4979	20.7	20.0	20.0	0.997
New cylinder response					
Baseline Corr As found:	80.9	Prev response:	79.70	*% change:	1.5%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.012395	AF Intercept:	-0.178253
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4917	82.6	80.0	79.6	1.005
Mid point	4959	41.3	40.0	39.5	1.012
Low point	4979	20.7	20.0	19.5	1.028
As left zero	5000	0.0	0.0	0.0	----
As left span	4917	82.6	80.0	78.0	1.025
SO2 Scrubber Check	4919	81.3	813.0	0.1	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	1.015
Date of last converter efficiency test:					

Notes: Inlet filter change completed after as founds. Adjusted the span. Scrubber check passed.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	October 8, 2025	Previous Calibration:	September 8, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:33	End Time (MST):	15:17
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	12113311966

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999972		≥ 0.995
80.0	79.6	1.0046	Slope	0.997956		$0.90 - 1.10$
40.0	39.5	1.0121	Intercept	-0.297954		± 3
20.0	19.5	1.0276				


TRS Calibration Curve



TRS Calibration Plot

Date: October 8, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: October 8, 2025 Last Cal Date: September 8, 2025
Start time (MST): 10:33 End time (MST): 15:17
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: September 5, 2027
Cal Gas Cylinder #: CC738239
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 1200326167
Converter make: CD Nova Converter serial #: 2022-221
Analyzer Range: 0 - 100 ppb Converter Temp: 315 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004245	1.005106	Backgd or Offset:	2.01	2.01
Calibration intercept:	-0.318056	-0.098195	Coeff or Slope:	0.977	0.973

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4917	82.6	80.0	81.0	0.984
As found Mid point	4959	41.3	40.0	40.3	0.985
As found Low point	4979	20.7	20.0	19.8	0.997
New cylinder response					
Baseline Corr As found:	81.3	Prev response:	79.98	*% change:	1.6%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.017826	AF Intercept:	-0.418317
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999987	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4917	82.6	80.0	80.2	0.997
Mid point	4959	41.3	40.0	40.2	0.994
Low point	4979	20.7	20.0	20.1	0.997
As left zero	5000	0.0	0.0	0.0	----
As left span	4917	82.6	80.0	79.3	1.008
SO2 Scrubber Check	4919	81.3	813.0	-0.1	----
Date of last scrubber change:	January 25, 2024		Ave Corr Factor		0.996
Date of last converter efficiency test:	November 7, 2024		107.9% efficiency		

Notes: Inlet filter change completed after as founds. Adjusted the span. Scrubber check passed.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	October 8, 2025	Previous Calibration:	September 8, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:33	End Time (MST):	15:17
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999991		≥0.995
80.0	80.2	0.9970	Slope	1.005106		0.90 - 1.10
40.0	40.2	0.9944	Intercept	-0.098195		+/-3
20.0	20.1	0.9970				



H₂S Calibration Plot

Date: October 8, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay
Calibration Date: October 20, 2025
Start time (MST): 11:45
Reason: Routine
Station number: AMS 01
Last Cal Date: September 3, 2025
End time (MST): 14:59

Calibration Standards

Gas Cert Reference: CC418809
CH₄ Cal Gas Conc. 497.2 ppm
C₃H₈ Cal Gas Conc. 205.3 ppm
Removed Gas Cert: NA
Removed CH₄ Conc. 497.2 ppm
Removed C₃H₈ Conc. 205.3 ppm
Diff between cyl (CH₄):
Calibrator Model: Teledyne API T700
Zero Air Gen model: Teledyne API T701
Cal Gas Expiry Date: March 10, 2031
CH₄ Equiv Conc. 1061.8 ppm
Removed Gas Expiry: NA
CH₄ Equiv Conc. 1061.8 ppm
Diff between cyl (THC):
Diff between cyl (NM):
Serial Number: 3565
Serial Number: 146

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm
Analyzer serial #: 1193585648
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.49E-04	2.49E-04	NMHC SP Ratio:	4.95E-05	4.95E-05
CH ₄ Retention time:	15.0	15.0	NMHC Peak Area:	185423	185423
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	17.27	17.19	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.19	Prev response	17.31	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	17.27	17.21	1.004
Mid point	4959	40.7	8.64	8.62	1.003
Low point	4979	20.3	4.31	4.34	0.994
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	17.27	17.26	1.001
Average Correction Factor					1.000

Notes: Changed the inlet filter after as founds. No adjustments made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	9.18	9.15	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.15	Prev response	9.24	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	9.18	9.19	0.999
Mid point	4959	40.7	4.60	4.60	1.000
Low point	4979	20.3	2.29	2.32	0.986
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	9.18	9.24	0.994
Average Correction Factor					0.995

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	8.09	8.04	1.006
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.04	Prev response	8.07	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	8.09	8.02	1.009
Mid point	4959	40.7	4.05	4.02	1.007
Low point	4979	20.3	2.02	2.01	1.002
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	8.09	8.02	1.008
Average Correction Factor					1.006

Calibration Statistics

	<i>Start</i>	<i>Finish</i>
THC Cal Slope:	1.001304	0.995697
THC Cal Offset:	0.024732	0.018346
CH ₄ Cal Slope:	0.998414	0.991109
CH ₄ Cal Offset:	0.001869	0.006464
NMHC Cal Slope:	1.003352	0.999352
NMHC Cal Offset:	0.023863	0.011683

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 3, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:45	End Time (MST):	14:59
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999993	≥ 0.995
17.27	17.21	1.0035	Slope	0.995697	$0.90 - 1.10$
8.64	8.62	1.0026	Intercept	0.018346	± 0.5
4.31	4.34	0.9939			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 3, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:45	End Time (MST):	14:59
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999997	≥ 0.995
8.09	8.02	1.0086	Slope	0.991109	$0.90 - 1.10$
4.05	4.02	1.0066	Intercept	0.006464	± 0.5
2.02	2.01	1.0024			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 3, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:45	End Time (MST):	14:59
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999986	≥ 0.995
9.18	9.19	0.9995	Slope	0.999352	$0.90 - 1.10$
4.60	4.60	1.0000	Intercept	0.011683	± 0.5
2.29	2.32	0.9864			



NMHC Calibration Plot

Date: October 20, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay
 Station number: AMS 01
 Calibration Date: October 14, 2025
 Last Cal Date: September 11, 2025
 Start time (MST): 10:01
 End time (MST): 12:57
 Reason: Removal

Calibration Standards

NO Gas Cylinder #: CC335700
 NOX Cal Gas Conc: 59.40 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 59.40 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T700
 ZAG make/model: Teledyne API T701
 Cal Gas Expiry Date: September 1, 2032
 NO Cal Gas Conc: 59.20 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 59.20 ppm
 NO gas Diff:
 Serial Number: 3565
 Serial Number: 146

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.3	0.8	-0.6	----	----
AF High point	4932	67.6	803.1	800.4	2.7	796.3	791.9	4.4	1.0089	1.0117
AF Mid point	4966	33.8	401.5	400.2	1.4	397.9	396.0	1.9	1.0099	1.0126
AF Low point	4983	16.9	200.8	200.1	0.7	199.6	197.4	2.2	1.0074	1.0178
New cyl resp										
Previous Response	NO _x = 802.9 ppb		NO = 799.1 ppb		* = > +/-5% change initiates investigation			*Percent Change	NO _x = -0.9%	
Baseline Corr 1st pt	NO _x = 796.0 ppb		NO = 791.1 ppb		<u>As Found Statistics</u>			*Percent Change	NO = -1.0%	
Baseline Corr 2nd pt	NO _x = 397.6 ppb		NO = 395.2 ppb		As found	NO _x r ² : 0.999999		Nx SI: 0.990989	Nx Int: 0.340	
Baseline Corr 3rd pt	NO _x = 199.3 ppb		NO = 196.6 ppb		As found	NO r ² : 0.999998		NO SI: 0.988854	NO Int: 0.260	
					As found	NO ₂ r ² : 0.999994		NO2 SI: 0.998455	NO ₂ Int: -1.048	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	-0.6	----	----
As found high GPT point	792.8	388.3	407.2	405.8	1.0035	99.7%
As found mid GPT point	792.8	586.0	209.5	207.8	1.0082	99.2%
As found low GPT point	792.8	688.3	107.2	105.6	1.0152	98.5%



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 7117

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.198	
NOX coeff or slope:	1.200	
NO2 coeff or slope:	1.000	

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	-4.1	
NOX bkgnd or offset:	-3.1	
Reaction cell Press:	8.5	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999542	
NO _x Cal Offset:	0.160000	
NO Cal Slope:	0.999277	
NO Cal Offset:	-0.740000	
NO ₂ Cal Slope:	1.002291	
NO ₂ Cal Offset:	0.364020	

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero										
High point										
Mid point										
Low point										
As left zero										
As left span										

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero						
High GPT point						
Mid GPT point						
Low GPT point						

Average Correction Factor


Notes: Removing the instrument. Performing as founds.

Calibration Performed By: Rene Chamberland

NO_x Calibration Plot

Date: October 14, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay
Station number: AMS 01
Calibration Date: October 15, 2025
Last Cal Date: NA
Start time (MST): 10:50
End time (MST): 15:37
Reason: Install

Calibration Standards

NO Gas Cylinder #: CC335700
NOX Cal Gas Conc: 59.40 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.40 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: September 1, 2032
NO Cal Gas Conc: 59.20 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.20 ppm
NO gas Diff:
Serial Number: 3565
Serial Number: 146

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x =	NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>			*Percent Change	NO =	NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found NO _x r ² :			Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found NO r ² :			NO SI:	NO Int:	
					As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	---	--	--	---	--	--

As Found GPT zero
As found high GPT point
As found mid GPT point
As found low GPT point



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12400232072

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:		0.819	NO bkgnd or offset:		0.6
NOX coeff or slope:		0.994	NOX bkgnd or offset:		0.7
NO2 coeff or slope:		1.000	Reaction cell Press:		126.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:		0.999769
NO _x Cal Offset:		0.980000
NO Cal Slope:		1.002890
NO Cal Offset:		-0.380000
NO ₂ Cal Slope:		0.984647
NO ₂ Cal Offset:		1.657543

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4932	67.6	803.1	800.4	2.7	802.9	802.0	0.8	1.0002	0.9980
Mid point	4966	33.8	401.5	400.2	1.4	404.4	402.3	2.1	0.9929	0.9948
Low point	4983	16.9	200.8	200.1	0.7	201.7	198.9	2.7	0.9954	1.0060
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
As left span	4932	67.6	803.1	400.3	402.8	792.1	400.3	391.8	1.0139	1.0000
Average Correction Factor									0.9962	0.9996

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	797.7	400.9	399.5	393.7	1.0147	98.5%
Mid GPT point	797.7	588.0	212.4	212.8	0.9981	100.2%
Low GPT point	797.7	693.1	107.3	108.3	0.9908	100.9%
Average Correction Factor					1.0012	99.9%

Notes: Installing a new instrument. Changed the inlet filter and sample pump. Adjusted the zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	NA
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:50	End Time (MST):	15:37
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232072

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999984		<i>≥0.995</i>
803.1	802.9	1.0002	Slope	0.999769		<i>0.90 - 1.10</i>
401.5	404.4	0.9929	Intercept	0.980000		<i>+/-20</i>
200.8	201.7	0.9954				





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	NA
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:50	End Time (MST):	15:37
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232072

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999888		<i>≥0.995</i>
399.5	393.7	1.0147	Slope	0.984647		<i>0.90 - 1.10</i>
212.4	212.8	0.9981	Intercept	1.657543		<i>+/-20</i>
107.3	108.3	0.9908				





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	NA
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:50	End Time (MST):	15:37
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232072

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999989		≥ 0.995
800.4	802.0	0.9980	Slope	1.002890		$0.90 - 1.10$
400.2	402.3	0.9948	Intercept	-0.380000		± 20
200.1	198.9	1.0060				



NO_x Calibration Plot

Date: October 15, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: October 6, 2025 Last Cal Date: September 2, 2025
Start time (MST): 10:38 End time (MST): 13:27
Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 1107
Analyzer Range: 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001429	1.001714	Backgd or Offset:	5.6	5.6
Calibration intercept:	-0.400000	-0.800000	Coeff or Slope:	1.006	1.006

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/Ic-AFzero) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.8	----
As found High point	5000	863.1	400.0	399.6	0.999
As found Mid point					
As found Low point					
Baseline Corr As found:	400.4	Previous response	400.2	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	5000	863.1	400.0	400.2	1.000
Mid point	5000	744.0	200.0	199.3	1.004
Low point	5000	651.7	100.0	98.6	1.014
As left zero	5000	0.0	0.0	0.1	----
As left span	5000	863.1	400.0	403.6	0.991
Average Correction Factor					1.006

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	October 6, 2025	Previous Calibration:	September 2, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:38	End Time (MST):	13:27
Analyzer make:	Teledyne API T400	Analyzer serial #:	1107

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999986		≥0.995
400.0	400.2	0.9995	Slope	1.001714		0.90 - 1.10
200.0	199.3	1.0035	Intercept	-0.800000		+/- 5
100.0	98.6	1.0142				



O₃ Calibration Plot

Date: October 6, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
Calibration Date: October 22, 2025 Last Cal Date: September 10, 2025
Start time (MST): 11:28 End time (MST): 13:27

Analyzer Make: Teledyne API T640 S/N: 322
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	7.3	7.6	7.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	734.6	737.19	734.6	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	5.428	4.927	<input checked="" type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37		37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA:	7.1	PM w/ HEPA:	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.1	10.7	10.7	<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: October 22, 2025
Date Disposable Filter Changed: August 20, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 20, 2025
Date RH/T Sensor Cleaned: October 22, 2025

Notes: Flow, temperature, and pressure were verified. Flow was adjusted. Leak check passed. Optical chamber and RH/T sensor cleaned. PMT peak test verified.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

Nt - NOX - NH3 Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
NOX Cal Date:	October 9, 2025	Last Cal Date:	September 4, 2025
Start time (MST):	10:32	End time (MST):	15:21
NH3 Cal Date:	October 10, 2025	Last Cal Date:	September 5, 2025
Start time (MST):	10:19	End time (MST):	14:08
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	59.40	ppm	NO Gas Cylinder #:	CC335700
NO Cal Gas Conc:	59.20	ppm	NO Cal Gas Expiry:	September 1, 2032
Removed NOX Conc:	59.40	ppm	Removed Cylinder #:	NA
Removed NO Conc:	59.20	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	77.80	ppm	NH3 Gas Cylinder #:	CC711249
			NH3 Cal Gas Expiry:	December 31, 2025
Removed NH3 Conc:	77.80	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	API T700		Serial Number:	3565
ZAG make/model:	API T701		Serial Number:	146

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	475
Converter model:	API T501	Converter serial #:	824
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.20
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	526

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	0.984	0.961	Nt coefficient:	0.997	0.975
NOX coefficient:	0.989	0.967	NO bkgnd:	-3.0	-3.0
NO2 coefficient:	1.000	1.000	NOX bkgnd:	-2.8	-2.8
NH3 coefficient:	0.987	0.987	Nt bkgnd:	-1.9	-1.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997535	0.998731
NO _x Cal Offset:	-0.960000	-1.880000
NO Cal Slope:	0.998878	0.999249
NO Cal Offset:	-2.800000	-2.780000
NO ₂ Cal Slope:	0.997147	1.004249
NO ₂ Cal Offset:	-0.284956	0.465952
NH3 Cal Slope:	0.998912	1.000028
NH3 Cal Offset:	6.194453	3.573310
Nt Cal Slope:	1.002428	1.003803
Nt Cal Offset:	6.567914	3.958139



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

NO_x / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO _x Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>	Baseline corr NO Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.7	0.7	0.6	----	----
As found span	4932	67.6	803.2	800.4	803.2	824.6	815.9	824.6	0.9740	0.9811
AF GPT span	4932	67.6	803.2	-----	803.2	804.2	-----	803.3	0.9987	-----
new NO cyl rp										

Baseline Corr As Fd	Nt = 824 ppb	NO _x = 823.9 ppb	NO = 815.2 ppb	*Percent Change	Nt _(NO) = 1.5%
Previous Response	Nt = 811.67 ppb	NO _x = 800.2 ppb	NO = 796.7 ppb	*Percent Change	NO _x = 2.9%
**NO _x Δ (NO to GPT response) = -2.5%				*Percent Change	NO = 2.3%
* = > +/-2% difference initiates investigation				* = > +/-5% change initiates investigation	

NO_x / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	-0.2	----	----
High point	4932	67.6	803.1	800.4	803.1	801.4	798.6	802.0	1.0021	1.0022
Mid point	4966	33.8	401.5	400.2	401.5	397.3	395.0	398.1	1.0107	1.0131
Low point	4983	16.9	200.8	200.1	200.8	197.5	195.0	198.8	1.0166	1.0261
Average Correction Factor									1.0098	1.0138

GPT Calibration Data

O ₃ Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found zero	----	----	0.0	0.0	----	----
Calibration zero	----	----	0.0	0.0	----	----
High GPT point (400 ppb O ₃)	791.6	392.2	402.1	404.3	0.9946	100.5%
Mid GPT point (200 ppb O ₃)	791.6	588.8	205.5	206.3	0.9961	100.4%
Low GPT point (100 ppb O ₃)	791.6	690.1	104.2	106.1	0.9821	101.8%
Average Correction Factor					0.9909	100.9%



Wood Buffalo Environmental Association NH₃ - N_T Calibration Report

NH₃ As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH3 Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.1	-1.1	-0.1	----	----
AF High point	2931	69.4	1799.5	0.0	1799.5	1808.2	6.9	1801.3	0.995	0.999
AF Mid point										
AF Low point										
new NH3 cyl rp										
Baseline Corr As Fd	Nt = 1809.3 ppb		NH3 = 1801.4 ppb					*Percent Change	Nt _(NH3) = -0.1%	
Previous Response	Nt = 1810.5 ppb		NH3 = 1803.8 ppb			* = > +/-5% change initiates investigation		*Percent Change	NH3 = -0.1%	

NH₃ Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
High point	2931	69.4	1799.8	0.0	1799.8	1808.2	6.9	1801.3	0.995	0.999
Mid point	2961	38.6	1001.0	0.0	1001.0	1010.8	4.3	1006.4	0.990	0.995
Low point	2981	19.3	500.5	0.0	500.5	510.9	2.8	508.1	0.980	0.985
Average Correction Factor									0.9884	0.9930
NH3 Previous Converter Efficiency =		98.7 %								
NH3 Current Converter Efficiency =		98.7 %								

Notes:

Changed the inlet filter after as founds. Adjusted the NOx, NO, NT span.

Calibration Performed By:

Rene Chamberland



Wood Buffalo Environmental Association


Nt Calibration Summary

Station Information

Calibration Date:	October 10, 2025	Previous Calibration:	September 4, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:32	End Time (MST):	15:21
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999973	≥ 0.995
1799.8	1808.2	0.9953	Slope	1.003803	0.90 - 1.10
1001.0	1010.8	0.9903	Intercept	3.958139	+/-20
500.5	510.9	0.9797			





Wood Buffalo Environmental Association


NH₃ Calibration Summary

Station Information

Calibration Date:	October 10, 2025	Previous Calibration:	September 4, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:32	End Time (MST):	15:21
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999979	≥0.995
1799.8	1801.3	0.9992	Slope	1.000028	0.90 - 1.10
1001.0	1006.4	0.9947	Intercept	3.573310	+/-20
500.5	508.1	0.9851			





Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 9, 2025	Previous Calibration:	September 4, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:32	End Time (MST):	15:21
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999973	≥0.995
803.1	801.4	1.0021	Slope	0.998731	0.90 - 1.10
401.5	397.3	1.0107	Intercept	-1.880000	+/-20
200.8	197.5	1.0166			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 9, 2025	Previous Calibration:	September 4, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:32	End Time (MST):	15:21
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999946	≥ 0.995
800.4	798.6	1.0022	Slope	0.999249	$0.90 - 1.10$
400.2	395.0	1.0131	Intercept	-2.780000	± 20
200.1	195.0	1.0261			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 9, 2025	Previous Calibration:	September 4, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:32	End Time (MST):	15:21
Analyzer make:	API T201	Analyzer serial #:	475

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999983	≥ 0.995
402.1	404.3	0.9946	Slope	1.004249	0.90 - 1.10
205.5	206.3	0.9961	Intercept	0.465952	+/-20
104.2	106.1	0.9821			



NO_x Calibration Plot

Date: October 9, 2025


Location: Bertha Ganter-Fort McKay



NH₃ Calibration Plot

Date: October 10, 2025

Location: Bertha Ganter-Fort McKay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Mildred Lake Station number: AMS 02
Calibration Date: October 30, 2025 Last Cal Date: September 8, 2025
Start time (MST): 10:08 End time (MST): 14:44
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.99 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: EB0112903
Removed Cal Gas Conc: 50.99 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 1185
Zero Air Gen Model: Teledyne API T701 Serial Number: 4891

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1404901075
Analyzer Range: 0-1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007336	1.001740	Backgd or Offset:	24.8	24.2
Calibration intercept:	-1.470446	-0.210939	Coeff or Slope:	0.762	0.762

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.6	----
As found High point	4913	78.6	803.0	801.0	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	801.6	Previous response	807.4	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4913	78.4	801.0	802.0	0.999
Mid point	4961	39.2	399.8	401.1	0.997
Low point	4980	19.6	199.9	198.9	1.005
As left zero	5000	0.0	0.0	0.1	----
As left span	4913	78.4	801.0	804.0	0.996
Average Correction Factor:					1.000

Notes: Changed sample inlet filter after as founds. Adjusted zero.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 30, 2025	Previous Calibration:	September 8, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:08	End Time (MST):	14:44
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999994	≥0.995
801.0	802.0	0.9987	Slope	1.001740	0.90 - 1.10
399.8	401.1	0.9967	Intercept	-0.210939	+/-30
199.9	198.9	1.0049			



SO2 Calibration Plot

Date: October 30, 2025

Location: Mildred Lake





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Mildred Lake
Calibration Date: October 22, 2025
Start time (MST): 9:10
Reason: Routine

Station number: AMS 02
Last Cal Date: September 25, 2025
End time (MST): 13:29

Calibration Standards

Cal Gas Concentration: 4.75 ppm
Cal Gas Cylinder #: CC700774
Removed Cal Gas Conc: 4.75 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 1185
Serial Number: 4891

Analyzer Information

Analyzer make: Thermo 43iQTL
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12333331546
Converter serial #: 2023-267
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997268	1.000125	Backgd or Offset:	1.39	1.39
Calibration intercept:	0.100000	0.000000	Coeff or Slope:	0.957	0.957

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4916	84.2	80.0	80.4	0.996
As found Mid point	4958	42.1	40.0	40.5	0.990
As found Low point	4979	21.1	20.0	20.1	1.000
New cylinder response					
Baseline Corr As found:	80.3	Prev response:	79.87	*% change:	0.5%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.004554	AF Intercept:	0.120000
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999983	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4916	84.2	80.0	80.1	0.999
Mid point	4958	42.1	40.0	39.8	1.005
Low point	4979	21.1	20.0	20.0	1.000
As left zero	5000	0.0	0.0	0.1	----
As left span	4916	84.2	80.0	80.4	0.995
SO2 Scrubber Check	4922	78.4	783.9	0.0	----
Date of last scrubber change:		July 16, 2024		Ave Corr Factor	1.001
Date of last converter efficiency test:		NA			

Notes: Changed sample inlet filter after multipoint as founds. SO2 scrubber check done and passed. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	October 22, 2025	Previous Calibration:	September 25, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	9:10	End Time (MST):	13:29
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331546

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999983		≥ 0.995
80.0	80.1	0.9986	Slope	1.000125		$0.90 - 1.10$
40.0	39.8	1.0049	Intercept	0.000000		± 3
20.0	20.0	0.9999				



H2S Calibration Plot

Date: October 22, 2025

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Mildred Lake
 Calibration Date: October 30, 2025
 Start time (MST): 10:08
 Reason: Routine

Station number: AMS 02
 Last Cal Date: September 24, 2025
 End time (MST): 14:44

Calibration Standards

Gas Cert Reference:	EB0112903	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	503.1 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	503.1 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
Zero Air Gen model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1180320040
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	4.77E-04	5.10E-04	NMHC SP Ratio:	7.77E-05
CH ₄ Retention time:	15.9	15.9	NMHC Peak Area:	113819
Zero Chromatogram:	ON	ON	Flat Baseline:	ON

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4922	78.4	16.73	16.13	1.037
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.13	Prev response	16.61	*% change	-3.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	
High point	4922	78.4	16.73	16.75	0.999
Mid point	4961	39.2	8.37	8.42	0.994
Low point	4980	19.6	4.18	4.10	1.021
As left zero	5000	0.0	0.00	0.00	----
As left span	4922	78.4	16.73	16.80	0.996
Average Correction Factor					1.004

Notes: Changed sample inlet filter and nitrogen cylinder after as founds. Adjusted zero and span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4922	78.4	8.84	8.82	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.82	Prev response	8.78	*% change	0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4922	78.4	8.84	8.81	1.004
Mid point	4961	39.2	4.42	4.45	0.993
Low point	4980	19.6	2.21	2.17	1.021
As left zero	5000	0.0	0.00	0.00	----
As left span	4922	78.4	8.84	8.78	1.007
Average Correction Factor					1.006

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4922	78.4	7.89	7.31	1.079
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.31	Prev response	7.83	*% change	-7.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4922	78.4	7.89	7.94	0.993
Mid point	4961	39.2	3.94	3.97	0.994
Low point	4980	19.6	1.97	1.93	1.020
As left zero	5000	0.0	0.00	0.00	----
As left span	4922	78.4	7.89	8.02	0.984
Average Correction Factor					1.003

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.994593	1.003248
THC Cal Offset:	-0.031065	-0.026669
CH ₄ Cal Slope:	0.997093	1.008814
CH ₄ Cal Offset:	-0.032325	-0.021527
NMHC Cal Slope:	0.992144	0.998283
NMHC Cal Offset:	0.002860	-0.005142

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 30, 2025	Previous Calibration:	September 24, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:08	End Time (MST):	14:44
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999944	≥ 0.995
16.73	16.75	0.9989	Slope	1.003248	$0.90 - 1.10$
8.37	8.42	0.9936	Intercept	-0.026669	± 0.5
4.18	4.10	1.0205			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 30, 2025	Previous Calibration:	September 24, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:08	End Time (MST):	14:44
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999948	≥ 0.995
7.89	7.94	0.9934	Slope	1.008814	0.90 - 1.10
3.94	3.97	0.9943	Intercept	-0.021527	± 0.5
1.97	1.93	1.0203			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 30, 2025	Previous Calibration:	September 24, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:08	End Time (MST):	14:44
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999920	<i>≥0.995</i>
8.84	8.81	1.0038	Slope	0.998283	<i>0.90 - 1.10</i>
4.42	4.45	0.9930	Intercept	-0.005142	<i>+/-0.5</i>
2.21	2.17	1.0208			



NMHC Calibration Plot

Date: October 30, 2025

Location: Mildred Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
Calibration Date: October 22, 2025 Last Cal Date: September 8, 2025
Start time (MST): 6:43 End time (MST): 8:05
Reason: Removal

Calibration Standards

Cal Gas Concentration: 50.87 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC446753
Removed Cal Gas Conc: 50.87 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3808
Zero Air Gen Model: Teledyne API T701 Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1327300932
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000929		Backgd or Offset:	27.5	27.5
Calibration intercept:	0.454838		Coeff or Slope:	0.875	0.875

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.6	----
As found High point	4921	78.6	799.7	792.4	1.010
As found Mid point	4961	39.3	399.8	396.6	1.010
As found Low point	4980	19.6	199.4	198.6	1.007
New cylinder response					
Baseline Corr As found:	791.8	Previous response	800.9	*% change	-1.2%
Baseline Corr 2nd AF pt:	396.0	AF Slope:	0.989856	AF Intercept:	0.852689
Baseline Corr 3rd AF pt:	198.0	AF Correlation:	0.999999	* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor:

Notes: Removed Due to lamp intensity not being stable and going from 83-102%.

Calibration Performed By: Melissa Lemay

SO2 Calibration Plot

Date: October 22, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: October 22, 2025
Start time (MST): 8:40
Reason: Install

Station number: AMS 04
Last Cal Date:
End time (MST): 10:44

Calibration Standards

Cal Gas Concentration: 50.87 ppm
Cal Gas Cylinder #: CC446753
Removed Cal Gas Conc: 50.87 ppm
Removed Gas Cyl #:
Calibrator Model: Teledyne API T700
Zero Air Gen Model: Teledyne API T701

Cal Gas Exp Date: March 10, 2031
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i-LTE
Analyzer Range: 0-1000ppb

Serial Number: 1410661331

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		0.997583	Backgd or Offset:		1.9
Calibration intercept:		1.875213	Coeff or Slope:		1.013

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/Ic-AFzero) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4921	78.6	799.7	798.9	1.001
Mid point	4961	39.3	399.8	401.3	0.996
Low point	4980	19.6	199.4	202.8	0.983
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	78.6	799.7	799.8	1.000
Average Correction Factor:					0.994

Notes: Install Due to old Thermo 43i not working properly. Zero and Span Adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 22, 2025	Previous Calibration:	
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	8:40	End Time (MST):	10:44
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1410661331

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999977	≥0.995
799.7	798.9	1.0011	Slope	0.997583	0.90 - 1.10
399.8	401.3	0.9963	Intercept	1.875213	+/-30
199.4	202.8	0.9834			



SO2 Calibration Plot

Date: October 22, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: October 3, 2025
Start time (MST): 5:55
Reason: Routine

Station number: AMS 04
Last Cal Date: September 3, 2025
End time (MST): 10:22

Calibration Standards

Cal Gas Concentration: 4.80 ppm
Cal Gas Cylinder #: DT0037528
Removed Cal Gas Conc: 4.80 ppm
Removed Gas Cyl #:
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i-LTE
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1008841400
Converter serial #: 2022-200
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998433	0.993002	Backgd or Offset:	1.86	1.86
Calibration intercept:	0.138223	-0.021732	Coeff or Slope:	1.077	1.077

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4917	83.3	80.0	80.7	0.991
As found Mid point	4958	41.7	40.0	40.4	0.991
As found Low point	4979	20.8	20.0	20.1	0.993
New cylinder response					
Baseline Corr As found:	80.7	Prev response:	79.98	*% change:	0.9%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.009436	AF Intercept:	-0.021789
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999999	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4917	83.3	80.0	79.4	1.007
Mid point	4958	41.7	40.0	39.7	1.008
Low point	4979	20.8	20.0	19.8	1.009
As left zero	5000	0.0	0.0	0.1	----
As left span	4917	83.3	80.0	78.0	1.025
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23		Ave Corr Factor		1.008
Date of last converter efficiency test:					

Notes: Sox scrubber checked after calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	October 3, 2025	Previous Calibration:	September 3, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:55	End Time (MST):	10:22
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999999		≥0.995
80.0	79.4	1.0071	Slope	0.993002		0.90 - 1.10
40.0	39.7	1.0084	Intercept	-0.021732		+/-3
20.0	19.8	1.0085				



H₂S Calibration Plot

Date: October 3, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Buffalo Viewpoint
 Calibration Date: October 7, 2025
 Start time (MST): 6:00
 Reason: Routine

Station number: AMS 04
 Last Cal Date: September 8, 2025
 End time (MST): 8:48

Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3808
Zero Air Gen model:	Teledyne API T701	Serial Number:	362

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1180320038
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.35E-04	3.41E-04	NMHC SP Ratio:	5.73E-04
CH ₄ Retention time:	15.4	15.2	NMHC Peak Area:	153813
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	16.64	16.42	1.013
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.42	Prev response	16.62	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	16.64	16.65	0.999
Mid point	4961	39.3	8.32	8.26	1.007
Low point	4980	19.6	4.15	4.12	1.008
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	16.64	16.64	1.000
Average Correction Factor					1.005

Notes:

Span adjusted. No maintenance done.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	8.82	8.71	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.71	Prev response	8.82	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	8.82	8.82	1.001
Mid point	4961	39.3	4.41	4.40	1.003
Low point	4980	19.6	2.20	2.20	0.998
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	8.82	8.80	1.002
Average Correction Factor					1.000

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	7.82	7.70	1.015
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.70	Prev response	7.80	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	7.82	7.83	0.998
Mid point	4961	39.3	3.91	3.86	1.012
Low point	4980	19.6	1.95	1.91	1.019
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	7.82	7.84	0.997
Average Correction Factor					1.010

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.001007	1.001164
THC Cal Offset:	-0.029740	-0.027136
CH ₄ Cal Slope:	1.001396	1.003179
CH ₄ Cal Offset:	-0.025908	-0.027504
NMHC Cal Slope:	1.000584	0.999132
NMHC Cal Offset:	-0.004032	0.000568

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 7, 2025	Previous Calibration:	September 8, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	6:00	End Time (MST):	8:48
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999980	≥ 0.995
16.64	16.65	0.9991	Slope	1.001164	$0.90 - 1.10$
8.32	8.26	1.0070	Intercept	-0.027136	± 0.5
4.15	4.12	1.0076			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 7, 2025	Previous Calibration:	September 8, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	6:00	End Time (MST):	8:48
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999927	<i>≥0.995</i>
7.82	7.83	0.9978	Slope	1.003179	<i>0.90 - 1.10</i>
3.91	3.86	1.0124	Intercept	-0.027504	<i>+/-0.5</i>
1.95	1.91	1.0189			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 7, 2025	Previous Calibration:	September 8, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	6:00	End Time (MST):	8:48
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999998	<i>≥0.995</i>
8.82	8.82	1.0005	Slope	0.999132	<i>0.90 - 1.10</i>
4.41	4.40	1.0025	Intercept	0.000568	<i>+/-0.5</i>
2.20	2.20	0.9979			



NMHC Calibration Plot

Date: October 7, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: October 23, 2025
Start time (MST): 7:50
Reason: Cylinder Change

Station number: AMS 04
Last Cal Date: October 7, 2025
End time (MST): 9:20

Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3808
Zero Air Gen model:	Teledyne API T701	Serial Number:	362

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1180320038
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.41E-04	3.41E-04	NMHC SP Ratio:	5.81E-04
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	151711
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	16.64	16.56	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.56	Prev response	16.62	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	16.64	16.51	1.007
Average Correction Factor					

Notes:

Hydrogen Cylinder Change.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	8.82	8.78	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.78	Prev response	8.82	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	8.82	8.73	1.011
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	7.82	7.79	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.79	Prev response	7.80	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	7.82	7.79	1.003
Average Correction Factor					

Calibration Statistics

	<u>Start</u>	
THC Cal Slope:	1.001007	
THC Cal Offset:	-0.029740	
CH ₄ Cal Slope:	1.001396	
CH ₄ Cal Offset:	-0.025908	
NMHC Cal Slope:	1.000584	
NMHC Cal Offset:	-0.004032	


Finish

Calibration Performed By: Melissa Lemay

NMHC Calibration Plot

Date: October 23, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 721

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.920	0.927	NO bkgnd or offset:	-9.8	-0.6
NOX coeff or slope:	0.910	0.918	NOX bkgnd or offset:	-9.7	-0.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.8	4.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001578	0.998122
NO _x Cal Offset:	1.187429	0.646580
NO Cal Slope:	1.001914	0.998738
NO Cal Offset:	-0.113474	-0.754295
NO ₂ Cal Slope:	0.993706	0.992875
NO ₂ Cal Offset:	-1.794650	0.593151

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.5	0.1	----	----
High point	4918	81.8	800.0	798.4	1.6	798.2	796.5	1.7	1.0023	1.0024
Mid point	4959	40.9	400.0	399.2	0.8	401.7	398.6	3.2	0.9958	1.0015
Low point	4980	20.4	199.5	199.1	0.4	200.1	197.3	2.7	0.9970	1.0091
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	0.0	----	----
As left span	4918	81.8	800.0	402.0	800.0	788.3	402.0	386.3	1.0149	1.0000
Average Correction Factor									0.9984	1.0043

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	793.0	401.4	393.2	390.1	1.0080	99.2%
Mid GPT point	793.0	600.8	193.8	195.4	0.9920	100.8%
Low GPT point	793.0	696.7	97.9	96.9	1.0107	98.9%
Average Correction Factor					1.0036	99.7%

Notes: No Maintenance done. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 1, 2025	Previous Calibration:	September 25, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:53	End Time (MST):	10:46
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.5	----	Correlation Coefficient	0.999984	≥0.995
800.0	798.2	1.0023	Slope	0.998122	0.90 - 1.10
400.0	401.7	0.9958	Intercept	0.646580	+/-20
199.5	200.1	0.9970			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 1, 2025	Previous Calibration:	September 25, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:53	End Time (MST):	10:46
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999910	≥0.995
393.2	390.1	1.0080	Slope	0.992875	0.90 - 1.10
193.8	195.4	0.9920	Intercept	0.593151	+/-20
97.9	96.9	1.0107			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 1, 2025	Previous Calibration:	September 25, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	5:53	End Time (MST):	10:46
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.5	----	Correlation Coefficient	0.999997	≥ 0.995
798.4	796.5	1.0024	Slope	0.998738	$0.90 - 1.10$
399.2	398.6	1.0015	Intercept	-0.754295	± 20
199.1	197.3	1.0091			



NO_x Calibration Plot

Date: October 1, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: October 7, 2025
Start time (MST): 8:47
Reason: Routine

Station number: AMS 04
Last Cal Date: September 26, 2025
End time (MST): 11:02

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: Teledyne API T400
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 2961

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999114	1.002086	Backgd or Offset:	-1.0	-1.0
Calibration intercept:	0.880000	0.060000	Coeff or Slope:	1.022	1.022

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	5000	1013.8	400.0	400.2	0.999
As found Mid point					
As found Low point					
Baseline Corr As found:	400.4	Previous response	400.5	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	5000	1011.7	400.0	401.0	0.998
Mid point	5000	832.5	200.0	200.3	0.999
Low point	5000	719.2	100.0	100.2	0.998
As left zero	5000	0.0	0.0	-0.4	----
As left span	5000	1013.5	400.0	403.0	0.993
Average Correction Factor					0.998

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	October 7, 2025	Previous Calibration:	September 26, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	8:47	End Time (MST):	11:02
Analyzer make:	Teledyne API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999999	≥0.995
400.0	401.0	0.9975	Slope	1.002086	0.90 - 1.10
200.0	200.3	0.9985	Intercept	0.060000	+/- 5
100.0	100.2	0.9980			



O₃ Calibration Plot

Date: October 7, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
Calibration Date: October 23, 2025 Last Cal Date: September 26, 2025
Start time (MST): 7:16 End time (MST): 7:38

Analyzer Make: Teledyne API T640 S/N: 321
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	5.6	5.2	5.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	724.4	726.3	724.4	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.06	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	----	37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 3.8		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 30-Jan-27
Lot No.: 100128-050-051

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: September 26, 2025
Date Disposable Filter Changed: September 26, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: September 26, 2025
Date RH/T Sensor Cleaned: September 26, 2025

No adjustments done.

Notes:

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS05
MANNIX
OCTOBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Mannix
Calibration Date: October 17, 2025
Start time (MST): 10:42
Reason: Routine

Station number: AMS 05
Last Cal Date: September 15, 2025
End time (MST): 14:15

Calibration Standards

Cal Gas Concentration: 50.06 ppm
Cal Gas Cylinder #: CC308040
Removed Cal Gas Conc: 50.06 ppm
Removed Gas Cyl #:
Calibrator Model: API T700
Zero Air Gen Model: API T701

Cal Gas Exp Date: October 22, 2032
Rem Gas Exp Date: October 22, 2032
Diff between cyl:
Serial Number: 5470
Serial Number: 361

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 1000 ppb

Serial Number: 1008841399

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001595	1.001237	Backgd or Offset:	10.1	11.1
Calibration intercept:	-0.657151	-0.856898	Coeff or Slope:	0.939	0.946

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4920	79.9	800.0	799.9	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	800.0	Previous response	800.6	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	79.9	800.0	800.5	0.999
Mid point	4960	40.0	400.5	400.0	1.001
Low point	4980	20.0	200.2	198.3	1.010
As left zero	5000	0.0	0.0	0.3	----
As left span	4920	79.9	800.0	799.9	1.000
Average Correction Factor:					1.003

Notes: Changed the inlet filter after as founds. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 15, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:42	End Time (MST):	14:15
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999991	≥0.995
800.0	800.5	0.9993	Slope	1.001237	0.90 - 1.10
400.5	400.0	1.0012	Intercept	-0.856898	+/-30
200.2	198.3	1.0098			



SO2 Calibration Plot

Date: October 17, 2025

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Mannix
Calibration Date: October 6, 2025
Start time (MST): 9:54
Reason: Routine

Station number: AMS 05
Last Cal Date: September 3, 2025
End time (MST): 14:10

Calibration Standards

Cal Gas Concentration: 4.96 ppm
Cal Gas Cylinder #: DT0037363
Removed Cal Gas Conc: 4.96 ppm
Removed Gas Cyl #: N/A
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: N/A
Diff between cyl:
Serial Number: 5470
Serial Number: 361

Analyzer Information

Analyzer make: Thermo 43iQ
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1200326169
Converter serial #: 2022-225
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005691	1.005833	Backgd or Offset:	1.25	1.25
Calibration intercept:	0.082267	0.002268	Coeff or Slope:	1.017	1.017

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4919	80.6	80.0	81.2	0.986
As found Mid point	4960	40.3	40.0	40.6	0.987
As found Low point	4980	20.2	20.0	20.1	1.002
New cylinder response					
Baseline Corr As found:	81.1	Prev response:	80.50	*% change:	0.7%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	1.015555	AF Intercept:	-0.037934
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999982	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4919	80.6	80.0	80.4	0.995
Mid point	4960	40.3	40.0	40.4	0.989
Low point	4980	20.2	20.0	19.9	1.007
As left zero	5000	0.0	0.0	0.2	----
As left span	4919	80.6	80.0	79.3	1.008
SO2 Scrubber Check	4920	80.3	803.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	0.997
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	October 6, 2025	Previous Calibration:	September 3, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	9:54	End Time (MST):	14:10
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1200326169

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999968		≥0.995
80.0	80.4	0.9945	Slope	1.005833		0.90 - 1.10
40.0	40.4	0.9895	Intercept	0.002268		+/-3
20.0	19.9	1.0069				



H₂S Calibration Plot

Date: October 6, 2025

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Mannix
 Calibration Date: October 17, 2025
 Start time (MST): 10:42
 Reason: Routine

Station number: AMS 05
 Last Cal Date: September 15, 2025
 End time (MST): 14:15

Calibration Standards

Gas Cert Reference: CC308040
 CH₄ Cal Gas Conc. 500.3 ppm
 C₃H₈ Cal Gas Conc. 199.0 ppm
 Removed Gas Cert:
 Removed CH₄ Conc. 500.3 ppm
 Removed C₃H₈ Conc. 199.0 ppm
 Diff between cyl (CH₄):
 Calibrator Model: API T700
 Zero Air Gen model: API T701

Cal Gas Expiry Date: October 22, 2032
 CH₄ Equiv Conc. 1047.6 ppm
 Removed Gas Expiry:
 CH₄ Equiv Conc. 1047.6 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 5470
 Serial Number: 361

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 15005164381
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.68E-04	2.61E-04	NMHC SP Ratio:	5.04E-05	4.91E-05
CH ₄ Retention time:	13.9	13.7	NMHC Peak Area:	173725	173301
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.9	16.74	17.14	0.977
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.14	Prev response	16.78	*% change	2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.9	16.74	16.69	1.003
Mid point	4960	40.0	8.38	8.41	0.997
Low point	4980	20.0	4.19	4.18	1.003
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	16.74	16.76	0.999
Average Correction Factor					1.001

Notes:

Changed the inlet filter after as founds. Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.9	8.75	8.95	0.978
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.95	Prev response	8.77	*% change	2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.9	8.75	8.74	1.001
Mid point	4960	40.0	4.38	4.40	0.996
Low point	4980	20.0	2.19	2.18	1.003
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	8.75	8.78	0.996
Average Correction Factor					1.000

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.9	7.99	8.20	0.975
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.20	Prev response	8.01	*% change	2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.9	7.99	7.96	1.005
Mid point	4960	40.0	4.00	4.01	0.999
Low point	4980	20.0	2.00	2.00	1.003
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	7.99	7.97	1.003
Average Correction Factor					1.002

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.001732	0.997173
THC Cal Offset:	0.007862	0.010273
CH ₄ Cal Slope:	1.000925	0.995208
CH ₄ Cal Offset:	0.005623	0.006632
NMHC Cal Slope:	1.002456	0.999308
NMHC Cal Offset:	0.002039	0.002841

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 15, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:42	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	15005164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999987	≥ 0.995
16.74	16.69	1.0032	Slope	0.997173	$0.90 - 1.10$
8.38	8.41	0.9971	Intercept	0.010273	± 0.5
4.19	4.18	1.0032			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 15, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:42	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	15005164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999987	<i>≥0.995</i>
7.99	7.96	1.0050	Slope	0.995208	<i>0.90 - 1.10</i>
4.00	4.01	0.9986	Intercept	0.006632	<i>+/-0.5</i>
2.00	2.00	1.0031			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 15, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:42	End Time (MST):	14:15
Analyzer make:	Thermo 55i	Analyzer serial #:	15005164381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999989	≥ 0.995
8.75	8.74	1.0012	Slope	0.999308	$0.90 - 1.10$
4.38	4.40	0.9959	Intercept	0.002841	± 0.5
2.19	2.18	1.0032			



NMHC Calibration Plot

Date: October 17, 2025

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Mannix
Calibration Date: October 24, 2025
Start time (MST): 9:55
Reason: Cylinder Change

Station number: AMS 05
Last Cal Date: October 17, 2025
End time (MST): 11:45

Calibration Standards

Gas Cert Reference: CC308040
CH₄ Cal Gas Conc. 500.3 ppm
C₃H₈ Cal Gas Conc. 199.0 ppm
Removed Gas Cert:
Removed CH₄ Conc. 500.3 ppm
Removed C₃H₈ Conc. 199.0 ppm
Diff between cyl (CH₄):
Calibrator Model: API T700
Zero Air Gen model: API T701

Cal Gas Expiry Date: October 22, 2032
CH₄ Equiv Conc. 1047.6 ppm
Removed Gas Expiry:
CH₄ Equiv Conc. 1047.6 ppm
Diff between cyl (THC):
Diff between cyl (NM):
Serial Number: 5470
Serial Number: 361

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 15005164381
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.61E-04	2.61E-04	NMHC SP Ratio:	4.91E-05	4.91E-05
CH ₄ Retention time:	13.7	13.7	NMHC Peak Area:	173301	173301
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.9	16.74	16.96	0.987
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.96	Prev response	16.70	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	16.74	17.00	0.985
Average Correction Factor					

Notes:

Changed the N2 cylinder after as founds.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.9	8.75	8.90	0.983
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.90	Prev response	8.74	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	8.75	8.92	0.981
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.9	7.99	8.07	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.07	Prev response	7.96	*% change	1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	7.99	8.07	0.990
Average Correction Factor					

Calibration Statistics

	<u>Start</u>	
THC Cal Slope:	0.997173	
THC Cal Offset:	0.010273	
CH ₄ Cal Slope:	0.995208	
CH ₄ Cal Offset:	0.006632	
NMHC Cal Slope:	0.999308	
NMHC Cal Offset:	0.002841	


Finish

Calibration Performed By: Max Farrell

NMHC Calibration Plot

Date: October 24, 2025

Location: Mannix





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: October 22, 2025
Start time (MST): 8:32
Reason: Routine

Station number: AMS 06
Last Cal Date: September 11, 2025
End time (MST): 12:15

Calibration Standards

Cal Gas Concentration: 50.08 ppm
Cal Gas Cylinder #: CC255448
Removed Cal Gas Conc: 50.08 ppm
Removed Gas Cyl #:
Calibrator Model: API T700
Zero Air Gen Model: API T701

Cal Gas Exp Date: October 22, 2032
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 1160290013

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000135	1.002395	Backgd or Offset:	18.2	18.1
Calibration intercept:	1.019135	1.378488	Coeff or Slope:	0.907	0.919

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4920.2	79.8	799.3	791.9	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	792.0	Previous response	800.4	*% change	-1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920.2	79.8	799.3	802.1	0.996
Mid point	4960.1	39.9	399.6	402.2	0.994
Low point	4980	20.0	200.3	203.6	0.984
As left zero	5000	0.0	0.0	0.2	----
As left span	4919.7	80.3	804.3	800.5	1.005
Average Correction Factor:					0.991

Notes: Changed the inlet filter after as founds. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 22, 2025	Previous Calibration:	September 11, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:32	End Time (MST):	12:15
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999990	≥0.995
799.3	802.1	0.9965	Slope	1.002395	0.90 - 1.10
399.6	402.2	0.9936	Intercept	1.378488	+/-30
200.3	203.6	0.9839			



SO2 Calibration Plot

Date: October 22, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: October 7, 2025
Start time (MST): 9:24
Reason: Routine

Station number: AMS 06
Last Cal Date: September 10, 2025
End time (MST): 14:00

Calibration Standards

Cal Gas Concentration: 4.760 ppm
Cal Gas Cylinder #: DT0014585
Removed Cal Gas Conc: 4.760 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1218153358
Converter serial #: 517
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993969	1.000686	Backgd or Offset:	1.99	1.99
Calibration intercept:	0.200000	0.140000	Coeff or Slope:	1.146	1.146

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4916	84.0	80.0	82.1	0.975
As found Mid point	4958	42.0	40.0	41.5	0.966
As found Low point	4979	21.0	20.0	20.6	0.975
New cylinder response					
Baseline Corr As found:	82.0	Prev response:	79.69	*% change:	2.8%
Baseline Corr 2nd AF pt:	41.4	AF Slope:	1.025982	AF Intercept:	0.180000
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999968	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4916	84.0	80.0	80.1	0.998
Mid point	4958	42.0	40.0	40.4	0.990
Low point	4979	21.0	20.0	19.9	1.005
As left zero	5000	0.0	0.0	0.4	----
As left span	4916	84.0	80.0	79.5	1.006
SO2 Scrubber Check				0.1	
Date of last scrubber change:		Monday, December 20, 2021		Ave Corr Factor	0.998
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	October 7, 2025	Previous Calibration:	September 10, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:24	End Time (MST):	14:00
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999963		≥ 0.995
80.0	80.1	0.9984	Slope	1.000686		$0.90 - 1.10$
40.0	40.4	0.9897	Intercept	0.140000		± 3
20.0	19.9	1.0046				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Patricia McInnes
 Calibration Date: October 22, 2025
 Start time (MST): 8:32
 Reason: Routine

Station number: AMS 06
 Last Cal Date: September 11, 2025
 End time (MST): 12:15

Calibration Standards

Gas Cert Reference: CC255448
 CH₄ Cal Gas Conc. 501.4 ppm
 C₃H₈ Cal Gas Conc. 199.3 ppm
 Removed Gas Cert:
 Removed CH₄ Conc. 501.4 ppm
 Removed C₃H₈ Conc. 199.3 ppm
 Diff between cyl (CH₄):
 Calibrator Model: API T700
 Zero Air Gen model: API T701

Cal Gas Expiry Date: October 22, 2032
 CH₄ Equiv Conc. 1049.5 ppm
 Removed Gas Expiry:
 CH₄ Equiv Conc. 1049.5 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 3566
 Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1118148494
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.61E-04	2.60E-04	NMHC SP Ratio:	5.23E-05
CH ₄ Retention time:	14.4	14.2	NMHC Peak Area:	167771
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				5.19E-05
				169079
				OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.75	16.74	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.74	Prev response	16.75	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.75	16.78	0.998
Mid point	4960	39.9	8.37	8.35	1.003
Low point	4980	20.0	4.20	4.18	1.003
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.75	16.75	1.000
Average Correction Factor					1.001

Notes: Changed the inlet filter after as founds. Changed the H2 cylinder after as founds. Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	8.75	8.82	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.82	Prev response	8.77	*% change	0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.75	8.78	0.996
Mid point	4960.1	39.9	4.37	4.40	0.994
Low point	4980	20.0	2.19	2.24	0.981
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.75	8.81	0.993
Average Correction Factor					0.991

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	8.00	7.91	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.91	Prev response	7.98	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.00	8.00	1.000
Mid point	4960.1	39.9	4.00	3.95	1.012
Low point	4980	20.0	2.01	1.95	1.029
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.00	7.95	1.007
Average Correction Factor					1.014

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.999600	1.002063
THC Cal Offset:	0.010600	-0.016203
CH ₄ Cal Slope:	0.999220	1.001703
CH ₄ Cal Offset:	-0.011799	-0.031995
NMHC Cal Slope:	1.000182	1.002391
NMHC Cal Offset:	0.021998	0.015792

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 22, 2025	Previous Calibration:	September 11, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:32	End Time (MST):	12:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999994		≥ 0.995
16.75	16.78	0.9981	Slope	1.002063		0.90 - 1.10
8.37	8.35	1.0026	Intercept	-0.016203		+/-0.5
4.20	4.18	1.0033				





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 22, 2025	Previous Calibration:	September 11, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:32	End Time (MST):	12:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999925	≥ 0.995
8.00	8.00	1.0000	Slope	1.001703	$0.90 - 1.10$
4.00	3.95	1.0119	Intercept	-0.031995	± 0.5
2.01	1.95	1.0290			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 22, 2025	Previous Calibration:	September 11, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:32	End Time (MST):	12:15
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999982	<i>≥0.995</i>
8.75	8.78	0.9964	Slope	1.002391	<i>0.90 - 1.10</i>
4.37	4.40	0.9942	Intercept	0.015792	<i>+/-0.5</i>
2.19	2.24	0.9809			



NMHC Calibration Plot

Date: October 22, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1172750022

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.887	0.893	NO bkgnd or offset:	3.5	3.6
NOX coeff or slope:	0.993	0.997	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.9	160.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.993669	0.998122
NO _x Cal Offset:	3.252938	2.911651
NO Cal Slope:	0.997225	0.999055
NO Cal Offset:	0.493970	1.153160
NO ₂ Cal Slope:	0.996503	1.004883
NO ₂ Cal Offset:	-0.455577	1.204514

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.8	0.5	0.4	----	----
High point	4935	64.6	803.7	799.8	3.9	803.8	799.8	3.9	0.9999	1.0000
Mid point	4968	32.3	401.8	399.9	1.9	405.7	401.2	4.6	0.9904	0.9966
Low point	4984	16.2	201.5	200.5	1.0	205.7	202.0	3.8	0.9797	0.9928
As left zero	5000	0.0	0.0	0.0	0.0	0.5	0.3	0.3	----	----
As left span	4935	64.6	803.7	394.9	408.8	805.5	394.9	410.6	0.9978	1.0000
Average Correction Factor									0.9900	0.9965

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.4	----	----
High GPT point	797.0	393.4	407.5	410.3	0.9931	100.7%
Mid GPT point	797.0	600.4	200.5	203.0	0.9876	101.3%
Low GPT point	797.0	701.7	99.2	101.7	0.9752	102.5%
Average Correction Factor					0.9853	101.5%

Notes: Changed the inlet filter after as founds. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 2, 2025	Previous Calibration:	September 4, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:00	End Time (MST):	14:05
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.8	----	Correlation Coefficient	0.999966	≥0.995
803.7	803.8	0.9999	Slope	0.998122	0.90 - 1.10
401.8	405.7	0.9904	Intercept	2.911651	+/-20
201.5	205.7	0.9797			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 2, 2025	Previous Calibration:	September 4, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:00	End Time (MST):	14:05
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999983	≥0.995
407.5	410.3	0.9931	Slope	1.004883	0.90 - 1.10
200.5	203.0	0.9876	Intercept	1.204514	+/-20
99.2	101.7	0.9752			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 2, 2025	Previous Calibration:	September 4, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:00	End Time (MST):	14:05
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999997	≥ 0.995
799.8	799.8	1.0000	Slope	0.999055	0.90 - 1.10
399.9	401.2	0.9966	Intercept	1.153160	+/-20
200.5	202.0	0.9928			



NO_x Calibration Plot

Date: October 2, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: October 3, 2025
Start time (MST): 10:00
Reason: Routine

Station number: AMS 06
Last Cal Date: September 16, 2025
End time (MST): 13:15

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1300156234

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001800	1.002886	Backgd or Offset:	-0.5	-0.5
Calibration intercept:	-0.040000	0.220000	Coeff or Slope:	1.020	1.020

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-0.5	----
As found High point	5000	1031.0	400.0	401.0	0.996
As found Mid point					
As found Low point					
Baseline Corr As found:	401.5	Previous response	400.7	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	-0.1	----
High point	5000	1031.0	400.0	401.3	0.997
Mid point	5000	821.4	200.0	200.7	0.997
Low point	5000	699.5	100.0	101.0	0.990
As left zero	5000	800.0	0.0	0.2	----
As left span	5000	1031.0	400.0	402.3	0.994
Average Correction Factor					0.994

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	October 3, 2025	Previous Calibration:	September 16, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:00	End Time (MST):	13:15
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999996	≥ 0.995
400.0	401.3	0.9968	Slope	1.002886	$0.90 - 1.10$
200.0	200.7	0.9965	Intercept	0.220000	± 5
100.0	101.0	0.9901			



O₃ Calibration Plot

Date: October 3, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Patricia McInnes
Calibration Date: October 3, 2025
Start time (MST): 13:30
Station number: AMS 06
Last Cal Date: September 16, 2025
End time (MST): 14:22
Analyzer Make: API T640
Particulate Fraction: PM2.5
S/N: 1547
Flow Meter Make/Model: Alicat FP-25BT
Temp/RH standard: Alicat FP-25BT
S/N: 388755

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	17.1	16.8	17.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	721.60	720.50	721.60	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	4.94	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	30	----	30	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 5.4		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 07-16-2026
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: September 16, 2025
Date Disposable Filter Changed: September 16, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: April 4, 2025
Date RH/T Sensor Cleaned: April 4, 2025

Quarterly calibration completed last month. No adjustments made.

Notes:

Calibration by: Max Farrell



Wood Buffalo Environmental Association

Nt - NOX - NH3 Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	October 8, 2025	Last Cal Date:	September 18, 2025
Start time (MST):	8:47	End time (MST):	14:50
NH3 Cal Date:	October 8, 2025	Last Cal Date:	September 18, 2025
Start time (MST):	14:55	End time (MST):	15:05
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	62.20	ppm	NO Gas Cylinder #:	DT0036234
NO Cal Gas Conc:	61.90	ppm	NO Cal Gas Expiry:	July 22, 2032
Removed NOX Conc:	62.20	ppm	Removed Cylinder #:	N/A
Removed NO Conc:	61.90	ppm	Removed cyl Expiry:	N/A
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	75.0	ppm	NH3 Gas Cylinder #:	CC709372
Removed NH3 Conc:	75.0	ppm	NH3 Cal Gas Expiry:	December 31, 2025
NH3 gas Diff:			Removed Cylinder #:	
Calibrator Model:	API T700		Removed cyl Expiry:	
ZAG make/model:	API T701		Serial Number:	3566
			Serial Number:	4602

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	215
Converter model:	API T501	Converter serial #:	147
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.20
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	25.4

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	1.033	1.033	Nt coefficient:	1.027	1.027
NOX coefficient:	1.028	1.028	NO bkgrnd:	0.2	0.2
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	-0.1	-0.1
NH3 coefficient:	0.956	0.956	Nt bkgrnd:	1.7	1.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002769	1.005416
NO _x Cal Offset:	1.651917	1.471046
NO Cal Slope:	1.007657	1.004684
NO Cal Offset:	-1.307779	-0.067219
NO ₂ Cal Slope:	0.997044	0.988675
NO ₂ Cal Offset:	-0.069871	1.369665
NH3 Cal Slope:	0.995041	0.992894
NH3 Cal Offset:	6.993465	6.488525
Nt Cal Slope:	0.998870	0.996556
Nt Cal Offset:	7.386406	7.069139



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

NO_x / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO _x Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>	Baseline corr NO Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	-0.5	----	----
As found span	4935	64.6	803.7	799.8	803.7	808.2	798.1	810.6	0.9944	1.0021
AF GPT span										
new NO cyl rp										

Baseline Corr As Fd Nt = 811.1 ppb NO_x = 808.3 ppb NO = 798.3 ppb
 Previous Response Nt = 810.17 ppb NO_x = 807.6 ppb NO = 804.6 ppb

*Percent Change Nt_(NO) = 0.1%
 *Percent Change NO_x = 0.1%
 *Percent Change NO = -0.8%

**NO_x Δ (NO to GPT response) =

* = > +/-2% difference initiates investigation

* = > +/-5% change initiates investigation

NO_x / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.4	-0.2	-0.2	----	----
High point	4935	64.6	803.7	799.8	803.7	808.4	802.9	808.5	0.9942	0.9962
Mid point	4968	32.3	401.8	399.9	401.8	407.7	403.3	407.5	0.9855	0.9914
Low point	4984	16.2	201.5	200.5	201.5	204.0	200.5	205.8	0.9878	1.0002
Average Correction Factor									0.9892	0.9959

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found zero	----	----	0.0	0.1	----	----
Calibration zero	----	----	0.0	0.6	----	----
High GPT point (400 ppb O3)	800.7	393.3	411.3	407.3	1.0098	99.0%
Mid GPT point (200 ppb O3)	800.7	602.5	202.1	202.5	0.9979	100.2%
Low GPT point (100 ppb O3)	800.7	702.0	102.6	102.9	0.9969	100.3%
Average Correction Factor					1.0015	99.9%



Wood Buffalo Environmental Association NH₃ - N_T Calibration Report

NH₃ As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH ₃ concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH ₃ concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH ₃ Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.1	-0.4	----	----
AF High point	3416	84.0	1799.0	0.0	1799.0	1795.5	7.0	1788.5	1.002	1.006
AF Mid point										
AF Low point										
new NH ₃ cyl rp										
Baseline Corr As Fd	Nt = 1796.0 ppb	NH ₃ = 1788.9 ppb							*Percent Change	Nt _(NH₃) = -0.5%
Previous Response	Nt = 1804.4 ppb	NH ₃ = 1797.1 ppb							*Percent Change	NH ₃ = -0.5%

* = > +/-5% change initiates investigation

NH₃ Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH ₃ concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH ₃ concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH ₃ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	-0.2	0.4	-0.6	----	----
High point	3416	84.0	1799.0	0.0	1799.0	1795.5	7.0	1788.5	1.002	1.006
Mid point	3453	46.7	1000.3	0.0	1000.3	1008.1	4.3	1003.6	0.992	0.997
Low point	3477	23.3	499.0	0.0	499.0	511.8	2.5	509.3	0.975	0.980
Average Correction Factor									0.9897	0.9941
NH ₃ Previous Converter Efficiency =	95.6 %									
NH ₃ Current Converter Efficiency =	95.6 %									

Notes:

Changed the inlet filter after as founds. No adjustment made.

Calibration Performed By:

Max Farrell



Wood Buffalo Environmental Association


Nt Calibration Summary

Station Information

Calibration Date:	October 8, 2025	Previous Calibration:	September 18, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:47	End Time (MST):	14:50
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999917	≥ 0.995
1799.0	1795.5	1.0020	Slope	0.996556	$0.90 - 1.10$
1000.3	1008.1	0.9922	Intercept	7.069139	± 20
499.0	511.8	0.9749			





Wood Buffalo Environmental Association


NH₃ Calibration Summary

Station Information

Calibration Date:	October 8, 2025	Previous Calibration:	September 18, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:47	End Time (MST):	14:50
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.6	----	Correlation Coefficient	0.999921	≥ 0.995
1799.0	1788.5	1.0059	Slope	0.992894	0.90 - 1.10
1000.3	1003.6	0.9967	Intercept	6.488525	+/-20
499.0	509.3	0.9797			





Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 8, 2025	Previous Calibration:	September 18, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:47	End Time (MST):	14:50
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999979	≥0.995
803.7	808.4	0.9942	Slope	1.005416	0.90 - 1.10
401.8	407.7	0.9855	Intercept	1.471046	+/-20
201.5	204.0	0.9878			





Wood Buffalo Environmental Association

NO Calibration Summary


Station Information

Calibration Date:	October 8, 2025	Previous Calibration:	September 18, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:47	End Time (MST):	14:50
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999989	≥ 0.995
799.8	802.9	0.9962	Slope	1.004684	0.90 - 1.10
399.9	403.3	0.9914	Intercept	-0.067219	+/-20
200.5	200.5	1.0002			

NO Calibration Curve





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 8, 2025	Previous Calibration:	September 18, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	8:47	End Time (MST):	14:50
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999968	≥ 0.995
411.3	407.3	1.0098	Slope	0.988675	0.90 - 1.10
202.1	202.5	0.9979	Intercept	1.369665	+/-20
102.6	102.9	0.9969			



NO_x Calibration Plot

Date: October 8, 2025


Location: Patricia McInnes



NH₃ Calibration Plot

Date: October 8, 2025

Location: Patricia McInnes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Athabasca Valley Station number: AMS07
Calibration Date: October 9, 2025 Last Cal Date: September 12, 2025
Start time (MST): 9:35 End time (MST): 14:51
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.06 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC320556
Removed Cal Gas Conc: 50.06 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3805
Zero Air Gen Model: API 701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i-LTE Serial Number: 1507864683
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998074	0.996572	Backgd or Offset:	2.57	2.66
Calibration intercept:	2.844516	1.544588	Coeff or Slope:	0.829	0.843

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4920	79.8	799.0	793.9	1.007
As found Mid point	4960	39.9	399.5	402.5	0.993
As found Low point	4980	20.0	200.2	203.2	0.986
New cylinder response					
Baseline Corr As found:	793.7	Previous response	800.3	*% change	-0.8%
Baseline Corr 2nd AF pt:	402.3	AF Slope:	0.992409	AF Intercept:	2.925296
Baseline Corr 3rd AF pt:	203.0	AF Correlation:	0.999932	* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	79.8	799.0	797.1	1.002
Mid point	4960	39.9	399.5	400.4	0.998
Low point	4980	20.0	200.2	202.4	0.989
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	79.8	799.0	795.8	1.004
Average Correction Factor:					0.996

Notes: Pump changed out after MPAF. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 9, 2025	Previous Calibration:	September 12, 2025
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	9:35	End Time (MST):	14:51
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999987	≥0.995
799.0	797.1	1.0024	Slope	0.996572	0.90 - 1.10
399.5	400.4	0.9977	Intercept	1.544588	+/-30
200.2	202.4	0.9893			



SO2 Calibration Plot

Date: October 9, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Athabasca Valley
Calibration Date: October 23, 2025
Start time (MST): 10:14
Reason: Routine

Station number: AMS07
Last Cal Date: September 10, 2025
End time (MST): 14:35

Calibration Standards

Cal Gas Concentration: 5.25 ppm
Cal Gas Cylinder #: CC504080
Removed Cal Gas Conc: 5.25 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API T701H

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3805
Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1180540018
Converter serial #: 551
Converter Temp: 840 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.012751	1.004105	Backgd or Offset:	2.7	2.7
Calibration intercept:	-0.062166	-0.162212	Coeff or Slope:	0.949	0.884

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4925	75.5	79.3	84.8	0.933
As found Mid point	4962	37.7	39.6	42.5	0.927
As found Low point	4981	18.9	19.8	21.1	0.932
New cylinder response					
Baseline Corr As found:	85.0	Prev response:	80.22	*% change:	5.6%
Baseline Corr 2nd AF pt:	42.7	AF Slope:	1.072598	AF Intercept:	-0.142283
Baseline Corr 3rd AF pt:	21.3	AF Correlation:	0.999989	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4925	75.5	79.3	79.5	0.998
Mid point	4962	37.7	39.6	39.6	1.000
Low point	4981	18.9	19.9	19.7	1.008
As left zero	5000	0.0	0.0	0.0	----
As left span	4925	75.5	79.3	79.3	1.000
SO2 Scrubber Check	4920	79.2	792.1	0.1	----
Date of last scrubber change:		8-Aug-25		Ave Corr Factor	1.002
Date of last converter efficiency test:		Friday, April 22, 2022			

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	October 23, 2025	Previous Calibration:	September 10, 2025
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:14	End Time (MST):	14:35
Analyzer make:	Thermo 43i LTE	Analyzer serial #:	1180540018

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999997		≥ 0.995
79.3	79.5	0.9976	Slope	1.004105		$0.90 - 1.10$
39.6	39.6	1.0003	Intercept	-0.162212		± 3
19.9	19.7	1.0080				

TRS Calibration Curve







Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: MacKay River
Calibration Date: October 15, 2025
Start time (MST): 7:06
Reason: Routine

Station number: AMS 20
Last Cal Date: September 4, 2025
End time (MST): 11:00

Calibration Standards

Cal Gas Concentration: 5.12 ppm
Cal Gas Cylinder #: CC515997
Removed Cal Gas Conc: 5.12 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API 701

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 5706
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1236656117
Converter serial #: 2022-226
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002025	1.003598	Backgd or Offset:	3.77	3.77
Calibration intercept:	-0.040503	-0.020530	Coeff or Slope:	1.071	1.071

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4922	78.1	80.0	81.5	0.980
As found Mid point	4961	39.0	39.9	40.5	0.984
As found Low point	4980	19.5	20.0	20.2	0.984
New cylinder response					
Baseline Corr As found:	81.6	Prev response:	80.09	*% change:	1.8%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.020463	AF Intercept:	-0.160294
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999996	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4922	78.1	80.0	80.2	0.997
Mid point	4961	39.0	39.9	40.1	0.996
Low point	4980	19.5	20.0	20.1	0.994
As left zero	5000	0.0	0.0	0.1	----
As left span	4922	78.1	80.0	79.0	1.012
SO ₂ Scrubber Check	4982	81.3	802.8	0.0	----
Date of last scrubber change:	25-May-23		Ave Corr Factor		0.996
Date of last converter efficiency test:					

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 4, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	7:06	End Time (MST):	11:00
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999996		≥ 0.995
80.0	80.2	0.9972	Slope	1.003598		0.90 - 1.10
39.9	40.1	0.9959	Intercept	-0.020530		+/-3
20.0	20.1	0.9935				



H₂S Calibration Plot

Date: October 15, 2025

Location: MacKay River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Athabasca Valley
 Calibration Date: October 9, 2025
 Start time (MST): 9:35
 Reason: Routine

Station number: AMS 07
 Last Cal Date: September 12, 2025
 End time (MST): 14:51

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	Monday, March 10, 2031
CH ₄ Cal Gas Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3805
Zero Air Gen model:	Teledyne API T701H	Serial Number:	198

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1331259520
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.72E-04	2.72E-04	NMHC SP Ratio:	5.19E-05
CH ₄ Retention time:	14.4	14.4	NMHC Peak Area:	173304
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.91	16.98	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.98	Prev response	16.97	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.91	16.90	1.001
Mid point	4960	39.9	8.46	8.50	0.995
Low point	4980	20.0	4.24	4.32	0.981
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.91	16.95	0.998
Average Correction Factor					0.992

Notes:

No adjustments made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	9.00	9.06	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.06	Prev response	9.05	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.8	9.00	9.01	0.999
Mid point	4960	39.9	4.50	4.54	0.992
Low point	4980	20.0	2.26	2.33	0.970
As left zero	5000	0.0	0.00	0.00	0.998
As left span	4920	79.8	9.00	9.02	0.998
Average Correction Factor					0.987

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	7.92	7.92	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.92	Prev response	7.92	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.8	7.92	7.89	1.004
Mid point	4960	39.9	3.96	3.96	0.999
Low point	4980	20.0	1.98	2.00	0.993
As left zero	5000	0.0	0.00	0.00	0.998
As left span	4920	79.8	7.92	7.93	0.999
Average Correction Factor					0.999

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.001796	0.997742
THC Cal Offset:	0.025640	0.044648
CH ₄ Cal Slope:	0.999605	0.995779
CH ₄ Cal Offset:	0.002461	0.011467
NMHC Cal Slope:	1.003432	0.999216
NMHC Cal Offset:	0.023580	0.033182

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 9, 2025	Previous Calibration:	September 12, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:35	End Time (MST):	14:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999968	≥ 0.995
16.91	16.90	1.0008	Slope	0.997742	$0.90 - 1.10$
8.46	8.50	0.9949	Intercept	0.044648	± 0.5
4.24	4.32	0.9808			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 9, 2025	Previous Calibration:	September 12, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:35	End Time (MST):	14:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999990	<i>≥0.995</i>
7.92	7.89	1.0036	Slope	0.995779	<i>0.90 - 1.10</i>
3.96	3.96	0.9993	Intercept	0.011467	<i>+/-0.5</i>
1.98	2.00	0.9935			





Wood Buffalo Environmental Association


NMHC Calibration Summary


Station Information

Calibration Date:	October 9, 2025	Previous Calibration:	September 12, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:35	End Time (MST):	14:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999935	<i>≥0.995</i>
9.00	9.01	0.9986	Slope	0.999216	<i>0.90 - 1.10</i>
4.50	4.54	0.9916	Intercept	0.033182	<i>+/-0.5</i>
2.26	2.33	0.9699			







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Athabasca Valley
 Calibration Date: October 23, 2025
 Start time (MST): 14:30
 Reason: Cylinder Change

Station number: AMS 07
 Last Cal Date: October 9, 2025
 End time (MST): 15:22

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	Monday, March 10, 2031
CH ₄ Cal Gas Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3805
Zero Air Gen model:	Teledyne API T701H	Serial Number:	198

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1331259520
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.72E-04	2.72E-04	NMHC SP Ratio:	5.19E-05
CH ₄ Retention time:	14.4	14.4	NMHC Peak Area:	173304
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.91	16.94	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.94	Prev response	16.97	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.91	16.82	1.005
Average Correction Factor					

Notes:

H2 cylinder changed out after as founds.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	9.00	9.01	0.998
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.01	Prev response	9.05	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.00	8.95	1.006
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	7.92	7.93	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.93	Prev response	7.92	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	7.92	7.88	1.005
Average Correction Factor					

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.001796	
THC Cal Offset:	0.025640	
CH ₄ Cal Slope:	0.999605	
CH ₄ Cal Offset:	0.002461	
NMHC Cal Slope:	1.003432	
NMHC Cal Offset:	0.023580	

Calibration Performed By: Aswin Sasi Kumar

NMHC Calibration Plot

Date: October 23, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1160120024

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997995	0.996059
NO _x Cal Offset:	1.871926	2.251931
NO Cal Slope:	0.999699	0.997028
NO Cal Offset:	1.591942	1.951944
NO ₂ Cal Slope:	0.999879	1.000683
NO ₂ Cal Offset:	0.292038	0.151290

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.917	0.917	NO bkgnd or offset:	6.4	6.4
NOX coeff or slope:	1.002	1.002	NOX bkgnd or offset:	6.6	6.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	185.1	185.1

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
High point	4933	66.8	803.0	800.3	2.7	800.7	798.5	2.2	1.0028	1.0022
Mid point	4966	33.4	401.5	400.2	1.3	404.2	403.2	1.1	0.9934	0.9925
Low point	4983	16.7	200.7	200.1	0.7	203.6	202.4	1.2	0.9860	0.9885
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
As left span	4933	66.8	803.0	398.7	404.3	800.7	398.7	402.0	1.0028	1.0000
Average Correction Factor									0.9941	0.9944

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	796.9	399.2	400.4	400.7	0.9992	100.1%
Mid GPT point	796.9	599.7	199.9	200.4	0.9974	100.3%
Low GPT point	796.9	698.8	100.8	100.9	0.9987	100.1%
Average Correction Factor					0.9984	100.2%

Notes:

No adjustments made.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 9, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:35	End Time (MST):	14:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999966	≥0.995
803.0	800.7	1.0028	Slope	0.996059	0.90 - 1.10
401.5	404.2	0.9934	Intercept	2.251931	+/-20
200.7	203.6	0.9860			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 9, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:35	End Time (MST):	14:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999999	≥ 0.995
400.4	400.7	0.9992	Slope	1.000683	$0.90 - 1.10$
199.9	200.4	0.9974	Intercept	0.151290	± 20
100.8	100.9	0.9987			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 9, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	9:35	End Time (MST):	14:01
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999967	≥ 0.995
800.3	798.5	1.0022	Slope	0.997028	$0.90 - 1.10$
400.2	403.2	0.9925	Intercept	1.951944	± 20
200.1	202.4	0.9885			



NO_x Calibration Plot

Date: October 17, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Athabasca Valley
Calibration Date: October 1, 2025
Start time (MST): 10:00
Reason: Removal

Station number: AMS07
Last Cal Date: September 19, 2025
End time (MST): 12:05

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: T700
ZAG Make/Model: T701H

Serial Number: 3805
Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1152220023

Start
Calibration slope: 1.002600
Calibration intercept: 1.120000

Finish

Start
Backgd or Offset: -1.8
Coeff or Slope: 1.556

Finish
N/A
N/A

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
					<i>Limit = 0.90-1.10</i>
As found zero	5000	NA	0.0	-0.6	----
As found High point	5000	1725.3	400.0	402.0	0.994
As found Mid point	5000	1172.8	200.0	201.1	0.992
As found Low point	5000	921.2	100.0	101.3	0.981
Baseline Corr As found:	402.6	Previous response	402.2	*% change	0.1%
Baseline Corr 2nd AF pt:	201.7	AF Slope:	1.005543	AF Intercept:	-0.020000
Baseline Corr 3rd AF pt:	101.9	AF Correlation:	0.999989	* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero))
					<i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Notes:

Removal calibration.


Calibration Performed By:

Aswin Sasi Kumar

O₃ Calibration Plot

Date: October 1, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Athabasca Valley
Calibration Date: October 1, 2025
Start time (MST): 12:10
Reason: Install

Station number: AMS07
Last Cal Date: N/A
End time (MST): 14:55

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: T700
ZAG Make/Model: T701H

Serial Number: 3805
Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1507964700

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	N/A	1.005086	Backgd or Offset:	N/A	-1.8
Calibration intercept:	N/A	-0.240000	Coeff or Slope:	N/A	1.556

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	NA	0.0	-0.7	----
High point	5000	1705.1	400.0	401.5	0.996
Mid point	5000	1172.8	200.0	201.2	0.994
Low point	5000	921.2	100.0	100.6	0.994
As left zero	5000	NA	0.0	0.0	----
As left span	5000	1582.6	400.0	406.0	0.985
Average Correction Factor					0.995

Notes: Install calibration. Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	October 1, 2025	Previous Calibration:	N/A
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	12:10	End Time (MST):	14:55
Analyzer make:	Thermo 49i	Analyzer serial #:	1507964700

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.7	----	Correlation Coefficient	0.999993	≥0.995
400.0	401.5	0.9963	Slope	1.005086	0.90 - 1.10
200.0	201.2	0.9940	Intercept	-0.240000	+/- 5
100.0	100.6	0.9940			



O₃ Calibration Plot

Date: October 1, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Athabasca Valley Station number: AMS 07
Calibration Date: October 9, 2025 Last Cal Date: September 19, 2025
Start time (MST): 13:30 End time (MST): 14:45

Analyzer Make: API T640 S/N: 2235
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388748
Temp/RH standard: Alicat FP-25BT S/N: 388748

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	16.4	15.4	16.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	740.3	739.84	740.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.03	4.85	5.03	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	32	----	32	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 1.2		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: July 16, 2026
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.0	11.0	11.0	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: October 9, 2025
Date Disposable Filter Changed: October 9, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 9, 2025
Date RH/T Sensor Cleaned: October 9, 2025

Notes: Temp, pressure and flow checked. Leak check passed. Quarterly and annual checks done.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name: Athabasca Valley Station number: AMS 07
Calibration Date: October 24, 2025 Last Cal Date: September 5, 2025
Start time (MST): 10:02 End time (MST): 14:15
Reason: Routine

Calibration Standards

Cal Gas Concentration: 2,953 ppm Cal Gas Exp Date: September 30, 2029
Cal Gas Cylinder #: T1TWKRN
Removed Cal Gas Conc: 2,953 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3805
ZAG Make/Model: Teledyne API 701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 48i-TLE Analyzer serial #: 1408761381
Analyzer Range: 0 - 50 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995870	0.998381	Backgd or Offset:	5.731
Calibration intercept:	0.227985	0.144014	Coeff or Slope:	1.074

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4932	67.8	40.0	40.2	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	40.08	Prev response:	40.11	*% change:	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4932	67.8	40.0	40.0	1.001
Mid point	4966	33.9	20.0	20.3	0.987
Low point	4983	16.9	10.0	10.2	0.978
As left zero	5000	0.0	0.0	0.0	----
As left span	4932	67.8	40.0	40.0	1.002
Average Correction Factor					0.988

Notes: Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

CO Calibration Summary


Station Information

Calibration Date:	October 24, 2025	Previous Calibration:	September 5, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:02	End Time (MST):	14:15
Analyzer make:	Thermo 48i-TLE	Analyzer serial #:	1408761381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999921	≥ 0.995
40.0	40.0	1.0006	Slope	0.998381	$0.90 - 1.10$
20.0	20.3	0.9868	Intercept	0.144014	± 1.5
10.0	10.2	0.9776			


CO Calibration Curve



CO Calibration Plot

Date: October 24, 2025

Location: Athabasca Valley





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
Calibration Date: October 16 2025 Last Cal Date: September 16, 2025
Start time (MST): 10:30 End time (MST): 13:48
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.84 ppm Cal Gas Exp Date: January 6, 2030
Cal Gas Cylinder #: CC196697
Removed Cal Gas Conc: 49.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3810
Zero Air Gen Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43i-TLE Serial Number: 1236656116
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003816	0.999773	Backgd or Offset:	1.9	1.9
Calibration intercept:	-0.145224	-0.603699	Coeff or Slope:	1.029	1.029

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4920	80.3	800.4	799.3	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	799.1	Previous response	803.3	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4920	80.3	800.4	800.0	1.000
Mid point	4960	40.2	400.7	399.7	1.002
Low point	4980	20.1	200.4	198.7	1.008
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	80.3	800.4	799.5	1.001
Average Correction Factor:					1.004

Notes: Changed out inlet filter after as founds. No adjustments made.

Calibration Performed By: Jeremy cardinal



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 16 2025	Previous Calibration:	September 16, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	10:30	End Time (MST):	13:48
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999994	≥0.995
800.4	800.0	1.0005	Slope	0.999773	0.90 - 1.10
400.7	399.7	1.0025	Intercept	-0.603699	+/-30
200.4	198.7	1.0083			



SO2 Calibration Plot

Date: October 16 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
Calibration Date: October 23, 2025 Last Cal Date: September 15 2025
Start time (MST): 11:36 End time (MST): 14:58
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: August 28, 2027
Cal Gas Cylinder #: SA7549
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3810
ZAG Make/Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153461
Converter make: CDN-101 Converter serial #: 630
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001244	0.997099	Backgd or Offset:	2.8
Calibration intercept:	0.061962	0.082015	Coeff or Slope:	1.158

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4917	82.6	80.0	80.9	0.986
As found Mid point	4959	41.3	40.0	40.5	0.982
As found Low point	4979	20.7	20.0	20.0	0.992
New cylinder response					
Baseline Corr As found:	81.1	Prev response:	80.12	*% change:	1.2%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.014968	AF Intercept:	-0.218287
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999990	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4917	82.6	80.0	79.7	1.003
Mid point	4959	41.3	40.0	40.2	0.994
Low point	4979	20.7	20.0	20.0	1.002
As left zero	5000	0.0	0.0	0.3	----
As left span	4917	82.6	80.0	81.3	0.984
SO2 Scrubber Check	4920	80.3	803.0	-0.1	----
Date of last scrubber change:	May 15, 2025			Ave Corr Factor	1.000
Date of last converter efficiency test:					

Notes: No adjustments made.

Calibration Performed By: jermey cardinal



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	October 23, 2025	Previous Calibration:	September 15 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	11:36	End Time (MST):	14:58
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999974		≥ 0.995
80.0	79.7	1.0033	Slope	0.997099		$0.90 - 1.10$
40.0	40.2	0.9944	Intercept	0.082015		± 3
20.0	20.0	1.0019				

TRS Calibration Curve







Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan
Station number: AMS 08
Calibration Date: October 17, 2025
Last Cal Date: September 17, 2025
Start time (MST): 7:28
End time (MST): 12:28
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC358100
NOX Cal Gas Conc: 60.10 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.10 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 197

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.0	-0.2	-0.7	----	----
AF High point	4933	66.7	801.8	800.4	1.3	741.4	736.2	5.3	1.0800	1.0870
AF Mid point										
AF Low point										
New cyl resp										

Previous Response NO_x = 803.5 ppb NO = 800.1 ppb
Baseline Corr 1st pt NO_x = 742.4 ppb NO = 736.4 ppb
Baseline Corr 2nd pt NO_x = NA ppb NO = NA ppb
Baseline Corr 3rd pt NO_x = NA ppb NO = NA ppb

* = > +/-5% change initiates investigation

*Percent Change NO_x = -8.2%

As Found Statistics

*Percent Change NO = -8.7%

As found NO_x r²:
As found NO r²:
As found NO₂ r²:

Nx SI: Nx Int:
NO SI: NO Int:
NO₂ SI: NO₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12124313137

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003197	0.998321
NO _x Cal Offset:	-0.885527	-0.825795
NO Cal Slope:	1.000955	0.999327
NO Cal Offset:	-1.065293	-1.145428
NO ₂ Cal Slope:	1.008496	1.005011
NO ₂ Cal Offset:	0.885220	-1.625691

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.743	0.806	NO bkgnd or offset:	0.8	0.8
NOX coeff or slope:	0.998	0.996	NOX bkgnd or offset:	1.7	1.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	149.3	159.2

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.1	-0.7	----	----
High point	4933	66.7	801.8	800.4	1.3	799.7	799.2	0.6	1.0026	1.0016
Mid point	4967	33.3	400.2	399.6	0.7	398.5	397.9	1.7	1.0044	1.0042
Low point	4983	16.7	200.7	200.4	0.3	199.7	197.9	1.8	1.0052	1.0126
As left zero	5000	0.0	0.0	0.0	0.0	-0.8	-0.1	-0.7	----	----
As left span	4933	66.7	801.8	384.1	417.7	787.7	384.1	403.6	1.0179	1.0000
Average Correction Factor									1.0041	1.0061

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.7	----	----
High GPT point	794.8	384.3	411.8	412.4	0.9986	100.1%
Mid GPT point	794.8	606.8	189.3	189.2	1.0007	99.9%
Low GPT point	794.8	698.8	97.3	94.6	1.0289	97.2%
Average Correction Factor					1.0094	99.1%

Notes: Changed filter after as found. Performed span adjustment.

Calibration Performed By: Mrgan Voyageur, Jerney Cardinal



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 17, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	7:28	End Time (MST):	12:28
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12124313137

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.8	----	Correlation Coefficient	1.000000	≥0.995
801.8	799.7	1.0026	Slope	0.998321	0.90 - 1.10
400.2	398.5	1.0044	Intercept	-0.825795	+/-20
200.7	199.7	1.0052			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 17, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	7:28	End Time (MST):	12:28
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12124313137

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.7	----	Correlation Coefficient	0.999960	≥0.995
411.8	412.4	0.9986	Slope	1.005011	0.90 - 1.10
189.3	189.2	1.0007	Intercept	-1.625691	+/-20
97.3	94.6	1.0289			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 17, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	7:28	End Time (MST):	12:28
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12124313137

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999992	≥ 0.995
800.4	799.2	1.0016	Slope	0.999327	0.90 - 1.10
399.6	397.9	1.0042	Intercept	-1.145428	+/-20
200.4	197.9	1.0126			



NO_x Calibration Plot

Date: October 17, 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Fort Chipewyan
Calibration Date: October 23, 2025
Start time (MST): 8:09
Reason: Routine

Station number: AMS 08
Last Cal Date: September 16, 2025
End time (MST): 11:04

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Serial Number: 3810
Serial Number: 135

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1152220026

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997800	1.002629	Backgd or Offset:	-0.3	-2.6
Calibration intercept:	1.060000	0.740000	Coeff or Slope:	1.029	0.995

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	5000	968.7	400.0	400.3	0.999
As found Mid point					
As found Low point					
Baseline Corr As found:	400.6	Previous response	400.2	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	5000	968.7	400.0	401.5	0.996
Mid point	5000	820.5	200.0	201.5	0.993
Low point	5000	720.0	100.0	101.7	0.983
As left zero					
As left span					

Average Correction Factor 0.991

Notes: Changed Filter after as founds. Adjustments made to Zero and then span high point. As lefts not captured.

Calibration Performed By: Jerney cardinal



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 16, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	8:09	End Time (MST):	11:04
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220026

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999988	≥0.995
400.0	401.5	0.9963	Slope	1.002629	0.90 - 1.10
200.0	201.5	0.9926	Intercept	0.740000	+/- 5
100.0	101.7	0.9833			



O₃ Calibration Plot

Date: October 23 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
Calibration Date: October 30, 2025 Last Cal Date: September 23, 2025
Start time (MST): 13:44 End time (MST): 15:35

Analyzer Make: Teledyne API T640 S/N: 319
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388744
Temp/RH standard: Alicat FP-25BT S/N: 388744

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	0.4	-0.80	0.4	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	742.8	745.50	742.8	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.03	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	30%		30%	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 0.90		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☐ Alignment Factor On : ☐

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: 10-Jun-24
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: August 19, 2025
Date Disposable Filter Changed: August 19, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 29, 2024
Date RH/T Sensor Cleaned: August 29, 2024

Notes:

No adjustments made.

Calibration by: Morgan Voyageur



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Barge Landing
Calibration Date: October 3, 2025
Start time (MST): 9:37
Reason: Routine

Station number: AMS 09
Last Cal Date: September 15, 2025
End time (MST): 12:28

Calibration Standards

Cal Gas Concentration: 50.56 ppm
Cal Gas Cylinder #: CC705748
Removed Cal Gas Conc: 50.56 ppm
Removed Gas Cyl #: NA
Calibrator Model: API T700
Zero Air Gen Model: APIT701

Cal Gas Exp Date: October 9, 2032
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3812
Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 1118148498

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000887	0.995670	Backgd or Offset:	11.1	11.5
Calibration intercept:	0.001592	0.401753	Coeff or Slope:	0.985	0.985

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4921	79.1	799.8	797.3	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.2	Previous response	800.6	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	79.1	799.8	796.5	1.004
Mid point	4961	39.5	399.4	398.5	1.002
Low point	4980	19.8	200.2	200.0	1.001
As left zero	5000	0.0	0.0	-0.1	----
As left span	4921	79.1	799.8	797.5	1.003
Average Correction Factor:					1.003

Notes: Sample inlet filter was changed after as founds. Adjusted zero only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 3, 2025	Previous Calibration:	September 15, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:37	End Time (MST):	12:28
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999999	≥0.995
799.8	796.5	1.0042	Slope	0.995670	0.90 - 1.10
399.4	398.5	1.0022	Intercept	0.401753	+/-30
200.2	200.0	1.0011			



SO2 Calibration Plot

Date: October 3, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Barge Landing Station number: AMS 09
Calibration Date: October 8, 2025 Last Cal Date: September 3, 2025
Start time (MST): 8:41 End time (MST): 13:21
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.17 ppm Cal Gas Exp Date: August 22, 2026
Cal Gas Cylinder #: CC511415
Removed Cal Gas Conc: 5.17 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3812
ZAG Make/Model: API T701 Serial Number: 5613

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 1203169744
Converter make: CDN-101 Converter serial #: 519
Analyzer Range: 0 - 100 ppb Converter Temp: 830 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998261	1.004685	Backgd or Offset:	2.070	2.260
Calibration intercept:	-0.220501	-0.320371	Coeff or Slope:	0.798	0.871

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4923	77.4	80.0	78.2	1.021
As found Mid point	4961	38.7	40.0	38.0	1.048
As found Low point	4981	19.3	20.0	18.3	1.079
New cylinder response					
Baseline Corr As found:	78.4	Prev response:	79.68	*% change:	-1.6%
Baseline Corr 2nd AF pt:	38.2	AF Slope:	0.982569	AF Intercept:	-0.821159
Baseline Corr 3rd AF pt:	18.5	AF Correlation:	0.999697	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4923	77.4	80.0	80.5	0.994
Mid point	4961	38.7	40.0	39.0	1.026
Low point	4981	19.3	20.0	19.9	1.003
As left zero	5000	0.0	0.0	0.5	----
As left span	4923	77.4	80.0	78.2	1.024
SO2 Scrubber Check	4920	80.2	802.0	-0.1	----
Date of last scrubber change:	8-Jul-25		Ave Corr Factor		1.008
Date of last converter efficiency test:					

Notes: Changed inlet filter after as founds. Scrubber check passed and done after calibrator zero. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	October 8, 2025	Previous Calibration:	September 3, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	8:41	End Time (MST):	13:21
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1203169744

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999691		≥ 0.995
80.0	80.5	0.9943	Slope	1.004685		$0.90 - 1.10$
40.0	39.0	1.0263	Intercept	-0.320371		± 3
20.0	19.9	1.0030				


TRS Calibration Curve



TRS Calibration Plot

Date: October 8, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Barge Landing
 Calibration Date: October 3, 2025
 Start time (MST): 9:37
 Reason: Routine

Station number: AMS 09
 Last Cal Date: September 15, 2025
 End time (MST): 12:28

Calibration Standards

Gas Cert Reference: CC705748
 CH₄ Cal Gas Conc. 505.6 ppm
 C₃H₈ Cal Gas Conc. 204.8 ppm
 Removed Gas Cert: CC151285
 Removed CH₄ Conc. 505.6 ppm
 Removed C₃H₈ Conc. 204.8 ppm
 Diff between cyl (CH₄):
 Calibrator Model: API T700
 Zero Air Gen model: APIT701

Cal Gas Expiry Date: October 9, 2032
 CH₄ Equiv Conc. 1068.8 ppm
 Removed Gas Expiry: January 5, 2025
 CH₄ Equiv Conc. 1068.8 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 3812
 Serial Number: 5613

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1193585650
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.24E-04	2.26E-04	NMHC SP Ratio: 5.16E+05	5.30E-05
CH ₄ Retention time:	14.4	14.4	NMHC Peak Area: 172810	168197
Zero Chromatogram:	OFF	OFF	Flat Baseline: OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	16.91	16.65	1.016
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.65	Prev response	16.92	*% change	-1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	16.91	16.94	0.998
Mid point	4961	39.5	8.44	8.39	1.006
Low point	4980	19.8	4.23	4.16	1.017
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	16.91	16.93	0.998
Average Correction Factor					1.007

Notes:

Sample inlet filter was changed after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.1	8.91	8.72	1.022
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.72	Prev response	8.94	*% change	-2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.1	8.91	8.95	0.996
Mid point	4961	39.5	4.45	4.43	1.005
Low point	4980	19.8	2.23	2.19	1.018
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	8.91	8.93	0.997
Average Correction Factor					1.006

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.1	8.00	7.93	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.93	Prev response	7.98	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.1	8.00	8.00	1.000
Mid point	4961	39.5	3.99	3.97	1.007
Low point	4980	19.8	2.00	1.97	1.016
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	8.00	8.00	1.000
Average Correction Factor					1.008

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.002991	1.003134
THC Cal Offset:	-0.037963	-0.045773
CH ₄ Cal Slope:	1.001017	1.001330
CH ₄ Cal Offset:	-0.022190	-0.018787
NMHC Cal Slope:	1.004994	1.005356
NMHC Cal Offset:	-0.016174	-0.027585

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 3, 2025	Previous Calibration:	September 15, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:37	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999966	≥ 0.995
16.91	16.94	0.9981	Slope	1.003134	$0.90 - 1.10$
8.44	8.39	1.0060	Intercept	-0.045773	± 0.5
4.23	4.16	1.0175			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 3, 2025	Previous Calibration:	September 15, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:37	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999974		<i>≥0.995</i>
8.00	8.00	0.9997	Slope	1.001330		<i>0.90 - 1.10</i>
3.99	3.97	1.0068	Intercept	-0.018787		<i>+/-0.5</i>
2.00	1.97	1.0164				






Wood Buffalo Environmental Association NMHC Calibration Summary

Station Information

Calibration Date:	October 3, 2025	Previous Calibration:	September 15, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	9:37	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999956	≥ 0.995
8.91	8.95	0.9960	Slope	1.005356	$0.90 - 1.10$
4.45	4.43	1.0049	Intercept	-0.027585	± 0.5
2.23	2.19	1.0184			



NMHC Calibration Plot

Date: October 3, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Barge Landing
 Calibration Date: October 24, 2025
 Start time (MST): 9:40
 Reason: Cylinder Change

Station number: AMS 09
 Last Cal Date: October 3, 2025
 End time (MST): 11:19

Calibration Standards

Gas Cert Reference: CC705748
 CH₄ Cal Gas Conc. 505.6 ppm
 C₃H₈ Cal Gas Conc. 204.8 ppm
 Removed Gas Cert: CC151285
 Removed CH₄ Conc. 505.6 ppm
 Removed C₃H₈ Conc. 204.8 ppm
 Diff between cyl (CH₄):
 Calibrator Model: API T700
 Zero Air Gen model: APIT701

Cal Gas Expiry Date: October 9, 2032
 CH₄ Equiv Conc. 1068.8 ppm
 Removed Gas Expiry: January 5, 2025
 CH₄ Equiv Conc. 1068.8 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 3812
 Serial Number: 5613

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1193585650
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.24E-04	2.26E-04	NMHC SP Ratio: 5.16E+05	5.30E-05
CH ₄ Retention time:	14.4	14.4	NMHC Peak Area: 172810	168197
Zero Chromatogram:	OFF	OFF	Flat Baseline: OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	16.91	16.92	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.92	Prev response	16.92	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	16.91	16.86	1.003
Average Correction Factor					

Notes:

N₂/ H₂ was changed after as founds. No adjustmend made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	8.91	8.97	0.994
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.97	Prev response	8.93	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	8.91	8.92	0.999
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	8.00	7.95	1.006
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.95	Prev response	7.99	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	8.00	7.95	1.006
Average Correction Factor					

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.003134	
THC Cal Offset:	-0.045773	
CH ₄ Cal Slope:	1.001330	
CH ₄ Cal Offset:	-0.018787	
NMHC Cal Slope:	1.005356	
NMHC Cal Offset:	-0.027585	

Calibration Performed By: Jan Castro

NMHC Calibration Plot

Date: October 24, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Barge Landing
Station number: AMS 09
Calibration Date: October 20, 2025
Last Cal Date: September 9, 2025
Start time (MST): 10:16
End time (MST): 14:22
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1KDH
NOX Cal Gas Conc: 47.38 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.38 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: Api T701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 46.94 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 46.94 ppm
NO gas Diff:
Serial Number: 3812
Serial Number: 4888

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.2	-0.2	----	----
AF High point	4915	85.3	808.3	800.7	7.5	810.9	800.4	10.5	0.9962	1.0002
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 810.0 ppb	NO = 802.4 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.2%
Baseline Corr 1st pt	NO _x = 811.3 ppb	NO = 800.6 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -0.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1426262593

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.103	1.103	NO bkgnd or offset:	10.1	10.1
NOX coeff or slope:	0.999	0.999	NOX bkgnd or offset:	10.4	10.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	176.2	176.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001255	0.998597
NO _x Cal Offset:	0.758641	0.098064
NO Cal Slope:	1.002550	0.998568
NO Cal Offset:	-0.383252	-0.664019
NO ₂ Cal Slope:	1.000667	0.999032
NO ₂ Cal Offset:	-0.442357	0.158236

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
High point	4915	85.3	808.3	800.7	7.5	807.1	799.3	7.8	1.0014	1.0018
Mid point	4957	42.6	403.7	400.0	3.7	403.3	398.3	5.0	1.0010	1.0042
Low point	4979	21.3	201.8	200.0	1.9	202.0	198.4	3.6	0.9991	1.0078
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
As left span	4915	85.3	808.3	419.2	389.1	805.0	419.2	385.7	1.0040	1.0000
Average Correction Factor									1.0005	1.0046

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	794.6	416.7	385.4	385.0	1.0011	99.9%
Mid GPT point	794.6	612.5	189.6	189.8	0.9990	100.1%
Low GPT point	794.6	705.1	97.0	97.4	0.9960	100.4%
Average Correction Factor					0.9987	100.1%

Notes:

Inlet filter changed after as founds. No adjustment.

Calibration Performed By:

Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 9, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:16	End Time (MST):	14:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999999	≥0.995
808.3	807.1	1.0014	Slope	0.998597	0.90 - 1.10
403.7	403.3	1.0010	Intercept	0.098064	+/-20
201.8	202.0	0.9991			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 9, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:16	End Time (MST):	14:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999996	≥0.995
385.4	385.0	1.0011	Slope	0.999032	0.90 - 1.10
189.6	189.8	0.9990	Intercept	0.158236	+/-20
97.0	97.4	0.9960			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 9, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:16	End Time (MST):	14:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999997	≥ 0.995
800.7	799.3	1.0018	Slope	0.998568	$0.90 - 1.10$
400.0	398.3	1.0042	Intercept	-0.664019	± 20
200.0	198.4	1.0078			



NO_x Calibration Plot

Date: October 20, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Barge Landing Station number: AMS 09
Calibration Date: October 20, 2025 Last Cal Date: September 15, 2025
Start time (MST): 12:29 End time (MST): 13:21

Analyzer Make: API T640 S/N: 2237
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388754
Temp/RH standard: Alicat FP-25 S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	3.40	3.43	3.40	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	736.30	734.44	736.30	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.96	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	32	----	32	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	5.60	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: January 30, 2027
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	9.2	10.8	10.8	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ October 20, 2025
Date Disposable Filter Changed: _____ October 20, 2025

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ July 22, 2025
Date RH/T Sensor Cleaned: _____ July 22, 2025

Notes: Verified flow, pressure, temperature and pump power. No adjustment needed . Leak check passed.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS11 LOWER CAMP OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Lower Camp Station number: AMS 11
Calibration Date: October 14, 2025 Last Cal Date: September 2, 2025
Start time (MST): 10:34 End time (MST): 15:01
Reason: Routine

Calibration Standards

Cal Gas Concentration: 48.75 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC741503
Removed Cal Gas Conc: 48.75 ppm Rem Gas Exp Date:
Removed Gas Cyl #: CC741503 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3811
Zero Air Gen Model: Teledyne API T701 Serial Number: 196

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 100841398
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996537	1.003586	Backgd or Offset:	18.2	18.4
Calibration intercept:	-0.159945	-0.402252	Coeff or Slope:	1.045	1.057

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4932	82.2	799.2	790.5	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	790.2	Previous response	796.3	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.5	----
High point	4932	82.2	799.2	802.4	0.996
Mid point	4971	41.2	400.7	400.4	1.001
Low point	4996	20.6	200.2	200.2	1.000
As left zero	5000	0.0	0.0	0.5	----
As left span	4932	82.2	799.2	802.4	0.996
Average Correction Factor:					0.999

Notes: Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	October 14, 2025	Previous Calibration:	September 2, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:34	End Time (MST):	15:01
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999991	≥0.995
799.2	802.4	0.9960	Slope	1.003586	0.90 - 1.10
400.7	400.4	1.0008	Intercept	-0.402252	+/-30
200.2	200.2	0.9999			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: October 14, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Lower Camp Station number: AMS 11
Calibration Date: October 8, 2025 Last Cal Date: September 23, 2025
Start time (MST): 11:14 End time (MST): 15:29
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.83 ppm Cal Gas Exp Date: August 28, 2028
Cal Gas Cylinder #: CC737863
Removed Cal Gas Conc: 4.83 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3811
ZAG Make/Model: API T701H Serial Number: 196

Analyzer Information

Analyzer make: Thermo 43iQ Analyzer serial #: 1203169745
Converter make: Global G150 Converter serial #: 2022-223
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004138	1.012008	Backgd or Offset:	2.5
Calibration intercept:	0.148573	0.048287	Coeff or Slope:	0.808

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4923	82.8	79.9	80.8	0.988
As found Mid point	4967	41.5	40.0	40.5	0.986
As found Low point	4999	20.8	20.0	20.4	0.976
New cylinder response					
Baseline Corr As found:	80.9	Prev response:	80.37	*% change:	0.7%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.011719	AF Intercept:	0.008389
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999991	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4923	82.8	79.9	80.8	0.989
Mid point	4967	41.5	40.0	40.7	0.983
Low point	4999	20.8	20.0	20.4	0.981
As left zero	5000	0.0	0.0	0.1	----
As left span	4923	82.8	79.9	80.5	0.992
SO2 Scrubber Check	4932	82.2	819.7	0.2	----
Date of last scrubber change:				Ave Corr Factor	0.984
Date of last converter efficiency test:					

Notes: Changed sample inlet filter after as founds. Ran scrubber check after calibrator zero and it passed.
No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	October 8, 2025	Previous Calibration:	September 23, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	11:14	End Time (MST):	15:29
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999982	≥0.995
79.9	80.8	0.9888	Slope	1.012008	0.90 - 1.10
40.0	40.7	0.9833	Intercept	0.048287	+/-3
20.0	20.4	0.9811			



H₂S Calibration Plot

Date: October 8, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH4 / NMHC Calibration Report

Station Information

Station Name: Lower Camp
 Calibration Date: October 14, 2025
 Start time (MST): 10:34
 Reason: Routine

Station number: AMS 11
 Last Cal Date: September 2, 2025
 End time (MST): 15:01

Calibration Standards

Gas Cert Reference:	CC741503	Cal Gas Expiry Date:	October 9, 2032
CH4 Cal Gas Conc.	504.8 ppm	CH4 Equiv Conc.	1071.9 ppm
C3H8 Cal Gas Conc.	206.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	504.8 ppm	CH4 Equiv Conc.	1071.9 ppm
Removed C3H8 Conc.	206.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3811
Zero Air Gen model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1118148495
 NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.65E-04	2.82E-04	NMHC SP Ratio:	4.62E-05
CH4 Retention time:	14.8	15.0	NMHC Peak Area:	201460
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	82.2	17.57	16.94	1.038
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.94	Prev response	17.58	*% change	-3.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	82.2	17.57	17.65	0.996
Mid point	4971	41.2	8.81	8.76	1.006
Low point	4996	20.6	4.40	4.36	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	17.57	17.49	1.005
				Average Correction Factor	1.004

Notes: No alarms were detected during the investigation. Diagnostics were consistent, and the chromatograms showed no significant discrepancies. Changed sample inlet filter after as founds. Adjusted zero and span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4932	82.2	9.30	9.10	1.022
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.10	Prev response	9.33	*% change	-2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4932	82.2	9.30	9.33	0.997
Mid point	4971	41.2	4.66	4.63	1.006
Low point	4996	20.6	2.33	2.30	1.012
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	9.30	9.25	1.005
Average Correction Factor					1.005

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4932	82.2	8.28	7.84	1.056
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.84	Prev response	8.25	*% change	-5.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4932	82.2	8.28	8.33	0.994
Mid point	4971	41.2	4.15	4.12	1.007
Low point	4996	20.6	2.07	2.06	1.006
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	8.28	8.24	1.004
Average Correction Factor					1.002

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.002581	1.004855
THC Cal Offset:	-0.037176	-0.042179
CH ₄ Cal Slope:	0.996443	1.006101
CH ₄ Cal Offset:	0.007460	-0.019544
NMHC Cal Slope:	1.008057	1.004041
NMHC Cal Offset:	-0.044437	-0.022835

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 14, 2025	Previous Calibration:	September 2, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:34	End Time (MST):	15:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999961	≥ 0.995
17.57	17.65	0.9957	Slope	1.004855	$0.90 - 1.10$
8.81	8.76	1.0063	Intercept	-0.042179	± 0.5
4.40	4.36	1.0097			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 14, 2025	Previous Calibration:	September 2, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:34	End Time (MST):	15:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999947	<i>≥0.995</i>
8.28	8.33	0.9940	Slope	1.006101	<i>0.90 - 1.10</i>
4.15	4.12	1.0069	Intercept	-0.019544	<i>+/-0.5</i>
2.07	2.06	1.0063			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 14, 2025	Previous Calibration:	September 2, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	10:34	End Time (MST):	15:01
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148495

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999968	<i>≥0.995</i>
9.30	9.33	0.9968	Slope	1.004041	<i>0.90 - 1.10</i>
4.66	4.63	1.0059	Intercept	-0.022835	<i>+/-0.5</i>
2.33	2.30	1.0124			



NMHC Calibration Plot

Date: October 14, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH4 / NMHC Calibration Report

Station Information

Station Name: Lower Camp
Calibration Date: October 15, 2025
Start time (MST): 10:55
Reason: Cylinder Change

Station number: AMS 11
Last Cal Date: October 14, 2025
End time (MST): 12:55

Calibration Standards

Gas Cert Reference:	CC741503	Cal Gas Expiry Date:	October 9, 2032
CH4 Cal Gas Conc.	504.8 ppm	CH4 Equiv Conc.	1071.9 ppm
C3H8 Cal Gas Conc.	206.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	504.8 ppm	CH4 Equiv Conc.	1071.9 ppm
Removed C3H8 Conc.	206.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3811
Zero Air Gen model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1118148495
NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.82E-04	2.82E-04	NMHC SP Ratio:	4.79E-05
CH4 Retention time:	15.0	15.0	NMHC Peak Area:	194321
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	82.2	17.57	17.64	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.64	Prev response	17.61	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.01	----
High point	4932	82.2	17.57	17.47	1.006
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.006

Notes:

Changed H2 cylinder.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4932	82.2	9.30	9.34	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.34	Prev response	9.31	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	82.2	9.30	9.26	1.004
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.004

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4932	82.2	8.28	8.30	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.30	Prev response	8.31	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.01	----
High point	4932	82.2	8.28	8.21	1.008
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 1.008

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.004855	0.993779
THC Cal Offset:	-0.042179	0.005000
CH ₄ Cal Slope:	1.006101	0.991733
CH ₄ Cal Offset:	-0.019544	0.005000
NMHC Cal Slope:	1.004041	0.995600
NMHC Cal Offset:	-0.022835	0.000000


Calibration Performed By:

Mohammed Kashif

NMHC Calibration Plot

Date: October 15, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Station Information

Station Name: Lower Camp Station Number: AMS 11
Calibration Date: October 15, 2025 Prev Cal Date: September 24, 2024
Start Time (MST): 13:04 End Time (MST): 15:56
Tower Height (m): 10.0 Reason: Routine

Wind Speed Calibration

Sensor make/model: Met One 010C-1 Serial Number: X16480
WS Calibrator: MetOne 053 Serial Number: R10866

Shaft RPM (Hz)	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.5%
400	39.4	39.4	0.1%
600	58.6	58.7	0.2%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	0.999997	0.999999	≥ 0.9995
Calculated slope	1.000125	0.998854	$0.98 - 1.02$
Calculated intercept	-0.035284	-0.028010	± 2

Wind Direction Calibration

Sensor make/model: Met One 020C-1 Serial Number: X16480
As Found Declination (deg east of True North): 14 As Left Declination (deg east of True North): 13
Solar noon (MST): 13:11 Calc Declination*: 13.3 Degrees
WD Calibrator: Met One 040 ** - calculated declination as per NOAA website*

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 360° FS) <i>Limit = +/- 1%</i>
10	12.5	0.7%
90	91.2	0.3%
180	180.0	0.0%
270	272.9	0.8%
350	351.4	0.4%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	0.999987	0.999965	≥ 0.9995
Calculated slope	0.993333	1.000422	$0.97 - 1.03$
Calculated intercept	-0.195681	-1.663187	± 5

Notes: Bearings still good. Cross arm was aligned using compass after calibration.

Calibration Performed By: Mohammed Kashif



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13 FORT MCKAY SOUTH OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Fort McKay South Station number: AMS 13
Calibration Date: October 14, 2025 Last Cal Date: September 10, 2025
Start time (MST): 9:42 End time (MST): 12:39
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.55 ppm Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #: CC260812
Removed Cal Gas Conc: 50.55 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T750 Serial Number: 281
Zero Air Gen Model: Teledyne API 751H Serial Number: 321

Analyzer Information

Analyzer make: Teledyne API T100 Serial Number: 599
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007647	1.003832	Backgd or Offset:	111.3	111.3
Calibration intercept:	-2.914135	-3.054307	Coeff or Slope:	0.662	0.658

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4921	79.1	799.7	805.4	0.992
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	805.8	Previous response	802.9	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.4	----
High point	4921	79.1	799.7	801.0	0.998
Mid point	4960	39.5	399.4	396.6	1.007
Low point	4980	19.8	200.2	195.2	1.026
As left zero	5000	0.0	0.0	-0.4	----
As left span	4921	79.1	799.7	802.0	0.997

Average Correction Factor: 1.010

Notes: Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	October 14, 2025	Previous Calibration:	September 10, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:42	End Time (MST):	12:39
Analyzer make:	Teledyne API T100	Analyzer serial #:	599

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	0.999950	≥0.995
799.7	801.0	0.9984	Slope	1.003832	0.90 - 1.10
399.4	396.6	1.0070	Intercept	-3.054307	+/-30
200.2	195.2	1.0255			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: October 14, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort McKay South
Calibration Date: October 7, 2025
Start time (MST): 8:43
Reason: Routine

Station number: AMS 13
Last Cal Date: September 4, 2025
End time (MST): 12:33

Calibration Standards

Cal Gas Concentration: 4.88 ppm
Cal Gas Cylinder #: CC500241
Removed Cal Gas Conc: 4.88 ppm
Removed Gas Cyl #: CC500241
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Exp Date: September 5, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 2657
Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1180540017
Converter serial #: 521
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008570	0.999674	Backgd or Offset:	3.36	3.36
Calibration intercept:	-0.018403	-0.058418	Coeff or Slope:	1.09	1.09

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4918	81.6	79.6	78.4	1.019
As found Mid point	4959	40.8	39.8	39.0	1.026
As found Low point	4980	20.4	19.9	19.4	1.037
New cylinder response					
Baseline Corr As found:	78.2	Prev response:	80.31	*% change:	-2.7%
Baseline Corr 2nd AF pt:	38.8	AF Slope:	0.982886	AF Intercept:	0.001558
Baseline Corr 3rd AF pt:	19.2	AF Correlation:	0.999970	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4918	81.6	79.6	79.7	0.999
Mid point	4959	40.8	39.8	39.6	1.006
Low point	4980	20.4	19.9	19.6	1.016
As left zero	5000	0.0	0.0	0.2	----
As left span	4918	81.6	79.6	79.8	0.998
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:	7-Aug-25		Ave Corr Factor		1.007
Date of last converter efficiency test:					

Notes: Changed inlet filter after as found. SO2 scrubber check after calibrator zero and passed. No adjustment.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	October 7, 2025	Previous Calibration:	September 4, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:43	End Time (MST):	12:33
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999952		≥ 0.995
79.6	79.7	0.9993	Slope	0.999674		$0.90 - 1.10$
39.8	39.6	1.0056	Intercept	-0.058418		± 3
19.9	19.6	1.0158				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Fort McKay South
 Calibration Date: October 14, 2025
 Start time (MST): 9:42
 Reason: Routine

Station number: AMS 13
 Last Cal Date: September 10, 2025
 End time (MST): 12:39

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	Friday, December 29, 2028
CH ₄ Cal Gas Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T750	Serial Number:	281
Zero Air Gen model:	Teledyne API 751H	Serial Number:	321

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1181490018
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.04E-04	3.00E-04	NMHC SP Ratio:	4.59E-05
CH ₄ Retention time:	16.20	16.20	NMHC Peak Area:	197639
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	17.05	17.16	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.16	Prev response	17.01	*% change	0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	17.05	17.11	0.996
Mid point	4960	39.5	8.51	8.50	1.001
Low point	4980	19.8	4.27	4.18	1.022
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	17.05	17.15	0.994
Average Correction Factor					1.006

Notes:

Changed inlet filter after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.1	9.08	9.07	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.07	Prev response	9.06	*% change	0.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.1	9.08	9.10	0.998
Mid point	4960	39.5	4.53	4.53	1.002
Low point	4980	19.8	2.27	2.22	1.024
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	9.08	9.11	0.997
Average Correction Factor					1.008

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.1	7.97	8.10	0.984
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.10	Prev response	7.95	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.1	7.97	8.01	0.995
Mid point	4960	39.5	3.98	3.98	1.000
Low point	4980	19.8	1.99	1.96	1.019
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	7.97	8.04	0.991
Average Correction Factor					1.005

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.000317	1.005364
THC Cal Offset:	-0.037504	-0.049894
CH ₄ Cal Slope:	1.001350	1.007187
CH ₄ Cal Offset:	-0.023945	-0.023539
NMHC Cal Slope:	0.999738	1.003878
NMHC Cal Offset:	-0.014359	-0.026555

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 14, 2025	Previous Calibration:	September 10, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:42	End Time (MST):	12:39
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999957	<i>≥0.995</i>
17.05	17.11	0.9965	Slope	1.005364	<i>0.90 - 1.10</i>
8.51	8.50	1.0011	Intercept	-0.049894	<i>+/-0.5</i>
4.27	4.18	1.0216			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 14, 2025	Previous Calibration:	September 10, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:42	End Time (MST):	12:39
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999959	≥ 0.995
7.97	8.01	0.9946	Slope	1.007187	$0.90 - 1.10$
3.98	3.98	1.0000	Intercept	-0.023539	± 0.5
1.99	1.96	1.0191			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 14, 2025	Previous Calibration:	September 10, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:42	End Time (MST):	12:39
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999956	<i>≥0.995</i>
9.08	9.10	0.9981	Slope	1.003878	<i>0.90 - 1.10</i>
4.53	4.53	1.0021	Intercept	-0.026555	<i>+/-0.5</i>
2.27	2.22	1.0238			



NMHC Calibration Plot

Date: October 14, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort McKay South
Station number: AMS 13
Calibration Date: October 20, 2025
Last Cal Date: September 23, 2025
Start time (MST): 10:05
End time (MST): 14:34
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2UP1RP
NOX Cal Gas Conc: 48.25 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.25 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne APIT701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 47.88 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.88 ppm
NO gas Diff:
Serial Number: 2657
Serial Number: 1118

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.0	-0.3	-0.7	----	----
AF High point	4917	83.5	805.7	799.5	6.2	745.4	740.3	5.1	1.0794	1.0796
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 802.9 ppb	NO = 798.4 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = -7.6%
Baseline Corr 1st pt	NO _x = 746.4 ppb	NO = 740.6 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -7.8%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12300522720

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001033	1.002566
NO _x Cal Offset:	-3.631831	-3.472366
NO Cal Slope:	1.004878	1.006494
NO Cal Offset:	-4.970210	-5.010714
NO ₂ Cal Slope:	0.982117	0.998355
NO ₂ Cal Offset:	-1.008395	0.110833

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.144	1.240	NO bkgnd or offset:	3.1	8.1
NOX coeff or slope:	0.997	0.997	NOX bkgnd or offset:	3.4	8.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	145.5	345.0

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	----	----
High point	4917	83.5	805.7	799.5	6.2	806.3	802.5	3.9	0.9992	0.9963
Mid point	4958	41.8	403.4	400.3	3.1	398.4	394.4	4.1	1.0125	1.0149
Low point	4979	20.9	201.7	200.1	1.5	195.8	192.2	3.6	1.0301	1.0413
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.1	----	----
As left span	4917	83.5	805.7	397.9	407.8	802.2	397.9	404.3	1.0044	1.0000
Average Correction Factor									1.0139	1.0175

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	797.8	395.4	408.6	407.8	1.0019	99.8%
Mid GPT point	797.8	600.1	203.9	204.2	0.9984	100.2%
Low GPT point	797.8	700.1	103.9	103.6	1.0027	99.7%
Average Correction Factor					1.0010	99.9%

Notes:

Adjusted zero and span.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 23, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:05	End Time (MST):	14:34
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999913	≥0.995
805.7	806.3	0.9992	Slope	1.002566	0.90 - 1.10
403.4	398.4	1.0125	Intercept	-3.472366	+/-20
201.7	195.8	1.0301			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 23, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:05	End Time (MST):	14:34
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999996	≥ 0.995
408.6	407.8	1.0019	Slope	0.998355	$0.90 - 1.10$
203.9	204.2	0.9984	Intercept	0.110833	± 20
103.9	103.6	1.0027			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 23, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:05	End Time (MST):	14:34
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	-0.1	----	Correlation Coefficient	0.999825	≥ 0.995
799.5	802.5	0.9963	Slope	1.006494	0.90 - 1.10
400.3	394.4	1.0149	Intercept	-5.010714	+/-20
200.1	192.2	1.0413			



NO_x Calibration Plot

Date: October 20, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort McKay South
Station number: AMS 13
Calibration Date: October 21, 2025
Last Cal Date: October 20, 2025
Start time (MST): 8:40
End time (MST): 13:37
Reason: Maintenance

Calibration Standards

NO Gas Cylinder #: T2UP1RP
NOX Cal Gas Conc: 48.25 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.25 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne APIT701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 47.88 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.88 ppm
NO gas Diff:
Serial Number: 2657
Serial Number: 1118

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
AF High point	4917	83.5	805.7	799.5	6.2	804.3	798.7	5.6	1.0015	1.0009
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 804.3 ppb	NO = 799.7 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.0%
Baseline Corr 1st pt	NO _x = 804.5 ppb	NO = 798.8 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -0.1%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12300522720

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.240	1.126
NOX coeff or slope:	0.997	0.998
NO2 coeff or slope:	1.000	1.000

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	8.1	3.0
NOX bkgnd or offset:	8.1	3.2
Reaction cell Press:	345.0	162.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002566	1.001701
NO _x Cal Offset:	-3.472366	-3.392379
NO Cal Slope:	1.006494	1.002536
NO Cal Offset:	-5.010714	-4.350297
NO ₂ Cal Slope:	0.998355	0.988647
NO ₂ Cal Offset:	0.110833	1.598208

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	0.0	----	----
High point	4917	83.5	805.7	799.5	6.2	805.5	799.6	5.8	1.0002	0.9999
Mid point	4958	41.8	403.4	400.3	3.1	398.7	394.3	4.4	1.0118	1.0152
Low point	4979	20.9	201.7	200.1	1.5	195.3	192.0	3.3	1.0327	1.0424
As left zero	5000	0.0	0.0	0.0	0.0	0.3	0.4	-0.1	----	----
As left span	4917	83.5	805.7	397.9	407.8	802.2	397.9	404.3	1.0044	1.0000
Average Correction Factor									1.0149	1.0192

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	799.1	398.1	407.2	402.8	1.0109	98.9%
Mid GPT point	799.1	601.3	204.0	205.7	0.9916	100.8%
Low GPT point	799.1	700.5	104.8	105.7	0.9913	100.9%
Average Correction Factor					0.9979	100.2%

Notes:

Adjusted zero and span.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 21, 2025	Previous Calibration:	October 20, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:40	End Time (MST):	13:37
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999914	≥0.995
805.7	805.5	1.0002	Slope	1.001701	0.90 - 1.10
403.4	398.7	1.0118	Intercept	-3.392379	+/-20
201.7	195.3	1.0327			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 21, 2025	Previous Calibration:	October 20, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:40	End Time (MST):	13:37
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999880	≥0.995
407.2	402.8	1.0109	Slope	0.988647	0.90 - 1.10
204.0	205.7	0.9916	Intercept	1.598208	+/-20
104.8	105.7	0.9913			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 21, 2025	Previous Calibration:	October 20, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	8:40	End Time (MST):	13:37
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999852	≥ 0.995
799.5	799.6	0.9999	Slope	1.002536	$0.90 - 1.10$
400.3	394.3	1.0152	Intercept	-4.350297	± 20
200.1	192.0	1.0424			



NO_x Calibration Plot

Date: October 21, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Fort McKay South Station number: AMS 13
Calibration Date: October 6, 2025 Last Cal Date: September 2, 2025
Start time (MST): 9:59 End time (MST): 12:54
Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T750 Serial Number: 281
ZAG Make/Model: Teledyne API 751H Serial Number: 321

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 7413
Analyzer Range: 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001257	1.001486	Backgd or Offset:	-3.6	-3.6
Calibration intercept:	1.680000	1.140000	Coeff or Slope:	1.030	1.030

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	0.6	----
As found High point	5000	966.7	400.0	401.7	0.997
As found Mid point					
As found Low point					
Baseline Corr As found:	401.1	Previous response	402.2	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.6	----
High point	5000	970.7	400.0	401.5	0.996
Mid point	5000	792.7	200.0	201.6	0.992
Low point	5000	677.1	100.0	101.9	0.981
As left zero	5000	800.0	0.0	0.8	----
As left span	5000	970.7	400.0	405.5	0.986
Average Correction Factor					0.990

Notes: Sample inlet filter was changed after as founds. No adjustment.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	October 6, 2025	Previous Calibration:	September 2, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:59	End Time (MST):	12:54
Analyzer make:	Teledyne API T400	Analyzer serial #:	7413

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999991	≥ 0.995
400.0	401.5	0.9963	Slope	1.001486	$0.90 - 1.10$
200.0	201.6	0.9921	Intercept	1.140000	± 5
100.0	101.9	0.9814			



O₃ Calibration Plot

Date: October 6, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay South Station number: AMS 13
Calibration Date: October 30, 2025 Last Cal Date: September 23, 2025
Start time (MST): 12:14 End time (MST): 12:33

Analyzer Make: Teledyne API T640 S/N: 1335
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388746
Temp/RH standard: Alicat FP-25 S/N: 388746

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	1.8	1.31	1.8	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	740.3	742.64	740.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.98	5.04	4.98	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	43	----	43	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	3.7	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: September 23, 2025
Date Disposable Filter Changed: September 23, 2025

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: June 2, 2025
Date RH/T Sensor Cleaned: June 2, 2025

Notes: PMT peak test adjusted.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC
OCTOBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Anzac Station number: AMS 14
Calibration Date: October 16, 2025 Last Cal Date: September 8, 2025
Start time (MST): 8:46 End time (MST): 12:16
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.32 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC462030
Removed Cal Gas Conc: 50.32 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3060
Zero Air Gen Model: API T701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 0710321322
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.013794	1.013696	Backgd or Offset:	25.5	25.7
Calibration intercept:	-1.021986	-1.262634	Coeff or Slope:	1.111	1.111

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4937	79.6	798.4	805.3	0.992
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	805.1	Previous response	808.4	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4937	79.6	798.4	809.1	0.987
Mid point	4977	39.7	398.2	400.7	0.994
Low point	4992	19.7	197.8	198.7	0.995
As left zero	5000	0.0	0.0	0.4	----
As left span	4937	79.6	798.4	808.1	0.988
Average Correction Factor:					0.992

Notes: Changed the sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 16, 2025	Previous Calibration:	September 8, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	8:46	End Time (MST):	12:16
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.0	----	Correlation Coefficient	0.999984	≥0.995
798.4	809.1	0.9868	Slope	1.013696	0.90 - 1.10
398.2	400.7	0.9938	Intercept	-1.262634	+/-30
197.8	198.7	0.9955			



SO2 Calibration Plot

Date: October 16, 2025

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Anzac Station number: AMS 14
Calibration Date: October 28, 2025 Last Cal Date: September 25, 2025
Start time (MST): 9:50 End time (MST): 15:22
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.15 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC510379
Removed Cal Gas Conc: 5.15 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3060
ZAG Make/Model: API 701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153582
Converter make: CD Nova CDN-101 Converter serial #: 631
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001176	1.001464	Backgd or Offset:	2.25 2.31
Calibration intercept:	0.001442	-0.058652	Coeff or Slope:	0.957 0.990

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4922	77.7	80.0	78.8	1.014
As found Mid point	4961	38.9	40.1	39.2	1.020
As found Low point	4981	19.4	20.0	19.3	1.030
New cylinder response					
Baseline Corr As found:	78.9	Prev response:	80.13	*% change:	-1.6%
Baseline Corr 2nd AF pt:	39.3	AF Slope:	0.986796	AF Intercept:	-0.258585
Baseline Corr 3rd AF pt:	19.4	AF Correlation:	0.999982	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4922	77.7	80.0	80.1	0.999
Mid point	4961	38.9	40.1	40.1	0.999
Low point	4981	19.4	20.0	19.7	1.014
As left zero	5000	0.0	0.0	0.3	----
As left span	4922	77.7	80.0	78.7	1.017
SO2 Scrubber Check	4921	79.5	794.9	0.0	----
Date of last scrubber change:	28-May-25		Ave Corr Factor		1.004
Date of last converter efficiency test:	September 16, 2025		103.4% efficiency		

Notes: Changed sample inlet filter after as founds. Ran scrubber check after calibrator zero and it passed.
Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	October 28, 2025	Previous Calibration:	September 25, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	9:50	End Time (MST):	15:22
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999975		≥ 0.995
80.0	80.1	0.9988	Slope	1.001464		$0.90 - 1.10$
40.1	40.1	0.9988	Intercept	-0.058652		± 3
20.0	19.7	1.0138				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Anzac
Calibration Date: October 16, 2025
Start time (MST): 8:46
Reason: Routine

Station number: AMS 14
Last Cal Date: September 8, 2025
End time (MST): 12:16

Calibration Standards

Gas Cert Reference:	CC462030	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	505.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	204.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	505.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	204.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
Zero Air Gen model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1331259521
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.99E-04	2.98E-04	NMHC SP Ratio:	5.88E-05
CH ₄ Retention time:	15.1	15.1	NMHC Peak Area:	152010
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4937	79.6	16.96	16.75	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.75	Prev response	16.88	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4937	79.6	16.96	16.99	0.998
Mid point	4977	39.7	8.46	8.29	1.020
Low point	4992	19.7	4.20	4.07	1.032
As left zero	5000	0.0	0.00	0.00	----
As left span	4937	79.6	16.96	16.91	1.003
Average Correction Factor					1.003

Notes: Changed the sample inlet filter and H₂ cylinder after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4937	79.6	8.94	8.83	1.013
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.83	Prev response	8.95	*% change	-1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4937	79.6	8.94	8.95	1.000
Mid point	4977	39.7	4.46	4.41	1.012
Low point	4992	19.7	2.21	2.17	1.022
As left zero	5000	0.0	0.00	0.00	----
As left span	4937	79.6	8.94	8.90	1.005
Average Correction Factor					1.011

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4937	79.6	8.02	7.93	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.93	Prev response	7.94	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4937	79.6	8.02	8.05	0.996
Mid point	4977	39.7	4.00	3.89	1.029
Low point	4992	19.7	1.99	1.90	1.044
As left zero	5000	0.0	0.00	0.00	----
As left span	4937	79.6	8.02	8.01	1.001
Average Correction Factor					1.023

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000314	1.003531
THC Cal Offset:	-0.080469	-0.092035
CH ₄ Cal Slope:	0.996730	1.005653
CH ₄ Cal Offset:	-0.052231	-0.061970
NMHC Cal Slope:	1.003798	1.001628
NMHC Cal Offset:	-0.029037	-0.030065

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 16, 2025	Previous Calibration:	September 8, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	8:46	End Time (MST):	12:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999836	≥ 0.995
16.96	16.99	0.9981	Slope	1.003531	$0.90 - 1.10$
8.46	8.29	1.0199	Intercept	-0.092035	± 0.5
4.20	4.07	1.0322			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 16, 2025	Previous Calibration:	September 8, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	8:46	End Time (MST):	12:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			Limits
0.00	0.00	----	Correlation Coefficient	0.999653		≥0.995
8.02	8.05	0.9965	Slope	1.005653		0.90 - 1.10
4.00	3.89	1.0290	Intercept	-0.061970		+/-0.5
1.99	1.90	1.0443				





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 16, 2025	Previous Calibration:	September 8, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	8:46	End Time (MST):	12:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999942	≥ 0.995
8.94	8.95	0.9995	Slope	1.001628	$0.90 - 1.10$
4.46	4.41	1.0118	Intercept	-0.030065	± 0.5
2.21	2.17	1.0216			



NMHC Calibration Plot

Date: October 16, 2025

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Anzac
Station number: AMS 14
Calibration Date: October 21, 2025
Last Cal Date: September 10, 2025
Start time (MST): 10:27
End time (MST): 15:18
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0037092
NOX Cal Gas Conc: 60.7 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.70 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T700H
Cal Gas Expiry Date: May 16, 2031
NO Cal Gas Conc: 60.40 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.40 ppm
NO gas Diff:
Serial Number: 3060
Serial Number: 357

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
AF High point	4934	66.3	804.8	800.9	4.0	805.2	801.2	4.0	0.9995	0.9994
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 803.4 ppb	NO = 801.4 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = 0.2%			
Baseline Corr 1st pt	NO _x = 805.2 ppb	NO = 801.3 ppb	<u>As Found Statistics</u>			*Percent Change	NO = 0.0%			
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:	NO Int:			
			As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430008

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.435	1.435	NO bkgnd or offset:	3.9	3.9
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	157.5	156.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998607	0.999346
NO _x Cal Offset:	-0.289967	-0.650114
NO Cal Slope:	1.003538	1.003009
NO Cal Offset:	-2.329840	-2.169606
NO ₂ Cal Slope:	0.994357	0.996014
NO ₂ Cal Offset:	-1.018568	-0.692405

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.2	----	----
High point	4934	66.3	804.8	800.9	4.0	803.9	802.0	2.0	1.0012	0.9986
Mid point	4985	33.2	401.6	399.6	2.0	400.9	398.2	2.7	1.0017	1.0035
Low point	5004	16.7	201.9	200.9	1.0	199.7	196.6	3.2	1.0110	1.0219
As left zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.2	----	----
As left span	4934	66.3	804.8	408.9	395.9	804.4	408.9	395.5	1.0005	1.0000
Average Correction Factor									1.0046	1.0080

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	799.3	406.8	396.5	394.8	1.0042	99.6%
Mid GPT point	799.3	605.6	197.7	195.3	1.0122	98.8%
Low GPT point	799.3	704.1	99.2	97.5	1.0172	98.3%
Average Correction Factor					1.0112	98.9%

Notes:

Sample inlet filter changed after as founds. No adjustments made.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 21, 2025	Previous Calibration:	September 10, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:27	End Time (MST):	15:18
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999991	≥0.995
804.8	803.9	1.0012	Slope	0.999346	0.90 - 1.10
401.6	400.9	1.0017	Intercept	-0.650114	+/-20
201.9	199.7	1.0110			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 21, 2025	Previous Calibration:	September 10, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:27	End Time (MST):	15:18
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999973	≥0.995
396.5	394.8	1.0042	Slope	0.996014	0.90 - 1.10
197.7	195.3	1.0122	Intercept	-0.692405	+/-20
99.2	97.5	1.0172			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 21, 2025	Previous Calibration:	September 10, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:27	End Time (MST):	15:18
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999961	≥ 0.995
800.9	802.0	0.9986	Slope	1.003009	0.90 - 1.10
399.6	398.2	1.0035	Intercept	-2.169606	+/-20
200.9	196.6	1.0219			



NO_x Calibration Plot

Date: October 21, 2025

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Anzac
Calibration Date: October 1, 2025
Start time (MST): 10:14
Reason: Routine

Station number: AMS 14
Last Cal Date: September 3, 2025
End time (MST): 13:14

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: API T700
ZAG Make/Model: API 701H

Serial Number: 3060
Serial Number: 357

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1426262595

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001743	1.001171	Backgd or Offset:	1.6	1.6
Calibration intercept:	0.020000	0.620000	Coeff or Slope:	1.712	1.712

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.9	----
As found High point	5000	935.9	400.0	400.6	1.001
As found Mid point					
As found Low point					
Baseline Corr As found:	399.7	Previous response	400.7	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.9	----
High point	5000	935.9	400.0	401.3	0.997
Mid point	5000	817.5	200.0	200.5	0.998
Low point	5000	722.8	100.0	100.6	0.994
As left zero	5000	0.0	0.0	0.3	----
As left span	5000	935.9	400.0	404.3	0.989
Average Correction Factor					0.996

Notes: Sample inlet filter changed after asfound. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	October 1, 2025	Previous Calibration:	September 3, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:14	End Time (MST):	13:14
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.9	----	Correlation Coefficient	0.999997	≥0.995
400.0	401.3	0.9968	Slope	1.001171	0.90 - 1.10
200.0	200.5	0.9975	Intercept	0.620000	+/- 5
100.0	100.6	0.9940			



O₃ Calibration Plot

Date: October 1, 2025

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Anzac
Calibration Date: October 6, 2025
Start time (MST): 14:05
Station number: AMS 14
Last Cal Date: September 16, 2025
End time (MST): 14:19
Analyzer Make: AP T640
S/N: 825
Particulate Fraction: PM2.5
Flow Meter Make/Model: Alicat FP-25BT
S/N: 388749
Temp/RH standard: Alicat FP-25BT
S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	5.5	3.66	5.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	718.1	719.25	718.1	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.020	4.95	5.020	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	-----	37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 3.7		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: Expiry Date:
Lot No.:

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: August 26, 2025
Date Disposable Filter Changed: August 26, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 26, 2025
Date RH/T Sensor Cleaned: August 26, 2025

Notes: No adjustments made. Leak check passed. Head cleaned

Calibration by: Mohammed Kashif



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS17
WAPASU
OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Wapasu Station number: AMS17
Calibration Date: October 15, 2025 Last Cal Date: September 15, 2025
Start time (MST): 10:17 End time (MST): 13:18
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.12 ppm Cal Gas Exp Date: April 9, 2033
Cal Gas Cylinder #: CC422255
Removed Cal Gas Conc: 50.12 ppm Rem Gas Exp Date:
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 2449
Zero Air Gen Model: Teledyne API 701H Serial Number: 1238

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1218153459
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011892	1.007363	Backgd or Offset:	14.2	15.8
Calibration intercept:	-1.009539	-2.875904	Coeff or Slope:	1.133	1.133

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.5	----
As found High point	4920	79.8	799.9	803.3	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	802.8	Previous response	808.5	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.6	----
High point	4920	79.8	799.9	804.0	0.995
Mid point	4960	39.9	400.0	399.2	1.002
Low point	4980	20.0	200.5	196.6	1.020
As left zero	5000	0.0	0.0	-0.8	----
As left span	4920	79.8	799.9	803.2	0.996
Average Correction Factor:					1.006

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 15, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:17	End Time (MST):	13:18
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.6	----	Correlation Coefficient	0.999963	≥0.995
799.9	804.0	0.9950	Slope	1.007363	0.90 - 1.10
400.0	399.2	1.0019	Intercept	-2.875904	+/-30
200.5	196.6	1.0197			



SO2 Calibration Plot

Date: October 15, 2025

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Wapasu
Calibration Date: October 20, 2025
Start time (MST): 10:00
Reason: Routine
Station number: AMS 17
Last Cal Date: September 24, 2025
End time (MST): 14:33

Calibration Standards

Cal Gas Concentration: 4.77 ppm
Cal Gas Cylinder #: DT20029267
Removed Cal Gas Conc: 4.77 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API T701H
Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: Thermo 450i
Converter make: CD Nova
Analyzer Range: 0 - 100 ppb
Analyzer serial #: 1218153583
Converter serial #: N/A
Converter Temp: 340 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001068	0.989072	Backgd or Offset:	13.1	13.1
Calibration intercept:	0.000213	0.020210	Coeff or Slope:	1.099	1.099

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4921	83.9	80.0	80.3	0.995
As found Mid point	4961	41.9	39.9	39.9	0.999
As found Low point	4980	21.0	20.0	19.7	1.012
New cylinder response					
Baseline Corr As found:	80.4	Prev response:	80.05	*% change:	0.4%
Baseline Corr 2nd AF pt:	40.0	AF Slope:	1.006689	AF Intercept:	-0.269437
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999979	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4916	83.9	80.0	79.3	1.009
Mid point	4958	41.9	40.0	39.5	1.012
Low point	4979	21.0	20.0	19.5	1.027
As left zero	5000	0.0	0.0	0.4	----
As left span	4916	83.9	80.0	76.5	1.046
SO2 Scrubber Check	4921	79.4	793.9	-0.1	----
Date of last scrubber change:		N/A		Ave Corr Factor	1.016
Date of last converter efficiency test:		N/A			

Notes:

No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 24, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:00	End Time (MST):	14:33
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999940	≥ 0.995
80.0	79.3	1.0094	Slope	0.989072	$0.90 - 1.10$
40.0	39.5	1.0120	Intercept	0.020210	± 3
20.0	19.5	1.0274			



H₂S Calibration Plot

Date: October 20, 2025

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Wapasu Station number: AMS17
Calibration Date: October 15, 2025 Last Cal Date: September 9, 2025
Start time (MST): 10:17 End time (MST): 13:18
Reason: Routine

Calibration Standards

Gas Cert Reference: CC422255 Cal Gas Expiry Date: April 9, 2033
CH4 Cal Gas Conc. 499.6 ppm CH4 Equiv Conc. 1059.0 ppm
C3H8 Cal Gas Conc. 203.4 ppm
Removed Gas Cert: n/a Removed Gas Expiry:
Removed CH4 Conc. 499.6 ppm CH4 Equiv Conc. 1059.0 ppm
Removed C3H8 Conc. 203.4 ppm Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 2449
ZAG Make/Model: Teledyne API 701H Serial Number: 1238

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1218153352
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.013335	1.000596	Background:	3.280
Calibration intercept:	-0.086557	-0.100133	Coefficient:	4.477
				4.376

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	-0.04	----
As found High point	4920	79.8	16.90	17.28	0.976
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.32	Previous response	17.04	*% change	1.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	-0.07	----
High point	4920	79.8	16.90	16.84	1.004
Mid point	4960	39.9	8.45	8.30	1.018
Low point	4980	20.0	4.24	4.13	1.026
As left zero	5000	0.0	0.00	-0.03	----
As left span	4921	79.4	16.81	16.95	0.992
Average Correction Factor					1.016

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 9, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:17	End Time (MST):	13:18
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.07	----	Correlation Coefficient	0.999967	≥ 0.995
16.90	16.84	1.0036	Slope	1.000596	$0.90 - 1.10$
8.45	8.30	1.0180	Intercept	-0.100133	± 1.5
4.24	4.13	1.0259			


THC Calibration Curve



THC Calibration Plot

Date: October 15, 2025

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Wapasu
Station number: AMS 17
Calibration Date: October 30, 2025
Last Cal Date: September 25, 2025
Start time (MST): 10:47
End time (MST): 16:00
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0045177
NOX Cal Gas Conc: 61.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 61.30 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: July 19, 2032
NO Cal Gas Conc: 61.00 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 61.00 ppm
NO gas Diff:
Serial Number: 2449
Serial Number: 359

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
AF High point	4934	65.6	804.3	800.4	3.9	779.3	775.7	3.6	1.0316	1.0317
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 803.5 ppb	NO = 800.0 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -3.1%	
Baseline Corr 1st pt	NO _x = 779.7 ppb	NO = 775.8 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -3.1%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo Scientific 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153460

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.827	0.855	NO bkgnd or offset:	2.8	2.9
NOX coeff or slope:	0.994	0.997	NOX bkgnd or offset:	3.1	3.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	191.0	195.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000474	1.000389
NO _x Cal Offset:	-1.180012	-1.400017
NO Cal Slope:	1.002753	1.002210
NO Cal Offset:	-2.580010	-2.640011
NO ₂ Cal Slope:	0.997456	0.999619
NO ₂ Cal Offset:	-0.113201	-0.134470

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.3	----	----
High point	4934	65.6	804.3	800.4	3.9	804.1	801.0	3.2	1.0003	0.9992
Mid point	4967	32.8	402.1	400.2	2.0	399.5	396.6	2.9	1.0066	1.0090
Low point	4983	16.4	201.1	200.1	1.0	199.1	195.6	3.5	1.0100	1.0230
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
As left span	4934	65.6	804.3	398.6	405.7	805.6	398.6	406.9	0.9984	1.0000
Average Correction Factor									1.0056	1.0104

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.3	----	----
High GPT point	799.0	395.4	407.5	407.1	1.0011	99.9%
Mid GPT point	799.0	599.4	203.5	203.6	0.9997	100.0%
Low GPT point	799.0	701.7	101.2	101.1	1.0013	99.9%
Average Correction Factor					1.0007	99.9%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 30, 2025	Previous Calibration:	September 25, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:47	End Time (MST):	16:00
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999987	≥0.995
804.3	804.1	1.0003	Slope	1.000389	0.90 - 1.10
402.1	399.5	1.0066	Intercept	-1.400017	+/-20
201.1	199.1	1.0100			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 30, 2025	Previous Calibration:	September 25, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:47	End Time (MST):	16:00
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999999	≥0.995
407.5	407.1	1.0011	Slope	0.999619	0.90 - 1.10
203.5	203.6	0.9997	Intercept	-0.134470	+/-20
101.2	101.1	1.0013			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 30, 2025	Previous Calibration:	September 25, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:47	End Time (MST):	16:00
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999949	≥ 0.995
800.4	801.0	0.9992	Slope	1.002210	$0.90 - 1.10$
400.2	396.6	1.0090	Intercept	-2.640011	± 20
200.1	195.6	1.0230			



NO_x Calibration Plot

Date: October 30, 2025

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Wapasu Station number: AMS17
Calibration Date: October 2, 2025 Last Cal Date: September 4, 2025
Start time (MST): 10:02 End time (MST): 13:02
Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: API T700 Serial Number: 2449
ZAG Make/Model: API T701H Serial Number: 359

Analyzer Information

Analyzer make: Thermo Scientific 49i Analyzer serial #: 1501663734
Analyzer Range: 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003629	1.010714	Backgd or Offset:	0.1	0.1
Calibration intercept:	-1.260000	-1.700000	Coeff or Slope:	1.043	1.043

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.5	----
As found High point	5000	1104.7	400.0	400.8	0.997
As found Mid point					
As found Low point					
Baseline Corr As found:	401.3	Previous response	400.2	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	5000	1104.7	400.0	403.2	0.992
Mid point	5000	917.3	200.0	200.0	1.000
Low point	5000	797.9	100.0	97.8	1.022
As left zero	5000	0.0	0.0	-0.2	----
As left span	5000	1104.7	400.0	408.5	0.979
Average Correction Factor					1.005

Notes: No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	October 2, 2025	Previous Calibration:	September 4, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	10:02	End Time (MST):	13:02
Analyzer make:	Thermo Scientific 49i	Analyzer serial #:	1501663734

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999944	≥0.995
400.0	403.2	0.9921	Slope	1.010714	0.90 - 1.10
200.0	200.0	1.0000	Intercept	-1.700000	+/- 5
100.0	97.8	1.0225			



O₃ Calibration Plot

Date: October 2, 2025

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Wapasu Station number: AMS 17
Calibration Date: October 30, 2025 Last Cal Date: September 25, 2025
Start time (MST): 13:35 End time (MST): 15:17

Analyzer Make: Teledyne API T640 S/N: 1183
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-0.80	-1.00	-0.80	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	718.40	719.30	718.40	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.07	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39	----	39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 7.5		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 6, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: September 25, 2025
Date Disposable Filter Changed: October 30, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 21, 2025
Date RH/T Sensor Cleaned: July 21, 2025

Notes:

Monthly checks done. DFU filter changed out.

Calibration by: Aswin Sasi Kumar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: October 23, 2025 Last Cal Date: September 18, 2025
Start time (MST): 11:15 End time (MST): 14:30
Reason: Routine

Calibration Standards

Cal Gas Concentration: 51.22 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC417455
Removed Cal Gas Conc: 51.22 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: CC417455 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 2658
Zero Air Gen Model: Teledyne API 701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1501301453
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008031	1.000714	Backgd or Offset:	25.9	25.9
Calibration intercept:	-3.345346	-1.637815	Coeff or Slope:	0.816	0.816

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4921	78.1	800.2	796.8	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.1	Previous response	803.3	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	78.1	800.2	799.9	1.000
Mid point	4960	39.1	400.6	398.6	1.005
Low point	4981	19.5	199.7	196.5	1.016
As left zero	5000	0.0	0.0	0.0	----
As left span	4921	78.1	800.2	799.7	1.001
Average Correction Factor:					1.007

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 18, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:15	End Time (MST):	14:30
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999981	≥0.995
800.2	799.9	1.0004	Slope	1.000714	0.90 - 1.10
400.6	398.6	1.0050	Intercept	-1.637815	+/-30
199.7	196.5	1.0165			



SO2 Calibration Plot

Date: October 23, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS18
Calibration Date: October 29, 2025 Last Cal Date: September 17, 2025
Start time (MST): 10:21 End time (MST): 15:45
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.86 ppm Cal Gas Exp Date: May 9, 2027
Cal Gas Cylinder #: CC523103
Removed Cal Gas Conc: 4.86 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
ZAG Make/Model: Teledyne API T701 Serial Number: 360

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359
Converter make: CD Nova CDN-101 Converter serial #: 555
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010081	0.992077	Backgd or Offset:	2.94
Calibration intercept:	-0.099139	-0.018635	Coeff or Slope:	1.181

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4917	82.3	80.0	83.5	0.959
As found Mid point	4958	41.2	40.1	41.7	0.963
As found Low point	4979	20.6	20.0	20.3	0.991
New cylinder response					
Baseline Corr As found:	83.4	Prev response:	80.71	*% change:	3.2%
Baseline Corr 2nd AF pt:	41.6	AF Slope:	1.045081	AF Intercept:	-0.199906
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999926	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4917	82.3	80.0	79.7	1.004
Mid point	4958	41.2	40.1	39.2	1.022
Low point	4979	20.6	20.0	19.5	1.027
As left zero	5000	0.0	0.0	0.3	----
As left span	4917	82.3	80.0	77.6	1.031
SO2 Scrubber Check	4923	77.1	771.0	-0.2	----
Date of last scrubber change:	17-Dec-21		Ave Corr Factor		1.018
Date of last converter efficiency test:					

Notes: No adjustment made. SOX scrubber tested post filter change, no issues to note.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	October 29, 2025	Previous Calibration:	September 17, 2025
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:21	End Time (MST):	15:45
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999775		≥ 0.995
80.0	79.7	1.0038	Slope	0.992077		$0.90 - 1.10$
40.1	39.2	1.0218	Intercept	-0.018635		± 3
20.0	19.5	1.0269				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Stony Mountain
 Calibration Date: October 23, 2025
 Start time (MST): 11:15
 Reason: Routine

Station number: AMS 18
 Last Cal Date: September 18, 2025
 End time (MST): 14:31

Calibration Standards

Gas Cert Reference:	XC026809B	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
Zero Air Gen model:	Teledyne API T701	Serial Number:	4890

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1170050130
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.65E-04	2.68E-04	NMHC SP Ratio:	4.88E-05
CH ₄ Retention time:	14.8	14.8	NMHC Peak Area:	183174
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.01	----
As found High point	4921	78.1	16.82	16.45	1.023
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.44	Prev response	16.78	*% change	-2.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.02	----
High point	4921	78.1	16.82	16.85	0.998
Mid point	4960	39.1	8.42	8.40	1.003
Low point	4981	19.5	4.20	4.17	1.008
As left zero	5000	0.0	0.00	0.02	----
As left span	4921	78.1	16.82	16.80	1.001
Average Correction Factor					1.003

Notes: Changed the inlet filter and the N₂ cylinder after as founds. Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.1	8.93	8.64	1.034
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.64	Prev response	8.90	*% change	-3.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.1	8.93	8.96	0.997
Mid point	4960	39.1	4.47	4.45	1.004
Low point	4981	19.5	2.23	2.21	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	8.93	8.94	0.999
Average Correction Factor					1.004

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.01	----
As found High point	4921	78.1	7.89	7.81	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.80	Prev response	7.88	*% change	-1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.02	----
High point	4921	78.1	7.89	7.89	0.999
Mid point	4960	39.1	3.95	3.94	1.001
Low point	4981	19.5	1.97	1.96	1.005
As left zero	5000	0.0	0.00	0.02	----
As left span	4921	78.1	7.89	7.86	1.004
Average Correction Factor					1.002

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.000104	1.001524
THC Cal Offset:	-0.044554	-0.012726
CH ₄ Cal Slope:	1.001469	0.999248
CH ₄ Cal Offset:	-0.023484	0.004620
NMHC Cal Slope:	0.999092	1.003865
NMHC Cal Offset:	-0.021076	-0.018145

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 18, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:15	End Time (MST):	14:31
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.02	----	Correlation Coefficient	0.999985	≥ 0.995
16.82	16.85	0.9982	Slope	1.001524	$0.90 - 1.10$
8.42	8.40	1.0026	Intercept	-0.012726	± 0.5
4.20	4.17	1.0075			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 18, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:15	End Time (MST):	14:31
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.02	----	Correlation Coefficient	0.999989	≥ 0.995
7.89	7.89	0.9994	Slope	0.999248	$0.90 - 1.10$
3.95	3.94	1.0013	Intercept	0.004620	± 0.5
1.97	1.96	1.0045			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 18, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:15	End Time (MST):	14:31
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999979		<i>≥0.995</i>
8.93	8.96	0.9969	Slope	1.003865		<i>0.90 - 1.10</i>
4.47	4.45	1.0040	Intercept	-0.018145		<i>+/-0.5</i>
2.23	2.21	1.0102				



NMHC Calibration Plot

Date: October 23, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Stony Mountain
Station number: AMS 18
Calibration Date: October 31, 2025
Last Cal Date: September 22, 2025
Start time (MST): 10:21
End time (MST): 15:18
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0045516
NO_x Cal Gas Conc: 60.30 ppm
Removed Cylinder #: N/A
Removed Gas NO_x Conc: 60.30 ppm
NO_x gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API 701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 60.10 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 60.10 ppm
NO gas Diff:
Serial Number: 2658
Serial Number: 4890

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
AF High point	4933	66.6	803.3	800.6	2.7	802.7	798.3	4.4	1.0007	1.0029
AF Mid point										
AF Low point										
New cyl resp										

Previous Respo 4933 NO_x = 801.8 ppb NO = 799.4 ppb * = > +/-5% change initiates investigation *Percent Change NO_x = 0.1%
Baseline Corr 1st pt NO_x = 802.7 ppb NO = 798.3 ppb *Percent Change NO = -0.1%
Baseline Corr 2nd pt NO_x = NA ppb NO = NA ppb
Baseline Corr 3rd pt NO_x = NA ppb NO = NA ppb

As Found Statistics

As found NO_x r²: Nx SI: Nx Int:
As found NO r²: NO SI: NO Int:
As found NO₂ r²: NO₂ SI: NO₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	--	--

As Found GPT zero
As found high GPT point
As found mid GPT point
As found low GPT point



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo Scientific 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1501663731

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998994	0.999094
NO _x Cal Offset:	-0.608232	-0.868433
NO Cal Slope:	1.000822	0.999139
NO Cal Offset:	-1.809363	-1.969984
NO ₂ Cal Slope:	1.000029	0.998329
NO ₂ Cal Offset:	1.371521	0.878481

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.060	1.059	NO bkgnd or offset:	10.4	10.4
NOX coeff or slope:	0.999	0.999	NOX bkgnd or offset:	10.5	10.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	220.7	218.9

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.3	-0.1	----	----
High point	4933	66.6	803.3	800.6	2.7	802.2	799.1	3.1	1.0013	1.0019
Mid point	4967	33.3	401.6	400.2	1.3	399.8	396.7	3.1	1.0044	1.0089
Low point	4983	16.6	200.2	199.5	0.7	198.1	195.2	3.0	1.0107	1.0223
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
As left span	4933	66.6	803.3	386.4	416.9	803.2	386.4	416.9	1.0001	1.0000
Average Correction Factor									1.0055	1.0110

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	798.6	385.3	416.0	415.7	1.0006	99.9%
Mid GPT point	798.6	597.4	203.9	204.8	0.9954	100.5%
Low GPT point	798.6	699.1	102.2	103.9	0.9833	101.7%
Average Correction Factor					0.9931	100.7%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 31, 2025	Previous Calibration:	September 22, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:21	End Time (MST):	15:18
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999992	≥0.995
803.3	802.2	1.0013	Slope	0.999094	0.90 - 1.10
401.6	399.8	1.0044	Intercept	-0.868433	+/-20
200.2	198.1	1.0107			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 31, 2025	Previous Calibration:	September 22, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:21	End Time (MST):	15:18
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999975	≥0.995
416.0	415.7	1.0006	Slope	0.998329	0.90 - 1.10
203.9	204.8	0.9954	Intercept	0.878481	+/-20
102.2	103.9	0.9833			





Wood Buffalo Environmental Association

NO Calibration Summary


Station Information

Calibration Date:	October 31, 2025	Previous Calibration:	September 22, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:21	End Time (MST):	15:18
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999963	≥ 0.995
800.6	799.1	1.0019	Slope	0.999139	$0.90 - 1.10$
400.2	396.7	1.0089	Intercept	-1.969984	± 20
199.5	195.2	1.0223			


NO Calibration Curve



NO_x Calibration Plot

Date: October 31, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: October 22, 2025 Last Cal Date: September 11, 2025
Start time (MST): 11:10 End time (MST): 14:34
Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
ZAG Make/Model: Teledyne API 701H Serial Number: 4890

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 825
Analyzer Range: 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003086	0.995086	Backgd or Offset:	2.3	2.3
Calibration intercept:	0.460000	-0.040000	Coeff or Slope:	0.991	0.991

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	NA	0.0	-0.1	----
As found High point	4888	1138.1	400.0	398.7	1.003
As found Mid point					
As found Low point					
Baseline Corr As found:	398.8	Previous response	401.7	*% change	-0.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	NA	0.0	-0.5	----
High point	4888	1138.1	400.0	397.4	1.007
Mid point	4888	884.5	200.0	200.3	0.999
Low point	4888	741.4	100.0	99.2	1.008
As left zero	5000	NA	0.0	0.6	----
As left span	4812	1097.9	400.0	398.9	1.003
Average Correction Factor					1.004

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	October 22, 2025	Previous Calibration:	September 11, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	11:10	End Time (MST):	14:34
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.5	----	Correlation Coefficient	0.999972	≥0.995
400.0	397.4	1.0065	Slope	0.995086	0.90 - 1.10
200.0	200.3	0.9985	Intercept	-0.040000	+/- 5
100.0	99.2	1.0081			



O₃ Calibration Plot

Date: October 22, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: October 31, 2025 Last Cal Date: September 23, 2025
Start time (MST): 13:31 End time (MST): 15:07

Analyzer Make: API T640 S/N: 324
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2.3	1.7	2.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	699.3	699.80	699.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	4.98	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	32	----	32	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 5.1		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: August 28, 2025
Date Disposable Filter Changed: October 31, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: September 23, 2025
Date RH/T Sensor Cleaned: September 23, 2025

Notes: Flow, pressure and temp checked. Leak check passed. DFU filter changed out.

Calibration by: Aswin Sasi Kumar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS19 FIREBAG OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Firebag Station number: AMS 19
Calibration Date: October 6, 2025 Last Cal Date: September 15, 2025
Start time (MST): 11:06 End time (MST): 14:23
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.97 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC705799
Removed Cal Gas Conc: 50.97 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 1607
Zero Air Gen Model: Teledyne API T701H Serial Number: 201

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1410661308
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002290	1.002734	Backgd or Offset:	11.9	11.8
Calibration intercept:	0.620000	0.740000	Coeff or Slope:	1.030	1.015

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	4999	0.0	0.0	-0.5	----
As found High point	4922	78.4	799.2	810.0	0.986
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	810.5	Previous response	801.7	*% change	1.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	4999	0.0	0.0	0.0	----
High point	4922	78.4	799.2	802.0	0.997
Mid point	4961	39.2	399.6	401.1	0.996
Low point	4980	19.6	199.8	202.3	0.988
As left zero	4999	0.0	0.0	-0.4	----
As left span	4922	78.4	799.2	804.0	0.994
Average Correction Factor:					0.993

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 6, 2025	Previous Calibration:	September 15, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:06	End Time (MST):	14:23
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999994	≥0.995
799.2	802.0	0.9965	Slope	1.002734	0.90 - 1.10
399.6	401.1	0.9963	Intercept	0.740000	+/-30
199.8	202.3	0.9877			



SO2 Calibration Plot

Date: October 6, 2025

Location: Firebag





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Firebag Station number: AMS 19
Calibration Date: October 7, 2025 Last Cal Date: September 10, 2025
Start time (MST): 10:50 End time (MST): 16:34
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.29 ppm Cal Gas Exp Date: March 19, 2027
Cal Gas Cylinder #: DT0010492
Removed Cal Gas Conc: 5.29 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 1607
ZAG Make/Model: Teledyne API T701 Serial Number: 201

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680032
Converter make: Global Converter serial #: 2022-222
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001333	0.980615	Backgd or Offset:	2.64	2.66
Calibration intercept:	0.060000	0.060000	Coeff or Slope:	1.190	1.190

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4924	75.6	80.0	80.3	0.996
As found Mid point	4962	37.8	40.0	40.3	0.992
As found Low point	4981	18.9	20.0	20.2	0.990
New cylinder response					
Baseline Corr As found:	80.3	Prev response:	80.15	*% change:	0.2%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.003619	AF Intercept:	0.080000
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4924	75.6	80.0	78.5	1.019
Mid point	4962	37.8	40.0	39.3	1.018
Low point	4981	18.9	20.0	19.6	1.020
As left zero	5000	0.0	0.0	0.2	----
As left span	4924	75.6	80.0	77.6	1.031
SO2 Scrubber Check				0.0	
Date of last scrubber change:		18-Jan-23		Ave Corr Factor	1.019
Date of last converter efficiency test:		November 26, 2024		106.2% efficiency	

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	October 7, 2025	Previous Calibration:	September 10, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:50	End Time (MST):	16:34
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999998		≥ 0.995
80.0	78.5	1.0189	Slope	0.980615		$0.90 - 1.10$
40.0	39.3	1.0176	Intercept	0.060000		± 3
20.0	19.6	1.0202				



H2S Calibration Plot

Date: October 7, 2025

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Firebag Station number: AMS 19
Calibration Date: October 6, 2025 Last Cal Date: September 15, 2025
Start time (MST): 11:06 End time (MST): 14:23
Reason: Routine

Calibration Standards

Gas Cert Reference: CC705799 Cal Gas Expiry Date: October 9, 2032
CH4 Cal Gas Conc. 505.1 ppm CH4 Equiv Conc. 1066.9 ppm
C3H8 Cal Gas Conc. 204.3 ppm
Removed Gas Cert: Removed Gas Expiry:
Removed CH4 Conc. 505.1 ppm CH4 Equiv Conc. 1066.9 ppm
Removed C3H8 Conc. 204.3 ppm Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 1607
ZAG Make/Model: Teledyne API T701H Serial Number: 201

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1336160089
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994254	0.995149	Background:	2.30	2.16
Calibration intercept:	-0.025133	-0.015933	Coefficient:	3.841	3.860

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	4999	0.0	0.00	-0.16	----
As found High point	4922	78.4	16.73	16.55	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	16.71	Previous response	16.61	*% change	0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	4999	0.0	0.00	0.02	----
High point	4922	78.4	16.73	16.66	1.004
Mid point	4961	39.2	8.36	8.26	1.013
Low point	4980	19.6	4.18	4.13	1.012
As left zero	4999	0.0	0.00	-0.03	----
As left span	4922	78.4	16.73	16.79	0.996
Average Correction Factor					1.010

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	October 6, 2025	Previous Calibration:	September 15, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	11:06	End Time (MST):	14:23
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.02	----	Correlation Coefficient	0.999970	≥ 0.995
16.73	16.66	1.0042	Slope	0.995149	$0.90 - 1.10$
8.36	8.26	1.0127	Intercept	-0.015933	± 1.5
4.18	4.13	1.0123			


THC Calibration Curve



THC Calibration Plot

Date: October 6, 2025

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Firebag
Station number: AMS 19
Calibration Date: October 29, 2025
Last Cal Date: September 11, 2025
Start time (MST): 9:26
End time (MST): 14:06
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0044018
NO_x Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NO_x Conc: 48.90 ppm
NO_x gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.70 ppm
NO gas Diff:
Serial Number: 1607
Serial Number: 201

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
AF High point	4918	82.1	802.9	799.7	3.3	815.0	812.0	2.7	0.9850	0.9846
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 802.9 ppb	NO = 801.6 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 1.5%
Baseline Corr 1st pt	NO _x = 815.2 ppb	NO = 812.2 ppb	<u>As Found Statistics</u>	*Percent Change	NO = 1.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :	NO SI:	NO Int:
			As found NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1410661309

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997477	0.998175
NO _x Cal Offset:	1.940035	1.319743
NO Cal Slope:	1.001331	0.998416
NO Cal Offset:	0.899770	0.719909
NO ₂ Cal Slope:	0.996705	1.003548
NO ₂ Cal Offset:	-0.099258	-0.371604

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.938	0.922	NO bkgnd or offset:	4.7	4.6
NOX coeff or slope:	0.991	0.993	NOX bkgnd or offset:	4.7	4.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	165.1	163.0

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4918	82.1	802.9	799.7	3.3	802.0	798.7	3.1	1.0012	1.0012
Mid point	4959	41.1	402.0	400.3	1.6	403.6	400.9	2.7	0.9959	0.9985
Low point	4980	20.5	200.5	199.7	0.8	202.5	200.7	1.9	0.9901	0.9949
As left zero	5000	0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
As left span	4918	82.1	802.9	391.9	411.0	802.0	391.9	410.3	1.0012	1.0000
Average Correction Factor									0.9957	0.9982

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	797.6	391.4	409.5	410.9	0.9966	100.3%
Mid GPT point	797.6	596.7	204.2	203.9	1.0014	99.9%
Low GPT point	797.6	700.3	100.6	100.5	1.0008	99.9%
Average Correction Factor					0.9996	100.0%

Notes:

Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 29, 2025	Previous Calibration:	September 11, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:26	End Time (MST):	14:06
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999987	≥ 0.995
802.9	802.0	1.0012	Slope	0.998175	$0.90 - 1.10$
402.0	403.6	0.9959	Intercept	1.319743	± 20
200.5	202.5	0.9901			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 29, 2025	Previous Calibration:	September 11, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:26	End Time (MST):	14:06
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999993	≥0.995
409.5	410.9	0.9966	Slope	1.003548	0.90 - 1.10
204.2	203.9	1.0014	Intercept	-0.371604	+/-20
100.6	100.5	1.0008			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 29, 2025	Previous Calibration:	September 11, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	9:26	End Time (MST):	14:06
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999996	≥ 0.995
799.7	798.7	1.0012	Slope	0.998416	$0.90 - 1.10$
400.3	400.9	0.9985	Intercept	0.719909	± 20
199.7	200.7	0.9949			



NO_x Calibration Plot

Date: October 29, 2025

Location: Firebag





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS20 MACKAY RIVER OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: MacKay River Station number: AMS 20
Calibration Date: October 16, 2025 Last Cal Date: September 5, 2025
Start time (MST): 7:34 End time (MST): 10:06
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.15 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC409669
Removed Cal Gas Conc: 49.15 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 5706
Zero Air Gen Model: API 701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1501301450
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999303	0.995748	Backgd or Offset:	19.2	20.4
Calibration intercept:	0.183886	0.303078	Coeff or Slope:	0.923	0.932

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.8	----
As found High point	4919	81.4	800.1	794.3	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	793.5	Previous response	799.7	*% change	-0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4919	81.4	800.1	796.7	1.004
Mid point	4959	40.7	400.1	399.2	1.002
Low point	4980	20.3	199.5	199.2	1.002
As left zero	5000	0.0	0.0	0.2	----
As left span	4919	81.4	800.1	800.0	1.000
Average Correction Factor:					1.003

Notes: No Maintenance done. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 16, 2025	Previous Calibration:	September 5, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	7:34	End Time (MST):	10:06
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999998	≥0.995
800.1	796.7	1.0043	Slope	0.995748	0.90 - 1.10
400.1	399.2	1.0023	Intercept	0.303078	+/-30
199.5	199.2	1.0017			



SO2 Calibration Plot

Date: October 16, 2025

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: MacKay River Station number: AMS 20
Calibration Date: October 15, 2025 Last Cal Date: September 4, 2025
Start time (MST): 7:06 End time (MST): 11:00
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.12 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC515997
Removed Cal Gas Conc: 5.12 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 5706
ZAG Make/Model: API 701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1236656117
Converter make: Global Converter serial #: 2022-226
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002025	1.003598	Backgd or Offset:	3.77	3.77
Calibration intercept:	-0.040503	-0.020530	Coeff or Slope:	1.071	1.071

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4922	78.1	80.0	81.5	0.980
As found Mid point	4961	39.0	39.9	40.5	0.984
As found Low point	4980	19.5	20.0	20.2	0.984
New cylinder response					
Baseline Corr As found:	81.6	Prev response:	80.09	*% change:	1.8%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.020463	AF Intercept:	-0.160294
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999996	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4922	78.1	80.0	80.2	0.997
Mid point	4961	39.0	39.9	40.1	0.996
Low point	4980	19.5	20.0	20.1	0.994
As left zero	5000	0.0	0.0	0.1	----
As left span	4922	78.1	80.0	79.0	1.012
SO ₂ Scrubber Check	4982	81.3	802.8	0.0	----
Date of last scrubber change:	25-May-23		Ave Corr Factor		0.996
Date of last converter efficiency test:					

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 4, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	7:06	End Time (MST):	11:00
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999996	≥ 0.995
80.0	80.2	0.9972	Slope	1.003598	$0.90 - 1.10$
39.9	40.1	0.9959	Intercept	-0.020530	± 3
20.0	20.1	0.9935			


H₂S Calibration Curve



H₂S Calibration Plot

Date: October 15, 2025

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: MacKay River Station number: AMS 20
Calibration Date: October 16, 2025 Last Cal Date: September 5, 2025
Start time (MST): 7:34 End time (MST): 10:05
Reason: Routine

Calibration Standards

Gas Cert Reference: CC409669 Cal Gas Expiry Date: October 9, 2032
CH4 Cal Gas Conc. 505.1 ppm CH4 Equiv Conc. 1072.7 ppm
C3H8 Cal Gas Conc. 206.4 ppm
Removed Gas Cert: Removed Gas Expiry:
Removed CH4 Conc. 505.1 ppm CH4 Equiv Conc. 1072.7 ppm
Removed C3H8 Conc. 206.4 ppm Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 5706
ZAG Make/Model: API 701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1501663727
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001159	0.994146	Background:	2.930	2.990
Calibration intercept:	-0.101189	0.003372	Coefficient:	4.900	5.000

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.10	----
As found High point	4919	81.4	17.46	17.23	1.019
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.13	Previous response	17.38	*% change	-1.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.04	----
High point	4919	81.4	17.46	17.39	1.004
Mid point	4959	40.7	8.73	8.66	1.009
Low point	4980	20.3	4.35	4.30	1.013
As left zero	5000	0.0	0.00	-0.07	----
As left span	4919	81.4	17.46	17.43	1.002
Average Correction Factor					1.009

Notes: No Maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	October 16, 2025	Previous Calibration:	September 5, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	7:34	End Time (MST):	10:05
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.04	----	Correlation Coefficient	0.999974	≥ 0.995
17.46	17.39	1.0044	Slope	0.994146	$0.90 - 1.10$
8.73	8.66	1.0087	Intercept	0.003372	± 1.5
4.35	4.30	1.0135			


THC Calibration Curve



THC Calibration Plot

Date: October 16, 2025

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: MacKay River
Station number: AMS 20
Calibration Date: October 8, 2025
Last Cal Date: September 2, 2025
Start time (MST): 6:42
End time (MST): 9:53
Reason: As Found

Calibration Standards

NO Gas Cylinder #: DT0037393
NO_x Cal Gas Conc: 62.00 ppm
Removed Cylinder #:
Removed Gas NO_x Conc: 62.00 ppm
NO_x gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: July 22, 2032
NO Cal Gas Conc: 61.90 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 61.90 ppm
NO gas Diff:
Serial Number: 5706
Serial Number: 4888

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
AF High point	4935	64.6	801.1	799.8	1.3	801.3	800.4	0.9	0.9996	0.9993
AF Mid point	4968	32.3	400.5	399.9	0.6	402.8	400.1	2.7	0.9940	0.9994
AF Low point	4984	16.2	200.9	200.5	0.3	202.5	199.0	3.6	0.9915	1.0078

New cyl resp

Previous Response	NO _x = 802.7 ppb	NO = 803.7 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = -0.2%
Baseline Corr 1st pt	NO _x = 801.4 ppb	NO = 800.4 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -0.4%
Baseline Corr 2nd pt	NO _x = 402.9 ppb	NO = 400.1 ppb	As found	NO _x r ² : 0.999989	Nx SI: 0.999984	Nx Int: 1.013
Baseline Corr 3rd pt	NO _x = 202.6 ppb	NO = 199.0 ppb	As found	NO r ² : 0.999994	NO SI: 1.001454	NO Int: -0.686
			As found	NO ₂ r ² : 0.999948	NO ₂ SI: 0.997895	NO ₂ Int: -1.339

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	-0.1	----	----
As found high GPT point	798.3	417.5	382.1	380.7	1.0037	99.6%
As found mid GPT point	798.3	604.8	194.8	192.1	1.0140	98.6%
As found low GPT point	798.3	701.0	98.6	96.0	1.0270	97.4%



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1505164379

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.016	1.016
NOX coeff or slope:	1.000	1.000
NO2 coeff or slope:	1.000	1.000

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	2.8	2.8
NOX bkgnd or offset:	3.0	3.0
Reaction cell Press:	158.8	158.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000283	
NO _x Cal Offset:	1.332750	
NO Cal Slope:	1.005655	
NO Cal Offset:	-0.607201	
NO ₂ Cal Slope:	0.997997	
NO ₂ Cal Offset:	-1.292479	

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero
High point
Mid point
Low point
As left zero
As left span

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero
High GPT point
Mid GPT point
Low GPT point

Average Correction Factor


Notes: As founds for Dryer to be replaced.

Calibration Performed By: Melissa Lemay

NO_x Calibration Plot

Date: October 8, 2025

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: MacKay River
Station number: AMS 20
Calibration Date: October 9, 2025
Last Cal Date: October 8, 2025
Start time (MST): 6:46
End time (MST): 10:53
Reason: As Found

Calibration Standards

NO Gas Cylinder #: DT0037393
NOX Cal Gas Conc: 62.00 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 62.00 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: July 22, 2032
NO Cal Gas Conc: 61.90 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 61.90 ppm
NO gas Diff:
Serial Number: 5706
Serial Number: 4888

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation		*Percent Change		NO _x =	NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>		*Percent Change		NO =	NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:		Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:		NO Int:	
					As found	NO ₂ r ² :	NO2 SI:		NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1505164379

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.016	1.033	NO bkgnd or offset:	2.8	2.9
NOX coeff or slope:	1.000	0.988	NOX bkgnd or offset:	3.0	3.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	155.8	155.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000283	0.998314
NO _x Cal Offset:	1.332750	0.773119
NO Cal Slope:	1.005655	1.001682
NO Cal Offset:	-0.607201	-0.666357
NO ₂ Cal Slope:	0.997997	0.994991
NO ₂ Cal Offset:	-1.292479	-1.666155

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
High point	4935	64.6	801.1	799.8	1.3	799.8	800.5	-0.8	1.0016	0.9991
Mid point	4968	32.3	400.5	399.9	0.6	401.9	400.5	1.4	0.9965	0.9984
Low point	4984	16.2	200.9	200.5	0.3	201.6	198.9	2.7	0.9964	1.0083
As left zero	5000	0.0	0.0	0.0	0.0	0.3	0.4	0.0	----	----
As left span	4935	64.6	801.1	414.8	386.3	798.7	414.8	384.0	1.0030	1.0000
Average Correction Factor									0.9982	1.0019

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	798.4	414.7	385.0	382.2	1.0073	99.3%
Mid GPT point	798.4	603.9	195.8	192.4	1.0176	98.3%
Low GPT point	798.4	699.3	100.4	96.6	1.0393	96.2%
Average Correction Factor					1.0214	97.9%

Notes: Calibration after dryer was replaced. Span adjusted.

Calibration Performed By:

Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 9, 2025	Previous Calibration:	October 8, 2025
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	6:46	End Time (MST):	10:53
Analyzer make:	Thermo 42i	6:50:00 AM	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999991	≥0.995
801.1	799.8	1.0016	Slope	0.998314	0.90 - 1.10
400.5	401.9	0.9965	Intercept	0.773119	+/-20
200.9	201.6	0.9964			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 9, 2025	Previous Calibration:	October 8, 2025
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	6:46	End Time (MST):	10:53
Analyzer make:	Thermo 42i		1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999921	≥0.995
385.0	382.2	1.0073	Slope	0.994991	0.90 - 1.10
195.8	192.4	1.0176	Intercept	-1.666155	+/-20
100.4	96.6	1.0393			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 9, 2025	Previous Calibration:	October 8, 2025
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	6:46	End Time (MST):	10:53
Analyzer make:	Thermo 42i		1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.0	----	Correlation Coefficient	0.999993	≥ 0.995
799.8	800.5	0.9991	Slope	1.001682	$0.90 - 1.10$
399.9	400.5	0.9984	Intercept	-0.666357	± 20
200.5	198.9	1.0083			



NO_x Calibration Plot

Date: October 9, 2025

Location: MacKay River





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS21
CONKLIN
OCTOBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Conklin
Calibration Date: October 15, 2025
Start time (MST): 10:28
Reason: Routine

Station number: AMS 21
Last Cal Date: September 23, 2025
End time (MST): 13:32

Calibration Standards

Cal Gas Concentration: 50.34 ppm
Cal Gas Cylinder #: CC340840
Removed Cal Gas Conc: 50.34 ppm
Removed Gas Cyl #: NA
Calibrator Model: Teledyne API T700P
Zero Air Gen Model: Teledyne API T701H

Cal Gas Exp Date: October 9, 2032
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 2656
Serial Number: 355

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 1428701363

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001290	0.994307	Backgd or Offset:	29.3	29.9
Calibration intercept:	2.496595	2.718110	Coeff or Slope:	0.891	0.891

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4921	79.5	800.3	796.5	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	796.3	Previous response	803.9	*% change	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4921	79.5	800.3	797.6	1.003
Mid point	4960	39.8	400.7	401.4	0.998
Low point	4980	19.9	200.4	205.0	0.977
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	79.5	800.3	798.2	1.003
Average Correction Factor:					0.993

Notes: Sample inlet filter was changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 23, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:28	End Time (MST):	13:32
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999954	≥0.995
800.3	797.6	1.0034	Slope	0.994307	0.90 - 1.10
400.7	401.4	0.9983	Intercept	2.718110	+/-30
200.4	205.0	0.9774			



SO2 Calibration Plot

Date: October 15, 2025

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Conklin
Calibration Date: October 7, 2025
Start time (MST): 9:01
Reason: Routine

Station number: AMS 21
Last Cal Date: September 24, 2025
End time (MST): 15:07

Calibration Standards

Cal Gas Concentration: 5.14 ppm
Cal Gas Cylinder #: CC501204
Removed Cal Gas Conc: 5.14 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne T700P
ZAG Make/Model: Teledyne T701H

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 2656
Serial Number: 355

Analyzer Information

Analyzer make: Thermo 43i-QTL
Converter make: CD-Nova 101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12228021058
Converter serial #: 565
Converter Temp: 825 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997811	0.994549	Backgd or Offset:	2.8	2.8
Calibration intercept:	0.318383	0.158394	Coeff or Slope:	1.467	1.470

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4922	78.4	80.6	78.8	1.024
As found Mid point	4961	39.2	40.3	40.0	1.010
As found Low point	4980	19.6	20.2	20.4	0.993
New cylinder response					
Baseline Corr As found:	78.7	Prev response:	80.73	*% change:	-2.6%
Baseline Corr 2nd AF pt:	39.9	AF Slope:	0.974696	AF Intercept:	0.458415
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999901	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4922	78.4	80.6	80.3	1.004
Mid point	4961	39.2	40.3	40.2	1.002
Low point	4980	19.6	20.2	20.3	0.993
As left zero	5000	0.0	0.0	0.2	----
As left span	4922	78.4	80.6	80.6	1.000
SO2 Scrubber Check	4921	79.5	794.9	0.1	----
Date of last scrubber change:	August 6, 2025		Ave Corr Factor		1.000
Date of last converter efficiency test:	October 7, 2025		98.8% efficiency		

Notes: Sample inlet filter was changed after multipoint as founds. Converter efficiency test done. SO2 scrubber check done and passed. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	October 7, 2025	Previous Calibration:	September 24, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	9:01	End Time (MST):	15:07
Analyzer make:	Thermo 43i-QTL	Analyzer serial #:	12228021058

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999996		≥ 0.995
80.6	80.3	1.0036	Slope	0.994549		$0.90 - 1.10$
40.3	40.2	1.0024	Intercept	0.158394		± 3
20.2	20.3	0.9926				


TRS Calibration Curve



TRS Calibration Plot

Date: October 7, 2025

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Conklin
 Calibration Date: October 15, 2025
 Start time (MST): 10:28
 Reason: Routine

Station number: AMS 21
 Last Cal Date: September 23, 2025
 End time (MST): 13:32

Calibration Standards

Gas Cert Reference:	CC340840	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	503.8 ppm	CH ₄ Equiv Conc.	1067.6 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	503.8 ppm	CH ₄ Equiv Conc.	1067.6 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700P	Serial Number:	2656
Zero Air Gen model:	Teledyne API T701H	Serial Number:	355

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1180320039
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.28E-04	2.34E-04	NMHC SP Ratio:	4.84E-05
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	185126
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.5	16.97	16.55	1.026
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.55	Prev response	16.81	*% change	-1.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.5	16.97	16.96	1.000
Mid point	4960	39.8	8.50	8.51	0.998
Low point	4980	19.9	4.25	4.34	0.979
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	16.97	16.97	1.000
Average Correction Factor					0.993

Notes: Sample inlet filter and H₂/N₂ cylinder was changed after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.5	8.96	8.74	1.025
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.74	Prev response	8.89	*% change	-1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.5	8.96	8.94	1.002
Mid point	4960	39.8	4.49	4.50	0.996
Low point	4980	19.9	2.24	2.30	0.976
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	8.96	8.95	1.002
Average Correction Factor					0.991

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.5	8.01	7.81	1.026
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.81	Prev response	7.91	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.5	8.01	8.02	0.999
Mid point	4960	39.8	4.01	4.01	1.000
Low point	4980	19.9	2.01	2.04	0.983
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	8.01	8.02	0.999
Average Correction Factor					0.994

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.987939	0.997750
THC Cal Offset:	0.044532	0.040887
CH ₄ Cal Slope:	0.986167	0.999822
CH ₄ Cal Offset:	0.014206	0.011577
NMHC Cal Slope:	0.988528	0.995912
NMHC Cal Offset:	0.031728	0.029509

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 23, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:28	End Time (MST):	13:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999965	≥ 0.995
16.97	16.96	1.0005	Slope	0.997750	$0.90 - 1.10$
8.50	8.51	0.9984	Intercept	0.040887	± 0.5
4.25	4.34	0.9790			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 23, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:28	End Time (MST):	13:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999976	<i>≥0.995</i>
8.01	8.02	0.9987	Slope	0.999822	<i>0.90 - 1.10</i>
4.01	4.01	1.0004	Intercept	0.011577	<i>+/-0.5</i>
2.01	2.04	0.9829			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 23, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:28	End Time (MST):	13:32
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999947	<i>≥0.995</i>
8.96	8.94	1.0021	Slope	0.995912	<i>0.90 - 1.10</i>
4.49	4.50	0.9964	Intercept	0.029509	<i>+/-0.5</i>
2.24	2.30	0.9756			



NMHC Calibration Plot

Date: October 15, 2025

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Conklin
Station number: AMS 21
Calibration Date: October 8, 2025
Last Cal Date: September 16, 2025
Start time (MST): 8:30
End time (MST): 12:35
Reason: Routine

Calibration Standards

NO Gas Cylinder #: SA18828
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700P
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 2659
Serial Number: 953

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
AF High point	4918	82.0	802.0	800.3	1.6	807.2	803.7	3.4	0.9933	0.9954
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 802.7 ppb	NO = 800.8 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.6%	
Baseline Corr 1st pt	NO _x = 807.4 ppb	NO = 804.0 ppb				<u>As Found Statistics</u>		*Percent Change	NO = 0.4%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153356

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.648	0.648	NO bkgnd or offset:	3.7	3.7
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	3.6	3.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	165.6	163.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998174	1.006212
NO _x Cal Offset:	2.188064	2.868139
NO Cal Slope:	0.999434	1.007774
NO Cal Offset:	0.888035	1.568103
NO ₂ Cal Slope:	0.999907	0.999764
NO ₂ Cal Offset:	-0.013101	0.063265

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	-0.2	0.1	----	----
High point	4918	82.0	802.0	800.3	1.6	808.7	807.3	1.4	0.9917	0.9914
Mid point	4959	41.0	401.0	400.2	0.8	406.8	405.5	1.3	0.9857	0.9868
Low point	4980	20.5	200.5	200.1	0.4	208.1	205.1	3.0	0.9633	0.9754
As left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
As left span	4918	82.0	802.0	390.3	411.7	810.0	390.3	419.7	0.9901	1.0000
Average Correction Factor									0.9802	0.9845

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	808.5	392.0	418.1	418.2	0.9999	100.0%
Mid GPT point	808.5	601.9	208.2	208.0	1.0012	99.9%
Low GPT point	808.5	703.5	106.6	106.8	0.9985	100.2%
Average Correction Factor					0.9998	100.0%

Notes:

Sample inlet filter was changed after as founds. No adjustment made.

Calibration Performed By:

Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 8, 2025	Previous Calibration:	September 16, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	8:30	End Time (MST):	12:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999938	≥0.995
802.0	808.7	0.9917	Slope	1.006212	0.90 - 1.10
401.0	406.8	0.9857	Intercept	2.868139	+/-20
200.5	208.1	0.9633			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 8, 2025	Previous Calibration:	September 16, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	8:30	End Time (MST):	12:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999999	≥ 0.995
418.1	418.2	0.9999	Slope	0.999764	$0.90 - 1.10$
208.2	208.0	1.0012	Intercept	0.063265	± 20
106.6	106.8	0.9985			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 8, 2025	Previous Calibration:	September 16, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	8:30	End Time (MST):	12:35
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999978	≥ 0.995
800.3	807.3	0.9914	Slope	1.007774	$0.90 - 1.10$
400.2	405.5	0.9868	Intercept	1.568103	± 20
200.1	205.1	0.9754			



NO_x Calibration Plot

Date: October 8, 2025

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Conklin
Calibration Date: October 2, 2025
Start time (MST): 10:49
Reason: Routine

Station number: AMS 21
Last Cal Date: September 3, 2025
End time (MST): 13:24

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700P
ZAG Make/Model: Teledyne API T701H

Serial Number: 2656
Serial Number: 355

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1300156233

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998771	1.007229	Backgd or Offset:	0.0	-0.2
Calibration intercept:	0.940000	2.260000	Coeff or Slope:	1.229	1.229

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-0.5	----
As found High point	5000	1176.6	400.0	402.0	0.994
As found Mid point					
As found Low point					
Baseline Corr As found:	402.5	Previous response	400.4	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.1	----
High point	5000	1175.9	400.0	404.0	0.990
Mid point	5000	939.1	200.0	205.0	0.976
Low point	5000	809.5	100.0	105.0	0.952
As left zero	5000	800.0	0.0	0.5	----
As left span	5000	1173.8	400.0	409.0	0.978
Average Correction Factor					0.973

Notes: Sample inlet filter was changed after as founds. Adjusted zero only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	October 2, 2025	Previous Calibration:	September 3, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:49	End Time (MST):	13:24
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156233

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999868	≥ 0.995
400.0	404.0	0.9901	Slope	1.007229	$0.90 - 1.10$
200.0	205.0	0.9756	Intercept	2.260000	± 5
100.0	105.0	0.9524			



O₃ Calibration Plot

Date: October 2, 2025

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Conklin
Calibration Date: October 8, 2025
Start time (MST): 10:49
Station number: AMS 21
Last Cal Date: September 23, 2025
End time (MST): 11:02
Analyzer Make: API T640
Particulate Fraction: PM2.5
S/N: 1266
Flow Meter Make/Model: Alicat FP-25BT
Temp/RH standard: Alicat FP-25BT
S/N: 388754
S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	10.90	10.00	10.90	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	705.80	707.70	705.80	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.12	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	39	----	39	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 5.20		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: July 16, 2026
Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: August 6, 2025
Date Disposable Filter Changed: August 6, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: June 4, 2025
Date RH/T Sensor Cleaned: June 30, 2025

Notes: Verified flow, pressure, temperature and pump power. Leak check passed. No adjustment needed.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER
OCTOBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: October 24, 2025 Last Cal Date: September 22, 2025
Start time (MST): 11:32 End time (MST): 14:55
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.11 ppm Cal Gas Exp Date: January 18, 2029
Cal Gas Cylinder #: CC281519
Removed Cal Gas Conc: 50.11 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3806
Zero Air Gen Model: Teledyne API T701 Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1152430006
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999707	1.001164	Backgd or Offset:	26.4	26.9
Calibration intercept:	0.624279	0.064348	Coeff or Slope:	1.000	1.010

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4920	79.8	799.8	792.5	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	792.3	Previous response	800.2	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4920	79.8	799.8	800.8	0.999
Mid point	4960	39.9	399.9	400.6	0.998
Low point	4980	20.0	200.4	200.3	1.001
As left zero	5000	0.0	0.0	0.5	----
As left span	4920	79.8	799.8	800.4	0.999
Average Correction Factor:					0.999

Notes: Changed the inlet filter after as founds. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 24, 2025	Previous Calibration:	September 22, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:32	End Time (MST):	14:55
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995
799.8	800.8	0.9987	Slope	1.001164	0.90 - 1.10
399.9	400.6	0.9982	Intercept	0.064348	+/-30
200.4	200.3	1.0007			



SO2 Calibration Plot

Date: October 24, 2025

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: October 28, 2025 Last Cal Date: September 18, 2025
Start time (MST): 11:39 End time (MST): 16:28
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.02 ppm Cal Gas Exp Date: November 15, 2026
Cal Gas Cylinder #: CC424047
Removed Cal Gas Conc: 5.02 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3806
ZAG Make/Model: Teledyne API T701 Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1151680031
Converter make: CDN-101 Converter serial #: 620
Analyzer Range: 0 - 100 ppb Converter Temp: 850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003379	0.961388	Backgd or Offset:	3.43	3.39
Calibration intercept:	0.140666	0.459964	Coeff or Slope:	1.180	1.180

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4920	79.7	80.0	79.7	1.007
As found Mid point	4960	39.8	40.0	40.0	1.004
As found Low point	4980	19.9	20.0	20.1	1.004
New cylinder response					
Baseline Corr As found:	79.5	Prev response:	80.43	*% change:	-1.2%
Baseline Corr 2nd AF pt:	39.8	AF Slope:	0.993382	AF Intercept:	0.240460
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999998	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4920	79.7	80.0	77.2	1.037
Mid point	4960	39.8	40.0	39.3	1.017
Low point	4980	19.9	20.0	19.6	1.019
As left zero	5000	0.0	0.0	0.7	----
As left span	4920	79.7	80.0	76.0	1.053
SO2 Scrubber Check	4920	79.8	798.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.024
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Scrubber test performed, no issues. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	October 28, 2025	Previous Calibration:	September 18, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:39	End Time (MST):	16:28
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999924		≥ 0.995
80.0	77.2	1.0366	Slope	0.961388		$0.90 - 1.10$
40.0	39.3	1.0168	Intercept	0.459964		± 3
20.0	19.6	1.0194				


TRS Calibration Curve



TRS Calibration Plot

Date: October 28, 2025

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Janvier
 Calibration Date: October 24, 2025
 Start time (MST): 11:32
 Reason: Routine

Station number: AMS 22
 Last Cal Date: September 22, 2025
 End time (MST): 14:55

Calibration Standards

Gas Cert Reference: CC281519
 CH₄ Cal Gas Conc. 502.8 ppm
 C₃H₈ Cal Gas Conc. 208.4 ppm
 Removed Gas Cert: NA
 Removed CH₄ Conc. 502.8 ppm
 Removed C₃H₈ Conc. 208.4 ppm
 Diff between cyl (CH₄):
 Calibrator Model: Teledyne API 700
 Zero Air Gen model: Teledyne API 701

Cal Gas Expiry Date: January 18, 2029
 CH₄ Equiv Conc. 1075.9 ppm
 Removed Gas Expiry: NA
 CH₄ Equiv Conc. 1075.9 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 3806
 Serial Number: 691

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1317958219
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.54E-04	2.58E-04	NMHC SP Ratio:	6.13E-05
CH ₄ Retention time:	11.6	11.6	NMHC Peak Area:	149196
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				147887
				OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	17.17	17.04	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.04	Prev response	17.17	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	17.17	17.13	1.002
Mid point	4960	39.9	8.59	8.49	1.011
Low point	4980	20.0	4.30	4.25	1.012
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	17.17	17.10	1.004
Average Correction Factor					1.009

Notes:

Changed the inlet filter after as founds. Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	9.15	9.11	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.11	Prev response	9.23	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.15	9.10	1.005
Mid point	4960	39.9	4.57	4.54	1.007
Low point	4980	20.0	2.29	2.28	1.004
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.15	9.10	1.006
Average Correction Factor					1.005

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	8.03	7.93	1.013
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.93	Prev response	7.94	*% change	-0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.03	8.02	1.000
Mid point	4960	39.9	4.01	3.95	1.016
Low point	4980	20.0	2.01	1.97	1.022
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.03	8.00	1.003
Average Correction Factor					1.013

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.002326	0.998025
THC Cal Offset:	-0.044604	-0.031787
CH ₄ Cal Slope:	0.992807	1.000840
CH ₄ Cal Offset:	-0.032144	-0.029357
NMHC Cal Slope:	1.010515	0.995131
NMHC Cal Offset:	-0.012060	-0.001228

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 24, 2025	Previous Calibration:	September 22, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:32	End Time (MST):	14:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999976	<i>≥0.995</i>
17.17	17.13	1.0023	Slope	0.998025	<i>0.90 - 1.10</i>
8.59	8.49	1.0111	Intercept	-0.031787	<i>+/-0.5</i>
4.30	4.25	1.0124			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 24, 2025	Previous Calibration:	September 22, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:32	End Time (MST):	14:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			Limits
0.00	0.00	----	Correlation Coefficient	0.999920		≥0.995
8.03	8.02	1.0001	Slope	1.000840		0.90 - 1.10
4.01	3.95	1.0155	Intercept	-0.029357		+/-0.5
2.01	1.97	1.0220				





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 24, 2025	Previous Calibration:	September 22, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:32	End Time (MST):	14:55
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			Limits
0.00	0.00	----	Correlation Coefficient	0.999999		≥ 0.995
9.15	9.10	1.0047	Slope	0.995131		0.90 - 1.10
4.57	4.54	1.0067	Intercept	-0.001228		± 0.5
2.29	2.28	1.0041				



NMHC Calibration Plot

Date: October 24, 2025

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Janvier
Station number: AMS 22
Calibration Date: October 23, 2025
Last Cal Date: September 19, 2025
Start time (MST): 11:31
End time (MST): 15:53
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0047765
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: March 11, 2031
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3806
Serial Number: 691

As Found Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
AF High point	4918	82.0	802.0	800.3	1.6	802.5	796.5	6.0	0.9991	1.0045
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 803.7 ppb	NO = 800.2 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -0.1%	
Baseline Corr 1st pt	NO _x = 802.7 ppb	NO = 796.7 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -0.4%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1229254994

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.012	1.012	NO bkgnd or offset:	2.7	2.8
NOX coeff or slope:	0.997	0.997	NOX bkgnd or offset:	2.9	2.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	172.6	172.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001535	1.000138
NO _x Cal Offset:	0.504115	-0.255953
NO Cal Slope:	1.001288	0.999346
NO Cal Offset:	-1.116032	-1.536101
NO ₂ Cal Slope:	1.004012	1.002701
NO ₂ Cal Offset:	1.010979	0.717283

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
High point	4918	82.0	802.0	800.3	1.6	801.9	799.0	2.9	1.0001	1.0017
Mid point	4960	41.0	400.9	400.1	0.8	400.6	397.5	3.1	1.0007	1.0065
Low point	4980	20.5	200.5	200.1	0.4	200.1	197.0	3.1	1.0018	1.0155
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
As left span	4918	82.0	802.0	395.8	406.2	801.2	395.8	405.3	1.0009	1.0000
Average Correction Factor									1.0009	1.0079

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	796.2	395.3	402.5	404.0	0.9964	100.4%
Mid GPT point	796.2	595.7	202.1	203.7	0.9923	100.8%
Low GPT point	796.2	695.0	102.8	104.6	0.9832	101.7%
Average Correction Factor					0.9906	100.9%

Notes:

Inlet filter was changed after as founds. No adjustments made.

Calibration Performed By:

Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 19, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:31	End Time (MST):	15:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1229254994

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	1.000000	≥0.995
802.0	801.9	1.0001	Slope	1.000138	0.90 - 1.10
400.9	400.6	1.0007	Intercept	-0.255953	+/-20
200.5	200.1	1.0018			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 19, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:31	End Time (MST):	15:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1229254994

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999985	≥0.995
402.5	404.0	0.9964	Slope	1.002701	0.90 - 1.10
202.1	203.7	0.9923	Intercept	0.717283	+/-20
102.8	104.6	0.9832			





Wood Buffalo Environmental Association

NO Calibration Summary


Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 19, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:31	End Time (MST):	15:53
Analyzer make:	Thermo 42i	Analyzer serial #:	1229254994

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999985	≥ 0.995
800.3	799.0	1.0017	Slope	0.999346	$0.90 - 1.10$
400.1	397.5	1.0065	Intercept	-1.536101	± 20
200.1	197.0	1.0155			


NO Calibration Curve



NO_x Calibration Plot

Date: October 23, 2025

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Janvier
Calibration Date: October 17, 2025
Start time (MST): 11:45
Reason: Routine

Station number: AMS 22
Last Cal Date: September 9, 2025
End time (MST): 14:31

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Serial Number: 3806
Serial Number: 691

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1227254861

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000400	0.998829	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	0.980000	0.980000	Coeff or Slope:	1.553	1.553

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-0.1	----
As found High point	5000	926.2	400.0	399.1	1.002
As found Mid point					
As found Low point					
Baseline Corr As found:	399.2	Previous response	401.1	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.1	----
High point	5000	926.2	400.0	400.0	1.000
Mid point	5000	768.9	200.0	201.4	0.993
Low point	5000	666.4	100.0	101.6	0.984
As left zero	5000	800.0	0.0	0.2	----
As left span	5000	926.2	400.0	404.4	0.989
Average Correction Factor					0.992

Notes: Changed the inlet filter after the as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 9, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	11:45	End Time (MST):	14:31
Analyzer make:	Thermo 49i	Analyzer serial #:	1227254861

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999977	≥0.995
400.0	400.0	1.0000	Slope	0.998829	0.90 - 1.10
200.0	201.4	0.9930	Intercept	0.980000	+/- 5
100.0	101.6	0.9843			



O₃ Calibration Plot

Date: October 17, 2025

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: October 28, 2025 Last Cal Date: September 19, 2025
Start time (MST): 12:18 End time (MST): 13:44

Analyzer Make: Teledyne API T640 S/N: 325
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	1.6	1.02	1.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	721.4	722.64	721.4	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	4.99	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	38	----	38	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 4.1		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 6, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input checked="" type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: August 26, 2025
Date Disposable Filter Changed: August 26, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: May 28, 2025
Date RH/T Sensor Cleaned: August 26, 2025

Notes:

Verified flow, temperature, and pressure. Leak check passed.

Calibration by: Rene Chamberland



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23 FORT HILLS OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Fort Hills Station number: AMS 23
Calibration Date: October 10, 2025 Last Cal Date: September 11, 2025
Start time (MST): 6:40 End time (MST): 9:36
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.35 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC484463
Removed Cal Gas Conc: 50.35 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 1222
Zero Air Gen Model: API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1160290012
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998675	0.995531	Backgd or Offset:	18.5	18.5
Calibration intercept:	0.040494	0.239920	Coeff or Slope:	1.063	1.063

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4921	79.4	799.5	795.3	1.006
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	795.0	Previous response	798.5	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4921	79.4	799.5	796.4	1.004
Mid point	4960	39.7	399.8	397.7	1.005
Low point	4980	19.8	199.4	198.9	1.002
As left zero	5000	0.0	0.0	0.4	----
As left span	4921	79.4	799.5	800.5	0.999
Average Correction Factor:					1.004

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 10, 2025	Previous Calibration:	September 11, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:40	End Time (MST):	9:36
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999999	≥0.995
799.5	796.4	1.0039	Slope	0.995531	0.90 - 1.10
399.8	397.7	1.0053	Intercept	0.239920	+/-30
199.4	198.9	1.0025			



SO2 Calibration Plot

Date: October 10, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort Hills
Calibration Date: October 2, 2025
Start time (MST): 6:32
Reason: Routine

Station number: AMS 23
Last Cal Date: September 12, 2025
End time (MST): 10:42

Calibration Standards

Cal Gas Concentration: 4.84 ppm
Cal Gas Cylinder #: DT0021910
Removed Cal Gas Conc: 4.84 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 1222
Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1300156232
Converter serial #: 594
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.012537	1.011534	Backgd or Offset:	1.94	1.96
Calibration intercept:	-0.058229	-0.298114	Coeff or Slope:	0.985	0.990

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4917	82.6	80.0	79.2	1.008
As found Mid point	4959	41.3	40.0	39.5	1.009
As found Low point	4979	20.7	20.0	19.5	1.022
New cylinder response					
Baseline Corr As found:	79.3	Prev response:	80.91	*% change:	-2.0%
Baseline Corr 2nd AF pt:	39.6	AF Slope:	0.992809	AF Intercept:	-0.217865
Baseline Corr 3rd AF pt:	19.6	AF Correlation:	0.999986	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4917	82.6	80.0	80.7	0.991
Mid point	4959	41.3	40.0	40.1	0.997
Low point	4979	20.7	20.0	19.6	1.022
As left zero	5000	0.0	0.0	0.0	----
As left span	4917	82.6	80.0	82.9	0.965
SO2 Scrubber Check	4920	80.3	803.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.003
Date of last converter efficiency test:		March 13, 2024		110.3% efficiency	

Notes: SOx scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	October 2, 2025	Previous Calibration:	September 12, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:32	End Time (MST):	10:42
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999933	≥ 0.995
80.0	80.7	0.9909	Slope	1.011534	$0.90 - 1.10$
40.0	40.1	0.9969	Intercept	-0.298114	± 3
20.0	19.6	1.0224			

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Fort Hills
 Calibration Date: October 10, 2025
 Start time (MST): 6:40
 Reason: Routine

Station number: AMS 23
 Last Cal Date: September 11, 2025
 End time (MST): 9:36

Calibration Standards

Gas Cert Reference:	CC484463	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	504.3 ppm	CH ₄ Equiv Conc.	1065.6 ppm
C ₃ H ₈ Cal Gas Conc.	204.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.3 ppm	CH ₄ Equiv Conc.	1065.6 ppm
Removed C ₃ H ₈ Conc.	204.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	1222
Zero Air Gen model:	API T701	Serial Number:	1117

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 12227620777
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.66E-04	3.70E-04	NMHC SP Ratio:	5.66E-05
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	157300
Zero Chromatogram:	OFF	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.04	----
As found High point	4921	79.4	16.92	17.03	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.99	Prev response	16.79	*% change	1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.4	16.92	16.89	1.002
Mid point	4960	39.7	8.46	8.34	1.014
Low point	4980	19.8	4.22	4.18	1.011
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	16.92	16.91	1.000
Average Correction Factor					1.009

Notes: Zero Chromatogram turned on. Zero and Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.4	8.91	9.00	0.990
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.00	Prev response	8.85	*% change	1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.4	8.91	8.89	1.003
Mid point	4960	39.7	4.46	4.40	1.013
Low point	4980	19.8	2.22	2.21	1.004
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	8.91	8.90	1.001
Average Correction Factor					1.007

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.04	Limit = 0.90-1.10
As found High point	4921	79.4	8.01	8.03	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.99	Prev response	7.94	*% change	0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.4	8.01	8.00	1.001
Mid point	4960	39.7	4.00	3.94	1.016
Low point	4980	19.8	2.00	1.96	1.018
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	8.01	8.01	1.000
Average Correction Factor					1.012

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.993891	0.998099
THC Cal Offset:	-0.029794	-0.035182
CH ₄ Cal Slope:	0.994366	0.999561
CH ₄ Cal Offset:	-0.023801	-0.024995
NMHC Cal Slope:	0.993580	0.996953
NMHC Cal Offset:	-0.006192	-0.010587

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 10, 2025	Previous Calibration:	September 11, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:40	End Time (MST):	9:36
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999955		≥ 0.995
16.92	16.89	1.0020	Slope	0.998099		0.90 - 1.10
8.46	8.34	1.0143	Intercept	-0.035182		+/-0.5
4.22	4.18	1.0107				





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 10, 2025	Previous Calibration:	September 11, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:40	End Time (MST):	9:36
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999934	<i>≥0.995</i>
8.01	8.00	1.0011	Slope	0.999561	<i>0.90 - 1.10</i>
4.00	3.94	1.0156	Intercept	-0.024995	<i>+/-0.5</i>
2.00	1.96	1.0184			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 10, 2025	Previous Calibration:	September 11, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:40	End Time (MST):	9:36
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999967	≥ 0.995
8.91	8.89	1.0026	Slope	0.996953	0.90 - 1.10
4.46	4.40	1.0129	Intercept	-0.010587	± 0.5
2.22	2.21	1.0044			



NMHC Calibration Plot

Date: October 10, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Fort Hills
 Calibration Date: October 26, 2025
 Start time (MST): 7:40
 Reason: Maintenance

Station number: AMS 23
 Last Cal Date: October 10, 2025
 End time (MST): 9:59

Calibration Standards

Gas Cert Reference:	CC484463	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	504.3 ppm	CH ₄ Equiv Conc.	1065.6 ppm
C ₃ H ₈ Cal Gas Conc.	204.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.3 ppm	CH ₄ Equiv Conc.	1065.6 ppm
Removed C ₃ H ₈ Conc.	204.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	1222
Zero Air Gen model:	API T701	Serial Number:	1117

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 12227620777
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.70E-04	3.73E-04	NMHC SP Ratio:	5.69E-05
CH ₄ Retention time:	15.2	15.6	NMHC Peak Area:	157134
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				5.65E-05
				158335
				OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.06	----
As found High point	4921	79.4	16.92	16.64	1.020
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.58	Prev response	16.79	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.4	16.92	16.90	1.001
Mid point	4960	39.7	8.46	8.35	1.014
Low point	4980	19.8	4.22	4.16	1.014
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	16.92	16.94	0.999
Average Correction Factor					1.010

Notes: Baseline is dipping Chromatograms slightly moved. Zero and Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.4	8.91	9.05	0.984
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.05	Prev response	8.85	*% change	2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.4	8.91	8.88	1.004
Mid point	4960	39.7	4.46	4.40	1.013
Low point	4980	19.8	2.22	2.21	1.007
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	8.91	8.88	1.004
Average Correction Factor					1.008

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.06	Limit = 0.90-1.10
As found High point	4921	79.4	8.01	7.59	1.063
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.53	Prev response	7.94	*% change	-5.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.4	8.01	8.02	0.999
Mid point	4960	39.7	4.00	3.95	1.015
Low point	4980	19.8	2.00	1.96	1.021
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	8.01	8.06	0.994
Average Correction Factor					1.011

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.993891	0.998924
THC Cal Offset:	-0.029794	-0.040784
CH ₄ Cal Slope:	0.994366	1.002272
CH ₄ Cal Offset:	-0.023801	-0.029992
NMHC Cal Slope:	0.993580	0.995569
NMHC Cal Offset:	-0.006192	-0.010192

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 26, 2025	Previous Calibration:	October 10, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:40	End Time (MST):	9:59
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999954	≥ 0.995
16.92	16.90	1.0014	Slope	0.998924	$0.90 - 1.10$
8.46	8.35	1.0137	Intercept	-0.040784	± 0.5
4.22	4.16	1.0137			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 26, 2025	Previous Calibration:	October 10, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:40	End Time (MST):	9:59
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999916	<i>≥0.995</i>
8.01	8.02	0.9987	Slope	1.002272	<i>0.90 - 1.10</i>
4.00	3.95	1.0145	Intercept	-0.029992	<i>+/-0.5</i>
2.00	1.96	1.0210			






Wood Buffalo Environmental Association NMHC Calibration Summary

Station Information

Calibration Date:	October 26, 2025	Previous Calibration:	October 10, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:40	End Time (MST):	9:59
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999977	<i>≥0.995</i>
8.91	8.88	1.0042	Slope	0.995569	<i>0.90 - 1.10</i>
4.46	4.40	1.0129	Intercept	-0.010192	<i>+/-0.5</i>
2.22	2.21	1.0071			



NMHC Calibration Plot

Date: October 26, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Fort Hills Station number: AMS 23
 Calibration Date: October 29, 2025 Last Cal Date: October 26, 2025
 Start time (MST): 7:28 End time (MST): 10:05
 Reason: Maintenance Increase Carrier Pressure

Calibration Standards

Gas Cert Reference: CC484463 Cal Gas Expiry Date: October 9, 2032
 CH₄ Cal Gas Conc. 504.3 ppm CH₄ Equiv Conc. 1065.6 ppm
 C₃H₈ Cal Gas Conc. 204.1 ppm
 Removed Gas Cert: Removed Gas Expiry:
 Removed CH₄ Conc. 504.3 ppm CH₄ Equiv Conc. 1065.6 ppm
 Removed C₃H₈ Conc. 204.1 ppm Diff between cyl (THC):
 Diff between cyl (CH₄): Diff between cyl (NM):
 Calibrator Model: API T700 Serial Number: 1222
 Zero Air Gen model: API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 12227620777
 THC Range: 0 - 20 ppm NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.73E-04	3.54E-04	NMHC SP Ratio:	5.65E-05
CH ₄ Retention time:	15.6	15.0	NMHC Peak Area:	158335
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.4	16.92	17.01	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.01	Prev response	16.86	*% change	0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.4	16.92	16.83	1.005
Mid point	4960	39.7	8.46	8.26	1.025
Low point	4980	19.8	4.22	4.12	1.024
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	16.92	16.85	1.004
Average Correction Factor					1.018

Notes: Unstable zero. Carrier pressure increased from 25.5-26.5PSI. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.4	8.91	8.92	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.92	Prev response	8.86	*% change	0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.4	8.91	8.84	1.008
Mid point	4960	39.7	4.46	4.37	1.021
Low point	4980	19.8	2.22	2.20	1.012
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	8.91	8.86	1.006
Average Correction Factor					1.013

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.4	8.01	8.09	0.990
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.09	Prev response	8.00	*% change	1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.4	8.01	7.99	1.003
Mid point	4960	39.7	4.00	3.89	1.029
Low point	4980	19.8	2.00	1.92	1.038
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	8.01	7.98	1.003
Average Correction Factor					1.023

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.998924	0.995176
THC Cal Offset:	-0.040784	-0.062798
CH ₄ Cal Slope:	1.002272	0.998720
CH ₄ Cal Offset:	-0.029992	-0.047801
NMHC Cal Slope:	0.995569	0.991991
NMHC Cal Offset:	-0.010192	-0.014998

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 29, 2025	Previous Calibration:	October 26, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:28	End Time (MST):	10:05
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999888	≥ 0.995
16.92	16.83	1.0053	Slope	0.995176	$0.90 - 1.10$
8.46	8.26	1.0247	Intercept	-0.062798	± 0.5
4.22	4.12	1.0242			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 29, 2025	Previous Calibration:	October 26, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:28	End Time (MST):	10:05
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999774	<i>≥0.995</i>
8.01	7.99	1.0027	Slope	0.998720	<i>0.90 - 1.10</i>
4.00	3.89	1.0294	Intercept	-0.047801	<i>+/-0.5</i>
2.00	1.92	1.0380			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 29, 2025	Previous Calibration:	October 26, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:28	End Time (MST):	10:05
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999951	<i>≥0.995</i>
8.91	8.84	1.0077	Slope	0.991991	<i>0.90 - 1.10</i>
4.46	4.37	1.0206	Intercept	-0.014998	<i>+/-0.5</i>
2.22	2.20	1.0122			



NMHC Calibration Plot

Date: October 29, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Hills
Station number: AMS 23
Calibration Date: October 20, 2025
Last Cal Date: September 9, 2025
Start time (MST): 6:33
End time (MST): 11:07
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC358149
NOX Cal Gas Conc: 60.30 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 60.30 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.10 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 60.10 ppm
NO gas Diff:
Serial Number: 1222
Serial Number: 1117

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.3	0.0	-0.3	----	----
AF High point	4934	66.3	799.5	796.9	2.7	806.0	802.0	4.0	0.9916	0.9936
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 800.3 ppb	NO = 797.4 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.7%	
Baseline Corr 1st pt	NO _x = 806.3 ppb	NO = 802.0 ppb				<u>As Found Statistics</u>		*Percent Change	NO = 0.6%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430007

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.975	0.975	NO bkgnd or offset:	2.8	2.8
NOX coeff or slope:	0.992	0.992	NOX bkgnd or offset:	3.2	3.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	147.2	147.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001155	1.005887
NO _x Cal Offset:	-0.133957	-0.015171
NO Cal Slope:	1.002620	1.008514
NO Cal Offset:	-1.532494	-1.513884
NO ₂ Cal Slope:	0.998169	0.997104
NO ₂ Cal Offset:	-0.341510	-0.955955

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.2	----	----
High point	4934	66.3	799.5	796.9	2.7	804.5	803.3	1.4	0.9938	0.9920
Mid point	4967	33.2	400.4	399.0	1.3	401.8	399.1	2.7	0.9965	0.9999
Low point	4983	16.6	200.2	199.5	0.7	202.1	198.8	3.3	0.9907	1.0038
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
As left span	4934	66.3	799.5	407.4	392.1	805.1	407.4	397.7	0.9931	1.0000
Average Correction Factor									0.9936	0.9985

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	799.9	403.0	399.6	398.1	1.0036	99.6%
Mid GPT point	799.9	602.0	200.6	197.8	1.0139	98.6%
Low GPT point	799.9	701.1	101.5	100.0	1.0145	98.6%
Average Correction Factor					1.0107	98.9%

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 9, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:33	End Time (MST):	11:07
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999996	≥0.995
799.5	804.5	0.9938	Slope	1.005887	0.90 - 1.10
400.4	401.8	0.9965	Intercept	-0.015171	+/-20
200.2	202.1	0.9907			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 9, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:33	End Time (MST):	11:07
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999971	≥0.995
399.6	398.1	1.0036	Slope	0.997104	0.90 - 1.10
200.6	197.8	1.0139	Intercept	-0.955955	+/-20
101.5	100.0	1.0145			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 20, 2025	Previous Calibration:	September 9, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	6:33	End Time (MST):	11:07
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999977	≥ 0.995
796.9	803.3	0.9920	Slope	1.008514	0.90 - 1.10
399.0	399.1	0.9999	Intercept	-1.513884	+/-20
199.5	198.8	1.0038			



NO_x Calibration Plot

Date: October 20, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Hills Station number: AMS 23
Calibration Date: October 20, 2025 Last Cal Date: September 11, 2025
Start time (MST): 7:35 End time (MST): 8:35

Analyzer Make: API T640 S/N: 320
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388744
Temp/RH standard: Alicat FP-25BT S/N: 388744

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	2.0	1.3	2.0	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	734.6	733.7	734.6	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	4.76	5.08	<input checked="" type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	59	----	62	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	2.8	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 16-Jul-26
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.1	11.4	11.4	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ October 20, 2025
Date Disposable Filter Changed: _____ October 20, 2025

Post- maintenance Zero Verification: PM w/ HEPA: _____ 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ October 20, 2025
Date RH/T Sensor Cleaned: _____ October 20, 2025

Notes: Flow was adjusted. Leak Check, Flow and PMT checked before and after cleaning.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25 WASKŌW OHCI PIMÂTISIWIN OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS 25
Calibration Date: October 24, 2025 Last Cal Date: September 23, 2025
Start time (MST): 6:15 End time (MST): 9:12
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.70 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC342445
Removed Cal Gas Conc: 49.70 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 621
Zero Air Gen Model: API T701 Serial Number: 4765

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1118148497
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001884	0.997669	Backgd or Offset:	10.5	11.1
Calibration intercept:	0.027877	0.327490	Coeff or Slope:	1.038	1.046

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4920	80.5	800.1	792.2	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	791.9	Previous response	801.6	*% change	-1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.5	800.1	798.2	1.002
Mid point	4960	40.2	399.6	399.7	1.000
Low point	4980	20.1	199.8	199.6	1.001
As left zero	5000	0.0	0.0	0.3	----
As left span	4920	80.5	800.1	801.3	0.998
Average Correction Factor:					1.001

Notes: No maintenance done. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 24, 2025	Previous Calibration:	September 23, 2025
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS 25
Start Time (MST):	6:15	End Time (MST):	9:12
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999998	≥0.995
800.1	798.2	1.0024	Slope	0.997669	0.90 - 1.10
399.6	399.7	0.9997	Intercept	0.327490	+/-30
199.8	199.6	1.0010			



SO2 Calibration Plot

Date: October 24, 2025

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS 25
Calibration Date: October 21, 2025 Last Cal Date: September 10, 2025
Start time (MST): 8:50 End time (MST): 11:23
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC517099
Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 747
ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
Converter make: Global G-150 Converter serial #: 2024-287
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>	
Calibration slope:	0.984518	0.987943	Backgd or Offset:	3.32	3.41
Calibration intercept:	0.302144	0.582216	Coeff or Slope:	1.102	1.129

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.5	----
High point	4919	80.5	80.0	79.6	1.005
Mid point	4960	40.3	40.1	40.2	0.996
Low point	4980	20.1	20.0	20.4	0.979
As left zero	5000	0.0	0.0	0.7	----
As left span	4920	80.0	800.0	807.6	0.991
SO2 Scrubber Check	4920	80.0	800.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.994
Date of last converter efficiency test:		February 12, 2025		111.0%	efficiency

Notes: No as founds as external converter would not turn on after power spike. Converter changed out..
SOx Scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	October 21, 2025	Previous Calibration:	September 10, 2025
Station Name:	Waskow ohci Pimatisiwin	Station Number:	10:47:00 AM
Start Time (MST):	8:50	End Time (MST):	11:23
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999995		≥ 0.995
80.0	79.6	1.0053	Slope	0.987943		$0.90 - 1.10$
40.1	40.2	0.9964	Intercept	0.582216		± 3
20.0	20.4	0.9794				



H₂S Calibration Plot

Date: October 21, 2025

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS 25
Calibration Date: October 27, 2025 Last Cal Date: October 21, 2025
Start time (MST): 6:46 End time (MST): 10:40
Reason: Maintenance Calibration to lower zero

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC517099
Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 747
ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
Converter make: Global G-150 Converter serial #: 2024-287
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.987943	1.000081	Backgd or Offset:	3.41
Calibration intercept:	0.582216	0.182210	Coeff or Slope:	1.129

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4919	80.5	80.0	81.0	0.990
As found Mid point	4960	40.3	40.1	40.5	0.994
As found Low point	4980	20.1	20.0	20.4	0.989
New cylinder response					
Baseline Corr As found:	80.8	Prev response:	79.64	*% change:	1.4%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.009360	AF Intercept:	0.182325
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4919	80.5	80.0	80.1	0.999
Mid point	4960	40.3	40.1	40.4	0.991
Low point	4980	20.1	20.0	20.3	0.984
As left zero	5000	0.0	0.0	0.3	----
As left span	4920	80.0	800.0	812.0	0.985
SO2 Scrubber Check	4920	80.0	800.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	0.992
Date of last converter efficiency test:		February 12, 2025		111.0% efficiency	

Notes: Converter replaced last calibration. Zero is high. Calibration to lower zero. Zero adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	October 27, 2025	Previous Calibration:	October 21, 2025
Station Name:	Waskow ohci Pimatisiwin	Station Number:	10:47:00 AM
Start Time (MST):	6:46	End Time (MST):	10:40
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999974		≥ 0.995
80.0	80.1	0.9991	Slope	1.000081		$0.90 - 1.10$
40.1	40.4	0.9915	Intercept	0.182210		± 3
20.0	20.3	0.9842				



H₂S Calibration Plot

Date: October 27, 2025

Location: Waskow ohci Pimatisiwin





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27
JACKFISH 2/3
OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Jackfish 2/3 Station number: AMS 27
Calibration Date: October 3, 2025 Last Cal Date: September 9, 2025
Start time (MST): 11:24 End time (MST): 14:49
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.58 ppm Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #: SG9133974BAL
Removed Cal Gas Conc: 50.58 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 5252
Zero Air Gen Model: API 701 Serial Number: 268

Analyzer Information

Analyzer make: Thermo 43iQ-TL Serial Number: 12124313138
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005262	1.001333	Backgd or Offset:	9.3
Calibration intercept:	0.502931	0.202353	Coeff or Slope:	0.907

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4913	78.9	799.4	800.4	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	800.2	Previous response	804.2	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4913	78.9	799.4	800.6	0.999
Mid point	4955	39.5	400.0	401.1	0.997
Low point	4971	19.7	199.7	199.9	0.999
As left zero	5000	0.0	0.0	0.1	----
As left span	4913	78.9	799.4	803.5	0.995
Average Correction Factor:					0.998

Notes: Changed the sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 3, 2025	Previous Calibration:	September 9, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:24	End Time (MST):	14:49
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999999	≥0.995
799.4	800.6	0.9986	Slope	1.001333	0.90 - 1.10
400.0	401.1	0.9973	Intercept	0.202353	+/-30
199.7	199.9	0.9988			



SO2 Calibration Plot

Date: October 3, 2025

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Jackfish 2/3 Station number: AMS 27
Calibration Date: October 24, 2025 Last Cal Date: September 18, 2025
Start time (MST): 9:22 End time (MST): 14:35
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.87 ppm Cal Gas Exp Date: September 5, 2027
Cal Gas Cylinder #: CC523090
Removed Cal Gas Conc: 4.87 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 5252
ZAG Make/Model: Teledyne API T701H Serial Number: 268

Analyzer Information

Analyzer make: Thermo 43iQ Analyzer serial #: 12228021055
Converter make: Global G150 Converter serial #: 2022-195
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.015588	1.004522	Backgd or Offset:	3.26
Calibration intercept:	0.115904	0.196003	Coeff or Slope:	1.228

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4911	82.0	80.0	80.3	0.996
As found Mid point	4950	41.0	40.0	40.1	0.998
As found Low point	4972	20.5	20.0	19.9	1.005
New cylinder response					
Baseline Corr As found:	80.3	Prev response:	81.34	*% change:	-1.3%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	1.004666	AF Intercept:	-0.084021
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4911	82.0	80.0	80.6	0.992
Mid point	4950	41.0	40.0	40.2	0.995
Low point	4972	20.5	20.0	20.4	0.980
As left zero	5000	0.0	0.0	0.3	----
As left span	4911	82.0	80.0	86.0	0.930
SO2 Scrubber Check	4915	78.9	790.0	0.1	----
Date of last scrubber change:	21-Feb-25		Ave Corr Factor		0.989
Date of last converter efficiency test:	April 23, 2025		91.4% efficiency		

Notes: Changed sample inlet filter after as founds. Scrubber check passed. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	October 24, 2025	Previous Calibration:	September 18, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	9:22	End Time (MST):	14:35
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12228021055

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999986	≥ 0.995
80.0	80.6	0.9923	Slope	1.004522	$0.90 - 1.10$
40.0	40.2	0.9952	Intercept	0.196003	± 3
20.0	20.4	0.9802			



H₂S Calibration Plot

Date: October 24, 2025

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Jackfish 2/3
Station number: AMS 27
Calibration Date: October 23, 2025
Last Cal Date: September 19, 2025
Start time (MST): 10:38
End time (MST): 16:00
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC757838
NOX Cal Gas Conc: 60.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.30 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 60.20 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.20 ppm
NO gas Diff:
Serial Number: 5252
Serial Number: 268

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
AF High point	4924	66.3	801.1	799.8	1.3	723.2	719.4	3.8	1.1078	1.1118
AF Mid point										
AF Low point										
New cyl resp										

Previous Response NO_x = 800.1 ppb NO = 797.6 ppb
Baseline Corr 1st pt NO_x = 723.2 ppb NO = 719.4 ppb
Baseline Corr 2nd pt NO_x = NA ppb NO = NA ppb
Baseline Corr 3rd pt NO_x = NA ppb NO = NA ppb

* = > +/-5% change initiates investigation

*Percent Change NO_x = -10.6%

As Found Statistics

*Percent Change NO = -10.9%

As found NO_x r²:
As found NO r²:
As found NO₂ r²:

Nx SI: Nx Int:
NO SI: NO Int:
NO₂ SI: NO₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153357

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.379	1.524	NO bkgnd or offset:	5.6	6.1
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	5.7	6.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	173.2	179.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997431	0.997189
NO _x Cal Offset:	1.070038	0.630204
NO Cal Slope:	0.997915	0.999286
NO Cal Offset:	-0.569212	-1.049205
NO ₂ Cal Slope:	1.005946	0.999326
NO ₂ Cal Offset:	0.227357	0.407685

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.2	0.2	----	----
High point	4924	66.3	801.1	799.8	1.3	799.4	798.8	0.7	1.0022	1.0013
Mid point	4958	33.2	401.1	400.4	0.7	400.6	398.5	2.0	1.0012	1.0049
Low point	4976	16.6	200.5	200.2	0.3	201.0	197.7	3.4	0.9975	1.0124
As left zero	5000	0.0	0.0	0.0	0.0	1.3	0.1	1.2	----	----
As left span	4924	66.3	801.1	365.6	435.5	797.3	365.6	431.7	1.0048	1.0000
Average Correction Factor									1.0003	1.0062

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	795.0	364.3	432.0	432.2	0.9996	100.0%
Mid GPT point	795.0	576.1	220.2	220.1	1.0006	99.9%
Low GPT point	795.0	682.2	114.1	115.0	0.9924	100.8%
Average Correction Factor					0.9975	100.2%

Notes: An investigation was conducted, diagnostics were consistent with previous calibration results except for a slight decrease in flow (from 0.52 L/min to 0.49 L/min). No issues were identified with the setup, and the flow remained within acceptable limits. Changed sample inlet filter after as founds. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 19, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	10:38	End Time (MST):	16:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995
801.1	799.4	1.0022	Slope	0.997189	0.90 - 1.10
401.1	400.6	1.0012	Intercept	0.630204	+/-20
200.5	201.0	0.9975			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 19, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	10:38	End Time (MST):	16:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999995	≥ 0.995
432.0	432.2	0.9996	Slope	0.999326	$0.90 - 1.10$
220.2	220.1	1.0006	Intercept	0.407685	± 20
114.1	115.0	0.9924			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 19, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	10:38	End Time (MST):	16:00
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999989	≥ 0.995
799.8	798.8	1.0013	Slope	0.999286	$0.90 - 1.10$
400.4	398.5	1.0049	Intercept	-1.049205	± 20
200.2	197.7	1.0124			



NO_x Calibration Plot

Date: October 23, 2025

Location: Jackfish 2/3





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Station Information

Station Name:	Jackfish	Station Number:	AMS 27
Calibration Date:	October 3, 2025	Prev Cal Date:	October 15, 2024
Start Time (MST):	12:45	End Time (MST):	15:07
Tower Height (m):	10.0	Reason:	Routine

Wind Speed Calibration

Sensor make/model:	Met One 010C-1	Serial Number:	X16480
WS Calibrator:	MetOne 053	Serial Number:	R10866

Shaft RPM (Hz)	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.4	0.1%
600	58.6	58.7	0.2%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	0.999999	0.999999	≥ 0.9995
Calculated slope	0.998137	0.998121	$0.98 - 1.02$
Calculated intercept	0.030641	0.030255	± 2

Wind Direction Calibration

Sensor make/model:	Met One 020C-1	Serial Number:	X16480
As Found Declination (deg east of True North):	<u>13</u>	As Left Declination (deg east of True North):	<u>13</u>
Solar noon (MST):	1:12:55 PM	Calc Declination*:	12.78 Degrees
WD Calibrator:	Met One 040	* - calculated declination as per NOAA website	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 360° FS) <i>Limit = +/- 1%</i>
0	2.1	0.6%
90	91.6	0.5%
180	180.6	0.2%
270	270.6	0.2%
355	357.1	0.6%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	0.999977	0.999987	≥ 0.9995
Calculated slope	0.994058	1.000975	$0.97 - 1.03$
Calculated intercept	0.184524	-1.597477	± 5

Notes: Bearings still good. Cross arm was aligned using compass after calibration.

Calibration Performed By: Mohammed Kashif



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS29
SURMONT 2
OCTOBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: October 24, 2025 Last Cal Date: September 5, 2025
Start time (MST): 10:13 End time (MST): 14:34
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.95 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC356229
Removed Cal Gas Conc: 49.95 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 5472
Zero Air Gen Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1170050150
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001760	0.996831	Backgd or Offset:	14.3
Calibration intercept:	-1.340582	-0.740911	Coeff or Slope:	0.940

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4920	80.1	800.2	804.0	0.995
As found Mid point	4960	40.0	399.6	400.6	0.997
As found Low point	4980	20.0	199.8	199.3	1.002
New cylinder response					
Baseline Corr As found:	804.2	Previous response	800.3	*% change	0.5%
Baseline Corr 2nd AF pt:	400.8	AF Slope:	1.005444	AF Intercept:	-0.879437
Baseline Corr 3rd AF pt:	199.5	AF Correlation:	0.999997	* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4920	80.1	800.2	797.0	1.004
Mid point	4960	40.0	399.6	397.9	1.004
Low point	4980	20.0	199.8	197.5	1.012
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	80.1	800.2	796.8	1.004
Average Correction Factor:					1.007

Notes: Changed sample inlet filter and external pump after as founds. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 24, 2025	Previous Calibration:	September 5, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:13	End Time (MST):	14:34
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999996	≥0.995
800.2	797.0	1.0040	Slope	0.996831	0.90 - 1.10
399.6	397.9	1.0043	Intercept	-0.740911	+/-30
199.8	197.5	1.0116			



SO2 Calibration Plot

Date: October 24, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: October 2, 2025 Last Cal Date: September 9, 2025
Start time (MST): 9:29 End time (MST): 14:42
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.750 ppm Cal Gas Exp Date: August 28, 2027
Cal Gas Cylinder #: CC737848
Removed Cal Gas Conc: 4.750 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 5472
ZAG Make/Model: Teledyne API T701 Serial Number: 4428

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 1200326170
Converter make: Global Converter serial #: 2022-220
Analyzer Range: 0 - 100 ppb Converter Temp: 325.0 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.988594	0.996738	Backgd or Offset:	0.82
Calibration intercept:	-0.020475	-0.080477	Coeff or Slope:	1.040

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4916	84.2	80.0	80.8	0.990
As found Mid point	4958	42.1	40.0	40.3	0.992
As found Low point	4979	21.1	20.0	20.2	0.990
New cylinder response					
Baseline Corr As found:	80.8	Prev response:	79.05	*% change:	2.2%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.010026	AF Intercept:	-0.020484
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999998	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4916	84.2	80.0	79.7	1.004
Mid point	4958	42.1	40.0	39.7	1.007
Low point	4979	21.1	20.0	19.8	1.010
As left zero	5000	0.0	0.0	0.2	----
As left span	4916	84.2	80.0	78.4	1.020
SO2 Scrubber Check	4919	81.3	813.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	1.007
Date of last converter efficiency test:		December 5, 2024		108.1% efficiency	

Notes: Changed sample inlet filter after as founds. Ran SOx scrubber check after cal zero, passed. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H2S Calibration Summary


Station Information

Calibration Date:	October 2, 2025	Previous Calibration:	September 9, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:29	End Time (MST):	14:42
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999995		≥ 0.995
80.0	79.7	1.0036	Slope	0.996738		$0.90 - 1.10$
40.0	39.7	1.0074	Intercept	-0.080477		± 3
20.0	19.8	1.0100				


H₂S Calibration Curve



H2S Calibration Plot

Date: October 2, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Gas Date: October 24, 2025 Last Cal Date: September 5, 2025
Start time (MST): 10:12 End time (MST): 14:34
Reason: Routine

Calibration Standards

Gas Cert Reference: CC356229 Cal Gas Expiry Date: October 9, 2032
CH4 Cal Gas Conc. 503.7 ppm CH4 Equiv Conc. 1066.9 ppm
C3H8 Cal Gas Conc. 204.8 ppm
Removed Gas Cert: NA Removed Gas Expiry: NA
Removed CH4 Conc. 503.7 ppm CH4 Equiv Conc. 1066.9 ppm
Removed C3H8 Conc. 204.8 ppm Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 5472
ZAG Make/Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1170050149
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004862	0.998037	Background:	3.68
Calibration intercept:	-0.055723	-0.023967	Coefficient:	4.114

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	-0.13	----
As found High point	4920	80.1	17.09	16.81	1.009
As found Mid point	4960	40.0	8.54	8.26	1.017
As found Low point	4980	20.0	4.27	4.01	1.031
New cylinder response					
Baseline Corr As found:	16.94	Previous response	17.12	*% change	-1.0%
Baseline Corr 2nd AF pt:	8.39	AF Slope:	0.992642	AF Intercept:	-0.182390
Baseline Corr 3rd AF pt:	4.14	AF Correlation:	0.999961	* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.03	----
High point	4920	80.1	17.09	17.06	1.002
Mid point	4960	40.0	8.54	8.47	1.008
Low point	4980	20.0	4.27	4.18	1.021
As left zero	5000	0.0	0.00	0.02	----
As left span	4920	80.1	17.09	17.12	0.998
Average Correction Factor					1.010

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	October 24, 2025	Previous Calibration:	September 5, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:12	End Time (MST):	14:34
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.03	----	Correlation Coefficient	0.999956	≥ 0.995
17.09	17.06	1.0019	Slope	0.998037	$0.90 - 1.10$
8.54	8.47	1.0077	Intercept	-0.023967	± 1.5
4.27	4.18	1.0207			

THC Calibration Curve



THC Calibration Plot

Date: October 24, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Surmont 2
Station number: AMS 29
Calibration Date: October 3, 2025
Last Cal Date: September 4, 2025
Start time (MST): 9:36
End time (MST): 16:30
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC218007
NOX Cal Gas Conc: 60.20 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.20 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 5472
Serial Number: 4428

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0	0.0	0.0	0.0	-0.5	-0.3	-0.2	----	----
AF High point	4933	66.7	803.1	800.4	2.7	813.0	805.0	7.9	0.9872	0.9940
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 801.4 ppb	NO = 797.7 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = 1.5%
Baseline Corr 1st pt	NO _x = 813.5 ppb	NO = 805.3 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.9%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1170050148

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998556	0.984898
NO _x Cal Offset:	-0.568938	2.354041
NO Cal Slope:	0.999656	0.989195
NO Cal Offset:	-2.488482	-0.025951
NO ₂ Cal Slope:	1.005325	1.001984
NO ₂ Cal Offset:	-0.127845	0.056024

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.042	1.036	NO bkgnd or offset:	1.5	1.3
NOX coeff or slope:	0.996	0.992	NOX bkgnd or offset:	1.9	1.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	151.0	149.8

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOX Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4933	66.7	803.1	800.4	2.7	790.4	790.0	0.4	1.0161	1.0132
Mid point	4967	33.3	400.9	399.6	1.3	403.7	400.6	3.1	0.9931	0.9974
Low point	4983	16.7	201.1	200.4	0.7	199.2	194.6	4.6	1.0094	1.0299
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
As left span	4933	66.7	803.1	433.1	370.0	803.0	433.1	369.7	1.0001	1.0000
Average Correction Factor									1.0062	1.0135

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	796.0	422.7	376.0	376.5	0.9986	100.1%
Mid GPT point	796.0	622.2	176.5	177.7	0.9931	100.7%
Low GPT point	796.0	710.3	88.4	88.1	1.0030	99.7%
Average Correction Factor					0.9982	100.2%

Notes:

Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 3, 2025	Previous Calibration:	September 4, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:36	End Time (MST):	16:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999831	≥0.995
803.1	790.4	1.0161	Slope	0.984898	0.90 - 1.10
400.9	403.7	0.9931	Intercept	2.354041	+/-20
201.1	199.2	1.0094			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 3, 2025	Previous Calibration:	September 4, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:36	End Time (MST):	16:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999987	≥0.995
376.0	376.5	0.9986	Slope	1.001984	0.90 - 1.10
176.5	177.7	0.9931	Intercept	0.056024	+/-20
88.4	88.1	1.0030			





Wood Buffalo Environmental Association

NO Calibration Summary


Station Information

Calibration Date:	October 3, 2025	Previous Calibration:	September 4, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	9:36	End Time (MST):	16:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999869	≥ 0.995
800.4	790.0	1.0132	Slope	0.989195	$0.90 - 1.10$
399.6	400.6	0.9974	Intercept	-0.025951	± 20
200.4	194.6	1.0299			


NO Calibration Curve



NO_x Calibration Plot

Date: October 3, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: October 24, 2025 Last Cal Date: September 9, 2025
Start time (MST): 10:25 End time (MST): 12:15

Analyzer Make: API T640 S/N: 2236
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	7.3	7.55	7.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	704.0	702.4	704.0	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.952	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	50	----	50	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 3.1		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: July 16, 2026
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.7	10.9	10.9	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: October 24, 2025
Date Disposable Filter Changed: October 24, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: April 14, 2025
Date RH/T Sensor Cleaned: July 15, 2025

Notes: Verified temperature, pressure and flow. Leak check passed. Noted high pump PW% change, will replace next month.

Calibration by: Braiden Boutilier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS30 ELLS RIVER OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Ells River Station number: AMS 30
Calibration Date: October 14, 2025 Last Cal Date: September 8, 2025
Start time (MST): 9:49 End time (MST): 12:28
Reason: Routine

Calibration Standards

Cal Gas Concentration: 48.75 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC350110
Removed Cal Gas Conc: 48.75 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3061
Zero Air Gen Model: API T701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1008841397
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998444	0.994713	Backgd or Offset:	10.6	10.6
Calibration intercept:	-3.132072	-2.552091	Coeff or Slope:	0.991	0.991

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4918	82.0	799.5	793.6	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	794.0	Previous response	795.1	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4918	82.0	799.5	794.3	1.007
Mid point	4959	41.0	399.8	393.0	1.017
Low point	4980	20.5	199.9	194.1	1.030
As left zero	5000	0.0	0.0	-0.1	----
As left span	4918	82.0	799.5	796.6	1.004
Average Correction Factor:					1.018

Notes: Sample inlet filter was changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 14, 2025	Previous Calibration:	September 8, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:49	End Time (MST):	12:28
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999947	≥0.995
799.5	794.3	1.0065	Slope	0.994713	0.90 - 1.10
399.8	393.0	1.0172	Intercept	-2.552091	+/-30
199.9	194.1	1.0296			



SO2 Calibration Plot

Date: October 14, 2025

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Ells River Station number: AMS 30
Calibration Date: October 1, 2025 Last Cal Date: September 15, 2025
Start time (MST): 9:55 End time (MST): 13:38
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.99 ppm Cal Gas Exp Date: November 15, 2026
Cal Gas Cylinder #: CC505806
Removed Cal Gas Conc: 4.99 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3061
ZAG Make/Model: API 701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43iq TL Analyzer serial #: 12426335708
Converter make: CDN- 101 Converter serial #: 562
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998054	1.000049	Backgd or Offset:	1.5	1.19
Calibration intercept:	-0.360734	0.019440	Coeff or Slope:	1.057	1.095

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4920	80.2	80.0	76.0	1.048
As found Mid point	4960	40.1	40.0	38.0	1.042
As found Low point	4980	20.0	20.0	18.6	1.051
New cylinder response					
Baseline Corr As found:	76.4	Prev response:	79.52	*% change:	-4.1%
Baseline Corr 2nd AF pt:	38.4	AF Slope:	0.955079	AF Intercept:	-0.381437
Baseline Corr 3rd AF pt:	19.0	AF Correlation:	0.999989	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.2	80.0	80.0	1.000
Mid point	4960	40.1	40.0	40.2	0.995
Low point	4980	20.0	20.0	19.9	1.003
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	80.2	80.0	81.2	0.986
SO2 Scrubber Check	4918	82.0	820.0	0.1	----
Date of last scrubber change:	13-Aug-25		Ave Corr Factor		1.000
Date of last converter efficiency test:					

Notes: Sample inlet filter was changed after multipoint as founds. Adjusted zero and span. SO2 scrubber check done and passed.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	October 1, 2025	Previous Calibration:	September 15, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:55	End Time (MST):	13:38
Analyzer make:	Thermo 43iq TL	Analyzer serial #:	12426335708

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999990		≥ 0.995
80.0	80.0	1.0005	Slope	1.000049		$0.90 - 1.10$
40.0	40.2	0.9955	Intercept	0.019440		± 3
20.0	19.9	1.0030				

TRS Calibration Curve



TRS Calibration Plot

Date: October 1, 2025

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Ells River
Calibration Date: October 28, 2025
Start time (MST): 8:30
Reason: Removal

Station number: AMS 30
Last Cal Date: October 1, 2025
End time (MST): 9:46

Calibration Standards

Cal Gas Concentration: 4.99 ppm
Cal Gas Cylinder #: CC505806
Removed Cal Gas Conc: 4.99 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API 701H

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43iq TL
Converter make: CDN- 101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12426335708
Converter serial #: 562
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000049		Backgd or Offset:	1.2	NA
Calibration intercept:	0.019440		Coeff or Slope:	1.095	NA

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	80.2	80.0	81.3	0.984
As found Mid point	4960	40.1	40.0	40.4	0.991
As found Low point	4980	20.0	20.0	20.0	0.998
New cylinder response					
Baseline Corr As found:	81.3	Prev response:	80.06	*% change:	1.5%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.016610	AF Intercept:	-0.160260
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999981	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					
SO2 Scrubber Check					
Date of last scrubber change:	13-Aug-25			Ave Corr Factor	<input type="text"/>
Date of last converter efficiency test:					


Notes: Removal calibrations done to do further troubleshooting at the shop.

Calibration Performed By: Jan Castro

TRS Calibration Plot

Date: October 28, 2025

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Ells River
Calibration Date: October 30, 2025
Start time (MST): 9:25
Reason: Install

Station number: AMS 30
Last Cal Date:
End time (MST): 11:41

Calibration Standards

Cal Gas Concentration: 4.99 ppm
Cal Gas Cylinder #: CC505806
Removed Cal Gas Conc: 4.99 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API 701H

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i-LTE
Converter make: CDN- 101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1170050152
Converter serial #: CDN606
Converter Temp: 860 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		0.997762	Backgd or Offset:		3.83
Calibration intercept:		-0.100519	Coeff or Slope:		1.140

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4920	80.2	80.0	79.9	1.002
Mid point	4960	40.1	40.0	39.5	1.013
Low point	4980	20.0	20.0	19.9	1.003
As left zero	5000	0.0	0.0	-0.2	----
As left span	4920	80.2	80.0	81.7	0.980
SO2 Scrubber Check	5000	80.0	787.4	-0.1	----
Date of last scrubber change:	13-Aug-25		Ave Corr Factor		1.006
Date of last converter efficiency test:					

Notes: Install of new analyzer and converter. All lines replaced. Sox scrubber checked after calibrator zero. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	October 30, 2025	Previous Calibration:	
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:25	End Time (MST):	11:41
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999958		≥ 0.995
80.0	79.9	1.0017	Slope	0.997762		$0.90 - 1.10$
40.0	39.5	1.0131	Intercept	-0.100519		± 3
20.0	19.9	1.0030				


TRS Calibration Curve



TRS Calibration Plot

Date: October 30, 2025

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Ells River
Calibration Date: October 14, 2025
Start time (MST): 9:49
Reason: Routine

Station number: AMS 30
Last Cal Date: September 8, 2025
End time (MST): 12:28

Calibration Standards

Gas Cert Reference: CC350110
CH₄ Cal Gas Conc. 496.6 ppm
C₃H₈ Cal Gas Conc. 207.2 ppm
Removed Gas Cert: NA
Removed CH₄ Conc. 496.6 ppm
Removed C₃H₈ Conc. 207.2 ppm
Diff between cyl (CH₄):
Calibrator Model: API T700
Zero Air Gen model: API T701H

Cal Gas Expiry Date: Monday, March 10, 2031
CH₄ Equiv Conc. 1066.4 ppm
Removed Gas Expiry: NA
CH₄ Equiv Conc. 1066.4 ppm
Diff between cyl (THC):
Diff between cyl (NM):
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1152430011
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.12E-04	3.12E-04	NMHC SP Ratio:	5.89E-05
CH ₄ Retention time:	17.4	17.4	NMHC Peak Area:	157597
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	17.49	17.43	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.43	Prev response	17.48	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	82.0	17.49	17.43	1.003
Mid point	4959	41.0	8.74	8.63	1.014
Low point	4980	20.5	4.37	4.25	1.028
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	17.49	17.37	1.007
Average Correction Factor					1.015

Notes:

Sample inlet filter was changed after as founds. No adjustment made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	9.34	9.31	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.31	Prev response	9.37	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	82.0	9.34	9.33	1.001
Mid point	4959	41.0	4.67	4.62	1.010
Low point	4980	20.5	2.34	2.29	1.018
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	9.34	9.27	1.008
Average Correction Factor					1.010

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	8.14	8.12	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.12	Prev response	8.12	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	82.0	8.14	8.10	1.005
Mid point	4959	41.0	4.07	4.00	1.017
Low point	4980	20.5	2.04	1.96	1.040
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	8.14	8.10	1.006
Average Correction Factor					1.021

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.003028	0.998290
THC Cal Offset:	-0.059225	-0.060226
CH ₄ Cal Slope:	1.001618	0.997226
CH ₄ Cal Offset:	-0.041319	-0.036919
NMHC Cal Slope:	1.004280	0.999437
NMHC Cal Offset:	-0.018506	-0.023707

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	October 14, 2025	Previous Calibration:	September 8, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:49	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999944	<i>≥0.995</i>
17.49	17.43	1.0032	Slope	0.998290	<i>0.90 - 1.10</i>
8.74	8.63	1.0136	Intercept	-0.060226	<i>+/-0.5</i>
4.37	4.25	1.0282			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	October 14, 2025	Previous Calibration:	September 8, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:49	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999905	<i>≥0.995</i>
8.14	8.10	1.0050	Slope	0.997226	<i>0.90 - 1.10</i>
4.07	4.00	1.0173	Intercept	-0.036919	<i>+/-0.5</i>
2.04	1.96	1.0398			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	October 14, 2025	Previous Calibration:	September 8, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	9:49	End Time (MST):	12:28
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999967	≥ 0.995
9.34	9.33	1.0015	Slope	0.999437	0.90 - 1.10
4.67	4.62	1.0105	Intercept	-0.023707	± 0.5
2.34	2.29	1.0183			



NMHC Calibration Plot

Date: October 14, 2025

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Ells River
Station number: AMS 30
Calibration Date: October 6, 2025
Last Cal Date: September 4, 2025
Start time (MST): 10:01
End time (MST): 14:02
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0027487
NO_x Cal Gas Conc: 59.30 ppm
Removed Cylinder #: NA
Removed Gas NO_x Conc: 59.30 ppm
NO_x gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.10 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.10 ppm
NO gas Diff:
Serial Number: 3061
Serial Number: 358

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.3	0.0	-0.2	----	----
AF High point	4932	67.7	803.0	800.3	2.7	796.0	788.1	7.9	1.0084	1.0154
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 801.1 ppb	NO = 796.6 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -0.6%	
Baseline Corr 1st pt	NO _x = 796.3 ppb	NO = 788.1 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -1.1%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 710321429

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000644	0.999406
NO _x Cal Offset:	-2.399282	-2.239384
NO Cal Slope:	1.000320	1.000405
NO Cal Offset:	-3.900988	-3.180739
NO ₂ Cal Slope:	1.003151	0.999890
NO ₂ Cal Offset:	0.343802	0.136766

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.335	1.356	NO bkgnd or offset:	15.6	15.6
NOX coeff or slope:	0.995	0.992	NOX bkgnd or offset:	15.8	15.8
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	202.6	206.5

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
High point	4932	67.7	803.0	800.3	2.7	801.3	799.1	2.1	1.0021	1.0015
Mid point	4966	33.8	400.9	399.5	1.4	397.5	394.7	2.8	1.0085	1.0122
Low point	4983	16.9	200.4	199.8	0.7	195.7	193.5	2.2	1.0242	1.0324
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
As left span	4932	67.7	803.0	429.6	373.4	802.1	429.6	372.6	1.0011	1.0000
Average Correction Factor									1.0116	1.0154

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	794.9	426.3	371.3	371.2	1.0003	100.0%
Mid GPT point	794.9	612.3	185.3	185.8	0.9974	100.3%
Low GPT point	794.9	702.3	95.3	95.5	0.9980	100.2%
Average Correction Factor					0.9985	100.1%

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 6, 2025	Previous Calibration:	September 4, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:01	End Time (MST):	14:02
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999964	≥0.995
803.0	801.3	1.0021	Slope	0.999406	0.90 - 1.10
400.9	397.5	1.0085	Intercept	-2.239384	+/-20
200.4	195.7	1.0242			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 6, 2025	Previous Calibration:	September 4, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:01	End Time (MST):	14:02
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999997	≥0.995
371.3	371.2	1.0003	Slope	0.999890	0.90 - 1.10
185.3	185.8	0.9974	Intercept	0.136766	+/-20
95.3	95.5	0.9980			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 6, 2025	Previous Calibration:	September 4, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:01	End Time (MST):	14:02
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999923	≥ 0.995
800.3	799.1	1.0015	Slope	1.000405	$0.90 - 1.10$
399.5	394.7	1.0122	Intercept	-3.180739	± 20
199.8	193.5	1.0324			



NO_x Calibration Plot

Date: October 6, 2025

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Ells River Station number: AMS 30
Calibration Date: October 14, 2025 Last Cal Date: September 2, 2025
Start time (MST): 11:13 End time (MST): 11:32

Analyzer Make: API T640 S/N: 875
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	6.20	5.68	6.20	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	731.10	733.10	731.10	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.04	5.03	5.04	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	32	----	32	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 1.60		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: January 30, 2027
Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: September 2, 2025
Date Disposable Filter Changed: September 2, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: June 18, 2025
Date RH/T Sensor Cleaned: June 18, 2025

Notes: Verified flow, temperature, pump power and pressure. No adjustment made. Leak check passed.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS31
BLACKROD
OCTOBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Blackrod Station number: AMS 31
Calibration Date: October 30, 2025 Last Cal Date: September 8, 2025
Start time (MST): 10:51 End time (MST): 13:40
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.25 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC327023
Removed Cal Gas Conc: 50.25 ppm Rem Gas Exp Date: N/A
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Model: Teledyne T700 Serial Number: 1220
Zero Air Gen Model: Teledyne N701H Serial Number: 72

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1160290014
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010356	1.003523	Backgd or Offset:	37.7	39.7
Calibration intercept:	-0.011908	-0.682143	Coeff or Slope:	0.980	0.997

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.9	----
As found High point	4920	79.6	800.0	788.9	1.015
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	788.0	Previous response	808.3	*% change	-2.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	4921	79.5	798.9	801.4	0.997
Mid point	4960	39.8	400.0	400.0	1.000
Low point	4980	19.9	200.0	200.0	1.000
As left zero	5000	0.0	0.0	-0.3	----
As left span	4921	79.5	798.9	801.0	0.997
Average Correction Factor:					0.999

Notes: Sample inlet filter was changed after as founds. Adjusted zero and span.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 30, 2025	Previous Calibration:	September 8, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	10:51	End Time (MST):	13:40
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290014

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999998	≥0.995
798.9	801.4	0.9969	Slope	1.003523	0.90 - 1.10
400.0	400.0	1.0000	Intercept	-0.682143	+/-30
200.0	200.0	1.0000			



SO2 Calibration Plot

Date: October 30, 2025

Location: Blackrod





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Blackrod
Calibration Date: October 9, 2025
Start time (MST): 10:15
Reason: Routine

Station number: AMS 31
Last Cal Date: September 17, 2025
End time (MST): 14:46

Calibration Standards

Cal Gas Concentration: 5.42 ppm
Cal Gas Cylinder #: DT0016926
Removed Cal Gas Conc: 5.42 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API N701H

Cal Gas Exp Date: March 19, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 1220
Serial Number: 72

Analyzer Information

Analyzer make: Thermo 43iQTL
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12228021056
Converter serial #: 2023-266
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003618	1.006763	Backgd or Offset: 2.48	2.54
Calibration intercept:	-0.140377	-0.000490	Coeff or Slope: 0.921	0.944

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4926	73.8	80.0	78.6	1.018
As found Mid point	4963	36.9	40.0	39.4	1.015
As found Low point	4982	18.5	20.1	19.6	1.023
New cylinder response					
Baseline Corr As found:	78.6	Prev response:	80.15	*% change:	-2.0%
Baseline Corr 2nd AF pt:	39.4	AF Slope:	0.983044	AF Intercept:	-0.020004
Baseline Corr 3rd AF pt:	19.6	AF Correlation:	0.999994	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4926	73.8	80.0	80.5	0.994
Mid point	4963	36.9	40.0	40.4	0.990
Low point	4982	18.5	20.1	20.1	0.998
As left zero	5000	0.0	0.0	0.1	----
As left span	4926	73.8	80.0	80.4	0.995
SO2 Scrubber Check	4921	79.5	794.9	0.0	----
Date of last scrubber change:	27-Aug-25		Ave Corr Factor		0.994
Date of last converter efficiency test:	October 9, 2025		97.9% efficiency		

Notes: Sample inlet filter was changed after multipoint as founds. Converter efficiency done. SO2 scrubber check done and passed. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	October 9, 2025	Previous Calibration:	September 17, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	10:15	End Time (MST):	14:46
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12228021056

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999993		≥ 0.995
80.0	80.5	0.9938	Slope	1.006763		$0.90 - 1.10$
40.0	40.4	0.9901	Intercept	-0.000490		± 3
20.1	20.1	0.9976				



H₂S Calibration Plot

Date: October 9, 2025

Location: Blackrod





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Blackrod
Station number: AMS 31
Calibration Date: October 21, 2025
Last Cal Date: September 9, 2025
Start time (MST): 10:12
End time (MST): 14:33
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0035071
NOX Cal Gas Conc: 59.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.30 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API N701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.10 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.10 ppm
NO gas Diff:
Serial Number: 1220
Serial Number: 72

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.4	-0.1	----	----
AF High point	4932	67.7	803.0	800.3	2.7	719.2	711.4	7.7	1.1157	1.1243
AF Mid point										
AF Low point										
New cyl resp										

Previous Response NO_x = 805.1 ppb NO = 802.1 ppb
Baseline Corr 1st pt NO_x = 719.7 ppb NO = 711.8 ppb
Baseline Corr 2nd pt NO_x = NA ppb NO = NA ppb
Baseline Corr 3rd pt NO_x = NA ppb NO = NA ppb

* = > +/-5% change initiates investigation

*Percent Change NO_x = -11.9%

*Percent Change NO = -12.7%

As Found Statistics

As found NO_x r²:
As found NO r²:
As found NO₂ r²:
Nx SI:
NO SI:
NO₂ SI:
Nx Int:
NO Int:
NO₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1426262592

Calibration Statistics

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.059	1.205	NO bkgnd or offset:	15.8	17.9	NO _x Cal Slope:	1.003173	1.000210
NOX coeff or slope:	0.997	0.995	NOX bkgnd or offset:	15.9	18.1	NO _x Cal Offset:	-0.437111	1.903333
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	143.6	141.2	NO Cal Slope:	1.003814	1.001497
						NO Cal Offset:	-1.198337	0.862201
						NO ₂ Cal Slope:	1.004641	0.974019
						NO ₂ Cal Offset:	-0.688005	3.859945

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.3	0.1	----	----
High point	4932	67.7	803.0	800.3	2.7	804.0	801.8	2.3	0.9987	0.9981
Mid point	4966	33.8	400.9	399.5	1.4	404.0	401.4	2.6	0.9923	0.9953
Low point	4983	16.9	200.4	199.8	0.7	204.3	202.2	2.1	0.9811	0.9879
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
As left span	4825	67.7	820.5	421.1	381.9	823.1	421.1	401.9	0.9969	1.0000
Average Correction Factor									0.9907	0.9938

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	815.4	414.6	403.5	393.6	1.0252	97.5%
Mid GPT point	815.4	618.7	199.4	204.3	0.9761	102.5%
Low GPT point	815.4	714.2	103.9	105.9	0.9812	101.9%
Average Correction Factor					0.9941	100.6%

Notes: Sample inlet filter was changed after as founds. Adjusted span only. Used 2nd NO point because of drift.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 21, 2025	Previous Calibration:	September 9, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	10:12	End Time (MST):	14:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999972	≥0.995
803.0	804.0	0.9987	Slope	1.000210	0.90 - 1.10
400.9	404.0	0.9923	Intercept	1.903333	+/-20
200.4	204.3	0.9811			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 21, 2025	Previous Calibration:	September 9, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	10:12	End Time (MST):	14:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999236	≥ 0.995
403.5	393.6	1.0252	Slope	0.974019	$0.90 - 1.10$
199.4	204.3	0.9761	Intercept	3.859945	± 20
103.9	105.9	0.9812			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 21, 2025	Previous Calibration:	September 9, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	10:12	End Time (MST):	14:33
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262592

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999990	≥ 0.995
800.3	801.8	0.9981	Slope	1.001497	$0.90 - 1.10$
399.5	401.4	0.9953	Intercept	0.862201	± 20
199.8	202.2	0.9879			



NO_x Calibration Plot

Date: October 21, 2025

Location: Blackrod





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS33 MONDAY CREEK OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Monday Creek Station number: AMS 33
Calibration Date: October 23, 2025 Last Cal Date: September 25, 2025
Start time (MST): 10:05 End time (MST): 12:46
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.62 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: EB0008522
Removed Cal Gas Conc: 50.62 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne T700 Serial Number: 3253
Zero Air Gen Model: Teledyne T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1152430005
Analyzer Range: 0- 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003381	0.996658	Backgd or Offset:	31.0	31.0
Calibration intercept:	-0.137802	0.442201	Coeff or Slope:	0.970	0.970

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4921	79.1	800.8	798.0	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	798.3	Previous response	803.4	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4921	79.1	800.8	798.1	1.003
Mid point	4961	39.5	399.9	400.0	1.000
Low point	4980	19.8	200.5	200.0	1.002
As left zero	5000	0.0	0.0	-0.2	----
As left span	4921	79.1	800.8	800.3	1.001
Average Correction Factor:					1.002

Notes: Sample inlet filter was changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 25, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	10:05	End Time (MST):	12:46
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430005

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999996	≥0.995
800.8	798.1	1.0034	Slope	0.996658	0.90 - 1.10
399.9	400.0	0.9996	Intercept	0.442201	+/-30
200.5	200.0	1.0023			



SO2 Calibration Plot

Date: October 23, 2025

Location: Monday Creek





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Monday Creek
Calibration Date: October 22, 2025
Start time (MST): 9:59
Reason: Routine

Station number: AMS 33
Last Cal Date: September 10, 2025
End time (MST): 14:13

Calibration Standards

Cal Gas Concentration: 5.05 ppm
Cal Gas Cylinder #: DT0014831
Removed Cal Gas Conc: 5.05 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne T701H

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3253
Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL
Converter make: Global 150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12333331547
Converter serial #: 2022-196
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006712	0.996282	Backgd or Offset:	1.3	1.3
Calibration intercept:	0.118386	0.358401	Coeff or Slope:	1.062	1.062

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4921	79.2	80.0	81.5	0.983
As found Mid point	4960	39.6	40.0	40.7	0.986
As found Low point	4980	19.8	20.0	20.3	0.992
New cylinder response					
Baseline Corr As found:	81.4	Prev response:	80.64	*% change:	0.9%
Baseline Corr 2nd AF pt:	40.6	AF Slope:	1.017842	AF Intercept:	0.036375
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999994	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4921	79.2	80.0	80.0	1.000
Mid point	4960	39.6	40.0	40.3	0.993
Low point	4980	19.8	20.0	20.3	0.985
As left zero	5000	0.0	0.0	0.2	----
As left span	4921	79.2	80.0	79.7	1.004
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:		11-Apr-24		Ave Corr Factor	0.993
Date of last converter efficiency test:		October 22, 2025		100.8%	efficiency

Notes: Sample inlet filter was changed after multipoint as founds. Converter efficiency done. SO2 scrubber check done and passed. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

H2S Calibration Summary


Station Information

Calibration Date:	October 22, 2025	Previous Calibration:	September 10, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	9:59	End Time (MST):	14:13
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331547

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999996		≥ 0.995
80.0	80.0	0.9999	Slope	0.996282		$0.90 - 1.10$
40.0	40.3	0.9925	Intercept	0.358401		± 3
20.0	20.3	0.9852				


H₂S Calibration Curve



H2S Calibration Plot

Date: October 22, 2025

Location: Monday Creek





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Monday Creek
Station number: AMS 33
Calibration Date: October 29, 2025
Last Cal Date: September 18, 2025
Start time (MST): 10:18
End time (MST): 14:22
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC755290
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: March 11, 2031
NO Cal Gas Conc: 48.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.70 ppm
NO gas Diff:
Serial Number: 3253
Serial Number: 832

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
AF High point	4918	82.1	802.9	799.6	3.3	797.5	795.3	2.2	1.0065	1.0053
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 798.3 ppb	NO = 793.7 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.1%
Baseline Corr 1st pt	NO _x = 797.7 ppb	NO = 795.4 ppb	<u>As Found Statistics</u>	*Percent Change	NO = 0.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :	NO SI:	NO Int:
			As found NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12426335704

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.993621	0.995970
NO _x Cal Offset:	0.470640	0.670358
NO Cal Slope:	0.992614	0.997903
NO Cal Offset:	-0.049056	-0.049822
NO ₂ Cal Slope:	0.994046	0.967535
NO ₂ Cal Offset:	0.760617	1.063691

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.045	1.045	NO bkgnd or offset:	1.4	1.4
NOX coeff or slope:	0.999	0.999	NOX bkgnd or offset:	1.5	1.5
NO2 coeff or slope:	0.990	0.990	Reaction cell Press:	138.5	136.2

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4918	82.1	802.9	799.6	3.3	800.0	797.8	2.2	1.0037	1.0023
Mid point	4959	41.1	401.9	400.3	1.6	401.4	399.8	1.6	1.0014	1.0013
Low point	4979	20.5	200.5	199.7	0.8	201.0	198.9	2.1	0.9976	1.0040
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
As left span	4918	82.1	802.9	400.0	402.9	783.0	400.0	383.0	1.0254	1.0000
Average Correction Factor									1.0009	1.0025

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	794.5	395.1	402.7	390.0	1.0325	96.9%
Mid GPT point	794.5	603.4	194.4	190.2	1.0220	97.8%
Low GPT point	794.5	701.6	96.2	94.8	1.0146	98.6%
Average Correction Factor					1.0230	97.8%

Notes:

Sample inlet filter was changed after as founds. No adjustment made.

Calibration Performed By:

Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 29, 2025	Previous Calibration:	September 18, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	10:18	End Time (MST):	14:22
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12426335704

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999997	≥0.995
802.9	800.0	1.0037	Slope	0.995970	0.90 - 1.10
401.9	401.4	1.0014	Intercept	0.670358	+/-20
200.5	201.0	0.9976			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 29, 2025	Previous Calibration:	September 18, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	10:18	End Time (MST):	14:22
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12426335704

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999962	≥0.995
402.7	390.0	1.0325	Slope	0.967535	0.90 - 1.10
194.4	190.2	1.0220	Intercept	1.063691	+/-20
96.2	94.8	1.0146			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 29, 2025	Previous Calibration:	September 18, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	10:18	End Time (MST):	14:22
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12426335704

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999999	≥ 0.995
799.6	797.8	1.0023	Slope	0.997903	0.90 - 1.10
400.3	399.8	1.0013	Intercept	-0.049822	+/-20
199.7	198.9	1.0040			



NO_x Calibration Plot

Date: October 29, 2025

Location: Monday Creek





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS501
LEISMER
OCTOBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Leismer Station number: AMS 501
Calibration Date: September 11, 2025 Last Cal Date:
Start time (MST): 11:30 End time (MST): 11:35
Reason: Install

Calibration Standards

Cal Gas Concentration: 50.52 ppm Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #: CC274266
Removed Cal Gas Conc: 50.52 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 2659
Zero Air Gen Model: Teledyne API T701 Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1160290011
Analyzer Range: 0-1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.000229	Backgd or Offset:	NA	19.8
Calibration intercept:	NA	0.064055	Coeff or Slope:	NA	0.971

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.6	----
High point	4921	79.2	800.2	801.1	0.999
Mid point	4960	39.6	400.2	398.9	1.003
Low point	4980	19.8	200.1	200.4	0.998
As left zero	5000	0.0	0.0	0.9	----
As left span	4921	79.2	800.2	797.6	1.003
Average Correction Factor:					1.000

Notes: Install Calibration. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	September 11, 2025	Previous Calibration:	
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	11:30	End Time (MST):	11:35
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999992	≥0.995
800.2	801.1	0.9989	Slope	1.000229	0.90 - 1.10
400.2	398.9	1.0031	Intercept	0.064055	+/-30
200.1	200.4	0.9983			



SO2 Calibration Plot

Date: September 11, 2025

Location: Leismer





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Leismer
Calibration Date: September 11, 2025
Start time (MST): 8:38
Reason: Install

Station number: AMS 501
Last Cal Date: NA
End time (MST): 11:32

Calibration Standards

Cal Gas Concentration: 4.89 ppm
Cal Gas Cylinder #: CC737971
Removed Cal Gas Conc: 4.89 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne 750
ZAG Make/Model: Teledyne 751H

Cal Gas Exp Date: September 5, 2027
Rem Gas Exp Date: NA
Diff between cyl: 282
Serial Number: 282
Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1180540020
Converter serial #: 2022-218
Converter Temp: 325.0 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	NA	1.014452	Backgd or Offset:	NA	3.78
Calibration intercept:	NA	-0.719189	Coeff or Slope:	NA	1.200

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	4918	81.8	80.0	80.7	0.991
Mid point	4959	40.9	40.0	39.5	1.013
Low point	4980	20.4	19.9	19.2	1.039
As left zero	5000	0.0	0.0	-0.2	----
As left span	4918	81.8	80.0	82.4	0.971
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.014
Date of last converter efficiency test:					

Notes:

Install Calibration. Adjusted span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	September 11, 2025	Previous Calibration:	NA
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	8:38	End Time (MST):	11:32
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999868		≥ 0.995
80.0	80.7	0.9914	Slope	1.014452		$0.90 - 1.10$
40.0	39.5	1.0127	Intercept	-0.719189		± 3
19.9	19.2	1.0390				



H2S Calibration Plot

Date: September 11, 2025

Location: Leismer



0



Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Leismer
Calibration Date: September 29, 2025
Start time (MST): 11:46
Reason: Maintenance

Station number: AMS 501
Last Cal Date: September 11, 2025
End time (MST): 16:17

Calibration Standards

Cal Gas Concentration: 4.89 ppm
Cal Gas Cylinder #: CC737971
Removed Cal Gas Conc: 4.89 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne 750
ZAG Make/Model: Teledyne 751H

Cal Gas Exp Date: September 5, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 282
Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1180540020
Converter serial #: 2022-218
Converter Temp: 325.0 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.014452	1.007306	Backgd or Offset:	3.78	3.71
Calibration intercept:	-0.719189	-0.319160	Coeff or Slope:	1.200	1.177

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4918	81.8	80.0	81.3	0.980
As found Mid point	4959	40.9	40.0	40.4	0.983
As found Low point	4980	20.4	19.9	19.9	0.988
New cylinder response					
Baseline Corr As found:	81.6	Prev response:	80.44	*% change:	1.4%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.020446	AF Intercept:	-0.378939
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999996	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4918	81.8	80.0	80.4	0.995
Mid point	4959	40.9	40.0	39.7	1.008
Low point	4980	20.4	19.9	19.8	1.008
As left zero	5000	0.0	0.0	-0.1	----
As left span	4918	81.8	80.0	80.3	0.996
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.003
Date of last converter efficiency test:					

Notes:

Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	September 29, 2025	Previous Calibration:	September 11, 2025
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	11:46	End Time (MST):	16:17
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999970	≥ 0.995
80.0	80.4	0.9951	Slope	1.007306	$0.90 - 1.10$
40.0	39.7	1.0076	Intercept	-0.319160	± 3
19.9	19.8	1.0076			



H2S Calibration Plot

Date: September 29, 2025

Location: Leismer



#

0



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Leismer
Station number: AMS 501
Calibration Date: September 12, 2025
Last Cal Date: NA
Start time (MST): 9:19
End time (MST): 13:06
Reason: Install

Calibration Standards

NO Gas Cylinder #: DT0022706
NOX Cal Gas Conc: 60.20 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.20 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API 701
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.10 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.10 ppm
NO gas Diff:
Serial Number: 3252
Serial Number: 4427

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = NA	
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>			*Percent Change	NO = NA	
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :		NO SI:	NO Int:	
					As found	NO ₂ r ² :		NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12400232071

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	NA	0.982
NOX coeff or slope:	NA	0.986
NO2 coeff or slope:	NA	1.000

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	NA	0.4
NOX bkgnd or offset:	NA	0.6
Reaction cell Press:	NA	185.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	NA	1.000452
NO _x Cal Offset:	NA	0.312513
NO Cal Slope:	NA	1.001420
NO Cal Offset:	NA	-1.068581
NO ₂ Cal Slope:	NA	0.992751
NO ₂ Cal Offset:	NA	2.025430

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4933	66.6	801.9	800.6	1.3	802.6	801.3	1.3	0.9992	0.9991
Mid point	4967	33.3	400.9	400.2	0.7	401.1	398.9	2.2	0.9995	1.0034
Low point	4983	16.6	199.9	199.5	0.3	200.9	197.9	3.1	0.9949	1.0083
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
As left span	4933	66.6	801.9	379.6	422.3	795.2	379.6	415.6	1.0085	1.0000
Average Correction Factor									0.9979	1.0036

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	795.6	367.8	429.1	426.8	1.0055	99.5%
Mid GPT point	795.6	599.1	197.8	200.5	0.9867	101.3%
Low GPT point	795.6	693.0	103.9	106.4	0.9768	102.4%
Average Correction Factor					0.9897	101.1%

Notes:

Install Calibration. Adjusted zero and span.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	September 12, 2025	Previous Calibration:	NA
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	9:19	End Time (MST):	13:06
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999998	≥0.995
801.9	802.6	0.9992	Slope	1.000452	0.90 - 1.10
400.9	401.1	0.9995	Intercept	0.312513	+/-20
199.9	200.9	0.9949			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	September 12, 2025	Previous Calibration:	NA
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	9:19	End Time (MST):	13:06
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999885	≥ 0.995
429.1	426.8	1.0055	Slope	0.992751	$0.90 - 1.10$
197.8	200.5	0.9867	Intercept	2.025430	± 20
103.9	106.4	0.9768			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	September 12, 2025	Previous Calibration:	NA
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	9:19	End Time (MST):	13:06
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999991	≥ 0.995
800.6	801.3	0.9991	Slope	1.001420	$0.90 - 1.10$
400.2	398.9	1.0034	Intercept	-1.068581	± 20
199.5	197.9	1.0083			



NO_x Calibration Plot

Date: September 12, 2025

Location: Leismer





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Station Information

Station Name:	Leismer	Station Number:	AMS 501
Calibration Date:	September 11, 2025	Prev Cal Date:	
Start Time (MST):	9:03	End Time (MST):	10:09
Tower Height (m):	10.0	Reason:	Install

Wind Speed Calibration

Sensor make/model:	Met One 010C-1	Serial Number:	Y18362
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM (Hz)	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.2	0.2%
400	39.4	39.4	0.1%
600	58.6	58.6	0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	0.999999	1.000000	≥ 0.9995
Calculated slope	0.998868	0.999465	$0.98 - 1.02$
Calculated intercept	0.028930	-0.013446	± 2

Wind Direction Calibration

Sensor make/model:	Met One 020C-1	Serial Number:	R14654
As Found Declination (deg east of True North):	<u>NA</u>	As Left Declination (deg east of True North):	<u>13</u>
Solar noon (MST):	13:22	Calc Declination*:	13.08 Degrees
WD Calibrator:	Met One 040	<i>* - calculated declination as per NOAA website</i>	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 360° FS) <i>Limit = +/- 1%</i>
10	12.2	0.6%
90	90.7	0.2%
180	178.8	-0.3%
270	268.9	-0.3%
350	348.9	-0.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	0.999926	0.999986	≥ 0.9995
Calculated slope	1.000769	1.009838	$0.97 - 1.03$
Calculated intercept	-1.799252	-1.669816	± 5

Notes: Install Calibrations.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Leismer Station number: AMS 501
Calibration Date: October 22, 2025 Last Cal Date: September 11, 2025
Start time (MST): 8:34 End time (MST): 11:29
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.52 ppm Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #: CC274266
Removed Cal Gas Conc: 50.52 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 2659
Zero Air Gen Model: Teledyne API T701 Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1160290011
Analyzer Range: 0-1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000229	1.007371	Backgd or Offset:	19.8	21.4
Calibration intercept:	0.064055	-0.136096	Coeff or Slope:	0.971	0.994

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.7	----
As found High point	4921	79.2	800.2	784.4	1.021
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	783.7	Previous response	800.5	*% change	-2.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4921	79.2	800.2	806.2	0.993
Mid point	4960	39.6	400.2	402.2	0.995
Low point	4980	19.8	200.1	202.0	0.990
As left zero	5000	0.0	0.0	-0.1	----
As left span	4921	79.2	800.2	807.2	0.991
Average Correction Factor:					0.993

Notes:

Adjusted zero and span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 22, 2025	Previous Calibration:	September 11, 2025
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	8:34	End Time (MST):	11:29
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999997	≥0.995
800.2	806.2	0.9926	Slope	1.007371	0.90 - 1.10
400.2	402.2	0.9949	Intercept	-0.136096	+/-30
200.1	202.0	0.9904			



SO2 Calibration Plot

Date: October 22, 2025

Location: Leismer





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Leismer
Calibration Date: October 9, 2025
Start time (MST): 8:33
Reason: Routine

Station number: AMS 501
Last Cal Date: September 29, 2025
End time (MST): 12:16

Calibration Standards

Cal Gas Concentration: 4.89 ppm
Cal Gas Cylinder #: CC737971
Removed Cal Gas Conc: 4.89 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne 750
ZAG Make/Model: Teledyne 751H

Cal Gas Exp Date: September 5, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 282
Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1180540020
Converter serial #: 2022-218
Converter Temp: 325.0 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007306	1.008591	Backgd or Offset:	3.71	3.71
Calibration intercept:	-0.319160	-0.239137	Coeff or Slope:	1.177	1.177

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4918	81.8	80.0	80.5	0.990
As found Mid point	4959	40.9	40.0	39.7	1.000
As found Low point	4980	20.4	19.9	19.7	0.997
New cylinder response					
Baseline Corr As found:	80.8	Prev response:	80.27	*% change:	0.7%
Baseline Corr 2nd AF pt:	40.0	AF Slope:	1.010020	AF Intercept:	-0.439122
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999970	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4918	81.8	80.0	80.6	0.993
Mid point	4959	40.9	40.0	39.8	1.005
Low point	4980	20.4	19.9	19.9	1.002
As left zero	5000	0.0	0.0	-0.1	----
As left span	4918	81.8	80.0	80.9	0.989
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.000
Date of last converter efficiency test:					

Notes:

No adjustment.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	October 9, 2025	Previous Calibration:	September 29, 2025
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	8:33	End Time (MST):	12:16
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999962		≥ 0.995
80.0	80.6	0.9926	Slope	1.008591		$0.90 - 1.10$
40.0	39.8	1.0051	Intercept	-0.239137		± 3
19.9	19.9	1.0025				



H2S Calibration Plot

Date: October 9, 2025

Location: Leismer





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Leismer
Station number: AMS 501
Calibration Date: October 15, 2025
Last Cal Date: September 12, 2025
Start time (MST): 8:21
End time (MST): 12:38
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0022706
NOX Cal Gas Conc: 60.20 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.20 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API 701
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.10 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.10 ppm
NO gas Diff:
Serial Number: 3252
Serial Number: 4427

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.4	0.3	0.1	----	----
AF High point	4933	66.6	801.9	800.6	1.3	802.8	801.1	1.7	0.9994	0.9997
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x =	802.6 ppb	NO =	800.7 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x =	0.0%
Baseline Corr 1st pt	NO _x =	802.4 ppb	NO =	800.8 ppb	<u>As Found Statistics</u>			*Percent Change	NO =	0.0%
Baseline Corr 2nd pt	NO _x =	NA ppb	NO =	NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:		
Baseline Corr 3rd pt	NO _x =	NA ppb	NO =	NA ppb	As found	NO r ² :	NO SI:	NO Int:		
					As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:		

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12400232071

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.982	0.975	NO bkgnd or offset:	0.4	0.4
NOX coeff or slope:	0.986	0.991	NOX bkgnd or offset:	0.6	0.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	185.8	181.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000452	0.998058
NO _x Cal Offset:	0.312513	1.452094
NO Cal Slope:	1.001420	0.998965
NO Cal Offset:	-1.068581	-0.409102
NO ₂ Cal Slope:	0.992751	0.990038
NO ₂ Cal Offset:	2.025430	1.083301

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.6	0.4	0.1	----	----
High point	4933	66.6	801.9	800.6	1.3	801.4	799.8	1.6	1.0007	1.0010
Mid point	4967	33.3	400.9	400.2	0.7	402.0	398.9	3.1	0.9973	1.0034
Low point	4983	16.6	199.9	199.5	0.3	201.8	198.2	3.5	0.9905	1.0068
As left zero	5000	0.0	0.0	0.0	0.0	0.5	0.4	0.0	----	----
As left span	4933	66.6	801.9	384.3	417.6	792.9	384.3	408.6	1.0114	1.0000
Average Correction Factor									0.9961	1.0037

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	795.8	380.6	416.5	413.4	1.0076	99.2%
Mid GPT point	795.8	579.6	217.5	215.6	1.0090	99.1%
Low GPT point	795.8	685.4	111.7	113.6	0.9836	101.7%
Average Correction Factor					1.0000	100.0%

Notes:

Adjusted span only.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 12, 2025
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	8:21	End Time (MST):	12:38
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999995	≥0.995
801.9	801.4	1.0007	Slope	0.998058	0.90 - 1.10
400.9	402.0	0.9973	Intercept	1.452094	+/-20
199.9	201.8	0.9905			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 12, 2025
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	8:21	End Time (MST):	12:38
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999943	≥0.995
416.5	413.4	1.0076	Slope	0.990038	0.90 - 1.10
217.5	215.6	1.0090	Intercept	1.083301	+/-20
111.7	113.6	0.9836			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 12, 2025
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	8:21	End Time (MST):	12:38
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999995	≥ 0.995
800.6	799.8	1.0010	Slope	0.998965	$0.90 - 1.10$
400.2	398.9	1.0034	Intercept	-0.409102	± 20
199.5	198.2	1.0068			



NO_x Calibration Plot

Date: October 15, 2025

Location: Leismer





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS505 SAWBONES BAY OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Sawbones Bay Station number: AMS 505
Calibration Date: October 29, 2025 Last Cal Date: September 25, 2025
Start time (MST): 9:43 End time (MST): 12:34
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.05 ppm Cal Gas Exp Date: April 9, 2033
Cal Gas Cylinder #: EB0063977
Removed Cal Gas Conc: 50.05 ppm Rem Gas Exp Date:
Removed Gas Cyl #: EB0063977 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 5112
Zero Air Gen Model: Teledyne API T701 Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 710321323
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000774	1.003636	Backgd or Offset:	20.7	20.7
Calibration intercept:	-0.954928	-1.255492	Coeff or Slope:	1.111	1.111

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.7	----
As found High point	4920	79.8	798.8	802.6	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	801.9	Previous response	798.5	*% change	0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4920	79.8	798.8	800.9	0.997
Mid point	4960	39.9	399.4	399.4	1.000
Low point	4980	20.0	200.2	198.4	1.009
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	79.8	798.8	801.7	0.996
Average Correction Factor:					1.002

Notes: Changed inlet filter after as founds. Adjusted zero only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 29, 2025	Previous Calibration:	September 25, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:43	End Time (MST):	12:34
Analyzer make:	Thermo 43i	Analyzer serial #:	710321323

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999992	≥0.995
798.8	800.9	0.9974	Slope	1.003636	0.90 - 1.10
399.4	399.4	1.0000	Intercept	-1.255492	+/-30
200.2	198.4	1.0091			



SO2 Calibration Plot

Date: October 29, 2025

Location: Sawbones Bay





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Sawbones Bay Station number: AMS 505
Calibration Date: October 28, 2025 Last Cal Date: September 19, 2025
Start time (MST): 9:01 End time (MST): 13:10
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.26 ppm Cal Gas Exp Date: March 19, 2027
Cal Gas Cylinder #: DT0034141
Removed Cal Gas Conc: 5.26 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T750 Serial Number: 282
ZAG Make/Model: Teledyne API T751H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 12113311965
Converter make: Global 150 Converter serial #: 2022-224
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992596	1.002887	Backgd or Offset:	0.920	0.920
Calibration intercept:	0.080000	0.120000	Coeff or Slope:	1.105	1.105

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4924	76.0	80.0	80.6	0.993
As found Mid point	4962	38.0	40.0	40.5	0.990
As found Low point	4981	19.0	20.0	20.0	1.004
New cylinder response					
Baseline Corr As found:	80.5	Prev response:	79.44	*% change:	1.3%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.008033	AF Intercept:	0.040000
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999981	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4924	76.0	80.0	80.3	0.996
Mid point	4962	38.0	40.0	40.3	0.992
Low point	4981	19.0	20.0	20.0	0.999
As left zero	5000	0.0	0.0	0.3	----
As left span	4924	76.0	80.0	80.2	0.997
SO2 Scrubber Check	4922	77.8	778.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.996
Date of last converter efficiency test:					

Notes: Changed inlet filter after as founds. Scrubber test was done after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	October 28, 2025	Previous Calibration:	September 19, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:01	End Time (MST):	13:10
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311965

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999988		≥0.995
80.0	80.3	0.9957	Slope	1.002887		0.90 - 1.10
40.0	40.3	0.9920	Intercept	0.120000		+/-3
20.0	20.0	0.9994				



H2S Calibration Plot

Date: October 28, 2025

Location: Sawbones Bay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Sawbones Bay
Station number: AMS 505
Calibration Date: October 23, 2025
Last Cal Date: September 18, 2025
Start time (MST): 8:24
End time (MST): 12:28
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0009786
NOX Cal Gas Conc: 60.10 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.10 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 5112
Serial Number: 690

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.5	-0.4	-0.1	----	----
AF High point	4933	66.7	801.8	800.4	1.3	798.9	795.2	3.7	1.0030	1.0061
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 802.8 ppb	NO = 800.3 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -0.4%	
Baseline Corr 1st pt	NO _x = 799.4 ppb	NO = 795.6 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -0.6%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 4259

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.949	0.949	NO bkgnd or offset:	0.9	0.9
NOX coeff or slope:	0.949	0.949	NOX bkgnd or offset:	1.4	1.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	3.7	3.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003271	0.999337
NO _x Cal Offset:	-1.589599	-1.510057
NO Cal Slope:	1.002516	0.998333
NO Cal Offset:	-2.189877	-2.250374
NO ₂ Cal Slope:	1.000956	1.002647
NO ₂ Cal Offset:	-0.173633	0.583578

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.3	-0.1	----	----
High point	4933	66.7	801.8	800.4	1.3	800.2	797.8	2.5	1.0020	1.0033
Mid point	4967	33.3	400.2	399.6	0.7	398.3	395.8	2.5	1.0049	1.0095
Low point	4983	16.7	200.7	200.4	0.3	197.6	195.8	1.8	1.0159	1.0236
As left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
As left span	4933	66.7	801.8	389.2	412.6	796.3	389.2	407.1	1.0069	1.0000
Average Correction Factor									1.0076	1.0121

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	794.7	388.2	407.8	409.1	0.9969	100.3%
Mid GPT point	794.7	599.2	196.8	198.5	0.9916	100.8%
Low GPT point	794.7	696.8	99.2	100.6	0.9864	101.4%
Average Correction Factor					0.9916	100.8%

Notes:

Changed inlet filter after as founds. No adjustment.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 18, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:24	End Time (MST):	12:28
Analyzer make:	API T200	Analyzer serial #:	4259

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999989	≥ 0.995
801.8	800.2	1.0020	Slope	0.999337	$0.90 - 1.10$
400.2	398.3	1.0049	Intercept	-1.510057	± 20
200.7	197.6	1.0159			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 18, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:24	End Time (MST):	12:28
Analyzer make:	API T200	Analyzer serial #:	4259

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999987	≥ 0.995
407.8	409.1	0.9969	Slope	1.002647	$0.90 - 1.10$
196.8	198.5	0.9916	Intercept	0.583578	± 20
99.2	100.6	0.9864			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 18, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	8:24	End Time (MST):	12:28
Analyzer make:	API T200	Analyzer serial #:	4259

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999973	≥ 0.995
800.4	797.8	1.0033	Slope	0.998333	$0.90 - 1.10$
399.6	395.8	1.0095	Intercept	-2.250374	± 20
200.4	195.8	1.0236			



NO_x Calibration Plot

Date: October 23, 2025

Location: Sawbones Bay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS507
KIRBY SOUTH
OCTOBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Kirby South Station number: AMS 507
Calibration Date: October 23, 2025 Last Cal Date: September 18, 2025
Start time (MST): 11:30 End time (MST): 14:35
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.74 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC255918
Removed Cal Gas Conc: 50.74 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 2445
Zero Air Gen Model: Teledyne API T701H Serial Number: 880

Analyzer Information

Analyzer make: Thermo 43iQ Serial Number: 1182340007
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003149	1.003849	Backgd or Offset:	28.9	29.6
Calibration intercept:	-0.832026	0.048008	Coeff or Slope:	1.091	1.112

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4921	78.8	799.7	788.7	1.014
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	788.7	Previous response	801.4	*% change	-1.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4921	78.8	799.7	803.0	0.996
Mid point	4961	39.4	399.8	401.0	0.997
Low point	4980	19.7	199.9	200.8	0.996
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	78.8	799.7	804.0	0.995
Average Correction Factor:					0.996

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 18, 2025
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	11:30	End Time (MST):	14:35
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999999	≥0.995
799.7	803.0	0.9959	Slope	1.003849	0.90 - 1.10
399.8	401.0	0.9970	Intercept	0.048008	+/-30
199.9	200.8	0.9957			



SO2 Calibration Plot

Date: October 23, 2025

Location: Kirby South





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Kirby South Station number: AMS 507
Calibration Date: October 15, 2025 Last Cal Date: September 18, 2025
Start time (MST): 10:05 End time (MST): 16:20
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.05 ppm Cal Gas Exp Date: November 15, 2026
Cal Gas Cylinder #: DT0019762
Removed Cal Gas Conc: 5.05 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: n/a Diff between cyl:
Calibrator Make/Model: Teledyne API T750 Serial Number: 281
ZAG Make/Model: Teledyne API T751H Serial Number: 530

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1150840012
Converter make: Global Converter serial #: 2022-197
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002672	1.002386	Backgd or Offset:	1.69	1.69
Calibration intercept:	0.060000	0.120000	Coeff or Slope:	1.052	1.052

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4921	79.2	80.0	81.6	0.981
As found Mid point	4960	39.6	40.0	41.2	0.973
As found Low point	4980	19.8	20.0	20.5	0.980
New cylinder response					
Baseline Corr As found:	81.5	Prev response:	80.27	*% change:	1.5%
Baseline Corr 2nd AF pt:	41.1	AF Slope:	1.019245	AF Intercept:	0.180000
Baseline Corr 3rd AF pt:	20.4	AF Correlation:	0.999976	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4921	79.2	80.0	80.3	0.996
Mid point	4960	39.6	40.0	40.2	0.995
Low point	4980	19.8	20.0	20.2	0.990
As left zero	5000	0.0	0.0	0.1	----
As left span	4922	78.4	79.2	81.9	0.967
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:		July 25, 2023		Ave Corr Factor	0.994
Date of last converter efficiency test:		September 18, 2025		102.4% efficiency	

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 18, 2025
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	10:05	End Time (MST):	16:20
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	1.000000		≥ 0.995
80.0	80.3	0.9962	Slope	1.002386		$0.90 - 1.10$
40.0	40.2	0.9949	Intercept	0.120000		± 3
20.0	20.2	0.9900				



H2S Calibration Plot

Date: October 15, 2025

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Kirby South
Calibration Date: October 23, 2025
Start time (MST): 11:30
Reason: Routine

Station number: AMS 507
Last Cal Date: September 18, 2025
End time (MST): 14:35

Calibration Standards

Gas Cert Reference: CC255918
CH4 Cal Gas Conc. 506.4 ppm
C3H8 Cal Gas Conc. 205.0 ppm
Removed Gas Cert:
Removed CH4 Conc. 506.4 ppm
Removed C3H8 Conc. 205.0 ppm
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Expiry Date: October 9, 2032
CH4 Equiv Conc. 1070.2 ppm
Removed Gas Expiry:
CH4 Equiv Conc. 1070.2 ppm
Diff between cyl:
Serial Number: 2445
Serial Number: 880

Analyzer Information

Analyzer make: Thermo 51i-LT
Analyzer Range: 0 - 20 ppm

Analyzer serial #: 1182340005

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001956	0.997247	Background:	1.81	1.78
Calibration intercept:	-0.038431	0.016568	Coefficient:	3.589	3.524

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.07	----
As found High point	4921	78.8	16.87	17.14	0.988
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.07	Previous response	16.86	*% change	1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.05	
High point	4921	78.8	16.87	16.85	1.001
Mid point	4961	39.4	8.43	8.42	1.001
Low point	4980	19.7	4.22	4.18	1.009
As left zero	5000	0.0	0.00	0.02	----
As left span	4921	78.8	16.87	16.88	0.999
Average Correction Factor					1.004

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	October 23, 2025	Previous Calibration:	September 18, 2025
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	11:30	End Time (MST):	14:35
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1182340005

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.05	----	Correlation Coefficient	0.999978	≥ 0.995
16.87	16.85	1.0010	Slope	0.997247	$0.90 - 1.10$
8.43	8.42	1.0014	Intercept	0.016568	± 1.5
4.22	4.18	1.0092			


THC Calibration Curve



THC Calibration Plot

Date: October 23, 2025

Location: Kirby South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Kirby South
Station number: AMS 507
Calibration Date: October 15, 2025
Last Cal Date: September 17, 2025
Start time (MST): 10:05
End time (MST): 16:30
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0019572
NOX Cal Gas Conc: 60.00 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.00 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 59.90 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.90 ppm
NO gas Diff:
Serial Number: 2445
Serial Number: 880

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
AF High point	4933	66.8	801.6	800.3	1.3	835.0	827.0	7.7	0.9599	0.9677
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 800.8 ppb	NO = 798.0 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = 4.1%
Baseline Corr 1st pt	NO _x = 835.1 ppb	NO = 827.0 ppb		*Percent Change	NO = 3.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb			

As Found Statistics

As found	NO _x r ² :	Nx SI:	Nx Int:
As found	NO r ² :	NO SI:	NO Int:
As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1173480006

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001898	1.001884
NO _x Cal Offset:	-2.373614	-0.493586
NO Cal Slope:	1.001214	0.999158
NO Cal Offset:	-3.273654	-0.953623
NO ₂ Cal Slope:	0.994238	0.982296
NO ₂ Cal Offset:	1.340861	0.418066

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.745	0.723	NO bkgnd or offset:	7.6	7.4
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	7.8	7.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	145.9	141.7

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.0	----	----
High point	4933	66.8	801.6	800.3	1.3	803.0	799.3	4.1	0.9983	1.0012
Mid point	4967	33.4	400.8	400.1	0.7	400.7	398.1	2.6	1.0002	1.0051
Low point	4983	16.7	200.4	200.1	0.3	199.6	197.9	1.6	1.0040	1.0109
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
As left span	4933	66.8	801.6	410.0	391.6	792.8	410.0	382.8	1.0111	1.0000
Average Correction Factor									1.0009	1.0058

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	795.1	408.7	387.7	380.6	1.0187	98.2%
Mid GPT point	795.1	620.6	175.8	175.1	1.0042	99.6%
Low GPT point	795.1	705.9	90.5	88.5	1.0230	97.8%
Average Correction Factor					1.0153	98.5%

Notes:

Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 17, 2025
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	10:05	End Time (MST):	16:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999997	≥0.995
801.6	803.0	0.9983	Slope	1.001884	0.90 - 1.10
400.8	400.7	1.0002	Intercept	-0.493586	+/-20
200.4	199.6	1.0040			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 17, 2025
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	10:05	End Time (MST):	16:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999935	≥ 0.995
387.7	380.6	1.0187	Slope	0.982296	0.90 - 1.10
175.8	175.1	1.0042	Intercept	0.418066	+/-20
90.5	88.5	1.0230			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	October 15, 2025	Previous Calibration:	September 17, 2025
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	10:05	End Time (MST):	16:30
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999990	≥ 0.995
800.3	799.3	1.0012	Slope	0.999158	0.90 - 1.10
400.1	398.1	1.0051	Intercept	-0.953623	+/-20
200.1	197.9	1.0109			



NO_x Calibration Plot

Date: October 15, 2025

Location: Kirby South





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS511
BLACKGOLD
OCTOBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

November 28, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Blackgold
Calibration Date: September 29, 2025
Start time (MST): 9:33
Reason: Install

Station number: AMS 511
Last Cal Date: N/A
End time (MST): 12:30

Calibration Standards

Cal Gas Concentration: 49.37 ppm
Cal Gas Cylinder #: CC303094
Removed Cal Gas Conc: 49.37 ppm
Removed Gas Cyl #: N/A
Calibrator Model: Teledyne API T750
Zero Air Gen Model: Teledyne API 701

Cal Gas Exp Date: January 5, 2029
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 281
Serial Number: 4865

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 1173410001

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	N/A	1.001459	Backgd or Offset:	N/A	14.4
Calibration intercept:	N/A	0.098845	Coeff or Slope:	N/A	1.157

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4919	81.0	799.8	801.4	0.998
Mid point	4959	40.5	399.9	399.7	1.001
Low point	4980	20.3	200.4	201.3	0.996
As left zero	5000	0.0	0.0	0.6	----
As left span	4919	81.0	799.8		

Average Correction Factor: 0.998

Notes: Used a portable calibrator for the calibration. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	September 29, 2025	Previous Calibration:	N/A
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	9:33	End Time (MST):	12:30
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999997	≥0.995
799.8	801.4	0.9980	Slope	1.001459	0.90 - 1.10
399.9	399.7	1.0006	Intercept	0.098845	+/-30
200.4	201.3	0.9957			



SO2 Calibration Plot

Date: September 29, 2025

Location: Blackgold





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Blackgold Station number: AMS 511
Calibration Date: October 17, 2025 Last Cal Date: September 29, 2025
Start time (MST): 10:47 End time (MST): 12:25
Reason: As Found

Calibration Standards

Cal Gas Concentration: 49.37 ppm Cal Gas Exp Date: January 5, 2029
Cal Gas Cylinder #: CC303094
Removed Cal Gas Conc: 49.37 ppm Rem Gas Exp Date:
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Model: Teledyne API T750 Serial Number: 281
Zero Air Gen Model: Teledyne API 701 Serial Number: 2659

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1173410001
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001459		Backgd or Offset:	14.400	
Calibration intercept:	0.098845		Coeff or Slope:	1.157	

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.4	----
As found High point	4919	81.0	799.8	795.8	1.006
As found Mid point	4959	40.5	399.9	397.6	1.007
As found Low point	4980	20.3	200.4	197.4	1.017
New cylinder response					
Baseline Corr As found:	795.4	Previous response	801.1	*% change	-0.7%
Baseline Corr 2nd AF pt:	397.2	AF Slope:	0.995426	AF Intercept:	-0.639045
Baseline Corr 3rd AF pt:	197.0	AF Correlation:	0.999990	* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor:


Notes: As founds only.

Calibration Performed By: Mohammed Kashif

SO2 Calibration Plot

Date: October 17, 2025

Location: Blackgold





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Blackgold Station number: AMS 511
Calibration Date: October 17, 2025 Last Cal Date: September 29, 2025
Start time (MST): 14:15 End time (MST): 16:41
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.37 ppm Cal Gas Exp Date: January 5, 2029
Cal Gas Cylinder #: CC303094
Removed Cal Gas Conc: 49.37 ppm Rem Gas Exp Date:
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Model: Teledyne API T750 Serial Number: 953
Zero Air Gen Model: Teledyne API 701 Serial Number: 2659

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1173410001
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001459	1.006702	Backgd or Offset:	14.4	14.7
Calibration intercept:	0.098845	-0.861427	Coeff or Slope:	1.157	1.163

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.6	----
High point	4919	81.0	799.8	805.2	0.993
Mid point	4959	40.5	399.9	400.5	0.999
Low point	4980	20.3	200.4	199.8	1.003
As left zero	5000	0.0	0.0	0.7	----
As left span	4919	81.0	799.8	807.9	0.990
Average Correction Factor:					0.998

Notes: Sample inlet filter and ZAG were changed. Adjusted span only

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 29, 2025
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	14:15	End Time (MST):	16:41
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.6	----	Correlation Coefficient	0.999984	≥0.995
799.8	805.2	0.9933	Slope	1.006702	0.90 - 1.10
399.9	400.5	0.9986	Intercept	-0.861427	+/-30
200.4	199.8	1.0032			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: October 17, 2025

Location: Blackgold





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Blackgold
Calibration Date: September 29, 2025
Start time (MST): 9:16
Reason: Install

Station number: AMS 511
Last Cal Date: N/A
End time (MST): 12:09

Calibration Standards

Cal Gas Concentration: 5.139 ppm
Cal Gas Cylinder #: CC511397
Removed Cal Gas Conc: 5.139 ppm
Removed Gas Cyl #: N/A
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date: N/A
Diff between cyl:
Serial Number: 5112
Serial Number: 4865

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1336160090
Converter serial #: 2025-299
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	N/A	1.003489	Backgd or Offset:	N/A	4.03
Calibration intercept:	N/A	-0.219202	Coeff or Slope:	N/A	1.367

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4922	77.8	80.0	80.2	0.997
Mid point	4961	38.9	40.0	39.7	1.007
Low point	4981	19.5	20.0	19.6	1.022
As left zero	5000	0.0	0.0	0.1	----
As left span	4922	77.8	80.0	79.5	1.006
SO2 Scrubber Check				-0.2	
Date of last scrubber change:				Ave Corr Factor	1.009
Date of last converter efficiency test:					

Notes: Install calibration. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	September 29, 2025	Previous Calibration:	N/A
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	9:16	End Time (MST):	12:09
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999927		≥0.995
80.0	80.2	0.9971	Slope	1.003489		0.90 - 1.10
40.0	39.7	1.0071	Intercept	-0.219202		+/-3
20.0	19.6	1.0225				



H₂S Calibration Plot

Date: September 29, 2025

Location: Blackgold





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Blackgold
Calibration Date: October 30, 2025
Start time (MST): 10:18
Reason: Routine

Station number: AMS 511
Last Cal Date: September 29, 2025
End time (MST): 15:09

Calibration Standards

Cal Gas Concentration: 5.139 ppm
Cal Gas Cylinder #: CC511397
Removed Cal Gas Conc: 5.139 ppm
Removed Gas Cyl #: N/A
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date: N/A
Diff between cyl:
Serial Number: 953
Serial Number: 2659

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1336160090
Converter serial #: 2025-299
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003489	0.987624	Backgd or Offset:	4.03	4.01
Calibration intercept:	-0.219202	0.061038	Coeff or Slope:	1.367	1.367

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4922	77.8	80.0	79.5	1.007
As found Mid point	4961	38.9	40.0	39.5	1.015
As found Low point	4981	19.5	20.0	19.9	1.012
New cylinder response					
Baseline Corr As found:	79.4	Prev response:	80.03	*% change:	-0.8%
Baseline Corr 2nd AF pt:	39.4	AF Slope:	0.992915	AF Intercept:	0.000860
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999983	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4922	77.8	80.0	79.1	1.011
Mid point	4961	38.9	40.0	39.5	1.012
Low point	4981	19.5	20.0	19.7	1.017
As left zero	5000	0.0	0.0	0.3	----
As left span	4922	77.8	80.0	78.0	1.025
SO2 Scrubber Check	4919	81.0	810.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.013
Date of last converter efficiency test:					

Notes: Changed sample inlet filter after as founds. Ran scrubber check after calibrator zero and it passed.
No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	October 30, 2025	Previous Calibration:	September 29, 2025
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	10:18	End Time (MST):	15:09
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999986	≥ 0.995
80.0	79.1	1.0109	Slope	0.987624	$0.90 - 1.10$
40.0	39.5	1.0122	Intercept	0.061038	± 3
20.0	19.7	1.0173			


H₂S Calibration Curve



H₂S Calibration Plot

Date: October 30, 2025

Location: Blackgold





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Blackgold
Calibration Date: September 29, 2025
Start time (MST): 9:33
Reason: Install
Station number: AMS 511
Last Cal Date: N/A
End time (MST): 12:30

Calibration Standards

Gas Cert Reference: CC303094
CH4 Cal Gas Conc. 499.2 ppm
C3H8 Cal Gas Conc. 202.7 ppm
Removed Gas Cert: N/A
Removed CH4 Conc. 499.2 ppm
Removed C3H8 Conc. 202.7 ppm
Calibrator Make/Model: Teledyne API T750
ZAG Make/Model: Teledyne API 701
Cal Gas Expiry Date: January 5, 2029
CH4 Equiv Conc. 1056.6 ppm
Removed Gas Expiry: N/A
CH4 Equiv Conc. 1056.6 ppm
Diff between cyl:
Serial Number: 281
Serial Number: 4865

Analyzer Information

Analyzer make: Thermo 51i
Analyzer Range: 0 - 20 ppm
Analyzer serial #: 12426335705

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	N/A	1.002035	Background:	N/A	0.870
Calibration intercept:	N/A	0.022887	Coefficient:	N/A	2.150

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/Ic-AFzero) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4919	81.0	17.12	17.16	0.997
Mid point	4959	40.5	8.56	8.62	0.993
Low point	4980	20.3	4.29	4.34	0.989
As left zero	5000	0.0	0.00	0.00	----
As left span	4919	81.0	17.12		

Average Correction Factor 0.993

Notes: Used a portable calibrator for the calibration. Adjusted zero and span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	September 29, 2025	Previous Calibration:	N/A
Station Name:	2025-09-29	Station Number:	N/A
Start Time (MST):	9:33	End Time (MST):	12:30
Analyzer make:	Thermo 51i	Analyzer serial #:	12426335705

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999991	≥ 0.995
17.12	17.16	0.9975	Slope	1.002035	$0.90 - 1.10$
8.56	8.62	0.9931	Intercept	0.022887	± 1.5
4.29	4.34	0.9886			


THC Calibration Curve



THC Calibration Plot

Date: September 29, 2025

Location: Blackgold





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Blackgold
Calibration Gas Date: October 17, 2025
Start time (MST): 10:47
Reason: As Found

Station number: AMS 511
Last Cal Date: September 29, 2025
End time (MST): 12:25

Calibration Standards

Gas Cert Reference: CC303094
CH4 Cal Gas Conc. 499.2 ppm
C3H8 Cal Gas Conc. 202.7 ppm
Removed Gas Cert: N/A
Removed CH4 Conc. 499.2 ppm
Removed C3H8 Conc. 202.7 ppm
Calibrator Make/Model: Teledyne API T750
ZAG Make/Model: Teledyne API 701

Cal Gas Expiry Date: January 5, 2029
CH4 Equiv Conc. 1056.6 ppm
Removed Gas Expiry: N/A
CH4 Equiv Conc. 1056.6 ppm
Diff between cyl:
Serial Number: 281
Serial Number: 2659

Analyzer Information

Analyzer make: Thermo 51i
Analyzer Range: 0 - 20 ppm

Analyzer serial #: 12426335705

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002035		Background:	0.870000	
Calibration intercept:	0.022887		Coefficient:	2.150000	

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	-0.06	----
As found High point	4919	81.0	17.12	16.98	1.004
As found Mid point	4959	40.5	8.56	8.45	1.005
As found Low point	4980	20.3	4.29	4.18	1.011
New cylinder response					
Baseline Corr As found:	17.04	Previous response	17.18	*% change	-0.8%
Baseline Corr 2nd AF pt:	8.51	AF Slope:	0.996182	AF Intercept:	-0.076270
Baseline Corr 3rd AF pt:	4.24	AF Correlation:	0.999997	* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Notes:


As founds only.

Calibration Performed By: Mohammed Kashif

THC Calibration Plot

Date: October 17, 2025

Location: Blackgold





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Blackgold
Calibration Date: October 17, 2025
Start time (MST): 14:15
Reason: Routine

Station number: AMS 511
Last Cal Date: September 29, 2025
End time (MST): 16:41

Calibration Standards

Gas Cert Reference: CC303094
CH4 Cal Gas Conc. 499.2 ppm
C3H8 Cal Gas Conc. 202.7 ppm
Removed Gas Cert: N/A
Removed CH4 Conc. 499.2 ppm
Removed C3H8 Conc. 202.7 ppm
Calibrator Make/Model: Teledyne API T750
ZAG Make/Model: Teledyne API 701

Cal Gas Expiry Date: January 5, 2029
CH4 Equiv Conc. 1056.6 ppm
Removed Gas Expiry: N/A
CH4 Equiv Conc. 1056.6 ppm
Diff between cyl:
Serial Number: 953
Serial Number: 2659

Analyzer Information

Analyzer make: Thermo 51i
Analyzer Range: 0 - 20 ppm

Analyzer serial #: 12426335705

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002035	1.006133	Background:	0.87
Calibration intercept:	0.022887	-0.074314	Coefficient:	2.150
				0.96
				2.390

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/Ic-AFzero)
As found zero					Limit = 0.90-1.10
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	-0.09	----
High point	4919	81.0	17.12	17.13	0.999
Mid point	4959	40.5	8.56	8.59	0.997
Low point	4980	20.3	4.29	4.22	1.016
As left zero	5000	0.0	0.00	-0.11	----
As left span	4919	81.0	17.12	17.32	0.989
Average Correction Factor					1.004

Notes: Sample inlet filter and ZAG were changed. The sample pressure was adjusted to approximately 18 psi, which likely caused the observed background shift. Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	October 17, 2025	Previous Calibration:	September 29, 2025
Station Name:	2025-09-29	Station Number:	N/A
Start Time (MST):	14:15	End Time (MST):	16:41
Analyzer make:	Thermo 51i	Analyzer serial #:	12426335705

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.09	----	Correlation Coefficient	0.999978	≥ 0.995
17.12	17.13	0.9994	Slope	1.006133	$0.90 - 1.10$
8.56	8.59	0.9966	Intercept	-0.074314	± 1.5
4.29	4.22	1.0158			

THC Calibration Curve



THC Calibration Plot

Date: October 17, 2025

Location: Blackgold





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Blackgold
Station number: AMS 511
Calibration Date: September 26, 2025
Last Cal Date: N/A
Start time (MST): 9:00
End time (MST): 13:30
Reason: Install

Calibration Standards

NO Gas Cylinder #: T0F8P52
NOX Cal Gas Conc: 47.43 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 47.43 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: August 16, 2026
NO Cal Gas Conc: 47.43 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 47.43 ppm
NO gas Diff:
Serial Number: 5112
Serial Number: 4865

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x =	NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>			*Percent Change	NO =	NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :		NO SI:	NO Int:	
					As found	NO ₂ r ² :		NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 7029

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	N/A	1.056	NO bkgnd or offset:	N/A	0.2
NOX coeff or slope:	N/A	1.052	NOX bkgnd or offset:	N/A	0.4
NO2 coeff or slope:	N/A	1.000	Reaction cell Press:	N/A	4.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	N/A	1.003246
NO _x Cal Offset:	N/A	-4.057433
NO Cal Slope:	N/A	1.005046
NO Cal Offset:	N/A	-4.537452
NO ₂ Cal Slope:	N/A	0.986228
NO ₂ Cal Offset:	N/A	-1.042553

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
High point	4916	84.3	799.6	799.6	0.0	800.7	802.0	-1.1	0.9987	0.9970
Mid point	4958	42.2	400.3	400.3	0.0	393.9	393.5	0.4	1.0163	1.0173
Low point	4979	21.1	200.2	200.2	0.0	193.9	193.7	0.2	1.0323	1.0333
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
As left span	4916	84.3	0.0	423.0	376.6	783.8	423.0	360.8	0.0000	1.0000
Average Correction Factor									1.0157	1.0159

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	797.0	410.0	387.0	381.2	1.0152	98.5%
Mid GPT point	797.0	598.2	198.8	194.5	1.0221	97.8%
Low GPT point	797.0	698.1	98.9	95.3	1.0378	96.4%
Average Correction Factor					1.0250	97.6%

Notes: Install calibration. Adjusted the span.

Calibration Performed By: Max Farrell



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	September 26, 2025	Previous Calibration:	N/A
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	9:00	End Time (MST):	13:30
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999876	≥0.995
799.6	800.7	0.9987	Slope	1.003246	0.90 - 1.10
400.3	393.9	1.0163	Intercept	-4.057433	+/-20
200.2	193.9	1.0323			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	September 26, 2025	Previous Calibration:	N/A
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	9:00	End Time (MST):	13:30
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999958	≥ 0.995
387.0	381.2	1.0152	Slope	0.986228	0.90 - 1.10
198.8	194.5	1.0221	Intercept	-1.042553	+/-20
98.9	95.3	1.0378			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	September 26, 2025	Previous Calibration:	N/A
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	9:00	End Time (MST):	13:30
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999847	≥ 0.995
799.6	802.0	0.9970	Slope	1.005046	$0.90 - 1.10$
400.3	393.5	1.0173	Intercept	-4.537452	± 20
200.2	193.7	1.0333			



NO_x Calibration Plot

Date: September 26, 2025

Location: Blackgold





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Blackgold
Station number: AMS 511
Calibration Date: October 7, 2025
Last Cal Date: September 26, 2025
Start time (MST): 10:47
End time (MST): 15:52
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T0F8P52
NOX Cal Gas Conc: 47.43 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 47.43 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: August 16, 2026
NO Cal Gas Conc: 47.43 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 47.43 ppm
NO gas Diff:
Serial Number: 5112
Serial Number: 4865

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.1	----	----
AF High point	4916	84.3	799.6	799.6	0.0	765.8	767.1	-1.4	1.0439	1.0421
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 798.2 ppb	NO = 799.1 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = -4.2%			
Baseline Corr 1st pt	NO _x = 766.0 ppb	NO = 767.3 ppb	<u>As Found Statistics</u>			*Percent Change	NO = -4.1%			
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:	NO Int:			
			As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 7029

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.056	1.082	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.052	1.078	NOX bkgnd or offset:	0.4	0.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.8	4.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.003246	1.002533
NO _x Cal Offset:	-4.057433	-2.958117
NO Cal Slope:	1.005046	1.003662
NO Cal Offset:	-4.537452	-3.578225
NO ₂ Cal Slope:	0.986228	0.993905
NO ₂ Cal Offset:	-1.042553	-0.629244

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	0.0	----	----
High point	4916	84.3	799.6	799.6	0.0	800.4	800.8	-0.5	0.9990	0.9985
Mid point	4958	42.2	400.3	400.3	0.0	396.2	396.4	-0.1	1.0104	1.0099
Low point	4979	21.1	200.2	200.2	0.0	195.2	193.6	1.6	1.0254	1.0339
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
As left span	4916	84.3	799.6	389.3	410.3	779.9	389.3	390.6	1.0253	1.0000
Average Correction Factor									1.0116	1.0141

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	793.3	392.5	400.8	398.0	1.0070	99.3%
Mid GPT point	793.3	595.3	198.0	196.0	1.0102	99.0%
Low GPT point	793.3	689.4	103.9	101.9	1.0196	98.1%
Average Correction Factor					1.0123	98.8%

Notes:

Sample inlet filter changed after as founds. Adjusted span only.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	October 7, 2025	Previous Calibration:	September 26, 2025
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	10:47	End Time (MST):	15:52
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999936	≥0.995
799.6	800.4	0.9990	Slope	1.002533	0.90 - 1.10
400.3	396.2	1.0104	Intercept	-2.958117	+/-20
200.2	195.2	1.0254			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	October 7, 2025	Previous Calibration:	September 26, 2025
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	10:47	End Time (MST):	15:52
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999988	≥ 0.995
400.8	398.0	1.0070	Slope	0.993905	$0.90 - 1.10$
198.0	196.0	1.0102	Intercept	-0.629244	± 20
103.9	101.9	1.0196			





Wood Buffalo Environmental Association

NO Calibration Summary


Station Information

Calibration Date:	October 7, 2025	Previous Calibration:	September 26, 2025
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	10:47	End Time (MST):	15:52
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999904	≥ 0.995
799.6	800.8	0.9985	Slope	1.003662	$0.90 - 1.10$
400.3	396.4	1.0099	Intercept	-3.578225	± 20
200.2	193.6	1.0339			


NO Calibration Curve



NO_x Calibration Plot

Date: October 7, 2025

Location: Blackgold





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Station Information

Station Name:	Blackgold	Station Number:	AMS 511
Calibration Date:	September 26, 2025	Prev Cal Date:	N/A
Start Time (MST):	12:02	End Time (MST):	13:04
Tower Height (m):	10.0	Reason:	Install

Wind Speed Calibration

Sensor make/model:	Met One 010C-1	Serial Number:	B10017
WS Calibrator:	MetOne 053	Serial Number:	P15103

Shaft RPM (Hz)	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.2	0.2%
400	39.4	39.4	0.1%
600	58.6	58.7	0.2%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)		0.999999	≥ 0.9995
Calculated slope		0.998949	$0.98 - 1.02$
Calculated intercept		-0.013169	± 2

Wind Direction Calibration

Sensor make/model:	Met One 020C-1	Serial Number:	D13603
As Found Declination (deg east of True North):	<u>N/A</u>	As Left Declination (deg east of True North):	<u>14</u>
Solar noon (MST):		Calc Declination*:	Degrees
WD Calibrator:	Met One 040		<small>* - calculated declination as per NOAA website</small>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 360° FS) <i>Limit = +/- 1%</i>
10	10.3	0.1%
90	89.4	-0.2%
180	179.2	-0.2%
270	272.4	0.7%
350	352.8	0.8%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)		0.999968	≥ 0.9995
Calculated slope		0.990632	$0.97 - 1.03$
Calculated intercept		0.873846	± 5

Notes: Install calibration. Used compass to align the crossarm. Bearing are in good condition.

Calibration Performed By: Max Farrell



End of Report