



**WOOD BUFFALO
ENVIRONMENTAL ASSOCIATION**

Unit 3 - 805 Memorial Drive
Fort McMurray, AB T9K 0K4
P: 780.799.4420 E: info@wbea.org
wbea.org

Wood Buffalo Environmental Association

NOVEMBER 2025 MONTHLY CALIBRATION REPORT

CONTINUOUS MONITORING

December 22, 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS01 BERTHA GANTER - FORT MCKAY NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: November 3, 2025 Last Cal Date: October 20, 2025
Start time (MST): 10:36 End time (MST): 14:00
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.21 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC418809
Removed Cal Gas Conc: 49.21 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3565
Zero Air Gen Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1501301448
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000408	0.996496	Backgd or Offset:	21.6	21.5
Calibration intercept:	-0.613104	-0.293182	Coeff or Slope:	0.879	0.875

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4918	81.3	800.3	802.2	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	802.4	Previous response	800.0	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4918	81.3	800.3	797.6	1.003
Mid point	4959	40.7	400.6	398.0	1.007
Low point	4979	20.3	199.8	198.9	1.005
As left zero	5000	0.0	0.0	0.0	----
As left span	4918	81.3	800.3	799.8	1.001
Average Correction Factor:					1.005

Notes: Changed the inlet filter after as founds. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 3, 2025	Previous Calibration:	October 20, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:36	End Time (MST):	14:00
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301448

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999997	≥0.995
800.3	797.6	1.0033	Slope	0.996496	0.90 - 1.10
400.6	398.0	1.0065	Intercept	-0.293182	+/-30
199.8	198.9	1.0046			



SO2 Calibration Plot

Date: November 3, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: November 4, 2025 Last Cal Date: October 8, 2025
Start time (MST): 11:12 End time (MST): 15:14
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: September 5, 2027
Cal Gas Cylinder #: CC738239
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 12113311966
Converter make: CD Nova Converter serial #: 580
Analyzer Range: 0 - 100 ppb Converter Temp: 850 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997956	1.003817	Backgd or Offset:	2.13	2.13
Calibration intercept:	-0.297954	-0.178082	Coeff or Slope:	1.119	1.119

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4917	82.6	80.0	80.8	0.988
As found Mid point	4959	41.3	40.0	40.4	0.987
As found Low point	4979	20.7	20.0	20.1	0.992
New cylinder response					
Baseline Corr As found:	80.9	Prev response:	79.50	*% change:	1.7%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	1.012110	AF Intercept:	-0.118267
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999998	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4917	82.6	80.0	80.1	0.998
Mid point	4959	41.3	40.0	40.0	0.999
Low point	4979	20.7	20.0	19.8	1.012
As left zero	5000	0.0	0.0	0.0	----
As left span	4917	82.6	80.0	79.4	1.007
SO2 Scrubber Check	4919	81.3	813.0	0.1	----
Date of last scrubber change:	December 17, 2021			Ave Corr Factor	1.003
Date of last converter efficiency test:					

Notes: Inlet filter change completed after as founds. No adjustments made. Scrubber check passed.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	November 4, 2025	Previous Calibration:	October 8, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:12	End Time (MST):	15:14
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	12113311966

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999992		≥ 0.995
80.0	80.1	0.9983	Slope	1.003817		$0.90 - 1.10$
40.0	40.0	0.9994	Intercept	-0.178082		± 3
20.0	19.8	1.0121				


TRS Calibration Curve



TRS Calibration Plot

Date: November 4, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: November 4, 2025 Last Cal Date: October 8, 2025
Start time (MST): 11:12 End time (MST): 15:14
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: September 5, 2027
Cal Gas Cylinder #: CC738239
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Thermo 43iQ-TL Analyzer serial #: 1200326167
Converter make: CD Nova Converter serial #: 2022-221
Analyzer Range: 0 - 100 ppb Converter Temp: 315 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005106	1.001244	Backgd or Offset:	2.01	2.01
Calibration intercept:	-0.098195	-0.138038	Coeff or Slope:	0.973	0.973

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4917	82.6	80.0	80.4	0.992
As found Mid point	4959	41.3	40.0	40.1	0.992
As found Low point	4979	20.7	20.0	20.0	0.992
New cylinder response					
Baseline Corr As found:	80.6	Prev response:	80.27	*% change:	0.4%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.007965	AF Intercept:	-0.198226
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	1.000000	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4917	82.6	80.0	79.9	1.001
Mid point	4959	41.3	40.0	40.0	0.999
Low point	4979	20.7	20.0	19.8	1.012
As left zero	5000	0.0	0.0	0.0	----
As left span	4917	82.6	80.0	79.7	1.003
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:	January 25, 2024			Ave Corr Factor	1.004
Date of last converter efficiency test:	November 7, 2024			107.9% efficiency	

Notes: Inlet filter change completed after as founds. No adjustments made. Scrubber check passed.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	November 4, 2025	Previous Calibration:	October 8, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:12	End Time (MST):	15:14
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	1200326167

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999991		≥0.995
80.0	79.9	1.0008	Slope	1.001244		0.90 - 1.10
40.0	40.0	0.9994	Intercept	-0.138038		+/-3
20.0	19.8	1.0121				



H₂S Calibration Plot

Date: November 4, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
Calibration Date:	November 3, 2025	Last Cal Date:	October 20, 2025
Start time (MST):	10:36	End time (MST):	14:00
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC418809	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.3 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1061.8 ppm
Removed C ₃ H ₈ Conc.	205.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3565
Zero Air Gen model:	Teledyne API T701	Serial Number:	146

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648
THC Range:	0 - 20 ppm	NMHC/CH ₄ Range:	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.49E-04	2.51E-04	NMHC SP Ratio:	4.95E-05
CH ₄ Retention time:	15.0	15.0	NMHC Peak Area:	185423
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	81.3	17.27	17.17	1.006
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.17	Prev response	17.21	*% change	-0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	81.3	17.27	17.26	1.000
Mid point	4959	40.7	8.64	8.64	1.000
Low point	4979	20.3	4.31	4.33	0.996
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	17.27	17.25	1.001
Average Correction Factor					0.999

Notes: Changed the inlet filter after as founds. Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4918	81.3	9.18	9.14	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.14	Prev response	9.19	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4918	81.3	9.18	9.16	1.002
Mid point	4959	40.7	4.60	4.61	0.997
Low point	4979	20.3	2.29	2.31	0.992
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	9.18	9.16	1.003
Average Correction Factor					0.997

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4918	81.3	8.09	8.03	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.03	Prev response	8.02	*% change	0.1%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4918	81.3	8.09	8.10	0.998
Mid point	4959	40.7	4.05	4.03	1.004
Low point	4979	20.3	2.02	2.02	1.000
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	81.3	8.09	8.09	0.999
Average Correction Factor					1.001

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.995697	0.999410
THC Cal Offset:	0.018346	0.007543
CH ₄ Cal Slope:	0.991109	1.001776
CH ₄ Cal Offset:	0.006464	-0.006527
NMHC Cal Slope:	0.999352	0.997326
NMHC Cal Offset:	0.011683	0.014070

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 3, 2025	Previous Calibration:	October 20, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:36	End Time (MST):	14:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999999	≥ 0.995
17.27	17.26	1.0002	Slope	0.999410	$0.90 - 1.10$
8.64	8.64	1.0002	Intercept	0.007543	± 0.5
4.31	4.33	0.9959			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 3, 2025	Previous Calibration:	October 20, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:36	End Time (MST):	14:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999990		≥ 0.995
8.09	8.10	0.9981	Slope	1.001776		0.90 - 1.10
4.05	4.03	1.0038	Intercept	-0.006527		± 0.5
2.02	2.02	1.0005				





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 3, 2025	Previous Calibration:	October 20, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:36	End Time (MST):	14:00
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585648

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999988	≥ 0.995
9.18	9.16	1.0021	Slope	0.997326	$0.90 - 1.10$
4.60	4.61	0.9969	Intercept	0.014070	± 0.5
2.29	2.31	0.9920			



NMHC Calibration Plot

Date: November 3, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay
Station number: AMS 01
Calibration Date: November 10, 2025
Last Cal Date: October 15, 2025
Start time (MST): 10:55
End time (MST): 15:46
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC335700
NOX Cal Gas Conc: 59.40 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.40 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: September 1, 2032
NO Cal Gas Conc: 59.20 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.20 ppm
NO gas Diff:
Serial Number: 3565
Serial Number: 146

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.4	0.1	0.2	----	----
AF High point	4932	67.6	803.1	800.4	2.7	801.6	798.2	3.5	1.0024	1.0029
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 803.9 ppb	NO = 802.3 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -0.3%	
Baseline Corr 1st pt	NO _x = 801.2 ppb	NO = 798.1 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -0.5%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12400232072

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.819	0.823
NOX coeff or slope:	0.994	0.998
NO2 coeff or slope:	1.000	1.000

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	0.6	0.6
NOX bkgnd or offset:	0.7	0.7
Reaction cell Press:	126.0	124.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999769	0.997834
NO _x Cal Offset:	0.980000	0.060000
NO Cal Slope:	1.002890	0.999178
NO Cal Offset:	-0.380000	-1.580000
NO ₂ Cal Slope:	0.984647	0.966258
NO ₂ Cal Offset:	1.657543	0.376088

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOX concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOX concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.2	0.2	----	----
High point	4932	67.6	803.1	800.4	2.7	801.4	799.0	2.3	1.0021	1.0017
Mid point	4966	33.8	401.5	400.2	1.4	401.0	397.5	3.5	1.0014	1.0068
Low point	4983	16.9	200.8	200.1	0.7	199.9	196.5	3.4	1.0044	1.0183
As left zero										
As left span										

Average Correction Factor

1.0026	1.0089
--------	--------

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	790.1	385.1	407.7	394.0	1.0348	96.6%
Mid GPT point	790.1	581.6	211.2	205.2	1.0293	97.2%
Low GPT point	790.1	688.1	104.7	101.3	1.0336	96.7%
Average Correction Factor					1.0325	96.8%

Notes: Changed the inlet filter after as founds. Adjusted the span. Removing the instrument because of the unstable GPT response.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 10, 2025	Previous Calibration:	October 15, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:55	End Time (MST):	15:46
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232072

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995
803.1	801.4	1.0021	Slope	0.997834	0.90 - 1.10
401.5	401.0	1.0014	Intercept	0.060000	+/-20
200.8	199.9	1.0044			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 10, 2025	Previous Calibration:	October 15, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:55	End Time (MST):	15:46
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232072

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999991	≥0.995
407.7	394.0	1.0348	Slope	0.966258	0.90 - 1.10
211.2	205.2	1.0293	Intercept	0.376088	+/-20
104.7	101.3	1.0336			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 10, 2025	Previous Calibration:	October 15, 2025
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	10:55	End Time (MST):	15:46
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232072

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999977	≥ 0.995
800.4	799.0	1.0017	Slope	0.999178	$0.90 - 1.10$
400.2	397.5	1.0068	Intercept	-1.580000	± 20
200.1	196.5	1.0183			



NO_x Calibration Plot

Date: November 10, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay
Station number: AMS 01
Calibration Date: November 11, 2025
Last Cal Date: NA
Start time (MST): 12:00
End time (MST): 16:02
Reason: Install

Calibration Standards

NO Gas Cylinder #: CC335700
NOX Cal Gas Conc: 59.40 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.40 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: September 1, 2032
NO Cal Gas Conc: 59.20 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.20 ppm
NO gas Diff:
Serial Number: 3565
Serial Number: 146

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation		*Percent Change		NO _x =	NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>		*Percent Change		NO =	NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:		Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:		NO Int:	
					As found	NO ₂ r ² :	NO2 SI:		NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1336160088

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.752		NO bkgnd or offset:	2.2	
NOX coeff or slope:	0.994		NOX bkgnd or offset:	2.3	
NO2 coeff or slope:	0.999		Reaction cell Press:	160.3	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:		0.999343
NO _x Cal Offset:		-0.220000
NO Cal Slope:		1.001148
NO Cal Offset:		-1.620000
NO ₂ Cal Slope:		0.995090
NO ₂ Cal Offset:		-1.054014

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.0	----	----
High point	4932	67.6	803.1	800.4	2.7	802.5	800.5	2.0	1.0007	0.9999
Mid point	4966	33.8	401.5	400.2	1.4	400.8	398.1	2.6	1.0019	1.0053
Low point	4983	16.9	200.8	200.1	0.7	200.3	197.3	3.0	1.0024	1.0142
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	0.2	----	----
As left span	4932	67.6	803.1	372.6	430.5	806.4	372.6	433.6	0.9959	1.0000
Average Correction Factor									1.0016	1.0064

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	797.9	398.8	401.8	399.4	1.0060	99.4%
Mid GPT point	797.9	595.1	205.5	202.7	1.0138	98.6%
Low GPT point	797.9	695.9	104.7	102.2	1.0245	97.6%
Average Correction Factor					1.0148	98.5%

Notes: Installing a new analyzer. Adjusted the zero and span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 11, 2025	Previous Calibration:	NA
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	12:00	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.0	----	Correlation Coefficient	1.000000	≥0.995
803.1	802.5	1.0007	Slope	0.999343	0.90 - 1.10
401.5	400.8	1.0019	Intercept	-0.220000	+/-20
200.8	200.3	1.0024			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 11, 2025	Previous Calibration:	NA
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	12:00	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999966	<i>≥0.995</i>
401.8	399.4	1.0060	Slope	0.995090	<i>0.90 - 1.10</i>
205.5	202.7	1.0138	Intercept	-1.054014	<i>+/-20</i>
104.7	102.2	1.0245			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 11, 2025	Previous Calibration:	NA
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	12:00	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1336160088

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999983	≥ 0.995
800.4	800.5	0.9999	Slope	1.001148	$0.90 - 1.10$
400.2	398.1	1.0053	Intercept	-1.620000	± 20
200.1	197.3	1.0142			



NO_x Calibration Plot

Date: November 11, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: November 17, 2025 Last Cal Date: October 6, 2025
Start time (MST): 11:40 End time (MST): 13:07
Reason: Removal

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 1107
Analyzer Range: 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001714		Backgd or Offset:	5.6	
Calibration intercept:	-0.800000		Coeff or Slope:	1.006	

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-2.6	----
As found High point	5000	863.1	400.0	396.8	1.002
As found Mid point	5000	744.0	200.0	196.2	1.006
As found Low point	5000	651.7	100.0	97.0	1.004
Baseline Corr As found:	399.4	Previous response	399.9	*% change	-0.1%
Baseline Corr 2nd AF pt:	198.8	AF Slope:	0.998457	AF Intercept:	-2.880000
Baseline Corr 3rd AF pt:	99.6	AF Correlation:	0.999994	* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor


Notes: Removing the instrument. Multi-point as founds completed.

Calibration Performed By: Rene Chamberland

O₃ Calibration Plot

Date: November 17, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Bertha Ganter-Fort McKay Station number: AMS 01
Calibration Date: November 17, 2025 Last Cal Date: NA
Start time (MST): 13:30 End time (MST): 15:48
Reason: Install

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700 Serial Number: 3565
ZAG Make/Model: Teledyne API T701 Serial Number: 146

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 7045
Analyzer Range: 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		1.000514	Backgd or Offset:		1.4
Calibration intercept:		0.560000	Coeff or Slope:		1.039

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	5000	863.1	400.0	400.4	0.999
Mid point	5000	744.0	200.0	201.4	0.993
Low point	5000	651.7	100.0	100.6	0.994
As left zero	5000	0.0	0.0	1.1	----
As left span	5000	863.1	400.0	404.1	0.990
Average Correction Factor					0.995

Notes: Installing a new instrument. Changed the inlet filter. Adjusted the span only.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	NA
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	13:30	End Time (MST):	15:48
Analyzer make:	Teledyne API T400	Analyzer serial #:	7045

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999991	≥0.995
400.0	400.4	0.9990	Slope	1.000514	0.90 - 1.10
200.0	201.4	0.9930	Intercept	0.560000	+/- 5
100.0	100.6	0.9940			



O₃ Calibration Plot

Date: November 17, 2025

Location: Bertha Ganter-Fort McKay





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
Calibration Date: November 14, 2025 Last Cal Date: October 22, 2025
Start time (MST): 14:06 End time (MST): 14:55

Analyzer Make: Teledyne API T640 S/N: 322
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752
Temp/RH standard: Alicat FP-25BT S/N: 388752

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	0.5	-1.0	0.5	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	723.6	725.83	723.6	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.07	4.913	5.07	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	50		50	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA:	7.1	PM w/ HEPA:	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11.0			<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: October 22, 2025
Date Disposable Filter Changed: August 20, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 20, 2025
Date RH/T Sensor Cleaned: October 22, 2025

Notes: Removing the instrument. Flow, temperature, and pressure were verified. Leak check passed. PMT peak test verified.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay - Bertha Ganter Station number: AMS 01
Calibration Date: November 14, 2025 Last Cal Date: NA
Start time (MST): 15:00 End time (MST): 15:47

Analyzer Make: Teledyne API T640 S/N: 323
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388752
Temp/RH standard: Alicat FP-25BT S/N: 388752

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-3	-3.6	-3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	728.1	727.09	728.1	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	4.946	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	40		40	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA:	0.4	PM w/ HEPA:	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test			10.8	<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: NA
Date Disposable Filter Changed: NA

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: NA
Date RH/T Sensor Cleaned: NA

Notes: Installing a new instrument. Flow, temperature, and pressure were verified. Leak check passed. PMT peak test completed.

Calibration by: Rene Chamberland



Wood Buffalo Environmental Association

Nt - NOX - NH3 Calibration Report

Station Information

Station Name:	Bertha Ganter-Fort McKay	Station number:	AMS 01
NOX Cal Date:	November 17, 2025	Last Cal Date:	N/A
Start time (MST):	11:30	End time (MST):	16:25
NH3 Cal Date:	November 17, 2025	Last Cal Date:	N/A
Start time (MST):	16:50	End time (MST):	19:02
Reason:	Install		

Calibration Standards

NOX Cal Gas Conc:	59.40	ppm	NO Gas Cylinder #:	CC335700
NO Cal Gas Conc:	59.20	ppm	NO Cal Gas Expiry:	September 1, 2032
Removed NOX Conc:	59.40	ppm	Removed Cylinder #:	NA
Removed NO Conc:	59.20	ppm	Removed cyl Expiry:	NA
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	77.80	ppm	NH3 Gas Cylinder #:	CC711249
			NH3 Cal Gas Expiry:	December 31, 2025
Removed NH3 Conc:	77.80	ppm	Removed Cylinder #:	NA
NH3 gas Diff:			Removed cyl Expiry:	NA
Calibrator Model:	API T750		Serial Number:	281
ZAG make/model:	API T751H		Serial Number:	530

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	152
Converter model:	API T501	Converter serial #:	484
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.90
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	543

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:		1.093	Nt coefficient:		1.136
NOX coefficient:		1.122	NO bkgnd:		-0.1
NO2 coefficient:		1.000	NOX bkgnd:		0.0
NH3 coefficient:		0.970	Nt bkgnd:		0.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:		1.005206
NO _x Cal Offset:		0.220000
NO Cal Slope:		1.004175
NO Cal Offset:		0.820000
NO ₂ Cal Slope:		0.983228
NO ₂ Cal Offset:		0.140606
NH3 Cal Slope:		1.001563
NH3 Cal Offset:		1.006481
Nt Cal Slope:		1.004131
Nt Cal Offset:		1.512338



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

NO_x / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO _x Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>	Baseline corr NO Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>
-----------	------------------------------	--------------------------------	---	--	--	--	---	---	---	--

As found zero

As found span

AF GPT span

new NO cyl rp

Baseline Corr As Fd Nt = NA ppb NO_x = NA ppb NO = NA ppb

*Percent Change Nt_(NO) = NA

Previous Response Nt = NA ppb NO_x = NA ppb NO = NA ppb

*Percent Change NO_x = NA

**NO_x Δ (NO to GPT response) =

*Percent Change NO = NA

* = > +/-2% difference initiates investigation

* = > +/-5% change initiates investigation

NO_x / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.3	----	----
High point	4932	67.6	803.1	800.4	803.1	806.5	803.0	807.2	0.9958	0.9967
Mid point	4966	33.8	401.5	400.2	401.5	406.9	406.8	405.5	0.9868	0.9838
Low point	4983	16.9	200.8	200.1	200.8	199.9	199.7	198.7	1.0044	1.0020
Average Correction Factor									0.9957	0.9942

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found zero						
Calibration zero	----	----	0.0	0.0	----	----
High GPT point (400 ppb O3)	801.0	406.1	397.6	391.8	1.0148	98.5%
Mid GPT point (200 ppb O3)	801.0	608.9	194.8	189.3	1.0291	97.2%
Low GPT point (100 ppb O3)	801.0	706.0	97.7	98.0	0.9970	100.3%
Average Correction Factor					1.0136	98.7%



Wood Buffalo Environmental Association NH₃ - N_T Calibration Report

NH₃ As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH3 Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.0	----	----
AF High point										
AF Mid point										
AF Low point										
new NH3 cyl rp										
Baseline Corr As Fd	Nt =	NA ppb	NH3 =	NA ppb					*Percent Change	Nt _(NH3) = NA
Previous Response	Nt =	NA ppb	NH3 =	NA ppb					*Percent Change	NH3 = NA

* = > +/-5% change initiates investigation

NH₃ Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH3 concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH3 concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH3 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.0	----	----
High point	2931	69.4	1799.8	0.0	1799.8	1805.2	5.2	1800.2	0.997	1.000
Mid point	2961	38.6	1001.0	0.0	1001.0	1014.6	3.1	1011.4	0.987	0.990
Low point	2981	19.3	500.5	0.0	500.5	500.9	2.0	498.9	0.999	1.003
Average Correction Factor									0.9943	0.9976
NH3 Previous Converter Efficiency =			NA %							
NH3 Current Converter Efficiency =			97.0 %							

Notes: Installing a new instrument and converter. Calibrating with a portable calibrator/ZAG. Adjusted the NOx, NO, NT zero and spans. Adjusted the NH3 span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


Nt Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	N/A
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:30	End Time (MST):	16:25
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999951	<i>≥0.995</i>
1799.8	1805.2	0.9970	Slope	1.004131	<i>0.90 - 1.10</i>
1001.0	1014.6	0.9866	Intercept	1.512338	<i>+/-20</i>
500.5	500.9	0.9992			





Wood Buffalo Environmental Association


NH₃ Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	N/A
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:30	End Time (MST):	16:25
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999952	<i>≥0.995</i>
1799.8	1800.2	0.9998	Slope	1.001563	<i>0.90 - 1.10</i>
1001.0	1011.4	0.9897	Intercept	1.006481	<i>+/-20</i>
500.5	498.9	1.0032			





Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	N/A
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:30	End Time (MST):	16:25
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999958	<i>≥0.995</i>
803.1	806.5	0.9958	Slope	1.005206	<i>0.90 - 1.10</i>
401.5	406.9	0.9868	Intercept	0.220000	<i>+/-20</i>
200.8	199.9	1.0044			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	N/A
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:30	End Time (MST):	16:25
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999933	≥ 0.995
800.4	803.0	0.9967	Slope	1.004175	0.90 - 1.10
400.2	406.8	0.9838	Intercept	0.820000	+/-20
200.1	199.7	1.0020			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information


Calibration Date:	November 17, 2025	Previous Calibration:	N/A
Station Name:	Bertha Ganter-Fort McKay	Station Number:	AMS 01
Start Time (MST):	11:30	End Time (MST):	16:25
Analyzer make:	API T201	Analyzer serial #:	152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999887	<i>≥0.995</i>
397.6	391.8	1.0148	Slope	0.983228	<i>0.90 - 1.10</i>
194.8	189.3	1.0291	Intercept	0.140606	<i>+/-20</i>
97.7	98.0	0.9970			









WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS02 MILDRED LAKE NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Mildred Lake Station number: AMS 02
Calibration Date: November 6, 2025 Last Cal Date: October 30, 2025
Start time (MST): 11:05 End time (MST): 14:50
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.99 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: EB0112903
Removed Cal Gas Conc: 50.99 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 1185
Zero Air Gen Model: Teledyne API T701 Serial Number: 4891

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1404901075
Analyzer Range: 0-1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001740	0.999500	Backgd or Offset:	24.2	24.3
Calibration intercept:	-0.210939	-0.351922	Coeff or Slope:	0.762	0.762

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4913	78.6	803.0	794.7	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	794.8	Previous response	804.2	*% change	-1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4913	78.4	801.0	800.3	1.001
Mid point	4961	39.2	399.8	399.4	1.001
Low point	4980	19.6	199.9	198.7	1.006
As left zero	5000	0.0	0.0	0.0	----
As left span	4913	78.4	801.0	801.0	1.000
Average Correction Factor:					1.003

Notes: Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 30, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	11:05	End Time (MST):	14:50
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1404901075

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999998	≥0.995
801.0	800.3	1.0008	Slope	0.999500	0.90 - 1.10
399.8	399.4	1.0009	Intercept	-0.351922	+/-30
199.9	198.7	1.0059			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: November 6, 2025

Location: Mildred Lake





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Mildred Lake
Calibration Date: November 19, 2025
Start time (MST): 10:03
Reason: Routine

Station number: AMS 02
Last Cal Date: October 22, 2025
End time (MST): 14:50

Calibration Standards

Cal Gas Concentration: 4.75 ppm
Cal Gas Cylinder #: CC700774
Removed Cal Gas Conc: 4.75 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 1185
Serial Number: 4891

Analyzer Information

Analyzer make: Thermo 43iQTL
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12333331546
Converter serial #: 2023-267
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000125	0.998125	Backgd or Offset:	1.39	1.39
Calibration intercept:	0.000000	0.120000	Coeff or Slope:	0.957	0.957

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4916	84.2	80.0	79.8	1.002
As found Mid point	4958	42.1	40.0	40.1	0.997
As found Low point	4979	21.1	20.0	20.1	0.995
New cylinder response					
Baseline Corr As found:	79.8	Prev response:	80.00	*% change:	-0.3%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	0.997268	AF Intercept:	0.100000
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999991	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4916	84.2	80.0	79.9	1.001
Mid point	4958	42.1	40.0	40.2	0.995
Low point	4979	21.1	20.0	20.0	1.000
As left zero	5000	0.0	0.0	0.2	----
As left span	4916	84.2	80.0	79.9	1.001
SO2 Scrubber Check	4922	78.4	783.9	0.0	----
Date of last scrubber change:		July 16, 2024		Ave Corr Factor	0.999
Date of last converter efficiency test:		NA			

Notes: Changed sample inlet filter after multipoint as founds. SO2 scrubber check done and passed. No adjustments made.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	November 19, 2025	Previous Calibration:	October 22, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	10:03	End Time (MST):	14:50
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331546

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999990		≥ 0.995
80.0	79.9	1.0011	Slope	0.998125		$0.90 - 1.10$
40.0	40.2	0.9949	Intercept	0.120000		± 3
20.0	20.0	0.9999				



H2S Calibration Plot

Date: November 19, 2025

Location: Mildred Lake





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Mildred Lake
 Calibration Date: November 6, 2025
 Start time (MST): 11:05
 Reason: Routine

Station number: AMS 02
 Last Cal Date: October 30, 2025
 End time (MST): 14:50

Calibration Standards

Gas Cert Reference:	EB0112903	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	503.1 ppm	CH ₄ Equiv Conc.	1067.1 ppm
C ₃ H ₈ Cal Gas Conc.	205.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	503.1 ppm	CH ₄ Equiv Conc.	1067.1 ppm
Removed C ₃ H ₈ Conc.	205.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	1185
Zero Air Gen model:	Teledyne API T701	Serial Number:	4891

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1180320040
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	5.10E-04	5.10E-04	NMHC SP Ratio:	7.72E-05
CH ₄ Retention time:	15.9	15.9	NMHC Peak Area:	114549
Zero Chromatogram:	ON	ON	Flat Baseline:	ON

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4922	78.4	16.73	16.73	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.73	Prev response	16.76	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	
High point	4922	78.4	16.73	16.70	1.002
Mid point	4961	39.2	8.37	8.30	1.008
Low point	4980	19.6	4.18	4.14	1.011
As left zero	5000	0.0	0.00	0.00	----
As left span	4922	78.4	16.73	16.74	0.999
Average Correction Factor					1.007

Notes: Changed sample inlet filter and hydrogen cylinder after as founds. No adjustments made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4922	78.4	8.84	8.84	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.84	Prev response	8.82	*% change	0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4922	78.4	8.84	8.78	1.007
Mid point	4961	39.2	4.42	4.43	0.997
Low point	4980	19.6	2.21	2.23	0.994
As left zero	5000	0.0	0.00	0.00	1.000
As left span	4922	78.4	8.84	8.84	1.000
Average Correction Factor					0.999

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4922	78.4	7.89	7.89	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.89	Prev response	7.94	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4922	78.4	7.89	7.93	0.995
Mid point	4961	39.2	3.94	3.87	1.019
Low point	4980	19.6	1.97	1.91	1.031
As left zero	5000	0.0	0.00	0.00	1.000
As left span	4922	78.4	7.89	7.90	0.998
Average Correction Factor					1.015

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.003248	0.998576
THC Cal Offset:	-0.026669	-0.025466
CH ₄ Cal Slope:	1.008814	1.006365
CH ₄ Cal Offset:	-0.021527	-0.046325
NMHC Cal Slope:	0.998283	0.992261
NMHC Cal Offset:	-0.005142	0.020659

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 30, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	11:05	End Time (MST):	14:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999987	≥ 0.995
16.73	16.70	1.0019	Slope	0.998576	$0.90 - 1.10$
8.37	8.30	1.0080	Intercept	-0.025466	± 0.5
4.18	4.14	1.0112			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 30, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	11:05	End Time (MST):	14:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999804	≥ 0.995
7.89	7.93	0.9953	Slope	1.006365	$0.90 - 1.10$
3.94	3.87	1.0192	Intercept	-0.046325	± 0.5
1.97	1.91	1.0315			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 30, 2025
Station Name:	Mildred Lake	Station Number:	AMS 02
Start Time (MST):	11:05	End Time (MST):	14:50
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320040

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999966	<i>≥0.995</i>
8.84	8.78	1.0072	Slope	0.992261	<i>0.90 - 1.10</i>
4.42	4.43	0.9973	Intercept	0.020659	<i>+/-0.5</i>
2.21	2.23	0.9937			



NMHC Calibration Plot

Date: November 6, 2025

Location: Mildred Lake





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS04 BUFFALO VIEWPOINT NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: November 17, 2025
Start time (MST): 7:16
Reason: Routine

Station number: AMS 04
Last Cal Date: October 22, 2025
End time (MST): 10:03

Calibration Standards

Cal Gas Concentration: 50.87 ppm
Cal Gas Cylinder #: CC446753
Removed Cal Gas Conc: 50.87 ppm
Removed Gas Cyl #:
Calibrator Model: Teledyne API T700
Zero Air Gen Model: Teledyne API T701

Cal Gas Exp Date: March 10, 2031
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i-LTE
Analyzer Range: 0-1000ppb

Serial Number: 1410661331

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997583	1.000741	Backgd or Offset:	1.85	1.62
Calibration intercept:	1.875213	1.695669	Coeff or Slope:	1.013	1.016

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4921	78.6	799.7	796.8	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.1	Previous response	799.7	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4921	78.6	799.7	801.3	0.998
Mid point	4961	39.3	399.8	402.5	0.993
Low point	4980	19.6	199.4	202.8	0.983
As left zero	5000	0.0	0.0	0.2	----
As left span	4921	78.6	799.7	801.6	0.998
Average Correction Factor:					0.992

Notes: Zero and Span adjusted. No Maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	October 22, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:16	End Time (MST):	10:03
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1410661331

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999984	≥0.995
799.7	801.3	0.9981	Slope	1.000741	0.90 - 1.10
399.8	402.5	0.9933	Intercept	1.695669	+/-30
199.4	202.8	0.9834			



SO2 Calibration Plot

Date: November 17, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Calibration Date: November 21, 2025
Start time (MST): 6:48
Reason: Routine

Station number: AMS 04
Last Cal Date: October 3, 2025
End time (MST): 10:35

Calibration Standards

Cal Gas Concentration: 4.80 ppm
Cal Gas Cylinder #: DT0037528
Removed Cal Gas Conc: 4.80 ppm
Removed Gas Cyl #:
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 3808
Serial Number: 362

Analyzer Information

Analyzer make: Thermo 43i-LTE
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1008841400
Converter serial #: 2022-200
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.993002	0.978001	Backgd or Offset:	1.86	1.86
Calibration intercept:	-0.021732	0.178189	Coeff or Slope:	1.077	1.077

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4917	83.3	80.0	79.9	1.002
As found Mid point	4958	41.7	40.0	40.2	0.998
As found Low point	4979	20.8	20.0	20.0	1.003
New cylinder response					
Baseline Corr As found:	79.8	Prev response:	79.38	*% change:	0.5%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	0.998292	AF Intercept:	0.118171
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4917	83.3	80.0	78.3	1.021
Mid point	4958	41.7	40.0	39.5	1.014
Low point	4979	20.8	20.0	19.7	1.014
As left zero	5000	0.0	0.0	0.3	----
As left span	4917	83.3	80.0	78.0	1.025
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:	16-May-23		Ave Corr Factor		1.016
Date of last converter efficiency test:					

Notes: Sox scrubber checked after calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	November 21, 2025	Previous Calibration:	October 3, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	6:48	End Time (MST):	10:35
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1008841400

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999988		≥0.995
80.0	78.3	1.0212	Slope	0.978001		0.90 - 1.10
40.0	39.5	1.0135	Intercept	0.178189		+/-3
20.0	19.7	1.0136				



H₂S Calibration Plot

Date: November 21, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Buffalo Viewpoint	Station number:	AMS 04
Calibration Date:	November 17, 2025	Last Cal Date:	October 7, 2025
Start time (MST):	7:16	End time (MST):	10:04
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC446753	Cal Gas Expiry Date:	March 10, 2031
CH ₄ Cal Gas Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
C ₃ H ₈ Cal Gas Conc.	204.0 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	497.2 ppm	CH ₄ Equiv Conc.	1058.2 ppm
Removed C ₃ H ₈ Conc.	204.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3808
Zero Air Gen model:	Teledyne API T701	Serial Number:	362

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038
THC Range:	0 - 20 ppm	NMHC/CH ₄ Range:	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.41E-04	3.49E-04	NMHC SP Ratio:	5.81E-04
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	151711
Zero Chromatogram:	OFF	OFF	Flat Baseline:	ON

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	16.64	16.22	1.026
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.22	Prev response	16.63	*% change	-2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	16.64	16.62	1.001
Mid point	4961	39.3	8.32	8.24	1.010
Low point	4980	19.6	4.15	4.10	1.012
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	16.64	16.67	0.998
Average Correction Factor					1.008

Notes:

Nitrogen Cylinder Changed. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	8.82	8.58	1.028
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.58	Prev response	8.81	*% change	-2.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	8.82	8.83	0.999
Mid point	4961	39.3	4.41	4.41	1.001
Low point	4980	19.6	2.20	2.20	0.999
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	8.82	8.84	0.998
Average Correction Factor					1.000

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	78.6	7.82	7.64	1.023
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.64	Prev response	7.81	*% change	-2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.6	7.82	7.79	1.003
Mid point	4961	39.3	3.91	3.83	1.020
Low point	4980	19.6	1.95	1.90	1.028
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.6	7.82	7.83	0.998
Average Correction Factor					1.017

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.001164	0.999427
THC Cal Offset:	-0.027136	-0.032747
CH ₄ Cal Slope:	1.003179	0.998253
CH ₄ Cal Offset:	-0.027504	-0.031916
NMHC Cal Slope:	0.999132	1.000506
NMHC Cal Offset:	0.000568	-0.001231

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	October 7, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:16	End Time (MST):	10:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999975		≥ 0.995
16.64	16.62	1.0010	Slope	0.999427		0.90 - 1.10
8.32	8.24	1.0097	Intercept	-0.032747		± 0.5
4.15	4.10	1.0123				





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	October 7, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:16	End Time (MST):	10:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999903	≥ 0.995
7.82	7.79	1.0029	Slope	0.998253	0.90 - 1.10
3.91	3.83	1.0198	Intercept	-0.031916	± 0.5
1.95	1.90	1.0280			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	October 7, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:16	End Time (MST):	10:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320038

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999999	<i>≥0.995</i>
8.82	8.83	0.9994	Slope	1.000506	<i>0.90 - 1.10</i>
4.41	4.41	1.0010	Intercept	-0.001231	<i>+/-0.5</i>
2.20	2.20	0.9992			



NMHC Calibration Plot

Date: November 17, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Buffalo Viewpoint
Station number: AMS 04
Calibration Date: November 5, 2025
Last Cal Date: October 1, 2025
Start time (MST): 7:10
End time (MST): 11:54
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC324979
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 3, 2032
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3808
Serial Number: 362

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.0	----	----
AF High point	4918	81.8	800.0	798.4	1.6	734.1	735.9	-1.8	1.0903	1.0854
AF Mid point										
AF Low point										
New cyl resp										

Previous Response NO_x = 799.2 ppb NO = 796.6 ppb * = > +/-5% change initiates investigation *Percent Change NO_x = -8.9%
Baseline Corr 1st pt NO_x = 733.8 ppb NO = 735.6 ppb As Found Statistics *Percent Change NO = -8.3%
Baseline Corr 2nd pt NO_x = NA ppb NO = NA ppb As found NO_x r²: Nx SI: Nx Int:
Baseline Corr 3rd pt NO_x = NA ppb NO = NA ppb As found NO r²: NO SI: NO Int:
As found NO₂ r²: NO₂ SI: NO₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 721

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.927	1.003	NO bkgnd or offset:	-0.6	-0.6
NOX coeff or slope:	0.918	0.996	NOX bkgnd or offset:	-0.7	-0.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.8	4.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998122	0.996779
NO _x Cal Offset:	0.646580	1.166469
NO Cal Slope:	0.998738	0.997937
NO Cal Offset:	-0.754295	-0.174526
NO ₂ Cal Slope:	0.992875	0.991741
NO ₂ Cal Offset:	0.593151	1.486706

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.5	0.4	0.1	----	----
High point	4918	81.8	800.0	798.4	1.6	798.3	796.7	1.7	1.0022	1.0021
Mid point	4959	40.9	400.0	399.2	0.8	400.2	398.4	1.8	0.9995	1.0020
Low point	4980	20.4	199.5	199.1	0.4	200.7	197.6	3.1	0.9940	1.0075
As left zero	5000	0.0	0.0	0.9	-0.9	1.0	0.9	0.1	----	----
As left span	4918	81.8	800.0	400.7	800.0	788.5	400.7	387.8	1.0146	1.0000
Average Correction Factor									0.9986	1.0039

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	792.1	402.1	391.6	389.4	1.0057	99.4%
Mid GPT point	792.1	597.4	196.3	196.3	1.0002	100.0%
Low GPT point	792.1	697.0	96.7	99.2	0.9752	102.5%
Average Correction Factor					0.9937	100.7%

Notes: Reaction Cell and orifice were changed out end of September. Diagnostics are similar to last months calibration. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 5, 2025	Previous Calibration:	October 1, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:10	End Time (MST):	11:54
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999997	≥0.995
800.0	798.3	1.0022	Slope	0.996779	0.90 - 1.10
400.0	400.2	0.9995	Intercept	1.166469	+/-20
199.5	200.7	0.9940			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 5, 2025	Previous Calibration:	October 1, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:10	End Time (MST):	11:54
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999936	≥0.995
391.6	389.4	1.0057	Slope	0.991741	0.90 - 1.10
196.3	196.3	1.0002	Intercept	1.486706	+/-20
96.7	99.2	0.9752			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 5, 2025	Previous Calibration:	October 1, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	7:10	End Time (MST):	11:54
Analyzer make:	Teledyne API T200	Analyzer serial #:	721

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999997	≥ 0.995
798.4	796.7	1.0021	Slope	0.997937	$0.90 - 1.10$
399.2	398.4	1.0020	Intercept	-0.174526	± 20
199.1	197.6	1.0075			



NO_x Calibration Plot

Date: November 5, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
Calibration Date: November 17, 2025 Last Cal Date: October 7, 2025
Start time (MST): 10:01 End time (MST): 12:22
Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700 Serial Number: 3808
ZAG Make/Model: Teledyne API T701 Serial Number: 362

Analyzer Information

Analyzer make: Teledyne API T400 Analyzer serial #: 2961
Analyzer Range: 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002086	0.999943	Backgd or Offset:	-1.0	-1.0
Calibration intercept:	0.060000	0.460000	Coeff or Slope:	1.022	1.008

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	5000	1017.5	400.0	405.0	0.988
As found Mid point					
As found Low point					
Baseline Corr As found:	405.0	Previous response	400.9	*% change	1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	5000	1017.6	400.0	400.2	1.000
Mid point	5000	834.5	200.0	201.0	0.995
Low point	5000	720.3	100.0	100.3	0.997
As left zero	5000	0.0	0.0	0.2	----
As left span	5000	1013.5	400.0	400.6	0.999
Average Correction Factor					0.997

Notes: Span adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	October 7, 2025
Station Name:	Buffalo Viewpoint	Station Number:	AMS 04
Start Time (MST):	10:01	End Time (MST):	12:22
Analyzer make:	Teledyne API T400	Analyzer serial #:	2961

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999995	≥ 0.995
400.0	400.2	0.9995	Slope	0.999943	$0.90 - 1.10$
200.0	201.0	0.9950	Intercept	0.460000	± 5
100.0	100.3	0.9970			



O₃ Calibration Plot

Date: November 17, 2025

Location: Buffalo Viewpoint





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Buffalo Viewpoint Station number: AMS 04
Calibration Date: November 21, 2025 Last Cal Date: October 23, 2025
Start time (MST): 9:35 End time (MST): 9:54

Analyzer Make: Teledyne API T640 S/N: 321
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388753
Temp/RH standard: Alicat FP-25BT S/N: 388753

Monthly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Measured</u>	<u>As left</u>	<u>Adjusted</u>	<u>(Limits)</u>
T (°C)	-1.3	-2.5	-1.3	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	716.9	718.9	716.9	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.06	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	37	----	37	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 2.4		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 30-Jan-27
Lot No.: 100128-050-051

<u>Parameter</u>	<u>As found</u>	<u>Post maintenance</u>	<u>As left</u>	<u>Adjusted</u>	<u>(Limits)</u>
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: September 26, 2025
Date Disposable Filter Changed: September 26, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: September 26, 2025
Date RH/T Sensor Cleaned: September 26, 2025

No adjustments done.

Notes:

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS05 MANNIX NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Mannix
Calibration Date: November 12, 2025
Start time (MST): 10:30
Reason: Routine

Station number: AMS 05
Last Cal Date: October 17, 2025
End time (MST): 13:40

Calibration Standards

Cal Gas Concentration: 50.06 ppm
Cal Gas Cylinder #: CC308040
Removed Cal Gas Conc: 50.06 ppm
Removed Gas Cyl #:
Calibrator Model: API T700
Zero Air Gen Model: API T701

Cal Gas Exp Date: October 22, 2032
Rem Gas Exp Date: October 22, 2032
Diff between cyl:
Serial Number: 5470
Serial Number: 361

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 1000 ppb

Serial Number: 1008841399

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001237	1.004936	Backgd or Offset:	11.1	11.0
Calibration intercept:	-0.856898	-1.177173	Coeff or Slope:	0.946	0.944

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	79.9	800.0	799.2	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	799.2	Previous response	800.1	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4920	79.9	800.0	803.6	0.995
Mid point	4960	40.0	400.5	400.1	1.001
Low point	4980	20.0	200.2	199.0	1.006
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	79.9	800.0	803.1	0.996
Average Correction Factor:					1.001

Notes: Inlet filter changed after as founds. Span adjusted.

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	November 12, 2025	Previous Calibration:	October 17, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:30	End Time (MST):	13:40
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841399

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999986	≥0.995
800.0	803.6	0.9955	Slope	1.004936	0.90 - 1.10
400.5	400.1	1.0009	Intercept	-1.177173	+/-30
200.2	199.0	1.0062			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: November 12, 2025

Location: Mannix





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Mannix
Calibration Date: November 14, 2025
Start time (MST): 10:46
Reason: Routine

Station number: AMS 05
Last Cal Date: October 6, 2025
End time (MST): 15:34

Calibration Standards

Cal Gas Concentration: 4.96 ppm
Cal Gas Cylinder #: DT0037363
Removed Cal Gas Conc: 4.96 ppm
Removed Gas Cyl #: N/A
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: N/A
Diff between cyl:
Serial Number: 5470
Serial Number: 361

Analyzer Information

Analyzer make: Thermo 43iQ
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1200326169
Converter serial #: 2022-225
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005833	1.005408	Backgd or Offset:	1.23	1.26
Calibration intercept:	0.002268	0.242153	Coeff or Slope:	1.017	1.040

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4919	80.6	80.0	78.9	1.015
As found Mid point	4960	40.3	40.0	39.8	1.007
As found Low point	4980	20.2	20.0	19.7	1.022
New cylinder response					
Baseline Corr As found:	78.8	Prev response:	80.43	*% change:	-2.1%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.986535	AF Intercept:	0.102574
Baseline Corr 3rd AF pt:	19.6	AF Correlation:	0.999969	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4919	80.6	80.0	80.5	0.993
Mid point	4960	40.3	40.0	40.7	0.982
Low point	4980	20.2	20.0	20.4	0.982
As left zero	5000	0.0	0.0	0.3	----
As left span	4919	80.6	80.0	79.9	1.001
SO2 Scrubber Check	4920	80.3	803.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	0.986
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. Span adjustment made.

Calibration Performed By: Param Kaur



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	November 14, 2025	Previous Calibration:	October 6, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:46	End Time (MST):	15:34
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1200326169

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999969		≥0.995
80.0	80.5	0.9933	Slope	1.005408		0.90 - 1.10
40.0	40.7	0.9822	Intercept	0.242153		+/-3
20.0	20.4	0.9822				



H₂S Calibration Plot

Date: November 14, 2025

Location: Mannix





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Mannix Station number: AMS 05
 Calibration Date: November 12, 2025 Last Cal Date: October 17, 2025
 Start time (MST): 10:30 End time (MST): 13:40
 Reason: Routine

Calibration Standards

Gas Cert Reference: CC308040 Cal Gas Expiry Date: October 22, 2032
 CH₄ Cal Gas Conc. 500.3 ppm CH₄ Equiv Conc. 1047.6 ppm
 C₃H₈ Cal Gas Conc. 199.0 ppm
 Removed Gas Cert: Removed Gas Expiry:
 Removed CH₄ Conc. 500.3 ppm CH₄ Equiv Conc. 1047.6 ppm
 Removed C₃H₈ Conc. 199.0 ppm Diff between cyl (THC):
 Diff between cyl (CH₄): Diff between cyl (NM):
 Calibrator Model: API T700 Serial Number: 5470
 Zero Air Gen model: API T701 Serial Number: 361

Analyzer Information

Analyzer make: Thermo 55i Analyzer serial #: 15005164381
 THC Range: 0 - 20 ppm NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.61E-04	2.65E-04	NMHC SP Ratio:	4.91E-05	4.99E-05
CH ₄ Retention time:	13.7	13.7	NMHC Peak Area:	173301	175327
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.9	16.74	16.46	1.017
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.46	Prev response	16.70	*% change	-1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.9	16.74	16.82	0.995
Mid point	4960	40.0	8.38	8.42	0.995
Low point	4980	20.0	4.19	4.20	0.997
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	16.74	16.82	0.995
Average Correction Factor					0.996

Notes: Inlet filter changed after as founds. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.9	8.75	8.59	1.018
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.59	Prev response	8.74	*% change	-1.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.9	8.75	8.78	0.996
Mid point	4960	40.0	4.38	4.38	1.000
Low point	4980	20.0	2.19	2.18	1.004
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	8.75	8.81	0.993
Average Correction Factor					1.000

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.9	7.99	7.87	1.016
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.87	Prev response	7.96	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.9	7.99	8.04	0.994
Mid point	4960	40.0	4.00	4.04	0.990
Low point	4980	20.0	2.00	2.02	0.989
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.9	7.99	8.01	0.998
Average Correction Factor					0.991

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.997173	1.005029
THC Cal Offset:	0.010273	-0.002544
CH ₄ Cal Slope:	0.995208	1.005829
CH ₄ Cal Offset:	0.006632	0.007214
NMHC Cal Slope:	0.999308	1.003997
NMHC Cal Offset:	0.002841	-0.009358

Calibration Performed By: Devin Russell



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 12, 2025	Previous Calibration:	October 17, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:30	End Time (MST):	13:40
Analyzer make:	Thermo 55i	Analyzer serial #:	15005164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	1.000000	<i>≥0.995</i>
16.74	16.82	0.9951	Slope	1.005029	<i>0.90 - 1.10</i>
8.38	8.42	0.9951	Intercept	-0.002544	<i>+/-0.5</i>
4.19	4.20	0.9967			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 12, 2025	Previous Calibration:	October 17, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:30	End Time (MST):	13:40
Analyzer make:	Thermo 55i	Analyzer serial #:	15005164381

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999994	<i>≥0.995</i>
7.99	8.04	0.9940	Slope	1.005829	<i>0.90 - 1.10</i>
4.00	4.04	0.9900	Intercept	0.007214	<i>+/-0.5</i>
2.00	2.02	0.9892			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 12, 2025	Previous Calibration:	October 17, 2025
Station Name:	Mannix	Station Number:	AMS 05
Start Time (MST):	10:30	End Time (MST):	13:40
Analyzer make:	Thermo 55i	Analyzer serial #:	15005164381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999995	≥ 0.995
8.75	8.78	0.9965	Slope	1.003997	$0.90 - 1.10$
4.38	4.38	0.9995	Intercept	-0.009358	± 0.5
2.19	2.18	1.0041			



NMHC Calibration Plot

Date: November 12, 2025

Location: Mannix





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS06
PATRICIA MCINNES
NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Patricia McInnes Station number: AMS 06
Calibration Date: November 10, 2025 Last Cal Date: October 22, 2025
Start time (MST): 9:35 End time (MST): 12:17
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.08 ppm Cal Gas Exp Date: October 22, 2032
Cal Gas Cylinder #: CC255448
Removed Cal Gas Conc: 50.08 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 3566
Zero Air Gen Model: API T701 Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1160290013
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002395	0.999065	Backgd or Offset:	18.1	18.3
Calibration intercept:	1.378488	1.785433	Coeff or Slope:	0.919	0.919

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4920	79.8	799.3	798.3	1.002
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	798.1	Previous response	802.6	*% change	-0.6%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4920	79.8	799.3	799.7	1.000
Mid point	4960.1	39.9	399.6	401.5	0.995
Low point	4980	20.0	200.3	203.6	0.984
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	79.8	799.3	800.0	0.999
Average Correction Factor:					0.993

Notes: Changed the inlet filter after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 10, 2025	Previous Calibration:	October 22, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:35	End Time (MST):	12:17
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290013

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999984	≥0.995
799.3	799.7	0.9995	Slope	0.999065	0.90 - 1.10
399.6	401.5	0.9954	Intercept	1.785433	+/-30
200.3	203.6	0.9839			



SO2 Calibration Plot

Date: November 10, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: November 5, 2025
Start time (MST): 9:23
Reason: Routine

Station number: AMS 06
Last Cal Date: October 7, 2025
End time (MST): 13:44

Calibration Standards

Cal Gas Concentration: 4.760 ppm
Cal Gas Cylinder #: DT0014585
Removed Cal Gas Conc: 4.760 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1218153358
Converter serial #: 517
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000686	1.003687	Backgd or Offset:	1.99	1.99
Calibration intercept:	0.140000	0.260000	Coeff or Slope:	1.146	1.146

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4916	84.0	80.0	82.9	0.967
As found Mid point	4958	42.0	40.0	41.9	0.959
As found Low point	4979	21.0	20.0	21.0	0.961
New cylinder response					
Baseline Corr As found:	82.7	Prev response:	80.16	*% change:	3.1%
Baseline Corr 2nd AF pt:	41.7	AF Slope:	1.034128	AF Intercept:	0.320000
Baseline Corr 3rd AF pt:	20.8	AF Correlation:	0.999978	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4916	84.0	80.0	80.4	0.995
Mid point	4958	42.0	40.0	40.8	0.980
Low point	4979	21.0	20.0	20.0	1.000
As left zero	5000	0.0	0.0	0.4	----
As left span	4916	84.0	80.0	80.0	1.000
SO2 Scrubber Check				0.1	
Date of last scrubber change:		Monday, December 20, 2021		Ave Corr Factor	0.991
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Ran a SO2 scrubber check after calibrator zero. No adjustments made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	November 5, 2025	Previous Calibration:	October 7, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:23	End Time (MST):	13:44
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1218153358

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999918		≥ 0.995
80.0	80.4	0.9946	Slope	1.003687		$0.90 - 1.10$
40.0	40.8	0.9800	Intercept	0.260000		± 3
20.0	20.0	0.9996				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: November 10, 2025
Start time (MST): 9:35
Reason: Routine

Station number: AMS 06
Last Cal Date: October 22, 2025
End time (MST): 12:17

Calibration Standards

Gas Cert Reference: CC255448
CH₄ Cal Gas Conc. 501.4 ppm
C₃H₈ Cal Gas Conc. 199.3 ppm
Removed Gas Cert:
Removed CH₄ Conc. 501.4 ppm
Removed C₃H₈ Conc. 199.3 ppm
Diff between cyl (CH₄):
Calibrator Model: API T700
Zero Air Gen model: API T701

Cal Gas Expiry Date: October 22, 2032
CH₄ Equiv Conc. 1049.5 ppm
Removed Gas Expiry:
CH₄ Equiv Conc. 1049.5 ppm
Diff between cyl (THC):
Diff between cyl (NM):
Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1118148494
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.60E-04	2.65E-04	NMHC SP Ratio:	5.19E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	169079
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.75	16.44	1.019
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.44	Prev response	16.77	*% change	-2.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.75	16.75	1.000
Mid point	4960	39.9	8.37	8.37	1.001
Low point	4980	20.0	4.20	4.21	0.996
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.75	16.80	0.997
Average Correction Factor					0.999

Notes:

Changed the inlet filter after as founds. Adjusted the span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	8.75	8.63	1.014
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.63	Prev response	8.78	*% change	-1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.75	8.74	1.000
Mid point	4960.1	39.9	4.37	4.40	0.994
Low point	4980	20.0	2.19	2.24	0.981
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.75	8.77	0.997
Average Correction Factor					0.992

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	8.00	7.81	1.025
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.81	Prev response	7.98	*% change	-2.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.00	8.00	1.000
Mid point	4960.1	39.9	4.00	3.97	1.009
Low point	4980	20.0	2.01	1.98	1.014
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.00	8.03	0.997
Average Correction Factor					1.008

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.002063	0.999409
THC Cal Offset:	-0.016203	0.004998
CH ₄ Cal Slope:	1.001703	1.000363
CH ₄ Cal Offset:	-0.031995	-0.017801
NMHC Cal Slope:	1.002391	0.998314
NMHC Cal Offset:	0.015792	0.023400

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 10, 2025	Previous Calibration:	October 22, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:35	End Time (MST):	12:17
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999998	≥ 0.995
16.75	16.75	1.0002	Slope	0.999409	$0.90 - 1.10$
8.37	8.37	1.0011	Intercept	0.004998	± 0.5
4.20	4.21	0.9964			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 10, 2025	Previous Calibration:	October 22, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:35	End Time (MST):	12:17
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999972	≥ 0.995
8.00	8.00	1.0003	Slope	1.000363	$0.90 - 1.10$
4.00	3.97	1.0091	Intercept	-0.017801	± 0.5
2.01	1.98	1.0140			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 10, 2025	Previous Calibration:	October 22, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:35	End Time (MST):	12:17
Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999967	<i>≥0.995</i>
8.75	8.74	1.0004	Slope	0.998314	<i>0.90 - 1.10</i>
4.37	4.40	0.9936	Intercept	0.023400	<i>+/-0.5</i>
2.19	2.24	0.9809			



NMHC Calibration Plot

Date: November 10, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
Calibration Date:	November 24, 2025	Last Cal Date:	November 10, 2025
Start time (MST):	10:16	End time (MST):	11:52
Reason:	Cylinder Change		

Calibration Standards

Gas Cert Reference:	CC255448	Cal Gas Expiry Date:	October 22, 2032
CH ₄ Cal Gas Conc.	501.4 ppm	CH ₄ Equiv Conc.	1049.5 ppm
C ₃ H ₈ Cal Gas Conc.	199.3 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	501.4 ppm	CH ₄ Equiv Conc.	1049.5 ppm
Removed C ₃ H ₈ Conc.	199.3 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3566
Zero Air Gen model:	API T701	Serial Number:	4602

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1118148494
THC Range:	0 - 20 ppm	NMHC/CH ₄ Range:	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.60E-04	2.65E-04	NMHC SP Ratio:	5.19E-05
CH ₄ Retention time:	14.2	14.2	NMHC Peak Area:	169079
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				166553
				OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.75	16.64	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.64	Prev response	16.74	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.75	16.75	1.000
Average Correction Factor					

Notes: Changed the N2 cylinder after as founds. No adjustment made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	8.75	8.68	1.008
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.68	Prev response	8.76	*% change	-0.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.75	8.70	1.006
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	8.00	7.96	1.006
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.96	Prev response	7.99	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.00	8.05	0.994
Average Correction Factor					

Calibration Statistics

	<u>Start</u>	
THC Cal Slope:	0.999409	
THC Cal Offset:	0.004998	
CH ₄ Cal Slope:	1.000363	
CH ₄ Cal Offset:	-0.017801	
NMHC Cal Slope:	0.998314	
NMHC Cal Offset:	0.023400	


Finish

Calibration Performed By: Jan Castro

NMHC Calibration Plot

Date: November 24, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Patricia McInnes
Station number: AMS 06
Calibration Date: November 14, 2025
Last Cal Date: October 2, 2025
Start time (MST): 10:12
End time (MST): 15:31
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0036234
NOX Cal Gas Conc: 62.2 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 62.20 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: July 22, 2032
NO Cal Gas Conc: 61.90 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 61.90 ppm
NO gas Diff:
Serial Number: 3566
Serial Number: 4602

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.1	-0.2	----	----
AF High point	4935	64.6	803.7	799.8	3.9	811.2	802.3	8.7	0.9906	0.9970
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 805.1 ppb	NO = 800.2 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.8%
Baseline Corr 1st pt	NO _x = 811.3 ppb	NO = 802.2 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1172750022

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998122	1.001848
NO _x Cal Offset:	2.911651	2.350913
NO Cal Slope:	0.999055	1.002112
NO Cal Offset:	1.153160	0.533026
NO ₂ Cal Slope:	1.004883	0.998443
NO ₂ Cal Offset:	1.204514	0.071833

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.893	0.893	NO bkgnd or offset:	3.6	3.6
NOX coeff or slope:	0.997	0.997	NOX bkgnd or offset:	3.9	3.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	160.9	159.3

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.4	0.3	0.2	----	----
High point	4935	64.6	803.7	799.8	3.9	806.3	801.6	4.5	0.9968	0.9978
Mid point	4968	32.3	401.8	399.9	1.9	406.6	402.3	4.3	0.9882	0.9939
Low point	4984	16.2	201.5	200.5	1.0	205.7	201.1	4.6	0.9797	0.9973
As left zero	5000	0.0	0.0	0.0	0.0	0.7	0.4	0.3	----	----
As left span	4935	64.6	803.7	398.5	405.2	805.1	398.5	406.6	0.9982	1.0000
Average Correction Factor									0.9882	0.9963

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	799.7	395.8	407.8	407.3	1.0012	99.9%
Mid GPT point	799.7	599.2	204.4	204.0	1.0018	99.8%
Low GPT point	799.7	699.0	104.6	104.4	1.0017	99.8%
Average Correction Factor					1.0016	99.8%

Notes:

Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 14, 2025	Previous Calibration:	October 2, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:12	End Time (MST):	15:31
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999971	≥0.995
803.7	806.3	0.9968	Slope	1.001848	0.90 - 1.10
401.8	406.6	0.9882	Intercept	2.350913	+/-20
201.5	205.7	0.9797			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 14, 2025	Previous Calibration:	October 2, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:12	End Time (MST):	15:31
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999999	≥0.995
407.8	407.3	1.0012	Slope	0.998443	0.90 - 1.10
204.4	204.0	1.0018	Intercept	0.071833	+/-20
104.6	104.4	1.0017			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 14, 2025	Previous Calibration:	October 2, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	10:12	End Time (MST):	15:31
Analyzer make:	Thermo 42i	Analyzer serial #:	1172750022

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999996	≥ 0.995
799.8	801.6	0.9978	Slope	1.002112	$0.90 - 1.10$
399.9	402.3	0.9939	Intercept	0.533026	± 20
200.5	201.1	0.9973			



NO_x Calibration Plot

Date: November 14, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Patricia McInnes
Calibration Date: November 10, 2025
Start time (MST): 12:17
Reason: Routine

Station number: AMS 06
Last Cal Date: October 3, 2025
End time (MST): 15:04

Calibration Standards

O3 generation mode: Photometer
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Serial Number: 3566
Serial Number: 4602

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1300156234

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002886	1.004000	Backgd or Offset:	-0.5	-1.1
Calibration intercept:	0.220000	1.000000	Coeff or Slope:	1.020	1.020

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-0.4	----
As found High point	5000	1049.8	400.0	401.6	0.995
As found Mid point					
As found Low point					
Baseline Corr As found:	402.0	Previous response	401.4	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.4	----
High point	5000	1048.9	400.0	402.0	0.995
Mid point	5000	834.0	200.0	203.0	0.985
Low point	5000	706.1	100.0	101.4	0.986
As left zero	5000	800.0	0.0	0.4	----
As left span	5000	1049.9	400.0	404.0	0.990
Average Correction Factor					0.989

Notes: Changed the inlet filter after as founds. Adjusted zero only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	November 10, 2025	Previous Calibration:	October 3, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	12:17	End Time (MST):	15:04
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156234

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999976	≥ 0.995
400.0	402.0	0.9950	Slope	1.004000	$0.90 - 1.10$
200.0	203.0	0.9852	Intercept	1.000000	± 5
100.0	101.4	0.9862			



O₃ Calibration Plot

Date: November 10, 2025

Location: Patricia McInnes





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Patricia McInnes Station number: AMS 06
Calibration Date: November 18, 2025 Last Cal Date: October 3, 2025
Start time (MST): 14:59 End time (MST): 15:28

Analyzer Make: API T640 S/N: 1547
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-1.40	-1.99	-1.40	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	726.20	725.24	726.20	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.89	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	33	----	33	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 83.70		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 07-16-2026
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: September 16, 2025
Date Disposable Filter Changed: September 16, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: April 4, 2025
Date RH/T Sensor Cleaned: April 4, 2025

Verified flow, pressure, temperature and pump power. No adjustment needed. Leak check passed.

Notes:

Calibration by: Jan Castro



Wood Buffalo Environmental Association

Nt - NOX - NH3 Calibration Report

Station Information

Station Name:	Patricia McInnes	Station number:	AMS 06
NOX Cal Date:	November 18, 2025	Last Cal Date:	October 8, 2025
Start time (MST):	9:53	End time (MST):	13:48
NH3 Cal Date:	November 18, 2025	Last Cal Date:	October 8, 2025
Start time (MST):	14:10	End time (MST):	16:45
Reason:	Routine		

Calibration Standards

NOX Cal Gas Conc:	62.20	ppm	NO Gas Cylinder #:	DT0036234
NO Cal Gas Conc:	61.90	ppm	NO Cal Gas Expiry:	July 22, 2032
Removed NOX Conc:	62.20	ppm	Removed Cylinder #:	N/A
Removed NO Conc:	61.90	ppm	Removed cyl Expiry:	N/A
NOX gas Diff:			NO gas Diff:	
NH3 Cal Gas Conc:	75.0	ppm	NH3 Gas Cylinder #:	CC709372
Removed NH3 Conc:	75.0	ppm	NH3 Cal Gas Expiry:	December 31, 2025
NH3 gas Diff:			Removed Cylinder #:	
Calibrator Model:	API T700		Removed cyl Expiry:	
ZAG make/model:	API T701		Serial Number:	3566
			Serial Number:	4602

Analyzer Information

Analyzer model:	API T201	Analyzer serial #:	215
Converter model:	API T501	Converter serial #:	147
NH3 Range (ppb):	0 - 2000 ppb	Reaction cell Press:	6.20
NOX Range (ppb):	0 - 1000 ppb	Sample Flow:	25.4

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coefficient:	1.033	1.033	Nt coefficient:	1.027	1.007
NOX coefficient:	1.028	1.028	NO bkgrnd:	0.2	0.2
NO2 coefficient:	1.000	1.000	NOX bkgrnd:	-0.1	-0.1
NH3 coefficient:	0.956	0.950	Nt bkgrnd:	1.7	1.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.005416	1.008420
NO _x Cal Offset:	1.471046	1.789202
NO Cal Slope:	1.004684	1.012790
NO Cal Offset:	-0.067219	-0.429575
NO ₂ Cal Slope:	0.988675	0.990408
NO ₂ Cal Offset:	1.369665	-1.214169
NH3 Cal Slope:	0.992894	0.990998
NH3 Cal Offset:	6.488525	5.402369
Nt Cal Slope:	0.996556	0.994860
Nt Cal Offset:	7.069139	5.667320



Wood Buffalo Environmental Association

NO_x - NO - NO₂ Calibration Report

NO_x / NO / Nt As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	Baseline corr NO _x Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>	Baseline corr NO Correction factor (Cc/Ic) <i>Limit = 0.9 - 1.0</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.2	0.0	-0.5	----	----
As found span	4935	64.6	803.7	799.8	803.7	817.9	806.2	822.3	0.9826	0.9921
AF GPT span										
new NO cyl rp										

Baseline Corr As Fd Nt = 822.8 ppb NO_x = 817.7 ppb NO = 806.2 ppb

Previous Response Nt = 807.99 ppb NO_x = 809.5 ppb NO = 803.5 ppb

****NO_x Δ (NO to GPT response) =**

*** = > +/-2% difference initiates investigation**

***Percent Change Nt(NO) = 1.8%**

***Percent Change NO_x = 1.0%**

***Percent Change NO = 0.3%**

*** = > +/-5% change initiates investigation**

NO_x / NO / Nt Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated Nt concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated Nt concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.3	0.2	0.3	----	----
High point	4935	64.6	803.7	799.8	803.7	811.6	810.1	803.0	0.9903	0.9873
Mid point	4968	32.3	401.8	399.9	401.8	407.4	403.7	407.4	0.9862	0.9905
Low point	4984	16.2	201.5	200.5	201.5	206.7	202.4	206.9	0.9749	0.9908
Average Correction Factor									0.9838	0.9895

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found zero	----	----	0.0	0.2	----	----
Calibration zero	----	----	0.0	0.1	----	----
High GPT point (400 ppb O3)	811.3	399.8	415.4	411.9	1.0084	99.2%
Mid GPT point (200 ppb O3)	811.3	602.9	212.3	205.3	1.0340	96.7%
Low GPT point (100 ppb O3)	811.3	705.5	109.7	108.1	1.0146	98.6%
Average Correction Factor					1.0190	98.1%



Wood Buffalo Environmental Association NH₃ - N_T Calibration Report

NH₃ As Found Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH ₃ concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH ₃ concentration (ppb) (Ic)	Baseline corr Nt Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>	Baseline corr NH ₃ Correction factor (Cc/(Ic-zero)) <i>Limit = 0.9 - 1.1</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.5	0.2	-0.7	----	----
AF High point	3416	84.0	1799.0	0.0	1799.0	1791.2	7.3	1784.0	1.004	1.008
AF Mid point										
AF Low point										
new NH ₃ cyl rp										
Baseline Corr As Fd	Nt = 1791.7 ppb	NH ₃ = 1784.7 ppb							*Percent Change	Nt _(NH₃) = -0.5%
Previous Response	Nt = 1799.9 ppb	NH ₃ = 1792.7 ppb				* = > +/-5% change initiates investigation			*Percent Change	NH ₃ = -0.5%

NH₃ Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated Nt concentration (ppb) (Cc)	Calculated NOX concentration (ppb) (Cc)	Calculated NH ₃ concentration (ppb) (Cc)	Indicated Nt concentration (ppb) (Ic)	Indicated NOX concentration (ppb) (Ic)	Indicated NH ₃ concentration (ppb) (Ic)	Nt Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NH ₃ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibration zero	5000	0.0	0.0	0.0	0.0	0.3	0.3	0.0	----	----
High point	3416	84.0	1799.0	0.0	1799.0	1791.2	7.3	1784.0	1.004	1.008
Mid point	3453	46.7	1000.3	0.0	1000.3	1006.5	4.2	1002.3	0.994	0.998
Low point	3477	23.3	499.0	0.0	499.0	506.0	2.1	503.9	0.986	0.990
Average Correction Factor									0.9948	0.9989
NH ₃ Previous Converter Efficiency =			95.6 %							
NH ₃ Current Converter Efficiency =			95.0 %							

Notes:

Changed the inlet filter after as founds. Adjusted TNx and NH₃ span.

Calibration Performed By:

Jan Castro



Wood Buffalo Environmental Association


Nt Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 8, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:53	End Time (MST):	13:48
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999946	≥ 0.995
1799.0	1791.2	1.0044	Slope	0.994860	0.90 - 1.10
1000.3	1006.5	0.9938	Intercept	5.667320	+/-20
499.0	506.0	0.9861			





Wood Buffalo Environmental Association


NH₃ Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 8, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:53	End Time (MST):	13:48
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999945	<i>≥0.995</i>
1799.0	1784.0	1.0084	Slope	0.990998	<i>0.90 - 1.10</i>
1000.3	1002.3	0.9980	Intercept	5.402369	<i>+/-20</i>
499.0	503.9	0.9902			





Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 8, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:53	End Time (MST):	13:48
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.3	----	Correlation Coefficient	0.999984	<i>≥0.995</i>
803.7	811.6	0.9903	Slope	1.008420	<i>0.90 - 1.10</i>
401.8	407.4	0.9862	Intercept	1.789202	<i>+/-20</i>
201.5	206.7	0.9749			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 8, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:53	End Time (MST):	13:48
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999996	≥ 0.995
799.8	810.1	0.9873	Slope	1.012790	0.90 - 1.10
399.9	403.7	0.9905	Intercept	-0.429575	+/-20
200.5	202.4	0.9908			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 8, 2025
Station Name:	Patricia McInnes	Station Number:	AMS 06
Start Time (MST):	9:53	End Time (MST):	13:48
Analyzer make:	API T201	Analyzer serial #:	215

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.1	----	Correlation Coefficient	0.999793	<i>≥0.995</i>
415.4	411.9	1.0084	Slope	0.990408	<i>0.90 - 1.10</i>
212.3	205.3	1.0340	Intercept	-1.214169	<i>+/-20</i>
109.7	108.1	1.0146			



NO_x Calibration Plot

Date: November 18, 2025


Location: Patricia McInnes



NH₃ Calibration Plot

Date: November 18, 2025

Location: Patricia McInnes





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS07 ATHABASCA VALLEY NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Athabasca Valley Station number: AMS07
Calibration Date: November 21, 2025 Last Cal Date: October 9, 2025
Start time (MST): 10:38 End time (MST): 14:38
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.06 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC320556
Removed Cal Gas Conc: 50.06 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3805
Zero Air Gen Model: API 701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i-LTE Serial Number: 1507864683
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996572	1.000892	Backgd or Offset:	2.66	2.73
Calibration intercept:	1.544588	2.083907	Coeff or Slope:	0.843	0.859

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4920	79.8	799.0	784.2	1.019
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	784.2	Previous response	797.8	*% change	-1.7%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4920	79.8	799.0	800.3	0.998
Mid point	4960	39.9	399.5	404.4	0.988
Low point	4980	20.0	200.2	203.5	0.984
As left zero	5000	0.0	0.0	0.1	----
As left span	4920	79.8	799.0	801.1	0.997
Average Correction Factor:					0.990

Notes: Replaced sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Parampreet Kaur



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 21, 2025	Previous Calibration:	October 9, 2025
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:38	End Time (MST):	14:38
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1507864683

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999962	≥0.995
799.0	800.3	0.9984	Slope	1.000892	0.90 - 1.10
399.5	404.4	0.9879	Intercept	2.083907	+/-30
200.2	203.5	0.9840			



SO2 Calibration Plot

Date: November 21, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Athabasca Valley Station number: AMS07
Calibration Date: November 19, 2025 Last Cal Date: October 23, 2025
Start time (MST): 10:45 End time (MST): 15:22
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.25 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC504080
Removed Cal Gas Conc: 5.25 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3805
ZAG Make/Model: API T701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 43i LTE Analyzer serial #: 1180540018
Converter make: CDN-101 Converter serial #: 551
Analyzer Range: 0 - 100 ppb Converter Temp: 840 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004105	0.999347	Backgd or Offset:	2.7	2.7
Calibration intercept:	-0.162212	0.077880	Coeff or Slope:	0.884	0.884

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4925	75.5	79.3	80.7	0.981
As found Mid point	4962	37.7	39.6	40.4	0.977
As found Low point	4981	18.9	19.8	20.4	0.968
New cylinder response					
Baseline Corr As found:	80.8	Prev response:	79.43	*% change:	1.7%
Baseline Corr 2nd AF pt:	40.5	AF Slope:	1.018380	AF Intercept:	0.037718
Baseline Corr 3rd AF pt:	20.5	AF Correlation:	0.999986	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4925	75.5	79.3	79.3	1.000
Mid point	4962	37.7	39.6	39.8	0.995
Low point	4981	18.9	19.9	19.8	1.003
As left zero	5000	0.0	0.0	0.0	----
As left span	4925	75.5	79.3	79.7	0.995
SO2 Scrubber Check	4920	79.2	792.1	0.0	----
Date of last scrubber change:	8-Aug-25		Ave Corr Factor		0.999
Date of last converter efficiency test:	Friday, April 22, 2022				

Notes: Pump changed out after as founds. No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	November 19, 2025	Previous Calibration:	October 23, 2025
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:45	End Time (MST):	15:22
Analyzer make:	Thermo 43i LTE	Analyzer serial #:	1180540018

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999990	≥ 0.995
79.3	79.3	1.0002	Slope	0.999347	$0.90 - 1.10$
39.6	39.8	0.9952	Intercept	0.077880	± 3
19.9	19.8	1.0029			

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name:	Athabasca Valley	Station number:	AMS 07
Calibration Date:	November 21, 2025	Last Cal Date:	October 9, 2025
Start time (MST):	10:38	End time (MST):	14:38
Reason:	Routine		

Calibration Standards

Gas Cert Reference:	CC320556	Cal Gas Expiry Date:	Monday, March 10, 2031
CH ₄ Cal Gas Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	496.0 ppm	CH ₄ Equiv Conc.	1059.8 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	3805
Zero Air Gen model:	Teledyne API T701H	Serial Number:	198

Analyzer Information

Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520
THC Range:	0 - 20 ppm	NMHC/CH ₄ Range:	0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.72E-04	2.73E-04	NMHC SP Ratio:	5.19E-05
CH ₄ Retention time:	14.4	14.4	NMHC Peak Area:	173304
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	16.91	16.79	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.79	Prev response	16.92	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	16.91	16.89	1.001
Mid point	4960	39.9	8.46	8.46	1.000
Low point	4980	20.0	4.24	4.28	0.992
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	16.91	16.83	1.005
Average Correction Factor					0.998

Notes: Replaced sample inlet filter after as founds. Adjusted span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	9.00	8.91	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.91	Prev response	9.02	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.8	9.00	8.98	1.002
Mid point	4960	39.9	4.50	4.52	0.996
Low point	4980	20.0	2.26	2.29	0.986
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.00	8.95	1.005
Average Correction Factor					0.995

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	7.92	7.88	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.88	Prev response	7.89	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4920	79.8	7.92	7.91	1.000
Mid point	4960	39.9	3.96	3.94	1.004
Low point	4980	20.0	1.98	1.99	0.997
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	7.92	7.88	1.004
Average Correction Factor					1.001

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.997742	0.997815
THC Cal Offset:	0.044648	0.019859
CH ₄ Cal Slope:	0.995779	0.999244
CH ₄ Cal Offset:	0.011467	-0.000539
NMHC Cal Slope:	0.999216	0.996990
NMHC Cal Offset:	0.033182	0.019197

Calibration Performed By: Parampreet Kaur



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 21, 2025	Previous Calibration:	October 9, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:38	End Time (MST):	14:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999993	≥ 0.995
16.91	16.89	1.0014	Slope	0.997815	$0.90 - 1.10$
8.46	8.46	0.9996	Intercept	0.019859	± 0.5
4.24	4.28	0.9916			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 21, 2025	Previous Calibration:	October 9, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:38	End Time (MST):	14:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999994	<i>≥0.995</i>
7.92	7.91	1.0003	Slope	0.999244	<i>0.90 - 1.10</i>
3.96	3.94	1.0038	Intercept	-0.000539	<i>+/-0.5</i>
1.98	1.99	0.9975			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 21, 2025	Previous Calibration:	October 9, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:38	End Time (MST):	14:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259520

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999979	≥ 0.995
9.00	8.98	1.0020	Slope	0.996990	$0.90 - 1.10$
4.50	4.52	0.9964	Intercept	0.019197	± 0.5
2.26	2.29	0.9864			



NMHC Calibration Plot

Date: November 21, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Athabasca Valley
Station number: AMS 07
Calibration Date: November 7, 2025
Last Cal Date: October 17, 2025
Start time (MST): 10:55
End time (MST): 16:15
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0033919
NOX Cal Gas Conc: 60.10 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 60.10 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.90 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 59.90 ppm
NO gas Diff:
Serial Number: 3805
Serial Number: 198

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
AF High point	4933	66.8	803.0	800.3	2.7	807.2	801.8	5.4	0.9948	0.9981
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 802.1 ppb	NO = 799.9 ppb	<i>* = > +/-5% change initiates investigation</i>		*Percent Change	NO _x = 0.6%
Baseline Corr 1st pt	NO _x = 807.2 ppb	NO = 801.8 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.2%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1160120024

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.996059	0.998479
NO _x Cal Offset:	2.251931	2.151910
NO Cal Slope:	0.997028	0.998642
NO Cal Offset:	1.951944	1.811941
NO ₂ Cal Slope:	1.000683	1.004105
NO ₂ Cal Offset:	0.151290	0.749299

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.917	0.909	NO bkgnd or offset:	6.4	6.3
NOX coeff or slope:	1.002	1.004	NOX bkgnd or offset:	6.6	6.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	185.1	185.7

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
High point	4933	66.8	803.0	800.3	2.7	802.5	799.8	2.7	1.0006	1.0006
Mid point	4966	33.4	401.5	400.2	1.3	405.1	403.4	1.7	0.9912	0.9920
Low point	4983	16.7	200.7	200.1	0.7	204.1	202.6	1.5	0.9836	0.9876
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
As left span	4933	66.8	803.0	398.9	404.1	803.8	398.9	405.0	0.9990	1.0000
Average Correction Factor									0.9918	0.9934

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	797.5	398.6	401.6	403.5	0.9952	100.5%
Mid GPT point	797.5	599.7	200.5	202.6	0.9895	101.1%
Low GPT point	797.5	699.0	101.2	103.1	0.9813	101.9%
Average Correction Factor					0.9887	101.1%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 17, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:55	End Time (MST):	16:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999963	≥ 0.995
803.0	802.5	1.0006	Slope	0.998479	$0.90 - 1.10$
401.5	405.1	0.9912	Intercept	2.151910	± 20
200.7	204.1	0.9836			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 17, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:55	End Time (MST):	16:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999979	≥0.995
401.6	403.5	0.9952	Slope	1.004105	0.90 - 1.10
200.5	202.6	0.9895	Intercept	0.749299	+/-20
101.2	103.1	0.9813			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 17, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	10:55	End Time (MST):	16:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1160120024

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999974	≥ 0.995
800.3	799.8	1.0006	Slope	0.998642	$0.90 - 1.10$
400.2	403.4	0.9920	Intercept	1.811941	± 20
200.1	202.6	0.9876			



NO_x Calibration Plot

Date: November 7, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Athabasca Valley
Calibration Date: November 4, 2025
Start time (MST): 10:09
Reason: Routine

Station number: AMS07
Last Cal Date: October 1, 2025
End time (MST): 12:59

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: T700
ZAG Make/Model: T701H

Serial Number: 3805
Serial Number: 198

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1507964700

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.005086	1.006714	Backgd or Offset:	-1.8	-1.8
Calibration intercept:	-0.240000	0.100000	Coeff or Slope:	1.556	1.556

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	NA	0.0	-0.2	----
As found High point	5000	1705.1	400.0	401.8	0.995
As found Mid point					
As found Low point					
Baseline Corr As found:	402.0	Previous response	401.8	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	NA	0.0	0.0	----
High point	5000	1705.1	400.0	402.6	0.994
Mid point	5000	1172.8	200.0	201.9	0.991
Low point	5000	921.2	100.0	100.6	0.994
As left zero	5000	NA	0.0	-0.2	----
As left span	5000	1582.6	400.0	406.0	0.985
Average Correction Factor					0.993

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	November 4, 2025	Previous Calibration:	October 1, 2025
Station Name:	Athabasca Valley	Station Number:	AMS07
Start Time (MST):	10:09	End Time (MST):	12:59
Analyzer make:	Thermo 49i	Analyzer serial #:	1507964700

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999997	≥0.995
400.0	402.6	0.9935	Slope	1.006714	0.90 - 1.10
200.0	201.9	0.9906	Intercept	0.100000	+/- 5
100.0	100.6	0.9940			



O₃ Calibration Plot

Date: November 4, 2025

Location: Athabasca Valley





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Athabasca Valley Station number: AMS 07
Calibration Date: November 24, 2025 Last Cal Date: October 9, 2025
Start time (MST): 13:01 End time (MST): 14:11

Analyzer Make: API T640 S/N: 2235
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-4.70	-4.96	-4.70	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	741.06	740.66	741.06	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.03	5.00	5.03	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	35	----	35	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 8.20		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: July 16, 2026
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.90	11.10	11.10	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: November 24, 2025
Date Disposable Filter Changed: November 24, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: October 9, 2025
Date RH/T Sensor Cleaned: October 9, 2025

Notes: Temp, pressure and flow checked. Leak check passed. Quarterly checks done.

Calibration by: Jan Castro



Wood Buffalo Environmental Association

CO Calibration Report

Station Information

Station Name: Athabasca Valley Station number: AMS 07
Calibration Date: November 4, 2025 Last Cal Date: October 24, 2025
Start time (MST): 13:01 End time (MST): 16:08
Reason: Routine

Calibration Standards

Cal Gas Concentration: 2,953 ppm Cal Gas Exp Date: September 30, 2029
Cal Gas Cylinder #: T1TWKRN
Removed Cal Gas Conc: 2,953 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3805
ZAG Make/Model: Teledyne API 701H Serial Number: 198

Analyzer Information

Analyzer make: Thermo 48i-TLE Analyzer serial #: 1408761381
Analyzer Range: 0 - 50 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998381	0.996326	Backgd or Offset:	5.170
Calibration intercept:	0.144014	0.179992	Coeff or Slope:	1.077

CO As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4932	67.8	40.0	40.0	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	39.94	Prev response:	40.12	*% change:	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CO Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4932	67.8	40.0	40.0	1.002
Mid point	4966	33.9	20.0	20.3	0.986
Low point	4983	16.9	10.0	10.2	0.976
As left zero	5000	0.0	0.0	0.0	----
As left span	4932	67.8	40.0	39.9	1.003
Average Correction Factor					0.988

Notes: No adjustments needed.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


CO Calibration Summary

Station Information

Calibration Date:	November 4, 2025	Previous Calibration:	October 24, 2025
Station Name:	Athabasca Valley	Station Number:	AMS 07
Start Time (MST):	13:01	End Time (MST):	16:08
Analyzer make:	Thermo 48i-TLE	Analyzer serial #:	1408761381

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999900	≥0.995
40.0	40.0	1.0021	Slope	0.996326	0.90 - 1.10
20.0	20.3	0.9858	Intercept	0.179992	+/-1.5
10.0	10.2	0.9757			



CO Calibration Plot

Date: November 4, 2025

Location: Athabasca Valley





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS08 FORT CHIPEWYAN NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
Calibration Date: November 18 2025 Last Cal Date: October 16 2025
Start time (MST): 10:09 End time (MST): 12:37
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.84 ppm Cal Gas Exp Date: January 6, 2030
Cal Gas Cylinder #: CC196697
Removed Cal Gas Conc: 49.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3810
Zero Air Gen Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43i-TLE Serial Number: 1236656116
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999773	1.007484	Backgd or Offset:	1.9	1.9
Calibration intercept:	-0.603699	0.194751	Coeff or Slope:	1.029	1.029

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4920	80.3	800.4	808.0	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	807.9	Previous response	799.6	*% change	1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4920	80.3	800.4	807.0	0.992
Mid point	4960	40.2	400.7	402.7	0.995
Low point	4980	20.1	200.4	202.7	0.988
As left zero	5000	0.0	0.0	0.3	----
As left span	4920	80.3	800.4	805.0	0.994
Average Correction Factor:					0.992

Notes: Changed out inlet filter after as founds. No adjustments made.

Calibration Performed By: Jeremy cardinal



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 18 2025	Previous Calibration:	October 16 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	10:09	End Time (MST):	12:37
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1236656116

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999994	≥0.995
800.4	807.0	0.9918	Slope	1.007484	0.90 - 1.10
400.7	402.7	0.9950	Intercept	0.194751	+/-30
200.4	202.7	0.9884			



SO2 Calibration Plot

Date: November 18 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
Calibration Date: November 13 2025 Last Cal Date: October 23 2025
Start time (MST): 9:38 End time (MST): 1:26
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.84 ppm Cal Gas Exp Date: August 28, 2027
Cal Gas Cylinder #: SA7549
Removed Cal Gas Conc: 4.84 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 3810
ZAG Make/Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153461
Converter make: CDN-101 Converter serial #: 630
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997099	1.013110	Backgd or Offset:	2.8
Calibration intercept:	0.082015	-0.078292	Coeff or Slope:	1.158

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4917	82.6	80.0	83.6	0.954
As found Mid point	4959	41.3	40.0	41.8	0.952
As found Low point	4979	20.7	20.0	20.7	0.959
New cylinder response					
Baseline Corr As found:	83.8	Prev response:	79.81	*% change:	4.8%
Baseline Corr 2nd AF pt:	42.0	AF Slope:	1.048562	AF Intercept:	-0.218910
Baseline Corr 3rd AF pt:	20.9	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4917	82.6	80.0	80.8	0.990
Mid point	4959	41.3	40.0	40.7	0.982
Low point	4979	20.7	20.0	20.2	0.992
As left zero	5000	0.0	0.0	0.3	----
As left span	4917	82.6	80.0	82.4	0.970
SO2 Scrubber Check	4920	80.3	803.0	0.1	----
Date of last scrubber change:	May 15, 2025			Ave Corr Factor	0.988
Date of last converter efficiency test:					

Notes: No adjustments made.

Calibration Performed By: jermey cardinal



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	November 13 2025	Previous Calibration:	October 23 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	9:38	End Time (MST):	1:26
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153461

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999969	≥ 0.995
80.0	80.8	0.9896	Slope	1.013110	$0.90 - 1.10$
40.0	40.7	0.9822	Intercept	-0.078292	± 3
20.0	20.2	0.9920			

TRS Calibration Curve







Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan
Station number: AMS 08
Calibration Date: November 12 2025
Last Cal Date: October 17 2025
Start time (MST): 11:14
End time (MST): 3:23
Reason: As Found

Calibration Standards

NO Gas Cylinder #: CC358100
NOX Cal Gas Conc: 60.10 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.10 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 197

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-1.0	-0.3	-0.8	----	----
AF High point	4933	66.7	801.8	800.4	1.3	749.7	746.4	3.3	1.0680	1.0720
AF Mid point	4967	33.3	400.2	399.6	0.7	372.8	371.2	1.6	1.0707	1.0756
AF Low point	4983	16.7	200.7	200.4	0.3	187.1	184.8	2.2	1.0672	1.0827

New cyl resp

Previous Response	NO _x = 799.6 ppb	NO = 798.8 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = -6.5%
Baseline Corr 1st pt	NO _x = 750.7 ppb	NO = 746.7 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -7.0%
Baseline Corr 2nd pt	NO _x = 373.8 ppb	NO = 371.5 ppb	As found	NO _x r ² : 0.999998	Nx SI: 0.936090	Nx Int: -1.127
Baseline Corr 3rd pt	NO _x = 188.1 ppb	NO = 185.1 ppb	As found	NO r ² : 0.999992	NO SI: 0.933465	NO Int: -1.287
			As found	NO ₂ r ² : 0.999908	NO ₂ SI: 1.005307	NO ₂ Int: -2.395

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	-0.8	----	----
As found high GPT point	743.9	368.1	377.1	377.8	0.9982	100.2%
As found mid GPT point	743.9	548.6	196.6	193.9	1.0141	98.6%
As found low GPT point	743.9	644.8	100.4	97.3	1.0322	96.9%



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12124313137

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.806	0.806
NOX coeff or slope:	0.996	0.996
NO2 coeff or slope:	1.000	1.000

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	0.8	0.8
NOX bkgnd or offset:	1.6	1.6
Reaction cell Press:	159.2	159.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998321	
NO _x Cal Offset:	-0.825795	
NO Cal Slope:	0.999327	
NO Cal Offset:	-1.145428	
NO ₂ Cal Slope:	1.005011	
NO ₂ Cal Offset:	-1.625691	

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero
High point
Mid point
Low point
As left zero
As left span

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero
High GPT point
Mid GPT point
Low GPT point

Average Correction Factor


Notes: As found calibration for pump replacement.

Calibration Performed By: Jermey cardinal

NO_x Calibration Plot

Date: November 12, 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Chipewyan
Station number: AMS 08
Calibration Date: November 17, 2025
Last Cal Date: November 12, 2025
Start time (MST): 9:45
End time (MST): 14:04
Reason: Maintenance

Calibration Standards

NO Gas Cylinder #: CC358100
NOX Cal Gas Conc: 60.10 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.10 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 3810
Serial Number: 197

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x =	NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>			*Percent Change	NO =	NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :		Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :		NO SI:	NO Int:	
					As found	NO ₂ r ² :		NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12124313137

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.806	0.806	NO bkgnd or offset:	0.8	0.8
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	1.6	1.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	159.2	159.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998321	1.009983
NO _x Cal Offset:	-0.825795	-0.565530
NO Cal Slope:	0.999327	1.008966
NO Cal Offset:	-1.145428	-0.745227
NO ₂ Cal Slope:	1.005011	1.008746
NO ₂ Cal Offset:	-1.625691	0.862590

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-1.1	-0.3	-0.8	----	----
High point	4933	66.7	801.8	800.4	1.3	809.0	807.0	1.6	0.9911	0.9919
Mid point	4967	33.3	400.2	399.6	0.7	403.8	402.5	1.7	0.9912	0.9927
Low point	4983	16.7	200.7	200.4	0.3	202.8	200.8	2.0	0.9898	0.9980
As left zero	5000	0.0	0.0	0.0	0.0	-0.6	-0.3	-0.4	----	----
As left span	4933	66.7	801.8	416.1	385.7	810.0	416.1	394.3	0.9899	1.0000
Average Correction Factor									0.9907	0.9942

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.8	----	----
High GPT point	806.0	410.5	396.8	400.4	0.9911	100.9%
Mid GPT point	806.0	595.5	211.8	215.0	0.9853	101.5%
Low GPT point	806.0	701.7	105.6	109.4	0.9656	103.6%
Average Correction Factor					0.9806	102.0%

Notes: Changed filer after as found. Performed span adjustment.

Calibration Performed By: Jermey cardinal



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	November 12, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	9:45	End Time (MST):	14:04
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12124313137

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-1.1	----	Correlation Coefficient	0.999998		≥0.995
801.8	809.0	0.9911	Slope	1.009983		0.90 - 1.10
400.2	403.8	0.9912	Intercept	-0.565530		+/-20
200.7	202.8	0.9898				





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	November 12, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	9:45	End Time (MST):	14:04
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12124313137

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.8	----	Correlation Coefficient	0.999915		≥ 0.995
396.8	400.4	0.9911	Slope	1.008746		$0.90 - 1.10$
211.8	215.0	0.9853	Intercept	0.862590		± 20
105.6	109.4	0.9656				





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	November 12, 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	9:45	End Time (MST):	14:04
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12124313137

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999998		≥ 0.995
800.4	807.0	0.9919	Slope	1.008966		0.90 - 1.10
399.6	402.5	0.9927	Intercept	-0.745227		+/-20
200.4	200.8	0.9980				



NO_x Calibration Plot

Date: November 17 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Fort Chipewyan
Calibration Date: November 10 2025
Start time (MST): 9:53
Reason: Routine

Station number: AMS 08
Last Cal Date: october 23 2025
End time (MST): 12:28

Calibration Standards

O3 generation mode: Photometer
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Serial Number: 3810
Serial Number: 135

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1152220026

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002629	1.000143	Backgd or Offset:	-2.6	-2.6
Calibration intercept:	0.740000	1.400000	Coeff or Slope:	0.995	0.995

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	5000	968.7	400.0	401.1	0.997
As found Mid point					
As found Low point					
Baseline Corr As found:	401.1	Previous response	401.8	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	5000	968.7	400.0	400.9	0.998
Mid point	5000	820.5	200.0	202.1	0.990
Low point	5000	720.0	100.0	102.3	0.978
As left zero					
As left span					

Average Correction Factor 0.988

Notes:

Calibration Performed By: Jermey cardinal



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	November 10 2025	Previous Calibration:	october 23 2025
Station Name:	Fort Chipewyan	Station Number:	AMS 08
Start Time (MST):	9:53	End Time (MST):	12:28
Analyzer make:	Thermo 49i	Analyzer serial #:	1152220026

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999971	≥ 0.995
400.0	400.9	0.9978	Slope	1.000143	$0.90 - 1.10$
200.0	202.1	0.9896	Intercept	1.400000	± 5
100.0	102.3	0.9775			



O₃ Calibration Plot

Date: November 10 2025

Location: Fort Chipewyan





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Chipewyan Station number: AMS 08
Calibration Date: November 18 2025 Last Cal Date: October 30 2025
Start time (MST): 2:00 End time (MST): 2:28

Analyzer Make: Teledyne API T640 S/N: 319
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388744
Temp/RH standard: Alicat FP-25BT S/N: 388744

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-6.1	-6.58	-6.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	734.8	736.70	734.8	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.97	4.98	4.97	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	34%		33%	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 4.50		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☐ Alignment Factor On : ☐

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: 10-Jun-24
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: August 19, 2025
Date Disposable Filter Changed: August 19, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: August 29, 2024
Date RH/T Sensor Cleaned: August 29, 2024

Notes:

No adjustments made.

Calibration by: Morgan Voyageur



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS09 BARGE LANDING NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Barge Landing Station number: AMS 09
Calibration Date: November 18, 2025 Last Cal Date: October 3, 2025
Start time (MST): 10:36 End time (MST): 13:33
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.56 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC705748
Removed Cal Gas Conc: 50.56 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3812
Zero Air Gen Model: APIT701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1118148498
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995670	1.002401	Backgd or Offset:	11.5	12.2
Calibration intercept:	0.401753	-0.478253	Coeff or Slope:	0.985	1.044

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4921	79.1	799.8	758.5	1.054
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	758.7	Previous response	796.8	*% change	-5.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4921	79.1	799.8	801.5	0.998
Mid point	4961	39.5	399.4	399.8	0.999
Low point	4980	19.8	200.2	199.5	1.004
As left zero	5000	0.0	0.0	0.2	----
As left span	4921	79.1	799.8	804.6	0.994
Average Correction Factor:					1.000

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 3, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:36	End Time (MST):	13:33
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148498

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999997	≥0.995
799.8	801.5	0.9979	Slope	1.002401	0.90 - 1.10
399.4	399.8	0.9990	Intercept	-0.478253	+/-30
200.2	199.5	1.0036			



SO2 Calibration Plot

Date: November 18, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Barge Landing
Calibration Date: November 4, 2025
Start time (MST): 10:20
Reason: Routine

Station number: AMS 09
Last Cal Date: October 8, 2025
End time (MST): 15:06

Calibration Standards

Cal Gas Concentration: 5.17 ppm
Cal Gas Cylinder #: CC511415
Removed Cal Gas Conc: 5.17 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: August 22, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3812
Serial Number: 5613

Analyzer Information

Analyzer make: Thermo 43iQ-TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1203169744
Converter serial #: 519
Converter Temp: 830 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.004685	1.005119	Backgd or Offset:	2.260	1.780
Calibration intercept:	-0.320371	-0.260552	Coeff or Slope:	0.871	0.690

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4923	77.4	80.0	83.8	0.953
As found Mid point	4961	38.7	40.0	40.9	0.974
As found Low point	4981	19.3	20.0	19.8	0.998
New cylinder response					
Baseline Corr As found:	84.0	Prev response:	80.10	*% change:	4.6%
Baseline Corr 2nd AF pt:	41.1	AF Slope:	1.052231	AF Intercept:	-0.759785
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999786	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4923	77.4	80.0	80.2	0.998
Mid point	4961	38.7	40.0	40.0	1.001
Low point	4981	19.3	20.0	19.7	1.013
As left zero	5000	0.0	0.0	0.2	----
As left span	4923	77.4	80.0	80.7	0.992
SO2 Scrubber Check	4920	80.2	802.0	0.1	----
Date of last scrubber change:	8-Jul-25		Ave Corr Factor		1.004
Date of last converter efficiency test:					

Notes: Changed inlet filter after as founds. Scrubber beads replaced and passed. It was done after calibrator zero. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	November 4, 2025	Previous Calibration:	October 8, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:20	End Time (MST):	15:06
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1203169744

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999996	≥ 0.995
80.0	80.2	0.9980	Slope	1.005119	$0.90 - 1.10$
40.0	40.0	1.0006	Intercept	-0.260552	± 3
20.0	19.7	1.0131			


TRS Calibration Curve



TRS Calibration Plot

Date: November 4, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Barge Landing
 Calibration Date: November 18, 2025
 Start time (MST): 10:36
 Reason: Routine

Station number: AMS 09
 Last Cal Date: October 3, 2025
 End time (MST): 13:33

Calibration Standards

Gas Cert Reference:	CC705748	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	505.6 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	204.8 ppm		
Removed Gas Cert:	CC151285	Removed Gas Expiry:	January 5, 2025
Removed CH ₄ Conc.	505.6 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	204.8 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3812
Zero Air Gen model:	APIT701	Serial Number:	5613

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1193585650
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.26E-04	2.40E-04	NMHC SP Ratio:	5.30E-05
CH ₄ Retention time:	14.4	14.4	NMHC Peak Area:	168197
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				5.85E-05
				152305
				OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	16.91	15.68	1.078
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	15.68	Prev response	16.92	*% change	-7.9%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	16.91	16.91	1.000
Mid point	4961	39.5	8.44	8.37	1.009
Low point	4980	19.8	4.23	4.14	1.023
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	16.91	16.90	1.000
Average Correction Factor					1.011

Notes:

Sample inlet filter was changed after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.1	8.91	8.15	1.093
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.15	Prev response	8.93	*% change	-9.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.1	8.91	8.90	1.001
Mid point	4961	39.5	4.45	4.42	1.007
Low point	4980	19.8	2.23	2.18	1.025
As left zero	5000	0.0	0.00	0.00	1.000
As left span	4921	79.1	8.91	8.91	1.011
Average Correction Factor					1.011

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.1	8.00	7.53	1.063
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.53	Prev response	7.99	*% change	-6.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.1	8.00	8.00	0.999
Mid point	4961	39.5	3.99	3.95	1.011
Low point	4980	19.8	2.00	1.96	1.022
As left zero	5000	0.0	0.00	0.00	1.001
As left span	4921	79.1	8.00	7.99	1.011
Average Correction Factor					1.011

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.003134	1.001268
THC Cal Offset:	-0.045773	-0.052971
CH ₄ Cal Slope:	1.001330	1.001961
CH ₄ Cal Offset:	-0.018787	-0.026992
NMHC Cal Slope:	1.005356	1.001018
NMHC Cal Offset:	-0.027585	-0.026179

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 3, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:36	End Time (MST):	13:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999955	≥ 0.995
16.91	16.91	1.0002	Slope	1.001268	$0.90 - 1.10$
8.44	8.37	1.0089	Intercept	-0.052971	± 0.5
4.23	4.14	1.0234			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 3, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:36	End Time (MST):	13:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999944	<i>≥0.995</i>
8.00	8.00	0.9993	Slope	1.001961	<i>0.90 - 1.10</i>
3.99	3.95	1.0108	Intercept	-0.026992	<i>+/-0.5</i>
2.00	1.96	1.0221			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 3, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:36	End Time (MST):	13:33
Analyzer make:	Thermo 55i	Analyzer serial #:	1193585650

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999960	≥ 0.995
8.91	8.90	1.0006	Slope	1.001018	0.90 - 1.10
4.45	4.42	1.0068	Intercept	-0.026179	± 0.5
2.23	2.18	1.0245			



NMHC Calibration Plot

Date: November 18, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Barge Landing
Station number: AMS 09
Calibration Date: November 13, 2025
Last Cal Date: October 20, 2025
Start time (MST): 10:03
End time (MST): 14:34
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T2Y1KDH
NOX Cal Gas Conc: 47.38 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 47.38 ppm
NOX gas Diff:
Calibrator Model: API T750
ZAG make/model: Api 751H
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 46.94 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 46.94 ppm
NO gas Diff:
Serial Number: 281
Serial Number: 530

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
AF High point	4915	85.3	808.3	800.7	7.5	794.5	787.7	6.8	1.0173	1.0166
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 807.2 ppb	NO = 798.9 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = -1.6%			
Baseline Corr 1st pt	NO _x = 794.5 ppb	NO = 787.7 ppb	<u>As Found Statistics</u>			*Percent Change	NO = -1.4%			
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:	NO Int:			
			As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1426262593

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.103	1.103	NO bkgnd or offset:	10.1	10.1
NOX coeff or slope:	0.999	0.999	NOX bkgnd or offset:	10.4	10.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	176.2	176.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998597	0.996914
NO _x Cal Offset:	0.098064	0.918075
NO Cal Slope:	0.998568	0.998810
NO Cal Offset:	-0.664019	0.076317
NO ₂ Cal Slope:	0.999032	1.003032
NO ₂ Cal Offset:	0.158236	-0.820375

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
High point	4915	85.3	808.3	800.7	7.5	806.2	799.9	6.3	1.0025	1.0011
Mid point	4957	42.6	403.7	400.0	3.7	403.9	399.5	4.4	0.9995	1.0012
Low point	4979	21.3	201.8	200.0	1.9	203.0	199.8	3.2	0.9942	1.0008
As left zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.1	-0.2	----	----
As left span	4915	85.3	808.3	441.2	367.1	839.6	441.2	398.3	0.9627	1.0000
Average Correction Factor									0.9988	1.0010

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	793.1	398.5	402.1	402.9	0.9980	100.2%
Mid GPT point	793.1	595.5	205.1	204.5	1.0030	99.7%
Low GPT point	793.1	694.1	106.5	105.3	1.0115	98.9%
Average Correction Factor					1.0041	99.6%

Notes:

Inlet filter changed after as founds. Adjusted span only.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 13, 2025	Previous Calibration:	October 20, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:03	End Time (MST):	14:34
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999994	≥ 0.995
808.3	806.2	1.0025	Slope	0.996914	$0.90 - 1.10$
403.7	403.9	0.9995	Intercept	0.918075	± 20
201.8	203.0	0.9942			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 13, 2025	Previous Calibration:	October 20, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:03	End Time (MST):	14:34
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999985	≥ 0.995
402.1	402.9	0.9980	Slope	1.003032	$0.90 - 1.10$
205.1	204.5	1.0030	Intercept	-0.820375	± 20
106.5	105.3	1.0115			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 13, 2025	Previous Calibration:	October 20, 2025
Station Name:	Barge Landing	Station Number:	AMS 09
Start Time (MST):	10:03	End Time (MST):	14:34
Analyzer make:	Thermo 42i	Analyzer serial #:	1426262593

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	1.000000	≥ 0.995
800.7	799.9	1.0011	Slope	0.998810	$0.90 - 1.10$
400.0	399.5	1.0012	Intercept	0.076317	± 20
200.0	199.8	1.0008			



NO_x Calibration Plot

Date: November 13, 2025

Location: Barge Landing





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Barge Landing Station number: AMS 09
Calibration Date: November 14, 2025 Last Cal Date: October 20, 2025
Start time (MST): 12:15 End time (MST): 12:38

Analyzer Make: API T640 S/N: 2237
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388754
Temp/RH standard: Alicat FP-25 S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-2.40	-2.89	-2.40	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	724.00	722.97	724.00	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.99	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	32	----	32	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	0.60	PM w/ HEPA: _____	0.00	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: January 30, 2027
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ October 20, 2025
Date Disposable Filter Changed: _____ October 20, 2025

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ July 22, 2025
Date RH/T Sensor Cleaned: _____ July 22, 2025

Notes: Verified flow, pressure, temperature and pump power. No adjustment needed . Leak check passed.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS11 LOWER CAMP NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Lower Camp Station number: AMS 11
Calibration Date: November 24, 2025 Last Cal Date: October 14, 2025
Start time (MST): 12:07 End time (MST): 14:02
Reason: As Found

Calibration Standards

Cal Gas Concentration: 48.75 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC741503
Removed Cal Gas Conc: 48.75 ppm Rem Gas Exp Date:
Removed Gas Cyl #: CC741503 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3811
Zero Air Gen Model: Teledyne API T701 Serial Number: 196

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 100841398
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003586		Backgd or Offset:	18.4	
Calibration intercept:	-0.402252		Coeff or Slope:	1.057	

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4932	82.2	799.2	783.9	1.020
As found Mid point	4971	41.2	400.7	394.3	1.017
As found Low point	4996	20.6	200.2	197.8	1.013
New cylinder response					
Baseline Corr As found:	783.8	Previous response	801.6	*% change	-2.3%
Baseline Corr 2nd AF pt:	394.2	AF Slope:	0.980341	AF Intercept:	0.884077
Baseline Corr 3rd AF pt:	197.7	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor:


Notes: As founds only.

Calibration Performed By: Mohammed Kashif

SO2 Calibration Plot

Date: November 24, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Lower Camp Station number: AMS 11
Calibration Date: November 24, 2025 Last Cal Date: October 14, 2025
Start time (MST): 16:45 End time (MST): 19:14
Reason: Maintenance

Calibration Standards

Cal Gas Concentration: 48.75 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC741503
Removed Cal Gas Conc: 48.75 ppm Rem Gas Exp Date:
Removed Gas Cyl #: CC741503 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3811
Zero Air Gen Model: Teledyne API T701 Serial Number: 4428

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 100841398
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003586	1.001427	Backgd or Offset:	18.4	18.9
Calibration intercept:	-0.402252	-0.821507	Coeff or Slope:	1.057	1.029

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4932	82.2	799.2	800.2	0.999
Mid point	4971	41.2	400.7	399.1	1.004
Low point	4996	20.6	200.2	199.6	1.003
As left zero	5000	0.0	0.0	-0.1	----
As left span	4932	82.2	799.2	801.9	0.997
Average Correction Factor:					1.002

Notes: Changed the sample inlet filter, analyzer fan, and ZAG. Adjusted the zero and span.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 24, 2025	Previous Calibration:	October 14, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	16:45	End Time (MST):	19:14
Analyzer make:	Thermo 43i	Analyzer serial #:	100841398

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999992	≥0.995
799.2	800.2	0.9987	Slope	1.001427	0.90 - 1.10
400.7	399.1	1.0041	Intercept	-0.821507	+/-30
200.2	199.6	1.0029			



SO2 Calibration Plot

Date: November 24, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Lower Camp Station number: AMS 11
Calibration Date: November 25, 2025 Last Cal Date: October 8, 2025
Start time (MST): 11:56 End time (MST): 16:00
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.83 ppm Cal Gas Exp Date: August 28, 2028
Cal Gas Cylinder #: CC737863
Removed Cal Gas Conc: 4.83 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3811
ZAG Make/Model: API T701H Serial Number: 196

Analyzer Information

Analyzer make: Thermo 43iQ Analyzer serial #: 1203169745
Converter make: Global G150 Converter serial #: 2022-223
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.012008	1.022591	Backgd or Offset:	2.5
Calibration intercept:	0.048287	0.128071	Coeff or Slope:	0.808

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4923	82.8	79.9	81.6	0.979
As found Mid point	4967	41.5	40.0	41.1	0.974
As found Low point	4999	20.8	20.0	20.6	0.972
New cylinder response					
Baseline Corr As found:	81.6	Prev response:	80.90	*% change:	0.9%
Baseline Corr 2nd AF pt:	41.1	AF Slope:	1.021020	AF Intercept:	0.108037
Baseline Corr 3rd AF pt:	20.6	AF Correlation:	0.999990	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4923	82.8	79.9	81.8	0.977
Mid point	4967	41.5	40.0	41.1	0.974
Low point	4999	20.8	20.0	20.6	0.972
As left zero	5000	0.0	0.0	0.1	----
As left span	4923	82.8	79.9	81.5	0.980
SO2 Scrubber Check	4932	82.2	819.7	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.974
Date of last converter efficiency test:					

Notes: Changed sample inlet filter after as founds. Ran scrubber check after calibrator zero and it passed.
No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	November 25, 2025	Previous Calibration:	October 8, 2025
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	11:56	End Time (MST):	16:00
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1203169745

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999999		≥0.995
79.9	81.8	0.9767	Slope	1.022591		0.90 - 1.10
40.0	41.1	0.9737	Intercept	0.128071		+/-3
20.0	20.6	0.9715				



H₂S Calibration Plot

Date: November 25, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH4 / NMHC Calibration Report

Station Information

Station Name: Lower Camp
Calibration Date: November 24, 2025
Start time (MST): 12:07
Reason: As Found

Station number: AMS 11
Last Cal Date: NA
End time (MST): 14:02

Calibration Standards

Gas Cert Reference:	CC741503	Cal Gas Expiry Date:	October 9, 2032
CH4 Cal Gas Conc.	504.8 ppm	CH4 Equiv Conc.	1071.9 ppm
C3H8 Cal Gas Conc.	206.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	504.8 ppm	CH4 Equiv Conc.	1071.9 ppm
Removed C3H8 Conc.	206.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH4):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3811
Zero Air Gen model:	API T701	Serial Number:	196

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1118148495
NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	2.82E-04	NA	NMHC SP Ratio:	4.79E-05	NA
CH4 Retention time:	15.0	NA	NMHC Peak Area:	194321	NA
Zero Chromatogram:	OFF		Flat Baseline:	OFF	

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4932	82.2	17.57	17.28	1.017
As found Mid point	4971	41.2	8.81	8.55	1.031
As found Low point	4996	20.6	4.40	4.23	1.042
New cylinder response					
Baseline Corr AF:	17.27	Prev response	17.61	*% change	-2.0%
Baseline Corr 2nd AF:	8.55	AF Slope:	0.984222	AF Intercept:	-0.061390
Baseline Corr 3rd AF:	4.23	AF Correlation:	0.999929	* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Notes:

AS founds only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4932	82.2	9.30	9.06	1.026
As found Mid point	4971	41.2	4.66	4.47	1.042
As found Low point	4996	20.6	2.33	2.22	1.049
New cylinder response					
Baseline Corr AF:	9.06	Prev response	9.31	*% change	-2.8%
Baseline Corr 2nd AF:	4.47	AF Slope:	0.975243	AF Intercept:	-0.033088
Baseline Corr 3rd AF:	2.22	AF Correlation:	0.999920	* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4932	82.2	8.28	8.22	1.007
As found Mid point	4971	41.2	4.15	4.07	1.020
As found Low point	4996	20.6	2.07	2.01	1.033
New cylinder response					
Baseline Corr AF:	8.22	Prev response	8.31	*% change	-1.1%
Baseline Corr 2nd AF:	4.07	AF Slope:	0.994280	AF Intercept:	-0.028701
Baseline Corr 3rd AF:	2.01	AF Correlation:	0.999932	* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor

Calibration Statistics

	<u>Start</u>	
THC Cal Slope:	1.004855	
THC Cal Offset:	-0.042179	
CH ₄ Cal Slope:	1.006101	
CH ₄ Cal Offset:	-0.019544	
NMHC Cal Slope:	1.004041	
NMHC Cal Offset:	-0.022835	

Finish


Calibration Performed By:

Mohammed Kashif

NMHC Calibration Plot

Date: November 24, 2025

Location: Lower Camp





Wood Buffalo Environmental Association

THC / CH4 / NMHC Calibration Report

Station Information

Station Name: Lower Camp
 Calibration Date: November 24, 2025
 Start time (MST): 16:45
 Reason: Install

Station number: AMS 11
 Last Cal Date: NA
 End time (MST): 19:14

Calibration Standards

Gas Cert Reference:	CC741503	Cal Gas Expiry Date:	October 9, 2032
CH4 Cal Gas Conc.	504.8 ppm	CH4 Equiv Conc.	1071.9 ppm
C3H8 Cal Gas Conc.	206.2 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH4 Conc.	504.8 ppm	CH4 Equiv Conc.	1071.9 ppm
Removed C3H8 Conc.	206.2 ppm	Diff between cyl (THC):	
Diff between cyl (CH4):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3811
Zero Air Gen model:	API T701	Serial Number:	4428

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1152430012
 NMHC/CH4 Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH4 SP Ratio:	NA	2.18E-04	NMHC SP Ratio:	NA
CH4 Retention time:	NA	13.2	NMHC Peak Area:	NA
Zero Chromatogram:		OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	82.2	17.57	17.59	0.999
Mid point	4971	41.2	8.81	8.93	0.987
Low point	4996	20.6	4.40	4.49	0.980
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	17.57	17.32	1.015
Average Correction Factor					0.989

Notes: Install calibration. Change sample inlet filter and ZAG. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	82.2	9.30	9.32	0.997
Mid point	4971	41.2	4.66	4.84	0.962
Low point	4996	20.6	2.33	2.45	0.951
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	9.30	9.06	1.026
Average Correction Factor					0.970

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	NA	Prev response	NA	*% change	NA
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4932	82.2	8.28	8.27	1.001
Mid point	4971	41.2	4.15	4.08	1.016
Low point	4996	20.6	2.07	2.04	1.015
As left zero	5000	0.0	0.00	0.00	----
As left span	4932	82.2	8.28	8.26	1.002
Average Correction Factor					1.011

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:		1.000264
THC Cal Offset:		0.054905
CH ₄ Cal Slope:		0.999194
CH ₄ Cal Offset:		-0.023512
NMHC Cal Slope:		1.000982
NMHC Cal Offset:		0.078618

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 24, 2025	Previous Calibration:	NA
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	16:45	End Time (MST):	19:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999945	<i>≥0.995</i>
17.57	17.59	0.9988	Slope	1.000264	<i>0.90 - 1.10</i>
8.81	8.93	0.9867	Intercept	0.054905	<i>+/-0.5</i>
4.40	4.49	0.9803			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 24, 2025	Previous Calibration:	NA
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	16:45	End Time (MST):	19:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999933		≥ 0.995
8.28	8.27	1.0011	Slope	0.999194		$0.90 - 1.10$
4.15	4.08	1.0160	Intercept	-0.023512		± 0.5
2.07	2.04	1.0151				





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 24, 2025	Previous Calibration:	NA
Station Name:	Lower Camp	Station Number:	AMS 11
Start Time (MST):	16:45	End Time (MST):	19:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430012

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999553	<i>≥0.995</i>
9.30	9.32	0.9970	Slope	1.000982	<i>0.90 - 1.10</i>
4.66	4.84	0.9622	Intercept	0.078618	<i>+/-0.5</i>
2.33	2.45	0.9512			



NMHC Calibration Plot

Date: November 24, 2025

Location: Lower Camp





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS13 FORT MCKAY SOUTH NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Fort McKay South
Calibration Date: November 17, 2025
Start time (MST): 12:38
Reason: Install

Station number: AMS 13
Last Cal Date: NA
End time (MST): 14:40

Calibration Standards

Cal Gas Concentration: 50.55 ppm
Cal Gas Cylinder #: CC260812
Removed Cal Gas Conc: 50.55 ppm
Removed Gas Cyl #: NA
Calibrator Model: Teledyne API T750
Zero Air Gen Model: Teledyne API 751H

Cal Gas Exp Date: December 29, 2028
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 281
Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i-TLE
Analyzer Range: 0 - 1000 ppb

Serial Number: 1331259320

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:		0.998301	Backgd or Offset:		4.23
Calibration intercept:		0.305308	Coeff or Slope:		0.969

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Previous response	NA	*% change	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4921	79.1	799.7	798.8	1.001
Mid point	4960	39.5	399.4	398.5	1.002
Low point	4980	19.8	200.2	200.5	0.998
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	79.1	799.7	799.6	1.000

Average Correction Factor: 1.001

Notes: Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	NA
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	12:38	End Time (MST):	14:40
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1331259320

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999999	≥0.995
799.7	798.8	1.0011	Slope	0.998301	0.90 - 1.10
399.4	398.5	1.0022	Intercept	0.305308	+/-30
200.2	200.5	0.9984			



SO2 Calibration Plot

Date: November 17, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Fort McKay South
Calibration Date: November 17, 2025
Start time (MST): 11:01
Reason: Removal

Station number: AMS 13
Last Cal Date: October 14, 2025
End time (MST): 12:24

Calibration Standards

Cal Gas Concentration: 50.55 ppm
Cal Gas Cylinder #: CC260812
Removed Cal Gas Conc: 50.55 ppm
Removed Gas Cyl #: NA
Calibrator Model: Teledyne API T750
Zero Air Gen Model: Teledyne API 751H

Cal Gas Exp Date: December 29, 2028
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 281
Serial Number: 321

Analyzer Information

Analyzer make: Teledyne API T100
Analyzer Range: 0 - 1000 ppb

Serial Number: 599

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003832		Backgd or Offset:	111.3	
Calibration intercept:	-3.054307		Coeff or Slope:	0.658	

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4921	79.1	799.7	801.7	0.998
As found Mid point	4960	39.5	399.4	396.2	1.008
As found Low point	4980	19.8	200.2	194.6	1.029
New cylinder response					
Baseline Corr As found:	801.6	Previous response	799.7	*% change	0.2%
Baseline Corr 2nd AF pt:	396.1	AF Slope:	1.004412	AF Intercept:	-3.202386
Baseline Corr 3rd AF pt:	194.5	AF Correlation:	0.999922	* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor:


Notes: Removal calibration.

Calibration Performed By: Sean Bala

SO2 Calibration Plot

Date: November 17, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort McKay South
Calibration Date: November 24, 2025
Start time (MST): 9:37
Reason: Routine

Station number: AMS 13
Last Cal Date: October 7, 2025
End time (MST): 13:26

Calibration Standards

Cal Gas Concentration: 4.88 ppm
Cal Gas Cylinder #: CC500241
Removed Cal Gas Conc: 4.88 ppm
Removed Gas Cyl #: CC500241
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Exp Date: September 5, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 2657
Serial Number: 1118

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1180540017
Converter serial #: 521
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.999674	0.999960	Backgd or Offset:	3.36	3.36
Calibration intercept:	-0.058418	0.081584	Coeff or Slope:	1.09	1.09

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4918	81.6	79.6	79.6	1.004
As found Mid point	4959	40.8	39.8	39.6	1.013
As found Low point	4980	20.4	19.9	19.7	1.026
New cylinder response					
Baseline Corr As found:	79.3	Prev response:	79.56	*% change:	-0.3%
Baseline Corr 2nd AF pt:	39.3	AF Slope:	0.996947	AF Intercept:	0.061579
Baseline Corr 3rd AF pt:	19.4	AF Correlation:	0.999958	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4918	81.6	79.6	79.8	0.998
Mid point	4959	40.8	39.8	39.9	0.998
Low point	4980	20.4	19.9	19.7	1.011
As left zero	5000	0.0	0.0	0.3	----
As left span	4918	81.6	79.6	80.1	0.994
SO2 Scrubber Check	4921	79.1	791.0	0.0	----
Date of last scrubber change:	7-Aug-25		Ave Corr Factor		1.002
Date of last converter efficiency test:					

Notes: Changed inlet filter after as found. SO2 scrubber check after calibrator zero and passed. No adjustment.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information


Calibration Date:	November 24, 2025	Previous Calibration:	October 7, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:37	End Time (MST):	13:26
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1180540017

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999961		≥ 0.995
79.6	79.8	0.9981	Slope	0.999960		$0.90 - 1.10$
39.8	39.9	0.9981	Intercept	0.081584		± 3
19.9	19.7	1.0106				

TRS Calibration Curve







Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Fort McKay South
Calibration Date: November 19, 2025
Start time (MST): 9:57
Reason: Routine

Station number: AMS 13
Last Cal Date: October 14, 2025
End time (MST): 12:38

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	Friday, December 29, 2028
CH ₄ Cal Gas Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T750	Serial Number:	281
Zero Air Gen model:	Teledyne API 751H	Serial Number:	321

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1181490018
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.00E-04	3.07E-04	NMHC SP Ratio:	4.62E-05
CH ₄ Retention time:	16.20	16.00	NMHC Peak Area:	196731
Zero Chromatogram:	ON	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	17.05	16.88	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.88	Prev response	17.09	*% change	-1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	17.05	16.99	1.003
Mid point	4960	39.5	8.51	8.45	1.008
Low point	4980	19.8	4.27	4.18	1.022
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	17.05	16.99	1.004
Average Correction Factor					1.011

Notes:

Changed inlet filter after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	9.08	9.07	1.001
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.07	Prev response	9.09	*% change	-0.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	9.08	9.07	1.001
Mid point	4960	39.5	4.53	4.50	1.008
Low point	4980	19.8	2.27	2.23	1.021
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	9.08	9.05	1.003
Average Correction Factor					1.010

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	16	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	196813
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	7.97	7.81	1.021
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.81	Prev response	8.00	*% change	-2.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	7.97	7.92	1.006
Mid point	4960	39.5	3.98	3.95	1.009
Low point	4980	19.8	1.99	1.95	1.023
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.1	7.97	7.94	1.004
Average Correction Factor					1.012

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	1.005364	0.997962
THC Cal Offset:	-0.049894	-0.038699
CH ₄ Cal Slope:	1.007187	0.995809
CH ₄ Cal Offset:	-0.023539	-0.016138
NMHC Cal Slope:	1.003878	0.999826
NMHC Cal Offset:	-0.026555	-0.022962

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 19, 2025	Previous Calibration:	October 14, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:57	End Time (MST):	12:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999976	<i>≥0.995</i>
17.05	16.99	1.0034	Slope	0.997962	<i>0.90 - 1.10</i>
8.51	8.45	1.0078	Intercept	-0.038699	<i>+/-0.5</i>
4.27	4.18	1.0221			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 19, 2025	Previous Calibration:	October 14, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:57	End Time (MST):	12:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			Limits
0.00	0.00	----	Correlation Coefficient	0.999979		≥ 0.995
7.97	7.92	1.0055	Slope	0.995809		$0.90 - 1.10$
3.98	3.95	1.0086	Intercept	-0.016138		± 0.5
1.99	1.95	1.0233				





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 19, 2025	Previous Calibration:	October 14, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:57	End Time (MST):	12:38
Analyzer make:	Thermo 55i	Analyzer serial #:	1181490018

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999971	≥ 0.995
9.08	9.07	1.0015	Slope	0.999826	$0.90 - 1.10$
4.53	4.50	1.0077	Intercept	-0.022962	± 0.5
2.27	2.23	1.0210			



NMHC Calibration Plot

Date: November 19, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Fort McKay South
 Calibration Date: November 27, 2025
 Start time (MST): 10:45
 Reason: Cylinder Change

Station number: AMS 13
 Last Cal Date: November 19, 2025
 End time (MST): 12:02

Calibration Standards

Gas Cert Reference:	CC260812	Cal Gas Expiry Date:	Friday, December 29, 2028
CH ₄ Cal Gas Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
C ₃ H ₈ Cal Gas Conc.	208.7 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	
Removed CH ₄ Conc.	503.6 ppm	CH ₄ Equiv Conc.	1077.5 ppm
Removed C ₃ H ₈ Conc.	208.7 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T701	Serial Number:	1118
Zero Air Gen model:	Teledyne API 700	Serial Number:	2657

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1181490018
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.07E-04	3.07E-04	NMHC SP Ratio:	4.61E-05	4.61E-05
CH ₄ Retention time:	16.00	16.00	NMHC Peak Area:	196813	196813
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	17.05	17.11	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.11	Prev response	16.97	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	17.05	17.12	0.996
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 0.996

Notes:

N2 change, no change to response.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	9.08	9.09	0.999
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.09	Prev response	9.05	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	9.08	9.09	0.998
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 0.998

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	16	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	196813
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.1	7.97	8.02	0.993
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.02	Prev response	7.92	*% change	1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.1	7.97	8.03	0.993
Mid point					
Low point					
As left zero					
As left span					

Average Correction Factor 0.993

Calibration Statistics


	<u>Start</u>	<u>Finish</u>
THC Cal Slope:	0.997962	1.004511
THC Cal Offset:	-0.038699	0.000000
CH ₄ Cal Slope:	0.995809	1.007432
CH ₄ Cal Offset:	-0.016138	0.000000
NMHC Cal Slope:	0.999826	1.001618
NMHC Cal Offset:	-0.022962	0.000000

Calibration Performed By: Ryan Power

NMHC Calibration Plot

Date: November 27, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort McKay South
Station number: AMS 13
Calibration Date: November 12, 2025
Last Cal Date: October 21, 2025
Start time (MST): 9:50
End time (MST): 14:06
Reason: Maintenance

Calibration Standards

NO Gas Cylinder #: T2UP1RP
NO_x Cal Gas Conc: 48.25 ppm
Removed Cylinder #: NA
Removed Gas NO_x Conc: 48.25 ppm
NO_x gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne APIT701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 47.88 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 47.88 ppm
NO gas Diff:
Serial Number: 2657
Serial Number: 1118

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.2	----	----
AF High point	4917	83.5	805.7	799.5	6.2	814.6	804.0	10.7	0.9888	0.9944
AF Mid point										
AF Low point										
New cyl resp										

Previous Response NO_x = 803.7 ppb NO = 797.2 ppb
Baseline Corr 1st pt NO_x = 814.8 ppb NO = 804.0 ppb
Baseline Corr 2nd pt NO_x = NA ppb NO = NA ppb
Baseline Corr 3rd pt NO_x = NA ppb NO = NA ppb

* = > +/-5% change initiates investigation

*Percent Change NO_x = 1.4%

*Percent Change NO = 0.8%

As Found Statistics

As found NO_x r²:
As found NO r²:
As found NO₂ r²:
Nx SI:
NO SI:
NO₂ SI:
Nx Int:
NO Int:
NO₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	--	--

As Found GPT zero
As found high GPT point
As found mid GPT point
As found low GPT point



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12300522720

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.126	1.126
NOX coeff or slope:	0.998	0.998
NO2 coeff or slope:	1.000	1.000

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	3.0	3.0
NOX bkgnd or offset:	3.2	3.2
Reaction cell Press:	162.1	162.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001701	1.004510
NO _x Cal Offset:	-3.392379	-2.733107
NO Cal Slope:	1.002536	1.002736
NO Cal Offset:	-4.350297	-3.770487
NO ₂ Cal Slope:	0.988647	0.984281
NO ₂ Cal Offset:	1.598208	2.392924

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.2	0.3	-0.1	----	----
High point	4917	83.5	805.7	799.5	6.2	808.3	800.2	8.1	0.9968	0.9991
Mid point	4958	41.8	403.4	400.3	3.1	400.3	394.9	5.3	1.0077	1.0137
Low point	4979	20.9	201.7	200.1	1.5	197.4	193.3	4.1	1.0217	1.0354
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.3	-0.1	----	----
As left span	4917	83.5	805.7	401.0	404.7	798.2	401.0	397.3	1.0094	1.0000
Average Correction Factor									1.0087	1.0161

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	798.4	398.9	405.7	400.3	1.0134	98.7%
Mid GPT point	798.4	601.1	203.5	204.3	0.9960	100.4%
Low GPT point	798.4	700.8	103.8	106.8	0.9717	102.9%
Average Correction Factor					0.9937	100.7%

Notes:

Adjusted span only.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 12, 2025	Previous Calibration:	October 21, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:50	End Time (MST):	14:06
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999938	≥0.995
805.7	808.3	0.9968	Slope	1.004510	0.90 - 1.10
403.4	400.3	1.0077	Intercept	-2.733107	+/-20
201.7	197.4	1.0217			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 12, 2025	Previous Calibration:	October 21, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:50	End Time (MST):	14:06
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999817	≥0.995
405.7	400.3	1.0134	Slope	0.984281	0.90 - 1.10
203.5	204.3	0.9960	Intercept	2.392924	+/-20
103.8	106.8	0.9717			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 12, 2025	Previous Calibration:	October 21, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	9:50	End Time (MST):	14:06
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12300522720

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999880	≥ 0.995
799.5	800.2	0.9991	Slope	1.002736	$0.90 - 1.10$
400.3	394.9	1.0137	Intercept	-3.770487	± 20
200.1	193.3	1.0354			



NO_x Calibration Plot

Date: November 12, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Fort McKay South
Calibration Date: November 3, 2025
Start time (MST): 10:25
Reason: Routine

Station number: AMS 13
Last Cal Date: October 6, 2025
End time (MST): 13:30

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T750
ZAG Make/Model: Teledyne API 751H

Serial Number: 281
Serial Number: 321

Analyzer Information

Analyzer make: Teledyne API T400
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 7413

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001486	1.002943	Backgd or Offset:	-3.6	-2.8
Calibration intercept:	1.140000	1.260000	Coeff or Slope:	1.030	1.025

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	0.6	----
As found High point	5000	984.4	400.0	404.4	0.991
As found Mid point					
As found Low point					
Baseline Corr As found:	403.8	Previous response	401.7	*% change	0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.5	----
High point	5000	984.4	400.0	402.0	0.995
Mid point	5000	799.8	200.0	202.4	0.988
Low point	5000	682.8	100.0	102.2	0.978
As left zero	5000	800.0	0.0	-0.1	----
As left span	5000	984.4	400.0	404.3	0.989
Average Correction Factor					0.987

Notes: Sample inlet filter was changed after as founds. Adjusted zero and span.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	November 3, 2025	Previous Calibration:	October 6, 2025
Station Name:	Fort McKay South	Station Number:	AMS 13
Start Time (MST):	10:25	End Time (MST):	13:30
Analyzer make:	Teledyne API T400	Analyzer serial #:	7413

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999983	≥0.995
400.0	402.0	0.9950	Slope	1.002943	0.90 - 1.10
200.0	202.4	0.9881	Intercept	1.260000	+/- 5
100.0	102.2	0.9785			



O₃ Calibration Plot

Date: November 3, 2025

Location: Fort McKay South





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort McKay South Station number: AMS 13
Calibration Date: November 24, 2025 Last Cal Date: October 30, 2025
Start time (MST): 14:10 End time (MST): 14:25

Analyzer Make: Teledyne API T640 S/N: 1335
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25 S/N: 388746
Temp/RH standard: Alicat FP-25 S/N: 388746

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-4.6	-5.12	-4.6	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	737.6	739.93	737.6	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	5.05	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	44	----	44	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	4.0	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: June 10, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: September 23, 2025
Date Disposable Filter Changed: September 23, 2025

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: June 2, 2025
Date RH/T Sensor Cleaned: June 2, 2025

Notes: No adjustments.

Calibration by: Sean Bala



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS14
ANZAC
NOVEMBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Anzac Station number: AMS 14
Calibration Date: November 17, 2025 Last Cal Date: October 16, 2025
Start time (MST): 11:47 End time (MST): 15:16
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.32 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC462030
Removed Cal Gas Conc: 50.32 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3060
Zero Air Gen Model: API T701H Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 0710321322
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.013696	1.008288	Backgd or Offset:	25.7	25.6
Calibration intercept:	-1.262634	-0.727248	Coeff or Slope:	1.111	1.111

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.5	----
As found High point	4937	79.6	798.4	801.5	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	801.0	Previous response	808.1	*% change	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4937	79.6	798.4	805.4	0.991
Mid point	4977	39.7	398.2	398.7	0.999
Low point	4992	19.7	197.8	198.6	0.996
As left zero	5000	0.0	0.0	0.4	----
As left span	4937	79.6	798.4	806.4	0.990
Average Correction Factor:					0.995

Notes: Changed the sample inlet filter after as founds. No adjustments made.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	October 16, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:47	End Time (MST):	15:16
Analyzer make:	Thermo 43i	Analyzer serial #:	0710321322

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999981	≥0.995
798.4	805.4	0.9914	Slope	1.008288	0.90 - 1.10
398.2	398.7	0.9988	Intercept	-0.727248	+/-30
197.8	198.6	0.9960			



SO2 Calibration Plot

Date: November 17, 2025

Location: Anzac





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Anzac
Calibration Date: November 26, 2025
Start time (MST): 10:13
Reason: Routine

Station number: AMS 14
Last Cal Date: October 28, 2025
End time (MST): 15:38

Calibration Standards

Cal Gas Concentration: 5.15 ppm
Cal Gas Cylinder #: CC510379
Removed Cal Gas Conc: 5.15 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API 701H

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3060
Serial Number: 357

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: CD Nova CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1218153582
Converter serial #: 631
Converter Temp: 800 degC

Calibration slope: Start 1.001464 Finish 0.985897
Calibration intercept: -0.058652 0.061300

Backgd or Offset: Start 2.31 Finish 2.28
Coeff or Slope: 0.990 0.990

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4922	77.7	80.0	79.9	1.000
As found Mid point	4961	38.9	40.1	40.1	0.997
As found Low point	4981	19.4	20.0	19.5	1.019
New cylinder response					
Baseline Corr As found:	80.0	Prev response:	80.09	*% change:	-0.1%
Baseline Corr 2nd AF pt:	40.2	AF Slope:	1.001363	AF Intercept:	-0.218739
Baseline Corr 3rd AF pt:	19.6	AF Correlation:	0.999961	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4922	77.7	80.0	78.9	1.014
Mid point	4961	38.9	40.1	39.7	1.009
Low point	4981	19.4	20.0	19.6	1.019
As left zero	5000	0.0	0.0	0.3	----
As left span	4922	77.7	80.0	78.7	1.017
SO2 Scrubber Check	4921	79.5	794.9	0.0	----
Date of last scrubber change:	28-May-25		Ave Corr Factor		1.014
Date of last converter efficiency test:	September 16, 2025		103.4% efficiency		

Notes: Changed sample inlet filter after as founds. Ran scrubber check after calibrator zero and it passed.
No adjustments made.

Calibration Performed By: Parampreet Kaur and Mohammed Kashif



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	November 26, 2025	Previous Calibration:	October 28, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:13	End Time (MST):	15:38
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153582

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999986	≥ 0.995
80.0	78.9	1.0140	Slope	0.985897	$0.90 - 1.10$
40.1	39.7	1.0089	Intercept	0.061300	± 3
20.0	19.6	1.0190			


TRS Calibration Curve



TRS Calibration Plot

Date: November 26, 2025

Location: Anzac





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Anzac
Calibration Date: November 17, 2025
Start time (MST): 11:47
Reason: Routine

Station number: AMS 14
Last Cal Date: October 16, 2025
End time (MST): 15:16

Calibration Standards

Gas Cert Reference:	CC462030	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	505.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
C ₃ H ₈ Cal Gas Conc.	204.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	505.3 ppm	CH ₄ Equiv Conc.	1068.8 ppm
Removed C ₃ H ₈ Conc.	204.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	3060
Zero Air Gen model:	API 701H	Serial Number:	357

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1331259521
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.98E-04	2.96E-04	NMHC SP Ratio:	5.95E-05
CH ₄ Retention time:	15.1	15.1	NMHC Peak Area:	150178
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4937	79.6	16.96	17.15	0.989
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.15	Prev response	16.93	*% change	1.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4937	79.6	16.96	16.92	1.003
Mid point	4977	39.7	8.46	8.24	1.026
Low point	4992	19.7	4.20	4.20	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4937	79.6	16.96	16.89	1.004
Average Correction Factor					1.004

Notes: Changed the sample inlet filter and N₂ cylinder after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4937	79.6	8.94	9.09	0.983
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.09	Prev response	8.93	*% change	1.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4937	79.6	8.94	8.91	1.004
Mid point	4977	39.7	4.46	4.37	1.019
Low point	4992	19.7	2.21	2.24	0.988
As left zero	5000	0.0	0.00	0.00	----
As left span	4937	79.6	8.94	8.91	1.004
Average Correction Factor					1.004

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4937	79.6	8.02	8.06	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.06	Prev response	8.00	*% change	0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4937	79.6	8.02	8.01	1.001
Mid point	4977	39.7	4.00	3.87	1.033
Low point	4992	19.7	1.99	1.96	1.016
As left zero	5000	0.0	0.00	0.00	----
As left span	4937	79.6	8.02	7.98	1.005
Average Correction Factor					1.017

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.003531	0.996009
THC Cal Offset:	-0.092035	-0.035840
CH ₄ Cal Slope:	1.005653	0.998784
CH ₄ Cal Offset:	-0.061970	-0.037171
NMHC Cal Slope:	1.001628	0.994020
NMHC Cal Offset:	-0.030065	0.000134

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	October 16, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:47	End Time (MST):	15:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999821	≥ 0.995
16.96	16.92	1.0026	Slope	0.996009	$0.90 - 1.10$
8.46	8.24	1.0259	Intercept	-0.035840	± 0.5
4.20	4.20	1.0010			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	October 16, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:47	End Time (MST):	15:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999696		≥ 0.995
8.02	8.01	1.0007	Slope	0.998784		$0.90 - 1.10$
4.00	3.87	1.0333	Intercept	-0.037171		± 0.5
1.99	1.96	1.0160				





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	October 16, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:47	End Time (MST):	15:16
Analyzer make:	Thermo 55i	Analyzer serial #:	1331259521

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999876	<i>≥0.995</i>
8.94	8.91	1.0038	Slope	0.994020	<i>0.90 - 1.10</i>
4.46	4.37	1.0195	Intercept	0.000134	<i>+/-0.5</i>
2.21	2.24	0.9884			



NMHC Calibration Plot

Date: November 17, 2025

Location: Anzac





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Anzac
Station number: AMS 14
Calibration Date: November 6, 2025
Last Cal Date: October 21, 2025
Start time (MST): 10:46
End time (MST): 18:22
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0037092
NOX Cal Gas Conc: 60.7 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.70 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T700H
Cal Gas Expiry Date: May 16, 2031
NO Cal Gas Conc: 60.40 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.40 ppm
NO gas Diff:
Serial Number: 3060
Serial Number: 357

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
AF High point	4952	66.4	803.1	799.2	4.0	810.8	803.8	7.0	0.9906	0.9941
AF Mid point	4984	33.2	401.7	399.7	2.0	403.7	399.0	4.7	0.9950	1.0015
AF Low point	5002	16.6	200.8	199.8	1.0	201.4	197.0	4.4	0.9969	1.0136

New cyl resp

Previous Response	NO _x = 802.0 ppb	NO = 799.4 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = 1.1%
Baseline Corr 1st pt	NO _x = 810.8 ppb	NO = 803.9 ppb	<u>As Found Statistics</u>		*Percent Change	NO = 0.6%
Baseline Corr 2nd pt	NO _x = 403.7 ppb	NO = 399.1 ppb	As found	NO _x r ² : 0.999993	Nx SI: 1.009833	Nx Int: -0.876
Baseline Corr 3rd pt	NO _x = 201.4 ppb	NO = 197.1 ppb	As found	NO r ² : 0.999968	NO SI: 1.007155	NO Int: -2.236
			As found	NO ₂ r ² : 0.999941	NO ₂ SI: 0.996016	NO ₂ Int: -1.314

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	0.1	----	----
As found high GPT point	802.6	409.8	396.8	394.5	1.0058	99.4%
As found mid GPT point	802.6	611.7	194.9	192.3	1.0134	98.7%
As found low GPT point	802.6	707.4	99.2	95.9	1.0341	96.7%



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430008

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.435	1.347	NO bkgnd or offset:	3.9	3.7
NOX coeff or slope:	0.996	1.002	NOX bkgnd or offset:	3.9	3.7
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	156.3	150.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999346	0.997154
NO _x Cal Offset:	-0.650114	-0.395907
NO Cal Slope:	1.003009	0.999461
NO Cal Offset:	-2.169606	-2.095797
NO ₂ Cal Slope:	0.996014	1.002788
NO ₂ Cal Offset:	-0.692405	0.577852

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.3	0.2	0.1	----	----
High point	4952	66.4	803.1	799.2	4.0	800.9	797.9	3.1	1.0028	1.0016
Mid point	4984	33.2	401.7	399.7	2.0	399.5	395.9	3.7	1.0054	1.0096
Low point	5002	16.6	200.8	199.8	1.0	199.3	195.5	3.8	1.0074	1.0219
As left zero	5000	0.0	0.0	0.0	0.0	0.3	0.1	0.2	----	----
As left span	4952	66.4	803.1	409.6	393.5	799.3	409.6	389.7	1.0048	1.0000
Average Correction Factor									1.0052	1.0110

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	794.0	409.8	388.2	389.6	0.9963	100.4%
Mid GPT point	794.0	608.0	190.0	191.3	0.9930	100.7%
Low GPT point	794.0	701.6	96.4	97.7	0.9864	101.4%
Average Correction Factor					0.9919	100.8%

Notes:

Sample inlet filter and pump was changed after as founds. Adjusted span only.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 21, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:46	End Time (MST):	18:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999996	≥0.995
803.1	800.9	1.0028	Slope	0.997154	0.90 - 1.10
401.7	399.5	1.0054	Intercept	-0.395907	+/-20
200.8	199.3	1.0074			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 21, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:46	End Time (MST):	18:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999993	≥ 0.995
388.2	389.6	0.9963	Slope	1.002788	$0.90 - 1.10$
190.0	191.3	0.9930	Intercept	0.577852	± 20
96.4	97.7	0.9864			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 21, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	10:46	End Time (MST):	18:22
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430008

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999962	≥ 0.995
799.2	797.9	1.0016	Slope	0.999461	$0.90 - 1.10$
399.7	395.9	1.0096	Intercept	-2.095797	± 20
199.8	195.5	1.0219			



NO_x Calibration Plot

Date: November 6, 2025

Location: Anzac





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Anzac
Calibration Date: November 4, 2025
Start time (MST): 11:36
Reason: Routine

Station number: AMS 14
Last Cal Date: October 1, 2025
End time (MST): 16:28

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: API T700
ZAG Make/Model: API 701H

Serial Number: 3060
Serial Number: 357

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1426262595

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001171	0.996943	Backgd or Offset:	1.6	0.5
Calibration intercept:	0.620000	0.460000	Coeff or Slope:	1.712	1.689

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted
					Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.6	----
As found High point	5000	941.3	400.0	401.4	0.998
As found Mid point	5000	820.3	200.0	200.1	1.003
As found Low point	5000	724.3	100.0	99.9	1.007
Baseline Corr As found:	400.8	Previous response	401.1	*% change	-0.1%
Baseline Corr 2nd AF pt:	199.5	AF Slope:	1.002514	AF Intercept:	0.060000
Baseline Corr 3rd AF pt:	99.3	AF Correlation:	0.999991	* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic- AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	5000	941.3	400.0	399.2	1.002
Mid point	5000	820.3	200.0	199.7	1.002
Low point	5000	724.3	100.0	100.6	0.994
As left zero	5000	0.0	0.0	0.0	----
As left span	5000	941.3	400.0	399.9	1.000
Average Correction Factor					0.999

Notes: The sample inlet filter was replaced following the multipoint as-found checks. The photometer lamp voltage was adjusted to increase cell intensities as part of maintenance. Zero and span were adjusted.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	November 4, 2025	Previous Calibration:	October 1, 2025
Station Name:	Anzac	Station Number:	AMS 14
Start Time (MST):	11:36	End Time (MST):	16:28
Analyzer make:	Thermo 49i	Analyzer serial #:	1426262595

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999997	≥0.995
400.0	399.2	1.0020	Slope	0.996943	0.90 - 1.10
200.0	199.7	1.0015	Intercept	0.460000	+/- 5
100.0	100.6	0.9940			



O₃ Calibration Plot

Date: November 4, 2025

Location: Anzac





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Anzac
Calibration Date: November 17, 2025
Start time (MST): 15:46
Station number: AMS 14
Last Cal Date: October 6, 2025
End time (MST): 15:58
Analyzer Make: AP T640
S/N: 825
Particulate Fraction: PM2.5
Flow Meter Make/Model: Alicat FP-25BT
S/N: 388749
Temp/RH standard: Alicat FP-25BT
S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	1.88	2.2	1.88	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	709.65	708.6	709.65	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.950	5.01	4.950	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	38	-----	38	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	11.0	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: _____ Expiry Date: _____
Lot No.: _____

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: _____ August 26, 2025
Date Disposable Filter Changed: _____ August 26, 2025

Post- maintenance Zero Verification: PM w/ HEPA: _____ <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: _____ August 26, 2025
Date RH/T Sensor Cleaned: _____ August 26, 2025

Notes: No adjustments made. Leak check passed. Head cleaned

Calibration by: Mohammed Kashif



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS17
WAPASU
NOVEMBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Wapasu Station number: AMS17
Calibration Date: November 20, 2025 Last Cal Date: October 15, 2025
Start time (MST): 11:34 End time (MST): 14:50
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.12 ppm Cal Gas Exp Date: April 9, 2033
Cal Gas Cylinder #: CC422255
Removed Cal Gas Conc: 50.12 ppm Rem Gas Exp Date:
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 2449
Zero Air Gen Model: Teledyne API 701H Serial Number: 1238

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1218153459
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007363	1.006606	Backgd or Offset:	15.8	15.8
Calibration intercept:	-2.875904	-2.636066	Coeff or Slope:	1.133	1.133

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.6	----
As found High point	4920	79.8	799.9	803.3	0.995
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	803.9	Previous response	803.0	*% change	0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	4920	79.8	799.9	804.0	0.995
Mid point	4960	39.9	400.0	398.1	1.005
Low point	4980	20.0	200.5	197.3	1.016
As left zero	5000	0.0	0.0	-0.4	----
As left span	4920	79.8	799.9	806.0	0.992
Average Correction Factor:					1.005

Notes: No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 15, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:34	End Time (MST):	14:50
Analyzer make:	Thermo 43i	Analyzer serial #:	1218153459

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999959	≥0.995
799.9	804.0	0.9950	Slope	1.006606	0.90 - 1.10
400.0	398.1	1.0047	Intercept	-2.636066	+/-30
200.5	197.3	1.0161			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: November 20, 2025

Location: Wapasu





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Wapasu
Calibration Date: November 24, 2025
Start time (MST): 9:35
Reason: Routine

Station number: AMS 17
Last Cal Date: October 20, 2025
End time (MST): 14:12

Calibration Standards

Cal Gas Concentration: 4.77 ppm
Cal Gas Cylinder #: DT20029267
Removed Cal Gas Conc: 4.77 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API T701H

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 2449
Serial Number: 359

Analyzer Information

Analyzer make: Thermo 450i
Converter make: CD Nova
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1218153583
Converter serial #: N/A
Converter Temp: 340 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.989072	0.989933	Backgd or Offset:	13.1	12.9
Calibration intercept:	0.020210	-0.159925	Coeff or Slope:	1.099	1.099

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.4	----
As found High point	4921	83.9	80.0	81.0	0.992
As found Mid point	4961	41.9	39.9	40.7	0.991
As found Low point	4980	21.0	20.0	20.3	1.007
New cylinder response					
Baseline Corr As found:	80.6	Prev response:	79.11	*% change:	1.9%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.009261	AF Intercept:	0.290586
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999982	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4916	83.9	80.0	79.2	1.011
Mid point	4958	41.9	40.0	39.2	1.020
Low point	4979	21.0	20.0	19.6	1.022
As left zero	5000	0.0	0.0	0.5	----
As left span	4916	83.9	80.0	77.9	1.027
SO2 Scrubber Check	4921	79.4	793.9	0.1	----
Date of last scrubber change:		N/A		Ave Corr Factor	1.018
Date of last converter efficiency test:		N/A			

Notes: Zero adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	November 24, 2025	Previous Calibration:	October 20, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	9:35	End Time (MST):	14:12
Analyzer make:	Thermo 450i	Analyzer serial #:	1218153583

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999974	≥0.995
80.0	79.2	1.0106	Slope	0.989933	0.90 - 1.10
40.0	39.2	1.0197	Intercept	-0.159925	+/-3
20.0	19.6	1.0221			


H₂S Calibration Curve



H₂S Calibration Plot

Date: November 24, 2025

Location: Wapasu





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Wapasu
Calibration Date: November 20, 2025
Start time (MST): 11:34
Reason: Routine
Station number: AMS17
Last Cal Date: October 15, 2025
End time (MST): 14:50

Calibration Standards

Gas Cert Reference: CC422255
CH4 Cal Gas Conc. 499.6 ppm
C3H8 Cal Gas Conc. 203.4 ppm
Removed Gas Cert: n/a
Removed CH4 Conc. 499.6 ppm
Removed C3H8 Conc. 203.4 ppm
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API 701H
Cal Gas Expiry Date: April 9, 2033
CH4 Equiv Conc. 1059.0 ppm
Removed Gas Expiry: 1059.0 ppm
CH4 Equiv Conc. 1059.0 ppm
Diff between cyl:
Serial Number: 2449
Serial Number: 1238

Analyzer Information

Analyzer make: Thermo 51i-LT
Analyzer Range: 0 - 20 ppm
Analyzer serial #: 1218153352

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000596	1.005532	Background:	3.200
Calibration intercept:	-0.100133	-0.141145	Coefficient:	4.376

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	-0.10	----
As found High point	4920	79.8	16.90	16.86	0.997
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	16.95	Previous response	16.81	*% change	0.8%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	-0.08	----
High point	4920	79.8	16.90	16.90	1.000
Mid point	4960	39.9	8.45	8.28	1.021
Low point	4980	20.0	4.24	4.09	1.035
As left zero	5000	0.0	0.00	-1.02	----
As left span	4921	79.4	16.81	16.80	1.001

Average Correction Factor 1.019

Notes:

Span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 15, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:34	End Time (MST):	14:50
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1218153352

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.08	----	Correlation Coefficient	0.999920	≥ 0.995
16.90	16.90	1.0001	Slope	1.005532	$0.90 - 1.10$
8.45	8.28	1.0210	Intercept	-0.141145	± 1.5
4.24	4.09	1.0351			


THC Calibration Curve



THC Calibration Plot

Date: November 20, 2025

Location: Wapasu





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Wapasu
Station number: AMS 17
Calibration Date: November 18, 2025
Last Cal Date: October 30, 2025
Start time (MST): 10:35
End time (MST): 15:43
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0045177
NOX Cal Gas Conc: 61.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 61.30 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: July 19, 2032
NO Cal Gas Conc: 61.00 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 61.00 ppm
NO gas Diff:
Serial Number: 2449
Serial Number: 359

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.1	-0.3	----	----
AF High point	4934	65.6	804.3	800.4	3.9	773.3	769.0	4.3	1.0396	1.0407
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 803.2 ppb	NO = 799.5 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = -3.8%
Baseline Corr 1st pt	NO _x = 773.7 ppb	NO = 769.1 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -4.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo Scientific 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153460

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.855	0.885	NO bkgnd or offset:	2.9	3.0
NOX coeff or slope:	0.997	0.997	NOX bkgnd or offset:	3.2	3.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	195.1	198.1

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000389	1.001654
NO _x Cal Offset:	-1.400017	-2.120010
NO Cal Slope:	1.002210	1.002467
NO Cal Offset:	-2.640011	-3.180009
NO ₂ Cal Slope:	0.999619	0.999452
NO ₂ Cal Offset:	-0.134470	-1.155064

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.3	0.0	-0.3	----	----
High point	4934	65.6	804.3	800.4	3.9	804.6	800.9	3.8	0.9997	0.9994
Mid point	4967	32.8	402.1	400.2	2.0	399.3	396.0	3.3	1.0071	1.0105
Low point	4983	16.4	201.1	200.1	1.0	197.8	194.5	3.3	1.0166	1.0288
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	0.0	-0.3	----	----
As left span	4934	65.6	804.3	392.8	411.5	803.0	392.8	409.9	1.0016	1.0000
Average Correction Factor									1.0078	1.0129

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.3	----	----
High GPT point	796.6	395.2	405.3	404.5	1.0021	99.8%
Mid GPT point	796.6	598.2	202.3	200.3	1.0102	99.0%
Low GPT point	796.6	699.2	101.3	99.5	1.0185	98.2%
Average Correction Factor					1.0102	99.0%

Notes:

Span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 30, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:35	End Time (MST):	15:43
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999976	≥0.995
804.3	804.6	0.9997	Slope	1.001654	0.90 - 1.10
402.1	399.3	1.0071	Intercept	-2.120010	+/-20
201.1	197.8	1.0166			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 30, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:35	End Time (MST):	15:43
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999978	≥0.995
405.3	404.5	1.0021	Slope	0.999452	0.90 - 1.10
202.3	200.3	1.0102	Intercept	-1.155064	+/-20
101.3	99.5	1.0185			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 30, 2025
Station Name:	Wapasu	Station Number:	AMS 17
Start Time (MST):	10:35	End Time (MST):	15:43
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1218153460

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999928	≥ 0.995
800.4	800.9	0.9994	Slope	1.002467	$0.90 - 1.10$
400.2	396.0	1.0105	Intercept	-3.180009	± 20
200.1	194.5	1.0288			



NO_x Calibration Plot

Date: November 18, 2025

Location: Wapasu





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Wapasu Station number: AMS17
Calibration Date: November 5, 2025 Last Cal Date: October 2, 2025
Start time (MST): 11:39 End time (MST): 14:33
Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: API T700 Serial Number: 2449
ZAG Make/Model: API T701H Serial Number: 359

Analyzer Information

Analyzer make: Thermo Scientific 49i Analyzer serial #: 1501663734
Analyzer Range: 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.010714	1.006029	Backgd or Offset:	0.1	0.1
Calibration intercept:	-1.700000	-1.380000	Coeff or Slope:	1.043	1.043

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.5	----
As found High point	5000	1104.7	400.0	403.1	0.991
As found Mid point					
As found Low point					
Baseline Corr As found:	403.6	Previous response	402.6	*% change	0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.4	----
High point	5000	1104.7	400.0	401.5	0.996
Mid point	5000	917.3	200.0	199.4	1.003
Low point	5000	797.9	100.0	98.2	1.018
As left zero	5000	0.0	0.0	-0.4	----
As left span	5000	1104.7	400.0	407.9	0.981
Average Correction Factor					1.006

Notes: No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	November 5, 2025	Previous Calibration:	October 2, 2025
Station Name:	Wapasu	Station Number:	AMS17
Start Time (MST):	11:39	End Time (MST):	14:33
Analyzer make:	Thermo Scientific 49i	Analyzer serial #:	1501663734

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	0.999973	≥0.995
400.0	401.5	0.9963	Slope	1.006029	0.90 - 1.10
200.0	199.4	1.0030	Intercept	-1.380000	+/- 5
100.0	98.2	1.0183			



O₃ Calibration Plot

Date: November 5, 2025

Location: Wapasu





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Wapasu Station number: AMS 17
Calibration Date: November 24, 2025 Last Cal Date: October 30, 2025
Start time (MST): 13:44 End time (MST): 14:10

Analyzer Make: Teledyne API T640 S/N: 1183
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-5.60	-5.90	-5.60	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	715.80	716.50	715.80	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.99	5.02	4.99	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	38	----	38	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 7.5		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 6, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: September 25, 2025
Date Disposable Filter Changed: October 30, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: July 21, 2025
Date RH/T Sensor Cleaned: July 21, 2025

Notes:

Monthly checks done.

Calibration by: Aswin Sasi Kumar



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Station Information

Station Name:	Wapasu	Station Number:	AMS 17
Calibration Date:	November 18, 2025	Prev Cal Date:	October 21, 2024
Start Time (MST):	10:47	End Time (MST):	13:13
Tower Height (m):	9.5	Reason:	Routine

Wind Speed Calibration

Sensor make/model:	Met One 010C-1	Serial Number:	N14664
WS Calibrator:	MetOne 053	Serial Number:	CA 05230

Shaft RPM (Hz)	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.1	-0.3%
400	39.4	39.1	-0.7%
600	58.6	58.9	0.6%
800	77.8	77.5	-0.3%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)		0.999968	≥ 0.9995
Calculated slope		1.000443	$0.98 - 1.02$
Calculated intercept		0.028279	± 2

Wind Direction Calibration

Sensor make/model:	Met One 020C-1	Serial Number:	P19942
As Found Declination (deg east of True North):		As Left Declination (deg east of True North):	
Solar noon (MST):	12:09	Calc Declination*:	13.09 Degrees
WD Calibrator:	Met One 040		<small>* - calculated declination as per NOAA website</small>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 360° FS) <i>Limit = +/- 1%</i>
10	9.3	-0.2%
90	90.2	0.1%
180	180.0	0.0%
270	270.2	0.1%
350	350.8	0.2%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)		0.999998	≥ 0.9995
Calculated slope		0.996562	$0.97 - 1.03$
Calculated intercept		0.519161	± 5

Notes: No issues with bearings on both sensors.

Calibration Performed By: Aswin Sasi Kumar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS18 STONY MOUNTAIN NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: November 26, 2025 Last Cal Date: October 23, 2025
Start time (MST): 10:25 End time (MST): 14:05
Reason: Routine

Calibration Standards

Cal Gas Concentration: 51.22 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC417455
Removed Cal Gas Conc: 51.22 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: CC417455 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 2658
Zero Air Gen Model: Teledyne API 701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43i Serial Number: JC1501301453
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000714	1.001599	Backgd or Offset:	25.9	25.9
Calibration intercept:	-1.637815	-1.597660	Coeff or Slope:	0.816	0.816

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4921	78.1	800.2	796.8	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	796.8	Previous response	799.1	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4921	78.1	800.2	800.9	0.999
Mid point	4960	39.1	400.6	398.6	1.005
Low point	4981	19.5	199.7	196.5	1.016
As left zero	5000	0.0	0.0	0.3	----
As left span	4921	78.1	800.2	799.7	1.001
Average Correction Factor:					1.007

Notes: Changed the inlet filter after as founds. No adjustments made.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 26, 2025	Previous Calibration:	October 23, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:25	End Time (MST):	14:05
Analyzer make:	Thermo 43i	Analyzer serial #:	JC1501301453

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999972	≥0.995
800.2	800.9	0.9991	Slope	1.001599	0.90 - 1.10
400.6	398.6	1.0050	Intercept	-1.597660	+/-30
199.7	196.5	1.0165			



SO2 Calibration Plot

Date: November 26, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS18
Calibration Date: November 27, 2025 Last Cal Date: October 29, 2025
Start time (MST): 10:30 End time (MST): 15:46
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.86 ppm Cal Gas Exp Date: May 9, 2027
Cal Gas Cylinder #: CC523103
Removed Cal Gas Conc: 4.86 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
ZAG Make/Model: Teledyne API T701 Serial Number: 4890

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1218153359
Converter make: CD Nova CDN-101 Converter serial #: 555
Analyzer Range: 0 - 100 ppb Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.992077	0.998224	Backgd or Offset:	2.94	2.94
Calibration intercept:	-0.018635	-0.058902	Coeff or Slope:	1.181	1.181

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4917	82.3	80.0	82.5	0.971
As found Mid point	4958	41.2	40.1	41.2	0.975
As found Low point	4979	20.6	20.0	20.3	0.991
New cylinder response					
Baseline Corr As found:	82.4	Prev response:	79.35	*% change:	3.7%
Baseline Corr 2nd AF pt:	41.1	AF Slope:	1.031511	AF Intercept:	-0.099658
Baseline Corr 3rd AF pt:	20.2	AF Correlation:	0.999970	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4917	82.3	80.0	80.0	1.000
Mid point	4958	41.2	40.1	39.7	1.009
Low point	4979	20.6	20.0	19.6	1.022
As left zero	5000	0.0	0.0	0.5	----
As left span	4917	82.3	80.0	79.4	1.008
SO2 Scrubber Check	4923	77.1	771.0	0.1	----
Date of last scrubber change:	17-Dec-21		Ave Corr Factor		1.010
Date of last converter efficiency test:					

Notes: No adjustment made. SOX scrubber tested. No issues to note.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	November 27, 2025	Previous Calibration:	October 29, 2025
Station Name:	Stony Mountain	Station Number:	AMS18
Start Time (MST):	10:30	End Time (MST):	15:46
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1218153359

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999907		≥ 0.995
80.0	80.0	1.0001	Slope	0.998224		$0.90 - 1.10$
40.1	39.7	1.0089	Intercept	-0.058902		± 3
20.0	19.6	1.0217				


TRS Calibration Curve



TRS Calibration Plot

Date: November 27, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Stony Mountain
Calibration Date: November 26, 2025
Start time (MST): 10:25
Reason: Routine

Station number: AMS 18
Last Cal Date: October 23, 2025
End time (MST): 14:04

Calibration Standards

Gas Cert Reference:	XC026809B	Cal Gas Expiry Date:	January 12, 2029
CH ₄ Cal Gas Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
C ₃ H ₈ Cal Gas Conc.	207.9 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	504.9 ppm	CH ₄ Equiv Conc.	1076.6 ppm
Removed C ₃ H ₈ Conc.	207.9 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700	Serial Number:	2658
Zero Air Gen model:	Teledyne API T701	Serial Number:	4890

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1170050130
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.68E-04	2.67E-04	NMHC SP Ratio:	5.04E-05
CH ₄ Retention time:	14.8	14.8	NMHC Peak Area:	177024
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.02	----
As found High point	4921	78.1	16.82	16.91	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.89	Prev response	16.83	*% change	0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.02	----
High point	4921	78.1	16.82	16.75	1.004
Mid point	4960	39.1	8.42	8.33	1.011
Low point	4981	19.5	4.20	4.13	1.018
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	16.82	16.59	1.014
Average Correction Factor					1.011

Notes: Changed the inlet filter and the H2 cylinder after as founds. Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	78.1	8.93	9.04	0.988
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.04	Prev response	8.95	*% change	1.0%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	78.1	8.93	8.89	1.005
Mid point	4960	39.1	4.47	4.42	1.012
Low point	4981	19.5	2.23	2.19	1.020
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	8.93	8.82	1.013
Average Correction Factor					1.012

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.02	Limit = 0.90-1.10
As found High point	4921	78.1	7.89	7.87	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.85	Prev response	7.89	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.02	----
High point	4921	78.1	7.89	7.87	1.003
Mid point	4960	39.1	3.95	3.91	1.010
Low point	4981	19.5	1.97	1.94	1.015
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	78.1	7.89	7.78	1.014
Average Correction Factor					1.009

Calibration Statistics

	Start	Finish
THC Cal Slope:	1.001524	0.996109
THC Cal Offset:	-0.012726	-0.026123
CH ₄ Cal Slope:	0.999248	0.996234
CH ₄ Cal Offset:	0.004620	-0.005977
NMHC Cal Slope:	1.003865	0.995793
NMHC Cal Offset:	-0.018145	-0.019347

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 26, 2025	Previous Calibration:	October 23, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:25	End Time (MST):	14:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.02	----	Correlation Coefficient	0.999970	≥ 0.995
16.82	16.75	1.0039	Slope	0.996109	$0.90 - 1.10$
8.42	8.33	1.0114	Intercept	-0.026123	± 0.5
4.20	4.13	1.0178			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 26, 2025	Previous Calibration:	October 23, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:25	End Time (MST):	14:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			Limits
0.00	0.02	----	Correlation Coefficient	0.999965		≥0.995
7.89	7.87	1.0029	Slope	0.996234		0.90 - 1.10
3.95	3.91	1.0097	Intercept	-0.005977		+/-0.5
1.97	1.94	1.0154				






Wood Buffalo Environmental Association NMHC Calibration Summary

Station Information

Calibration Date:	November 26, 2025	Previous Calibration:	October 23, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:25	End Time (MST):	14:04
Analyzer make:	Thermo 55i	Analyzer serial #:	1170050130

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999977	<i>≥0.995</i>
8.93	8.89	1.0051	Slope	0.995793	<i>0.90 - 1.10</i>
4.47	4.42	1.0124	Intercept	-0.019347	<i>+/-0.5</i>
2.23	2.19	1.0199			



NMHC Calibration Plot

Date: November 26, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Stony Mountain
Station number: AMS 18
Calibration Date: November 6, 2025
Last Cal Date: October 31, 2025
Start time (MST): 12:10
End time (MST): 17:21
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0045516
NOX Cal Gas Conc: 60.30 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 60.30 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API 701
Cal Gas Expiry Date: November 17, 2026
NO Cal Gas Conc: 60.10 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 60.10 ppm
NO gas Diff:
Serial Number: 2658
Serial Number: 4890

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero										
AF High point										
AF Mid point										
AF Low point										
New cyl resp										
Previous Respo 4933	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation		*Percent Change		NO _x =	NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>		*Percent Change		NO =	NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:		Nx Int:	
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:		NO Int:	
					As found	NO ₂ r ² :	NO2 SI:		NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo Scientific 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1501663731

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.059	1.117	NO bkgnd or offset:	10.4	11.2
NOX coeff or slope:	0.999	1.000	NOX bkgnd or offset:	10.6	14.0
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	218.9	219.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999094	0.999010
NO _x Cal Offset:	-0.868433	-2.488659
NO Cal Slope:	0.999139	1.000380
NO Cal Offset:	-1.969984	-2.329568
NO ₂ Cal Slope:	0.998329	1.001371
NO ₂ Cal Offset:	0.878481	-1.144685

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.8	0.0	-0.8	----	----
High point	4933	66.6	803.3	800.6	2.7	801.0	799.9	1.3	1.0028	1.0009
Mid point	4967	33.3	401.6	400.2	1.3	397.3	396.4	0.9	1.0108	1.0097
Low point	4983	16.6	200.2	199.5	0.7	196.2	195.3	0.9	1.0204	1.0218
As left zero	5000	0.0	0.0	0.0	0.0	-2.3	0.0	-2.3	----	----
As left span	4933	66.6	803.3	387.7	415.6	801.3	387.7	413.6	1.0024	1.0000
Average Correction Factor									1.0113	1.0108

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.8	----	----
High GPT point	796.3	384.6	414.4	413.9	1.0011	99.9%
Mid GPT point	796.3	596.6	202.4	201.5	1.0043	99.6%
Low GPT point	796.3	696.4	102.6	101.1	1.0145	98.6%
Average Correction Factor					1.0066	99.3%

Notes:

No asfound done. Mode valve replaced. Zero and span adjusted.

Calibration Performed By:

Aswin Sasi Kumar



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 31, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	12:10	End Time (MST):	17:21
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.8	----	Correlation Coefficient	0.999978	≥0.995
803.3	801.0	1.0028	Slope	0.999010	0.90 - 1.10
401.6	397.3	1.0108	Intercept	-2.488659	+/-20
200.2	196.2	1.0204			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 31, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	12:10	End Time (MST):	17:21
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1501663731

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.8	----	Correlation Coefficient	0.999996	≥0.995
414.4	413.9	1.0011	Slope	1.001371	0.90 - 1.10
202.4	201.5	1.0043	Intercept	-1.144685	+/-20
102.6	101.1	1.0145			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 31, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	12:10	End Time (MST):	17:21
Analyzer make:	Thermo Scientific 42i	Analyzer serial #:	1501663731

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999960	≥ 0.995
800.6	799.9	1.0009	Slope	1.000380	$0.90 - 1.10$
400.2	396.4	1.0097	Intercept	-2.329568	± 20
199.5	195.3	1.0218			



NO_x Calibration Plot

Date: November 6, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: November 25, 2025 Last Cal Date: October 22, 2025
Start time (MST): 10:10 End time (MST): 13:32
Reason: Routine

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700 Serial Number: 2658
ZAG Make/Model: Teledyne API 701H Serial Number: 4890

Analyzer Information

Analyzer make: API T400 Analyzer serial #: 825
Analyzer Range: 0 - 500 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995086	0.999286	Backgd or Offset:	2.3	3.1
Calibration intercept:	-0.040000	0.500000	Coeff or Slope:	0.991	1.010

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	NA	0.0	0.9	----
As found High point	4888	1138.1	400.0	394.2	1.017
As found Mid point					
As found Low point					
Baseline Corr As found:	393.3	Previous response	398.0	*% change	-1.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	NA	0.0	0.2	----
High point	4888	1138.1	400.0	399.8	1.001
Mid point	4888	884.5	200.0	201.3	0.994
Low point	4888	741.4	100.0	100.2	0.998
As left zero	5000	NA	0.0	0.5	----
As left span	4812	1097.9	400.0	403.0	0.993
Average Correction Factor					0.997

Notes: Zero and span adjusted.

Calibration Performed By: Aswin Sasi Kumar



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	November 25, 2025	Previous Calibration:	October 22, 2025
Station Name:	Stony Mountain	Station Number:	AMS 18
Start Time (MST):	10:10	End Time (MST):	13:32
Analyzer make:	API T400	Analyzer serial #:	825

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999986	≥0.995
400.0	399.8	1.0005	Slope	0.999286	0.90 - 1.10
200.0	201.3	0.9935	Intercept	0.500000	+/- 5
100.0	100.2	0.9980			



O₃ Calibration Plot

Date: November 25, 2025

Location: Stony Mountain





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Stony Mountain Station number: AMS 18
Calibration Date: November 27, 2025 Last Cal Date: October 31, 2025
Start time (MST): 14:23 End time (MST): 15:47

Analyzer Make: API T640 S/N: 324
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

<u>Parameter</u>	<u>As found</u>	<u>Measured</u>	<u>As left</u>	<u>Adjusted</u>	<u>(Limits)</u>
T (°C)	-9.7	-8.7	-9.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	710.8	699.80	710.8	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	4.98	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	38	----	38	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 3.0		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 10, 2024
Lot No.: 100128-050-042

<u>Parameter</u>	<u>As found</u>	<u>Post maintenance</u>	<u>As left</u>	<u>Adjusted</u>	<u>(Limits)</u>
PMT Peak Test	10.1	10.1	11.1	<input checked="" type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: November 27, 2025
Date Disposable Filter Changed: October 31, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: September 23, 2025
Date RH/T Sensor Cleaned: September 23, 2025

Notes: Flow, pressure and temp checked. Leak check passed. Quarterly cleaning done. PMT adjusted.

Calibration by: Aswin Sasi Kumar



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS19 FIREBAG NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Firebag Station number: AMS 19
Calibration Date: November 13, 2025 Last Cal Date: October 6, 2025
Start time (MST): 12:07 End time (MST): 15:16
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.97 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC705799
Removed Cal Gas Conc: 50.97 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 1607
Zero Air Gen Model: Teledyne API T701H Serial Number: 201

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1410661308
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002734	0.999159	Backgd or Offset:	11.8	13.3
Calibration intercept:	0.740000	0.140000	Coeff or Slope:	1.015	1.001

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	4999	0.0	0.0	1.4	----
As found High point	4922	78.4	799.2	812.0	0.986
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	810.6	Previous response	802.1	*% change	1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	4999	0.0	0.0	-0.1	----
High point	4922	78.4	799.2	798.7	1.001
Mid point	4961	39.2	399.6	399.1	1.001
Low point	4980	19.6	199.8	200.3	0.998
As left zero	4999	0.0	0.0	-0.1	----
As left span	4922	78.4	799.2	801.0	0.998
Average Correction Factor:					1.000

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	November 13, 2025	Previous Calibration:	October 6, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	12:07	End Time (MST):	15:16
Analyzer make:	Thermo 43i	Analyzer serial #:	1410661308

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999999	≥0.995
799.2	798.7	1.0006	Slope	0.999159	0.90 - 1.10
399.6	399.1	1.0013	Intercept	0.140000	+/-30
199.8	200.3	0.9975			

SO₂ Calibration Curve



SO2 Calibration Plot

Date: November 13, 2025

Location: Firebag





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Firebag
Calibration Date: November 10, 2025
Start time (MST): 10:41
Reason: Routine

Station number: AMS 19
Last Cal Date: October 7, 2025
End time (MST): 16:32

Calibration Standards

Cal Gas Concentration: 5.29 ppm
Cal Gas Cylinder #: DT0010492
Removed Cal Gas Conc: 5.29 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701

Cal Gas Exp Date: March 19, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 1607
Serial Number: 201

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1151680032
Converter serial #: 2022-222
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.980615	0.995046	Backgd or Offset:	2.66	2.87
Calibration intercept:	0.060000	0.080000	Coeff or Slope:	1.190	1.211

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4924	75.6	80.0	79.7	1.005
As found Mid point	4962	37.8	40.0	40.2	0.997
As found Low point	4981	18.9	20.0	20.1	1.000
New cylinder response					
Baseline Corr As found:	79.6	Prev response:	78.49	*% change:	1.4%
Baseline Corr 2nd AF pt:	40.1	AF Slope:	0.995189	AF Intercept:	0.200000
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999983	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4924	75.6	80.0	79.7	1.004
Mid point	4962	37.8	40.0	39.7	1.007
Low point	4981	18.9	20.0	20.2	0.990
As left zero	5000	0.0	0.0	0.1	----
As left span	4924	75.6	80.0	78.5	1.019
SO2 Scrubber Check				0.1	
Date of last scrubber change:		18-Jan-23		Ave Corr Factor	1.000
Date of last converter efficiency test:		November 26, 2024		106.2%	efficiency

Notes: Changed sample inlet filter after as founds. Adjusted zero and span. Scrubber check completed after as left zero.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	November 10, 2025	Previous Calibration:	October 7, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:41	End Time (MST):	16:32
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680032

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999975	≥ 0.995
80.0	79.7	1.0036	Slope	0.995046	$0.90 - 1.10$
40.0	39.7	1.0074	Intercept	0.080000	± 3
20.0	20.2	0.9899			



H2S Calibration Plot

Date: November 10, 2025

Location: Firebag





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Firebag Station number: AMS 19
Calibration Date: November 13, 2025 Last Cal Date: October 6, 2025
Start time (MST): 12:07 End time (MST): 15:16
Reason: Routine

Calibration Standards

Gas Cert Reference: CC705799 Cal Gas Expiry Date: October 9, 2032
CH4 Cal Gas Conc. 505.1 ppm CH4 Equiv Conc. 1066.9 ppm
C3H8 Cal Gas Conc. 204.3 ppm
Removed Gas Cert: Removed Gas Expiry:
Removed CH4 Conc. 505.1 ppm CH4 Equiv Conc. 1066.9 ppm
Removed C3H8 Conc. 204.3 ppm Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 1607
ZAG Make/Model: Teledyne API T701H Serial Number: 201

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1336160089
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995149	1.005458	Background:	2.16
Calibration intercept:	-0.015933	-0.096135	Coefficient:	3.860

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	4999	0.0	0.00	-0.10	Limit = 0.90-1.10
As found High point	4922	78.4	16.73	16.47	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	16.57	Previous response	16.63	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	4999	0.0	0.00	-0.08	Limit = 0.95-1.05
High point	4922	78.4	16.73	16.74	0.999
Mid point	4961	39.2	8.36	8.28	1.010
Low point	4980	19.6	4.18	4.12	1.016
As left zero	4999	0.0	0.00	-0.12	----
As left span	4922	78.4	16.73	16.73	1.000
Average Correction Factor					1.009

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	November 13, 2025	Previous Calibration:	October 6, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	12:07	End Time (MST):	15:16
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1336160089

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.08	----	Correlation Coefficient	0.999990	≥ 0.995
16.73	16.74	0.9994	Slope	1.005458	$0.90 - 1.10$
8.36	8.28	1.0102	Intercept	-0.096135	± 1.5
4.18	4.12	1.0162			


THC Calibration Curve



THC Calibration Plot

Date: November 13, 2025

Location: Firebag





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Firebag
Station number: AMS 19
Calibration Date: November 12, 2025
Last Cal Date: October 29, 2025
Start time (MST): 10:50
End time (MST): 16:02
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0044018
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.70 ppm
NO gas Diff:
Serial Number: 1607
Serial Number: 201

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.1	----	----
AF High point	4918	82.1	802.9	799.7	3.3	806.0	800.0	6.0	0.9960	0.9993
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 802.8 ppb	NO = 799.1 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.4%	
Baseline Corr 1st pt	NO _x = 806.2 ppb	NO = 800.2 ppb				<u>As Found Statistics</u>		*Percent Change	NO = 0.1%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1410661309

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.922	0.922	NO bkgnd or offset:	4.6	4.6
NOX coeff or slope:	0.993	0.993	NOX bkgnd or offset:	4.6	4.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	163.0	163.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998175	0.999427
NO _x Cal Offset:	1.319743	1.179966
NO Cal Slope:	0.998416	1.000087
NO Cal Offset:	0.719909	0.459895
NO ₂ Cal Slope:	1.003548	0.999118
NO ₂ Cal Offset:	-0.371604	-0.454709

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.2	0.1	----	----
High point	4918	82.1	802.9	799.7	3.3	803.0	799.8	2.9	0.9999	0.9998
Mid point	4959	41.1	402.0	400.3	1.6	403.6	401.3	2.3	0.9959	0.9975
Low point	4980	20.5	200.5	199.7	0.8	202.8	200.7	2.1	0.9886	0.9949
As left zero	5000	0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
As left span	4918	82.1	802.9	393.6	409.3	805.0	393.6	411.7	0.9974	1.0000
Average Correction Factor									0.9948	0.9974

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	799.1	391.8	410.6	410.0	1.0014	99.9%
Mid GPT point	799.1	597.2	205.2	204.4	1.0038	99.6%
Low GPT point	799.1	700.4	102.0	100.8	1.0117	98.8%
Average Correction Factor					1.0057	99.4%

Notes:

Changed sample inlet filter after as founds. No adjustments made.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 12, 2025	Previous Calibration:	October 29, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:50	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999988	≥0.995
802.9	803.0	0.9999	Slope	0.999427	0.90 - 1.10
402.0	403.6	0.9959	Intercept	1.179966	+/-20
200.5	202.8	0.9886			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 12, 2025	Previous Calibration:	October 29, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:50	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999991	≥0.995
410.6	410.0	1.0014	Slope	0.999118	0.90 - 1.10
205.2	204.4	1.0038	Intercept	-0.454709	+/-20
102.0	100.8	1.0117			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 12, 2025	Previous Calibration:	October 29, 2025
Station Name:	Firebag	Station Number:	AMS 19
Start Time (MST):	10:50	End Time (MST):	16:02
Analyzer make:	Thermo 42i	Analyzer serial #:	1410661309

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999997	≥ 0.995
799.7	799.8	0.9998	Slope	1.000087	$0.90 - 1.10$
400.3	401.3	0.9975	Intercept	0.459895	± 20
199.7	200.7	0.9949			



NO_x Calibration Plot

Date: November 12, 2025

Location: Firebag





Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Station Information

Station Name: Firebag Station Number: AMS 19
Calibration Date: November 12, 2025 Prev Cal Date: October 9, 2024
Start Time (MST): 11:33 End Time (MST): 15:40 Removal
Tower Height (m): 10.0 Reason:

Wind Speed Calibration

Sensor make/model: Met One 010C-1 Serial Number: W15276
WS Calibrator: MetOne 053 Serial Number: CA 03845

Shaft RPM (Hz)	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0		---
200	20.2		
400	39.4		
600	58.6		
800	77.8		

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	1.000000		≥ 0.9995
Calculated slope	1.000759		$0.98 - 1.02$
Calculated intercept	-0.016136		± 2

Wind Direction Calibration

Sensor make/model: Met One 020C-1 Serial Number: P22885
As Found Declination (deg east of True North): 13 As Left Declination (deg east of True North): 13
Solar noon (MST): 13:10 Calc Declination*: 13 Degrees
WD Calibrator: Met One 040
* - calculated declination as per NOAA website

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 360° FS) <i>Limit = +/- 1%</i>
10	16.5	1.8%
90	92.2	0.6%
180	183.2	0.9%
270	284.3	4.0%
350	368.7	5.2%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	0.999977	0.999547	≥ 0.9995
Calculated slope	1.001824	0.958186	$0.97 - 1.03$
Calculated intercept	-1.849285	-1.085588	± 5

Notes: Verified WD sensor, removed due to inaccurate readings.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Station Information

Station Name:	Firebag	Station Number:	AMS 19
Calibration Date:	November 12, 2025	Prev Cal Date:	October 9, 2024
Start Time (MST):	11:33	End Time (MST):	15:40
Tower Height (m):	10.0	Reason:	Routine

Wind Speed Calibration

Sensor make/model:	Met One 010C-1	Serial Number:	W15276
WS Calibrator:	MetOne 053	Serial Number:	CA 03845

Shaft RPM (Hz)	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.4%
400	39.4	39.4	0.1%
600	58.6	58.6	0.1%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	1.000000	1.000000	≥ 0.9995
Calculated slope	1.000759	0.999307	$0.98 - 1.02$
Calculated intercept	-0.016136	-0.025223	± 2

Wind Direction Calibration

Sensor make/model:	Met One 020C-1	Serial Number:	P22885
As Found Declination (deg east of True North):	<u>13</u>	As Left Declination (deg east of True North):	<u>13</u>
Solar noon (MST):	13:10	Calc Declination*:	13 Degrees
WD Calibrator:	Met One 040	* - calculated declination as per NOAA website	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 360° FS) <i>Limit = +/- 1%</i>
10	10.6	0.2%
90	92.2	0.6%
180	182.0	0.6%
270	274.9	1.4%
350	353.2	0.9%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)	0.999977	0.999974	≥ 0.9995
Calculated slope	1.001824	0.990825	$0.97 - 1.03$
Calculated intercept	-1.849285	-0.912774	± 5

Notes: Replaced wind direction sensor. Verified WS readings. Aligned with previous marking and compass.

Calibration Performed By: Braiden Boutilier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS20 MACKAY RIVER NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: MacKay River Station number: AMS 20
Calibration Date: November 10, 2025 Last Cal Date: October 16, 2025
Start time (MST): 7:26 End time (MST): 10:06
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.15 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC409669
Removed Cal Gas Conc: 49.15 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 5706
Zero Air Gen Model: API 701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1501301450
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995748	0.997334	Backgd or Offset:	20.4	20.4
Calibration intercept:	0.303078	-0.376941	Coeff or Slope:	0.932	0.932

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4919	81.4	800.1	796.6	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	796.6	Previous response	797.0	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4919	81.4	800.1	797.9	1.003
Mid point	4959	40.7	400.1	398.3	1.005
Low point	4980	20.3	199.5	198.1	1.007
As left zero	5000	0.0	0.0	0.1	----
As left span	4919	81.4	800.1	798.4	1.002
Average Correction Factor:					1.005

Notes: Changed Averaging time from 60 to 90 sec. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 10, 2025	Previous Calibration:	October 16, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	7:26	End Time (MST):	10:06
Analyzer make:	Thermo 43i	Analyzer serial #:	1501301450

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999998	≥0.995
800.1	797.9	1.0028	Slope	0.997334	0.90 - 1.10
400.1	398.3	1.0045	Intercept	-0.376941	+/-30
199.5	198.1	1.0073			



SO2 Calibration Plot

Date: November 10, 2025

Location: MacKay River





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: MacKay River Station number: AMS 20
Calibration Date: November 6, 2025 Last Cal Date: October 15, 2025
Start time (MST): 6:55 End time (MST): 10:52
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.12 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC515997
Removed Cal Gas Conc: 5.12 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 5706
ZAG Make/Model: API 701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 43i TLE Analyzer serial #: 1236656117
Converter make: Global Converter serial #: 2022-226
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003598	1.006030	Backgd or Offset:	3.77
Calibration intercept:	-0.020530	0.019450	Coeff or Slope:	1.071

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4922	78.1	80.0	81.5	0.980
As found Mid point	4961	39.0	39.9	40.7	0.979
As found Low point	4980	19.5	20.0	20.2	0.984
New cylinder response					
Baseline Corr As found:	81.6	Prev response:	80.24	*% change:	1.7%
Baseline Corr 2nd AF pt:	40.8	AF Slope:	1.020747	AF Intercept:	-0.120224
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999998	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4922	78.1	80.0	80.5	0.993
Mid point	4961	39.0	39.9	40.1	0.996
Low point	4980	19.5	20.0	20.2	0.989
As left zero	5000	0.0	0.0	0.1	----
As left span	4922	78.1	80.0	79.8	1.002
SO2 Scrubber Check	4982	81.3	802.8	0.0	----
Date of last scrubber change:	25-May-23			Ave Corr Factor	0.993
Date of last converter efficiency test:					

Notes: Sox scrubber checked after the calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 15, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	6:55	End Time (MST):	10:52
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1236656117

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999995	≥ 0.995
80.0	80.5	0.9935	Slope	1.006030	$0.90 - 1.10$
39.9	40.1	0.9959	Intercept	0.019450	± 3
20.0	20.2	0.9886			


H₂S Calibration Curve



H₂S Calibration Plot

Date: November 6, 2025

Location: MacKay River





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: MacKay River Station number: AMS 20
Calibration Date: November 10, 2025 Last Cal Date: October 16, 2025
Start time (MST): 7:28 End time (MST): 10:05
Reason: Routine

Calibration Standards

Gas Cert Reference: CC409669 Cal Gas Expiry Date: October 9, 2032
CH4 Cal Gas Conc. 505.1 ppm CH4 Equiv Conc. 1072.7 ppm
C3H8 Cal Gas Conc. 206.4 ppm
Removed Gas Cert: Removed Gas Expiry:
Removed CH4 Conc. 505.1 ppm CH4 Equiv Conc. 1072.7 ppm
Removed C3H8 Conc. 206.4 ppm Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 5706
ZAG Make/Model: API 701 Serial Number: 4888

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1501663727
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994146	0.994780	Background:	2.990	2.990
Calibration intercept:	0.003372	-0.067227	Coefficient:	5.000	5.000

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	-0.01	----
As found High point	4919	81.4	17.46	17.38	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.39	Previous response	17.36	*% change	0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	-0.03	----
High point	4919	81.4	17.46	17.32	1.008
Mid point	4959	40.7	8.73	8.61	1.015
Low point	4980	20.3	4.35	4.23	1.030
As left zero	5000	0.0	0.00	-0.12	----
As left span	4919	81.4	17.46	17.39	1.004
Average Correction Factor					1.018

Notes: No Maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	November 10, 2025	Previous Calibration:	October 16, 2025
Station Name:	MacKay River	Station Number:	AMS 20
Start Time (MST):	7:28	End Time (MST):	10:05
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1501663727

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.03	----	Correlation Coefficient	0.999983	≥ 0.995
17.46	17.32	1.0083	Slope	0.994780	$0.90 - 1.10$
8.73	8.61	1.0146	Intercept	-0.067227	± 1.5
4.35	4.23	1.0298			


THC Calibration Curve



THC Calibration Plot

Date: November 10, 2025

Location: MacKay River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: MacKay River
Station number: AMS 20
Calibration Date: November 7, 2025
Last Cal Date: October 9, 2025
Start time (MST): 8:14
End time (MST): 12:14
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0037393
NOX Cal Gas Conc: 62.00 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 62.00 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: July 22, 2032
NO Cal Gas Conc: 61.90 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 61.90 ppm
NO gas Diff:
Serial Number: 5706
Serial Number: 4888

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.1	0.3	-0.2	----	----
AF High point	4935	64.6	801.1	799.8	1.3	786.8	792.7	-5.9	1.0183	1.0094
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 800.5 ppb	NO = 800.5 ppb								NO _x = -1.8%
Baseline Corr 1st pt	NO _x = 786.7 ppb	NO = 792.4 ppb								NO = -1.0%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb								
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb								
As Found Statistics										
						As found NO _x r ² :				
						As found NO r ² :				
						As found NO ₂ r ² :				

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1505164379

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.033	1.043	NO bkgnd or offset:	2.9	2.9
NOX coeff or slope:	0.988	1.000	NOX bkgnd or offset:	3.1	3.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	148.6	148.6

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998314	0.997714
NO _x Cal Offset:	0.773119	0.433494
NO Cal Slope:	1.001682	0.997982
NO Cal Offset:	-0.666357	0.353813
NO ₂ Cal Slope:	0.994991	1.005007
NO ₂ Cal Offset:	-1.666155	-0.185055

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.2	----	----
High point	4935	64.6	801.1	799.8	1.3	799.3	798.3	1.0	1.0023	1.0019
Mid point	4968	32.3	400.5	399.9	0.6	400.9	400.0	0.8	0.9990	0.9996
Low point	4984	16.2	200.9	200.5	0.3	200.7	200.3	0.4	1.0009	1.0012
As left zero	5000	0.0	0.0	0.0	0.0	0.1	0.2	-0.1	----	----
As left span	4935	64.6	801.1	416.4	384.7	802.6	416.4	386.2	0.9981	1.0000
Average Correction Factor									1.0007	1.0009

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	794.2	414.8	380.7	382.5	0.9953	100.5%
Mid GPT point	794.2	602.5	193.0	193.5	0.9974	100.3%
Low GPT point	794.2	697.4	98.1	98.6	0.9948	100.5%
Average Correction Factor					0.9958	100.4%

Notes: No Maintenance done. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 9, 2025
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	8:14	End Time (MST):	12:14
Analyzer make:	Thermo 42i		1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999997	≥0.995
801.1	799.3	1.0023	Slope	0.997714	0.90 - 1.10
400.5	400.9	0.9990	Intercept	0.433494	+/-20
200.9	200.7	1.0009			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 9, 2025
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	8:14	End Time (MST):	12:14
Analyzer make:	Thermo 42i	6:50:00 AM	1505164379

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999998	≥0.995
380.7	382.5	0.9953	Slope	1.005007	0.90 - 1.10
193.0	193.5	0.9974	Intercept	-0.185055	+/-20
98.1	98.6	0.9948			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 9, 2025
Station Name:	Mackay River	Station Number:	AMS 20
Start Time (MST):	8:14	End Time (MST):	12:14
Analyzer make:	Thermo 42i	6:50:00 AM	1505164379

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999999	≥ 0.995
799.8	798.3	1.0019	Slope	0.997982	$0.90 - 1.10$
399.9	400.0	0.9996	Intercept	0.353813	± 20
200.5	200.3	1.0012			



NO_x Calibration Plot

Date: November 7, 2025

Location: MacKay River





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS21
CONKLIN
NOVEMBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Conklin
Calibration Date: November 4, 2025
Start time (MST): 10:51
Reason: Routine

Station number: AMS 21
Last Cal Date: October 15, 2025
End time (MST): 13:51

Calibration Standards

Cal Gas Concentration: 50.34 ppm
Cal Gas Cylinder #: CC340840
Removed Cal Gas Conc: 50.34 ppm
Removed Gas Cyl #: NA
Calibrator Model: Teledyne API T700P
Zero Air Gen Model: Teledyne API T701H

Cal Gas Exp Date: October 9, 2032
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 2656
Serial Number: 355

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0 - 1000 ppb

Serial Number: 1428701363

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994307	1.002589	Backgd or Offset:	29.9	30.2
Calibration intercept:	2.718110	2.316463	Coeff or Slope:	0.891	0.905

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4921	79.5	800.3	790.3	1.013
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	790.4	Previous response	798.5	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4921	79.5	800.3	804.0	0.995
Mid point	4960	39.8	400.7	404.0	0.992
Low point	4980	19.9	200.4	206.2	0.972
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	79.5	800.3	804.0	0.995

Average Correction Factor: 0.986

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 4, 2025	Previous Calibration:	October 15, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:51	End Time (MST):	13:51
Analyzer make:	Thermo 43i	Analyzer serial #:	1428701363

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999959	≥0.995
800.3	804.0	0.9954	Slope	1.002589	0.90 - 1.10
400.7	404.0	0.9919	Intercept	2.316463	+/-30
200.4	206.2	0.9717			



SO2 Calibration Plot

Date: November 4, 2025

Location: Conklin





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Conklin
Calibration Date: November 26, 2025
Start time (MST): 10:13
Reason: Routine

Station number: AMS 21
Last Cal Date: October 7, 2025
End time (MST): 14:04

Calibration Standards

Cal Gas Concentration: 5.14 ppm
Cal Gas Cylinder #: CC501204
Removed Cal Gas Conc: 5.14 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne T700P
ZAG Make/Model: Teledyne T701H

Cal Gas Exp Date: January 3, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 2656
Serial Number: 355

Analyzer Information

Analyzer make: Thermo 43i-QTL
Converter make: CD-Nova 101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12228021058
Converter serial #: 565
Converter Temp: 825 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994549	1.005185	Backgd or Offset:	2.8	3.3
Calibration intercept:	0.158394	-0.041619	Coeff or Slope:	1.470	1.494

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4922	78.4	80.6	79.3	1.020
As found Mid point	4961	39.2	40.3	40.0	1.015
As found Low point	4980	19.6	20.2	20.2	1.013
New cylinder response					
Baseline Corr As found:	79.0	Prev response:	80.31	*% change:	-1.7%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.979942	AF Intercept:	0.398417
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999991	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4922	78.4	80.6	81.0	0.995
Mid point	4961	39.2	40.3	40.4	0.997
Low point	4980	19.6	20.2	20.2	0.998
As left zero	5000	0.0	0.0	0.0	----
As left span	4922	78.4	80.6	80.8	0.997
SO2 Scrubber Check	4921	79.5	794.9	0.0	----
Date of last scrubber change:	August 6, 2025		Ave Corr Factor		0.997
Date of last converter efficiency test:	October 7, 2025				

Notes: Sample inlet filter was changed after multipoint as founds. SO2 scrubber check done and passed.
Adjusted zero and span.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	November 26, 2025	Previous Calibration:	October 7, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:13	End Time (MST):	14:04
Analyzer make:	Thermo 43i-QTL	Analyzer serial #:	12228021058

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999998	≥ 0.995
80.6	81.0	0.9949	Slope	1.005185	$0.90 - 1.10$
40.3	40.4	0.9974	Intercept	-0.041619	± 3
20.2	20.2	0.9975			


TRS Calibration Curve



TRS Calibration Plot

Date: November 26, 2025

Location: Conklin





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Conklin
 Calibration Date: November 4, 2025
 Start time (MST): 10:51
 Reason: Routine

Station number: AMS 21
 Last Cal Date: October 15, 2025
 End time (MST): 13:51

Calibration Standards

Gas Cert Reference:	CC340840	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	503.8 ppm	CH ₄ Equiv Conc.	1067.6 ppm
C ₃ H ₈ Cal Gas Conc.	205.0 ppm		
Removed Gas Cert:	NA	Removed Gas Expiry:	NA
Removed CH ₄ Conc.	503.8 ppm	CH ₄ Equiv Conc.	1067.6 ppm
Removed C ₃ H ₈ Conc.	205.0 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	Teledyne API T700P	Serial Number:	2656
Zero Air Gen model:	Teledyne API T701H	Serial Number:	355

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1180320039
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.34E-04	2.34E-04	NMHC SP Ratio:	4.97E-05
CH ₄ Retention time:	15.2	15.2	NMHC Peak Area:	180338
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				5.03E-05
				178091
				OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.5	16.97	16.90	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.90	Prev response	16.98	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.5	16.97	16.95	1.001
Mid point	4960	39.8	8.50	8.50	1.000
Low point	4980	19.9	4.25	4.34	0.978
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	16.97	16.98	1.000
Average Correction Factor					0.993

Notes:

Sample inlet filter was changed after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.5	8.96	8.85	1.013
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.85	Prev response	8.96	*% change	-1.2%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.5	8.96	8.96	1.001
Mid point	4960	39.8	4.49	4.50	0.997
Low point	4980	19.9	2.24	2.30	0.975
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	8.96	8.94	1.002
Average Correction Factor					0.991

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.5	8.01	8.04	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.04	Prev response	8.02	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.5	8.01	8.00	1.002
Mid point	4960	39.8	4.01	4.00	1.003
Low point	4980	19.9	2.01	2.04	0.982
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.5	8.01	8.03	0.997
Average Correction Factor					0.996

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.997750	0.997002
THC Cal Offset:	0.040887	0.041693
CH ₄ Cal Slope:	0.999822	0.996483
CH ₄ Cal Offset:	0.011577	0.014784
NMHC Cal Slope:	0.995912	0.997326
NMHC Cal Offset:	0.029509	0.026709

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 4, 2025	Previous Calibration:	October 15, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:51	End Time (MST):	13:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999960		≥ 0.995
16.97	16.95	1.0011	Slope	0.997002		0.90 - 1.10
8.50	8.50	0.9998	Intercept	0.041693		± 0.5
4.25	4.34	0.9783				





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 4, 2025	Previous Calibration:	October 15, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:51	End Time (MST):	13:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999964	<i>≥0.995</i>
8.01	8.00	1.0017	Slope	0.996483	<i>0.90 - 1.10</i>
4.01	4.00	1.0034	Intercept	0.014784	<i>+/-0.5</i>
2.01	2.04	0.9820			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 4, 2025	Previous Calibration:	October 15, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:51	End Time (MST):	13:51
Analyzer make:	Thermo 55i	Analyzer serial #:	1180320039

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999951	<i>≥0.995</i>
8.96	8.96	1.0006	Slope	0.997326	<i>0.90 - 1.10</i>
4.49	4.50	0.9970	Intercept	0.026709	<i>+/-0.5</i>
2.24	2.30	0.9751			



NMHC Calibration Plot

Date: November 4, 2025

Location: Conklin





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Conklin
Station number: AMS 21
Calibration Date: November 25, 2025
Last Cal Date: October 8, 2025
Start time (MST): 10:20
End time (MST): 14:52
Reason: Routine

Calibration Standards

NO Gas Cylinder #: SA18828
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700P
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: November 3, 2031
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 2659
Serial Number: 953

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.4	0.1	----	----
AF High point	4918	82.0	802.0	800.3	1.6	824.2	821.1	3.2	0.9727	0.9742
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 809.8 ppb	NO = 808.1 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = 1.8%			
Baseline Corr 1st pt	NO _x = 824.5 ppb	NO = 821.5 ppb	<u>As Found Statistics</u>			*Percent Change	NO = 1.6%			
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:			
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :			NO SI:	NO Int:			
			As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:			

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153356

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.648	0.631	NO bkgnd or offset:	3.7	3.7
NOX coeff or slope:	0.994	0.994	NOX bkgnd or offset:	3.6	3.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	163.8	165.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.006212	0.999101
NO _x Cal Offset:	2.868139	2.688082
NO Cal Slope:	1.007774	1.001134
NO Cal Offset:	1.568103	0.868050
NO ₂ Cal Slope:	0.999764	0.996440
NO ₂ Cal Offset:	0.063265	-0.836124

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	-0.2	0.2	----	----
High point	4918	82.0	802.0	800.3	1.6	803.0	802.0	1.6	0.9987	0.9979
Mid point	4959	41.0	401.0	400.2	0.8	403.4	400.7	2.8	0.9940	0.9987
Low point	4980	20.5	200.5	200.1	0.4	206.5	203.1	3.4	0.9708	0.9850
As left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
As left span	4918	82.0	802.0	389.7	412.3	800.5	389.7	410.9	1.0018	1.0000
Average Correction Factor									0.9878	0.9939

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	801.0	389.1	413.5	411.9	1.0040	99.6%
Mid GPT point	801.0	597.4	205.2	202.7	1.0125	98.8%
Low GPT point	801.0	697.9	104.7	102.8	1.0189	98.1%
Average Correction Factor					1.0118	98.8%

Notes:

Sample inlet filter was changed after as founds. Span adjustment made.

Calibration Performed By:

Parampreet Kaur



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 25, 2025	Previous Calibration:	October 8, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:20	End Time (MST):	14:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999942	≥0.995
802.0	803.0	0.9987	Slope	0.999101	0.90 - 1.10
401.0	403.4	0.9940	Intercept	2.688082	+/-20
200.5	206.5	0.9708			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 25, 2025	Previous Calibration:	October 8, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:20	End Time (MST):	14:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999968	≥ 0.995
413.5	411.9	1.0040	Slope	0.996440	$0.90 - 1.10$
205.2	202.7	1.0125	Intercept	-0.836124	± 20
104.7	102.8	1.0189			





Wood Buffalo Environmental Association

NO Calibration Summary


Station Information

Calibration Date:	November 25, 2025	Previous Calibration:	October 8, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	10:20	End Time (MST):	14:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153356

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999984	≥ 0.995
800.3	802.0	0.9979	Slope	1.001134	$0.90 - 1.10$
400.2	400.7	0.9987	Intercept	0.868050	± 20
200.1	203.1	0.9850			


NO Calibration Curve



NO_x Calibration Plot

Date: November 25, 2025

Location: Conklin





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Conklin
Calibration Date: November 5, 2025
Start time (MST): 11:36
Reason: Routine

Station number: AMS 21
Last Cal Date: October 2, 2025
End time (MST): 14:08

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700P
ZAG Make/Model: Teledyne API T701H

Serial Number: 2656
Serial Number: 355

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb

Analyzer serial #: 1300156233

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007229	1.007857	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	2.260000	1.300000	Coeff or Slope:	1.229	1.229

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	-0.1	----
As found High point	5000	1186.1	400.0	401.5	0.996
As found Mid point					
As found Low point					
Baseline Corr As found:	401.6	Previous response	405.2	*% change	-0.9%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.4	----
High point	5000	1185.2	400.0	404.0	0.990
Mid point	5000	940.0	200.0	203.3	0.984
Low point	5000	809.0	100.0	103.0	0.971
As left zero	5000	800.0	0.0	0.5	----
As left span	5000	1180.3	400.0	408.0	0.980
Average Correction Factor					0.982

Notes: Sample inlet filter was changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	November 5, 2025	Previous Calibration:	October 2, 2025
Station Name:	Conklin	Station Number:	AMS 21
Start Time (MST):	11:36	End Time (MST):	14:08
Analyzer make:	Thermo 49i	Analyzer serial #:	1300156233

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999977	≥0.995
400.0	404.0	0.9901	Slope	1.007857	0.90 - 1.10
200.0	203.3	0.9838	Intercept	1.300000	+/- 5
100.0	103.0	0.9709			



O₃ Calibration Plot

Date: November 5, 2025

Location: Conklin





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Conklin
Calibration Date: November 26, 2025
Start time (MST): 11:58
Station number: AMS 21
Last Cal Date: October 8, 2025
End time (MST): 12:53
Analyzer Make: API T640
Particulate Fraction: PM2.5
S/N: 1266
Flow Meter Make/Model: Alicat FP-25BT
Temp/RH standard: Alicat FP-25BT
S/N: 388754
S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-5.00	-4.79	-5.00	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	716.40	719.01	716.40	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.02	5.12	5.02	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	42	----	42	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 6.80		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: July 16, 2026
Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.70	10.90	10.90	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: November 26, 2025
Date Disposable Filter Changed: November 26, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.00 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: June 4, 2025
Date RH/T Sensor Cleaned: June 30, 2025

Notes: Verified flow, pressure, temperature and pump power. Leak check passed. No adjustment needed.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS22
JANVIER
NOVEMBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: November 7, 2025 Last Cal Date: October 24, 2025
Start time (MST): 12:56 End time (MST): 16:14
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.11 ppm Cal Gas Exp Date: January 18, 2029
Cal Gas Cylinder #: CC281519
Removed Cal Gas Conc: 50.11 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 3806
Zero Air Gen Model: Teledyne API T701 Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1152430006
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001164	0.997920	Backgd or Offset:	26.9	26.7
Calibration intercept:	0.064348	0.824696	Coeff or Slope:	1.010	1.019

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4920	79.8	799.8	790.5	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	790.4	Previous response	800.8	*% change	-1.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4920	79.8	799.8	798.6	1.001
Mid point	4960	39.9	399.9	400.5	0.998
Low point	4980	20.0	200.4	201.0	0.997
As left zero	5000	0.0	0.0	0.5	----
As left span	4920	79.8	799.8	799.5	1.000
Average Correction Factor:					0.999

Notes: Changed the inlet filter after as founds. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 24, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:56	End Time (MST):	16:14
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999998	≥0.995
799.8	798.6	1.0015	Slope	0.997920	0.90 - 1.10
399.9	400.5	0.9985	Intercept	0.824696	+/-30
200.4	201.0	0.9972			



SO2 Calibration Plot

Date: November 7, 2025

Location: Janvier





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Janvier
Calibration Date: November 19, 2025
Start time (MST): 12:24
Reason: Routine
Station number: AMS 22
Last Cal Date: October 28, 2025
End time (MST): 17:00

Calibration Standards

Cal Gas Concentration: 5.02 ppm
Cal Gas Cylinder #: CC424047
Removed Cal Gas Conc: 5.02 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701
Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3806
Serial Number: 691

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb
Analyzer serial #: 1151680031
Converter serial #: 620
Converter Temp: 850 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.961388	0.968816	Backgd or Offset:	3.39
Calibration intercept:	0.459964	0.500050	Coeff or Slope:	1.180

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4920	79.7	80.0	78.3	1.026
As found Mid point	4960	39.8	40.0	39.4	1.022
As found Low point	4980	19.9	20.0	19.7	1.030
New cylinder response					
Baseline Corr As found:	78.0	Prev response:	77.39	*% change:	0.8%
Baseline Corr 2nd AF pt:	39.1	AF Slope:	0.975245	AF Intercept:	0.300095
Baseline Corr 3rd AF pt:	19.4	AF Correlation:	0.999992	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4920	79.7	80.0	77.9	1.027
Mid point	4960	39.8	40.0	39.5	1.012
Low point	4980	19.9	20.0	19.8	1.009
As left zero	5000	0.0	0.0	0.7	----
As left span	4920	79.7	80.0	76.8	1.042
SO2 Scrubber Check	4920	79.8	798.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.016
Date of last converter efficiency test:					

Notes: Changed the inlet filter after as founds. Scrubber test performed, no issues. Adjusted the span.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	November 19, 2025	Previous Calibration:	October 28, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:24	End Time (MST):	17:00
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1151680031

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999966	≥ 0.995
80.0	77.9	1.0273	Slope	0.968816	$0.90 - 1.10$
40.0	39.5	1.0117	Intercept	0.500050	± 3
20.0	19.8	1.0091			


TRS Calibration Curve



TRS Calibration Plot

Date: November 19, 2025

Location: Janvier





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Janvier
 Calibration Date: November 7, 2025
 Start time (MST): 12:56
 Reason: Routine

Station number: AMS 22
 Last Cal Date: October 24, 2025
 End time (MST): 16:14

Calibration Standards

Gas Cert Reference: CC281519
 CH₄ Cal Gas Conc. 502.8 ppm
 C₃H₈ Cal Gas Conc. 208.4 ppm
 Removed Gas Cert: NA
 Removed CH₄ Conc. 502.8 ppm
 Removed C₃H₈ Conc. 208.4 ppm
 Diff between cyl (CH₄):
 Calibrator Model: Teledyne API 700
 Zero Air Gen model: Teledyne API 701

Cal Gas Expiry Date: January 18, 2029
 CH₄ Equiv Conc. 1075.9 ppm
 Removed Gas Expiry: NA
 CH₄ Equiv Conc. 1075.9 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 3806
 Serial Number: 691

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1317958219
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	2.58E-04	2.56E-04	NMHC SP Ratio:	6.19E-05
CH ₄ Retention time:	11.6	11.6	NMHC Peak Area:	147887
Zero Chromatogram:	OFF	OFF	Flat Baseline:	OFF
				148671

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4920	79.8	17.17	17.36	0.989
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.36	Prev response	17.11	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	17.17	17.19	0.999
Mid point	4960	39.9	8.59	8.54	1.006
Low point	4980	20.0	4.30	4.26	1.010
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	17.17	17.16	1.001
Average Correction Factor					1.005

Notes:

Changed the inlet filter after as founds. Adjusted the span.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	9.15	9.23	0.991
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.23	Prev response	9.10	*% change	1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	9.15	9.15	1.000
Mid point	4960	39.9	4.57	4.57	1.001
Low point	4980	20.0	2.29	2.29	1.001
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	9.15	9.13	1.002
Average Correction Factor					1.001

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4920	79.8	8.03	8.12	0.988
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.12	Prev response	8.00	*% change	1.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4920	79.8	8.03	8.04	0.998
Mid point	4960	39.9	4.01	3.97	1.012
Low point	4980	20.0	2.01	1.97	1.020
As left zero	5000	0.0	0.00	0.00	----
As left span	4920	79.8	8.03	8.03	0.999
Average Correction Factor					1.010

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.998025	1.001560
THC Cal Offset:	-0.031787	-0.030598
CH ₄ Cal Slope:	1.000840	1.002976
CH ₄ Cal Offset:	-0.029357	-0.028360
NMHC Cal Slope:	0.995131	0.999867
NMHC Cal Offset:	-0.001228	-0.001436

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 24, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:56	End Time (MST):	16:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999983	≥ 0.995
17.17	17.19	0.9990	Slope	1.001560	$0.90 - 1.10$
8.59	8.54	1.0057	Intercept	-0.030598	± 0.5
4.30	4.26	1.0102			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 24, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:56	End Time (MST):	16:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999932	≥0.995
8.03	8.04	0.9981	Slope	1.002976	0.90 - 1.10
4.01	3.97	1.0117	Intercept	-0.028360	+/-0.5
2.01	1.97	1.0204			






Wood Buffalo Environmental Association NMHC Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 24, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:56	End Time (MST):	16:14
Analyzer make:	Thermo 55i	Analyzer serial #:	1317958219

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	1.000000	<i>≥0.995</i>
9.15	9.15	1.0002	Slope	0.999867	<i>0.90 - 1.10</i>
4.57	4.57	1.0005	Intercept	-0.001436	<i>+/-0.5</i>
2.29	2.29	1.0015			



NMHC Calibration Plot

Date: November 7, 2025

Location: Janvier





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Janvier
Station number: AMS 22
Calibration Date: November 18, 2025
Last Cal Date: October 23, 2025
Start time (MST): 13:03
End time (MST): 17:17
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0047765
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: March 11, 2031
NO Cal Gas Conc: 48.80 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.80 ppm
NO gas Diff:
Serial Number: 3806
Serial Number: 691

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	0.0	----	----
AF High point	4918	82.0	802.0	800.3	1.6	805.6	796.7	9.0	0.9951	1.0043
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 801.8 ppb	NO = 798.3 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = 0.5%	
Baseline Corr 1st pt	NO _x = 805.9 ppb	NO = 796.9 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -0.2%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1229254994

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.012	1.012	NO bkgnd or offset:	2.8	2.8
NOX coeff or slope:	0.997	0.997	NOX bkgnd or offset:	2.9	2.9
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	172.9	173.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000138	1.002746
NO _x Cal Offset:	-0.255953	-0.095882
NO Cal Slope:	0.999346	1.001974
NO Cal Offset:	-1.536101	-1.656036
NO ₂ Cal Slope:	1.002701	1.001732
NO ₂ Cal Offset:	0.717283	0.338633

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
High point	4918	82.0	802.0	800.3	1.6	804.1	801.0	3.3	0.9973	0.9992
Mid point	4960	41.0	400.9	400.1	0.8	401.9	398.5	3.4	0.9975	1.0040
Low point	4980	20.5	200.5	200.1	0.4	200.8	197.2	3.7	0.9984	1.0145
As left zero	5000	0.0	0.0	0.0	0.0	0.0	-0.1	0.1	----	----
As left span	4918	82.0	802.0	400.1	401.9	802.0	400.1	401.9	1.0000	1.0000
Average Correction Factor									0.9977	1.0059

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.1	----	----
High GPT point	798.4	399.0	401.0	402.0	0.9976	100.2%
Mid GPT point	798.4	597.0	203.0	203.7	0.9968	100.3%
Low GPT point	798.4	696.4	103.6	104.5	0.9918	100.8%
Average Correction Factor					0.9954	100.5%

Notes:

Inlet filter was changed after as founds. No adjustments made.

Calibration Performed By:

Rene Chamberland



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 23, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	13:03	End Time (MST):	17:17
Analyzer make:	Thermo 42i	Analyzer serial #:	1229254994

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	1.000000	≥0.995
802.0	804.1	0.9973	Slope	1.002746	0.90 - 1.10
400.9	401.9	0.9975	Intercept	-0.095882	+/-20
200.5	200.8	0.9984			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 23, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	13:03	End Time (MST):	17:17
Analyzer make:	Thermo 42i	Analyzer serial #:	1229254994

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999998	≥ 0.995
401.0	402.0	0.9976	Slope	1.001732	$0.90 - 1.10$
203.0	203.7	0.9968	Intercept	0.338633	± 20
103.6	104.5	0.9918			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 23, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	13:03	End Time (MST):	17:17
Analyzer make:	Thermo 42i	Analyzer serial #:	1229254994

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999983	≥ 0.995
800.3	801.0	0.9992	Slope	1.001974	$0.90 - 1.10$
400.1	398.5	1.0040	Intercept	-1.656036	± 20
200.1	197.2	1.0145			



NO_x Calibration Plot

Date: November 18, 2025

Location: Janvier





Wood Buffalo Environmental Association

O₃ Calibration Report

Station Information

Station Name: Janvier
Calibration Date: November 13, 2025
Start time (MST): 12:47
Reason: Routine
Station number: AMS 22
Last Cal Date: October 17, 2025
End time (MST): 15:26

Calibration Standards

O₃ generation mode: Photometer
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H
Serial Number: 3806
Serial Number: 691

Analyzer Information

Analyzer make: Thermo 49i
Analyzer Range: 0 - 500 ppb
Analyzer serial #: 1227254861

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998829	0.999800	Backgd or Offset:	-0.2	-0.2
Calibration intercept:	0.980000	0.860000	Coeff or Slope:	1.553	1.553

O₃ As Found Data

Set Point	Dilution air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	800.0	0.0	0.1	----
As found High point	5000	926.2	400.0	398.7	1.004
As found Mid point					
As found Low point					
Baseline Corr As found:	398.6	Previous response	400.5	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

O₃ Calibration Data

Set Point	Total air flow rate (sccm)	Calibrator Lamp Voltage Drive (mV)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	800.0	0.0	0.2	----
High point	5000	926.2	400.0	400.4	0.999
Mid point	5000	768.9	200.0	201.3	0.994
Low point	5000	666.4	100.0	101.4	0.986
As left zero	5000	800.0	0.0	0.2	----
As left span	5000	926.2	400.0	405.0	0.988
Average Correction Factor					0.993

Notes: Changed the inlet filter after the as founds. No adjustments made.

Calibration Performed By: Rene Chamberland



Wood Buffalo Environmental Association


O₃ Calibration Summary

Station Information

Calibration Date:	November 13, 2025	Previous Calibration:	October 17, 2025
Station Name:	Janvier	Station Number:	AMS 22
Start Time (MST):	12:47	End Time (MST):	15:26
Analyzer make:	Thermo 49i	Analyzer serial #:	1227254861

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999987	≥0.995
400.0	400.4	0.9990	Slope	0.999800	0.90 - 1.10
200.0	201.3	0.9935	Intercept	0.860000	+/- 5
100.0	101.4	0.9862			



O₃ Calibration Plot

Date: November 13, 2025

Location: Janvier





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Janvier Station number: AMS 22
Calibration Date: November 18, 2025 Last Cal Date: October 28, 2025
Start time (MST): 13:45 End time (MST): 15:09

Analyzer Make: Teledyne API T640 S/N: 325
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388749
Temp/RH standard: Alicat FP-25BT S/N: 388749

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-0.1	-0.68	-0.1	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	715.6	716.52	715.6	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.01	4.988	5.01	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	40	----	40	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 3.6		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: October 6, 2024
Lot No.: 100128-050-042

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	11	10.9	10.9	<input type="checkbox"/>	10.9 +/- 0.5

Date Optical Chamber Cleaned: November 18, 2025
Date Disposable Filter Changed: November 18, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: May 28, 2025
Date RH/T Sensor Cleaned: November 18, 2025

Notes: Verified flow, temperature, and pressure. Leak checks passed. PMT peak tests completed. Optical chamber and RH/T sensor cleaned. Disposable filter changed.

Calibration by: Rene Chamberland



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS23 FORT HILLS NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Fort Hills Station number: AMS 23
Calibration Date: November 19, 2025 Last Cal Date: October 10, 2025
Start time (MST): 9:25 End time (MST): 12:06
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.35 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC484463
Removed Cal Gas Conc: 50.35 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 1222
Zero Air Gen Model: API T701 Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1160290012
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.995531	0.995831	Backgd or Offset:	18.5	18.5
Calibration intercept:	0.239920	-0.539973	Coeff or Slope:	1.063	1.063

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.5	----
As found High point	4921	79.4	799.5	794.5	1.007
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	794.0	Previous response	796.2	*% change	-0.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.5	----
High point	4921	79.4	799.5	796.7	1.004
Mid point	4960	39.7	399.8	395.4	1.011
Low point	4980	19.8	199.4	198.1	1.007
As left zero	5000	0.0	0.0	0.8	----
As left span	4921	79.4	799.5	797.1	1.003
Average Correction Factor:					1.007

Notes: No maintenance or adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 19, 2025	Previous Calibration:	October 10, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	9:25	End Time (MST):	12:06
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290012

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999980	≥0.995
799.5	796.7	1.0035	Slope	0.995831	0.90 - 1.10
399.8	395.4	1.0111	Intercept	-0.539973	+/-30
199.4	198.1	1.0065			



SO2 Calibration Plot

Date: November 19, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Fort Hills
Calibration Date: November 14, 2025
Start time (MST): 8:34
Reason: Routine

Station number: AMS 23
Last Cal Date: October 2, 2025
End time (MST): 12:02

Calibration Standards

Cal Gas Concentration: 4.84 ppm
Cal Gas Cylinder #: DT0021910
Removed Cal Gas Conc: 4.84 ppm
Removed Gas Cyl #:
Calibrator Make/Model: API T700
ZAG Make/Model: API T701

Cal Gas Exp Date: August 28, 2027
Rem Gas Exp Date:
Diff between cyl:
Serial Number: 1222
Serial Number: 1117

Analyzer Information

Analyzer make: Thermo 43i TLE
Converter make: CDN-101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1300156232
Converter serial #: 594
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.011534	0.998532	Backgd or Offset:	1.96	1.92
Calibration intercept:	-0.298114	-0.018119	Coeff or Slope:	0.990	0.965

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4917	82.6	80.0	81.8	0.978
As found Mid point	4959	41.3	40.0	40.7	0.982
As found Low point	4979	20.7	20.0	20.3	0.987
New cylinder response					
Baseline Corr As found:	81.8	Prev response:	80.59	*% change:	1.5%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.023546	AF Intercept:	-0.118469
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999990	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4917	82.6	80.0	79.9	1.001
Mid point	4959	41.3	40.0	39.7	1.007
Low point	4979	20.7	20.0	20.1	0.997
As left zero	5000	0.0	0.0	0.0	----
As left span	4917	82.6	80.0	80.5	0.993
SO2 Scrubber Check	4920	80.3	803.0	0.1	----
Date of last scrubber change:				Ave Corr Factor	1.002
Date of last converter efficiency test:		March 13, 2024		110.3% efficiency	

Notes: SOx scrubber checked after the calibrator zero. Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	November 14, 2025	Previous Calibration:	October 2, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	8:34	End Time (MST):	12:02
Analyzer make:	Thermo 43i TLE	Analyzer serial #:	1300156232

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999984	≥ 0.995
80.0	79.9	1.0008	Slope	0.998532	$0.90 - 1.10$
40.0	39.7	1.0070	Intercept	-0.018119	± 3
20.0	20.1	0.9970			


TRS Calibration Curve



TRS Calibration Plot

Date: November 14, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Fort Hills
 Calibration Date: November 19, 2025
 Start time (MST): 9:25
 Reason: Routine

Station number: AMS 23
 Last Cal Date: October 29, 2025
 End time (MST): 12:06

Calibration Standards

Gas Cert Reference:	CC484463	Cal Gas Expiry Date:	October 9, 2032
CH ₄ Cal Gas Conc.	504.3 ppm	CH ₄ Equiv Conc.	1065.6 ppm
C ₃ H ₈ Cal Gas Conc.	204.1 ppm		
Removed Gas Cert:		Removed Gas Expiry:	
Removed CH ₄ Conc.	504.3 ppm	CH ₄ Equiv Conc.	1065.6 ppm
Removed C ₃ H ₈ Conc.	204.1 ppm	Diff between cyl (THC):	
Diff between cyl (CH ₄):		Diff between cyl (NM):	
Calibrator Model:	API T700	Serial Number:	1222
Zero Air Gen model:	API T701	Serial Number:	1117

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 12227620777
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.54E-04	3.62E-04	NMHC SP Ratio:	5.68E-05
CH ₄ Retention time:	15.0	15.0	NMHC Peak Area:	157420
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				156181
				OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4921	79.4	16.92	16.92	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	16.92	Prev response	16.78	*% change	0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4921	79.4	16.92	16.93	1.000
Mid point	4960	39.7	8.46	8.32	1.016
Low point	4980	19.8	4.22	4.15	1.017
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	16.92	16.83	1.005
Average Correction Factor					1.011

Notes: Nitrogen Cylinder Changed out. Span adjusted.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.4	8.91	8.95	0.996
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.95	Prev response	8.83	*% change	1.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.4	8.91	8.89	1.003
Mid point	4960	39.7	4.46	4.41	1.011
Low point	4980	19.8	2.22	2.21	1.005
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	8.91	8.80	1.013
Average Correction Factor					1.006

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4921	79.4	8.01	7.97	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	7.97	Prev response	7.95	*% change	0.3%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.00	0.00	Limit = 0.95-1.05
High point	4921	79.4	8.01	8.04	0.997
Mid point	4960	39.7	4.00	3.91	1.023
Low point	4980	19.8	2.00	1.94	1.030
As left zero	5000	0.0	0.00	0.00	----
As left span	4921	79.4	8.01	8.03	0.998
Average Correction Factor					1.017

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.995176	1.000896
THC Cal Offset:	-0.062798	-0.056377
CH ₄ Cal Slope:	0.998720	1.004713
CH ₄ Cal Offset:	-0.047801	-0.046789
NMHC Cal Slope:	0.991991	0.997197
NMHC Cal Offset:	-0.014998	-0.009788

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 19, 2025	Previous Calibration:	October 29, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	9:25	End Time (MST):	12:06
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.00	----	Correlation Coefficient	0.999913	≥ 0.995
16.92	16.93	0.9996	Slope	1.000896	$0.90 - 1.10$
8.46	8.32	1.0165	Intercept	-0.056377	± 0.5
4.22	4.15	1.0166			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 19, 2025	Previous Calibration:	October 29, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	9:25	End Time (MST):	12:06
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999778	<i>≥0.995</i>
8.01	8.04	0.9966	Slope	1.004713	<i>0.90 - 1.10</i>
4.00	3.91	1.0231	Intercept	-0.046789	<i>+/-0.5</i>
2.00	1.94	1.0300			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 19, 2025	Previous Calibration:	October 29, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	9:25	End Time (MST):	12:06
Analyzer make:	Thermo 55i	Analyzer serial #:	12227620777

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999976	<i>≥0.995</i>
8.91	8.89	1.0025	Slope	0.997197	<i>0.90 - 1.10</i>
4.46	4.41	1.0113	Intercept	-0.009788	<i>+/-0.5</i>
2.22	2.21	1.0049			



NMHC Calibration Plot

Date: November 19, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Fort Hills
Station number: AMS 23
Calibration Date: November 20, 2025
Last Cal Date: October 20, 2025
Start time (MST): 7:27
End time (MST): 11:47
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC358149
NOX Cal Gas Conc: 60.30 ppm
Removed Cylinder #:
Removed Gas NOX Conc: 60.30 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.10 ppm
Removed Gas Exp Date:
Removed Gas NO Conc: 60.10 ppm
NO gas Diff:
Serial Number: 1222
Serial Number: 1117

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.2	----	----
AF High point	4934	66.3	799.5	796.9	2.7	812.9	806.2	6.6	0.9833	0.9883
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 804.2 ppb	NO = 802.1 ppb				* = > +/-5% change initiates investigation			*Percent Change	NO _x = 1.1%
Baseline Corr 1st pt	NO _x = 813.1 ppb	NO = 806.3 ppb				<u>As Found Statistics</u>			*Percent Change	NO = 0.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found NO _x r ² :			Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found NO r ² :			NO SI:	NO Int:
						As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1152430007

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.975	0.960	NO bkgnd or offset:	2.8	2.8
NOX coeff or slope:	0.992	0.993	NOX bkgnd or offset:	3.2	3.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	147.2	147.2

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.005887	0.998968
NO _x Cal Offset:	-0.015171	-0.093276
NO Cal Slope:	1.008514	0.998603
NO Cal Offset:	-1.513884	-1.331243
NO ₂ Cal Slope:	0.997104	1.001135
NO ₂ Cal Offset:	-0.955955	-0.460255

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.2	-0.2	----	----
High point	4934	66.3	799.5	796.9	2.7	799.0	795.7	3.4	1.0007	1.0015
Mid point	4967	33.2	400.4	399.0	1.3	398.8	394.9	3.8	1.0040	1.0105
Low point	4983	16.6	200.2	199.5	0.7	200.5	197.4	3.1	0.9986	1.0109
As left zero	5000	0.0	0.0	0.0	0.0	0.0	0.1	-0.1	----	----
As left span	4934	66.3	799.5	403.8	395.7	795.2	403.8	391.4	1.0054	1.0000
Average Correction Factor									1.0011	1.0076

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	790.9	401.2	392.4	392.5	0.9996	100.0%
Mid GPT point	790.9	594.7	198.9	198.4	1.0023	99.8%
Low GPT point	790.9	692.6	101.0	100.4	1.0055	99.5%
Average Correction Factor					1.0025	99.8%

Notes: Span adjusted. No maintenance done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 20, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:27	End Time (MST):	11:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999995	≥0.995
799.5	799.0	1.0007	Slope	0.998968	0.90 - 1.10
400.4	398.8	1.0040	Intercept	-0.093276	+/-20
200.2	200.5	0.9986			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 20, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:27	End Time (MST):	11:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999998	≥ 0.995
392.4	392.5	0.9996	Slope	1.001135	$0.90 - 1.10$
198.9	198.4	1.0023	Intercept	-0.460255	± 20
101.0	100.4	1.0055			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 20, 2025
Station Name:	Fort Hills	Station Number:	AMS 23
Start Time (MST):	7:27	End Time (MST):	11:47
Analyzer make:	Thermo 42i	Analyzer serial #:	1152430007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999973	≥ 0.995
796.9	795.7	1.0015	Slope	0.998603	$0.90 - 1.10$
399.0	394.9	1.0105	Intercept	-1.331243	± 20
199.5	197.4	1.0109			



NO_x Calibration Plot

Date: November 20, 2025

Location: Fort Hills





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Fort Hills Station number: AMS 23
Calibration Date: November 19, 2025 Last Cal Date: October 20, 2025
Start time (MST): 8:08 End time (MST): 9:18

Analyzer Make: API T640 S/N: 320
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388744
Temp/RH standard: Alicat FP-25BT S/N: 388744

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-5.9	-6.4	-5.9	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	732.7	731.8	732.7	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.00	5.03	5.00	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	78	----	75	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: _____	6.0	PM w/ HEPA: _____	0.0	<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: 16-Jul-26
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	8.9	10.7	10.7	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: November 19, 2025
Date Disposable Filter Changed: November 19, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: November 19, 2025
Date RH/T Sensor Cleaned: November 19, 2025

Notes: No adjustments done. Leak Check, Flow and PMT checked before and after cleaning.

Calibration by: Melissa Lemay



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS25 WASKŌW OHCI PIMÂTISIWIN NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS 25
Calibration Date: November 24, 2025 Last Cal Date: October 24, 2025
Start time (MST): 8:05 End time (MST): 11:12
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.70 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC342445
Removed Cal Gas Conc: 49.70 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: API T700 Serial Number: 621
Zero Air Gen Model: API T701 Serial Number: 4765

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1118148497
Analyzer Range: 0-1000ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997669	0.999156	Backgd or Offset:	11.1	11.6
Calibration intercept:	0.327490	-0.592733	Coeff or Slope:	1.046	1.056

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.5	----
As found High point	4920	80.5	800.1	791.2	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	790.7	Previous response	798.6	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.3	----
High point	4920	80.5	800.1	799.0	1.001
Mid point	4960	40.2	399.6	398.4	1.003
Low point	4980	20.1	199.8	198.8	1.005
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	80.5	800.1	799.4	1.001
Average Correction Factor:					1.003

Notes: No maintenance done. Zero and Span adjusted.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 24, 2025	Previous Calibration:	October 24, 2025
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS 25
Start Time (MST):	8:05	End Time (MST):	11:12
Analyzer make:	Thermo 43i	Analyzer serial #:	1118148497

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999999	≥0.995
800.1	799.0	1.0014	Slope	0.999156	0.90 - 1.10
399.6	398.4	1.0029	Intercept	-0.592733	+/-30
199.8	198.8	1.0050			



SO2 Calibration Plot

Date: November 24, 2025

Location: Waskow ohci Pimatisiwin





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Waskow ohci Pimatisiwin Station number: AMS 25
Calibration Date: November 18, 2025 Last Cal Date: October 27, 2025
Start time (MST): 6:42 End time (MST): 11:17
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.97 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC517099
Removed Cal Gas Conc: 4.97 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 747
ZAG Make/Model: API T701 Serial Number: 261

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050146
Converter make: Global G-150 Converter serial #: 2024-287
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.000081	1.004077	Backgd or Offset:	3.66
Calibration intercept:	0.182210	0.142303	Coeff or Slope:	1.129

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4919	80.5	80.0	81.0	0.988
As found Mid point	4960	40.3	40.1	40.7	0.984
As found Low point	4980	20.1	20.0	20.3	0.984
New cylinder response					
Baseline Corr As found:	81.0	Prev response:	80.21	*% change:	1.0%
Baseline Corr 2nd AF pt:	40.7	AF Slope:	1.012075	AF Intercept:	0.062230
Baseline Corr 3rd AF pt:	20.3	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4919	80.5	80.0	80.5	0.994
Mid point	4960	40.3	40.1	40.3	0.994
Low point	4980	20.1	20.0	20.3	0.984
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	80.0	800.0	817.0	0.979
SO2 Scrubber Check	4920	80.0	800.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	0.991
Date of last converter efficiency test:		February 12, 2025		111.0% efficiency	

Notes: Sox scrubber checked after calibrator zero. No adjustments done.

Calibration Performed By: Melissa Lemay



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 27, 2025
Station Name:	Waskow ohci Pimatisiwin	Station Number:	AMS 25
Start Time (MST):	6:42	End Time (MST):	11:17
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050146

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999996		≥0.995
80.0	80.5	0.9941	Slope	1.004077		0.90 - 1.10
40.1	40.3	0.9939	Intercept	0.142303		+/-3
20.0	20.3	0.9842				


H₂S Calibration Curve



H₂S Calibration Plot

Date: November 18, 2025

Location: Waskow ohci Pimatisiwin





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS27
JACKFISH 2/3
NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Jackfish 2/3 Station number: AMS 27
Calibration Date: November 3, 2025 Last Cal Date: October 3, 2025
Start time (MST): 11:34 End time (MST): 15:05
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.58 ppm Cal Gas Exp Date: December 29, 2028
Cal Gas Cylinder #: SG9133974BAL
Removed Cal Gas Conc: 50.58 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 5252
Zero Air Gen Model: API 701 Serial Number: 268

Analyzer Information

Analyzer make: Thermo 43iQ-TL Serial Number: 12124313138
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.001333	1.002161	Backgd or Offset:	9.3
Calibration intercept:	0.202353	0.362717	Coeff or Slope:	0.907

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4913	78.9	799.4	790.2	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	790.0	Previous response	800.7	*% change	-1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4913	78.9	799.4	801.6	0.997
Mid point	4955	39.5	400.0	400.9	0.998
Low point	4971	19.7	199.7	200.9	0.994
As left zero	5000	0.0	0.0	-0.1	----
As left span	4913	78.9	799.4	806.4	0.991
Average Correction Factor:					0.996

Notes: Changed the sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 3, 2025	Previous Calibration:	October 3, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	11:34	End Time (MST):	15:05
Analyzer make:	Thermo 43iQ-TL	Analyzer serial #:	12124313138

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999999	≥0.995
799.4	801.6	0.9973	Slope	1.002161	0.90 - 1.10
400.0	400.9	0.9978	Intercept	0.362717	+/-30
199.7	200.9	0.9938			



SO2 Calibration Plot

Date: November 3, 2025

Location: Jackfish 2/3





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Jackfish 2/3
Calibration Date: November 18, 2025
Start time (MST): 13:27
Reason: Routine

Station number: AMS 27
Last Cal Date: October 24, 2025
End time (MST): 22:36

Calibration Standards

Cal Gas Concentration: 4.87 ppm
Cal Gas Cylinder #: CC523090
Removed Cal Gas Conc: 4.87 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API T701H

Cal Gas Exp Date: September 5, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 5252
Serial Number: 268

Analyzer Information

Analyzer make: Thermo 43iQ
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12228021055
Converter serial #: 2022-195
Converter Temp: 325 degC

Calibration slope: Start 1.004522 Finish 1.003521
Calibration intercept: 0.196003 -0.043977

Backgd or Offset: Start 3.26 Finish 3.39
Coeff or Slope: 1.228 1.071

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4911	82.0	80.0	80.1	1.001
As found Mid point	4950	41.0	40.0	39.9	1.008
As found Low point	4972	20.5	20.0	20.1	1.005
New cylinder response					
Baseline Corr As found:	79.9	Prev response:	80.54	*% change:	-0.8%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.998949	AF Intercept:	0.116027
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999986	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4911	82.0	80.0	80.3	0.996
Mid point	4950	41.0	40.0	39.8	1.005
Low point	4972	20.5	20.0	20.3	0.985
As left zero	5000	0.0	0.0	0.1	----
As left span	4911	82.0	80.0	80.7	0.991
SO2 Scrubber Check	4915	78.9	790.0	-0.1	----
Date of last scrubber change:	18-Nov-25		Ave Corr Factor		0.995
Date of last converter efficiency test:	April 23, 2025		91.4% efficiency		

Notes: Changed sample inlet filter after as founds. Performed troubleshooting to investigate elevated "as-left" span values. As part of the corrective actions, replaced the valve tubing, scrubber tube, and scrubber beads, and ensured all fittings were properly tightened. The scrubber check was completed successfully. Adjusted both zero and span.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	November 18, 2025	Previous Calibration:	October 24, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	13:27	End Time (MST):	22:36
Analyzer make:	Thermo 43iQ	Analyzer serial #:	12228021055

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999949	≥0.995
80.0	80.3	0.9960	Slope	1.003521	0.90 - 1.10
40.0	39.8	1.0052	Intercept	-0.043977	+/-3
20.0	20.3	0.9851			



H₂S Calibration Plot

Date: November 18, 2025

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Jackfish 2/3
Station number: AMS 27
Calibration Date: November 12, 2025
Last Cal Date: October 23, 2025
Start time (MST): 13:16
End time (MST): 16:45
Reason: As Found

Calibration Standards

NO Gas Cylinder #: CC757838
NOX Cal Gas Conc: 60.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.30 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 60.20 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.20 ppm
NO gas Diff:
Serial Number: 5252
Serial Number: 268

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
AF High point	4924	66.3	801.1	799.8	1.3	738.0	732.5	5.6	1.0854	1.0919
AF Mid point	4958	33.2	401.1	400.4	0.7	369.5	365.3	4.1	1.0852	1.0962
AF Low point	4976	16.6	200.5	200.2	0.3	185.3	181.3	4.0	1.0814	1.1040
New cyl resp										
Previous Response	NO _x = 799.5 ppb	NO = 798.2 ppb	* = > +/-5% change initiates investigation				*Percent Change		NO _x = -8.3%	
Baseline Corr 1st pt	NO _x = 738.1 ppb	NO = 732.5 ppb	<u>As Found Statistics</u>				*Percent Change		NO = -9.0%	
Baseline Corr 2nd pt	NO _x = 369.6 ppb	NO = 365.3 ppb	As found NO _x r ² : 0.999999				Nx SI: 0.921038		Nx Int: 0.185	
Baseline Corr 3rd pt	NO _x = 185.4 ppb	NO = 181.3 ppb	As found NO r ² : 0.999990				NO SI: 0.916508		NO Int: -1.094	
			As found NO ₂ r ² : 0.999885				NO ₂ SI: 1.000631		NO ₂ Int: 1.472	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	-0.1	----	----
As found high GPT point	727.0	353.1	375.2	376.5	0.9966	100.3%
As found mid GPT point	727.0	532.4	195.9	197.1	0.9941	100.6%
As found low GPT point	727.0	626.7	101.6	105.6	0.9624	103.9%



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153357

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997189	
NO _x Cal Offset:	0.630204	
NO Cal Slope:	0.999286	
NO Cal Offset:	-1.049205	
NO ₂ Cal Slope:	0.999326	
NO ₂ Cal Offset:	0.407685	

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.524	1.524	NO bkgnd or offset:	6.1	6.1
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	6.2	6.2
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	179.7	179.7

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero
High point
Mid point
Low point
As left zero
As left span

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero
High GPT point
Mid GPT point
Low GPT point

Average Correction Factor

Notes:

As founds only.


Calibration Performed By:

Mohammed Kashif

NO_x Calibration Plot

Date: November 12, 2025

Location: Jackfish 2/3





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Jackfish 2/3
Station number: AMS 27
Calibration Date: November 13, 2025
Last Cal Date: October 23, 2025
Start time (MST): 8:56
End time (MST): 13:36
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC757838
NOX Cal Gas Conc: 60.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.30 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 60.20 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.20 ppm
NO gas Diff:
Serial Number: 5252
Serial Number: 268

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
-----------	------------------------------	--------------------------------	---	--	---	--	---	--	--	---

As found zero
AF High point
AF Mid point
AF Low point
New cyl resp

Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>		*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found	NO r ² :	NO SI:	NO Int:
					As found	NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	---	--	--	---	--	--

As Found GPT zero
As found high GPT point
As found mid GPT point
As found low GPT point



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1218153357

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.524	1.513	NO bkgnd or offset:	6.1	6.1
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	6.2	6.1
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	179.7	176.4

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.997189	0.998771
NO _x Cal Offset:	0.630204	0.050109
NO Cal Slope:	0.999286	0.998442
NO Cal Offset:	-1.049205	-1.228850
NO ₂ Cal Slope:	0.999326	1.004627
NO ₂ Cal Offset:	0.407685	1.152046

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	-0.1	0.2	----	----
High point	4924	66.3	801.1	799.8	1.3	800.3	798.1	2.3	1.0010	1.0021
Mid point	4958	33.2	401.1	400.4	0.7	400.4	397.4	3.0	1.0017	1.0076
Low point	4976	16.6	200.5	200.2	0.3	200.4	197.9	2.6	1.0005	1.0114
As left zero	5000	0.0	0.0	0.0	0.0	0.9	0.1	0.8	----	----
As left span	4924	66.3	801.1	381.8	419.3	797.7	381.8	415.9	1.0043	1.0000
Average Correction Factor									1.0011	1.0071

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.2	----	----
High GPT point	792.6	383.7	410.2	412.7	0.9940	100.6%
Mid GPT point	792.6	582.4	211.5	214.3	0.9871	101.3%
Low GPT point	792.6	679.5	114.4	117.0	0.9780	102.2%
Average Correction Factor					0.9864	101.4%

Notes:

Changed sample inlet filter and pump. Adjusted span only.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 13, 2025	Previous Calibration:	October 23, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	8:56	End Time (MST):	13:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	1.000000	≥0.995
801.1	800.3	1.0010	Slope	0.998771	0.90 - 1.10
401.1	400.4	1.0017	Intercept	0.050109	+/-20
200.5	200.4	1.0005			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 13, 2025	Previous Calibration:	October 23, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	8:56	End Time (MST):	13:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999973	≥0.995
410.2	412.7	0.9940	Slope	1.004627	0.90 - 1.10
211.5	214.3	0.9871	Intercept	1.152046	+/-20
114.4	117.0	0.9780			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 13, 2025	Previous Calibration:	October 23, 2025
Station Name:	Jackfish 2/3	Station Number:	AMS 27
Start Time (MST):	8:56	End Time (MST):	13:36
Analyzer make:	Thermo 42i	Analyzer serial #:	1218153357

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999989	≥ 0.995
799.8	798.1	1.0021	Slope	0.998442	$0.90 - 1.10$
400.4	397.4	1.0076	Intercept	-1.228850	± 20
200.2	197.9	1.0114			



NO_x Calibration Plot

Date: November 13, 2025

Location: Jackfish 2/3





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS29 SURMONT 2 NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: November 3, 2025 Last Cal Date: October 24, 2025
Start time (MST): 10:52 End time (MST): 14:01
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.95 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC356229
Removed Cal Gas Conc: 49.95 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 5472
Zero Air Gen Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1170050150
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996831	1.001461	Backgd or Offset:	14.5	14.5
Calibration intercept:	-0.740911	-1.461001	Coeff or Slope:	0.940	0.944

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.4	----
As found High point	4919.9	80.1	800.2	796.2	1.005
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	796.6	Previous response	796.9	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.3	----
High point	4920	80.1	800.2	801.0	0.999
Mid point	4960	40.0	399.6	397.2	1.006
Low point	4980	20.0	199.8	197.3	1.013
As left zero	5000	0.0	0.0	0.2	----
As left span	4920	80.1	800.2	802.0	0.998
Average Correction Factor:					1.006

Notes: Changed sample inlet filter and external pump after as founds. Adjusted span.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 3, 2025	Previous Calibration:	October 24, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:52	End Time (MST):	14:01
Analyzer make:	Thermo 43i	Analyzer serial #:	1170050150

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.3	----	Correlation Coefficient	0.999976	≥0.995
800.2	801.0	0.9990	Slope	1.001461	0.90 - 1.10
399.6	397.2	1.0060	Intercept	-1.461001	+/-30
199.8	197.3	1.0127			



SO2 Calibration Plot

Date: November 3, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: November 4, 2025 Last Cal Date: October 2, 2025
Start time (MST): 10:40 End time (MST): 15:48
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.750 ppm Cal Gas Exp Date: August 28, 2027
Cal Gas Cylinder #: CC737848
Removed Cal Gas Conc: 4.750 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 5472
ZAG Make/Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 43iQ-TLE Analyzer serial #: 1200326170
Converter make: Global Converter serial #: 2022-220
Analyzer Range: 0 - 100 ppb Converter Temp: 325.0 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996738	0.976306	Backgd or Offset:	0.82
Calibration intercept:	-0.080477	0.059532	Coeff or Slope:	1.040

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.0	----
As found High point	4916	84.2	80.0	79.6	1.005
As found Mid point	4958	42.1	40.0	39.7	1.007
As found Low point	4979	21.1	20.0	19.8	1.010
New cylinder response					
Baseline Corr As found:	79.6	Prev response:	79.65	*% change:	-0.1%
Baseline Corr 2nd AF pt:	39.7	AF Slope:	0.995452	AF Intercept:	-0.060477
Baseline Corr 3rd AF pt:	19.8	AF Correlation:	0.999997	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4916	84.2	80.0	78.2	1.023
Mid point	4958	42.1	40.0	39.0	1.025
Low point	4979	21.1	20.0	19.6	1.020
As left zero	5000	0.0	0.0	0.1	----
As left span	4916	84.2	80.0	77.6	1.031
SO2 Scrubber Check	4919	81.3	813.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.023
Date of last converter efficiency test:		December 5, 2024		108.1%	efficiency

Notes: Changed sample inlet filter after as founds. Ran SOx scrubber check after cal zero, passed. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H2S Calibration Summary


Station Information

Calibration Date:	November 4, 2025	Previous Calibration:	October 2, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:40	End Time (MST):	15:48
Analyzer make:	Thermo 43iQ-TLE	Analyzer serial #:	1200326170

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999995		≥ 0.995
80.0	78.2	1.0228	Slope	0.976306		$0.90 - 1.10$
40.0	39.0	1.0255	Intercept	0.059532		± 3
20.0	19.6	1.0203				


H₂S Calibration Curve



H2S Calibration Plot

Date: November 4, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: November 3, 2025 Last Cal Date: October 24, 2025
Start time (MST): 10:52 End time (MST): 14:01
Reason: Routine

Calibration Standards

Gas Cert Reference: CC356229 Cal Gas Expiry Date: October 9, 2032
CH4 Cal Gas Conc. 503.7 ppm CH4 Equiv Conc. 1066.9 ppm
C3H8 Cal Gas Conc. 204.8 ppm
Removed Gas Cert: NA Removed Gas Expiry: NA
Removed CH4 Conc. 503.7 ppm CH4 Equiv Conc. 1066.9 ppm
Removed C3H8 Conc. 204.8 ppm Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 5472
ZAG Make/Model: Teledyne API T701 Serial Number: 135

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1170050149
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.998037	0.994853	Background:	3.56	3.54
Calibration intercept:	-0.023967	0.015835	Coefficient:	4.113	4.117

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	-0.01	----
As found High point	4920	80.1	17.09	17.02	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.03	Previous response	17.03	*% change	0.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.04	----
High point	4920	80.1	17.09	17.03	1.004
Mid point	4960	40.0	8.54	8.50	1.004
Low point	4980	20.0	4.27	4.23	1.008
As left zero	5000	0.0	0.00	0.01	----
As left span	4920	80.1	17.09	17.09	1.000
Average Correction Factor					1.005

Notes: Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	November 3, 2025	Previous Calibration:	October 24, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:52	End Time (MST):	14:01
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1170050149

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.04	----	Correlation Coefficient	0.999991	≥ 0.995
17.09	17.03	1.0036	Slope	0.994853	$0.90 - 1.10$
8.54	8.50	1.0041	Intercept	0.015835	± 1.5
4.27	4.23	1.0079			


THC Calibration Curve



THC Calibration Plot

Date: November 3, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Surmont 2
Station number: AMS 29
Calibration Date: November 7, 2025
Last Cal Date: October 3, 2025
Start time (MST): 10:41
End time (MST): 15:55
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC218007
NOX Cal Gas Conc: 60.20 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.20 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 5472
Serial Number: 135

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0	0.0	0.0	0.0	-0.3	-0.1	-0.2	----	----
AF High point	4933	66.7	803.1	800.4	2.7	789.5	785.8	3.7	1.0169	1.0185
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 793.3 ppb	NO = 791.8 ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = -0.4%
Baseline Corr 1st pt	NO _x = 789.8 ppb	NO = 785.9 ppb	<u>As Found Statistics</u>	*Percent Change	NO = -0.7%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found NO r ² :	NO SI:	NO Int:
			As found NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1170050148

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.036	1.051	NO bkgnd or offset:	1.3	1.3
NOX coeff or slope:	0.992	0.991	NOX bkgnd or offset:	1.5	1.5
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	149.8	149.0

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.984898	0.999895
NO _x Cal Offset:	2.354041	-0.889026
NO Cal Slope:	0.989195	1.001300
NO Cal Offset:	-0.025951	-1.889053
NO ₂ Cal Slope:	1.001984	1.008426
NO ₂ Cal Offset:	0.056024	0.022619

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
High point	4933	66.7	803.1	800.4	2.7	802.0	800.0	1.8	1.0014	1.0006
Mid point	4967	33.3	400.9	399.6	1.3	401.2	398.9	2.3	0.9993	1.0017
Low point	4983	16.7	201.1	200.4	0.7	198.3	195.8	2.5	1.0140	1.0236
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
As left span	4933	66.7	803.1	411.3	391.8	803.0	411.3	392.2	1.0001	1.0000
Average Correction Factor									1.0049	1.0086

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	795.9	429.2	369.4	372.0	0.9929	100.7%
Mid GPT point	795.9	626.6	172.0	175.0	0.9827	101.8%
Low GPT point	795.9	710.3	88.3	88.1	1.0019	99.8%
Average Correction Factor					0.9925	100.8%

Notes:

Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 3, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:41	End Time (MST):	15:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999984	≥0.995
803.1	802.0	1.0014	Slope	0.999895	0.90 - 1.10
400.9	401.2	0.9993	Intercept	-0.889026	+/-20
201.1	198.3	1.0140			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 3, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:41	End Time (MST):	15:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999953	≥0.995
369.4	372.0	0.9929	Slope	1.008426	0.90 - 1.10
172.0	175.0	0.9827	Intercept	0.022619	+/-20
88.3	88.1	1.0019			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 3, 2025
Station Name:	Surmont 2	Station Number:	AMS 29
Start Time (MST):	10:41	End Time (MST):	15:55
Analyzer make:	Thermo 42i	Analyzer serial #:	1170050148

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999963	≥ 0.995
800.4	800.0	1.0006	Slope	1.001300	$0.90 - 1.10$
399.6	398.9	1.0017	Intercept	-1.889053	± 20
200.4	195.8	1.0236			



NO_x Calibration Plot

Date: November 7, 2025

Location: Surmont 2





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Surmont 2 Station number: AMS 29
Calibration Date: November 3, 2025 Last Cal Date: October 24, 2025
Start time (MST): 10:59 End time (MST): 12:19

Analyzer Make: API T640 S/N: 2236
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	4.7	4.43	4.7	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	707.3	706.03	707.3	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	5.03	4.968	5.03	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	52	----	33	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 1.4		PM w/ HEPA: 0.0		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.9 Expiry Date: July 16, 2026
Lot No.: 100128-050-050

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test	10.8	10.9	10.9	<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: November 3, 2025
Date Disposable Filter Changed: November 3, 2025

Post- maintenance Zero Verification: PM w/ HEPA: 0.0 <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: April 14, 2025
Date RH/T Sensor Cleaned: July 15, 2025

Notes: Verified temperature, pressure and flow. Leak check passed. Swapped pump after as founds.

Calibration by: Braiden Boutilier



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Station Information

Station Name:	Surmont	Station Number:	AMS 29
Calibration Date:	November 17, 2025	Prev Cal Date:	August 27, 2025
Start Time (MST):	11:45	End Time (MST):	14:07
Tower Height (m):	10.0	Reason:	Removal

Wind Speed Calibration

Sensor make/model:	Met One 010C-1	Serial Number:	W15275
WS Calibrator:	MetOne 053	Serial Number:	05231

Shaft RPM (Hz)	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.4%
400	39.4	39.4	0.1%
600	58.6	58.9	0.7%
800	77.8	77.8	0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)		0.999991	≥ 0.9995
Calculated slope		0.997542	$0.98 - 1.02$
Calculated intercept		-0.023819	± 2

Wind Direction Calibration

Sensor make/model:	Met One 020C-1	Serial Number:	U11347
As Found Declination (deg east of True North):	<u>13</u>	As Left Declination (deg east of True North):	<u>13</u>
Solar noon (MST):	12:25	Calc Declination*:	13 Degrees
WD Calibrator:	Met One 040		<small>* - calculated declination as per NOAA website</small>

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 360° FS) <i>Limit = +/- 1%</i>
10		
90		
180		
270		
350		

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)			≥ 0.9995
Calculated slope			$0.97 - 1.03$
Calculated intercept			± 5

Notes: Removing sensor due to inconsistent uptime. Replaced WS cable.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

Wind Speed/Direction Calibration Report

Station Information

Station Name:	Surmont	Station Number:	AMS 29
Calibration Date:	November 17, 2025	Prev Cal Date:	August 27, 2025
Start Time (MST):	11:45	End Time (MST):	14:07
Tower Height (m):	10.0	Reason:	Install

Wind Speed Calibration

Sensor make/model:	Met One 010C-1	Serial Number:	B4129
WS Calibrator:	MetOne 053	Serial Number:	05231

Shaft RPM (Hz)	Calculated Speed (K/hr) (Cv)	Indicated Speed (K/hr) (Iv)	% Error <i>Limit = +/- 1.5%</i>
0	0.0	0.0	---
200	20.2	20.3	0.4%
400	39.4	39.4	0.1%
600	58.6	58.5	0.0%
800	77.8	77.7	-0.1%

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)		0.999999	≥ 0.9995
Calculated slope		1.001155	$0.98 - 1.02$
Calculated intercept		-0.057643	± 2

Wind Direction Calibration

Sensor make/model:	Met One 020C-1	Serial Number:	U11347
As Found Declination (deg east of True North):	<u>13</u>	As Left Declination (deg east of True North):	<u>13</u>
Solar noon (MST):	12:25	Calc Declination*:	13 Degrees
WD Calibrator:	Met One 040	* - calculated declination as per NOAA website	

Physical Direction (Degrees) (Cv)	Indicated Direction (Degrees) (Iv)	% Error (based on 360° FS) <i>Limit = +/- 1%</i>
10		
90		
180		
270		
350		

	<u>Start</u>	<u>Finish</u>	<u>Limits</u>
Correl Coeff (r^2)			≥ 0.9995
Calculated slope			$0.97 - 1.03$
Calculated intercept			± 5

Notes: Installed new WS sensor due to inconsistent uptime. Aligned using previous marking and compass.

Calibration Performed By: Braiden Boutilier



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS30 ELLS RIVER NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Ells River Station number: AMS 30
Calibration Date: November 3, 2025 Last Cal Date: October 14, 2025
Start time (MST): 10:40 End time (MST): 13:37
Reason: Routine

Calibration Standards

Cal Gas Concentration: 48.75 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC350110
Removed Cal Gas Conc: 48.75 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: API T700 Serial Number: 3061
Zero Air Gen Model: API T701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1008841397
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.994713	0.994199	Backgd or Offset:	10.6	10.6
Calibration intercept:	-2.552091	-2.172107	Coeff or Slope:	0.991	0.991

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.3	----
As found High point	4918	82.0	799.5	791.6	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	791.9	Previous response	792.7	*% change	-0.1%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4918	82.0	799.5	793.6	1.007
Mid point	4959	41.0	399.8	394.8	1.013
Low point	4980	20.5	199.9	193.8	1.031
As left zero	5000	0.0	0.0	-0.3	----
As left span	4918	82.0	799.5	795.1	1.006
Average Correction Factor:					1.017

Notes: Sample inlet filter was changed after as founds. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 3, 2025	Previous Calibration:	October 14, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:40	End Time (MST):	13:37
Analyzer make:	Thermo 43i	Analyzer serial #:	1008841397

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999961	≥0.995
799.5	793.6	1.0074	Slope	0.994199	0.90 - 1.10
399.8	394.8	1.0125	Intercept	-2.172107	+/-30
199.9	193.8	1.0312			



SO2 Calibration Plot

Date: November 3, 2025

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Ells River Station number: AMS 30
Calibration Date: November 7, 2025 Last Cal Date: October 30, 2025
Start time (MST): 10:42 End time (MST): 14:32
Reason: Routine

Calibration Standards

Cal Gas Concentration: 4.99 ppm Cal Gas Exp Date: November 15, 2026
Cal Gas Cylinder #: CC505806
Removed Cal Gas Conc: 4.99 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 3061
ZAG Make/Model: API 701H Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i-LTE Analyzer serial #: 1170050152
Converter make: CDN- 101 Converter serial #: CDN606
Analyzer Range: 0 - 100 ppb Converter Temp: 860 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997762	1.002476	Backgd or Offset:	3.8
Calibration intercept:	-0.100519	-0.240504	Coeff or Slope:	1.140

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.5	----
As found High point	4920	80.2	80.0	73.2	1.086
As found Mid point	4960	40.1	40.0	33.4	1.181
As found Low point	4980	20.0	20.0	14.3	1.349
New cylinder response					
Baseline Corr As found:	73.7	Prev response:	79.76	*% change:	-8.2%
Baseline Corr 2nd AF pt:	33.9	AF Slope:	0.931975	AF Intercept:	-2.522708
Baseline Corr 3rd AF pt:	14.8	AF Correlation:	0.996587	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4920	80.2	80.0	80.1	0.999
Mid point	4960	40.1	40.0	39.7	1.008
Low point	4980	20.0	20.0	19.7	1.013
As left zero	5000	0.0	0.0	-0.2	----
As left span	4920	80.2	80.0	80.0	1.000
SO2 Scrubber Check	4918	82.0	820.0	0.0	----
Date of last scrubber change:	13-Aug-25			Ave Corr Factor	1.007
Date of last converter efficiency test:					

Notes: Sample inlet filter, capillary and o-ring was changed after multipoint as founds. SO2 scrubber check done and passed. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 30, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:42	End Time (MST):	14:32
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999981		≥ 0.995
80.0	80.1	0.9992	Slope	1.002476		$0.90 - 1.10$
40.0	39.7	1.0080	Intercept	-0.240504		± 3
20.0	19.7	1.0132				


TRS Calibration Curve



TRS Calibration Plot

Date: November 7, 2025

Location: Ells River





Wood Buffalo Environmental Association

TRS Calibration Report

Station Information

Station Name: Ells River
Calibration Date: November 17, 2025
Start time (MST): 11:35
Reason: Maintenance

Station number: AMS 30
Last Cal Date: November 7, 2025
End time (MST): 14:36

Calibration Standards

Cal Gas Concentration: 4.99 ppm
Cal Gas Cylinder #: CC505806
Removed Cal Gas Conc: 4.99 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: API T700
ZAG Make/Model: API 701H

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 43i-LTE
Converter make: CDN- 101
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1170050152
Converter serial #: 632
Converter Temp: 800 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002476	1.016613	Backgd or Offset:	4.1	4.13
Calibration intercept:	-0.240504	-0.560357	Coeff or Slope:	1.256	1.256

TRS As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero					
As found High point					
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	NA	Prev response:	NA	*% change:	NA
Baseline Corr 2nd AF pt:	NA	AF Slope:	NA	AF Intercept:	NA
Baseline Corr 3rd AF pt:	NA	AF Correlation:	NA	* = > +/-5% change initiates investigation	

TRS Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.2	----
High point	4920	80.2	80.0	81.0	0.988
Mid point	4960	40.1	40.0	39.9	1.003
Low point	4980	20.0	20.0	19.4	1.029
As left zero	5000	0.0	0.0	-0.3	----
As left span	4920	80.2	80.0	80.6	0.993
SO2 Scrubber Check	4918	82.0	820.0	0.0	----
Date of last scrubber change:	13-Aug-25		Ave Corr Factor		1.007
Date of last converter efficiency test:					

Notes: No MPAF as the converter is failing. Converter was changed before calibrator zero. SO2 scrubber check done and passed. No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

TRS Calibration Summary


Station Information

Calibration Date:	November 17, 2025	Previous Calibration:	November 7, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	11:35	End Time (MST):	14:36
Analyzer make:	Thermo 43i-LTE	Analyzer serial #:	1170050152

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999910		≥ 0.995
80.0	81.0	0.9881	Slope	1.016613		$0.90 - 1.10$
40.0	39.9	1.0030	Intercept	-0.560357		± 3
20.0	19.4	1.0289				


TRS Calibration Curve



TRS Calibration Plot

Date: November 17, 2025

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Ells River
 Calibration Date: November 3, 2025
 Start time (MST): 10:40
 Reason: Routine

Station number: AMS 30
 Last Cal Date: October 14, 2025
 End time (MST): 13:37

Calibration Standards

Gas Cert Reference: CC350110
 CH₄ Cal Gas Conc. 496.6 ppm
 C₃H₈ Cal Gas Conc. 207.2 ppm
 Removed Gas Cert: NA
 Removed CH₄ Conc. 496.6 ppm
 Removed C₃H₈ Conc. 207.2 ppm
 Diff between cyl (CH₄):
 Calibrator Model: API T700
 Zero Air Gen model: API T701H

Cal Gas Expiry Date: Monday, March 10, 2031
 CH₄ Equiv Conc. 1066.4 ppm
 Removed Gas Expiry: NA
 CH₄ Equiv Conc. 1066.4 ppm
 Diff between cyl (THC):
 Diff between cyl (NM):
 Serial Number: 3061
 Serial Number: 358

Analyzer Information

Analyzer make: Thermo 55i
 THC Range: 0 - 20 ppm

Analyzer serial #: 1152430011
 NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.12E-04	3.15E-04	NMHC SP Ratio:	5.89E-05
CH ₄ Retention time:	17.4	17.6	NMHC Peak Area:	157597
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF
				5.95E-05
				156872
				OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	17.49	17.31	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.31	Prev response	17.40	*% change	-0.5%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	82.0	17.49	17.45	1.002
Mid point	4959	41.0	8.74	8.65	1.011
Low point	4980	20.5	4.37	4.26	1.026
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	17.49	17.46	1.002
Average Correction Factor					1.013

Notes:

Sample inlet filter was changed after as founds. Adjusted span only.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4918	82.0	9.34	9.25	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.25	Prev response	9.32	*% change	-0.8%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	82.0	9.34	9.33	1.002
Mid point	4959	41.0	4.67	4.64	1.007
Low point	4980	20.5	2.34	2.30	1.014
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	9.34	9.33	1.002
Average Correction Factor					1.008

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	Limit = 0.90-1.10
As found High point	4918	82.0	8.14	8.05	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.05	Prev response	8.08	*% change	-0.4%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	0.00	----
High point	4918	82.0	8.14	8.13	1.002
Mid point	4959	41.0	4.07	4.01	1.015
Low point	4980	20.5	2.04	1.96	1.039
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	8.14	8.13	1.001
Average Correction Factor					1.019

Calibration Statistics

	Start	Finish
THC Cal Slope:	0.998290	0.999473
THC Cal Offset:	-0.060226	-0.056026
CH ₄ Cal Slope:	0.997226	0.999948
CH ₄ Cal Offset:	-0.036919	-0.039119
NMHC Cal Slope:	0.999437	0.999071
NMHC Cal Offset:	-0.023707	-0.016707

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


THC Calibration Summary

Station Information

Calibration Date:	November 3, 2025	Previous Calibration:	October 14, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:40	End Time (MST):	13:37
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999953	<i>≥0.995</i>
17.49	17.45	1.0020	Slope	0.999473	<i>0.90 - 1.10</i>
8.74	8.65	1.0110	Intercept	-0.056026	<i>+/-0.5</i>
4.37	4.26	1.0258			





Wood Buffalo Environmental Association


CH₄ Calibration Summary

Station Information

Calibration Date:	November 3, 2025	Previous Calibration:	October 14, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:40	End Time (MST):	13:37
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.00	0.00	----	Correlation Coefficient	0.999894	<i>≥0.995</i>
8.14	8.13	1.0024	Slope	0.999948	<i>0.90 - 1.10</i>
4.07	4.01	1.0152	Intercept	-0.039119	<i>+/-0.5</i>
2.04	1.96	1.0392			





Wood Buffalo Environmental Association


NMHC Calibration Summary

Station Information

Calibration Date:	November 3, 2025	Previous Calibration:	October 14, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	10:40	End Time (MST):	13:37
Analyzer make:	Thermo 55i	Analyzer serial #:	1152430011

Calibration Data


Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.00	0.00	----	Correlation Coefficient	0.999985	≥ 0.995
9.34	9.33	1.0017	Slope	0.999071	$0.90 - 1.10$
4.67	4.64	1.0072	Intercept	-0.016707	± 0.5
2.34	2.30	1.0143			



NMHC Calibration Plot

Date: November 3, 2025

Location: Ells River





Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

Station Information

Station Name: Ells River
Calibration Date: November 14, 2025
Start time (MST): 10:19
Reason: Cylinder Change

Station number: AMS 30
Last Cal Date: November 3, 2025
End time (MST): 11:49

Calibration Standards

Gas Cert Reference: CC350110
CH₄ Cal Gas Conc. 496.6 ppm
C₃H₈ Cal Gas Conc. 207.2 ppm
Removed Gas Cert: NA
Removed CH₄ Conc. 496.6 ppm
Removed C₃H₈ Conc. 207.2 ppm
Diff between cyl (CH₄):
Calibrator Model: API T700
Zero Air Gen model: API T701H

Cal Gas Expiry Date: Monday, March 10, 2031
CH₄ Equiv Conc. 1066.4 ppm
Removed Gas Expiry: NA
CH₄ Equiv Conc. 1066.4 ppm
Diff between cyl (THC):
Diff between cyl (NM):
Serial Number: 3061
Serial Number: 358

Analyzer Information

Analyzer make: Thermo 55i
THC Range: 0 - 20 ppm

Analyzer serial #: 1152430011
NMHC/CH₄ Range: 0 - 10 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
CH ₄ SP Ratio:	3.12E-04	3.15E-04	NMHC SP Ratio:	5.89E-05
CH ₄ Retention time:	17.4	17.6	NMHC Peak Area:	157597
Zero Chromatogram:	ON	ON	Flat Baseline:	OFF

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	17.49	17.32	1.010
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	17.32	Prev response	17.42	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.95-1.05</i>
Calibrator zero					
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	17.49	17.34	1.008
Average Correction Factor					

Notes:

H2/ N2 was changed after as founds. No adjustment made.



Wood Buffalo Environmental Association

THC / CH₄ / NMHC Calibration Report

NMHC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	9.34	9.26	1.009
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	9.26	Prev response	9.32	*% change	-0.7%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

NMHC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	9.34	9.26	1.010
Average Correction Factor					

CH₄ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFZero))
As found zero	5000	0.0	0.00	0.00	----
As found High point	4918	82.0	8.14	8.06	1.011
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr AF:	8.06	Prev response	8.10	*% change	-0.6%
Baseline Corr 2nd AF:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

CH₄ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
Calibrator zero					----
High point					
Mid point					
Low point					
As left zero	5000	0.0	0.00	0.00	----
As left span	4918	82.0	8.14	8.09	1.007
Average Correction Factor					

Calibration Statistics

	<u>Start</u>	
THC Cal Slope:	0.999473	
THC Cal Offset:	-0.056026	
CH ₄ Cal Slope:	0.999948	
CH ₄ Cal Offset:	-0.039119	
NMHC Cal Slope:	0.999071	
NMHC Cal Offset:	-0.016707	


Finish

Calibration Performed By: Jan Castro

NMHC Calibration Plot

Date: November 14, 2025

Location: Ells River





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Ells River
Station number: AMS 30
Calibration Date: November 14, 2025
Last Cal Date: October 6, 2025
Start time (MST): 11:49
End time (MST): 15:58
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0027487
NOX Cal Gas Conc: 59.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.30 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.10 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.10 ppm
NO gas Diff:
Serial Number: 3061
Serial Number: 358

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.3	-0.2	-0.1	----	----
AF High point	4932	67.7	803.0	800.3	2.7	793.4	788.4	5.0	1.0117	1.0148
AF Mid point										
AF Low point										
New cyl resp										

Previous Response NO_x = 800.3 ppb NO = 797.4 ppb
Baseline Corr 1st pt NO_x = 793.7 ppb NO = 788.6 ppb
Baseline Corr 2nd pt NO_x = NA ppb NO = NA ppb
Baseline Corr 3rd pt NO_x = NA ppb NO = NA ppb

* = > +/-5% change initiates investigation

*Percent Change NO_x = -0.8%

As Found Statistics

*Percent Change NO = -1.1%

As found NO_x r²: Nx SI: Nx Int:

As found NO r²: NO SI: NO Int:

As found NO₂ r²: NO₂ SI: NO₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	--	--

As Found GPT zero
As found high GPT point
As found mid GPT point
As found low GPT point



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 710321429

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.356	1.375	NO bkgnd or offset:	15.6	16.1
NOX coeff or slope:	0.992	0.994	NOX bkgnd or offset:	15.8	16.3
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	206.5	200.5

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999406	1.001740
NO _x Cal Offset:	-2.239384	-2.258752
NO Cal Slope:	1.000405	1.001748
NO Cal Offset:	-3.180739	-3.100468
NO ₂ Cal Slope:	0.999890	0.999891
NO ₂ Cal Offset:	0.136766	-0.713116

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	0.1	-0.3	----	----
High point	4932	67.7	803.0	800.3	2.7	803.1	800.3	2.9	0.9998	1.0000
Mid point	4966	33.8	400.9	399.5	1.4	398.4	395.1	3.2	1.0062	1.0112
Low point	4983	16.9	200.4	199.8	0.7	196.4	194.1	2.2	1.0206	1.0292
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	0.2	-0.3	----	----
As left span	4932	67.7	803.0	426.5	376.5	802.6	426.5	376.1	1.0005	1.0000
Average Correction Factor									1.0089	1.0134

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.3	----	----
High GPT point	796.6	418.8	380.5	380.2	1.0008	99.9%
Mid GPT point	796.6	613.6	185.7	184.0	1.0093	99.1%
Low GPT point	796.6	702.0	97.3	96.7	1.0063	99.4%
Average Correction Factor					1.0055	99.5%

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 14, 2025	Previous Calibration:	October 6, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	11:49	End Time (MST):	15:58
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999970	≥0.995
803.0	803.1	0.9998	Slope	1.001740	0.90 - 1.10
400.9	398.4	1.0062	Intercept	-2.258752	+/-20
200.4	196.4	1.0206			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 14, 2025	Previous Calibration:	October 6, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	11:49	End Time (MST):	15:58
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999983	≥0.995
380.5	380.2	1.0008	Slope	0.999891	0.90 - 1.10
185.7	184.0	1.0093	Intercept	-0.713116	+/-20
97.3	96.7	1.0063			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 14, 2025	Previous Calibration:	October 6, 2025
Station Name:	Ells River	Station Number:	AMS 30
Start Time (MST):	11:49	End Time (MST):	15:58
Analyzer make:	Thermo 42i	Analyzer serial #:	710321429

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999926	≥ 0.995
800.3	800.3	1.0000	Slope	1.001748	$0.90 - 1.10$
399.5	395.1	1.0112	Intercept	-3.100468	± 20
199.8	194.1	1.0292			



NO_x Calibration Plot

Date: November 14, 2025

Location: Ells River





Wood Buffalo Environmental Association

T640 PM_{2.5} CALIBRATION

Version-01-2024

Station Information

Station Name: Ells River Station number: AMS 30
Calibration Date: November 14, 2025 Last Cal Date: October 14, 2025
Start time (MST): 14:28 End time (MST): 14:56

Analyzer Make: API T640 S/N: 875
Particulate Fraction: PM2.5

Flow Meter Make/Model: Alicat FP-25BT S/N: 388754
Temp/RH standard: Alicat FP-25BT S/N: 388754

Monthly Calibration Test

Parameter	As found	Measured	As left	Adjusted	(Limits)
T (°C)	-3.00	-3.47	-3.00	<input type="checkbox"/>	+/- 2 °C
P (mmHg)	721.20	722.70	721.20	<input type="checkbox"/>	+/- 10 mmHg
Flow (LPM)	4.98	5.05	4.98	<input type="checkbox"/>	+/- 0.25 LPM
PW% (pump)	34	----	34	<input type="checkbox"/>	>80%
Zero Verification	PM w/o HEPA: 0.20		PM w/ HEPA: 0.00		<0.2 ug/m3

Note: this leak check will be completed before the quarterly work and will serve as the pre maintenance leak check

PM Inlet observation : Inlet Head Clean ☒ Alignment Factor On : ☒

Quarterly Calibration Test

SPAN DUST Refractive Index: 10.90 Expiry Date: January 30, 2027
Lot No.: 100128-050-040

Parameter	As found	Post maintenance	As left	Adjusted	(Limits)
PMT Peak Test				<input type="checkbox"/>	+/- 0.5

Date Optical Chamber Cleaned: September 2, 2025
Date Disposable Filter Changed: September 2, 2025

Post- maintenance Zero Verification: PM w/ HEPA: <0.2 ug/m3

Annual Maintenance

Date Sample Tube Cleaned: June 18, 2025
Date RH/T Sensor Cleaned: June 18, 2025

Notes: Verified flow, temperature, pump power and pressure. No adjustment made. Leak check passed.

Calibration by: Jan Castro



WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS31 BLACKROD NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Blackrod Station number: AMS 31
Calibration Date: November 27, 2025 Last Cal Date: October 30, 2025
Start time (MST): 11:17 End time (MST): 14:07
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.25 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: CC327023
Removed Cal Gas Conc: 50.25 ppm Rem Gas Exp Date: N/A
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Model: Teledyne T700 Serial Number: 1220
Zero Air Gen Model: Teledyne N701H Serial Number: 72

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1160290014
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003523	1.000405	Backgd or Offset:	39.7	40.4
Calibration intercept:	-0.682143	-0.041621	Coeff or Slope:	0.997	1.019

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4921	79.5	798.9	782.5	1.021
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	782.2	Previous response	801.0	*% change	-2.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.5	----
High point	4921	79.5	798.9	799.4	0.999
Mid point	4960	39.8	400.0	400.0	1.000
Low point	4980	19.9	200.0	199.4	1.003
As left zero	5000	0.0	0.0	0.2	----
As left span	4921	79.5	798.9	798.6	1.000
Average Correction Factor:					1.001

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	November 27, 2025	Previous Calibration:	October 30, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	11:17	End Time (MST):	14:07
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290014

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.5	----	Correlation Coefficient	0.999998	≥0.995
798.9	799.4	0.9994	Slope	1.000405	0.90 - 1.10
400.0	400.0	1.0000	Intercept	-0.041621	+/-30
200.0	199.4	1.0030			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: November 27, 2025

Location: Blackrod





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Blackrod
Calibration Date: November 12, 2025
Start time (MST): 11:00
Reason: Routine

Station number: AMS 31
Last Cal Date: October 9, 2025
End time (MST): 14:32

Calibration Standards

Cal Gas Concentration: 5.42 ppm
Cal Gas Cylinder #: DT0016926
Removed Cal Gas Conc: 5.42 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne API N701H

Cal Gas Exp Date: March 19, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 1220
Serial Number: 72

Analyzer Information

Analyzer make: Thermo 43iQTL
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12228021056
Converter serial #: 2023-266
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006763	1.006617	Backgd or Offset: 2.54	2.54
Calibration intercept:	-0.000490	0.179606	Coeff or Slope: 0.944	0.944

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4926	73.8	80.0	80.0	1.001
As found Mid point	4963	36.9	40.0	40.4	0.993
As found Low point	4982	18.5	20.1	20.1	1.003
New cylinder response					
Baseline Corr As found:	79.9	Prev response:	80.54	*% change:	-0.8%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	0.999333	AF Intercept:	0.159668
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999972	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4926	73.8	80.0	80.6	0.993
Mid point	4963	36.9	40.0	40.8	0.980
Low point	4982	18.5	20.1	20.1	0.998
As left zero	5000	0.0	0.0	0.2	----
As left span	4926	73.8	80.0	80.7	0.991
SO2 Scrubber Check	4921	79.5	794.9	-0.1	----
Date of last scrubber change:	27-Aug-25		Ave Corr Factor		
Date of last converter efficiency test:	October 9, 2025		0.990		

Notes: Sample inlet filter was changed after multipoint as founds. SO2 scrubber check done and passed.
No adjustment made.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


H₂S Calibration Summary

Station Information

Calibration Date:	November 12, 2025	Previous Calibration:	October 9, 2025
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	11:00	End Time (MST):	14:32
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12228021056

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999941		≥ 0.995
80.0	80.6	0.9926	Slope	1.006617		$0.90 - 1.10$
40.0	40.8	0.9804	Intercept	0.179606		± 3
20.1	20.1	0.9976				



H₂S Calibration Plot

Date: November 12, 2025

Location: Blackrod





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Blackrod
 Station number: AMS 31
 Calibration Date: November 12, 2025
 Last Cal Date: October 21, 2025
 Start time (MST): 11:00
 End time (MST): 12:43
 Reason: Removal

Calibration Standards

NO Gas Cylinder #: DT0035071
 NOX Cal Gas Conc: 59.30 ppm
 Removed Cylinder #: NA
 Removed Gas NOX Conc: 59.30 ppm
 NOX gas Diff:
 Calibrator Model: Teledyne API T750
 ZAG make/model: Teledyne API 751H
 Cal Gas Expiry Date: January 9, 2032
 NO Cal Gas Conc: 59.10 ppm
 Removed Gas Exp Date: NA
 Removed Gas NO Conc: 59.10 ppm
 NO gas Diff:
 Serial Number: 281
 Serial Number: 530

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.3	0.1	----	----
AF High point	4932	67.7	803.0	800.3	2.7	803.0	799.4	3.5	0.9997	1.0007
AF Mid point	4966	33.8	400.9	399.5	1.4	409.4	405.9	3.5	0.9787	0.9836
AF Low point	4983	16.9	200.4	199.8	0.7	200.9	198.9	1.9	0.9967	1.0028

New cyl resp

Previous Response	NO _x = 805.0 ppb	NO = 802.3 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = -0.2%
Baseline Corr 1st pt	NO _x = 803.2 ppb	NO = 799.7 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -0.3%
Baseline Corr 2nd pt	NO _x = 409.6 ppb	NO = 406.2 ppb	As found	NO _x r ² : 0.999851	Nx SI: 1.001243	Nx Int: 1.766
Baseline Corr 3rd pt	NO _x = 201.1 ppb	NO = 199.2 ppb	As found	NO r ² : 0.999894	NO SI: 1.000463	NO Int: 0.924
			As found	NO ₂ r ² : 1.000000	NO ₂ SI: 1.069070	NO ₂ Int: 0.100

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero	----	----	0.0	0.1	----	----
As found high GPT point	785.6	409.1	379.2	405.5	0.9352	106.9%
As found mid GPT point	785.6					
As found low GPT point	785.6					



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1426262592

Instrument Settings

	<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.205	
NOX coeff or slope:	0.995	
NO2 coeff or slope:	1.000	

	<u>Start</u>	<u>Finish</u>
NO bkgnd or offset:	17.9	
NOX bkgnd or offset:	18.1	
Reaction cell Press:	141.2	

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.000210	
NO _x Cal Offset:	1.903333	
NO Cal Slope:	1.001497	
NO Cal Offset:	0.862201	
NO ₂ Cal Slope:	0.974019	
NO ₂ Cal Offset:	3.859945	

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
-----------	---------------------------	-----------------------------	---	--	---	--	---------------------------------------	--	---	--

Cal zero
High point
Mid point
Low point
As left zero
As left span

Average Correction Factor

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	--	---------------------------------------	---	--	---	--

Cal zero
High GPT point
Mid GPT point
Low GPT point

Average Correction Factor


Notes: Removal calibrations done to do further troubleshooting at the shop.

Calibration Performed By: Jan Castro

NO_x Calibration Plot

Date: November 12, 2025

Location: Blackrod





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Blackrod
Station number: AMS 31
Calibration Date: November 13, 2025
Last Cal Date:
Start time (MST): 10:39
End time (MST): 14:15
Reason: Install

Calibration Standards

NO Gas Cylinder #: DT0035071
NOX Cal Gas Conc: 59.30 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 59.30 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API N701H
Cal Gas Expiry Date: January 9, 2032
NO Cal Gas Conc: 59.10 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.10 ppm
NO gas Diff:
Serial Number: 1220
Serial Number: 72

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
-----------	------------------------------	--------------------------------	---	--	---	--	---	--	--	---

As found zero
AF High point
AF Mid point
AF Low point
New cyl resp

Previous Response	NO _x = NA	ppb	NO = NA	ppb	* = > +/-5% change initiates investigation	*Percent Change	NO _x = NA
Baseline Corr 1st pt	NO _x = NA	ppb	NO = NA	ppb	<u>As Found Statistics</u>	*Percent Change	NO = NA
Baseline Corr 2nd pt	NO _x = NA	ppb	NO = NA	ppb	As found NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA	ppb	NO = NA	ppb	As found NO r ² :	NO SI:	NO Int:
					As found NO ₂ r ² :	NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
-------------------	---	--	--	---	--	--

As Found GPT zero
As found high GPT point
As found mid GPT point
As found low GPT point



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1173480006

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:		0.967	NO bkgnd or offset:		2.3
NOX coeff or slope:		0.993	NOX bkgnd or offset:		2.3
NO2 coeff or slope:		1.000	Reaction cell Press:		145.3

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:		1.000142
NO _x Cal Offset:		-0.297959
NO Cal Slope:		1.001672
NO Cal Offset:		-1.199111
NO ₂ Cal Slope:		0.998566
NO ₂ Cal Offset:		-1.057289

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.0	0.0	0.0	----	----
High point	4932	67.7	803.0	800.3	2.7	802.8	801.0	1.8	1.0002	0.9991
Mid point	4966	33.8	400.9	399.5	1.4	400.9	398.4	2.5	1.0000	1.0028
Low point	4983	16.9	200.4	199.8	0.7	199.6	197.7	2.0	1.0042	1.0104
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.1	0.1	----	----
As left span	4932	67.7	803.0	407.6	395.4	795.8	407.6	388.2	1.0090	1.0000
Average Correction Factor									1.0015	1.0041

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	797.0	409.8	389.9	388.8	1.0029	99.7%
Mid GPT point	797.0	600.0	199.7	198.0	1.0086	99.1%
Low GPT point	797.0	695.3	104.4	102.0	1.0236	97.7%
Average Correction Factor					1.0117	98.9%

Notes: Install calibrations. Sample inlet filter was changed before calibrator zero. Adjusted zero and span.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 13, 2025	Previous Calibration:	
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	10:39	End Time (MST):	14:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999999	≥ 0.995
803.0	802.8	1.0002	Slope	1.000142	$0.90 - 1.10$
400.9	400.9	1.0000	Intercept	-0.297959	± 20
200.4	199.6	1.0042			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 13, 2025	Previous Calibration:	
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	10:39	End Time (MST):	14:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<i>Limits</i>
0.0	0.0	----	Correlation Coefficient	0.999964	<i>≥0.995</i>
389.9	388.8	1.0029	Slope	0.998566	<i>0.90 - 1.10</i>
199.7	198.0	1.0086	Intercept	-1.057289	<i>+/-20</i>
104.4	102.0	1.0236			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 13, 2025	Previous Calibration:	
Station Name:	Blackrod	Station Number:	AMS 31
Start Time (MST):	10:39	End Time (MST):	14:15
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999990	≥ 0.995
800.3	801.0	0.9991	Slope	1.001672	$0.90 - 1.10$
399.5	398.4	1.0028	Intercept	-1.199111	± 20
199.8	197.7	1.0104			



NO_x Calibration Plot

Date: November 13, 2025

Location: Blackrod





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS33 MONDAY CREEK NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Monday Creek Station number: AMS 33
Calibration Date: November 6, 2025 Last Cal Date: October 23, 2025
Start time (MST): 10:46 End time (MST): 13:34
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.62 ppm Cal Gas Exp Date: March 10, 2031
Cal Gas Cylinder #: EB0008522
Removed Cal Gas Conc: 50.62 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Model: Teledyne T700 Serial Number: 3253
Zero Air Gen Model: Teledyne T701H Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1152430005
Analyzer Range: 0- 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996658	0.999672	Backgd or Offset:	31.0	31.1
Calibration intercept:	0.442201	0.861611	Coeff or Slope:	0.970	0.983

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4921	79.1	800.8	797.2	1.004
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.3	Previous response	798.6	*% change	-0.2%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4921	79.1	800.8	801.0	1.000
Mid point	4961	39.5	399.9	401.0	0.997
Low point	4980	19.8	200.5	202.0	0.992
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	79.1	800.8	802.0	0.998
Average Correction Factor:					0.996

Notes: Sample inlet filter was changed after as founds. Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 23, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	10:46	End Time (MST):	13:34
Analyzer make:	Thermo 43i	Analyzer serial #:	1152430005

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999996	≥0.995
800.8	801.0	0.9997	Slope	0.999672	0.90 - 1.10
399.9	401.0	0.9972	Intercept	0.861611	+/-30
200.5	202.0	0.9924			



SO2 Calibration Plot

Date: November 6, 2025

Location: Monday Creek





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Monday Creek
Calibration Date: November 20, 2025
Start time (MST): 11:08
Reason: Routine

Station number: AMS 33
Last Cal Date: October 22, 2025
End time (MST): 14:48

Calibration Standards

Cal Gas Concentration: 5.05 ppm
Cal Gas Cylinder #: DT0014831
Removed Cal Gas Conc: 5.05 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne API T700
ZAG Make/Model: Teledyne T701H

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 3253
Serial Number: 832

Analyzer Information

Analyzer make: Thermo 43iQTL
Converter make: Global 150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 12333331547
Converter serial #: 2022-196
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.996282	1.000426	Backgd or Offset:	1.3	1.3
Calibration intercept:	0.358401	0.238404	Coeff or Slope:	1.062	1.082

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.2	----
As found High point	4921	79.2	80.0	80.0	1.002
As found Mid point	4960	39.6	40.0	40.1	1.002
As found Low point	4980	19.8	20.0	20.3	0.995
New cylinder response					
Baseline Corr As found:	79.8	Prev response:	80.05	*% change:	-0.3%
Baseline Corr 2nd AF pt:	39.9	AF Slope:	0.996997	AF Intercept:	0.258405
Baseline Corr 3rd AF pt:	20.1	AF Correlation:	0.999996	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4921	79.2	80.0	80.3	0.996
Mid point	4960	39.6	40.0	40.1	0.997
Low point	4980	19.8	20.0	20.4	0.980
As left zero	5000	0.0	0.0	0.4	----
As left span	4921	79.2	80.0	80.2	0.997
SO2 Scrubber Check	4921	79.1	791.0	0.1	----
Date of last scrubber change:	11-Apr-24		Ave Corr Factor		0.991
Date of last converter efficiency test:	October 22, 2025				

Notes: Sample inlet filter was changed after multipoint as founds. SO2 scrubber check done and passed.
Adjusted span only.

Calibration Performed By: Jan Castro



Wood Buffalo Environmental Association

H2S Calibration Summary


Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 22, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	11:08	End Time (MST):	14:48
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12333331547

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999986		≥ 0.995
80.0	80.3	0.9961	Slope	1.000426		$0.90 - 1.10$
40.0	40.1	0.9975	Intercept	0.238404		± 3
20.0	20.4	0.9803				


H₂S Calibration Curve



H2S Calibration Plot

Date: November 20, 2025

Location: Monday Creek





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Monday Creek
Station number: AMS 33
Calibration Date: November 6, 2025
Last Cal Date: October 29, 2025
Start time (MST): 10:46
End time (MST): 14:55
Reason: Routine

Calibration Standards

NO Gas Cylinder #: CC755290
NOX Cal Gas Conc: 48.90 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 48.90 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T750
ZAG make/model: Teledyne API 751H
Cal Gas Expiry Date: March 11, 2031
NO Cal Gas Conc: 48.70 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 48.70 ppm
NO gas Diff:
Serial Number: 281
Serial Number: 530

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
AF High point	4918	82.1	802.9	799.6	3.3	792.9	785.8	3.6	1.0125	1.0175
AF Mid point										
AF Low point										
New cyl resp										

Previous Response	NO _x = 800.4 ppb	NO = 797.9 ppb	* = > +/-5% change initiates investigation		*Percent Change	NO _x = -0.9%
Baseline Corr 1st pt	NO _x = 793.0 ppb	NO = 785.9 ppb	<u>As Found Statistics</u>		*Percent Change	NO = -1.5%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb	As found	NO _x r ² :	Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb	As found	NO r ² :	NO SI:	NO Int:
			As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12426335704

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.995970	0.989581
NO _x Cal Offset:	0.670358	1.890234
NO Cal Slope:	0.997903	0.990857
NO Cal Offset:	-0.049822	1.390612
NO ₂ Cal Slope:	0.967535	0.991773
NO ₂ Cal Offset:	1.063691	1.628365

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.045	1.045	NO bkgnd or offset:	1.4	1.4
NOX coeff or slope:	0.999	0.998	NOX bkgnd or offset:	1.5	1.5
NO2 coeff or slope:	0.990	0.990	Reaction cell Press:	136.2	136.8

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	NOx Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.2	0.0	----	----
High point	4918	82.1	802.9	799.6	3.3	795.3	793.0	2.3	1.0096	1.0084
Mid point	4959	41.1	401.9	400.3	1.6	401.0	398.6	2.4	1.0024	1.0043
Low point	4979	20.5	200.5	199.7	0.8	202.2	201.0	1.2	0.9916	0.9935
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	0.0	----	----
As left span	4918	82.1	802.9	428.0	374.9	785.0	428.0	357.0	1.0228	1.0000
Average Correction Factor									1.0012	1.0020

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	NO2 Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	789.1	397.2	395.2	392.5	1.0068	99.3%
Mid GPT point	789.1	596.7	195.7	197.3	0.9918	100.8%
Low GPT point	789.1	694.6	97.8	99.7	0.9808	102.0%
Average Correction Factor					0.9931	100.7%

Notes:

Sample inlet filter was changed after as founds. No adjustment made.

Calibration Performed By:

Jan Castro



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 29, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	10:46	End Time (MST):	14:55
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12426335704

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999968	≥0.995
802.9	795.3	1.0096	Slope	0.989581	0.90 - 1.10
401.9	401.0	1.0024	Intercept	1.890234	+/-20
200.5	202.2	0.9916			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 29, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	10:46	End Time (MST):	14:55
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12426335704

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999911	≥0.995
395.2	392.5	1.0068	Slope	0.991773	0.90 - 1.10
195.7	197.3	0.9918	Intercept	1.628365	+/-20
97.8	99.7	0.9808			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 29, 2025
Station Name:	Monday Creek	Station Number:	AMS 33
Start Time (MST):	10:46	End Time (MST):	14:55
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12426335704

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999981	≥ 0.995
799.6	793.0	1.0084	Slope	0.990857	$0.90 - 1.10$
400.3	398.6	1.0043	Intercept	1.390612	± 20
199.7	201.0	0.9935			



NO_x Calibration Plot

Date: November 6, 2025

Location: Monday Creek





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

**AMS501
LEISMER
NOVEMBER 2025**

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Leismer
Calibration Date: November 7, 2025
Start time (MST): 10:36
Reason: Routine

Station number: AMS 501
Last Cal Date: October 22, 2025
End time (MST): 13:12

Calibration Standards

Cal Gas Concentration: 50.52 ppm
Cal Gas Cylinder #: CC274266
Removed Cal Gas Conc: 50.52 ppm
Removed Gas Cyl #: NA
Calibrator Model: Teledyne API T700
Zero Air Gen Model: Teledyne API T701

Cal Gas Exp Date: December 29, 2028
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 2659
Serial Number: 4427

Analyzer Information

Analyzer make: Thermo 43i
Analyzer Range: 0-1000 ppb

Serial Number: 1160290011

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.007371	0.997145	Backgd or Offset:	21.4	21.4
Calibration intercept:	-0.136096	0.444060	Coeff or Slope:	0.994	0.994

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4921	79.2	800.2	798.0	1.003
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	797.9	Previous response	806.0	*% change	-1.0%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4921	79.2	800.2	798.5	1.002
Mid point	4960	39.6	400.2	398.8	1.003
Low point	4980	19.8	200.1	200.7	0.997
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	79.2	800.2	803.0	0.997
Average Correction Factor:					1.001

Notes: No adjustment.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 7, 2025	Previous Calibration:	October 22, 2025
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	10:36	End Time (MST):	13:12
Analyzer make:	Thermo 43i	Analyzer serial #:	1160290011

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		Limits
0.0	0.2	----	Correlation Coefficient	0.999997	≥0.995
800.2	798.5	1.0021	Slope	0.997145	0.90 - 1.10
400.2	398.8	1.0034	Intercept	0.444060	+/-30
200.1	200.7	0.9968			



SO2 Calibration Plot

Date: November 7, 2025

Location: Leismer





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Leismer
Calibration Date: November 20, 2025
Start time (MST): 9:11
Reason: Routine

Station number: AMS 501
Last Cal Date: October 9, 2025
End time (MST): 12:54

Calibration Standards

Cal Gas Concentration: 4.89 ppm
Cal Gas Cylinder #: CC737971
Removed Cal Gas Conc: 4.89 ppm
Removed Gas Cyl #: NA
Calibrator Make/Model: Teledyne 750
ZAG Make/Model: Teledyne 751H

Cal Gas Exp Date: September 5, 2027
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 282
Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: Global G150
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1180540020
Converter serial #: 2022-218
Converter Temp: 325.0 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.008591	0.987880	Backgd or Offset:	3.71
Calibration intercept:	-0.239137	-0.039505	Coeff or Slope:	1.177

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.2	----
As found High point	4918	81.8	80.0	78.3	1.019
As found Mid point	4959	40.9	40.0	39.2	1.015
As found Low point	4980	20.4	19.9	19.7	1.002
New cylinder response					
Baseline Corr As found:	78.5	Prev response:	80.45	*% change:	-2.5%
Baseline Corr 2nd AF pt:	39.4	AF Slope:	0.980025	AF Intercept:	-0.039638
Baseline Corr 3rd AF pt:	19.9	AF Correlation:	0.999980	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	-0.1	----
High point	4918	81.8	80.0	79.0	1.013
Mid point	4959	40.9	40.0	39.4	1.015
Low point	4980	20.4	19.9	19.8	1.008
As left zero	5000	0.0	0.0	-0.1	----
As left span	4918	81.8	80.0	79.1	1.011
SO2 Scrubber Check	4921	79.2	800.2	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.012
Date of last converter efficiency test:	November 20, 2025				

Notes: No adjustment.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association


H2S Calibration Summary

Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 9, 2025
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	9:11	End Time (MST):	12:54
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1180540020

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999992		≥ 0.995
80.0	79.0	1.0127	Slope	0.987880		0.90 - 1.10
40.0	39.4	1.0153	Intercept	-0.039505		+/-3
19.9	19.8	1.0076				



H2S Calibration Plot

Date: November 20, 2025

Location: Leismer





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Leismer
Station number: AMS 501
Calibration Date: November 6, 2025
Last Cal Date: October 15, 2025
Start time (MST): 9:39
End time (MST): 13:48
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0022706
NOX Cal Gas Conc: 60.20 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.20 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API 701
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.10 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.10 ppm
NO gas Diff:
Serial Number: 3252
Serial Number: 4427

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	0.2	0.3	-0.1	----	----
AF High point	4933	66.6	801.9	800.6	1.3	800.6	795.5	5.2	1.0019	1.0068
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x =	801.8 ppb	NO =	799.4 ppb	* = > +/-5% change initiates investigation			*Percent Change	NO _x = -0.2%	
Baseline Corr 1st pt	NO _x =	800.4 ppb	NO =	795.2 ppb	<u>As Found Statistics</u>			*Percent Change	NO = -0.5%	
Baseline Corr 2nd pt	NO _x =	NA ppb	NO =	NA ppb	As found NO _x r ² :			Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x =	NA ppb	NO =	NA ppb	As found NO r ² :			NO SI:	NO Int:	
					As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42iQ
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 12400232071

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.975	0.975	NO bkgnd or offset:	0.4	0.4
NOX coeff or slope:	0.991	0.991	NOX bkgnd or offset:	0.6	0.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	181.7	180.9

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.998058	0.996675
NO _x Cal Offset:	1.452094	1.211956
NO Cal Slope:	0.998965	0.995610
NO Cal Offset:	-0.409102	0.090507
NO ₂ Cal Slope:	0.990038	0.986982
NO ₂ Cal Offset:	1.083301	1.590215

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.4	0.4	0.0	----	----
High point	4933	66.6	801.9	800.6	1.3	800.3	797.5	2.8	1.0020	1.0039
Mid point	4967	33.3	400.9	400.2	0.7	400.5	397.9	2.6	1.0010	1.0059
Low point	4983	16.6	199.9	199.5	0.3	201.7	198.8	2.9	0.9910	1.0038
As left zero	5000	0.0	0.0	0.0	0.0	0.4	0.4	0.0	----	----
As left span	4933	66.6	801.9	384.3	417.6	792.9	384.3	408.6	1.0114	1.0000
Average Correction Factor									0.9980	1.0045

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	795.7	376.2	420.8	416.3	1.0109	98.9%
Mid GPT point	795.7	580.3	216.7	215.7	1.0048	99.5%
Low GPT point	795.7	685.1	111.9	114.1	0.9810	101.9%
Average Correction Factor					0.9989	100.1%

Notes:

No adjustment.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 15, 2025
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	9:39	End Time (MST):	13:48
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999993	≥0.995
801.9	800.3	1.0020	Slope	0.996675	0.90 - 1.10
400.9	400.5	1.0010	Intercept	1.211956	+/-20
199.9	201.7	0.9910			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 15, 2025
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	9:39	End Time (MST):	13:48
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999924	≥0.995
420.8	416.3	1.0109	Slope	0.986982	0.90 - 1.10
216.7	215.7	1.0048	Intercept	1.590215	+/-20
111.9	114.1	0.9810			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 6, 2025	Previous Calibration:	October 15, 2025
Station Name:	Leismer	Station Number:	AMS 501
Start Time (MST):	9:39	End Time (MST):	13:48
Analyzer make:	Thermo 42iQ	Analyzer serial #:	12400232071

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999998	≥ 0.995
800.6	797.5	1.0039	Slope	0.995610	$0.90 - 1.10$
400.2	397.9	1.0059	Intercept	0.090507	± 20
199.5	198.8	1.0038			



NO_x Calibration Plot

Date: November 6, 2025

Location: Leismer





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS505 SAWBONES BAY NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Sawbones Bay Station number: AMS 505
Calibration Date: November 27, 2025 Last Cal Date: October 29, 2025
Start time (MST): 9:44 End time (MST): 12:40
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.05 ppm Cal Gas Exp Date: April 9, 2033
Cal Gas Cylinder #: EB0063977
Removed Cal Gas Conc: 50.05 ppm Rem Gas Exp Date:
Removed Gas Cyl #: EB0063977 Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 5112
Zero Air Gen Model: Teledyne API T701 Serial Number: 690

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 710321323
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003636	1.003809	Backgd or Offset:	20.7	21.7
Calibration intercept:	-1.255492	-0.415883	Coeff or Slope:	1.111	1.128

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.5	----
As found High point	4920	79.8	798.8	789.1	1.012
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	789.6	Previous response	800.5	*% change	-1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4920	79.8	798.8	801.7	0.996
Mid point	4960	39.9	399.4	400.3	0.998
Low point	4980	20.0	200.2	200.0	1.001
As left zero	5000	0.0	0.0	0.0	----
As left span	4920	79.8	798.8	802.0	0.996
Average Correction Factor:					0.998

Notes: Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

SO₂ Calibration Summary


Station Information

Calibration Date:	November 27, 2025	Previous Calibration:	October 29, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:44	End Time (MST):	12:40
Analyzer make:	Thermo 43i	Analyzer serial #:	710321323

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999998	≥0.995
798.8	801.7	0.9964	Slope	1.003809	0.90 - 1.10
399.4	400.3	0.9978	Intercept	-0.415883	+/-30
200.2	200.0	1.0010			


SO₂ Calibration Curve



SO2 Calibration Plot

Date: November 27, 2025

Location: Sawbones Bay





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Sawbones Bay Station number: AMS 505
Calibration Date: November 26, 2025 Last Cal Date: October 28, 2025
Start time (MST): 9:21 End time (MST): 13:09
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.26 ppm Cal Gas Exp Date: March 19, 2027
Cal Gas Cylinder #: DT0034141
Removed Cal Gas Conc: 5.26 ppm Rem Gas Exp Date: NA
Removed Gas Cyl #: NA Diff between cyl:
Calibrator Make/Model: Teledyne API T750 Serial Number: 282
ZAG Make/Model: Teledyne API T751H Serial Number: 321

Analyzer Information

Analyzer make: Thermo 43iQTL Analyzer serial #: 12113311965
Converter make: Global 150 Converter serial #: 2022-224
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002887	0.991595	Backgd or Offset:	0.920	0.920
Calibration intercept:	0.120000	0.140000	Coeff or Slope:	1.105	1.105

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.1	----
As found High point	4924	76.0	80.0	80.8	0.991
As found Mid point	4962	38.0	40.0	40.4	0.992
As found Low point	4981	19.0	20.0	20.1	0.999
New cylinder response					
Baseline Corr As found:	80.7	Prev response:	80.30	*% change:	0.5%
Baseline Corr 2nd AF pt:	40.3	AF Slope:	1.010035	AF Intercept:	0.020000
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999995	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.2	----
High point	4924	76.0	80.0	79.5	1.006
Mid point	4962	38.0	40.0	39.6	1.009
Low point	4981	19.0	20.0	20.0	0.999
As left zero	5000	0.0	0.0	0.3	----
As left span	4924	76.0	80.0	78.7	1.016
SO2 Scrubber Check	4922	77.8	778.0	0.0	----
Date of last scrubber change:				Ave Corr Factor	1.005
Date of last converter efficiency test:					

Notes: Changed inlet filter after as founds. Scrubber test was done after calibrator zero. No adjustment made.

Calibration Performed By: Sean Bala



Wood Buffalo Environmental Association

H2S Calibration Summary


Station Information

Calibration Date:	November 26, 2025	Previous Calibration:	October 28, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:21	End Time (MST):	13:09
Analyzer make:	Thermo 43iQTL	Analyzer serial #:	12113311965

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.2	----	Correlation Coefficient	0.999987		≥ 0.995
80.0	79.5	1.0057	Slope	0.991595		$0.90 - 1.10$
40.0	39.6	1.0095	Intercept	0.140000		± 3
20.0	20.0	0.9994				


H₂S Calibration Curve



H2S Calibration Plot

Date: November 26, 2025

Location: Sawbones Bay





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Sawbones Bay
Station number: AMS 505
Calibration Date: November 25, 2025
Last Cal Date: October 23, 2025
Start time (MST): 9:24
End time (MST): 13:40
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0009786
NOX Cal Gas Conc: 60.10 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.10 ppm
NOX gas Diff:
Calibrator Model: API T700
ZAG make/model: API T701H
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 60.00 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 60.00 ppm
NO gas Diff:
Serial Number: 5112
Serial Number: 690

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.7	-0.5	-0.1	----	----
AF High point	4933	66.7	801.8	800.4	1.3	789.6	785.7	3.9	1.0145	1.0181
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 799.7 ppb	NO = 796.9 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -1.2%	
Baseline Corr 1st pt	NO _x = 790.3 ppb	NO = 786.2 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -1.4%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 4259

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.949	0.964	NO bkgnd or offset:	0.9	0.9
NOX coeff or slope:	0.949	0.962	NOX bkgnd or offset:	1.4	1.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	3.8	3.7

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	0.999337	1.002061
NO _x Cal Offset:	-1.510057	-0.890158
NO Cal Slope:	0.998333	1.003202
NO Cal Offset:	-2.250374	-1.630038
NO ₂ Cal Slope:	1.002647	0.999814
NO ₂ Cal Offset:	0.583578	0.456557

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.4	-0.3	-0.1	----	----
High point	4933	66.7	801.8	800.4	1.3	802.7	801.9	0.8	0.9989	0.9982
Mid point	4967	33.3	400.2	399.6	0.7	400.2	399.0	1.3	1.0001	1.0014
Low point	4983	16.7	200.7	200.4	0.3	199.6	197.8	1.7	1.0057	1.0132
As left zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
As left span	4933	66.7	801.8	396.5	405.3	799.4	396.5	402.9	1.0030	1.0000
Average Correction Factor									1.0016	1.0043

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.1	----	----
High GPT point	799.0	394.2	406.1	406.3	0.9996	100.0%
Mid GPT point	799.0	609.8	190.5	191.1	0.9970	100.3%
Low GPT point	799.0	704.9	95.4	96.5	0.9890	101.1%
Average Correction Factor					0.9952	100.5%

Notes:

Changed inlet filter after as founds. Adjusted span only.

Calibration Performed By:

Sean Bala



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 25, 2025	Previous Calibration:	October 23, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:24	End Time (MST):	13:40
Analyzer make:	API T200	Analyzer serial #:	4259

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.4	----	Correlation Coefficient	0.999998	≥ 0.995
801.8	802.7	0.9989	Slope	1.002061	$0.90 - 1.10$
400.2	400.2	1.0001	Intercept	-0.890158	± 20
200.7	199.6	1.0057			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 25, 2025	Previous Calibration:	October 23, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:24	End Time (MST):	13:40
Analyzer make:	API T200	Analyzer serial #:	4259

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999991	≥ 0.995
406.1	406.3	0.9996	Slope	0.999814	$0.90 - 1.10$
190.5	191.1	0.9970	Intercept	0.456557	± 20
95.4	96.5	0.9890			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 25, 2025	Previous Calibration:	October 23, 2025
Station Name:	Sawbones Bay	Station Number:	AMS 505
Start Time (MST):	9:24	End Time (MST):	13:40
Analyzer make:	API T200	Analyzer serial #:	4259

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.3	----	Correlation Coefficient	0.999987	≥ 0.995
800.4	801.9	0.9982	Slope	1.003202	$0.90 - 1.10$
399.6	399.0	1.0014	Intercept	-1.630038	± 20
200.4	197.8	1.0132			



NO_x Calibration Plot

Date: November 25, 2025

Location: Sawbones Bay





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS507 KIRBY SOUTH NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Kirby South Station number: AMS 507
Calibration Date: November 21, 2025 Last Cal Date: October 23, 2025
Start time (MST): 8:49 End time (MST): 12:32
Reason: Routine

Calibration Standards

Cal Gas Concentration: 50.74 ppm Cal Gas Exp Date: October 9, 2032
Cal Gas Cylinder #: CC255918
Removed Cal Gas Conc: 50.74 ppm Rem Gas Exp Date:
Removed Gas Cyl #: Diff between cyl:
Calibrator Model: Teledyne API T700 Serial Number: 2445
Zero Air Gen Model: Teledyne API T701H Serial Number: 880

Analyzer Information

Analyzer make: Thermo 43iQ Serial Number: 1182340007
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.003849	1.001506	Backgd or Offset:	29.6	30.1
Calibration intercept:	0.048008	-0.232030	Coeff or Slope:	1.112	1.112

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.3	----
As found High point	4921	78.8	799.7	800.0	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	799.7	Previous response	802.8	*% change	-0.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.0	----
High point	4921	78.8	799.7	801.0	0.998
Mid point	4961	39.4	399.8	399.4	1.001
Low point	4980	19.7	199.9	200.2	0.999
As left zero	5000	0.0	0.0	0.0	----
As left span	4921	78.8	799.7	802.0	0.997
Average Correction Factor:					0.999

Notes: Changed sample inlet filter after as founds. Adjusted zero.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 21, 2025	Previous Calibration:	October 23, 2025
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	8:49	End Time (MST):	12:32
Analyzer make:	Thermo 43iQ	Analyzer serial #:	1182340007

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999998	≥0.995
799.7	801.0	0.9984	Slope	1.001506	0.90 - 1.10
399.8	399.4	1.0010	Intercept	-0.232030	+/-30
199.9	200.2	0.9986			



SO2 Calibration Plot

Date: November 21, 2025

Location: Kirby South





Wood Buffalo Environmental Association

H2S Calibration Report

Station Information

Station Name: Kirby South
Calibration Date: November 20, 2025
Start time (MST): 12:15
Reason: Routine

Station number: AMS 507
Last Cal Date: October 15, 2025
End time (MST): 17:44

Calibration Standards

Cal Gas Concentration: 5.05 ppm
Cal Gas Cylinder #: DT0019762
Removed Cal Gas Conc: 5.05 ppm
Removed Gas Cyl #: n/a
Calibrator Make/Model: Teledyne API T750
ZAG Make/Model: Teledyne API T751H

Cal Gas Exp Date: November 15, 2026
Rem Gas Exp Date: NA
Diff between cyl:
Serial Number: 281
Serial Number: 530

Analyzer Information

Analyzer make: Thermo 43i-TLE
Converter make: Global
Analyzer Range: 0 - 100 ppb

Analyzer serial #: 1150840012
Converter serial #: 2022-197
Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.002386	0.995957	Backgd or Offset:	1.69	1.70
Calibration intercept:	0.120000	0.120000	Coeff or Slope:	1.052	1.052

H2S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	-0.1	----
As found High point	4921	79.2	80.0	81.1	0.985
As found Mid point	4960	39.6	40.0	40.3	0.990
As found Low point	4980	19.8	20.0	19.9	1.000
New cylinder response					
Baseline Corr As found:	81.2	Prev response:	80.30	*% change:	1.1%
Baseline Corr 2nd AF pt:	40.4	AF Slope:	1.016102	AF Intercept:	-0.260000
Baseline Corr 3rd AF pt:	20.0	AF Correlation:	0.999982	* = > +/-5% change initiates investigation	

H2S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4921	79.2	80.0	79.8	1.002
Mid point	4960	39.6	40.0	39.9	1.002
Low point	4980	19.8	20.0	20.1	0.995
As left zero	5000	0.0	0.0	0.1	----
As left span	4921	79.2	80.0	81.2	0.985
SO2 Scrubber Check	4919	80.0	800.2	0.0	----
Date of last scrubber change:	July 25, 2023			Ave Corr Factor	1.000
Date of last converter efficiency test:	September 18, 2025			102.4% efficiency	

Notes: Changed sample inlet filter and external pump after as founds. No adjustments made.

Calibration Performed By: Braiden Boutillier



Wood Buffalo Environmental Association

H2S Calibration Summary


Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 15, 2025
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	12:15	End Time (MST):	17:44
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1150840012

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation			<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999998		≥ 0.995
80.0	79.8	1.0024	Slope	0.995957		$0.90 - 1.10$
40.0	39.9	1.0024	Intercept	0.120000		± 3
20.0	20.1	0.9949				


H₂S Calibration Curve



H2S Calibration Plot

Date: November 20, 2025

Location: Kirby South





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Kirby South Station number: AMS 507
Calibration Gas Date: November 21, 2025 Last Cal Date: October 23, 2025
Start time (MST): 8:49 End time (MST): 12:32
Reason: Routine

Calibration Standards

Gas Cert Reference: CC255918 Cal Gas Expiry Date: October 9, 2032
CH4 Cal Gas Conc. 506.4 ppm CH4 Equiv Conc. 1070.2 ppm
C3H8 Cal Gas Conc. 205.0 ppm
Removed Gas Cert: Removed Gas Expiry:
Removed CH4 Conc. 506.4 ppm CH4 Equiv Conc. 1070.2 ppm
Removed C3H8 Conc. 205.0 ppm Diff between cyl:
Calibrator Make/Model: Teledyne API T700 Serial Number: 2445
ZAG Make/Model: Teledyne API T701H Serial Number: 880

Analyzer Information

Analyzer make: Thermo 51i-LT Analyzer serial #: 1182340005
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	0.997247	0.997911	Background:	1.78	1.85
Calibration intercept:	0.016568	-0.018831	Coefficient:	3.524	3.489

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.00	-0.07	----
As found High point	4921	78.8	16.87	17.00	0.988
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.07	Previous response	16.84	*% change	1.4%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.00	0.01	
High point	4921	78.8	16.87	16.82	1.003
Mid point	4961	39.4	8.43	8.40	1.004
Low point	4980	19.7	4.22	4.15	1.016
As left zero	5000	0.0	0.00	-0.02	----
As left span	4921	78.8	16.87	16.98	0.993
Average Correction Factor					1.008

Notes: Changed sample inlet filter after as founds. Adjusted zero and span.

Calibration Performed By: Braiden Boutilier



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	November 21, 2025	Previous Calibration:	October 23, 2025
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	8:49	End Time (MST):	12:32
Analyzer make:	Thermo 51i-LT	Analyzer serial #:	1182340005

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	0.01	----	Correlation Coefficient	0.999984	≥ 0.995
16.87	16.82	1.0027	Slope	0.997911	$0.90 - 1.10$
8.43	8.40	1.0038	Intercept	-0.018831	± 1.5
4.22	4.15	1.0163			


THC Calibration Curve



THC Calibration Plot

Date: November 21, 2025

Location: Kirby South





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Kirby South
Station number: AMS 507
Calibration Date: November 20, 2025
Last Cal Date: October 15, 2025
Start time (MST): 12:15
End time (MST): 17:52
Reason: Routine

Calibration Standards

NO Gas Cylinder #: DT0019572
NOX Cal Gas Conc: 60.00 ppm
Removed Cylinder #: NA
Removed Gas NOX Conc: 60.00 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701H
Cal Gas Expiry Date: January 5, 2032
NO Cal Gas Conc: 59.90 ppm
Removed Gas Exp Date: NA
Removed Gas NO Conc: 59.90 ppm
NO gas Diff:
Serial Number: 2445
Serial Number: 880

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NOx concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO2 concentration (ppb) (Cc)	Indicated NOx concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NOx Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.1	----	----
AF High point	4933	66.8	801.6	800.3	1.3	802.0	795.9	6.6	0.9994	1.0055
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 802.6 ppb	NO = 798.7 ppb				* = > +/-5% change initiates investigation			*Percent Change	NO _x = -0.1%
Baseline Corr 1st pt	NO _x = 802.1 ppb	NO = 795.9 ppb				<u>As Found Statistics</u>			*Percent Change	NO = -0.3%
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found NO _x r ² :			Nx SI:	Nx Int:
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found NO r ² :			NO SI:	NO Int:
						As found NO ₂ r ² :			NO2 SI:	NO ₂ Int:

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO2 concentration (ppb) (Cc)	Indicated NO2 concentration (ppb) (Ic)	Baseline Adjusted NO2 Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Thermo 42i
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 1173480006

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.001884	1.003352
NO _x Cal Offset:	-0.493586	-1.433601
NO Cal Slope:	0.999158	1.001885
NO Cal Offset:	-0.953623	-2.733647
NO ₂ Cal Slope:	0.982296	0.973364
NO ₂ Cal Offset:	0.418066	0.733853

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	0.723	0.727	NO bkgnd or offset:	7.4	7.4
NOX coeff or slope:	0.996	0.996	NOX bkgnd or offset:	7.5	7.6
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	141.7	142.0

Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	0.1	0.1	0.0	----	----
High point	4933	66.8	801.6	800.3	1.3	804.0	801.0	2.7	0.9971	0.9991
Mid point	4967	33.4	400.8	400.1	0.7	398.9	395.2	3.7	1.0048	1.0125
Low point	4983	16.7	200.4	200.1	0.3	198.8	195.9	2.9	1.0080	1.0213
As left zero	5000	0.0	0.0	0.0	0.0	0.2	0.2	0.0	----	----
As left span	4933	66.8	801.6	400.7	400.9	795.0	400.7	394.4	1.0083	1.0000
Average Correction Factor									1.0033	1.0110

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	0.0	----	----
High GPT point	801.0	413.1	389.2	379.4	1.0259	97.5%
Mid GPT point	801.0	624.0	178.3	174.2	1.0237	97.7%
Low GPT point	801.0	707.5	94.8	94.1	1.0078	99.2%
Average Correction Factor					1.0192	98.1%

Notes:

Changed sample inlet filter after as founds. Adjusted span.

Calibration Performed By:

Braiden Boutilier



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 15, 2025
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	12:15	End Time (MST):	17:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999979	≥0.995
801.6	804.0	0.9971	Slope	1.003352	0.90 - 1.10
400.8	398.9	1.0048	Intercept	-1.433601	+/-20
200.4	198.8	1.0080			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 15, 2025
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	12:15	End Time (MST):	17:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999978	≥0.995
389.2	379.4	1.0259	Slope	0.973364	0.90 - 1.10
178.3	174.2	1.0237	Intercept	0.733853	+/-20
94.8	94.1	1.0078			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 15, 2025
Station Name:	Kirby South	Station Number:	AMS 507
Start Time (MST):	12:15	End Time (MST):	17:52
Analyzer make:	Thermo 42i	Analyzer serial #:	1173480006

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999932	≥ 0.995
800.3	801.0	0.9991	Slope	1.001885	$0.90 - 1.10$
400.1	395.2	1.0125	Intercept	-2.733647	± 20
200.1	195.9	1.0213			



NO_x Calibration Plot

Date: November 20, 2025

Location: Kirby South





WOOD BUFFALO ENVIRONMENTAL ASSOCIATION

CONTINUOUS AMBIENT AIR QUALITY
MONITORING PROGRAM
CALIBRATION REPORT

AMS511 BLACKGOLD NOVEMBER 2025

Operations, Data Collection,
QA/QC, Data Validation and Reporting by:
Wood Buffalo Environmental Association
Fort McMurray, Alberta

December 22, 2025



Wood Buffalo Environmental Association

SO₂ Calibration Report

Station Information

Station Name: Blackgold Station number: AMS 511
Calibration Date: November 20, 2025 Last Cal Date: October 17, 2025
Start time (MST): 9:29 End time (MST): 13:08
Reason: Routine

Calibration Standards

Cal Gas Concentration: 49.37 ppm Cal Gas Exp Date: January 5, 2029
Cal Gas Cylinder #: CC303094
Removed Cal Gas Conc: 49.37 ppm Rem Gas Exp Date:
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Model: Teledyne API T750 Serial Number: 953
Zero Air Gen Model: Teledyne API 701 Serial Number: 2659

Analyzer Information

Analyzer make: Thermo 43i Serial Number: 1173410001
Analyzer Range: 0 - 1000 ppb

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006702	1.003515	Backgd or Offset:	14.7	15.2
Calibration intercept:	-0.861427	-1.320664	Coeff or Slope:	1.163	1.163

SO₂ As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90-1.10</i>
As found zero	5000	0.0	0.0	0.6	----
As found High point	4919	81.0	799.8	800.7	1.000
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	800.1	Previous response	804.3	*% change	-0.5%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

SO₂ Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Calibrator zero	5000	0.0	0.0	0.1	----
High point	4919	81.0	799.8	802.0	0.997
Mid point	4959	40.5	399.9	399.3	1.002
Low point	4980	20.3	200.4	198.4	1.010
As left zero	5000	0.0	0.0	0.2	----
As left span	4919	81.0	799.8	800.9	0.999
Average Correction Factor:					1.003

Notes: Sample inlet filter changed after as founds. Adjusted zero only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association


SO₂ Calibration Summary

Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 17, 2025
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	9:29	End Time (MST):	13:08
Analyzer make:	Thermo 43i	Analyzer serial #:	1173410001

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.1	----	Correlation Coefficient	0.999986	≥0.995
799.8	802.0	0.9972	Slope	1.003515	0.90 - 1.10
399.9	399.3	1.0016	Intercept	-1.320664	+/-30
200.4	198.4	1.0102			



SO2 Calibration Plot

Date: November 20, 2025

Location: Blackgold





Wood Buffalo Environmental Association

H₂S Calibration Report

Station Information

Station Name: Blackgold Station number: AMS 511
Calibration Date: November 27, 2025 Last Cal Date: October 30, 2025
Start time (MST): 11:55 End time (MST): 16:26
Reason: Routine

Calibration Standards

Cal Gas Concentration: 5.139 ppm Cal Gas Exp Date: January 3, 2026
Cal Gas Cylinder #: CC511397
Removed Cal Gas Conc: 5.139 ppm Rem Gas Exp Date: N/A
Removed Gas Cyl #: N/A Diff between cyl:
Calibrator Make/Model: API T700 Serial Number: 953
ZAG Make/Model: API T701 Serial Number: 2659

Analyzer Information

Analyzer make: Thermo 43i-TLE Analyzer serial #: 1336160090
Converter make: Global G150 Converter serial #: 2025-299
Analyzer Range: 0 - 100 ppb Converter Temp: 325 degC

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>	
Calibration slope:	0.987624	1.005349	Backgd or Offset:	4.01	4.10
Calibration intercept:	0.061038	0.140712	Coeff or Slope:	1.367	1.415

H₂S As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.0	0.2	----
As found High point	4922	77.8	80.0	78.2	1.025
As found Mid point	4961	38.9	40.0	39.1	1.028
As found Low point	4981	19.5	20.0	19.5	1.038
New cylinder response					
Baseline Corr As found:	78.0	Prev response:	79.04	*% change:	-1.3%
Baseline Corr 2nd AF pt:	38.9	AF Slope:	0.976331	AF Intercept:	0.081253
Baseline Corr 3rd AF pt:	19.3	AF Correlation:	0.999989	* = > +/-5% change initiates investigation	

H₂S Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)
Calibrator zero	5000	0.0	0.0	0.4	----
High point	4922	77.8	80.0	80.7	0.991
Mid point	4961	38.9	40.0	40.1	0.997
Low point	4981	19.5	20.0	20.1	0.997
As left zero	5000	0.0	0.0	0.4	----
As left span	4922	77.8	80.0	79.6	1.005
SO2 Scrubber Check	4919	81.0	810.0	-0.1	----
Date of last scrubber change:				Ave Corr Factor	0.995
Date of last converter efficiency test:					

Notes: Changed sample inlet filter after as founds. Ran scrubber check after calibrator zero and it passed.
Adjusted span only.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

H₂S Calibration Summary


Station Information

Calibration Date:	November 27, 2025	Previous Calibration:	October 30, 2025
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	11:55	End Time (MST):	16:26
Analyzer make:	Thermo 43i-TLE	Analyzer serial #:	1336160090

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.4	----	Correlation Coefficient	0.999947	≥ 0.995
80.0	80.7	0.9909	Slope	1.005349	$0.90 - 1.10$
40.0	40.1	0.9971	Intercept	0.140712	± 3
20.0	20.1	0.9970			


H₂S Calibration Curve



H₂S Calibration Plot

Date: November 27, 2025

Location: Blackgold





Wood Buffalo Environmental Association

THC Calibration Report

Station Information

Station Name: Blackgold Station number: AMS 511
Calibration Date: November 20, 2025 Last Cal Date: October 17, 2025
Start time (MST): 9:29 End time (MST): 13:08
Reason: Routine

Calibration Standards

Gas Cert Reference: CC303094 Cal Gas Expiry Date: January 5, 2029
CH4 Cal Gas Conc. 499.2 ppm CH4 Equiv Conc. 1056.6 ppm
C3H8 Cal Gas Conc. 202.7 ppm
Removed Gas Cert: N/A Removed Gas Expiry: N/A
Removed CH4 Conc. 499.2 ppm CH4 Equiv Conc. 1056.6 ppm
Removed C3H8 Conc. 202.7 ppm Diff between cyl:
Calibrator Make/Model: Teledyne API T750 Serial Number: 953
ZAG Make/Model: Teledyne API 701 Serial Number: 2659

Analyzer Information

Analyzer make: Thermo 51i Analyzer serial #: 12426335705
Analyzer Range: 0 - 20 ppm

	<u>Start</u>	<u>Finish</u>	<u>Start</u>	<u>Finish</u>
Calibration slope:	1.006133	1.000796	Background:	0.96
Calibration intercept:	-0.074314	-0.030085	Coefficient:	2.390

THC As Found Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Baseline Adjusted Correction factor (Cc/(Ic-AFzero))
As found zero	5000	0.0	0.00	-0.15	----
As found High point	4919	81.0	17.12	17.77	0.955
As found Mid point					
As found Low point					
New cylinder response					
Baseline Corr As found:	17.92	Previous response	17.15	*% change	4.3%
Baseline Corr 2nd AF pt:	NA	AF Slope:		AF Intercept:	
Baseline Corr 3rd AF pt:	NA	AF Correlation:		* = > +/-5% change initiates investigation	

THC Calibration Data

Set Point	Dilution air flow rate (sccm)	Source gas flow rate (sccm)	Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic) Limit = 0.95-1.05
Calibrator zero	5000	0.0	0.00	-0.01	----
High point	4919	81.0	17.12	17.11	1.000
Mid point	4959	40.5	8.56	8.53	1.004
Low point	4980	20.3	4.29	4.24	1.012
As left zero	5000	0.0	0.00	-0.01	----
As left span	4919	81.0	17.12	17.24	0.993
Average Correction Factor					1.005

Notes: Sample inlet filter changed after as founds. Adjusted zero and span.

Calibration Performed By: Mohammed Kashif



Wood Buffalo Environmental Association

THC Calibration Summary


Station Information

Calibration Date:	November 20, 2025	Previous Calibration:	October 17, 2025
Station Name:	2025-09-29	Station Number:	N/A
Start Time (MST):	9:29	End Time (MST):	13:08
Analyzer make:	Thermo 51i	Analyzer serial #:	12426335705

Calibration Data

Calculated Concentration (ppm) (Cc)	Indicated Concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.00	-0.01	----	Correlation Coefficient	0.999992	≥ 0.995
17.12	17.11	1.0004	Slope	1.000796	$0.90 - 1.10$
8.56	8.53	1.0036	Intercept	-0.030085	± 1.5
4.29	4.24	1.0124			


THC Calibration Curve



THC Calibration Plot

Date: November 20, 2025

Location: Blackgold





Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Station Information

Station Name: Blackgold
Station number: AMS 511
Calibration Date: November 19, 2025
Last Cal Date: October 7, 2025
Start time (MST): 9:26
End time (MST): 15:52
Reason: Routine

Calibration Standards

NO Gas Cylinder #: T0F8P52
NOX Cal Gas Conc: 47.43 ppm
Removed Cylinder #: N/A
Removed Gas NOX Conc: 47.43 ppm
NOX gas Diff:
Calibrator Model: Teledyne API T700
ZAG make/model: Teledyne API T701
Cal Gas Expiry Date: August 16, 2026
NO Cal Gas Conc: 47.43 ppm
Removed Gas Exp Date: N/A
Removed Gas NO Conc: 47.43 ppm
NO gas Diff:
Serial Number: 2659
Serial Number: 953

As Found Dilution Calibration Data

Set Point	Dilution flow rate (sccm)	Source gas flow rate (sccm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO _x Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Baseline Adjusted NO Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>
As found zero	5000	0.0	0.0	0.0	0.0	-0.2	-0.1	-0.1	----	----
AF High point	4916	84.3	799.6	799.6	0.0	794.1	790.6	3.5	1.0067	1.0113
AF Mid point										
AF Low point										
New cyl resp										
Previous Response	NO _x = 798.7 ppb	NO = 799.0 ppb				* = > +/-5% change initiates investigation		*Percent Change	NO _x = -0.6%	
Baseline Corr 1st pt	NO _x = 794.3 ppb	NO = 790.7 ppb				<u>As Found Statistics</u>		*Percent Change	NO = -1.0%	
Baseline Corr 2nd pt	NO _x = NA ppb	NO = NA ppb				As found	NO _x r ² :	Nx SI:	Nx Int:	
Baseline Corr 3rd pt	NO _x = NA ppb	NO = NA ppb				As found	NO r ² :	NO SI:	NO Int:	
						As found	NO ₂ r ² :	NO ₂ SI:	NO ₂ Int:	

As Found GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	Baseline Adjusted NO ₂ Correction factor (Cc/(Ic-AFzero)) <i>Limit = 0.90 - 1.10</i>	Converter Efficiency <i>Limit = 96-104%</i>
As Found GPT zero						
As found high GPT point						
As found mid GPT point						
As found low GPT point						



Wood Buffalo Environmental Association

NO_x \ NO \ NO₂ Calibration Report

Analyzer Information

Analyzer Make: Teledyne API T200
NOX Range (ppb): 0 - 1000 ppb

Serial Number: 7029

Instrument Settings

	<u>Start</u>	<u>Finish</u>		<u>Start</u>	<u>Finish</u>
NO coeff or slope:	1.082	1.087	NO bkgnd or offset:	0.2	0.2
NOX coeff or slope:	1.078	1.089	NOX bkgnd or offset:	0.4	0.4
NO2 coeff or slope:	1.000	1.000	Reaction cell Press:	4.7	4.8

Calibration Statistics

	<u>Start</u>	<u>Finish</u>
NO _x Cal Slope:	1.002533	1.003235
NO _x Cal Offset:	-2.958117	-2.378695
NO Cal Slope:	1.003662	1.003049
NO Cal Offset:	-3.578225	-2.438750
NO ₂ Cal Slope:	0.993905	1.001718
NO ₂ Cal Offset:	-0.629244	-0.187507

Dilution Calibration Data

Set Point	Dilution flow rate (scm)	Source gas flow rate (scm)	Calculated NO _x concentration (ppb) (Cc)	Calculated NO concentration (ppb) (Cc)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO _x concentration (ppb) (Ic)	Indicated NO concentration (ppb) (Ic)	Indicated NO ₂ concentration (ppb) (Ic)	NO _x Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	NO Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>
Cal zero	5000	0.0	0.0	0.0	0.0	-0.1	0.0	-0.2	----	----
High point	4916	84.3	799.6	799.6	0.0	801.1	800.8	0.8	0.9982	0.9985
Mid point	4958	42.2	400.3	400.3	0.0	397.7	398.0	-0.3	1.0066	1.0058
Low point	4979	21.1	200.2	200.2	0.0	196.4	195.8	0.6	1.0191	1.0222
As left zero	5000	0.0	0.0	0.0	0.0	-0.1	-0.1	0.0	----	----
As left span	4916	84.3	799.6	383.2	416.4	791.5	383.2	408.4	1.0103	1.0000
Average Correction Factor									1.0079	1.0089

GPT Calibration Data

O3 Setpoint (ppb)	Indicated NO Reference concentration (ppb)	Indicated NO Drop concentration (ppb)	Calculated NO ₂ concentration (ppb) (Cc)	Indicated NO ₂ concentration (ppb) (Ic)	NO ₂ Correction factor (Cc/Ic) <i>Limit = 0.95-1.05</i>	Converter Efficiency <i>Limit = 96-104%</i>
Cal zero	----	----	0.0	-0.2	----	----
High GPT point	794.1	386.4	407.7	408.4	0.9983	100.2%
Mid GPT point	794.1	584.3	209.8	209.4	1.0019	99.8%
Low GPT point	794.1	683.9	110.2	110.6	0.9964	100.4%
Average Correction Factor					0.9989	100.1%

Notes:

Sample inlet filter changed after as founds. Adjusted span only.

Calibration Performed By:

Mohammed Kashif



Wood Buffalo Environmental Association


NO_x Calibration Summary

Station Information

Calibration Date:	November 19, 2025	Previous Calibration:	October 7, 2025
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	9:26	End Time (MST):	15:52
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.1	----	Correlation Coefficient	0.999962	≥0.995
799.6	801.1	0.9982	Slope	1.003235	0.90 - 1.10
400.3	397.7	1.0066	Intercept	-2.378695	+/-20
200.2	196.4	1.0191			





Wood Buffalo Environmental Association


NO₂ Calibration Summary

Station Information

Calibration Date:	November 19, 2025	Previous Calibration:	October 7, 2025
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	9:26	End Time (MST):	15:52
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	-0.2	----	Correlation Coefficient	0.999994	≥0.995
407.7	408.4	0.9983	Slope	1.001718	0.90 - 1.10
209.8	209.4	1.0019	Intercept	-0.187507	+/-20
110.2	110.6	0.9964			





Wood Buffalo Environmental Association


NO Calibration Summary

Station Information

Calibration Date:	November 19, 2025	Previous Calibration:	October 7, 2025
Station Name:	Blackgold	Station Number:	AMS 511
Start Time (MST):	9:26	End Time (MST):	15:52
Analyzer make:	Teledyne API T200	Analyzer serial #:	7029

Calibration Data


Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation		<u>Limits</u>
0.0	0.0	----	Correlation Coefficient	0.999958	≥ 0.995
799.6	800.8	0.9985	Slope	1.003049	$0.90 - 1.10$
400.3	398.0	1.0058	Intercept	-2.438750	± 20
200.2	195.8	1.0222			



NO_x Calibration Plot

Date: November 19, 2025

Location: Blackgold





End of Report